

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: NOVAK,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211943  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 04/01/2022 02:56:46

---

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:45	
Control Test		0.081	02:45	
Air Blank		0.000	02:46	
Subject Sample	# 1	0.133	02:49	Volume Not Met
Breath Volume		0.324	02:49	
Air Blank		0.000	02:50	
Air Blank		0.000	02:51	
Subject Sample	# 2	0.161	02:55	Volume Not Met
Breath Volume		0.429	02:55	
Air Blank		0.000	02:55	
Control Test		0.079	02:56	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:56	

---

Data Download Date/Time: 04/11/2022 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: NOVAK,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211943  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: P630815865860

DATE/TIME OF TEST: 04/01/2022 03:07:53

---

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.079	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.165	03:01
Breath Volume	2.054	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.162	03:06
Breath Volume	2.074	03:06
Air Blank	0.000	03:06
Control Test	0.077	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

---

Data Download Date/Time: 04/11/2022 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MERRILL,JACOB  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211943  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: BOURIS

DATE/TIME OF TEST: 04/10/2022 01:39:45

---

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.081	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.129	01:35
Breath Volume		1.621	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.121	01:38
Breath Volume		1.730	01:38
Air Blank		0.000	01:38
Control Test		0.080	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

---

Data Download Date/Time: 04/11/2022 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: NOVAK,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211943  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LAUGHLIN

DATE/TIME OF TEST: 04/20/2022 01:13:15

---

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.078	01:07	
Air Blank	0.000	01:08	
Subject Sample # 1	0.000	01:08	
Breath Volume	1.230	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample # 2		01:11	Interferent Detect
Breath Volume	1.945	01:11	
Air Blank		01:12	Purge Fail
Air Blank	0.000	01:13	

---

Data Download Date/Time: 05/09/2022 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: NOVAK,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211943  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LAUGHLIN

DATE/TIME OF TEST: 04/20/2022 01:22:02

---

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.082	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1 0.000	01:16	
Breath Volume	1.570	01:16	
Air Blank	0.000	01:17	
Air Blank	0.000	01:19	
Subject Sample	# 2 0.000	01:20	
Breath Volume	1.453	01:20	
Air Blank	0.000	01:20	
Control Test	0.082	01:20	
Air Blank	0.000	01:21	
Diagnostics Check	Fail	01:21	Analytical Stability Test
Air Blank	0.000	01:22	

---

Data Download Date/Time: 05/09/2022 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 04/12/2022 01:18:15

---

Diagnostics Check		OK	01:06	
Air Blank		0.000	01:06	
Control Test		0.080	01:07	
Air Blank		0.000	01:07	
Subject Sample	# 1	0.146	01:10	Volume Not Met
Breath Volume		0.105	01:10	
Air Blank		0.000	01:11	
Air Blank		0.000	01:13	
Subject Sample	# 2	0.107	01:16	Volume Not Met
Breath Volume		0.105	01:16	
Air Blank		0.000	01:17	
Control Test		0.080	01:17	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

---

Data Download Date/Time: 05/10/2022 10:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JABLONSKI,MARK,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: JABLONSKI,MARK  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 12:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: SOTO DIAZ

DATE/TIME OF TEST: 04/16/2022 05:22:17

---

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.087	05:17
Breath Volume	1.164	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.083	05:20
Breath Volume	1.128	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

---

Data Download Date/Time: 05/10/2022 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: COLUNGA

DATE/TIME OF TEST: 04/16/2022 22:52:59

---

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.081	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.129	22:48
Breath Volume	2.148	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.120	22:51
Breath Volume	2.714	22:51
Air Blank	0.000	22:52
Control Test	0.078	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

---

Data Download Date/Time: 05/10/2022 10:47



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: ROJAS RODRIGUEZ

DATE/TIME OF TEST: 04/17/2022 02:32:11

---

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.160	02:26
Breath Volume	2.386	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.164	02:30
Breath Volume	3.003	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

---

Data Download Date/Time: 05/10/2022 10:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JIMENEZ, WILLIAM,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: JIMENEZ, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:48

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 04/23/2022 12:09:52

---

Diagnostics Check	OK	12:09	
Air Blank		12:09	RFI Detect
Air Blank	0.000	12:09	

---

Data Download Date/Time: 05/10/2022 10:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JIMENEZ, WILLIAM,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: JIMENEZ, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:48

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 04/23/2022 12:18:43

---

Diagnostics Check	OK	12:11
Air Blank	0.000	12:12
Control Test	0.081	12:12
Air Blank	0.000	12:13
Subject Sample # 1	0.185	12:13
Breath Volume	1.582	12:13
Air Blank	0.000	12:14
Air Blank	0.000	12:15
Subject Sample # 2	0.184	12:16
Breath Volume	1.480	12:16
Air Blank	0.000	12:17
Control Test	0.080	12:18
Air Blank	0.000	12:18
Diagnostics Check	OK	12:18

---

Data Download Date/Time: 05/10/2022 10:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIDSON,CHRISTOPHER,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: DAVIDSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: BUI  
ARREST TIME: 17:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: KILGO

DATE/TIME OF TEST: 04/23/2022 19:54:15

---

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:43	
Control Test	0.081	19:43	
Air Blank	0.000	19:44	
Subject Sample	# 1	19:47	No Sample Provided
Breath Volume	0.000	19:47	
Air Blank	0.000	19:47	
Air Blank	0.000	19:49	
Subject Sample	# 2	19:52	No Sample Provided
Breath Volume	0.000	19:52	
Air Blank	0.000	19:53	
Control Test	0.081	19:53	
Air Blank	0.000	19:54	
Diagnostics Check	OK	19:54	

---

Data Download Date/Time: 05/10/2022 10:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JIMENEZ,WILLIAM,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: JIMENEZ,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 04/24/2022 08:19:40

---

Diagnostics Check	OK	08:16	
Air Blank	0.000	08:16	
Control Test	0.081	08:17	
Air Blank	0.000	08:17	
Subject Sample	# 1	08:18	Subject Test Refused
Breath Volume	0.000	08:18	
Air Blank	0.000	08:18	
Control Test	0.081	08:19	
Air Blank	0.000	08:19	
Diagnostics Check	OK	08:19	

---

Data Download Date/Time: 05/10/2022 10:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: DEFIBAUGH

DATE/TIME OF TEST: 04/02/2022 03:39:21

---

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.163	03:34
Breath Volume	2.863	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample # 2	0.164	03:37
Breath Volume	3.394	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

---

Data Download Date/Time: 04/19/2022 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: BOUTIN

DATE/TIME OF TEST: 04/02/2022 04:03:26

---

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.127	03:58
Breath Volume	1.816	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.115	04:01
Breath Volume	2.292	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

---

Data Download Date/Time: 04/19/2022 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CLARK,I  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: AQUINO MIESES

DATE/TIME OF TEST: 04/02/2022 05:49:48

---

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.080	05:43
Air Blank	0.000	05:43
Subject Sample # 1	0.072	05:45
Breath Volume	2.687	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.074	05:48
Breath Volume	2.367	05:48
Air Blank	0.000	05:48
Control Test	0.080	05:49
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

---

Data Download Date/Time: 04/19/2022 12:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREENLEE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 04/02/2022 18:24:43

---

Diagnostics Check	OK	18:17
Air Blank	0.000	18:17
Control Test	0.080	18:18
Air Blank	0.000	18:18
Subject Sample # 1	0.199	18:20
Breath Volume	1.992	18:20
Air Blank	0.000	18:20
Air Blank	0.000	18:22
Subject Sample # 2	0.192	18:23
Breath Volume	1.433	18:23
Air Blank	0.000	18:23
Control Test	0.078	18:24
Air Blank	0.000	18:24
Diagnostics Check	OK	18:24

---

Data Download Date/Time: 04/19/2022 12:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HERNANDEZ,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: DUNNE

DATE/TIME OF TEST: 04/03/2022 03:06:10

---

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.165	03:01
Breath Volume	1.343	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.173	03:04
Breath Volume	1.289	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time: 04/19/2022 12:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CARDENAS OVIEDO

DATE/TIME OF TEST: 04/03/2022 03:22:36

---

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.079	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.266	03:17
Breath Volume		1.648	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.267	03:20
Breath Volume		2.039	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

---

Data Download Date/Time: 04/19/2022 12:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: GARCIA CASTRO

DATE/TIME OF TEST: 04/04/2022 03:34:08

---

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.081	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.186	03:29
Breath Volume		2.023	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:31
Subject Sample	# 2	0.196	03:32
Breath Volume		2.015	03:32
Air Blank		0.000	03:33
Control Test		0.080	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

---

Data Download Date/Time: 04/19/2022 12:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNYDER,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: AGUILAR SANTOS

DATE/TIME OF TEST: 04/04/2022 06:13:37

---

Diagnostics Check	OK	06:07
Air Blank	0.000	06:07
Control Test	0.081	06:07
Air Blank	0.000	06:08
Subject Sample # 1	0.068	06:08
Breath Volume	3.015	06:08
Air Blank	0.000	06:09
Air Blank	0.000	06:11
Subject Sample # 2	0.064	06:12
Breath Volume	1.953	06:12
Air Blank	0.000	06:12
Control Test	0.080	06:13
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

---

Data Download Date/Time: 04/19/2022 12:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/05/2022 22:11:01

---

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.081	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.180	22:06
Breath Volume	1.875	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.176	22:09
Breath Volume	1.429	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:11

---

Data Download Date/Time: 04/19/2022 12:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 04/06/2022 03:56:53

---

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.203	03:52
Breath Volume		2.136	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.200	03:55
Breath Volume		2.597	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

---

Data Download Date/Time: 04/19/2022 12:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: NAVA MONTOYA

DATE/TIME OF TEST: 04/08/2022 03:12:18

---

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.096	03:07
Breath Volume	4.449	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.095	03:10
Breath Volume	3.574	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

---

Data Download Date/Time: 04/19/2022 12:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: ABDUSAMED

DATE/TIME OF TEST: 04/08/2022 04:54:16

---

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.080	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.170	04:44
Breath Volume		1.468	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.200	04:49
Breath Volume		1.503	04:49
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 3	0.185	04:52
Breath Volume		1.253	04:52
Air Blank		0.000	04:53
Control Test		0.078	04:53
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

---

Data Download Date/Time: 04/19/2022 12:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: COMM VEHIC  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: CANALES CASTRO

DATE/TIME OF TEST: 04/08/2022 23:24:25

---

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.081	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.129	23:19
Breath Volume	1.386	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:21
Subject Sample # 2	0.131	23:22
Breath Volume	1.875	23:22
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

---

Data Download Date/Time: 04/19/2022 12:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 04/09/2022 02:52:13

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.127	02:47
Breath Volume	2.460	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.122	02:50
Breath Volume	1.988	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

---

Data Download Date/Time: 04/19/2022 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: VELAZQUEZ TORT

DATE/TIME OF TEST: 04/09/2022 03:06:30

---

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.081	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.115	03:01
Breath Volume	2.371	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.108	03:04
Breath Volume	1.835	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time: 04/19/2022 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNOX,STEPHANIE,R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 06:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 04/09/2022 08:00:51

---

Diagnostics Check	OK	07:53
Air Blank	0.000	07:54
Control Test	0.081	07:54
Air Blank	0.000	07:55
Subject Sample # 1	0.072	07:56
Breath Volume	1.312	07:56
Air Blank	0.000	07:56
Air Blank	0.000	07:58
Subject Sample # 2	0.069	07:59
Breath Volume	1.941	07:59
Air Blank	0.000	07:59
Control Test	0.081	08:00
Air Blank	0.000	08:00
Diagnostics Check	OK	08:00

---

Data Download Date/Time: 04/19/2022 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: EARLYWINE,G  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: SOTO TOLENTINO

DATE/TIME OF TEST: 04/09/2022 23:26:48

---

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.081	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.226	23:22
Breath Volume	2.042	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:24
Subject Sample # 2	0.225	23:25
Breath Volume	1.792	23:25
Air Blank	0.000	23:25
Control Test	0.081	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

---

Data Download Date/Time: 04/19/2022 12:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FISHER,T  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: CARDENAS-LEON

DATE/TIME OF TEST: 04/09/2022 23:58:58

---

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.081	23:48	
Air Blank	0.000	23:48	
Subject Sample	# 1 0.007	23:51	Volume Not Met
Breath Volume	0.050	23:51	
Air Blank	0.000	23:52	
Air Blank	0.000	23:54	
Subject Sample	# 2	23:57	No Sample Provided
Breath Volume	0.000	23:57	
Air Blank	0.000	23:58	
Control Test	0.079	23:58	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

---

Data Download Date/Time: 04/19/2022 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RIVERA,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: DERMODY

DATE/TIME OF TEST: 04/10/2022 04:14:17

---

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.081	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.159	04:09
Breath Volume	1.871	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.158	04:12
Breath Volume	1.753	04:12
Air Blank	0.000	04:13
Control Test	0.081	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

---

Data Download Date/Time: 04/19/2022 12:30



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILLIAMS,P  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: COMM VEHIC  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: CONSTANT

DATE/TIME OF TEST: 04/10/2022 05:15:56

---

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.081	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.116	05:11
Breath Volume		1.210	05:11
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.108	05:14
Breath Volume		1.480	05:14
Air Blank		0.000	05:14
Control Test		0.080	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

---

Data Download Date/Time: 04/19/2022 12:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: DALY

DATE/TIME OF TEST: 04/10/2022 19:36:40

---

Diagnostics Check	OK	19:29
Air Blank	0.000	19:30
Control Test	0.081	19:30
Air Blank	0.000	19:31
Subject Sample # 1	0.273	19:31
Breath Volume	1.515	19:31
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample # 2	0.278	19:34
Breath Volume	1.558	19:34
Air Blank	0.000	19:35
Control Test	0.081	19:36
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

---

Data Download Date/Time: 04/19/2022 12:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: REID,W  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: RAMOS GUZMAN

DATE/TIME OF TEST: 04/11/2022 01:49:58

---

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.081	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.168	01:45
Breath Volume		2.660	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:47
Subject Sample	# 2	0.167	01:48
Breath Volume		3.308	01:48
Air Blank		0.000	01:49
Control Test		0.082	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

---

Data Download Date/Time: 04/19/2022 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/11/2022 04:38:01

---

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:26	
Control Test		0.081	04:26	
Air Blank		0.000	04:27	
Subject Sample	# 1	0.131	04:30	Volume Not Met
Breath Volume		0.121	04:30	
Air Blank		0.000	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 2	0.159	04:33	
Breath Volume		2.417	04:33	
Air Blank		0.000	04:33	
Air Blank		0.000	04:35	
Subject Sample	# 3	0.160	04:36	
Breath Volume		1.410	04:36	
Air Blank		0.000	04:37	
Control Test		0.081	04:37	
Air Blank		0.000	04:37	
Diagnostics Check		OK	04:38	

---

Data Download Date/Time: 04/19/2022 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAAR,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: CARRICO

DATE/TIME OF TEST: 04/11/2022 21:39:09

---

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.082	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.091	21:34
Breath Volume	1.195	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.090	21:37
Breath Volume	1.300	21:37
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

---

Data Download Date/Time: 04/19/2022 12:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTER,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: PEREZ REYES

DATE/TIME OF TEST: 04/13/2022 02:32:50

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.000	02:28
Breath Volume	2.304	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.000	02:31
Breath Volume	2.859	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

---

Data Download Date/Time: 04/19/2022 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GARNER,W  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: RATLIFF

DATE/TIME OF TEST: 04/14/2022 02:28:24

---

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.171	02:23
Breath Volume	2.687	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.172	02:26
Breath Volume	2.605	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

---

Data Download Date/Time: 04/19/2022 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MIRAS

DATE/TIME OF TEST: 04/14/2022 03:56:08

---

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.080	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.170	03:47
Breath Volume	3.171	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.143	03:51
Breath Volume	1.378	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 3	0.156	03:54
Breath Volume	2.167	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time: 04/19/2022 12:31



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FARLYWINE,G  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: RAMIREZ NARANJO

DATE/TIME OF TEST: 04/14/2022 05:20:22

---

Diagnostics Check	OK	05:13
Air Blank	0.000	05:13
Control Test	0.080	05:13
Air Blank	0.000	05:14
Subject Sample # 1	0.051	05:15
Breath Volume	2.375	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.051	05:18
Breath Volume	2.464	05:18
Air Blank	0.000	05:19
Control Test	0.080	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

---

Data Download Date/Time: 04/19/2022 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DURANCEAU,R  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ORDAZ

DATE/TIME OF TEST: 04/15/2022 03:56:40

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.007	03:52
Breath Volume	1.886	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.000	03:55
Breath Volume	3.273	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time: 04/19/2022 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PEREZ-RANGEL

DATE/TIME OF TEST: 04/15/2022 23:56:57

---

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.146	23:52
Breath Volume	3.906	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.143	23:55
Breath Volume	3.144	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

---

Data Download Date/Time: 04/19/2022 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CROGG,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: RODRIGUEZ SANCHEZ

DATE/TIME OF TEST: 04/16/2022 02:42:45

---

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.126	02:37
Breath Volume	1.144	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample # 2	0.119	02:41
Breath Volume	1.355	02:41
Air Blank	0.000	02:41
Control Test	0.078	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

---

Data Download Date/Time: 04/19/2022 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAAR,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SCHOULER

DATE/TIME OF TEST: 04/16/2022 03:48:46

---

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.129	03:44
Breath Volume	3.335	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.116	03:47
Breath Volume	2.339	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

---

Data Download Date/Time: 04/19/2022 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,ANGELA,V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SWEENEY,JEFFERY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 14:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: RIVERA ANDUJAR

DATE/TIME OF TEST: 04/16/2022 17:21:39

---

Diagnostics Check	OK	17:14
Air Blank	0.000	17:15
Control Test	0.081	17:15
Air Blank	0.000	17:16
Subject Sample # 1	0.000	17:17
Breath Volume	4.332	17:17
Air Blank	0.000	17:17
Air Blank	0.000	17:19
Subject Sample # 2	0.000	17:20
Breath Volume	4.339	17:20
Air Blank	0.000	17:20
Control Test	0.081	17:21
Air Blank	0.000	17:21
Diagnostics Check	OK	17:21

---

Data Download Date/Time: 04/19/2022 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: DEESON

DATE/TIME OF TEST: 04/17/2022 01:18:23

---

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.000	01:14
Breath Volume		1.859	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.000	01:17
Breath Volume		2.257	01:17
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

---

Data Download Date/Time: 04/19/2022 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ROBERTS,J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 04/17/2022 04:42:47

---

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:34	
Control Test	0.079	04:35	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:38	No Sample Provided
Breath Volume	0.000	04:38	
Air Blank	0.000	04:39	
Air Blank	0.000	04:41	
Subject Sample	# 2	04:41	Subject Test Refused
Breath Volume	0.000	04:41	
Air Blank	0.000	04:41	
Control Test	0.079	04:42	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

---

Data Download Date/Time: 04/19/2022 12:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: LANDRY,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: CORDOVA

DATE/TIME OF TEST: 04/17/2022 21:27:10

---

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.081	21:21
Air Blank	0.000	21:21
Subject Sample # 1	0.000	21:22
Breath Volume	1.871	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.000	21:25
Breath Volume	1.867	21:25
Air Blank	0.000	21:26
Control Test	0.081	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

---

Data Download Date/Time: 04/19/2022 12:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENDY,HALEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MONTOUTE,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:56

SUBJECT'S LAST NAME: SOBERANES PALMA

DATE/TIME OF TEST: 04/18/2022 10:23:45

---

Diagnostics Check	OK	10:16
Air Blank	0.000	10:17
Control Test	0.080	10:17
Air Blank	0.000	10:18
Subject Sample # 1	0.292	10:18
Breath Volume	2.894	10:18
Air Blank	0.000	10:19
Air Blank	0.000	10:21
Subject Sample # 2	0.282	10:22
Breath Volume	2.207	10:22
Air Blank	0.000	10:22
Control Test	0.079	10:23
Air Blank	0.000	10:23
Diagnostics Check	OK	10:23

---

Data Download Date/Time: 04/19/2022 12:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HERNANDEZ,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: HO

DATE/TIME OF TEST: 04/18/2022 22:56:01

---

Diagnostics Check		OK	22:48
Air Blank		0.000	22:48
Control Test		0.080	22:48
Air Blank		0.000	22:49
Subject Sample	# 1	0.161	22:51
Breath Volume		1.769	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:53
Subject Sample	# 2	0.150	22:54
Breath Volume		1.695	22:54
Air Blank		0.000	22:55
Control Test		0.080	22:55
Air Blank		0.000	22:55
Diagnostics Check		OK	22:56

---

Data Download Date/Time: 04/19/2022 12:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARPER,DARIN,W  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: HARPER,HARPER  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: CARTIER

DATE/TIME OF TEST: 04/01/2022 06:08:56

---

Diagnostics Check		OK	06:01
Air Blank		0.000	06:02
Control Test		0.081	06:02
Air Blank		0.000	06:03
Subject Sample	# 1	0.144	06:04
Breath Volume		1.230	06:04
Air Blank		0.000	06:04
Air Blank		0.000	06:06
Subject Sample	# 2	0.153	06:07
Breath Volume		3.242	06:07
Air Blank		0.000	06:07
Control Test		0.080	06:08
Air Blank		0.000	06:08
Diagnostics Check		OK	06:08

---

Data Download Date/Time: 04/26/2022 11:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: WILKINSON,RYAN  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: SCHOLZ

DATE/TIME OF TEST: 04/03/2022 03:59:18

---

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.175	03:53
Breath Volume	1.867	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.169	03:57
Breath Volume	1.992	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

---

Data Download Date/Time: 04/26/2022 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PARKER,CARL  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 04/03/2022 22:05:58

---

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:02	
Control Test	0.081	22:03	
Air Blank	0.000	22:03	
Subject Sample	# 1	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:05	
Control Test	0.080	22:05	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

---

Data Download Date/Time: 04/26/2022 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPELA,JORDAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: CAPELA,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 04/04/2022 04:47:06

---

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.139	04:42
Breath Volume	1.644	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.139	04:45
Breath Volume	1.250	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

---

Data Download Date/Time: 04/26/2022 11:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARPER,DARIN,W  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: HARPER,DARIN  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MADRIGAL

DATE/TIME OF TEST: 04/06/2022 03:04:32

---

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.082	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.258	03:00
Breath Volume		1.988	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:02
Subject Sample	# 2	0.267	03:02
Breath Volume		2.078	03:02
Air Blank		0.000	03:03
Control Test		0.077	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

---

Data Download Date/Time: 04/26/2022 11:52



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONG,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: ROSSIC  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 09:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:11

SUBJECT'S LAST NAME: STAHLI

DATE/TIME OF TEST: 04/06/2022 11:36:28

---

Diagnostics Check	OK	11:33	
Air Blank	0.000	11:33	
Control Test	0.081	11:33	
Air Blank	0.000	11:34	
Subject Sample	# 1	11:35	Subject Test Refused
Breath Volume	0.000	11:35	
Air Blank	0.000	11:35	
Control Test	0.080	11:35	
Air Blank	0.000	11:36	
Diagnostics Check	OK	11:36	

---

Data Download Date/Time: 04/26/2022 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PARKER,CARL  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 04/07/2022 00:24:08

---

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.082	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:22	Subject Test Refused
Breath Volume	0.000	00:22	
Air Blank	0.000	00:23	
Control Test	0.081	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

---

Data Download Date/Time: 04/26/2022 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPELA,JORDAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: WILKINSON,RYAN  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/08/2022 03:24:34

---

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.081	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Control Test	0.081	03:23	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

---

Data Download Date/Time: 04/26/2022 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEEHAN,SHANE,P  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: JOHNSON,D  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 04/08/2022 22:20:13

---

Diagnostics Check	OK	22:19	
Air Blank		22:19	RFI Detect
Air Blank	0.000	22:20	

---

Data Download Date/Time: 04/26/2022 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEEHAN,SHANE,P  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: JOHNSON,D  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 04/08/2022 22:27:07

---

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:23	
Control Test	0.082	22:23	
Air Blank	0.000	22:24	
Subject Sample	# 1	22:25	Subject Test Refused
Breath Volume	0.000	22:25	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

---

Data Download Date/Time: 04/26/2022 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERGMAN,A,W  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: STAMPS,B  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 04/09/2022 02:50:04

---

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.082	02:49	
Air Blank		02:49	RFI Detect
Air Blank	0.000	02:50	

---

Data Download Date/Time: 04/26/2022 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: QUIROGA-JOYA,ANJELICA  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211543  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FENNELL

DATE/TIME OF TEST: 04/12/2022 03:23:14

---

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.081	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:21	RFI Detect
Breath Volume	0.363	03:21	
Air Blank	0.000	03:22	
Control Test	0.081	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

---

Data Download Date/Time: 04/26/2022 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: DORSEY,JAXON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211543  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/29/2022 19:50:55

---

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:47	
Control Test	0.080	19:47	
Air Blank	0.000	19:48	
Subject Sample	# 1	19:49	RFI Detect
Breath Volume	0.593	19:49	
Air Blank	0.000	19:49	
Control Test	0.080	19:50	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

---

Data Download Date/Time: 05/02/2022 14:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: DORSEY,JAXON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211543  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/29/2022 20:02:03

---

Diagnostics Check	OK	19:52	
Air Blank	0.000	19:53	
Control Test	0.079	19:53	
Air Blank	0.000	19:54	
Subject Sample # 1	0.183	19:55	
Breath Volume	1.187	19:55	
Air Blank	0.000	19:56	
Air Blank	0.000	19:58	
Subject Sample # 2		20:00	RFI Detect
Breath Volume	0.667	20:00	
Air Blank	0.000	20:01	
Control Test	0.079	20:01	
Air Blank	0.000	20:01	
Diagnostics Check	OK	20:02	

---

Data Download Date/Time: 05/02/2022 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: DORSEY,JAXON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211543  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/29/2022 20:11:47

---

Diagnostics Check	OK	20:05
Air Blank	0.000	20:05
Control Test	0.079	20:05
Air Blank	0.000	20:06
Subject Sample # 1	0.178	20:07
Breath Volume	1.675	20:07
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.182	20:10
Breath Volume	1.875	20:10
Air Blank	0.000	20:10
Control Test	0.078	20:11
Air Blank	0.000	20:11
Diagnostics Check	OK	20:11

---

Data Download Date/Time: 05/02/2022 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,FREDERICK,A  
BREATH TEST OPERATOR AGENCY: Kennedy Space Center  
ARRESTING OFFICER: SMITH,ERIC  
ARRESTING OFFICER AGENCY: Kennedy Space Center  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: MCCUNE

DATE/TIME OF TEST: 04/23/2022 22:50:57

---

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:49	
Control Test	0.081	22:49	
Air Blank		22:50	Purge Fail
Air Blank	0.000	22:50	

---

Data Download Date/Time: 05/18/2022 11:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,FREDERICK,A  
BREATH TEST OPERATOR AGENCY: Kennedy Space Center  
ARRESTING OFFICER: SMITH,ERIC  
ARRESTING OFFICER AGENCY: Kennedy Space Center  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: MCCUNE

DATE/TIME OF TEST: 04/23/2022 22:56:49

---

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:54	
Control Test	0.080	22:54	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:55	
Control Test	0.080	22:56	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

---

Data Download Date/Time: 05/18/2022 11:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: CASAS,JAVIER  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: CARDENAS SILVA

DATE/TIME OF TEST: 04/01/2022 04:28:04

---

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.079	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:27	
Control Test	0.079	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:28	

---

Data Download Date/Time: 04/29/2022 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DE LA RUA,BRYAN,A  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: WILLIAMS,DAVID  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 04/06/2022 03:04:56

---

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.079	03:02	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.080	03:04	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COON,REBECCA,L  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: CASAS,JAVIER  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 04/09/2022 01:46:49

---

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Subject Sample # 1	0.178	01:33	
Breath Volume	1.203	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample # 2	0.139	01:39	Volume Not Met
Breath Volume	0.570	01:39	
Air Blank	0.000	01:40	
Air Blank	0.000	01:41	
Subject Sample # 3	0.159	01:44	Volume Not Met
Breath Volume	0.140	01:44	
Air Blank	0.000	01:45	
Control Test	0.078	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COON,REBECCA,L  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: CASAS,JAVIER  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 04/09/2022 02:04:58

---

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:49	
Control Test		0.077	01:49	
Air Blank		0.000	01:50	
Subject Sample	# 1	0.154	01:53	Volume Not Met
Breath Volume		0.132	01:53	
Air Blank		0.000	01:54	
Air Blank		0.000	01:55	
Subject Sample	# 2	0.161	01:58	
Breath Volume		1.621	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:00	
Subject Sample	# 3	0.160	02:03	
Breath Volume		1.136	02:03	
Air Blank		0.000	02:04	
Control Test		0.077	02:04	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

---

Data Download Date/Time: 04/29/2022 13:00



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIMONEY,RYAN,P  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: BUGALLO,GABRIEL  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: RODRIGUES

DATE/TIME OF TEST: 04/10/2022 03:52:06

---

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.079	03:50	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:50	Subject Test Refused
Breath Volume	0.000	03:50	
Air Blank	0.000	03:51	
Control Test	0.080	03:51	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

---

Data Download Date/Time: 04/29/2022 13:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DAVID,R  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: WILLIAMS,DAVID  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: GRANADOS MEJIA

DATE/TIME OF TEST: 04/11/2022 05:49:28

---

Diagnostics Check		OK	05:43
Air Blank		0.000	05:43
Control Test		0.079	05:44
Air Blank		0.000	05:44
Subject Sample	# 1	0.123	05:44
Breath Volume		1.972	05:44
Air Blank		0.000	05:45
Air Blank		0.000	05:47
Subject Sample	# 2	0.116	05:47
Breath Volume		1.156	05:47
Air Blank		0.000	05:48
Control Test		0.079	05:48
Air Blank		0.000	05:49
Diagnostics Check		OK	05:49

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER,KIANA,P  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: WILLIAMS,DAVID  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: LOWER

DATE/TIME OF TEST: 04/12/2022 00:13:56

---

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:02	
Control Test		0.079	00:02	
Air Blank		0.000	00:02	
Subject Sample	# 1	0.121	00:06	Volume Not Met
Breath Volume		0.160	00:06	
Air Blank		0.000	00:06	
Air Blank		0.000	00:08	
Subject Sample	# 2	0.131	00:08	
Breath Volume		1.640	00:08	
Air Blank		0.000	00:09	
Air Blank		0.000	00:11	
Subject Sample	# 3	0.120	00:12	
Breath Volume		1.324	00:12	
Air Blank		0.000	00:13	
Control Test		0.079	00:13	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER,KIANA,P  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: HARRISON,DANIELLE  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: KHOURIGOU

DATE/TIME OF TEST: 04/12/2022 03:34:20

---

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:22	
Control Test		0.080	03:23	
Air Blank		0.000	03:23	
Subject Sample	# 1	0.143	03:26	Volume Not Met
Breath Volume		0.628	03:26	
Air Blank		0.000	03:27	
Air Blank		0.000	03:29	
Subject Sample	# 2	0.143	03:32	Volume Not Met
Breath Volume		0.496	03:32	
Air Blank		0.000	03:33	
Control Test		0.078	03:33	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

---

Data Download Date/Time: 04/29/2022 13:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,DANIEL,X  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: WILLIAMS,DAVID  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BRACKEN

DATE/TIME OF TEST: 04/13/2022 01:58:51

---

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.079	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1	01:57	Subject Test Refused
Breath Volume	0.000	01:57	
Air Blank	0.000	01:57	
Control Test	0.080	01:58	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZIADIE,ANDREW,C  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: MILLER,BRYAN  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 10:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:38

SUBJECT'S LAST NAME: VRANCKEN

DATE/TIME OF TEST: 04/13/2022 11:32:10

---

Diagnostics Check	OK	11:29	
Air Blank	0.000	11:29	
Control Test	0.079	11:30	
Air Blank	0.000	11:30	
Subject Sample	# 1	11:30	Subject Test Refused
Breath Volume	0.000	11:30	
Air Blank	0.000	11:31	
Control Test	0.079	11:31	
Air Blank	0.000	11:32	
Diagnostics Check	OK	11:32	

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASAS,JAVIER,  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: GOLDEN,LISA  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/13/2022 23:07:59

---

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:05	
Control Test	0.079	23:05	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:06	Subject Test Refused
Breath Volume	0.000	23:06	
Air Blank	0.000	23:07	
Control Test	0.079	23:07	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: GRAHAM,AUSTIN  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PEREZ DOMINGUEZ

DATE/TIME OF TEST: 04/14/2022 01:11:45

---

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:05	
Subject Sample # 1	0.230	01:06	
Breath Volume	2.046	01:06	
Air Blank	0.000	01:07	
Air Blank	0.000	01:09	
Subject Sample # 2	0.221	01:09	
Breath Volume	1.378	01:09	
Air Blank	0.000	01:10	
Control Test	0.074	01:11	Control Outside Tolerance
Air Blank	0.000	01:11	

---

Data Download Date/Time: 04/29/2022 13:00



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,DANIEL,X  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: WILLIAMS,DAVID  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SHULMAN

DATE/TIME OF TEST: 04/17/2022 03:10:45

---

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.080	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,DANIEL,X  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: WILLIAMS,DAVID  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ABUJAMRA

DATE/TIME OF TEST: 04/18/2022 03:20:04

---

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:08	
Control Test		0.080	03:08	
Air Blank		0.000	03:09	
Subject Sample	# 1	0.130	03:12	Volume Not Met
Breath Volume		0.082	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:14	
Subject Sample	# 2	0.133	03:18	Volume Not Met
Breath Volume		0.351	03:18	
Air Blank		0.000	03:19	
Control Test		0.075	03:19	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:20	

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,DANIEL,X  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: WILLIAMS,DAVID  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ABUJAMRA

DATE/TIME OF TEST: 04/18/2022 03:23:46

---

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.073	03:23	Control Outside Tolerance
Air Blank	0.000	03:23	

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,DANIEL,X  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: WILLIAMS,DAVID  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ABUJAMRA

DATE/TIME OF TEST: 04/18/2022 03:34:51

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.076	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.154	03:29
Breath Volume	1.750	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.147	03:33
Breath Volume	1.988	03:33
Air Blank	0.000	03:33
Control Test	0.076	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNETTE,ASHLEY,  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: SAAVEDRA,ALONSO  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 10:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:24

SUBJECT'S LAST NAME: SANTANA MANZANARES

DATE/TIME OF TEST: 04/18/2022 11:04:07

---

Diagnostics Check	OK	10:55
Air Blank	0.000	10:56
Control Test	0.079	10:56
Air Blank	0.000	10:57
Subject Sample # 1	0.127	10:59
Breath Volume	1.457	10:59
Air Blank	0.000	11:00
Air Blank	0.000	11:01
Subject Sample # 2	0.139	11:02
Breath Volume	1.738	11:02
Air Blank	0.000	11:03
Control Test	0.079	11:03
Air Blank	0.000	11:04
Diagnostics Check	OK	11:04

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HORNE,ASHTON,D  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: WILLIAMS,DAVID  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 04/19/2022 22:16:56

---

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.172	22:12
Breath Volume	1.433	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:14
Subject Sample # 2	0.170	22:15
Breath Volume	1.929	22:15
Air Blank	0.000	22:16
Control Test	0.078	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

---

Data Download Date/Time: 04/29/2022 13:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIMONEY,RYAN,P  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: CASAS,JAVIER  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 04/22/2022 03:43:20

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.160	03:38
Breath Volume	2.867	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.156	03:41
Breath Volume	2.636	03:41
Air Blank	0.000	03:42
Control Test	0.076	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,DANIEL,N  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: WILLIAMS,DAVID  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SEMCHENKO

DATE/TIME OF TEST: 04/26/2022 03:53:54

---

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.079	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1	0.167	03:45	Volume Not Met
Breath Volume		0.890	03:45	
Air Blank		0.000	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 2	0.165	03:48	
Breath Volume		2.214	03:48	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 3	0.169	03:52	
Breath Volume		2.148	03:52	
Air Blank		0.000	03:52	
Control Test		0.078	03:53	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

---

Data Download Date/Time: 04/29/2022 13:00



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,DANIEL,C  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: HARRISON,DANIELLE  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: DENMARK

DATE/TIME OF TEST: 04/27/2022 03:26:53

---

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.079	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.079	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

---

Data Download Date/Time: 04/29/2022 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,HALEY,M  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: CASAS,JAVIER  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: DOLL

DATE/TIME OF TEST: 04/30/2022 00:43:13

---

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.079	00:28	
Air Blank	0.000	00:29	
Subject Sample # 1	0.292	00:31	
Breath Volume	1.625	00:31	
Air Blank	0.000	00:32	
Air Blank	0.000	00:33	
Subject Sample # 2	0.075	00:36	Volume Not Met
Breath Volume	0.292	00:36	
Air Blank	0.000	00:37	
Air Blank	0.000	00:39	
Subject Sample # 3		00:41	Subject Test Refused
Breath Volume	0.675	00:41	
Air Blank	0.000	00:42	
Control Test	0.077	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

---

Data Download Date/Time: 05/18/2022 16:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROJAS,GUILLERMO,A  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: ROJAS,GUILLERMO  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CARBALLO ALMENARES

DATE/TIME OF TEST: 04/04/2022 02:10:55

---

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.081	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.249	02:06
Breath Volume	1.960	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample # 2	0.232	02:09
Breath Volume	4.066	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

---

Data Download Date/Time: 04/27/2022 10:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERO,NATALIE,  
BREATH TEST OPERATOR AGENCY: OTHER LOCAL  
ARRESTING OFFICER: ROMAN,ANTONIO  
ARRESTING OFFICER AGENCY: OTHER LOCAL  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: IZAGUIRRE

DATE/TIME OF TEST: 04/06/2022 03:28:02

---

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.081	03:25	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Control Test	0.081	03:27	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:28	

---

Data Download Date/Time: 04/27/2022 10:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: DOMINGUEZ,JOSE  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: FEBRUARY

DATE/TIME OF TEST: 04/08/2022 22:28:25

---

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.081	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.151	22:23
Breath Volume	1.628	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.150	22:26
Breath Volume	2.203	22:26
Air Blank	0.000	22:27
Control Test	0.079	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

---

Data Download Date/Time: 04/27/2022 10:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,ARGENIS,A  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: ARGENIS,PEREZ  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RENOIR

DATE/TIME OF TEST: 04/09/2022 03:16:25

---

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.315	03:08
Breath Volume	2.156	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.343	03:11
Breath Volume	1.539	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 3	0.339	03:14
Breath Volume	2.804	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

---

Data Download Date/Time: 04/27/2022 10:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,ALEX,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: JOHNSON,ALEX  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 15:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: NUNEZ ARMIJO

DATE/TIME OF TEST: 04/10/2022 16:08:23

---

Diagnostics Check		OK	16:00
Air Blank		0.000	16:00
Control Test		0.081	16:01
Air Blank		0.000	16:01
Subject Sample	# 1	0.256	16:02
Breath Volume		1.898	16:02
Air Blank		0.000	16:03
Air Blank		0.000	16:04
Subject Sample	# 2	0.250	16:06
Breath Volume		1.691	16:06
Air Blank		0.000	16:07
Control Test		0.080	16:07
Air Blank		0.000	16:08
Diagnostics Check		OK	16:08

---

Data Download Date/Time: 04/27/2022 10:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROJAS,GUILLERMO,A  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: ROJAS,GUILLERMO  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/11/2022 03:29:27

---

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.081	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.178	03:24
Breath Volume	1.378	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:26
Subject Sample # 2	0.166	03:27
Breath Volume	2.433	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time: 04/27/2022 10:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALBO,JOSHUA,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: ESTEVEZ,JOSE  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: MATUS

DATE/TIME OF TEST: 04/11/2022 20:54:24

---

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.081	20:48
Air Blank	0.000	20:48
Subject Sample # 1	0.182	20:49
Breath Volume	1.230	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.185	20:52
Breath Volume	1.589	20:52
Air Blank	0.000	20:53
Control Test	0.081	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

---

Data Download Date/Time: 04/27/2022 10:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COMESANA,ANGELIKA,N  
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD  
ARRESTING OFFICER: LORENZO,FRANCISCO  
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: COMESANA

DATE/TIME OF TEST: 04/13/2022 01:08:43

---

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.081	01:06	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.081	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

---

Data Download Date/Time: 04/27/2022 10:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAXIMILIEN, RONALD,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: MAXIMILIEN, RONALD  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:09

SUBJECT'S LAST NAME: GARCIA ROSALES

DATE/TIME OF TEST: 04/13/2022 06:34:51

---

Diagnostics Check	OK	06:31	
Air Blank	0.000	06:32	
Control Test	0.081	06:32	
Air Blank	0.000	06:32	
Subject Sample	# 1	06:33	RFI Detect
Breath Volume	0.000	06:33	
Air Blank	0.000	06:33	
Control Test	0.081	06:34	
Air Blank	0.000	06:34	
Diagnostics Check	OK	06:34	

---

Data Download Date/Time: 04/27/2022 10:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAXIMILIEN, RONALD,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: MAXIMILIEN, RONALD  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/13/2022 06:43:07

---

Diagnostics Check	OK	06:40	
Air Blank	0.000	06:40	
Control Test	0.081	06:41	
Air Blank	0.000	06:41	
Subject Sample	# 1	06:41	Subject Test Refused
Breath Volume	0.000	06:41	
Air Blank	0.000	06:42	
Control Test	0.081	06:42	
Air Blank	0.000	06:43	
Diagnostics Check	OK	06:43	

---

Data Download Date/Time: 04/27/2022 10:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALBO, JOSHUA,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: KEMP, TREVOR  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/16/2022 22:18:45

---

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:11
Subject Sample # 1	0.337	22:13
Breath Volume	1.742	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.324	22:17
Breath Volume	1.156	22:17
Air Blank	0.000	22:17
Control Test	0.079	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

---

Data Download Date/Time: 04/27/2022 10:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FALCON,IVAN,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: FALCON,IVAN  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: GLICKEL

DATE/TIME OF TEST: 04/17/2022 01:57:01

---

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.080	01:54	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:56	
Control Test	0.080	01:56	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:57	

---

Data Download Date/Time: 04/27/2022 10:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTANA,YUDIEL,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: SANTANA,YUDIEL  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ALEXIS

DATE/TIME OF TEST: 04/17/2022 03:34:06

---

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	RFI Detect
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.082	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

---

Data Download Date/Time: 04/27/2022 10:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTANA,YUDIEL,S  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: SANTANA,YUDIEL  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ALEXIS

DATE/TIME OF TEST: 04/17/2022 03:43:39

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.167	03:38
Breath Volume	1.519	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.164	03:42
Breath Volume	2.261	03:42
Air Blank	0.000	03:42
Control Test	0.076	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time: 04/27/2022 10:41



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POST,JACOB,S  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: POST,JACOB  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 04/18/2022 22:00:13

---

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.081	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.136	21:55
Breath Volume	1.582	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:57
Subject Sample # 2	0.143	21:58
Breath Volume	2.417	21:58
Air Blank	0.000	21:59
Control Test	0.081	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

---

Data Download Date/Time: 04/27/2022 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA,HENRY,E  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: GARCIA,HENRY  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: CARABALLO

DATE/TIME OF TEST: 04/21/2022 00:40:33

---

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.081	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.175	00:34
Breath Volume		1.449	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.181	00:38
Breath Volume		1.929	00:38
Air Blank		0.000	00:39
Control Test		0.082	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

---

Data Download Date/Time: 04/27/2022 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIERRA,ALIEK,O  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: SIERRA,ALIEK  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 04/23/2022 10:56:06

---

Diagnostics Check	OK	10:53	
Air Blank	0.000	10:53	
Control Test	0.081	10:54	
Air Blank	0.000	10:54	
Subject Sample	# 1	10:54	Subject Test Refused
Breath Volume	0.000	10:54	
Air Blank	0.000	10:55	
Control Test	0.081	10:55	
Air Blank	0.000	10:56	
Diagnostics Check	OK	10:56	

---

Data Download Date/Time: 04/27/2022 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MERRILL,JACOB  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: FRANCISCO PEDRO

DATE/TIME OF TEST: 04/14/2022 00:08:52

---

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.078	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.199	00:04
Breath Volume	2.539	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.195	00:07
Breath Volume	1.894	00:07
Air Blank	0.000	00:07
Control Test	0.077	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

---

Data Download Date/Time: 05/09/2022 10:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: REYES,MICHAEL  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: WOODWORTH

DATE/TIME OF TEST: 04/02/2022 03:57:34

---

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.103	03:52
Breath Volume	3.363	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.102	03:55
Breath Volume	3.214	03:55
Air Blank	0.000	03:56
Control Test	0.077	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

---

Data Download Date/Time: 04/11/2022 14:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BARRETT,CODY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 04/02/2022 19:34:06

---

Diagnostics Check		OK	19:26
Air Blank		0.000	19:26
Control Test		0.081	19:27
Air Blank		0.000	19:27
Subject Sample	# 1	0.243	19:28
Breath Volume		1.296	19:28
Air Blank		0.000	19:29
Air Blank		0.000	19:31
Subject Sample	# 2	0.229	19:32
Breath Volume		1.109	19:32
Air Blank		0.000	19:33
Control Test		0.077	19:33
Air Blank		0.000	19:34
Diagnostics Check		OK	19:34

---

Data Download Date/Time: 04/11/2022 14:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BARRETT,CODY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 04/03/2022 00:47:59

---

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.081	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.170	00:42
Breath Volume	1.136	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:44
Subject Sample # 2	0.187	00:46
Breath Volume	1.773	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

---

Data Download Date/Time: 04/11/2022 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: COOK,AUSTIN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: DANG

DATE/TIME OF TEST: 04/03/2022 03:13:22

---

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.157	03:08
Breath Volume	2.878	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.154	03:11
Breath Volume	2.113	03:11
Air Blank	0.000	03:12
Control Test	0.077	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

---

Data Download Date/Time: 04/11/2022 14:34



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: COTTOM,CALEB  
ARRESTING OFFICER AGENCY: Palmetto PD  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 04/08/2022 22:18:59

---

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.082	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.176	22:14
Breath Volume	1.968	22:14
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.167	22:17
Breath Volume	1.941	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

---

Data Download Date/Time: 04/11/2022 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERRILL,JACOB,M  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MERRILL,JACOB  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: MOMENTHY

DATE/TIME OF TEST: 04/10/2022 05:43:26

---

Diagnostics Check	OK	05:33
Air Blank	0.000	05:34
Control Test	0.082	05:34
Air Blank	0.000	05:35
Subject Sample # 1	0.183	05:38
Breath Volume	2.253	05:38
Air Blank	0.000	05:39
Air Blank	0.000	05:40
Subject Sample # 2	0.185	05:41
Breath Volume	2.019	05:41
Air Blank	0.000	05:42
Control Test	0.079	05:42
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

---

Data Download Date/Time: 04/11/2022 14:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402231696  
DRY GAS EXPIRY DATE: 10/07/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: MENDOZA GUIFARRO

DATE/TIME OF TEST: 04/03/2022 01:29:14

---

Diagnostics Check	OK	01:28	
Air Blank		01:28	RFI Detect
Air Blank	0.000	01:29	

---

Data Download Date/Time: 04/21/2022 21:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402231696  
DRY GAS EXPIRY DATE: 10/07/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: MENDOZA GUIFARRO

DATE/TIME OF TEST: 04/03/2022 01:37:42

---

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.176	01:33
Breath Volume	2.914	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.174	01:36
Breath Volume	2.695	01:36
Air Blank	0.000	01:36
Control Test	0.080	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

---

Data Download Date/Time: 04/21/2022 21:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402231696  
DRY GAS EXPIRY DATE: 10/07/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CREESE

DATE/TIME OF TEST: 04/08/2022 04:16:17

---

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.098	04:11
Breath Volume	2.734	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.096	04:14
Breath Volume	2.019	04:14
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time: 04/21/2022 21:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: GARARD,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402231696  
DRY GAS EXPIRY DATE: 10/07/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BOYKIN

DATE/TIME OF TEST: 04/16/2022 03:11:09

---

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.081	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.170	03:06
Breath Volume		2.066	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.169	03:09
Breath Volume		2.832	03:09
Air Blank		0.000	03:10
Control Test		0.081	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

---

Data Download Date/Time: 04/21/2022 21:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: GARARD,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402231696  
DRY GAS EXPIRY DATE: 10/07/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 04/17/2022 05:21:40

---

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.080	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.244	05:16
Breath Volume		1.800	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.243	05:19
Breath Volume		1.746	05:19
Air Blank		0.000	05:20
Control Test		0.080	05:21
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

---

Data Download Date/Time: 04/21/2022 21:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HURT,ALEXANDER,C  
BREATH TEST OPERATOR AGENCY: Holmes Beach PD  
ARRESTING OFFICER: SIRIGNANO,JESSICA  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 04/12/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175970  
DRY GAS EXPIRY DATE: 08/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/18/2022 01:30:00

---

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.077	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.086	01:25
Breath Volume	1.218	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.082	01:28
Breath Volume	1.285	01:28
Air Blank	0.000	01:29
Control Test	0.076	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:30

---

Data Download Date/Time: 05/15/2022 07:38



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANHORN, MICHAEL, T  
BREATH TEST OPERATOR AGENCY: Holmes Beach PD  
ARRESTING OFFICER: VANHORN, MICHAEL  
ARRESTING OFFICER AGENCY: Holmes Beach PD  
LAST AGENCY INSPECTION: 04/12/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175970  
DRY GAS EXPIRY DATE: 08/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: RUIZ DEL VIZO

DATE/TIME OF TEST: 04/27/2022 16:02:09

---

Diagnostics Check	OK	15:54
Air Blank	0.000	15:55
Control Test	0.077	15:55
Air Blank	0.000	15:56
Subject Sample # 1	0.376	15:57
Breath Volume	2.585	15:57
Air Blank	0.000	15:58
Air Blank	0.000	16:00
Subject Sample # 2	0.372	16:00
Breath Volume	2.527	16:00
Air Blank	0.000	16:01
Control Test	0.076	16:01
Air Blank	0.000	16:02
Diagnostics Check	OK	16:02

---

Data Download Date/Time: 05/15/2022 07:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 3024022365  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: DOW

DATE/TIME OF TEST: 04/01/2022 22:40:12

---

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.135	22:35
Breath Volume	1.566	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.143	22:38
Breath Volume	2.078	22:38
Air Blank	0.000	22:39
Control Test	0.081	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

---

Data Download Date/Time: 04/11/2022 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER,RICHARD,F  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SNYDER,RICHARD  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 3024022365  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 04/02/2022 22:30:28

---

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.109	22:25
Breath Volume	2.609	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.108	22:28
Breath Volume	2.523	22:28
Air Blank	0.000	22:29
Control Test	0.079	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

---

Data Download Date/Time: 04/11/2022 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SIRIGNANO,JESSICA  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 3024022365  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: LOEBELL

DATE/TIME OF TEST: 04/17/2022 20:36:19

---

Diagnostics Check	OK	20:29
Air Blank	0.000	20:30
Control Test	0.080	20:30
Air Blank	0.000	20:31
Subject Sample # 1	0.133	20:31
Breath Volume	2.347	20:31
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.126	20:34
Breath Volume	1.644	20:34
Air Blank	0.000	20:35
Control Test	0.078	20:35
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

---

Data Download Date/Time: 05/10/2022 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 3024022365  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ROBLERO-SARGENTO

DATE/TIME OF TEST: 04/18/2022 00:47:30

---

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.253	00:42
Breath Volume		1.960	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.252	00:45
Breath Volume		2.453	00:45
Air Blank		0.000	00:46
Control Test		0.077	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

---

Data Download Date/Time: 05/10/2022 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANGIERI,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: LANGIERI,JOSHUA  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 3024022365  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: SHURMUR

DATE/TIME OF TEST: 04/27/2022 04:15:20

---

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.148	04:10
Breath Volume	1.500	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.155	04:13
Breath Volume	1.558	04:13
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

---

Data Download Date/Time: 05/10/2022 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBEL,NEIL,F  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SMITH,CALE  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1327124  
DRY GAS EXPIRY DATE: 05/14/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/13/2022 21:39:16

---

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.078	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.040	21:34
Breath Volume	2.000	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.039	21:37
Breath Volume	2.195	21:37
Air Blank	0.000	21:38
Control Test	0.078	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

---

Data Download Date/Time: 05/13/2022 13:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LOGAN,ZACHARY  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ALMENDAREZ VELA

DATE/TIME OF TEST: 04/01/2022 23:48:58

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.078	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.000	23:44
Breath Volume	2.941	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.000	23:47
Breath Volume	3.398	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

---

Data Download Date/Time: 04/11/2022 10:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GOMEZ GARCIA

DATE/TIME OF TEST: 04/10/2022 02:47:10

---

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.169	02:42
Breath Volume	2.675	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.173	02:45
Breath Volume	2.683	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

---

Data Download Date/Time: 04/11/2022 10:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: MCGOWAN

DATE/TIME OF TEST: 04/16/2022 00:53:01

---

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:40	
Control Test		0.079	00:40	
Air Blank		0.000	00:40	
Subject Sample	# 1	0.188	00:44	Slope Not Level
Breath Volume		1.480	00:44	
Air Blank		0.000	00:44	
Air Blank		0.000	00:46	
Subject Sample	# 2	0.192	00:48	
Breath Volume		1.515	00:48	
Air Blank		0.000	00:48	
Air Blank		0.000	00:50	
Subject Sample	# 3		00:51	Subject Test Refused
Breath Volume		0.000	00:51	
Air Blank		0.000	00:52	
Control Test		0.077	00:52	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:53	

---

Data Download Date/Time: 05/10/2022 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: DIAZ ANTUNEZ

DATE/TIME OF TEST: 04/30/2022 03:32:18

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.226	03:27
Breath Volume	1.148	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.236	03:30
Breath Volume	1.980	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

---

Data Download Date/Time: 05/10/2022 10:56



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

INSTRUMENT REGISTERED TO: FHP TROOP L  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRIQUEZ ZUNIGA,MARCIO,E  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 03/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DESKIN

DATE/TIME OF TEST: 04/02/2022 22:32:17

---

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.143	22:27
Breath Volume	2.718	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.155	22:30
Breath Volume	2.714	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 04/28/2022 12:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: ROCHEFORD

DATE/TIME OF TEST: 04/01/2022 21:29:57

---

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.077	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.094	21:25
Breath Volume	1.441	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.104	21:28
Breath Volume	2.261	21:28
Air Blank	0.000	21:29
Control Test	0.077	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

---

Data Download Date/Time: 04/25/2022 19:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: CASSANI

DATE/TIME OF TEST: 04/03/2022 22:16:49

---

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.077	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.210	22:11
Breath Volume	1.199	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.202	22:15
Breath Volume	1.863	22:15
Air Blank	0.000	22:15
Control Test	0.076	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

---

Data Download Date/Time: 04/25/2022 19:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 04/13/2022 00:10:23

---

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.077	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.044	00:05
Breath Volume	2.722	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.043	00:08
Breath Volume	2.941	00:08
Air Blank	0.000	00:09
Control Test	0.078	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

---

Data Download Date/Time: 04/25/2022 19:07



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: TAMAYO

DATE/TIME OF TEST: 04/15/2022 20:56:18

---

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.077	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.253	20:51
Breath Volume	3.585	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.237	20:54
Breath Volume	1.882	20:54
Air Blank	0.000	20:55
Control Test	0.076	20:55
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

---

Data Download Date/Time: 04/25/2022 19:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: GALARZA

DATE/TIME OF TEST: 04/22/2022 01:29:44

---

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.077	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.147	01:24
Breath Volume	1.281	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.163	01:28
Breath Volume	1.750	01:28
Air Blank	0.000	01:28
Control Test	0.077	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

---

Data Download Date/Time: 04/25/2022 19:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/26/2022 02:57:24

---

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:44	
Control Test		0.077	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1	0.209	02:48	Volume Not Met
Breath Volume		0.062	02:48	
Air Blank		0.000	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 2	0.211	02:52	
Breath Volume		1.554	02:52	
Air Blank		0.000	02:53	
Air Blank		0.000	02:55	
Subject Sample	# 3	0.218	02:55	
Breath Volume		2.714	02:55	
Air Blank		0.000	02:56	
Control Test		0.074	02:56	Control Outside Tolerance
Air Blank		0.000	02:57	

---

Data Download Date/Time: 05/18/2022 20:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/26/2022 03:08:10

---

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.077	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.198	03:02
Breath Volume	1.375	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.196	03:06
Breath Volume	2.398	03:06
Air Blank	0.000	03:07
Control Test	0.075	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

---

Data Download Date/Time: 05/18/2022 20:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/25/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: CHIRICO

DATE/TIME OF TEST: 04/26/2022 23:28:39

---

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.077	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.215	23:23
Breath Volume	2.347	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.213	23:26
Breath Volume	1.410	23:26
Air Blank	0.000	23:27
Control Test	0.076	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

---

Data Download Date/Time: 05/18/2022 20:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/01/2022 01:41:36

---

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.164	01:36
Breath Volume	1.250	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.181	01:39
Breath Volume	1.750	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

---

Data Download Date/Time: 04/27/2022 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: BOGCESS

DATE/TIME OF TEST: 04/02/2022 02:41:03

---

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:26	
Control Test		0.080	02:26	
Air Blank		0.000	02:27	
Subject Sample	# 1	0.096	02:30	Volume Not Met
Breath Volume		0.140	02:30	
Air Blank		0.000	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 2	0.120	02:35	
Breath Volume		1.238	02:35	
Air Blank		0.000	02:36	
Air Blank		0.000	02:38	
Subject Sample	# 3	0.122	02:39	
Breath Volume		1.445	02:39	
Air Blank		0.000	02:40	
Control Test		0.076	02:40	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:41	

---

Data Download Date/Time: 04/27/2022 14:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WHETSTONE

DATE/TIME OF TEST: 04/23/2022 02:16:35

---

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.113	02:09
Breath Volume		1.460	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.123	02:14
Breath Volume		1.738	02:14
Air Blank		0.000	02:15
Control Test		0.080	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

---

Data Download Date/Time: 04/27/2022 14:20



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: CUARTERO,LESLIE  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 04/23/2022 05:10:04

---

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.079	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.163	05:05
Breath Volume		1.675	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.169	05:08
Breath Volume		1.613	05:08
Air Blank		0.000	05:09
Control Test		0.078	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:10

---

Data Download Date/Time: 04/27/2022 14:20



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERUZ,GABRIEL,S  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: TERUZ,GABRIEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 3024023571  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: CONSUEGRA TAMAYO

DATE/TIME OF TEST: 04/29/2022 00:43:41

---

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.081	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.218	00:39
Breath Volume	1.554	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.227	00:42
Breath Volume	1.613	00:42
Air Blank	0.000	00:42
Control Test	0.080	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

---

Data Download Date/Time: 05/26/2022 12:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: TEJADA

DATE/TIME OF TEST: 04/01/2022 00:44:37

---

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.078	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:43	
Control Test	0.077	00:44	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

---

Data Download Date/Time: 04/20/2022 07:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO,GEORGE,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: PACHECO,GEORGE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: FUMERO RODRIGUEZ

DATE/TIME OF TEST: 04/01/2022 06:18:02

---

Diagnostics Check		OK	06:10
Air Blank		0.000	06:10
Control Test		0.078	06:11
Air Blank		0.000	06:11
Subject Sample	# 1	0.262	06:13
Breath Volume		2.062	06:13
Air Blank		0.000	06:14
Air Blank		0.000	06:15
Subject Sample	# 2	0.265	06:16
Breath Volume		1.691	06:16
Air Blank		0.000	06:17
Control Test		0.077	06:17
Air Blank		0.000	06:17
Diagnostics Check		OK	06:18

---

Data Download Date/Time: 04/20/2022 07:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO,GEORGE,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: PACHECO,GEORGE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: MORALES MORILLO

DATE/TIME OF TEST: 04/02/2022 02:38:33

---

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:25	
Control Test		0.077	02:25	
Air Blank		0.000	02:26	
Subject Sample	# 1	0.142	02:29	Volume Not Met
Breath Volume		1.066	02:29	
Air Blank		0.000	02:30	
Air Blank		0.000	02:32	
Subject Sample	# 2	0.190	02:33	No .020 agreement
Breath Volume		2.093	02:33	
Air Blank		0.000	02:34	
Air Blank		0.000	02:36	
Subject Sample	# 3	0.151	02:36	No .020 agreement
Breath Volume		1.453	02:36	
Air Blank		0.000	02:37	
Control Test		0.077	02:37	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

---

Data Download Date/Time: 04/20/2022 07:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO,GEORGE,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: PACHECO,GEORGE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: MORALES MORILLO

DATE/TIME OF TEST: 04/02/2022 02:49:02

---

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.077	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.828	02:47	
Air Blank	0.000	02:48	
Control Test	0.076	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:49	

---

Data Download Date/Time: 04/20/2022 07:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: LOUIS,W  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 04/02/2022 21:55:15

---

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.077	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.172	21:50
Breath Volume	2.539	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.172	21:53
Breath Volume	3.222	21:53
Air Blank	0.000	21:54
Control Test	0.078	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

---

Data Download Date/Time: 04/20/2022 07:50



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYER,ALYSSA,J  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: RODRIGUEZ,JOSHUA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: REYESLOPEZ

DATE/TIME OF TEST: 04/03/2022 00:52:27

---

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:50	
Control Test	0.078	00:50	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:51	
Control Test	0.077	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

---

Data Download Date/Time: 04/20/2022 07:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,F  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: CAIAFFA

DATE/TIME OF TEST: 04/10/2022 00:03:43

---

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.079	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.079	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

---

Data Download Date/Time: 04/20/2022 07:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 04/15/2022 04:22:40

---

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:11	
Control Test		0.078	04:11	
Air Blank		0.000	04:12	
Subject Sample	# 1	0.174	04:15	Volume Not Met
Breath Volume		0.144	04:15	
Air Blank		0.000	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 2	0.190	04:20	Volume Not Met
Breath Volume		0.281	04:20	
Air Blank		0.000	04:21	
Control Test		0.076	04:22	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:22	

---

Data Download Date/Time: 04/20/2022 07:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,F  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 04/15/2022 04:37:47

---

Diagnostics Check		OK	04:28
Air Blank		0.000	04:29
Control Test		0.076	04:29
Air Blank		0.000	04:30
Subject Sample	# 1	0.178	04:32
Breath Volume		1.718	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.188	04:36
Breath Volume		1.660	04:36
Air Blank		0.000	04:36
Control Test		0.076	04:37
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

---

Data Download Date/Time: 04/20/2022 07:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: DELGADO,M  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MILIAN

DATE/TIME OF TEST: 04/17/2022 01:32:01

---

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.077	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:31	
Control Test	0.075	01:31	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:32	

---

Data Download Date/Time: 04/20/2022 07:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: DELGADO,M  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 04/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: SEGURA

DATE/TIME OF TEST: 04/22/2022 01:35:04

---

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.078	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.100	01:30
Breath Volume		1.730	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.097	01:33
Breath Volume		1.246	01:33
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:35

---

Data Download Date/Time: 05/05/2022 03:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,F  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 04/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: GADEA ALTAMIRANO

DATE/TIME OF TEST: 04/23/2022 04:19:39

---

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.077	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.182	04:14
Breath Volume	1.250	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.184	04:18
Breath Volume	1.261	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

---

Data Download Date/Time: 05/05/2022 03:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOPEZ,JONATHAN  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: OLVERA GALLARDO

DATE/TIME OF TEST: 04/15/2022 22:06:34

---

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.081	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.184	22:01
Breath Volume	1.582	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.199	22:04
Breath Volume	2.824	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

---

Data Download Date/Time: 05/25/2022 10:20



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: KAMPIRE

DATE/TIME OF TEST: 04/16/2022 00:34:27

---

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.080	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.240	00:29
Breath Volume		1.437	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:31
Subject Sample	# 2	0.237	00:32
Breath Volume		1.910	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

---

Data Download Date/Time: 05/25/2022 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ARMSTRONG,BRITTANY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: DAWKINS JR

DATE/TIME OF TEST: 04/16/2022 03:40:20

---

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.110	03:33
Breath Volume		1.996	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.097	03:38
Breath Volume		1.632	03:38
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

---

Data Download Date/Time: 05/25/2022 10:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GUZMAN,JOSE  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 04/19/2022 23:57:11

---

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.000	23:52
Breath Volume	1.761	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.000	23:55
Breath Volume	1.378	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

---

Data Download Date/Time: 05/25/2022 10:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: DIAZ MORALES

DATE/TIME OF TEST: 04/20/2022 00:52:07

---

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.000	00:47
Breath Volume	3.148	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.000	00:50
Breath Volume	2.859	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

---

Data Download Date/Time: 05/25/2022 10:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GOTSCHALL,BROOKE  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: OBAS

DATE/TIME OF TEST: 04/21/2022 21:11:22

---

Diagnostics Check	OK	20:59	
Air Blank	0.000	21:00	
Control Test	0.080	21:00	
Air Blank	0.000	21:01	
Subject Sample # 1	0.158	21:04	Volume Not Met
Breath Volume	0.000	21:04	
Air Blank	0.000	21:05	
Air Blank	0.000	21:06	
Subject Sample # 2		21:10	No Sample Provided
Breath Volume	0.000	21:10	
Air Blank	0.000	21:10	
Control Test	0.079	21:10	
Air Blank	0.000	21:11	
Diagnostics Check	OK	21:11	

---

Data Download Date/Time: 05/25/2022 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: NIKOLOV

DATE/TIME OF TEST: 04/22/2022 22:48:32

---

Diagnostics Check		OK	22:42
Air Blank		0.000	22:42
Control Test		0.080	22:42
Air Blank		0.000	22:43
Subject Sample	# 1	0.272	22:43
Breath Volume		2.296	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:46
Subject Sample	# 2	0.274	22:46
Breath Volume		1.695	22:46
Air Blank		0.000	22:47
Control Test		0.078	22:47
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

---

Data Download Date/Time: 05/25/2022 10:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GUZMAN,JOSE  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 04/27/2022 03:47:51

---

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.081	03:37	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:40	No Sample Provided
Breath Volume	0.000	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 2	03:46	No Sample Provided
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Control Test	0.081	03:47	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

---

Data Download Date/Time: 05/25/2022 10:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POPICK,SCOTT,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: POPICK,SCOTT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 06:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: CASTOR

DATE/TIME OF TEST: 04/30/2022 07:47:23

---

Diagnostics Check	OK	07:41
Air Blank	0.000	07:41
Control Test	0.081	07:41
Air Blank	0.000	07:42
Subject Sample # 1	0.180	07:42
Breath Volume	1.671	07:42
Air Blank	0.000	07:43
Air Blank	0.000	07:45
Subject Sample # 2	0.176	07:45
Breath Volume	2.343	07:45
Air Blank	0.000	07:46
Control Test	0.079	07:46
Air Blank	0.000	07:47
Diagnostics Check	OK	07:47

---

Data Download Date/Time: 05/25/2022 10:21



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOHM,JONATHAN,  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: BOHM,JONATHAN  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: OSULLIVAN

DATE/TIME OF TEST: 04/21/2022 03:55:12

---

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:54	
Control Test	0.081	03:54	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

---

Data Download Date/Time: 04/28/2022 17:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOHM,JONATHAN,  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: BOHM,JONATHAN  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: JOYCE

DATE/TIME OF TEST: 04/22/2022 02:17:56

---

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.188	02:13
Breath Volume		2.136	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.177	02:16
Breath Volume		1.578	02:16
Air Blank		0.000	02:17
Control Test		0.081	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

---

Data Download Date/Time: 04/28/2022 17:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HULIN

DATE/TIME OF TEST: 04/07/2022 03:36:28

---

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:34	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:35	
Control Test	0.079	03:35	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

---

Data Download Date/Time: 04/28/2022 19:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: AVSHALOMOV

DATE/TIME OF TEST: 04/09/2022 01:59:21

---

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.119	01:54
Breath Volume	1.695	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:56
Subject Sample # 2	0.116	01:57
Breath Volume	1.121	01:57
Air Blank	0.000	01:58
Control Test	0.082	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

---

Data Download Date/Time: 04/28/2022 19:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: GRIER

DATE/TIME OF TEST: 04/09/2022 04:17:03

---

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.236	04:12
Breath Volume	1.597	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.242	04:15
Breath Volume	1.894	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:17

---

Data Download Date/Time: 04/28/2022 19:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: BOHM,JONATHAN  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/13/2022 00:27:13

---

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.081	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.100	00:22
Breath Volume		1.566	00:22
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.103	00:25
Breath Volume		1.765	00:25
Air Blank		0.000	00:26
Control Test		0.081	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

---

Data Download Date/Time: 04/28/2022 19:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: PRESTERA

DATE/TIME OF TEST: 04/14/2022 00:27:10

---

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.126	00:22
Breath Volume	1.457	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.127	00:25
Breath Volume	1.585	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

---

Data Download Date/Time: 04/28/2022 19:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: MARTINSON

DATE/TIME OF TEST: 04/15/2022 00:09:35

---

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.388	00:04
Breath Volume		1.261	00:04
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.380	00:07
Breath Volume		1.886	00:07
Air Blank		0.000	00:08
Control Test		0.077	00:08
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

---

Data Download Date/Time: 04/28/2022 19:30



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: MONTAS

DATE/TIME OF TEST: 04/16/2022 03:02:11

---

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

---

Data Download Date/Time: 04/28/2022 19:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MICHEL

DATE/TIME OF TEST: 04/20/2022 03:22:15

---

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.081	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

---

Data Download Date/Time: 04/28/2022 19:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/29/2022 04:12:15

---

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.169	04:07
Breath Volume	1.281	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.172	04:10
Breath Volume	1.632	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time: 06/06/2022 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: JAY

DATE/TIME OF TEST: 04/30/2022 03:44:56

---

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample	# 2	0.000	03:40
Breath Volume	1.144	03:40	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 3	0.000	03:43
Breath Volume	1.562	03:43	
Air Blank	0.000	03:44	
Control Test	0.080	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

---

Data Download Date/Time: 06/06/2022 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ,J.,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: HERNANDEZ,J.  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/25/2022 02:33:33

---

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:21	
Control Test		0.078	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 1	0.184	02:23	
Breath Volume		1.210	02:23	
Air Blank		0.000	02:23	
Air Blank		0.000	02:25	
Subject Sample	# 2	0.165	02:28	Volume Not Met
Breath Volume		0.683	02:28	
Air Blank		0.000	02:29	
Air Blank		0.000	02:31	
Subject Sample	# 3	0.179	02:31	
Breath Volume		1.242	02:31	
Air Blank		0.000	02:32	
Control Test		0.076	02:32	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

---

Data Download Date/Time: 05/05/2022 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOWNSEND,ROBERT,B  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: GRAY,J  
ARRESTING OFFICER AGENCY: MARCO ISLAND PD  
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 14:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/25/2022 16:45:03

---

Diagnostics Check	OK	16:38
Air Blank	0.000	16:38
Control Test	0.078	16:38
Air Blank	0.000	16:39
Subject Sample # 1	0.163	16:40
Breath Volume	2.468	16:40
Air Blank	0.000	16:40
Air Blank	0.000	16:42
Subject Sample # 2	0.160	16:43
Breath Volume	2.121	16:43
Air Blank	0.000	16:44
Control Test	0.078	16:44
Air Blank	0.000	16:44
Diagnostics Check	OK	16:45

---

Data Download Date/Time: 05/05/2022 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LANDRY,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: JAMISON

DATE/TIME OF TEST: 04/22/2022 00:21:23

---

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:15
Subject Sample # 1	0.160	00:16
Breath Volume	3.289	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.153	00:19
Breath Volume	2.597	00:19
Air Blank	0.000	00:20
Control Test	0.078	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

---

Data Download Date/Time: 05/17/2022 13:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTER,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: DUNNE

DATE/TIME OF TEST: 04/22/2022 02:41:38

---

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.190	02:36
Breath Volume	2.625	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.181	02:40
Breath Volume	1.742	02:40
Air Blank	0.000	02:40
Control Test	0.078	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

---

Data Download Date/Time: 05/17/2022 13:27



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,ANGELA,V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ODONNELL,TYLER  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 07:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:47

SUBJECT'S LAST NAME: MARTINEZ PUPO

DATE/TIME OF TEST: 04/22/2022 08:16:40

---

Diagnostics Check	OK	08:09
Air Blank	0.000	08:09
Control Test	0.079	08:10
Air Blank	0.000	08:10
Subject Sample # 1	0.021	08:11
Breath Volume	1.738	08:11
Air Blank	0.000	08:12
Air Blank	0.000	08:14
Subject Sample # 2	0.021	08:15
Breath Volume	1.257	08:15
Air Blank	0.000	08:15
Control Test	0.079	08:16
Air Blank	0.000	08:16
Diagnostics Check	OK	08:16

---

Data Download Date/Time: 05/17/2022 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: CONYERS

DATE/TIME OF TEST: 04/22/2022 22:34:14

---

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:19	
Control Test		0.079	22:19	
Air Blank		0.000	22:20	
Subject Sample	# 1		22:23	No Sample Provided
Breath Volume		0.000	22:23	
Air Blank		0.000	22:23	
Air Blank		0.000	22:25	
Subject Sample	# 2	0.256	22:27	
Breath Volume		1.437	22:27	
Air Blank		0.000	22:27	
Air Blank		0.000	22:29	
Subject Sample	# 3	0.268	22:32	
Breath Volume		1.792	22:32	
Air Blank		0.000	22:33	
Control Test		0.078	22:33	
Air Blank		0.000	22:34	
Diagnostics Check		OK	22:34	

---

Data Download Date/Time: 05/17/2022 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARRY,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/22/2022 23:24:34

---

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:10	
Control Test	0.078	23:10	
Air Blank	0.000	23:11	
Subject Sample	# 1 0.478	23:11	
Breath Volume	1.984	23:11	
Air Blank	0.000	23:12	
Air Blank	0.000	23:14	
Subject Sample	# 2	23:17	No Sample Provided
Breath Volume	0.000	23:17	
Air Blank	0.000	23:18	
Air Blank	0.000	23:20	
Subject Sample	# 3	23:23	No Sample Provided
Breath Volume	0.000	23:23	
Air Blank	0.000	23:23	
Control Test	0.079	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

---

Data Download Date/Time: 05/17/2022 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARBER,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: OSARIO

DATE/TIME OF TEST: 04/23/2022 00:28:15

---

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.145	00:23
Breath Volume	2.121	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.144	00:26
Breath Volume	1.898	00:26
Air Blank	0.000	00:27
Control Test	0.078	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

---

Data Download Date/Time: 05/17/2022 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GONZALEZ,C  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: MUSTIAN

DATE/TIME OF TEST: 04/23/2022 03:44:53

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.185	03:40
Breath Volume	2.656	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.181	03:43
Breath Volume	2.011	03:43
Air Blank	0.000	03:43
Control Test	0.078	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time: 05/17/2022 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: EARLYWINE,G  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: BOWDEN

DATE/TIME OF TEST: 04/23/2022 06:25:23

---

Diagnostics Check	OK	06:21	
Air Blank	0.000	06:21	
Control Test	0.079	06:22	
Air Blank	0.000	06:22	
Subject Sample	# 1	06:24	Subject Test Refused
Breath Volume	0.000	06:24	
Air Blank	0.000	06:24	
Control Test	0.080	06:24	
Air Blank	0.000	06:25	
Diagnostics Check	OK	06:25	

---

Data Download Date/Time: 05/17/2022 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER,MELONIE,E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CRAVER,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: TERWILLIGER

DATE/TIME OF TEST: 04/23/2022 17:11:47

---

Diagnostics Check	OK	17:05
Air Blank	0.000	17:06
Control Test	0.079	17:06
Air Blank	0.000	17:07
Subject Sample # 1	0.000	17:07
Breath Volume	3.570	17:07
Air Blank	0.000	17:08
Air Blank	0.000	17:10
Subject Sample # 2	0.000	17:10
Breath Volume	2.992	17:10
Air Blank	0.000	17:10
Control Test	0.079	17:11
Air Blank	0.000	17:11
Diagnostics Check	OK	17:11

---

Data Download Date/Time: 05/17/2022 13:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILSON,N  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/24/2022 03:22:02

---

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.129	03:17
Breath Volume	3.097	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.135	03:20
Breath Volume	3.242	03:20
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

---

Data Download Date/Time: 05/17/2022 13:28



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DE LIMA,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: CANARTE

DATE/TIME OF TEST: 04/24/2022 05:09:37

---

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.175	05:04
Breath Volume	2.035	05:04
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.179	05:08
Breath Volume	2.167	05:08
Air Blank	0.000	05:08
Control Test	0.078	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

---

Data Download Date/Time: 05/17/2022 13:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILLIAMS,P  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: GONZALEZ PEREZ

DATE/TIME OF TEST: 04/24/2022 23:23:12

---

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.080	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.257	23:16
Breath Volume		1.265	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.267	23:21
Breath Volume		1.992	23:21
Air Blank		0.000	23:22
Control Test		0.079	23:22
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

---

Data Download Date/Time: 05/17/2022 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HERNANDEZ,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: NAVA NOLAZCO

DATE/TIME OF TEST: 04/25/2022 05:51:23

---

Diagnostics Check		OK	05:43
Air Blank		0.000	05:43
Control Test		0.079	05:43
Air Blank		0.000	05:44
Subject Sample	# 1	0.154	05:46
Breath Volume		2.406	05:46
Air Blank		0.000	05:47
Air Blank		0.000	05:48
Subject Sample	# 2	0.141	05:49
Breath Volume		1.605	05:49
Air Blank		0.000	05:50
Control Test		0.077	05:50
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

---

Data Download Date/Time: 05/17/2022 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILILLO,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: TRAMEL

DATE/TIME OF TEST: 04/26/2022 01:06:16

---

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.180	01:01
Breath Volume	2.042	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.187	01:04
Breath Volume	2.039	01:04
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

---

Data Download Date/Time: 05/17/2022 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTER,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: MENDOSA

DATE/TIME OF TEST: 04/26/2022 21:54:56

---

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.079	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.175	21:49
Breath Volume	2.320	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.181	21:53
Breath Volume	2.882	21:53
Air Blank	0.000	21:54
Control Test	0.079	21:54
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

---

Data Download Date/Time: 05/17/2022 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY, JANICE, F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RODRIGUEZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: BOLTON

DATE/TIME OF TEST: 04/27/2022 00:20:49

---

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:14	
Control Test	0.079	00:15	
Air Blank	0.000	00:15	
Subject Sample # 1	0.000	00:17	
Breath Volume	2.867	00:17	
Air Blank	0.000	00:17	
Air Blank	0.000	00:19	
Subject Sample # 2	0.000	00:20	
Breath Volume	2.746	00:20	
Air Blank		00:20	Improper Sample
Air Blank	0.000	00:20	

---

Data Download Date/Time: 05/17/2022 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RODRIGUEZ,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: BOLTON

DATE/TIME OF TEST: 04/27/2022 00:32:59

---

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.079	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.000	00:28
Breath Volume	3.125	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.000	00:31
Breath Volume	2.390	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

---

Data Download Date/Time: 05/17/2022 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: SMITH JR

DATE/TIME OF TEST: 04/27/2022 01:43:54

---

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.257	01:39
Breath Volume	3.179	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample # 2	0.262	01:42
Breath Volume	3.132	01:42
Air Blank	0.000	01:42
Control Test	0.078	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

---

Data Download Date/Time: 05/17/2022 13:29



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/29/2022 00:35:21

---

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.096	00:30
Breath Volume	2.136	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.094	00:33
Breath Volume	2.316	00:33
Air Blank	0.000	00:34
Control Test	0.079	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

---

Data Download Date/Time: 05/17/2022 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CLARK,I  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SAUBAN

DATE/TIME OF TEST: 04/29/2022 22:02:28

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.000	21:57
Breath Volume	3.144	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.000	22:01
Breath Volume	2.871	22:01
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

---

Data Download Date/Time: 05/17/2022 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREENLEE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CROSBY,J  
ARRESTING OFFICER AGENCY: Seminole Res (Hills)  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 09:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 04/30/2022 10:29:56

---

Diagnostics Check	OK	10:23
Air Blank	0.000	10:23
Control Test	0.079	10:24
Air Blank	0.000	10:24
Subject Sample # 1	0.117	10:25
Breath Volume	2.527	10:25
Air Blank	0.000	10:26
Air Blank	0.000	10:27
Subject Sample # 2	0.120	10:28
Breath Volume	2.816	10:28
Air Blank	0.000	10:29
Control Test	0.079	10:29
Air Blank	0.000	10:29
Diagnostics Check	OK	10:29

---

Data Download Date/Time: 05/17/2022 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CHOI,H  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: GUZMAN PEREZ

DATE/TIME OF TEST: 04/30/2022 20:24:23

---

Diagnostics Check		OK	20:17
Air Blank		0.000	20:18
Control Test		0.080	20:18
Air Blank		0.000	20:19
Subject Sample	# 1	0.209	20:19
Breath Volume		2.808	20:19
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.212	20:22
Breath Volume		2.390	20:22
Air Blank		0.000	20:23
Control Test		0.078	20:23
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

---

Data Download Date/Time: 05/17/2022 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 04/30/2022 21:00:09

---

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.079	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.220	20:55
Breath Volume	2.000	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:57
Subject Sample # 2	0.220	20:58
Breath Volume	1.480	20:58
Air Blank	0.000	20:59
Control Test	0.079	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

---

Data Download Date/Time: 05/17/2022 13:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHEY,REX,M  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: WARD,SHANE  
ARRESTING OFFICER AGENCY: Hardee County SO  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/23/2022 17:40:41

---

Diagnostics Check	Fail	17:40	Internal Printer Test
Air Blank	0.000	17:40	

---

Data Download Date/Time: 04/28/2022 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHEY, REX, M  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: WARD, SHANE  
ARRESTING OFFICER AGENCY: Hardee County SO  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/23/2022 17:49:58

---

Diagnostics Check	OK	17:43
Air Blank	0.000	17:44
Control Test	0.077	17:44
Air Blank	0.000	17:44
Subject Sample # 1	0.194	17:45
Breath Volume	2.402	17:45
Air Blank	0.000	17:46
Air Blank	0.000	17:47
Subject Sample # 2	0.180	17:48
Breath Volume	1.675	17:48
Air Blank	0.000	17:48
Control Test	0.076	17:49
Air Blank	0.000	17:49
Diagnostics Check	OK	17:49

---

Data Download Date/Time: 04/28/2022 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SALAS PEREZ

DATE/TIME OF TEST: 04/02/2022 01:29:12

---

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.079	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:22	No Sample Provided
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Air Blank	0.000	01:24	
Subject Sample	# 2	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Control Test	0.079	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

---

Data Download Date/Time: 04/19/2022 11:38



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LANDRY,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: FUNSCH

DATE/TIME OF TEST: 04/02/2022 04:32:44

---

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.152	04:28
Breath Volume		2.007	04:28
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.154	04:31
Breath Volume		2.589	04:31
Air Blank		0.000	04:31
Control Test		0.078	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

---

Data Download Date/Time: 04/19/2022 11:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: COMM VEHIC  
ARREST TIME: 05:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: PINCKNEY

DATE/TIME OF TEST: 04/02/2022 06:27:50

---

Diagnostics Check	OK	06:21
Air Blank	0.000	06:21
Control Test	0.078	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.074	06:23
Breath Volume	2.050	06:23
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.073	06:26
Breath Volume	2.261	06:26
Air Blank	0.000	06:26
Control Test	0.078	06:27
Air Blank	0.000	06:27
Diagnostics Check	OK	06:27

---

Data Download Date/Time: 04/19/2022 11:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTER,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 04/03/2022 01:52:54

---

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.079	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.176	01:48
Breath Volume		2.058	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 2	0.170	01:51
Breath Volume		1.300	01:51
Air Blank		0.000	01:51
Control Test		0.079	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

---

Data Download Date/Time: 04/19/2022 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 04/03/2022 03:48:27

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.146	03:42
Breath Volume	1.316	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.143	03:46
Breath Volume	1.945	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

---

Data Download Date/Time: 04/19/2022 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: JENKINS,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SENORY

DATE/TIME OF TEST: 04/03/2022 04:00:35

---

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.078	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1 0.111	03:59	Volume Not Met
Breath Volume	0.109	03:59	
Air Blank		03:59	Sequence Aborted
Air Blank	0.000	04:00	

---

Data Download Date/Time: 04/19/2022 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: JENKINS,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: SENORY

DATE/TIME OF TEST: 04/03/2022 04:33:56

---

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.078	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.119	04:32
Breath Volume	1.515	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:33

---

Data Download Date/Time: 04/19/2022 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: JENKINS,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: SENORY

DATE/TIME OF TEST: 04/03/2022 05:06:44

---

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.078	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.123	05:01
Breath Volume	2.972	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.110	05:05
Breath Volume	2.753	05:05
Air Blank	0.000	05:05
Control Test	0.078	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

---

Data Download Date/Time: 04/19/2022 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILILLO,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: PERDOMO PONCE

DATE/TIME OF TEST: 04/06/2022 03:25:13

---

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.203	03:20
Breath Volume	2.519	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:22
Subject Sample # 2	0.202	03:23
Breath Volume	2.398	03:23
Air Blank	0.000	03:24
Control Test	0.077	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

---

Data Download Date/Time: 04/19/2022 11:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: LOPEZ GOMEZ

DATE/TIME OF TEST: 04/06/2022 04:16:27

---

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.078	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.270	04:11
Breath Volume	2.371	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.256	04:14
Breath Volume	1.875	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time: 04/19/2022 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTER,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: VELAZQUEZ PEREZ

DATE/TIME OF TEST: 04/07/2022 01:36:10

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.156	01:31
Breath Volume	1.945	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:33
Subject Sample # 2	0.150	01:34
Breath Volume	1.828	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

---

Data Download Date/Time: 04/19/2022 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LANDRY,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 04/07/2022 22:48:47

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.227	22:42
Breath Volume	1.488	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.231	22:46
Breath Volume	3.261	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

---

Data Download Date/Time: 04/19/2022 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MEINKE

DATE/TIME OF TEST: 04/09/2022 02:11:27

---

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.103	02:06
Breath Volume	2.863	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.099	02:09
Breath Volume	2.058	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

---

Data Download Date/Time: 04/19/2022 11:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LEISCH,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: ALEJANDRO

DATE/TIME OF TEST: 04/09/2022 02:57:13

---

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.079	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.150	02:52
Breath Volume	1.597	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.155	02:55
Breath Volume	1.796	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

---

Data Download Date/Time: 04/19/2022 11:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: BENCK

DATE/TIME OF TEST: 04/09/2022 05:00:35

---

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.106	04:56
Breath Volume	2.312	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.105	04:58
Breath Volume	1.746	04:58
Air Blank	0.000	04:59
Control Test	0.077	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

---

Data Download Date/Time: 04/19/2022 11:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILILLO,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: BLISS

DATE/TIME OF TEST: 04/10/2022 01:30:00

---

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.099	01:25
Breath Volume	2.527	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.099	01:28
Breath Volume	2.386	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:30

---

Data Download Date/Time: 04/19/2022 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FISHER,T  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: PONS PUPO

DATE/TIME OF TEST: 04/10/2022 04:14:44

---

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.091	04:10
Breath Volume		2.613	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.095	04:13
Breath Volume		3.210	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

---

Data Download Date/Time: 04/19/2022 11:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ,C  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: GARCIA DIAZ

DATE/TIME OF TEST: 04/10/2022 04:38:25

---

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.138	04:33
Breath Volume	1.179	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.151	04:36
Breath Volume	1.605	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

---

Data Download Date/Time: 04/19/2022 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILLIAMS,P  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 04/11/2022 00:29:33

---

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.081	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.088	00:24
Breath Volume	2.195	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.085	00:27
Breath Volume	1.687	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

---

Data Download Date/Time: 04/19/2022 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: MCATEER

DATE/TIME OF TEST: 04/11/2022 01:53:24

---

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.080	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.225	01:48
Breath Volume		2.726	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.225	01:51
Breath Volume		3.242	01:51
Air Blank		0.000	01:52
Control Test		0.079	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

---

Data Download Date/Time: 04/19/2022 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ,C  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ILLISCAVITCH

DATE/TIME OF TEST: 04/11/2022 03:33:57

---

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.205	03:28
Breath Volume		2.203	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.218	03:32
Breath Volume		1.890	03:32
Air Blank		0.000	03:32
Control Test		0.079	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

---

Data Download Date/Time: 04/19/2022 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILLIAMS,P  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: LANDRY

DATE/TIME OF TEST: 04/11/2022 03:54:11

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.124	03:49
Breath Volume	1.628	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:51
Subject Sample # 2	0.122	03:52
Breath Volume	2.253	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

---

Data Download Date/Time: 04/19/2022 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAAR,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: FOUNTAIN

DATE/TIME OF TEST: 04/15/2022 23:45:55

---

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:41	No Sample Provided
Breath Volume	0.000	23:41	
Air Blank	0.000	23:42	
Air Blank	0.000	23:44	
Subject Sample	# 2	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:44	
Control Test	0.079	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

---

Data Download Date/Time: 04/19/2022 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DOBSON,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/16/2022 03:41:04

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.178	03:36
Breath Volume	1.894	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:38
Subject Sample # 2	0.162	03:39
Breath Volume	1.984	03:39
Air Blank	0.000	03:40
Control Test	0.077	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:41

---

Data Download Date/Time: 04/19/2022 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 04/16/2022 04:18:16

---

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.078	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.147	04:13
Breath Volume	1.347	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.144	04:16
Breath Volume	2.269	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

---

Data Download Date/Time: 04/19/2022 11:40



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RODRIGUEZ,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 04/17/2022 04:33:39

---

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.141	04:28
Breath Volume	2.449	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.143	04:31
Breath Volume	2.480	04:31
Air Blank	0.000	04:32
Control Test	0.078	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

---

Data Download Date/Time: 04/19/2022 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: EARLYWINE,G  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 04/19/2022 04:20:20

---

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.164	04:15
Breath Volume	2.308	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.158	04:18
Breath Volume	1.699	04:18
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

---

Data Download Date/Time: 04/19/2022 11:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 04/22/2022 04:02:05

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.146	03:57
Breath Volume	2.445	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 2	0.140	04:00
Breath Volume	1.273	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

---

Data Download Date/Time: 05/17/2022 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: ARCIA

DATE/TIME OF TEST: 04/23/2022 04:38:24

---

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.148	04:33
Breath Volume		1.796	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:35
Subject Sample	# 2	0.149	04:36
Breath Volume		1.718	04:36
Air Blank		0.000	04:37
Control Test		0.079	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

---

Data Download Date/Time: 05/17/2022 12:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: CEBULA

DATE/TIME OF TEST: 04/23/2022 06:10:18

---

Diagnostics Check	OK	06:06	
Air Blank	0.000	06:06	
Control Test	0.080	06:06	
Air Blank	0.000	06:07	
Subject Sample	# 1	06:08	Subject Test Refused
Breath Volume	0.000	06:08	
Air Blank	0.000	06:09	
Control Test	0.080	06:09	
Air Blank	0.000	06:10	
Diagnostics Check	OK	06:10	

---

Data Download Date/Time: 05/17/2022 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 04/23/2022 23:32:12

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.167	23:27
Breath Volume	1.261	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:29
Subject Sample # 2	0.169	23:30
Breath Volume	1.218	23:30
Air Blank	0.000	23:31
Control Test	0.078	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

---

Data Download Date/Time: 05/17/2022 12:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MEYERS,J  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BECERRA

DATE/TIME OF TEST: 04/24/2022 04:39:21

---

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.099	04:34
Breath Volume	1.664	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.097	04:37
Breath Volume	1.394	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

---

Data Download Date/Time: 05/17/2022 12:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VANTREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: CHRISTIANI

DATE/TIME OF TEST: 04/25/2022 00:04:12

---

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.079	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.207	23:59
Breath Volume		1.207	23:59
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.203	00:02
Breath Volume		1.808	00:02
Air Blank		0.000	00:03
Control Test		0.079	00:03
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

---

Data Download Date/Time: 05/17/2022 12:43



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GRECO,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DUBOSE

DATE/TIME OF TEST: 04/30/2022 02:52:10

---

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.079	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.209	02:47
Breath Volume		2.734	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:49
Subject Sample	# 2	0.206	02:50
Breath Volume		2.750	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

---

Data Download Date/Time: 05/17/2022 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAITE,KENNETH,J  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: WAITE,KENNETH  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/01/2022 02:43:38

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.175	02:39
Breath Volume	3.816	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.173	02:41
Breath Volume	3.246	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

---

Data Download Date/Time: 04/07/2022 09:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: CHLEBANOWSKI,RYAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 04/02/2022 19:56:47

---

Diagnostics Check	OK	19:50
Air Blank	0.000	19:50
Control Test	0.079	19:50
Air Blank	0.000	19:51
Subject Sample # 1	0.139	19:52
Breath Volume	1.632	19:52
Air Blank	0.000	19:53
Air Blank	0.000	19:54
Subject Sample # 2	0.137	19:55
Breath Volume	1.460	19:55
Air Blank	0.000	19:55
Control Test	0.079	19:56
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

---

Data Download Date/Time: 04/07/2022 09:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: CHLEBANOWSKI,RYAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: TORO

DATE/TIME OF TEST: 04/08/2022 02:26:29

---

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.080	02:24	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:25	Subject Test Refused
Breath Volume	0.000	02:25	
Air Blank	0.000	02:25	
Control Test	0.080	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

---

Data Download Date/Time: 05/10/2022 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILROY,ZACHARY,I  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: GILROY,ZACHARY  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: PIEDRA

DATE/TIME OF TEST: 04/08/2022 02:37:16

---

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.080	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

---

Data Download Date/Time: 05/10/2022 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAITE,KENNETH,J  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: GALBO,NICHOLAS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: FRITZBERG

DATE/TIME OF TEST: 04/09/2022 04:38:41

---

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:36	
Control Test	0.081	04:36	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:37	
Control Test	0.081	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

---

Data Download Date/Time: 05/10/2022 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: GRUBA,JACK  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SANTIAGO CRUZ

DATE/TIME OF TEST: 04/10/2022 02:19:22

---

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.181	02:14
Breath Volume	1.460	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.185	02:17
Breath Volume	2.304	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time: 05/10/2022 15:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TELLIER, JEREMEY, J  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: HERNANDEZ, LEONARDO  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: PENA VASQUEZ

DATE/TIME OF TEST: 04/18/2022 05:13:39

---

Diagnostics Check		OK	05:07
Air Blank		0.000	05:07
Control Test		0.080	05:08
Air Blank		0.000	05:08
Subject Sample	# 1	0.213	05:09
Breath Volume		2.023	05:09
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.207	05:11
Breath Volume		1.949	05:11
Air Blank		0.000	05:12
Control Test		0.079	05:13
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

---

Data Download Date/Time: 05/10/2022 15:45



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODDY,AARON,A  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HERNANDEZ,LEONARDO  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 16:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: RASPOLICH

DATE/TIME OF TEST: 04/21/2022 17:40:48

---

Diagnostics Check	OK	17:26	
Air Blank	0.000	17:26	
Control Test	0.080	17:27	
Air Blank	0.000	17:27	
Subject Sample	# 1	17:30	No Sample Provided
Breath Volume	0.000	17:30	
Air Blank	0.000	17:31	
Air Blank	0.000	17:33	
Subject Sample	# 2	0.314	17:35
Breath Volume	1.628	17:35	
Air Blank	0.000	17:36	
Air Blank	0.000	17:38	
Subject Sample	# 3	0.297	17:39
Breath Volume	1.687	17:39	
Air Blank	0.000	17:39	
Control Test	0.078	17:40	
Air Blank	0.000	17:40	
Diagnostics Check	OK	17:40	

---

Data Download Date/Time: 05/10/2022 15:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: CHLEBANOWSKI,RYAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: RAIGOZA

DATE/TIME OF TEST: 04/23/2022 02:40:03

---

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.249	02:31
Breath Volume	3.070	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.226	02:34
Breath Volume	1.812	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 3	0.229	02:38
Breath Volume	3.242	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:40

---

Data Download Date/Time: 05/10/2022 15:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODDY,AARON,A  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: RODDY,AARON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 16:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: DUHAME

DATE/TIME OF TEST: 04/23/2022 17:20:02

---

Diagnostics Check	OK	17:16	
Air Blank	0.000	17:16	
Control Test	0.080	17:17	
Air Blank	0.000	17:17	
Subject Sample	# 1	17:18	Subject Test Refused
Breath Volume	0.000	17:18	
Air Blank	0.000	17:19	
Control Test	0.080	17:19	
Air Blank	0.000	17:19	
Diagnostics Check	OK	17:20	

---

Data Download Date/Time: 05/10/2022 15:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MARTIN-REYES,JUAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: CLATON

DATE/TIME OF TEST: 04/23/2022 20:18:18

---

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.118	20:13
Breath Volume	1.125	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:15
Subject Sample # 2	0.119	20:16
Breath Volume	1.183	20:16
Air Blank	0.000	20:17
Control Test	0.079	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

---

Data Download Date/Time: 05/10/2022 15:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMSON, BRANDEE, E  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: WILLIAMSON, BRANDEE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/25/2022 21:36:58

---

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.164	21:32
Breath Volume	1.277	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:34
Subject Sample # 2	0.167	21:35
Breath Volume	1.738	21:35
Air Blank	0.000	21:36
Control Test	0.077	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

---

Data Download Date/Time: 05/10/2022 15:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILROY,ZACHARY,I  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: GILROY,ZACHARY  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: GALLI

DATE/TIME OF TEST: 04/30/2022 01:14:46

---

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.080	01:12	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:13	
Control Test	0.081	01:14	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

---

Data Download Date/Time: 05/10/2022 15:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,JORGE,L  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: HERNANDEZ,LEONARDO  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: EMPER

DATE/TIME OF TEST: 04/30/2022 02:20:05

---

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.080	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.319	02:15
Breath Volume		2.902	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:17
Subject Sample	# 2	0.309	02:18
Breath Volume		2.171	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

---

Data Download Date/Time: 05/10/2022 15:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FIERRO,ALEXANDER,D  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: FIERRO,ALEXANDER  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: GALE

DATE/TIME OF TEST: 04/30/2022 02:45:01

---

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:45	

---

Data Download Date/Time: 05/10/2022 15:45



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATKINS,MICHAEL,W  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: BRABLC,M  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:02

SUBJECT'S LAST NAME: THERIAULT

DATE/TIME OF TEST: 04/30/2022 08:34:48

---

Diagnostics Check		OK	08:22	
Air Blank		0.000	08:23	
Control Test		0.080	08:23	
Air Blank		0.000	08:24	
Subject Sample	# 1	0.172	08:27	Volume Not Met
Breath Volume		0.031	08:27	
Air Blank		0.000	08:28	
Air Blank		0.000	08:29	
Subject Sample	# 2	0.184	08:32	Volume Not Met
Breath Volume		0.128	08:32	
Air Blank		0.000	08:33	
Control Test		0.077	08:34	
Air Blank		0.000	08:34	
Diagnostics Check		OK	08:34	

---

Data Download Date/Time: 05/10/2022 15:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARINO, JUAN CARLOS,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: O'NEILL, CONNOR  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: BOTELHO II

DATE/TIME OF TEST: 04/01/2022 19:41:46

---

Diagnostics Check	OK	19:38	
Air Blank	0.000	19:39	
Control Test	0.078	19:39	
Air Blank	0.000	19:40	
Subject Sample	# 1	19:40	Subject Test Refused
Breath Volume	0.000	19:40	
Air Blank	0.000	19:40	
Control Test	0.078	19:41	
Air Blank	0.000	19:41	
Diagnostics Check	OK	19:41	

---

Data Download Date/Time: 04/07/2022 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MARINO,JUAN CARLOS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: JENSEN

DATE/TIME OF TEST: 04/02/2022 00:57:37

---

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.079	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:56	Subject Test Refused
Breath Volume	0.000	00:56	
Air Blank	0.000	00:56	
Control Test	0.078	00:57	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

---

Data Download Date/Time: 04/07/2022 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JOEL,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HRADECKY,CORBIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: PARRADO

DATE/TIME OF TEST: 04/02/2022 02:10:23

---

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:07	
Control Test	0.078	02:08	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Control Test	0.079	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

---

Data Download Date/Time: 04/07/2022 15:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HRADECKY,CORBIN,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: ONEILL,CONNOR  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/06/2022 21:47:11

---

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:42	
Control Test	0.078	21:43	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:45	RFI Detect
Breath Volume	0.882	21:45	
Air Blank	0.000	21:46	
Control Test	0.078	21:46	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

---

Data Download Date/Time: 04/07/2022 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HRADECKY,CORBIN,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: ONEILL,CONNOR  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/06/2022 21:59:42

---

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:48	
Control Test	0.077	21:49	
Air Blank	0.000	21:49	
Subject Sample # 1	0.356	21:53	
Breath Volume	1.207	21:53	
Air Blank	0.000	21:54	
Air Blank	0.000	21:55	
Subject Sample # 2		21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:58	
Control Test	0.074	21:59	Control Outside Tolerance
Air Blank	0.000	21:59	

---

Data Download Date/Time: 04/07/2022 15:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MORALES, ALEXANDER  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 04/09/2022 05:00:02

---

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.078	04:57	
Air Blank	0.000	04:58	
Subject Sample	# 1	04:58	Subject Test Refused
Breath Volume	0.000	04:58	
Air Blank	0.000	04:59	
Control Test	0.079	04:59	
Air Blank	0.000	04:59	
Diagnostics Check	OK	05:00	

---

Data Download Date/Time: 05/11/2022 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILL,THOMAS,C  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: GUERRA,WILLIAM  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 04/10/2022 23:05:26

---

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:54	
Control Test	0.078	22:54	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:58	No Sample Provided
Breath Volume	0.000	22:58	
Air Blank	0.000	22:58	
Air Blank	0.000	23:00	
Subject Sample	# 2	23:04	No Sample Provided
Breath Volume	0.000	23:04	
Air Blank	0.000	23:04	
Control Test	0.078	23:04	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

---

Data Download Date/Time: 05/11/2022 11:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: WARE ROGER, BRANDON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: NASON

DATE/TIME OF TEST: 04/13/2022 23:51:35

---

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:49	
Control Test	0.079	23:49	
Air Blank	0.000	23:49	
Subject Sample	# 1	23:50	Subject Test Refused
Breath Volume	0.000	23:50	
Air Blank	0.000	23:50	
Control Test	0.078	23:50	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

---

Data Download Date/Time: 05/11/2022 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HRADECKY,CORBIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 04/22/2022 04:48:04

---

Diagnostics Check		OK	04:34	
Air Blank		0.000	04:35	
Control Test		0.078	04:35	
Air Blank		0.000	04:36	
Subject Sample	# 1	0.088	04:39	Volume Not Met
Breath Volume		0.035	04:39	
Air Blank		0.000	04:39	
Air Blank		0.000	04:41	
Subject Sample	# 2	0.122	04:43	
Breath Volume		1.941	04:43	
Air Blank		0.000	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 3	0.125	04:46	
Breath Volume		1.511	04:46	
Air Blank		0.000	04:47	
Control Test		0.076	04:47	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:48	

---

Data Download Date/Time: 05/11/2022 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILROY,ZACHARY,I  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: FOSTER,MEGHAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: RODENBERG

DATE/TIME OF TEST: 04/23/2022 23:38:59

---

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.078	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.208	23:34
Breath Volume	1.941	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.213	23:37
Breath Volume	2.523	23:37
Air Blank	0.000	23:38
Control Test	0.078	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

---

Data Download Date/Time: 05/11/2022 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HILL, THOMAS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: SARMIENTO

DATE/TIME OF TEST: 04/24/2022 01:51:04

---

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.078	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:50	
Control Test	0.078	01:50	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:51	

---

Data Download Date/Time: 05/11/2022 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNDORFF,TERRY,D  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MUNDORFF,TERRY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/27/2022 18:47:21

---

Diagnostics Check	OK	18:40
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.219	18:42
Breath Volume	1.890	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:44
Subject Sample # 2	0.218	18:45
Breath Volume	1.582	18:45
Air Blank	0.000	18:46
Control Test	0.078	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

---

Data Download Date/Time: 05/06/2022 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRINGTON,A,M  
BREATH TEST OPERATOR AGENCY: Belleview PD  
ARRESTING OFFICER: HARRINGTON,A  
ARRESTING OFFICER AGENCY: Belleview PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1355293  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: ROLESON

DATE/TIME OF TEST: 04/01/2022 02:49:46

---

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.079	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.174	02:45
Breath Volume		2.503	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.170	02:48
Breath Volume		2.550	02:48
Air Blank		0.000	02:48
Control Test		0.079	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: ORCHARD,DUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1355293  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 04/01/2022 21:27:28

---

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.109	21:22
Breath Volume	3.390	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.105	21:25
Breath Volume	2.660	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1355293  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/02/2022 02:13:29

---

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.080	02:08
Breath Volume	1.503	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.082	02:11
Breath Volume	2.523	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

---

Data Download Date/Time: 04/27/2022 15:48



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,JUSTIN,G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,JUSTIN  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1355293  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MERRILL

DATE/TIME OF TEST: 04/02/2022 04:34:42

---

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.205	04:29
Breath Volume	3.460	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:31
Subject Sample # 2	0.190	04:32
Breath Volume	4.195	04:32
Air Blank	0.000	04:33
Control Test	0.078	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANZA,CHRISTIAN,E  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: LANZA,CHRISTIAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1355293  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: BUCKMAN

DATE/TIME OF TEST: 04/03/2022 18:33:08

---

Diagnostics Check	OK	18:30	
Air Blank	0.000	18:30	
Control Test	0.080	18:30	
Air Blank	0.000	18:31	
Subject Sample	# 1	18:31	Subject Test Refused
Breath Volume	0.000	18:31	
Air Blank	0.000	18:32	
Control Test	0.080	18:32	
Air Blank	0.000	18:33	
Diagnostics Check	OK	18:33	

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: RUSSO,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1355293  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: EASTON

DATE/TIME OF TEST: 04/05/2022 05:07:49

---

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.080	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.165	05:03
Breath Volume	2.132	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:05
Subject Sample # 2	0.167	05:06
Breath Volume	2.179	05:06
Air Blank	0.000	05:06
Control Test	0.077	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1355293  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: KAMILOVIC

DATE/TIME OF TEST: 04/06/2022 05:46:47

---

Diagnostics Check		OK	05:40
Air Blank		0.000	05:40
Control Test		0.079	05:41
Air Blank		0.000	05:41
Subject Sample	# 1	0.000	05:42
Breath Volume		2.460	05:42
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.000	05:45
Breath Volume		1.691	05:45
Air Blank		0.000	05:45
Control Test		0.079	05:46
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KREITENSTEIN, JONATHAN, A  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: KREITENSTEIN, JONATHAN  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1355293  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: VERNON

DATE/TIME OF TEST: 04/07/2022 03:44:36

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.232	03:39
Breath Volume	2.074	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.237	03:42
Breath Volume	1.566	03:42
Air Blank	0.000	03:43
Control Test	0.077	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: RUSSO,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 04/08/2022 21:12:31

---

Diagnostics Check	OK	21:02
Air Blank	0.000	21:02
Control Test	0.080	21:03
Air Blank	0.000	21:03
Subject Sample # 1	0.103	21:05
Breath Volume	1.613	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.107	21:10
Breath Volume	1.652	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

---

Data Download Date/Time: 04/27/2022 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 04/08/2022 23:54:36

---

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.080	23:52	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:53	
Control Test	0.080	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: RUSSO,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 04/09/2022 01:09:32

---

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.157	01:04
Breath Volume		1.398	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.157	01:07
Breath Volume		1.273	01:07
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

---

Data Download Date/Time: 04/27/2022 15:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 04/09/2022 05:14:28

---

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:11	
Control Test	0.081	05:11	
Air Blank	0.000	05:12	
Subject Sample	# 1	05:13	Subject Test Refused
Breath Volume	0.000	05:13	
Air Blank	0.000	05:13	
Control Test	0.081	05:13	
Air Blank	0.000	05:14	
Diagnostics Check	OK	05:14	

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,MARIO,  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: RODRIGUEZ,MARIO  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: RATCLIFF

DATE/TIME OF TEST: 04/09/2022 23:33:13

---

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.081	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.131	23:28
Breath Volume	1.425	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.133	23:31
Breath Volume	1.730	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

---

Data Download Date/Time: 04/27/2022 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOWER,CHEYENNE,M  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: MEYERS,ERIC  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 04/10/2022 00:46:36

---

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.081	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.215	00:41
Breath Volume		1.144	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:43
Subject Sample	# 2	0.233	00:44
Breath Volume		2.671	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FENTON,CODY,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: FENTON,CODY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: METOYER

DATE/TIME OF TEST: 04/10/2022 02:17:21

---

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.219	02:12
Breath Volume		1.375	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:14
Subject Sample	# 2	0.200	02:15
Breath Volume		1.867	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILASENOR,DAVID,L  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: CESPEDES,ROXANA  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: MARSAC

DATE/TIME OF TEST: 04/10/2022 03:17:28

---

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:05	
Control Test		0.081	03:06	
Air Blank		0.000	03:06	
Subject Sample	# 1	0.133	03:09	Volume Not Met
Breath Volume		0.070	03:09	
Air Blank		0.000	03:10	
Air Blank		0.000	03:12	
Subject Sample	# 2	0.191	03:15	Volume Not Met
Breath Volume		0.308	03:15	
Air Blank		0.000	03:16	
Control Test		0.079	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

---

Data Download Date/Time: 04/27/2022 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLASENOR,DAVID,L  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: CESPEDES,ROXANA  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: MARSAC

DATE/TIME OF TEST: 04/10/2022 03:29:42

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.078	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.212	03:24
Breath Volume	2.488	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:26
Subject Sample # 2	0.223	03:27
Breath Volume	1.250	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THEIS,ZACHARY,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: THEIS,ZACHARY  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1355293  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 14:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: STAVOLA

DATE/TIME OF TEST: 04/10/2022 15:28:19

---

Diagnostics Check	OK	15:18	
Air Blank	0.000	15:18	
Control Test	0.081	15:18	
Air Blank	0.000	15:19	
Subject Sample	# 1 0.142	15:22	Volume Not Met
Breath Volume	1.011	15:22	
Air Blank	0.000	15:23	
Air Blank	0.000	15:24	
Subject Sample	# 2	15:26	Subject Test Refused
Breath Volume	0.640	15:26	
Air Blank	0.000	15:27	
Control Test	0.078	15:27	
Air Blank	0.000	15:28	
Diagnostics Check	OK	15:28	

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SOMMERS

DATE/TIME OF TEST: 04/11/2022 01:23:55

---

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.081	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.107	01:17
Breath Volume	2.085	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.108	01:22
Breath Volume	1.371	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

---

Data Download Date/Time: 04/27/2022 15:48



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,DILLON,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BROWN,DILLON  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: HEAD

DATE/TIME OF TEST: 04/12/2022 00:42:14

---

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.081	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:35	No Sample Provided
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Air Blank	0.000	00:37	
Subject Sample	# 2	00:40	No Sample Provided
Breath Volume	0.000	00:40	
Air Blank	0.000	00:41	
Control Test	0.081	00:41	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: GALLEGOS VELASCO

DATE/TIME OF TEST: 04/15/2022 02:28:53

---

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.172	02:24
Breath Volume	2.042	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.164	02:27
Breath Volume	2.667	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

---

Data Download Date/Time: 04/27/2022 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELLS,JOSIE,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WELLS,JOSIE  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: HENDRICKSON

DATE/TIME OF TEST: 04/15/2022 22:59:10

---

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.080	22:56	
Air Blank	0.000	22:57	
Subject Sample	# 1	22:57	Subject Test Refused
Breath Volume	0.000	22:57	
Air Blank	0.000	22:58	
Control Test	0.080	22:58	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

---

Data Download Date/Time: 04/27/2022 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,DILLON,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BROWN,DILLON  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 04/16/2022 22:59:55

---

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.216	22:55
Breath Volume	2.515	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.220	22:58
Breath Volume	1.890	22:58
Air Blank	0.000	22:58
Control Test	0.077	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LISS,CODY,S  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: LISS,CODY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1355293  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MILLAN

DATE/TIME OF TEST: 04/17/2022 01:15:09

---

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.080	01:04	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:08	No Sample Provided
Breath Volume	0.000	01:08	
Air Blank	0.000	01:08	
Air Blank	0.000	01:10	
Subject Sample	# 2	01:13	No Sample Provided
Breath Volume	0.000	01:13	
Air Blank	0.000	01:14	
Control Test	0.080	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

---

Data Download Date/Time: 04/27/2022 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARIAS,W,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: ARIAS,W  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: HEHIR

DATE/TIME OF TEST: 04/17/2022 16:32:47

---

Diagnostics Check	OK	16:23
Air Blank	0.000	16:24
Control Test	0.080	16:24
Air Blank	0.000	16:25
Subject Sample # 1	0.099	16:25
Breath Volume	1.265	16:25
Air Blank	0.000	16:26
Air Blank	0.000	16:27
Subject Sample # 2	0.092	16:30
Breath Volume	1.636	16:30
Air Blank	0.000	16:31
Control Test	0.079	16:32
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

---

Data Download Date/Time: 04/27/2022 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,DILLON,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BROWN,DILLON  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/17/2022 20:02:30

---

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:51	
Control Test	0.080	19:51	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:55	No Sample Provided
Breath Volume	0.000	19:55	
Air Blank	0.000	19:55	
Air Blank	0.000	19:57	
Subject Sample	# 2	20:01	No Sample Provided
Breath Volume	0.000	20:01	
Air Blank	0.000	20:01	
Control Test	0.079	20:01	
Air Blank	0.000	20:02	
Diagnostics Check	OK	20:02	

---

Data Download Date/Time: 04/27/2022 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: CONNORS

DATE/TIME OF TEST: 04/20/2022 03:01:44

---

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.130	02:56
Breath Volume	2.355	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.124	03:00
Breath Volume	1.269	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

---

Data Download Date/Time: 04/27/2022 15:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: CESPEDES,R  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: SANTIAGO CARRERA

DATE/TIME OF TEST: 04/20/2022 04:23:55

---

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:12	
Control Test		0.080	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.123	04:16	Volume Not Met
Breath Volume		0.160	04:16	
Air Blank		0.000	04:17	
Air Blank		0.000	04:18	
Subject Sample	# 2	0.160	04:22	Volume Not Met
Breath Volume		0.019	04:22	
Air Blank		0.000	04:22	
Control Test		0.079	04:23	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: CESPEDES,R  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: SANTIAGO CARRERA

DATE/TIME OF TEST: 04/20/2022 04:38:55

---

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:27	
Control Test		0.079	04:27	
Air Blank		0.000	04:28	
Subject Sample	# 1	0.153	04:31	Volume Not Met
Breath Volume		0.019	04:31	
Air Blank		0.000	04:32	
Air Blank		0.000	04:34	
Subject Sample	# 2	0.000	04:37	Volume Not Met
Breath Volume		0.019	04:37	
Air Blank		0.000	04:37	
Control Test		0.077	04:38	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:38	

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EASLEY,M,A  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: PATTERSON,H  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 04/21/2022 03:44:22

---

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.167	03:39
Breath Volume		1.707	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.169	03:42
Breath Volume		1.761	03:42
Air Blank		0.000	03:43
Control Test		0.078	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

---

Data Download Date/Time: 04/27/2022 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 04/23/2022 05:55:12

---

Diagnostics Check	OK	05:48
Air Blank	0.000	05:49
Control Test	0.081	05:49
Air Blank	0.000	05:50
Subject Sample # 1	0.082	05:50
Breath Volume	2.722	05:50
Air Blank	0.000	05:51
Air Blank	0.000	05:53
Subject Sample # 2	0.077	05:53
Breath Volume	2.320	05:53
Air Blank	0.000	05:54
Control Test	0.080	05:54
Air Blank	0.000	05:55
Diagnostics Check	OK	05:55

---

Data Download Date/Time: 04/27/2022 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNDORFF,TERRY,D  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MUNDORFF,TERRY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MASSARO

DATE/TIME OF TEST: 04/23/2022 21:02:54

---

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.081	20:56
Air Blank	0.000	20:56
Subject Sample # 1	0.124	20:57
Breath Volume	1.871	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.120	21:01
Breath Volume	1.585	21:01
Air Blank	0.000	21:01
Control Test	0.079	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

---

Data Download Date/Time: 04/27/2022 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWER,DALTON,J  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: OWER,DALTON  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: LUCK-BAHRUTH

DATE/TIME OF TEST: 04/26/2022 23:55:21

---

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.190	23:50
Breath Volume	2.101	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:52
Subject Sample # 2	0.184	23:53
Breath Volume	1.460	23:53
Air Blank	0.000	23:54
Control Test	0.077	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

---

Data Download Date/Time: 04/27/2022 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PENA,R,S  
BREATH TEST OPERATOR AGENCY: Groveland PD  
ARRESTING OFFICER: PENA,R  
ARRESTING OFFICER AGENCY: Groveland PD  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 10:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:05

SUBJECT'S LAST NAME: DARBY

DATE/TIME OF TEST: 04/12/2022 11:37:15

---

Diagnostics Check	OK	11:25	
Air Blank	0.000	11:26	
Control Test	0.079	11:26	
Air Blank	0.000	11:27	
Subject Sample	# 1 0.206	11:30	Volume Not Met
Breath Volume	0.316	11:30	
Air Blank	0.000	11:30	
Air Blank	0.000	11:32	
Subject Sample	# 2	11:35	No Sample Provided
Breath Volume	0.000	11:35	
Air Blank	0.000	11:36	
Control Test	0.078	11:36	
Air Blank	0.000	11:37	
Diagnostics Check	OK	11:37	

---

Data Download Date/Time: 04/18/2022 09:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PENA,R,S  
BREATH TEST OPERATOR AGENCY: Groveland PD  
ARRESTING OFFICER: PENA,R  
ARRESTING OFFICER AGENCY: Groveland PD  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: BACAB CORTES

DATE/TIME OF TEST: 04/16/2022 23:36:58

---

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.080	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:35	No Sample Provided
Breath Volume	0.000	23:35	
Air Blank		23:35	RFI Detect
Air Blank	0.000	23:36	
Control Test	0.080	23:36	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

---

Data Download Date/Time: 04/18/2022 09:22



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PENA,R,S  
BREATH TEST OPERATOR AGENCY: Groveland PD  
ARRESTING OFFICER: PENA,R  
ARRESTING OFFICER AGENCY: Groveland PD  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: BACAB CORTES

DATE/TIME OF TEST: 04/16/2022 23:50:04

---

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.080	23:39	
Air Blank	0.000	23:39	
Subject Sample	# 1	23:43	No Sample Provided
Breath Volume	0.000	23:43	
Air Blank	0.000	23:43	
Air Blank	0.000	23:45	
Subject Sample	# 2	23:48	No Sample Provided
Breath Volume	0.000	23:48	
Air Blank	0.000	23:49	
Control Test	0.079	23:49	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:50	

---

Data Download Date/Time: 04/18/2022 09:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMAHON,KENNETH,J  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: MCMAHON,KENNETH  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: HILLSON

DATE/TIME OF TEST: 04/08/2022 20:47:48

---

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.082	20:42
Air Blank	0.000	20:42
Subject Sample # 1	0.162	20:43
Breath Volume	2.332	20:43
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.166	20:46
Breath Volume	1.343	20:46
Air Blank	0.000	20:46
Control Test	0.081	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

---

Data Download Date/Time: 05/18/2022 10:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMAHON,KENNETH,J  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: PINO,RICHARD  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: HOLDING

DATE/TIME OF TEST: 04/09/2022 03:59:47

---

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.082	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.131	03:55
Breath Volume	3.160	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.127	03:58
Breath Volume	1.808	03:58
Air Blank	0.000	03:58
Control Test	0.082	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

---

Data Download Date/Time: 05/18/2022 10:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOMASEGOVICH,KODY,D  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: SEGREE,SCOUT  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 04/13/2022 21:52:56

---

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.082	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.000	21:48
Breath Volume	2.507	21:48
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.000	21:51
Breath Volume	1.601	21:51
Air Blank	0.000	21:52
Control Test	0.082	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

---

Data Download Date/Time: 05/18/2022 10:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MLINAC,MADISON,A  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: MASSEY,SIDNEY  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: KEEFE-UNDERWOOD

DATE/TIME OF TEST: 04/20/2022 18:43:39

---

Diagnostics Check	OK	18:37
Air Blank	0.000	18:38
Control Test	0.082	18:38
Air Blank	0.000	18:38
Subject Sample # 1	0.000	18:39
Breath Volume	1.507	18:39
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.000	18:42
Breath Volume	2.230	18:42
Air Blank	0.000	18:42
Control Test	0.082	18:43
Air Blank	0.000	18:43
Diagnostics Check	OK	18:43

---

Data Download Date/Time: 05/18/2022 10:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MLINAC,MADISON,A  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: COTNER,STEVEN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/21/2022 04:18:34

---

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:06	
Control Test		0.082	04:07	
Air Blank		0.000	04:07	
Subject Sample	# 1	0.167	04:10	Volume Not Met
Breath Volume		1.195	04:10	
Air Blank		0.000	04:11	
Air Blank		0.000	04:13	
Subject Sample	# 2	0.184	04:13	
Breath Volume		1.910	04:13	
Air Blank		0.000	04:14	
Air Blank		0.000	04:16	
Subject Sample	# 3	0.187	04:17	
Breath Volume		3.257	04:17	
Air Blank		0.000	04:17	
Control Test		0.080	04:17	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

---

Data Download Date/Time: 05/18/2022 10:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMAHON,KENNETH,J  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: KENT,ROBERT  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: FAGOT

DATE/TIME OF TEST: 04/24/2022 03:44:28

---

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.082	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.142	03:40
Breath Volume	1.726	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.139	03:42
Breath Volume	1.105	03:42
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time: 05/18/2022 10:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMAHON,KENNETH,J  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: PINO,RICHARD  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: OVERLADE

DATE/TIME OF TEST: 04/24/2022 04:03:22

---

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.097	03:58
Breath Volume	1.917	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.097	04:01
Breath Volume	2.910	04:01
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

---

Data Download Date/Time: 05/18/2022 10:59



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GENTRY, BENJAMIN, W  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: DONOVAN, M.  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: ALUISIO

DATE/TIME OF TEST: 04/25/2022 16:56:48

---

Diagnostics Check	OK	16:49
Air Blank	0.000	16:49
Control Test	0.080	16:50
Air Blank	0.000	16:50
Subject Sample # 1	0.347	16:51
Breath Volume	1.535	16:51
Air Blank	0.000	16:51
Air Blank	0.000	16:53
Subject Sample # 2	0.334	16:55
Breath Volume	2.148	16:55
Air Blank	0.000	16:55
Control Test	0.079	16:56
Air Blank	0.000	16:56
Diagnostics Check	OK	16:56

---

Data Download Date/Time: 05/18/2022 10:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DUFOUR,NOAH,I

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: DUFOUR,NOAH

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 20:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: GONZALEZ GARCIA

DATE/TIME OF TEST: 04/25/2022 21:23:19

---

Diagnostics Check		OK	21:16
Air Blank		0.000	21:16
Control Test		0.081	21:17
Air Blank		0.000	21:17
Subject Sample	# 1	0.224	21:18
Breath Volume		2.500	21:18
Air Blank		0.000	21:19
Air Blank		0.000	21:21
Subject Sample	# 2	0.217	21:21
Breath Volume		1.863	21:21
Air Blank		0.000	21:22
Control Test		0.077	21:22
Air Blank		0.000	21:23
Diagnostics Check		OK	21:23

---

Data Download Date/Time: 05/18/2022 10:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORALES,KARISSA,C  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: MORALES,KARISSA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 04/25/2022 21:54:54

---

Diagnostics Check		OK	21:43	
Air Blank		0.000	21:44	
Control Test		0.081	21:44	
Air Blank		0.000	21:45	
Subject Sample	# 1	0.047	21:48	Volume Not Met
Breath Volume		0.460	21:48	
Air Blank		0.000	21:49	
Air Blank		0.000	21:50	
Subject Sample	# 2	0.046	21:54	Volume Not Met
Breath Volume		0.183	21:54	
Air Blank			21:54	Improper Sample
Air Blank		0.000	21:54	

---

Data Download Date/Time: 05/18/2022 10:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: ATWELL

DATE/TIME OF TEST: 04/28/2022 22:40:53

---

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1 0.179	22:29	
Breath Volume	1.464	22:29	
Air Blank	0.000	22:30	
Air Blank	0.000	22:32	
Subject Sample	# 2 0.143	22:35	Volume Not Met
Breath Volume	0.476	22:35	
Air Blank	0.000	22:36	
Air Blank	0.000	22:37	
Subject Sample	# 3 0.162	22:39	
Breath Volume	1.273	22:39	
Air Blank	0.000	22:39	
Control Test	0.079	22:40	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

---

Data Download Date/Time: 05/18/2022 10:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MLINAC,MADISON,A  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: MLINAC,MADISON  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TINSLEY

DATE/TIME OF TEST: 04/30/2022 02:56:54

---

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.082	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.165	02:52
Breath Volume		4.652	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.165	02:55
Breath Volume		4.355	02:55
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

---

Data Download Date/Time: 05/18/2022 10:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORREA,JONATHAN,E  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: CORREA,JONATHAN  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02021080A1  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: DORSETT

DATE/TIME OF TEST: 04/09/2022 02:01:41

---

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.110	01:57
Breath Volume		1.335	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.109	02:00
Breath Volume		1.910	02:00
Air Blank		0.000	02:00
Control Test		0.080	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

---

Data Download Date/Time: 04/29/2022 18:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: TUSHA,JYLTON  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02021080A1  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: JEUNELOT

DATE/TIME OF TEST: 04/10/2022 19:16:44

---

Diagnostics Check		OK	19:10
Air Blank		0.000	19:10
Control Test		0.082	19:11
Air Blank		0.000	19:11
Subject Sample	# 1	0.236	19:12
Breath Volume		4.898	19:12
Air Blank		0.000	19:12
Air Blank		0.000	19:14
Subject Sample	# 2	0.230	19:15
Breath Volume		4.019	19:15
Air Blank		0.000	19:15
Control Test		0.080	19:16
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

---

Data Download Date/Time: 04/29/2022 18:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,SCOTT,T  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: EARLY,JACQUELINE  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02021080A1  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 04/14/2022 01:55:08

---

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.155	01:50
Breath Volume	2.117	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:52
Subject Sample # 2	0.154	01:53
Breath Volume	1.808	01:53
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

---

Data Download Date/Time: 04/29/2022 18:20



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,SCOTT,B  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: CARTER,JARRED  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02021080A1  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 05:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: ARGUETA

DATE/TIME OF TEST: 04/14/2022 06:34:06

---

Diagnostics Check	OK	06:27
Air Blank	0.000	06:27
Control Test	0.081	06:28
Air Blank	0.000	06:28
Subject Sample # 1	0.145	06:29
Breath Volume	1.800	06:29
Air Blank	0.000	06:30
Air Blank	0.000	06:32
Subject Sample # 2	0.132	06:32
Breath Volume	1.414	06:32
Air Blank	0.000	06:33
Control Test	0.077	06:33
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

---

Data Download Date/Time: 04/29/2022 18:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORREA,JONATHAN,E  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: CORREA,JONATHAN  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02021080A1  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MALINIS

DATE/TIME OF TEST: 04/16/2022 01:55:00

---

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.081	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.132	01:50
Breath Volume	1.148	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.142	01:53
Breath Volume	1.152	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:55

---

Data Download Date/Time: 04/29/2022 18:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHIRSHIKOVA,ALINA,A  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: LEAR,TREY  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02021080A1  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MCANALLEN

DATE/TIME OF TEST: 04/16/2022 21:35:13

---

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.080	21:24
Air Blank	0.000	21:24
Subject Sample # 1	0.369	21:25
Breath Volume	2.488	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:27
Subject Sample # 2	0.328	21:28
Breath Volume	1.390	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 3	0.332	21:33
Breath Volume	1.507	21:33
Air Blank	0.000	21:34
Control Test	0.077	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

---

Data Download Date/Time: 04/29/2022 18:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: JAROUSSE,DANIELLE  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02021080A1  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BREWINGTON

DATE/TIME OF TEST: 04/17/2022 01:27:48

---

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.080	01:21	
Air Blank	0.000	01:22	
Subject Sample # 1	0.203	01:22	
Breath Volume	2.082	01:22	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample # 2	0.200	01:25	
Breath Volume	1.425	01:25	
Air Blank	0.000	01:26	
Control Test	0.071	01:27	Control Outside Tolerance
Air Blank	0.000	01:27	

---

Data Download Date/Time: 04/29/2022 18:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: JAROUSSE,DANIELLE  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02021080A1  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: BRUCH

DATE/TIME OF TEST: 04/24/2022 05:52:09

---

Diagnostics Check	OK	05:45
Air Blank	0.000	05:46
Control Test	0.081	05:46
Air Blank	0.000	05:47
Subject Sample # 1	0.118	05:47
Breath Volume	3.644	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.117	05:50
Breath Volume	2.125	05:50
Air Blank	0.000	05:51
Control Test	0.081	05:51
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

---

Data Download Date/Time: 04/29/2022 18:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDDY,D,  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: RAYA,Z  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1346773  
DRY GAS EXPIRY DATE: 07/08/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:51

SUBJECT'S LAST NAME: ALAMILLA

DATE/TIME OF TEST: 04/09/2022 12:35:03

---

Diagnostics Check		OK	12:23	
Air Blank		0.000	12:23	
Control Test		0.079	12:24	
Air Blank		0.000	12:24	
Subject Sample	# 1	0.142	12:27	Volume Not Met
Breath Volume		0.066	12:27	
Air Blank		0.000	12:28	
Air Blank		0.000	12:30	
Subject Sample	# 2	0.140	12:30	
Breath Volume		1.750	12:30	
Air Blank		0.000	12:31	
Air Blank		0.000	12:33	
Subject Sample	# 3	0.141	12:33	
Breath Volume		1.453	12:33	
Air Blank		0.000	12:34	
Control Test		0.078	12:34	
Air Blank		0.000	12:34	
Diagnostics Check		OK	12:35	

---

Data Download Date/Time: 04/26/2022 11:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORRESTER, TRAVIS, D  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: SANTOS, PAUL  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1346773  
DRY GAS EXPIRY DATE: 07/08/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MCNEIR

DATE/TIME OF TEST: 04/11/2022 03:41:41

---

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.077	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:34	No Sample Provided
Breath Volume	0.000	03:34	
Air Blank	0.000	03:35	
Air Blank	0.000	03:37	
Subject Sample	# 2	03:40	No Sample Provided
Breath Volume	0.000	03:40	
Air Blank	0.000	03:40	
Control Test	0.079	03:41	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

---

Data Download Date/Time: 04/26/2022 11:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORRESTER, TRAVIS, D  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: SCHULTZ, JEREMY  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1346773  
DRY GAS EXPIRY DATE: 07/08/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: TUFO

DATE/TIME OF TEST: 04/29/2022 04:27:50

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.173	04:22
Breath Volume	2.351	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.175	04:26
Breath Volume	2.894	04:26
Air Blank	0.000	04:26
Control Test	0.078	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

---

Data Download Date/Time: 06/04/2022 08:33



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: BARRERA,ALEXANDER  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 04/02/2022 01:29:05

---

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.079	01:26	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Control Test	0.079	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

---

Data Download Date/Time: 04/04/2022 17:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: BARRERA,ALEXANDER  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/03/2022 03:26:46

---

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.140	03:21
Breath Volume	2.121	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.142	03:25
Breath Volume	2.949	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

---

Data Download Date/Time: 04/04/2022 17:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: VAUGHN,T  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: PLUMLEY

DATE/TIME OF TEST: 04/03/2022 04:57:31

---

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.217	04:52
Breath Volume	1.679	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.216	04:55
Breath Volume	2.050	04:55
Air Blank	0.000	04:56
Control Test	0.079	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

---

Data Download Date/Time: 04/04/2022 17:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHROCK,L,B  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HIGGINS,G  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 06:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: CARRANZA CANTILLANO

DATE/TIME OF TEST: 04/03/2022 08:11:53

---

Diagnostics Check	OK	08:03
Air Blank	0.000	08:04
Control Test	0.079	08:04
Air Blank	0.000	08:05
Subject Sample # 1	0.155	08:06
Breath Volume	1.546	08:06
Air Blank	0.000	08:06
Air Blank	0.000	08:08
Subject Sample # 2	0.151	08:10
Breath Volume	1.273	08:10
Air Blank	0.000	08:10
Control Test	0.076	08:11
Air Blank	0.000	08:11
Diagnostics Check	OK	08:11

---

Data Download Date/Time: 04/04/2022 17:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAYWARD,KYLE,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: KWAK,PETER  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: ORAVETZ

DATE/TIME OF TEST: 04/03/2022 16:54:19

---

Diagnostics Check	OK	16:48
Air Blank	0.000	16:48
Control Test	0.079	16:49
Air Blank	0.000	16:49
Subject Sample # 1	0.162	16:49
Breath Volume	2.300	16:49
Air Blank	0.000	16:50
Air Blank	0.000	16:52
Subject Sample # 2	0.159	16:52
Breath Volume	1.910	16:52
Air Blank	0.000	16:53
Control Test	0.077	16:53
Air Blank	0.000	16:54
Diagnostics Check	OK	16:54

---

Data Download Date/Time: 04/04/2022 17:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PUTTICK,EMILY,E  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: ALMAGUER,BRANDON  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: BORGES

DATE/TIME OF TEST: 04/04/2022 01:05:53

---

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:02	
Control Test	0.079	01:02	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:04	
Control Test	0.080	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

---

Data Download Date/Time: 04/04/2022 17:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREEN,J,A  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: GREEN,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 04/07/2022 01:29:23

---

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.085	01:24
Breath Volume	1.859	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.078	01:27
Breath Volume	1.855	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

---

Data Download Date/Time: 04/28/2022 22:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,SCOTT,T  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MANDILE,JEREMY  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 04/14/2022 02:48:09

---

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.078	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1 0.148	02:42	
Breath Volume	1.773	02:42	
Air Blank	0.000	02:43	
Air Blank	0.000	02:45	
Subject Sample	# 2 0.149	02:46	
Breath Volume	1.769	02:46	
Air Blank	0.000	02:46	
Control Test	0.079	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	Fail	02:47	Internal Printer Test
Air Blank	0.000	02:48	

---

Data Download Date/Time: 04/28/2022 22:44



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREEN,J,A  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: MANDILE,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 04/14/2022 02:57:39

---

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.145	02:52
Breath Volume	1.632	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.145	02:56
Breath Volume	1.195	02:56
Air Blank	0.000	02:56
Control Test	0.077	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

---

Data Download Date/Time: 04/28/2022 22:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREEN,J,A  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: GREEN,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BIANCIELLA  
DATE/TIME OF TEST: 04/15/2022 22:06:29

---

Diagnostics Check	Fail	22:06	Internal Printer Test
Air Blank	0.000	22:06	

---

Data Download Date/Time: 04/28/2022 22:45





# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREEN,J,A  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: GREEN,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BIANCIELLA

DATE/TIME OF TEST: 04/15/2022 22:32:26

---

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.146	22:27
Breath Volume	1.511	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.149	22:30
Breath Volume	1.476	22:30
Air Blank	0.000	22:31
Control Test	0.078	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 04/28/2022 22:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREEN,JASON,  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: ELLIS,ALEX  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: OLIVERA

DATE/TIME OF TEST: 04/16/2022 02:33:55

---

Diagnostics Check	Fail	02:33	Internal Printer Test
Air Blank	0.000	02:33	

---

Data Download Date/Time: 04/28/2022 22:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREEN,J,A  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: GREEN,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: PAIVA

DATE/TIME OF TEST: 04/29/2022 01:11:29

---

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.208	01:06
Breath Volume		1.664	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.203	01:09
Breath Volume		1.550	01:09
Air Blank		0.000	01:10
Control Test		0.081	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

---

Data Download Date/Time: 06/08/2022 16:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GAZMURI,KEYSER  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 04/03/2022 18:58:04

---

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.082	18:52
Air Blank	0.000	18:53
Subject Sample # 1	0.000	18:53
Breath Volume	2.203	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.000	18:56
Breath Volume	2.031	18:56
Air Blank	0.000	18:57
Control Test	0.082	18:57
Air Blank	0.000	18:57
Diagnostics Check	OK	18:58

---

Data Download Date/Time: 04/13/2022 10:38



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GUERRA,A  
ARRESTING OFFICER AGENCY: Hallandale Beach PD  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: HOLDA

DATE/TIME OF TEST: 04/04/2022 03:11:03

---

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.082	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.181	03:06
Breath Volume	1.281	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:08
Subject Sample # 2	0.185	03:09
Breath Volume	1.343	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:11

---

Data Download Date/Time: 04/13/2022 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: RIVERA RIOS

DATE/TIME OF TEST: 04/09/2022 02:06:31

---

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:52	
Control Test	0.082	01:52	
Air Blank	0.000	01:53	
Subject Sample # 1	0.165	01:55	
Breath Volume	1.382	01:55	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample # 2	0.137	02:00	Slope Not Level
Breath Volume	1.332	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:03	
Subject Sample # 3	0.151	02:04	
Breath Volume	1.359	02:04	
Air Blank	0.000	02:05	
Control Test	0.078	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

---

Data Download Date/Time: 04/13/2022 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: POPICK,SCOTT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BALTAZAR GONZALEZ

DATE/TIME OF TEST: 04/09/2022 04:39:05

---

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.162	04:34
Breath Volume	2.210	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.163	04:37
Breath Volume	1.781	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

---

Data Download Date/Time: 04/13/2022 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: EKWEREKWU,CHUKWUEMEK  
A  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Male

OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 04/09/2022 05:42:19

---

Diagnostics Check	OK	05:35
Air Blank	0.000	05:36
Control Test	0.082	05:36
Air Blank	0.000	05:37
Subject Sample # 1	0.135	05:37
Breath Volume	2.433	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.130	05:40
Breath Volume	2.039	05:40
Air Blank	0.000	05:41
Control Test	0.081	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

---

Data Download Date/Time: 04/13/2022 10:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HALES,ANDREW  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/09/2022 21:07:17

---

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.078	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.194	21:02
Breath Volume	1.890	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.185	21:05
Breath Volume	1.421	21:05
Air Blank	0.000	21:06
Control Test	0.078	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

---

Data Download Date/Time: 04/13/2022 10:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ARMSTRONG,BRITTANY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: LOPEZ VELASQUEZ

DATE/TIME OF TEST: 04/09/2022 23:38:46

---

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.082	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.177	23:34
Breath Volume	2.285	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.174	23:37
Breath Volume	1.609	23:37
Air Blank	0.000	23:37
Control Test	0.078	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

---

Data Download Date/Time: 04/13/2022 10:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GAZMURI,KEYSER  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: DONALDSON-MOSS

DATE/TIME OF TEST: 04/10/2022 00:06:53

---

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:54	
Control Test		0.081	23:54	
Air Blank		0.000	23:54	
Subject Sample	# 3	0.115	23:57	Volume Not Met
Breath Volume		0.589	23:57	
Air Blank		0.000	23:58	
Air Blank		0.000	00:00	
Subject Sample	# 1	0.109	00:02	
Breath Volume		1.152	00:02	
Air Blank		0.000	00:02	
Air Blank		0.000	00:04	
Subject Sample	# 2	0.123	00:05	
Breath Volume		1.246	00:05	
Air Blank		0.000	00:05	
Control Test		0.080	00:06	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

---

Data Download Date/Time: 04/13/2022 10:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA,NIVIA,M  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: GASCA,RAQUEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1237851  
DRY GAS EXPIRY DATE: 11/21/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: SMILICH

DATE/TIME OF TEST: 04/01/2022 16:26:45

---

Diagnostics Check	OK	16:17	
Air Blank	0.000	16:17	
Control Test	0.082	16:17	
Air Blank	0.000	16:18	
Subject Sample	# 1 0.203	16:21	Volume Not Met
Breath Volume	0.210	16:21	
Air Blank	0.000	16:22	
Air Blank	0.000	16:24	
Subject Sample	# 2	16:24	Subject Test Refused
Breath Volume	0.191	16:24	
Air Blank	0.000	16:25	
Control Test	0.078	16:26	
Air Blank	0.000	16:26	
Diagnostics Check	OK	16:26	

---

Data Download Date/Time: 04/05/2022 10:25



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROCA,LUIS,C  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: ROCA,LUIS  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1237851  
DRY GAS EXPIRY DATE: 11/21/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 04/11/2022 14:03:50

---

Diagnostics Check	OK	13:56
Air Blank	0.000	13:57
Control Test	0.082	13:57
Air Blank	0.000	13:58
Subject Sample # 1	0.175	13:58
Breath Volume	2.250	13:58
Air Blank	0.000	13:59
Air Blank	0.000	14:01
Subject Sample # 2	0.178	14:02
Breath Volume	2.218	14:02
Air Blank	0.000	14:02
Control Test	0.081	14:03
Air Blank	0.000	14:03
Diagnostics Check	OK	14:03

---

Data Download Date/Time: 05/30/2022 09:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: INNOCENT,L,  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: RAMIREZ,W  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1237851  
DRY GAS EXPIRY DATE: 11/21/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BARTOLON

DATE/TIME OF TEST: 04/15/2022 22:55:42

---

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:53	
Control Test	0.082	22:53	
Air Blank	0.000	22:54	
Subject Sample	# 1	22:54	Subject Test Refused
Breath Volume	0.000	22:54	
Air Blank	0.000	22:54	
Control Test	0.081	22:55	
Air Blank	0.000	22:55	
Diagnostics Check	OK	22:55	

---

Data Download Date/Time: 05/30/2022 09:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SINCORE,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SINCORE,MICHAEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1237851  
DRY GAS EXPIRY DATE: 11/21/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: GARCIAHERNANDEZ

DATE/TIME OF TEST: 04/17/2022 18:12:40

---

Diagnostics Check	OK	18:05
Air Blank	0.000	18:06
Control Test	0.082	18:06
Air Blank	0.000	18:07
Subject Sample # 1	0.192	18:07
Breath Volume	2.523	18:07
Air Blank	0.000	18:08
Air Blank	0.000	18:10
Subject Sample # 2	0.188	18:10
Breath Volume	2.300	18:10
Air Blank	0.000	18:11
Control Test	0.081	18:12
Air Blank	0.000	18:12
Diagnostics Check	OK	18:12

---

Data Download Date/Time: 05/30/2022 09:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILPIN,SCHELISA,  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: GILPIN,SCHELISA  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1237851  
DRY GAS EXPIRY DATE: 11/21/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 04/21/2022 19:07:44

---

Diagnostics Check	OK	18:54	
Air Blank	0.000	18:54	
Control Test	0.082	18:55	
Air Blank	0.000	18:55	
Subject Sample	# 1 0.244	18:57	
Breath Volume	1.800	18:57	
Air Blank	0.000	18:58	
Air Blank	0.000	18:59	
Subject Sample	# 2 0.239	19:02	Volume Not Met
Breath Volume	0.363	19:02	
Air Blank	0.000	19:04	
Air Blank	0.000	19:05	
Subject Sample	# 3 0.243	19:05	
Breath Volume	1.800	19:05	
Air Blank	0.000	19:06	
Control Test	0.077	19:07	
Air Blank	0.000	19:07	
Diagnostics Check	OK	19:07	

---

Data Download Date/Time: 05/30/2022 09:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILPIN,SCHELISA,  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: GILPIN,SCHELISA  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1237851  
DRY GAS EXPIRY DATE: 11/21/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: AMESQUITA

DATE/TIME OF TEST: 04/23/2022 21:17:39

---

Diagnostics Check	OK	21:14	
Air Blank	0.000	21:15	
Control Test	0.081	21:15	
Air Blank	0.000	21:16	
Subject Sample	# 1	21:16	Subject Test Refused
Breath Volume	0.000	21:16	
Air Blank	0.000	21:16	
Control Test	0.081	21:17	
Air Blank	0.000	21:17	
Diagnostics Check	OK	21:17	

---

Data Download Date/Time: 05/30/2022 09:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: ACKERMAN,ELVIS  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1237851  
DRY GAS EXPIRY DATE: 11/21/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: AGUILAR GUZMAN

DATE/TIME OF TEST: 04/28/2022 21:41:07

---

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.081	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.216	21:36
Breath Volume	2.339	21:36
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.218	21:39
Breath Volume	2.628	21:39
Air Blank	0.000	21:40
Control Test	0.080	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

---

Data Download Date/Time: 05/30/2022 09:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,RAUL,  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: PEREZ,RAUL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1237851  
DRY GAS EXPIRY DATE: 11/21/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: JOSE PASCUAL

DATE/TIME OF TEST: 04/29/2022 00:44:00

---

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:35	
Control Test	0.081	00:36	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:39	No Sample Provided
Breath Volume	0.000	00:39	
Air Blank	0.000	00:40	
Air Blank	0.000	00:42	
Subject Sample	# 2	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:43	
Control Test	0.082	00:43	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:44	

---

Data Download Date/Time: 05/30/2022 09:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROCA,LUIS,C  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: ROCA,LUIS  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1237851  
DRY GAS EXPIRY DATE: 11/21/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/30/2022 23:23:59

---

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.082	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:23	
Control Test	0.082	23:23	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

---

Data Download Date/Time: 05/30/2022 09:45



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/15/2022 05:09:38

---

Diagnostics Check	OK	05:09	
Air Blank		05:09	RFI Detect
Air Blank	0.000	05:09	

---

Data Download Date/Time: 05/25/2022 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/15/2022 05:26:19

---

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:12	
Control Test	0.081	05:12	
Air Blank	0.000	05:13	
Subject Sample # 1	0.131	05:15	
Breath Volume	1.167	05:15	
Air Blank	0.000	05:16	
Air Blank	0.000	05:17	
Subject Sample # 2	0.122	05:20	Volume Not Met
Breath Volume	0.531	05:20	
Air Blank	0.000	05:21	
Air Blank	0.000	05:23	
Subject Sample # 3	0.131	05:24	
Breath Volume	1.378	05:24	
Air Blank	0.000	05:25	
Control Test	0.078	05:25	
Air Blank	0.000	05:26	
Diagnostics Check	OK	05:26	

---

Data Download Date/Time: 05/25/2022 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HERNANDEZ,REYNA  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SIDDIQUI

DATE/TIME OF TEST: 04/17/2022 02:20:11

---

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.189	02:15
Breath Volume	1.578	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.172	02:18
Breath Volume	1.890	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

---

Data Download Date/Time: 05/25/2022 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/17/2022 03:45:42

---

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.153	03:41
Breath Volume	2.386	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.151	03:44
Breath Volume	2.308	03:44
Air Blank	0.000	03:44
Control Test	0.079	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

---

Data Download Date/Time: 05/25/2022 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: FACCINI,DAVID  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SEQUEIRA

DATE/TIME OF TEST: 04/18/2022 00:31:52

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.166	00:27
Breath Volume	1.902	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.162	00:30
Breath Volume	2.308	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

---

Data Download Date/Time: 05/25/2022 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LAU,MATTHEW  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 09:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: HICE

DATE/TIME OF TEST: 04/23/2022 11:26:01

---

Diagnostics Check	OK	11:19	
Air Blank	0.000	11:20	
Control Test	0.081	11:20	
Air Blank	0.000	11:20	
Subject Sample # 1	0.000	11:21	
Breath Volume	2.347	11:21	
Air Blank	0.000	11:22	
Air Blank	0.000	11:24	
Subject Sample # 2	0.000	11:24	
Breath Volume	2.105	11:24	
Air Blank	0.000	11:25	
Control Test	0.081	11:25	
Air Blank		11:25	RFI Detect
Air Blank	0.000	11:26	

---

Data Download Date/Time: 05/25/2022 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LAU,MATTHEW  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 09:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: HICE

DATE/TIME OF TEST: 04/23/2022 11:36:00

---

Diagnostics Check	OK	11:28
Air Blank	0.000	11:29
Control Test	0.081	11:29
Air Blank	0.000	11:30
Subject Sample # 1	0.000	11:31
Breath Volume	2.015	11:31
Air Blank	0.000	11:32
Air Blank	0.000	11:34
Subject Sample # 2	0.000	11:34
Breath Volume	2.000	11:34
Air Blank	0.000	11:35
Control Test	0.081	11:35
Air Blank	0.000	11:35
Diagnostics Check	OK	11:36

---

Data Download Date/Time: 05/25/2022 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GUZMAN,JOSE  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: BARRERA GUZMAN

DATE/TIME OF TEST: 04/26/2022 00:45:42

---

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.081	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.139	00:41
Breath Volume	2.785	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.136	00:44
Breath Volume	2.425	00:44
Air Blank	0.000	00:44
Control Test	0.079	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

---

Data Download Date/Time: 05/25/2022 10:15



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GUZMAN,JOSE  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: FLITE

DATE/TIME OF TEST: 04/26/2022 04:41:20

---

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:30	
Control Test	0.081	04:30	
Air Blank	0.000	04:31	
Subject Sample	# 1 0.134	04:34	Volume Not Met
Breath Volume	0.101	04:34	
Air Blank	0.000	04:35	
Air Blank	0.000	04:36	
Subject Sample	# 2	04:39	No Sample Provided
Breath Volume	0.000	04:39	
Air Blank	0.000	04:40	
Control Test	0.077	04:40	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

---

Data Download Date/Time: 05/25/2022 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: POPICK,SCOTT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BELLO

DATE/TIME OF TEST: 04/29/2022 23:19:06

---

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.119	23:13
Breath Volume	1.835	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.117	23:17
Breath Volume	2.085	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

---

Data Download Date/Time: 05/25/2022 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/30/2022 00:30:42

---

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.043	00:25
Breath Volume	1.582	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.043	00:29
Breath Volume	1.570	00:29
Air Blank	0.000	00:29
Control Test	0.080	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

---

Data Download Date/Time: 05/25/2022 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ARMSTRONG,BRITTANY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 04/30/2022 01:50:49

---

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.183	01:46
Breath Volume	1.742	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.180	01:49
Breath Volume	2.679	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

---

Data Download Date/Time: 05/25/2022 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BOHM,JONATHON  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: LESINA

DATE/TIME OF TEST: 04/01/2022 02:02:48

---

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.080	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.195	01:57
Breath Volume	1.523	01:57
Air Blank	0.000	01:58
Air Blank	0.000	01:59
Subject Sample # 2	0.194	02:00
Breath Volume	2.617	02:00
Air Blank	0.000	02:01
Control Test	0.078	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time: 04/13/2022 12:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ARMSTRONG,BRITTNEY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: DUPUIS

DATE/TIME OF TEST: 04/02/2022 01:19:55

---

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.112	01:15
Breath Volume	3.199	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:17
Subject Sample # 2	0.108	01:18
Breath Volume	2.343	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

---

Data Download Date/Time: 04/13/2022 12:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: OSBOURNE, TIFFANY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SANTOS ABAD

DATE/TIME OF TEST: 04/02/2022 03:12:00

---

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.078	03:07
Breath Volume	1.468	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample # 2	0.075	03:10
Breath Volume	1.968	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:12

---

Data Download Date/Time: 04/13/2022 12:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: POPICK,SCOTT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 04/02/2022 04:46:00

---

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.101	04:41
Breath Volume	2.476	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:43
Subject Sample # 2	0.100	04:44
Breath Volume	2.542	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:46

---

Data Download Date/Time: 04/13/2022 12:46



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: POPICK,SCOTT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: RAMIREZ ROJAS

DATE/TIME OF TEST: 04/03/2022 03:54:59

---

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.180	03:49
Breath Volume	1.507	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.197	03:53
Breath Volume	1.382	03:53
Air Blank	0.000	03:54
Control Test	0.077	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

---

Data Download Date/Time: 04/13/2022 12:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUZMAN,JOSE,M  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GUZMAN,JOSE  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: NICHOLSON

DATE/TIME OF TEST: 04/04/2022 06:07:04

---

Diagnostics Check	OK	05:57
Air Blank	0.000	05:58
Control Test	0.081	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.074	06:02
Breath Volume	1.921	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:04
Subject Sample # 2	0.071	06:05
Breath Volume	1.667	06:05
Air Blank	0.000	06:06
Control Test	0.077	06:06
Air Blank	0.000	06:06
Diagnostics Check	OK	06:07

---

Data Download Date/Time: 04/13/2022 12:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BERMAN,GARY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:11

SUBJECT'S LAST NAME: BOYLE

DATE/TIME OF TEST: 04/07/2022 09:43:27

---

Diagnostics Check	OK	09:37
Air Blank	0.000	09:37
Control Test	0.080	09:37
Air Blank	0.000	09:38
Subject Sample # 1	0.000	09:39
Breath Volume	1.480	09:39
Air Blank	0.000	09:39
Air Blank	0.000	09:41
Subject Sample # 2	0.000	09:42
Breath Volume	1.882	09:42
Air Blank	0.000	09:42
Control Test	0.081	09:42
Air Blank	0.000	09:43
Diagnostics Check	OK	09:43

---

Data Download Date/Time: 04/13/2022 12:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HALES,ANDREW  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 04/09/2022 08:44:03

---

Diagnostics Check	OK	08:37
Air Blank	0.000	08:37
Control Test	0.081	08:38
Air Blank	0.000	08:38
Subject Sample # 1	0.201	08:39
Breath Volume	2.714	08:39
Air Blank	0.000	08:40
Air Blank	0.000	08:41
Subject Sample # 2	0.197	08:42
Breath Volume	2.320	08:42
Air Blank	0.000	08:43
Control Test	0.080	08:43
Air Blank	0.000	08:43
Diagnostics Check	OK	08:44

---

Data Download Date/Time: 04/13/2022 12:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GREISHEIMER,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: YASSINE

DATE/TIME OF TEST: 04/09/2022 10:18:20

---

Diagnostics Check	OK	10:11
Air Blank	0.000	10:11
Control Test	0.080	10:11
Air Blank	0.000	10:12
Subject Sample # 1	0.162	10:13
Breath Volume	1.542	10:13
Air Blank	0.000	10:14
Air Blank	0.000	10:15
Subject Sample # 2	0.161	10:16
Breath Volume	1.804	10:16
Air Blank	0.000	10:17
Control Test	0.080	10:17
Air Blank	0.000	10:18
Diagnostics Check	OK	10:18

---

Data Download Date/Time: 04/13/2022 12:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOPEZ,BRYAN  
ARRESTING OFFICER AGENCY: Hallandale Beach PD  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: VASILENKA

DATE/TIME OF TEST: 04/09/2022 19:55:57

---

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.080	19:46
Air Blank	0.000	19:47
Subject Sample # 1	0.289	19:48
Breath Volume	2.507	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:50
Subject Sample # 2	0.258	19:51
Breath Volume	1.300	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:53
Subject Sample # 3	0.285	19:54
Breath Volume	1.851	19:54
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

---

Data Download Date/Time: 04/13/2022 12:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: DIAZ,SUZANNE  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/09/2022 20:41:24

---

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:40	
Control Test	0.072	20:40	Control Outside Tolerance
Air Blank	0.000	20:41	

---

Data Download Date/Time: 04/13/2022 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HALES,ANDREW  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/09/2022 20:55:14

---

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:47	
Control Test	0.076	20:48	
Air Blank	0.000	20:48	
Subject Sample	# 1 0.218	20:50	
Breath Volume	2.226	20:50	
Air Blank	0.000	20:51	
Air Blank	0.000	20:52	
Subject Sample	# 2 0.210	20:53	
Breath Volume	2.023	20:53	
Air Blank	0.000	20:54	
Control Test	0.073	20:54	Control Outside Tolerance
Air Blank	0.000	20:55	

---

Data Download Date/Time: 04/13/2022 13:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/22/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 04/01/2022 04:07:39

---

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.081	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.128	04:02
Breath Volume		2.218	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.133	04:06
Breath Volume		1.652	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

---

Data Download Date/Time: 04/29/2022 01:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,WILLIAM,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/22/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 04/02/2022 05:08:05

---

Diagnostics Check		OK	05:00
Air Blank		0.000	05:01
Control Test		0.080	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.076	05:02
Breath Volume		2.671	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.071	05:06
Breath Volume		2.355	05:06
Air Blank		0.000	05:07
Control Test		0.082	05:07
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

---

Data Download Date/Time: 04/29/2022 01:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/22/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: COELHO

DATE/TIME OF TEST: 04/07/2022 00:45:10

---

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.083	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.086	00:40
Breath Volume	2.562	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.081	00:43
Breath Volume	1.738	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

---

Data Download Date/Time: 04/29/2022 01:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: COLLINS, THOMAS  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 03/22/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: POTTER

DATE/TIME OF TEST: 04/09/2022 00:21:31

---

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.081	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.163	00:16
Breath Volume	1.136	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:18
Subject Sample # 2	0.174	00:19
Breath Volume	1.867	00:19
Air Blank	0.000	00:20
Control Test	0.082	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

---

Data Download Date/Time: 04/29/2022 01:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,WILLIAM,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/22/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: BARTLETT

DATE/TIME OF TEST: 04/09/2022 02:34:54

---

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.082	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.077	02:30
Breath Volume	1.734	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:32
Subject Sample # 2	0.077	02:33
Breath Volume	1.800	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time: 04/29/2022 01:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/22/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 04/17/2022 02:23:58

---

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:19	
Control Test	0.081	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

---

Data Download Date/Time: 04/29/2022 01:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CHACON FONTANA

DATE/TIME OF TEST: 04/30/2022 05:18:21

---

Diagnostics Check		OK	05:11
Air Blank		0.000	05:12
Control Test		0.082	05:12
Air Blank		0.000	05:13
Subject Sample	# 1	0.063	05:13
Breath Volume		1.128	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.071	05:16
Breath Volume		1.351	05:16
Air Blank		0.000	05:17
Control Test		0.082	05:17
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

---

Data Download Date/Time: 06/03/2022 11:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MOSCHETTO,ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 03/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03521080A3  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: RODRIGUEZ MORALES

DATE/TIME OF TEST: 04/02/2022 00:31:51

---

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:29	
Subject Sample	# 1 0.343	00:30	
Breath Volume	3.214	00:30	
Air Blank		00:31	Purge Fail
Air Blank	0.000	00:31	

---

Data Download Date/Time: 04/07/2022 16:06



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MOSCHETTO,ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 03/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03521080A3  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: RODRIGUEZ MORALES

DATE/TIME OF TEST: 04/02/2022 00:40:29

---

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.075	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.340	00:35
Breath Volume	3.050	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:37
Subject Sample # 2	0.333	00:38
Breath Volume	3.015	00:38
Air Blank	0.000	00:39
Control Test	0.077	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

---

Data Download Date/Time: 04/07/2022 16:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MOSCHETTO,ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 03/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03521080A3  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: WEDDING

DATE/TIME OF TEST: 04/07/2022 20:38:27

---

Diagnostics Check	OK	20:33	
Air Blank	0.000	20:33	
Control Test	0.078	20:34	
Air Blank	0.000	20:34	
Subject Sample	# 1	20:37	Subject Test Refused
Breath Volume	0.000	20:37	
Air Blank	0.000	20:37	
Control Test	0.081	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

---

Data Download Date/Time: 04/21/2022 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MOSCHETTO,ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 03/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03521080A3  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: WEDDING

DATE/TIME OF TEST: 04/07/2022 20:46:19

---

Diagnostics Check	OK	20:39
Air Blank	0.000	20:40
Control Test	0.080	20:40
Air Blank	0.000	20:41
Subject Sample # 1	0.074	20:41
Breath Volume	1.738	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.075	20:44
Breath Volume	2.121	20:44
Air Blank	0.000	20:45
Control Test	0.080	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

---

Data Download Date/Time: 04/21/2022 13:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN,JAMES,  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: ALLEN,JAMES  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 03/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03521080A3  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 04/08/2022 03:55:42

---

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.173	03:50
Breath Volume	2.191	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:52
Subject Sample # 2	0.180	03:53
Breath Volume	1.144	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

---

Data Download Date/Time: 04/21/2022 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MOSCHETTO,ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 03/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03521080A3  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 04/10/2022 01:44:59

---

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.050	01:40
Breath Volume	2.449	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample # 2	0.046	01:43
Breath Volume	2.941	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

---

Data Download Date/Time: 04/21/2022 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARTON, JEREMY,  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: BARTON, JEREMY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 03/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03521080A3  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 04/15/2022 03:03:12

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.081	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.170	02:58
Breath Volume	1.636	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample # 2	0.165	03:01
Breath Volume	1.878	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

---

Data Download Date/Time: 04/21/2022 13:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MOSCHETTO,ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 03/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03521080A3  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: RASPPERRY

DATE/TIME OF TEST: 04/15/2022 23:34:31

---

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.091	23:29
Breath Volume	3.078	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.094	23:32
Breath Volume	2.425	23:32
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

---

Data Download Date/Time: 04/21/2022 13:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MOSCHETTO,ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 03/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03521080A3  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: BEGAY

DATE/TIME OF TEST: 04/16/2022 21:38:10

---

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.080	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.301	21:33
Breath Volume	2.863	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:35
Subject Sample # 2	0.296	21:36
Breath Volume	3.015	21:36
Air Blank	0.000	21:37
Control Test	0.078	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

---

Data Download Date/Time: 04/21/2022 13:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,VICTOR,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RIVERA,VICTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03521080A3  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: BEASLEY JR

DATE/TIME OF TEST: 04/07/2022 03:05:06

---

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.078	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.130	02:58
Breath Volume		1.210	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:00
Subject Sample	# 2	0.140	03:03
Breath Volume		1.875	03:03
Air Blank		0.000	03:04
Control Test		0.075	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

---

Data Download Date/Time: 04/21/2022 21:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNCH,CHAD,J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: JUSTICE,JUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 04/09/2022 01:48:07

---

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.081	01:45	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:46	RFI Detect
Breath Volume	0.031	01:46	
Air Blank	0.000	01:47	
Control Test	0.080	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

---

Data Download Date/Time: 04/19/2022 12:15



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNCH,CHAD,J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: JUSTICE,JUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 04/09/2022 01:54:31

---

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:53	
Control Test	0.080	01:53	
Air Blank		01:54	RFI Detect
Air Blank		01:54	RFI Detect

---

Data Download Date/Time: 04/19/2022 12:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNCH,CHAD,J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LYNCH,CHAD  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 16:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 04/20/2022 17:40:16

---

Diagnostics Check	OK	17:33
Air Blank	0.000	17:33
Control Test	0.080	17:33
Air Blank	0.000	17:34
Subject Sample # 1	0.081	17:35
Breath Volume	1.242	17:35
Air Blank	0.000	17:36
Air Blank	0.000	17:38
Subject Sample # 2	0.077	17:38
Breath Volume	1.234	17:38
Air Blank	0.000	17:39
Control Test	0.080	17:39
Air Blank	0.000	17:40
Diagnostics Check	OK	17:40

---

Data Download Date/Time: 05/04/2022 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNCH,CHAD,J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: FOLSOM,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: MCCARY

DATE/TIME OF TEST: 04/20/2022 19:08:20

---

Diagnostics Check	OK	19:00	
Air Blank	0.000	19:01	
Control Test	0.080	19:01	
Air Blank	0.000	19:02	
Subject Sample	# 1 0.198	19:02	
Breath Volume	1.453	19:02	
Air Blank	0.000	19:03	
Air Blank	0.000	19:05	
Subject Sample	# 2 0.208	19:06	
Breath Volume	2.503	19:06	
Air Blank	0.000	19:07	
Control Test	0.079	19:07	
Air Blank		19:07	RFI Detect
Air Blank	0.000	19:08	

---

Data Download Date/Time: 05/04/2022 15:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ALEX,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: PINTO,JOSE  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 04/09/2022 23:25:20

---

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.081	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:24	
Control Test	0.082	23:24	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

---

Data Download Date/Time: 04/25/2022 14:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JENKINS,ANTHONY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HERNANDEZ,STEVEN  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13921080A2  
DRY GAS EXPIRY DATE: 07/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: ORTIZ JIMENEZ

DATE/TIME OF TEST: 04/01/2022 22:12:52

---

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.078	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.199	22:08
Breath Volume	1.781	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.192	22:11
Breath Volume	1.414	22:11
Air Blank	0.000	22:11
Control Test	0.080	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

---

Data Download Date/Time: 04/19/2022 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JENKINS,ANTHONY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: JENKINS,ANTHONY  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13921080A2  
DRY GAS EXPIRY DATE: 07/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: ANDARA MEJIA

DATE/TIME OF TEST: 04/03/2022 20:52:21

---

Diagnostics Check	OK	20:44
Air Blank	0.000	20:45
Control Test	0.079	20:45
Air Blank	0.000	20:46
Subject Sample # 1	0.249	20:47
Breath Volume	3.113	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:49
Subject Sample # 2	0.246	20:50
Breath Volume	2.699	20:50
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

---

Data Download Date/Time: 04/19/2022 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JENKINS,ANTHONY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTER,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13921080A2  
DRY GAS EXPIRY DATE: 07/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 05:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: BATIR

DATE/TIME OF TEST: 04/04/2022 06:13:03

---

Diagnostics Check	OK	06:06
Air Blank	0.000	06:06
Control Test	0.077	06:07
Air Blank	0.000	06:07
Subject Sample # 1	0.115	06:08
Breath Volume	1.820	06:08
Air Blank	0.000	06:09
Air Blank	0.000	06:10
Subject Sample # 2	0.113	06:11
Breath Volume	1.878	06:11
Air Blank	0.000	06:12
Control Test	0.082	06:12
Air Blank	0.000	06:12
Diagnostics Check	OK	06:13

---

Data Download Date/Time: 04/19/2022 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARBOGAST, JOSHUA, S  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SELF, GRANT  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: BUI  
ARREST TIME: 20:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: WIRTH

DATE/TIME OF TEST: 04/01/2022 20:55:11

---

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:52	
Control Test	0.079	20:52	
Air Blank	0.000	20:52	
Subject Sample	# 1	20:53	Subject Test Refused
Breath Volume	0.000	20:53	
Air Blank	0.000	20:54	
Control Test	0.078	20:54	RFI Detect
Air Blank	0.000	20:55	

---

Data Download Date/Time: 04/21/2022 13:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WILLIAMS,DANIEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: BUI  
ARREST TIME: 14:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 04/02/2022 14:52:15

---

Diagnostics Check	OK	14:46
Air Blank	0.000	14:46
Control Test	0.077	14:46
Air Blank	0.000	14:47
Subject Sample # 1	0.032	14:47
Breath Volume	2.769	14:47
Air Blank	0.000	14:48
Air Blank	0.000	14:50
Subject Sample # 2	0.031	14:50
Breath Volume	2.644	14:50
Air Blank	0.000	14:51
Control Test	0.077	14:51
Air Blank	0.000	14:52
Diagnostics Check	OK	14:52

---

Data Download Date/Time: 05/04/2022 11:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: STINSON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: BUI  
ARREST TIME: 17:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: BROGDEN

DATE/TIME OF TEST: 04/04/2022 18:18:51

---

Diagnostics Check	OK	18:10	
Air Blank	0.000	18:10	
Control Test	0.077	18:10	
Air Blank	0.000	18:11	
Subject Sample	# 1	18:14	No Sample Provided
Breath Volume	0.000	18:14	
Air Blank	0.000	18:15	
Air Blank	0.000	18:17	
Subject Sample	# 2	18:17	Subject Test Refused
Breath Volume	0.000	18:17	
Air Blank	0.000	18:17	
Control Test	0.078	18:18	
Air Blank	0.000	18:18	
Diagnostics Check	OK	18:18	

---

Data Download Date/Time: 05/04/2022 11:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,DANIEL,S

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: WILLIAMS,DANIEL

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 16:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: BREED

DATE/TIME OF TEST: 04/10/2022 17:14:50

---

Diagnostics Check	OK	17:03	
Air Blank	0.000	17:03	
Control Test	0.078	17:04	
Air Blank	0.000	17:04	
Subject Sample	# 1	17:07	No Sample Provided
Breath Volume	0.000	17:07	
Air Blank	0.000	17:08	
Air Blank	0.000	17:10	
Subject Sample	# 2	17:13	No Sample Provided
Breath Volume	0.000	17:13	
Air Blank	0.000	17:13	
Control Test	0.078	17:14	
Air Blank	0.000	17:14	
Diagnostics Check	OK	17:14	

---

Data Download Date/Time: 05/04/2022 11:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 15421080A1
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S	DRY GAS EXPIRY DATE: 08/05/2023
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: WILLIAMS,DANIEL	ARREST TIME: 16:44
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 03/02/2022	OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: B630817671010      DATE/TIME OF TEST: 04/10/2022      17:20:26

---

Diagnostics Check	OK	17:17	
Air Blank	0.000	17:17	
Control Test	0.078	17:18	
Air Blank	0.000	17:18	
Subject Sample	# 1	17:19	Subject Test Refused
Breath Volume	0.000	17:19	
Air Blank	0.000	17:19	
Control Test	0.078	17:19	
Air Blank	0.000	17:20	
Diagnostics Check	OK	17:20	

---

Data Download Date/Time: 05/04/2022 11:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALTAIS,ANDREW,W  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MALTAIS,ANDREW  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: BUI  
ARREST TIME: 17:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: THORPE

DATE/TIME OF TEST: 04/15/2022 18:13:45

---

Diagnostics Check	OK	18:09	
Air Blank	0.000	18:09	
Control Test	0.078	18:09	
Air Blank	0.000	18:10	
Subject Sample	# 1	18:12	Subject Test Refused
Breath Volume	0.000	18:12	
Air Blank	0.000	18:12	
Control Test	0.078	18:13	
Air Blank	0.000	18:13	
Diagnostics Check	OK	18:13	

---

Data Download Date/Time: 05/04/2022 11:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOTT,CALEB,E  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: LOTT,CALEB  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 14:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:46

SUBJECT'S LAST NAME: PITT

DATE/TIME OF TEST: 04/05/2022 15:12:31

---

Diagnostics Check	OK	15:06
Air Blank	0.000	15:06
Control Test	0.081	15:07
Air Blank	0.000	15:07
Subject Sample # 1	0.000	15:08
Breath Volume	3.171	15:08
Air Blank	0.000	15:08
Air Blank	0.000	15:10
Subject Sample # 2	0.000	15:11
Breath Volume	3.507	15:11
Air Blank	0.000	15:11
Control Test	0.080	15:11
Air Blank	0.000	15:12
Diagnostics Check	OK	15:12

---

Data Download Date/Time: 04/27/2022 07:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALTMAN,BRANDON,W  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: GILLARD,A  
ARRESTING OFFICER AGENCY: Sebring PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LOCKETT

DATE/TIME OF TEST: 04/09/2022 03:45:29

---

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:34	
Control Test		0.083	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.140	03:38	Volume Not Met
Breath Volume		0.601	03:38	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2	0.133	03:43	Volume Not Met
Breath Volume		0.203	03:43	
Air Blank		0.000	03:44	
Control Test		0.080	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

---

Data Download Date/Time: 04/27/2022 07:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALTMAN,BRANDON,W  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: GILLARD,A  
ARRESTING OFFICER AGENCY: Sebring PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LOCKETT

DATE/TIME OF TEST: 04/09/2022 03:55:25

---

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.135	03:50
Breath Volume		1.722	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.134	03:53
Breath Volume		2.035	03:53
Air Blank		0.000	03:54
Control Test		0.081	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

---

Data Download Date/Time: 04/27/2022 07:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOTT,CALEB,E  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: ADAIR,RYAN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DEMERS

DATE/TIME OF TEST: 04/11/2022 03:28:20

---

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:16	
Control Test		0.083	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1	0.210	03:17	
Breath Volume		2.355	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2	0.210	03:23	Volume Not Met
Breath Volume		0.039	03:23	
Air Blank		0.000	03:24	
Air Blank		0.000	03:25	
Subject Sample	# 3	0.201	03:26	
Breath Volume		2.707	03:26	
Air Blank		0.000	03:27	
Control Test		0.079	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

---

Data Download Date/Time: 04/27/2022 07:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINISH,BARRY,A  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: CARRANZA,RUBEN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 04/15/2022 22:54:56

---

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.082	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.129	22:49
Breath Volume	2.171	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.126	22:53
Breath Volume	2.328	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

---

Data Download Date/Time: 04/27/2022 07:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOTT,CALEB,E  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: PALMERTON,KENNETH  
ARRESTING OFFICER AGENCY: Sebring PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 04/23/2022 01:34:39

---

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.082	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Control Test	0.081	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

---

Data Download Date/Time: 04/27/2022 07:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALTMAN,BRANDON,W  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: SALVADOR,OFFICER  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 04/23/2022 22:12:08

---

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.082	22:08	
Air Blank	0.000	22:09	
Subject Sample	# 1	22:10	Subject Test Refused
Breath Volume	0.000	22:10	
Air Blank	0.000	22:11	
Control Test	0.081	22:11	
Air Blank	0.000	22:12	
Diagnostics Check	OK	22:12	

---

Data Download Date/Time: 04/27/2022 07:38



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALTMAN,BRANDON,W  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: PALMERTON,K  
ARRESTING OFFICER AGENCY: Sebring PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 04/24/2022 03:43:37

---

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.082	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.081	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

---

Data Download Date/Time: 04/27/2022 07:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOTT,CALEB,E  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: ALLRED,AARON  
ARRESTING OFFICER AGENCY: Lake Placid PD  
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: KEYSER

DATE/TIME OF TEST: 04/29/2022 01:02:47

---

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.082	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.163	00:58
Breath Volume	1.437	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.161	01:01
Breath Volume	1.582	01:01
Air Blank	0.000	01:01
Control Test	0.081	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

---

Data Download Date/Time: 06/03/2022 11:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROE,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: ROE,ROBERT  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 10:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: CRAM

DATE/TIME OF TEST: 04/29/2022 11:20:18

---

Diagnostics Check	OK	11:14
Air Blank	0.000	11:14
Control Test	0.082	11:14
Air Blank	0.000	11:15
Subject Sample # 1	0.000	11:15
Breath Volume	3.222	11:15
Air Blank	0.000	11:16
Air Blank	0.000	11:18
Subject Sample # 2	0.000	11:18
Breath Volume	3.121	11:18
Air Blank	0.000	11:19
Control Test	0.080	11:19
Air Blank	0.000	11:20
Diagnostics Check	OK	11:20

---

Data Download Date/Time: 06/03/2022 11:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROE,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: ROE,ROBERT  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 14:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: PITT

DATE/TIME OF TEST: 04/29/2022 14:44:05

---

Diagnostics Check	OK	14:38
Air Blank	0.000	14:38
Control Test	0.082	14:38
Air Blank	0.000	14:39
Subject Sample # 1	0.000	14:39
Breath Volume	2.941	14:39
Air Blank	0.000	14:40
Air Blank	0.000	14:42
Subject Sample # 2	0.000	14:42
Breath Volume	1.925	14:42
Air Blank	0.000	14:43
Control Test	0.081	14:43
Air Blank	0.000	14:44
Diagnostics Check	OK	14:44

---

Data Download Date/Time: 06/03/2022 11:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROE,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: ROE,ROBERT  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 15:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 04/29/2022 15:57:38

---

Diagnostics Check	OK	15:51
Air Blank	0.000	15:52
Control Test	0.081	15:52
Air Blank	0.000	15:52
Subject Sample # 1	0.000	15:53
Breath Volume	1.667	15:53
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.000	15:56
Breath Volume	1.808	15:56
Air Blank	0.000	15:56
Control Test	0.080	15:57
Air Blank	0.000	15:57
Diagnostics Check	OK	15:57

---

Data Download Date/Time: 06/03/2022 11:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GLASCO,GREGORY,S  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: GLASCO,GREGORY  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 04/29/2022 19:49:58

---

Diagnostics Check	OK	19:44
Air Blank	0.000	19:44
Control Test	0.082	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.000	19:45
Breath Volume	1.335	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.000	19:48
Breath Volume	1.820	19:48
Air Blank	0.000	19:49
Control Test	0.081	19:49
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

---

Data Download Date/Time: 06/03/2022 11:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINISH,BARRY,A  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: HINISH,BARRY  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: FLOREZ

DATE/TIME OF TEST: 04/01/2022 21:55:15

---

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.000	21:50
Breath Volume	1.773	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.000	21:53
Breath Volume	2.023	21:53
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

---

Data Download Date/Time: 04/27/2022 07:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINISH,BARRY,A  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: CARRANZA,RUBEN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 04/02/2022 03:27:49

---

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.079	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.285	03:26	
Air Blank	0.000	03:26	
Control Test	0.079	03:27	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

---

Data Download Date/Time: 04/27/2022 07:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINISH,BARRY,A  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: HINISH,BARRY  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: GREGIS

DATE/TIME OF TEST: 04/06/2022 19:08:19

---

Diagnostics Check	OK	19:01
Air Blank	0.000	19:01
Control Test	0.079	19:02
Air Blank	0.000	19:02
Subject Sample # 1	0.000	19:03
Breath Volume	1.253	19:03
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.000	19:06
Breath Volume	1.910	19:06
Air Blank	0.000	19:07
Control Test	0.078	19:07
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

---

Data Download Date/Time: 04/27/2022 07:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINISH,BARRY,A  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: HINISH,BARRY  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: MONTIEL

DATE/TIME OF TEST: 04/06/2022 19:22:24

---

Diagnostics Check	OK	19:15
Air Blank	0.000	19:15
Control Test	0.079	19:16
Air Blank	0.000	19:16
Subject Sample # 1	0.000	19:17
Breath Volume	1.984	19:17
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.000	19:20
Breath Volume	1.468	19:20
Air Blank	0.000	19:21
Control Test	0.080	19:21
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

---

Data Download Date/Time: 04/27/2022 07:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINISH,BARRY,A  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: CARRANZA,RUBEN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 04/08/2022 03:41:47

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.118	03:36
Breath Volume	1.585	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.110	03:40
Breath Volume	1.363	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

---

Data Download Date/Time: 04/27/2022 07:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINISH,BARRY,A  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: CARRANZA,RUBEN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: DEL VALLE VELEZ

DATE/TIME OF TEST: 04/16/2022 03:26:27

---

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.180	03:21
Breath Volume	2.132	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.172	03:24
Breath Volume	1.593	03:24
Air Blank	0.000	03:25
Control Test	0.078	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

---

Data Download Date/Time: 04/27/2022 07:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINISH,BARRY,A  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: HINISH,BARRY  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: GREGIS

DATE/TIME OF TEST: 04/20/2022 19:14:09

---

Diagnostics Check	OK	19:07
Air Blank	0.000	19:08
Control Test	0.080	19:08
Air Blank	0.000	19:09
Subject Sample # 1	0.000	19:09
Breath Volume	2.113	19:09
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample # 2	0.000	19:12
Breath Volume	2.058	19:12
Air Blank	0.000	19:13
Control Test	0.079	19:13
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

---

Data Download Date/Time: 04/27/2022 07:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINISH,BARRY,A  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: CARRANZA,RUBEN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: MCCANN

DATE/TIME OF TEST: 04/21/2022 03:14:55

---

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.079	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.158	03:09
Breath Volume		2.550	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.153	03:13
Breath Volume		2.839	03:13
Air Blank		0.000	03:13
Control Test		0.075	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

---

Data Download Date/Time: 04/27/2022 07:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINISH,BARRY,A  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: CARRANZA,RUBEN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: HORTERT

DATE/TIME OF TEST: 04/22/2022 03:45:32

---

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.141	03:40
Breath Volume	2.777	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.144	03:43
Breath Volume	2.996	03:43
Air Blank	0.000	03:44
Control Test	0.075	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

---

Data Download Date/Time: 04/27/2022 07:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DITTMN, LAVERNE, T  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: BATES, JOBIE  
ARRESTING OFFICER AGENCY: Arcadia PD  
LAST AGENCY INSPECTION: 03/24/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 22521080A1  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: MENDIOLA

DATE/TIME OF TEST: 04/01/2022 00:38:11

---

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.079	00:22	
Air Blank	0.000	00:23	
Subject Sample	# 1 0.114	00:25	
Breath Volume	1.171	00:25	
Air Blank	0.000	00:26	
Air Blank	0.000	00:28	
Subject Sample	# 2 0.099	00:31	Volume Not Met
Breath Volume	0.683	00:31	
Air Blank	0.000	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 3	00:36	Subject Test Refused
Breath Volume	0.164	00:36	
Air Blank	0.000	00:37	
Control Test	0.078	00:37	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

---

Data Download Date/Time: 04/30/2022 09:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VEGA,JASON,  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: HAEUSSLER,BRETT  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 03/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A1  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 15:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: RODRIGUEZ-MARTINEZ

DATE/TIME OF TEST: 04/02/2022 16:57:42

---

Diagnostics Check		OK	16:51
Air Blank		0.000	16:51
Control Test		0.079	16:51
Air Blank		0.000	16:52
Subject Sample	# 1	0.216	16:52
Breath Volume		1.937	16:52
Air Blank		0.000	16:53
Air Blank		0.000	16:55
Subject Sample	# 2	0.216	16:55
Breath Volume		1.894	16:55
Air Blank		0.000	16:56
Control Test		0.079	16:57
Air Blank		0.000	16:57
Diagnostics Check		OK	16:57

---

Data Download Date/Time: 04/30/2022 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DANIEL,KYLE,J  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: SMITH,DUSTIN  
ARRESTING OFFICER AGENCY: Hardee County SO  
LAST AGENCY INSPECTION: 03/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A1  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 04/07/2022 04:35:07

---

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.140	04:30
Breath Volume		1.527	04:30
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.140	04:33
Breath Volume		1.566	04:33
Air Blank		0.000	04:34
Control Test		0.079	04:34
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

---

Data Download Date/Time: 04/30/2022 09:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: POVEDA,CARLOS  
ARRESTING OFFICER AGENCY: DeSoto County SO  
LAST AGENCY INSPECTION: 03/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A1  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DE JESUS CHAVEZ

DATE/TIME OF TEST: 04/26/2022 00:01:58

---

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.079	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:01	
Control Test	0.079	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

---

Data Download Date/Time: 04/30/2022 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: SANTANA,ELIO  
ARRESTING OFFICER AGENCY: DeSoto County SO  
LAST AGENCY INSPECTION: 03/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A1  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: PALAFOX

DATE/TIME OF TEST: 04/27/2022 05:54:26

---

Diagnostics Check	OK	05:51	
Air Blank	0.000	05:52	
Control Test	0.080	05:52	
Air Blank	0.000	05:52	
Subject Sample	# 1	05:53	Subject Test Refused
Breath Volume	0.000	05:53	
Air Blank	0.000	05:53	
Control Test	0.080	05:53	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:54	

---

Data Download Date/Time: 04/30/2022 09:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,J,G  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HOWARD,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211543  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: FULTON

DATE/TIME OF TEST: 04/06/2022 23:24:25

---

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.077	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.200	23:19
Breath Volume		1.863	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.206	23:22
Breath Volume		1.546	23:22
Air Blank		0.000	23:23
Control Test		0.076	23:23
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

---

Data Download Date/Time: 04/28/2022 22:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,J,G  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: WARD,G  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211543  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: KARAS

DATE/TIME OF TEST: 04/07/2022 01:02:45

---

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.076	00:56	
Air Blank	0.000	00:57	
Subject Sample # 1	0.175	00:58	
Breath Volume	2.175	00:58	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample # 2	0.192	01:01	
Breath Volume	3.164	01:01	
Air Blank	0.000	01:01	
Control Test	0.074	01:02	Control Outside Tolerance
Air Blank	0.000	01:02	

---

Data Download Date/Time: 04/28/2022 22:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,J,G  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HOWARD,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211543  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 14:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: VALENCIA ROJAS

DATE/TIME OF TEST: 04/07/2022 16:01:37

---

Diagnostics Check	OK	15:54	
Air Blank	0.000	15:54	
Control Test	0.078	15:55	
Air Blank	0.000	15:55	
Subject Sample	# 1 0.146	15:57	
Breath Volume	2.968	15:57	
Air Blank	0.000	15:57	
Air Blank	0.000	15:59	
Subject Sample	# 2 0.164	15:59	
Breath Volume	2.320	15:59	
Air Blank	0.000	16:00	
Control Test	0.074	16:01	Control Outside Tolerance
Air Blank	0.000	16:01	

---

Data Download Date/Time: 04/28/2022 22:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,J,G  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HOWARD,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211543  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: HEILENMAN

DATE/TIME OF TEST: 04/16/2022 00:31:32

---

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.078	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.211	00:26
Breath Volume		3.421	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.207	00:29
Breath Volume		3.460	00:29
Air Blank		0.000	00:30
Control Test		0.076	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

---

Data Download Date/Time: 04/28/2022 22:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,J,G  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HOWARD,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211543  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: PIRIZ

DATE/TIME OF TEST: 04/19/2022 20:16:35

---

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.078	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.000	20:12
Breath Volume	1.457	20:12
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.000	20:15
Breath Volume	1.695	20:15
Air Blank	0.000	20:15
Control Test	0.077	20:15
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

---

Data Download Date/Time: 04/28/2022 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,JOHNATHAN,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: BROWN,JOHNATHAN  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211543  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 04/23/2022 22:15:48

---

Diagnostics Check	OK	22:07
Air Blank	0.000	22:08
Control Test	0.078	22:08
Air Blank	0.000	22:09
Subject Sample # 1	0.185	22:09
Breath Volume	1.402	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.172	22:14
Breath Volume	1.105	22:14
Air Blank	0.000	22:14
Control Test	0.077	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

---

Data Download Date/Time: 04/28/2022 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEYMANS,MORNE,  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HEYMANS,MORNE  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211543  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/24/2022 03:42:35

---

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:31	
Control Test		0.078	03:31	
Air Blank		0.000	03:32	
Subject Sample	# 1	0.116	03:35	Volume Not Met
Breath Volume		0.000	03:35	
Air Blank		0.000	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 2	0.129	03:41	Volume Not Met
Breath Volume		0.097	03:41	
Air Blank		0.000	03:41	
Control Test		0.076	03:41	
Air Blank		0.000	03:42	
Diagnostics Check		OK	03:42	

---

Data Download Date/Time: 04/28/2022 22:30