

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LYBRAND,CHRISTOPHER,J

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: JENNINGS,ANTHONY

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: O325341-10

DRY GAS EXPIRY DATE: 01/18/2024

VIOLATION CODE: DUI

ARREST TIME: 15:05

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: MONK

DATE/TIME OF TEST: 05/01/2022 16:31:54

Diagnostics Check	OK	16:25
Air Blank	0.000	16:25
Control Test	0.080	16:26
Air Blank	0.000	16:26
Subject Sample # 1	0.238	16:27
Breath Volume	1.941	16:27
Air Blank	0.000	16:28
Air Blank	0.000	16:29
Subject Sample # 2	0.242	16:30
Breath Volume	1.703	16:30
Air Blank	0.000	16:30
Control Test	0.079	16:31
Air Blank	0.000	16:31
Diagnostics Check	OK	16:31

Data Download Date/Time: 05/23/2022 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GLENECK,JORDAN,A

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: HANSFORD,CARLDERIUS

ARRESTING OFFICER AGENCY: Ft. Walton Beach PD

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: O325341-10

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 01:32

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GAMBLE

DATE/TIME OF TEST: 05/02/2022 03:08:52

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.145	03:04
Breath Volume	2.058	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.134	03:07
Breath Volume	1.726	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 05/23/2022 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: O325341-10
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ATKINS,JAMES	ARREST TIME: 04:50
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Female
LAST AGENCY INSPECTION: 04/27/2022	OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: BURRELL	DATE/TIME OF TEST: 05/03/2022	05:45:05
------------------------------	-------------------------------	----------

Diagnostics Check	OK	05:38
Air Blank	0.000	05:39
Control Test	0.081	05:39
Air Blank	0.000	05:40
Subject Sample # 1	0.000	05:40
Breath Volume	3.179	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.000	05:43
Breath Volume	3.996	05:43
Air Blank	0.000	05:44
Control Test	0.080	05:44
Air Blank	0.000	05:45
Diagnostics Check	OK	05:45

Data Download Date/Time: 05/23/2022 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WARF,KYLE,J

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: WARF,KYLE

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: O325341-10

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 01:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: SPRINGFLOAT

DATE/TIME OF TEST: 05/05/2022 02:34:27

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.081	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:29	No Sample Provided
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Air Blank	0.000	02:31	
Subject Sample	# 2	02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:33	
Control Test	0.080	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 05/23/2022 15:26

Subject Test Electronic Data

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: GRIFFIS

DATE/TIME OF TEST: 05/07/2022 03:49:28

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:47	
Control Test	0.082	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:48	
Control Test	0.080	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 05/23/2022 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 05/07/2022 04:57:11

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.105	04:52
Breath Volume	1.246	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.107	04:55
Breath Volume	1.929	04:55
Air Blank	0.000	04:56
Control Test	0.081	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 05/23/2022 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HACKETT,TRAVIS
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: OZUNA

DATE/TIME OF TEST: 05/08/2022 04:51:23

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.082	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.144	04:46
Breath Volume	3.386	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.142	04:49
Breath Volume	3.550	04:49
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 05/23/2022 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELTON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELTON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 06:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:24

SUBJECT'S LAST NAME: LOFTON

DATE/TIME OF TEST: 05/08/2022 07:51:26

Diagnostics Check	OK	07:45
Air Blank	0.000	07:45
Control Test	0.081	07:45
Air Blank	0.000	07:46
Subject Sample # 1	0.060	07:47
Breath Volume	1.601	07:47
Air Blank	0.000	07:47
Air Blank	0.000	07:49
Subject Sample # 2	0.061	07:49
Breath Volume	1.988	07:49
Air Blank	0.000	07:50
Control Test	0.082	07:50
Air Blank	0.000	07:51
Diagnostics Check	OK	07:51

Data Download Date/Time: 05/23/2022 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: CORTES MENDOZA

DATE/TIME OF TEST: 05/12/2022 03:47:40

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.014	03:42
Breath Volume	1.195	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.014	03:46
Breath Volume	2.230	03:46
Air Blank	0.000	03:46
Control Test	0.081	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 05/23/2022 15:26

Subject Test Electronic Data

DATE/TIME OF TEST: 05/13/2022 22:12:23

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:04	
Control Test	0.082	22:04	
Air Blank	0.000	22:05	
Subject Sample	# 1	22:08	No Sample Provided
Breath Volume	0.000	22:08	
Air Blank	0.000	22:08	
Air Blank	0.000	22:10	
Subject Sample	# 2	22:11	Subject Test Refused
Breath Volume	0.000	22:11	
Air Blank	0.000	22:11	
Control Test	0.080	22:11	
Air Blank	0.000	22:12	
Diagnostics Check	OK	22:12	

Data Download Date/Time: 05/23/2022 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: PABLO HERNANDEZ

DATE/TIME OF TEST: 05/13/2022 23:06:33

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:04	
Control Test	0.080	23:04	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:05	RFI Detect
Breath Volume	0.000	23:05	
Air Blank	0.000	23:05	
Control Test	0.080	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 05/23/2022 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: PABLO HERNANDEZ

DATE/TIME OF TEST: 05/13/2022 23:19:51

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:10	
Control Test	0.080	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1	23:14	No Sample Provided
Breath Volume	0.000	23:14	
Air Blank	0.000	23:15	
Air Blank	0.000	23:17	
Subject Sample	# 2	23:18	Subject Test Refused
Breath Volume	0.000	23:18	
Air Blank	0.000	23:18	
Control Test	0.081	23:19	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 05/23/2022 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAHR,DANIEL,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HAHR,DANIEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: BUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: FERNANDES COSTA

DATE/TIME OF TEST: 05/14/2022 20:53:21

Diagnostics Check	OK	20:41	
Air Blank	0.000	20:42	
Control Test	0.081	20:42	
Air Blank	0.000	20:43	
Subject Sample	# 1	20:46	No Sample Provided
Breath Volume	0.000	20:46	
Air Blank	0.000	20:46	
Air Blank	0.000	20:48	
Subject Sample	# 2	20:51	Volume Not Met
Breath Volume	0.289	20:51	
Air Blank	0.000	20:52	
Control Test	0.081	20:52	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

Data Download Date/Time: 05/23/2022 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: O325341-12
BREATH TEST OPERATOR: HAHR,DANIEL,E	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: HAHR,DANIEL	ARREST TIME: 19:00
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Female
LAST AGENCY INSPECTION: 04/27/2022	OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: FERNANDES COSTA	DATE/TIME OF TEST: 05/14/2022 21:34:58
--------------------------------------	--

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.160	21:28
Breath Volume	1.566	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:30
Subject Sample # 2	0.165	21:33
Breath Volume	1.875	21:33
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLENECK,JORDAN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GLENECK,JORDAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: HEIGES

DATE/TIME OF TEST: 05/15/2022 04:52:46

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.145	04:47
Breath Volume	1.972	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.153	04:51
Breath Volume	3.203	04:51
Air Blank	0.000	04:51
Control Test	0.082	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MALTAIS, ANDREW, W

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: MALTAIS, ANDREW

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: O325341-12

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: BUI

ARREST TIME: 13:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: MCCURLEY

DATE/TIME OF TEST: 05/15/2022 15:47:13

Diagnostics Check	OK	15:40
Air Blank	0.000	15:41
Control Test	0.081	15:41
Air Blank	0.000	15:42
Subject Sample # 1	0.129	15:42
Breath Volume	4.511	15:42
Air Blank	0.000	15:43
Air Blank	0.000	15:45
Subject Sample # 2	0.130	15:45
Breath Volume	4.703	15:45
Air Blank	0.000	15:46
Control Test	0.080	15:46
Air Blank	0.000	15:47
Diagnostics Check	OK	15:47

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMATEER, TRACY, D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PALMATEER, TRACY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: A140072993450

DATE/TIME OF TEST: 05/15/2022 21:53:22

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.080	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.121	21:48
Breath Volume	1.429	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.125	21:51
Breath Volume	1.496	21:51
Air Blank	0.000	21:52
Control Test	0.081	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,LARRY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: HOEFlickER

DATE/TIME OF TEST: 05/18/2022 05:31:11

Diagnostics Check	OK	05:24
Air Blank	0.000	05:25
Control Test	0.081	05:25
Air Blank	0.000	05:26
Subject Sample # 1	0.125	05:26
Breath Volume	4.988	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.116	05:29
Breath Volume	1.945	05:29
Air Blank	0.000	05:30
Control Test	0.078	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER, LANDON, L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ROUSH, ERIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 10:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:21

SUBJECT'S LAST NAME: NUTTING

DATE/TIME OF TEST: 05/18/2022 11:55:05

Diagnostics Check	OK	11:44
Air Blank	0.000	11:45
Control Test	0.079	11:45
Air Blank	0.000	11:46
Subject Sample # 1	0.343	11:48
Breath Volume	2.019	11:48
Air Blank	0.000	11:49
Air Blank	0.000	11:50
Subject Sample # 2	0.345	11:53
Breath Volume	1.472	11:53
Air Blank	0.000	11:54
Control Test	0.078	11:54
Air Blank	0.000	11:54
Diagnostics Check	OK	11:55

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MALTAIS,ANDREW,W

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: MALTAIS,ANDREW

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: O325341-12

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: BUI

ARREST TIME: 17:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/18/2022 19:06:31

Diagnostics Check	OK	19:00
Air Blank	0.000	19:00
Control Test	0.081	19:01
Air Blank	0.000	19:01
Subject Sample # 1	0.045	19:02
Breath Volume	3.496	19:02
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.045	19:05
Breath Volume	3.929	19:05
Air Blank	0.000	19:05
Control Test	0.079	19:05
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MURSCH,THEODORE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: LANIER

DATE/TIME OF TEST: 05/19/2022 23:49:10

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:41	
Control Test	0.081	23:41	
Air Blank	0.000	23:41	
Subject Sample	# 1	23:44	No Sample Provided
Breath Volume	0.000	23:44	
Air Blank	0.000	23:45	
Air Blank	0.000	23:47	
Subject Sample	# 2	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:48	
Control Test	0.081	23:48	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HART,JOSEPH,A

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: HANSFORD,CARLDERIUS

ARRESTING OFFICER AGENCY: Ft. Walton Beach PD

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: O325341-12

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 02:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ELJAHMI

DATE/TIME OF TEST: 05/20/2022 03:54:51

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.140	03:50
Breath Volume	2.710	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.140	03:53
Breath Volume	2.519	03:53
Air Blank	0.000	03:53
Control Test	0.079	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: O325341-12
BREATH TEST OPERATOR: CARTER,LARRY,J	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ATKINS,JAMES	ARREST TIME: 04:22
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Female
LAST AGENCY INSPECTION: 04/27/2022	OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: THOMPSON DATE/TIME OF TEST: 05/21/2022 05:31:01

Diagnostics Check	OK	05:24
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:26
Subject Sample # 1	0.195	05:26
Breath Volume	3.273	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.195	05:29
Breath Volume	2.902	05:29
Air Blank	0.000	05:30
Control Test	0.079	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:31

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SHELTON,ERIC,L

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: ROGERS,GERALD

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: O325341-12

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 13:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 13:46

SUBJECT'S LAST NAME: SAN AGUSTIN-VALDEZ

DATE/TIME OF TEST: 05/21/2022 14:12:54

Diagnostics Check	OK	14:06
Air Blank	0.000	14:07
Control Test	0.084	14:07
Air Blank	0.000	14:08
Subject Sample # 1	0.231	14:08
Breath Volume	1.636	14:08
Air Blank	0.000	14:09
Air Blank	0.000	14:10
Subject Sample # 2	0.238	14:11
Breath Volume	1.371	14:11
Air Blank	0.000	14:11
Control Test	0.080	14:12
Air Blank	0.000	14:12
Diagnostics Check	OK	14:12

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SHELTON,ERIC,L

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: SHELTON,ERIC

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: O325341-12

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 17:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: LAMALIE

DATE/TIME OF TEST: 05/21/2022 18:44:32

Diagnostics Check	OK	18:38
Air Blank	0.000	18:39
Control Test	0.080	18:39
Air Blank	0.000	18:39
Subject Sample # 1	0.000	18:40
Breath Volume	1.488	18:40
Air Blank	0.000	18:40
Air Blank	0.000	18:42
Subject Sample # 2	0.000	18:43
Breath Volume	1.492	18:43
Air Blank	0.000	18:43
Control Test	0.080	18:43
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VASQUEZ,GEORGE,D

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: HANSFORD,CARLDERIUS

ARRESTING OFFICER AGENCY: Ft. Walton Beach PD

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: O325341-12

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 03:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LOPEZ PALOMINO

DATE/TIME OF TEST: 05/22/2022 04:09:53

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.202	04:05
Breath Volume	1.964	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.207	04:08
Breath Volume	3.039	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 05/23/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HACKETT,TRAVIS
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: PINTO

DATE/TIME OF TEST: 05/22/2022 04:26:30

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.079	04:24	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Control Test	0.079	04:25	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 05/23/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: O325341-12
BREATH TEST OPERATOR: SHELTON,ERIC,L	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: WILLIAMS,SETH	ARREST TIME: 05:00
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD	GENDER: Male
LAST AGENCY INSPECTION: 04/27/2022	OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: BETANCOURTH GIRON	DATE/TIME OF TEST: 05/22/2022 06:22:32
--	--

Diagnostics Check	OK	06:19	
Air Blank	0.000	06:20	
Control Test	0.080	06:20	
Air Blank	0.000	06:20	
Subject Sample	# 1	06:21	Subject Test Refused
Breath Volume	0.000	06:21	
Air Blank	0.000	06:21	
Control Test	0.079	06:21	
Air Blank	0.000	06:22	
Diagnostics Check	OK	06:22	

Data Download Date/Time: 05/23/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,LARRY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ABEL,ANTHONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/23/2022 01:20:50

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.182	01:16
Breath Volume	1.128	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:18
Subject Sample # 2	0.184	01:19
Breath Volume	1.539	01:19
Air Blank	0.000	01:19
Control Test	0.079	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 05/23/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: HOLCOMB

DATE/TIME OF TEST: 05/23/2022 22:17:28

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:03	
Control Test	0.079	22:03	
Air Blank	0.000	22:03	
Subject Sample	# 1 0.331	22:06	Volume Not Met
Breath Volume	0.156	22:06	
Air Blank	0.000	22:07	
Air Blank	0.000	22:09	
Subject Sample	# 2 0.372	22:09	
Breath Volume	1.675	22:09	
Air Blank	0.000	22:10	
Air Blank	0.000	22:12	
Subject Sample	# 3 0.331	22:15	Volume Not Met
Breath Volume	1.007	22:15	
Air Blank	0.000	22:16	
Control Test	0.080	22:16	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:17	

Data Download Date/Time: 06/28/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MALLIN

DATE/TIME OF TEST: 05/28/2022 02:52:58

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.231	02:47
Breath Volume	1.890	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.219	02:51
Breath Volume	1.226	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 06/28/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: O325341-12
BREATH TEST OPERATOR: GLENECK,JORDAN,A	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GLENECK,JORDAN	ARREST TIME: 03:02
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Female
LAST AGENCY INSPECTION: 05/23/2022	OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: CLAY	DATE/TIME OF TEST: 05/28/2022 03:49:57
---------------------------	--

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.162	03:45
Breath Volume	1.554	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:47
Subject Sample # 2	0.149	03:48
Breath Volume	1.167	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 06/28/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLENECK,JORDAN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURKE,RYAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 05/28/2022 04:32:15

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:29	
Control Test	0.081	04:30	
Air Blank	0.000	04:30	
Subject Sample	# 1	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Control Test	0.079	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 06/28/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLENECK,JORDAN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURKE,RYAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-12
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 05/28/2022 04:50:52

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:38	
Control Test	0.081	04:39	
Air Blank	0.000	04:39	
Subject Sample	# 1 0.118	04:42	Volume Not Met
Breath Volume	0.410	04:42	
Air Blank	0.000	04:43	
Air Blank	0.000	04:45	
Subject Sample	# 2 0.118	04:46	
Breath Volume	1.179	04:46	
Air Blank	0.000	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 3 0.126	04:49	
Breath Volume	1.691	04:49	
Air Blank	0.000	04:49	
Control Test	0.079	04:50	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 06/28/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: O325341-12
BREATH TEST OPERATOR: GLENECK,JORDAN,A	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GLENECK,JORDAN	ARREST TIME: 04:45
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 05/23/2022	OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: VALVERI DATE/TIME OF TEST: 05/29/2022 05:42:02

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.081	05:34
Air Blank	0.000	05:35
Subject Sample # 1	0.218	05:37
Breath Volume	1.328	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:39
Subject Sample # 2	0.236	05:40
Breath Volume	1.265	05:40
Air Blank	0.000	05:41
Control Test	0.079	05:41
Air Blank	0.000	05:41
Diagnostics Check	OK	05:42

Data Download Date/Time: 06/28/2022 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LINCOLN,A,

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: LINCOLN,A

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: O325341-12

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 05:02

GENDER: Female

OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: SHIVERS

DATE/TIME OF TEST: 05/29/2022 06:16:30

Diagnostics Check	OK	06:13	
Air Blank	0.000	06:13	
Control Test	0.080	06:14	
Air Blank	0.000	06:14	
Subject Sample	# 1	06:15	Subject Test Refused
Breath Volume	0.000	06:15	
Air Blank	0.000	06:15	
Control Test	0.080	06:15	
Air Blank	0.000	06:16	
Diagnostics Check	OK	06:16	

Data Download Date/Time: 06/28/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: O325341-12
BREATH TEST OPERATOR: LAPEE,WYATT,T	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: LAPEE,WYATT	ARREST TIME: 06:58
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 05/23/2022	OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: FREGOLLE	DATE/TIME OF TEST: 05/29/2022	08:44:16
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	08:40	
Air Blank	0.000	08:41	
Control Test	0.081	08:41	
Air Blank	0.000	08:42	
Subject Sample	# 1	08:42	Subject Test Refused
Breath Volume	0.000	08:42	
Air Blank	0.000	08:43	
Control Test	0.080	08:43	
Air Blank	0.000	08:44	
Diagnostics Check	OK	08:44	

Data Download Date/Time: 06/28/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WARF,KYLE,J

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: WARF,KYLE

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: O325341-12

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 23:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: REDMOND

DATE/TIME OF TEST: 05/30/2022 00:24:13

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.134	00:17
Breath Volume	1.156	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.151	00:22
Breath Volume	1.500	00:22
Air Blank	0.000	00:23
Control Test	0.082	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 06/28/2022 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ESNEULM,MICHAEL,J

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: ESNEUL,MICHAEL

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: O325341-12

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 23:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: FREDERICK

DATE/TIME OF TEST: 05/30/2022 01:09:13

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.080	01:05	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:08	
Control Test	0.078	01:08	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 06/28/2022 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,ALLISON,T
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: LECUYER,JASON
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 05/14/2022 05:17:31

Diagnostics Check	OK	05:10
Air Blank	0.000	05:11
Control Test	0.076	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.175	05:13
Breath Volume	2.941	05:13
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.183	05:16
Breath Volume	2.757	05:16
Air Blank	0.000	05:16
Control Test	0.076	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 05/22/2022 20:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RINALDI,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: O'STEEN,SEAN
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: KNIGHTON

DATE/TIME OF TEST: 05/20/2022 04:05:52

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.078	03:58	
Air Blank	0.000	03:58	
Subject Sample	# 1	04:01	No Sample Provided
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 2	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.078	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 05/22/2022 20:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RINALDI,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: IMBODEN,WESLEY
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 05/20/2022 04:14:19

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:11	
Control Test	0.079	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:13	
Control Test	0.078	04:13	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 05/22/2022 20:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER, MONICA, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MILLENDER, MONICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/03/2022 21:18:58

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:13
Subject Sample # 1	0.164	21:14
Breath Volume	2.320	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.162	21:17
Breath Volume	2.417	21:17
Air Blank	0.000	21:18
Control Test	0.078	21:18
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 05/05/2022 07:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANNING,CARL,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MANNING,CARL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MCCLUNG

DATE/TIME OF TEST: 05/07/2022 23:10:30

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.080	23:03	
Air Blank	0.000	23:04	
Subject Sample	# 1 0.197	23:06	
Breath Volume	1.988	23:06	
Air Blank	0.000	23:07	
Air Blank	0.000	23:08	
Subject Sample	# 2 0.197	23:09	
Breath Volume	1.421	23:09	
Air Blank		23:10	RFI Detect
Air Blank	0.000	23:10	

Data Download Date/Time: 06/03/2022 14:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/09/2022 03:31:22

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.000	03:26
Breath Volume	2.066	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.000	03:30
Breath Volume	1.835	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 06/03/2022 14:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANNING,CARL,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MANNING,CARL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CAUSIE

DATE/TIME OF TEST: 05/29/2022 01:50:29

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.000	01:46
Breath Volume	1.210	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.000	01:49
Breath Volume	1.753	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 06/03/2022 14:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSTICE,JOSEPH,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JUSTICE,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: STALLWORTH

DATE/TIME OF TEST: 05/03/2022 17:34:41

Diagnostics Check	OK	17:28
Air Blank	0.000	17:29
Control Test	0.082	17:29
Air Blank	0.000	17:30
Subject Sample # 1	0.000	17:30
Breath Volume	1.691	17:30
Air Blank	0.000	17:30
Air Blank	0.000	17:32
Subject Sample # 2	0.000	17:33
Breath Volume	1.265	17:33
Air Blank	0.000	17:33
Control Test	0.082	17:34
Air Blank	0.000	17:34
Diagnostics Check	OK	17:34

Data Download Date/Time: 05/04/2022 07:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:32

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 05/07/2022 08:59:40

Diagnostics Check	OK	08:53
Air Blank	0.000	08:53
Control Test	0.082	08:53
Air Blank	0.000	08:54
Subject Sample # 1	0.125	08:54
Breath Volume	1.222	08:54
Air Blank	0.000	08:55
Air Blank	0.000	08:57
Subject Sample # 2	0.128	08:58
Breath Volume	1.140	08:58
Air Blank	0.000	08:58
Control Test	0.081	08:59
Air Blank	0.000	08:59
Diagnostics Check	OK	08:59

Data Download Date/Time: 06/03/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLGOOD,JEFFREY,S
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ALLGOOD,JEFFREY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: BUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: MESEROLL

DATE/TIME OF TEST: 05/07/2022 19:11:49

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.082	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.139	19:07
Breath Volume	2.156	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:09
Subject Sample # 2	0.132	19:10
Breath Volume	1.382	19:10
Air Blank	0.000	19:10
Control Test	0.081	19:11
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 06/03/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANNING,CARL,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MANNING,CARL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 10:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MCCLUNG

DATE/TIME OF TEST: 05/07/2022 22:59:40

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:57	
Control Test	0.082	22:58	
Air Blank	0.000	22:58	
Subject Sample	# 1	22:59	RFI Detect
Breath Volume	0.000	22:59	
Air Blank		22:59	RFI Detect
Air Blank		22:59	RFI Detect

Data Download Date/Time: 06/03/2022 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT, COLTEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALEXANDER, C
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 05/14/2022 03:59:14

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.082	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:58	
Control Test	0.081	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 06/03/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT,COLTEN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALEXANDER,C
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 05/14/2022 04:12:47

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.082	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.244	04:08
Breath Volume	1.515	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.259	04:11
Breath Volume	2.414	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 06/03/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHAR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BARGE

DATE/TIME OF TEST: 05/15/2022 22:27:52

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.081	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.000	22:23
Breath Volume	1.410	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.000	22:26
Breath Volume	1.691	22:26
Air Blank	0.000	22:26
Control Test	0.081	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 06/03/2022 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLSOM, DAVID, A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FOLSOM, DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 05/22/2022 23:27:14

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.081	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.238	23:22
Breath Volume	1.902	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.234	23:25
Breath Volume	1.523	23:25
Air Blank	0.000	23:26
Control Test	0.076	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 06/03/2022 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:59

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 05/23/2022 14:28:03

Diagnostics Check	OK	14:21
Air Blank	0.000	14:21
Control Test	0.082	14:22
Air Blank	0.000	14:22
Subject Sample # 1	0.000	14:23
Breath Volume	2.273	14:23
Air Blank	0.000	14:24
Air Blank	0.000	14:26
Subject Sample # 2	0.000	14:26
Breath Volume	1.820	14:26
Air Blank	0.000	14:27
Control Test	0.082	14:27
Air Blank	0.000	14:27
Diagnostics Check	OK	14:28

Data Download Date/Time: 06/03/2022 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: BODDA

DATE/TIME OF TEST: 05/26/2022 00:31:51

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.082	00:20	
Air Blank	0.000	00:21	
Subject Sample	# 1 0.091	00:24	Volume Not Met
Breath Volume	0.226	00:24	
Air Blank	0.000	00:25	
Air Blank	0.000	00:27	
Subject Sample	# 2 0.094	00:30	Volume Not Met
Breath Volume	0.015	00:30	
Air Blank	0.000	00:30	
Control Test	0.080	00:31	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 06/03/2022 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEIB,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: MCCARTY

DATE/TIME OF TEST: 05/27/2022 17:37:36

Diagnostics Check	OK	17:31
Air Blank	0.000	17:31
Control Test	0.082	17:32
Air Blank	0.000	17:32
Subject Sample # 1	0.205	17:33
Breath Volume	4.082	17:33
Air Blank	0.000	17:33
Air Blank	0.000	17:35
Subject Sample # 2	0.192	17:36
Breath Volume	3.445	17:36
Air Blank	0.000	17:36
Control Test	0.080	17:36
Air Blank	0.000	17:37
Diagnostics Check	OK	17:37

Data Download Date/Time: 06/03/2022 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: REASE

DATE/TIME OF TEST: 05/28/2022 01:50:26

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:47	
Control Test	0.081	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.081	01:49	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 06/03/2022 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLURD, TRACEY, L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BURKHEAD, ELI
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: BUI
ARREST TIME: 18:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: MCCLURD

DATE/TIME OF TEST: 05/28/2022 20:09:51

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:58	
Control Test	0.082	19:59	
Air Blank	0.000	19:59	
Subject Sample	# 1	20:02	No Sample Provided
Breath Volume	0.000	20:02	
Air Blank	0.000	20:03	
Air Blank	0.000	20:05	
Subject Sample	# 2	20:08	No Sample Provided
Breath Volume	0.000	20:08	
Air Blank	0.000	20:08	
Control Test	0.082	20:09	
Air Blank	0.000	20:09	
Diagnostics Check	OK	20:09	

Data Download Date/Time: 06/03/2022 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLGOOD,JEFFREY,S
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ALLGOOD,JEFFREY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: BUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: PENTON

DATE/TIME OF TEST: 05/28/2022 22:03:23

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:00	
Control Test	0.083	22:01	
Air Blank	0.000	22:01	
Subject Sample	# 1	22:02	Subject Test Refused
Breath Volume	0.000	22:02	
Air Blank	0.000	22:02	
Control Test	0.081	22:02	
Air Blank	0.000	22:03	
Diagnostics Check	OK	22:03	

Data Download Date/Time: 06/03/2022 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, TONY, D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BURKHEAD, ELI
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: BUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 05/29/2022 20:54:49

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:50	
Control Test	0.082	20:50	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:53	Subject Test Refused
Breath Volume	0.000	20:53	
Air Blank	0.000	20:53	
Control Test	0.082	20:54	
Air Blank	0.000	20:54	
Diagnostics Check	OK	20:54	

Data Download Date/Time: 06/03/2022 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGALADO,JORGE,I
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: REGALADO,JORGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 05/30/2022 04:18:37

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.082	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.174	04:13
Breath Volume	1.218	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.175	04:16
Breath Volume	1.433	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 06/03/2022 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 10:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:01

SUBJECT'S LAST NAME: TRAWICK

DATE/TIME OF TEST: 05/20/2022 11:33:34

Diagnostics Check	OK	11:21	
Air Blank	0.000	11:22	
Control Test	0.085	11:22	
Air Blank	0.000	11:23	
Subject Sample	# 1 0.170	11:26	Volume Not Met
Breath Volume	0.593	11:26	
Air Blank	0.000	11:27	
Air Blank	0.000	11:28	
Subject Sample	# 2 0.187	11:31	Volume Not Met
Breath Volume	0.175	11:31	
Air Blank	0.000	11:32	
Control Test	0.077	11:32	
Air Blank	0.000	11:33	
Diagnostics Check	OK	11:33	

Data Download Date/Time: 06/14/2022 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,JR,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SOLOMON,JR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: DRIVER

DATE/TIME OF TEST: 05/26/2022 21:20:09

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:11	
Control Test	0.080	21:11	
Air Blank	0.000	21:11	
Subject Sample	# 1	21:15	No Sample Provided
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Air Blank	0.000	21:17	
Subject Sample	# 2	21:18	Subject Test Refused
Breath Volume	0.000	21:18	
Air Blank	0.000	21:19	
Control Test	0.079	21:19	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 06/14/2022 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PASLEY,JACOB
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 08:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:39

SUBJECT'S LAST NAME: KEATOR

DATE/TIME OF TEST: 05/28/2022 11:02:53

Diagnostics Check	OK	10:59	
Air Blank	0.000	11:00	
Control Test	0.080	11:00	
Air Blank	0.000	11:00	
Subject Sample	# 1	11:01	Interferent Detect
Breath Volume	1.847	11:01	
Air Blank	0.000	11:01	
Control Test	0.081	11:02	
Air Blank	0.000	11:02	
Diagnostics Check	OK	11:02	

Data Download Date/Time: 06/14/2022 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYO,RICHARD,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WEBB,JAMES
ARRESTING OFFICER AGENCY: Chipley PD
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: STECENKO

DATE/TIME OF TEST: 05/28/2022 18:28:08

Diagnostics Check	OK	18:21
Air Blank	0.000	18:21
Control Test	0.079	18:22
Air Blank	0.000	18:22
Subject Sample # 1	0.231	18:23
Breath Volume	1.957	18:23
Air Blank	0.000	18:24
Air Blank	0.000	18:26
Subject Sample # 2	0.239	18:26
Breath Volume	2.734	18:26
Air Blank	0.000	18:27
Control Test	0.078	18:27
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 06/14/2022 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,ERIC,C
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: FOWLER,ERIC
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: BUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: NUNNELEE

DATE/TIME OF TEST: 05/28/2022 19:41:27

Diagnostics Check	OK	19:32
Air Blank	0.000	19:32
Control Test	0.078	19:32
Air Blank	0.000	19:33
Subject Sample # 1	0.180	19:33
Breath Volume	3.542	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.155	19:36
Breath Volume	1.890	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 3	0.175	19:39
Breath Volume	2.847	19:39
Air Blank	0.000	19:40
Control Test	0.077	19:40
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 06/14/2022 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,JERRY,D
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: PETERSON,GARRICK
ARRESTING OFFICER AGENCY: Chipley PD
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 05/29/2022 03:35:13

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.265	03:30
Breath Volume	1.410	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.274	03:33
Breath Volume	1.718	03:33
Air Blank	0.000	03:34
Control Test	0.075	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 06/14/2022 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,JERRY,D
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: PASLEY,JACOB
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/29/2022 21:43:31

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:38
Subject Sample # 1	0.000	21:39
Breath Volume	4.546	21:39
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.000	21:42
Breath Volume	1.464	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:42
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 06/14/2022 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO,KENNETH,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PRIETO,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: CLOWER

DATE/TIME OF TEST: 05/02/2022 13:16:22

Diagnostics Check	OK	13:10
Air Blank	0.000	13:10
Control Test	0.079	13:10
Air Blank	0.000	13:11
Subject Sample # 1	0.000	13:12
Breath Volume	2.039	13:12
Air Blank	0.000	13:12
Air Blank	0.000	13:14
Subject Sample # 2	0.000	13:15
Breath Volume	2.628	13:15
Air Blank	0.000	13:15
Control Test	0.079	13:15
Air Blank	0.000	13:16
Diagnostics Check	OK	13:16

Data Download Date/Time: 05/10/2022 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DUGAN,CHAD
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: NOONEY

DATE/TIME OF TEST: 05/04/2022 03:43:36

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.079	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample	# 2	03:42	No Sample Provided
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.079	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 05/10/2022 12:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TOBY,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CLEMENTS,TOBY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: RUMORE

DATE/TIME OF TEST: 05/04/2022 05:24:21

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:21	
Control Test	0.080	05:22	
Air Blank	0.000	05:22	
Subject Sample	# 1	05:22	Subject Test Refused
Breath Volume	0.000	05:22	
Air Blank	0.000	05:23	
Control Test	0.079	05:23	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 05/10/2022 12:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEYES,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KEYES,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/13/2022 20:01:19

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:56
Subject Sample # 1	0.017	19:56
Breath Volume	2.355	19:56
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.017	19:59
Breath Volume	3.148	19:59
Air Blank	0.000	20:00
Control Test	0.079	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 06/02/2022 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,DAVID,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HUGHES,DAVID
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: GLENN

DATE/TIME OF TEST: 05/14/2022 19:05:26

Diagnostics Check	OK	18:59
Air Blank	0.000	18:59
Control Test	0.079	18:59
Air Blank	0.000	19:00
Subject Sample # 1	0.120	19:00
Breath Volume	2.796	19:00
Air Blank	0.000	19:01
Air Blank	0.000	19:03
Subject Sample # 2	0.116	19:03
Breath Volume	3.351	19:03
Air Blank	0.000	19:04
Control Test	0.079	19:04
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

Data Download Date/Time: 06/02/2022 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WENK,ANTHONY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 05/15/2022 03:07:06

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:04	
Control Test	0.079	03:05	
Air Blank	0.000	03:05	
Subject Sample	# 1	03:05	Subject Test Refused
Breath Volume	0.000	03:05	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 06/02/2022 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,JORDI,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAEZ,JORDI
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/18/2022 04:05:00

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.080	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:04	
Control Test	0.079	04:04	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 06/02/2022 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CLAYTON,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CLAYTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 05/20/2022 17:03:17

Diagnostics Check	OK	16:51	
Air Blank	0.000	16:52	
Control Test	0.079	16:52	
Air Blank	0.000	16:53	
Subject Sample	# 1	16:56	No Sample Provided
Breath Volume	0.000	16:56	
Air Blank	0.000	16:56	
Air Blank	0.000	16:58	
Subject Sample	# 2	17:01	No Sample Provided
Breath Volume	0.000	17:01	
Air Blank	0.000	17:02	
Control Test	0.080	17:02	
Air Blank	0.000	17:03	
Diagnostics Check	OK	17:03	

Data Download Date/Time: 06/02/2022 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BIERMAN,CODY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: DIAZ CAMPOSECO

DATE/TIME OF TEST: 05/21/2022 00:36:20

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:35	
		00:35	RFI Detect
Air Blank	0.000	00:36	

Data Download Date/Time: 06/02/2022 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BIERMAN,CODY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: DIAZ CAMPOSECO

DATE/TIME OF TEST: 05/21/2022 00:44:36

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.263	00:40
Breath Volume	2.351	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.269	00:42
Breath Volume	3.050	00:42
Air Blank	0.000	00:43
Control Test	0.078	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 06/02/2022 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TOBY,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CLEMENTS,TOBY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: TUFENKIAN

DATE/TIME OF TEST: 05/22/2022 01:26:12

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.127	01:21
Breath Volume	1.277	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.129	01:24
Breath Volume	1.460	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 06/02/2022 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,JOSHUA,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CABRAL,JOHNATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: MISIAK

DATE/TIME OF TEST: 05/22/2022 18:21:23

Diagnostics Check	OK	18:14
Air Blank	0.000	18:14
Control Test	0.079	18:15
Air Blank	0.000	18:15
Subject Sample # 1	0.000	18:17
Breath Volume	5.710	18:17
Air Blank	0.000	18:17
Air Blank	0.000	18:19
Subject Sample # 2	0.000	18:20
Breath Volume	3.113	18:20
Air Blank	0.000	18:20
Control Test	0.079	18:20
Air Blank	0.000	18:21
Diagnostics Check	OK	18:21

Data Download Date/Time: 06/02/2022 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,JORDI,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CLEMENTS,TOBY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ORTIZ VIDAL

DATE/TIME OF TEST: 05/26/2022 03:13:39

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.190	03:09
Breath Volume	2.195	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.202	03:12
Breath Volume	1.339	03:12
Air Blank	0.000	03:12
Control Test	0.078	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 06/02/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TOBY,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CLEMENTS,TOBY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 05/26/2022 21:18:40

Diagnostics Check	OK	21:12
Air Blank	0.000	21:12
Control Test	0.079	21:12
Air Blank	0.000	21:13
Subject Sample # 1	0.172	21:13
Breath Volume	1.140	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.161	21:16
Breath Volume	1.605	21:16
Air Blank	0.000	21:17
Control Test	0.078	21:18
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 06/02/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, LEIGHA, M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: JONES, LEIGHA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: DUKES

DATE/TIME OF TEST: 05/27/2022 16:05:11

Diagnostics Check	OK	15:59
Air Blank	0.000	15:59
Control Test	0.079	16:00
Air Blank	0.000	16:00
Subject Sample # 1	0.000	16:00
Breath Volume	1.800	16:00
Air Blank	0.000	16:01
Air Blank	0.000	16:03
Subject Sample # 2	0.000	16:03
Breath Volume	1.937	16:03
Air Blank	0.000	16:04
Control Test	0.079	16:04
Air Blank	0.000	16:05
Diagnostics Check	OK	16:05

Data Download Date/Time: 06/02/2022 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,BRITTON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: DREW,BRITTON
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: BUI
ARREST TIME: 15:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: COMFORT

DATE/TIME OF TEST: 05/28/2022 16:43:33

Diagnostics Check	OK	16:37
Air Blank	0.000	16:38
Control Test	0.079	16:38
Air Blank	0.000	16:38
Subject Sample # 1	0.010	16:39
Breath Volume	2.304	16:39
Air Blank	0.000	16:39
Air Blank	0.000	16:41
Subject Sample # 2	0.007	16:42
Breath Volume	2.578	16:42
Air Blank	0.000	16:42
Control Test	0.076	16:42
Air Blank	0.000	16:43
Diagnostics Check	OK	16:43

Data Download Date/Time: 06/02/2022 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUIS,NORMAND,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: GOEN

DATE/TIME OF TEST: 05/03/2022 20:55:05

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:43	
Control Test	0.081	20:43	
Air Blank	0.000	20:44	
Subject Sample	# 1 0.415	20:44	No .020 agreement
Breath Volume	2.187	20:44	
Air Blank	0.000	20:45	
Air Blank	0.000	20:47	
Subject Sample	# 2 0.389	20:48	No .020 agreement
Breath Volume	1.992	20:48	
Air Blank	0.000	20:48	
Air Blank	0.000	20:50	
Subject Sample	# 3	20:53	No Sample Provided
Breath Volume	0.000	20:53	
Air Blank	0.000	20:54	
Control Test	0.079	20:54	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

Data Download Date/Time: 05/24/2022 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,GABRIEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: FLINN,AARON
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 05/06/2022 02:23:06

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.082	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:22	
Control Test	0.081	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 05/24/2022 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLONNA, TYLER, S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: COLONNA, TYLER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: KYPFER

DATE/TIME OF TEST: 05/09/2022 22:56:36

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.081	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.210	22:51
Breath Volume	1.496	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.208	22:54
Breath Volume	2.621	22:54
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 05/24/2022 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEYES,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PETERS,KRISTY
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: GALDAMEZ

DATE/TIME OF TEST: 05/22/2022 04:56:28

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.088	04:51
Breath Volume	1.597	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.095	04:54
Breath Volume	1.230	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 05/24/2022 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,GABRIEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: FLINN,AARON
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: AYALA AREVALO

DATE/TIME OF TEST: 05/30/2022 00:20:13

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.081	00:18	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:19	
Control Test	0.081	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 06/29/2022 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNING,CADEN,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BLAKE,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 06:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: BURDEN

DATE/TIME OF TEST: 05/30/2022 10:09:02

Diagnostics Check	OK	10:06	
Air Blank	0.000	10:06	
Control Test	0.082	10:06	
Air Blank	0.000	10:07	
Subject Sample	# 1	10:07	Subject Test Refused
Breath Volume	0.000	10:07	
Air Blank	0.000	10:08	
Control Test	0.081	10:08	
Air Blank	0.000	10:08	
Diagnostics Check	OK	10:09	

Data Download Date/Time: 06/29/2022 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNING,CADEN,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUPLCHIN,ANDREW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/31/2022 07:51:34

Diagnostics Check	OK	07:44
Air Blank	0.000	07:44
Control Test	0.080	07:44
Air Blank	0.000	07:45
Subject Sample # 1	0.091	07:46
Breath Volume	2.859	07:46
Air Blank	0.000	07:47
Air Blank	0.000	07:49
Subject Sample # 2	0.089	07:50
Breath Volume	3.230	07:50
Air Blank	0.000	07:50
Control Test	0.080	07:50
Air Blank	0.000	07:51
Diagnostics Check	OK	07:51

Data Download Date/Time: 06/29/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRS,RUSSELL,B
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: BARRS,RUSSELL
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 05/05/2022 10:21:15

Diagnostics Check	OK	10:15
Air Blank	0.000	10:15
Control Test	0.079	10:15
Air Blank	0.000	10:16
Subject Sample # 1	0.000	10:16
Breath Volume	2.500	10:16
Air Blank	0.000	10:17
Air Blank	0.000	10:19
Subject Sample # 2	0.000	10:19
Breath Volume	3.136	10:19
Air Blank	0.000	10:20
Control Test	0.078	10:20
Air Blank	0.000	10:21
Diagnostics Check	OK	10:21

Data Download Date/Time: 05/31/2022 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURR,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DURR,KYLE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 05/11/2022 00:53:25

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.245	00:48
Breath Volume	2.218	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.256	00:51
Breath Volume	2.679	00:51
Air Blank	0.000	00:52
Control Test	0.078	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 05/31/2022 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRBY, CHAD, W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KIRBY, CHAD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: CHAVARRIA-MARTINEZ

DATE/TIME OF TEST: 05/22/2022 18:48:28

Diagnostics Check	OK	18:38
Air Blank	0.000	18:39
Control Test	0.079	18:39
Air Blank	0.000	18:40
Subject Sample # 1	0.282	18:41
Breath Volume	1.156	18:41
Air Blank	0.000	18:41
Air Blank	0.000	18:43
Subject Sample # 2	0.246	18:43
Breath Volume	1.855	18:43
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 3	0.261	18:46
Breath Volume	1.898	18:46
Air Blank	0.000	18:47
Control Test	0.079	18:47
Air Blank	0.000	18:48
Diagnostics Check	OK	18:48

Data Download Date/Time: 05/31/2022 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,GERALD,V
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: ROUNTREE,JASON
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: HERNANDEZ-LOPEZ

DATE/TIME OF TEST: 05/27/2022 19:35:04

Diagnostics Check	OK	19:28
Air Blank	0.000	19:29
Control Test	0.079	19:29
Air Blank	0.000	19:29
Subject Sample # 1	0.262	19:30
Breath Volume	2.945	19:30
Air Blank	0.000	19:31
Air Blank	0.000	19:32
Subject Sample # 2	0.249	19:33
Breath Volume	2.878	19:33
Air Blank	0.000	19:34
Control Test	0.078	19:34
Air Blank	0.000	19:34
Diagnostics Check	OK	19:35

Data Download Date/Time: 05/31/2022 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,GERALD,V
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: JOHNSTON,JOSEPH
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 05/28/2022 20:33:33

Diagnostics Check	OK	20:27
Air Blank	0.000	20:27
Control Test	0.079	20:27
Air Blank	0.000	20:28
Subject Sample # 1	0.095	20:29
Breath Volume	2.351	20:29
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.096	20:31
Breath Volume	3.503	20:31
Air Blank	0.000	20:32
Control Test	0.078	20:32
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 05/31/2022 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBLE,LANCE,W
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: DAVENPORT,LUKE
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: BUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: GAMBLE

DATE/TIME OF TEST: 05/29/2022 17:07:11

Diagnostics Check	OK	17:03	
Air Blank	0.000	17:03	
Control Test	0.079	17:04	
Air Blank	0.000	17:04	
Subject Sample	# 1	17:05	Subject Test Refused
Breath Volume	0.000	17:05	
Air Blank	0.000	17:06	
Control Test	0.079	17:06	
Air Blank	0.000	17:07	
Diagnostics Check	OK	17:07	

Data Download Date/Time: 05/31/2022 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOONE, RODNEY, E
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: CLINE, MITCHELL
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: BUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 05/29/2022 22:25:26

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.206	22:17
Breath Volume	4.210	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.184	22:20
Breath Volume	1.574	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 3	0.207	22:23
Breath Volume	3.937	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 05/31/2022 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,ALBERT,W
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: KIRKLIN,LEYON
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 05/08/2022 02:32:01

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.078	02:18	
Air Blank	0.000	02:18	
Subject Sample	# 1 0.152	02:21	Volume Not Met
Breath Volume	0.097	02:21	
Air Blank	0.000	02:22	
Air Blank	0.000	02:24	
Subject Sample	# 2 0.156	02:24	
Breath Volume	1.863	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 3 0.149	02:30	Volume Not Met
Breath Volume	0.195	02:30	
Air Blank	0.000	02:31	
Control Test	0.071	02:31	Control Outside Tolerance
Air Blank	0.000	02:32	

Data Download Date/Time: 06/01/2022 21:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEGEL,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RIEGEL,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: FIGUEROA ACOSTA

DATE/TIME OF TEST: 05/08/2022 21:04:33

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.079	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.078	21:00
Breath Volume	2.367	21:00
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.074	21:02
Breath Volume	2.917	21:02
Air Blank	0.000	21:03
Control Test	0.076	21:03
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 06/01/2022 21:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCKLES,SCHIEFER,S
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: BUCKLES,SCHIEFER
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 05/22/2022 02:16:56

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.077	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.000	02:12
Breath Volume	3.742	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.000	02:15
Breath Volume	2.031	02:15
Air Blank	0.000	02:16
Control Test	0.075	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 06/01/2022 21:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAURES,BRANDYN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAURES,BRANDYN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/01/2022 18:46:49

Diagnostics Check	OK	18:38
Air Blank	0.000	18:39
Control Test	0.081	18:39
Air Blank	0.000	18:40
Subject Sample # 1	0.198	18:41
Breath Volume	1.429	18:41
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.198	18:45
Breath Volume	1.667	18:45
Air Blank	0.000	18:45
Control Test	0.080	18:46
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 05/27/2022 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRBY,CHAD,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KIRBY,CHAD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 201893
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: KOLAK

DATE/TIME OF TEST: 05/07/2022 19:08:16

Diagnostics Check	OK	18:56	
Air Blank	0.000	18:57	
Control Test	0.082	18:57	
Air Blank	0.000	18:57	
Subject Sample	# 1 0.000	19:01	Volume Not Met
Breath Volume	0.636	19:01	
Air Blank	0.000	19:01	
Air Blank	0.000	19:03	
Subject Sample	# 2 0.000	19:03	
Breath Volume	2.761	19:03	
Air Blank	0.000	19:04	
Air Blank	0.000	19:06	
Subject Sample	# 3 0.000	19:06	
Breath Volume	3.097	19:06	
Air Blank	0.000	19:07	
Control Test	0.081	19:07	
Air Blank	0.000	19:08	
Diagnostics Check	OK	19:08	

Data Download Date/Time: 05/27/2022 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WILLS

DATE/TIME OF TEST: 05/10/2022 01:28:42

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:24	
Control Test	0.081	01:24	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.081	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 05/27/2022 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: NASH,J
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: KEARNS

DATE/TIME OF TEST: 05/11/2022 22:28:58

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.082	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.256	22:23
Breath Volume	1.535	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.242	22:27
Breath Volume	1.742	22:27
Air Blank	0.000	22:27
Control Test	0.080	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 05/27/2022 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSKO,KRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 05/12/2022 01:29:26

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.080	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:28	Subject Test Refused
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Control Test	0.080	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 05/27/2022 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY, MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ROWAN

DATE/TIME OF TEST: 05/12/2022 21:29:47

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.081	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.009	21:25
Breath Volume	1.464	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.009	21:28
Breath Volume	1.851	21:28
Air Blank	0.000	21:28
Control Test	0.081	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 05/27/2022 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSKO,KRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: NAMIRR

DATE/TIME OF TEST: 05/13/2022 00:35:02

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:29	
Control Test	0.080	00:30	
Air Blank	0.000	00:30	
Subject Sample	# 1 0.218	00:33	Volume Not Met
Breath Volume	0.687	00:33	
Air Blank		00:34	Purge Fail
Air Blank	0.000	00:35	

Data Download Date/Time: 05/27/2022 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSKO,KRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: NAMIRR

DATE/TIME OF TEST: 05/13/2022 00:42:31

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.078	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1 0.213	00:41	Volume Not Met
Breath Volume	0.710	00:41	
Air Blank		00:42	Purge Fail
Air Blank	0.000	00:42	

Data Download Date/Time: 05/27/2022 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURR,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DURR,KYLE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 10:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: ROSENDALE

DATE/TIME OF TEST: 05/13/2022 11:38:59

Diagnostics Check	OK	11:32
Air Blank	0.000	11:33
Control Test	0.081	11:33
Air Blank	0.000	11:33
Subject Sample # 1	0.197	11:34
Breath Volume	2.214	11:34
Air Blank	0.000	11:35
Air Blank	0.000	11:36
Subject Sample # 2	0.188	11:37
Breath Volume	1.792	11:37
Air Blank	0.000	11:38
Control Test	0.078	11:38
Air Blank	0.000	11:38
Diagnostics Check	OK	11:38

Data Download Date/Time: 05/27/2022 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELAN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PHELAN,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MAYEUX

DATE/TIME OF TEST: 05/15/2022 21:43:42

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.082	21:37
Air Blank	0.000	21:37
Subject Sample # 1	0.096	21:38
Breath Volume	1.550	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.099	21:41
Breath Volume	1.257	21:41
Air Blank	0.000	21:42
Control Test	0.079	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 05/27/2022 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 05/18/2022 02:37:43

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.095	02:32
Breath Volume	1.222	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.100	02:36
Breath Volume	1.652	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 05/27/2022 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEATR,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: MATHEWS

DATE/TIME OF TEST: 05/24/2022 22:54:28

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.081	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.054	22:49
Breath Volume	2.429	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.054	22:52
Breath Volume	2.132	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 05/27/2022 17:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: JOHNS,E
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 04/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: GARICA

DATE/TIME OF TEST: 05/26/2022 00:28:56

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.081	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.000	00:24
Breath Volume	1.820	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.000	00:27
Breath Volume	3.218	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 05/27/2022 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNSED, DAVID, J
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: BURNSED, DAVID
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 201893
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: BUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 05/27/2022 22:24:27

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.145	22:19
Breath Volume	4.281	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.151	22:22
Breath Volume	3.628	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 06/28/2022 03:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PARISI,T
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: MESKO

DATE/TIME OF TEST: 05/28/2022 21:40:09

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:37	
Control Test	0.081	21:38	
Air Blank	0.000	21:38	
Subject Sample	# 1	21:38	Subject Test Refused
Breath Volume	0.000	21:38	
Air Blank	0.000	21:39	
Control Test	0.081	21:39	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

Data Download Date/Time: 06/28/2022 03:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 05/01/2022 04:29:48

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 1 0.189	04:28	
Breath Volume	2.406	04:28	
Air Blank		04:29	Purge Fail
Air Blank	0.000	04:29	

Data Download Date/Time: 05/20/2022 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 05/01/2022 04:38:57

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.078	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.192	04:34
Breath Volume	1.996	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:36
Subject Sample # 2	0.189	04:37
Breath Volume	1.949	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 05/20/2022 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 07:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:23

SUBJECT'S LAST NAME: ROSAS

DATE/TIME OF TEST: 05/05/2022 07:55:28

Diagnostics Check	OK	07:55	
Air Blank		07:55	RFI Detect
Air Blank		07:55	RFI Detect

Data Download Date/Time: 05/20/2022 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 07:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:23

SUBJECT'S LAST NAME: GENTRY

DATE/TIME OF TEST: 05/05/2022 08:05:21

Diagnostics Check	OK	07:58
Air Blank	0.000	07:58
Control Test	0.081	07:58
Air Blank	0.000	07:59
Subject Sample # 1	0.270	08:00
Breath Volume	2.507	08:00
Air Blank	0.000	08:01
Air Blank	0.000	08:02
Subject Sample # 2	0.255	08:03
Breath Volume	2.386	08:03
Air Blank	0.000	08:04
Control Test	0.077	08:04
Air Blank	0.000	08:05
Diagnostics Check	OK	08:05

Data Download Date/Time: 05/20/2022 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PEREZ GONZALEZ

DATE/TIME OF TEST: 05/07/2022 23:20:03

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:10	
Control Test	0.081	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1	23:14	No Sample Provided
Breath Volume	0.000	23:14	
Air Blank	0.000	23:15	
Air Blank	0.000	23:17	
Subject Sample	# 2	23:18	Subject Test Refused
Breath Volume	0.000	23:18	
Air Blank	0.000	23:19	
Control Test	0.081	23:19	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 05/20/2022 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/08/2022 16:26:55

Diagnostics Check	OK	16:20	
Air Blank	0.000	16:21	
Control Test	0.080	16:21	
Air Blank	0.000	16:22	
Subject Sample	# 1 0.313	16:25	Volume Not Met
Breath Volume	0.019	16:25	
Air Blank		16:26	Purge Fail
Air Blank	0.000	16:26	

Data Download Date/Time: 05/20/2022 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/08/2022 16:31:25

Diagnostics Check	OK	16:30	
Air Blank	0.000	16:30	
Control Test	0.071	16:30	Control Outside Tolerance
Air Blank	0.000	16:31	

Data Download Date/Time: 05/20/2022 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/08/2022 16:40:12

Diagnostics Check	OK	16:34	
Air Blank	0.000	16:34	
Control Test	0.076	16:35	
Air Blank	0.000	16:35	
Subject Sample	# 1 0.283	16:38	Volume Not Met
Breath Volume	0.046	16:38	
Air Blank		16:39	Purge Fail
Air Blank	0.000	16:40	

Data Download Date/Time: 05/20/2022 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 05/11/2022 17:33:03

Diagnostics Check	OK	17:26
Air Blank	0.000	17:26
Control Test	0.082	17:27
Air Blank	0.000	17:27
Subject Sample # 1	0.242	17:28
Breath Volume	1.839	17:28
Air Blank	0.000	17:29
Air Blank	0.000	17:30
Subject Sample # 2	0.243	17:31
Breath Volume	1.933	17:31
Air Blank	0.000	17:32
Control Test	0.079	17:32
Air Blank	0.000	17:32
Diagnostics Check	OK	17:33

Data Download Date/Time: 05/20/2022 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,BRETT,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WAGNER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MINOR

DATE/TIME OF TEST: 05/15/2022 00:45:19

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.082	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.235	00:40
Breath Volume	1.742	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:42
Subject Sample # 2	0.230	00:43
Breath Volume	1.960	00:43
Air Blank	0.000	00:44
Control Test	0.078	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 05/20/2022 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RADCLIFF, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: ABSHER

DATE/TIME OF TEST: 05/24/2022 18:33:19

Diagnostics Check	OK	18:30	
Air Blank	0.000	18:30	
Control Test	0.082	18:31	
Air Blank	0.000	18:31	
Subject Sample	# 1	18:31	Subject Test Refused
Breath Volume	0.000	18:31	
Air Blank	0.000	18:32	
Control Test	0.082	18:32	
Air Blank	0.000	18:33	
Diagnostics Check	OK	18:33	

Data Download Date/Time: 06/03/2022 15:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: BERNARDES

DATE/TIME OF TEST: 05/27/2022 23:36:10

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.081	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Control Test	0.080	23:35	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 06/03/2022 15:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: STAGGS

DATE/TIME OF TEST: 05/29/2022 00:58:50

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.082	00:56	
Air Blank	0.000	00:57	
Subject Sample	# 1	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Control Test	0.081	00:58	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 06/03/2022 15:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COX,DAVID,S

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: PASLEY,JACOB

ARRESTING OFFICER AGENCY: Washington County SO

LAST AGENCY INSPECTION: 04/09/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1206004

DRY GAS EXPIRY DATE: 09/20/2022

VIOLATION CODE: DUI

ARREST TIME: 01:35

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: SCHARMEN

DATE/TIME OF TEST: 05/02/2022 02:48:16

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.000	02:44
Breath Volume	1.660	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.000	02:46
Breath Volume	1.472	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 05/20/2022 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: PRATHER

DATE/TIME OF TEST: 05/03/2022 21:58:57

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.052	21:54
Breath Volume	1.730	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:56
Subject Sample # 2	0.051	21:57
Breath Volume	1.660	21:57
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 05/20/2022 15:01

Subject Test Electronic Data

DATE/TIME OF TEST: 05/06/2022 23:12:27

Subject Test Electronic Data

DATE/TIME OF TEST: 05/08/2022 03:27:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,JOHNNY,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WHITAKER,JOHNNY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: STANTON

DATE/TIME OF TEST: 05/08/2022 22:17:28

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.051	22:12
Breath Volume	2.707	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.048	22:15
Breath Volume	1.167	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 05/20/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,JOHNNY,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WHITAKER,JOHNNY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 05/11/2022 18:06:55

Diagnostics Check	OK	18:04	
Air Blank	0.000	18:04	
Control Test	0.080	18:04	
Air Blank	0.000	18:05	
Subject Sample	# 1	18:05	RFI Detect
Breath Volume	0.000	18:05	
Air Blank	0.000	18:06	
Control Test	0.080	18:06	
Air Blank	0.000	18:06	
Diagnostics Check	OK	18:06	

Data Download Date/Time: 05/20/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,JOHNNY,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WHITAKER,JOHNNY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 05/11/2022 18:17:06

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.080	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.169	18:12
Breath Volume	1.210	18:12
Air Blank	0.000	18:13
Air Blank	0.000	18:14
Subject Sample # 2	0.156	18:15
Breath Volume	2.437	18:15
Air Blank	0.000	18:16
Control Test	0.080	18:16
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

Data Download Date/Time: 05/20/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,BRANDON,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FRENCH,JASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 05/14/2022 18:55:32

Diagnostics Check	OK	18:44	
Air Blank	0.000	18:44	
Control Test	0.081	18:44	
Air Blank	0.000	18:45	
Subject Sample	# 1	18:48	No Sample Provided
Breath Volume	0.000	18:48	
Air Blank	0.000	18:49	
Air Blank	0.000	18:51	
Subject Sample	# 2	18:54	No Sample Provided
Breath Volume	0.000	18:54	
Air Blank	0.000	18:54	
Control Test	0.080	18:54	
Air Blank	0.000	18:55	
Diagnostics Check	OK	18:55	

Data Download Date/Time: 05/20/2022 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARDY,JOE,E

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: HARDY,JOE

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 04/09/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1206004

DRY GAS EXPIRY DATE: 09/20/2022

VIOLATION CODE: DUI

ARREST TIME: 17:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 05/15/2022 20:53:44

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:49	
Control Test	0.080	20:49	
Air Blank	0.000	20:50	
Subject Sample	# 1 0.261	20:52	
Breath Volume	1.468	20:52	
Air Blank		20:52	Purge Fail
Air Blank		20:53	Purge Fail

Data Download Date/Time: 05/20/2022 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARDY,JOE,E

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: HARDY,JOE

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 04/09/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1206004

DRY GAS EXPIRY DATE: 09/20/2022

VIOLATION CODE: DUI

ARREST TIME: 17:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 05/15/2022 20:58:15

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:57	
Control Test	0.038	20:57	Control Outside Tolerance
Air Blank	0.000	20:58	

Data Download Date/Time: 05/20/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/13/2022 02:23:09

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.078	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.184	02:18
Breath Volume	2.355	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.183	02:21
Breath Volume	1.394	02:21
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 06/08/2022 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SHAFFER,ROBIN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: PURDY

DATE/TIME OF TEST: 05/14/2022 00:30:17

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.077	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.000	00:25
Breath Volume	2.269	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.000	00:28
Breath Volume	3.156	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 06/08/2022 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: WATSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: SCHIERER

DATE/TIME OF TEST: 05/28/2022 16:00:13

Diagnostics Check	OK	15:53
Air Blank	0.000	15:54
Control Test	0.078	15:54
Air Blank	0.000	15:55
Subject Sample # 1	0.155	15:55
Breath Volume	2.507	15:55
Air Blank	0.000	15:56
Air Blank	0.000	15:58
Subject Sample # 2	0.147	15:58
Breath Volume	2.878	15:58
Air Blank	0.000	15:59
Control Test	0.079	15:59
Air Blank	0.000	16:00
Diagnostics Check	OK	16:00

Data Download Date/Time: 06/08/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON, WESLEY, A
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: HENDERSON, WESLEY
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/07/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 05/29/2022 16:26:05

Diagnostics Check	OK	16:21	
Air Blank	0.000	16:21	
Control Test	0.077	16:22	
Air Blank	0.000	16:22	
Subject Sample # 1	0.000	16:24	
Breath Volume	2.417	16:24	
Air Blank		16:25	Purge Fail
Air Blank	0.000	16:26	

Data Download Date/Time: 06/08/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON, WESLEY, A
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: HENDERSON, WESLEY
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 05/29/2022 16:36:05

Diagnostics Check	OK	16:29
Air Blank	0.000	16:30
Control Test	0.077	16:30
Air Blank	0.000	16:31
Subject Sample # 1	0.000	16:31
Breath Volume	3.562	16:31
Air Blank	0.000	16:32
Air Blank	0.000	16:34
Subject Sample # 2	0.000	16:34
Breath Volume	4.605	16:34
Air Blank	0.000	16:35
Control Test	0.079	16:35
Air Blank	0.000	16:35
Diagnostics Check	OK	16:36

Data Download Date/Time: 06/08/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS, DAVID, A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HICKS, JASON
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: PRATHER

DATE/TIME OF TEST: 05/02/2022 23:44:06

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:33	
Control Test	0.079	23:33	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:37	No Sample Provided
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample	# 2	23:42	No Sample Provided
Breath Volume	0.000	23:42	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 05/23/2022 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: GARZA-MCLEAN

DATE/TIME OF TEST: 05/14/2022 03:53:01

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.079	03:42	
Air Blank	0.000	03:42	
Subject Sample	# 1 0.261	03:45	Volume Not Met
Breath Volume	0.972	03:45	
Air Blank	0.000	03:46	
Air Blank	0.000	03:48	
Subject Sample	# 2 0.224	03:51	Volume Not Met
Breath Volume	0.945	03:51	
Air Blank	0.000	03:52	
Control Test	0.075	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 05/23/2022 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: GARZA-MCLEAN

DATE/TIME OF TEST: 05/14/2022 04:01:30

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:55	
Control Test	0.077	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 1 0.273	03:56	
Breath Volume	1.242	03:56	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 2	03:59	Slope Not Met
Breath Volume	0.886	03:59	
Air Blank	0.000	04:00	
Control Test	0.077	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 05/23/2022 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RICKY,.
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PEREZ,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: BENDER

DATE/TIME OF TEST: 05/08/2022 16:01:18

Diagnostics Check	OK	15:52
Air Blank	0.000	15:53
Control Test	0.080	15:53
Air Blank	0.000	15:53
Subject Sample # 1	0.000	15:55
Breath Volume	1.625	15:55
Air Blank	0.000	15:55
Air Blank	0.000	15:57
Subject Sample # 2	0.000	15:59
Breath Volume	1.347	15:59
Air Blank	0.000	16:00
Control Test	0.080	16:00
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

Data Download Date/Time: 05/21/2022 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,ERIC,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GREGORY,OMAR
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:52

SUBJECT'S LAST NAME: NICKS

DATE/TIME OF TEST: 05/10/2022 13:26:41

Diagnostics Check	OK	13:20
Air Blank	0.000	13:21
Control Test	0.079	13:21
Air Blank	0.000	13:21
Subject Sample # 1	0.357	13:22
Breath Volume	2.832	13:22
Air Blank	0.000	13:22
Air Blank	0.000	13:24
Subject Sample # 2	0.358	13:25
Breath Volume	3.425	13:25
Air Blank	0.000	13:25
Control Test	0.078	13:26
Air Blank	0.000	13:26
Diagnostics Check	OK	13:26

Data Download Date/Time: 05/21/2022 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRCHNER,MATTHEW,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: FORMELIO,J
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: HARRY

DATE/TIME OF TEST: 05/13/2022 21:05:54

Diagnostics Check	OK	20:59
Air Blank	0.000	20:59
Control Test	0.079	21:00
Air Blank	0.000	21:00
Subject Sample # 1	0.161	21:01
Breath Volume	1.953	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:03
Subject Sample # 2	0.171	21:04
Breath Volume	2.714	21:04
Air Blank	0.000	21:05
Control Test	0.078	21:05
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 05/21/2022 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA, ADAEL,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: COLGAN, ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/21/2022 21:35:17

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.080	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.158	21:30
Breath Volume	1.945	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:32
Subject Sample # 2	0.165	21:33
Breath Volume	1.929	21:33
Air Blank	0.000	21:34
Control Test	0.078	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 06/30/2022 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYFORD,GREGG,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MOJICA,JASMINE
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: REILLY

DATE/TIME OF TEST: 05/22/2022 18:57:36

Diagnostics Check	OK	18:51
Air Blank	0.000	18:51
Control Test	0.080	18:52
Air Blank	0.000	18:52
Subject Sample # 1	0.173	18:53
Breath Volume	2.671	18:53
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.177	18:56
Breath Volume	1.828	18:56
Air Blank	0.000	18:56
Control Test	0.079	18:57
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 06/30/2022 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPLEBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: PIERREVIL

DATE/TIME OF TEST: 05/23/2022 05:43:01

Diagnostics Check	OK	05:36
Air Blank	0.000	05:37
Control Test	0.080	05:37
Air Blank	0.000	05:37
Subject Sample # 1	0.090	05:38
Breath Volume	1.820	05:38
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample # 2	0.092	05:41
Breath Volume	2.117	05:41
Air Blank	0.000	05:42
Control Test	0.079	05:42
Air Blank	0.000	05:42
Diagnostics Check	OK	05:43

Data Download Date/Time: 06/30/2022 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GHOBRIAL,MARTIN,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: ZAMFINO,ANDREW
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: MUSELLA

DATE/TIME OF TEST: 05/26/2022 20:20:54

Diagnostics Check	OK	20:14
Air Blank	0.000	20:15
Control Test	0.079	20:15
Air Blank	0.000	20:15
Subject Sample # 1	0.288	20:16
Breath Volume	2.218	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:18
Subject Sample # 2	0.279	20:19
Breath Volume	2.007	20:19
Air Blank	0.000	20:20
Control Test	0.077	20:20
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 06/30/2022 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRCHNER,MATTHEW,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: THERIAULT,THEDDY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: VINITSKY

DATE/TIME OF TEST: 05/29/2022 22:18:30

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.079	22:15	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:17	Sequence Aborted
Breath Volume	0.152	22:17	
Air Blank	0.000	22:18	

Data Download Date/Time: 06/30/2022 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,KYLE,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PATTERSON,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: HEEN

DATE/TIME OF TEST: 05/30/2022 17:08:43

Diagnostics Check	OK	17:02
Air Blank	0.000	17:02
Control Test	0.079	17:03
Air Blank	0.000	17:03
Subject Sample # 1	0.183	17:04
Breath Volume	2.917	17:04
Air Blank	0.000	17:04
Air Blank	0.000	17:06
Subject Sample # 2	0.186	17:07
Breath Volume	3.000	17:07
Air Blank	0.000	17:07
Control Test	0.079	17:08
Air Blank	0.000	17:08
Diagnostics Check	OK	17:08

Data Download Date/Time: 06/30/2022 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERBER, OLIVER, B
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: JARRELL, CHRISTIAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GALANOS

DATE/TIME OF TEST: 05/20/2022 23:06:02

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.093	23:01
Breath Volume	3.277	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.093	23:04
Breath Volume	3.003	23:04
Air Blank	0.000	23:05
Control Test	0.078	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:06

Data Download Date/Time: 06/10/2022 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,ANDREW,H
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: JARRELL,CHRISTIAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 05/29/2022 19:40:07

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.079	19:34
Air Blank	0.000	19:35
Subject Sample # 1	0.092	19:35
Breath Volume	3.109	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.092	19:38
Breath Volume	4.074	19:38
Air Blank	0.000	19:39
Control Test	0.079	19:39
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 06/10/2022 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIPFORD,JASON,V
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LIPFORD,JASON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 05/29/2022 21:26:42

Diagnostics Check	OK	21:23	
Air Blank	0.000	21:24	
Control Test	0.078	21:24	
Air Blank	0.000	21:24	
Subject Sample	# 1	21:26	Sequence Aborted
Breath Volume	0.000	21:26	
Air Blank	0.000	21:26	

Data Download Date/Time: 06/10/2022 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GODBLOT

DATE/TIME OF TEST: 05/19/2022 21:23:15

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:20	
Control Test	0.080	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:21	Subject Test Refused
Breath Volume	0.000	21:21	
Air Blank	0.000	21:22	
Control Test	0.079	21:22	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 06/06/2022 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTIS,LAYNE,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: CURTIS,LAYNE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 05/04/2022 02:47:17

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.079	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:40	No Sample Provided
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2	02:45	No Sample Provided
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 05/11/2022 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,COREY,H
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: STONE,COREY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: CHENETTE

DATE/TIME OF TEST: 05/06/2022 00:52:56

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.081	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:45	No Sample Provided
Breath Volume	0.000	00:45	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 2	00:51	No Sample Provided
Breath Volume	0.000	00:51	
Air Blank	0.000	00:52	
Control Test	0.081	00:52	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 05/11/2022 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAXTON,RYAN,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PAXTON,RYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 05/06/2022 23:38:25

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.175	23:33
Breath Volume	2.429	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.171	23:36
Breath Volume	2.148	23:36
Air Blank	0.000	23:37
Control Test	0.078	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 05/11/2022 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 05/07/2022 00:13:46

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:11	
Control Test	0.080	00:11	
Air Blank	0.000	00:12	
Subject Sample	# 1	00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:12	
Control Test	0.079	00:13	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 05/11/2022 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RUSSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: OSBOURNE

DATE/TIME OF TEST: 05/07/2022 03:20:01

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.080	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:13	No Sample Provided
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample	# 2	03:18	No Sample Provided
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 05/11/2022 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,AUSTIN,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COON,AUSTIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: HIBBERT

DATE/TIME OF TEST: 05/07/2022 15:18:31

Diagnostics Check	OK	15:15	
Air Blank	0.000	15:16	
Control Test	0.080	15:16	
Air Blank	0.000	15:16	
Subject Sample	# 1	15:17	Subject Test Refused
Breath Volume	0.000	15:17	
Air Blank	0.000	15:17	
Control Test	0.080	15:17	
Air Blank	0.000	15:18	
Diagnostics Check	OK	15:18	

Data Download Date/Time: 05/11/2022 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSNYAK,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOSNYAK,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: YORDAN

DATE/TIME OF TEST: 05/07/2022 22:38:03

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.211	22:33
Breath Volume	3.046	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:35
Subject Sample # 2	0.212	22:36
Breath Volume	3.113	22:36
Air Blank	0.000	22:37
Control Test	0.078	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:38

Data Download Date/Time: 05/11/2022 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 08:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:33

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 05/10/2022 10:00:40

Diagnostics Check	OK	09:54
Air Blank	0.000	09:54
Control Test	0.081	09:55
Air Blank	0.000	09:55
Subject Sample # 1	0.000	09:56
Breath Volume	2.664	09:56
Air Blank	0.000	09:57
Air Blank	0.000	09:59
Subject Sample # 2	0.000	09:59
Breath Volume	2.582	09:59
Air Blank	0.000	09:59
Control Test	0.080	10:00
Air Blank	0.000	10:00
Diagnostics Check	OK	10:00

Data Download Date/Time: 05/11/2022 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,BRITTANY,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FISHER,BRITTANY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: STENGER

DATE/TIME OF TEST: 05/11/2022 01:17:28

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:10	No Sample Provided
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample	# 2	01:16	No Sample Provided
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.080	01:16	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 05/11/2022 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: LAMONES

DATE/TIME OF TEST: 05/21/2022 04:13:20

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:05	No Sample Provided
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Air Blank	0.000	04:07	
Subject Sample	# 2	0.191	04:08
Breath Volume	1.941	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample	# 3	0.180	04:11
Breath Volume	1.886	04:11	
Air Blank	0.000	04:12	
Control Test	0.078	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 06/15/2022 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SHIPMAN

DATE/TIME OF TEST: 05/22/2022 04:20:59

Diagnostics Check	Fail	04:20	Voltage/Current Test
Air Blank	0.000	04:20	

Data Download Date/Time: 06/15/2022 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SHIPMAN

DATE/TIME OF TEST: 05/22/2022 04:33:15

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.080	04:23	
Air Blank	0.000	04:23	
Subject Sample	# 1 0.156	04:27	Volume Not Met
Breath Volume	0.121	04:27	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample	# 2	04:31	RFI Detect
Breath Volume	0.058	04:31	
Air Blank	0.000	04:32	
Control Test	0.079	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 06/15/2022 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HECKMAN,SHELBY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LAGUNES MARTINEZ

DATE/TIME OF TEST: 05/22/2022 04:37:29

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.078	04:36	RFI Detect
Air Blank	0.000	04:37	

Data Download Date/Time: 06/15/2022 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HECKMAN,SHELBY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LAGUNES MARTINEZ

DATE/TIME OF TEST: 05/22/2022 04:39:57

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
		04:39	RFI Detect
Air Blank		04:39	RFI Detect

Data Download Date/Time: 06/15/2022 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HECKMAN,SHELBY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LAGUNES MARTINEZ

DATE/TIME OF TEST: 05/22/2022 04:41:49

Diagnostics Check	OK	04:41	
Air Blank		04:41	RFI Detect
Air Blank		04:41	RFI Detect

Data Download Date/Time: 06/15/2022 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 05/29/2022 02:36:44

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.147	02:32
Breath Volume	1.414	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.140	02:35
Breath Volume	1.500	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 06/15/2022 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LETSON,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: DARLINGTON

DATE/TIME OF TEST: 05/04/2022 01:52:49

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.078	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.000	01:48
Breath Volume	1.128	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.000	01:51
Breath Volume	1.507	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BRAVE,STEVE
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: BACON

DATE/TIME OF TEST: 05/05/2022 20:07:58

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.080	20:02
Air Blank	0.000	20:02
Subject Sample # 1	0.116	20:03
Breath Volume	1.410	20:03
Air Blank	0.000	20:04
Air Blank	0.000	20:05
Subject Sample # 2	0.118	20:06
Breath Volume	1.750	20:06
Air Blank	0.000	20:07
Control Test	0.080	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TAPANES,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: VANDENBERGHE

DATE/TIME OF TEST: 05/06/2022 23:29:40

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.079	23:22
Air Blank	0.000	23:22
Subject Sample # 1	0.150	23:24
Breath Volume	2.191	23:24
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.142	23:28
Breath Volume	2.152	23:28
Air Blank	0.000	23:28
Control Test	0.079	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LETSON,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: BEARDSLEY

DATE/TIME OF TEST: 05/08/2022 00:39:39

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.115	00:34
Breath Volume	1.980	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.116	00:38
Breath Volume	2.703	00:38
Air Blank	0.000	00:38
Control Test	0.079	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMANN,HANS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 05/09/2022 22:56:38

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.081	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.099	22:52
Breath Volume	1.468	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.100	22:55
Breath Volume	1.136	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMANN,HANS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/10/2022 20:48:41

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:46	
Control Test	0.079	20:46	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:47	Subject Test Refused
Breath Volume	0.000	20:47	
Air Blank	0.000	20:47	
Control Test	0.080	20:48	
Air Blank	0.000	20:48	
Diagnostics Check	OK	20:48	

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HOYT,DANIELLE
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 05/14/2022 02:23:27

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.169	02:18
Breath Volume	1.714	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.167	02:21
Breath Volume	1.636	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMANN,HANS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: HANRAHAN

DATE/TIME OF TEST: 05/15/2022 02:19:30

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.000	02:15
Breath Volume	1.988	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.000	02:18
Breath Volume	1.585	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMANN,HANES
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: FRANK

DATE/TIME OF TEST: 05/15/2022 19:02:16

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.079	18:56
Air Blank	0.000	18:56
Subject Sample # 1	0.194	18:57
Breath Volume	2.050	18:57
Air Blank	0.000	18:58
Air Blank	0.000	18:59
Subject Sample # 2	0.192	19:00
Breath Volume	2.062	19:00
Air Blank	0.000	19:01
Control Test	0.079	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMANN,HANES
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: FRANK

DATE/TIME OF TEST: 05/15/2022 19:50:32

Diagnostics Check	OK	19:43
Air Blank	0.000	19:44
Control Test	0.081	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.172	19:45
Breath Volume	1.660	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.166	19:48
Breath Volume	1.777	19:48
Air Blank	0.000	19:49
Control Test	0.079	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SILVA,JEAN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: MOCERINO

DATE/TIME OF TEST: 05/16/2022 16:49:11

Diagnostics Check	OK	16:42
Air Blank	0.000	16:43
Control Test	0.080	16:43
Air Blank	0.000	16:44
Subject Sample # 1	0.000	16:44
Breath Volume	1.839	16:44
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.000	16:47
Breath Volume	1.765	16:47
Air Blank	0.000	16:48
Control Test	0.081	16:48
Air Blank	0.000	16:49
Diagnostics Check	OK	16:49

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: SYSK

DATE/TIME OF TEST: 05/16/2022 18:57:05

Diagnostics Check	OK	18:50
Air Blank	0.000	18:50
Control Test	0.079	18:51
Air Blank	0.000	18:51
Subject Sample # 1	0.000	18:52
Breath Volume	2.832	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.000	18:55
Breath Volume	1.371	18:55
Air Blank	0.000	18:56
Control Test	0.079	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WARE,G
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: BARRY

DATE/TIME OF TEST: 05/16/2022 20:05:38

Diagnostics Check	OK	19:53	
Air Blank	0.000	19:53	
Control Test	0.080	19:54	
Air Blank	0.000	19:54	
Subject Sample	# 1 0.054	19:57	Volume Not Met
Breath Volume	0.597	19:57	
Air Blank	0.000	19:58	
Air Blank	0.000	20:00	
Subject Sample	# 2 0.071	20:00	
Breath Volume	1.566	20:00	
Air Blank	0.000	20:01	
Air Blank	0.000	20:03	
Subject Sample	# 3 0.065	20:04	
Breath Volume	1.843	20:04	
Air Blank	0.000	20:04	
Control Test	0.079	20:05	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 05/23/2022 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN, DEMERIS, P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WELLMAN, A
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 05/17/2022 21:33:14

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:22	
Control Test	0.080	21:22	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:26	No Sample Provided
Breath Volume	0.000	21:26	
Air Blank	0.000	21:26	
Air Blank	0.000	21:28	
Subject Sample	# 2	21:31	Volume Not Met
Breath Volume	0.878	21:31	
Air Blank	0.000	21:32	
Control Test	0.080	21:32	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 05/23/2022 05:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN, DEMERIS, P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WELLMAN, A
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 05/17/2022 21:48:38

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.078	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1 0.196	21:37	
Breath Volume	1.433	21:37	
Air Blank	0.000	21:38	
Air Blank	0.000	21:40	
Subject Sample	# 2 0.183	21:43	Volume Not Met
Breath Volume	0.703	21:43	
Air Blank	0.000	21:44	
Air Blank	0.000	21:46	
Subject Sample	# 3 0.185	21:47	
Breath Volume	2.457	21:47	
Air Blank	0.000	21:47	
Control Test	0.077	21:48	
Air Blank	0.000	21:48	
Diagnostics Check	OK	21:48	

Data Download Date/Time: 05/23/2022 05:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALE,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ABBOTT,M
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: MORMAN

DATE/TIME OF TEST: 05/20/2022 09:53:45

Diagnostics Check	OK	09:47
Air Blank	0.000	09:47
Control Test	0.079	09:48
Air Blank	0.000	09:48
Subject Sample # 1	0.100	09:49
Breath Volume	1.417	09:49
Air Blank	0.000	09:49
Air Blank	0.000	09:51
Subject Sample # 2	0.098	09:52
Breath Volume	1.523	09:52
Air Blank	0.000	09:52
Control Test	0.080	09:53
Air Blank	0.000	09:53
Diagnostics Check	OK	09:53

Data Download Date/Time: 05/23/2022 05:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAUL,N
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMANN,H
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: YANEZ

DATE/TIME OF TEST: 05/22/2022 01:26:22

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.080	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.094	01:21
Breath Volume	2.488	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.089	01:24
Breath Volume	2.296	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 05/23/2022 05:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCASTER,CODY,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: PAQUETTE

DATE/TIME OF TEST: 05/22/2022 15:21:27

Diagnostics Check	OK	15:09
Air Blank	0.000	15:09
Control Test	0.080	15:10
Air Blank	0.000	15:10
Subject Sample # 1	0.296	15:12
Breath Volume	1.761	15:12
Air Blank	0.000	15:13
Air Blank	0.000	15:14
Subject Sample # 2	0.319	15:15
Breath Volume	2.359	15:15
Air Blank	0.000	15:16
Air Blank	0.000	15:18
Subject Sample # 3	0.321	15:19
Breath Volume	2.367	15:19
Air Blank	0.000	15:20
Control Test	0.077	15:20
Air Blank	0.000	15:21
Diagnostics Check	OK	15:21

Data Download Date/Time: 05/23/2022 05:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: PERMENTER,NICHOLAS
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 05/23/2022 19:53:49

Diagnostics Check	OK	19:47
Air Blank	0.000	19:47
Control Test	0.079	19:48
Air Blank	0.000	19:48
Subject Sample # 1	0.095	19:49
Breath Volume	1.867	19:49
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.092	19:52
Breath Volume	1.953	19:52
Air Blank	0.000	19:52
Control Test	0.079	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 06/29/2022 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMANN,HANS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: ERB

DATE/TIME OF TEST: 05/24/2022 01:05:22

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.191	00:59
Breath Volume	1.199	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.198	01:03
Breath Volume	1.175	01:03
Air Blank	0.000	01:04
Control Test	0.077	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 06/29/2022 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARM,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:18

SUBJECT'S LAST NAME: FLANNIGAN

DATE/TIME OF TEST: 05/24/2022 09:41:53

Diagnostics Check	OK	09:39	
Air Blank	0.000	09:39	
Control Test	0.079	09:39	
Air Blank	0.000	09:40	
Subject Sample	# 1	09:40	Subject Test Refused
Breath Volume	0.000	09:40	
Air Blank	0.000	09:40	
Control Test	0.078	09:41	
Air Blank	0.000	09:41	
Diagnostics Check	OK	09:41	

Data Download Date/Time: 06/29/2022 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAMOS,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BANVILLE

DATE/TIME OF TEST: 05/26/2022 22:18:03

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.081	22:11
Air Blank	0.000	22:11
Subject Sample # 1	0.121	22:13
Breath Volume	1.527	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.123	22:16
Breath Volume	1.656	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:18

Data Download Date/Time: 06/29/2022 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMANN,HANS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SUTCLIFFE

DATE/TIME OF TEST: 05/28/2022 01:14:56

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.081	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.087	01:10
Breath Volume	1.371	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.084	01:13
Breath Volume	1.339	01:13
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 06/29/2022 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BECKMAN

DATE/TIME OF TEST: 05/28/2022 02:59:08

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.181	02:54
Breath Volume	2.171	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:56
Subject Sample # 2	0.177	02:57
Breath Volume	1.941	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 06/29/2022 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: FROEHLICH,TYLER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/28/2022 04:04:28

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.014	03:59
Breath Volume	1.343	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample # 2	0.012	04:02
Breath Volume	1.316	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 06/29/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMANN,HANS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BILBRO

DATE/TIME OF TEST: 05/29/2022 02:19:17

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.100	02:14
Breath Volume	1.296	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.106	02:17
Breath Volume	1.476	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 06/29/2022 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCDONALD,A
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 05/29/2022 07:11:34

Diagnostics Check	OK	07:08	
Air Blank	0.000	07:09	
Control Test	0.081	07:09	
Air Blank	0.000	07:09	
Subject Sample	# 1	07:10	Subject Test Refused
Breath Volume	0.000	07:10	
Air Blank	0.000	07:10	
Control Test	0.079	07:10	
Air Blank	0.000	07:11	
Diagnostics Check	OK	07:11	

Data Download Date/Time: 06/29/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GUINANE,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: VODAR

DATE/TIME OF TEST: 05/29/2022 14:22:55

Diagnostics Check	OK	14:15
Air Blank	0.000	14:15
Control Test	0.080	14:16
Air Blank	0.000	14:16
Subject Sample # 1	0.260	14:17
Breath Volume	1.628	14:17
Air Blank	0.000	14:18
Air Blank	0.000	14:19
Subject Sample # 2	0.264	14:21
Breath Volume	1.324	14:21
Air Blank	0.000	14:21
Control Test	0.077	14:22
Air Blank	0.000	14:22
Diagnostics Check	OK	14:22

Data Download Date/Time: 06/29/2022 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RICH,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: CHAMBERLIN

DATE/TIME OF TEST: 05/29/2022 23:03:45

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.132	22:59
Breath Volume	1.125	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.129	23:02
Breath Volume	1.378	23:02
Air Blank	0.000	23:02
Control Test	0.079	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 06/29/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCASTER,CODY,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: WALDBEESSOR

DATE/TIME OF TEST: 05/30/2022 16:30:11

Diagnostics Check	OK	16:22
Air Blank	0.000	16:23
Control Test	0.079	16:23
Air Blank	0.000	16:24
Subject Sample # 1	0.136	16:25
Breath Volume	2.500	16:25
Air Blank	0.000	16:25
Air Blank	0.000	16:27
Subject Sample # 2	0.137	16:28
Breath Volume	2.832	16:28
Air Blank	0.000	16:29
Control Test	0.079	16:29
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 06/29/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 05/01/2022 23:00:50

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.080	22:55
Air Blank	0.000	22:55
Subject Sample # 1	0.134	22:56
Breath Volume	2.949	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:58
Subject Sample # 2	0.132	22:59
Breath Volume	3.289	22:59
Air Blank	0.000	22:59
Control Test	0.079	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 05/03/2022 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TITUS,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: MERRITT

DATE/TIME OF TEST: 05/03/2022 18:01:39

Diagnostics Check	OK	17:53
Air Blank	0.000	17:54
Control Test	0.079	17:54
Air Blank	0.000	17:54
Subject Sample # 1	0.256	17:56
Breath Volume	1.250	17:56
Air Blank	0.000	17:57
Air Blank	0.000	17:58
Subject Sample # 2	0.269	18:00
Breath Volume	1.496	18:00
Air Blank	0.000	18:00
Control Test	0.079	18:01
Air Blank	0.000	18:01
Diagnostics Check	OK	18:01

Data Download Date/Time: 05/23/2022 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: VIEIRA,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: BELLAMY

DATE/TIME OF TEST: 05/04/2022 02:46:44

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.079	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 05/23/2022 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: NAUERTZ

DATE/TIME OF TEST: 05/04/2022 20:12:01

Diagnostics Check	OK	20:09	
Air Blank	0.000	20:09	
Control Test	0.079	20:09	
Air Blank	0.000	20:10	
Subject Sample	# 1	20:10	Subject Test Refused
Breath Volume	0.000	20:10	
Air Blank	0.000	20:11	
Control Test	0.079	20:11	
Air Blank	0.000	20:11	
Diagnostics Check	OK	20:12	

Data Download Date/Time: 05/23/2022 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOX,JAMES,M
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: HOLLIS,BOBBY
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: LANDRY

DATE/TIME OF TEST: 05/07/2022 23:05:35

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.080	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.158	23:00
Breath Volume	3.363	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.158	23:03
Breath Volume	3.558	23:03
Air Blank	0.000	23:04
Control Test	0.078	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 05/23/2022 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,RANDY,K
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: HUNTER,DONALD
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 05/08/2022 00:09:40

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.080	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:09	Sequence Aborted
Breath Volume	0.000	00:09	
Air Blank	0.000	00:09	

Data Download Date/Time: 05/23/2022 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,RANDY,K
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: HUNTER,DONALD
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 05/08/2022 00:16:13

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:13	
Control Test	0.079	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:15	
Control Test	0.080	00:15	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 05/23/2022 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/13/2022 22:41:10

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.079	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.237	22:36
Breath Volume	1.300	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.240	22:39
Breath Volume	1.527	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 05/23/2022 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/15/2022 00:21:51

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.079	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Control Test	0.079	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 05/23/2022 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BRANDT

DATE/TIME OF TEST: 05/15/2022 03:47:33

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:45	
Control Test	0.079	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 05/23/2022 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: JOHNSON,A
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: MCCALED

DATE/TIME OF TEST: 05/17/2022 23:11:54

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.079	23:09	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Control Test	0.079	23:11	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 05/23/2022 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, VICTORIA, S
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: RICH, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: SELLARS

DATE/TIME OF TEST: 05/18/2022 23:58:23

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.079	23:56	
Air Blank	0.000	23:56	
Subject Sample	# 1	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:57	
Control Test	0.079	23:57	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 05/23/2022 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 05/19/2022 21:15:28

Diagnostics Check	OK	21:01	
Air Blank	0.000	21:02	
Control Test	0.079	21:02	
Air Blank	0.000	21:03	
Subject Sample	# 1 0.293	21:03	
Breath Volume	1.824	21:03	
Air Blank	0.000	21:04	
Air Blank	0.000	21:06	
Subject Sample	# 2	21:09	No Sample Provided
Breath Volume	0.000	21:09	
Air Blank	0.000	21:10	
Air Blank	0.000	21:12	
Subject Sample	# 3 0.277	21:13	
Breath Volume	2.859	21:13	
Air Blank	0.000	21:14	
Control Test	0.079	21:14	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:15	

Data Download Date/Time: 05/23/2022 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WELLS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: JUANEZ

DATE/TIME OF TEST: 05/22/2022 02:19:39

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.000	02:15
Breath Volume	1.351	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.000	02:18
Breath Volume	1.320	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 05/23/2022 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GITTO,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: D525074642080

DATE/TIME OF TEST: 05/24/2022 21:16:04

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.078	21:09
Air Blank	0.000	21:09
Subject Sample # 1	0.157	21:11
Breath Volume	1.789	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:13
Subject Sample # 2	0.167	21:14
Breath Volume	1.558	21:14
Air Blank	0.000	21:15
Control Test	0.077	21:15
Air Blank	0.000	21:15
Diagnostics Check	OK	21:16

Data Download Date/Time: 06/13/2022 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: D'ADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TITUS,A
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ROOSEN

DATE/TIME OF TEST: 05/25/2022 23:55:52

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.080	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.000	23:51
Breath Volume	2.070	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.000	23:54
Breath Volume	1.941	23:54
Air Blank	0.000	23:54
Control Test	0.079	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 06/13/2022 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: RICH,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 05/28/2022 03:15:02

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.150	03:10
Breath Volume	1.386	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:12
Subject Sample # 2	0.164	03:13
Breath Volume	1.144	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:15

Data Download Date/Time: 06/13/2022 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: REID,REID
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/19/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: HESS

DATE/TIME OF TEST: 05/29/2022 17:54:57

Diagnostics Check	OK	17:48
Air Blank	0.000	17:49
Control Test	0.080	17:49
Air Blank	0.000	17:49
Subject Sample # 1	0.029	17:50
Breath Volume	3.308	17:50
Air Blank	0.000	17:51
Air Blank	0.000	17:52
Subject Sample # 2	0.028	17:53
Breath Volume	1.851	17:53
Air Blank	0.000	17:54
Control Test	0.080	17:54
Air Blank	0.000	17:54
Diagnostics Check	OK	17:54

Data Download Date/Time: 06/13/2022 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: D'ADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CRANE,R
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: STANAGE

DATE/TIME OF TEST: 05/29/2022 21:13:23

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:11	
Control Test	0.079	21:11	
Air Blank	0.000	21:11	
Subject Sample	# 1	21:12	Subject Test Refused
Breath Volume	0.000	21:12	
Air Blank	0.000	21:12	
Control Test	0.079	21:12	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 06/13/2022 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: SENTENERI

DATE/TIME OF TEST: 05/03/2022 00:08:49

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.119	00:04
Breath Volume	1.449	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:06
Subject Sample # 2	0.121	00:07
Breath Volume	1.425	00:07
Air Blank	0.000	00:07
Control Test	0.079	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 05/24/2022 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/09/2022 02:03:25

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.081	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.228	01:58
Breath Volume	2.031	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:00
Subject Sample # 2	0.217	02:01
Breath Volume	2.128	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 05/24/2022 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEUSER,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: PAONESSA,DOMINIC
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 05/15/2022 22:16:52

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.081	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.203	22:11
Breath Volume	2.937	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.189	22:15
Breath Volume	2.441	22:15
Air Blank	0.000	22:15
Control Test	0.079	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 05/24/2022 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LOPEZ,C
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: MOJICA JR

DATE/TIME OF TEST: 05/17/2022 19:25:44

Diagnostics Check	OK	19:19
Air Blank	0.000	19:19
Control Test	0.082	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.120	19:21
Breath Volume	2.312	19:21
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.131	19:24
Breath Volume	2.007	19:24
Air Blank	0.000	19:24
Control Test	0.081	19:25
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 06/09/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 05/21/2022 02:09:21

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.082	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.140	02:04
Breath Volume	1.566	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.148	02:07
Breath Volume	1.179	02:07
Air Blank	0.000	02:08
Control Test	0.081	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 06/09/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 05/21/2022 02:41:45

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.010	02:37
Breath Volume	1.296	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.011	02:40
Breath Volume	1.417	02:40
Air Blank	0.000	02:40
Control Test	0.082	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 06/09/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MUTCHERSON JR

DATE/TIME OF TEST: 05/21/2022 03:34:07

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.146	03:29
Breath Volume	2.519	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.156	03:32
Breath Volume	2.320	03:32
Air Blank	0.000	03:33
Control Test	0.081	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: SERRANO ESTRADA

DATE/TIME OF TEST: 05/21/2022 22:50:51

Diagnostics Check	OK	22:44
Air Blank	0.000	22:45
Control Test	0.081	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.241	22:46
Breath Volume	1.597	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.257	22:49
Breath Volume	1.683	22:49
Air Blank	0.000	22:49
Control Test	0.081	22:50
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: HERNANDEZ GOMEZ

DATE/TIME OF TEST: 05/22/2022 02:01:59

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.081	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.138	01:57
Breath Volume	1.894	01:57
Air Blank	0.000	01:58
Air Blank	0.000	01:59
Subject Sample # 2	0.136	02:00
Breath Volume	3.246	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURANCEAU,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: RAMIREZ VAZQUEZ

DATE/TIME OF TEST: 05/22/2022 04:31:38

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.082	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.086	04:27
Breath Volume	1.777	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.084	04:30
Breath Volume	1.308	04:30
Air Blank	0.000	04:30
Control Test	0.081	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: CABRERA JAULAR

DATE/TIME OF TEST: 05/23/2022 01:36:20

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.216	01:30
Breath Volume	2.285	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.205	01:34
Breath Volume	1.320	01:34
Air Blank	0.000	01:35
Control Test	0.081	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 05/23/2022 04:30:48

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.082	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.197	04:26
Breath Volume	2.242	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.211	04:29
Breath Volume	3.195	04:29
Air Blank	0.000	04:29
Control Test	0.081	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/27/2022 03:23:57

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.180	03:19
Breath Volume	1.691	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.190	03:22
Breath Volume	1.792	03:22
Air Blank	0.000	03:23
Control Test	0.081	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: BESWICK

DATE/TIME OF TEST: 05/28/2022 02:14:53

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.117	02:10
Breath Volume	1.597	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 2	0.112	02:13
Breath Volume	1.835	02:13
Air Blank	0.000	02:13
Control Test	0.082	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIERMEK,T
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BYREDDY

DATE/TIME OF TEST: 05/28/2022 04:32:33

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.082	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.177	04:27
Breath Volume	3.281	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.176	04:30
Breath Volume	3.714	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: COMM VEHIC
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: BANDA-BANDA

DATE/TIME OF TEST: 05/28/2022 22:58:27

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.082	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.115	22:52
Breath Volume	1.554	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:54
Subject Sample # 2	0.111	22:56
Breath Volume	1.296	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/28/2022 23:26:17

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.081	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.000	23:21
Breath Volume	3.171	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.000	23:24
Breath Volume	3.207	23:24
Air Blank	0.000	23:25
Control Test	0.081	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 05/29/2022 05:53:53

Diagnostics Check	OK	05:46
Air Blank	0.000	05:46
Control Test	0.082	05:47
Air Blank	0.000	05:47
Subject Sample # 1	0.153	05:49
Breath Volume	2.347	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.158	05:52
Breath Volume	1.757	05:52
Air Blank	0.000	05:52
Control Test	0.081	05:53
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DALTON,MEGAN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:37

SUBJECT'S LAST NAME: KATHI

DATE/TIME OF TEST: 05/29/2022 11:09:15

Diagnostics Check	OK	11:01
Air Blank	0.000	11:02
Control Test	0.082	11:02
Air Blank	0.000	11:02
Subject Sample # 1	0.040	11:04
Breath Volume	1.621	11:04
Air Blank	0.000	11:05
Air Blank	0.000	11:07
Subject Sample # 2	0.039	11:07
Breath Volume	1.589	11:07
Air Blank	0.000	11:08
Control Test	0.081	11:08
Air Blank	0.000	11:09
Diagnostics Check	OK	11:09

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES,F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: ROSS ROBLERO

DATE/TIME OF TEST: 05/30/2022 04:49:24

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.081	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.194	04:44
Breath Volume	1.921	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.204	04:47
Breath Volume	2.035	04:47
Air Blank	0.000	04:48
Control Test	0.081	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 06/09/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: ZERO TOL
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/04/2022 00:30:49

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.102	00:26
Breath Volume	1.867	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.098	00:29
Breath Volume	2.320	00:29
Air Blank	0.000	00:29
Control Test	0.079	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 05/23/2022 21:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: MENDOZA-CRUZ

DATE/TIME OF TEST: 05/05/2022 22:46:32

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.158	22:41
Breath Volume	3.261	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.149	22:44
Breath Volume	3.292	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 05/23/2022 21:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/07/2022 23:43:19

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:36	
Control Test	0.080	23:36	
Air Blank	0.000	23:37	
Subject Sample # 1	0.108	23:38	
Breath Volume	1.691	23:38	
Air Blank	0.000	23:38	
Air Blank	0.000	23:40	
Subject Sample # 2	0.102	23:41	
Breath Volume	1.171	23:41	
Air Blank	0.000	23:41	
Control Test	0.080	23:42	
Air Blank	0.000	23:42	
Diagnostics Check	Fail	23:42	Temperature Regulation Test
Air Blank	0.000	23:43	

Data Download Date/Time: 05/23/2022 21:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: NADER

DATE/TIME OF TEST: 05/12/2022 01:17:41

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1 0.196	01:10	Volume Not Met
Breath Volume	0.082	01:10	
Air Blank	0.000	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 2 0.225	01:16	Volume Not Met
Breath Volume	0.070	01:16	
Air Blank	0.000	01:16	
Control Test	0.079	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 05/23/2022 21:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: NETTELS

DATE/TIME OF TEST: 05/19/2022 00:18:13

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.313	00:13
Breath Volume	2.457	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:15
Subject Sample # 2	0.306	00:16
Breath Volume	2.460	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 05/23/2022 21:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 05/20/2022 01:58:54

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.190	01:54
Breath Volume	2.136	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.190	01:57
Breath Volume	1.429	01:57
Air Blank	0.000	01:57
Control Test	0.077	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 05/23/2022 21:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: DUKES

DATE/TIME OF TEST: 05/23/2022 02:02:44

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.267	01:58
Breath Volume	2.910	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.261	02:01
Breath Volume	3.105	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 05/23/2022 21:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAMP,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: STAMP,KYLE
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: BUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 05/01/2022 21:47:14

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:44	
Control Test	0.079	21:44	
Air Blank	0.000	21:45	
Subject Sample	# 1	21:45	RFI Detect
Breath Volume	0.000	21:45	
Air Blank	0.000	21:46	
Control Test	0.080	21:46	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

Data Download Date/Time: 06/03/2022 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAMP,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: STAMP,KYLE
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: BUI
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:32

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 05/01/2022 21:54:40

Diagnostics Check	OK	21:49	
Air Blank	0.000	21:49	
Control Test	0.079	21:49	
Air Blank	0.000	21:50	
Subject Sample	# 1 0.141	21:50	
Breath Volume	3.019	21:50	
Air Blank	0.000	21:51	
Air Blank	0.000	21:53	
Subject Sample	# 2 0.141	21:53	
Breath Volume	2.804	21:53	
Air Blank		21:54	RFI Detect
Air Blank	0.000	21:54	

Data Download Date/Time: 06/03/2022 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETTS,J,A
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 05/07/2022 03:55:15

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.097	03:50
Breath Volume	4.070	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.097	03:53
Breath Volume	4.304	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 06/03/2022 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETTS,J,A
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GALUSHA

DATE/TIME OF TEST: 05/08/2022 01:10:16

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.102	01:03
Breath Volume	3.390	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.099	01:08
Breath Volume	2.824	01:08
Air Blank	0.000	01:09
Control Test	0.081	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 06/03/2022 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,JAMES,E
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BLACKMORE

DATE/TIME OF TEST: 05/08/2022 03:11:46

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.158	03:07
Breath Volume	1.246	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.166	03:10
Breath Volume	1.628	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 06/03/2022 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: DORAN,RYAN
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GARRIS

DATE/TIME OF TEST: 05/10/2022 20:33:42

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:31	
Control Test	0.080	20:31	
Air Blank	0.000	20:32	
Subject Sample	# 1	20:32	Subject Test Refused
Breath Volume	0.000	20:32	
Air Blank	0.000	20:32	
Control Test	0.080	20:33	
Air Blank	0.000	20:33	
Diagnostics Check	OK	20:33	

Data Download Date/Time: 06/03/2022 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SIMMONS,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DOHERTY

DATE/TIME OF TEST: 05/12/2022 04:12:11

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.135	04:07
Breath Volume	2.839	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.140	04:10
Breath Volume	2.835	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 06/03/2022 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: KUCZKA,JACOB
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: T610421893450

DATE/TIME OF TEST: 05/14/2022 05:12:40

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.172	05:08
Breath Volume	2.710	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.154	05:11
Breath Volume	2.558	05:11
Air Blank	0.000	05:11
Control Test	0.081	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 06/03/2022 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWNSBERGER, BLAINE, T
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: FAUST, JOSHUA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: CERVANTES

DATE/TIME OF TEST: 05/17/2022 04:56:00

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.079	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.170	04:51
Breath Volume	3.917	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:53
Subject Sample # 2	0.166	04:54
Breath Volume	4.363	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:56

Data Download Date/Time: 06/03/2022 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: SOTO,ADRIAN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 05/20/2022 04:51:35

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.162	04:46
Breath Volume	1.621	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.161	04:49
Breath Volume	1.722	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 06/03/2022 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,JAMES,E
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: HUGHES,JAMES
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: ARREAGA

DATE/TIME OF TEST: 05/23/2022 01:22:03

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.079	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.205	01:17
Breath Volume	1.890	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.205	01:20
Breath Volume	2.582	01:20
Air Blank	0.000	01:21
Control Test	0.077	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:22

Data Download Date/Time: 06/03/2022 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,JAMES,E
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: BUEHLER,A
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: H220729791910

DATE/TIME OF TEST: 05/23/2022 04:06:51

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.075	04:02
Breath Volume	1.800	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.071	04:05
Breath Volume	3.343	04:05
Air Blank	0.000	04:05
Control Test	0.078	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 06/03/2022 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,JAMES,E
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: BUEHLER,A
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 05/23/2022 04:16:47

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.078	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.072	04:12
Breath Volume	3.191	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample # 2	0.074	04:15
Breath Volume	3.171	04:15
Air Blank	0.000	04:15
Control Test	0.078	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 06/03/2022 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: COODY,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 05/28/2022 19:58:15

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.080	19:51
Air Blank	0.000	19:52
Subject Sample # 1	0.105	19:53
Breath Volume	1.226	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:55
Subject Sample # 2	0.102	19:56
Breath Volume	2.019	19:56
Air Blank	0.000	19:57
Control Test	0.080	19:57
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 06/03/2022 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HOGAN,JEREMY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/09/2022 23:57:22

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.077	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.161	23:52
Breath Volume	1.535	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.162	23:55
Breath Volume	1.601	23:55
Air Blank	0.000	23:56
Control Test	0.077	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 06/14/2022 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,WILLIAM,B
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: TURNER,WILLIAM
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 05/16/2022 22:04:00

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:03	
Control Test	0.069	22:03	Control Outside Tolerance
Air Blank	0.000	22:04	

Data Download Date/Time: 06/14/2022 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,WILLIAM,B
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: TURNER,WILLIAM
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 05/16/2022 22:08:13

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:07	
Control Test	0.072	22:07	Control Outside Tolerance
Air Blank	0.000	22:08	

Data Download Date/Time: 06/14/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER, WILLIAM, B
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: TURNER, WILLIAM
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 05/16/2022 22:29:07

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.075	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 1 0.178	22:18	Volume Not Met
Breath Volume	0.000	22:18	
Air Blank	0.000	22:19	
Air Blank	0.000	22:21	
Subject Sample	# 2 0.179	22:22	
Breath Volume	1.589	22:22	
Air Blank	0.000	22:22	
Air Blank	0.000	22:24	
Subject Sample	# 3	22:27	Subject Test Refused
Breath Volume	0.132	22:27	
Air Blank	0.000	22:28	
Control Test	0.077	22:28	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

Data Download Date/Time: 06/14/2022 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MALOY,EMILY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/18/2022 06:04:17

Diagnostics Check	OK	05:58	
Air Blank	0.000	05:59	
Control Test	0.077	05:59	
Air Blank	0.000	05:59	
Subject Sample	# 1	06:02	Subject Test Refused
Breath Volume	0.062	06:02	
Air Blank	0.000	06:03	
Control Test	0.077	06:03	
Air Blank	0.000	06:04	
Diagnostics Check	OK	06:04	

Data Download Date/Time: 06/14/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MALOY,EMILY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/18/2022 06:19:42

Diagnostics Check	OK	06:11
Air Blank	0.000	06:11
Control Test	0.078	06:12
Air Blank	0.000	06:12
Subject Sample # 1	0.177	06:14
Breath Volume	3.031	06:14
Air Blank	0.000	06:15
Air Blank	0.000	06:17
Subject Sample # 2	0.185	06:18
Breath Volume	1.140	06:18
Air Blank	0.000	06:18
Control Test	0.076	06:19
Air Blank	0.000	06:19
Diagnostics Check	OK	06:19

Data Download Date/Time: 06/14/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORTER,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PORTER,ALEXANDER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 10:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:24

SUBJECT'S LAST NAME: SWEARINGEN

DATE/TIME OF TEST: 05/19/2022 11:02:05

Diagnostics Check	OK	10:55
Air Blank	0.000	10:56
Control Test	0.078	10:56
Air Blank	0.000	10:56
Subject Sample # 1	0.000	10:57
Breath Volume	1.601	10:57
Air Blank	0.000	10:58
Air Blank	0.000	11:00
Subject Sample # 2	0.000	11:00
Breath Volume	1.984	11:00
Air Blank	0.000	11:01
Control Test	0.078	11:01
Air Blank	0.000	11:02
Diagnostics Check	OK	11:02

Data Download Date/Time: 06/14/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MCDONALD,FRANK
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/20/2022 14:51:52

Diagnostics Check	OK	14:45
Air Blank	0.000	14:45
Control Test	0.076	14:46
Air Blank	0.000	14:46
Subject Sample # 1	0.270	14:47
Breath Volume	4.050	14:47
Air Blank	0.000	14:47
Air Blank	0.000	14:49
Subject Sample # 2	0.252	14:50
Breath Volume	2.175	14:50
Air Blank	0.000	14:50
Control Test	0.075	14:51
Air Blank	0.000	14:51
Diagnostics Check	OK	14:51

Data Download Date/Time: 06/14/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY,ALISHA,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: HALE,CALEB
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MAULDEN

DATE/TIME OF TEST: 05/28/2022 00:01:30

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.011	23:57
Breath Volume	1.570	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.011	00:00
Breath Volume	1.660	00:00
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 06/14/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: VIZI, SAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HUGHEY

DATE/TIME OF TEST: 05/01/2022 03:17:29

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.201	03:13
Breath Volume	1.554	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.201	03:16
Breath Volume	1.855	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 05/03/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNOVA,CASSANDRA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KAVANAGH,SHAWN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 05/01/2022 19:06:49

Diagnostics Check	OK	18:59	
Air Blank	0.000	18:59	
Control Test	0.079	18:59	
Air Blank	0.000	19:00	
Subject Sample	# 1 0.207	19:00	
Breath Volume	2.906	19:00	
Air Blank	0.000	19:01	
Air Blank	0.000	19:03	
Subject Sample	# 2	19:06	Sequence Aborted
Breath Volume	1.636	19:06	
Air Blank	0.000	19:06	

Data Download Date/Time: 05/03/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNOVA,CASSANDRA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KAVANAGH,SHAWN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 05/01/2022 19:14:50

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.078	19:09
Air Blank	0.000	19:10
Subject Sample # 1	0.201	19:10
Breath Volume	2.886	19:10
Air Blank	0.000	19:11
Air Blank	0.000	19:12
Subject Sample # 2	0.205	19:13
Breath Volume	1.242	19:13
Air Blank	0.000	19:13
Control Test	0.078	19:14
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

Data Download Date/Time: 05/03/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNOVA,CASSANDRA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 05/01/2022 22:19:07

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.000	22:14
Breath Volume	3.949	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.000	22:17
Breath Volume	2.253	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 05/03/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINEZ,F
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: HANGEN

DATE/TIME OF TEST: 05/01/2022 23:25:55

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.271	23:21
Breath Volume	2.843	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.258	23:24
Breath Volume	2.761	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 05/03/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATO,MARK,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PANTOJA VAZQUEZ,ABRAHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: BOATWRIGHT

DATE/TIME OF TEST: 05/02/2022 17:56:04

Diagnostics Check	OK	17:49
Air Blank	0.000	17:50
Control Test	0.080	17:50
Air Blank	0.000	17:51
Subject Sample # 1	0.000	17:51
Breath Volume	1.773	17:51
Air Blank	0.000	17:52
Air Blank	0.000	17:54
Subject Sample # 2	0.000	17:54
Breath Volume	1.492	17:54
Air Blank	0.000	17:55
Control Test	0.080	17:55
Air Blank	0.000	17:55
Diagnostics Check	OK	17:56

Data Download Date/Time: 05/03/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATO,MARK,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DURHAM,KARI
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: SHEENE JR.

DATE/TIME OF TEST: 05/02/2022 19:20:16

Diagnostics Check	OK	19:13
Air Blank	0.000	19:14
Control Test	0.080	19:14
Air Blank	0.000	19:15
Subject Sample # 1	0.000	19:15
Breath Volume	2.789	19:15
Air Blank	0.000	19:16
Air Blank	0.000	19:18
Subject Sample # 2	0.000	19:18
Breath Volume	2.121	19:18
Air Blank	0.000	19:19
Control Test	0.080	19:19
Air Blank	0.000	19:20
Diagnostics Check	OK	19:20

Data Download Date/Time: 05/03/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: LUCIANO

DATE/TIME OF TEST: 05/05/2022 22:06:38

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.079	22:00
Air Blank	0.000	22:00
Subject Sample # 1	0.234	22:01
Breath Volume	1.414	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:03
Subject Sample # 2	0.241	22:05
Breath Volume	2.148	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 06/02/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNOVA,CASSANDRA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CANNOVA,CASSANDRA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: SCHRAMM

DATE/TIME OF TEST: 05/05/2022 22:49:14

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.159	22:44
Breath Volume	2.339	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.142	22:47
Breath Volume	1.390	22:47
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 06/02/2022 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MATTHES,LEA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: LABONTE

DATE/TIME OF TEST: 05/06/2022 19:03:13

Diagnostics Check	OK	18:52	
Air Blank	0.000	18:52	
Control Test	0.081	18:52	
Air Blank	0.000	18:53	
Subject Sample	# 1 0.131	18:56	Volume Not Met
Breath Volume	0.371	18:56	
Air Blank	0.000	18:57	
Air Blank	0.000	18:59	
Subject Sample	# 2	19:01	Subject Test Refused
Breath Volume	0.230	19:01	
Air Blank	0.000	19:02	
Control Test	0.080	19:02	
Air Blank	0.000	19:03	
Diagnostics Check	OK	19:03	

Data Download Date/Time: 06/02/2022 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MATTHES,LEA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: LABONTE

DATE/TIME OF TEST: 05/06/2022 19:11:04

Diagnostics Check	OK	19:05	
Air Blank	0.000	19:05	
Control Test	0.080	19:06	
Air Blank	0.000	19:06	
Subject Sample	# 1	19:09	Subject Test Refused
Breath Volume	0.187	19:09	
Air Blank	0.000	19:10	
Control Test	0.080	19:10	
Air Blank	0.000	19:10	
Diagnostics Check	OK	19:11	

Data Download Date/Time: 06/02/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/06/2022 19:23:27

Diagnostics Check	OK	19:15
Air Blank	0.000	19:15
Control Test	0.079	19:15
Air Blank	0.000	19:16
Subject Sample # 1	0.248	19:18
Breath Volume	1.167	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.248	19:21
Breath Volume	1.382	19:21
Air Blank	0.000	19:22
Control Test	0.078	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 06/02/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: LOUK

DATE/TIME OF TEST: 05/07/2022 00:51:09

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.000	00:46
Breath Volume	2.429	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:49
Breath Volume	3.328	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 06/02/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,ALEXANDER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: ZAMPINO

DATE/TIME OF TEST: 05/07/2022 04:32:37

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.102	04:27
Breath Volume	1.363	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.111	04:31
Breath Volume	1.707	04:31
Air Blank	0.000	04:31
Control Test	0.080	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 06/02/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MATTHES,LEA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BURNSIDE

DATE/TIME OF TEST: 05/08/2022 01:38:36

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.097	01:34
Breath Volume	2.093	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.093	01:37
Breath Volume	2.835	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 06/02/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP,RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/08/2022 04:09:14

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.183	04:04
Breath Volume	4.285	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.183	04:07
Breath Volume	4.125	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 06/02/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,JILLIAN,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CARROLL,JILLIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: KRASKA

DATE/TIME OF TEST: 05/08/2022 21:18:26

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.080	21:12
Air Blank	0.000	21:13
Subject Sample # 1	0.149	21:13
Breath Volume	1.199	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.157	21:16
Breath Volume	1.554	21:16
Air Blank	0.000	21:17
Control Test	0.080	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGIEL,ADRIAN,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GOMEZ,JESSE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: TSCHAPPAT

DATE/TIME OF TEST: 05/10/2022 03:31:05

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.184	03:26
Breath Volume	3.031	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.179	03:29
Breath Volume	3.222	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MADURO,VICTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 05/11/2022 22:26:46

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.000	22:22
Breath Volume	2.136	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.000	22:25
Breath Volume	2.097	22:25
Air Blank	0.000	22:25
Control Test	0.080	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP,RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/12/2022 02:13:38

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.166	02:09
Breath Volume	2.738	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.169	02:12
Breath Volume	2.625	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRONK,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FRONK,MATTHEW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: SHEEHAN

DATE/TIME OF TEST: 05/13/2022 00:51:48

Diagnostics Check	OK	00:51	
Air Blank		00:51	RFI Detect
Air Blank	0.000	00:51	

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRONK,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FRONK,MATTHEW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: SHEEHAN

DATE/TIME OF TEST: 05/13/2022 01:08:39

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:54	
Control Test	0.080	00:54	
Air Blank	0.000	00:54	
Subject Sample	# 1 0.172	00:55	
Breath Volume	2.070	00:55	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample	# 2 0.183	01:01	Volume Not Met
Breath Volume	0.035	01:01	
Air Blank	0.000	01:02	
Air Blank	0.000	01:03	
Subject Sample	# 3 0.162	01:07	
Breath Volume	1.613	01:07	
Air Blank	0.000	01:07	
Control Test	0.077	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BIROS

DATE/TIME OF TEST: 05/13/2022 03:27:27

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.128	03:22
Breath Volume	2.789	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.129	03:26
Breath Volume	1.769	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: DAGOSTA

DATE/TIME OF TEST: 05/14/2022 01:45:18

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.158	01:40
Breath Volume	2.695	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.149	01:43
Breath Volume	2.136	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RODRIGUEZ,KELVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: NARANJO ESPINOSA

DATE/TIME OF TEST: 05/14/2022 18:33:54

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.080	18:26
Air Blank	0.000	18:27
Subject Sample # 1	0.130	18:28
Breath Volume	2.437	18:28
Air Blank	0.000	18:29
Air Blank	0.000	18:31
Subject Sample # 2	0.136	18:32
Breath Volume	1.894	18:32
Air Blank	0.000	18:33
Control Test	0.080	18:33
Air Blank	0.000	18:33
Diagnostics Check	OK	18:33

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MADURO,VICTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: RABELL

DATE/TIME OF TEST: 05/14/2022 23:11:42

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.205	23:06
Breath Volume	1.269	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:08
Subject Sample # 2	0.212	23:10
Breath Volume	1.351	23:10
Air Blank	0.000	23:10
Control Test	0.080	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNOVA,CASSANDRA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CANNOVA,CASSANDRA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: GADUS

DATE/TIME OF TEST: 05/21/2022 01:50:19

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.078	01:50	RFI Detect
Air Blank		01:50	RFI Detect

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNOVA,CASSANDRA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CANNOVA,CASSANDRA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: GADUS

DATE/TIME OF TEST: 05/21/2022 01:59:28

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.081	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.044	01:55
Breath Volume	3.613	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.042	01:58
Breath Volume	3.644	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 05/21/2022 03:34:22

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.054	03:29
Breath Volume	1.875	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.049	03:32
Breath Volume	1.230	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: SANTER

DATE/TIME OF TEST: 05/26/2022 03:34:56

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.197	03:30
Breath Volume	3.445	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.195	03:33
Breath Volume	3.476	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,STEPHEN,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LEONARD,STEPHEN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: GRESH

DATE/TIME OF TEST: 05/26/2022 08:50:29

Diagnostics Check	OK	08:47	
Air Blank	0.000	08:48	
Control Test	0.080	08:48	
Air Blank	0.000	08:48	
Subject Sample	# 1	08:49	Subject Test Refused
Breath Volume	0.000	08:49	
Air Blank	0.000	08:49	
Control Test	0.080	08:49	
Air Blank	0.000	08:50	
Diagnostics Check	OK	08:50	

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: DAUGHERTY

DATE/TIME OF TEST: 05/26/2022 19:55:15

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.080	19:50
Air Blank	0.000	19:50
Subject Sample # 1	0.278	19:50
Breath Volume	1.886	19:50
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.285	19:53
Breath Volume	1.656	19:53
Air Blank	0.000	19:54
Control Test	0.079	19:54
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: GARCIA IRIZARRY

DATE/TIME OF TEST: 05/27/2022 00:21:13

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.143	00:16
Breath Volume	3.484	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.135	00:19
Breath Volume	3.046	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ARD

DATE/TIME OF TEST: 05/27/2022 01:43:46

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.000	01:39
Breath Volume	3.195	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.000	01:42
Breath Volume	2.613	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: MORALES CARTAGENA

DATE/TIME OF TEST: 05/28/2022 03:25:17

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.159	03:20
Breath Volume	3.574	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.155	03:23
Breath Volume	2.992	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL, CHASE,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HARRISON, WILLIE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SOUTER

DATE/TIME OF TEST: 05/29/2022 01:12:04

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.184	01:07
Breath Volume	1.191	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.189	01:10
Breath Volume	1.378	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:12

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SATTERLEE, ANDREW, S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GOULD, PATRICIA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:08

SUBJECT'S LAST NAME: BRIOC III

DATE/TIME OF TEST: 05/30/2022 14:19:04

Diagnostics Check	OK	14:12
Air Blank	0.000	14:13
Control Test	0.080	14:13
Air Blank	0.000	14:13
Subject Sample # 1	0.000	14:14
Breath Volume	2.750	14:14
Air Blank	0.000	14:14
Air Blank	0.000	14:16
Subject Sample # 2	0.000	14:17
Breath Volume	2.828	14:17
Air Blank	0.000	14:18
Control Test	0.080	14:18
Air Blank	0.000	14:18
Diagnostics Check	OK	14:19

Data Download Date/Time: 06/02/2022 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: PALAGIN

DATE/TIME OF TEST: 05/06/2022 20:55:02

Diagnostics Check	OK	20:48
Air Blank	0.000	20:49
Control Test	0.078	20:49
Air Blank	0.000	20:49
Subject Sample # 1	0.140	20:50
Breath Volume	4.542	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:52
Subject Sample # 2	0.142	20:53
Breath Volume	4.281	20:53
Air Blank	0.000	20:54
Control Test	0.077	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:55

Data Download Date/Time: 05/30/2022 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DOUVILLIER

DATE/TIME OF TEST: 05/07/2022 03:42:11

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.150	03:37
Breath Volume	2.351	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.152	03:40
Breath Volume	2.714	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 05/30/2022 22:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BRATTON

DATE/TIME OF TEST: 05/08/2022 23:12:30

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.078	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.067	23:08
Breath Volume	1.742	23:08
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.065	23:11
Breath Volume	2.179	23:11
Air Blank	0.000	23:11
Control Test	0.078	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 05/30/2022 22:04