

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: INFANTE

DATE/TIME OF TEST: 05/05/2022 05:19:40

Diagnostics Check	OK	05:12
Air Blank	0.000	05:13
Control Test	0.081	05:13
Air Blank	0.000	05:14
Subject Sample # 1	0.196	05:15
Breath Volume	3.140	05:15
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.190	05:18
Breath Volume	3.132	05:18
Air Blank	0.000	05:18
Control Test	0.080	05:19
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 05/31/2022 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 05/05/2022 23:56:40

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.153	23:52
Breath Volume	2.363	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.167	23:55
Breath Volume	1.453	23:55
Air Blank	0.000	23:55
Control Test	0.079	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 05/31/2022 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GOBELI,KIRBY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 05/06/2022 00:11:30

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.189	00:06
Breath Volume	1.230	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:08
Subject Sample # 2	0.191	00:09
Breath Volume	1.917	00:09
Air Blank	0.000	00:10
Control Test	0.079	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 05/31/2022 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN,JONATHAN,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: CAMPBELL,T

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 23:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SLOAN

DATE/TIME OF TEST: 05/08/2022 00:17:32

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.081	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.191	00:12
Breath Volume	1.757	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.192	00:15
Breath Volume	1.859	00:15
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 05/31/2022 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,R,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MOHR,J

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 20:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BENNEFIELD III

DATE/TIME OF TEST: 05/13/2022 21:38:28

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:37	
Control Test	0.073	21:37	Control Outside Tolerance
Air Blank	0.000	21:38	

Data Download Date/Time: 05/31/2022 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,R,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MOHR,J

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 20:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BENNEFIELD III

DATE/TIME OF TEST: 05/13/2022 21:42:45

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.073	21:42	Control Outside Tolerance
Air Blank	0.000	21:42	

Data Download Date/Time: 05/31/2022 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BACKSTROM,BENJAMIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:43

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: O'KEEFE

DATE/TIME OF TEST: 05/16/2022 02:43:22

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.175	02:38
Breath Volume	1.359	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.168	02:41
Breath Volume	1.355	02:41
Air Blank	0.000	02:42
Control Test	0.078	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 05/31/2022 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN,JONATHAN,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LOGAN,JONATHAN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 03:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: COOKSEY

DATE/TIME OF TEST: 05/22/2022 04:12:29

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.155	04:07
Breath Volume	1.246	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.151	04:10
Breath Volume	1.316	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 05/31/2022 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: FORBES,OMAR

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 21:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 05/29/2022 22:32:25

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Control Test	0.080	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 05/31/2022 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCDONALD,COLE,M

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: NEELY,BRANDON

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 22:42

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 05/30/2022 23:09:56

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.000	23:05
Breath Volume	1.937	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.000	23:08
Breath Volume	1.753	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 05/31/2022 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: PODUSENKO

DATE/TIME OF TEST: 05/01/2022 20:16:19

Diagnostics Check	OK	20:08
Air Blank	0.000	20:09
Control Test	0.079	20:09
Air Blank	0.000	20:10
Subject Sample # 1	0.173	20:10
Breath Volume	1.156	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.191	20:14
Breath Volume	2.226	20:14
Air Blank	0.000	20:15
Control Test	0.080	20:15
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,ROHAN,G
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: REMY,LUCKENSON
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: INSLEY

DATE/TIME OF TEST: 05/02/2022 04:49:54

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:35	
Control Test	0.079	04:36	
Air Blank	0.000	04:36	
Subject Sample	# 1 0.185	04:39	Volume Not Met
Breath Volume	0.070	04:39	
Air Blank	0.000	04:40	
Air Blank	0.000	04:42	
Subject Sample	# 2 0.213	04:42	
Breath Volume	3.484	04:42	
Air Blank	0.000	04:43	
Air Blank	0.000	04:45	
Subject Sample	# 3 0.211	04:48	
Breath Volume	1.582	04:48	
Air Blank		04:49	Purge Fail
Air Blank		04:49	Purge Fail

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,ROHAN,G
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: REMY,LUCKENSON
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: INSLEY

DATE/TIME OF TEST: 05/02/2022 04:56:48

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:55	
Control Test	0.000	04:56	Control Outside Tolerance
Air Blank		04:56	Ambient Fail

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,ROHAN,G
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: REMY,LUCKENSON
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: INSLEY

DATE/TIME OF TEST: 05/02/2022 05:00:17

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.000	04:59	Control Outside Tolerance
Air Blank	0.000	05:00	

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SIDOTI,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 05/02/2022 18:13:19

Diagnostics Check	OK	18:11	
Air Blank	0.000	18:12	
Control Test	0.027	18:12	Interferent Detect
Air Blank	0.000	18:13	

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SIDOTI,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 05/02/2022 18:16:04

Diagnostics Check	OK	18:14	
Air Blank	0.000	18:15	
Control Test	0.041	18:15	Control Outside Tolerance
Air Blank	0.000	18:16	

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: TECK OSAETA

DATE/TIME OF TEST: 05/04/2022 22:37:04

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:36	
Control Test	0.063	22:36	Control Outside Tolerance
Air Blank	0.000	22:37	

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CARSON, TISHA
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 05/08/2022 00:28:14

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:27	
Control Test	0.071	00:27	Control Outside Tolerance
Air Blank	0.000	00:28	

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CARSON, TISHA
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 05/08/2022 00:31:13

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.074	00:30	Control Outside Tolerance
Air Blank	0.000	00:31	

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: EDMOND,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: MEJIA MARTINEZ

DATE/TIME OF TEST: 05/08/2022 23:06:06

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:05	
Control Test	0.072	23:05	Control Outside Tolerance
Air Blank	0.000	23:06	

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: EDMOND,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: MEJIA MARTINEZ

DATE/TIME OF TEST: 05/08/2022 23:09:01

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:08	
Control Test	0.073	23:08	Control Outside Tolerance
Air Blank	0.000	23:09	

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,CORY,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ECKERT,K
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GIRARD

DATE/TIME OF TEST: 05/14/2022 03:27:48

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.198	03:22
Breath Volume	1.613	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.185	03:26
Breath Volume	1.464	03:26
Air Blank	0.000	03:26
Control Test	0.078	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: ALBERTS

DATE/TIME OF TEST: 05/14/2022 19:02:34

Diagnostics Check	OK	18:56
Air Blank	0.000	18:56
Control Test	0.079	18:56
Air Blank	0.000	18:57
Subject Sample # 1	0.382	18:57
Breath Volume	1.816	18:57
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.366	19:00
Breath Volume	1.101	19:00
Air Blank	0.000	19:01
Control Test	0.077	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DAVIDSON,BRADLEY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 05/14/2022 22:12:03

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.079	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.236	22:06
Breath Volume	1.800	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.231	22:10
Breath Volume	1.824	22:10
Air Blank	0.000	22:11
Control Test	0.078	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:12

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEACH,B,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: URAL,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CHUMAKOVA

DATE/TIME OF TEST: 05/15/2022 05:25:48

Diagnostics Check	OK	05:22	
Air Blank	0.000	05:22	
Control Test	0.079	05:22	
Air Blank	0.000	05:23	
Subject Sample	# 1 0.250	05:24	
Breath Volume	2.414	05:24	
Air Blank		05:25	Purge Fail
Air Blank	0.000	05:25	

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEACH,B,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ZUNIGA,STEVENS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: AGUILAR TOMA

DATE/TIME OF TEST: 05/15/2022 05:49:32

Diagnostics Check	OK	05:37	
Air Blank	0.000	05:37	
Control Test	0.077	05:38	
Air Blank	0.000	05:38	
Subject Sample	# 1 0.194	05:41	Volume Not Met
Breath Volume	0.277	05:41	
Air Blank	0.000	05:42	
Air Blank	0.000	05:44	
Subject Sample	# 2 0.172	05:44	
Breath Volume	2.609	05:44	
Air Blank	0.000	05:45	
Air Blank	0.000	05:47	
Subject Sample	# 3 0.181	05:47	
Breath Volume	3.332	05:47	
Air Blank	0.000	05:48	
Control Test	0.075	05:48	
Air Blank	0.000	05:49	
Diagnostics Check	OK	05:49	

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING, JONATHAN, L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: HENRY, CRAIG
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MCAFEE

DATE/TIME OF TEST: 05/15/2022 22:05:27

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.135	22:00
Breath Volume	2.746	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.134	22:03
Breath Volume	2.312	22:03
Air Blank	0.000	22:04
Control Test	0.077	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JAYDA,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SIN,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/16/2022 01:37:36

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.219	01:33
Breath Volume	1.882	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.227	01:35
Breath Volume	1.449	01:35
Air Blank	0.000	01:36
Control Test	0.077	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 05/16/2022 02:28:47

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.079	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.110	02:24
Breath Volume	1.390	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.100	02:27
Breath Volume	1.589	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 05/18/2022 21:56:59

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.079	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.159	21:52
Breath Volume	1.593	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.158	21:55
Breath Volume	1.253	21:55
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/19/2022 02:20:22

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.201	02:15
Breath Volume	1.437	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.199	02:18
Breath Volume	1.156	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/20/2022 02:08:34

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:04	
Control Test	0.080	02:04	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:07	
Control Test	0.081	02:07	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,PATRICK,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MORRIS,HOWARD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: ISME

DATE/TIME OF TEST: 05/20/2022 07:50:54

Diagnostics Check	OK	07:44
Air Blank	0.000	07:44
Control Test	0.079	07:45
Air Blank	0.000	07:45
Subject Sample # 1	0.064	07:46
Breath Volume	2.066	07:46
Air Blank	0.000	07:46
Air Blank	0.000	07:48
Subject Sample # 2	0.061	07:49
Breath Volume	1.421	07:49
Air Blank	0.000	07:49
Control Test	0.078	07:50
Air Blank	0.000	07:50
Diagnostics Check	OK	07:50

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: CONO

DATE/TIME OF TEST: 05/21/2022 00:08:58

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.079	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.177	00:04
Breath Volume	1.691	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:06
Subject Sample # 2	0.176	00:07
Breath Volume	1.507	00:07
Air Blank	0.000	00:08
Control Test	0.080	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLIGAN,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MIGUEL REGINO

DATE/TIME OF TEST: 05/22/2022 00:33:16

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.110	00:28
Breath Volume	1.742	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.111	00:31
Breath Volume	1.605	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SOMMERS,EDWARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/05/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/22/2022 01:51:33

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.189	01:46
Breath Volume	2.558	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.198	01:49
Breath Volume	4.019	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 05/23/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: NEMEC

DATE/TIME OF TEST: 05/23/2022 18:38:27

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.080	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.131	18:33
Breath Volume	1.781	18:33
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.121	18:36
Breath Volume	1.429	18:36
Air Blank	0.000	18:37
Control Test	0.079	18:37
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 06/08/2022 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: SAINT FLEUR

DATE/TIME OF TEST: 05/24/2022 21:23:18

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:20	
Control Test	0.079	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:21	Subject Test Refused
Breath Volume	0.000	21:21	
Air Blank	0.000	21:22	
Control Test	0.078	21:22	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 06/08/2022 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 05/25/2022 02:47:03

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.079	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.080	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 06/08/2022 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: CHESSER

DATE/TIME OF TEST: 05/27/2022 22:13:24

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:08
Subject Sample # 1	0.333	22:08
Breath Volume	1.792	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:10
Subject Sample # 2	0.342	22:11
Breath Volume	1.921	22:11
Air Blank	0.000	22:12
Control Test	0.079	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 06/08/2022 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JAYDA,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 05/28/2022 00:32:15

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.080	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:30	Subject Test Refused
Breath Volume	0.156	00:30	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 06/08/2022 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JAYDA,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 05/28/2022 00:51:45

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.079	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.281	00:46
Breath Volume	1.480	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.285	00:49
Breath Volume	1.757	00:49
Air Blank	0.000	00:50
Control Test	0.077	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 06/08/2022 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DZUGAN,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DZUGAN,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: LOPEZ JUAREZ

DATE/TIME OF TEST: 05/28/2022 04:19:34

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.176	04:14
Breath Volume	2.062	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.173	04:17
Breath Volume	1.468	04:17
Air Blank	0.000	04:18
Control Test	0.075	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 06/08/2022 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: STEPHENS MOBERLY

DATE/TIME OF TEST: 05/29/2022 03:29:37

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.164	03:24
Breath Volume	1.671	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.157	03:28
Breath Volume	1.109	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 06/08/2022 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZUNIGA, STEVENS, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: ZUNIGA, STEVENS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: MENDEZ CARDONA

DATE/TIME OF TEST: 05/29/2022 06:43:24

Diagnostics Check	OK	06:31	
Air Blank	0.000	06:31	
Control Test	0.080	06:32	
Air Blank	0.000	06:32	
Subject Sample	# 1 0.133	06:35	Volume Not Met
Breath Volume	0.097	06:35	
Air Blank	0.000	06:36	
Air Blank	0.000	06:38	
Subject Sample	# 2 0.168	06:41	Volume Not Met
Breath Volume	0.121	06:41	
Air Blank	0.000	06:42	
Control Test	0.076	06:42	
Air Blank	0.000	06:43	
Diagnostics Check	OK	06:43	

Data Download Date/Time: 06/08/2022 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: PAZ-RODRIGUEZ

DATE/TIME OF TEST: 05/30/2022 04:01:51

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.193	03:57
Breath Volume	2.675	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 2	0.190	04:00
Breath Volume	2.835	04:00
Air Blank	0.000	04:00
Control Test	0.077	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 06/08/2022 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLINGAME,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BURLINGAME,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: CAMEAU

DATE/TIME OF TEST: 05/30/2022 16:09:49

Diagnostics Check	OK	16:02
Air Blank	0.000	16:03
Control Test	0.080	16:03
Air Blank	0.000	16:03
Subject Sample # 1	0.130	16:04
Breath Volume	3.250	16:04
Air Blank	0.000	16:05
Air Blank	0.000	16:07
Subject Sample # 2	0.127	16:08
Breath Volume	2.804	16:08
Air Blank	0.000	16:08
Control Test	0.080	16:09
Air Blank	0.000	16:09
Diagnostics Check	OK	16:09

Data Download Date/Time: 06/08/2022 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANPELT,J,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WEISS,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 05/30/2022 21:34:14

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:28
Subject Sample # 1	0.255	21:29
Breath Volume	1.394	21:29
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.249	21:32
Breath Volume	1.621	21:32
Air Blank	0.000	21:33
Control Test	0.078	21:33
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 06/08/2022 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/01/2022 01:52:30

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.088	01:47
Breath Volume	1.718	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.103	01:50
Breath Volume	1.343	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 05/26/2022 00:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,PAUL,S
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GONZALES,PAUL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CASALLAS

DATE/TIME OF TEST: 05/01/2022 02:06:59

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.078	01:59	
Air Blank	0.000	01:59	
Subject Sample # 1	0.078	02:02	
Breath Volume	2.433	02:02	
Air Blank	0.000	02:03	
Air Blank	0.000	02:04	
Subject Sample # 2	0.077	02:05	
Breath Volume	1.917	02:05	
Air Blank	0.000	02:06	
Control Test	0.074	02:06	Control Outside Tolerance
Air Blank	0.000	02:06	

Data Download Date/Time: 05/26/2022 00:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,PAUL,S
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GONZALES,PAUL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CASALLAS

DATE/TIME OF TEST: 05/01/2022 02:19:59

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.077	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.080	02:15
Breath Volume	2.589	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.077	02:18
Breath Volume	2.207	02:18
Air Blank	0.000	02:19
Control Test	0.076	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 05/26/2022 00:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,PAUL,S
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GONZALES,PAUL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: REA

DATE/TIME OF TEST: 05/02/2022 01:51:29

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.078	01:39	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:43	No Sample Provided
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Air Blank	0.000	01:45	
Subject Sample	# 2	0.154	01:46
Breath Volume	1.417	01:46	
Air Blank	0.000	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 3	0.157	01:49
Breath Volume	1.566	01:49	
Air Blank	0.000	01:50	
Control Test	0.076	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 05/26/2022 00:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: STINSON

DATE/TIME OF TEST: 05/06/2022 00:36:31

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.240	00:31
Breath Volume	2.925	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.235	00:34
Breath Volume	2.667	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 05/26/2022 00:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/06/2022 04:14:25

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.103	04:09
Breath Volume	3.925	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.098	04:12
Breath Volume	3.710	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 05/26/2022 00:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,HENRY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: COWART

DATE/TIME OF TEST: 05/06/2022 19:58:35

Diagnostics Check	OK	19:50
Air Blank	0.000	19:51
Control Test	0.079	19:51
Air Blank	0.000	19:52
Subject Sample # 1	0.053	19:53
Breath Volume	1.332	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.054	19:57
Breath Volume	1.289	19:57
Air Blank	0.000	19:57
Control Test	0.079	19:57
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: IVANKO

DATE/TIME OF TEST: 05/06/2022 21:27:06

Diagnostics Check	OK	21:19
Air Blank	0.000	21:20
Control Test	0.079	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.000	21:22
Breath Volume	2.699	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.000	21:25
Breath Volume	3.003	21:25
Air Blank	0.000	21:26
Control Test	0.078	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: BISET-MORALES

DATE/TIME OF TEST: 05/06/2022 23:48:26

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:41	No Sample Provided
Breath Volume	0.000	23:41	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample	# 2	23:47	No Sample Provided
Breath Volume	0.000	23:47	
Air Blank	0.000	23:47	
Control Test	0.079	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,HENRY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 05/07/2022 20:04:05

Diagnostics Check	OK	19:56
Air Blank	0.000	19:57
Control Test	0.079	19:57
Air Blank	0.000	19:58
Subject Sample # 1	0.057	19:59
Breath Volume	1.914	19:59
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.054	20:02
Breath Volume	1.781	20:02
Air Blank	0.000	20:03
Control Test	0.078	20:03
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: SHEEHAN

DATE/TIME OF TEST: 05/07/2022 21:58:05

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.079	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.036	21:53
Breath Volume	2.574	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:55
Subject Sample # 2	0.034	21:56
Breath Volume	2.785	21:56
Air Blank	0.000	21:57
Control Test	0.078	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: FRAHS

DATE/TIME OF TEST: 05/08/2022 21:27:47

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.079	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.215	21:23
Breath Volume	2.585	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.212	21:26
Breath Volume	2.078	21:26
Air Blank	0.000	21:26
Control Test	0.079	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,S,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,S
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 05/11/2022 17:50:43

Diagnostics Check	OK	17:43
Air Blank	0.000	17:43
Control Test	0.079	17:44
Air Blank	0.000	17:44
Subject Sample # 1	0.000	17:46
Breath Volume	2.304	17:46
Air Blank	0.000	17:46
Air Blank	0.000	17:48
Subject Sample # 2	0.000	17:49
Breath Volume	1.851	17:49
Air Blank	0.000	17:49
Control Test	0.077	17:50
Air Blank	0.000	17:50
Diagnostics Check	OK	17:50

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,HENRY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 05/11/2022 21:31:03

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:16	
Control Test	0.079	21:16	
Air Blank	0.000	21:17	
Subject Sample	# 1	21:20	No Sample Provided
Breath Volume	0.000	21:20	
Air Blank	0.000	21:20	
Air Blank	0.000	21:22	
Subject Sample	# 2	0.211	21:25
Breath Volume	1.621	21:25	
Air Blank	0.000	21:26	
Air Blank	0.000	21:28	
Subject Sample	# 3	0.205	21:29
Breath Volume	1.375	21:29	
Air Blank	0.000	21:30	
Control Test	0.079	21:30	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:31	

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIARS,KELSEY,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VIARS,KELSEY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: GOHSTAT

DATE/TIME OF TEST: 05/12/2022 04:25:29

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.080	04:12	
Air Blank	0.000	04:13	
Subject Sample # 1	0.178	04:14	
Breath Volume	1.195	04:14	
Air Blank	0.000	04:15	
Air Blank	0.000	04:17	
Subject Sample # 2	0.163	04:20	Volume Not Met
Breath Volume	0.390	04:20	
Air Blank	0.000	04:21	
Air Blank	0.000	04:22	
Subject Sample # 3	0.183	04:23	
Breath Volume	1.683	04:23	
Air Blank	0.000	04:24	
Control Test	0.077	04:24	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: M635201053910

DATE/TIME OF TEST: 05/13/2022 03:03:48

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.148	02:58
Breath Volume	1.714	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.141	03:02
Breath Volume	1.109	03:02
Air Blank	0.000	03:02
Control Test	0.078	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,S,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,S
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: HAUCK

DATE/TIME OF TEST: 05/17/2022 18:51:27

Diagnostics Check	OK	18:43
Air Blank	0.000	18:44
Control Test	0.079	18:44
Air Blank	0.000	18:44
Subject Sample # 1	0.239	18:45
Breath Volume	4.519	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample # 2	0.240	18:49
Breath Volume	2.621	18:49
Air Blank	0.000	18:50
Control Test	0.077	18:50
Air Blank	0.000	18:51
Diagnostics Check	OK	18:51

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,LAMON,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: ARROYAVE

DATE/TIME OF TEST: 05/18/2022 01:11:26

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.078	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.184	01:06
Breath Volume	2.820	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.176	01:09
Breath Volume	3.109	01:09
Air Blank	0.000	01:10
Control Test	0.076	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 05/20/2022 04:34:10

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.179	04:29
Breath Volume	1.117	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.174	04:32
Breath Volume	1.558	04:32
Air Blank	0.000	04:33
Control Test	0.078	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: MEGAHAN

DATE/TIME OF TEST: 05/21/2022 00:14:06

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.079	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.000	00:09
Breath Volume	2.917	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.000	00:12
Breath Volume	2.472	00:12
Air Blank	0.000	00:13
Control Test	0.078	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 05/26/2022 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 05/21/2022 03:30:36

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.157	03:25
Breath Volume	3.281	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.153	03:29
Breath Volume	2.710	03:29
Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 05/26/2022 01:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: EILER

DATE/TIME OF TEST: 05/22/2022 00:38:30

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.325	00:33
Breath Volume	3.660	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.324	00:36
Breath Volume	2.742	00:36
Air Blank	0.000	00:37
Control Test	0.078	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 05/26/2022 01:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,PAUL,S
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GONZALES,PAUL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: POLLEY

DATE/TIME OF TEST: 05/22/2022 21:05:16

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.079	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.115	21:00
Breath Volume	3.347	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.111	21:03
Breath Volume	3.910	21:03
Air Blank	0.000	21:04
Control Test	0.076	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 05/26/2022 01:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,S,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,S
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: KRUG

DATE/TIME OF TEST: 05/23/2022 07:09:38

Diagnostics Check	OK	07:02
Air Blank	0.000	07:03
Control Test	0.079	07:03
Air Blank	0.000	07:04
Subject Sample # 1	0.138	07:04
Breath Volume	3.371	07:04
Air Blank	0.000	07:05
Air Blank	0.000	07:07
Subject Sample # 2	0.141	07:08
Breath Volume	3.679	07:08
Air Blank	0.000	07:08
Control Test	0.077	07:09
Air Blank	0.000	07:09
Diagnostics Check	OK	07:09

Data Download Date/Time: 05/26/2022 01:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220213
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: DYER

DATE/TIME OF TEST: 05/05/2022 00:31:08

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.143	00:26
Breath Volume	2.738	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.149	00:29
Breath Volume	2.371	00:29
Air Blank	0.000	00:30
Control Test	0.076	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 05/10/2022 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220213
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: GUZMAN MEDINA

DATE/TIME OF TEST: 05/11/2022 20:53:37

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.081	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.239	20:48
Breath Volume	2.007	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.234	20:51
Breath Volume	1.691	20:51
Air Blank	0.000	20:52
Control Test	0.079	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 06/20/2022 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220213
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: VANHOLTEN

DATE/TIME OF TEST: 05/12/2022 19:01:34

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.080	18:55
Air Blank	0.000	18:56
Subject Sample # 1	0.000	18:57
Breath Volume	4.015	18:57
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample # 2	0.000	19:00
Breath Volume	2.574	19:00
Air Blank	0.000	19:00
Control Test	0.080	19:00
Air Blank	0.000	19:01
Diagnostics Check	OK	19:01

Data Download Date/Time: 06/20/2022 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: DANKE

DATE/TIME OF TEST: 05/14/2022 23:18:33

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.132	23:13
Breath Volume	2.359	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.116	23:16
Breath Volume	1.417	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 06/20/2022 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: GRAVE ACABAL

DATE/TIME OF TEST: 05/15/2022 18:17:33

Diagnostics Check	OK	18:11
Air Blank	0.000	18:11
Control Test	0.080	18:11
Air Blank	0.000	18:12
Subject Sample # 1	0.279	18:12
Breath Volume	2.132	18:12
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.273	18:15
Breath Volume	2.265	18:15
Air Blank	0.000	18:16
Control Test	0.079	18:16
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

Data Download Date/Time: 06/20/2022 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: BOLEN

DATE/TIME OF TEST: 05/16/2022 19:25:56

Diagnostics Check	OK	19:19
Air Blank	0.000	19:20
Control Test	0.080	19:20
Air Blank	0.000	19:20
Subject Sample # 1	0.000	19:21
Breath Volume	3.195	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.000	19:24
Breath Volume	4.113	19:24
Air Blank	0.000	19:24
Control Test	0.079	19:25
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 06/20/2022 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 05/28/2022 01:25:58

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.103	01:21
Breath Volume	2.210	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.099	01:24
Breath Volume	2.019	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 06/20/2022 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/06/2022 03:24:11

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.078	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.146	03:19
Breath Volume	1.792	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.140	03:22
Breath Volume	2.511	03:22
Air Blank	0.000	03:23
Control Test	0.077	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 05/18/2022 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: CANADY

DATE/TIME OF TEST: 05/24/2022 23:18:05

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.079	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.067	23:13
Breath Volume	3.742	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:15
Subject Sample # 2	0.068	23:16
Breath Volume	3.882	23:16
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:18

Data Download Date/Time: 06/27/2022 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,ANDREW,E
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: MCCRAY,CHELSEA
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 04/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/11/2022 04:14:28

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.081	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.176	04:08
Breath Volume	2.531	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.169	04:12
Breath Volume	1.804	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 05/12/2022 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MEDINA,ALEXIS,

BREATH TEST OPERATOR AGENCY: HIALEAH PD

ARRESTING OFFICER: HERNANDEZ,OSDANY

ARRESTING OFFICER AGENCY: HIALEAH PD

LAST AGENCY INSPECTION: 04/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1224982

DRY GAS EXPIRY DATE: 10/31/2022

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BRICENO ROMERO

DATE/TIME OF TEST: 05/11/2022 04:43:46

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:41	
Control Test	0.076	04:41	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:42	Subject Test Refused
Breath Volume	0.000	04:42	
Air Blank	0.000	04:42	
Control Test	0.076	04:43	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 06/23/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAMAYO,KEVIN,X
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GIL,GEORGE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ROLLE

DATE/TIME OF TEST: 05/02/2022 00:27:09

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.142	00:22
Breath Volume	3.300	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.138	00:25
Breath Volume	2.691	00:25
Air Blank	0.000	00:26
Control Test	0.077	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 06/01/2022 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 05/12/2022 20:59:52

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.078	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.210	20:54
Breath Volume	1.664	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.218	20:58
Breath Volume	1.234	20:58
Air Blank	0.000	20:58
Control Test	0.077	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 05/23/2022 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: EASSA

DATE/TIME OF TEST: 05/14/2022 18:07:12

Diagnostics Check	OK	18:00
Air Blank	0.000	18:01
Control Test	0.078	18:01
Air Blank	0.000	18:02
Subject Sample # 1	0.210	18:02
Breath Volume	3.757	18:02
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 2	0.201	18:05
Breath Volume	2.906	18:05
Air Blank	0.000	18:06
Control Test	0.079	18:06
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

Data Download Date/Time: 05/23/2022 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: POPPLETON

DATE/TIME OF TEST: 05/20/2022 23:35:41

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.077	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.183	23:31
Breath Volume	3.117	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.173	23:34
Breath Volume	1.769	23:34
Air Blank	0.000	23:34
Control Test	0.077	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 05/23/2022 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: POSTON

DATE/TIME OF TEST: 05/28/2022 00:08:04

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.078	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.116	00:03
Breath Volume	3.414	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:05
Subject Sample # 2	0.109	00:06
Breath Volume	3.003	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:08

Data Download Date/Time: 06/19/2022 15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: THOMAS,MEGAN,A

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: THOMAS,MEGAN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 05/02/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG129201

DRY GAS EXPIRY DATE: 10/19/2023

VIOLATION CODE: BUI

ARREST TIME: 13:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: THIEME

DATE/TIME OF TEST: 05/07/2022 16:35:01

Diagnostics Check	OK	16:28
Air Blank	0.000	16:29
Control Test	0.079	16:29
Air Blank	0.000	16:30
Subject Sample # 1	0.098	16:30
Breath Volume	3.460	16:30
Air Blank	0.000	16:31
Air Blank	0.000	16:33
Subject Sample # 2	0.101	16:33
Breath Volume	3.371	16:33
Air Blank	0.000	16:34
Control Test	0.079	16:34
Air Blank	0.000	16:34
Diagnostics Check	OK	16:35

Data Download Date/Time: 06/16/2022 17:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NORTH,DUANE,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: LAWRENCE,JOSHUA

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 05/02/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG129201

DRY GAS EXPIRY DATE: 10/19/2023

VIOLATION CODE: BUI

ARREST TIME: 16:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: HEIDER-WEST

DATE/TIME OF TEST: 05/07/2022 17:37:56

Diagnostics Check	OK	17:35	
Air Blank	0.000	17:35	
Control Test	0.079	17:35	
Air Blank	0.000	17:36	
Subject Sample	# 1	17:36	Subject Test Refused
Breath Volume	0.000	17:36	
Air Blank	0.000	17:37	
Control Test	0.079	17:37	
Air Blank	0.000	17:37	
Diagnostics Check	OK	17:37	

Data Download Date/Time: 06/16/2022 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ARAFA,YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/01/2022 01:18:27

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.177	01:13
Breath Volume	2.222	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample # 2	0.179	01:16
Breath Volume	2.554	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 05/17/2022 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 05/01/2022 02:45:14

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.087	02:40
Breath Volume	2.058	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.084	02:43
Breath Volume	1.582	02:43
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 05/17/2022 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: BASFORD,NICOLE,M	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BOWER,RACHEL	ARREST TIME: 14:56
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Female
LAST AGENCY INSPECTION: 04/19/2022	OBSERVATION PERIOD BEGAN: 14:56

SUBJECT'S LAST NAME: B656116727540 DATE/TIME OF TEST: 05/14/2022 15:28:01

Diagnostics Check	OK	15:16	
Air Blank	0.000	15:16	
Control Test	0.079	15:17	
Air Blank	0.000	15:17	
Subject Sample	# 1 0.121	15:20	Volume Not Met
Breath Volume	0.167	15:20	
Air Blank	0.000	15:21	
Air Blank	0.000	15:23	
Subject Sample	# 2 0.126	15:26	Volume Not Met
Breath Volume	0.519	15:26	
Air Blank	0.000	15:27	
Control Test	0.079	15:27	
Air Blank	0.000	15:27	
Diagnostics Check	OK	15:28	

Data Download Date/Time: 05/17/2022 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BASFORD,NICOLE,M

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: BOWER,RACHEL

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: BUI

ARREST TIME: 14:56

GENDER: Female

OBSERVATION PERIOD BEGAN: 14:56

SUBJECT'S LAST NAME: BARNHART

DATE/TIME OF TEST: 05/14/2022 15:35:55

Diagnostics Check	OK	15:32	
Air Blank	0.000	15:33	
Control Test	0.078	15:33	
Air Blank	0.000	15:34	
Subject Sample	# 1	15:34	Subject Test Refused
Breath Volume	0.000	15:34	
Air Blank	0.000	15:35	
Control Test	0.079	15:35	
Air Blank	0.000	15:35	
Diagnostics Check	OK	15:35	

Data Download Date/Time: 05/17/2022 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: BARTLETT,SHELTON,T	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BARTLETT,SHELTON	ARREST TIME: 15:29
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 04/19/2022	OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: HOOPER DATE/TIME OF TEST: 05/14/2022 15:56:10

Diagnostics Check	OK	15:49
Air Blank	0.000	15:49
Control Test	0.078	15:50
Air Blank	0.000	15:50
Subject Sample # 1	0.262	15:51
Breath Volume	3.019	15:51
Air Blank	0.000	15:52
Air Blank	0.000	15:54
Subject Sample # 2	0.267	15:54
Breath Volume	2.113	15:54
Air Blank	0.000	15:55
Control Test	0.078	15:55
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

Data Download Date/Time: 05/17/2022 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NICHOLS, HEATH, C

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: PRITCHARD, THOMAS

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: BUI

ARREST TIME: 15:53

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: MARENGO

DATE/TIME OF TEST: 05/14/2022 16:25:42

Diagnostics Check	OK	16:22	
Air Blank	0.000	16:23	
Control Test	0.079	16:23	
Air Blank	0.000	16:24	
Subject Sample	# 1	16:24	Subject Test Refused
Breath Volume	0.000	16:24	
Air Blank	0.000	16:24	
Control Test	0.079	16:25	
Air Blank	0.000	16:25	
Diagnostics Check	OK	16:25	

Data Download Date/Time: 05/17/2022 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: NICHOLS, HEATH, C	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: OLIVER, WILLIAM	ARREST TIME: 17:05
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 04/19/2022	OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: DUVAL	DATE/TIME OF TEST: 05/14/2022 17:33:41
----------------------------	--

Diagnostics Check	OK	17:30	
Air Blank	0.000	17:31	
Control Test	0.079	17:31	
Air Blank	0.000	17:32	
Subject Sample	# 1	17:32	Subject Test Refused
Breath Volume	0.000	17:32	
Air Blank	0.000	17:32	
Control Test	0.078	17:33	
Air Blank	0.000	17:33	
Diagnostics Check	OK	17:33	

Data Download Date/Time: 05/17/2022 10:43

Subject Test Electronic Data

DATE/TIME OF TEST: 05/28/2022 17:17:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: BASFORD,NICOLE,M	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BASFORD,NICOLE	ARREST TIME: 17:12
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/16/2022	OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: PATE	DATE/TIME OF TEST: 05/30/2022	19:42:38
---------------------------	-------------------------------	----------

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.135	19:37
Breath Volume	2.867	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.141	19:41
Breath Volume	2.964	19:41
Air Blank	0.000	19:41
Control Test	0.079	19:42
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 06/21/2022 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: BASFORD, TRAVIS, L	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BASFORD, TRAVIS	ARREST TIME: 17:38
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/16/2022	OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: JOHNSON	DATE/TIME OF TEST: 05/30/2022	22:01:57
------------------------------	-------------------------------	----------

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.079	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.138	21:57
Breath Volume	1.855	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.132	22:00
Breath Volume	1.265	22:00
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 06/21/2022 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN-
ROSQUETE,STEPHANNE,J
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: BROWN-
ROSQUETE,STEPHANNE
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022

VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male

OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: PAZ

DATE/TIME OF TEST: 05/02/2022 00:15:13

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.080	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.196	00:10
Breath Volume	1.839	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.194	00:13
Breath Volume	1.515	00:13
Air Blank	0.000	00:14
Control Test	0.076	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 05/05/2022 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGODA,JOSHUA,S
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: DARIAS,SOFIA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: CUEVAS

DATE/TIME OF TEST: 05/13/2022 20:46:13

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.080	20:38
Air Blank	0.000	20:38
Subject Sample # 1	0.192	20:39
Breath Volume	1.953	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.185	20:44
Breath Volume	1.359	20:44
Air Blank	0.000	20:45
Control Test	0.079	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 06/15/2022 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: LINS,GIOVANE
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: SUTHERLAND

DATE/TIME OF TEST: 05/14/2022 05:52:59

Diagnostics Check	OK	05:49	
Air Blank	0.000	05:50	
Control Test	0.080	05:50	
Air Blank	0.000	05:51	
Subject Sample	# 1	05:51	Subject Test Refused
Breath Volume	0.000	05:51	
Air Blank	0.000	05:52	
Control Test	0.080	05:52	
Air Blank	0.000	05:52	
Diagnostics Check	OK	05:52	

Data Download Date/Time: 06/15/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE,ALEXANDRA,P
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: PEREZ,OMAR
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: JOMARRON JR

DATE/TIME OF TEST: 05/23/2022 20:51:57

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.080	20:46
Air Blank	0.000	20:46
Subject Sample # 1	0.148	20:47
Breath Volume	1.339	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:49
Subject Sample # 2	0.165	20:50
Breath Volume	1.718	20:50
Air Blank	0.000	20:50
Control Test	0.078	20:51
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 06/15/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: SOCARRAS,PEDRO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 05/26/2022 21:03:11

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.238	20:58
Breath Volume	2.199	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:00
Subject Sample # 2	0.235	21:01
Breath Volume	1.835	21:01
Air Blank	0.000	21:02
Control Test	0.079	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 06/15/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCLJOSE,F
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: WOODSTOCK,KENDRA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 05/27/2022 04:28:10

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.000	04:22
Breath Volume	1.882	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.000	04:26
Breath Volume	2.769	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 06/15/2022 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SCHERRER

DATE/TIME OF TEST: 05/27/2022 23:39:09

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:36	
Control Test	0.080	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:37	Subject Test Refused
Breath Volume	0.000	23:37	
Air Blank	0.000	23:38	
Control Test	0.080	23:38	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 06/15/2022 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGODA,JOSHUA,S
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: FREDERICQUE

DATE/TIME OF TEST: 05/28/2022 23:50:36

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.080	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 1 0.182	23:39	Volume Not Met
Breath Volume	0.593	23:39	
Air Blank	0.000	23:40	
Air Blank	0.000	23:42	
Subject Sample	# 2 0.222	23:43	
Breath Volume	1.765	23:43	
Air Blank	0.000	23:44	
Air Blank	0.000	23:46	
Subject Sample	# 3 0.210	23:48	
Breath Volume	1.707	23:48	
Air Blank	0.000	23:49	
Control Test	0.077	23:49	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 06/15/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: LOPEZPARRA

DATE/TIME OF TEST: 05/27/2022 23:30:20

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.185	23:24
Breath Volume	1.699	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.182	23:28
Breath Volume	1.300	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 07/05/2022 08:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VAZQUEZ,V
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: KNOTTS

DATE/TIME OF TEST: 05/01/2022 00:22:53

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.078	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.174	00:14
Breath Volume	2.039	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.196	00:18
Breath Volume	2.468	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 3	0.184	00:21
Breath Volume	2.046	00:21
Air Blank	0.000	00:21
Control Test	0.076	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 05/17/2022 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: COMM VEHIC
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 05/01/2022 02:29:50

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.078	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.107	02:24
Breath Volume	2.277	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.110	02:28
Breath Volume	2.312	02:28
Air Blank	0.000	02:28
Control Test	0.078	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 05/17/2022 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/01/2022 02:42:19

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.078	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.165	02:37
Breath Volume	3.492	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.174	02:40
Breath Volume	3.851	02:40
Air Blank	0.000	02:41
Control Test	0.077	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 05/17/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MATUTE BUESO

DATE/TIME OF TEST: 05/01/2022 04:15:32

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.079	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.105	04:10
Breath Volume	2.132	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.100	04:13
Breath Volume	1.210	04:13
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 05/17/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: EDINGER

DATE/TIME OF TEST: 05/01/2022 05:02:45

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.077	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.230	04:57
Breath Volume	3.257	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.214	05:01
Breath Volume	2.187	05:01
Air Blank	0.000	05:01
Control Test	0.077	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 05/17/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: UBAH

DATE/TIME OF TEST: 05/01/2022 07:57:21

Diagnostics Check	OK	07:45	
Air Blank	0.000	07:46	
Control Test	0.078	07:46	
Air Blank	0.000	07:47	
Subject Sample	# 1 0.065	07:50	Volume Not Met
Breath Volume	0.480	07:50	
Air Blank	0.000	07:50	
Air Blank	0.000	07:52	
Subject Sample	# 2 0.066	07:55	Volume Not Met
Breath Volume	1.066	07:55	
Air Blank	0.000	07:56	
Control Test	0.077	07:56	
Air Blank	0.000	07:57	
Diagnostics Check	OK	07:57	

Data Download Date/Time: 05/17/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: UBAH

DATE/TIME OF TEST: 05/01/2022 08:08:23

Diagnostics Check	OK	08:01
Air Blank	0.000	08:02
Control Test	0.076	08:02
Air Blank	0.000	08:02
Subject Sample # 1	0.070	08:03
Breath Volume	1.757	08:03
Air Blank	0.000	08:03
Air Blank	0.000	08:05
Subject Sample # 2	0.067	08:06
Breath Volume	1.730	08:06
Air Blank	0.000	08:07
Control Test	0.077	08:07
Air Blank	0.000	08:08
Diagnostics Check	OK	08:08

Data Download Date/Time: 05/17/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: ALVARADO MORALES

DATE/TIME OF TEST: 05/01/2022 20:26:36

Diagnostics Check	OK	20:19
Air Blank	0.000	20:20
Control Test	0.078	20:20
Air Blank	0.000	20:21
Subject Sample # 1	0.137	20:21
Breath Volume	1.542	20:21
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.146	20:24
Breath Volume	1.500	20:24
Air Blank	0.000	20:25
Control Test	0.078	20:26
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 05/17/2022 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: VIDAL

DATE/TIME OF TEST: 05/01/2022 23:50:42

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.078	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.175	23:46
Breath Volume	1.917	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.164	23:49
Breath Volume	1.289	23:49
Air Blank	0.000	23:49
Control Test	0.078	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 05/17/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 05/02/2022 02:33:30

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.229	02:28
Breath Volume	1.304	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample # 2	0.233	02:31
Breath Volume	1.640	02:31
Air Blank	0.000	02:32
Control Test	0.077	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 05/17/2022 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 05/02/2022 03:27:47

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.150	03:23
Breath Volume	1.386	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.156	03:26
Breath Volume	1.761	03:26
Air Blank	0.000	03:26
Control Test	0.077	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 05/17/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: BABIN

DATE/TIME OF TEST: 05/02/2022 05:50:35

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.079	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.125	05:46
Breath Volume	2.023	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.122	05:49
Breath Volume	1.968	05:49
Air Blank	0.000	05:49
Control Test	0.078	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 05/17/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: RUDOLPH

DATE/TIME OF TEST: 05/03/2022 23:00:50

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.080	22:55
Air Blank	0.000	22:55
Subject Sample # 1	0.154	22:56
Breath Volume	1.250	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.153	22:59
Breath Volume	1.464	22:59
Air Blank	0.000	22:59
Control Test	0.078	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 05/17/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MOQADDEM

DATE/TIME OF TEST: 05/04/2022 01:50:52

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.249	01:45
Breath Volume	1.625	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.244	01:49
Breath Volume	1.867	01:49
Air Blank	0.000	01:49
Control Test	0.078	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 05/17/2022 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JOHNSON,B
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 05/04/2022 20:31:03

Diagnostics Check	OK	20:24
Air Blank	0.000	20:24
Control Test	0.079	20:25
Air Blank	0.000	20:25
Subject Sample # 1	0.147	20:26
Breath Volume	3.093	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:28
Subject Sample # 2	0.151	20:29
Breath Volume	3.457	20:29
Air Blank	0.000	20:30
Control Test	0.078	20:30
Air Blank	0.000	20:30
Diagnostics Check	OK	20:31

Data Download Date/Time: 05/17/2022 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: LEORDEAN

DATE/TIME OF TEST: 05/05/2022 02:36:29

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.079	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Control Test	0.078	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 05/17/2022 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BOYKINS

DATE/TIME OF TEST: 05/06/2022 01:38:31

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.079	01:35	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.078	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 05/17/2022 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBERTS,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: BORGES ALBINO

DATE/TIME OF TEST: 05/06/2022 05:04:47

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.100	05:00
Breath Volume	3.417	05:00
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.107	05:03
Breath Volume	3.425	05:03
Air Blank	0.000	05:03
Control Test	0.079	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 05/17/2022 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: GIFFORD

DATE/TIME OF TEST: 05/06/2022 23:40:24

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.173	23:33
Breath Volume	1.679	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:35
Subject Sample # 2	0.174	23:38
Breath Volume	2.164	23:38
Air Blank	0.000	23:39
Control Test	0.078	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 05/17/2022 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: TREJO

DATE/TIME OF TEST: 05/07/2022 00:51:34

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.079	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.219	00:46
Breath Volume	3.562	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.217	00:49
Breath Volume	2.851	00:49
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 05/17/2022 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 05/07/2022 03:00:21

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.077	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.185	02:55
Breath Volume	2.296	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.186	02:58
Breath Volume	2.335	02:58
Air Blank	0.000	02:59
Control Test	0.077	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 05/17/2022 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 05/07/2022 03:27:39

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.078	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.127	03:22
Breath Volume	1.742	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.124	03:26
Breath Volume	1.726	03:26
Air Blank	0.000	03:26
Control Test	0.078	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 05/17/2022 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: CONTRERAS MEDRANO

DATE/TIME OF TEST: 05/07/2022 05:24:11

Diagnostics Check	OK	05:16
Air Blank	0.000	05:17
Control Test	0.078	05:17
Air Blank	0.000	05:18
Subject Sample # 1	0.176	05:19
Breath Volume	2.289	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.173	05:22
Breath Volume	2.371	05:22
Air Blank	0.000	05:23
Control Test	0.076	05:23
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

Data Download Date/Time: 05/17/2022 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: MARTINEZ JEREZ

DATE/TIME OF TEST: 05/07/2022 19:12:05

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.077	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.323	19:07
Breath Volume	1.210	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:09
Subject Sample # 2	0.323	19:10
Breath Volume	1.652	19:10
Air Blank	0.000	19:11
Control Test	0.077	19:11
Air Blank	0.000	19:11
Diagnostics Check	OK	19:12

Data Download Date/Time: 05/17/2022 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/07/2022 23:35:15

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.060	23:30
Breath Volume	1.292	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.054	23:33
Breath Volume	2.523	23:33
Air Blank	0.000	23:34
Control Test	0.077	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 05/17/2022 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: FERAZZOLI

DATE/TIME OF TEST: 05/08/2022 02:12:29

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.166	02:07
Breath Volume	1.175	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.175	02:10
Breath Volume	1.792	02:10
Air Blank	0.000	02:11
Control Test	0.077	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 05/17/2022 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/08/2022 03:56:01

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.077	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.284	03:51
Breath Volume	2.636	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.293	03:54
Breath Volume	2.726	03:54
Air Blank	0.000	03:55
Control Test	0.076	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:56

Data Download Date/Time: 05/17/2022 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WINTER,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: LAPOLLO

DATE/TIME OF TEST: 05/08/2022 04:12:04

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.077	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.193	04:07
Breath Volume	2.636	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.197	04:10
Breath Volume	2.742	04:10
Air Blank	0.000	04:11
Control Test	0.076	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:12

Data Download Date/Time: 05/17/2022 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 05/08/2022 04:24:51

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.077	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.043	04:20
Breath Volume	1.113	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.046	04:23
Breath Volume	1.351	04:23
Air Blank	0.000	04:23
Control Test	0.078	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 05/17/2022 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: MILLIEN

DATE/TIME OF TEST: 05/08/2022 06:14:59

Diagnostics Check	OK	06:08
Air Blank	0.000	06:08
Control Test	0.077	06:08
Air Blank	0.000	06:09
Subject Sample # 1	0.153	06:10
Breath Volume	2.574	06:10
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.153	06:13
Breath Volume	1.527	06:13
Air Blank	0.000	06:14
Control Test	0.078	06:14
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

Data Download Date/Time: 05/17/2022 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: HERNANDEZ GONZALEZ

DATE/TIME OF TEST: 05/08/2022 23:59:11

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.078	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.196	23:50
Breath Volume	1.250	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.228	23:54
Breath Volume	1.523	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:56
Subject Sample # 3	0.202	23:57
Breath Volume	1.675	23:57
Air Blank	0.000	23:58
Control Test	0.078	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 05/17/2022 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LONGMIRE,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 05/09/2022 18:13:44

Diagnostics Check	OK	18:06
Air Blank	0.000	18:07
Control Test	0.079	18:07
Air Blank	0.000	18:08
Subject Sample # 1	0.267	18:09
Breath Volume	3.054	18:09
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.263	18:12
Breath Volume	3.250	18:12
Air Blank	0.000	18:12
Control Test	0.078	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 05/17/2022 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PARKER,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: O'CONNELL

DATE/TIME OF TEST: 05/09/2022 23:32:20

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:24	
Control Test	0.080	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 1	23:28	No Sample Provided
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Air Blank	0.000	23:30	
Subject Sample	# 2	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 05/17/2022 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY, JANICE, F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: TORRES ROCAFORT

DATE/TIME OF TEST: 05/12/2022 02:19:57

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.148	02:14
Breath Volume	3.578	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.153	02:18
Breath Volume	4.226	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 05/17/2022 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEYERS,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: HANNA

DATE/TIME OF TEST: 05/12/2022 03:03:17

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.265	02:58
Breath Volume	1.652	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.268	03:01
Breath Volume	1.324	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 05/17/2022 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: NAJARRO

DATE/TIME OF TEST: 05/14/2022 03:34:13

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.139	03:29
Breath Volume	2.238	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.145	03:32
Breath Volume	2.550	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 05/17/2022 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: JOSE GARCIA

DATE/TIME OF TEST: 05/14/2022 04:57:24

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.186	04:52
Breath Volume	2.703	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.188	04:55
Breath Volume	2.960	04:55
Air Blank	0.000	04:56
Control Test	0.077	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 05/17/2022 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: PASCUAL HERNANDEZ

DATE/TIME OF TEST: 05/14/2022 23:17:52

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.186	23:13
Breath Volume	2.121	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:15
Subject Sample # 2	0.192	23:16
Breath Volume	1.347	23:16
Air Blank	0.000	23:16
Control Test	0.079	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 05/17/2022 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: MATOS TOVAR

DATE/TIME OF TEST: 05/15/2022 02:02:55

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.111	01:58
Breath Volume	1.742	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.111	02:01
Breath Volume	1.648	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 05/17/2022 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BROWNLEE

DATE/TIME OF TEST: 05/16/2022 01:44:39

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.117	01:39
Breath Volume	2.476	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.116	01:43
Breath Volume	2.039	01:43
Air Blank	0.000	01:43
Control Test	0.078	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 05/17/2022 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BORROR

DATE/TIME OF TEST: 05/16/2022 02:33:10

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.156	02:28
Breath Volume	1.117	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.167	02:31
Breath Volume	2.132	02:31
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 05/17/2022 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: PLA

DATE/TIME OF TEST: 05/16/2022 05:07:39

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.079	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.141	05:02
Breath Volume	1.714	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.137	05:06
Breath Volume	1.777	05:06
Air Blank	0.000	05:06
Control Test	0.079	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 05/17/2022 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: TRIVETTE

DATE/TIME OF TEST: 05/16/2022 06:14:28

Diagnostics Check	OK	06:07
Air Blank	0.000	06:08
Control Test	0.079	06:08
Air Blank	0.000	06:08
Subject Sample # 1	0.000	06:09
Breath Volume	1.636	06:09
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.000	06:13
Breath Volume	1.476	06:13
Air Blank	0.000	06:13
Control Test	0.078	06:13
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

Data Download Date/Time: 05/17/2022 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 05/17/2022 02:38:25

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.079	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1 0.198	02:26	No .020 agreement
Breath Volume	1.156	02:26	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample	# 2 0.220	02:31	No .020 agreement
Breath Volume	2.007	02:31	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 3 0.188	02:36	Volume Not Met
Breath Volume	0.070	02:36	
Air Blank	0.000	02:37	
Control Test	0.078	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 05/17/2022 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 05/17/2022 02:49:46

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.076	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.208	02:43
Breath Volume	2.101	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.201	02:48
Breath Volume	1.925	02:48
Air Blank	0.000	02:48
Control Test	0.076	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 05/17/2022 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCKHALT, WILLIAM, H
BREATH TEST OPERATOR AGENCY: Liberty County SO
ARRESTING OFFICER: BUCKHALT, WILLIAM
ARRESTING OFFICER AGENCY: Liberty County SO
LAST AGENCY INSPECTION: 05/16/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: BROUSSARD

DATE/TIME OF TEST: 05/27/2022 23:56:06

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.078	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.105	23:51
Breath Volume	2.277	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.108	23:54
Breath Volume	2.515	23:54
Air Blank	0.000	23:55
Control Test	0.077	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 06/09/2022 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SWEET

DATE/TIME OF TEST: 05/03/2022 02:07:37

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.079	02:02	
Air Blank	0.000	02:02	
Subject Sample	# 1 0.000	02:05	Volume Not Met
Breath Volume	0.617	02:05	
Air Blank		02:06	RFI Detect
Air Blank	0.000	02:06	
Control Test	0.080	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 06/03/2022 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SWEET

DATE/TIME OF TEST: 05/03/2022 02:16:09

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.079	02:11	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:14	RFI Detect
Breath Volume	0.191	02:14	
Air Blank	0.000	02:15	
Control Test	0.080	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 06/03/2022 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: MORRELL

DATE/TIME OF TEST: 05/09/2022 03:31:17

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:29	
Control Test	0.080	03:29	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:30	RFI Detect
Breath Volume	0.039	03:30	
Air Blank		03:30	RFI Detect
Air Blank	0.000	03:31	

Data Download Date/Time: 06/03/2022 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STEPHENS, TONY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 05/01/2022 04:12:38

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.083	04:08
Breath Volume	1.300	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.082	04:11
Breath Volume	2.109	04:11
Air Blank	0.000	04:11
Control Test	0.081	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 05/18/2022 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STEPHENS, TONY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: JABLONSKI

DATE/TIME OF TEST: 05/01/2022 23:16:49

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.080	23:05	
Air Blank	0.000	23:06	
Subject Sample	# 1 0.137	23:09	Volume Not Met
Breath Volume	0.035	23:09	
Air Blank	0.000	23:10	
Air Blank	0.000	23:11	
Subject Sample	# 2 0.136	23:15	Volume Not Met
Breath Volume	0.050	23:15	
Air Blank	0.000	23:15	
Control Test	0.079	23:16	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 05/18/2022 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STEPHENS,TONY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: JABLONSKI

DATE/TIME OF TEST: 05/01/2022 23:34:03

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.078	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1 0.144	23:25	Volume Not Met
Breath Volume	0.769	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 2 0.127	23:28	
Breath Volume	2.839	23:28	
Air Blank	0.000	23:29	
Air Blank	0.000	23:31	
Subject Sample	# 3 0.126	23:32	
Breath Volume	3.226	23:32	
Air Blank	0.000	23:33	
Control Test	0.077	23:33	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 05/18/2022 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON, KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: UNGER

DATE/TIME OF TEST: 05/04/2022 19:31:30

Diagnostics Check	OK	19:17	
Air Blank	0.000	19:17	
Control Test	0.080	19:18	
Air Blank	0.000	19:18	
Subject Sample	# 1 0.205	19:21	
Breath Volume	1.125	19:21	
Air Blank	0.000	19:21	
Air Blank	0.000	19:23	
Subject Sample	# 2 0.199	19:26	Volume Not Met
Breath Volume	0.878	19:26	
Air Blank	0.000	19:27	
Air Blank	0.000	19:29	
Subject Sample	# 3 0.195	19:29	
Breath Volume	1.226	19:29	
Air Blank	0.000	19:30	
Control Test	0.077	19:30	
Air Blank	0.000	19:31	
Diagnostics Check	OK	19:31	

Data Download Date/Time: 05/18/2022 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,DALE,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: BEHRENS,BRENT
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/09/2022 12:59:24

Diagnostics Check	OK	12:52
Air Blank	0.000	12:53
Control Test	0.081	12:53
Air Blank	0.000	12:54
Subject Sample # 1	0.000	12:54
Breath Volume	1.343	12:54
Air Blank	0.000	12:55
Air Blank	0.000	12:57
Subject Sample # 2	0.000	12:58
Breath Volume	3.187	12:58
Air Blank	0.000	12:58
Control Test	0.081	12:58
Air Blank	0.000	12:59
Diagnostics Check	OK	12:59

Data Download Date/Time: 05/18/2022 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 05/11/2022 03:44:32

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.223	03:39
Breath Volume	1.667	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample # 2	0.225	03:42
Breath Volume	1.703	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 05/18/2022 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEPHENS, TONY, T

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: STEPHENS, TONY

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:16

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 05/15/2022 03:18:53

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.082	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.081	03:18	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 05/18/2022 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: RIGGALL,NICHOLAS

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 04/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/15/2022 03:58:30

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.082	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.215	03:52
Breath Volume	1.242	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.203	03:56
Breath Volume	1.281	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 05/19/2022 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFONTANT, SEDONA, S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: COLLINS, JOSHUA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 05/23/2022 04:09:34

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.132	04:04
Breath Volume	2.625	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.138	04:07
Breath Volume	2.964	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 06/27/2022 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STEPHENS, TONY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: COTTONGIM

DATE/TIME OF TEST: 05/27/2022 23:48:22

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.081	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.137	23:42
Breath Volume	1.308	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.134	23:46
Breath Volume	2.082	23:46
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 06/27/2022 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 05/28/2022 00:23:44

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.081	00:13	
Air Blank	0.000	00:13	
Subject Sample	# 1	00:16	No Sample Provided
Breath Volume	0.000	00:16	
Air Blank	0.000	00:17	
Air Blank	0.000	00:19	
Subject Sample	# 2	00:22	No Sample Provided
Breath Volume	0.000	00:22	
Air Blank	0.000	00:22	
Control Test	0.082	00:23	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 06/27/2022 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STEPHENS, TONY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 05/28/2022 02:51:58

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.081	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:51	
Control Test	0.082	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 06/27/2022 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG, S
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CRUZ-SANCHEZ

DATE/TIME OF TEST: 05/29/2022 01:32:03

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.081	01:28	
Air Blank	0.000	01:28	
Subject Sample	# 1 0.084	01:30	
Breath Volume	1.843	01:30	
Air Blank		01:30	RFI Detect
Air Blank	0.000	01:31	
Control Test	0.082	01:31	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 06/27/2022 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG, S
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CRUZ-SANCHEZ

DATE/TIME OF TEST: 05/29/2022 01:41:50

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.076	01:37
Breath Volume	1.402	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.082	01:40
Breath Volume	4.082	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 06/27/2022 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: BAYONNE

DATE/TIME OF TEST: 05/29/2022 20:24:46

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.081	20:19
Air Blank	0.000	20:19
Subject Sample # 1	0.168	20:20
Breath Volume	1.777	20:20
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.167	20:23
Breath Volume	1.570	20:23
Air Blank	0.000	20:23
Control Test	0.082	20:24
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 06/27/2022 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STEPHENS, TONY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/30/2022 03:02:09

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.081	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 06/27/2022 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROOKS,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GARCIA-MARTIN

DATE/TIME OF TEST: 05/02/2022 03:53:28

Diagnostics Check	OK	03:52	
Air Blank		03:53	RFI Detect
Air Blank	0.000	03:53	

Data Download Date/Time: 05/25/2022 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROOKS,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GARCIA-MARTIN

DATE/TIME OF TEST: 05/02/2022 04:03:31

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.202	03:59
Breath Volume	3.000	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.214	04:02
Breath Volume	2.445	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 05/25/2022 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1192013
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 08:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/07/2022 09:31:33

Diagnostics Check	OK	09:24
Air Blank	0.000	09:25
Control Test	0.081	09:25
Air Blank	0.000	09:26
Subject Sample # 1	0.116	09:27
Breath Volume	2.261	09:27
Air Blank	0.000	09:27
Air Blank	0.000	09:29
Subject Sample # 2	0.114	09:30
Breath Volume	2.554	09:30
Air Blank	0.000	09:30
Control Test	0.081	09:30
Air Blank	0.000	09:31
Diagnostics Check	OK	09:31

Data Download Date/Time: 05/25/2022 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: COLEY

DATE/TIME OF TEST: 05/07/2022 04:04:25

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.078	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.175	03:59
Breath Volume	2.378	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.177	04:02
Breath Volume	2.621	04:02
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 06/04/2022 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HALPIN

DATE/TIME OF TEST: 05/26/2022 23:27:56

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.079	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.217	23:23
Breath Volume	1.753	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.212	23:26
Breath Volume	1.246	23:26
Air Blank	0.000	23:27
Control Test	0.075	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 06/04/2022 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 05/02/2022 00:34:21

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.079	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.187	00:27
Breath Volume	1.468	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.204	00:32
Breath Volume	2.410	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 05/17/2022 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: FARNEY

DATE/TIME OF TEST: 05/07/2022 02:58:37

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.036	02:54
Breath Volume	2.121	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.035	02:57
Breath Volume	2.507	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 05/17/2022 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: FEASER

DATE/TIME OF TEST: 05/15/2022 00:21:03

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.183	00:15
Breath Volume	2.585	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.195	00:19
Breath Volume	2.019	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:21

Data Download Date/Time: 05/17/2022 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 05/08/2022 20:08:01

Diagnostics Check	OK	20:05	
Air Blank	0.000	20:05	
Control Test	0.079	20:05	
Air Blank	0.000	20:06	
Subject Sample	# 1	20:06	Subject Test Refused
Breath Volume	0.000	20:06	
Air Blank	0.000	20:07	
Control Test	0.079	20:07	
Air Blank	0.000	20:07	
Diagnostics Check	OK	20:08	

Data Download Date/Time: 05/21/2022 07:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 05/13/2022 00:56:15

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.261	00:51
Breath Volume	1.515	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.264	00:54
Breath Volume	1.566	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 05/21/2022 07:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ARRIZA,D
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 05/15/2022 02:24:10

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.080	02:22	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.079	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 05/21/2022 07:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: PELLETIER

DATE/TIME OF TEST: 05/16/2022 17:52:41

Diagnostics Check	OK	17:49	
Air Blank	0.000	17:49	
Control Test	0.080	17:50	
Air Blank	0.000	17:50	
Subject Sample	# 1	17:51	Subject Test Refused
Breath Volume	0.000	17:51	
Air Blank	0.000	17:51	
Control Test	0.079	17:52	
Air Blank	0.000	17:52	
Diagnostics Check	OK	17:52	

Data Download Date/Time: 05/21/2022 07:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COURINGTON,R
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: ADDISON

DATE/TIME OF TEST: 05/18/2022 23:56:10

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.000	23:51
Breath Volume	1.691	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.000	23:54
Breath Volume	1.136	23:54
Air Blank	0.000	23:55
Control Test	0.078	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 05/21/2022 07:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 05/20/2022 20:58:38

Diagnostics Check	OK	20:50
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.167	20:53
Breath Volume	1.164	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:55
Subject Sample # 2	0.170	20:57
Breath Volume	1.101	20:57
Air Blank	0.000	20:57
Control Test	0.079	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 05/21/2022 07:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: KNEISS

DATE/TIME OF TEST: 05/28/2022 22:23:59

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.093	22:19
Breath Volume	2.371	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:21
Subject Sample # 2	0.097	22:22
Breath Volume	1.617	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 06/19/2022 05:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: DESHLER

DATE/TIME OF TEST: 05/29/2022 23:14:04

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.080	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Control Test	0.079	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 06/19/2022 05:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: MARSH

DATE/TIME OF TEST: 05/31/2022 17:33:43

Diagnostics Check	OK	17:26
Air Blank	0.000	17:27
Control Test	0.079	17:27
Air Blank	0.000	17:28
Subject Sample # 1	0.312	17:28
Breath Volume	1.542	17:28
Air Blank	0.000	17:29
Air Blank	0.000	17:31
Subject Sample # 2	0.301	17:32
Breath Volume	1.753	17:32
Air Blank	0.000	17:32
Control Test	0.078	17:33
Air Blank	0.000	17:33
Diagnostics Check	OK	17:33

Data Download Date/Time: 06/19/2022 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: MONCALIERI

DATE/TIME OF TEST: 05/31/2022 18:10:55

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:06	
Control Test	0.078	18:07	
Air Blank	0.000	18:07	
Subject Sample	# 1	18:09	Subject Test Refused
Breath Volume	0.000	18:09	
Air Blank	0.000	18:10	
Control Test	0.079	18:10	
Air Blank	0.000	18:10	
Diagnostics Check	OK	18:10	

Data Download Date/Time: 06/19/2022 05:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE, CHARLES,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CASTILLO REYES

DATE/TIME OF TEST: 05/01/2022 03:32:52

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.216	03:27
Breath Volume	2.281	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.216	03:31
Breath Volume	2.710	03:31
Air Blank	0.000	03:31
Control Test	0.077	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 05/24/2022 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CEPHUS,JAMIN,I
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: CEPHUS,JAMIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 04/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:16

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 05/09/2022 11:45:27

Diagnostics Check	OK	11:38
Air Blank	0.000	11:39
Control Test	0.080	11:39
Air Blank	0.000	11:40
Subject Sample # 1	0.053	11:40
Breath Volume	2.039	11:40
Air Blank	0.000	11:41
Air Blank	0.000	11:43
Subject Sample # 2	0.051	11:43
Breath Volume	2.570	11:43
Air Blank	0.000	11:44
Control Test	0.080	11:44
Air Blank	0.000	11:45
Diagnostics Check	OK	11:45

Data Download Date/Time: 05/24/2022 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 05/01/2022 04:09:52

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.081	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	04:02	No Sample Provided
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 2	0.239	04:05
Breath Volume	2.542	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 3	0.225	04:08
Breath Volume	2.363	04:08	
Air Blank	0.000	04:08	
Control Test	0.079	04:09	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ORTEGA,R
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: KHOURY

DATE/TIME OF TEST: 05/02/2022 02:38:40

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.081	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:37	
Control Test	0.081	02:38	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTENSEN,S,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: O'CONNELL,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/02/2022 03:20:00

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:13	No Sample Provided
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample	# 2	03:18	No Sample Provided
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Control Test	0.081	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:04

SUBJECT'S LAST NAME: CASKIE

DATE/TIME OF TEST: 05/03/2022 08:37:54

Diagnostics Check	OK	08:24	
Air Blank	0.000	08:24	
Control Test	0.079	08:25	
Air Blank	0.000	08:25	
Subject Sample	# 1	08:28	No Sample Provided
Breath Volume	0.000	08:28	
Air Blank	0.000	08:29	
Air Blank	0.000	08:31	
Subject Sample	# 2	0.180	08:32
Breath Volume	1.589	08:32	
Air Blank	0.000	08:33	
Air Blank	0.000	08:35	
Subject Sample	# 3	0.171	08:36
Breath Volume	1.886	08:36	
Air Blank	0.000	08:36	
Control Test	0.076	08:37	
Air Blank	0.000	08:37	
Diagnostics Check	OK	08:37	

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,CARLOS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 05/05/2022 00:34:02

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.187	00:28
Breath Volume	2.136	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.186	00:32
Breath Volume	1.250	00:32
Air Blank	0.000	00:33
Control Test	0.077	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:34

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: COGHILL

DATE/TIME OF TEST: 05/05/2022 01:53:21

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.228	01:48
Breath Volume	1.632	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.228	01:51
Breath Volume	1.652	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/05/2022 02:12:04

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.080	02:07
Breath Volume	1.160	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.072	02:10
Breath Volume	1.207	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:12

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: ZEIGEL

DATE/TIME OF TEST: 05/06/2022 00:02:02

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 2	23:55	No Sample Provided
Breath Volume	0.000	23:55	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample	# 1	00:00	No Sample Provided
Breath Volume	0.000	00:00	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/06/2022 02:40:50

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.179	02:35
Breath Volume	1.222	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.170	02:39
Breath Volume	1.167	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:16

SUBJECT'S LAST NAME: BURNES

DATE/TIME OF TEST: 05/06/2022 12:43:00

Diagnostics Check	OK	12:36
Air Blank	0.000	12:37
Control Test	0.081	12:37
Air Blank	0.000	12:38
Subject Sample # 1	0.000	12:38
Breath Volume	3.437	12:38
Air Blank	0.000	12:39
Air Blank	0.000	12:41
Subject Sample # 2	0.000	12:41
Breath Volume	2.914	12:41
Air Blank	0.000	12:42
Control Test	0.079	12:42
Air Blank	0.000	12:42
Diagnostics Check	OK	12:43

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: FEEHAN

DATE/TIME OF TEST: 05/06/2022 20:50:06

Diagnostics Check	OK	20:43
Air Blank	0.000	20:44
Control Test	0.081	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.000	20:45
Breath Volume	2.019	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.000	20:48
Breath Volume	1.910	20:48
Air Blank	0.000	20:49
Control Test	0.080	20:49
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 05/07/2022 00:30:48

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.081	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.205	00:25
Breath Volume	1.574	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.192	00:29
Breath Volume	1.410	00:29
Air Blank	0.000	00:29
Control Test	0.080	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 05/07/2022 04:16:32

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.081	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.196	04:10
Breath Volume	1.121	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.193	04:14
Breath Volume	1.160	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: AYALA ZAMORA

DATE/TIME OF TEST: 05/07/2022 06:48:21

Diagnostics Check	OK	06:42
Air Blank	0.000	06:42
Control Test	0.081	06:42
Air Blank	0.000	06:43
Subject Sample # 1	0.164	06:43
Breath Volume	1.300	06:43
Air Blank	0.000	06:44
Air Blank	0.000	06:46
Subject Sample # 2	0.163	06:46
Breath Volume	1.714	06:46
Air Blank	0.000	06:47
Control Test	0.079	06:47
Air Blank	0.000	06:48
Diagnostics Check	OK	06:48

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: W300541813080

DATE/TIME OF TEST: 05/07/2022 07:56:31

Diagnostics Check	OK	07:39	
Air Blank	0.000	07:40	
Control Test	0.080	07:40	
Air Blank	0.000	07:40	
Subject Sample	# 1 0.162	07:43	
Breath Volume	1.195	07:43	
Air Blank	0.000	07:44	
Air Blank	0.000	07:46	
Subject Sample	# 2 0.161	07:49	Volume Not Met
Breath Volume	0.996	07:49	
Air Blank	0.000	07:50	
Air Blank	0.000	07:51	
Subject Sample	# 3 0.114	07:54	Volume Not Met
Breath Volume	1.066	07:54	
Air Blank	0.000	07:55	
Control Test	0.061	07:55	Control Outside Tolerance
Air Blank	0.000	07:56	

Data Download Date/Time: 05/31/2022 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/07/2022 11:46:57

Diagnostics Check	OK	11:40
Air Blank	0.000	11:41
Control Test	0.080	11:41
Air Blank	0.000	11:41
Subject Sample # 1	0.000	11:42
Breath Volume	2.050	11:42
Air Blank	0.000	11:43
Air Blank	0.000	11:45
Subject Sample # 2	0.000	11:45
Breath Volume	1.492	11:45
Air Blank	0.000	11:46
Control Test	0.079	11:46
Air Blank	0.000	11:46
Diagnostics Check	OK	11:46

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 05/07/2022 23:38:11

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.081	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.171	23:33
Breath Volume	2.230	23:33
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.169	23:36
Breath Volume	1.984	23:36
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: PEREZ MANTILLA

DATE/TIME OF TEST: 05/08/2022 00:45:50

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.081	00:35	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:38	No Sample Provided
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 2	00:44	No Sample Provided
Breath Volume	0.000	00:44	
Air Blank	0.000	00:44	
Control Test	0.081	00:45	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 05/08/2022 03:56:01

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.150	03:51
Breath Volume	1.164	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.150	03:54
Breath Volume	1.792	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:56

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: MCKEEVER

DATE/TIME OF TEST: 05/09/2022 17:39:31

Diagnostics Check	OK	17:33
Air Blank	0.000	17:33
Control Test	0.080	17:33
Air Blank	0.000	17:34
Subject Sample # 1	0.390	17:35
Breath Volume	2.242	17:35
Air Blank	0.000	17:35
Air Blank	0.000	17:37
Subject Sample # 2	0.392	17:37
Breath Volume	2.335	17:37
Air Blank	0.000	17:38
Control Test	0.079	17:38
Air Blank	0.000	17:39
Diagnostics Check	OK	17:39

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLINE,NICOLE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: OCONNELL,BRIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: HARGROVE

DATE/TIME OF TEST: 05/11/2022 02:56:35

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.225	02:51
Breath Volume	3.183	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.224	02:54
Breath Volume	1.363	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/12/2022 01:53:45

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.135	01:49
Breath Volume	1.226	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.128	01:52
Breath Volume	1.402	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 05/12/2022 02:32:56

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:24	
Control Test	0.081	02:24	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:28	No Sample Provided
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample	# 2	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:32	
Control Test	0.081	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 05/12/2022 02:57:50

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.289	02:51
Breath Volume	1.429	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 2	0.283	02:56
Breath Volume	1.910	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 05/12/2022 04:28:14

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:24	
Control Test	0.081	04:24	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:27	
Control Test	0.081	04:27	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCOULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/12/2022 14:08:36

Diagnostics Check	OK	14:00
Air Blank	0.000	14:00
Control Test	0.080	14:01
Air Blank	0.000	14:01
Subject Sample # 1	0.141	14:04
Breath Volume	1.222	14:04
Air Blank	0.000	14:04
Air Blank	0.000	14:06
Subject Sample # 2	0.137	14:07
Breath Volume	1.671	14:07
Air Blank	0.000	14:07
Control Test	0.078	14:08
Air Blank	0.000	14:08
Diagnostics Check	OK	14:08

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: DAVIS JR

DATE/TIME OF TEST: 05/13/2022 01:30:31

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.082	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.166	01:26
Breath Volume	2.035	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.164	01:28
Breath Volume	1.296	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 05/13/2022 02:34:29

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.081	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.173	02:28
Breath Volume	1.628	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.169	02:32
Breath Volume	1.566	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: OLIVIERI

DATE/TIME OF TEST: 05/13/2022 03:26:11

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.140	03:21
Breath Volume	1.507	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.139	03:24
Breath Volume	1.445	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: HACKMAN

DATE/TIME OF TEST: 05/13/2022 23:45:46

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.081	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.207	23:41
Breath Volume	1.218	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.206	23:44
Breath Volume	1.253	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: PLANK

DATE/TIME OF TEST: 05/14/2022 00:40:14

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.080	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Control Test	0.081	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 05/14/2022 03:33:50

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.075	03:29
Breath Volume	2.187	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.070	03:32
Breath Volume	1.734	03:32
Air Blank	0.000	03:32
Control Test	0.081	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 05/14/2022 04:05:22

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.081	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:58	No Sample Provided
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Air Blank	0.000	04:00	
Subject Sample	# 2	04:03	No Sample Provided
Breath Volume	0.000	04:03	
Air Blank	0.000	04:04	
Control Test	0.081	04:04	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 05/31/2022 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MULVANEY

DATE/TIME OF TEST: 05/15/2022 01:51:30

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.245	01:46
Breath Volume	1.585	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.233	01:49
Breath Volume	1.441	01:49
Air Blank	0.000	01:50
Control Test	0.082	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CONIGLIO

DATE/TIME OF TEST: 05/17/2022 03:07:33

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.206	03:02
Breath Volume	1.492	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.219	03:05
Breath Volume	1.101	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 05/17/2022 20:53:20

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.081	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.215	20:48
Breath Volume	1.539	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.215	20:51
Breath Volume	1.511	20:51
Air Blank	0.000	20:52
Control Test	0.080	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/17/2022 22:56:44

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:48	
Control Test	0.080	22:48	
Air Blank	0.000	22:49	
Subject Sample	# 1	22:52	No Sample Provided
Breath Volume	0.000	22:52	
Air Blank	0.000	22:52	
Air Blank	0.000	22:54	
Subject Sample	# 2	22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:55	
Control Test	0.081	22:56	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: SEARCH

DATE/TIME OF TEST: 05/18/2022 00:11:41

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.214	00:07
Breath Volume	2.359	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.204	00:10
Breath Volume	2.031	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MARION

DATE/TIME OF TEST: 05/18/2022 01:11:16

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.255	01:06
Breath Volume	2.308	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.257	01:09
Breath Volume	2.679	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 05/18/2022 02:44:28

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.188	02:39
Breath Volume	1.238	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.180	02:42
Breath Volume	1.589	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: WESCHE

DATE/TIME OF TEST: 05/18/2022 22:52:36

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.081	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.147	22:48
Breath Volume	1.777	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.147	22:51
Breath Volume	1.464	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: WEIS

DATE/TIME OF TEST: 05/19/2022 01:16:26

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.080	01:05	
Air Blank	0.000	01:05	
Subject Sample	# 1 0.243	01:08	Volume Not Met
Breath Volume	0.613	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample	# 2 0.279	01:11	
Breath Volume	1.816	01:11	
Air Blank	0.000	01:12	
Air Blank	0.000	01:14	
Subject Sample	# 3 0.278	01:14	
Breath Volume	1.839	01:14	
Air Blank	0.000	01:15	
Control Test	0.079	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: O'CONNELL,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 05/19/2022 03:47:34

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.164	03:42
Breath Volume	1.382	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.146	03:45
Breath Volume	1.421	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: FILSON

DATE/TIME OF TEST: 05/20/2022 01:26:07

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.081	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.194	01:21
Breath Volume	2.492	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.193	01:24
Breath Volume	1.750	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: BONACOR

DATE/TIME OF TEST: 05/20/2022 01:46:19

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.081	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:42	No Sample Provided
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample	# 2	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Control Test	0.081	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WILLIAMS JR

DATE/TIME OF TEST: 05/21/2022 03:28:18

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.154	03:23
Breath Volume	1.234	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.155	03:26
Breath Volume	1.175	03:26
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 05/21/2022 06:37:32

Diagnostics Check	OK	06:22	
Air Blank	0.000	06:23	
Control Test	0.081	06:23	
Air Blank	0.000	06:24	
Subject Sample	# 1 0.136	06:27	Volume Not Met
Breath Volume	0.804	06:27	
Air Blank	0.000	06:27	
Air Blank	0.000	06:29	
Subject Sample	# 2 0.143	06:32	
Breath Volume	2.574	06:32	
Air Blank	0.000	06:33	
Air Blank	0.000	06:34	
Subject Sample	# 3 0.148	06:35	
Breath Volume	2.414	06:35	
Air Blank	0.000	06:36	
Control Test	0.079	06:36	
Air Blank	0.000	06:37	
Diagnostics Check	OK	06:37	

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKASAL, TYVON, T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: JACKASAL, TYVON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: SUKIS

DATE/TIME OF TEST: 05/24/2022 20:21:31

Diagnostics Check	OK	20:14
Air Blank	0.000	20:14
Control Test	0.081	20:15
Air Blank	0.000	20:15
Subject Sample # 1	0.250	20:16
Breath Volume	1.621	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.252	20:19
Breath Volume	1.753	20:19
Air Blank	0.000	20:20
Control Test	0.080	20:20
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 05/26/2022 00:01:06

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.081	23:50	
Air Blank	0.000	23:50	
Subject Sample	# 1	23:54	No Sample Provided
Breath Volume	0.000	23:54	
Air Blank	0.000	23:54	
Air Blank	0.000	23:56	
Subject Sample	# 2	23:59	No Sample Provided
Breath Volume	0.000	23:59	
Air Blank	0.000	00:00	
Control Test	0.081	00:00	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 05/26/2022 01:11:58

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.081	01:03	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:07	No Sample Provided
Breath Volume	0.000	01:07	
Air Blank	0.000	01:08	
Air Blank	0.000	01:10	
Subject Sample	# 2	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:11	
Control Test	0.082	01:11	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: BAYNES

DATE/TIME OF TEST: 05/26/2022 03:02:05

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.081	02:53	
Air Blank	0.000	02:53	
Subject Sample	# 1	02:56	No Sample Provided
Breath Volume	0.000	02:56	
Air Blank	0.000	02:57	
Air Blank	0.000	02:59	
Subject Sample	# 2	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.081	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 05/31/2022 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHA,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: COZZE

DATE/TIME OF TEST: 05/26/2022 11:21:31

Diagnostics Check	OK	11:15
Air Blank	0.000	11:15
Control Test	0.081	11:15
Air Blank	0.000	11:16
Subject Sample # 1	0.000	11:17
Breath Volume	2.500	11:17
Air Blank	0.000	11:17
Air Blank	0.000	11:19
Subject Sample # 2	0.000	11:20
Breath Volume	2.585	11:20
Air Blank	0.000	11:20
Control Test	0.079	11:20
Air Blank	0.000	11:21
Diagnostics Check	OK	11:21

Data Download Date/Time: 05/31/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: SCINLARI

DATE/TIME OF TEST: 05/26/2022 22:19:32

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.128	22:15
Breath Volume	1.332	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.129	22:17
Breath Volume	1.937	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 05/31/2022 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: LAIL

DATE/TIME OF TEST: 05/26/2022 23:05:01

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.155	22:59
Breath Volume	1.292	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.155	23:03
Breath Volume	1.812	23:03
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:05

Data Download Date/Time: 05/31/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/27/2022 23:48:11

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.148	23:39
Breath Volume	1.472	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.171	23:42
Breath Volume	1.238	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 3	0.161	23:46
Breath Volume	1.507	23:46
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 05/31/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: CURELLA

DATE/TIME OF TEST: 05/28/2022 00:58:22

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.081	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.117	00:53
Breath Volume	1.652	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.116	00:56
Breath Volume	1.355	00:56
Air Blank	0.000	00:57
Control Test	0.078	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 05/31/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: PETTIGREW

DATE/TIME OF TEST: 05/28/2022 03:49:12

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.129	03:44
Breath Volume	2.781	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.125	03:47
Breath Volume	1.804	03:47
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 05/31/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,CLAYTON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PACE,J
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: PEEPLES

DATE/TIME OF TEST: 05/29/2022 21:40:33

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.080	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.247	21:35
Breath Volume	1.175	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.240	21:38
Breath Volume	1.375	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 05/31/2022 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,CLAYTON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,CLAYTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: CHICA

DATE/TIME OF TEST: 05/29/2022 22:18:39

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.223	22:13
Breath Volume	2.468	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.224	22:17
Breath Volume	2.746	22:17
Air Blank	0.000	22:17
Control Test	0.079	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 05/31/2022 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: O'CONNELL,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 05/30/2022 03:25:41

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:23	
Control Test	0.080	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 05/31/2022 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: TORVIK

DATE/TIME OF TEST: 05/31/2022 00:52:31

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.196	00:47
Breath Volume	2.542	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.198	00:50
Breath Volume	2.472	00:50
Air Blank	0.000	00:51
Control Test	0.078	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 05/31/2022 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: RIBEIRO SILVA

DATE/TIME OF TEST: 05/02/2022 04:56:21

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:53	
Control Test	0.081	04:54	
Air Blank	0.000	04:54	
Subject Sample	# 1	04:54	Subject Test Refused
Breath Volume	0.000	04:54	
Air Blank	0.000	04:55	
Control Test	0.081	04:55	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,ROHAN,J
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: REMY,LUCKENSON
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: INSLEY

DATE/TIME OF TEST: 05/02/2022 05:03:38

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.082	05:00	
Air Blank	0.000	05:00	
Subject Sample	# 1 0.206	05:02	
Breath Volume	1.574	05:02	
Air Blank		05:02	Purge Fail
Air Blank		05:03	Purge Fail

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SIDOTI,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 05/02/2022 18:15:02

Diagnostics Check	OK	18:13	
Air Blank	0.000	18:14	
Control Test	0.074	18:14	Control Outside Tolerance
Air Blank	0.000	18:15	

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SIDOTI,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 05/02/2022 18:21:18

Diagnostics Check	OK	18:16	
Air Blank	0.000	18:16	
Control Test	0.075	18:17	
Air Blank	0.000	18:17	
Subject Sample	# 1	18:19	Subject Test Refused
Breath Volume	0.191	18:19	
Air Blank	0.000	18:20	
Control Test	0.074	18:20	Control Outside Tolerance
Air Blank	0.000	18:21	

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: TECK OSAETA

DATE/TIME OF TEST: 05/04/2022 22:42:55

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.081	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.129	22:38
Breath Volume	1.468	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.128	22:41
Breath Volume	2.363	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CARSON, TISHA
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 05/08/2022 00:34:15

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.082	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.132	00:29
Breath Volume	2.007	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.132	00:32
Breath Volume	1.621	00:32
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: EDMOND,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: MEJIA MARTINEZ

DATE/TIME OF TEST: 05/08/2022 23:11:24

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.081	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.040	23:06
Breath Volume	2.648	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.039	23:09
Breath Volume	2.917	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: RITTER

DATE/TIME OF TEST: 05/18/2022 00:16:54

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.082	00:14	
Air Blank	0.000	00:14	
Subject Sample # 1	0.000	00:15	
Breath Volume	2.679	00:15	
Air Blank		00:15	Improper Sample
Air Blank	0.000	00:15	
Control Test	0.081	00:16	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: RITTER

DATE/TIME OF TEST: 05/18/2022 00:24:50

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.081	00:19	
Air Blank	0.000	00:20	
Subject Sample # 1	0.000	00:20	
Breath Volume	2.292	00:20	
Air Blank	0.000	00:21	
Air Blank	0.000	00:23	
Subject Sample # 2	0.000	00:23	
Breath Volume	2.597	00:23	
Air Blank	0.000	00:24	
		00:24	RFI Detect
Air Blank	0.000	00:24	

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,AB,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CHERNOFF 19039,R
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/19/2022 06:18:32

Diagnostics Check	OK	06:11
Air Blank	0.000	06:12
Control Test	0.081	06:12
Air Blank	0.000	06:13
Subject Sample # 1	0.220	06:13
Breath Volume	1.449	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:16
Subject Sample # 2	0.226	06:16
Breath Volume	2.285	06:16
Air Blank	0.000	06:17
Control Test	0.080	06:17
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: ARTEAGA

DATE/TIME OF TEST: 05/22/2022 22:06:33

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:02	
Control Test	0.082	22:02	
Air Blank	0.000	22:03	
Subject Sample # 1	0.198	22:05	
Breath Volume	3.875	22:05	
Air Blank		22:05	Purge Fail
Air Blank	0.000	22:06	

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: ARTEAGA

DATE/TIME OF TEST: 05/22/2022 22:17:10

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.078	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.185	22:12
Breath Volume	1.980	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:14
Subject Sample # 2	0.196	22:15
Breath Volume	2.234	22:15
Air Blank	0.000	22:16
Control Test	0.078	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFER,BROOKS,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WEISENBURGH,LOGAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 05/25/2022 05:17:54

Diagnostics Check	OK	05:17	
Air Blank		05:17	RFI Detect
Air Blank		05:17	RFI Detect

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFER,BROOKS,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WEISENBURGH,LOGAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 05/25/2022 05:26:49

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.081	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.000	05:22
Breath Volume	2.117	05:22
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.000	05:25
Breath Volume	2.699	05:25
Air Blank	0.000	05:25
Control Test	0.080	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/28/2022 03:53:05

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.083	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	RFI Detect
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/28/2022 04:02:39

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.082	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.182	03:57
Breath Volume	1.484	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.174	04:00
Breath Volume	1.230	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 05/31/2022 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: AMATRUDA,CHRIS,G

BREATH TEST OPERATOR AGENCY: Dade City PD

ARRESTING OFFICER: WATSON,WILLIAM

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 04/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 25721080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: TEJADA

DATE/TIME OF TEST: 05/01/2022 02:22:30

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.040	02:18
Breath Volume	1.859	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.043	02:20
Breath Volume	1.367	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 05/17/2022 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: AMATRUDA,CHRIS,G

BREATH TEST OPERATOR AGENCY: Dade City PD

ARRESTING OFFICER: WATSON,WILLIAM

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 04/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 25721080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:59

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/02/2022 00:39:43

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.207	00:34
Breath Volume	1.750	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.200	00:38
Breath Volume	2.742	00:38
Air Blank	0.000	00:38
Control Test	0.078	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 05/17/2022 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: CLIFTON, WYATT
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: RANDOLPH

DATE/TIME OF TEST: 05/29/2022 00:52:59

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.090	00:47
Breath Volume	2.734	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.092	00:51
Breath Volume	2.582	00:51
Air Blank	0.000	00:52
Control Test	0.079	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 06/08/2022 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/29/2022 01:14:50

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.078	01:12	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:13	
Control Test	0.078	01:14	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 06/03/2022 08:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: PUPO TERUEL

DATE/TIME OF TEST: 05/29/2022 23:37:51

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.077	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1 0.130	23:34	
Breath Volume	3.035	23:34	
Air Blank	0.000	23:35	
Air Blank		23:37	RFI Detect
Air Blank		23:37	RFI Detect
Air Blank	0.000	23:37	

Data Download Date/Time: 06/03/2022 08:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: PUPO TERUEL

DATE/TIME OF TEST: 05/29/2022 23:44:19

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:40	
Control Test	0.078	23:40	
Air Blank	0.000	23:40	
Subject Sample	# 1 0.126	23:41	
Breath Volume	3.660	23:41	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample	# 2 0.131	23:44	
Breath Volume	3.699	23:44	
Air Blank		23:44	RFI Detect
Air Blank		23:44	RFI Detect

Data Download Date/Time: 06/03/2022 08:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: PUPO TERUEL

DATE/TIME OF TEST: 05/30/2022 00:24:00

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.077	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.115	00:19
Breath Volume	3.531	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.118	00:22
Breath Volume	3.714	00:22
Air Blank	0.000	00:23
Control Test	0.076	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:24

Data Download Date/Time: 06/03/2022 08:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES-RIVERA, DAVID, E
BREATH TEST OPERATOR AGENCY: FDLE
ARRESTING OFFICER: TEST, TEST
ARRESTING OFFICER AGENCY: FDLE
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: ADMIN
ARREST TIME: 10:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:34

SUBJECT'S LAST NAME: SAMPLE

DATE/TIME OF TEST: 05/17/2022 12:00:35

Diagnostics Check	OK	11:54	
Air Blank	0.000	11:55	
Control Test	0.082	11:55	
Air Blank	0.000	11:55	
Subject Sample # 1	0.000	11:56	
Breath Volume	1.632	11:56	
Air Blank	0.000	11:56	
Air Blank	0.000	11:58	
Subject Sample # 2		11:59	Subject Test Refused
Breath Volume	0.000	11:59	
Air Blank	0.000	11:59	
Control Test	0.081	11:59	
Air Blank	0.000	12:00	
Diagnostics Check	OK	12:00	

Data Download Date/Time: 05/17/2022 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES-RIVERA,DAVID,E
BREATH TEST OPERATOR AGENCY: FDLE
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: FDLE
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: ADMIN
ARREST TIME: 14:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 05/17/2022 14:39:03

Diagnostics Check	OK	14:36	
Air Blank	0.000	14:36	
Control Test	0.082	14:36	
Air Blank	0.000	14:37	
Subject Sample	# 1	14:37	Subject Test Refused
Breath Volume	0.000	14:37	
Air Blank	0.000	14:38	
Control Test	0.082	14:38	
Air Blank	0.000	14:38	
Diagnostics Check	OK	14:39	

Data Download Date/Time: 05/17/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: GANDICA-ZAMBRANO

DATE/TIME OF TEST: 05/13/2022 03:27:53

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.078	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.150	03:23
Breath Volume	1.675	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.151	03:26
Breath Volume	1.613	03:26
Air Blank	0.000	03:26
Control Test	0.078	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 06/09/2022 02:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PULLES,A
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: VIVEROS-AGUIRRE

DATE/TIME OF TEST: 05/14/2022 02:51:25

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.028	02:46
Breath Volume	3.683	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.026	02:49
Breath Volume	3.957	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 06/09/2022 02:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESPIAN,MIGUEL,M
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: DESPIAN,MIGUEL
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: JEAN PHILIPPE

DATE/TIME OF TEST: 05/30/2022 20:47:13

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.080	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.144	20:42
Breath Volume	2.476	20:42
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.144	20:45
Breath Volume	3.007	20:45
Air Blank	0.000	20:46
Control Test	0.079	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 06/15/2022 15:09