

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,ANDREW,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PENA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: CALMO

DATE/TIME OF TEST: 06/03/2022 17:20:48

Diagnostics Check	OK	17:19	
Air Blank		17:20	RFI Detect
Air Blank	0.000	17:20	

Data Download Date/Time: 07/05/2022 06:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,ANDREW,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PENA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: CALMO

DATE/TIME OF TEST: 06/03/2022 17:30:17

Diagnostics Check	OK	17:23
Air Blank	0.000	17:24
Control Test	0.079	17:24
Air Blank	0.000	17:25
Subject Sample # 1	0.245	17:25
Breath Volume	2.281	17:25
Air Blank	0.000	17:26
Air Blank	0.000	17:28
Subject Sample # 2	0.237	17:28
Breath Volume	2.210	17:28
Air Blank	0.000	17:29
Control Test	0.077	17:29
Air Blank	0.000	17:30
Diagnostics Check	OK	17:30

Data Download Date/Time: 07/05/2022 06:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: ROSSON,LUCAS
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/05/2022 00:34:48

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:29	
Control Test	0.079	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.703	00:33	
Air Blank	0.000	00:33	
Control Test	0.080	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 07/05/2022 06:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEETZ,JOHN,W
BREATH TEST OPERATOR AGENCY: Calhoun County SO
ARRESTING OFFICER: ADKINS,D
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 12:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:14

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 06/21/2022 12:38:54

Diagnostics Check	OK	12:34	
Air Blank	0.000	12:35	
Control Test	0.081	12:35	
Air Blank	0.000	12:35	
Subject Sample	# 1	12:37	Subject Test Refused
Breath Volume	0.000	12:37	
Air Blank	0.000	12:38	
Control Test	0.082	12:38	
Air Blank	0.000	12:38	
Diagnostics Check	OK	12:38	

Data Download Date/Time: 06/22/2022 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWKINS,JUNIOR,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: ABASCAL,JEFFREY
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 05/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: VARONA

DATE/TIME OF TEST: 06/15/2022 23:49:44

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:46	
Control Test	0.078	23:46	
Air Blank	0.000	23:46	
Subject Sample	# 1	23:48	Subject Test Refused
Breath Volume	0.000	23:48	
Air Blank	0.000	23:48	
Control Test	0.077	23:49	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time: 06/23/2022 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTAGLIA,PAUL,M
BREATH TEST OPERATOR AGENCY: North Bay Village PD
ARRESTING OFFICER: BATTAGLIA,PAUL
ARRESTING OFFICER AGENCY: North Bay Village PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 06/04/2022 07:01:44

Diagnostics Check	OK	06:58	
Air Blank	0.000	06:59	
Control Test	0.079	06:59	
Air Blank	0.000	07:00	
Subject Sample	# 1	07:00	Subject Test Refused
Breath Volume	0.000	07:00	
Air Blank	0.000	07:00	
Control Test	0.079	07:01	
Air Blank	0.000	07:01	
Diagnostics Check	OK	07:01	

Data Download Date/Time: 06/09/2022 02:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRYCE-JOHNSON, ANESTIA, A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GIBSON, TRANICE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 06/08/2022 10:05:27

Diagnostics Check	OK	10:02	
Air Blank	0.000	10:03	
Control Test	0.078	10:03	
Air Blank	0.000	10:03	
Subject Sample	# 1	10:04	Subject Test Refused
Breath Volume	0.000	10:04	
Air Blank	0.000	10:04	
Control Test	0.079	10:04	
Air Blank	0.000	10:05	
Diagnostics Check	OK	10:05	

Data Download Date/Time: 06/09/2022 02:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEJERA,MARTHA,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: TEJERA,MARTHA
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 06/29/2022 23:59:09

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.079	23:56	
Air Blank	0.000	23:56	
Subject Sample	# 1	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:58	
Control Test	0.078	23:58	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 07/12/2022 04:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIL,ALEJANDRO,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OCEGUERA,YASMANI
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: KISSANE

DATE/TIME OF TEST: 06/02/2022 22:11:16

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:58	
Control Test	0.080	21:58	
Air Blank	0.000	21:59	
Subject Sample	# 1	22:02	No Sample Provided
Breath Volume	0.000	22:02	
Air Blank	0.000	22:03	
Air Blank	0.000	22:05	
Subject Sample	# 2	0.218 22:05	
Breath Volume	1.324	22:05	
Air Blank	0.000	22:06	
Air Blank	0.000	22:08	
Subject Sample	# 3	0.222 22:09	
Breath Volume	1.144	22:09	
Air Blank	0.000	22:10	
Control Test	0.079	22:10	
Air Blank	0.000	22:11	
Diagnostics Check	OK	22:11	

Data Download Date/Time: 06/03/2022 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,ARISBEL,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SERRANO,LUIS
ARRESTING OFFICER AGENCY: Opa Locka PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 06/19/2022 11:31:09

Diagnostics Check	OK	11:24
Air Blank	0.000	11:25
Control Test	0.080	11:25
Air Blank	0.000	11:26
Subject Sample # 1	0.000	11:26
Breath Volume	2.105	11:26
Air Blank	0.000	11:27
Air Blank	0.000	11:29
Subject Sample # 2	0.000	11:29
Breath Volume	2.355	11:29
Air Blank	0.000	11:30
Control Test	0.079	11:30
Air Blank	0.000	11:31
Diagnostics Check	OK	11:31

Data Download Date/Time: 07/12/2022 04:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEWIS,TIMOTHY,B

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: LEWIS,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1482048

DRY GAS EXPIRY DATE: 04/21/2024

VIOLATION CODE: DUI

ARREST TIME: 09:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 06/02/2022 10:12:51

Diagnostics Check	OK	10:06
Air Blank	0.000	10:07
Control Test	0.080	10:07
Air Blank	0.000	10:08
Subject Sample # 1	0.000	10:08
Breath Volume	2.566	10:08
Air Blank	0.000	10:09
Air Blank	0.000	10:11
Subject Sample # 2	0.000	10:11
Breath Volume	1.378	10:11
Air Blank	0.000	10:11
Control Test	0.079	10:12
Air Blank	0.000	10:12
Diagnostics Check	OK	10:12

Data Download Date/Time: 06/03/2022 15:15

Subject Test Electronic Data

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEWIS,TIMOTHY,B

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: LEWIS,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/03/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1482048

DRY GAS EXPIRY DATE: 04/21/2024

VIOLATION CODE: DUI

ARREST TIME: 22:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 06/10/2022 22:59:48

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:54	
Control Test	0.077	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:58	Subject Test Refused
Breath Volume	0.000	22:58	
Air Blank	0.000	22:58	
Control Test	0.079	22:59	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 07/06/2022 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1482048
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B	DRY GAS EXPIRY DATE: 04/21/2024
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: LEWIS,TIMOTHY	ARREST TIME: 23:13
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Male
LAST AGENCY INSPECTION: 06/03/2022	OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: NUWER DATE/TIME OF TEST: 06/13/2022 23:42:31

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.270	23:38
Breath Volume	2.203	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.270	23:41
Breath Volume	2.289	23:41
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 07/06/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEWIS,TIMOTHY,B

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: LEWIS,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/03/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1482048

DRY GAS EXPIRY DATE: 04/21/2024

VIOLATION CODE: DUI

ARREST TIME: 19:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: BINGHAM

DATE/TIME OF TEST: 06/18/2022 20:20:08

Diagnostics Check	OK	20:16	
Air Blank	0.000	20:17	
Control Test	0.082	20:17	
Air Blank	0.000	20:18	
Subject Sample	# 1	20:18	Subject Test Refused
Breath Volume	0.000	20:18	
Air Blank	0.000	20:19	
Control Test	0.080	20:19	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 07/06/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEWIS,TIMOTHY,B

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: LEWIS,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/03/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1482048

DRY GAS EXPIRY DATE: 04/21/2024

VIOLATION CODE: DUI

ARREST TIME: 07:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: CRILLY

DATE/TIME OF TEST: 06/23/2022 08:30:07

Diagnostics Check	OK	08:28	
Air Blank	0.000	08:29	
Control Test	0.082	08:29	Interferent Detect
Air Blank	0.000	08:30	

Data Download Date/Time: 07/06/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEWIS,TIMOTHY,B

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: LEWIS,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/03/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1482048

DRY GAS EXPIRY DATE: 04/21/2024

VIOLATION CODE: DUI

ARREST TIME: 07:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: CRILLY

DATE/TIME OF TEST: 06/23/2022 08:36:27

Diagnostics Check	OK	08:33	
Air Blank	0.000	08:34	
Control Test	0.081	08:34	
Air Blank	0.000	08:34	
Subject Sample	# 1	08:35	Subject Test Refused
Breath Volume	0.000	08:35	
Air Blank	0.000	08:35	
Control Test	0.077	08:35	
Air Blank	0.000	08:36	
Diagnostics Check	OK	08:36	

Data Download Date/Time: 07/06/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEWIS,TIMOTHY,B

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: LEWIS,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/03/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1482048

DRY GAS EXPIRY DATE: 04/21/2024

VIOLATION CODE: DUI

ARREST TIME: 20:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 06/23/2022 20:30:50

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:28	
Control Test	0.079	20:28	
Air Blank	0.000	20:29	
Subject Sample	# 1	20:29	Subject Test Refused
Breath Volume	0.000	20:29	
Air Blank	0.000	20:29	
Control Test	0.075	20:30	
Air Blank	0.000	20:30	
Diagnostics Check	OK	20:30	

Data Download Date/Time: 07/06/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: LOSTROH

DATE/TIME OF TEST: 06/01/2022 04:42:46

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.174	04:38
Breath Volume	2.203	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.172	04:41
Breath Volume	2.082	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 06/29/2022 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEAVER,NICOLE
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 06/01/2022 23:49:25

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.080	23:48	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:48	Sequence Aborted
Breath Volume	0.000	23:48	
Air Blank	0.000	23:49	

Data Download Date/Time: 06/29/2022 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEAVER,NICOLE
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 06/02/2022 00:04:11

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:55
Air Blank	0.000	23:55
Subject Sample # 2	0.340	23:56
Breath Volume	2.804	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 3	0.314	23:58
Breath Volume	1.386	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.335	00:02
Breath Volume	2.496	00:02
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 06/29/2022 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/03/2022 02:29:02

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.081	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Control Test	0.081	02:28	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 06/29/2022 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: ROEDER

DATE/TIME OF TEST: 06/03/2022 23:05:00

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.108	22:59
Breath Volume	1.375	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.114	23:03
Breath Volume	1.296	23:03
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:05

Data Download Date/Time: 06/29/2022 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,KAYLA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: BOLEN

DATE/TIME OF TEST: 06/03/2022 23:36:49

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.195	23:32
Breath Volume	1.359	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.197	23:35
Breath Volume	1.285	23:35
Air Blank	0.000	23:35
Control Test	0.079	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 06/29/2022 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CASUSKY

DATE/TIME OF TEST: 06/04/2022 02:46:06

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.172	02:39
Breath Volume	1.660	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.174	02:44
Breath Volume	1.582	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 06/29/2022 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/04/2022 03:38:01

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.080	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Control Test	0.080	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 06/29/2022 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHRISTIE,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: WONG

DATE/TIME OF TEST: 06/04/2022 05:38:08

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.080	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.161	05:32
Breath Volume	2.128	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.147	05:36
Breath Volume	1.890	05:36
Air Blank	0.000	05:37
Control Test	0.079	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 06/29/2022 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: PEPIN

DATE/TIME OF TEST: 06/04/2022 23:54:27

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.159	23:50
Breath Volume	2.195	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.157	23:52
Breath Volume	1.890	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 06/29/2022 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/05/2022 02:10:41

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.080	02:00	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:04	No Sample Provided
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample	# 2	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Control Test	0.080	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 06/29/2022 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/05/2022 02:43:02

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.132	02:38
Breath Volume	2.152	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.133	02:41
Breath Volume	1.941	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:43

Data Download Date/Time: 06/29/2022 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: KOZLOWSKI

DATE/TIME OF TEST: 06/05/2022 20:12:59

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:02	
Control Test	0.081	20:02	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:06	No Sample Provided
Breath Volume	0.000	20:06	
Air Blank	0.000	20:06	
Air Blank	0.000	20:08	
Subject Sample	# 2	20:11	No Sample Provided
Breath Volume	0.000	20:11	
Air Blank	0.000	20:12	
Control Test	0.081	20:12	
Air Blank	0.000	20:12	
Diagnostics Check	OK	20:12	

Data Download Date/Time: 06/29/2022 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: CIANCIOLO

DATE/TIME OF TEST: 06/07/2022 20:19:28

Diagnostics Check	OK	20:13
Air Blank	0.000	20:14
Control Test	0.081	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.000	20:15
Breath Volume	2.171	20:15
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.000	20:18
Breath Volume	2.054	20:18
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 06/29/2022 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKEIRIK,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: MILJURE

DATE/TIME OF TEST: 06/08/2022 00:27:46

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:23	
Control Test	0.080	00:23	
Air Blank	0.000	00:24	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:26	
Control Test	0.081	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 06/29/2022 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROZZELL,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: STOUT

DATE/TIME OF TEST: 06/08/2022 03:09:08

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.184	03:04
Breath Volume	1.480	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.190	03:07
Breath Volume	1.906	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 06/29/2022 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: PIMENTEL

DATE/TIME OF TEST: 06/08/2022 04:16:57

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.080	04:14	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Control Test	0.080	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 06/29/2022 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MUCA

DATE/TIME OF TEST: 06/09/2022 02:19:48

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.074	02:15
Breath Volume	1.660	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.078	02:18
Breath Volume	1.714	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 06/29/2022 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHRISTIE,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: HALLAM

DATE/TIME OF TEST: 06/10/2022 01:19:28

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.255	01:14
Breath Volume	2.398	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.261	01:17
Breath Volume	2.585	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 06/29/2022 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 06/10/2022 01:34:59

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.081	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 06/29/2022 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MARTY

DATE/TIME OF TEST: 06/11/2022 00:04:40

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Subject Sample # 1	0.099	00:00
Breath Volume	1.535	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.096	00:03
Breath Volume	1.480	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 06/29/2022 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 06/11/2022 01:38:45

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.080	01:30	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:33	No Sample Provided
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample	# 2	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.080	01:38	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 06/29/2022 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIS,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BENFIELD

DATE/TIME OF TEST: 06/11/2022 02:25:01

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.120	02:20
Breath Volume	1.125	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.117	02:23
Breath Volume	1.781	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:25

Data Download Date/Time: 06/29/2022 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SHOULTA,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: PEEPLES

DATE/TIME OF TEST: 06/12/2022 01:36:47

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.081	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.125	01:32
Breath Volume	2.554	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.117	01:35
Breath Volume	1.546	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 06/29/2022 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: WEDGEWORTH

DATE/TIME OF TEST: 06/12/2022 23:46:58

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:44	
Control Test	0.080	23:44	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:46	
Control Test	0.081	23:46	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 06/29/2022 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WHEELER,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/13/2022 05:57:44

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.081	05:52
Air Blank	0.000	05:52
Subject Sample # 1	0.227	05:53
Breath Volume	1.843	05:53
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.226	05:56
Breath Volume	2.445	05:56
Air Blank	0.000	05:56
Control Test	0.080	05:57
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 06/29/2022 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,CHRISTOPHER,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIS,BRITT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: DIAZ RAMOS

DATE/TIME OF TEST: 06/13/2022 07:06:41

Diagnostics Check	OK	07:03	
Air Blank	0.000	07:04	
Control Test	0.081	07:04	
Air Blank	0.000	07:05	
Subject Sample	# 1	07:05	Subject Test Refused
Breath Volume	0.000	07:05	
Air Blank	0.000	07:05	
Control Test	0.081	07:06	
Air Blank	0.000	07:06	
Diagnostics Check	OK	07:06	

Data Download Date/Time: 06/29/2022 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Belleair PD
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: BEERS

DATE/TIME OF TEST: 06/14/2022 19:18:06

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:15	
Control Test	0.081	19:15	
Air Blank	0.000	19:16	
Subject Sample	# 1	19:16	Subject Test Refused
Breath Volume	0.000	19:16	
Air Blank	0.000	19:17	
Control Test	0.080	19:17	
Air Blank	0.000	19:18	
Diagnostics Check	OK	19:18	

Data Download Date/Time: 06/29/2022 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEWTON,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 06/15/2022 03:58:26

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.121	03:53
Breath Volume	2.527	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.118	03:56
Breath Volume	1.710	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 06/29/2022 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVANOVICLM,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IVANOVICLM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 06/15/2022 05:45:24

Diagnostics Check	OK	05:37
Air Blank	0.000	05:38
Control Test	0.081	05:38
Air Blank	0.000	05:39
Subject Sample # 1	0.098	05:40
Breath Volume	1.449	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.095	05:43
Breath Volume	1.125	05:43
Air Blank	0.000	05:44
Control Test	0.080	05:44
Air Blank	0.000	05:45
Diagnostics Check	OK	05:45

Data Download Date/Time: 06/29/2022 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 06/16/2022 03:32:47

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.081	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1 0.159	03:23	Volume Not Met
Breath Volume	0.167	03:23	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 2 0.169	03:28	
Breath Volume	2.542	03:28	
Air Blank	0.000	03:28	
Air Blank	0.000	03:30	
Subject Sample	# 3 0.176	03:31	
Breath Volume	1.800	03:31	
Air Blank	0.000	03:31	
Control Test	0.078	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 06/29/2022 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FERGUSON,BRADDON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 06/16/2022 03:46:53

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.079	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.215	03:42
Breath Volume	1.195	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.225	03:45
Breath Volume	2.812	03:45
Air Blank	0.000	03:45
Control Test	0.080	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 06/29/2022 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BENTLEY,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: HERRING

DATE/TIME OF TEST: 06/17/2022 18:59:02

Diagnostics Check	OK	18:55	
Air Blank	0.000	18:56	
Control Test	0.081	18:56	
Air Blank	0.000	18:56	
Subject Sample	# 1	18:57	Subject Test Refused
Breath Volume	0.000	18:57	
Air Blank	0.000	18:58	
Control Test	0.081	18:58	
Air Blank	0.000	18:58	
Diagnostics Check	OK	18:59	

Data Download Date/Time: 06/29/2022 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: FINAN

DATE/TIME OF TEST: 06/17/2022 20:17:10

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.080	20:10
Air Blank	0.000	20:10
Subject Sample # 1	0.128	20:12
Breath Volume	2.070	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.127	20:15
Breath Volume	1.488	20:15
Air Blank	0.000	20:16
Control Test	0.079	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 06/29/2022 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/18/2022 00:48:42

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.231	00:44
Breath Volume	2.214	00:44
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.225	00:47
Breath Volume	1.636	00:47
Air Blank	0.000	00:47
Control Test	0.080	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 06/29/2022 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FERRETTI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: WIECKHORST

DATE/TIME OF TEST: 06/18/2022 01:05:12

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:53	
Control Test	0.080	00:53	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:57	No Sample Provided
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Air Blank	0.000	00:59	
Subject Sample	# 2	01:00	
Breath Volume	1.632	01:00	
Air Blank	0.000	01:01	
Air Blank	0.000	01:03	
Subject Sample	# 3	01:03	
Breath Volume	1.378	01:03	
Air Blank	0.000	01:04	
Control Test	0.080	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 06/29/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RADDER,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HUDNALL

DATE/TIME OF TEST: 06/18/2022 01:38:40

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.080	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.080	01:38	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 06/29/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIGGS,J
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MANEVAL

DATE/TIME OF TEST: 06/18/2022 04:12:12

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.081	04:10	
Air Blank	0.000	04:10	
Subject Sample	# 1	04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:11	
Control Test	0.081	04:11	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 06/29/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEWTON,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: ARANY

DATE/TIME OF TEST: 06/18/2022 20:15:05

Diagnostics Check	OK	20:12	
Air Blank	0.000	20:12	
Control Test	0.081	20:13	
Air Blank	0.000	20:13	
Subject Sample	# 1	20:13	Subject Test Refused
Breath Volume	0.000	20:13	
Air Blank	0.000	20:14	
Control Test	0.080	20:14	
Air Blank	0.000	20:15	
Diagnostics Check	OK	20:15	

Data Download Date/Time: 06/29/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/18/2022 21:02:49

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.126	20:58
Breath Volume	1.445	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.128	21:01
Breath Volume	1.906	21:01
Air Blank	0.000	21:01
Control Test	0.079	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 06/29/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/19/2022 01:47:13

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.059	01:42
Breath Volume	1.671	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.065	01:45
Breath Volume	1.359	01:45
Air Blank	0.000	01:46
Control Test	0.081	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 06/29/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,JIMMY,W
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REVILLA,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: NUTTER

DATE/TIME OF TEST: 06/19/2022 16:51:35

Diagnostics Check	OK	16:45
Air Blank	0.000	16:45
Control Test	0.080	16:46
Air Blank	0.000	16:46
Subject Sample # 1	0.170	16:47
Breath Volume	3.957	16:47
Air Blank	0.000	16:47
Air Blank	0.000	16:49
Subject Sample # 2	0.172	16:50
Breath Volume	3.738	16:50
Air Blank	0.000	16:50
Control Test	0.079	16:50
Air Blank	0.000	16:51
Diagnostics Check	OK	16:51

Data Download Date/Time: 06/29/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 06/04/2022 22:27:09

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:23	
Control Test	0.081	22:24	
Air Blank	0.000	22:24	
Subject Sample	# 1	22:25	Subject Test Refused
Breath Volume	0.000	22:25	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 06/23/2022 04:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: ZUCCHETTI

DATE/TIME OF TEST: 06/30/2022 00:54:36

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.081	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:53	
Control Test	0.078	00:54	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

Data Download Date/Time: 07/15/2022 12:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LECOCQ,MATTHEW,E

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: LECOCQ,MATTHEW

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 16:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: ROE

DATE/TIME OF TEST: 06/03/2022 17:07:37

Diagnostics Check	OK	17:00
Air Blank	0.000	17:01
Control Test	0.081	17:01
Air Blank	0.000	17:01
Subject Sample # 1	0.174	17:02
Breath Volume	2.710	17:02
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample # 2	0.156	17:06
Breath Volume	2.640	17:06
Air Blank	0.000	17:06
Control Test	0.081	17:07
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

Data Download Date/Time: 06/22/2022 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1349789
BREATH TEST OPERATOR: PRITCHARD,THOMAS,D	DRY GAS EXPIRY DATE: 07/15/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: PRITCHARD,THOMAS	ARREST TIME: 23:02
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/16/2022	OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ANTUNES	DATE/TIME OF TEST: 06/05/2022 23:23:52
------------------------------	--

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:18	
Control Test	0.082	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:22	
Control Test	0.081	23:23	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 06/22/2022 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LECOCQ,MATTHEW,E

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: LECOCQ,MATTHEW

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 16:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: DANA

DATE/TIME OF TEST: 06/08/2022 16:54:57

Diagnostics Check	OK	16:48
Air Blank	0.000	16:49
Control Test	0.081	16:49
Air Blank	0.000	16:50
Subject Sample # 1	0.106	16:50
Breath Volume	3.363	16:50
Air Blank	0.000	16:51
Air Blank	0.000	16:52
Subject Sample # 2	0.101	16:53
Breath Volume	3.414	16:53
Air Blank	0.000	16:54
Control Test	0.078	16:54
Air Blank	0.000	16:54
Diagnostics Check	OK	16:54

Data Download Date/Time: 06/22/2022 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MALAIS,ANDREW,W

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: MALTAIS,ANDREW

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 16:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 06/10/2022 17:36:28

Diagnostics Check	OK	17:29	
Air Blank	0.000	17:30	
Control Test	0.082	17:30	
Air Blank	0.000	17:30	
Subject Sample	# 1 0.000	17:31	
Breath Volume	4.625	17:31	
Air Blank	0.000	17:32	
Air Blank	0.000	17:34	
Subject Sample	# 2 0.000	17:34	
Breath Volume	4.332	17:34	
Air Blank	0.000	17:35	
Control Test	0.080	17:35	
Air Blank	0.000	17:35	
Diagnostics Check	Fail	17:36	Analytical Stability Test
Air Blank	0.000	17:36	

Data Download Date/Time: 06/22/2022 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MALTAIS,ANDREW,W

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: POWELL,MICHAEL

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 14:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/11/2022 14:35:39

Diagnostics Check	OK	14:34	
Air Blank		14:35	RFI Detect
Air Blank	0.000	14:35	

Data Download Date/Time: 06/22/2022 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1349789
BREATH TEST OPERATOR: MALTAIS,ANDREW,W	DRY GAS EXPIRY DATE: 07/15/2023
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: POWELL,MICHAEL	ARREST TIME: 14:06
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 05/16/2022	OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: SMITH	DATE/TIME OF TEST: 06/11/2022	14:46:34
----------------------------	-------------------------------	----------

Diagnostics Check	OK	14:37
Air Blank	0.000	14:37
Control Test	0.081	14:38
Air Blank	0.000	14:38
Subject Sample # 1	0.183	14:39
Breath Volume	3.781	14:39
Air Blank	0.000	14:39
Air Blank	0.000	14:41
Subject Sample # 2	0.157	14:42
Breath Volume	2.308	14:42
Air Blank	0.000	14:42
Air Blank	0.000	14:44
Subject Sample # 3	0.171	14:44
Breath Volume	2.785	14:44
Air Blank	0.000	14:45
Control Test	0.079	14:45
Air Blank	0.000	14:46
Diagnostics Check	OK	14:46

Data Download Date/Time: 06/22/2022 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: Yes
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1349789
BREATH TEST OPERATOR: MURSCH,THEODORE,R	DRY GAS EXPIRY DATE: 07/15/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: HAHR,DAN	ARREST TIME: 20:23
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/16/2022	OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: RODRIGUEZ	DATE/TIME OF TEST: 06/11/2022 20:14:49
--------------------------------	--

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.081	20:09
Air Blank	0.000	20:09
Subject Sample # 1	0.221	20:10
Breath Volume	2.000	20:10
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.215	20:13
Breath Volume	1.714	20:13
Air Blank	0.000	20:13
Control Test	0.079	20:14
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 06/22/2022 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,DANIEL,S

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: WILLIAMS,DANIEL

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 16:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: GARY

DATE/TIME OF TEST: 06/14/2022 17:01:18

Diagnostics Check	OK	16:54
Air Blank	0.000	16:54
Control Test	0.080	16:55
Air Blank	0.000	16:55
Subject Sample # 1	0.171	16:56
Breath Volume	1.140	16:56
Air Blank	0.000	16:57
Air Blank	0.000	16:58
Subject Sample # 2	0.177	16:59
Breath Volume	1.496	16:59
Air Blank	0.000	17:00
Control Test	0.080	17:00
Air Blank	0.000	17:01
Diagnostics Check	OK	17:01

Data Download Date/Time: 06/22/2022 09:00

Subject Test Electronic Data

DATE/TIME OF TEST: 06/15/2022 16:15:32

Printed: 13 September 2022
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,DANIEL,S

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: STINSON,ERIC

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 14:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: LECROY

DATE/TIME OF TEST: 06/17/2022 15:31:03

Diagnostics Check	OK	15:24
Air Blank	0.000	15:25
Control Test	0.082	15:25
Air Blank	0.000	15:25
Subject Sample # 1	0.263	15:26
Breath Volume	1.703	15:26
Air Blank	0.000	15:27
Air Blank	0.000	15:29
Subject Sample # 2	0.270	15:29
Breath Volume	2.000	15:29
Air Blank	0.000	15:30
Control Test	0.078	15:30
Air Blank	0.000	15:30
Diagnostics Check	OK	15:31

Data Download Date/Time: 06/22/2022 09:00

Subject Test Electronic Data

DATE/TIME OF TEST: 06/17/2022 17:00:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MURSCH,THEODORE,R

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: MURSCH,THEODORE

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 05/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 17:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: CASTELLANOS

DATE/TIME OF TEST: 06/17/2022 17:50:52

Diagnostics Check	OK	17:47	
Air Blank	0.000	17:48	
Control Test	0.082	17:48	
Air Blank	0.000	17:49	
Subject Sample	# 1	17:49	Subject Test Refused
Breath Volume	0.000	17:49	
Air Blank	0.000	17:49	
Control Test	0.081	17:50	
Air Blank	0.000	17:50	
Diagnostics Check	OK	17:50	

Data Download Date/Time: 06/22/2022 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LECOCQ,MATTHEW,E

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: LECOCQ,MATTHEW

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 16:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: BOE

DATE/TIME OF TEST: 06/18/2022 17:09:49

Diagnostics Check	OK	17:04	
Air Blank	0.000	17:04	
Control Test	0.081	17:05	
Air Blank	0.000	17:05	
Subject Sample	# 1 0.025	17:06	
Breath Volume	3.867	17:06	
Air Blank	0.000	17:06	
Air Blank	0.000	17:08	
Subject Sample	# 2 0.017	17:08	
Breath Volume	2.675	17:08	
Air Blank		17:09	RFI Detect
Air Blank	0.000	17:09	

Data Download Date/Time: 06/22/2022 09:00

Subject Test Electronic Data

DATE/TIME OF TEST: 06/18/2022 17:59:01

Printed: 13 September 2022
COBRA ©1997-2014 CMI, Inc.

Subject Test Electronic Data

DATE/TIME OF TEST: 06/19/2022 17:45:17

Printed: 13 September 2022
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MURSCH,THEODORE,R

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: PRITCHARD,THOMAS

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 05/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 20:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: EIMERS

DATE/TIME OF TEST: 06/19/2022 20:48:07

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:43	
Control Test	0.080	20:44	
Air Blank	0.000	20:44	
Subject Sample	# 1	20:46	Subject Test Refused
Breath Volume	0.000	20:46	
Air Blank	0.000	20:47	
Control Test	0.080	20:47	
Air Blank	0.000	20:48	
Diagnostics Check	OK	20:48	

Data Download Date/Time: 06/22/2022 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LECOCQ,MATTHEW,E

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: STINSON,ERIC

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 14:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 06/22/2022 15:27:13

Diagnostics Check	OK	15:21
Air Blank	0.000	15:21
Control Test	0.081	15:21
Air Blank	0.000	15:22
Subject Sample # 1	0.077	15:22
Breath Volume	2.687	15:22
Air Blank	0.000	15:23
Air Blank	0.000	15:25
Subject Sample # 2	0.073	15:25
Breath Volume	3.296	15:25
Air Blank	0.000	15:26
Control Test	0.080	15:26
Air Blank	0.000	15:27
Diagnostics Check	OK	15:27

Data Download Date/Time: 07/29/2022 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STINSON,ERIC,S

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: WILLIAMS,DANIEL

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 16:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: BARRAS

DATE/TIME OF TEST: 06/22/2022 16:53:35

Diagnostics Check	OK	16:47
Air Blank	0.000	16:47
Control Test	0.082	16:48
Air Blank	0.000	16:48
Subject Sample # 1	0.125	16:49
Breath Volume	1.234	16:49
Air Blank	0.000	16:49
Air Blank	0.000	16:51
Subject Sample # 2	0.120	16:52
Breath Volume	1.144	16:52
Air Blank	0.000	16:52
Control Test	0.081	16:52
Air Blank	0.000	16:53
Diagnostics Check	OK	16:53

Data Download Date/Time: 07/29/2022 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,DANIEL,S

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: POWELL,MICHAEL

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 13:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 13:37

SUBJECT'S LAST NAME: TANKSLEY

DATE/TIME OF TEST: 06/23/2022 14:05:08

Diagnostics Check	OK	13:58
Air Blank	0.000	13:59
Control Test	0.082	13:59
Air Blank	0.000	13:59
Subject Sample # 1	0.173	14:00
Breath Volume	3.515	14:00
Air Blank	0.000	14:01
Air Blank	0.000	14:02
Subject Sample # 2	0.181	14:03
Breath Volume	2.796	14:03
Air Blank	0.000	14:04
Control Test	0.080	14:04
Air Blank	0.000	14:05
Diagnostics Check	OK	14:05

Data Download Date/Time: 07/29/2022 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINSON, WILLIAM, T

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: STALLINGS, JOSHUA

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 16:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/25/2022 16:55:00

Diagnostics Check	OK	16:51	
Air Blank	0.000	16:52	
Control Test	0.083	16:52	
Air Blank	0.000	16:52	
Subject Sample	# 1	16:53	Subject Test Refused
Breath Volume	0.000	16:53	
Air Blank	0.000	16:54	
Control Test	0.084	16:54	
Air Blank	0.000	16:54	
Diagnostics Check	OK	16:55	

Data Download Date/Time: 07/29/2022 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1349789
BREATH TEST OPERATOR: POWELL,MICHAEL,T	DRY GAS EXPIRY DATE: 07/15/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: POWELL,MICHAEL	ARREST TIME: 14:08
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2022	OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: SALAZAR-NEGRON DATE/TIME OF TEST: 06/26/2022 14:43:21

Diagnostics Check	OK	14:36
Air Blank	0.000	14:37
Control Test	0.082	14:37
Air Blank	0.000	14:37
Subject Sample # 1	0.138	14:38
Breath Volume	1.179	14:38
Air Blank	0.000	14:39
Air Blank	0.000	14:41
Subject Sample # 2	0.155	14:41
Breath Volume	1.324	14:41
Air Blank	0.000	14:42
Control Test	0.083	14:42
Air Blank	0.000	14:43
Diagnostics Check	OK	14:43

Data Download Date/Time: 07/29/2022 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MALTAIS,ANDREW,W

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: MALTAIS,ANDREW

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 17:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 06/26/2022 17:55:02

Diagnostics Check	OK	17:52	
Air Blank	0.000	17:53	
Control Test	0.077	17:53	
Air Blank		17:54	Purge Fail
Air Blank	0.000	17:55	

Data Download Date/Time: 07/29/2022 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1349789
BREATH TEST OPERATOR: MALTAIS,ANDREW,W	DRY GAS EXPIRY DATE: 07/15/2023
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: MALTAIS,ANDREW	ARREST TIME: 17:15
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2022	OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: HENDERSON	DATE/TIME OF TEST: 06/26/2022	18:01:40
--------------------------------	-------------------------------	----------

Diagnostics Check	OK	17:57	
Air Blank	0.000	17:58	
Control Test	0.080	17:58	
Air Blank	0.000	17:59	
Subject Sample	# 1	18:00	Slope Not Met
Breath Volume	2.285	18:00	
Air Blank	0.000	18:00	
Control Test	0.085	18:01	
Air Blank	0.000	18:01	
Diagnostics Check	OK	18:01	

Data Download Date/Time: 07/29/2022 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MALTAIS,ANDREW,W

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: MALTAIS,ANDREW

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 17:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 06/26/2022 18:10:37

Diagnostics Check	OK	18:03
Air Blank	0.000	18:04
Control Test	0.081	18:04
Air Blank	0.000	18:05
Subject Sample # 1	0.024	18:05
Breath Volume	2.492	18:05
Air Blank	0.000	18:06
Air Blank	0.000	18:08
Subject Sample # 2	0.018	18:09
Breath Volume	1.574	18:09
Air Blank	0.000	18:09
Control Test	0.078	18:10
Air Blank	0.000	18:10
Diagnostics Check	OK	18:10

Data Download Date/Time: 07/29/2022 10:58

Subject Test Electronic Data

DATE/TIME OF TEST: 06/27/2022 16:39:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAM,DANIEL,S

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: WILLIAMS,DANIEL

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 18:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: MOLIQUE

DATE/TIME OF TEST: 06/27/2022 19:22:45

Diagnostics Check	OK	19:14	
Air Blank	0.000	19:14	
Control Test	0.082	19:14	
Air Blank	0.000	19:15	
Subject Sample	# 1	19:18	No Sample Provided
Breath Volume	0.000	19:18	
Air Blank	0.000	19:19	
Air Blank	0.000	19:21	
Subject Sample	# 2	19:21	Subject Test Refused
Breath Volume	0.000	19:21	
Air Blank	0.000	19:21	
Control Test	0.080	19:22	
Air Blank	0.000	19:22	
Diagnostics Check	OK	19:22	

Data Download Date/Time: 07/29/2022 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STINSON,ERIC,S

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: STINSON,ERIC

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 18:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 06/27/2022 19:28:32

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:25	
Control Test	0.081	19:25	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:27	Subject Test Refused
Breath Volume	0.000	19:27	
Air Blank	0.000	19:27	
Control Test	0.085	19:27	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:28	

Data Download Date/Time: 07/29/2022 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LECOCQ,MATTHEW,E

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: LECOCQ,MATTHEW

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: BUI

ARREST TIME: 17:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: BEDWELL

DATE/TIME OF TEST: 06/28/2022 17:59:09

Diagnostics Check	OK	17:52
Air Blank	0.000	17:53
Control Test	0.080	17:53
Air Blank	0.000	17:54
Subject Sample # 1	0.201	17:54
Breath Volume	2.492	17:54
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample # 2	0.200	17:57
Breath Volume	2.179	17:57
Air Blank	0.000	17:58
Control Test	0.079	17:58
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

Data Download Date/Time: 07/29/2022 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYCHOFF, COLLIN, T
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: WAYCHOFF, COLLIN
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SHOOK

DATE/TIME OF TEST: 06/25/2022 03:01:49

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.067	02:57
Breath Volume	1.789	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.066	03:00
Breath Volume	2.050	03:00
Air Blank	0.000	03:00
Control Test	0.077	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 07/12/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PEREZ AMAYA

DATE/TIME OF TEST: 06/05/2022 02:38:32

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.083	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.000	02:33
Breath Volume	2.628	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.000	02:37
Breath Volume	2.949	02:37
Air Blank	0.000	02:37
Control Test	0.082	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 06/10/2022 02:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,CARLOS,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WEINTRAUB,E
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/05/2022 04:53:04

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.081	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.062	04:48
Breath Volume	2.003	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.064	04:51
Breath Volume	2.898	04:51
Air Blank	0.000	04:52
Control Test	0.077	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:53

Data Download Date/Time: 06/10/2022 02:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JAVIER,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: RODRIGUEZ,DARIEN
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: ANDION

DATE/TIME OF TEST: 06/06/2022 09:01:15

Diagnostics Check	OK	08:58	
Air Blank	0.000	08:58	
Control Test	0.082	08:59	
Air Blank	0.000	08:59	
Subject Sample	# 1	08:59	Subject Test Refused
Breath Volume	0.000	08:59	
Air Blank	0.000	09:00	
Control Test	0.083	09:00	
Air Blank	0.000	09:01	
Diagnostics Check	OK	09:01	

Data Download Date/Time: 06/10/2022 02:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ORBETA,JORGE,

BREATH TEST OPERATOR AGENCY: South Miami PD

ARRESTING OFFICER: ORBETA,JORGE

ARRESTING OFFICER AGENCY: South Miami PD

LAST AGENCY INSPECTION: 05/18/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1380119

DRY GAS EXPIRY DATE: 09/22/2023

VIOLATION CODE: DUI

ARREST TIME: 02:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WILICK

DATE/TIME OF TEST: 06/09/2022 02:55:47

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.084	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.084	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 06/10/2022 02:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRBETA,JORGE,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 06/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ CARRION

DATE/TIME OF TEST: 06/10/2022 03:21:41

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.083	03:11	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:14	No Sample Provided
Breath Volume	0.000	03:14	
Air Blank	0.000	03:15	
Air Blank	0.000	03:17	
Subject Sample	# 2	03:20	No Sample Provided
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.084	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 07/26/2022 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ORBETA,JORGE,

BREATH TEST OPERATOR AGENCY: South Miami PD

ARRESTING OFFICER: ORBETA,JORGE

ARRESTING OFFICER AGENCY: South Miami PD

LAST AGENCY INSPECTION: 06/10/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1380119

DRY GAS EXPIRY DATE: 09/22/2023

VIOLATION CODE: DUI

ARREST TIME: 02:18

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ CARRION

DATE/TIME OF TEST: 06/10/2022 03:26:50

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.083	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.083	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 07/26/2022 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ORBETA,JORGE,

BREATH TEST OPERATOR AGENCY: South Miami PD

ARRESTING OFFICER: ORBETA,JORGE

ARRESTING OFFICER AGENCY: South Miami PD

LAST AGENCY INSPECTION: 06/10/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1380119

DRY GAS EXPIRY DATE: 09/22/2023

VIOLATION CODE: DUI

ARREST TIME: 02:18

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ CARRION

DATE/TIME OF TEST: 06/10/2022 03:43:07

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.083	03:42	
Air Blank		03:42	RFI Detect
Air Blank	0.000	03:43	

Data Download Date/Time: 07/26/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ORBETA,JORGE,

BREATH TEST OPERATOR AGENCY: South Miami PD

ARRESTING OFFICER: ORBETA,JORGE

ARRESTING OFFICER AGENCY: South Miami PD

LAST AGENCY INSPECTION: 06/10/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1380119

DRY GAS EXPIRY DATE: 09/22/2023

VIOLATION CODE: DUI

ARREST TIME: 02:18

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ CARRION

DATE/TIME OF TEST: 06/10/2022 03:52:15

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.084	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.175	03:47
Breath Volume	1.785	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.164	03:50
Breath Volume	2.402	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/26/2022 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: SARRIA HENRIQUEZ

DATE/TIME OF TEST: 06/12/2022 04:22:39

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.124	04:17
Breath Volume	1.156	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.133	04:21
Breath Volume	2.406	04:21
Air Blank	0.000	04:21
Control Test	0.082	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 07/26/2022 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOWNE,SHAWN,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CARRASCO,JONATHAN
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: BALDAN

DATE/TIME OF TEST: 06/17/2022 21:02:01

Diagnostics Check	OK	20:59	
Air Blank	0.000	20:59	
Control Test	0.080	20:59	
Air Blank	0.000	21:00	
Subject Sample	# 1	21:00	Subject Test Refused
Breath Volume	0.000	21:00	
Air Blank	0.000	21:01	
Control Test	0.080	21:01	
Air Blank	0.000	21:01	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 07/26/2022 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VILLEDA

DATE/TIME OF TEST: 06/19/2022 03:16:17

Diagnostics Check	OK	03:15	
Air Blank		03:15	RFI Detect
Air Blank	0.000	03:16	

Data Download Date/Time: 07/26/2022 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VILLEDA

DATE/TIME OF TEST: 06/19/2022 03:22:21

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.081	03:20	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.081	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 07/26/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RUBI

DATE/TIME OF TEST: 06/26/2022 23:39:00

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.206	23:34
Breath Volume	1.660	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.215	23:37
Breath Volume	1.859	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:39

Data Download Date/Time: 07/26/2022 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CARPENTER, RICHARD
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 06/02/2022 03:33:42

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.189	03:28
Breath Volume	2.156	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.187	03:31
Breath Volume	2.042	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 06/28/2022 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: MIGUEL ANDRES

DATE/TIME OF TEST: 06/05/2022 01:38:20

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.148	01:32
Breath Volume	1.925	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.146	01:36
Breath Volume	1.562	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 06/28/2022 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARBOGAST, JOSHUA, S
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BIONDO, NICK
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: BUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/10/2022 22:44:57

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.079	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.211	22:39
Breath Volume	1.796	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.201	22:43
Breath Volume	1.824	22:43
Air Blank	0.000	22:43
Control Test	0.078	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 06/28/2022 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 06/11/2022 04:07:00

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.078	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.182	04:02
Breath Volume	2.335	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.166	04:05
Breath Volume	2.878	04:05
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:07

Data Download Date/Time: 06/28/2022 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WEISS, MASON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/14/2022 18:37:17

Diagnostics Check	OK	18:30
Air Blank	0.000	18:31
Control Test	0.079	18:31
Air Blank	0.000	18:32
Subject Sample # 1	0.252	18:32
Breath Volume	1.406	18:32
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.263	18:35
Breath Volume	1.882	18:35
Air Blank	0.000	18:36
Control Test	0.079	18:36
Air Blank	0.000	18:37
Diagnostics Check	OK	18:37

Data Download Date/Time: 06/28/2022 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: MANKIN

DATE/TIME OF TEST: 06/15/2022 06:22:15

Diagnostics Check	OK	06:13
Air Blank	0.000	06:13
Control Test	0.079	06:14
Air Blank	0.000	06:14
Subject Sample # 1	0.188	06:17
Breath Volume	2.746	06:17
Air Blank	0.000	06:18
Air Blank	0.000	06:19
Subject Sample # 2	0.183	06:20
Breath Volume	1.312	06:20
Air Blank	0.000	06:21
Control Test	0.077	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 06/28/2022 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CLEARY, ALYSSA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: BASILIO

DATE/TIME OF TEST: 06/19/2022 04:45:35

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.218	04:40
Breath Volume	3.226	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.209	04:43
Breath Volume	2.765	04:43
Air Blank	0.000	04:44
Control Test	0.078	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 06/28/2022 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERRE,KRENID,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: HOEL,JEFFEREY
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 05/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: BOS

DATE/TIME OF TEST: 06/02/2022 01:19:35

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.078	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.171	01:15
Breath Volume	1.914	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.166	01:18
Breath Volume	2.140	01:18
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 06/14/2022 14:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 05/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 06/05/2022 03:56:45

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.159	03:51
Breath Volume	2.277	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.173	03:55
Breath Volume	3.101	03:55
Air Blank	0.000	03:55
Control Test	0.078	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 06/14/2022 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERRE,KRENID,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 05/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: TORRES MUNIZ

DATE/TIME OF TEST: 06/06/2022 19:28:57

Diagnostics Check	OK	19:22
Air Blank	0.000	19:22
Control Test	0.079	19:22
Air Blank	0.000	19:23
Subject Sample # 1	0.237	19:23
Breath Volume	2.531	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.225	19:27
Breath Volume	2.468	19:27
Air Blank	0.000	19:28
Control Test	0.077	19:28
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 06/14/2022 14:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BROWN,ANTUJUAN
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 05/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: HERNANDEZJIMENIZ

DATE/TIME OF TEST: 06/11/2022 11:47:35

Diagnostics Check	OK	11:41
Air Blank	0.000	11:41
Control Test	0.078	11:41
Air Blank	0.000	11:42
Subject Sample # 1	0.236	11:42
Breath Volume	1.894	11:42
Air Blank	0.000	11:43
Air Blank	0.000	11:45
Subject Sample # 2	0.243	11:46
Breath Volume	1.386	11:46
Air Blank	0.000	11:46
Control Test	0.078	11:46
Air Blank	0.000	11:47
Diagnostics Check	OK	11:47

Data Download Date/Time: 06/14/2022 14:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CARDOZA QUIJADA

DATE/TIME OF TEST: 06/27/2022 05:48:19

Diagnostics Check	OK	05:41
Air Blank	0.000	05:42
Control Test	0.078	05:42
Air Blank	0.000	05:42
Subject Sample # 1	0.098	05:43
Breath Volume	2.414	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.100	05:46
Breath Volume	2.449	05:46
Air Blank	0.000	05:47
Control Test	0.078	05:47
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 07/20/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILL,JULIA,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: GLOVER,LAURENCE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MARSH

DATE/TIME OF TEST: 06/18/2022 20:54:46

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.078	20:49
Air Blank	0.000	20:49
Subject Sample # 1	0.120	20:50
Breath Volume	3.488	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.120	20:53
Breath Volume	4.308	20:53
Air Blank	0.000	20:53
Control Test	0.078	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 06/30/2022 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,THOMAS,J
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: WEBB,THOMAS
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 06/20/2022 12:02:31

Diagnostics Check	OK	11:56
Air Blank	0.000	11:56
Control Test	0.079	11:57
Air Blank	0.000	11:57
Subject Sample # 1	0.000	11:58
Breath Volume	1.468	11:58
Air Blank	0.000	11:58
Air Blank	0.000	12:00
Subject Sample # 2	0.000	12:01
Breath Volume	1.492	12:01
Air Blank	0.000	12:01
Control Test	0.079	12:01
Air Blank	0.000	12:02
Diagnostics Check	OK	12:02

Data Download Date/Time: 06/30/2022 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,ROLF,E
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: GORDON,ROLF
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 06/03/2022 21:54:53

Diagnostics Check	OK	21:52	
Air Blank	0.000	21:53	
Control Test	0.079	21:53	
Air Blank	0.000	21:54	
Subject Sample	# 1	21:54	Sequence Aborted
Breath Volume	0.000	21:54	
Air Blank	0.000	21:54	

Data Download Date/Time: 06/30/2022 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,ROLF,E
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: GORDON,ROLF
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 06/03/2022 22:10:34

Diagnostics Check	OK	21:59	
Air Blank	0.000	21:59	
Control Test	0.078	22:00	
Air Blank	0.000	22:00	
Subject Sample	# 1	22:03	No Sample Provided
Breath Volume	0.000	22:03	
Air Blank	0.000	22:04	
Air Blank	0.000	22:06	
Subject Sample	# 2	22:09	Subject Test Refused
Breath Volume	0.000	22:09	
Air Blank	0.000	22:09	
Control Test	0.080	22:09	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

Data Download Date/Time: 06/30/2022 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, DAVID, H
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROSMINI, JOE
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BLACKMON

DATE/TIME OF TEST: 06/15/2022 22:19:54

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.076	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.031	22:15
Breath Volume	1.800	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.031	22:18
Breath Volume	2.320	22:18
Air Blank	0.000	22:19
Control Test	0.078	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 06/27/2022 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BLEVINS,A
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/12/2022 04:37:45

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.079	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.147	04:33
Breath Volume	1.718	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:35
Subject Sample # 2	0.139	04:36
Breath Volume	1.851	04:36
Air Blank	0.000	04:36
Control Test	0.079	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 07/19/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JAMES,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: NEDD

DATE/TIME OF TEST: 06/19/2022 07:54:12

Diagnostics Check	OK	07:47
Air Blank	0.000	07:48
Control Test	0.080	07:48
Air Blank	0.000	07:49
Subject Sample # 1	0.026	07:49
Breath Volume	1.570	07:49
Air Blank	0.000	07:50
Air Blank	0.000	07:52
Subject Sample # 2	0.026	07:52
Breath Volume	1.785	07:52
Air Blank	0.000	07:53
Control Test	0.080	07:53
Air Blank	0.000	07:54
Diagnostics Check	OK	07:54

Data Download Date/Time: 07/19/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORAN,ETHAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FORAN,ETHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: LOPEZ MARTINEZ

DATE/TIME OF TEST: 06/08/2022 15:08:31

Diagnostics Check	OK	15:01
Air Blank	0.000	15:01
Control Test	0.082	15:01
Air Blank	0.000	15:02
Subject Sample # 1	0.315	15:03
Breath Volume	1.136	15:03
Air Blank	0.000	15:04
Air Blank	0.000	15:05
Subject Sample # 2	0.302	15:06
Breath Volume	2.472	15:06
Air Blank	0.000	15:07
Control Test	0.080	15:07
Air Blank	0.000	15:08
Diagnostics Check	OK	15:08

Data Download Date/Time: 06/27/2022 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SSANDERS,DEON,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SANDERS,DEON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 06/12/2022 04:33:27

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.082	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Control Test	0.081	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 06/27/2022 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: STROHECKER,KYNDAL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 06/14/2022 04:15:10

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.007	04:10
Breath Volume	3.480	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.007	04:13
Breath Volume	2.664	04:13
Air Blank	0.000	04:14
Control Test	0.082	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 06/27/2022 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MURRELL

DATE/TIME OF TEST: 06/17/2022 05:47:54

Diagnostics Check	OK	05:46	
Air Blank	0.000	05:46	
Control Test	0.081	05:47	
Air Blank		05:47	RFI Detect
Air Blank	0.000	05:47	

Data Download Date/Time: 06/27/2022 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: \34520080A
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MURRELL

DATE/TIME OF TEST: 06/17/2022 05:54:08

Diagnostics Check	OK	05:51	
Air Blank	0.000	05:51	
Control Test	0.081	05:52	
Air Blank	0.000	05:52	
Subject Sample	# 1	05:52	RFI Detect
Breath Volume	0.000	05:52	
Air Blank	0.000	05:53	
Control Test	0.081	05:53	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:54	

Data Download Date/Time: 06/27/2022 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MURRELL

DATE/TIME OF TEST: 06/17/2022 06:03:15

Diagnostics Check	OK	05:56
Air Blank	0.000	05:56
Control Test	0.081	05:56
Air Blank	0.000	05:57
Subject Sample # 1	0.120	05:58
Breath Volume	2.246	05:58
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.115	06:01
Breath Volume	1.808	06:01
Air Blank	0.000	06:02
Control Test	0.082	06:02
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

Data Download Date/Time: 06/27/2022 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ROSA NUNEZ

DATE/TIME OF TEST: 06/21/2022 22:58:02

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.082	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.192	22:53
Breath Volume	3.640	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.173	22:56
Breath Volume	3.609	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:58

Data Download Date/Time: 06/27/2022 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 06/25/2022 22:12:20

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:57	
Control Test	0.081	21:58	
Air Blank	0.000	21:58	
Subject Sample	# 1 0.199	22:01	Volume Not Met
Breath Volume	1.382	22:01	
Air Blank	0.000	22:02	
Air Blank	0.000	22:04	
Subject Sample	# 2 0.204	22:05	
Breath Volume	1.519	22:05	
Air Blank	0.000	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 3 0.193	22:10	
Breath Volume	1.117	22:10	
Air Blank	0.000	22:11	
Control Test	0.082	22:11	
Air Blank	0.000	22:12	
Diagnostics Check	OK	22:12	

Data Download Date/Time: 06/27/2022 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,BOBBY,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MORALES,BOBBY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 06/04/2022 20:05:00

Diagnostics Check	OK	19:58
Air Blank	0.000	19:59
Control Test	0.079	19:59
Air Blank	0.000	20:00
Subject Sample # 1	0.068	20:00
Breath Volume	1.734	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.072	20:03
Breath Volume	3.574	20:03
Air Blank	0.000	20:04
Control Test	0.080	20:04
Air Blank	0.000	20:04
Diagnostics Check	OK	20:05

Data Download Date/Time: 06/30/2022 15:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,ANGIE,B
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: CAUSSEAU,DYLAN
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: RUBIN

DATE/TIME OF TEST: 06/04/2022 20:42:42

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.079	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.169	20:38
Breath Volume	3.195	20:38
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.178	20:41
Breath Volume	2.828	20:41
Air Blank	0.000	20:41
Control Test	0.079	20:42
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 06/30/2022 15:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: 1S07,SETH,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: GRAY,SETH
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 14:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/11/2022 16:32:22

Diagnostics Check	OK	16:26
Air Blank	0.000	16:26
Control Test	0.079	16:26
Air Blank	0.000	16:27
Subject Sample # 1	0.000	16:28
Breath Volume	3.230	16:28
Air Blank	0.000	16:28
Air Blank	0.000	16:30
Subject Sample # 2	0.000	16:31
Breath Volume	3.464	16:31
Air Blank	0.000	16:31
Control Test	0.080	16:31
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

Data Download Date/Time: 06/30/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLSON,SCOTT,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: OLSON,SCOTT
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 06/11/2022 20:56:51

Diagnostics Check	OK	20:52	
Air Blank	0.000	20:52	
Control Test	0.080	20:53	
Air Blank	0.000	20:53	
Subject Sample	# 1	20:55	Subject Test Refused
Breath Volume	0.000	20:55	
Air Blank	0.000	20:55	
Control Test	0.080	20:56	
Air Blank	0.000	20:56	
Diagnostics Check	OK	20:56	

Data Download Date/Time: 06/30/2022 15:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLSON,SCOTT,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: OLSON,SCOTT
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 06/11/2022 21:11:27

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.079	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.097	21:06
Breath Volume	4.957	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.098	21:09
Breath Volume	5.214	21:09
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 06/30/2022 15:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,NATHAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: PIENTA,DAVID
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: MILLETTE

DATE/TIME OF TEST: 06/19/2022 22:30:34

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:28	
Control Test	0.079	22:28	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:29	Subject Test Refused
Breath Volume	0.000	22:29	
Air Blank	0.000	22:29	
Control Test	0.079	22:29	
Air Blank	0.000	22:30	
Diagnostics Check	OK	22:30	

Data Download Date/Time: 06/30/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: BRINKLEY

DATE/TIME OF TEST: 06/22/2022 00:59:37

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.137	00:55
Breath Volume	2.738	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.122	00:58
Breath Volume	1.683	00:58
Air Blank	0.000	00:58
Control Test	0.080	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 06/30/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HARLEY

DATE/TIME OF TEST: 06/26/2022 23:12:14

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:00	
Control Test	0.079	23:01	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:04	No Sample Provided
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	
Air Blank	0.000	23:07	
Subject Sample	# 2	0.152	23:07
Breath Volume	1.992	23:07	
Air Blank	0.000	23:08	
Air Blank	0.000	23:10	
Subject Sample	# 3	0.154	23:10
Breath Volume	1.800	23:10	
Air Blank	0.000	23:11	
Control Test	0.080	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 06/30/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KRUPINSKI,LUKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: OCASIO

DATE/TIME OF TEST: 06/03/2022 21:03:21

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.079	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.243	20:58
Breath Volume	1.175	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.244	21:01
Breath Volume	2.105	21:01
Air Blank	0.000	21:02
Control Test	0.077	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 06/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 06/04/2022 03:48:42

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.151	03:44
Breath Volume	1.175	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.151	03:47
Breath Volume	1.414	03:47
Air Blank	0.000	03:47
Control Test	0.078	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 06/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CUEVAS,JOSE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: DARISO

DATE/TIME OF TEST: 06/04/2022 04:40:07

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.091	04:34
Breath Volume	1.300	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.084	04:38
Breath Volume	1.175	04:38
Air Blank	0.000	04:39
Control Test	0.078	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 06/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PREDELUS,SMITH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/04/2022 20:28:41

Diagnostics Check	OK	20:22
Air Blank	0.000	20:23
Control Test	0.078	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.177	20:24
Breath Volume	1.902	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:26
Subject Sample # 2	0.165	20:27
Breath Volume	1.218	20:27
Air Blank	0.000	20:27
Control Test	0.077	20:28
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 06/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: HARDING

DATE/TIME OF TEST: 06/05/2022 03:30:36

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.000	03:26
Breath Volume	2.121	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.000	03:29
Breath Volume	2.242	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 06/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 06/05/2022 19:00:17

Diagnostics Check	OK	18:54
Air Blank	0.000	18:54
Control Test	0.079	18:54
Air Blank	0.000	18:55
Subject Sample # 1	0.127	18:55
Breath Volume	2.082	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.132	18:58
Breath Volume	2.406	18:58
Air Blank	0.000	18:59
Control Test	0.077	18:59
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 06/13/2022 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 06/05/2022 21:01:47

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.079	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.200	20:57
Breath Volume	1.707	20:57
Air Blank	0.000	20:58
Air Blank	0.000	20:59
Subject Sample # 2	0.198	21:00
Breath Volume	1.675	21:00
Air Blank	0.000	21:00
Control Test	0.078	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 06/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHIPLEY,DANIEL
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BEATO

DATE/TIME OF TEST: 06/10/2022 23:12:29

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.215	23:08
Breath Volume	1.343	23:08
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.207	23:10
Breath Volume	1.914	23:10
Air Blank	0.000	23:11
Control Test	0.077	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 06/13/2022 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOSA,EMMANUEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GUEVARA HERNANDEZ

DATE/TIME OF TEST: 06/11/2022 21:52:38

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.079	21:44
Air Blank	0.000	21:45
Subject Sample # 1	0.256	21:47
Breath Volume	1.929	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.252	21:51
Breath Volume	2.472	21:51
Air Blank	0.000	21:51
Control Test	0.077	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 06/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRICKEY,ALEX
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: TATUM

DATE/TIME OF TEST: 06/12/2022 02:48:55

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.146	02:44
Breath Volume	2.179	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.147	02:47
Breath Volume	2.000	02:47
Air Blank	0.000	02:48
Control Test	0.078	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 06/13/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHOEN,ELLIS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: NINCHE HERNANDEZ

DATE/TIME OF TEST: 06/16/2022 03:34:58

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.122	03:30
Breath Volume	1.578	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.126	03:33
Breath Volume	1.464	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MACHADO NARDELLI

DATE/TIME OF TEST: 06/17/2022 02:49:12

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.096	02:44
Breath Volume	2.492	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.095	02:47
Breath Volume	2.253	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAUNDERS,HUNTER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 06/18/2022 02:01:26

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.035	01:57
Breath Volume	1.558	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.033	01:59
Breath Volume	2.355	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 06/19/2022 07:50:38

Diagnostics Check	OK	07:44
Air Blank	0.000	07:44
Control Test	0.079	07:44
Air Blank	0.000	07:45
Subject Sample # 1	0.139	07:45
Breath Volume	1.781	07:45
Air Blank	0.000	07:46
Air Blank	0.000	07:48
Subject Sample # 2	0.141	07:49
Breath Volume	2.484	07:49
Air Blank	0.000	07:49
Control Test	0.079	07:50
Air Blank	0.000	07:50
Diagnostics Check	OK	07:50

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILLS,DARREN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 06/23/2022 02:03:43

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.193	01:58
Breath Volume	1.375	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.180	02:02
Breath Volume	1.445	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: RIVERA AYALA

DATE/TIME OF TEST: 06/23/2022 03:25:41

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.160	03:21
Breath Volume	2.085	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample # 2	0.157	03:24
Breath Volume	1.714	03:24
Air Blank	0.000	03:24
Control Test	0.078	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DOYNO

DATE/TIME OF TEST: 06/24/2022 02:53:11

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.215	02:48
Breath Volume	1.878	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.206	02:51
Breath Volume	2.789	02:51
Air Blank	0.000	02:52
Control Test	0.077	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD,ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: RESENDIZ MARTINEZ

DATE/TIME OF TEST: 06/24/2022 04:07:26

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.110	04:03
Breath Volume	2.750	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.108	04:05
Breath Volume	2.574	04:05
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BOYD,RENEE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: WHITTINGTON

DATE/TIME OF TEST: 06/25/2022 00:39:09

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.120	00:34
Breath Volume	1.179	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.119	00:37
Breath Volume	1.371	00:37
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATSON,ISAIAH
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: HUGHES JR

DATE/TIME OF TEST: 06/25/2022 01:47:33

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:33	
Control Test	0.079	01:33	
Air Blank	0.000	01:34	
Subject Sample	# 1 0.138	01:37	Volume Not Met
Breath Volume	0.277	01:37	
Air Blank	0.000	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 2 0.141	01:40	
Breath Volume	1.109	01:40	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 3 0.145	01:46	Volume Not Met
Breath Volume	0.222	01:46	
Air Blank	0.000	01:46	
Control Test	0.078	01:46	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATSON,ISAIAH
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: HUGHES JR

DATE/TIME OF TEST: 06/25/2022 02:01:39

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.077	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1 0.130	01:54	Volume Not Met
Breath Volume	0.812	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:56	
Subject Sample	# 2 0.140	02:00	Volume Not Met
Breath Volume	0.218	02:00	
Air Blank	0.000	02:00	
Control Test	0.077	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CALERO,LUDEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: MICOURT

DATE/TIME OF TEST: 06/25/2022 04:19:56

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.218	04:15
Breath Volume	1.703	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.224	04:18
Breath Volume	1.601	04:18
Air Blank	0.000	04:19
Control Test	0.076	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STOWE, NICOLAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MALAYKHAN

DATE/TIME OF TEST: 06/26/2022 04:15:30

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:04	
Control Test	0.078	04:04	
Air Blank	0.000	04:05	
Subject Sample	# 1 0.218	04:08	Volume Not Met
Breath Volume	0.105	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 2 0.219	04:13	Volume Not Met
Breath Volume	0.160	04:13	
Air Blank	0.000	04:14	
Control Test	0.076	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIESENFELT,JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: ZAHRIYEH

DATE/TIME OF TEST: 06/29/2022 03:42:36

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.190	03:38
Breath Volume	1.808	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.183	03:41
Breath Volume	2.015	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUNT,FRANCIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: NORONHANOVAESPEREIRA

DATE/TIME OF TEST: 06/29/2022 23:08:05

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.000	23:03
Breath Volume	2.031	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.000	23:06
Breath Volume	1.863	23:06
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 07/05/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: SILVIO HERNANDEZ

DATE/TIME OF TEST: 06/30/2022 03:44:09

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.026	03:39
Breath Volume	1.511	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.024	03:42
Breath Volume	1.375	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 07/05/2022 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: WILLIAMS,ROBERT
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: EATON

DATE/TIME OF TEST: 06/11/2022 03:37:44

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.079	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 1 0.198	03:25	
Breath Volume	1.910	03:25	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2	03:30	No Sample Provided
Breath Volume	0.000	03:30	
Air Blank	0.000	03:31	
Air Blank	0.000	03:33	
Subject Sample	# 3	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.078	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 06/13/2022 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: WILLIAMS,ROBERT
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: DICKEY

DATE/TIME OF TEST: 06/25/2022 03:53:59

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.079	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:47	No Sample Provided
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Air Blank	0.000	03:49	
Subject Sample	# 2	03:52	No Sample Provided
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Control Test	0.079	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 07/11/2022 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAKE,PALMER,A
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: DILL

DATE/TIME OF TEST: 06/27/2022 23:42:33

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.277	23:37
Breath Volume	3.546	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.269	23:40
Breath Volume	2.671	23:40
Air Blank	0.000	23:41
Control Test	0.077	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 07/11/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: BOTET

DATE/TIME OF TEST: 06/01/2022 03:24:11

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.000	03:19
Breath Volume	1.996	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.000	03:22
Breath Volume	2.816	03:22
Air Blank	0.000	03:23
Control Test	0.081	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 06/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,BRIAN,G
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: VAZQUEZ,RAUL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MADERA

DATE/TIME OF TEST: 06/02/2022 00:57:43

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.138	00:52
Breath Volume	2.632	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.136	00:56
Breath Volume	2.476	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 06/13/2022 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LARMOND

DATE/TIME OF TEST: 06/03/2022 00:44:47

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:33	
Control Test	0.081	00:33	
Air Blank	0.000	00:34	
Subject Sample	# 1 0.179	00:37	Volume Not Met
Breath Volume	0.281	00:37	
Air Blank	0.000	00:38	
Air Blank	0.000	00:40	
Subject Sample	# 2 0.168	00:43	Volume Not Met
Breath Volume	0.214	00:43	
Air Blank	0.000	00:43	
Control Test	0.080	00:44	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 06/13/2022 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LARMOND

DATE/TIME OF TEST: 06/03/2022 01:05:54

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:52	
Subject Sample	# 1 0.172	00:55	Volume Not Met
Breath Volume	0.218	00:55	
Air Blank	0.000	00:56	
Air Blank	0.000	00:57	
Subject Sample	# 2 0.172	00:59	
Breath Volume	1.171	00:59	
Air Blank	0.000	00:59	
Air Blank	0.000	01:01	
Subject Sample	# 3	01:04	Subject Test Refused
Breath Volume	0.015	01:04	
Air Blank	0.000	01:04	
Control Test	0.079	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 06/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: LOGRONO SOLORZANO

DATE/TIME OF TEST: 06/04/2022 05:43:34

Diagnostics Check	OK	05:36
Air Blank	0.000	05:37
Control Test	0.081	05:37
Air Blank	0.000	05:38
Subject Sample # 1	0.145	05:38
Breath Volume	4.359	05:38
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample # 2	0.139	05:42
Breath Volume	4.242	05:42
Air Blank	0.000	05:42
Control Test	0.081	05:42
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time: 06/13/2022 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYBIRAL,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WYBIRAL,JUSTIN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 06/05/2022 02:16:28

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.189	02:11
Breath Volume	1.371	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.188	02:14
Breath Volume	2.410	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 06/13/2022 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MALDONADO,PRECIOUSLEE
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: VIOLA

DATE/TIME OF TEST: 06/05/2022 21:27:45

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.170	21:23
Breath Volume	1.511	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.167	21:26
Breath Volume	1.230	21:26
Air Blank	0.000	21:26
Control Test	0.080	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 06/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: LOPEZ-VEGA

DATE/TIME OF TEST: 06/08/2022 05:02:06

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.082	04:56
Air Blank	0.000	04:56
Subject Sample # 1	0.207	04:57
Breath Volume	2.308	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.187	05:00
Breath Volume	2.710	05:00
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 06/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/09/2022 23:07:42

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.081	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.000	23:03
Breath Volume	3.777	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.000	23:06
Breath Volume	3.738	23:06
Air Blank	0.000	23:06
Control Test	0.082	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 06/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: GARCIA-OTALORA

DATE/TIME OF TEST: 06/10/2022 23:46:42

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.082	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.077	23:42
Breath Volume	2.812	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.075	23:45
Breath Volume	2.792	23:45
Air Blank	0.000	23:45
Control Test	0.081	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 06/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/11/2022 01:46:03

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.149	01:41
Breath Volume	2.687	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.159	01:44
Breath Volume	2.601	01:44
Air Blank	0.000	01:45
Control Test	0.082	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:46

Data Download Date/Time: 06/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BOWEN,JAMES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 13:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 06/11/2022 15:55:32

Diagnostics Check	OK	15:49
Air Blank	0.000	15:49
Control Test	0.081	15:50
Air Blank	0.000	15:50
Subject Sample # 1	0.000	15:51
Breath Volume	1.679	15:51
Air Blank	0.000	15:51
Air Blank	0.000	15:53
Subject Sample # 2	0.000	15:54
Breath Volume	1.492	15:54
Air Blank	0.000	15:54
Control Test	0.081	15:54
Air Blank	0.000	15:55
Diagnostics Check	OK	15:55

Data Download Date/Time: 06/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SERRANOFigureOA

DATE/TIME OF TEST: 06/12/2022 03:23:15

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.081	03:12	
Air Blank	0.000	03:12	
Subject Sample	# 1 0.132	03:16	Volume Not Met
Breath Volume	0.121	03:16	
Air Blank	0.000	03:16	
Air Blank	0.000	03:18	
Subject Sample	# 2 0.123	03:21	Volume Not Met
Breath Volume	0.347	03:21	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 06/13/2022 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SERRANOFIGUEROA

DATE/TIME OF TEST: 06/12/2022 03:42:02

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.080	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1 0.094	03:34	Volume Not Met
Breath Volume	0.300	03:34	
Air Blank	0.000	03:35	
Air Blank	0.000	03:37	
Subject Sample	# 2 0.116	03:40	Volume Not Met
Breath Volume	0.621	03:40	
Air Blank	0.000	03:41	
Control Test	0.081	03:41	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 06/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PERSUAD,PATRICK
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ALVARADO CUEVAS

DATE/TIME OF TEST: 06/13/2022 02:32:37

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.066	02:27
Breath Volume	1.949	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.065	02:31
Breath Volume	1.941	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 06/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYLE,OWEN,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GAYLE,OWEN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 14:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 06/14/2022 15:31:14

Diagnostics Check	OK	15:22	
Air Blank	0.000	15:23	
Control Test	0.081	15:23	
Air Blank	0.000	15:24	
Subject Sample	# 1	15:27	No Sample Provided
Breath Volume	0.000	15:27	
Air Blank	0.000	15:27	
Air Blank	0.000	15:29	
Subject Sample	# 2	15:29	Subject Test Refused
Breath Volume	0.000	15:29	
Air Blank	0.000	15:30	
Control Test	0.081	15:30	
Air Blank	0.000	15:31	
Diagnostics Check	OK	15:31	

Data Download Date/Time: 07/12/2022 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/16/2022 01:02:02

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.051	00:57
Breath Volume	3.519	00:57
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.052	01:00
Breath Volume	3.218	01:00
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:02

Data Download Date/Time: 07/12/2022 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 06/17/2022 22:46:14

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.081	22:39	
Air Blank	0.000	22:40	
Subject Sample # 1	0.134	22:40	
Breath Volume	1.253	22:40	
Air Blank	0.000	22:41	
Air Blank	0.000	22:43	
Subject Sample # 2	0.141	22:44	
Breath Volume	1.484	22:44	
Air Blank	0.000	22:44	
Control Test	0.081	22:45	
Air Blank	0.000	22:45	
Diagnostics Check	Fail	22:45	Voltage/Current Test
Air Blank	0.000	22:46	

Data Download Date/Time: 07/12/2022 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 06/17/2022 23:01:53

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:54	
Control Test	0.081	22:54	
Air Blank	0.000	22:54	
Subject Sample	# 1 0.144	22:55	
Breath Volume	1.464	22:55	
Air Blank	0.000	22:56	
Air Blank	0.000	22:58	
Subject Sample	# 2 0.135	22:59	
Breath Volume	1.601	22:59	
Air Blank	0.000	23:00	
Control Test	0.081	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	Fail	23:01	Voltage/Current Test
Air Blank	0.000	23:01	

Data Download Date/Time: 07/12/2022 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HANSELL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: PEREZ MOLINA

DATE/TIME OF TEST: 06/18/2022 00:36:37

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.081	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1 0.179	00:31	
Breath Volume	1.289	00:31	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 2 0.178	00:34	
Breath Volume	1.183	00:34	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	Fail	00:36	Voltage/Current Test
Air Blank	0.000	00:36	

Data Download Date/Time: 07/12/2022 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BYNUM,CHRISTIE
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: GAMAS MORALES

DATE/TIME OF TEST: 06/04/2022 02:12:21

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.081	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:07	No Sample Provided
Breath Volume	0.000	02:07	
Air Blank	0.000	02:07	
Air Blank	0.000	02:09	
Subject Sample	# 2	02:11	Sequence Aborted
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	

Data Download Date/Time: 07/01/2022 08:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLBRIGHT,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: LAFLEUR,MATTHEW
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: CARRUBBA

DATE/TIME OF TEST: 06/04/2022 17:02:39

Diagnostics Check	OK	16:55
Air Blank	0.000	16:56
Control Test	0.080	16:56
Air Blank	0.000	16:57
Subject Sample # 1	0.164	16:58
Breath Volume	2.683	16:58
Air Blank	0.000	16:58
Air Blank	0.000	17:00
Subject Sample # 2	0.159	17:01
Breath Volume	2.343	17:01
Air Blank	0.000	17:01
Control Test	0.081	17:02
Air Blank	0.000	17:02
Diagnostics Check	OK	17:02

Data Download Date/Time: 07/01/2022 08:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BYNUM,CHRISTIE
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: ROUSE

DATE/TIME OF TEST: 06/10/2022 02:21:35

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.172	02:17
Breath Volume	4.222	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.176	02:20
Breath Volume	4.355	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 07/01/2022 08:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ROSA,EMILIO
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 06/14/2022 22:55:56

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:44	
Control Test	0.080	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:48	No Sample Provided
Breath Volume	0.000	22:48	
Air Blank	0.000	22:48	
Air Blank	0.000	22:50	
Subject Sample	# 2	0.329 22:51	
Breath Volume	2.437	22:51	
Air Blank	0.000	22:51	
Air Blank	0.000	22:53	
Subject Sample	# 3	0.345 22:54	
Breath Volume	2.582	22:54	
Air Blank	0.000	22:55	
Control Test	0.080	22:55	
Air Blank	0.000	22:55	
Diagnostics Check	OK	22:55	

Data Download Date/Time: 07/01/2022 08:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLSEN, IAN, N
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HATZEL, BRANDON
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 06/15/2022 10:00:39

Diagnostics Check	OK	09:45	
Air Blank	0.000	09:46	
Control Test	0.080	09:46	
Air Blank	0.000	09:46	
Subject Sample	# 1 0.125	09:47	
Breath Volume	1.308	09:47	
Air Blank	0.000	09:48	
Air Blank	0.000	09:50	
Subject Sample	# 2 0.113	09:53	Volume Not Met
Breath Volume	0.933	09:53	
Air Blank	0.000	09:54	
Air Blank	0.000	09:56	
Subject Sample	# 3 0.105	09:59	Volume Not Met
Breath Volume	0.500	09:59	
Air Blank	0.000	09:59	
Control Test	0.079	10:00	
Air Blank	0.000	10:00	
Diagnostics Check	OK	10:00	

Data Download Date/Time: 07/01/2022 08:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ROSA,EMILIO
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: DEANGELO

DATE/TIME OF TEST: 06/22/2022 21:24:18

Diagnostics Check	OK	21:17
Air Blank	0.000	21:18
Control Test	0.081	21:18
Air Blank	0.000	21:18
Subject Sample # 1	0.245	21:19
Breath Volume	2.046	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.260	21:22
Breath Volume	2.183	21:22
Air Blank	0.000	21:23
Control Test	0.080	21:23
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 07/01/2022 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ESCHBACH,DANIEL
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: MCGRAW

DATE/TIME OF TEST: 06/30/2022 13:35:06

Diagnostics Check	OK	13:28
Air Blank	0.000	13:29
Control Test	0.080	13:29
Air Blank	0.000	13:30
Subject Sample # 1	0.233	13:30
Breath Volume	2.800	13:30
Air Blank	0.000	13:31
Air Blank	0.000	13:33
Subject Sample # 2	0.245	13:33
Breath Volume	2.472	13:33
Air Blank	0.000	13:34
Control Test	0.081	13:34
Air Blank	0.000	13:35
Diagnostics Check	OK	13:35

Data Download Date/Time: 07/01/2022 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WEAVER,JACOB,T

BREATH TEST OPERATOR AGENCY: Titusville PD

ARRESTING OFFICER: PEPIATT,CAMERON

ARRESTING OFFICER AGENCY: Titusville PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 33321080A2

DRY GAS EXPIRY DATE: 01/05/2024

VIOLATION CODE: DUI

ARREST TIME: 14:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:34

SUBJECT'S LAST NAME: DELGRIPPO

DATE/TIME OF TEST: 06/30/2022 15:13:49

Diagnostics Check	OK	15:06
Air Blank	0.000	15:07
Control Test	0.079	15:07
Air Blank	0.000	15:08
Subject Sample # 1	0.169	15:09
Breath Volume	4.417	15:09
Air Blank	0.000	15:09
Air Blank	0.000	15:11
Subject Sample # 2	0.166	15:12
Breath Volume	2.960	15:12
Air Blank	0.000	15:12
Control Test	0.078	15:13
Air Blank	0.000	15:13
Diagnostics Check	OK	15:13

Data Download Date/Time: 08/09/2022 17:30