

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,CAMRYN,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HILL,CAMRYN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: TERRETT

DATE/TIME OF TEST: 06/04/2022 01:16:24

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.079	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	1.437	01:14	
Air Blank	0.000	01:15	
Control Test	0.080	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 06/22/2022 19:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ROBISON

DATE/TIME OF TEST: 06/09/2022 03:12:03

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.078	03:07	
Air Blank	0.000	03:08	
Subject Sample	# 1 0.187	03:10	
Breath Volume	1.343	03:10	
Air Blank		03:11	Purge Fail
Air Blank		03:12	Purge Fail

Data Download Date/Time: 06/22/2022 19:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: PRUNTY

DATE/TIME OF TEST: 06/02/2022 22:53:32

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.000	22:49
Breath Volume	4.117	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.000	22:52
Breath Volume	5.289	22:52
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 06/08/2022 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: MICKEY

DATE/TIME OF TEST: 06/03/2022 18:05:40

Diagnostics Check	OK	17:59
Air Blank	0.000	17:59
Control Test	0.079	18:00
Air Blank	0.000	18:00
Subject Sample # 1	0.168	18:01
Breath Volume	3.531	18:01
Air Blank	0.000	18:01
Air Blank	0.000	18:03
Subject Sample # 2	0.160	18:04
Breath Volume	2.910	18:04
Air Blank	0.000	18:04
Control Test	0.079	18:05
Air Blank	0.000	18:05
Diagnostics Check	OK	18:05

Data Download Date/Time: 06/08/2022 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 06/04/2022 00:38:33

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.078	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.138	00:33
Breath Volume	1.402	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.144	00:37
Breath Volume	3.972	00:37
Air Blank	0.000	00:37
Control Test	0.078	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 06/08/2022 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: JANSON

DATE/TIME OF TEST: 06/04/2022 02:09:34

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.104	02:05
Breath Volume	3.054	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.103	02:08
Breath Volume	2.878	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 06/08/2022 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BAKER,LOGAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 06/08/2022 23:36:18

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.081	23:29	
Air Blank	0.000	23:30	
Subject Sample	# 1 0.096	23:31	
Breath Volume	1.449	23:31	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample	# 2	23:34	Interferent Detect
Breath Volume	1.226	23:34	
Air Blank	0.000	23:35	
Control Test	0.083	23:35	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 06/09/2022 00:35:30

Diagnostics Check	OK	00:34	
Air Blank		00:35	Ambient Fail
Air Blank	0.000	00:35	

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 06/09/2022 00:37:53

Diagnostics Check	OK	00:36	
Air Blank		00:37	Ambient Fail
Air Blank	0.000	00:37	

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HAMMOND

DATE/TIME OF TEST: 06/14/2022 22:47:08

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.314	22:42
Breath Volume	4.433	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.328	22:45
Breath Volume	3.937	22:45
Air Blank	0.000	22:46
Control Test	0.078	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MARCANTONIO

DATE/TIME OF TEST: 06/16/2022 00:34:41

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.078	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.000	00:30
Breath Volume	3.875	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:33
Breath Volume	3.710	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 06/16/2022 22:48:52

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.000	22:44
Breath Volume	3.113	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.000	22:47
Breath Volume	2.992	22:47
Air Blank	0.000	22:47
Control Test	0.078	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,A
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MILM

DATE/TIME OF TEST: 06/25/2022 02:24:07

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.172	02:16
Breath Volume	1.816	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.173	02:22
Breath Volume	1.367	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MARTIN,STERLING
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: OPRTA

DATE/TIME OF TEST: 06/03/2022 00:12:32

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.081	00:09	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.167	00:11	
Air Blank	0.000	00:11	
Control Test	0.082	00:11	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 06/30/2022 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,R,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MORALES,B
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 06/04/2022 22:53:19

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.081	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.142	22:47
Breath Volume	1.765	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.129	22:51
Breath Volume	2.222	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 06/30/2022 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSSI,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ROSSI,ASHLEY
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: ALLBRITTON

DATE/TIME OF TEST: 06/10/2022 23:20:37

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.118	23:14
Breath Volume	1.117	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.126	23:18
Breath Volume	1.160	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 06/30/2022 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSSI,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ROSSI,ASHLEY
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: CHAPPELL

DATE/TIME OF TEST: 06/12/2022 00:10:54

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.014	00:06
Breath Volume	1.593	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.015	00:09
Breath Volume	3.203	00:09
Air Blank	0.000	00:09
Control Test	0.079	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 06/30/2022 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BROWN,LAVAUGHN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 06/25/2022 20:59:43

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:51	
Control Test	0.081	20:52	
Air Blank	0.000	20:52	
Subject Sample	# 1 0.215	20:54	
Breath Volume	1.312	20:54	
Air Blank	0.000	20:55	
Air Blank	0.000	20:56	
Subject Sample	# 2 0.200	20:58	
Breath Volume	1.921	20:58	
Air Blank	0.000	20:58	
Control Test	0.058	20:59	Control Outside Tolerance
Air Blank	0.000	20:59	

Data Download Date/Time: 06/30/2022 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BROWN,LAVAUGHN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 06/25/2022 21:12:27

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.079	21:03
Air Blank	0.000	21:04
Subject Sample # 1	0.233	21:05
Breath Volume	1.558	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.231	21:10
Breath Volume	2.097	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 06/30/2022 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ALTIDOR,JUDE
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: SIRMANS

DATE/TIME OF TEST: 06/26/2022 05:37:57

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.000	05:33
Breath Volume	2.996	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.000	05:36
Breath Volume	1.457	05:36
Air Blank	0.000	05:37
Control Test	0.080	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 06/30/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CORBITT,JAMEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: VERGARA AGUILAR

DATE/TIME OF TEST: 06/27/2022 05:37:38

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.080	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.211	05:33
Breath Volume	3.417	05:33
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.208	05:36
Breath Volume	2.472	05:36
Air Blank	0.000	05:36
Control Test	0.079	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 06/30/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WALTER,CRAIG
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 06/11/2022 04:08:41

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.081	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.207	04:04
Breath Volume	1.667	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.207	04:07
Breath Volume	2.335	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 07/06/2022 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: CASSEUS,JEAN
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: OSBORN

DATE/TIME OF TEST: 06/21/2022 21:47:10

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.081	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.266	21:42
Breath Volume	1.531	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.265	21:45
Breath Volume	1.636	21:45
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 07/06/2022 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: SALMON,MATTHEW
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/24/2022 22:12:13

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.081	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.164	22:07
Breath Volume	3.035	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.170	22:10
Breath Volume	2.980	22:10
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 07/06/2022 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: EVERING

DATE/TIME OF TEST: 06/12/2022 02:25:57

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.164	02:21
Breath Volume	1.691	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.179	02:24
Breath Volume	1.902	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 06/22/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: SAINT LOUIS

DATE/TIME OF TEST: 06/12/2022 05:12:50

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.079	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.237	05:05
Breath Volume	2.113	05:05
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.205	05:08
Breath Volume	1.750	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 3	0.222	05:11
Breath Volume	1.957	05:11
Air Blank	0.000	05:11
Control Test	0.079	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 06/22/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 05/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/18/2022 06:37:37

Diagnostics Check	OK	06:29
Air Blank	0.000	06:30
Control Test	0.080	06:30
Air Blank	0.000	06:31
Subject Sample # 1	0.178	06:32
Breath Volume	1.632	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:35
Subject Sample # 2	0.178	06:36
Breath Volume	1.320	06:36
Air Blank	0.000	06:36
Control Test	0.079	06:37
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

Data Download Date/Time: 06/22/2022 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/24/2022 22:02:48

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.209	21:57
Breath Volume	2.441	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.212	22:01
Breath Volume	1.769	22:01
Air Blank	0.000	22:01
Control Test	0.078	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 07/06/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 06/25/2022 03:25:42

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.148	03:20
Breath Volume	2.250	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.145	03:24
Breath Volume	3.000	03:24
Air Blank	0.000	03:24
Control Test	0.078	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 07/06/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

INSTRUMENT REGISTERED TO: LAKE HELEN P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSHWORTH,JOSEPH,D
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BERNAL,ANTHONY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: COCKRELL

DATE/TIME OF TEST: 06/22/2022 19:21:47

Diagnostics Check	OK	19:14
Air Blank	0.000	19:15
Control Test	0.080	19:15
Air Blank	0.000	19:16
Subject Sample # 1	0.209	19:16
Breath Volume	1.562	19:16
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample # 2	0.208	19:20
Breath Volume	2.027	19:20
Air Blank	0.000	19:20
Control Test	0.078	19:21
Air Blank	0.000	19:21
Diagnostics Check	OK	19:21

Data Download Date/Time: 07/11/2022 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGRAM,TRISTAN,N
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: INGRAM,TRISTAN
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 06/03/2022 03:17:08

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.229	03:11
Breath Volume	1.195	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.234	03:15
Breath Volume	2.027	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 06/27/2022 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: GHICA,VLAD
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: VARGAS DIAZ

DATE/TIME OF TEST: 06/04/2022 19:35:44

Diagnostics Check	OK	19:28
Air Blank	0.000	19:28
Control Test	0.079	19:28
Air Blank	0.000	19:29
Subject Sample # 1	0.183	19:31
Breath Volume	1.628	19:31
Air Blank	0.000	19:32
Air Blank	0.000	19:33
Subject Sample # 2	0.171	19:34
Breath Volume	1.507	19:34
Air Blank	0.000	19:34
Control Test	0.077	19:35
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 06/27/2022 16:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,KYLE,T
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: MATHIAS

DATE/TIME OF TEST: 06/07/2022 21:09:57

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.081	21:04
Air Blank	0.000	21:04
Subject Sample # 1	0.166	21:05
Breath Volume	2.105	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.162	21:08
Breath Volume	1.929	21:08
Air Blank	0.000	21:09
Control Test	0.081	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 06/27/2022 16:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,B
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: NOLAN,NICOLETTE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 06/17/2022 21:38:54

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.081	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.250	21:34
Breath Volume	3.550	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:36
Subject Sample # 2	0.249	21:37
Breath Volume	2.332	21:37
Air Blank	0.000	21:37
Control Test	0.078	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 06/27/2022 16:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/09/2022 23:58:03

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.114	23:53
Breath Volume	2.667	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.111	23:56
Breath Volume	2.050	23:56
Air Blank	0.000	23:57
Control Test	0.083	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:58

Data Download Date/Time: 07/19/2022 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAXEY,JORDAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MAXEY,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 06/11/2022 08:17:05

Diagnostics Check	OK	08:10
Air Blank	0.000	08:11
Control Test	0.080	08:11
Air Blank	0.000	08:12
Subject Sample # 1	0.000	08:12
Breath Volume	1.332	08:12
Air Blank	0.000	08:13
Air Blank	0.000	08:15
Subject Sample # 2	0.000	08:15
Breath Volume	1.156	08:15
Air Blank	0.000	08:16
Control Test	0.080	08:16
Air Blank	0.000	08:16
Diagnostics Check	OK	08:17

Data Download Date/Time: 07/19/2022 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MARCHIONDO,GUIDO
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: ANTONELLI

DATE/TIME OF TEST: 06/17/2022 00:53:55

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.079	00:52	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:53	RFI Detect
Breath Volume	0.000	00:53	
Air Blank		00:53	RFI Detect
Air Blank		00:53	RFI Detect

Data Download Date/Time: 07/19/2022 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD, N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 06/25/2022 04:11:30

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.081	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1 0.124	04:07	
Breath Volume	1.429	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 2 0.124	04:10	
Breath Volume	1.687	04:10	
Air Blank		04:11	RFI Detect
Air Blank		04:11	RFI Detect

Data Download Date/Time: 07/19/2022 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SODERHOLM,CHELSEA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: HARE

DATE/TIME OF TEST: 06/26/2022 19:34:22

Diagnostics Check	OK	19:27
Air Blank	0.000	19:28
Control Test	0.081	19:28
Air Blank	0.000	19:29
Subject Sample # 1	0.260	19:29
Breath Volume	1.269	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.256	19:32
Breath Volume	1.164	19:32
Air Blank	0.000	19:33
Control Test	0.079	19:33
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 07/19/2022 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: LOWERY,JOHN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: DEEL

DATE/TIME OF TEST: 06/02/2022 16:36:30

Diagnostics Check	OK	16:33	
Air Blank	0.000	16:34	
Control Test	0.080	16:34	
Air Blank	0.000	16:34	
Subject Sample	# 1	16:35	Subject Test Refused
Breath Volume	0.000	16:35	
Air Blank	0.000	16:35	
Control Test	0.079	16:35	
Air Blank	0.000	16:36	
Diagnostics Check	OK	16:36	

Data Download Date/Time: 06/22/2022 19:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CALEB,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WOODIN,STEPHEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: DOUCETTE

DATE/TIME OF TEST: 06/09/2022 23:08:23

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.127	23:04
Breath Volume	2.250	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.122	23:06
Breath Volume	2.148	23:06
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 06/22/2022 19:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: YETMAN

DATE/TIME OF TEST: 06/10/2022 19:10:01

Diagnostics Check	OK	19:03
Air Blank	0.000	19:03
Control Test	0.079	19:04
Air Blank	0.000	19:04
Subject Sample # 1	0.160	19:05
Breath Volume	1.582	19:05
Air Blank	0.000	19:06
Air Blank	0.000	19:07
Subject Sample # 2	0.164	19:08
Breath Volume	2.054	19:08
Air Blank	0.000	19:09
Control Test	0.079	19:09
Air Blank	0.000	19:09
Diagnostics Check	OK	19:10

Data Download Date/Time: 06/22/2022 19:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/16/2022 23:31:26

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.192	23:27
Breath Volume	2.703	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.202	23:29
Breath Volume	2.296	23:29
Air Blank	0.000	23:30
Control Test	0.078	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 06/22/2022 19:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBE,BRIAN,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28420080A1
DRY GAS EXPIRY DATE: 12/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/03/2022 00:59:24

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.082	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.237	00:54
Breath Volume	1.996	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.247	00:57
Breath Volume	2.218	00:57
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 06/25/2022 20:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JOSE,L
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: BRYANT,GAIGE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28420080A1
DRY GAS EXPIRY DATE: 12/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 06/06/2022 15:48:50

Diagnostics Check	OK	15:40
Air Blank	0.000	15:40
Control Test	0.082	15:41
Air Blank	0.000	15:41
Subject Sample # 1	0.077	15:44
Breath Volume	2.593	15:44
Air Blank	0.000	15:44
Air Blank	0.000	15:46
Subject Sample # 2	0.083	15:47
Breath Volume	2.550	15:47
Air Blank	0.000	15:47
Control Test	0.082	15:48
Air Blank	0.000	15:48
Diagnostics Check	OK	15:48

Data Download Date/Time: 06/25/2022 20:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: RADCLIFF,AARON
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 05/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: WAGLER

DATE/TIME OF TEST: 06/02/2022 22:01:21

Diagnostics Check	OK	21:49	
Air Blank	0.000	21:50	
Control Test	0.082	21:50	
Air Blank	0.000	21:50	
Subject Sample	# 1 0.181	21:54	Volume Not Met
Breath Volume	0.121	21:54	
Air Blank	0.000	21:54	
Air Blank	0.000	21:56	
Subject Sample	# 2 0.179	21:59	Volume Not Met
Breath Volume	0.035	21:59	
Air Blank	0.000	22:00	
Control Test	0.079	22:00	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 06/27/2022 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: BATCH,J
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 05/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: FAIRCHILD

DATE/TIME OF TEST: 06/04/2022 03:11:21

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.081	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1 0.229	03:04	Volume Not Met
Breath Volume	0.050	03:04	
Air Blank	0.000	03:04	
Air Blank	0.000	03:06	
Subject Sample	# 2 0.216	03:09	Volume Not Met
Breath Volume	0.304	03:09	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 06/27/2022 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: WESTON,A
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 05/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: LINARES

DATE/TIME OF TEST: 06/05/2022 03:59:04

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.185	03:54
Breath Volume	2.910	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.189	03:57
Breath Volume	3.097	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:59

Data Download Date/Time: 06/27/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: MORIARTY,RORY
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 05/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: HAMMONDS

DATE/TIME OF TEST: 06/10/2022 15:44:47

Diagnostics Check	OK	15:35
Air Blank	0.000	15:35
Control Test	0.080	15:36
Air Blank	0.000	15:36
Subject Sample # 1	0.301	15:37
Breath Volume	1.605	15:37
Air Blank	0.000	15:37
Air Blank	0.000	15:39
Subject Sample # 2	0.322	15:39
Breath Volume	1.460	15:39
Air Blank	0.000	15:40
Air Blank	0.000	15:42
Subject Sample # 3	0.324	15:43
Breath Volume	1.632	15:43
Air Blank	0.000	15:43
Control Test	0.080	15:44
Air Blank	0.000	15:44
Diagnostics Check	OK	15:44

Data Download Date/Time: 06/27/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: MCCARLEY,MATTHEW
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 05/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: KREINER

DATE/TIME OF TEST: 06/17/2022 05:16:22

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.081	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.231	05:11
Breath Volume	3.578	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.229	05:14
Breath Volume	3.765	05:14
Air Blank	0.000	05:15
Control Test	0.079	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 06/27/2022 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: MCCARLEY,MATTHEW
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 05/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: FAVREAU

DATE/TIME OF TEST: 06/25/2022 03:18:23

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.227	03:13
Breath Volume	2.449	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.230	03:16
Breath Volume	2.726	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 06/27/2022 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: FARRAGHER, THOMAS
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SMITHSON

DATE/TIME OF TEST: 06/28/2022 04:08:49

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.081	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.152	04:04
Breath Volume	2.097	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.155	04:07
Breath Volume	1.691	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 07/26/2022 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REASONER,DEBBIE,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BERRIDGE,KATIE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: EPPRIGHT

DATE/TIME OF TEST: 06/23/2022 23:49:33

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.176	23:44
Breath Volume	2.785	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.173	23:48
Breath Volume	2.679	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 06/29/2022 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARPER,JIMMIE,L

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: MORALES,K

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 21:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: CULLISON

DATE/TIME OF TEST: 06/24/2022 21:35:31

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:33	
Control Test	0.077	21:33	
Air Blank	0.000	21:33	
Subject Sample	# 1	21:34	Subject Test Refused
Breath Volume	0.000	21:34	
Air Blank	0.000	21:34	
Control Test	0.077	21:34	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

Data Download Date/Time: 06/29/2022 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARPER,J,L

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: CAMACHO,N

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 20:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: EKERN

DATE/TIME OF TEST: 06/26/2022 20:46:37

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.080	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.094	20:42
Breath Volume	1.445	20:42
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.087	20:45
Breath Volume	1.757	20:45
Air Blank	0.000	20:45
Control Test	0.079	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 06/29/2022 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,J,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CRASKE

DATE/TIME OF TEST: 06/08/2022 01:08:45

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:56	
Control Test	0.081	00:57	
Air Blank	0.000	00:57	
Subject Sample	# 1 0.186	00:58	
Breath Volume	1.335	00:58	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample	# 2	01:03	No Sample Provided
Breath Volume	0.000	01:03	
Air Blank	0.000	01:04	
Air Blank	0.000	01:06	
Subject Sample	# 3	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 06/20/2022 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,R,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: PENA,R
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CANCETTY

DATE/TIME OF TEST: 06/11/2022 03:07:35

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.127	03:02
Breath Volume	1.812	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.128	03:06
Breath Volume	1.957	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 06/20/2022 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,R,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: PENA,R
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: VELAZQUEZ PORTILLO

DATE/TIME OF TEST: 06/12/2022 16:11:30

Diagnostics Check	OK	16:04
Air Blank	0.000	16:04
Control Test	0.080	16:05
Air Blank	0.000	16:05
Subject Sample # 1	0.179	16:07
Breath Volume	3.425	16:07
Air Blank	0.000	16:07
Air Blank	0.000	16:09
Subject Sample # 2	0.183	16:09
Breath Volume	3.613	16:09
Air Blank	0.000	16:10
Control Test	0.080	16:10
Air Blank	0.000	16:11
Diagnostics Check	OK	16:11

Data Download Date/Time: 06/20/2022 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASIMIRO,B,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ROLLINS,S
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: DELABERRERA

DATE/TIME OF TEST: 06/12/2022 23:02:24

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.258	22:57
Breath Volume	2.777	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.266	23:00
Breath Volume	2.644	23:00
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 06/20/2022 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,J,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 05/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: SAVINO

DATE/TIME OF TEST: 06/15/2022 23:17:57

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:15	
Control Test	0.080	23:15	
Air Blank	0.000	23:16	
Subject Sample	# 1	23:16	Subject Test Refused
Breath Volume	0.000	23:16	
Air Blank	0.000	23:17	
Control Test	0.080	23:17	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 06/20/2022 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERIC,JEAN,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FREDERIC,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: MILTON

DATE/TIME OF TEST: 06/02/2022 06:46:49

Diagnostics Check	OK	06:40
Air Blank	0.000	06:40
Control Test	0.081	06:40
Air Blank	0.000	06:41
Subject Sample # 1	0.054	06:42
Breath Volume	2.554	06:42
Air Blank	0.000	06:42
Air Blank	0.000	06:44
Subject Sample # 2	0.056	06:45
Breath Volume	2.570	06:45
Air Blank	0.000	06:45
Control Test	0.081	06:46
Air Blank	0.000	06:46
Diagnostics Check	OK	06:46

Data Download Date/Time: 06/07/2022 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: SHUTT

DATE/TIME OF TEST: 06/02/2022 23:02:09

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:47	
Control Test	0.082	22:47	
Air Blank	0.000	22:48	
Subject Sample	# 1 0.205	22:49	
Breath Volume	1.109	22:49	
Air Blank	0.000	22:49	
Air Blank	0.000	22:51	
Subject Sample	# 2 0.152	22:54	Volume Not Met
Breath Volume	0.550	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:57	
Subject Sample	# 3 0.189	23:00	Volume Not Met
Breath Volume	0.015	23:00	
Air Blank	0.000	23:01	
Control Test	0.080	23:01	
Air Blank	0.000	23:02	
Diagnostics Check	OK	23:02	

Data Download Date/Time: 06/07/2022 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: SHUTT

DATE/TIME OF TEST: 06/02/2022 23:17:18

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.193	23:11
Breath Volume	1.343	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.193	23:15
Breath Volume	1.289	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 06/07/2022 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: COLGAN,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 06/04/2022 03:56:03

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.081	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.149	03:51
Breath Volume	1.218	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.148	03:54
Breath Volume	2.121	03:54
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:56

Data Download Date/Time: 06/07/2022 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: AGUILAR-DOMINGUEL

DATE/TIME OF TEST: 06/06/2022 03:56:15

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.081	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:50	No Sample Provided
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Air Blank	0.000	03:52	
Subject Sample	# 2	03:54	Subject Test Refused
Breath Volume	0.000	03:54	
Air Blank	0.000	03:55	
Control Test	0.081	03:55	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 06/07/2022 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,JAMES,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: SANDERS,JAMES
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 06/06/2022 16:23:29

Diagnostics Check	OK	16:20	
Air Blank	0.000	16:20	
Control Test	0.082	16:20	
Air Blank	0.000	16:21	
Subject Sample	# 1	16:22	Subject Test Refused
Breath Volume	0.000	16:22	
Air Blank	0.000	16:22	
Control Test	0.081	16:22	
Air Blank	0.000	16:23	
Diagnostics Check	OK	16:23	

Data Download Date/Time: 06/07/2022 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORES,JADISE,W
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 06/09/2022 11:14:19

Diagnostics Check	OK	11:11	
Air Blank	0.000	11:11	
Control Test	0.082	11:11	
Air Blank	0.000	11:12	
Subject Sample	# 1	11:12	Subject Test Refused
Breath Volume	0.000	11:12	
Air Blank	0.000	11:13	
Control Test	0.082	11:13	
Air Blank	0.000	11:14	
Diagnostics Check	OK	11:14	

Data Download Date/Time: 07/19/2022 23:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 06/09/2022 11:23:52

Diagnostics Check	OK	11:16
Air Blank	0.000	11:17
Control Test	0.081	11:17
Air Blank	0.000	11:18
Subject Sample # 1	0.000	11:18
Breath Volume	1.343	11:18
Air Blank	0.000	11:19
Air Blank	0.000	11:21
Subject Sample # 2	0.000	11:22
Breath Volume	1.687	11:22
Air Blank	0.000	11:22
Control Test	0.082	11:23
Air Blank	0.000	11:23
Diagnostics Check	OK	11:23

Data Download Date/Time: 07/19/2022 23:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 06/10/2022 03:05:19

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.082	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.130	03:00
Breath Volume	2.519	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.145	03:03
Breath Volume	1.496	03:03
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/19/2022 23:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: PEREZ GARCIA

DATE/TIME OF TEST: 06/11/2022 00:43:48

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.081	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.179	00:38
Breath Volume	2.390	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.180	00:42
Breath Volume	2.890	00:42
Air Blank	0.000	00:42
Control Test	0.081	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 07/19/2022 23:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRIQUEZ ZUNIGA,MARCIO,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 06/12/2022 03:44:59

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.217	03:40
Breath Volume	2.179	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:42
Subject Sample # 2	0.217	03:43
Breath Volume	1.507	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 07/19/2022 23:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: CURBELO

DATE/TIME OF TEST: 06/15/2022 02:48:34

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.234	02:43
Breath Volume	2.648	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.235	02:47
Breath Volume	3.359	02:47
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 07/19/2022 23:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: CERNUTO,MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: TRANBERG

DATE/TIME OF TEST: 06/17/2022 23:07:58

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.202	23:03
Breath Volume	1.730	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:05
Subject Sample # 2	0.219	23:06
Breath Volume	3.601	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 07/19/2022 23:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: GARCIA-CHAVIRA

DATE/TIME OF TEST: 06/20/2022 01:16:48

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.081	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.193	01:11
Breath Volume	1.226	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:13
Subject Sample # 2	0.201	01:15
Breath Volume	2.082	01:15
Air Blank	0.000	01:15
Control Test	0.081	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 07/19/2022 23:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/22/2022 17:40:19

Diagnostics Check	OK	17:33
Air Blank	0.000	17:33
Control Test	0.082	17:34
Air Blank	0.000	17:34
Subject Sample # 1	0.046	17:35
Breath Volume	1.351	17:35
Air Blank	0.000	17:36
Air Blank	0.000	17:38
Subject Sample # 2	0.045	17:38
Breath Volume	1.101	17:38
Air Blank	0.000	17:39
Control Test	0.081	17:39
Air Blank	0.000	17:40
Diagnostics Check	OK	17:40

Data Download Date/Time: 07/19/2022 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: DEWITT

DATE/TIME OF TEST: 06/26/2022 00:17:22

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.082	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.252	00:11
Breath Volume	1.246	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.246	00:15
Breath Volume	1.902	00:15
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 07/19/2022 23:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PEARLMAN,RYAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: DIAZDOMINGO

DATE/TIME OF TEST: 06/26/2022 01:27:27

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.082	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.146	01:22
Breath Volume	4.246	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.147	01:25
Breath Volume	2.093	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 07/19/2022 23:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: PEDRO

DATE/TIME OF TEST: 06/27/2022 20:32:07

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.082	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.202	20:27
Breath Volume	3.195	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.206	20:30
Breath Volume	3.035	20:30
Air Blank	0.000	20:31
Control Test	0.081	20:31
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 07/19/2022 23:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: BRABO LOPEZ

DATE/TIME OF TEST: 06/27/2022 23:12:11

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.081	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.092	23:07
Breath Volume	1.781	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:09
Subject Sample # 2	0.087	23:10
Breath Volume	1.464	23:10
Air Blank	0.000	23:11
Control Test	0.081	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 07/19/2022 23:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: \,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: TISON,BRIAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ORR

DATE/TIME OF TEST: 06/28/2022 01:41:27

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.082	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.169	01:32
Breath Volume	1.324	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.191	01:36
Breath Volume	2.230	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 3	0.191	01:39
Breath Volume	1.945	01:39
Air Blank	0.000	01:40
Control Test	0.078	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 07/19/2022 23:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:08

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/28/2022 09:37:41

Diagnostics Check	OK	09:31
Air Blank	0.000	09:32
Control Test	0.082	09:32
Air Blank	0.000	09:32
Subject Sample # 1	0.118	09:33
Breath Volume	2.574	09:33
Air Blank	0.000	09:33
Air Blank	0.000	09:35
Subject Sample # 2	0.114	09:36
Breath Volume	2.687	09:36
Air Blank	0.000	09:36
Control Test	0.081	09:37
Air Blank	0.000	09:37
Diagnostics Check	OK	09:37

Data Download Date/Time: 07/19/2022 23:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHSON,KAED,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MATHSON

DATE/TIME OF TEST: 06/04/2022 01:20:54

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.078	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.078	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 06/27/2022 16:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHSON,KAED,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MATHSON

DATE/TIME OF TEST: 06/04/2022 01:32:36

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.184	01:28
Breath Volume	2.183	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.183	01:31
Breath Volume	1.886	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 06/27/2022 16:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHNATHAN,F
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: YOUNG,JOHNATHAN
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 06:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: ENNIS

DATE/TIME OF TEST: 06/05/2022 07:07:37

Diagnostics Check	OK	06:58
Air Blank	0.000	06:58
Control Test	0.081	06:58
Air Blank	0.000	06:59
Subject Sample # 1	0.175	07:01
Breath Volume	1.765	07:01
Air Blank	0.000	07:01
Air Blank	0.000	07:03
Subject Sample # 2	0.159	07:06
Breath Volume	1.140	07:06
Air Blank	0.000	07:06
Control Test	0.082	07:06
Air Blank	0.000	07:07
Diagnostics Check	OK	07:07

Data Download Date/Time: 06/27/2022 16:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,EVERTON,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS,BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/09/2022 22:52:31

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.000	22:48
Breath Volume	1.218	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.000	22:51
Breath Volume	1.101	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 06/27/2022 16:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COUPE,JESSICA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 06/13/2022 06:16:32

Diagnostics Check	OK	06:10
Air Blank	0.000	06:11
Control Test	0.081	06:11
Air Blank	0.000	06:12
Subject Sample # 1	0.000	06:12
Breath Volume	2.800	06:12
Air Blank	0.000	06:12
Air Blank	0.000	06:14
Subject Sample # 2	0.000	06:15
Breath Volume	3.609	06:15
Air Blank	0.000	06:15
Control Test	0.081	06:15
Air Blank	0.000	06:16
Diagnostics Check	OK	06:16

Data Download Date/Time: 06/27/2022 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 06/14/2022 19:23:21

Diagnostics Check	OK	19:17
Air Blank	0.000	19:17
Control Test	0.080	19:17
Air Blank	0.000	19:18
Subject Sample # 1	0.214	19:18
Breath Volume	1.550	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.221	19:21
Breath Volume	1.199	19:21
Air Blank	0.000	19:22
Control Test	0.080	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 06/27/2022 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 06/14/2022 23:12:25

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.218	23:07
Breath Volume	2.039	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.228	23:10
Breath Volume	1.757	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 06/27/2022 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYE, DOUGLAS, W
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: SHANKLE, CHASE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: RAYE

DATE/TIME OF TEST: 06/15/2022 03:17:51

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.081	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.090	03:13
Breath Volume	1.835	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.085	03:16
Breath Volume	3.078	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 06/27/2022 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 06/17/2022 04:54:28

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.081	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.057	04:49
Breath Volume	1.250	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.060	04:52
Breath Volume	1.765	04:52
Air Blank	0.000	04:53
Control Test	0.082	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 06/27/2022 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 06/17/2022 22:07:15

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.081	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.076	22:02
Breath Volume	1.140	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.072	22:05
Breath Volume	1.640	22:05
Air Blank	0.000	22:06
Control Test	0.081	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 06/27/2022 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/18/2022 00:39:22

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.206	00:30
Breath Volume	1.476	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:32
Subject Sample # 2	0.183	00:33
Breath Volume	1.171	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 3	0.176	00:37
Breath Volume	1.906	00:37
Air Blank	0.000	00:38
Control Test	0.081	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 06/27/2022 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 06/18/2022 03:12:36

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.175	03:08
Breath Volume	1.750	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.182	03:11
Breath Volume	1.695	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 06/27/2022 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: MCCOY IV

DATE/TIME OF TEST: 06/23/2022 06:28:07

Diagnostics Check	OK	06:21
Air Blank	0.000	06:22
Control Test	0.081	06:22
Air Blank	0.000	06:23
Subject Sample # 1	0.083	06:23
Breath Volume	1.902	06:23
Air Blank	0.000	06:24
Air Blank	0.000	06:26
Subject Sample # 2	0.093	06:26
Breath Volume	2.585	06:26
Air Blank	0.000	06:27
Control Test	0.081	06:27
Air Blank	0.000	06:28
Diagnostics Check	OK	06:28

Data Download Date/Time: 06/27/2022 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/28/2022 06:13:10

Diagnostics Check	OK	06:05
Air Blank	0.000	06:06
Control Test	0.082	06:06
Air Blank	0.000	06:07
Subject Sample # 1	0.108	06:08
Breath Volume	2.472	06:08
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 2	0.109	06:11
Breath Volume	1.320	06:11
Air Blank	0.000	06:12
Control Test	0.082	06:12
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

Data Download Date/Time: 07/22/2022 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: SMINK

DATE/TIME OF TEST: 06/28/2022 23:59:56

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.081	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.021	23:55
Breath Volume	3.777	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.021	23:58
Breath Volume	3.386	23:58
Air Blank	0.000	23:59
Control Test	0.082	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 09:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:32

SUBJECT'S LAST NAME: BECKFORD

DATE/TIME OF TEST: 06/29/2022 09:56:07

Diagnostics Check	OK	09:53	
Air Blank	0.000	09:53	
Control Test	0.082	09:53	
Air Blank	0.000	09:54	
Subject Sample	# 1	09:54	Subject Test Refused
Breath Volume	0.000	09:54	
Air Blank	0.000	09:55	
Control Test	0.082	09:55	
Air Blank	0.000	09:56	
Diagnostics Check	OK	09:56	

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HALE,CALEB
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 06/15/2022 05:30:11

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:21	
Control Test	0.079	05:21	
Air Blank	0.000	05:22	
Subject Sample	# 1	05:25	No Sample Provided
Breath Volume	0.000	05:25	
Air Blank	0.000	05:25	
Air Blank	0.000	05:27	
Subject Sample	# 2	05:28	Subject Test Refused
Breath Volume	0.000	05:28	
Air Blank	0.000	05:29	
Control Test	0.077	05:29	
Air Blank	0.000	05:30	
Diagnostics Check	OK	05:30	

Data Download Date/Time: 07/25/2022 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HALE,CALEB
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 06/15/2022 05:40:28

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:34	
Control Test	0.078	05:34	
Air Blank	0.000	05:34	
Subject Sample	# 1 0.185	05:35	
Breath Volume	1.335	05:35	
Air Blank	0.000	05:36	
Air Blank	0.000	05:38	
Subject Sample	# 2	05:39	Subject Test Refused
Breath Volume	0.175	05:39	
Air Blank	0.000	05:39	
Control Test	0.079	05:39	
Air Blank	0.000	05:40	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 07/25/2022 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PAUGA,JOHN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: A
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: TOVAR SERNA

DATE/TIME OF TEST: 06/18/2022 01:41:50

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.078	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.112	01:37
Breath Volume	1.433	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.118	01:40
Breath Volume	2.851	01:40
Air Blank	0.000	01:40
Control Test	0.078	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 07/25/2022 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HALE,CALEB
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DOZIER

DATE/TIME OF TEST: 06/25/2022 01:23:48

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.198	01:18
Breath Volume	1.554	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.197	01:22
Breath Volume	3.156	01:22
Air Blank	0.000	01:22
Control Test	0.078	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 07/25/2022 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: STALNAKER,TEDDY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:02

SUBJECT'S LAST NAME: BRYSON

DATE/TIME OF TEST: 06/25/2022 09:29:13

Diagnostics Check	OK	09:22
Air Blank	0.000	09:23
Control Test	0.079	09:23
Air Blank	0.000	09:23
Subject Sample # 1	0.164	09:24
Breath Volume	2.011	09:24
Air Blank	0.000	09:24
Air Blank	0.000	09:26
Subject Sample # 2	0.147	09:27
Breath Volume	1.726	09:27
Air Blank	0.000	09:28
Control Test	0.079	09:28
Air Blank	0.000	09:29
Diagnostics Check	OK	09:29

Data Download Date/Time: 07/25/2022 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKPATRICK,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: KIRKPATRICK,MATTHEW
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: NARY

DATE/TIME OF TEST: 06/02/2022 00:46:11

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.078	00:33	
Air Blank	0.000	00:33	
Subject Sample	# 1 0.246	00:36	Volume Not Met
Breath Volume	0.574	00:36	
Air Blank	0.000	00:37	
Air Blank	0.000	00:39	
Subject Sample	# 2 0.239	00:41	
Breath Volume	2.570	00:41	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 3 0.231	00:44	
Breath Volume	2.371	00:44	
Air Blank	0.000	00:45	
Control Test	0.078	00:45	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 06/29/2022 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COMERFORD,NATHAN,S
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: COMERFORD,NATHAN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 06/02/2022 01:04:03

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.078	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:56	No Sample Provided
Breath Volume	0.000	00:56	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample	# 2	0.178	00:59
Breath Volume	1.386	00:59	
Air Blank	0.000	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 3	0.198	01:02
Breath Volume	1.433	01:02	
Air Blank	0.000	01:03	
Control Test	0.079	01:03	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 06/29/2022 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COMERFORD,NATHAN,S
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: COMERFORD,NATHAN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 06/11/2022 20:00:27

Diagnostics Check	OK	19:54
Air Blank	0.000	19:54
Control Test	0.078	19:55
Air Blank	0.000	19:55
Subject Sample # 1	0.252	19:56
Breath Volume	1.117	19:56
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.241	19:58
Breath Volume	1.445	19:58
Air Blank	0.000	19:59
Control Test	0.078	19:59
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 06/29/2022 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 06/07/2022 03:15:18

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.081	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:10	No Sample Provided
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Air Blank	0.000	03:13	
Subject Sample	# 2	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:14	
Control Test	0.081	03:14	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 06/23/2022 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 06/20/2022 20:20:55

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:10	
Control Test	0.082	20:11	
Air Blank	0.000	20:11	
Subject Sample	# 1 0.253	20:14	Volume Not Met
Breath Volume	0.164	20:14	
Air Blank	0.000	20:15	
Air Blank	0.000	20:17	
Subject Sample	# 2	20:19	Subject Test Refused
Breath Volume	0.000	20:19	
Air Blank	0.000	20:20	
Control Test	0.079	20:20	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 06/23/2022 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: CHARLES,JAMESLOO
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: LIZARDO

DATE/TIME OF TEST: 06/25/2022 23:42:02

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.081	23:39	
Air Blank	0.000	23:39	
Subject Sample	# 1	23:40	Subject Test Refused
Breath Volume	0.000	23:40	
Air Blank	0.000	23:41	
Control Test	0.081	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 07/22/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: VERGARA

DATE/TIME OF TEST: 06/27/2022 01:40:55

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.081	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.141	01:36
Breath Volume	2.660	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.132	01:39
Breath Volume	2.179	01:39
Air Blank	0.000	01:40
Control Test	0.081	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 07/22/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: CARTER,BENJAMIN
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: MEEKS

DATE/TIME OF TEST: 06/04/2022 19:27:27

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.080	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.412	19:22
Breath Volume	3.011	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.401	19:25
Breath Volume	3.257	19:25
Air Blank	0.000	19:26
Control Test	0.078	19:26
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 06/22/2022 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: GARRIDO,STEPHANIE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SLOVER

DATE/TIME OF TEST: 06/08/2022 22:02:56

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.080	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.000	21:58
Breath Volume	1.281	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.000	22:01
Breath Volume	2.511	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 06/22/2022 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: KESSINGER,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/09/2022 02:09:29

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.080	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:08	
Control Test	0.079	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 06/22/2022 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,JAMES,E
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: MARTINEZ,EZEQUIEL
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: GOOD

DATE/TIME OF TEST: 06/10/2022 03:55:26

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:52	
Control Test	0.081	03:53	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:54	Subject Test Refused
Breath Volume	0.000	03:54	
Air Blank	0.000	03:54	
Control Test	0.081	03:54	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 06/22/2022 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINDLEY,CARLTONJANE,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PRANOUSKES,DENNIS
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: DURHAM

DATE/TIME OF TEST: 06/10/2022 15:05:47

Diagnostics Check	OK	14:57	
Air Blank	0.000	14:57	
Control Test	0.081	14:58	
Air Blank	0.000	14:58	
Subject Sample	# 1	15:01	No Sample Provided
Breath Volume	0.000	15:01	
Air Blank	0.000	15:02	
Air Blank	0.000	15:04	
Subject Sample	# 2	15:05	RFI Detect
Breath Volume	0.000	15:05	
Air Blank		15:05	RFI Detect
Air Blank	0.000	15:05	

Data Download Date/Time: 06/22/2022 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: BULTRON,CARLOS
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: YEPEZ

DATE/TIME OF TEST: 06/11/2022 00:18:24

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.081	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.199	00:14
Breath Volume	3.132	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.197	00:16
Breath Volume	3.789	00:16
Air Blank	0.000	00:17
Control Test	0.081	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 06/22/2022 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,R,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MARKLUND,K
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: NIEDZWIECKI

DATE/TIME OF TEST: 06/12/2022 01:47:20

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.081	01:46	
Air Blank		01:46	RFI Detect
Air Blank	0.000	01:47	

Data Download Date/Time: 06/22/2022 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,R,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MARKLUND,K
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: NIEDZWIECKI

DATE/TIME OF TEST: 06/12/2022 01:55:20

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.195	01:50
Breath Volume	3.171	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.197	01:53
Breath Volume	3.589	01:53
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 06/22/2022 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUEGER, ELIZABETH, A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: SANCHEZ, ABISAI
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: NOBLES

DATE/TIME OF TEST: 06/15/2022 19:13:27

Diagnostics Check	OK	18:58	
Air Blank	0.000	18:58	
Control Test	0.081	18:59	
Air Blank	0.000	18:59	
Subject Sample	# 1 0.011	19:00	
Breath Volume	1.226	19:00	
Air Blank	0.000	19:01	
Air Blank	0.000	19:03	
Subject Sample	# 2 0.008	19:06	Volume Not Met
Breath Volume	0.113	19:06	
Air Blank	0.000	19:06	
Air Blank	0.000	19:08	
Subject Sample	# 3 0.000	19:12	Volume Not Met
Breath Volume	0.347	19:12	
Air Blank	0.000	19:12	
Control Test	0.080	19:12	
Air Blank	0.000	19:13	
Diagnostics Check	OK	19:13	

Data Download Date/Time: 06/22/2022 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY, WAYNE, G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WILSON, TEONDRE
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/18/2022 20:06:25

Diagnostics Check	OK	19:59
Air Blank	0.000	19:59
Control Test	0.081	19:59
Air Blank	0.000	20:00
Subject Sample # 1	0.180	20:01
Breath Volume	1.960	20:01
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.173	20:04
Breath Volume	1.945	20:04
Air Blank	0.000	20:05
Control Test	0.081	20:05
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 06/22/2022 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,JAMES,E
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: HALL,JESS
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 06/19/2022 00:49:04

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.080	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.222	00:44
Breath Volume	2.699	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:46
Subject Sample # 2	0.224	00:47
Breath Volume	2.835	00:47
Air Blank	0.000	00:48
Control Test	0.081	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:49

Data Download Date/Time: 06/22/2022 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLDHAM,TAYLOR,D
BREATH TEST OPERATOR AGENCY: Mascotte PD
ARRESTING OFFICER: SOTO,ADRIAN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 06/19/2022 08:04:45

Diagnostics Check	OK	07:56
Air Blank	0.000	07:57
Control Test	0.081	07:57
Air Blank	0.000	07:58
Subject Sample # 1	0.141	08:00
Breath Volume	1.867	08:00
Air Blank	0.000	08:00
Air Blank	0.000	08:02
Subject Sample # 2	0.142	08:03
Breath Volume	2.347	08:03
Air Blank	0.000	08:03
Control Test	0.079	08:04
Air Blank	0.000	08:04
Diagnostics Check	OK	08:04

Data Download Date/Time: 06/22/2022 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASIMIRO,B,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: CASIMIRO,B
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/21/2022 00:25:05

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.081	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:24	
Control Test	0.081	00:24	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 06/22/2022 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASIMIRO,B,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: CASIMIRO,B
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/21/2022 00:40:40

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.081	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.202	00:31
Breath Volume	1.921	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.176	00:35
Breath Volume	1.671	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 3	0.175	00:39
Breath Volume	1.726	00:39
Air Blank	0.000	00:39
Control Test	0.078	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 06/22/2022 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUEGER, ELIZABETH, A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: GOEHRING, JONATHAN
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: GAGNE

DATE/TIME OF TEST: 06/24/2022 19:40:44

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.081	19:34
Air Blank	0.000	19:35
Subject Sample # 1	0.196	19:36
Breath Volume	3.722	19:36
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.196	19:39
Breath Volume	3.156	19:39
Air Blank	0.000	19:39
Control Test	0.078	19:40
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 07/24/2022 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, WILLIAM, S
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCKENZIE, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: WISOCKY

DATE/TIME OF TEST: 06/29/2022 00:15:05

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.081	00:04	
Air Blank	0.000	00:04	
Subject Sample	# 1 0.205	00:07	Volume Not Met
Breath Volume	0.164	00:07	
Air Blank	0.000	00:08	
Air Blank	0.000	00:10	
Subject Sample	# 2 0.199	00:13	Volume Not Met
Breath Volume	0.183	00:13	
Air Blank	0.000	00:14	
Control Test	0.080	00:14	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 07/24/2022 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/29/2022 20:58:43

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:47	
Control Test	0.081	20:48	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:51	No Sample Provided
Breath Volume	0.000	20:51	
Air Blank	0.000	20:52	
Air Blank	0.000	20:54	
Subject Sample	# 2	20:57	No Sample Provided
Breath Volume	0.000	20:57	
Air Blank	0.000	20:57	
Control Test	0.081	20:58	
Air Blank	0.000	20:58	
Diagnostics Check	OK	20:58	

Data Download Date/Time: 07/24/2022 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JACOB,M
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: RANSOM,JACK
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: PERFORs

DATE/TIME OF TEST: 06/03/2022 04:16:07

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.079	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1 0.144	04:07	Volume Not Met
Breath Volume	0.570	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 2 0.168	04:11	
Breath Volume	2.910	04:11	
Air Blank	0.000	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 3 0.174	04:14	
Breath Volume	2.843	04:14	
Air Blank	0.000	04:15	
Control Test	0.076	04:15	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 06/14/2022 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,BILL,R
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: WOOLEYHAN,HEATHER
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/12/2022 05:50:22

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.080	05:43
Air Blank	0.000	05:43
Subject Sample # 1	0.157	05:45
Breath Volume	1.859	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.158	05:48
Breath Volume	1.914	05:48
Air Blank	0.000	05:49
Control Test	0.078	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 06/14/2022 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILLUE,MOLLY,S
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BILLUE,MOLLY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/18/2022 03:08:23

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.169	03:03
Breath Volume	2.359	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.172	03:06
Breath Volume	2.105	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 07/25/2022 18:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,KYLE,T
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BUSKEY

DATE/TIME OF TEST: 06/07/2022 03:36:01

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.075	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.149	03:31
Breath Volume	2.960	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.147	03:34
Breath Volume	1.550	03:34
Air Blank	0.000	03:35
Control Test	0.075	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:36

Data Download Date/Time: 06/19/2022 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,KYLE,T
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: MATHIAS

DATE/TIME OF TEST: 06/07/2022 20:32:21

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:31	
Control Test	0.074	20:31	Control Outside Tolerance
Air Blank	0.000	20:32	

Data Download Date/Time: 06/19/2022 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,KYLE,T
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: MATHIAS

DATE/TIME OF TEST: 06/07/2022 20:40:47

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:34	
Control Test	0.075	20:35	
Air Blank	0.000	20:35	
Subject Sample	# 1 0.173	20:36	
Breath Volume	1.699	20:36	
Air Blank	0.000	20:36	
Air Blank	0.000	20:38	
Subject Sample	# 2 0.170	20:39	
Breath Volume	1.746	20:39	
Air Blank	0.000	20:39	
Control Test	0.074	20:40	Control Outside Tolerance
Air Blank	0.000	20:40	

Data Download Date/Time: 06/19/2022 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,KYLE,T
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: MATHIAS

DATE/TIME OF TEST: 06/07/2022 20:43:53

Diagnostics Check	OK	20:42	
Air Blank	0.000	20:42	
Control Test	0.074	20:43	Control Outside Tolerance
Air Blank	0.000	20:43	

Data Download Date/Time: 06/19/2022 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: BINZ,STEPHEN
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: SYS CHECK
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 06/08/2022 10:02:00

Diagnostics Check	OK	09:56	
Air Blank	0.000	09:56	
Control Test	0.075	09:56	
Air Blank	0.000	09:57	
Subject Sample # 1	0.000	09:57	
Breath Volume	4.144	09:57	
Air Blank	0.000	09:58	
Air Blank	0.000	10:00	
Subject Sample # 2	0.000	10:00	
Breath Volume	3.570	10:00	
Air Blank	0.000	10:01	
Control Test	0.074	10:01	Control Outside Tolerance
Air Blank	0.000	10:02	

Data Download Date/Time: 06/20/2022 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ, STEHEN, D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: BINZ, STEPHEN
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: SYS CHECK
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 06/08/2022 10:09:56

Diagnostics Check	OK	10:04
Air Blank	0.000	10:04
Control Test	0.075	10:04
Air Blank	0.000	10:05
Subject Sample # 1	0.000	10:05
Breath Volume	3.457	10:05
Air Blank	0.000	10:06
Air Blank	0.000	10:08
Subject Sample # 2	0.000	10:08
Breath Volume	2.941	10:08
Air Blank	0.000	10:09
Control Test	0.075	10:09
Air Blank	0.000	10:09
Diagnostics Check	OK	10:09

Data Download Date/Time: 06/20/2022 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,B
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: NOLAN,NICOLETTE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 06/17/2022 21:04:41

Diagnostics Check	OK	21:03	
Air Blank		21:03	Ambient Fail
Air Blank	0.000	21:04	

Data Download Date/Time: 06/20/2022 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,B
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: NOLAN,NICOLETTE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 06/17/2022 21:07:41

Diagnostics Check	OK	21:06	
Air Blank	0.000	21:06	
Control Test	0.069	21:07	Control Outside Tolerance
Air Blank	0.000	21:07	

Data Download Date/Time: 06/20/2022 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: BINZ,STEHEN
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 06/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: SYS CHECK
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 06/20/2022 08:48:31

Diagnostics Check	OK	08:42
Air Blank	0.000	08:43
Control Test	0.079	08:43
Air Blank	0.000	08:43
Subject Sample # 1	0.000	08:44
Breath Volume	2.136	08:44
Air Blank	0.000	08:44
Air Blank	0.000	08:46
Subject Sample # 2	0.000	08:47
Breath Volume	2.152	08:47
Air Blank	0.000	08:47
Control Test	0.079	08:47
Air Blank	0.000	08:48
Diagnostics Check	OK	08:48

Data Download Date/Time: 06/20/2022 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STOKER,RONALD
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 06/04/2022 16:45:50

Diagnostics Check	OK	16:38
Air Blank	0.000	16:39
Control Test	0.080	16:39
Air Blank	0.000	16:39
Subject Sample # 1	0.204	16:40
Breath Volume	1.222	16:40
Air Blank	0.000	16:41
Air Blank	0.000	16:43
Subject Sample # 2	0.205	16:44
Breath Volume	1.695	16:44
Air Blank	0.000	16:44
Control Test	0.081	16:45
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 07/05/2022 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOVANOVICH, THOMAS, E
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MEISLAHN, ERIC
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 06/09/2022 15:06:18

Diagnostics Check	OK	14:55
Air Blank	0.000	14:56
Control Test	0.081	14:56
Air Blank	0.000	14:56
Subject Sample # 1	0.300	14:58
Breath Volume	2.125	14:58
Air Blank	0.000	14:58
Air Blank	0.000	15:00
Subject Sample # 2	0.277	15:01
Breath Volume	1.515	15:01
Air Blank	0.000	15:02
Air Blank	0.000	15:03
Subject Sample # 3	0.297	15:04
Breath Volume	2.144	15:04
Air Blank	0.000	15:05
Control Test	0.080	15:05
Air Blank	0.000	15:06
Diagnostics Check	OK	15:06

Data Download Date/Time: 07/05/2022 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCBRIDE,CRYSTA,C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MCBRIDE,CRYSTA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: COCHRANE

DATE/TIME OF TEST: 06/09/2022 23:35:02

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.080	23:29	
Air Blank	0.000	23:30	
Subject Sample	# 1 0.142	23:30	
Breath Volume	1.816	23:30	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample	# 2 0.146	23:34	
Breath Volume	1.785	23:34	
Air Blank		23:34	RFI Detect
Air Blank	0.000	23:35	

Data Download Date/Time: 07/05/2022 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINES,CHRISTOPHER,G
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MACINTYRE,NICHOLAS
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:39

SUBJECT'S LAST NAME: GOLDFARB

DATE/TIME OF TEST: 06/10/2022 09:03:42

Diagnostics Check	OK	09:02	
Air Blank	0.000	09:02	
Control Test	0.080	09:02	
Air Blank		09:03	RFI Detect
Air Blank	0.000	09:03	

Data Download Date/Time: 07/05/2022 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINES,CHRISTOPHER,G
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MACINTYRE,NICHOLAS
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:39

SUBJECT'S LAST NAME: GOLDFARB

DATE/TIME OF TEST: 06/10/2022 09:12:22

Diagnostics Check	OK	09:05
Air Blank	0.000	09:06
Control Test	0.080	09:06
Air Blank	0.000	09:07
Subject Sample # 1	0.000	09:07
Breath Volume	3.167	09:07
Air Blank	0.000	09:08
Air Blank	0.000	09:10
Subject Sample # 2	0.000	09:11
Breath Volume	3.476	09:11
Air Blank	0.000	09:11
Control Test	0.080	09:11
Air Blank	0.000	09:12
Diagnostics Check	OK	09:12

Data Download Date/Time: 07/05/2022 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVILA,NICOLE,N
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MILLS,JACOB
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: MEINCKE

DATE/TIME OF TEST: 06/24/2022 22:53:48

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.138	22:49
Breath Volume	3.511	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:51
Subject Sample # 2	0.132	22:52
Breath Volume	1.578	22:52
Air Blank	0.000	22:52
Control Test	0.080	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 07/05/2022 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFFNEY,CASEY,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: ROSA,JOSEPH
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CRIGGER

DATE/TIME OF TEST: 06/21/2022 03:35:44

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.082	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.000	03:31
Breath Volume	1.953	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.000	03:34
Breath Volume	2.535	03:34
Air Blank	0.000	03:34
Control Test	0.082	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 07/29/2022 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFFNEY,CASEY,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: GAFFNEY,CASEY
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: LARUE

DATE/TIME OF TEST: 06/25/2022 20:55:38

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.082	20:50
Air Blank	0.000	20:50
Subject Sample # 1	0.000	20:51
Breath Volume	3.773	20:51
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.000	20:54
Breath Volume	2.855	20:54
Air Blank	0.000	20:54
Control Test	0.081	20:55
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 07/29/2022 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: ROBISON,CASEY

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: A

DRY GAS EXPIRY DATE: 12/29/2021

VIOLATION CODE: DUI

ARREST TIME: 22:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 06/10/2022 22:36:03

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:34	
Control Test	0.079	22:34	
Air Blank		22:35	Purge Fail
Air Blank		22:36	Purge Fail

Data Download Date/Time: 06/20/2022 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: ROBISON,CASEY

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: A

DRY GAS EXPIRY DATE: 12/29/2021

VIOLATION CODE: DUI

ARREST TIME: 22:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 06/10/2022 22:39:14

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:38	
Control Test	0.009	22:38	Interferent Detect
Air Blank	0.000	22:39	

Data Download Date/Time: 06/20/2022 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: RAMOS DA SILVA

DATE/TIME OF TEST: 06/16/2022 02:17:14

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.082	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.155	02:12
Breath Volume	1.347	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.164	02:15
Breath Volume	1.816	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 07/05/2022 14:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MONTGOMERY,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/18/2022 03:08:14

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.203	03:02
Breath Volume	1.730	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.201	03:06
Breath Volume	2.765	03:06
Air Blank	0.000	03:07
Control Test	0.082	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 07/05/2022 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOSA,EMMANUEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: AMBROCIO RAMIREZ

DATE/TIME OF TEST: 06/25/2022 01:19:59

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.081	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.194	01:15
Breath Volume	2.335	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:17
Subject Sample # 2	0.198	01:18
Breath Volume	1.437	01:18
Air Blank	0.000	01:19
Control Test	0.078	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 07/05/2022 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUTCHINSON,JUSTIN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/25/2022 01:46:58

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.131	01:42
Breath Volume	2.730	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:44
Subject Sample # 2	0.127	01:45
Breath Volume	1.972	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 07/05/2022 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 06/26/2022 00:51:42

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.083	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.153	00:47
Breath Volume	3.261	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.155	00:50
Breath Volume	3.003	00:50
Air Blank	0.000	00:50
Control Test	0.081	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 07/05/2022 14:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOSA,EMMANUEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: BRAVO MUNOZ

DATE/TIME OF TEST: 06/26/2022 04:14:33

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.198	04:09
Breath Volume	3.097	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.197	04:12
Breath Volume	2.843	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 07/05/2022 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIDEAUX,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/26/2022 05:30:47

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:19	
Control Test	0.081	05:20	
Air Blank	0.000	05:20	
Subject Sample	# 1	05:23	No Sample Provided
Breath Volume	0.000	05:23	
Air Blank	0.000	05:24	
Air Blank	0.000	05:26	
Subject Sample	# 2	05:29	No Sample Provided
Breath Volume	0.000	05:29	
Air Blank	0.000	05:29	
Control Test	0.082	05:30	
Air Blank	0.000	05:30	
Diagnostics Check	OK	05:30	

Data Download Date/Time: 07/05/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCALES,BRANDON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 06/26/2022 23:12:00

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.081	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.136	23:07
Breath Volume	1.285	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.131	23:10
Breath Volume	1.699	23:10
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:12

Data Download Date/Time: 07/05/2022 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLARD,DALTON
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GONZALEZ CRUZ

DATE/TIME OF TEST: 06/27/2022 02:08:15

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.082	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.229	02:03
Breath Volume	2.621	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.231	02:06
Breath Volume	2.546	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 07/05/2022 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCCARTHY,MICHAEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 06/30/2022 23:02:29

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.237	22:57
Breath Volume	1.484	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.234	23:00
Breath Volume	1.960	23:00
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 07/05/2022 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: LEITCH,RICKFORD
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: OTHER
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BUITRAGO

DATE/TIME OF TEST: 06/14/2022 01:42:22

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:37
Breath Volume	2.937	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.000	01:41
Breath Volume	1.433	01:41
Air Blank	0.000	01:41
Control Test	0.078	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 07/13/2022 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIND,CHRISTIAN,G
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: LIND,CHRISTIAN
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: PINEDA DELCID

DATE/TIME OF TEST: 06/18/2022 03:04:15

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.152	02:59
Breath Volume	2.636	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.159	03:02
Breath Volume	2.578	03:02
Air Blank	0.000	03:03
Control Test	0.078	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 07/13/2022 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEITCH,RICKFORD,D

BREATH TEST OPERATOR AGENCY: Cocoa PD

ARRESTING OFFICER: FITZGERALD,MICHAEL

ARRESTING OFFICER AGENCY: Cocoa PD

LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 25721080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 05:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 06/18/2022 07:28:29

Diagnostics Check	OK	07:17	
Air Blank	0.000	07:17	
Control Test	0.079	07:17	
Air Blank	0.000	07:18	
Subject Sample	# 1	07:21	No Sample Provided
Breath Volume	0.000	07:21	
Air Blank	0.000	07:21	
Air Blank	0.000	07:23	
Subject Sample	# 2	07:27	No Sample Provided
Breath Volume	0.000	07:27	
Air Blank	0.000	07:27	
Control Test	0.080	07:27	
Air Blank	0.000	07:28	
Diagnostics Check	OK	07:28	

Data Download Date/Time: 07/13/2022 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: BENITEZ-SIFONTES,RAFAEL
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: MONTEJO-MENDOZA

DATE/TIME OF TEST: 06/24/2022 23:26:03

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.217	23:17
Breath Volume	2.300	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.253	23:21
Breath Volume	3.777	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 3	0.229	23:24
Breath Volume	1.472	23:24
Air Blank	0.000	23:25
Control Test	0.078	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:26

Data Download Date/Time: 07/13/2022 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL, RAVEN, N
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HARTZELL, EDWARD
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 06/25/2022 22:12:15

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.079	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.188	22:06
Breath Volume	1.382	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.180	22:10
Breath Volume	1.789	22:10
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 07/13/2022 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYNY,ERIC,C
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: BENITEZ,R
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 06/30/2022 23:21:46

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.137	23:17
Breath Volume	2.011	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.133	23:20
Breath Volume	1.808	23:20
Air Blank	0.000	23:20
Control Test	0.079	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 07/13/2022 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1175968
BREATH TEST OPERATOR: PENDER,DEREK,A	DRY GAS EXPIRY DATE: 07/22/2022
BREATH TEST OPERATOR AGENCY: Rockledge PD	VIOLATION CODE: DUI
ARRESTING OFFICER: LIVINGSTON,JOVON	ARREST TIME: 12:22
ARRESTING OFFICER AGENCY: Rockledge PD	GENDER: Female
LAST AGENCY INSPECTION: 05/31/2022	OBSERVATION PERIOD BEGAN: 13:12

SUBJECT'S LAST NAME: CONNEELY	DATE/TIME OF TEST: 06/11/2022	13:47:56
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	13:32	
Air Blank	0.000	13:33	
Control Test	0.077	13:33	
Air Blank	0.000	13:34	
Subject Sample	# 1 0.092	13:37	Volume Not Met
Breath Volume	0.429	13:37	
Air Blank	0.000	13:37	
Air Blank	0.000	13:39	
Subject Sample	# 2 0.091	13:42	
Breath Volume	1.597	13:42	
Air Blank	0.000	13:42	
Air Blank	0.000	13:44	
Subject Sample	# 3 0.086	13:46	
Breath Volume	1.441	13:46	
Air Blank	0.000	13:47	
Control Test	0.078	13:47	
Air Blank	0.000	13:47	
Diagnostics Check	OK	13:47	

Data Download Date/Time: 06/15/2022 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: WEAVER,JOCELYN
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/22/2022
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HERNANDEZ-VASQUEZ

DATE/TIME OF TEST: 06/15/2022 21:34:23

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.079	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1 0.208	21:22	
Breath Volume	1.933	21:22	
Air Blank	0.000	21:23	
Air Blank	0.000	21:25	
Subject Sample	# 2 0.178	21:28	Volume Not Met
Breath Volume	1.019	21:28	
Air Blank	0.000	21:29	
Air Blank	0.000	21:30	
Subject Sample	# 3 0.195	21:32	
Breath Volume	1.347	21:32	
Air Blank	0.000	21:33	
Control Test	0.077	21:33	
Air Blank	0.000	21:34	
Diagnostics Check	OK	21:34	

Data Download Date/Time: 06/21/2022 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: DUFT,KOYIE
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/22/2022
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: TRAVERS

DATE/TIME OF TEST: 06/17/2022 01:14:03

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.205	01:08
Breath Volume	1.386	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.198	01:12
Breath Volume	1.855	01:12
Air Blank	0.000	01:13
Control Test	0.077	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:14

Data Download Date/Time: 06/21/2022 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILBIN,SKYLER,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MICCIANTUONO,JAMES
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: BATTEN

DATE/TIME OF TEST: 06/14/2022 00:07:16

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.174	00:02
Breath Volume	2.589	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.169	00:05
Breath Volume	2.472	00:05
Air Blank	0.000	00:06
Control Test	0.076	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 06/29/2022 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTTO,ALISON,E
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: OTTO,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: TROUP

DATE/TIME OF TEST: 06/14/2022 08:02:39

Diagnostics Check	OK	07:56
Air Blank	0.000	07:56
Control Test	0.080	07:57
Air Blank	0.000	07:57
Subject Sample # 1	0.000	07:58
Breath Volume	1.152	07:58
Air Blank	0.000	07:58
Air Blank	0.000	08:00
Subject Sample # 2	0.000	08:01
Breath Volume	1.191	08:01
Air Blank	0.000	08:01
Control Test	0.079	08:02
Air Blank	0.000	08:02
Diagnostics Check	OK	08:02

Data Download Date/Time: 06/29/2022 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ,ANDREW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 06/17/2022 00:38:52

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.081	00:34
Breath Volume	1.250	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.078	00:37
Breath Volume	2.230	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 06/29/2022 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ,ANDREW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: BEGLEY

DATE/TIME OF TEST: 06/17/2022 20:51:24

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.080	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.000	20:46
Breath Volume	1.664	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.000	20:50
Breath Volume	1.265	20:50
Air Blank	0.000	20:50
Control Test	0.079	20:50
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 06/29/2022 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ,ANDREW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: DAGOSTINO

DATE/TIME OF TEST: 06/18/2022 21:26:20

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.080	21:21
Air Blank	0.000	21:21
Subject Sample # 1	0.000	21:22
Breath Volume	1.710	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.000	21:24
Breath Volume	2.199	21:24
Air Blank	0.000	21:25
Control Test	0.081	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 06/29/2022 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: WILLINGHAM,JARRETT
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: JEEVANDHAM

DATE/TIME OF TEST: 06/19/2022 05:09:01

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.148	05:04
Breath Volume	1.699	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:06
Subject Sample # 2	0.147	05:07
Breath Volume	1.875	05:07
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:09

Data Download Date/Time: 06/29/2022 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTTO,ALISON,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: OTTO,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: SALISBURY

DATE/TIME OF TEST: 06/19/2022 14:09:46

Diagnostics Check	OK	14:00
Air Blank	0.000	14:01
Control Test	0.080	14:01
Air Blank	0.000	14:02
Subject Sample # 1	0.000	14:04
Breath Volume	1.339	14:04
Air Blank	0.000	14:05
Air Blank	0.000	14:07
Subject Sample # 2	0.000	14:08
Breath Volume	1.367	14:08
Air Blank	0.000	14:08
Control Test	0.080	14:09
Air Blank	0.000	14:09
Diagnostics Check	OK	14:09

Data Download Date/Time: 06/29/2022 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ,ANDREW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: CORRIGAN

DATE/TIME OF TEST: 06/19/2022 20:06:47

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.080	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.334	20:02
Breath Volume	1.718	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:04
Subject Sample # 2	0.335	20:05
Breath Volume	1.679	20:05
Air Blank	0.000	20:05
Control Test	0.079	20:06
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 06/29/2022 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTTO,ALISON,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: OTTO,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 08:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: STAPLES

DATE/TIME OF TEST: 06/22/2022 10:40:40

Diagnostics Check	OK	10:34
Air Blank	0.000	10:35
Control Test	0.081	10:35
Air Blank	0.000	10:35
Subject Sample # 1	0.000	10:36
Breath Volume	1.378	10:36
Air Blank	0.000	10:36
Air Blank	0.000	10:38
Subject Sample # 2	0.000	10:39
Breath Volume	1.847	10:39
Air Blank	0.000	10:39
Control Test	0.080	10:40
Air Blank	0.000	10:40
Diagnostics Check	OK	10:40

Data Download Date/Time: 06/29/2022 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTTO,ALISON,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: OTTO,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: UPDIKE

DATE/TIME OF TEST: 06/27/2022 09:21:59

Diagnostics Check	OK	09:15
Air Blank	0.000	09:15
Control Test	0.081	09:16
Air Blank	0.000	09:16
Subject Sample # 1	0.000	09:17
Breath Volume	2.054	09:17
Air Blank	0.000	09:18
Air Blank	0.000	09:20
Subject Sample # 2	0.000	09:20
Breath Volume	1.460	09:20
Air Blank	0.000	09:21
Control Test	0.080	09:21
Air Blank	0.000	09:21
Diagnostics Check	OK	09:21

Data Download Date/Time: 06/29/2022 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTTO,ALISON,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: OTTO,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: CLOUATRE

DATE/TIME OF TEST: 06/28/2022 09:13:39

Diagnostics Check	OK	09:06
Air Blank	0.000	09:07
Control Test	0.080	09:07
Air Blank	0.000	09:08
Subject Sample # 1	0.000	09:09
Breath Volume	1.132	09:09
Air Blank	0.000	09:09
Air Blank	0.000	09:11
Subject Sample # 2	0.000	09:12
Breath Volume	1.273	09:12
Air Blank	0.000	09:12
Control Test	0.081	09:13
Air Blank	0.000	09:13
Diagnostics Check	OK	09:13

Data Download Date/Time: 06/29/2022 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPLAN,MICHAEL,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: KAPLAN,MICHAEL
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: BUSBY

DATE/TIME OF TEST: 06/28/2022 23:25:21

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.080	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.108	23:20
Breath Volume	1.703	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.107	23:23
Breath Volume	1.804	23:23
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 06/29/2022 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,A,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SUDMAN,B
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 06/29/2022 18:40:02

Diagnostics Check	OK	18:36	
Air Blank	0.000	18:37	
Control Test	0.080	18:37	
Air Blank	0.000	18:38	
Subject Sample	# 1	18:38	Subject Test Refused
Breath Volume	0.000	18:38	
Air Blank	0.000	18:39	
Control Test	0.080	18:39	
Air Blank	0.000	18:39	
Diagnostics Check	OK	18:40	

Data Download Date/Time: 07/28/2022 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINSKER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: PINSKER,JOSEPH
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 06/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BEYER

DATE/TIME OF TEST: 06/18/2022 01:12:28

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:08	
Control Test	0.079	01:09	
Air Blank	0.000	01:09	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 07/09/2022 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAL, ADAM, M
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: BEAL, ADAM
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 06/05/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: DILLARD

DATE/TIME OF TEST: 06/20/2022 00:06:26

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:55	
Control Test	0.080	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 2 0.141	23:59	Volume Not Met
Breath Volume	0.953	23:59	
Air Blank	0.000	23:59	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.150	00:04	Volume Not Met
Breath Volume	0.464	00:04	
Air Blank	0.000	00:05	
Control Test	0.079	00:05	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 07/09/2022 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAL, ADAM, M
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: BEAL, ADAM
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 06/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: DILLARD

DATE/TIME OF TEST: 06/20/2022 00:17:54

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.142	00:13
Breath Volume	1.234	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.160	00:16
Breath Volume	3.664	00:16
Air Blank	0.000	00:17
Control Test	0.078	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 07/09/2022 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACHELLER,ALEXANDRIA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SCHUCK,KYLE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: STACIOKAS

DATE/TIME OF TEST: 06/10/2022 03:16:51

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.081	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.165	03:10
Breath Volume	1.585	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:12
Subject Sample # 2	0.157	03:15
Breath Volume	1.914	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 06/27/2022 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DONALDSON,LAUREN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: MCALPIN

DATE/TIME OF TEST: 06/11/2022 20:07:45

Diagnostics Check	OK	19:58
Air Blank	0.000	19:59
Control Test	0.080	19:59
Air Blank	0.000	20:00
Subject Sample # 1	0.203	20:02
Breath Volume	1.328	20:02
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.189	20:06
Breath Volume	1.511	20:06
Air Blank	0.000	20:06
Control Test	0.079	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 06/27/2022 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HACKETT,BRADLEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 06/15/2022 02:42:45

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.222	02:37
Breath Volume	1.523	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.227	02:41
Breath Volume	1.707	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/27/2022 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACHELLER,ALEXANDRIA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BACHELLER,ALEXANDRIA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: OQUENDO

DATE/TIME OF TEST: 06/19/2022 22:02:13

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.172	21:57
Breath Volume	1.367	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.176	22:00
Breath Volume	1.250	22:00
Air Blank	0.000	22:01
Control Test	0.076	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 06/27/2022 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACHELLER,ALEXANDRIA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SCHOENER,ANDREW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PEREZ PALENCIA

DATE/TIME OF TEST: 06/23/2022 03:07:18

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.154	03:02
Breath Volume	1.144	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.152	03:05
Breath Volume	1.617	03:05
Air Blank	0.000	03:06
Control Test	0.076	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 06/27/2022 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:47

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: BHAGU

DATE/TIME OF TEST: 06/05/2022 02:03:19

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.079	02:00	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:02	
Control Test	0.079	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 06/20/2022 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: URIAS

DATE/TIME OF TEST: 06/06/2022 02:24:53

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:22	
Control Test	0.079	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:23	RFI Detect
Breath Volume	1.480	02:23	
Air Blank	0.000	02:23	
Control Test	0.079	02:24	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 06/20/2022 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: URIAS

DATE/TIME OF TEST: 06/06/2022 02:32:11

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.184	02:27
Breath Volume	1.910	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.188	02:30
Breath Volume	2.164	02:30
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 06/20/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: FERINO,WESLEY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: WHEATLEY

DATE/TIME OF TEST: 06/11/2022 02:34:16

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.210	02:29
Breath Volume	1.789	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.215	02:32
Breath Volume	2.945	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 06/20/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LANE,CHRISTINA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 20:08

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: SWALLEY

DATE/TIME OF TEST: 06/11/2022 20:28:19

Diagnostics Check	OK	20:20
Air Blank	0.000	20:20
Control Test	0.079	20:21
Air Blank	0.000	20:21
Subject Sample # 1	0.254	20:23
Breath Volume	1.472	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.270	20:26
Breath Volume	1.542	20:26
Air Blank	0.000	20:27
Control Test	0.079	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 06/20/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 23:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 06/12/2022 00:07:23

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.071	00:02
Breath Volume	1.515	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.071	00:05
Breath Volume	1.742	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 06/20/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: ZINSER

DATE/TIME OF TEST: 06/12/2022 01:23:24

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.300	01:16
Breath Volume	1.476	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.323	01:19
Breath Volume	1.917	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 3	0.319	01:21
Breath Volume	2.011	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 06/20/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ-VOL/.02 PC 4 S1,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: GONZALEZ LOPEZ

DATE/TIME OF TEST: 06/18/2022 20:41:22

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.080	20:36
Air Blank	0.000	20:36
Subject Sample # 1	0.124	20:37
Breath Volume	2.960	20:37
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.115	20:39
Breath Volume	1.378	20:39
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 06/20/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: FERINO,WESLEY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 22:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: PICOE

DATE/TIME OF TEST: 06/19/2022 22:38:18

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.080	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.228	22:31
Breath Volume	1.703	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.209	22:36
Breath Volume	2.425	22:36
Air Blank	0.000	22:37
Control Test	0.081	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 06/20/2022 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:29

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: WESTERMAN

DATE/TIME OF TEST: 06/25/2022 02:47:15

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.166	02:42
Breath Volume	2.234	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.166	02:45
Breath Volume	1.617	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 07/31/2022 18:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: CASTRO,JUAN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MICHEL

DATE/TIME OF TEST: 06/25/2022 03:16:23

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.093	03:10
Breath Volume	1.300	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.083	03:14
Breath Volume	1.882	03:14
Air Blank	0.000	03:15
Control Test	0.078	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 07/31/2022 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CANADY,CHRISTINA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 06/03/2022 03:02:25

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.089	02:57
Breath Volume	1.945	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.092	03:00
Breath Volume	2.210	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 06/22/2022 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIDEROA,KEVIN,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: HIDEROA,KEVIN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: OSTERBLOM

DATE/TIME OF TEST: 06/03/2022 23:44:02

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.080	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:43	Subject Test Refused
Breath Volume	0.000	23:43	
Air Blank		23:43	RFI Detect
Air Blank	0.000	23:44	

Data Download Date/Time: 06/22/2022 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DIAL,VICTORIA
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: ESTRADA CHAVEZ

DATE/TIME OF TEST: 06/05/2022 04:04:19

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.082	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.101	03:59
Breath Volume	2.917	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.104	04:02
Breath Volume	3.093	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 06/22/2022 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIDEROA,KEVIN,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: HIDEROA,KEVIN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 06/05/2022 05:16:27

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.082	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.221	05:11
Breath Volume	2.007	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.212	05:14
Breath Volume	1.750	05:14
Air Blank	0.000	05:15
Control Test	0.079	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 06/22/2022 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PONDER,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: PONDER,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: COSTANZO

DATE/TIME OF TEST: 06/05/2022 17:43:52

Diagnostics Check	OK	17:41	
Air Blank	0.000	17:41	
Control Test	0.080	17:41	
Air Blank	0.000	17:42	
Subject Sample	# 1	17:42	Subject Test Refused
Breath Volume	0.000	17:42	
Air Blank	0.000	17:42	
Control Test	0.080	17:43	
Air Blank	0.000	17:43	
Diagnostics Check	OK	17:43	

Data Download Date/Time: 06/22/2022 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HEARSEY,RYAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: RODRIQUEZ-TARR

DATE/TIME OF TEST: 06/09/2022 01:41:35

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.145	01:36
Breath Volume	1.234	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.142	01:39
Breath Volume	1.746	01:39
Air Blank	0.000	01:40
Control Test	0.081	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 06/22/2022 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DIAL,VICTORIA
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/09/2022 05:02:36

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.080	04:56
Air Blank	0.000	04:56
Subject Sample # 1	0.000	04:58
Breath Volume	2.378	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.000	05:01
Breath Volume	2.398	05:01
Air Blank	0.000	05:01
Control Test	0.082	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 06/22/2022 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISSEY,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: MORRISSEY,ZACHARY
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: RAMIREZ VELASQUEZ

DATE/TIME OF TEST: 06/09/2022 19:07:35

Diagnostics Check	OK	18:57
Air Blank	0.000	18:57
Control Test	0.081	18:57
Air Blank	0.000	18:58
Subject Sample # 1	0.265	18:59
Breath Volume	1.871	18:59
Air Blank	0.000	19:00
Air Blank	0.000	19:01
Subject Sample # 2	0.231	19:02
Breath Volume	1.765	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 3	0.253	19:05
Breath Volume	1.210	19:05
Air Blank	0.000	19:06
Control Test	0.075	19:07
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 06/22/2022 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HAMILTON,DWAYNE
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 06/09/2022 22:46:25

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.245	22:41
Breath Volume	1.589	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.240	22:44
Breath Volume	1.449	22:44
Air Blank	0.000	22:45
Control Test	0.080	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 06/22/2022 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CABRERA,DARIAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/10/2022 02:18:18

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.181	02:13
Breath Volume	3.894	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.183	02:16
Breath Volume	3.527	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 06/22/2022 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: GADSON

DATE/TIME OF TEST: 06/10/2022 02:48:49

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.063	02:44
Breath Volume	3.601	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.064	02:47
Breath Volume	3.750	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 06/22/2022 22:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REHDER,PATRICK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BAUMGARDNER

DATE/TIME OF TEST: 06/10/2022 04:21:54

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.168	04:17
Breath Volume	3.324	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.168	04:20
Breath Volume	2.808	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 06/22/2022 22:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SAUNDERS,HUNTER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 06/11/2022 02:42:47

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.154	02:38
Breath Volume	1.894	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.169	02:41
Breath Volume	2.273	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/22/2022 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: KRUPA,KYLE
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: LEDFORD

DATE/TIME OF TEST: 06/11/2022 17:56:43

Diagnostics Check	OK	17:48	
Air Blank	0.000	17:48	
Control Test	0.081	17:49	
Air Blank	0.000	17:49	
Subject Sample # 1	0.183	17:50	
Breath Volume	2.191	17:50	
Air Blank	0.000	17:50	
Air Blank	0.000	17:52	
Subject Sample # 2	0.160	17:53	
Breath Volume	1.230	17:53	
Air Blank	0.000	17:53	
Air Blank		17:55	RFI Detect
Air Blank	0.000	17:55	
Control Test	0.081	17:56	
Air Blank	0.000	17:56	
Diagnostics Check	OK	17:56	

Data Download Date/Time: 06/22/2022 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: KRUPA,KYLE
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: LEDFORD

DATE/TIME OF TEST: 06/11/2022 18:05:23

Diagnostics Check	OK	17:58
Air Blank	0.000	17:59
Control Test	0.080	17:59
Air Blank	0.000	18:00
Subject Sample # 1	0.168	18:00
Breath Volume	1.226	18:00
Air Blank	0.000	18:01
Air Blank	0.000	18:02
Subject Sample # 2	0.164	18:03
Breath Volume	1.640	18:03
Air Blank	0.000	18:04
Control Test	0.079	18:04
Air Blank	0.000	18:05
Diagnostics Check	OK	18:05

Data Download Date/Time: 06/22/2022 22:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,JERRY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 06/12/2022 02:47:08

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.082	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.238	02:42
Breath Volume	2.347	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.234	02:45
Breath Volume	3.308	02:45
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 06/22/2022 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERVAIS, ROSEMARY, C
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLAZINA, M
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: FRYER

DATE/TIME OF TEST: 06/30/2022 02:20:17

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.084	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.000	02:15
Breath Volume	3.503	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.000	02:18
Breath Volume	2.750	02:18
Air Blank	0.000	02:19
Control Test	0.083	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 07/16/2022 23:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SPEARS JR

DATE/TIME OF TEST: 06/03/2022 23:47:41

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.078	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.000	23:43
Breath Volume	1.289	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.000	23:46
Breath Volume	1.410	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 06/22/2022 19:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 06/22/2022 23:09:25

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.079	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.000	23:05
Breath Volume	2.546	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.000	23:08
Breath Volume	2.367	23:08
Air Blank	0.000	23:08
Control Test	0.079	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 07/14/2022 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORT,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WORBY,DAVID
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:57

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 06/24/2022 07:28:50

Diagnostics Check	OK	07:18
Air Blank	0.000	07:18
Control Test	0.080	07:18
Air Blank	0.000	07:19
Subject Sample # 1	0.269	07:20
Breath Volume	2.031	07:20
Air Blank	0.000	07:21
Air Blank	0.000	07:22
Subject Sample # 2	0.236	07:23
Breath Volume	1.851	07:23
Air Blank	0.000	07:24
Air Blank	0.000	07:26
Subject Sample # 3	0.235	07:27
Breath Volume	1.683	07:27
Air Blank	0.000	07:27
Control Test	0.076	07:28
Air Blank	0.000	07:28
Diagnostics Check	OK	07:28

Data Download Date/Time: 07/14/2022 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: ANGLIN

DATE/TIME OF TEST: 06/24/2022 23:33:17

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.235	23:28
Breath Volume	2.820	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.228	23:31
Breath Volume	2.535	23:31
Air Blank	0.000	23:32
Control Test	0.084	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 07/14/2022 19:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 06/25/2022 01:48:02

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.057	01:43
Breath Volume	3.000	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.063	01:46
Breath Volume	2.503	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:48

Data Download Date/Time: 07/14/2022 19:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: JUNG

DATE/TIME OF TEST: 06/25/2022 04:04:02

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:03	
Control Test	0.079	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 07/14/2022 19:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/25/2022 19:50:11

Diagnostics Check	OK	19:41	
Air Blank	0.000	19:41	
Control Test	0.079	19:42	
Air Blank	0.000	19:42	
Subject Sample	# 1	19:45	No Sample Provided
Breath Volume	0.000	19:45	
Air Blank	0.000	19:46	
Air Blank	0.000	19:48	
Subject Sample	# 2	19:48	Subject Test Refused
Breath Volume	0.000	19:48	
Air Blank	0.000	19:49	
Control Test	0.079	19:49	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 07/14/2022 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: LUKE

DATE/TIME OF TEST: 06/26/2022 00:41:44

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.129	00:37
Breath Volume	1.570	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:39
Subject Sample # 2	0.124	00:40
Breath Volume	1.718	00:40
Air Blank	0.000	00:40
Control Test	0.079	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: CONTTRERAS

DATE/TIME OF TEST: 06/26/2022 01:15:33

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:13	
Control Test	0.078	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Control Test	0.079	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: ALBEE

DATE/TIME OF TEST: 06/26/2022 01:33:18

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:30	
Control Test	0.079	01:30	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.079	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOODY,BRITTANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: FORTNEY

DATE/TIME OF TEST: 06/26/2022 02:01:51

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.082	01:57
Breath Volume	1.503	01:57
Air Blank	0.000	01:58
Air Blank	0.000	01:59
Subject Sample # 2	0.080	02:00
Breath Volume	1.671	02:00
Air Blank	0.000	02:00
Control Test	0.077	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: WHITNER

DATE/TIME OF TEST: 06/26/2022 04:32:37

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.098	04:27
Breath Volume	1.328	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample # 2	0.095	04:30
Breath Volume	1.261	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PHELPS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: HILBERT

DATE/TIME OF TEST: 06/26/2022 05:12:17

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.078	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.185	05:05
Breath Volume	1.140	05:05
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.185	05:10
Breath Volume	1.312	05:10
Air Blank	0.000	05:11
Control Test	0.078	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EALY,STEVEN,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MATTOX,DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 13:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: BRAIA

DATE/TIME OF TEST: 06/27/2022 14:07:13

Diagnostics Check	OK	14:00
Air Blank	0.000	14:01
Control Test	0.079	14:01
Air Blank	0.000	14:01
Subject Sample # 1	0.000	14:02
Breath Volume	1.335	14:02
Air Blank	0.000	14:03
Air Blank	0.000	14:05
Subject Sample # 2	0.000	14:05
Breath Volume	1.929	14:05
Air Blank	0.000	14:06
Control Test	0.080	14:06
Air Blank	0.000	14:07
Diagnostics Check	OK	14:07

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHRISTIE,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: LIVESAY

DATE/TIME OF TEST: 06/28/2022 00:44:27

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.079	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:43	
Control Test	0.079	00:43	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 06/29/2022 22:47:14

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.080	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.085	22:42
Breath Volume	1.226	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.083	22:45
Breath Volume	1.230	22:45
Air Blank	0.000	22:46
Control Test	0.078	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CAPUCHINO TOVAR

DATE/TIME OF TEST: 06/30/2022 03:52:30

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.079	03:50	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.078	03:51	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAMBROSIO,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: BAUM

DATE/TIME OF TEST: 06/22/2022 23:07:41

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:05	
Control Test	0.080	23:05	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:06	Subject Test Refused
Breath Volume	0.000	23:06	
Air Blank	0.000	23:06	
Control Test	0.079	23:07	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 07/14/2022 19:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ASSELIN

DATE/TIME OF TEST: 06/23/2022 01:03:22

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:01	
Control Test	0.080	01:01	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:02	
Control Test	0.079	01:02	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 07/14/2022 19:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 06/23/2022 03:05:11

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:02	
Control Test	0.079	03:03	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.080	03:04	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 07/14/2022 19:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 06/23/2022 03:30:01

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:28	Subject Test Refused
Breath Volume	0.000	03:28	
Air Blank	0.000	03:29	
Control Test	0.080	03:29	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 07/14/2022 19:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVANOVICLM,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IVANOVICLM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 06/23/2022 18:51:09

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:46	
Control Test	0.080	18:46	
Air Blank	0.000	18:47	
Subject Sample	# 1	18:49	Subject Test Refused
Breath Volume	0.000	18:49	
Air Blank	0.000	18:50	
Control Test	0.080	18:50	
Air Blank	0.000	18:51	
Diagnostics Check	OK	18:51	

Data Download Date/Time: 07/14/2022 19:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 06/23/2022 20:19:13

Diagnostics Check	OK	20:12
Air Blank	0.000	20:13
Control Test	0.080	20:13
Air Blank	0.000	20:14
Subject Sample # 1	0.064	20:14
Breath Volume	2.011	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.068	20:17
Breath Volume	2.253	20:17
Air Blank	0.000	20:18
Control Test	0.079	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MARTINELLI

DATE/TIME OF TEST: 06/24/2022 05:44:01

Diagnostics Check	OK	05:36
Air Blank	0.000	05:36
Control Test	0.080	05:36
Air Blank	0.000	05:37
Subject Sample # 1	0.127	05:39
Breath Volume	1.398	05:39
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample # 2	0.117	05:42
Breath Volume	1.914	05:42
Air Blank	0.000	05:43
Control Test	0.078	05:43
Air Blank	0.000	05:43
Diagnostics Check	OK	05:44

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: LAUX

DATE/TIME OF TEST: 06/24/2022 17:57:33

Diagnostics Check	OK	17:51
Air Blank	0.000	17:51
Control Test	0.080	17:51
Air Blank	0.000	17:52
Subject Sample # 1	0.180	17:53
Breath Volume	1.507	17:53
Air Blank	0.000	17:53
Air Blank	0.000	17:55
Subject Sample # 2	0.174	17:56
Breath Volume	1.445	17:56
Air Blank	0.000	17:56
Control Test	0.079	17:56
Air Blank	0.000	17:57
Diagnostics Check	OK	17:57

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: SYLVESTRI

DATE/TIME OF TEST: 06/24/2022 22:22:59

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.229	22:18
Breath Volume	1.128	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.222	22:21
Breath Volume	2.640	22:21
Air Blank	0.000	22:22
Control Test	0.079	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIS,BRITT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: SHORTER

DATE/TIME OF TEST: 06/24/2022 23:33:40

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.120	23:29
Breath Volume	1.414	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.129	23:32
Breath Volume	1.781	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTELLA,D
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: CISTONE

DATE/TIME OF TEST: 06/25/2022 01:51:24

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.078	01:47	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.214	01:49	
Air Blank	0.000	01:50	
Control Test	0.079	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BENTSCHNER,OLIVIA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: OJEDA-MARTINEZ

DATE/TIME OF TEST: 06/25/2022 18:47:04

Diagnostics Check	OK	18:44	
Air Blank	0.000	18:44	
Control Test	0.079	18:45	
Air Blank	0.000	18:45	
Subject Sample	# 1	18:45	Subject Test Refused
Breath Volume	0.000	18:45	
Air Blank	0.000	18:46	
Control Test	0.080	18:46	
Air Blank	0.000	18:46	
Diagnostics Check	OK	18:47	

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: AKERS

DATE/TIME OF TEST: 06/25/2022 21:53:57

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.080	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.204	21:49
Breath Volume	2.003	21:49
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.203	21:52
Breath Volume	1.867	21:52
Air Blank	0.000	21:53
Control Test	0.077	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: GRATER GERARDI

DATE/TIME OF TEST: 06/25/2022 23:45:39

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.000	23:41
Breath Volume	1.933	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.000	23:44
Breath Volume	2.289	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 06/26/2022 04:51:22

Diagnostics Check	OK	04:48	
Air Blank	0.000	04:48	
Control Test	0.080	04:49	
Air Blank	0.000	04:49	
Subject Sample	# 1	04:50	Subject Test Refused
Breath Volume	0.000	04:50	
Air Blank	0.000	04:50	
Control Test	0.080	04:50	
Air Blank	0.000	04:51	
Diagnostics Check	OK	04:51	

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 06/26/2022 05:07:22

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.079	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.075	05:02
Breath Volume	2.617	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.077	05:05
Breath Volume	1.953	05:05
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: GLOSS

DATE/TIME OF TEST: 06/26/2022 05:23:26

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:20	
Control Test	0.080	05:21	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:22	Subject Test Refused
Breath Volume	0.000	05:22	
Air Blank	0.000	05:22	
Control Test	0.079	05:22	
Air Blank	0.000	05:23	
Diagnostics Check	OK	05:23	

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: ACKAL

DATE/TIME OF TEST: 06/26/2022 17:59:22

Diagnostics Check	OK	17:51
Air Blank	0.000	17:51
Control Test	0.079	17:52
Air Blank	0.000	17:52
Subject Sample # 1	0.221	17:54
Breath Volume	2.054	17:54
Air Blank	0.000	17:54
Air Blank	0.000	17:56
Subject Sample # 2	0.219	17:57
Breath Volume	1.441	17:57
Air Blank	0.000	17:58
Control Test	0.079	17:58
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DUCKERS,ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: HARLOW

DATE/TIME OF TEST: 06/26/2022 19:11:28

Diagnostics Check	OK	19:03
Air Blank	0.000	19:04
Control Test	0.078	19:04
Air Blank	0.000	19:05
Subject Sample # 1	0.126	19:06
Breath Volume	2.117	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.126	19:09
Breath Volume	2.164	19:09
Air Blank	0.000	19:10
Control Test	0.078	19:10
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRENNAN,CONNOR
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 06/27/2022 00:18:39

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.079	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.042	00:14
Breath Volume	3.398	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.042	00:17
Breath Volume	3.351	00:17
Air Blank	0.000	00:17
Control Test	0.080	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DUNCAN,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MURCHIE

DATE/TIME OF TEST: 06/27/2022 01:13:03

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:10	
Control Test	0.079	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:12	
Control Test	0.080	01:12	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS, BRIDGETT, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HICKS, JASON
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: ZAMMCITARO

DATE/TIME OF TEST: 06/28/2022 03:32:19

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:29	
Control Test	0.079	03:30	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:31	
Control Test	0.080	03:31	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 07/14/2022 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/28/2022 22:02:42

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.070	21:57
Breath Volume	2.394	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.068	22:01
Breath Volume	2.578	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEATTY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BRUNELLE

DATE/TIME OF TEST: 06/29/2022 02:42:11

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.000	02:37
Breath Volume	2.734	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.000	02:40
Breath Volume	2.468	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEHR,GEORGE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 10:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 06/29/2022 11:46:12

Diagnostics Check	OK	11:40
Air Blank	0.000	11:40
Control Test	0.080	11:40
Air Blank	0.000	11:41
Subject Sample # 1	0.000	11:41
Breath Volume	2.496	11:41
Air Blank	0.000	11:42
Air Blank	0.000	11:44
Subject Sample # 2	0.000	11:44
Breath Volume	1.394	11:44
Air Blank	0.000	11:45
Control Test	0.080	11:45
Air Blank	0.000	11:46
Diagnostics Check	OK	11:46

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IFFT,GARRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IFFT,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: HEINE

DATE/TIME OF TEST: 06/29/2022 20:53:23

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.080	20:48
Air Blank	0.000	20:48
Subject Sample # 1	0.137	20:49
Breath Volume	3.402	20:49
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.135	20:51
Breath Volume	2.429	20:51
Air Blank	0.000	20:52
Control Test	0.080	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SAUVAGEOT

DATE/TIME OF TEST: 06/30/2022 01:04:02

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.112	00:59
Breath Volume	3.179	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:01
Subject Sample # 2	0.112	01:02
Breath Volume	2.742	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:04

Data Download Date/Time: 07/14/2022 19:06