

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LORENZO,FRANCISCO,  
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD  
ARRESTING OFFICER: REYES,ISRAEL  
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD  
LAST AGENCY INSPECTION: 06/01/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06022080A2  
DRY GAS EXPIRY DATE: 04/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: JIMENEZ GONZALEZ

DATE/TIME OF TEST: 06/02/2022 05:55:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:45 |
| Air Blank          | 0.000 | 05:45 |
| Control Test       | 0.080 | 05:46 |
| Air Blank          | 0.000 | 05:46 |
| Subject Sample # 1 | 0.249 | 05:48 |
| Breath Volume      | 2.269 | 05:48 |
| Air Blank          | 0.000 | 05:48 |
| Air Blank          | 0.000 | 05:50 |
| Subject Sample # 2 | 0.231 | 05:53 |
| Breath Volume      | 2.039 | 05:53 |
| Air Blank          | 0.000 | 05:54 |
| Control Test       | 0.079 | 05:54 |
| Air Blank          | 0.000 | 05:55 |
| Diagnostics Check  | OK    | 05:55 |

---

Data Download Date/Time: 06/03/2022 13:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LORENZO,FRANCISCO,  
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD  
ARRESTING OFFICER: LORENZO,FRANCISCO  
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD  
LAST AGENCY INSPECTION: 06/01/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06022080A2  
DRY GAS EXPIRY DATE: 04/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: LLERENA HERNANDEZ

DATE/TIME OF TEST: 06/11/2022 23:26:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:23 |                      |
| Air Blank         | 0.000 | 23:23 |                      |
| Control Test      | 0.080 | 23:24 |                      |
| Air Blank         | 0.000 | 23:24 |                      |
| Subject Sample    | # 1   | 23:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:24 |                      |
| Air Blank         | 0.000 | 23:25 |                      |
| Control Test      | 0.079 | 23:25 |                      |
| Air Blank         | 0.000 | 23:26 |                      |
| Diagnostics Check | OK    | 23:26 |                      |

---

Data Download Date/Time: 07/01/2022 12:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUBELA,CHANTEL,  
BREATH TEST OPERATOR AGENCY: DORAL PD  
ARRESTING OFFICER: CUBELA,CHANTEL  
ARRESTING OFFICER AGENCY: DORAL PD  
LAST AGENCY INSPECTION: 06/01/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06022080A2  
DRY GAS EXPIRY DATE: 04/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BAIDE

DATE/TIME OF TEST: 06/25/2022 04:53:45

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:50 |                      |
| Air Blank         | 0.000 | 04:51 |                      |
| Control Test      | 0.080 | 04:51 |                      |
| Air Blank         | 0.000 | 04:52 |                      |
| Subject Sample    | # 1   | 04:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:52 |                      |
| Air Blank         | 0.000 | 04:52 |                      |
| Control Test      | 0.079 | 04:53 |                      |
| Air Blank         | 0.000 | 04:53 |                      |
| Diagnostics Check | OK    | 04:53 |                      |

---

Data Download Date/Time: 07/01/2022 12:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LORENZO,FRANCISCO,  
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD  
ARRESTING OFFICER: LORENZO,FRANCISCO  
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD  
LAST AGENCY INSPECTION: 06/01/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06022080A2  
DRY GAS EXPIRY DATE: 04/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CURBELO

DATE/TIME OF TEST: 06/26/2022 02:01:15

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:58 |                      |
| Air Blank         | 0.000 | 01:58 |                      |
| Control Test      | 0.081 | 01:59 |                      |
| Air Blank         | 0.000 | 01:59 |                      |
| Subject Sample    | # 1   | 01:59 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:59 |                      |
| Air Blank         | 0.000 | 02:00 |                      |
| Control Test      | 0.080 | 02:00 |                      |
| Air Blank         | 0.000 | 02:01 |                      |
| Diagnostics Check | OK    | 02:01 |                      |

---

Data Download Date/Time: 07/01/2022 12:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORENO,FRANCISCO,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: MORENO,FRANCISCO  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/03/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: URBINA

DATE/TIME OF TEST: 06/20/2022 07:32:57

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 07:30 |                      |
| Air Blank         | 0.000 | 07:30 |                      |
| Control Test      | 0.078 | 07:30 |                      |
| Air Blank         | 0.000 | 07:31 |                      |
| Subject Sample    | # 1   | 07:31 | Subject Test Refused |
| Breath Volume     | 0.000 | 07:31 |                      |
| Air Blank         | 0.000 | 07:32 |                      |
| Control Test      | 0.078 | 07:32 |                      |
| Air Blank         | 0.000 | 07:32 |                      |
| Diagnostics Check | OK    | 07:32 |                      |

---

Data Download Date/Time: 07/06/2022 05:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,LEON,J  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: OCHOA,GUILLERMO  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 12:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DAVILA SOLANO

DATE/TIME OF TEST: 06/04/2022 02:31:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:23 |
| Air Blank          | 0.000 | 02:24 |
| Control Test       | 0.079 | 02:24 |
| Air Blank          | 0.000 | 02:25 |
| Subject Sample # 1 | 0.164 | 02:26 |
| Breath Volume      | 1.281 | 02:26 |
| Air Blank          | 0.000 | 02:27 |
| Air Blank          | 0.000 | 02:28 |
| Subject Sample # 2 | 0.165 | 02:29 |
| Breath Volume      | 1.597 | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Control Test       | 0.077 | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Diagnostics Check  | OK    | 02:31 |

---

Data Download Date/Time: 06/09/2022 02:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLISON,DAVID,J  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: ELLISON,DAVID  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: CORDERO HERNANDEZ

DATE/TIME OF TEST: 06/05/2022 21:50:38

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 21:38 |                |
| Air Blank         | 0.000     | 21:38 |                |
| Control Test      | 0.080     | 21:38 |                |
| Air Blank         | 0.000     | 21:39 |                |
| Subject Sample    | # 1 0.156 | 21:42 | Volume Not Met |
| Breath Volume     | 0.800     | 21:42 |                |
| Air Blank         | 0.000     | 21:43 |                |
| Air Blank         | 0.000     | 21:45 |                |
| Subject Sample    | # 2 0.168 | 21:45 |                |
| Breath Volume     | 1.234     | 21:45 |                |
| Air Blank         | 0.000     | 21:46 |                |
| Air Blank         | 0.000     | 21:48 |                |
| Subject Sample    | # 3 0.166 | 21:48 |                |
| Breath Volume     | 1.812     | 21:48 |                |
| Air Blank         | 0.000     | 21:49 |                |
| Control Test      | 0.079     | 21:50 |                |
| Air Blank         | 0.000     | 21:50 |                |
| Diagnostics Check | OK        | 21:50 |                |

---

Data Download Date/Time: 06/09/2022 02:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PREDELUS,J,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: PREDELUS,J  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 05/05/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:27

SUBJECT'S LAST NAME: PENA AMUE

DATE/TIME OF TEST: 06/06/2022 09:28:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 09:24 |                      |
| Air Blank         | 0.000 | 09:25 |                      |
| Control Test      | 0.080 | 09:25 |                      |
| Air Blank         | 0.000 | 09:25 |                      |
| Subject Sample    | # 1   | 09:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 09:27 |                      |
| Air Blank         | 0.000 | 09:27 |                      |
| Control Test      | 0.080 | 09:28 |                      |
| Air Blank         | 0.000 | 09:28 |                      |
| Diagnostics Check | OK    | 09:28 |                      |

---

Data Download Date/Time: 06/09/2022 02:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAVARRIA,DEGLYS,E  
BREATH TEST OPERATOR AGENCY: MEDLEY PD  
ARRESTING OFFICER: CHAVARRIA,DEGLYS  
ARRESTING OFFICER AGENCY: MEDLEY PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: RIVERA HERNANDEZ

DATE/TIME OF TEST: 06/10/2022 00:58:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:52 |
| Air Blank          | 0.000 | 00:52 |
| Control Test       | 0.080 | 00:52 |
| Air Blank          | 0.000 | 00:53 |
| Subject Sample # 1 | 0.000 | 00:53 |
| Breath Volume      | 2.007 | 00:53 |
| Air Blank          | 0.000 | 00:54 |
| Air Blank          | 0.000 | 00:56 |
| Subject Sample # 2 | 0.000 | 00:56 |
| Breath Volume      | 2.421 | 00:56 |
| Air Blank          | 0.000 | 00:57 |
| Control Test       | 0.080 | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Diagnostics Check  | OK    | 00:58 |

---

Data Download Date/Time: 07/12/2022 04:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAINTILIEEN,BOB,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SAINTILIEEN,BOB  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: LOPEZMURILLO

DATE/TIME OF TEST: 06/11/2022 08:39:06

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 08:34 |            |
| Air Blank         | 0.000     | 08:34 |            |
| Control Test      | 0.080     | 08:34 |            |
| Air Blank         | 0.000     | 08:35 |            |
| Subject Sample    | # 1 0.260 | 08:35 |            |
| Breath Volume     | 1.468     | 08:35 |            |
| Air Blank         | 0.000     | 08:36 |            |
| Air Blank         |           | 08:37 | RFI Detect |
| Air Blank         | 0.000     | 08:38 |            |
| Control Test      | 0.081     | 08:38 |            |
| Air Blank         |           | 08:38 | RFI Detect |
| Air Blank         |           | 08:39 | RFI Detect |

---

Data Download Date/Time: 07/12/2022 04:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAINTILIEEN,BOB,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SAINTILIEEN,BOB  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: LOPEZMURILLO

DATE/TIME OF TEST: 06/11/2022 08:48:20

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 08:45 |            |
| Air Blank         | 0.000 | 08:45 |            |
| Control Test      | 0.080 | 08:45 |            |
| Air Blank         | 0.000 | 08:46 |            |
| Subject Sample    | # 1   | 08:46 | RFI Detect |
| Breath Volume     | 0.000 | 08:46 |            |
| Air Blank         | 0.000 | 08:47 |            |
| Control Test      | 0.081 | 08:47 |            |
| Air Blank         | 0.000 | 08:48 |            |
| Diagnostics Check | OK    | 08:48 |            |

---

Data Download Date/Time: 07/12/2022 04:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAINTILIEN,BOB,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SAINTILIEN,BOB  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: LOPEZMURILLO

DATE/TIME OF TEST: 06/11/2022 08:52:28

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 08:51 |            |
| Air Blank         |       | 08:51 | RFI Detect |
| Air Blank         | 0.000 | 08:52 |            |

---

Data Download Date/Time: 07/12/2022 04:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAINTILIEEN,BOB,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SAINTILIEEN,BOB  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: LOPEZMURILLO

DATE/TIME OF TEST: 06/11/2022 09:01:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:55 |
| Air Blank          | 0.000 | 08:55 |
| Control Test       | 0.080 | 08:55 |
| Air Blank          | 0.000 | 08:56 |
| Subject Sample # 1 | 0.235 | 08:57 |
| Breath Volume      | 1.984 | 08:57 |
| Air Blank          | 0.000 | 08:57 |
| Air Blank          | 0.000 | 08:59 |
| Subject Sample # 2 | 0.246 | 08:59 |
| Breath Volume      | 2.855 | 08:59 |
| Air Blank          | 0.000 | 09:00 |
| Control Test       | 0.080 | 09:01 |
| Air Blank          | 0.000 | 09:01 |
| Diagnostics Check  | OK    | 09:01 |

---

Data Download Date/Time: 07/12/2022 04:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYER,ALYSSA,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: AZEEZ,HAKEEM  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: ISIDORE

DATE/TIME OF TEST: 06/13/2022 04:58:47

---

|                   |           |       |               |
|-------------------|-----------|-------|---------------|
| Diagnostics Check | OK        | 04:52 |               |
| Air Blank         | 0.000     | 04:52 |               |
| Control Test      | 0.079     | 04:53 |               |
| Air Blank         | 0.000     | 04:53 |               |
| Subject Sample    | # 1 0.233 | 04:54 |               |
| Breath Volume     | 2.277     | 04:54 |               |
| Air Blank         | 0.000     | 04:55 |               |
| Air Blank         | 0.000     | 04:56 |               |
| Subject Sample    | # 2       | 04:56 | Slope Not Met |
| Breath Volume     | 1.132     | 04:56 |               |
| Air Blank         | 0.000     | 04:57 |               |
| Control Test      | 0.076     | 04:58 |               |
| Air Blank         | 0.000     | 04:58 |               |
| Diagnostics Check | OK        | 04:58 |               |

---

Data Download Date/Time: 07/12/2022 04:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYER,ALYSSA,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: MOYER,ALYSSA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 06/18/2022 04:38:32

---

|                   |    |       |                 |
|-------------------|----|-------|-----------------|
| Diagnostics Check | OK | 04:37 |                 |
| Air Blank         |    | 04:38 | Improper Sample |
| Air Blank         |    | 04:38 | Improper Sample |

---

Data Download Date/Time: 07/12/2022 04:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYER,ALYSSA,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: MOYER,ALYSSA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 06/18/2022 04:41:07

---

|                   |    |       |                 |
|-------------------|----|-------|-----------------|
| Diagnostics Check | OK | 04:40 |                 |
| Air Blank         |    | 04:40 | Improper Sample |
| Air Blank         |    | 04:41 | Improper Sample |

---

Data Download Date/Time: 07/12/2022 04:21



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYER,ALYSSA,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: MOYER,ALYSSA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GURRERO

DATE/TIME OF TEST: 06/18/2022 04:44:11

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 04:43 |                 |
| Air Blank         |       | 04:43 | Improper Sample |
| Air Blank         | 0.000 | 04:44 |                 |

---

Data Download Date/Time: 07/12/2022 04:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAVARRIA,DEGLYS,E  
BREATH TEST OPERATOR AGENCY: MEDLEY PD  
ARRESTING OFFICER: CHAVARRIA,DEGLYS  
ARRESTING OFFICER AGENCY: MEDLEY PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: MOREIRA

DATE/TIME OF TEST: 06/18/2022 05:28:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:22 |
| Air Blank          | 0.000 | 05:23 |
| Control Test       | 0.078 | 05:23 |
| Air Blank          | 0.000 | 05:24 |
| Subject Sample # 1 | 0.000 | 05:24 |
| Breath Volume      | 2.050 | 05:24 |
| Air Blank          | 0.000 | 05:24 |
| Air Blank          | 0.000 | 05:26 |
| Subject Sample # 2 | 0.000 | 05:27 |
| Breath Volume      | 2.343 | 05:27 |
| Air Blank          | 0.000 | 05:27 |
| Control Test       | 0.079 | 05:28 |
| Air Blank          | 0.000 | 05:28 |
| Diagnostics Check  | OK    | 05:28 |

---

Data Download Date/Time: 07/12/2022 04:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYER,ALYSSA,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: MOYER,ALYSSA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 06/20/2022 00:10:18

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 00:09 |                           |
| Air Blank         | 0.000 | 00:09 |                           |
| Control Test      | 0.000 | 00:09 | Control Outside Tolerance |
| Air Blank         |       | 00:10 | Improper Sample           |

---

Data Download Date/Time: 07/12/2022 04:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYER,ALYSSA,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: MOYER,ALYSSA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 06/20/2022 00:14:19

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 00:13 |                 |
| Air Blank         |       | 00:13 | Improper Sample |
| Air Blank         | 0.000 | 00:14 |                 |

---

Data Download Date/Time: 07/12/2022 04:21

# **Florida Department of Law Enforcement**

## **Alcohol Testing Program**

### **Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYER,ALYSSA,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: MOYER,ALYSSA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 06/20/2022 00:18:52

---

|                   |    |       |                 |
|-------------------|----|-------|-----------------|
| Diagnostics Check | OK | 00:18 |                 |
| Air Blank         |    | 00:18 | Improper Sample |
| Air Blank         |    | 00:18 | Improper Sample |

---

Data Download Date/Time: 07/12/2022 04:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYER,ALYSSA,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: MOYER,ALYSSA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 06/20/2022 00:23:35

---

|                   |       |       |              |
|-------------------|-------|-------|--------------|
| Diagnostics Check | OK    | 00:22 |              |
| Air Blank         |       | 00:23 | Ambient Fail |
| Air Blank         | 0.000 | 00:23 |              |

---

Data Download Date/Time: 07/12/2022 04:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,  
BREATH TEST OPERATOR AGENCY: SWEETWATER PD  
ARRESTING OFFICER: FERNANDEZ,ORLANDO  
ARRESTING OFFICER AGENCY: SWEETWATER PD  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: LANZAS

DATE/TIME OF TEST: 06/25/2022 04:18:58

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:15 |                      |
| Air Blank         | 0.000 | 04:16 |                      |
| Control Test      | 0.080 | 04:16 |                      |
| Air Blank         | 0.000 | 04:17 |                      |
| Subject Sample    | # 1   | 04:17 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:17 |                      |
| Air Blank         | 0.000 | 04:18 |                      |
| Control Test      | 0.080 | 04:18 |                      |
| Air Blank         | 0.000 | 04:18 |                      |
| Diagnostics Check | OK    | 04:18 |                      |

---

Data Download Date/Time: 07/12/2022 04:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERO,NATALIE,  
BREATH TEST OPERATOR AGENCY: OTHER LOCAL  
ARRESTING OFFICER: RIVERO,NATALIE  
ARRESTING OFFICER AGENCY: OTHER LOCAL  
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 06/26/2022 02:17:07

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 02:05 |                |
| Air Blank         | 0.000     | 02:05 |                |
| Control Test      | 0.080     | 02:05 |                |
| Air Blank         | 0.000     | 02:06 |                |
| Subject Sample    | # 1 0.204 | 02:09 | Volume Not Met |
| Breath Volume     | 0.070     | 02:09 |                |
| Air Blank         | 0.000     | 02:10 |                |
| Air Blank         | 0.000     | 02:12 |                |
| Subject Sample    | # 2 0.182 | 02:15 | Volume Not Met |
| Breath Volume     | 0.066     | 02:15 |                |
| Air Blank         | 0.000     | 02:16 |                |
| Control Test      | 0.077     | 02:16 |                |
| Air Blank         | 0.000     | 02:17 |                |
| Diagnostics Check | OK        | 02:17 |                |

---

Data Download Date/Time: 07/12/2022 04:21



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOROZOV,S  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: KIMBLE

DATE/TIME OF TEST: 06/05/2022 04:20:22

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:17 |                      |
| Air Blank         | 0.000 | 04:17 |                      |
| Control Test      | 0.077 | 04:18 |                      |
| Air Blank         | 0.000 | 04:18 |                      |
| Subject Sample    | # 1   | 04:19 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:19 |                      |
| Air Blank         | 0.000 | 04:19 |                      |
| Control Test      | 0.077 | 04:19 |                      |
| Air Blank         | 0.000 | 04:20 |                      |
| Diagnostics Check | OK    | 04:20 |                      |

---

Data Download Date/Time: 06/23/2022 18:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEURER,D,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BELLINGER

DATE/TIME OF TEST: 06/13/2022 02:55:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.077 | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Subject Sample # 1 | 0.201 | 02:50 |
| Breath Volume      | 1.996 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Air Blank          | 0.000 | 02:53 |
| Subject Sample # 2 | 0.202 | 02:53 |
| Breath Volume      | 1.253 | 02:53 |
| Air Blank          | 0.000 | 02:54 |
| Control Test       | 0.078 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Diagnostics Check  | OK    | 02:55 |

---

Data Download Date/Time: 06/24/2022 17:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MARINO,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: N/A

DATE/TIME OF TEST: 06/13/2022 04:11:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:04 |
| Air Blank          | 0.000 | 04:05 |
| Control Test       | 0.078 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Subject Sample # 1 | 0.140 | 04:06 |
| Breath Volume      | 1.523 | 04:06 |
| Air Blank          | 0.000 | 04:07 |
| Air Blank          | 0.000 | 04:09 |
| Subject Sample # 2 | 0.138 | 04:10 |
| Breath Volume      | 1.437 | 04:10 |
| Air Blank          | 0.000 | 04:10 |
| Control Test       | 0.078 | 04:11 |
| Air Blank          | 0.000 | 04:11 |
| Diagnostics Check  | OK    | 04:11 |

---

Data Download Date/Time: 06/24/2022 17:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNS,D,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: DIAZ PORTILLIO

DATE/TIME OF TEST: 06/23/2022 00:15:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:09 |
| Air Blank          | 0.000 | 00:09 |
| Control Test       | 0.078 | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Subject Sample # 1 | 0.083 | 00:11 |
| Breath Volume      | 2.425 | 00:11 |
| Air Blank          | 0.000 | 00:11 |
| Air Blank          | 0.000 | 00:13 |
| Subject Sample # 2 | 0.081 | 00:14 |
| Breath Volume      | 1.578 | 00:14 |
| Air Blank          | 0.000 | 00:14 |
| Control Test       | 0.078 | 00:14 |
| Air Blank          | 0.000 | 00:15 |
| Diagnostics Check  | OK    | 00:15 |

---

Data Download Date/Time: 06/24/2022 17:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 06/26/2022 00:03:34

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:00 |                      |
| Air Blank         | 0.000 | 00:01 |                      |
| Control Test      | 0.077 | 00:01 |                      |
| Air Blank         | 0.000 | 00:02 |                      |
| Subject Sample    | # 1   | 00:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:02 |                      |
| Air Blank         | 0.000 | 00:02 |                      |
| Control Test      | 0.077 | 00:02 |                      |
| Air Blank         | 0.000 | 00:03 |                      |
| Diagnostics Check | OK    | 00:03 |                      |

---

Data Download Date/Time: 07/22/2022 12:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: RIVERA MANCIA

DATE/TIME OF TEST: 06/26/2022 03:22:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.077 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Subject Sample # 1 | 0.165 | 03:17 |
| Breath Volume      | 1.617 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 2 | 0.158 | 03:20 |
| Breath Volume      | 1.339 | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.078 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Diagnostics Check  | OK    | 03:22 |

---

Data Download Date/Time: 07/22/2022 12:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODERICK,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 06/26/2022 05:21:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:14 |
| Air Blank          | 0.000 | 05:15 |
| Control Test       | 0.078 | 05:15 |
| Air Blank          | 0.000 | 05:16 |
| Subject Sample # 1 | 0.148 | 05:16 |
| Breath Volume      | 1.746 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Air Blank          | 0.000 | 05:19 |
| Subject Sample # 2 | 0.145 | 05:20 |
| Breath Volume      | 1.769 | 05:20 |
| Air Blank          | 0.000 | 05:20 |
| Control Test       | 0.077 | 05:20 |
| Air Blank          | 0.000 | 05:21 |
| Diagnostics Check  | OK    | 05:21 |

---

Data Download Date/Time: 07/22/2022 12:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHISMAN,N  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SURRANO

DATE/TIME OF TEST: 06/01/2022 01:28:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:25 |                      |
| Air Blank         | 0.000 | 01:26 |                      |
| Control Test      | 0.078 | 01:26 |                      |
| Air Blank         | 0.000 | 01:26 |                      |
| Subject Sample    | # 1   | 01:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:27 |                      |
| Air Blank         | 0.000 | 01:27 |                      |
| Control Test      | 0.078 | 01:27 |                      |
| Air Blank         | 0.000 | 01:28 |                      |
| Diagnostics Check | OK    | 01:28 |                      |

---

Data Download Date/Time: 06/01/2022 09:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAGOMENI,G,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MORRISON,M  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: KELLAM

DATE/TIME OF TEST: 06/01/2022 02:52:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:44 |
| Air Blank          | 0.000 | 02:45 |
| Control Test       | 0.077 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Subject Sample # 1 | 0.000 | 02:47 |
| Breath Volume      | 1.953 | 02:47 |
| Air Blank          | 0.000 | 02:48 |
| Air Blank          | 0.000 | 02:50 |
| Subject Sample # 2 | 0.000 | 02:50 |
| Breath Volume      | 1.527 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Control Test       | 0.077 | 02:51 |
| Air Blank          | 0.000 | 02:52 |
| Diagnostics Check  | OK    | 02:52 |

---

Data Download Date/Time: 06/01/2022 09:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: SANTOS SANTIESTEBAN

DATE/TIME OF TEST: 06/01/2022 03:56:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:49 |
| Air Blank          | 0.000 | 03:50 |
| Control Test       | 0.076 | 03:50 |
| Air Blank          | 0.000 | 03:50 |
| Subject Sample # 1 | 0.218 | 03:51 |
| Breath Volume      | 1.160 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Air Blank          | 0.000 | 03:54 |
| Subject Sample # 2 | 0.216 | 03:55 |
| Breath Volume      | 1.300 | 03:55 |
| Air Blank          | 0.000 | 03:55 |
| Control Test       | 0.077 | 03:55 |
| Air Blank          | 0.000 | 03:56 |
| Diagnostics Check  | OK    | 03:56 |

---

Data Download Date/Time: 06/01/2022 09:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/02/2022 00:44:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Control Test       | 0.078 | 00:39 |
| Air Blank          | 0.000 | 00:39 |
| Subject Sample # 1 | 0.065 | 00:40 |
| Breath Volume      | 1.781 | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 2 | 0.065 | 00:43 |
| Breath Volume      | 2.113 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Control Test       | 0.078 | 00:44 |
| Air Blank          | 0.000 | 00:44 |
| Diagnostics Check  | OK    | 00:44 |

---

Data Download Date/Time: 06/23/2022 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ULYSSE,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 06/03/2022 01:59:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Control Test       | 0.078 | 01:53 |
| Air Blank          | 0.000 | 01:53 |
| Subject Sample # 1 | 0.000 | 01:54 |
| Breath Volume      | 3.472 | 01:54 |
| Air Blank          | 0.000 | 01:55 |
| Air Blank          | 0.000 | 01:57 |
| Subject Sample # 2 | 0.000 | 01:57 |
| Breath Volume      | 3.277 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Control Test       | 0.079 | 01:58 |
| Air Blank          | 0.000 | 01:58 |
| Diagnostics Check  | OK    | 01:59 |

---

Data Download Date/Time: 06/23/2022 17:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 06/03/2022 03:45:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:38 |
| Air Blank          | 0.000 | 03:39 |
| Control Test       | 0.077 | 03:39 |
| Air Blank          | 0.000 | 03:40 |
| Subject Sample # 1 | 0.175 | 03:41 |
| Breath Volume      | 2.429 | 03:41 |
| Air Blank          | 0.000 | 03:41 |
| Air Blank          | 0.000 | 03:43 |
| Subject Sample # 2 | 0.174 | 03:44 |
| Breath Volume      | 1.886 | 03:44 |
| Air Blank          | 0.000 | 03:44 |
| Control Test       | 0.077 | 03:44 |
| Air Blank          | 0.000 | 03:45 |
| Diagnostics Check  | OK    | 03:45 |

---

Data Download Date/Time: 06/23/2022 17:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: NEMETZ

DATE/TIME OF TEST: 06/03/2022 20:49:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:43 |
| Air Blank          | 0.000 | 20:44 |
| Control Test       | 0.077 | 20:44 |
| Air Blank          | 0.000 | 20:44 |
| Subject Sample # 1 | 0.098 | 20:45 |
| Breath Volume      | 1.988 | 20:45 |
| Air Blank          | 0.000 | 20:45 |
| Air Blank          | 0.000 | 20:47 |
| Subject Sample # 2 | 0.093 | 20:48 |
| Breath Volume      | 1.847 | 20:48 |
| Air Blank          | 0.000 | 20:48 |
| Control Test       | 0.077 | 20:49 |
| Air Blank          | 0.000 | 20:49 |
| Diagnostics Check  | OK    | 20:49 |

---

Data Download Date/Time: 06/23/2022 17:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 06/04/2022 02:34:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:32 |                      |
| Air Blank         | 0.000 | 02:32 |                      |
| Control Test      | 0.078 | 02:32 |                      |
| Air Blank         | 0.000 | 02:33 |                      |
| Subject Sample    | # 1   | 02:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:33 |                      |
| Air Blank         | 0.000 | 02:33 |                      |
| Control Test      | 0.078 | 02:34 |                      |
| Air Blank         | 0.000 | 02:34 |                      |
| Diagnostics Check | OK    | 02:34 |                      |

---

Data Download Date/Time: 06/23/2022 17:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CONRAD,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: SWAIN

DATE/TIME OF TEST: 06/04/2022 04:10:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:03 |
| Air Blank          | 0.000 | 04:04 |
| Control Test       | 0.077 | 04:04 |
| Air Blank          | 0.000 | 04:05 |
| Subject Sample # 1 | 0.038 | 04:05 |
| Breath Volume      | 1.597 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Air Blank          | 0.000 | 04:08 |
| Subject Sample # 2 | 0.038 | 04:09 |
| Breath Volume      | 1.992 | 04:09 |
| Air Blank          | 0.000 | 04:09 |
| Control Test       | 0.078 | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Diagnostics Check  | OK    | 04:10 |

---

Data Download Date/Time: 06/23/2022 17:34



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/05/2022 01:53:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Control Test       | 0.078 | 01:46 |
| Air Blank          | 0.000 | 01:46 |
| Subject Sample # 1 | 0.210 | 01:48 |
| Breath Volume      | 1.707 | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Subject Sample # 2 | 0.224 | 01:51 |
| Breath Volume      | 2.398 | 01:51 |
| Air Blank          | 0.000 | 01:52 |
| Control Test       | 0.076 | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Diagnostics Check  | OK    | 01:53 |

---

Data Download Date/Time: 06/23/2022 17:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/05/2022 03:38:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Control Test       | 0.077 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Subject Sample # 1 | 0.169 | 03:32 |
| Breath Volume      | 3.234 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Air Blank          | 0.000 | 03:35 |
| Subject Sample # 2 | 0.159 | 03:36 |
| Breath Volume      | 1.191 | 03:36 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.076 | 03:37 |
| Air Blank          | 0.000 | 03:38 |
| Diagnostics Check  | OK    | 03:38 |

---

Data Download Date/Time: 06/23/2022 17:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSES,L,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VO,S  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: BRUNO

DATE/TIME OF TEST: 06/05/2022 14:33:44

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 14:30 |                      |
| Air Blank         | 0.000 | 14:31 |                      |
| Control Test      | 0.077 | 14:31 |                      |
| Air Blank         | 0.000 | 14:32 |                      |
| Subject Sample    | # 1   | 14:32 | Subject Test Refused |
| Breath Volume     | 0.000 | 14:32 |                      |
| Air Blank         | 0.000 | 14:32 |                      |
| Control Test      | 0.077 | 14:33 |                      |
| Air Blank         | 0.000 | 14:33 |                      |
| Diagnostics Check | OK    | 14:33 |                      |

---

Data Download Date/Time: 06/23/2022 17:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CLARK,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DUPREE

DATE/TIME OF TEST: 06/06/2022 01:53:40

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:50 |                      |
| Air Blank         | 0.000 | 01:51 |                      |
| Control Test      | 0.078 | 01:51 |                      |
| Air Blank         | 0.000 | 01:52 |                      |
| Subject Sample    | # 1   | 01:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:52 |                      |
| Air Blank         | 0.000 | 01:52 |                      |
| Control Test      | 0.078 | 01:53 |                      |
| Air Blank         | 0.000 | 01:53 |                      |
| Diagnostics Check | OK    | 01:53 |                      |

---

Data Download Date/Time: 06/23/2022 17:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PETERSON,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: CLOUD

DATE/TIME OF TEST: 06/06/2022 04:13:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:07 |
| Air Blank          | 0.000 | 04:07 |
| Control Test       | 0.078 | 04:08 |
| Air Blank          | 0.000 | 04:08 |
| Subject Sample # 1 | 0.141 | 04:09 |
| Breath Volume      | 2.062 | 04:09 |
| Air Blank          | 0.000 | 04:09 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 2 | 0.135 | 04:11 |
| Breath Volume      | 1.582 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Control Test       | 0.078 | 04:12 |
| Air Blank          | 0.000 | 04:13 |
| Diagnostics Check  | OK    | 04:13 |

---

Data Download Date/Time: 06/23/2022 17:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HART,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: ROBISON

DATE/TIME OF TEST: 06/07/2022 01:40:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:34 |
| Air Blank          | 0.000 | 01:34 |
| Control Test       | 0.077 | 01:35 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 1 | 0.063 | 01:36 |
| Breath Volume      | 2.335 | 01:36 |
| Air Blank          | 0.000 | 01:36 |
| Air Blank          | 0.000 | 01:38 |
| Subject Sample # 2 | 0.064 | 01:39 |
| Breath Volume      | 2.039 | 01:39 |
| Air Blank          | 0.000 | 01:39 |
| Control Test       | 0.077 | 01:40 |
| Air Blank          | 0.000 | 01:40 |
| Diagnostics Check  | OK    | 01:40 |

---

Data Download Date/Time: 06/23/2022 17:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 06/07/2022 02:39:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.078 | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 1 | 0.220 | 02:35 |
| Breath Volume      | 2.808 | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Subject Sample # 2 | 0.226 | 02:38 |
| Breath Volume      | 3.304 | 02:38 |
| Air Blank          | 0.000 | 02:38 |
| Control Test       | 0.078 | 02:39 |
| Air Blank          | 0.000 | 02:39 |
| Diagnostics Check  | OK    | 02:39 |

---

Data Download Date/Time: 06/23/2022 17:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CLARK,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ESPOSITO

DATE/TIME OF TEST: 06/08/2022 03:07:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:00 |
| Air Blank          | 0.000 | 03:00 |
| Control Test       | 0.078 | 03:01 |
| Air Blank          | 0.000 | 03:01 |
| Subject Sample # 1 | 0.057 | 03:02 |
| Breath Volume      | 1.582 | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Air Blank          | 0.000 | 03:04 |
| Subject Sample # 2 | 0.057 | 03:05 |
| Breath Volume      | 1.433 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Control Test       | 0.078 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Diagnostics Check  | OK    | 03:07 |

---

Data Download Date/Time: 06/23/2022 17:35



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STANLEY,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: BARGER

DATE/TIME OF TEST: 06/09/2022 23:45:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:39 |
| Air Blank          | 0.000 | 23:39 |
| Control Test       | 0.078 | 23:40 |
| Air Blank          | 0.000 | 23:40 |
| Subject Sample # 1 | 0.117 | 23:41 |
| Breath Volume      | 2.375 | 23:41 |
| Air Blank          | 0.000 | 23:41 |
| Air Blank          | 0.000 | 23:43 |
| Subject Sample # 2 | 0.118 | 23:44 |
| Breath Volume      | 1.882 | 23:44 |
| Air Blank          | 0.000 | 23:44 |
| Control Test       | 0.078 | 23:44 |
| Air Blank          | 0.000 | 23:45 |
| Diagnostics Check  | OK    | 23:45 |

---

Data Download Date/Time: 06/23/2022 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: HOCHHAUSER

DATE/TIME OF TEST: 06/10/2022 04:16:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:13 |                      |
| Air Blank         | 0.000 | 04:14 |                      |
| Control Test      | 0.078 | 04:14 |                      |
| Air Blank         | 0.000 | 04:14 |                      |
| Subject Sample    | # 1   | 04:15 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:15 |                      |
| Air Blank         | 0.000 | 04:15 |                      |
| Control Test      | 0.078 | 04:15 |                      |
| Air Blank         | 0.000 | 04:16 |                      |
| Diagnostics Check | OK    | 04:16 |                      |

---

Data Download Date/Time: 06/23/2022 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 06/11/2022 04:12:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Control Test       | 0.078 | 04:06 |
| Air Blank          | 0.000 | 04:06 |
| Subject Sample # 1 | 0.118 | 04:07 |
| Breath Volume      | 1.226 | 04:07 |
| Air Blank          | 0.000 | 04:08 |
| Air Blank          | 0.000 | 04:10 |
| Subject Sample # 2 | 0.119 | 04:10 |
| Breath Volume      | 1.187 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Control Test       | 0.077 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Diagnostics Check  | OK    | 04:12 |

---

Data Download Date/Time: 06/23/2022 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MORLEY,R  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 06/11/2022 19:50:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:47 |                      |
| Air Blank         | 0.000 | 19:48 |                      |
| Control Test      | 0.076 | 19:48 |                      |
| Air Blank         | 0.000 | 19:49 |                      |
| Subject Sample    | # 1   | 19:49 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:49 |                      |
| Air Blank         | 0.000 | 19:49 |                      |
| Control Test      | 0.077 | 19:50 |                      |
| Air Blank         | 0.000 | 19:50 |                      |
| Diagnostics Check | OK    | 19:50 |                      |

---

Data Download Date/Time: 06/23/2022 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: GRIMALDO GOMEZ

DATE/TIME OF TEST: 06/12/2022 03:00:15

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:57 |                      |
| Air Blank         | 0.000 | 02:57 |                      |
| Control Test      | 0.077 | 02:58 |                      |
| Air Blank         | 0.000 | 02:58 |                      |
| Subject Sample    | # 1   | 02:58 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:58 |                      |
| Air Blank         | 0.000 | 02:59 |                      |
| Control Test      | 0.078 | 02:59 |                      |
| Air Blank         | 0.000 | 03:00 |                      |
| Diagnostics Check | OK    | 03:00 |                      |

---

Data Download Date/Time: 06/23/2022 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ULYSSE,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MCREYNOLDS

DATE/TIME OF TEST: 06/12/2022 05:16:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:10 |
| Air Blank          | 0.000 | 05:11 |
| Control Test       | 0.078 | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Subject Sample # 1 | 0.200 | 05:12 |
| Breath Volume      | 1.308 | 05:12 |
| Air Blank          | 0.000 | 05:12 |
| Air Blank          | 0.000 | 05:14 |
| Subject Sample # 2 | 0.199 | 05:15 |
| Breath Volume      | 1.597 | 05:15 |
| Air Blank          | 0.000 | 05:16 |
| Control Test       | 0.078 | 05:16 |
| Air Blank          | 0.000 | 05:16 |
| Diagnostics Check  | OK    | 05:16 |

---

Data Download Date/Time: 06/23/2022 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 06/13/2022 00:55:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:52 |                      |
| Air Blank         | 0.000 | 00:53 |                      |
| Control Test      | 0.078 | 00:53 |                      |
| Air Blank         | 0.000 | 00:54 |                      |
| Subject Sample    | # 1   | 00:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:54 |                      |
| Air Blank         | 0.000 | 00:54 |                      |
| Control Test      | 0.077 | 00:55 |                      |
| Air Blank         | 0.000 | 00:55 |                      |
| Diagnostics Check | OK    | 00:55 |                      |

---

Data Download Date/Time: 06/23/2022 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BALAIS

DATE/TIME OF TEST: 06/13/2022 02:11:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:04 |
| Air Blank          | 0.000 | 02:05 |
| Control Test       | 0.078 | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Subject Sample # 1 | 0.139 | 02:06 |
| Breath Volume      | 1.726 | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Air Blank          | 0.000 | 02:09 |
| Subject Sample # 2 | 0.135 | 02:09 |
| Breath Volume      | 1.890 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Control Test       | 0.078 | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Diagnostics Check  | OK    | 02:11 |

---

Data Download Date/Time: 06/23/2022 17:36



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ULYSSE,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: BRAUN

DATE/TIME OF TEST: 06/14/2022 01:02:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:56 |
| Air Blank          | 0.000 | 00:56 |
| Control Test       | 0.077 | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Subject Sample # 1 | 0.147 | 00:58 |
| Breath Volume      | 2.660 | 00:58 |
| Air Blank          | 0.000 | 00:58 |
| Air Blank          | 0.000 | 01:00 |
| Subject Sample # 2 | 0.154 | 01:01 |
| Breath Volume      | 2.460 | 01:01 |
| Air Blank          | 0.000 | 01:01 |
| Control Test       | 0.077 | 01:02 |
| Air Blank          | 0.000 | 01:02 |
| Diagnostics Check  | OK    | 01:02 |

---

Data Download Date/Time: 06/23/2022 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 06/14/2022 03:19:35

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:16 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Control Test      | 0.077 | 03:17 |                      |
| Air Blank         | 0.000 | 03:18 |                      |
| Subject Sample    | # 1   | 03:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:18 |                      |
| Air Blank         | 0.000 | 03:18 |                      |
| Control Test      | 0.077 | 03:19 |                      |
| Air Blank         | 0.000 | 03:19 |                      |
| Diagnostics Check | OK    | 03:19 |                      |

---

Data Download Date/Time: 06/23/2022 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTON,D,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: OSCHNER,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 10:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:09

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 06/14/2022 12:36:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:29 |
| Air Blank          | 0.000 | 12:29 |
| Control Test       | 0.077 | 12:30 |
| Air Blank          | 0.000 | 12:30 |
| Subject Sample # 1 | 0.187 | 12:31 |
| Breath Volume      | 3.062 | 12:31 |
| Air Blank          | 0.000 | 12:32 |
| Air Blank          | 0.000 | 12:34 |
| Subject Sample # 2 | 0.189 | 12:34 |
| Breath Volume      | 3.277 | 12:34 |
| Air Blank          | 0.000 | 12:35 |
| Control Test       | 0.077 | 12:35 |
| Air Blank          | 0.000 | 12:35 |
| Diagnostics Check  | OK    | 12:36 |

---

Data Download Date/Time: 06/23/2022 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: HERNANDEZ BAUTISTA

DATE/TIME OF TEST: 06/14/2022 22:36:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:29 |
| Air Blank          | 0.000 | 22:30 |
| Control Test       | 0.078 | 22:30 |
| Air Blank          | 0.000 | 22:30 |
| Subject Sample # 1 | 0.076 | 22:31 |
| Breath Volume      | 1.558 | 22:31 |
| Air Blank          | 0.000 | 22:32 |
| Air Blank          | 0.000 | 22:34 |
| Subject Sample # 2 | 0.068 | 22:34 |
| Breath Volume      | 1.207 | 22:34 |
| Air Blank          | 0.000 | 22:35 |
| Control Test       | 0.078 | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Diagnostics Check  | OK    | 22:36 |

---

Data Download Date/Time: 06/23/2022 17:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 06/16/2022 01:39:19

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:36 |                      |
| Air Blank         | 0.000 | 01:36 |                      |
| Control Test      | 0.078 | 01:37 |                      |
| Air Blank         | 0.000 | 01:37 |                      |
| Subject Sample    | # 1   | 01:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:37 |                      |
| Air Blank         | 0.000 | 01:38 |                      |
| Control Test      | 0.077 | 01:38 |                      |
| Air Blank         | 0.000 | 01:39 |                      |
| Diagnostics Check | OK    | 01:39 |                      |

---

Data Download Date/Time: 06/23/2022 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SATHER,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 06/16/2022 05:20:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:13 |
| Air Blank          | 0.000 | 05:13 |
| Control Test       | 0.077 | 05:13 |
| Air Blank          | 0.000 | 05:14 |
| Subject Sample # 1 | 0.135 | 05:15 |
| Breath Volume      | 1.164 | 05:15 |
| Air Blank          | 0.000 | 05:15 |
| Air Blank          | 0.000 | 05:17 |
| Subject Sample # 2 | 0.135 | 05:18 |
| Breath Volume      | 1.546 | 05:18 |
| Air Blank          | 0.000 | 05:19 |
| Control Test       | 0.078 | 05:19 |
| Air Blank          | 0.000 | 05:19 |
| Diagnostics Check  | OK    | 05:20 |

---

Data Download Date/Time: 06/23/2022 17:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/17/2022 01:10:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:04 |
| Air Blank          | 0.000 | 01:05 |
| Control Test       | 0.078 | 01:05 |
| Air Blank          | 0.000 | 01:06 |
| Subject Sample # 1 | 0.078 | 01:06 |
| Breath Volume      | 2.898 | 01:06 |
| Air Blank          | 0.000 | 01:07 |
| Air Blank          | 0.000 | 01:09 |
| Subject Sample # 2 | 0.079 | 01:09 |
| Breath Volume      | 2.148 | 01:09 |
| Air Blank          | 0.000 | 01:10 |
| Control Test       | 0.079 | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Diagnostics Check  | OK    | 01:10 |

---

Data Download Date/Time: 06/23/2022 17:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PETERSON,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BOUIE

DATE/TIME OF TEST: 06/17/2022 03:56:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:54 |                      |
| Air Blank         | 0.000 | 03:54 |                      |
| Control Test      | 0.077 | 03:54 |                      |
| Air Blank         | 0.000 | 03:55 |                      |
| Subject Sample    | # 1   | 03:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:55 |                      |
| Air Blank         | 0.000 | 03:56 |                      |
| Control Test      | 0.079 | 03:56 |                      |
| Air Blank         | 0.000 | 03:56 |                      |
| Diagnostics Check | OK    | 03:56 |                      |

---

Data Download Date/Time: 06/23/2022 17:37



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNSON,G  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PLESSALA

DATE/TIME OF TEST: 06/18/2022 02:43:12

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:40 |                      |
| Air Blank         | 0.000 | 02:40 |                      |
| Control Test      | 0.078 | 02:41 |                      |
| Air Blank         | 0.000 | 02:41 |                      |
| Subject Sample    | # 1   | 02:41 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:41 |                      |
| Air Blank         | 0.000 | 02:42 |                      |
| Control Test      | 0.078 | 02:42 |                      |
| Air Blank         | 0.000 | 02:43 |                      |
| Diagnostics Check | OK    | 02:43 |                      |

---

Data Download Date/Time: 06/23/2022 17:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ANSLEY,A  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 06/18/2022 21:46:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:39 |
| Air Blank          | 0.000 | 21:40 |
| Control Test       | 0.077 | 21:40 |
| Air Blank          | 0.000 | 21:41 |
| Subject Sample # 1 | 0.144 | 21:41 |
| Breath Volume      | 2.011 | 21:41 |
| Air Blank          | 0.000 | 21:42 |
| Air Blank          | 0.000 | 21:44 |
| Subject Sample # 2 | 0.148 | 21:44 |
| Breath Volume      | 2.000 | 21:44 |
| Air Blank          | 0.000 | 21:45 |
| Control Test       | 0.078 | 21:45 |
| Air Blank          | 0.000 | 21:46 |
| Diagnostics Check  | OK    | 21:46 |

---

Data Download Date/Time: 06/23/2022 17:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HAUSLER,T  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: BUI  
ARREST TIME: 20:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/18/2022 23:26:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:20 |
| Air Blank          | 0.000 | 23:20 |
| Control Test       | 0.078 | 23:21 |
| Air Blank          | 0.000 | 23:21 |
| Subject Sample # 1 | 0.069 | 23:22 |
| Breath Volume      | 2.273 | 23:22 |
| Air Blank          | 0.000 | 23:22 |
| Air Blank          | 0.000 | 23:24 |
| Subject Sample # 2 | 0.064 | 23:25 |
| Breath Volume      | 2.101 | 23:25 |
| Air Blank          | 0.000 | 23:25 |
| Control Test       | 0.078 | 23:26 |
| Air Blank          | 0.000 | 23:26 |
| Diagnostics Check  | OK    | 23:26 |

---

Data Download Date/Time: 06/23/2022 17:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 06/19/2022 01:12:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:08 |                      |
| Air Blank         | 0.000 | 01:08 |                      |
| Control Test      | 0.078 | 01:08 |                      |
| Air Blank         | 0.000 | 01:09 |                      |
| Subject Sample    | # 1   | 01:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:10 |                      |
| Air Blank         | 0.000 | 01:11 |                      |
| Control Test      | 0.078 | 01:11 |                      |
| Air Blank         | 0.000 | 01:12 |                      |
| Diagnostics Check | OK    | 01:12 |                      |

---

Data Download Date/Time: 06/23/2022 17:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/19/2022 03:04:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:57 |
| Air Blank          | 0.000 | 02:58 |
| Control Test       | 0.078 | 02:58 |
| Air Blank          | 0.000 | 02:59 |
| Subject Sample # 1 | 0.181 | 02:59 |
| Breath Volume      | 1.714 | 02:59 |
| Air Blank          | 0.000 | 03:00 |
| Air Blank          | 0.000 | 03:02 |
| Subject Sample # 2 | 0.179 | 03:02 |
| Breath Volume      | 1.929 | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Control Test       | 0.078 | 03:03 |
| Air Blank          | 0.000 | 03:04 |
| Diagnostics Check  | OK    | 03:04 |

---

Data Download Date/Time: 06/23/2022 17:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: ALMANZAR ROSA

DATE/TIME OF TEST: 06/19/2022 04:43:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:37 |
| Air Blank          | 0.000 | 04:37 |
| Control Test       | 0.078 | 04:37 |
| Air Blank          | 0.000 | 04:38 |
| Subject Sample # 1 | 0.150 | 04:38 |
| Breath Volume      | 2.218 | 04:38 |
| Air Blank          | 0.000 | 04:39 |
| Air Blank          | 0.000 | 04:41 |
| Subject Sample # 2 | 0.147 | 04:42 |
| Breath Volume      | 1.167 | 04:42 |
| Air Blank          | 0.000 | 04:42 |
| Control Test       | 0.077 | 04:43 |
| Air Blank          | 0.000 | 04:43 |
| Diagnostics Check  | OK    | 04:43 |

---

Data Download Date/Time: 06/23/2022 17:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 06/19/2022 22:12:48

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:09 |                      |
| Air Blank         | 0.000 | 22:10 |                      |
| Control Test      | 0.079 | 22:10 |                      |
| Air Blank         | 0.000 | 22:11 |                      |
| Subject Sample    | # 1   | 22:11 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:11 |                      |
| Air Blank         | 0.000 | 22:11 |                      |
| Control Test      | 0.078 | 22:12 |                      |
| Air Blank         | 0.000 | 22:12 |                      |
| Diagnostics Check | OK    | 22:12 |                      |

---

Data Download Date/Time: 06/23/2022 17:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: BOSS

DATE/TIME OF TEST: 06/19/2022 23:45:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:39 |
| Air Blank          | 0.000 | 23:39 |
| Control Test       | 0.078 | 23:39 |
| Air Blank          | 0.000 | 23:40 |
| Subject Sample # 1 | 0.253 | 23:40 |
| Breath Volume      | 1.507 | 23:40 |
| Air Blank          | 0.000 | 23:41 |
| Air Blank          | 0.000 | 23:43 |
| Subject Sample # 2 | 0.254 | 23:44 |
| Breath Volume      | 1.882 | 23:44 |
| Air Blank          | 0.000 | 23:44 |
| Control Test       | 0.078 | 23:45 |
| Air Blank          | 0.000 | 23:45 |
| Diagnostics Check  | OK    | 23:45 |

---

Data Download Date/Time: 06/23/2022 17:37



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ALSTON

DATE/TIME OF TEST: 06/20/2022 01:44:41

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:41 |                      |
| Air Blank         | 0.000 | 01:42 |                      |
| Control Test      | 0.078 | 01:42 |                      |
| Air Blank         | 0.000 | 01:43 |                      |
| Subject Sample    | # 1   | 01:43 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:43 |                      |
| Air Blank         | 0.000 | 01:43 |                      |
| Control Test      | 0.078 | 01:44 |                      |
| Air Blank         | 0.000 | 01:44 |                      |
| Diagnostics Check | OK    | 01:44 |                      |

---

Data Download Date/Time: 06/23/2022 17:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: GOBEN,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/20/2022 03:24:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:17 |
| Air Blank          | 0.000 | 03:17 |
| Control Test       | 0.078 | 03:18 |
| Air Blank          | 0.000 | 03:18 |
| Subject Sample # 1 | 0.171 | 03:19 |
| Breath Volume      | 1.886 | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 2 | 0.165 | 03:22 |
| Breath Volume      | 1.953 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Control Test       | 0.078 | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Diagnostics Check  | OK    | 03:24 |

---

Data Download Date/Time: 06/23/2022 17:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOCKERS,B  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 06/21/2022 05:39:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:33 |
| Air Blank          | 0.000 | 05:34 |
| Control Test       | 0.078 | 05:34 |
| Air Blank          | 0.000 | 05:34 |
| Subject Sample # 1 | 0.208 | 05:35 |
| Breath Volume      | 1.824 | 05:35 |
| Air Blank          | 0.000 | 05:35 |
| Air Blank          | 0.000 | 05:37 |
| Subject Sample # 2 | 0.218 | 05:38 |
| Breath Volume      | 3.082 | 05:38 |
| Air Blank          | 0.000 | 05:38 |
| Control Test       | 0.079 | 05:39 |
| Air Blank          | 0.000 | 05:39 |
| Diagnostics Check  | OK    | 05:39 |

---

Data Download Date/Time: 06/23/2022 17:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER,P,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STANLEY,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 13:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:43

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 06/21/2022 16:14:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:03 |
| Air Blank          | 0.000 | 16:03 |
| Control Test       | 0.077 | 16:04 |
| Air Blank          | 0.000 | 16:04 |
| Subject Sample # 1 | 0.302 | 16:05 |
| Breath Volume      | 2.714 | 16:05 |
| Air Blank          | 0.000 | 16:05 |
| Air Blank          | 0.000 | 16:07 |
| Subject Sample # 2 | 0.251 | 16:08 |
| Breath Volume      | 1.125 | 16:08 |
| Air Blank          | 0.000 | 16:08 |
| Air Blank          | 0.000 | 16:10 |
| Subject Sample # 3 | 0.257 | 16:12 |
| Breath Volume      | 1.937 | 16:12 |
| Air Blank          | 0.000 | 16:13 |
| Control Test       | 0.077 | 16:13 |
| Air Blank          | 0.000 | 16:13 |
| Diagnostics Check  | OK    | 16:14 |

---

Data Download Date/Time: 06/23/2022 17:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: MOONEY

DATE/TIME OF TEST: 06/21/2022 20:57:14

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 20:45 |                |
| Air Blank         | 0.000     | 20:46 |                |
| Control Test      | 0.077     | 20:46 |                |
| Air Blank         | 0.000     | 20:46 |                |
| Subject Sample    | # 1 0.109 | 20:50 | Volume Not Met |
| Breath Volume     | 0.218     | 20:50 |                |
| Air Blank         | 0.000     | 20:50 |                |
| Air Blank         | 0.000     | 20:52 |                |
| Subject Sample    | # 2 0.144 | 20:55 | Volume Not Met |
| Breath Volume     | 0.000     | 20:55 |                |
| Air Blank         | 0.000     | 20:56 |                |
| Control Test      | 0.077     | 20:56 |                |
| Air Blank         | 0.000     | 20:57 |                |
| Diagnostics Check | OK        | 20:57 |                |

---

Data Download Date/Time: 06/23/2022 17:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 06/21/2022 22:17:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:11 |
| Air Blank          | 0.000 | 22:11 |
| Control Test       | 0.078 | 22:12 |
| Air Blank          | 0.000 | 22:12 |
| Subject Sample # 1 | 0.018 | 22:13 |
| Breath Volume      | 2.492 | 22:13 |
| Air Blank          | 0.000 | 22:13 |
| Air Blank          | 0.000 | 22:15 |
| Subject Sample # 2 | 0.020 | 22:16 |
| Breath Volume      | 2.261 | 22:16 |
| Air Blank          | 0.000 | 22:16 |
| Control Test       | 0.077 | 22:17 |
| Air Blank          | 0.000 | 22:17 |
| Diagnostics Check  | OK    | 22:17 |

---

Data Download Date/Time: 06/23/2022 17:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JIMENEZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PASCUAL PEREZ

DATE/TIME OF TEST: 06/21/2022 23:12:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Control Test       | 0.077 | 23:06 |
| Air Blank          | 0.000 | 23:07 |
| Subject Sample # 1 | 0.111 | 23:07 |
| Breath Volume      | 2.582 | 23:07 |
| Air Blank          | 0.000 | 23:08 |
| Air Blank          | 0.000 | 23:10 |
| Subject Sample # 2 | 0.109 | 23:10 |
| Breath Volume      | 2.402 | 23:10 |
| Air Blank          | 0.000 | 23:11 |
| Control Test       | 0.078 | 23:11 |
| Air Blank          | 0.000 | 23:12 |
| Diagnostics Check  | OK    | 23:12 |

---

Data Download Date/Time: 06/23/2022 17:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 06/22/2022 00:13:00

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:09 |                      |
| Air Blank         | 0.000 | 00:10 |                      |
| Control Test      | 0.078 | 00:10 |                      |
| Air Blank         | 0.000 | 00:11 |                      |
| Subject Sample    | # 1   | 00:11 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:11 |                      |
| Air Blank         | 0.000 | 00:11 |                      |
| Control Test      | 0.077 | 00:12 |                      |
| Air Blank         |       | 00:12 | RFI Detect           |
| Air Blank         | 0.000 | 00:13 |                      |

---

Data Download Date/Time: 06/23/2022 17:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: WUSSICK

DATE/TIME OF TEST: 06/22/2022 04:32:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:29 |                      |
| Air Blank         | 0.000 | 04:29 |                      |
| Control Test      | 0.078 | 04:30 |                      |
| Air Blank         | 0.000 | 04:30 |                      |
| Subject Sample    | # 1   | 04:30 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:30 |                      |
| Air Blank         | 0.000 | 04:31 |                      |
| Control Test      | 0.078 | 04:31 |                      |
| Air Blank         | 0.000 | 04:32 |                      |
| Diagnostics Check | OK    | 04:32 |                      |

---

Data Download Date/Time: 06/23/2022 17:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 06/24/2022 01:58:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:55 |                      |
| Air Blank         | 0.000 | 01:56 |                      |
| Control Test      | 0.077 | 01:56 |                      |
| Air Blank         | 0.000 | 01:57 |                      |
| Subject Sample    | # 1   | 01:57 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:57 |                      |
| Air Blank         | 0.000 | 01:57 |                      |
| Control Test      | 0.077 | 01:58 |                      |
| Air Blank         | 0.000 | 01:58 |                      |
| Diagnostics Check | OK    | 01:58 |                      |

---

Data Download Date/Time: 07/22/2022 11:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JAMES, ALFONSO,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: SHANNON

DATE/TIME OF TEST: 06/24/2022 02:43:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:37 |
| Air Blank          | 0.000 | 02:37 |
| Control Test       | 0.077 | 02:38 |
| Air Blank          | 0.000 | 02:38 |
| Subject Sample # 1 | 0.146 | 02:39 |
| Breath Volume      | 2.507 | 02:39 |
| Air Blank          | 0.000 | 02:39 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 2 | 0.138 | 02:42 |
| Breath Volume      | 2.558 | 02:42 |
| Air Blank          | 0.000 | 02:42 |
| Control Test       | 0.078 | 02:43 |
| Air Blank          | 0.000 | 02:43 |
| Diagnostics Check  | OK    | 02:43 |

---

Data Download Date/Time: 07/22/2022 11:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNSON,G  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/24/2022 04:00:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Control Test       | 0.078 | 03:54 |
| Air Blank          | 0.000 | 03:55 |
| Subject Sample # 1 | 0.187 | 03:56 |
| Breath Volume      | 1.890 | 03:56 |
| Air Blank          | 0.000 | 03:56 |
| Air Blank          | 0.000 | 03:58 |
| Subject Sample # 2 | 0.194 | 03:59 |
| Breath Volume      | 2.371 | 03:59 |
| Air Blank          | 0.000 | 03:59 |
| Control Test       | 0.077 | 04:00 |
| Air Blank          | 0.000 | 04:00 |
| Diagnostics Check  | OK    | 04:00 |

---

Data Download Date/Time: 07/22/2022 11:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COTHRAN,D,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CAMACHO,F.  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:28

SUBJECT'S LAST NAME: ALBERTA

DATE/TIME OF TEST: 06/24/2022 09:55:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:48 |
| Air Blank          | 0.000 | 09:49 |
| Control Test       | 0.078 | 09:49 |
| Air Blank          | 0.000 | 09:50 |
| Subject Sample # 1 | 0.000 | 09:50 |
| Breath Volume      | 2.726 | 09:50 |
| Air Blank          | 0.000 | 09:51 |
| Air Blank          | 0.000 | 09:53 |
| Subject Sample # 2 | 0.000 | 09:53 |
| Breath Volume      | 4.250 | 09:53 |
| Air Blank          | 0.000 | 09:54 |
| Control Test       | 0.078 | 09:54 |
| Air Blank          | 0.000 | 09:55 |
| Diagnostics Check  | OK    | 09:55 |

---

Data Download Date/Time: 07/22/2022 11:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MAXWELL,T  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HOLDEN

DATE/TIME OF TEST: 06/24/2022 22:57:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:54 |                      |
| Air Blank         | 0.000 | 22:55 |                      |
| Control Test      | 0.078 | 22:55 |                      |
| Air Blank         | 0.000 | 22:56 |                      |
| Subject Sample    | # 1   | 22:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:56 |                      |
| Air Blank         | 0.000 | 22:56 |                      |
| Control Test      | 0.077 | 22:57 |                      |
| Air Blank         | 0.000 | 22:57 |                      |
| Diagnostics Check | OK    | 22:57 |                      |

---

Data Download Date/Time: 07/22/2022 11:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 06/25/2022 02:05:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:57 |
| Air Blank          | 0.000 | 01:57 |
| Control Test       | 0.078 | 01:58 |
| Air Blank          | 0.000 | 01:58 |
| Subject Sample # 1 | 0.079 | 02:00 |
| Breath Volume      | 1.445 | 02:00 |
| Air Blank          | 0.000 | 02:01 |
| Air Blank          | 0.000 | 02:03 |
| Subject Sample # 2 | 0.080 | 02:03 |
| Breath Volume      | 1.757 | 02:03 |
| Air Blank          | 0.000 | 02:04 |
| Control Test       | 0.077 | 02:04 |
| Air Blank          | 0.000 | 02:05 |
| Diagnostics Check  | OK    | 02:05 |

---

Data Download Date/Time: 07/22/2022 11:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 06/25/2022 05:12:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:06 |
| Air Blank          | 0.000 | 05:06 |
| Control Test       | 0.078 | 05:06 |
| Air Blank          | 0.000 | 05:07 |
| Subject Sample # 1 | 0.109 | 05:07 |
| Breath Volume      | 2.144 | 05:07 |
| Air Blank          | 0.000 | 05:08 |
| Air Blank          | 0.000 | 05:10 |
| Subject Sample # 2 | 0.105 | 05:10 |
| Breath Volume      | 1.832 | 05:10 |
| Air Blank          | 0.000 | 05:11 |
| Control Test       | 0.078 | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Diagnostics Check  | OK    | 05:12 |

---

Data Download Date/Time: 07/22/2022 11:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/25/2022 20:50:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:43 |
| Air Blank          | 0.000 | 20:44 |
| Control Test       | 0.078 | 20:44 |
| Air Blank          | 0.000 | 20:44 |
| Subject Sample # 1 | 0.151 | 20:45 |
| Breath Volume      | 1.351 | 20:45 |
| Air Blank          | 0.000 | 20:46 |
| Air Blank          | 0.000 | 20:47 |
| Subject Sample # 2 | 0.151 | 20:48 |
| Breath Volume      | 1.148 | 20:48 |
| Air Blank          | 0.000 | 20:49 |
| Control Test       | 0.076 | 20:49 |
| Air Blank          | 0.000 | 20:50 |
| Diagnostics Check  | OK    | 20:50 |

---

Data Download Date/Time: 07/22/2022 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SAVITSKIY

DATE/TIME OF TEST: 06/25/2022 21:52:40

---

|                   |       |       |                  |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK    | 21:48 |                  |
| Air Blank         | 0.000 | 21:49 |                  |
| Control Test      | 0.078 | 21:49 |                  |
| Air Blank         | 0.000 | 21:50 |                  |
| Subject Sample    | # 1   | 21:52 | Sequence Aborted |
| Breath Volume     | 0.000 | 21:52 |                  |
| Air Blank         | 0.000 | 21:52 |                  |

---

Data Download Date/Time: 07/22/2022 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SAVITSKIY

DATE/TIME OF TEST: 06/25/2022 22:06:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:00 |
| Air Blank          | 0.000 | 22:01 |
| Control Test       | 0.078 | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Subject Sample # 1 | 0.269 | 22:02 |
| Breath Volume      | 2.265 | 22:02 |
| Air Blank          | 0.000 | 22:03 |
| Air Blank          | 0.000 | 22:05 |
| Subject Sample # 2 | 0.269 | 22:05 |
| Breath Volume      | 2.531 | 22:05 |
| Air Blank          | 0.000 | 22:06 |
| Control Test       | 0.077 | 22:06 |
| Air Blank          | 0.000 | 22:06 |
| Diagnostics Check  | OK    | 22:06 |

---

Data Download Date/Time: 07/22/2022 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CLARK,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ECHAVARRIA

DATE/TIME OF TEST: 06/26/2022 03:43:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:36 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.077 | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Subject Sample # 1 | 0.168 | 03:38 |
| Breath Volume      | 1.417 | 03:38 |
| Air Blank          | 0.000 | 03:39 |
| Air Blank          | 0.000 | 03:41 |
| Subject Sample # 2 | 0.155 | 03:41 |
| Breath Volume      | 2.472 | 03:41 |
| Air Blank          | 0.000 | 03:42 |
| Control Test       | 0.077 | 03:42 |
| Air Blank          | 0.000 | 03:43 |
| Diagnostics Check  | OK    | 03:43 |

---

Data Download Date/Time: 07/22/2022 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: CRUM

DATE/TIME OF TEST: 06/26/2022 23:45:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:39 |
| Air Blank          | 0.000 | 23:39 |
| Control Test       | 0.078 | 23:40 |
| Air Blank          | 0.000 | 23:40 |
| Subject Sample # 1 | 0.185 | 23:41 |
| Breath Volume      | 3.265 | 23:41 |
| Air Blank          | 0.000 | 23:41 |
| Air Blank          | 0.000 | 23:43 |
| Subject Sample # 2 | 0.189 | 23:44 |
| Breath Volume      | 2.433 | 23:44 |
| Air Blank          | 0.000 | 23:44 |
| Control Test       | 0.078 | 23:45 |
| Air Blank          | 0.000 | 23:45 |
| Diagnostics Check  | OK    | 23:45 |

---

Data Download Date/Time: 07/22/2022 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 06/27/2022 03:34:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:31 |                      |
| Air Blank         | 0.000 | 03:32 |                      |
| Control Test      | 0.077 | 03:32 |                      |
| Air Blank         | 0.000 | 03:33 |                      |
| Subject Sample    | # 1   | 03:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:33 |                      |
| Air Blank         | 0.000 | 03:33 |                      |
| Control Test      | 0.077 | 03:34 |                      |
| Air Blank         | 0.000 | 03:34 |                      |
| Diagnostics Check | OK    | 03:34 |                      |

---

Data Download Date/Time: 07/22/2022 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: KOVACH

DATE/TIME OF TEST: 06/27/2022 04:46:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:40 |
| Air Blank          | 0.000 | 04:40 |
| Control Test       | 0.078 | 04:40 |
| Air Blank          | 0.000 | 04:41 |
| Subject Sample # 1 | 0.125 | 04:42 |
| Breath Volume      | 1.394 | 04:42 |
| Air Blank          | 0.000 | 04:42 |
| Air Blank          | 0.000 | 04:44 |
| Subject Sample # 2 | 0.124 | 04:45 |
| Breath Volume      | 1.195 | 04:45 |
| Air Blank          | 0.000 | 04:45 |
| Control Test       | 0.077 | 04:45 |
| Air Blank          | 0.000 | 04:46 |
| Diagnostics Check  | OK    | 04:46 |

---

Data Download Date/Time: 07/22/2022 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ECHEVERRIAVIDAL,X  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: VAN HORN

DATE/TIME OF TEST: 06/27/2022 05:31:07

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:28 |                      |
| Air Blank         | 0.000 | 05:28 |                      |
| Control Test      | 0.078 | 05:29 |                      |
| Air Blank         | 0.000 | 05:29 |                      |
| Subject Sample    | # 1   | 05:29 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:29 |                      |
| Air Blank         | 0.000 | 05:30 |                      |
| Control Test      | 0.078 | 05:30 |                      |
| Air Blank         | 0.000 | 05:31 |                      |
| Diagnostics Check | OK    | 05:31 |                      |

---

Data Download Date/Time: 07/22/2022 11:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 06/27/2022 23:06:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:00 |
| Air Blank          | 0.000 | 23:00 |
| Control Test       | 0.078 | 23:01 |
| Air Blank          | 0.000 | 23:01 |
| Subject Sample # 1 | 0.000 | 23:02 |
| Breath Volume      | 3.648 | 23:02 |
| Air Blank          | 0.000 | 23:02 |
| Air Blank          | 0.000 | 23:04 |
| Subject Sample # 2 | 0.000 | 23:05 |
| Breath Volume      | 3.691 | 23:05 |
| Air Blank          | 0.000 | 23:05 |
| Control Test       | 0.078 | 23:06 |
| Air Blank          | 0.000 | 23:06 |
| Diagnostics Check  | OK    | 23:06 |

---

Data Download Date/Time: 07/22/2022 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PETERSON,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: COATS

DATE/TIME OF TEST: 06/28/2022 00:51:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:48 |                      |
| Air Blank         | 0.000 | 00:49 |                      |
| Control Test      | 0.078 | 00:49 |                      |
| Air Blank         | 0.000 | 00:50 |                      |
| Subject Sample    | # 1   | 00:50 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:50 |                      |
| Air Blank         | 0.000 | 00:50 |                      |
| Control Test      | 0.079 | 00:51 |                      |
| Air Blank         | 0.000 | 00:51 |                      |
| Diagnostics Check | OK    | 00:51 |                      |

---

Data Download Date/Time: 07/22/2022 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CLARK,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: KOCI

DATE/TIME OF TEST: 06/28/2022 01:27:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:21 |
| Air Blank          | 0.000 | 01:21 |
| Control Test       | 0.077 | 01:21 |
| Air Blank          | 0.000 | 01:22 |
| Subject Sample # 1 | 0.000 | 01:22 |
| Breath Volume      | 2.808 | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Air Blank          | 0.000 | 01:25 |
| Subject Sample # 2 | 0.000 | 01:25 |
| Breath Volume      | 1.812 | 01:25 |
| Air Blank          | 0.000 | 01:26 |
| Control Test       | 0.078 | 01:26 |
| Air Blank          | 0.000 | 01:27 |
| Diagnostics Check  | OK    | 01:27 |

---

Data Download Date/Time: 07/22/2022 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/28/2022 03:20:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:14 |
| Air Blank          | 0.000 | 03:15 |
| Control Test       | 0.078 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Subject Sample # 1 | 0.000 | 03:16 |
| Breath Volume      | 2.152 | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Air Blank          | 0.000 | 03:18 |
| Subject Sample # 2 | 0.000 | 03:19 |
| Breath Volume      | 2.421 | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Control Test       | 0.078 | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Diagnostics Check  | OK    | 03:20 |

---

Data Download Date/Time: 07/22/2022 11:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MILLER,D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29821080A4  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: PERROTTA

DATE/TIME OF TEST: 06/28/2022 01:59:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Control Test       | 0.080 | 01:53 |
| Air Blank          | 0.000 | 01:54 |
| Subject Sample # 1 | 0.124 | 01:54 |
| Breath Volume      | 2.730 | 01:54 |
| Air Blank          | 0.000 | 01:55 |
| Air Blank          | 0.000 | 01:57 |
| Subject Sample # 2 | 0.127 | 01:57 |
| Breath Volume      | 3.351 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Control Test       | 0.079 | 01:58 |
| Air Blank          | 0.000 | 01:59 |
| Diagnostics Check  | OK    | 01:59 |

---

Data Download Date/Time: 07/22/2022 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29821080A4  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: KHASAWINAH

DATE/TIME OF TEST: 06/28/2022 22:20:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:13 |
| Air Blank          | 0.000 | 22:14 |
| Control Test       | 0.080 | 22:14 |
| Air Blank          | 0.000 | 22:15 |
| Subject Sample # 1 | 0.000 | 22:15 |
| Breath Volume      | 2.652 | 22:15 |
| Air Blank          | 0.000 | 22:16 |
| Air Blank          | 0.000 | 22:18 |
| Subject Sample # 2 | 0.000 | 22:18 |
| Breath Volume      | 2.742 | 22:18 |
| Air Blank          | 0.000 | 22:19 |
| Control Test       | 0.079 | 22:19 |
| Air Blank          | 0.000 | 22:19 |
| Diagnostics Check  | OK    | 22:20 |

---

Data Download Date/Time: 07/22/2022 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STANLEY,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 29821080A4  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 06/29/2022 00:05:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:58 |
| Air Blank          | 0.000 | 23:59 |
| Control Test       | 0.080 | 23:59 |
| Air Blank          | 0.000 | 23:59 |
| Subject Sample # 1 | 0.071 | 00:00 |
| Breath Volume      | 1.605 | 00:00 |
| Air Blank          | 0.000 | 00:01 |
| Air Blank          | 0.000 | 00:03 |
| Subject Sample # 2 | 0.070 | 00:04 |
| Breath Volume      | 1.839 | 00:04 |
| Air Blank          | 0.000 | 00:05 |
| Control Test       | 0.079 | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Diagnostics Check  | OK    | 00:05 |

---

Data Download Date/Time: 07/22/2022 12:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29821080A4  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: CHALABI

DATE/TIME OF TEST: 06/29/2022 04:08:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:02 |
| Air Blank          | 0.000 | 04:03 |
| Control Test       | 0.080 | 04:03 |
| Air Blank          | 0.000 | 04:04 |
| Subject Sample # 1 | 0.043 | 04:04 |
| Breath Volume      | 3.371 | 04:04 |
| Air Blank          | 0.000 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Subject Sample # 2 | 0.041 | 04:07 |
| Breath Volume      | 1.816 | 04:07 |
| Air Blank          | 0.000 | 04:08 |
| Control Test       | 0.079 | 04:08 |
| Air Blank          | 0.000 | 04:08 |
| Diagnostics Check  | OK    | 04:08 |

---

Data Download Date/Time: 07/22/2022 12:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ULYSSE,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29821080A4  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 06/29/2022 19:28:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:22 |
| Air Blank          | 0.000 | 19:22 |
| Control Test       | 0.080 | 19:23 |
| Air Blank          | 0.000 | 19:23 |
| Subject Sample # 1 | 0.000 | 19:24 |
| Breath Volume      | 2.820 | 19:24 |
| Air Blank          | 0.000 | 19:24 |
| Air Blank          | 0.000 | 19:26 |
| Subject Sample # 2 | 0.000 | 19:27 |
| Breath Volume      | 2.726 | 19:27 |
| Air Blank          | 0.000 | 19:27 |
| Control Test       | 0.079 | 19:28 |
| Air Blank          | 0.000 | 19:28 |
| Diagnostics Check  | OK    | 19:28 |

---

Data Download Date/Time: 07/22/2022 12:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29821080A4  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: TAMEZ GARCIA

DATE/TIME OF TEST: 06/29/2022 21:44:18

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:41 |                      |
| Air Blank         | 0.000 | 21:41 |                      |
| Control Test      | 0.080 | 21:42 |                      |
| Air Blank         | 0.000 | 21:42 |                      |
| Subject Sample    | # 1   | 21:42 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:42 |                      |
| Air Blank         | 0.000 | 21:43 |                      |
| Control Test      | 0.080 | 21:43 |                      |
| Air Blank         | 0.000 | 21:44 |                      |
| Diagnostics Check | OK    | 21:44 |                      |

---

Data Download Date/Time: 07/22/2022 12:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29821080A4  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/30/2022 01:42:29

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:39 |                      |
| Air Blank         | 0.000 | 01:40 |                      |
| Control Test      | 0.080 | 01:40 |                      |
| Air Blank         | 0.000 | 01:40 |                      |
| Subject Sample    | # 1   | 01:41 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:41 |                      |
| Air Blank         | 0.000 | 01:41 |                      |
| Control Test      | 0.079 | 01:41 |                      |
| Air Blank         | 0.000 | 01:42 |                      |
| Diagnostics Check | OK    | 01:42 |                      |

---

Data Download Date/Time: 07/22/2022 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,J,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS,W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MASELY

DATE/TIME OF TEST: 06/01/2022 03:38:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.078 | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 1 | 0.162 | 03:33 |
| Breath Volume      | 3.648 | 03:33 |
| Air Blank          | 0.000 | 03:34 |
| Air Blank          | 0.000 | 03:36 |
| Subject Sample # 2 | 0.165 | 03:37 |
| Breath Volume      | 4.914 | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.076 | 03:38 |
| Air Blank          | 0.000 | 03:38 |
| Diagnostics Check  | OK    | 03:38 |

---

Data Download Date/Time: 06/18/2022 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICKETT,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: GOETHE,R  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: EMERY

DATE/TIME OF TEST: 06/04/2022 04:06:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Control Test       | 0.078 | 04:01 |
| Air Blank          | 0.000 | 04:01 |
| Subject Sample # 1 | 0.212 | 04:02 |
| Breath Volume      | 1.953 | 04:02 |
| Air Blank          | 0.000 | 04:03 |
| Air Blank          | 0.000 | 04:04 |
| Subject Sample # 2 | 0.212 | 04:05 |
| Breath Volume      | 2.558 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Control Test       | 0.076 | 04:06 |
| Air Blank          | 0.000 | 04:06 |
| Diagnostics Check  | OK    | 04:06 |

---

Data Download Date/Time: 06/18/2022 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,J,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS,W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 06/04/2022 23:12:53

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:59 |                |
| Air Blank         | 0.000     | 22:59 |                |
| Control Test      | 0.078     | 23:00 |                |
| Air Blank         | 0.000     | 23:00 |                |
| Subject Sample    | # 1 0.190 | 23:03 | Volume Not Met |
| Breath Volume     | 0.195     | 23:03 |                |
| Air Blank         | 0.000     | 23:04 |                |
| Air Blank         | 0.000     | 23:06 |                |
| Subject Sample    | # 2 0.174 | 23:07 |                |
| Breath Volume     | 1.328     | 23:07 |                |
| Air Blank         | 0.000     | 23:07 |                |
| Air Blank         | 0.000     | 23:09 |                |
| Subject Sample    | # 3 0.174 | 23:11 |                |
| Breath Volume     | 1.398     | 23:11 |                |
| Air Blank         | 0.000     | 23:11 |                |
| Control Test      | 0.077     | 23:12 |                |
| Air Blank         | 0.000     | 23:12 |                |
| Diagnostics Check | OK        | 23:12 |                |

---

Data Download Date/Time: 06/18/2022 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,J,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS,W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: MCELROY

DATE/TIME OF TEST: 06/05/2022 03:06:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:57 |
| Air Blank          | 0.000 | 02:57 |
| Control Test       | 0.078 | 02:58 |
| Air Blank          | 0.000 | 02:58 |
| Subject Sample # 1 | 0.108 | 03:00 |
| Breath Volume      | 4.359 | 03:00 |
| Air Blank          | 0.000 | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Subject Sample # 2 | 0.104 | 03:04 |
| Breath Volume      | 1.761 | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Control Test       | 0.078 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Diagnostics Check  | OK    | 03:06 |

---

Data Download Date/Time: 06/18/2022 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICKETT,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HOLTON,G  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 06/10/2022 01:15:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:08 |
| Air Blank          | 0.000 | 01:08 |
| Control Test       | 0.078 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Subject Sample # 1 | 0.185 | 01:10 |
| Breath Volume      | 2.156 | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Air Blank          | 0.000 | 01:12 |
| Subject Sample # 2 | 0.184 | 01:13 |
| Breath Volume      | 2.410 | 01:13 |
| Air Blank          | 0.000 | 01:14 |
| Control Test       | 0.077 | 01:14 |
| Air Blank          | 0.000 | 01:15 |
| Diagnostics Check  | OK    | 01:15 |

---

Data Download Date/Time: 06/18/2022 13:37



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNEAD,C,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: GORDON,G  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WHEIHE

DATE/TIME OF TEST: 06/12/2022 03:22:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:15 |
| Air Blank          | 0.000 | 03:15 |
| Control Test       | 0.078 | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Subject Sample # 1 | 0.239 | 03:17 |
| Breath Volume      | 1.441 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 2 | 0.231 | 03:21 |
| Breath Volume      | 1.589 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.076 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Diagnostics Check  | OK    | 03:22 |

---

Data Download Date/Time: 06/18/2022 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON, M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: DILL, T  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: KUHN

DATE/TIME OF TEST: 06/14/2022 02:41:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.078 | 02:34 |
| Air Blank          | 0.000 | 02:35 |
| Subject Sample # 1 | 0.224 | 02:36 |
| Breath Volume      | 1.582 | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Air Blank          | 0.000 | 02:38 |
| Subject Sample # 2 | 0.239 | 02:39 |
| Breath Volume      | 3.835 | 02:39 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.076 | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Diagnostics Check  | OK    | 02:41 |

---

Data Download Date/Time: 06/18/2022 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,J,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: MILLAN,S  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 06/15/2022 00:03:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:56 |
| Air Blank          | 0.000 | 23:56 |
| Control Test       | 0.078 | 23:57 |
| Air Blank          | 0.000 | 23:57 |
| Subject Sample # 2 | 0.185 | 23:58 |
| Breath Volume      | 2.949 | 23:58 |
| Air Blank          | 0.000 | 23:59 |
| Air Blank          | 0.000 | 00:01 |
| Subject Sample # 1 | 0.180 | 00:01 |
| Breath Volume      | 3.304 | 00:01 |
| Air Blank          | 0.000 | 00:02 |
| Control Test       | 0.077 | 00:02 |
| Air Blank          | 0.000 | 00:03 |
| Diagnostics Check  | OK    | 00:03 |

---

Data Download Date/Time: 06/18/2022 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICKETT,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: GOETHE,R  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: RAWLS

DATE/TIME OF TEST: 06/19/2022 02:52:50

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:49 |                      |
| Air Blank         | 0.000 | 02:50 |                      |
| Control Test      | 0.079 | 02:50 |                      |
| Air Blank         | 0.000 | 02:51 |                      |
| Subject Sample    | # 1   | 02:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:51 |                      |
| Air Blank         | 0.000 | 02:51 |                      |
| Control Test      | 0.079 | 02:52 |                      |
| Air Blank         | 0.000 | 02:52 |                      |
| Diagnostics Check | OK    | 02:52 |                      |

---

Data Download Date/Time: 08/02/2022 20:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICKETT,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS,W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: LEMUS

DATE/TIME OF TEST: 06/19/2022 04:14:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:05 |
| Air Blank          | 0.000 | 04:05 |
| Control Test       | 0.078 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Subject Sample # 1 | 0.117 | 04:08 |
| Breath Volume      | 1.972 | 04:08 |
| Air Blank          | 0.000 | 04:09 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 2 | 0.123 | 04:13 |
| Breath Volume      | 2.250 | 04:13 |
| Air Blank          | 0.000 | 04:13 |
| Control Test       | 0.077 | 04:14 |
| Air Blank          | 0.000 | 04:14 |
| Diagnostics Check  | OK    | 04:14 |

---

Data Download Date/Time: 08/02/2022 20:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNEAD,C,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HOLTON,G  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 06/20/2022 22:48:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:42 |
| Air Blank          | 0.000 | 22:42 |
| Control Test       | 0.078 | 22:42 |
| Air Blank          | 0.000 | 22:43 |
| Subject Sample # 1 | 0.236 | 22:43 |
| Breath Volume      | 2.328 | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Air Blank          | 0.000 | 22:46 |
| Subject Sample # 2 | 0.219 | 22:46 |
| Breath Volume      | 1.726 | 22:46 |
| Air Blank          | 0.000 | 22:47 |
| Control Test       | 0.078 | 22:47 |
| Air Blank          | 0.000 | 22:48 |
| Diagnostics Check  | OK    | 22:48 |

---

Data Download Date/Time: 08/02/2022 20:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAL,JACOB,N  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: BARNES,D  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 06/24/2022 03:16:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:09 |
| Air Blank          | 0.000 | 03:09 |
| Control Test       | 0.079 | 03:10 |
| Air Blank          | 0.000 | 03:10 |
| Subject Sample # 1 | 0.000 | 03:11 |
| Breath Volume      | 2.894 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Air Blank          | 0.000 | 03:14 |
| Subject Sample # 2 | 0.000 | 03:14 |
| Breath Volume      | 2.308 | 03:14 |
| Air Blank          | 0.000 | 03:15 |
| Control Test       | 0.076 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Diagnostics Check  | OK    | 03:16 |

---

Data Download Date/Time: 08/02/2022 20:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLOUNT,J,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS,S  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: HANLY

DATE/TIME OF TEST: 06/24/2022 19:43:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:33 |
| Air Blank          | 0.000 | 19:33 |
| Control Test       | 0.077 | 19:34 |
| Air Blank          | 0.000 | 19:34 |
| Subject Sample # 1 | 0.246 | 19:35 |
| Breath Volume      | 2.191 | 19:35 |
| Air Blank          | 0.000 | 19:36 |
| Air Blank          | 0.000 | 19:38 |
| Subject Sample # 2 | 0.268 | 19:38 |
| Breath Volume      | 1.503 | 19:38 |
| Air Blank          | 0.000 | 19:39 |
| Air Blank          | 0.000 | 19:41 |
| Subject Sample # 3 | 0.263 | 19:42 |
| Breath Volume      | 2.191 | 19:42 |
| Air Blank          | 0.000 | 19:42 |
| Control Test       | 0.077 | 19:42 |
| Air Blank          | 0.000 | 19:43 |
| Diagnostics Check  | OK    | 19:43 |

---

Data Download Date/Time: 08/02/2022 20:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNEAD,CODY,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WEIGLE,R  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CHIVERS

DATE/TIME OF TEST: 06/26/2022 03:01:16

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 02:54 |                           |
| Air Blank         | 0.000     | 02:55 |                           |
| Control Test      | 0.078     | 02:55 |                           |
| Air Blank         | 0.000     | 02:55 |                           |
| Subject Sample    | # 1 0.137 | 02:56 |                           |
| Breath Volume     | 1.917     | 02:56 |                           |
| Air Blank         | 0.000     | 02:57 |                           |
| Air Blank         | 0.000     | 02:58 |                           |
| Subject Sample    | # 2 0.127 | 02:59 |                           |
| Breath Volume     | 1.789     | 02:59 |                           |
| Air Blank         | 0.000     | 03:00 |                           |
| Control Test      | 0.074     | 03:00 | Control Outside Tolerance |
| Air Blank         | 0.000     | 03:01 |                           |

---

Data Download Date/Time: 08/02/2022 20:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNEAD,CODY,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WEIGLE,R  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CHIVER

DATE/TIME OF TEST: 06/26/2022 03:18:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:12 |
| Air Blank          | 0.000 | 03:12 |
| Control Test       | 0.078 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Subject Sample # 1 | 0.125 | 03:14 |
| Breath Volume      | 1.539 | 03:14 |
| Air Blank          | 0.000 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Subject Sample # 2 | 0.126 | 03:17 |
| Breath Volume      | 1.781 | 03:17 |
| Air Blank          | 0.000 | 03:17 |
| Control Test       | 0.077 | 03:18 |
| Air Blank          | 0.000 | 03:18 |
| Diagnostics Check  | OK    | 03:18 |

---

Data Download Date/Time: 08/02/2022 20:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KONKEL,SCOTT,M  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: DUNCAN,SAVANA  
ARRESTING OFFICER AGENCY: Bradford County SO  
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/13/2022 01:19:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:12 |
| Air Blank          | 0.000 | 01:13 |
| Control Test       | 0.078 | 01:13 |
| Air Blank          | 0.000 | 01:14 |
| Subject Sample # 1 | 0.053 | 01:14 |
| Breath Volume      | 2.023 | 01:14 |
| Air Blank          | 0.000 | 01:15 |
| Air Blank          | 0.000 | 01:17 |
| Subject Sample # 2 | 0.048 | 01:17 |
| Breath Volume      | 1.839 | 01:17 |
| Air Blank          | 0.000 | 01:18 |
| Control Test       | 0.076 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Diagnostics Check  | OK    | 01:19 |

---

Data Download Date/Time: 08/03/2022 13:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARD,GLENN,B  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: WARD,GLENN  
ARRESTING OFFICER AGENCY: Bradford County SO  
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ROLFES

DATE/TIME OF TEST: 06/19/2022 03:40:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:34 |
| Air Blank          | 0.000 | 03:35 |
| Control Test       | 0.078 | 03:35 |
| Air Blank          | 0.000 | 03:35 |
| Subject Sample # 1 | 0.170 | 03:36 |
| Breath Volume      | 3.605 | 03:36 |
| Air Blank          | 0.000 | 03:37 |
| Air Blank          | 0.000 | 03:38 |
| Subject Sample # 2 | 0.156 | 03:39 |
| Breath Volume      | 2.972 | 03:39 |
| Air Blank          | 0.000 | 03:40 |
| Control Test       | 0.075 | 03:40 |
| Air Blank          | 0.000 | 03:40 |
| Diagnostics Check  | OK    | 03:40 |

---

Data Download Date/Time: 08/03/2022 13:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORENE,AIDAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: NORENE,AIDAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 06/01/2022 00:11:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Control Test       | 0.081 | 00:05 |
| Air Blank          | 0.000 | 00:06 |
| Subject Sample # 1 | 0.122 | 00:06 |
| Breath Volume      | 1.953 | 00:06 |
| Air Blank          | 0.000 | 00:07 |
| Air Blank          | 0.000 | 00:09 |
| Subject Sample # 2 | 0.128 | 00:09 |
| Breath Volume      | 1.625 | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Control Test       | 0.080 | 00:10 |
| Air Blank          | 0.000 | 00:11 |
| Diagnostics Check  | OK    | 00:11 |

---

Data Download Date/Time: 06/27/2022 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,JAMES,D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: DAVIS,JAMES  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BRONSTON

DATE/TIME OF TEST: 06/02/2022 02:13:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Control Test       | 0.080 | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Subject Sample # 1 | 0.091 | 02:08 |
| Breath Volume      | 2.613 | 02:08 |
| Air Blank          | 0.000 | 02:09 |
| Air Blank          | 0.000 | 02:11 |
| Subject Sample # 2 | 0.090 | 02:12 |
| Breath Volume      | 2.593 | 02:12 |
| Air Blank          | 0.000 | 02:12 |
| Control Test       | 0.079 | 02:12 |
| Air Blank          | 0.000 | 02:13 |
| Diagnostics Check  | OK    | 02:13 |

---

Data Download Date/Time: 06/27/2022 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CAREY,ANDREW  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 06/04/2022 00:26:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:19 |
| Air Blank          | 0.000 | 00:20 |
| Control Test       | 0.081 | 00:20 |
| Air Blank          | 0.000 | 00:20 |
| Subject Sample # 1 | 0.000 | 00:21 |
| Breath Volume      | 1.593 | 00:21 |
| Air Blank          | 0.000 | 00:21 |
| Air Blank          | 0.000 | 00:23 |
| Subject Sample # 2 | 0.000 | 00:24 |
| Breath Volume      | 1.566 | 00:24 |
| Air Blank          | 0.000 | 00:25 |
| Control Test       | 0.081 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Diagnostics Check  | OK    | 00:26 |

---

Data Download Date/Time: 06/27/2022 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOGETTE,MASON,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HOUNSHELL,LINDSEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/04/2022 07:34:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:28 |
| Air Blank          | 0.000 | 07:29 |
| Control Test       | 0.079 | 07:29 |
| Air Blank          | 0.000 | 07:29 |
| Subject Sample # 1 | 0.172 | 07:30 |
| Breath Volume      | 2.265 | 07:30 |
| Air Blank          | 0.000 | 07:30 |
| Air Blank          | 0.000 | 07:32 |
| Subject Sample # 2 | 0.183 | 07:33 |
| Breath Volume      | 2.453 | 07:33 |
| Air Blank          | 0.000 | 07:34 |
| Control Test       | 0.080 | 07:34 |
| Air Blank          | 0.000 | 07:34 |
| Diagnostics Check  | OK    | 07:34 |

---

Data Download Date/Time: 06/27/2022 14:25



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,KAYLAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: DAVIS,KAYLAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 16:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: LOHMILLER

DATE/TIME OF TEST: 06/04/2022 18:02:08

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 17:50 |                    |
| Air Blank         | 0.000 | 17:51 |                    |
| Control Test      | 0.080 | 17:51 |                    |
| Air Blank         | 0.000 | 17:52 |                    |
| Subject Sample    | # 1   | 17:55 | No Sample Provided |
| Breath Volume     | 0.000 | 17:55 |                    |
| Air Blank         | 0.000 | 17:55 |                    |
| Air Blank         | 0.000 | 17:57 |                    |
| Subject Sample    | # 2   | 18:00 | No Sample Provided |
| Breath Volume     | 0.000 | 18:00 |                    |
| Air Blank         | 0.000 | 18:01 |                    |
| Control Test      | 0.080 | 18:01 |                    |
| Air Blank         | 0.000 | 18:02 |                    |
| Diagnostics Check | OK    | 18:02 |                    |

Data Download Date/Time: 06/27/2022 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BREAZEAL,DENNIS  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: GRACE

DATE/TIME OF TEST: 06/05/2022 01:07:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:01 |
| Air Blank          | 0.000 | 01:01 |
| Control Test       | 0.080 | 01:01 |
| Air Blank          | 0.000 | 01:02 |
| Subject Sample # 1 | 0.121 | 01:02 |
| Breath Volume      | 3.085 | 01:02 |
| Air Blank          | 0.000 | 01:03 |
| Air Blank          | 0.000 | 01:05 |
| Subject Sample # 2 | 0.117 | 01:06 |
| Breath Volume      | 2.437 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Control Test       | 0.080 | 01:07 |
| Air Blank          | 0.000 | 01:07 |
| Diagnostics Check  | OK    | 01:07 |

---

Data Download Date/Time: 06/27/2022 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBIN,BRITTON,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBIN,BRITTON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: BUI  
ARREST TIME: 18:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 06/05/2022 20:21:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:14 |
| Air Blank          | 0.000 | 20:14 |
| Control Test       | 0.081 | 20:15 |
| Air Blank          | 0.000 | 20:15 |
| Subject Sample # 1 | 0.127 | 20:16 |
| Breath Volume      | 3.519 | 20:16 |
| Air Blank          | 0.000 | 20:17 |
| Air Blank          | 0.000 | 20:18 |
| Subject Sample # 2 | 0.123 | 20:19 |
| Breath Volume      | 3.781 | 20:19 |
| Air Blank          | 0.000 | 20:20 |
| Control Test       | 0.080 | 20:20 |
| Air Blank          | 0.000 | 20:21 |
| Diagnostics Check  | OK    | 20:21 |

---

Data Download Date/Time: 06/27/2022 14:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: CUPPLES,ZACHARY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: YAN

DATE/TIME OF TEST: 06/06/2022 20:10:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:04 |
| Air Blank          | 0.000 | 20:05 |
| Control Test       | 0.079 | 20:05 |
| Air Blank          | 0.000 | 20:06 |
| Subject Sample # 1 | 0.240 | 20:06 |
| Breath Volume      | 2.335 | 20:06 |
| Air Blank          | 0.000 | 20:07 |
| Air Blank          | 0.000 | 20:08 |
| Subject Sample # 2 | 0.236 | 20:09 |
| Breath Volume      | 2.320 | 20:09 |
| Air Blank          | 0.000 | 20:09 |
| Control Test       | 0.079 | 20:10 |
| Air Blank          | 0.000 | 20:10 |
| Diagnostics Check  | OK    | 20:10 |

---

Data Download Date/Time: 06/27/2022 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBIN,BRITTON,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BOWER,RACHEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: BUI  
ARREST TIME: 16:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: OSWALD

DATE/TIME OF TEST: 06/07/2022 18:16:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:09 |
| Air Blank          | 0.000 | 18:09 |
| Control Test       | 0.081 | 18:10 |
| Air Blank          | 0.000 | 18:10 |
| Subject Sample # 1 | 0.081 | 18:11 |
| Breath Volume      | 3.269 | 18:11 |
| Air Blank          | 0.000 | 18:12 |
| Air Blank          | 0.000 | 18:14 |
| Subject Sample # 2 | 0.073 | 18:14 |
| Breath Volume      | 1.687 | 18:14 |
| Air Blank          | 0.000 | 18:15 |
| Control Test       | 0.080 | 18:15 |
| Air Blank          | 0.000 | 18:16 |
| Diagnostics Check  | OK    | 18:16 |

---

Data Download Date/Time: 06/27/2022 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEAVER,COREY,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: ONEILL,BRAYDEN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SPONHEIMER

DATE/TIME OF TEST: 06/08/2022 03:34:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Control Test       | 0.080 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Subject Sample # 1 | 0.085 | 03:30 |
| Breath Volume      | 1.250 | 03:30 |
| Air Blank          | 0.000 | 03:30 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 2 | 0.081 | 03:33 |
| Breath Volume      | 1.558 | 03:33 |
| Air Blank          | 0.000 | 03:33 |
| Control Test       | 0.079 | 03:34 |
| Air Blank          | 0.000 | 03:34 |
| Diagnostics Check  | OK    | 03:34 |

---

Data Download Date/Time: 06/27/2022 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOHNER,KYLER,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MACPHERSON,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/08/2022 04:15:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:09 |
| Air Blank          | 0.000 | 04:09 |
| Control Test       | 0.082 | 04:10 |
| Air Blank          | 0.000 | 04:10 |
| Subject Sample # 1 | 0.000 | 04:11 |
| Breath Volume      | 2.019 | 04:11 |
| Air Blank          | 0.000 | 04:11 |
| Air Blank          | 0.000 | 04:13 |
| Subject Sample # 2 | 0.000 | 04:14 |
| Breath Volume      | 2.257 | 04:14 |
| Air Blank          | 0.000 | 04:14 |
| Control Test       | 0.080 | 04:15 |
| Air Blank          | 0.000 | 04:15 |
| Diagnostics Check  | OK    | 04:15 |

---

Data Download Date/Time: 06/27/2022 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVERA,DANTE,E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: TOLBERT,JOEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 06/08/2022 07:24:01

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 07:23 |            |
| Air Blank         |       | 07:23 | RFI Detect |
| Air Blank         | 0.000 | 07:24 |            |

---

Data Download Date/Time: 06/27/2022 14:25



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVERA,DANTE,E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: TOLBERT,JOEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 06/08/2022 07:33:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:27 |
| Air Blank          | 0.000 | 07:27 |
| Control Test       | 0.081 | 07:27 |
| Air Blank          | 0.000 | 07:28 |
| Subject Sample # 1 | 0.000 | 07:29 |
| Breath Volume      | 1.535 | 07:29 |
| Air Blank          | 0.000 | 07:30 |
| Air Blank          | 0.000 | 07:32 |
| Subject Sample # 2 | 0.000 | 07:32 |
| Breath Volume      | 1.687 | 07:32 |
| Air Blank          | 0.000 | 07:33 |
| Control Test       | 0.080 | 07:33 |
| Air Blank          | 0.000 | 07:33 |
| Diagnostics Check  | OK    | 07:33 |

---

Data Download Date/Time: 06/27/2022 14:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOHNER,KYLER,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: JONES,TYLER  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HELM

DATE/TIME OF TEST: 06/09/2022 04:13:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:11 |                      |
| Air Blank         | 0.000 | 04:11 |                      |
| Control Test      | 0.081 | 04:11 |                      |
| Air Blank         | 0.000 | 04:12 |                      |
| Subject Sample    | # 1   | 04:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:12 |                      |
| Air Blank         | 0.000 | 04:13 |                      |
| Control Test      | 0.080 | 04:13 |                      |
| Air Blank         | 0.000 | 04:13 |                      |
| Diagnostics Check | OK    | 04:13 |                      |

---

Data Download Date/Time: 06/27/2022 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEAVER,COREY,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: O'NEILL,BRAYDEN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: REGOT

DATE/TIME OF TEST: 06/09/2022 04:51:33

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 04:50 |            |
| Air Blank         |       | 04:51 | RFI Detect |
| Air Blank         | 0.000 | 04:51 |            |

---

Data Download Date/Time: 06/27/2022 14:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEAVER,COREY,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: O'NEILL,BRAYDEN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: REGOT

DATE/TIME OF TEST: 06/09/2022 04:56:48

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:53 |                      |
| Air Blank         | 0.000 | 04:54 |                      |
| Control Test      | 0.081 | 04:54 |                      |
| Air Blank         | 0.000 | 04:55 |                      |
| Subject Sample    | # 1   | 04:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:55 |                      |
| Air Blank         | 0.000 | 04:55 |                      |
| Control Test      | 0.081 | 04:56 |                      |
| Air Blank         | 0.000 | 04:56 |                      |
| Diagnostics Check | OK    | 04:56 |                      |

---

Data Download Date/Time: 06/27/2022 14:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEAVER,COREY,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: JONES,TYLER  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 06/11/2022 04:51:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:45 |
| Air Blank          | 0.000 | 04:45 |
| Control Test       | 0.080 | 04:46 |
| Air Blank          | 0.000 | 04:46 |
| Subject Sample # 1 | 0.107 | 04:47 |
| Breath Volume      | 3.832 | 04:47 |
| Air Blank          | 0.000 | 04:48 |
| Air Blank          | 0.000 | 04:49 |
| Subject Sample # 2 | 0.102 | 04:50 |
| Breath Volume      | 3.605 | 04:50 |
| Air Blank          | 0.000 | 04:51 |
| Control Test       | 0.079 | 04:51 |
| Air Blank          | 0.000 | 04:51 |
| Diagnostics Check  | OK    | 04:51 |

---

Data Download Date/Time: 06/27/2022 14:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: LUCAS, JOHN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/12/2022 03:44:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:37 |
| Air Blank          | 0.000 | 03:38 |
| Control Test       | 0.080 | 03:38 |
| Air Blank          | 0.000 | 03:38 |
| Subject Sample # 1 | 0.156 | 03:40 |
| Breath Volume      | 3.429 | 03:40 |
| Air Blank          | 0.000 | 03:40 |
| Air Blank          | 0.000 | 03:42 |
| Subject Sample # 2 | 0.151 | 03:43 |
| Breath Volume      | 3.343 | 03:43 |
| Air Blank          | 0.000 | 03:43 |
| Control Test       | 0.080 | 03:44 |
| Air Blank          | 0.000 | 03:44 |
| Diagnostics Check  | OK    | 03:44 |

---

Data Download Date/Time: 06/27/2022 14:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOHNER,KYLER,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: JONES,TYLER  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/16/2022 00:04:41

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:01 |                      |
| Air Blank         | 0.000 | 00:02 |                      |
| Control Test      | 0.080 | 00:02 |                      |
| Air Blank         | 0.000 | 00:02 |                      |
| Subject Sample    | # 1   | 00:03 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:03 |                      |
| Air Blank         | 0.000 | 00:03 |                      |
| Control Test      | 0.080 | 00:04 |                      |
| Air Blank         | 0.000 | 00:04 |                      |
| Diagnostics Check | OK    | 00:04 |                      |

---

Data Download Date/Time: 06/27/2022 14:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORSMAN,MICHAEL,S  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: FORSMAN,MICHAEL  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 16:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: CALLE-CALLE

DATE/TIME OF TEST: 06/16/2022 17:52:45

---

|                   |       |       |                  |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK    | 17:49 |                  |
| Air Blank         | 0.000 | 17:50 |                  |
| Control Test      | 0.081 | 17:50 |                  |
| Air Blank         | 0.000 | 17:50 |                  |
| Subject Sample    | # 1   | 17:52 | Sequence Aborted |
| Breath Volume     | 0.000 | 17:52 |                  |
| Air Blank         | 0.000 | 17:52 |                  |

---

Data Download Date/Time: 06/27/2022 14:26



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORSMAN,MICHAEL,S  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: FORSMAN,MICHAEL  
ARRESTING OFFICER AGENCY: Hampton PD  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 16:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: CALLE-CALLE

DATE/TIME OF TEST: 06/16/2022 18:01:17

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 17:57 |            |
| Air Blank         | 0.000 | 17:58 |            |
| Control Test      | 0.081 | 17:58 |            |
| Air Blank         | 0.000 | 17:59 |            |
| Subject Sample    | # 1   | 18:00 | RFI Detect |
| Breath Volume     | 0.000 | 18:00 |            |
| Air Blank         |       | 18:00 | RFI Detect |
| Air Blank         | 0.000 | 18:01 |            |

---

Data Download Date/Time: 06/27/2022 14:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOHNER,KYLER,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MACPHERSON,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: TREASE

DATE/TIME OF TEST: 06/17/2022 00:38:33

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:33 |                      |
| Air Blank         | 0.000 | 00:34 |                      |
| Control Test      | 0.080 | 00:34 |                      |
| Air Blank         | 0.000 | 00:34 |                      |
| Subject Sample    | # 1   | 00:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:37 |                      |
| Air Blank         | 0.000 | 00:37 |                      |
| Control Test      | 0.080 | 00:37 |                      |
| Air Blank         | 0.000 | 00:38 |                      |
| Diagnostics Check | OK    | 00:38 |                      |

---

Data Download Date/Time: 06/27/2022 14:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,JAMES,D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: WATERS,CALEB  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: DEWOLFE

DATE/TIME OF TEST: 06/18/2022 05:21:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:14 |
| Air Blank          | 0.000 | 05:15 |
| Control Test       | 0.080 | 05:15 |
| Air Blank          | 0.000 | 05:15 |
| Subject Sample # 1 | 0.182 | 05:16 |
| Breath Volume      | 3.730 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Air Blank          | 0.000 | 05:19 |
| Subject Sample # 2 | 0.178 | 05:20 |
| Breath Volume      | 3.414 | 05:20 |
| Air Blank          | 0.000 | 05:20 |
| Control Test       | 0.079 | 05:21 |
| Air Blank          | 0.000 | 05:21 |
| Diagnostics Check  | OK    | 05:21 |

---

Data Download Date/Time: 06/27/2022 14:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CAREY,ANDREW  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 06/19/2022 06:05:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:58 |
| Air Blank          | 0.000 | 05:58 |
| Control Test       | 0.081 | 05:59 |
| Air Blank          | 0.000 | 05:59 |
| Subject Sample # 1 | 0.093 | 06:00 |
| Breath Volume      | 1.546 | 06:00 |
| Air Blank          | 0.000 | 06:01 |
| Air Blank          | 0.000 | 06:02 |
| Subject Sample # 2 | 0.092 | 06:04 |
| Breath Volume      | 2.230 | 06:04 |
| Air Blank          | 0.000 | 06:04 |
| Control Test       | 0.080 | 06:05 |
| Air Blank          | 0.000 | 06:05 |
| Diagnostics Check  | OK    | 06:05 |

---

Data Download Date/Time: 06/27/2022 14:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD,J,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BOYD,J  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 12:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:28

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/19/2022 13:58:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:48 |
| Air Blank          | 0.000 | 13:49 |
| Control Test       | 0.081 | 13:49 |
| Air Blank          | 0.000 | 13:50 |
| Subject Sample # 1 | 0.270 | 13:50 |
| Breath Volume      | 1.230 | 13:50 |
| Air Blank          | 0.000 | 13:51 |
| Air Blank          | 0.000 | 13:52 |
| Subject Sample # 2 | 0.249 | 13:53 |
| Breath Volume      | 1.593 | 13:53 |
| Air Blank          | 0.000 | 13:54 |
| Air Blank          | 0.000 | 13:56 |
| Subject Sample # 3 | 0.254 | 13:56 |
| Breath Volume      | 1.636 | 13:56 |
| Air Blank          | 0.000 | 13:57 |
| Control Test       | 0.080 | 13:57 |
| Air Blank          | 0.000 | 13:58 |
| Diagnostics Check  | OK    | 13:58 |

---

Data Download Date/Time: 06/27/2022 14:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOHNER,KYLER,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: PEA,CLARISE  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 06/22/2022 01:44:51

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:41 |                      |
| Air Blank         | 0.000 | 01:41 |                      |
| Control Test      | 0.080 | 01:42 |                      |
| Air Blank         | 0.000 | 01:42 |                      |
| Subject Sample    | # 1   | 01:43 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:43 |                      |
| Air Blank         | 0.000 | 01:43 |                      |
| Control Test      | 0.080 | 01:44 |                      |
| Air Blank         | 0.000 | 01:44 |                      |
| Diagnostics Check | OK    | 01:44 |                      |

Data Download Date/Time: 06/27/2022 14:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLSON,JOSHUA,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MOORE,JUSTIN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 06/22/2022 20:41:10

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:37 |                      |
| Air Blank         | 0.000 | 20:37 |                      |
| Control Test      | 0.081 | 20:38 |                      |
| Air Blank         | 0.000 | 20:38 |                      |
| Subject Sample    | # 1   | 20:39 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:39 |                      |
| Air Blank         | 0.000 | 20:40 |                      |
| Control Test      | 0.080 | 20:40 |                      |
| Air Blank         | 0.000 | 20:41 |                      |
| Diagnostics Check | OK    | 20:41 |                      |

---

Data Download Date/Time: 06/27/2022 14:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNN,JAMISON,C  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: LYNN,JAMISON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 14:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:08

SUBJECT'S LAST NAME: KWAKENAT

DATE/TIME OF TEST: 06/23/2022 15:33:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:26 |
| Air Blank          | 0.000 | 15:26 |
| Control Test       | 0.080 | 15:26 |
| Air Blank          | 0.000 | 15:27 |
| Subject Sample # 1 | 0.374 | 15:28 |
| Breath Volume      | 4.511 | 15:28 |
| Air Blank          | 0.000 | 15:29 |
| Air Blank          | 0.000 | 15:30 |
| Subject Sample # 2 | 0.377 | 15:31 |
| Breath Volume      | 4.906 | 15:31 |
| Air Blank          | 0.000 | 15:32 |
| Control Test       | 0.079 | 15:32 |
| Air Blank          | 0.000 | 15:33 |
| Diagnostics Check  | OK    | 15:33 |

---

Data Download Date/Time: 06/27/2022 14:27



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,WILLIAM,E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HOUNSHELL,LINDSEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 08:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:18

SUBJECT'S LAST NAME: CORELLA

DATE/TIME OF TEST: 06/24/2022 09:54:29

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 09:53 |            |
| Air Blank         |       | 09:54 | RFI Detect |
| Air Blank         | 0.000 | 09:54 |            |

---

Data Download Date/Time: 06/27/2022 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,WILLIAM,E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HOUNSHELL,LINDSEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/14/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 08:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:18

SUBJECT'S LAST NAME: CORELLA

DATE/TIME OF TEST: 06/24/2022 10:00:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 09:57 |                      |
| Air Blank         | 0.000 | 09:58 |                      |
| Control Test      | 0.080 | 09:58 |                      |
| Air Blank         | 0.000 | 09:59 |                      |
| Subject Sample    | # 1   | 09:59 | Subject Test Refused |
| Breath Volume     | 0.000 | 09:59 |                      |
| Air Blank         | 0.000 | 09:59 |                      |
| Control Test      | 0.080 | 10:00 |                      |
| Air Blank         | 0.000 | 10:00 |                      |
| Diagnostics Check | OK    | 10:00 |                      |

---

Data Download Date/Time: 06/27/2022 14:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOLSOM, DAVID, A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: CUPPLES, ZACHARY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 06/29/2022 22:52:48

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:49 |                      |
| Air Blank         | 0.000 | 22:49 |                      |
| Control Test      | 0.080 | 22:50 |                      |
| Air Blank         | 0.000 | 22:50 |                      |
| Subject Sample    | # 1   | 22:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:51 |                      |
| Air Blank         | 0.000 | 22:51 |                      |
| Control Test      | 0.081 | 22:52 |                      |
| Air Blank         | 0.000 | 22:52 |                      |
| Diagnostics Check | OK    | 22:52 |                      |

---

Data Download Date/Time: 07/14/2022 18:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MUISE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: STEED

DATE/TIME OF TEST: 06/25/2022 03:06:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:00 |
| Air Blank          | 0.000 | 03:00 |
| Control Test       | 0.081 | 03:00 |
| Air Blank          | 0.000 | 03:01 |
| Subject Sample # 1 | 0.247 | 03:01 |
| Breath Volume      | 3.253 | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Air Blank          | 0.000 | 03:04 |
| Subject Sample # 2 | 0.233 | 03:05 |
| Breath Volume      | 2.214 | 03:05 |
| Air Blank          | 0.000 | 03:05 |
| Control Test       | 0.081 | 03:06 |
| Air Blank          | 0.000 | 03:06 |
| Diagnostics Check  | OK    | 03:06 |

---

Data Download Date/Time: 06/27/2022 14:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNDT,DAVID,L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MUNDT,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/25/2022 20:45:10

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 20:39 |            |
| Air Blank         | 0.000     | 20:39 |            |
| Control Test      | 0.081     | 20:39 |            |
| Air Blank         | 0.000     | 20:40 |            |
| Subject Sample    | # 1 0.215 | 20:40 |            |
| Breath Volume     | 1.945     | 20:40 |            |
| Air Blank         | 0.000     | 20:41 |            |
| Air Blank         | 0.000     | 20:43 |            |
| Subject Sample    | # 2 0.213 | 20:44 |            |
| Breath Volume     | 1.648     | 20:44 |            |
| Air Blank         |           | 20:44 | RFI Detect |
| Air Blank         | 0.000     | 20:45 |            |

---

Data Download Date/Time: 06/27/2022 14:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SPENCER, RAYMOND  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: SHOENS

DATE/TIME OF TEST: 06/26/2022 01:10:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:59 |
| Air Blank          | 0.000 | 01:00 |
| Control Test       | 0.081 | 01:00 |
| Air Blank          | 0.000 | 01:00 |
| Subject Sample # 1 | 0.139 | 01:03 |
| Breath Volume      | 1.117 | 01:03 |
| Air Blank          | 0.000 | 01:04 |
| Air Blank          | 0.000 | 01:06 |
| Subject Sample # 2 | 0.134 | 01:08 |
| Breath Volume      | 1.296 | 01:08 |
| Air Blank          | 0.000 | 01:09 |
| Control Test       | 0.081 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Diagnostics Check  | OK    | 01:10 |

---

Data Download Date/Time: 06/27/2022 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: CUPPLES,ZACHARY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 15:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 06/27/2022 16:36:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:29 |
| Air Blank          | 0.000 | 16:29 |
| Control Test       | 0.081 | 16:30 |
| Air Blank          | 0.000 | 16:30 |
| Subject Sample # 1 | 0.045 | 16:31 |
| Breath Volume      | 3.851 | 16:31 |
| Air Blank          | 0.000 | 16:31 |
| Air Blank          | 0.000 | 16:33 |
| Subject Sample # 2 | 0.043 | 16:34 |
| Breath Volume      | 3.218 | 16:34 |
| Air Blank          | 0.000 | 16:35 |
| Control Test       | 0.082 | 16:35 |
| Air Blank          | 0.000 | 16:35 |
| Diagnostics Check  | OK    | 16:36 |

---

Data Download Date/Time: 07/14/2022 18:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUNTER,REGEN,B  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: GUNTER,REGEN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: ANDERSEN

DATE/TIME OF TEST: 06/28/2022 22:19:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:16 |                      |
| Air Blank         | 0.000 | 22:17 |                      |
| Control Test      | 0.080 | 22:17 |                      |
| Air Blank         | 0.000 | 22:17 |                      |
| Subject Sample    | # 1   | 22:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:18 |                      |
| Air Blank         | 0.000 | 22:18 |                      |
| Control Test      | 0.080 | 22:18 |                      |
| Air Blank         | 0.000 | 22:19 |                      |
| Diagnostics Check | OK    | 22:19 |                      |

---

Data Download Date/Time: 07/14/2022 18:38



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOLSOM,DAVID,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: CUPPLES,ZACK  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 06/29/2022 23:04:25

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:01 |                      |
| Air Blank         | 0.000 | 23:02 |                      |
| Control Test      | 0.080 | 23:02 |                      |
| Air Blank         | 0.000 | 23:02 |                      |
| Subject Sample    | # 1   | 23:03 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:03 |                      |
| Air Blank         | 0.000 | 23:03 |                      |
| Control Test      | 0.081 | 23:03 |                      |
| Air Blank         | 0.000 | 23:04 |                      |
| Diagnostics Check | OK    | 23:04 |                      |

---

Data Download Date/Time: 07/14/2022 18:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOHNER,KYLER,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BICKNELL,BRANDON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: MCNEIL

DATE/TIME OF TEST: 06/30/2022 00:59:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:56 |                      |
| Air Blank         | 0.000 | 00:56 |                      |
| Control Test      | 0.081 | 00:57 |                      |
| Air Blank         | 0.000 | 00:57 |                      |
| Subject Sample    | # 1   | 00:58 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:58 |                      |
| Air Blank         | 0.000 | 00:58 |                      |
| Control Test      | 0.080 | 00:58 |                      |
| Air Blank         | 0.000 | 00:59 |                      |
| Diagnostics Check | OK    | 00:59 |                      |

---

Data Download Date/Time: 07/14/2022 18:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HENSON,JAY,C

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: MAYO,KEITH

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/05/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1346146

DRY GAS EXPIRY DATE: 07/07/2023

VIOLATION CODE: DUI

ARREST TIME: 22:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BRANDT

DATE/TIME OF TEST: 06/07/2022 23:33:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:30 |                      |
| Air Blank         | 0.000 | 23:30 |                      |
| Control Test      | 0.082 | 23:31 |                      |
| Air Blank         | 0.000 | 23:31 |                      |
| Subject Sample    | # 1   | 23:32 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:32 |                      |
| Air Blank         | 0.000 | 23:33 |                      |
| Control Test      | 0.082 | 23:33 |                      |
| Air Blank         | 0.000 | 23:33 |                      |
| Diagnostics Check | OK    | 23:33 |                      |

---

Data Download Date/Time: 06/14/2022 09:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANKS,SEAN,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BANKS,SEAN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 05/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 06/04/2022 23:12:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:06 |
| Air Blank          | 0.000 | 23:06 |
| Control Test       | 0.081 | 23:07 |
| Air Blank          | 0.000 | 23:07 |
| Subject Sample # 1 | 0.177 | 23:08 |
| Breath Volume      | 3.140 | 23:08 |
| Air Blank          | 0.000 | 23:08 |
| Air Blank          | 0.000 | 23:10 |
| Subject Sample # 2 | 0.168 | 23:11 |
| Breath Volume      | 2.062 | 23:11 |
| Air Blank          | 0.000 | 23:11 |
| Control Test       | 0.081 | 23:12 |
| Air Blank          | 0.000 | 23:12 |
| Diagnostics Check  | OK    | 23:12 |

---

Data Download Date/Time: 06/08/2022 12:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALZA, PAOLOGENE, R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: NEWMAN, SCOTT  
ARRESTING OFFICER AGENCY: Madison PD  
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: DVONTE

DATE/TIME OF TEST: 06/06/2022 01:43:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:36 |
| Air Blank          | 0.000 | 01:37 |
| Control Test       | 0.076 | 01:37 |
| Air Blank          | 0.000 | 01:38 |
| Subject Sample # 1 | 0.134 | 01:38 |
| Breath Volume      | 1.273 | 01:38 |
| Air Blank          | 0.000 | 01:39 |
| Air Blank          | 0.000 | 01:41 |
| Subject Sample # 2 | 0.144 | 01:42 |
| Breath Volume      | 2.371 | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.077 | 01:43 |
| Air Blank          | 0.000 | 01:43 |
| Diagnostics Check  | OK    | 01:43 |

---

Data Download Date/Time: 06/27/2022 12:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARKE,GREGORY,A  
BREATH TEST OPERATOR AGENCY: FL Dept of AG  
ARRESTING OFFICER: CLARKE,GREGORY  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: SICKLER JR

DATE/TIME OF TEST: 06/18/2022 01:14:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:08 |
| Air Blank          | 0.000 | 01:09 |
| Control Test       | 0.076 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Subject Sample # 1 | 0.180 | 01:10 |
| Breath Volume      | 1.878 | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Air Blank          | 0.000 | 01:12 |
| Subject Sample # 2 | 0.178 | 01:13 |
| Breath Volume      | 2.214 | 01:13 |
| Air Blank          | 0.000 | 01:13 |
| Control Test       | 0.075 | 01:14 |
| Air Blank          | 0.000 | 01:14 |
| Diagnostics Check  | OK    | 01:14 |

---

Data Download Date/Time: 06/27/2022 12:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NASH,JARRED,D  
BREATH TEST OPERATOR AGENCY: Madison County SO  
ARRESTING OFFICER: NASH,JARRED  
ARRESTING OFFICER AGENCY: Madison County SO  
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: WOOTEN

DATE/TIME OF TEST: 06/18/2022 22:41:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:35 |
| Air Blank          | 0.000 | 22:35 |
| Control Test       | 0.076 | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Subject Sample # 1 | 0.207 | 22:36 |
| Breath Volume      | 1.871 | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Air Blank          | 0.000 | 22:39 |
| Subject Sample # 2 | 0.191 | 22:39 |
| Breath Volume      | 2.507 | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Control Test       | 0.076 | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Diagnostics Check  | OK    | 22:41 |

---

Data Download Date/Time: 06/27/2022 12:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,DAVID,H  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: THACKER,M  
ARRESTING OFFICER AGENCY: Madison PD  
LAST AGENCY INSPECTION: 05/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 06/26/2022 00:35:11

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 00:21 |                    |
| Air Blank         | 0.000 | 00:22 |                    |
| Control Test      | 0.077 | 00:22 |                    |
| Air Blank         | 0.000 | 00:22 |                    |
| Subject Sample    | # 1   | 00:25 | No Sample Provided |
| Breath Volume     | 0.191 | 00:25 |                    |
| Air Blank         | 0.000 | 00:26 |                    |
| Air Blank         | 0.000 | 00:28 |                    |
| Subject Sample    | # 2   | 00:29 |                    |
| Breath Volume     | 2.449 | 00:29 |                    |
| Air Blank         | 0.000 | 00:30 |                    |
| Air Blank         | 0.000 | 00:32 |                    |
| Subject Sample    | # 3   | 00:33 |                    |
| Breath Volume     | 1.765 | 00:33 |                    |
| Air Blank         | 0.000 | 00:34 |                    |
| Control Test      | 0.077 | 00:34 |                    |
| Air Blank         | 0.000 | 00:35 |                    |
| Diagnostics Check | OK    | 00:35 |                    |

---

Data Download Date/Time: 06/27/2022 12:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLYN, WILLIAM, P  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: HOLMES, KYLE  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: OLVERA

DATE/TIME OF TEST: 06/01/2022 02:31:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:25 |
| Air Blank          | 0.000 | 02:25 |
| Control Test       | 0.082 | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Subject Sample # 1 | 0.188 | 02:27 |
| Breath Volume      | 1.500 | 02:27 |
| Air Blank          | 0.000 | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 2 | 0.177 | 02:30 |
| Breath Volume      | 1.789 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Control Test       | 0.081 | 02:31 |
| Air Blank          | 0.000 | 02:31 |
| Diagnostics Check  | OK    | 02:31 |

---

Data Download Date/Time: 06/28/2022 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLYN, WILLIAM, P  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ALLYN, WILLIAM  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CANNA

DATE/TIME OF TEST: 06/04/2022 03:05:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:58 |
| Air Blank          | 0.000 | 02:59 |
| Control Test       | 0.081 | 02:59 |
| Air Blank          | 0.000 | 03:00 |
| Subject Sample # 1 | 0.132 | 03:00 |
| Breath Volume      | 2.187 | 03:00 |
| Air Blank          | 0.000 | 03:01 |
| Air Blank          | 0.000 | 03:03 |
| Subject Sample # 2 | 0.136 | 03:03 |
| Breath Volume      | 2.710 | 03:03 |
| Air Blank          | 0.000 | 03:04 |
| Control Test       | 0.081 | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Diagnostics Check  | OK    | 03:05 |

---

Data Download Date/Time: 06/28/2022 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSE,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: HOUSE,STEVEN  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: STOUT

DATE/TIME OF TEST: 06/17/2022 01:15:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:08 |
| Air Blank          | 0.000 | 01:09 |
| Control Test       | 0.081 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Subject Sample # 1 | 0.169 | 01:10 |
| Breath Volume      | 1.625 | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Air Blank          | 0.000 | 01:12 |
| Subject Sample # 2 | 0.160 | 01:13 |
| Breath Volume      | 1.386 | 01:13 |
| Air Blank          | 0.000 | 01:14 |
| Control Test       | 0.081 | 01:14 |
| Air Blank          | 0.000 | 01:15 |
| Diagnostics Check  | OK    | 01:15 |

---

Data Download Date/Time: 06/28/2022 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAND,CURTIS,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: BLAND,CURTIS  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BURRESS

DATE/TIME OF TEST: 06/19/2022 03:32:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:26 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.082 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 1 | 0.123 | 03:27 |
| Breath Volume      | 2.601 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Air Blank          | 0.000 | 03:30 |
| Subject Sample # 2 | 0.127 | 03:30 |
| Breath Volume      | 2.878 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.081 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Diagnostics Check  | OK    | 03:32 |

---

Data Download Date/Time: 06/28/2022 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSE,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: SIPE,SEVEN  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/22/2022 04:39:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:32 |
| Air Blank          | 0.000 | 04:33 |
| Control Test       | 0.081 | 04:33 |
| Air Blank          | 0.000 | 04:33 |
| Subject Sample # 1 | 0.236 | 04:34 |
| Breath Volume      | 1.281 | 04:34 |
| Air Blank          | 0.000 | 04:35 |
| Air Blank          | 0.000 | 04:37 |
| Subject Sample # 2 | 0.242 | 04:38 |
| Breath Volume      | 1.988 | 04:38 |
| Air Blank          | 0.000 | 04:38 |
| Control Test       | 0.080 | 04:39 |
| Air Blank          | 0.000 | 04:39 |
| Diagnostics Check  | OK    | 04:39 |

---

Data Download Date/Time: 06/28/2022 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: PRICE,CARLON  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: PETTY

DATE/TIME OF TEST: 06/29/2022 23:12:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Control Test       | 0.081 | 23:06 |
| Air Blank          | 0.000 | 23:07 |
| Subject Sample # 1 | 0.172 | 23:07 |
| Breath Volume      | 3.078 | 23:07 |
| Air Blank          | 0.000 | 23:08 |
| Air Blank          | 0.000 | 23:10 |
| Subject Sample # 2 | 0.170 | 23:11 |
| Breath Volume      | 3.253 | 23:11 |
| Air Blank          | 0.000 | 23:11 |
| Control Test       | 0.081 | 23:12 |
| Air Blank          | 0.000 | 23:12 |
| Diagnostics Check  | OK    | 23:12 |

---

Data Download Date/Time: 07/11/2022 11:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: PRICE,CARLON  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: COPPALA

DATE/TIME OF TEST: 06/09/2022 00:30:28

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 00:25 |            |
| Air Blank         | 0.000     | 00:25 |            |
| Control Test      | 0.079     | 00:26 |            |
| Air Blank         | 0.000     | 00:26 |            |
| Subject Sample    | # 1 0.096 | 00:27 |            |
| Breath Volume     | 2.113     | 00:27 |            |
| Air Blank         | 0.000     | 00:27 |            |
| Air Blank         | 0.000     | 00:29 |            |
| Subject Sample    | # 2 0.095 | 00:30 |            |
| Breath Volume     | 1.675     | 00:30 |            |
| Air Blank         |           | 00:30 | RFI Detect |
| Air Blank         |           | 00:30 | RFI Detect |

---

Data Download Date/Time: 06/28/2022 14:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: MOSS,BOB  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PHELAN

DATE/TIME OF TEST: 06/09/2022 03:06:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:59 |
| Air Blank          | 0.000 | 02:59 |
| Control Test       | 0.077 | 02:59 |
| Air Blank          | 0.000 | 03:00 |
| Subject Sample # 1 | 0.132 | 03:01 |
| Breath Volume      | 1.472 | 03:01 |
| Air Blank          | 0.000 | 03:01 |
| Air Blank          | 0.000 | 03:03 |
| Subject Sample # 2 | 0.139 | 03:04 |
| Breath Volume      | 1.714 | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Control Test       | 0.076 | 03:05 |
| Air Blank          | 0.000 | 03:05 |
| Diagnostics Check  | OK    | 03:06 |

---

Data Download Date/Time: 06/28/2022 14:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: PRANTHER

DATE/TIME OF TEST: 06/09/2022 18:13:05

---

|                   |           |       |               |
|-------------------|-----------|-------|---------------|
| Diagnostics Check | OK        | 18:06 |               |
| Air Blank         | 0.000     | 18:06 |               |
| Control Test      | 0.080     | 18:07 |               |
| Air Blank         | 0.000     | 18:07 |               |
| Subject Sample    | # 1 0.348 | 18:08 |               |
| Breath Volume     | 1.488     | 18:08 |               |
| Air Blank         | 0.000     | 18:09 |               |
| Air Blank         | 0.000     | 18:10 |               |
| Subject Sample    | # 2       | 18:11 | Slope Not Met |
| Breath Volume     | 0.816     | 18:11 |               |
| Air Blank         | 0.000     | 18:12 |               |
| Control Test      | 0.080     | 18:12 |               |
| Air Blank         | 0.000     | 18:12 |               |
| Diagnostics Check | OK        | 18:13 |               |

---

Data Download Date/Time: 06/28/2022 14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: PRATHRN

DATE/TIME OF TEST: 06/09/2022 18:22:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:16 |
| Air Blank          | 0.000 | 18:16 |
| Control Test       | 0.079 | 18:16 |
| Air Blank          | 0.000 | 18:17 |
| Subject Sample # 1 | 0.346 | 18:18 |
| Breath Volume      | 1.511 | 18:18 |
| Air Blank          | 0.000 | 18:19 |
| Air Blank          | 0.000 | 18:20 |
| Subject Sample # 2 | 0.340 | 18:21 |
| Breath Volume      | 1.523 | 18:21 |
| Air Blank          | 0.000 | 18:22 |
| Control Test       | 0.075 | 18:22 |
| Air Blank          | 0.000 | 18:22 |
| Diagnostics Check  | OK    | 18:22 |

---

Data Download Date/Time: 06/28/2022 14:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: PRICE,CARLON  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: CHAVEZ MAYA

DATE/TIME OF TEST: 06/11/2022 00:13:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:07 |
| Air Blank          | 0.000 | 00:08 |
| Control Test       | 0.080 | 00:08 |
| Air Blank          | 0.000 | 00:09 |
| Subject Sample # 1 | 0.215 | 00:09 |
| Breath Volume      | 2.175 | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Air Blank          | 0.000 | 00:11 |
| Subject Sample # 2 | 0.208 | 00:12 |
| Breath Volume      | 1.515 | 00:12 |
| Air Blank          | 0.000 | 00:13 |
| Control Test       | 0.080 | 00:13 |
| Air Blank          | 0.000 | 00:13 |
| Diagnostics Check  | OK    | 00:13 |

---

Data Download Date/Time: 06/28/2022 14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: STALNAKER,DILLON  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: MULLINAX

DATE/TIME OF TEST: 06/11/2022 20:26:31

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 20:19 |                |
| Air Blank         | 0.000     | 20:19 |                |
| Control Test      | 0.079     | 20:20 |                |
| Air Blank         | 0.000     | 20:20 |                |
| Subject Sample    | # 1 0.242 | 20:23 | Volume Not Met |
| Breath Volume     | 0.929     | 20:23 |                |
| Air Blank         | 0.000     | 20:24 |                |
| Air Blank         |           | 20:26 | RFI Detect     |
| Air Blank         |           | 20:26 | RFI Detect     |
| Air Blank         |           | 20:26 | RFI Detect     |

---

Data Download Date/Time: 06/28/2022 14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: STALNAKER,DILLON  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: MULLINAX

DATE/TIME OF TEST: 06/11/2022 20:33:15

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:29 |                      |
| Air Blank         | 0.000 | 20:29 |                      |
| Control Test      | 0.077 | 20:29 |                      |
| Air Blank         | 0.000 | 20:30 |                      |
| Subject Sample    | # 1   | 20:31 | Subject Test Refused |
| Breath Volume     | 0.617 | 20:31 |                      |
| Air Blank         | 0.000 | 20:32 |                      |
| Control Test      | 0.076 | 20:32 |                      |
| Air Blank         | 0.000 | 20:33 |                      |
| Diagnostics Check | OK    | 20:33 |                      |

---

Data Download Date/Time: 06/28/2022 14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: PRICE,CARLON  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: ROMERO RAMOS

DATE/TIME OF TEST: 06/11/2022 21:22:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:14 |
| Air Blank          | 0.000 | 21:15 |
| Control Test       | 0.079 | 21:15 |
| Air Blank          | 0.000 | 21:16 |
| Subject Sample # 1 | 0.134 | 21:17 |
| Breath Volume      | 1.437 | 21:17 |
| Air Blank          | 0.000 | 21:18 |
| Air Blank          | 0.000 | 21:20 |
| Subject Sample # 2 | 0.145 | 21:20 |
| Breath Volume      | 1.511 | 21:20 |
| Air Blank          | 0.000 | 21:21 |
| Control Test       | 0.076 | 21:21 |
| Air Blank          | 0.000 | 21:21 |
| Diagnostics Check  | OK    | 21:22 |

---

Data Download Date/Time: 06/28/2022 14:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: WALKER,MARINA  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/11/2022 22:55:30

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:42 |                |
| Air Blank         | 0.000     | 22:42 |                |
| Control Test      | 0.080     | 22:43 |                |
| Air Blank         | 0.000     | 22:43 |                |
| Subject Sample    | # 1 0.225 | 22:46 | Volume Not Met |
| Breath Volume     | 0.726     | 22:46 |                |
| Air Blank         | 0.000     | 22:47 |                |
| Air Blank         | 0.000     | 22:49 |                |
| Subject Sample    | # 2 0.227 | 22:50 |                |
| Breath Volume     | 1.738     | 22:50 |                |
| Air Blank         | 0.000     | 22:50 |                |
| Air Blank         | 0.000     | 22:52 |                |
| Subject Sample    | # 3 0.237 | 22:53 |                |
| Breath Volume     | 1.894     | 22:53 |                |
| Air Blank         | 0.000     | 22:54 |                |
| Control Test      | 0.077     | 22:54 |                |
| Air Blank         | 0.000     | 22:55 |                |
| Diagnostics Check | OK        | 22:55 |                |

---

Data Download Date/Time: 06/28/2022 14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: HOUSE,STEVEN  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 06/16/2022 02:45:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:39 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.080 | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 1 | 0.213 | 02:41 |
| Breath Volume      | 1.882 | 02:41 |
| Air Blank          | 0.000 | 02:42 |
| Air Blank          | 0.000 | 02:43 |
| Subject Sample # 2 | 0.208 | 02:44 |
| Breath Volume      | 1.519 | 02:44 |
| Air Blank          | 0.000 | 02:44 |
| Control Test       | 0.078 | 02:45 |
| Air Blank          | 0.000 | 02:45 |
| Diagnostics Check  | OK    | 02:45 |

---

Data Download Date/Time: 06/28/2022 14:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: BUTTS,KATELYN  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: CHAMPAGNE

DATE/TIME OF TEST: 06/23/2022 22:28:52

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 22:15 |                           |
| Air Blank         | 0.000     | 22:16 |                           |
| Control Test      | 0.079     | 22:16 |                           |
| Air Blank         | 0.000     | 22:16 |                           |
| Subject Sample    | # 1 0.252 | 22:17 |                           |
| Breath Volume     | 1.183     | 22:17 |                           |
| Air Blank         | 0.000     | 22:18 |                           |
| Air Blank         | 0.000     | 22:20 |                           |
| Subject Sample    | # 2 0.217 | 22:23 | Volume Not Met            |
| Breath Volume     | 0.750     | 22:23 |                           |
| Air Blank         | 0.000     | 22:24 |                           |
| Air Blank         | 0.000     | 22:25 |                           |
| Subject Sample    | # 3 0.263 | 22:26 |                           |
| Breath Volume     | 1.660     | 22:26 |                           |
| Air Blank         | 0.000     | 22:27 |                           |
| Control Test      | 0.077     | 22:27 |                           |
| Air Blank         | 0.000     | 22:28 |                           |
| Diagnostics Check | Fail      | 22:28 | Analytical Stability Test |
| Air Blank         | 0.000     | 22:28 |                           |

Data Download Date/Time: 06/28/2022 14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 06/24/2022 03:12:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Control Test       | 0.080 | 03:06 |
| Air Blank          | 0.000 | 03:06 |
| Subject Sample # 1 | 0.207 | 03:07 |
| Breath Volume      | 2.156 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Air Blank          | 0.000 | 03:10 |
| Subject Sample # 2 | 0.200 | 03:10 |
| Breath Volume      | 1.824 | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Control Test       | 0.079 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Diagnostics Check  | OK    | 03:12 |

---

Data Download Date/Time: 06/28/2022 14:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: GORE,MATHEW  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG121002  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: BUI  
ARREST TIME: 19:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: ALVEREZ

DATE/TIME OF TEST: 06/24/2022 20:27:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:20 |
| Air Blank          | 0.000 | 20:21 |
| Control Test       | 0.078 | 20:21 |
| Air Blank          | 0.000 | 20:22 |
| Subject Sample # 1 | 0.144 | 20:22 |
| Breath Volume      | 1.757 | 20:22 |
| Air Blank          | 0.000 | 20:23 |
| Air Blank          | 0.000 | 20:25 |
| Subject Sample # 2 | 0.139 | 20:26 |
| Breath Volume      | 1.753 | 20:26 |
| Air Blank          | 0.000 | 20:27 |
| Control Test       | 0.079 | 20:27 |
| Air Blank          | 0.000 | 20:27 |
| Diagnostics Check  | OK    | 20:27 |

---

Data Download Date/Time: 06/28/2022 14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SIMMONS,JUSTIN,M

BREATH TEST OPERATOR AGENCY: Niceville PD

ARRESTING OFFICER: BRONSON,THOMAS

ARRESTING OFFICER AGENCY: Niceville PD

LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1283438

DRY GAS EXPIRY DATE: 02/25/2023

VIOLATION CODE: DUI

ARREST TIME: 04:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: POULIN

DATE/TIME OF TEST: 06/06/2022 05:53:45

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 05:52 |            |
| Air Blank         | 0.000 | 05:52 |            |
| Control Test      | 0.078 | 05:52 |            |
| Air Blank         |       | 05:53 | RFI Detect |
| Air Blank         | 0.000 | 05:53 |            |

---

Data Download Date/Time: 06/10/2022 18:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Niceville PD  
ARRESTING OFFICER: URSRY,STEPHAN  
ARRESTING OFFICER AGENCY: Niceville PD  
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1283438  
DRY GAS EXPIRY DATE: 02/25/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: POULIN

DATE/TIME OF TEST: 06/06/2022 06:07:53

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 05:56 |                    |
| Air Blank         | 0.000 | 05:56 |                    |
| Control Test      | 0.079 | 05:57 |                    |
| Air Blank         | 0.000 | 05:57 |                    |
| Subject Sample    | # 1   | 06:00 | No Sample Provided |
| Breath Volume     | 0.000 | 06:00 |                    |
| Air Blank         | 0.000 | 06:01 |                    |
| Air Blank         | 0.000 | 06:03 |                    |
| Subject Sample    | # 2   | 06:06 | No Sample Provided |
| Breath Volume     | 0.000 | 06:06 |                    |
| Air Blank         | 0.000 | 06:06 |                    |
| Control Test      | 0.078 | 06:07 |                    |
| Air Blank         | 0.000 | 06:07 |                    |
| Diagnostics Check | OK    | 06:07 |                    |

Data Download Date/Time: 06/10/2022 18:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUNGE,JACOB,L  
BREATH TEST OPERATOR AGENCY: Niceville PD  
ARRESTING OFFICER: NEWCOMB,WILLIAM  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1283438  
DRY GAS EXPIRY DATE: 02/25/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 06/10/2022 08:40:58

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 08:39 |            |
| Air Blank         | 0.000 | 08:40 |            |
| Control Test      | 0.073 | 08:40 | RFI Detect |
| Air Blank         | 0.000 | 08:40 |            |

---

Data Download Date/Time: 06/10/2022 18:13

# **Florida Department of Law Enforcement**

## **Alcohol Testing Program**

### **Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RUNGE,JACOB,L

BREATH TEST OPERATOR AGENCY: Niceville PD

ARRESTING OFFICER: NEWCOMB,WILLIAM

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1283438

DRY GAS EXPIRY DATE: 02/25/2023

VIOLATION CODE: DUI

ARREST TIME: 07:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 06/10/2022 08:44:56

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 08:44 |            |
| Air Blank         |       | 08:44 | RFI Detect |
| Air Blank         | 0.000 | 08:44 |            |

---

Data Download Date/Time: 06/10/2022 18:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUNGE,JACOB,L  
BREATH TEST OPERATOR AGENCY: Niceville PD  
ARRESTING OFFICER: NEWCOMB,WILLIAM  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1283438  
DRY GAS EXPIRY DATE: 02/25/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 06/10/2022 08:48:47

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 08:47 |            |
| Air Blank         |       | 08:48 | RFI Detect |
| Air Blank         | 0.000 | 08:48 |            |

---

Data Download Date/Time: 06/10/2022 18:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RUNGE,JACOB,L

BREATH TEST OPERATOR AGENCY: Niceville PD

ARRESTING OFFICER: NEWCOMB,WILLIAM

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1283438

DRY GAS EXPIRY DATE: 02/25/2023

VIOLATION CODE: DUI

ARREST TIME: 07:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 06/10/2022 08:58:27

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 08:54 |            |
| Air Blank         | 0.000     | 08:54 |            |
| Control Test      | 0.077     | 08:55 |            |
| Air Blank         | 0.000     | 08:55 |            |
| Subject Sample    | # 1 0.126 | 08:56 |            |
| Breath Volume     | 3.425     | 08:56 |            |
| Air Blank         |           | 08:56 | RFI Detect |
| Air Blank         | 0.000     | 08:57 |            |
| Control Test      | 0.077     | 08:57 |            |
| Air Blank         |           | 08:57 | RFI Detect |
| Air Blank         | 0.000     | 08:58 |            |

---

Data Download Date/Time: 06/10/2022 18:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RUNGE,JACOB,L

BREATH TEST OPERATOR AGENCY: Niceville PD

ARRESTING OFFICER: NEWCOMB,WILLIAM

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1283438

DRY GAS EXPIRY DATE: 02/25/2023

VIOLATION CODE: DUI

ARREST TIME: 07:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 06/10/2022 09:10:27

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 09:08 |            |
| Air Blank         | 0.000 | 09:09 |            |
| Control Test      | 0.078 | 09:09 |            |
| Air Blank         |       | 09:09 | RFI Detect |
| Air Blank         | 0.000 | 09:10 |            |

---

Data Download Date/Time: 06/10/2022 18:13

# **Florida Department of Law Enforcement**

## **Alcohol Testing Program**

### **Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RUNGE,JACOB,L

BREATH TEST OPERATOR AGENCY: Niceville PD

ARRESTING OFFICER: NEWCOMB,WILLIAM

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 1283438

DRY GAS EXPIRY DATE: 02/25/2023

VIOLATION CODE: DUI

ARREST TIME: 07:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 06/10/2022 09:42:57

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 09:41 |                           |
| Air Blank         | 0.000 | 09:42 |                           |
| Control Test      | 0.065 | 09:42 | Control Outside Tolerance |
| Air Blank         | 0.000 | 09:42 |                           |

---

Data Download Date/Time: 06/10/2022 18:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUNGE,JACOB,L  
BREATH TEST OPERATOR AGENCY: Niceville PD  
ARRESTING OFFICER: NEWCOMB,WILLIAM  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 05/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1283438  
DRY GAS EXPIRY DATE: 02/25/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 06/10/2022 09:53:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:46 |
| Air Blank          | 0.000 | 09:47 |
| Control Test       | 0.077 | 09:47 |
| Air Blank          | 0.000 | 09:47 |
| Subject Sample # 1 | 0.111 | 09:48 |
| Breath Volume      | 4.167 | 09:48 |
| Air Blank          | 0.000 | 09:49 |
| Air Blank          | 0.000 | 09:51 |
| Subject Sample # 2 | 0.111 | 09:51 |
| Breath Volume      | 4.675 | 09:51 |
| Air Blank          | 0.000 | 09:52 |
| Control Test       | 0.077 | 09:52 |
| Air Blank          | 0.000 | 09:53 |
| Diagnostics Check  | OK    | 09:53 |

---

Data Download Date/Time: 06/10/2022 18:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRONSON, THOMAS, T  
BREATH TEST OPERATOR AGENCY: Niceville PD  
ARRESTING OFFICER: MACBETH, MATTHEW  
ARRESTING OFFICER AGENCY: Niceville PD  
LAST AGENCY INSPECTION: 06/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1283438  
DRY GAS EXPIRY DATE: 02/25/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 06/17/2022 21:27:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:21 |
| Air Blank          | 0.000 | 21:22 |
| Control Test       | 0.077 | 21:22 |
| Air Blank          | 0.000 | 21:22 |
| Subject Sample # 1 | 0.232 | 21:23 |
| Breath Volume      | 2.019 | 21:23 |
| Air Blank          | 0.000 | 21:24 |
| Air Blank          | 0.000 | 21:25 |
| Subject Sample # 2 | 0.225 | 21:26 |
| Breath Volume      | 2.187 | 21:26 |
| Air Blank          | 0.000 | 21:26 |
| Control Test       | 0.077 | 21:27 |
| Air Blank          | 0.000 | 21:27 |
| Diagnostics Check  | OK    | 21:27 |

---

Data Download Date/Time: 06/24/2022 15:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAY,WILLIAM,P  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: MAY,WILLIAM  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 05/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: AVERY

DATE/TIME OF TEST: 06/04/2022 09:13:51

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 09:11 |                      |
| Air Blank         | 0.000 | 09:11 |                      |
| Control Test      | 0.081 | 09:11 |                      |
| Air Blank         | 0.000 | 09:12 |                      |
| Subject Sample    | # 1   | 09:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 09:12 |                      |
| Air Blank         | 0.000 | 09:12 |                      |
| Control Test      | 0.082 | 09:13 |                      |
| Air Blank         | 0.000 | 09:13 |                      |
| Diagnostics Check | OK    | 09:13 |                      |

---

Data Download Date/Time: 06/24/2022 09:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VENTRY, WILLIS, E  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: ADAMS, CLINT  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 05/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: MRAZ

DATE/TIME OF TEST: 06/05/2022 22:30:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:22 |
| Air Blank          | 0.000 | 22:22 |
| Control Test       | 0.081 | 22:23 |
| Air Blank          | 0.000 | 22:23 |
| Subject Sample # 1 | 0.160 | 22:25 |
| Breath Volume      | 1.355 | 22:25 |
| Air Blank          | 0.000 | 22:25 |
| Air Blank          | 0.000 | 22:27 |
| Subject Sample # 2 | 0.165 | 22:29 |
| Breath Volume      | 1.105 | 22:29 |
| Air Blank          | 0.000 | 22:29 |
| Control Test       | 0.081 | 22:30 |
| Air Blank          | 0.000 | 22:30 |
| Diagnostics Check  | OK    | 22:30 |

---

Data Download Date/Time: 06/24/2022 09:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GORDON,JON,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: GORDON,JON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 06/10/2022 22:19:49

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 22:08 |                    |
| Air Blank         | 0.000 | 22:08 |                    |
| Control Test      | 0.081 | 22:09 |                    |
| Air Blank         | 0.000 | 22:09 |                    |
| Subject Sample    | # 1   | 22:12 | No Sample Provided |
| Breath Volume     | 0.000 | 22:12 |                    |
| Air Blank         | 0.000 | 22:13 |                    |
| Air Blank         | 0.000 | 22:15 |                    |
| Subject Sample    | # 2   | 22:18 | No Sample Provided |
| Breath Volume     | 0.000 | 22:18 |                    |
| Air Blank         | 0.000 | 22:18 |                    |
| Control Test      | 0.082 | 22:19 |                    |
| Air Blank         | 0.000 | 22:19 |                    |
| Diagnostics Check | OK    | 22:19 |                    |

---

Data Download Date/Time: 06/24/2022 09:30



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLISON,STEVEN,L  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: WILLIAMS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 05/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: CHISHOLM

DATE/TIME OF TEST: 06/11/2022 05:14:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:08 |
| Air Blank          | 0.000 | 05:08 |
| Control Test       | 0.081 | 05:09 |
| Air Blank          | 0.000 | 05:09 |
| Subject Sample # 1 | 0.202 | 05:10 |
| Breath Volume      | 2.078 | 05:10 |
| Air Blank          | 0.000 | 05:10 |
| Air Blank          | 0.000 | 05:12 |
| Subject Sample # 2 | 0.197 | 05:13 |
| Breath Volume      | 1.917 | 05:13 |
| Air Blank          | 0.000 | 05:13 |
| Control Test       | 0.081 | 05:14 |
| Air Blank          | 0.000 | 05:14 |
| Diagnostics Check  | OK    | 05:14 |

---

Data Download Date/Time: 06/24/2022 09:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAY,WILLIAM,P  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: MAY,WILLIAM  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 05/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 06/18/2022 00:56:43

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 00:45 |                |
| Air Blank         | 0.000     | 00:45 |                |
| Control Test      | 0.081     | 00:45 |                |
| Air Blank         | 0.000     | 00:46 |                |
| Subject Sample    | # 1 0.186 | 00:49 | Volume Not Met |
| Breath Volume     | 0.035     | 00:49 |                |
| Air Blank         | 0.000     | 00:50 |                |
| Air Blank         | 0.000     | 00:52 |                |
| Subject Sample    | # 2 0.154 | 00:55 | Volume Not Met |
| Breath Volume     | 0.199     | 00:55 |                |
| Air Blank         | 0.000     | 00:55 |                |
| Control Test      | 0.080     | 00:56 |                |
| Air Blank         | 0.000     | 00:56 |                |
| Diagnostics Check | OK        | 00:56 |                |

---

Data Download Date/Time: 06/24/2022 09:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAY,WILLIAM,P  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: MAY,WILLIAM  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 05/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 06/18/2022 01:08:36

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 00:59 |                      |
| Air Blank         | 0.000     | 00:59 |                      |
| Control Test      | 0.080     | 01:00 |                      |
| Air Blank         | 0.000     | 01:00 |                      |
| Subject Sample    | # 1 0.171 | 01:03 |                      |
| Breath Volume     | 1.187     | 01:03 |                      |
| Air Blank         | 0.000     | 01:04 |                      |
| Air Blank         | 0.000     | 01:05 |                      |
| Subject Sample    | # 2       | 01:07 | Subject Test Refused |
| Breath Volume     | 0.296     | 01:07 |                      |
| Air Blank         | 0.000     | 01:07 |                      |
| Control Test      | 0.079     | 01:07 |                      |
| Air Blank         | 0.000     | 01:08 |                      |
| Diagnostics Check | OK        | 01:08 |                      |

---

Data Download Date/Time: 06/24/2022 09:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAY,WILLIAM,P  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: MAY,WILLIAM  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 05/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HEYMAN

DATE/TIME OF TEST: 06/19/2022 03:27:41

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:24 |                      |
| Air Blank         | 0.000 | 03:25 |                      |
| Control Test      | 0.082 | 03:25 |                      |
| Air Blank         | 0.000 | 03:26 |                      |
| Subject Sample    | # 1   | 03:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:26 |                      |
| Air Blank         | 0.000 | 03:26 |                      |
| Control Test      | 0.082 | 03:27 |                      |
| Air Blank         | 0.000 | 03:27 |                      |
| Diagnostics Check | OK    | 03:27 |                      |

---

Data Download Date/Time: 06/24/2022 09:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAY,WILLIAM,P  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: MAY,WILLIAM  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 05/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ALVARENGA MELGAR

DATE/TIME OF TEST: 06/19/2022 22:10:05

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:07 |                      |
| Air Blank         | 0.000 | 22:07 |                      |
| Control Test      | 0.080 | 22:07 |                      |
| Air Blank         | 0.000 | 22:08 |                      |
| Subject Sample    | # 1   | 22:08 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:08 |                      |
| Air Blank         | 0.000 | 22:09 |                      |
| Control Test      | 0.081 | 22:09 |                      |
| Air Blank         | 0.000 | 22:10 |                      |
| Diagnostics Check | OK    | 22:10 |                      |

---

Data Download Date/Time: 06/24/2022 09:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOROS,CHUWAN,  
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD  
ARRESTING OFFICER: BOROS,CHUWAN  
ARRESTING OFFICER AGENCY: DeFuniak Springs PD  
LAST AGENCY INSPECTION: 05/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/20/2022 23:20:15

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:14 |                      |
| Air Blank         | 0.000 | 23:15 |                      |
| Control Test      | 0.081 | 23:15 |                      |
| Air Blank         | 0.000 | 23:16 |                      |
| Subject Sample    | # 1   | 23:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:18 |                      |
| Air Blank         | 0.000 | 23:19 |                      |
| Control Test      | 0.083 | 23:19 |                      |
| Air Blank         | 0.000 | 23:20 |                      |
| Diagnostics Check | OK    | 23:20 |                      |

---

Data Download Date/Time: 06/24/2022 09:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOST,CRAIG,R  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: HANNAWALT,JAMES  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: INCORVATI

DATE/TIME OF TEST: 06/25/2022 23:42:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:36 |
| Air Blank          | 0.000 | 23:36 |
| Control Test       | 0.081 | 23:37 |
| Air Blank          | 0.000 | 23:37 |
| Subject Sample # 1 | 0.159 | 23:38 |
| Breath Volume      | 1.660 | 23:38 |
| Air Blank          | 0.000 | 23:39 |
| Air Blank          | 0.000 | 23:40 |
| Subject Sample # 2 | 0.167 | 23:41 |
| Breath Volume      | 1.621 | 23:41 |
| Air Blank          | 0.000 | 23:41 |
| Control Test       | 0.081 | 23:42 |
| Air Blank          | 0.000 | 23:42 |
| Diagnostics Check  | OK    | 23:42 |

---

Data Download Date/Time: 07/29/2022 10:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, ANDRES, L  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: ROVIROSA, ARMANDO  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 06/03/2022 18:19:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:11 |
| Air Blank          | 0.000 | 18:12 |
| Control Test       | 0.079 | 18:12 |
| Air Blank          | 0.000 | 18:12 |
| Subject Sample # 1 | 0.124 | 18:13 |
| Breath Volume      | 1.710 | 18:13 |
| Air Blank          | 0.000 | 18:14 |
| Air Blank          | 0.000 | 18:16 |
| Subject Sample # 2 | 0.111 | 18:18 |
| Breath Volume      | 2.757 | 18:18 |
| Air Blank          | 0.000 | 18:18 |
| Control Test       | 0.079 | 18:19 |
| Air Blank          | 0.000 | 18:19 |
| Diagnostics Check  | OK    | 18:19 |

---

Data Download Date/Time: 06/28/2022 14:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MURTAGH,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: MURTAGH,ROBERT  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: VARRI

DATE/TIME OF TEST: 06/03/2022 22:00:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:52 |                      |
| Air Blank         | 0.000 | 21:52 |                      |
| Control Test      | 0.079 | 21:52 |                      |
| Air Blank         | 0.000 | 21:53 |                      |
| Subject Sample    | # 1   | 21:56 | No Sample Provided   |
| Breath Volume     | 0.000 | 21:56 |                      |
| Air Blank         | 0.000 | 21:57 |                      |
| Air Blank         | 0.000 | 21:59 |                      |
| Subject Sample    | # 2   | 21:59 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:59 |                      |
| Air Blank         | 0.000 | 21:59 |                      |
| Control Test      | 0.079 | 22:00 |                      |
| Air Blank         | 0.000 | 22:00 |                      |
| Diagnostics Check | OK    | 22:00 |                      |

---

Data Download Date/Time: 06/28/2022 14:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 06/04/2022 01:56:18

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:52 |                      |
| Air Blank         | 0.000 | 01:53 |                      |
| Control Test      | 0.079 | 01:53 |                      |
| Air Blank         | 0.000 | 01:54 |                      |
| Subject Sample    | # 1   | 01:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:54 |                      |
| Air Blank         | 0.000 | 01:55 |                      |
| Control Test      | 0.079 | 01:55 |                      |
| Air Blank         | 0.000 | 01:56 |                      |
| Diagnostics Check | OK    | 01:56 |                      |

---

Data Download Date/Time: 06/28/2022 14:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: CASIAN,SKYLAR  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:09

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 06/04/2022 15:36:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:29 |
| Air Blank          | 0.000 | 15:30 |
| Control Test       | 0.077 | 15:30 |
| Air Blank          | 0.000 | 15:31 |
| Subject Sample # 1 | 0.180 | 15:31 |
| Breath Volume      | 3.027 | 15:31 |
| Air Blank          | 0.000 | 15:32 |
| Air Blank          | 0.000 | 15:34 |
| Subject Sample # 2 | 0.187 | 15:34 |
| Breath Volume      | 2.210 | 15:34 |
| Air Blank          | 0.000 | 15:35 |
| Control Test       | 0.078 | 15:35 |
| Air Blank          | 0.000 | 15:35 |
| Diagnostics Check  | OK    | 15:36 |

---

Data Download Date/Time: 06/28/2022 14:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: TYSON

DATE/TIME OF TEST: 06/04/2022 22:19:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:13 |
| Air Blank          | 0.000 | 22:14 |
| Control Test       | 0.079 | 22:14 |
| Air Blank          | 0.000 | 22:14 |
| Subject Sample # 1 | 0.000 | 22:15 |
| Breath Volume      | 3.234 | 22:15 |
| Air Blank          | 0.000 | 22:15 |
| Air Blank          | 0.000 | 22:17 |
| Subject Sample # 2 | 0.000 | 22:18 |
| Breath Volume      | 2.152 | 22:18 |
| Air Blank          | 0.000 | 22:18 |
| Control Test       | 0.079 | 22:19 |
| Air Blank          | 0.000 | 22:19 |
| Diagnostics Check  | OK    | 22:19 |

---

Data Download Date/Time: 06/28/2022 14:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BRAZELL

DATE/TIME OF TEST: 06/05/2022 03:23:13

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:20 |                      |
| Air Blank         | 0.000 | 03:20 |                      |
| Control Test      | 0.080 | 03:21 |                      |
| Air Blank         | 0.000 | 03:21 |                      |
| Subject Sample    | # 1   | 03:21 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:21 |                      |
| Air Blank         | 0.000 | 03:22 |                      |
| Control Test      | 0.080 | 03:22 |                      |
| Air Blank         | 0.000 | 03:23 |                      |
| Diagnostics Check | OK    | 03:23 |                      |

---

Data Download Date/Time: 06/28/2022 14:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 06/08/2022 23:50:40

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:47 |                      |
| Air Blank         | 0.000 | 23:48 |                      |
| Control Test      | 0.079 | 23:48 |                      |
| Air Blank         | 0.000 | 23:49 |                      |
| Subject Sample    | # 1   | 23:49 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:49 |                      |
| Air Blank         | 0.000 | 23:49 |                      |
| Control Test      | 0.080 | 23:50 |                      |
| Air Blank         | 0.000 | 23:50 |                      |
| Diagnostics Check | OK    | 23:50 |                      |

---

Data Download Date/Time: 06/28/2022 14:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 06/10/2022 02:47:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Control Test       | 0.079 | 02:41 |
| Air Blank          | 0.000 | 02:42 |
| Subject Sample # 1 | 0.146 | 02:42 |
| Breath Volume      | 4.054 | 02:42 |
| Air Blank          | 0.000 | 02:43 |
| Air Blank          | 0.000 | 02:45 |
| Subject Sample # 2 | 0.144 | 02:45 |
| Breath Volume      | 4.156 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Control Test       | 0.079 | 02:46 |
| Air Blank          | 0.000 | 02:47 |
| Diagnostics Check  | OK    | 02:47 |

---

Data Download Date/Time: 06/28/2022 14:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: GALVEZ DURAN

DATE/TIME OF TEST: 06/11/2022 00:02:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:56 |
| Air Blank          | 0.000 | 23:56 |
| Control Test       | 0.079 | 23:57 |
| Air Blank          | 0.000 | 23:57 |
| Subject Sample # 2 | 0.231 | 23:58 |
| Breath Volume      | 3.289 | 23:58 |
| Air Blank          | 0.000 | 23:58 |
| Air Blank          | 0.000 | 00:00 |
| Subject Sample # 1 | 0.229 | 00:00 |
| Breath Volume      | 3.382 | 00:00 |
| Air Blank          | 0.000 | 00:01 |
| Control Test       | 0.079 | 00:01 |
| Air Blank          | 0.000 | 00:02 |
| Diagnostics Check  | OK    | 00:02 |

---

Data Download Date/Time: 06/28/2022 14:53



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MURTAGH,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: MURTAGH,ROBERT  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 06/12/2022 00:02:08

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:58 |                      |
| Air Blank         | 0.000 | 23:59 |                      |
| Control Test      | 0.080 | 23:59 |                      |
| Air Blank         | 0.000 | 00:00 |                      |
| Subject Sample    | # 1   | 00:00 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:00 |                      |
| Air Blank         | 0.000 | 00:01 |                      |
| Control Test      | 0.080 | 00:01 |                      |
| Air Blank         | 0.000 | 00:02 |                      |
| Diagnostics Check | OK    | 00:02 |                      |

---

Data Download Date/Time: 06/28/2022 14:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CALCAGNO, MERCEDES, T  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: BUNTING, JESSICA  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: VELASQUEZ HERNANDEZ

DATE/TIME OF TEST: 06/12/2022 20:08:10

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 19:54 |                |
| Air Blank         | 0.000     | 19:54 |                |
| Control Test      | 0.080     | 19:55 |                |
| Air Blank         | 0.000     | 19:55 |                |
| Subject Sample    | # 1 0.216 | 19:58 | Volume Not Met |
| Breath Volume     | 0.199     | 19:58 |                |
| Air Blank         | 0.000     | 19:59 |                |
| Air Blank         | 0.000     | 20:01 |                |
| Subject Sample    | # 2 0.237 | 20:02 |                |
| Breath Volume     | 1.386     | 20:02 |                |
| Air Blank         | 0.000     | 20:02 |                |
| Air Blank         | 0.000     | 20:04 |                |
| Subject Sample    | # 3 0.222 | 20:06 |                |
| Breath Volume     | 1.585     | 20:06 |                |
| Air Blank         | 0.000     | 20:07 |                |
| Control Test      | 0.078     | 20:07 |                |
| Air Blank         | 0.000     | 20:08 |                |
| Diagnostics Check | OK        | 20:08 |                |

---

Data Download Date/Time: 06/28/2022 14:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: OSCEOLA

DATE/TIME OF TEST: 06/14/2022 03:11:02

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:08 |                      |
| Air Blank         | 0.000 | 03:08 |                      |
| Control Test      | 0.079 | 03:08 |                      |
| Air Blank         | 0.000 | 03:09 |                      |
| Subject Sample    | # 1   | 03:09 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:09 |                      |
| Air Blank         | 0.000 | 03:10 |                      |
| Control Test      | 0.079 | 03:10 |                      |
| Air Blank         | 0.000 | 03:10 |                      |
| Diagnostics Check | OK    | 03:11 |                      |

---

Data Download Date/Time: 06/28/2022 14:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 06/16/2022 19:50:00

|                    |       |       |                      |
|--------------------|-------|-------|----------------------|
| Diagnostics Check  | OK    | 19:41 |                      |
| Air Blank          | 0.000 | 19:42 |                      |
| Control Test       | 0.080 | 19:42 |                      |
| Air Blank          | 0.000 | 19:43 |                      |
| Subject Sample # 1 | 0.165 | 19:44 |                      |
| Breath Volume      | 2.355 | 19:44 |                      |
| Air Blank          | 0.000 | 19:44 |                      |
| Air Blank          | 0.000 | 19:46 |                      |
| Subject Sample # 2 |       | 19:48 | Subject Test Refused |
| Breath Volume      | 0.000 | 19:48 |                      |
| Air Blank          | 0.000 | 19:49 |                      |
| Control Test       | 0.079 | 19:49 |                      |
| Air Blank          | 0.000 | 19:49 |                      |
| Diagnostics Check  | OK    | 19:50 |                      |

Data Download Date/Time: 06/28/2022 14:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: COURSON

DATE/TIME OF TEST: 06/17/2022 00:19:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:13 |
| Air Blank          | 0.000 | 00:13 |
| Control Test       | 0.080 | 00:14 |
| Air Blank          | 0.000 | 00:14 |
| Subject Sample # 1 | 0.000 | 00:15 |
| Breath Volume      | 2.800 | 00:15 |
| Air Blank          | 0.000 | 00:15 |
| Air Blank          | 0.000 | 00:17 |
| Subject Sample # 2 | 0.000 | 00:18 |
| Breath Volume      | 2.539 | 00:18 |
| Air Blank          | 0.000 | 00:18 |
| Control Test       | 0.079 | 00:18 |
| Air Blank          | 0.000 | 00:19 |
| Diagnostics Check  | OK    | 00:19 |

---

Data Download Date/Time: 06/28/2022 14:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: MORALES BAUTISTA

DATE/TIME OF TEST: 06/17/2022 21:16:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:09 |
| Air Blank          | 0.000 | 21:10 |
| Control Test       | 0.080 | 21:10 |
| Air Blank          | 0.000 | 21:11 |
| Subject Sample # 1 | 0.058 | 21:11 |
| Breath Volume      | 2.562 | 21:11 |
| Air Blank          | 0.000 | 21:12 |
| Air Blank          | 0.000 | 21:14 |
| Subject Sample # 2 | 0.053 | 21:15 |
| Breath Volume      | 2.281 | 21:15 |
| Air Blank          | 0.000 | 21:15 |
| Control Test       | 0.080 | 21:16 |
| Air Blank          | 0.000 | 21:16 |
| Diagnostics Check  | OK    | 21:16 |

---

Data Download Date/Time: 06/28/2022 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 17:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 06/18/2022 17:54:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:46 |
| Air Blank          | 0.000 | 17:47 |
| Control Test       | 0.080 | 17:47 |
| Air Blank          | 0.000 | 17:48 |
| Subject Sample # 1 | 0.187 | 17:49 |
| Breath Volume      | 2.265 | 17:49 |
| Air Blank          | 0.000 | 17:50 |
| Air Blank          | 0.000 | 17:52 |
| Subject Sample # 2 | 0.180 | 17:53 |
| Breath Volume      | 3.277 | 17:53 |
| Air Blank          | 0.000 | 17:53 |
| Control Test       | 0.080 | 17:54 |
| Air Blank          | 0.000 | 17:54 |
| Diagnostics Check  | OK    | 17:54 |

---

Data Download Date/Time: 06/28/2022 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/18/2022 23:04:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:58 |
| Air Blank          | 0.000 | 22:59 |
| Control Test       | 0.079 | 22:59 |
| Air Blank          | 0.000 | 22:59 |
| Subject Sample # 1 | 0.292 | 23:00 |
| Breath Volume      | 1.820 | 23:00 |
| Air Blank          | 0.000 | 23:00 |
| Air Blank          | 0.000 | 23:02 |
| Subject Sample # 2 | 0.292 | 23:03 |
| Breath Volume      | 2.539 | 23:03 |
| Air Blank          | 0.000 | 23:03 |
| Control Test       | 0.078 | 23:04 |
| Air Blank          | 0.000 | 23:04 |
| Diagnostics Check  | OK    | 23:04 |

---

Data Download Date/Time: 06/28/2022 14:54



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: MICCO

DATE/TIME OF TEST: 06/20/2022 02:06:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:59 |
| Air Blank          | 0.000 | 02:00 |
| Control Test       | 0.079 | 02:00 |
| Air Blank          | 0.000 | 02:00 |
| Subject Sample # 1 | 0.189 | 02:01 |
| Breath Volume      | 2.390 | 02:01 |
| Air Blank          | 0.000 | 02:01 |
| Air Blank          | 0.000 | 02:03 |
| Subject Sample # 2 | 0.179 | 02:04 |
| Breath Volume      | 1.886 | 02:04 |
| Air Blank          | 0.000 | 02:05 |
| Control Test       | 0.079 | 02:05 |
| Air Blank          | 0.000 | 02:05 |
| Diagnostics Check  | OK    | 02:06 |

---

Data Download Date/Time: 06/28/2022 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLLOCK, STEVEN, D  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: POLLOCK, STEVEN  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: CLAY

DATE/TIME OF TEST: 06/24/2022 22:07:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:00 |
| Air Blank          | 0.000 | 22:01 |
| Control Test       | 0.080 | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Subject Sample # 1 | 0.133 | 22:02 |
| Breath Volume      | 2.769 | 22:02 |
| Air Blank          | 0.000 | 22:03 |
| Air Blank          | 0.000 | 22:05 |
| Subject Sample # 2 | 0.147 | 22:05 |
| Breath Volume      | 3.097 | 22:05 |
| Air Blank          | 0.000 | 22:06 |
| Control Test       | 0.080 | 22:06 |
| Air Blank          | 0.000 | 22:07 |
| Diagnostics Check  | OK    | 22:07 |

---

Data Download Date/Time: 06/28/2022 14:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLLOCK,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: POLLOCK,STEVEN  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: LEVI

DATE/TIME OF TEST: 06/25/2022 01:39:18

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:36 |                      |
| Air Blank         | 0.000 | 01:36 |                      |
| Control Test      | 0.079 | 01:37 |                      |
| Air Blank         | 0.000 | 01:37 |                      |
| Subject Sample    | # 1   | 01:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:37 |                      |
| Air Blank         | 0.000 | 01:38 |                      |
| Control Test      | 0.080 | 01:38 |                      |
| Air Blank         | 0.000 | 01:39 |                      |
| Diagnostics Check | OK    | 01:39 |                      |

---

Data Download Date/Time: 06/28/2022 14:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MURTAGH,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: MURTAGH,ROBERT  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 05/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: FRALIX

DATE/TIME OF TEST: 06/25/2022 03:38:22

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:33 |                      |
| Air Blank         | 0.000 | 03:34 |                      |
| Control Test      | 0.080 | 03:34 |                      |
| Air Blank         | 0.000 | 03:35 |                      |
| Subject Sample    | # 1   | 03:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:37 |                      |
| Air Blank         | 0.000 | 03:37 |                      |
| Control Test      | 0.080 | 03:37 |                      |
| Air Blank         | 0.000 | 03:38 |                      |
| Diagnostics Check | OK    | 03:38 |                      |

---

Data Download Date/Time: 06/28/2022 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WRIGHT, MARK, A  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: DOVAL, R  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: SEOK

DATE/TIME OF TEST: 06/28/2022 01:45:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:38 |
| Air Blank          | 0.000 | 01:38 |
| Control Test       | 0.079 | 01:38 |
| Air Blank          | 0.000 | 01:39 |
| Subject Sample # 1 | 0.197 | 01:40 |
| Breath Volume      | 1.902 | 01:40 |
| Air Blank          | 0.000 | 01:41 |
| Air Blank          | 0.000 | 01:43 |
| Subject Sample # 2 | 0.185 | 01:43 |
| Breath Volume      | 2.167 | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Control Test       | 0.078 | 01:44 |
| Air Blank          | 0.000 | 01:45 |
| Diagnostics Check  | OK    | 01:45 |

---

Data Download Date/Time: 06/28/2022 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,MISEI,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: ESQUIVEL,MISEI  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402345249  
DRY GAS EXPIRY DATE: 02/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: LOLLIS

DATE/TIME OF TEST: 06/29/2022 02:45:33

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:37 |                      |
| Air Blank         | 0.000 | 02:37 |                      |
| Control Test      | 0.079 | 02:37 |                      |
| Air Blank         | 0.000 | 02:38 |                      |
| Subject Sample    | # 1   | 02:41 | No Sample Provided   |
| Breath Volume     | 0.000 | 02:41 |                      |
| Air Blank         | 0.000 | 02:41 |                      |
| Air Blank         | 0.000 | 02:43 |                      |
| Subject Sample    | # 2   | 02:44 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:44 |                      |
| Air Blank         | 0.000 | 02:44 |                      |
| Control Test      | 0.079 | 02:44 |                      |
| Air Blank         | 0.000 | 02:45 |                      |
| Diagnostics Check | OK    | 02:45 |                      |

---

Data Download Date/Time: 07/13/2022 08:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: HELSETH,KEVIN  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A1  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PELLICIA

DATE/TIME OF TEST: 06/21/2022 01:18:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:07 |
| Air Blank          | 0.000 | 01:08 |
| Control Test       | 0.080 | 01:08 |
| Air Blank          | 0.000 | 01:08 |
| Subject Sample # 1 | 0.196 | 01:11 |
| Breath Volume      | 1.210 | 01:11 |
| Air Blank          | 0.000 | 01:12 |
| Air Blank          | 0.000 | 01:14 |
| Subject Sample # 2 | 0.195 | 01:17 |
| Breath Volume      | 1.476 | 01:17 |
| Air Blank          | 0.000 | 01:17 |
| Control Test       | 0.078 | 01:18 |
| Air Blank          | 0.000 | 01:18 |
| Diagnostics Check  | OK    | 01:18 |

---

Data Download Date/Time: 07/18/2022 17:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUBITSCHK,LEONARD,H  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: FALEIROS,JULIANO  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 08622080A1  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: EGAN

DATE/TIME OF TEST: 06/24/2022 22:02:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:55 |
| Air Blank          | 0.000 | 21:56 |
| Control Test       | 0.080 | 21:56 |
| Air Blank          | 0.000 | 21:57 |
| Subject Sample # 1 | 0.000 | 21:57 |
| Breath Volume      | 2.042 | 21:57 |
| Air Blank          | 0.000 | 21:58 |
| Air Blank          | 0.000 | 22:00 |
| Subject Sample # 2 | 0.000 | 22:00 |
| Breath Volume      | 2.187 | 22:00 |
| Air Blank          | 0.000 | 22:01 |
| Control Test       | 0.079 | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Diagnostics Check  | OK    | 22:02 |

---

Data Download Date/Time: 07/18/2022 17:45



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOMCZYK,STENLY,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: SIMPSON,PATRICK  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A1  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: DOLNICK

DATE/TIME OF TEST: 06/25/2022 00:03:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:55 |
| Air Blank          | 0.000 | 23:55 |
| Control Test       | 0.080 | 23:56 |
| Air Blank          | 0.000 | 23:56 |
| Subject Sample # 2 | 0.188 | 23:58 |
| Breath Volume      | 1.695 | 23:58 |
| Air Blank          | 0.000 | 23:59 |
| Air Blank          | 0.000 | 00:00 |
| Subject Sample # 1 | 0.173 | 00:01 |
| Breath Volume      | 1.902 | 00:01 |
| Air Blank          | 0.000 | 00:02 |
| Control Test       | 0.078 | 00:03 |
| Air Blank          | 0.000 | 00:03 |
| Diagnostics Check  | OK    | 00:03 |

---

Data Download Date/Time: 07/18/2022 17:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: HERNANDEZ,YAMILET  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A1  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/25/2022 20:57:50

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:48 |                      |
| Air Blank         | 0.000 | 20:48 |                      |
| Control Test      | 0.080 | 20:48 |                      |
| Air Blank         | 0.000 | 20:49 |                      |
| Subject Sample    | # 1   | 20:52 | No Sample Provided   |
| Breath Volume     | 0.000 | 20:52 |                      |
| Air Blank         | 0.000 | 20:53 |                      |
| Air Blank         | 0.000 | 20:54 |                      |
| Subject Sample    | # 2   | 20:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:56 |                      |
| Air Blank         | 0.000 | 20:56 |                      |
| Control Test      | 0.079 | 20:57 |                      |
| Air Blank         | 0.000 | 20:57 |                      |
| Diagnostics Check | OK    | 20:57 |                      |

---

Data Download Date/Time: 07/18/2022 17:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNOZ,CASANDRA,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: KUBITSCHK,LEONARD  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A1  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: FELDMAN

DATE/TIME OF TEST: 06/30/2022 03:00:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:53 |
| Air Blank          | 0.000 | 02:54 |
| Control Test       | 0.080 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Subject Sample # 1 | 0.117 | 02:56 |
| Breath Volume      | 4.339 | 02:56 |
| Air Blank          | 0.000 | 02:56 |
| Air Blank          | 0.000 | 02:58 |
| Subject Sample # 2 | 0.108 | 02:59 |
| Breath Volume      | 3.980 | 02:59 |
| Air Blank          | 0.000 | 03:00 |
| Control Test       | 0.080 | 03:00 |
| Air Blank          | 0.000 | 03:00 |
| Diagnostics Check  | OK    | 03:00 |

---

Data Download Date/Time: 07/18/2022 17:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HISS,J,M  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: MILLER,T  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GENOVESE

DATE/TIME OF TEST: 06/03/2022 03:22:57

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:20 |                      |
| Air Blank         | 0.000 | 03:20 |                      |
| Control Test      | 0.080 | 03:20 |                      |
| Air Blank         | 0.000 | 03:21 |                      |
| Subject Sample    | # 1   | 03:21 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:21 |                      |
| Air Blank         | 0.000 | 03:22 |                      |
| Control Test      | 0.081 | 03:22 |                      |
| Air Blank         | 0.000 | 03:22 |                      |
| Diagnostics Check | OK    | 03:22 |                      |

---

Data Download Date/Time: 06/17/2022 16:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AGUILAR,RODRIGO,  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: AGUILAR,RODRIGO  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: PEEK

DATE/TIME OF TEST: 06/03/2022 22:47:18

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 22:35 |                    |
| Air Blank         | 0.000 | 22:36 |                    |
| Control Test      | 0.080 | 22:36 |                    |
| Air Blank         | 0.000 | 22:36 |                    |
| Subject Sample    | # 1   | 22:40 | No Sample Provided |
| Breath Volume     | 0.000 | 22:40 |                    |
| Air Blank         | 0.000 | 22:40 |                    |
| Air Blank         | 0.000 | 22:42 |                    |
| Subject Sample    | # 2   | 0.114 | 22:45              |
| Breath Volume     | 0.449 | 22:45 | Volume Not Met     |
| Air Blank         | 0.000 | 22:46 |                    |
| Control Test      | 0.081 | 22:46 |                    |
| Air Blank         | 0.000 | 22:47 |                    |
| Diagnostics Check | OK    | 22:47 |                    |

---

Data Download Date/Time: 06/17/2022 16:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AGUILAR, RODRIGO,  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: AGUILAR, AGUILAR  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: PEEK

DATE/TIME OF TEST: 06/03/2022 23:04:03

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:01 |                      |
| Air Blank         | 0.000 | 23:01 |                      |
| Control Test      | 0.079 | 23:01 |                      |
| Air Blank         | 0.000 | 23:02 |                      |
| Subject Sample    | # 1   | 23:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:02 |                      |
| Air Blank         | 0.000 | 23:03 |                      |
| Control Test      | 0.079 | 23:03 |                      |
| Air Blank         | 0.000 | 23:03 |                      |
| Diagnostics Check | OK    | 23:04 |                      |

---

Data Download Date/Time: 06/17/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HISS,J,M  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: PEACH,K  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: BARRON

DATE/TIME OF TEST: 06/07/2022 04:31:05

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:28 |                      |
| Air Blank         | 0.000 | 04:28 |                      |
| Control Test      | 0.080 | 04:29 |                      |
| Air Blank         | 0.000 | 04:29 |                      |
| Subject Sample    | # 1   | 04:29 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:29 |                      |
| Air Blank         | 0.000 | 04:30 |                      |
| Control Test      | 0.081 | 04:30 |                      |
| Air Blank         | 0.000 | 04:31 |                      |
| Diagnostics Check | OK    | 04:31 |                      |

---

Data Download Date/Time: 06/17/2022 16:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HISS,J,M  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: GREELEY,C  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: KOHL

DATE/TIME OF TEST: 06/07/2022 23:35:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:28 |
| Air Blank          | 0.000 | 23:28 |
| Control Test       | 0.080 | 23:29 |
| Air Blank          | 0.000 | 23:29 |
| Subject Sample # 1 | 0.325 | 23:30 |
| Breath Volume      | 2.031 | 23:30 |
| Air Blank          | 0.000 | 23:31 |
| Air Blank          | 0.000 | 23:32 |
| Subject Sample # 2 | 0.312 | 23:33 |
| Breath Volume      | 2.062 | 23:33 |
| Air Blank          | 0.000 | 23:34 |
| Control Test       | 0.079 | 23:34 |
| Air Blank          | 0.000 | 23:34 |
| Diagnostics Check  | OK    | 23:35 |

---

Data Download Date/Time: 06/17/2022 16:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHILE,C,J  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: MOON,C  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/09/2022 20:07:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:00 |
| Air Blank          | 0.000 | 20:00 |
| Control Test       | 0.080 | 20:01 |
| Air Blank          | 0.000 | 20:01 |
| Subject Sample # 1 | 0.220 | 20:02 |
| Breath Volume      | 1.750 | 20:02 |
| Air Blank          | 0.000 | 20:03 |
| Air Blank          | 0.000 | 20:04 |
| Subject Sample # 2 | 0.219 | 20:05 |
| Breath Volume      | 1.769 | 20:05 |
| Air Blank          | 0.000 | 20:06 |
| Control Test       | 0.079 | 20:06 |
| Air Blank          | 0.000 | 20:06 |
| Diagnostics Check  | OK    | 20:07 |

---

Data Download Date/Time: 06/17/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEARSE,A,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: ANTOSIA,E  
ARRESTING OFFICER AGENCY: Sebastian PD  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ZEMANOVIC

DATE/TIME OF TEST: 06/11/2022 02:31:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:29 |                      |
| Air Blank         | 0.000 | 02:29 |                      |
| Control Test      | 0.081 | 02:29 |                      |
| Air Blank         | 0.000 | 02:30 |                      |
| Subject Sample    | # 1   | 02:30 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:30 |                      |
| Air Blank         | 0.000 | 02:31 |                      |
| Control Test      | 0.080 | 02:31 |                      |
| Air Blank         | 0.000 | 02:31 |                      |
| Diagnostics Check | OK    | 02:31 |                      |

---

Data Download Date/Time: 06/17/2022 16:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,.  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: MILLER,DEPUTY  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LEBLANC

DATE/TIME OF TEST: 06/12/2022 00:08:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:01 |
| Air Blank          | 0.000 | 00:01 |
| Control Test       | 0.081 | 00:02 |
| Air Blank          | 0.000 | 00:02 |
| Subject Sample # 1 | 0.264 | 00:03 |
| Breath Volume      | 2.246 | 00:03 |
| Air Blank          | 0.000 | 00:03 |
| Air Blank          | 0.000 | 00:05 |
| Subject Sample # 2 | 0.257 | 00:06 |
| Breath Volume      | 1.859 | 00:06 |
| Air Blank          | 0.000 | 00:07 |
| Control Test       | 0.080 | 00:07 |
| Air Blank          | 0.000 | 00:08 |
| Diagnostics Check  | OK    | 00:08 |

---

Data Download Date/Time: 06/17/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AYALA,KASSANDRA,  
BREATH TEST OPERATOR AGENCY: Fellsmere PD  
ARRESTING OFFICER: ROLLANDO,THOMAS  
ARRESTING OFFICER AGENCY: Fellsmere PD  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: OVALLE

DATE/TIME OF TEST: 06/12/2022 03:37:27

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:30 |                    |
| Air Blank         | 0.000 | 03:30 |                    |
| Control Test      | 0.080 | 03:31 |                    |
| Air Blank         | 0.000 | 03:31 |                    |
| Subject Sample    | # 1   | 03:34 | No Sample Provided |
| Breath Volume     | 0.000 | 03:34 |                    |
| Air Blank         | 0.000 | 03:35 |                    |
| Air Blank         |       | 03:37 | RFI Detect         |
| Air Blank         |       | 03:37 | RFI Detect         |
| Air Blank         |       | 03:37 | RFI Detect         |

---

Data Download Date/Time: 06/17/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AYALA,KASSANDRA,  
BREATH TEST OPERATOR AGENCY: Fellsmere PD  
ARRESTING OFFICER: ROLLANDO,THOMAS  
ARRESTING OFFICER AGENCY: Fellsmere PD  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: OVALLE

DATE/TIME OF TEST: 06/12/2022 03:48:13

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:45 |                      |
| Air Blank         | 0.000 | 03:45 |                      |
| Control Test      | 0.081 | 03:46 |                      |
| Air Blank         | 0.000 | 03:46 |                      |
| Subject Sample    | # 1   | 03:46 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:46 |                      |
| Air Blank         | 0.000 | 03:47 |                      |
| Control Test      | 0.080 | 03:47 |                      |
| Air Blank         | 0.000 | 03:48 |                      |
| Diagnostics Check | OK    | 03:48 |                      |

---

Data Download Date/Time: 06/17/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,.  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: HATTON,DEPUTY  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: LEBLANC

DATE/TIME OF TEST: 06/12/2022 22:12:50

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:10 |                      |
| Air Blank         | 0.000 | 22:10 |                      |
| Control Test      | 0.081 | 22:10 |                      |
| Air Blank         | 0.000 | 22:11 |                      |
| Subject Sample    | # 1   | 22:11 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:11 |                      |
| Air Blank         | 0.000 | 22:11 |                      |
| Control Test      | 0.080 | 22:12 |                      |
| Air Blank         | 0.000 | 22:12 |                      |
| Diagnostics Check | OK    | 22:12 |                      |

---

Data Download Date/Time: 06/17/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS,C,A  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: FINNEGAN,Z  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: CARABAJAL

DATE/TIME OF TEST: 06/13/2022 03:41:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:36 |                      |
| Air Blank         | 0.000 | 03:37 |                      |
| Control Test      | 0.081 | 03:37 |                      |
| Air Blank         | 0.000 | 03:38 |                      |
| Subject Sample    | # 1   | 03:40 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:40 |                      |
| Air Blank         | 0.000 | 03:41 |                      |
| Control Test      | 0.080 | 03:41 |                      |
| Air Blank         | 0.000 | 03:41 |                      |
| Diagnostics Check | OK    | 03:41 |                      |

---

Data Download Date/Time: 06/17/2022 16:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRANDY,KEVIN,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: MAGUIRE,M  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 10:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: VILLALPANDO

DATE/TIME OF TEST: 06/13/2022 11:27:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:21 |
| Air Blank          | 0.000 | 11:22 |
| Control Test       | 0.081 | 11:22 |
| Air Blank          | 0.000 | 11:22 |
| Subject Sample # 1 | 0.000 | 11:23 |
| Breath Volume      | 1.226 | 11:23 |
| Air Blank          | 0.000 | 11:23 |
| Air Blank          | 0.000 | 11:25 |
| Subject Sample # 2 | 0.000 | 11:26 |
| Breath Volume      | 1.390 | 11:26 |
| Air Blank          | 0.000 | 11:26 |
| Control Test       | 0.081 | 11:27 |
| Air Blank          | 0.000 | 11:27 |
| Diagnostics Check  | OK    | 11:27 |

---

Data Download Date/Time: 06/17/2022 16:38



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AYALA,KASSANDRA,  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: AYALA,KASSANDRA  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/15/2022 20:55:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:49 |
| Air Blank          | 0.000 | 20:49 |
| Control Test       | 0.081 | 20:49 |
| Air Blank          | 0.000 | 20:50 |
| Subject Sample # 1 | 0.009 | 20:50 |
| Breath Volume      | 2.332 | 20:50 |
| Air Blank          | 0.000 | 20:51 |
| Air Blank          | 0.000 | 20:53 |
| Subject Sample # 2 | 0.011 | 20:53 |
| Breath Volume      | 2.945 | 20:53 |
| Air Blank          | 0.000 | 20:54 |
| Control Test       | 0.081 | 20:54 |
| Air Blank          | 0.000 | 20:55 |
| Diagnostics Check  | OK    | 20:55 |

---

Data Download Date/Time: 06/17/2022 16:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDWARDS,SAMANTHA,K  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: CHOUINARD,JEFFREY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 05/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 07:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:59

SUBJECT'S LAST NAME: COLLADO

DATE/TIME OF TEST: 06/16/2022 09:33:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:27 |
| Air Blank          | 0.000 | 09:28 |
| Control Test       | 0.081 | 09:28 |
| Air Blank          | 0.000 | 09:29 |
| Subject Sample # 1 | 0.000 | 09:29 |
| Breath Volume      | 5.101 | 09:29 |
| Air Blank          | 0.000 | 09:30 |
| Air Blank          | 0.000 | 09:32 |
| Subject Sample # 2 | 0.000 | 09:32 |
| Breath Volume      | 4.082 | 09:32 |
| Air Blank          | 0.000 | 09:33 |
| Control Test       | 0.079 | 09:33 |
| Air Blank          | 0.000 | 09:33 |
| Diagnostics Check  | OK    | 09:33 |

---

Data Download Date/Time: 06/17/2022 16:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHILE,C,J  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: FRAKER,E  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HESSLER

DATE/TIME OF TEST: 06/17/2022 23:13:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:06 |
| Air Blank          | 0.000 | 23:07 |
| Control Test       | 0.081 | 23:07 |
| Air Blank          | 0.000 | 23:07 |
| Subject Sample # 1 | 0.142 | 23:08 |
| Breath Volume      | 1.199 | 23:08 |
| Air Blank          | 0.000 | 23:09 |
| Air Blank          | 0.000 | 23:10 |
| Subject Sample # 2 | 0.137 | 23:11 |
| Breath Volume      | 1.792 | 23:11 |
| Air Blank          | 0.000 | 23:12 |
| Control Test       | 0.081 | 23:12 |
| Air Blank          | 0.000 | 23:13 |
| Diagnostics Check  | OK    | 23:13 |

---

Data Download Date/Time: 07/20/2022 07:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,.  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: GRUBER,DEPUTY  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 17:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: PIPICELLI

DATE/TIME OF TEST: 06/20/2022 18:34:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:29 |
| Air Blank          | 0.000 | 18:29 |
| Control Test       | 0.081 | 18:29 |
| Air Blank          | 0.000 | 18:30 |
| Subject Sample # 1 | 0.000 | 18:30 |
| Breath Volume      | 2.578 | 18:30 |
| Air Blank          | 0.000 | 18:31 |
| Air Blank          | 0.000 | 18:33 |
| Subject Sample # 2 | 0.000 | 18:33 |
| Breath Volume      | 2.367 | 18:33 |
| Air Blank          | 0.000 | 18:34 |
| Control Test       | 0.081 | 18:34 |
| Air Blank          | 0.000 | 18:34 |
| Diagnostics Check  | OK    | 18:34 |

---

Data Download Date/Time: 07/20/2022 07:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,CODY,L  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: BROWN,CODY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 08:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:18

SUBJECT'S LAST NAME: COUNTS

DATE/TIME OF TEST: 06/21/2022 08:55:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:48 |
| Air Blank          | 0.000 | 08:49 |
| Control Test       | 0.081 | 08:49 |
| Air Blank          | 0.000 | 08:50 |
| Subject Sample # 1 | 0.000 | 08:51 |
| Breath Volume      | 1.789 | 08:51 |
| Air Blank          | 0.000 | 08:51 |
| Air Blank          | 0.000 | 08:53 |
| Subject Sample # 2 | 0.000 | 08:53 |
| Breath Volume      | 2.082 | 08:53 |
| Air Blank          | 0.000 | 08:54 |
| Control Test       | 0.079 | 08:54 |
| Air Blank          | 0.000 | 08:55 |
| Diagnostics Check  | OK    | 08:55 |

---

Data Download Date/Time: 07/20/2022 07:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEARSE,A,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: EVANS,H  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: UNDERHILL

DATE/TIME OF TEST: 06/21/2022 21:24:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:18 |
| Air Blank          | 0.000 | 21:18 |
| Control Test       | 0.081 | 21:18 |
| Air Blank          | 0.000 | 21:19 |
| Subject Sample # 1 | 0.000 | 21:20 |
| Breath Volume      | 3.617 | 21:20 |
| Air Blank          | 0.000 | 21:20 |
| Air Blank          | 0.000 | 21:22 |
| Subject Sample # 2 | 0.000 | 21:23 |
| Breath Volume      | 4.140 | 21:23 |
| Air Blank          | 0.000 | 21:23 |
| Control Test       | 0.080 | 21:23 |
| Air Blank          | 0.000 | 21:24 |
| Diagnostics Check  | OK    | 21:24 |

---

Data Download Date/Time: 07/20/2022 07:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOLF,JOSEPH,M  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: SAUCIER,M  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 06/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 06/22/2022 21:45:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:38 |
| Air Blank          | 0.000 | 21:38 |
| Control Test       | 0.080 | 21:39 |
| Air Blank          | 0.000 | 21:39 |
| Subject Sample # 1 | 0.284 | 21:40 |
| Breath Volume      | 1.832 | 21:40 |
| Air Blank          | 0.000 | 21:41 |
| Air Blank          | 0.000 | 21:43 |
| Subject Sample # 2 | 0.298 | 21:43 |
| Breath Volume      | 3.144 | 21:43 |
| Air Blank          | 0.000 | 21:44 |
| Control Test       | 0.079 | 21:44 |
| Air Blank          | 0.000 | 21:45 |
| Diagnostics Check  | OK    | 21:45 |

---

Data Download Date/Time: 07/20/2022 07:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: MCKENZIE, B  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 06/24/2022 03:02:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:56 |
| Air Blank          | 0.000 | 02:56 |
| Control Test       | 0.080 | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Subject Sample # 1 | 0.216 | 02:57 |
| Breath Volume      | 1.828 | 02:57 |
| Air Blank          | 0.000 | 02:58 |
| Air Blank          | 0.000 | 03:00 |
| Subject Sample # 2 | 0.212 | 03:00 |
| Breath Volume      | 1.273 | 03:00 |
| Air Blank          | 0.000 | 03:01 |
| Control Test       | 0.080 | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Diagnostics Check  | OK    | 03:02 |

---

Data Download Date/Time: 07/20/2022 07:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHILE,C,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: EVANS,M  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/29/2022 04:28:33

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 04:16 |                    |
| Air Blank         | 0.000 | 04:17 |                    |
| Control Test      | 0.081 | 04:17 |                    |
| Air Blank         | 0.000 | 04:17 |                    |
| Subject Sample    | # 1   | 04:20 | No Sample Provided |
| Breath Volume     | 0.000 | 04:20 |                    |
| Air Blank         | 0.000 | 04:21 |                    |
| Air Blank         | 0.000 | 04:23 |                    |
| Subject Sample    | # 2   | 0.187 | No .020 agreement  |
| Breath Volume     | 1.406 | 04:23 |                    |
| Air Blank         | 0.000 | 04:24 |                    |
| Air Blank         | 0.000 | 04:26 |                    |
| Subject Sample    | # 3   | 0.218 | No .020 agreement  |
| Breath Volume     | 1.710 | 04:26 |                    |
| Air Blank         | 0.000 | 04:27 |                    |
| Control Test      | 0.080 | 04:27 |                    |
| Air Blank         | 0.000 | 04:28 |                    |
| Diagnostics Check | OK    | 04:28 |                    |

Data Download Date/Time: 07/20/2022 07:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: EVANS, H  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/29/2022 05:24:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:14 |
| Air Blank          | 0.000 | 05:14 |
| Control Test       | 0.080 | 05:15 |
| Air Blank          | 0.000 | 05:15 |
| Subject Sample # 1 | 0.200 | 05:16 |
| Breath Volume      | 2.390 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Air Blank          | 0.000 | 05:18 |
| Subject Sample # 2 | 0.161 | 05:19 |
| Breath Volume      | 1.753 | 05:19 |
| Air Blank          | 0.000 | 05:20 |
| Air Blank          | 0.000 | 05:22 |
| Subject Sample # 3 | 0.161 | 05:22 |
| Breath Volume      | 1.828 | 05:22 |
| Air Blank          | 0.000 | 05:23 |
| Control Test       | 0.080 | 05:23 |
| Air Blank          | 0.000 | 05:23 |
| Diagnostics Check  | OK    | 05:24 |

---

Data Download Date/Time: 07/20/2022 07:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS,C,A  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: GREELEY,C  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 06/30/2022 00:31:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:24 |
| Air Blank          | 0.000 | 00:25 |
| Control Test       | 0.081 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Subject Sample # 1 | 0.149 | 00:26 |
| Breath Volume      | 3.078 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Air Blank          | 0.000 | 00:29 |
| Subject Sample # 2 | 0.145 | 00:29 |
| Breath Volume      | 3.406 | 00:29 |
| Air Blank          | 0.000 | 00:30 |
| Control Test       | 0.080 | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Diagnostics Check  | OK    | 00:31 |

---

Data Download Date/Time: 07/20/2022 07:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS,C,A  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: NUNLEY,B  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 06/30/2022 01:19:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:13 |
| Air Blank          | 0.000 | 01:13 |
| Control Test       | 0.080 | 01:14 |
| Air Blank          | 0.000 | 01:14 |
| Subject Sample # 1 | 0.133 | 01:15 |
| Breath Volume      | 3.625 | 01:15 |
| Air Blank          | 0.000 | 01:16 |
| Air Blank          | 0.000 | 01:17 |
| Subject Sample # 2 | 0.135 | 01:18 |
| Breath Volume      | 3.695 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Control Test       | 0.080 | 01:19 |
| Air Blank          | 0.000 | 01:19 |
| Diagnostics Check  | OK    | 01:19 |

---

Data Download Date/Time: 07/20/2022 07:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS,C,A  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: RISPOLI,M  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 06/30/2022 20:20:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:14 |
| Air Blank          | 0.000 | 20:14 |
| Control Test       | 0.081 | 20:14 |
| Air Blank          | 0.000 | 20:15 |
| Subject Sample # 1 | 0.000 | 20:15 |
| Breath Volume      | 3.175 | 20:15 |
| Air Blank          | 0.000 | 20:16 |
| Air Blank          | 0.000 | 20:18 |
| Subject Sample # 2 | 0.000 | 20:19 |
| Breath Volume      | 2.574 | 20:19 |
| Air Blank          | 0.000 | 20:19 |
| Control Test       | 0.081 | 20:20 |
| Air Blank          | 0.000 | 20:20 |
| Diagnostics Check  | OK    | 20:20 |

---

Data Download Date/Time: 07/20/2022 07:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEREZ,DAVID,A

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: ROUBICEK,JORDAN'

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211143

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/02/2022 00:46:30

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 00:40 |            |
| Air Blank         | 0.000     | 00:40 |            |
| Control Test      | 0.081     | 00:41 |            |
| Air Blank         | 0.000     | 00:41 |            |
| Subject Sample    | # 1 0.119 | 00:41 |            |
| Breath Volume     | 1.656     | 00:41 |            |
| Air Blank         | 0.000     | 00:42 |            |
| Air Blank         | 0.000     | 00:44 |            |
| Subject Sample    | # 2 0.120 | 00:44 |            |
| Breath Volume     | 1.292     | 00:44 |            |
| Air Blank         | 0.000     | 00:45 |            |
| Control Test      | 0.080     | 00:45 |            |
| Air Blank         |           | 00:46 | RFI Detect |
| Air Blank         | 0.000     | 00:46 |            |

---

Data Download Date/Time: 07/06/2022 15:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ, DAVID, A  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: ROUBICEK, JORDAN  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211143  
DRY GAS EXPIRY DATE: 07/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/02/2022 00:54:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Control Test       | 0.081 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Subject Sample # 1 | 0.121 | 00:50 |
| Breath Volume      | 1.390 | 00:50 |
| Air Blank          | 0.000 | 00:51 |
| Air Blank          | 0.000 | 00:52 |
| Subject Sample # 2 | 0.122 | 00:53 |
| Breath Volume      | 1.800 | 00:53 |
| Air Blank          | 0.000 | 00:53 |
| Control Test       | 0.080 | 00:54 |
| Air Blank          | 0.000 | 00:54 |
| Diagnostics Check  | OK    | 00:54 |

---

Data Download Date/Time: 07/06/2022 15:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PERRY, WAYNE, G

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: KRONDAHL, CARL

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211143

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:03

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: KUNJAN

DATE/TIME OF TEST: 06/03/2022 23:12:10

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 23:00 |                    |
| Air Blank         | 0.000 | 23:01 |                    |
| Control Test      | 0.081 | 23:01 |                    |
| Air Blank         | 0.000 | 23:01 |                    |
| Subject Sample    | # 1   | 23:05 | No Sample Provided |
| Breath Volume     | 0.000 | 23:05 |                    |
| Air Blank         | 0.000 | 23:05 |                    |
| Air Blank         | 0.000 | 23:07 |                    |
| Subject Sample    | # 2   | 23:10 | No Sample Provided |
| Breath Volume     | 0.000 | 23:10 |                    |
| Air Blank         | 0.000 | 23:11 |                    |
| Control Test      | 0.081 | 23:11 |                    |
| Air Blank         | 0.000 | 23:12 |                    |
| Diagnostics Check | OK    | 23:12 |                    |

---

Data Download Date/Time: 07/06/2022 15:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PERRY, WAYNE, G  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: KRONDAHL, CARL  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211143  
DRY GAS EXPIRY DATE: 07/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: KUNJAN

DATE/TIME OF TEST: 06/03/2022 23:23:42

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 23:15 |                      |
| Air Blank         | 0.000     | 23:15 |                      |
| Control Test      | 0.080     | 23:16 |                      |
| Air Blank         | 0.000     | 23:16 |                      |
| Subject Sample    | # 1 0.169 | 23:19 |                      |
| Breath Volume     | 1.222     | 23:19 |                      |
| Air Blank         | 0.000     | 23:20 |                      |
| Air Blank         | 0.000     | 23:22 |                      |
| Subject Sample    | # 2       | 23:22 | Subject Test Refused |
| Breath Volume     | 0.000     | 23:22 |                      |
| Air Blank         | 0.000     | 23:22 |                      |
| Control Test      | 0.079     | 23:23 |                      |
| Air Blank         | 0.000     | 23:23 |                      |
| Diagnostics Check | OK        | 23:23 |                      |

---

Data Download Date/Time: 07/06/2022 15:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PERRY, WAYNE, G

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: PERRY, WAYNE

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211143

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 06/06/2022 02:03:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:00 |                      |
| Air Blank         | 0.000 | 02:01 |                      |
| Control Test      | 0.082 | 02:01 |                      |
| Air Blank         | 0.000 | 02:01 |                      |
| Subject Sample    | # 1   | 02:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:02 |                      |
| Air Blank         | 0.000 | 02:02 |                      |
| Control Test      | 0.081 | 02:02 |                      |
| Air Blank         | 0.000 | 02:03 |                      |
| Diagnostics Check | OK    | 02:03 |                      |

---

Data Download Date/Time: 07/06/2022 15:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAPP,ERIC,G

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: RAPP,ERIC

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211143

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:05

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HARBAUGH

DATE/TIME OF TEST: 06/11/2022 23:53:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:46 |
| Air Blank          | 0.000 | 23:47 |
| Control Test       | 0.081 | 23:47 |
| Air Blank          | 0.000 | 23:47 |
| Subject Sample # 1 | 0.000 | 23:48 |
| Breath Volume      | 2.941 | 23:48 |
| Air Blank          | 0.000 | 23:49 |
| Air Blank          | 0.000 | 23:51 |
| Subject Sample # 2 | 0.000 | 23:51 |
| Breath Volume      | 3.054 | 23:51 |
| Air Blank          | 0.000 | 23:52 |
| Control Test       | 0.080 | 23:52 |
| Air Blank          | 0.000 | 23:52 |
| Diagnostics Check  | OK    | 23:53 |

---

Data Download Date/Time: 07/06/2022 15:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KRONDAHL,CARL,J  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: KRONDAHL,CARL  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13921080A2  
DRY GAS EXPIRY DATE: 07/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: PINKHAM

DATE/TIME OF TEST: 06/20/2022 00:02:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:54 |                      |
| Air Blank         | 0.000 | 23:54 |                      |
| Control Test      | 0.081 | 23:54 |                      |
| Air Blank         | 0.000 | 23:55 |                      |
| Subject Sample    | # 2   | 23:58 | No Sample Provided   |
| Breath Volume     | 0.000 | 23:58 |                      |
| Air Blank         | 0.000 | 23:59 |                      |
| Air Blank         | 0.000 | 00:01 |                      |
| Subject Sample    | # 1   | 00:01 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:01 |                      |
| Air Blank         | 0.000 | 00:01 |                      |
| Control Test      | 0.081 | 00:02 |                      |
| Air Blank         | 0.000 | 00:02 |                      |
| Diagnostics Check | OK    | 00:02 |                      |

---

Data Download Date/Time: 07/06/2022 15:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CRONK,A,

BREATH TEST OPERATOR AGENCY: Lady Lake PD

ARRESTING OFFICER: PETTIS,J

ARRESTING OFFICER AGENCY: Lady Lake PD

LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211143

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 04:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ZINIEWICZ

DATE/TIME OF TEST: 06/21/2022 05:43:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:36 |
| Air Blank          | 0.000 | 05:37 |
| Control Test       | 0.080 | 05:37 |
| Air Blank          | 0.000 | 05:38 |
| Subject Sample # 1 | 0.367 | 05:38 |
| Breath Volume      | 1.847 | 05:38 |
| Air Blank          | 0.000 | 05:39 |
| Air Blank          | 0.000 | 05:41 |
| Subject Sample # 2 | 0.368 | 05:41 |
| Breath Volume      | 3.636 | 05:41 |
| Air Blank          | 0.000 | 05:42 |
| Control Test       | 0.079 | 05:42 |
| Air Blank          | 0.000 | 05:43 |
| Diagnostics Check  | OK    | 05:43 |

---

Data Download Date/Time: 07/06/2022 15:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEREZ,DAVID,A

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: RAPP,ERIC

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211143

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: TABOR

DATE/TIME OF TEST: 06/22/2022 01:23:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:20 |                      |
| Air Blank         | 0.000 | 01:21 |                      |
| Control Test      | 0.081 | 01:21 |                      |
| Air Blank         | 0.000 | 01:21 |                      |
| Subject Sample    | # 1   | 01:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:22 |                      |
| Air Blank         | 0.000 | 01:22 |                      |
| Control Test      | 0.082 | 01:22 |                      |
| Air Blank         | 0.000 | 01:23 |                      |
| Diagnostics Check | OK    | 01:23 |                      |

---

Data Download Date/Time: 07/06/2022 15:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAPP,ERIC,G

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: RAPP,ERIC

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211143

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: KLING

DATE/TIME OF TEST: 06/24/2022 22:21:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:16 |                      |
| Air Blank         | 0.000 | 22:17 |                      |
| Control Test      | 0.081 | 22:17 |                      |
| Air Blank         | 0.000 | 22:17 |                      |
| Subject Sample    | # 1   | 22:19 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:19 |                      |
| Air Blank         | 0.000 | 22:20 |                      |
| Control Test      | 0.081 | 22:20 |                      |
| Air Blank         | 0.000 | 22:21 |                      |
| Diagnostics Check | OK    | 22:21 |                      |

---

Data Download Date/Time: 07/06/2022 15:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAPP,ERIC,G  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: COE,BARRY  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 05/31/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 211143  
DRY GAS EXPIRY DATE: 07/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: LANGSTON

DATE/TIME OF TEST: 06/25/2022 22:35:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:28 |
| Air Blank          | 0.000 | 22:28 |
| Control Test       | 0.080 | 22:28 |
| Air Blank          | 0.000 | 22:29 |
| Subject Sample # 1 | 0.225 | 22:30 |
| Breath Volume      | 2.781 | 22:30 |
| Air Blank          | 0.000 | 22:31 |
| Air Blank          | 0.000 | 22:33 |
| Subject Sample # 2 | 0.235 | 22:33 |
| Breath Volume      | 2.015 | 22:33 |
| Air Blank          | 0.000 | 22:34 |
| Control Test       | 0.080 | 22:34 |
| Air Blank          | 0.000 | 22:35 |
| Diagnostics Check  | OK    | 22:35 |

---

Data Download Date/Time: 07/06/2022 15:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIMITROV, OFFICER,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ALFALLA, OFFICER  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CIECHANOWSKI

DATE/TIME OF TEST: 06/01/2022 03:53:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Control Test       | 0.081 | 03:48 |
| Air Blank          | 0.000 | 03:48 |
| Subject Sample # 1 | 0.000 | 03:49 |
| Breath Volume      | 3.156 | 03:49 |
| Air Blank          | 0.000 | 03:49 |
| Air Blank          | 0.000 | 03:51 |
| Subject Sample # 2 | 0.000 | 03:52 |
| Breath Volume      | 3.664 | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Control Test       | 0.080 | 03:52 |
| Air Blank          | 0.000 | 03:53 |
| Diagnostics Check  | OK    | 03:53 |

---

Data Download Date/Time: 06/04/2022 22:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: DAVENPORT,DANIEL  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: XILOJ

DATE/TIME OF TEST: 06/01/2022 21:04:30

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 20:52 |                |
| Air Blank         | 0.000     | 20:52 |                |
| Control Test      | 0.080     | 20:52 |                |
| Air Blank         | 0.000     | 20:53 |                |
| Subject Sample    | # 1 0.181 | 20:56 | Volume Not Met |
| Breath Volume     | 0.183     | 20:56 |                |
| Air Blank         | 0.000     | 20:57 |                |
| Air Blank         | 0.000     | 20:59 |                |
| Subject Sample    | # 2 0.205 | 20:59 |                |
| Breath Volume     | 1.906     | 20:59 |                |
| Air Blank         | 0.000     | 21:00 |                |
| Air Blank         | 0.000     | 21:02 |                |
| Subject Sample    | # 3 0.205 | 21:03 |                |
| Breath Volume     | 1.531     | 21:03 |                |
| Air Blank         | 0.000     | 21:03 |                |
| Control Test      | 0.079     | 21:03 |                |
| Air Blank         | 0.000     | 21:04 |                |
| Diagnostics Check | OK        | 21:04 |                |

---

Data Download Date/Time: 06/04/2022 22:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURLEY,SHAUN,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PACHECO,OFFICER  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: FISCHER

DATE/TIME OF TEST: 06/02/2022 03:10:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:04 |
| Air Blank          | 0.000 | 03:04 |
| Control Test       | 0.080 | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Subject Sample # 1 | 0.206 | 03:05 |
| Breath Volume      | 1.468 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Air Blank          | 0.000 | 03:08 |
| Subject Sample # 2 | 0.193 | 03:08 |
| Breath Volume      | 1.367 | 03:08 |
| Air Blank          | 0.000 | 03:09 |
| Control Test       | 0.080 | 03:09 |
| Air Blank          | 0.000 | 03:10 |
| Diagnostics Check  | OK    | 03:10 |

---

Data Download Date/Time: 06/04/2022 22:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: NORTON,GREGORY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: STOOTS

DATE/TIME OF TEST: 06/02/2022 04:01:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:55 |
| Air Blank          | 0.000 | 03:55 |
| Control Test       | 0.080 | 03:55 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 1 | 0.123 | 03:56 |
| Breath Volume      | 2.046 | 03:56 |
| Air Blank          | 0.000 | 03:57 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 2 | 0.124 | 03:59 |
| Breath Volume      | 2.679 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Control Test       | 0.080 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Diagnostics Check  | OK    | 04:01 |

---

Data Download Date/Time: 06/04/2022 22:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDRADE,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: GONZALEZ,OFFICER  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 14:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: LEXELL

DATE/TIME OF TEST: 06/02/2022 17:08:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:02 |
| Air Blank          | 0.000 | 17:02 |
| Control Test       | 0.080 | 17:03 |
| Air Blank          | 0.000 | 17:03 |
| Subject Sample # 1 | 0.000 | 17:04 |
| Breath Volume      | 1.339 | 17:04 |
| Air Blank          | 0.000 | 17:04 |
| Air Blank          | 0.000 | 17:06 |
| Subject Sample # 2 | 0.000 | 17:07 |
| Breath Volume      | 2.332 | 17:07 |
| Air Blank          | 0.000 | 17:07 |
| Control Test       | 0.080 | 17:07 |
| Air Blank          | 0.000 | 17:08 |
| Diagnostics Check  | OK    | 17:08 |

---

Data Download Date/Time: 06/04/2022 22:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: DAVENPORT,DANIEL  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HIPOLITO

DATE/TIME OF TEST: 06/03/2022 03:22:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Control Test       | 0.080 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Subject Sample # 1 | 0.187 | 03:18 |
| Breath Volume      | 2.265 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 2 | 0.183 | 03:21 |
| Breath Volume      | 2.519 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Control Test       | 0.080 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Diagnostics Check  | OK    | 03:22 |

---

Data Download Date/Time: 06/04/2022 22:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/05/2022 01:19:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:11 |
| Air Blank          | 0.000 | 01:12 |
| Control Test       | 0.080 | 01:12 |
| Air Blank          | 0.000 | 01:13 |
| Subject Sample # 1 | 0.143 | 01:14 |
| Breath Volume      | 1.218 | 01:14 |
| Air Blank          | 0.000 | 01:15 |
| Air Blank          | 0.000 | 01:17 |
| Subject Sample # 2 | 0.145 | 01:18 |
| Breath Volume      | 1.875 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Control Test       | 0.081 | 01:19 |
| Air Blank          | 0.000 | 01:19 |
| Diagnostics Check  | OK    | 01:19 |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DEP,R  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MILLER,DEP R  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: GUTIERREZ PEREZ

DATE/TIME OF TEST: 06/05/2022 02:49:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:42 |
| Air Blank          | 0.000 | 02:43 |
| Control Test       | 0.081 | 02:43 |
| Air Blank          | 0.000 | 02:44 |
| Subject Sample # 1 | 0.031 | 02:45 |
| Breath Volume      | 2.968 | 02:45 |
| Air Blank          | 0.000 | 02:45 |
| Air Blank          | 0.000 | 02:47 |
| Subject Sample # 2 | 0.027 | 02:48 |
| Breath Volume      | 2.867 | 02:48 |
| Air Blank          | 0.000 | 02:48 |
| Control Test       | 0.079 | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Diagnostics Check  | OK    | 02:49 |

---

Data Download Date/Time: 07/01/2022 18:45



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACKLEY,DEP,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: GREGORIO,K  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ATHANAZE

DATE/TIME OF TEST: 06/05/2022 03:33:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.081 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Subject Sample # 1 | 0.136 | 03:28 |
| Breath Volume      | 3.699 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Air Blank          | 0.000 | 03:31 |
| Subject Sample # 2 | 0.138 | 03:32 |
| Breath Volume      | 4.253 | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.080 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Diagnostics Check  | OK    | 03:33 |

---

Data Download Date/Time: 07/01/2022 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,N,N  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: NELSON,A  
ARRESTING OFFICER AGENCY: Venice PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: JOPPIE

DATE/TIME OF TEST: 06/05/2022 23:10:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:04 |
| Air Blank          | 0.000 | 23:04 |
| Control Test       | 0.080 | 23:05 |
| Air Blank          | 0.000 | 23:05 |
| Subject Sample # 1 | 0.000 | 23:06 |
| Breath Volume      | 2.304 | 23:06 |
| Air Blank          | 0.000 | 23:06 |
| Air Blank          | 0.000 | 23:08 |
| Subject Sample # 2 | 0.000 | 23:08 |
| Breath Volume      | 2.343 | 23:08 |
| Air Blank          | 0.000 | 23:09 |
| Control Test       | 0.080 | 23:09 |
| Air Blank          | 0.000 | 23:10 |
| Diagnostics Check  | OK    | 23:10 |

---

Data Download Date/Time: 07/01/2022 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: KIRKENDALL

DATE/TIME OF TEST: 06/06/2022 01:01:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:55 |
| Air Blank          | 0.000 | 00:55 |
| Control Test       | 0.080 | 00:56 |
| Air Blank          | 0.000 | 00:56 |
| Subject Sample # 1 | 0.208 | 00:57 |
| Breath Volume      | 1.832 | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Air Blank          | 0.000 | 00:59 |
| Subject Sample # 2 | 0.207 | 01:00 |
| Breath Volume      | 2.167 | 01:00 |
| Air Blank          | 0.000 | 01:00 |
| Control Test       | 0.080 | 01:01 |
| Air Blank          | 0.000 | 01:01 |
| Diagnostics Check  | OK    | 01:01 |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DEP,R  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MILLER,DEP R  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 06/06/2022 02:32:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:26 |
| Air Blank          | 0.000 | 02:27 |
| Control Test       | 0.081 | 02:27 |
| Air Blank          | 0.000 | 02:27 |
| Subject Sample # 1 | 0.144 | 02:28 |
| Breath Volume      | 2.648 | 02:28 |
| Air Blank          | 0.000 | 02:28 |
| Air Blank          | 0.000 | 02:30 |
| Subject Sample # 2 | 0.135 | 02:31 |
| Breath Volume      | 2.765 | 02:31 |
| Air Blank          | 0.000 | 02:31 |
| Control Test       | 0.079 | 02:32 |
| Air Blank          | 0.000 | 02:32 |
| Diagnostics Check  | OK    | 02:32 |

---

Data Download Date/Time: 07/01/2022 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SURRAN,DEPUTY,C  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SHAW,ROBERT  
ARRESTING OFFICER AGENCY: Sarasota-Manatee Air  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 06/06/2022 08:29:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:23 |
| Air Blank          | 0.000 | 08:23 |
| Control Test       | 0.080 | 08:23 |
| Air Blank          | 0.000 | 08:24 |
| Subject Sample # 1 | 0.162 | 08:25 |
| Breath Volume      | 1.656 | 08:25 |
| Air Blank          | 0.000 | 08:25 |
| Air Blank          | 0.000 | 08:27 |
| Subject Sample # 2 | 0.151 | 08:28 |
| Breath Volume      | 1.281 | 08:28 |
| Air Blank          | 0.000 | 08:28 |
| Control Test       | 0.080 | 08:29 |
| Air Blank          | 0.000 | 08:29 |
| Diagnostics Check  | OK    | 08:29 |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERSON,CORY,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PIERSON,CORY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: TUERK

DATE/TIME OF TEST: 06/07/2022 00:05:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:56 |
| Air Blank          | 0.000 | 23:56 |
| Control Test       | 0.080 | 23:57 |
| Air Blank          | 0.000 | 23:57 |
| Subject Sample # 1 | 0.198 | 00:00 |
| Breath Volume      | 2.480 | 00:00 |
| Air Blank          | 0.000 | 00:00 |
| Air Blank          | 0.000 | 00:02 |
| Subject Sample # 2 | 0.193 | 00:03 |
| Breath Volume      | 1.652 | 00:03 |
| Air Blank          | 0.000 | 00:04 |
| Control Test       | 0.078 | 00:04 |
| Air Blank          | 0.000 | 00:05 |
| Diagnostics Check  | OK    | 00:05 |

---

Data Download Date/Time: 07/01/2022 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: CULBERT

DATE/TIME OF TEST: 06/07/2022 00:18:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:12 |
| Air Blank          | 0.000 | 00:13 |
| Control Test       | 0.079 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Subject Sample # 1 | 0.000 | 00:14 |
| Breath Volume      | 4.214 | 00:14 |
| Air Blank          | 0.000 | 00:15 |
| Air Blank          | 0.000 | 00:17 |
| Subject Sample # 2 | 0.007 | 00:17 |
| Breath Volume      | 4.445 | 00:17 |
| Air Blank          | 0.000 | 00:17 |
| Control Test       | 0.079 | 00:18 |
| Air Blank          | 0.000 | 00:18 |
| Diagnostics Check  | OK    | 00:18 |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: NORTON,DEPUTY G.  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/08/2022 00:19:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:13 |
| Air Blank          | 0.000 | 00:13 |
| Control Test       | 0.080 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Subject Sample # 1 | 0.000 | 00:15 |
| Breath Volume      | 1.382 | 00:15 |
| Air Blank          | 0.000 | 00:15 |
| Air Blank          | 0.000 | 00:17 |
| Subject Sample # 2 | 0.000 | 00:18 |
| Breath Volume      | 1.890 | 00:18 |
| Air Blank          | 0.000 | 00:18 |
| Control Test       | 0.079 | 00:18 |
| Air Blank          | 0.000 | 00:19 |
| Diagnostics Check  | OK    | 00:19 |

---

Data Download Date/Time: 07/01/2022 18:45



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERSON,CORY,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PIERSON,CORY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BATISTA PETROS

DATE/TIME OF TEST: 06/08/2022 02:34:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Control Test       | 0.079 | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Subject Sample # 1 | 0.132 | 02:30 |
| Breath Volume      | 2.277 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Air Blank          | 0.000 | 02:33 |
| Subject Sample # 2 | 0.131 | 02:33 |
| Breath Volume      | 2.328 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.080 | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Diagnostics Check  | OK    | 02:34 |

---

Data Download Date/Time: 07/01/2022 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: HANNAH

DATE/TIME OF TEST: 06/08/2022 03:54:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:51 |                      |
| Air Blank         | 0.000 | 03:51 |                      |
| Control Test      | 0.080 | 03:51 |                      |
| Air Blank         | 0.000 | 03:52 |                      |
| Subject Sample    | # 1   | 03:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:52 |                      |
| Air Blank         | 0.000 | 03:53 |                      |
| Control Test      | 0.080 | 03:53 |                      |
| Air Blank         | 0.000 | 03:54 |                      |
| Diagnostics Check | OK    | 03:54 |                      |

---

Data Download Date/Time: 07/01/2022 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOLADVANDI,OFC. M.,  
BREATH TEST OPERATOR AGENCY: North Port PD  
ARRESTING OFFICER: FOLADVANDI,OFC. M.  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/08/2022 22:56:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:49 |
| Air Blank          | 0.000 | 22:50 |
| Control Test       | 0.081 | 22:50 |
| Air Blank          | 0.000 | 22:50 |
| Subject Sample # 1 | 0.196 | 22:51 |
| Breath Volume      | 1.949 | 22:51 |
| Air Blank          | 0.000 | 22:52 |
| Air Blank          | 0.000 | 22:54 |
| Subject Sample # 2 | 0.189 | 22:54 |
| Breath Volume      | 1.574 | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Control Test       | 0.079 | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Diagnostics Check  | OK    | 22:56 |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DEP,R  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MILLER,DEP R  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SPANN

DATE/TIME OF TEST: 06/09/2022 02:57:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:52 |
| Air Blank          | 0.000 | 02:52 |
| Control Test       | 0.081 | 02:52 |
| Air Blank          | 0.000 | 02:53 |
| Subject Sample # 1 | 0.129 | 02:53 |
| Breath Volume      | 2.394 | 02:53 |
| Air Blank          | 0.000 | 02:54 |
| Air Blank          | 0.000 | 02:56 |
| Subject Sample # 2 | 0.127 | 02:56 |
| Breath Volume      | 2.269 | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Control Test       | 0.080 | 02:57 |
| Air Blank          | 0.000 | 02:57 |
| Diagnostics Check  | OK    | 02:57 |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DEP,R  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRUNO,OFC N  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: PASSALACQUA

DATE/TIME OF TEST: 06/09/2022 04:31:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Control Test       | 0.081 | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Subject Sample # 1 | 0.201 | 04:27 |
| Breath Volume      | 3.519 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Air Blank          | 0.000 | 04:29 |
| Subject Sample # 2 | 0.204 | 04:30 |
| Breath Volume      | 4.195 | 04:30 |
| Air Blank          | 0.000 | 04:30 |
| Control Test       | 0.079 | 04:30 |
| Air Blank          | 0.000 | 04:31 |
| Diagnostics Check  | OK    | 04:31 |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERSON,CORY,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PIERSON,CORY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: JUSTICE

DATE/TIME OF TEST: 06/10/2022 22:00:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:54 |
| Air Blank          | 0.000 | 21:54 |
| Control Test       | 0.081 | 21:54 |
| Air Blank          | 0.000 | 21:55 |
| Subject Sample # 1 | 0.190 | 21:56 |
| Breath Volume      | 2.832 | 21:56 |
| Air Blank          | 0.000 | 21:56 |
| Air Blank          | 0.000 | 21:58 |
| Subject Sample # 2 | 0.193 | 21:59 |
| Breath Volume      | 3.328 | 21:59 |
| Air Blank          | 0.000 | 21:59 |
| Control Test       | 0.080 | 22:00 |
| Air Blank          | 0.000 | 22:00 |
| Diagnostics Check  | OK    | 22:00 |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERSON,CORY,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: KNIGHT,DAVID  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: BROOM

DATE/TIME OF TEST: 06/10/2022 22:22:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:15 |
| Air Blank          | 0.000 | 22:15 |
| Control Test       | 0.081 | 22:15 |
| Air Blank          | 0.000 | 22:16 |
| Subject Sample # 1 | 0.239 | 22:17 |
| Breath Volume      | 2.308 | 22:17 |
| Air Blank          | 0.000 | 22:18 |
| Air Blank          | 0.000 | 22:20 |
| Subject Sample # 2 | 0.229 | 22:21 |
| Breath Volume      | 1.824 | 22:21 |
| Air Blank          | 0.000 | 22:22 |
| Control Test       | 0.079 | 22:22 |
| Air Blank          | 0.000 | 22:22 |
| Diagnostics Check  | OK    | 22:22 |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERSON,CORY,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PIETRON,MICHAEL  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: CHUPP

DATE/TIME OF TEST: 06/11/2022 00:48:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:41 |
| Air Blank          | 0.000 | 00:42 |
| Control Test       | 0.081 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 1 | 0.157 | 00:43 |
| Breath Volume      | 3.742 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Air Blank          | 0.000 | 00:46 |
| Subject Sample # 2 | 0.154 | 00:47 |
| Breath Volume      | 2.859 | 00:47 |
| Air Blank          | 0.000 | 00:47 |
| Control Test       | 0.081 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Diagnostics Check  | OK    | 00:48 |

---

Data Download Date/Time: 07/01/2022 18:45



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSS,MICHAEL,K  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SHAFFER,OFFICER  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 15:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: STERLING

DATE/TIME OF TEST: 06/11/2022 16:07:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:00 |
| Air Blank          | 0.000 | 16:01 |
| Control Test       | 0.081 | 16:01 |
| Air Blank          | 0.000 | 16:01 |
| Subject Sample # 1 | 0.216 | 16:02 |
| Breath Volume      | 1.789 | 16:02 |
| Air Blank          | 0.000 | 16:03 |
| Air Blank          | 0.000 | 16:05 |
| Subject Sample # 2 | 0.220 | 16:05 |
| Breath Volume      | 1.664 | 16:05 |
| Air Blank          | 0.000 | 16:06 |
| Control Test       | 0.080 | 16:06 |
| Air Blank          | 0.000 | 16:07 |
| Diagnostics Check  | OK    | 16:07 |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: DAVENPORT,DANIEL  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: TELLES

DATE/TIME OF TEST: 06/12/2022 02:28:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:22 |
| Air Blank          | 0.000 | 02:22 |
| Control Test       | 0.081 | 02:23 |
| Air Blank          | 0.000 | 02:23 |
| Subject Sample # 1 | 0.126 | 02:24 |
| Breath Volume      | 1.789 | 02:24 |
| Air Blank          | 0.000 | 02:25 |
| Air Blank          | 0.000 | 02:26 |
| Subject Sample # 2 | 0.128 | 02:27 |
| Breath Volume      | 1.566 | 02:27 |
| Air Blank          | 0.000 | 02:27 |
| Control Test       | 0.081 | 02:28 |
| Air Blank          | 0.000 | 02:28 |
| Diagnostics Check  | OK    | 02:28 |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: DAVENPORT,DANIEL  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/12/2022 21:57:20

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 21:45 |                |
| Air Blank         | 0.000     | 21:46 |                |
| Control Test      | 0.080     | 21:46 |                |
| Air Blank         | 0.000     | 21:47 |                |
| Subject Sample    | # 1 0.176 | 21:50 | Volume Not Met |
| Breath Volume     | 0.191     | 21:50 |                |
| Air Blank         | 0.000     | 21:50 |                |
| Air Blank         | 0.000     | 21:52 |                |
| Subject Sample    | # 2 0.173 | 21:55 | Volume Not Met |
| Breath Volume     | 0.199     | 21:55 |                |
| Air Blank         | 0.000     | 21:56 |                |
| Control Test      | 0.078     | 21:56 |                |
| Air Blank         | 0.000     | 21:57 |                |
| Diagnostics Check | OK        | 21:57 |                |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURLEY,SHAUN,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PACHECO,E  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: SHKAROVSKIY

DATE/TIME OF TEST: 06/13/2022 00:51:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:41 |
| Air Blank          | 0.000 | 00:41 |
| Control Test       | 0.080 | 00:41 |
| Air Blank          | 0.000 | 00:42 |
| Subject Sample # 1 | 0.290 | 00:43 |
| Breath Volume      | 3.390 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Air Blank          | 0.000 | 00:45 |
| Subject Sample # 2 | 0.260 | 00:46 |
| Breath Volume      | 2.304 | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Air Blank          | 0.000 | 00:48 |
| Subject Sample # 3 | 0.278 | 00:49 |
| Breath Volume      | 3.085 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Control Test       | 0.078 | 00:50 |
| Air Blank          | 0.000 | 00:51 |
| Diagnostics Check  | OK    | 00:51 |

---

Data Download Date/Time: 07/01/2022 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: LABOLLE

DATE/TIME OF TEST: 06/13/2022 01:16:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Control Test       | 0.079 | 01:10 |
| Air Blank          | 0.000 | 01:11 |
| Subject Sample # 1 | 0.170 | 01:12 |
| Breath Volume      | 3.312 | 01:12 |
| Air Blank          | 0.000 | 01:12 |
| Air Blank          | 0.000 | 01:14 |
| Subject Sample # 2 | 0.172 | 01:15 |
| Breath Volume      | 3.156 | 01:15 |
| Air Blank          | 0.000 | 01:15 |
| Control Test       | 0.079 | 01:16 |
| Air Blank          | 0.000 | 01:16 |
| Diagnostics Check  | OK    | 01:16 |

---

Data Download Date/Time: 07/01/2022 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SHAW,OFFICER  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:39

SUBJECT'S LAST NAME: ADREY

DATE/TIME OF TEST: 06/13/2022 07:26:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:20 |
| Air Blank          | 0.000 | 07:20 |
| Control Test       | 0.080 | 07:20 |
| Air Blank          | 0.000 | 07:21 |
| Subject Sample # 1 | 0.199 | 07:21 |
| Breath Volume      | 1.820 | 07:21 |
| Air Blank          | 0.000 | 07:22 |
| Air Blank          | 0.000 | 07:24 |
| Subject Sample # 2 | 0.200 | 07:24 |
| Breath Volume      | 1.773 | 07:24 |
| Air Blank          | 0.000 | 07:25 |
| Control Test       | 0.079 | 07:25 |
| Air Blank          | 0.000 | 07:25 |
| Diagnostics Check  | OK    | 07:26 |

---

Data Download Date/Time: 07/01/2022 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 06/15/2022 03:53:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Control Test       | 0.081 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Subject Sample # 1 | 0.147 | 03:48 |
| Breath Volume      | 2.949 | 03:48 |
| Air Blank          | 0.000 | 03:49 |
| Air Blank          | 0.000 | 03:51 |
| Subject Sample # 2 | 0.139 | 03:51 |
| Breath Volume      | 2.867 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Control Test       | 0.081 | 03:52 |
| Air Blank          | 0.000 | 03:53 |
| Diagnostics Check  | OK    | 03:53 |

---

Data Download Date/Time: 07/01/2022 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J.  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 06/17/2022 00:15:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:08 |
| Air Blank          | 0.000 | 00:09 |
| Control Test       | 0.081 | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Subject Sample # 1 | 0.022 | 00:10 |
| Breath Volume      | 2.218 | 00:10 |
| Air Blank          | 0.000 | 00:11 |
| Air Blank          | 0.000 | 00:13 |
| Subject Sample # 2 | 0.021 | 00:13 |
| Breath Volume      | 1.425 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Control Test       | 0.081 | 00:14 |
| Air Blank          | 0.000 | 00:15 |
| Diagnostics Check  | OK    | 00:15 |

---

Data Download Date/Time: 07/01/2022 18:46



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DEP,R  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: NELSON,OFC A  
ARRESTING OFFICER AGENCY: Venice PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HENDRICKS

DATE/TIME OF TEST: 06/19/2022 02:41:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Control Test       | 0.081 | 02:36 |
| Air Blank          | 0.000 | 02:36 |
| Subject Sample # 1 | 0.134 | 02:37 |
| Breath Volume      | 3.035 | 02:37 |
| Air Blank          | 0.000 | 02:37 |
| Air Blank          | 0.000 | 02:39 |
| Subject Sample # 2 | 0.135 | 02:40 |
| Breath Volume      | 3.160 | 02:40 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.080 | 02:41 |
| Air Blank          | 0.000 | 02:41 |
| Diagnostics Check  | OK    | 02:41 |

---

Data Download Date/Time: 07/01/2022 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 06/19/2022 02:56:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.080 | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Subject Sample # 1 | 0.146 | 02:50 |
| Breath Volume      | 1.328 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Air Blank          | 0.000 | 02:53 |
| Subject Sample # 2 | 0.149 | 02:54 |
| Breath Volume      | 1.140 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Control Test       | 0.080 | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Diagnostics Check  | OK    | 02:56 |

---

Data Download Date/Time: 07/01/2022 18:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PUFAHL,G,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PUFAHL,G  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 06:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:59

SUBJECT'S LAST NAME: LOPEZ CRUZ

DATE/TIME OF TEST: 06/20/2022 07:25:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:19 |
| Air Blank          | 0.000 | 07:19 |
| Control Test       | 0.080 | 07:20 |
| Air Blank          | 0.000 | 07:20 |
| Subject Sample # 1 | 0.146 | 07:21 |
| Breath Volume      | 1.421 | 07:21 |
| Air Blank          | 0.000 | 07:21 |
| Air Blank          | 0.000 | 07:23 |
| Subject Sample # 2 | 0.129 | 07:24 |
| Breath Volume      | 1.468 | 07:24 |
| Air Blank          | 0.000 | 07:24 |
| Control Test       | 0.080 | 07:25 |
| Air Blank          | 0.000 | 07:25 |
| Diagnostics Check  | OK    | 07:25 |

---

Data Download Date/Time: 07/01/2022 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SURRAN,DEPUTY,C  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: RODRIGUEZ,E  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 14:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 06/20/2022 15:30:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:23 |
| Air Blank          | 0.000 | 15:24 |
| Control Test       | 0.080 | 15:24 |
| Air Blank          | 0.000 | 15:24 |
| Subject Sample # 1 | 0.345 | 15:26 |
| Breath Volume      | 1.292 | 15:26 |
| Air Blank          | 0.000 | 15:26 |
| Air Blank          | 0.000 | 15:28 |
| Subject Sample # 2 | 0.358 | 15:29 |
| Breath Volume      | 1.496 | 15:29 |
| Air Blank          | 0.000 | 15:29 |
| Control Test       | 0.079 | 15:30 |
| Air Blank          | 0.000 | 15:30 |
| Diagnostics Check  | OK    | 15:30 |

---

Data Download Date/Time: 07/01/2022 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURLEY,SHAUN,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PIETRON,M  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: HOSKINS

DATE/TIME OF TEST: 06/21/2022 00:57:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:50 |
| Air Blank          | 0.000 | 00:51 |
| Control Test       | 0.080 | 00:51 |
| Air Blank          | 0.000 | 00:52 |
| Subject Sample # 1 | 0.209 | 00:52 |
| Breath Volume      | 1.343 | 00:52 |
| Air Blank          | 0.000 | 00:53 |
| Air Blank          | 0.000 | 00:55 |
| Subject Sample # 2 | 0.215 | 00:55 |
| Breath Volume      | 1.324 | 00:55 |
| Air Blank          | 0.000 | 00:56 |
| Control Test       | 0.079 | 00:56 |
| Air Blank          | 0.000 | 00:57 |
| Diagnostics Check  | OK    | 00:57 |

---

Data Download Date/Time: 07/01/2022 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURLEY,SHAUN,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PACHECO,E  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: TREMBLAY

DATE/TIME OF TEST: 06/21/2022 20:06:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:59 |
| Air Blank          | 0.000 | 20:00 |
| Control Test       | 0.080 | 20:00 |
| Air Blank          | 0.000 | 20:01 |
| Subject Sample # 1 | 0.146 | 20:01 |
| Breath Volume      | 2.617 | 20:01 |
| Air Blank          | 0.000 | 20:02 |
| Air Blank          | 0.000 | 20:04 |
| Subject Sample # 2 | 0.146 | 20:04 |
| Breath Volume      | 1.867 | 20:04 |
| Air Blank          | 0.000 | 20:05 |
| Control Test       | 0.079 | 20:05 |
| Air Blank          | 0.000 | 20:06 |
| Diagnostics Check  | OK    | 20:06 |

---

Data Download Date/Time: 07/01/2022 18:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: BEATTIE

DATE/TIME OF TEST: 06/22/2022 22:23:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:17 |
| Air Blank          | 0.000 | 22:17 |
| Control Test       | 0.081 | 22:17 |
| Air Blank          | 0.000 | 22:18 |
| Subject Sample # 1 | 0.150 | 22:19 |
| Breath Volume      | 1.628 | 22:19 |
| Air Blank          | 0.000 | 22:19 |
| Air Blank          | 0.000 | 22:21 |
| Subject Sample # 2 | 0.151 | 22:21 |
| Breath Volume      | 2.207 | 22:21 |
| Air Blank          | 0.000 | 22:22 |
| Control Test       | 0.080 | 22:22 |
| Air Blank          | 0.000 | 22:23 |
| Diagnostics Check  | OK    | 22:23 |

---

Data Download Date/Time: 07/01/2022 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARRISON,ANNA,L  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: GARRISON

DATE/TIME OF TEST: 06/23/2022 00:25:48

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:16 |                      |
| Air Blank         | 0.000 | 00:16 |                      |
| Control Test      | 0.080 | 00:16 |                      |
| Air Blank         | 0.000 | 00:17 |                      |
| Subject Sample    | # 1   | 00:20 | No Sample Provided   |
| Breath Volume     | 0.000 | 00:20 |                      |
| Air Blank         | 0.000 | 00:21 |                      |
| Air Blank         | 0.000 | 00:23 |                      |
| Subject Sample    | # 2   | 00:24 | Subject Test Refused |
| Breath Volume     | 0.035 | 00:24 |                      |
| Air Blank         | 0.000 | 00:24 |                      |
| Control Test      | 0.080 | 00:25 |                      |
| Air Blank         | 0.000 | 00:25 |                      |
| Diagnostics Check | OK    | 00:25 |                      |

---

Data Download Date/Time: 07/01/2022 18:46



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HOEFER

DATE/TIME OF TEST: 06/23/2022 03:26:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Control Test       | 0.081 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 1 | 0.000 | 03:22 |
| Breath Volume      | 2.941 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Air Blank          | 0.000 | 03:24 |
| Subject Sample # 2 | 0.000 | 03:25 |
| Breath Volume      | 2.617 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Control Test       | 0.081 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Diagnostics Check  | OK    | 03:26 |

---

Data Download Date/Time: 07/01/2022 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: WITKOSKY,OFFICER  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 14:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: GOESE

DATE/TIME OF TEST: 06/23/2022 15:55:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:49 |
| Air Blank          | 0.000 | 15:50 |
| Control Test       | 0.081 | 15:50 |
| Air Blank          | 0.000 | 15:50 |
| Subject Sample # 1 | 0.175 | 15:51 |
| Breath Volume      | 2.925 | 15:51 |
| Air Blank          | 0.000 | 15:51 |
| Air Blank          | 0.000 | 15:53 |
| Subject Sample # 2 | 0.175 | 15:53 |
| Breath Volume      | 3.019 | 15:53 |
| Air Blank          | 0.000 | 15:54 |
| Control Test       | 0.080 | 15:54 |
| Air Blank          | 0.000 | 15:55 |
| Diagnostics Check  | OK    | 15:55 |

---

Data Download Date/Time: 07/01/2022 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: WATSON,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 06/25/2022 00:30:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:24 |
| Air Blank          | 0.000 | 00:24 |
| Control Test       | 0.081 | 00:24 |
| Air Blank          | 0.000 | 00:25 |
| Subject Sample # 1 | 0.118 | 00:25 |
| Breath Volume      | 1.398 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Air Blank          | 0.000 | 00:28 |
| Subject Sample # 2 | 0.110 | 00:28 |
| Breath Volume      | 1.191 | 00:28 |
| Air Blank          | 0.000 | 00:29 |
| Control Test       | 0.081 | 00:29 |
| Air Blank          | 0.000 | 00:29 |
| Diagnostics Check  | OK    | 00:30 |

---

Data Download Date/Time: 07/01/2022 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY J,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J.  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: SPAULDING

DATE/TIME OF TEST: 06/26/2022 01:22:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:15 |
| Air Blank          | 0.000 | 01:16 |
| Control Test       | 0.081 | 01:16 |
| Air Blank          | 0.000 | 01:17 |
| Subject Sample # 1 | 0.375 | 01:17 |
| Breath Volume      | 1.789 | 01:17 |
| Air Blank          | 0.000 | 01:18 |
| Air Blank          | 0.000 | 01:20 |
| Subject Sample # 2 | 0.375 | 01:20 |
| Breath Volume      | 2.105 | 01:20 |
| Air Blank          | 0.000 | 01:21 |
| Control Test       | 0.080 | 01:21 |
| Air Blank          | 0.000 | 01:22 |
| Diagnostics Check  | OK    | 01:22 |

---

Data Download Date/Time: 07/01/2022 18:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NAMORS,J,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PIETRON,M  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: VARGAS ROMERO

DATE/TIME OF TEST: 06/26/2022 02:53:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:47 |
| Air Blank          | 0.000 | 02:48 |
| Control Test       | 0.081 | 02:48 |
| Air Blank          | 0.000 | 02:48 |
| Subject Sample # 1 | 0.087 | 02:49 |
| Breath Volume      | 3.207 | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 2 | 0.085 | 02:52 |
| Breath Volume      | 3.726 | 02:52 |
| Air Blank          | 0.000 | 02:52 |
| Control Test       | 0.081 | 02:53 |
| Air Blank          | 0.000 | 02:53 |
| Diagnostics Check  | OK    | 02:53 |

---

Data Download Date/Time: 07/01/2022 18:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERSTEN,DANIEL,N  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: NERSTEN',DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 06/26/2022 03:35:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:29 |
| Air Blank          | 0.000 | 03:29 |
| Control Test       | 0.081 | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Subject Sample # 1 | 0.127 | 03:30 |
| Breath Volume      | 2.960 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Air Blank          | 0.000 | 03:33 |
| Subject Sample # 2 | 0.122 | 03:33 |
| Breath Volume      | 2.441 | 03:33 |
| Air Blank          | 0.000 | 03:34 |
| Control Test       | 0.080 | 03:34 |
| Air Blank          | 0.000 | 03:35 |
| Diagnostics Check  | OK    | 03:35 |

---

Data Download Date/Time: 07/01/2022 18:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: WATSON,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 06/26/2022 04:30:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:24 |
| Air Blank          | 0.000 | 04:25 |
| Control Test       | 0.081 | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Subject Sample # 1 | 0.087 | 04:26 |
| Breath Volume      | 1.890 | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Air Blank          | 0.000 | 04:28 |
| Subject Sample # 2 | 0.083 | 04:29 |
| Breath Volume      | 1.839 | 04:29 |
| Air Blank          | 0.000 | 04:29 |
| Control Test       | 0.080 | 04:30 |
| Air Blank          | 0.000 | 04:30 |
| Diagnostics Check  | OK    | 04:30 |

---

Data Download Date/Time: 07/01/2022 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J.  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: SZARY

DATE/TIME OF TEST: 06/26/2022 20:00:46

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 19:46 |                    |
| Air Blank         | 0.000 | 19:47 |                    |
| Control Test      | 0.080 | 19:47 |                    |
| Air Blank         | 0.000 | 19:48 |                    |
| Subject Sample    | # 1   | 19:51 | No Sample Provided |
| Breath Volume     | 0.000 | 19:51 |                    |
| Air Blank         | 0.000 | 19:51 |                    |
| Air Blank         | 0.000 | 19:53 |                    |
| Subject Sample    | # 2   | 0.187 | No .020 agreement  |
| Breath Volume     | 2.355 | 19:55 |                    |
| Air Blank         | 0.000 | 19:56 |                    |
| Air Blank         | 0.000 | 19:58 |                    |
| Subject Sample    | # 3   | 0.164 | No .020 agreement  |
| Breath Volume     | 1.179 | 19:59 |                    |
| Air Blank         | 0.000 | 19:59 |                    |
| Control Test      | 0.079 | 20:00 |                    |
| Air Blank         | 0.000 | 20:00 |                    |
| Diagnostics Check | OK    | 20:00 |                    |

---

Data Download Date/Time: 07/01/2022 18:46



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERSON,CORY,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PIERSON,CORY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: PINA-SANTANA

DATE/TIME OF TEST: 06/26/2022 20:24:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:18 |
| Air Blank          | 0.000 | 20:19 |
| Control Test       | 0.078 | 20:19 |
| Air Blank          | 0.000 | 20:19 |
| Subject Sample # 1 | 0.298 | 20:20 |
| Breath Volume      | 1.300 | 20:20 |
| Air Blank          | 0.000 | 20:21 |
| Air Blank          | 0.000 | 20:22 |
| Subject Sample # 2 | 0.301 | 20:23 |
| Breath Volume      | 2.121 | 20:23 |
| Air Blank          | 0.000 | 20:23 |
| Control Test       | 0.078 | 20:24 |
| Air Blank          | 0.000 | 20:24 |
| Diagnostics Check  | OK    | 20:24 |

---

Data Download Date/Time: 07/01/2022 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: WATSON,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/27/2022 04:08:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:02 |
| Air Blank          | 0.000 | 04:02 |
| Control Test       | 0.080 | 04:02 |
| Air Blank          | 0.000 | 04:03 |
| Subject Sample # 1 | 0.208 | 04:04 |
| Breath Volume      | 1.980 | 04:04 |
| Air Blank          | 0.000 | 04:04 |
| Air Blank          | 0.000 | 04:06 |
| Subject Sample # 2 | 0.222 | 04:06 |
| Breath Volume      | 1.652 | 04:06 |
| Air Blank          | 0.000 | 04:07 |
| Control Test       | 0.080 | 04:07 |
| Air Blank          | 0.000 | 04:08 |
| Diagnostics Check  | OK    | 04:08 |

---

Data Download Date/Time: 07/01/2022 18:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MEISNER

DATE/TIME OF TEST: 06/27/2022 23:56:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:50 |
| Air Blank          | 0.000 | 23:51 |
| Control Test       | 0.080 | 23:51 |
| Air Blank          | 0.000 | 23:51 |
| Subject Sample # 1 | 0.000 | 23:52 |
| Breath Volume      | 2.039 | 23:52 |
| Air Blank          | 0.000 | 23:53 |
| Air Blank          | 0.000 | 23:55 |
| Subject Sample # 2 | 0.000 | 23:55 |
| Breath Volume      | 1.242 | 23:55 |
| Air Blank          | 0.000 | 23:55 |
| Control Test       | 0.081 | 23:56 |
| Air Blank          | 0.000 | 23:56 |
| Diagnostics Check  | OK    | 23:56 |

---

Data Download Date/Time: 07/01/2022 18:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DEP,R  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MILLER,DEP R  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/28/2022 02:02:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.081 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Subject Sample # 1 | 0.238 | 01:58 |
| Breath Volume      | 2.902 | 01:58 |
| Air Blank          | 0.000 | 01:58 |
| Air Blank          | 0.000 | 02:00 |
| Subject Sample # 2 | 0.239 | 02:01 |
| Breath Volume      | 2.964 | 02:01 |
| Air Blank          | 0.000 | 02:01 |
| Control Test       | 0.078 | 02:02 |
| Air Blank          | 0.000 | 02:02 |
| Diagnostics Check  | OK    | 02:02 |

---

Data Download Date/Time: 07/01/2022 18:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOLADVANDI, OFC. M.,  
BREATH TEST OPERATOR AGENCY: North Port PD  
ARRESTING OFFICER: FOLADVANDI, OFC. M.  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: O'DONNELL

DATE/TIME OF TEST: 06/29/2022 00:13:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:07 |
| Air Blank          | 0.000 | 00:07 |
| Control Test       | 0.080 | 00:07 |
| Air Blank          | 0.000 | 00:08 |
| Subject Sample # 1 | 0.174 | 00:08 |
| Breath Volume      | 3.007 | 00:08 |
| Air Blank          | 0.000 | 00:09 |
| Air Blank          | 0.000 | 00:11 |
| Subject Sample # 2 | 0.172 | 00:11 |
| Breath Volume      | 2.863 | 00:11 |
| Air Blank          | 0.000 | 00:12 |
| Control Test       | 0.080 | 00:12 |
| Air Blank          | 0.000 | 00:13 |
| Diagnostics Check  | OK    | 00:13 |

---

Data Download Date/Time: 07/01/2022 18:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SURRAN,DEPUTY,C  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: VALENTE,J  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: HAUERT

DATE/TIME OF TEST: 06/29/2022 13:38:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:31 |
| Air Blank          | 0.000 | 13:31 |
| Control Test       | 0.081 | 13:31 |
| Air Blank          | 0.000 | 13:32 |
| Subject Sample # 1 | 0.000 | 13:33 |
| Breath Volume      | 1.578 | 13:33 |
| Air Blank          | 0.000 | 13:34 |
| Air Blank          | 0.000 | 13:36 |
| Subject Sample # 2 | 0.000 | 13:37 |
| Breath Volume      | 1.699 | 13:37 |
| Air Blank          | 0.000 | 13:37 |
| Control Test       | 0.080 | 13:38 |
| Air Blank          | 0.000 | 13:38 |
| Diagnostics Check  | OK    | 13:38 |

---

Data Download Date/Time: 07/01/2022 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: DAVENPORT,DANIEL  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15421080A1  
DRY GAS EXPIRY DATE: 08/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: LOUGHNER

DATE/TIME OF TEST: 06/30/2022 00:45:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:39 |
| Air Blank          | 0.000 | 00:39 |
| Control Test       | 0.078 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Subject Sample # 1 | 0.266 | 00:40 |
| Breath Volume      | 1.937 | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 2 | 0.255 | 00:44 |
| Breath Volume      | 2.074 | 00:44 |
| Air Blank          | 0.000 | 00:44 |
| Control Test       | 0.080 | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Diagnostics Check  | OK    | 00:45 |

---

Data Download Date/Time: 07/01/2022 18:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACEBO,ANDRES,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: LUPL,V  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: DENNEN

DATE/TIME OF TEST: 06/05/2022 00:02:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:55 |
| Air Blank          | 0.000 | 23:56 |
| Control Test       | 0.080 | 23:56 |
| Air Blank          | 0.000 | 23:56 |
| Subject Sample # 2 | 0.142 | 23:58 |
| Breath Volume      | 1.933 | 23:58 |
| Air Blank          | 0.000 | 23:58 |
| Air Blank          | 0.000 | 00:00 |
| Subject Sample # 1 | 0.131 | 00:01 |
| Breath Volume      | 1.324 | 00:01 |
| Air Blank          | 0.000 | 00:01 |
| Control Test       | 0.079 | 00:02 |
| Air Blank          | 0.000 | 00:02 |
| Diagnostics Check  | OK    | 00:02 |

---

Data Download Date/Time: 06/26/2022 12:21



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACEBO,ANDRES,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: HARTMANSTORFER,C  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/05/2022 04:05:02

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:51 |                    |
| Air Blank         | 0.000 | 03:52 |                    |
| Control Test      | 0.079 | 03:52 |                    |
| Air Blank         | 0.000 | 03:53 |                    |
| Subject Sample    | # 1   | 03:56 | No Sample Provided |
| Breath Volume     | 0.000 | 03:56 |                    |
| Air Blank         | 0.000 | 03:56 |                    |
| Air Blank         | 0.000 | 03:58 |                    |
| Subject Sample    | # 2   | 0.233 | 03:59              |
| Breath Volume     | 1.492 | 03:59 |                    |
| Air Blank         | 0.000 | 04:00 |                    |
| Air Blank         | 0.000 | 04:02 |                    |
| Subject Sample    | # 3   | 0.252 | 04:03              |
| Breath Volume     | 1.324 | 04:03 |                    |
| Air Blank         | 0.000 | 04:04 |                    |
| Control Test      | 0.080 | 04:04 |                    |
| Air Blank         | 0.000 | 04:04 |                    |
| Diagnostics Check | OK    | 04:05 |                    |

---

Data Download Date/Time: 06/26/2022 12:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACEBO,ANDRES,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: RICHARDSON,B  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: JANOSIK

DATE/TIME OF TEST: 06/09/2022 23:26:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:19 |
| Air Blank          | 0.000 | 23:19 |
| Control Test       | 0.079 | 23:19 |
| Air Blank          | 0.000 | 23:20 |
| Subject Sample # 1 | 0.260 | 23:21 |
| Breath Volume      | 3.394 | 23:21 |
| Air Blank          | 0.000 | 23:22 |
| Air Blank          | 0.000 | 23:24 |
| Subject Sample # 2 | 0.252 | 23:24 |
| Breath Volume      | 2.835 | 23:24 |
| Air Blank          | 0.000 | 23:25 |
| Control Test       | 0.079 | 23:25 |
| Air Blank          | 0.000 | 23:26 |
| Diagnostics Check  | OK    | 23:26 |

---

Data Download Date/Time: 06/26/2022 12:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GIRARD,DFC  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CHEN

DATE/TIME OF TEST: 06/16/2022 03:47:37

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:30 |                |
| Air Blank         | 0.000     | 03:31 |                |
| Control Test      | 0.079     | 03:31 |                |
| Air Blank         | 0.000     | 03:32 |                |
| Subject Sample    | # 1 0.112 | 03:35 | Volume Not Met |
| Breath Volume     | 0.226     | 03:35 |                |
| Air Blank         | 0.000     | 03:35 |                |
| Air Blank         | 0.000     | 03:37 |                |
| Subject Sample    | # 2 0.141 | 03:40 |                |
| Breath Volume     | 1.308     | 03:40 |                |
| Air Blank         | 0.000     | 03:41 |                |
| Air Blank         | 0.000     | 03:42 |                |
| Subject Sample    | # 3 0.137 | 03:46 |                |
| Breath Volume     | 1.246     | 03:46 |                |
| Air Blank         | 0.000     | 03:46 |                |
| Control Test      | 0.078     | 03:47 |                |
| Air Blank         | 0.000     | 03:47 |                |
| Diagnostics Check | OK        | 03:47 |                |

---

Data Download Date/Time: 06/26/2022 12:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGAN,MELISSA,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CONANT,R  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: TRIPPETT

DATE/TIME OF TEST: 06/17/2022 20:12:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:04 |
| Air Blank          | 0.000 | 20:05 |
| Control Test       | 0.080 | 20:05 |
| Air Blank          | 0.000 | 20:05 |
| Subject Sample # 1 | 0.185 | 20:07 |
| Breath Volume      | 1.308 | 20:07 |
| Air Blank          | 0.000 | 20:07 |
| Air Blank          | 0.000 | 20:09 |
| Subject Sample # 2 | 0.184 | 20:11 |
| Breath Volume      | 1.718 | 20:11 |
| Air Blank          | 0.000 | 20:11 |
| Control Test       | 0.080 | 20:12 |
| Air Blank          | 0.000 | 20:12 |
| Diagnostics Check  | OK    | 20:12 |

---

Data Download Date/Time: 06/26/2022 12:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACEBO,ANDRES,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CONANT,R  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BAIN

DATE/TIME OF TEST: 06/19/2022 02:23:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:17 |
| Air Blank          | 0.000 | 02:17 |
| Control Test       | 0.080 | 02:17 |
| Air Blank          | 0.000 | 02:18 |
| Subject Sample # 1 | 0.000 | 02:19 |
| Breath Volume      | 3.515 | 02:19 |
| Air Blank          | 0.000 | 02:19 |
| Air Blank          | 0.000 | 02:21 |
| Subject Sample # 2 | 0.000 | 02:22 |
| Breath Volume      | 3.648 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Control Test       | 0.080 | 02:23 |
| Air Blank          | 0.000 | 02:23 |
| Diagnostics Check  | OK    | 02:23 |

---

Data Download Date/Time: 06/26/2022 12:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSSO,C,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: SCHWARTZ,DFC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: REBA

DATE/TIME OF TEST: 06/20/2022 23:12:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Control Test       | 0.079 | 23:06 |
| Air Blank          | 0.000 | 23:06 |
| Subject Sample # 1 | 0.159 | 23:08 |
| Breath Volume      | 2.562 | 23:08 |
| Air Blank          | 0.000 | 23:08 |
| Air Blank          | 0.000 | 23:10 |
| Subject Sample # 2 | 0.165 | 23:11 |
| Breath Volume      | 3.031 | 23:11 |
| Air Blank          | 0.000 | 23:12 |
| Control Test       | 0.079 | 23:12 |
| Air Blank          | 0.000 | 23:12 |
| Diagnostics Check  | OK    | 23:12 |

---

Data Download Date/Time: 06/26/2022 12:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACEBO,ANDRES,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BRASCH,D  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: MCGRAW

DATE/TIME OF TEST: 06/23/2022 05:37:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:29 |
| Air Blank          | 0.000 | 05:30 |
| Control Test       | 0.080 | 05:30 |
| Air Blank          | 0.000 | 05:31 |
| Subject Sample # 1 | 0.168 | 05:33 |
| Breath Volume      | 3.167 | 05:33 |
| Air Blank          | 0.000 | 05:33 |
| Air Blank          | 0.000 | 05:35 |
| Subject Sample # 2 | 0.172 | 05:36 |
| Breath Volume      | 3.191 | 05:36 |
| Air Blank          | 0.000 | 05:36 |
| Control Test       | 0.080 | 05:37 |
| Air Blank          | 0.000 | 05:37 |
| Diagnostics Check  | OK    | 05:37 |

---

Data Download Date/Time: 06/26/2022 12:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACEBO,A,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ACEBO,A  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: ADMIN  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 06/25/2022 21:47:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:39 |
| Air Blank          | 0.000 | 21:40 |
| Control Test       | 0.080 | 21:40 |
| Air Blank          | 0.000 | 21:41 |
| Subject Sample # 1 | 0.000 | 21:42 |
| Breath Volume      | 1.687 | 21:42 |
| Air Blank          | 0.000 | 21:42 |
| Air Blank          | 0.000 | 21:44 |
| Subject Sample # 2 | 0.000 | 21:46 |
| Breath Volume      | 1.378 | 21:46 |
| Air Blank          | 0.000 | 21:46 |
| Control Test       | 0.080 | 21:46 |
| Air Blank          | 0.000 | 21:47 |
| Diagnostics Check  | OK    | 21:47 |

---

Data Download Date/Time: 06/26/2022 12:22



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITMAN,JACELYN,A  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CULLIFER

DATE/TIME OF TEST: 06/25/2022 23:07:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:00 |
| Air Blank          | 0.000 | 23:01 |
| Control Test       | 0.080 | 23:01 |
| Air Blank          | 0.000 | 23:02 |
| Subject Sample # 1 | 0.210 | 23:02 |
| Breath Volume      | 1.414 | 23:02 |
| Air Blank          | 0.000 | 23:03 |
| Air Blank          | 0.000 | 23:04 |
| Subject Sample # 2 | 0.198 | 23:05 |
| Breath Volume      | 1.738 | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Control Test       | 0.080 | 23:06 |
| Air Blank          | 0.000 | 23:07 |
| Diagnostics Check  | OK    | 23:07 |

---

Data Download Date/Time: 06/26/2022 12:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: LITHERLAND,OFC  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 06/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: MATTIELLO

DATE/TIME OF TEST: 06/29/2022 21:21:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:13 |
| Air Blank          | 0.000 | 21:14 |
| Control Test       | 0.079 | 21:14 |
| Air Blank          | 0.000 | 21:14 |
| Subject Sample # 1 | 0.188 | 21:15 |
| Breath Volume      | 3.464 | 21:15 |
| Air Blank          | 0.000 | 21:16 |
| Air Blank          | 0.000 | 21:18 |
| Subject Sample # 2 | 0.168 | 21:19 |
| Breath Volume      | 1.585 | 21:19 |
| Air Blank          | 0.000 | 21:20 |
| Control Test       | 0.079 | 21:20 |
| Air Blank          | 0.000 | 21:21 |
| Diagnostics Check  | OK    | 21:21 |

---

Data Download Date/Time: 08/04/2022 09:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: FRIEND,DFC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: DELISI

DATE/TIME OF TEST: 06/30/2022 03:35:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.080 | 03:26 |
| Air Blank          | 0.000 | 03:26 |
| Subject Sample # 1 | 0.000 | 03:29 |
| Breath Volume      | 1.773 | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 2 | 0.000 | 03:33 |
| Breath Volume      | 1.535 | 03:33 |
| Air Blank          | 0.000 | 03:34 |
| Control Test       | 0.079 | 03:34 |
| Air Blank          | 0.000 | 03:35 |
| Diagnostics Check  | OK    | 03:35 |

---

Data Download Date/Time: 08/04/2022 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITMAN,JACELYN,A  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: REILLY,SEAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 06/30/2022 04:30:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Control Test       | 0.080 | 04:23 |
| Air Blank          | 0.000 | 04:24 |
| Subject Sample # 1 | 0.344 | 04:25 |
| Breath Volume      | 1.851 | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Air Blank          | 0.000 | 04:27 |
| Subject Sample # 2 | 0.348 | 04:28 |
| Breath Volume      | 1.593 | 04:28 |
| Air Blank          | 0.000 | 04:29 |
| Control Test       | 0.079 | 04:29 |
| Air Blank          | 0.000 | 04:29 |
| Diagnostics Check  | OK    | 04:30 |

---

Data Download Date/Time: 08/04/2022 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADKINS, JAKE, M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: REEVES, DAVID  
ARRESTING OFFICER AGENCY: Trenton PD  
LAST AGENCY INSPECTION: 05/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: HOOPER

DATE/TIME OF TEST: 06/04/2022 21:18:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:11 |
| Air Blank          | 0.000 | 21:11 |
| Control Test       | 0.081 | 21:11 |
| Air Blank          | 0.000 | 21:12 |
| Subject Sample # 1 | 0.249 | 21:13 |
| Breath Volume      | 1.359 | 21:13 |
| Air Blank          | 0.000 | 21:14 |
| Air Blank          | 0.000 | 21:16 |
| Subject Sample # 2 | 0.267 | 21:16 |
| Breath Volume      | 1.816 | 21:16 |
| Air Blank          | 0.000 | 21:17 |
| Control Test       | 0.078 | 21:17 |
| Air Blank          | 0.000 | 21:18 |
| Diagnostics Check  | OK    | 21:18 |

---

Data Download Date/Time: 07/01/2022 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,KEVIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DAVIS,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 06/16/2022 01:10:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:07 |                      |
| Air Blank         | 0.000 | 01:07 |                      |
| Control Test      | 0.080 | 01:08 |                      |
| Air Blank         | 0.000 | 01:08 |                      |
| Subject Sample    | # 1   | 01:08 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:08 |                      |
| Air Blank         | 0.000 | 01:09 |                      |
| Control Test      | 0.080 | 01:09 |                      |
| Air Blank         | 0.000 | 01:10 |                      |
| Diagnostics Check | OK    | 01:10 |                      |

---

Data Download Date/Time: 07/01/2022 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS, MICHAEL, T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: THOMAS, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 12:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SIXTO

DATE/TIME OF TEST: 06/18/2022 01:23:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:20 |                      |
| Air Blank         | 0.000 | 01:21 |                      |
| Control Test      | 0.081 | 01:21 |                      |
| Air Blank         | 0.000 | 01:22 |                      |
| Subject Sample    | # 1   | 01:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:22 |                      |
| Air Blank         | 0.000 | 01:22 |                      |
| Control Test      | 0.081 | 01:23 |                      |
| Air Blank         | 0.000 | 01:23 |                      |
| Diagnostics Check | OK    | 01:23 |                      |

---

Data Download Date/Time: 07/01/2022 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,BONNIE,S  
BREATH TEST OPERATOR AGENCY: Levy County SO  
ARRESTING OFFICER: MILLER,THEODORE  
ARRESTING OFFICER AGENCY: Levy County SO  
LAST AGENCY INSPECTION: 05/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: PIRL

DATE/TIME OF TEST: 06/18/2022 22:43:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Control Test       | 0.081 | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Subject Sample # 1 | 0.162 | 22:38 |
| Breath Volume      | 2.285 | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Subject Sample # 2 | 0.151 | 22:41 |
| Breath Volume      | 2.425 | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Control Test       | 0.079 | 22:42 |
| Air Blank          | 0.000 | 22:42 |
| Diagnostics Check  | OK    | 22:43 |

---

Data Download Date/Time: 07/01/2022 11:35



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,KEVIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DAVIS,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 05:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 06/21/2022 07:01:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:54 |
| Air Blank          | 0.000 | 06:54 |
| Control Test       | 0.080 | 06:55 |
| Air Blank          | 0.000 | 06:55 |
| Subject Sample # 1 | 0.164 | 06:56 |
| Breath Volume      | 2.398 | 06:56 |
| Air Blank          | 0.000 | 06:57 |
| Air Blank          | 0.000 | 06:58 |
| Subject Sample # 2 | 0.160 | 06:59 |
| Breath Volume      | 2.011 | 06:59 |
| Air Blank          | 0.000 | 07:00 |
| Control Test       | 0.078 | 07:00 |
| Air Blank          | 0.000 | 07:00 |
| Diagnostics Check  | OK    | 07:01 |

---

Data Download Date/Time: 07/01/2022 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,KEVIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DAVIS,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 06/25/2022 22:28:55

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:26 |                      |
| Air Blank         | 0.000 | 22:26 |                      |
| Control Test      | 0.080 | 22:26 |                      |
| Air Blank         | 0.000 | 22:27 |                      |
| Subject Sample    | # 1   | 22:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:27 |                      |
| Air Blank         | 0.000 | 22:27 |                      |
| Control Test      | 0.080 | 22:28 |                      |
| Air Blank         | 0.000 | 22:28 |                      |
| Diagnostics Check | OK    | 22:28 |                      |

---

Data Download Date/Time: 07/01/2022 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS, MICHAEL, T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: REEVES, DAVID  
ARRESTING OFFICER AGENCY: Trenton PD  
LAST AGENCY INSPECTION: 06/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: FULMER

DATE/TIME OF TEST: 06/27/2022 23:39:03

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:36 |                      |
| Air Blank         | 0.000 | 23:36 |                      |
| Control Test      | 0.077 | 23:36 |                      |
| Air Blank         | 0.000 | 23:37 |                      |
| Subject Sample    | # 1   | 23:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:37 |                      |
| Air Blank         | 0.000 | 23:38 |                      |
| Control Test      | 0.077 | 23:38 |                      |
| Air Blank         | 0.000 | 23:38 |                      |
| Diagnostics Check | OK    | 23:39 |                      |

---

Data Download Date/Time: 07/01/2022 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS, MICHAEL, T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: THOMAS, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 06/29/2022 04:56:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:49 |
| Air Blank          | 0.000 | 04:49 |
| Control Test       | 0.080 | 04:49 |
| Air Blank          | 0.000 | 04:50 |
| Subject Sample # 1 | 0.142 | 04:51 |
| Breath Volume      | 2.441 | 04:51 |
| Air Blank          | 0.000 | 04:51 |
| Air Blank          | 0.000 | 04:53 |
| Subject Sample # 2 | 0.161 | 04:54 |
| Breath Volume      | 2.777 | 04:54 |
| Air Blank          | 0.000 | 04:55 |
| Control Test       | 0.081 | 04:55 |
| Air Blank          | 0.000 | 04:56 |
| Diagnostics Check  | OK    | 04:56 |

---

Data Download Date/Time: 07/01/2022 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LUGO,OSCAR  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: VELAZQUEZHERNANDEZ

DATE/TIME OF TEST: 06/01/2022 03:33:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.082 | 03:27 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 1 | 0.086 | 03:28 |
| Breath Volume      | 3.960 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Air Blank          | 0.000 | 03:31 |
| Subject Sample # 2 | 0.085 | 03:31 |
| Breath Volume      | 3.675 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.082 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Diagnostics Check  | OK    | 03:33 |

---

Data Download Date/Time: 06/13/2022 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RIDEAUX,ADAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 06/02/2022 02:33:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:27 |
| Air Blank          | 0.000 | 02:28 |
| Control Test       | 0.082 | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 1 | 0.000 | 02:29 |
| Breath Volume      | 2.843 | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Air Blank          | 0.000 | 02:32 |
| Subject Sample # 2 | 0.000 | 02:32 |
| Breath Volume      | 2.605 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Control Test       | 0.082 | 02:33 |
| Air Blank          | 0.000 | 02:33 |
| Diagnostics Check  | OK    | 02:33 |

---

Data Download Date/Time: 06/13/2022 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: STOWE,NICOLAS  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 06/02/2022 03:46:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:40 |
| Air Blank          | 0.000 | 03:40 |
| Control Test       | 0.081 | 03:40 |
| Air Blank          | 0.000 | 03:41 |
| Subject Sample # 1 | 0.182 | 03:41 |
| Breath Volume      | 1.761 | 03:41 |
| Air Blank          | 0.000 | 03:42 |
| Air Blank          | 0.000 | 03:44 |
| Subject Sample # 2 | 0.180 | 03:44 |
| Breath Volume      | 2.250 | 03:44 |
| Air Blank          | 0.000 | 03:45 |
| Control Test       | 0.082 | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Diagnostics Check  | OK    | 03:46 |

---

Data Download Date/Time: 06/13/2022 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DIDOMENICO,GINO  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MARTINEZ CASANOVA

DATE/TIME OF TEST: 06/04/2022 03:45:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:38 |
| Air Blank          | 0.000 | 03:39 |
| Control Test       | 0.083 | 03:39 |
| Air Blank          | 0.000 | 03:39 |
| Subject Sample # 1 | 0.000 | 03:41 |
| Breath Volume      | 1.835 | 03:41 |
| Air Blank          | 0.000 | 03:41 |
| Air Blank          | 0.000 | 03:43 |
| Subject Sample # 2 | 0.000 | 03:44 |
| Breath Volume      | 2.457 | 03:44 |
| Air Blank          | 0.000 | 03:44 |
| Control Test       | 0.082 | 03:45 |
| Air Blank          | 0.000 | 03:45 |
| Diagnostics Check  | OK    | 03:45 |

---

Data Download Date/Time: 06/13/2022 14:35



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LEONARD,WAKENS  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: GOUR

DATE/TIME OF TEST: 06/04/2022 04:22:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Control Test       | 0.082 | 04:16 |
| Air Blank          | 0.000 | 04:17 |
| Subject Sample # 1 | 0.216 | 04:17 |
| Breath Volume      | 1.429 | 04:17 |
| Air Blank          | 0.000 | 04:18 |
| Air Blank          | 0.000 | 04:20 |
| Subject Sample # 2 | 0.223 | 04:20 |
| Breath Volume      | 3.074 | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Control Test       | 0.082 | 04:21 |
| Air Blank          | 0.000 | 04:22 |
| Diagnostics Check  | OK    | 04:22 |

---

Data Download Date/Time: 06/13/2022 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BRAZEE,COLE  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 06/05/2022 00:37:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.082 | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Subject Sample # 1 | 0.135 | 00:32 |
| Breath Volume      | 2.718 | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Air Blank          | 0.000 | 00:35 |
| Subject Sample # 2 | 0.132 | 00:35 |
| Breath Volume      | 3.191 | 00:35 |
| Air Blank          | 0.000 | 00:36 |
| Control Test       | 0.081 | 00:36 |
| Air Blank          | 0.000 | 00:37 |
| Diagnostics Check  | OK    | 00:37 |

---

Data Download Date/Time: 06/13/2022 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANER, OSVALDO,  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BRUNY, AINALEM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:23

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/07/2022 07:59:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:49 |
| Air Blank          | 0.000 | 07:49 |
| Control Test       | 0.082 | 07:50 |
| Air Blank          | 0.000 | 07:50 |
| Subject Sample # 1 | 0.317 | 07:52 |
| Breath Volume      | 2.925 | 07:52 |
| Air Blank          | 0.000 | 07:52 |
| Air Blank          | 0.000 | 07:54 |
| Subject Sample # 2 | 0.304 | 07:57 |
| Breath Volume      | 1.312 | 07:57 |
| Air Blank          | 0.000 | 07:58 |
| Control Test       | 0.081 | 07:58 |
| Air Blank          | 0.000 | 07:58 |
| Diagnostics Check  | OK    | 07:59 |

---

Data Download Date/Time: 06/13/2022 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: IRELAND,DAVID  
ARRESTING OFFICER AGENCY: Edgewood PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: WOJCIK

DATE/TIME OF TEST: 06/08/2022 01:13:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:07 |
| Air Blank          | 0.000 | 01:08 |
| Control Test       | 0.081 | 01:08 |
| Air Blank          | 0.000 | 01:08 |
| Subject Sample # 1 | 0.175 | 01:09 |
| Breath Volume      | 1.289 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Air Blank          | 0.000 | 01:11 |
| Subject Sample # 2 | 0.178 | 01:12 |
| Breath Volume      | 1.285 | 01:12 |
| Air Blank          | 0.000 | 01:12 |
| Control Test       | 0.081 | 01:12 |
| Air Blank          | 0.000 | 01:13 |
| Diagnostics Check  | OK    | 01:13 |

---

Data Download Date/Time: 06/13/2022 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GONZALEZ,JOEL  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 06/08/2022 06:32:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:24 |
| Air Blank          | 0.000 | 06:25 |
| Control Test       | 0.082 | 06:25 |
| Air Blank          | 0.000 | 06:25 |
| Subject Sample # 1 | 0.055 | 06:27 |
| Breath Volume      | 1.800 | 06:27 |
| Air Blank          | 0.000 | 06:28 |
| Air Blank          | 0.000 | 06:30 |
| Subject Sample # 2 | 0.055 | 06:30 |
| Breath Volume      | 1.687 | 06:30 |
| Air Blank          | 0.000 | 06:31 |
| Control Test       | 0.081 | 06:31 |
| Air Blank          | 0.000 | 06:32 |
| Diagnostics Check  | OK    | 06:32 |

---

Data Download Date/Time: 06/13/2022 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CHEN,QI  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: ARBOITE

DATE/TIME OF TEST: 06/09/2022 06:03:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:57 |
| Air Blank          | 0.000 | 05:57 |
| Control Test       | 0.083 | 05:57 |
| Air Blank          | 0.000 | 05:58 |
| Subject Sample # 1 | 0.121 | 05:58 |
| Breath Volume      | 1.953 | 05:58 |
| Air Blank          | 0.000 | 05:59 |
| Air Blank          | 0.000 | 06:01 |
| Subject Sample # 2 | 0.113 | 06:02 |
| Breath Volume      | 1.714 | 06:02 |
| Air Blank          | 0.000 | 06:02 |
| Control Test       | 0.082 | 06:03 |
| Air Blank          | 0.000 | 06:03 |
| Diagnostics Check  | OK    | 06:03 |

---

Data Download Date/Time: 06/13/2022 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DAUGHERTY,KENYATTA  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 06/09/2022 21:47:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:41 |
| Air Blank          | 0.000 | 21:41 |
| Control Test       | 0.081 | 21:41 |
| Air Blank          | 0.000 | 21:42 |
| Subject Sample # 1 | 0.234 | 21:42 |
| Breath Volume      | 3.308 | 21:42 |
| Air Blank          | 0.000 | 21:43 |
| Air Blank          | 0.000 | 21:45 |
| Subject Sample # 2 | 0.236 | 21:45 |
| Breath Volume      | 3.285 | 21:45 |
| Air Blank          | 0.000 | 21:46 |
| Control Test       | 0.081 | 21:46 |
| Air Blank          | 0.000 | 21:47 |
| Diagnostics Check  | OK    | 21:47 |

---

Data Download Date/Time: 06/13/2022 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LIESENFELT,JASON  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: NOLASCO RAMIREZ

DATE/TIME OF TEST: 06/10/2022 04:03:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Control Test       | 0.081 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 1 | 0.181 | 03:59 |
| Breath Volume      | 2.324 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Air Blank          | 0.000 | 04:02 |
| Subject Sample # 2 | 0.170 | 04:02 |
| Breath Volume      | 2.648 | 04:02 |
| Air Blank          | 0.000 | 04:03 |
| Control Test       | 0.081 | 04:03 |
| Air Blank          | 0.000 | 04:03 |
| Diagnostics Check  | OK    | 04:03 |

---

Data Download Date/Time: 06/13/2022 14:35



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KING,ARTHUR  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: CELIS

DATE/TIME OF TEST: 06/11/2022 05:18:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:12 |
| Air Blank          | 0.000 | 05:12 |
| Control Test       | 0.082 | 05:13 |
| Air Blank          | 0.000 | 05:13 |
| Subject Sample # 1 | 0.000 | 05:14 |
| Breath Volume      | 1.496 | 05:14 |
| Air Blank          | 0.000 | 05:14 |
| Air Blank          | 0.000 | 05:16 |
| Subject Sample # 2 | 0.000 | 05:17 |
| Breath Volume      | 2.027 | 05:17 |
| Air Blank          | 0.000 | 05:17 |
| Control Test       | 0.081 | 05:17 |
| Air Blank          | 0.000 | 05:18 |
| Diagnostics Check  | OK    | 05:18 |

---

Data Download Date/Time: 06/13/2022 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: IRELAND,DAVID  
ARRESTING OFFICER AGENCY: Edgewood PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/12/2022 02:03:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:56 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.082 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Subject Sample # 1 | 0.218 | 01:59 |
| Breath Volume      | 1.730 | 01:59 |
| Air Blank          | 0.000 | 01:59 |
| Air Blank          | 0.000 | 02:01 |
| Subject Sample # 2 | 0.212 | 02:02 |
| Breath Volume      | 1.296 | 02:02 |
| Air Blank          | 0.000 | 02:02 |
| Control Test       | 0.081 | 02:03 |
| Air Blank          | 0.000 | 02:03 |
| Diagnostics Check  | OK    | 02:03 |

---

Data Download Date/Time: 06/13/2022 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KISSANE,BRIAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/16/2022 02:04:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:57 |
| Air Blank          | 0.000 | 01:57 |
| Control Test       | 0.082 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Subject Sample # 1 | 0.154 | 01:59 |
| Breath Volume      | 2.074 | 01:59 |
| Air Blank          | 0.000 | 02:00 |
| Air Blank          | 0.000 | 02:01 |
| Subject Sample # 2 | 0.150 | 02:02 |
| Breath Volume      | 1.449 | 02:02 |
| Air Blank          | 0.000 | 02:03 |
| Control Test       | 0.082 | 02:03 |
| Air Blank          | 0.000 | 02:03 |
| Diagnostics Check  | OK    | 02:04 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARWOOD,ARMANDO  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MOSES

DATE/TIME OF TEST: 06/17/2022 01:31:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:24 |
| Air Blank          | 0.000 | 01:24 |
| Control Test       | 0.082 | 01:25 |
| Air Blank          | 0.000 | 01:25 |
| Subject Sample # 1 | 0.270 | 01:26 |
| Breath Volume      | 3.179 | 01:26 |
| Air Blank          | 0.000 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Subject Sample # 2 | 0.261 | 01:29 |
| Breath Volume      | 2.859 | 01:29 |
| Air Blank          | 0.000 | 01:30 |
| Control Test       | 0.081 | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Diagnostics Check  | OK    | 01:31 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SOSA,EMMANUEL  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CASTILLORAMIREZ

DATE/TIME OF TEST: 06/18/2022 01:27:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:19 |
| Air Blank          | 0.000 | 01:20 |
| Control Test       | 0.081 | 01:20 |
| Air Blank          | 0.000 | 01:21 |
| Subject Sample # 1 | 0.215 | 01:23 |
| Breath Volume      | 1.726 | 01:23 |
| Air Blank          | 0.000 | 01:24 |
| Air Blank          | 0.000 | 01:25 |
| Subject Sample # 2 | 0.220 | 01:26 |
| Breath Volume      | 1.671 | 01:26 |
| Air Blank          | 0.000 | 01:26 |
| Control Test       | 0.080 | 01:27 |
| Air Blank          | 0.000 | 01:27 |
| Diagnostics Check  | OK    | 01:27 |

---

Data Download Date/Time: 07/05/2022 14:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BAKSH,SAEED  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: DONCILIO

DATE/TIME OF TEST: 06/18/2022 02:46:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:40 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.081 | 02:41 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 1 | 0.177 | 02:42 |
| Breath Volume      | 1.593 | 02:42 |
| Air Blank          | 0.000 | 02:42 |
| Air Blank          | 0.000 | 02:44 |
| Subject Sample # 2 | 0.180 | 02:45 |
| Breath Volume      | 2.003 | 02:45 |
| Air Blank          | 0.000 | 02:45 |
| Control Test       | 0.081 | 02:46 |
| Air Blank          | 0.000 | 02:46 |
| Diagnostics Check  | OK    | 02:46 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DANIEL,CONNELL  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ODEUS

DATE/TIME OF TEST: 06/18/2022 03:52:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:44 |
| Air Blank          | 0.000 | 03:44 |
| Control Test       | 0.082 | 03:45 |
| Air Blank          | 0.000 | 03:45 |
| Subject Sample # 1 | 0.200 | 03:47 |
| Breath Volume      | 1.664 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Air Blank          | 0.000 | 03:50 |
| Subject Sample # 2 | 0.192 | 03:51 |
| Breath Volume      | 1.738 | 03:51 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.080 | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Diagnostics Check  | OK    | 03:52 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GORDON,CURTIS  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 06/18/2022 10:05:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:55 |
| Air Blank          | 0.000 | 09:56 |
| Control Test       | 0.081 | 09:56 |
| Air Blank          | 0.000 | 09:57 |
| Subject Sample # 1 | 0.204 | 09:58 |
| Breath Volume      | 3.070 | 09:58 |
| Air Blank          | 0.000 | 09:58 |
| Air Blank          | 0.000 | 10:00 |
| Subject Sample # 2 | 0.183 | 10:01 |
| Breath Volume      | 1.355 | 10:01 |
| Air Blank          | 0.000 | 10:01 |
| Air Blank          | 0.000 | 10:03 |
| Subject Sample # 3 | 0.183 | 10:03 |
| Breath Volume      | 1.539 | 10:03 |
| Air Blank          | 0.000 | 10:04 |
| Control Test       | 0.081 | 10:04 |
| Air Blank          | 0.000 | 10:05 |
| Diagnostics Check  | OK    | 10:05 |

---

Data Download Date/Time: 07/05/2022 14:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BAKSH,SAEED  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: USECHE

DATE/TIME OF TEST: 06/19/2022 02:17:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:09 |
| Air Blank          | 0.000 | 02:09 |
| Control Test       | 0.082 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Subject Sample # 1 | 0.134 | 02:12 |
| Breath Volume      | 1.660 | 02:12 |
| Air Blank          | 0.000 | 02:13 |
| Air Blank          | 0.000 | 02:15 |
| Subject Sample # 2 | 0.134 | 02:15 |
| Breath Volume      | 2.613 | 02:15 |
| Air Blank          | 0.000 | 02:16 |
| Control Test       | 0.081 | 02:16 |
| Air Blank          | 0.000 | 02:17 |
| Diagnostics Check  | OK    | 02:17 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LIESENFELT,JASON  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: DATIZ SANCHEZ

DATE/TIME OF TEST: 06/20/2022 05:07:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:00 |
| Air Blank          | 0.000 | 05:00 |
| Control Test       | 0.081 | 05:01 |
| Air Blank          | 0.000 | 05:01 |
| Subject Sample # 1 | 0.216 | 05:02 |
| Breath Volume      | 1.121 | 05:02 |
| Air Blank          | 0.000 | 05:03 |
| Air Blank          | 0.000 | 05:05 |
| Subject Sample # 2 | 0.230 | 05:06 |
| Breath Volume      | 2.625 | 05:06 |
| Air Blank          | 0.000 | 05:06 |
| Control Test       | 0.081 | 05:07 |
| Air Blank          | 0.000 | 05:07 |
| Diagnostics Check  | OK    | 05:07 |

---

Data Download Date/Time: 07/05/2022 14:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KING,ARTHUR  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BRINK

DATE/TIME OF TEST: 06/21/2022 04:07:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Control Test       | 0.082 | 04:01 |
| Air Blank          | 0.000 | 04:02 |
| Subject Sample # 1 | 0.219 | 04:02 |
| Breath Volume      | 1.378 | 04:02 |
| Air Blank          | 0.000 | 04:03 |
| Air Blank          | 0.000 | 04:05 |
| Subject Sample # 2 | 0.223 | 04:05 |
| Breath Volume      | 1.140 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Control Test       | 0.082 | 04:06 |
| Air Blank          | 0.000 | 04:07 |
| Diagnostics Check  | OK    | 04:07 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LITTLE,ELAINE  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 17:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: AMIC

DATE/TIME OF TEST: 06/21/2022 18:41:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:34 |
| Air Blank          | 0.000 | 18:35 |
| Control Test       | 0.082 | 18:35 |
| Air Blank          | 0.000 | 18:35 |
| Subject Sample # 1 | 0.158 | 18:36 |
| Breath Volume      | 1.457 | 18:36 |
| Air Blank          | 0.000 | 18:37 |
| Air Blank          | 0.000 | 18:39 |
| Subject Sample # 2 | 0.149 | 18:39 |
| Breath Volume      | 1.539 | 18:39 |
| Air Blank          | 0.000 | 18:40 |
| Control Test       | 0.081 | 18:40 |
| Air Blank          | 0.000 | 18:41 |
| Diagnostics Check  | OK    | 18:41 |

---

Data Download Date/Time: 07/05/2022 14:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LIESENFELT,JASON  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: RESEDNIZ MARTINEZ

DATE/TIME OF TEST: 06/23/2022 02:49:36

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 02:37 |                |
| Air Blank         | 0.000     | 02:37 |                |
| Control Test      | 0.082     | 02:37 |                |
| Air Blank         | 0.000     | 02:38 |                |
| Subject Sample    | # 1 0.103 | 02:41 | Volume Not Met |
| Breath Volume     | 0.570     | 02:41 |                |
| Air Blank         | 0.000     | 02:42 |                |
| Air Blank         | 0.000     | 02:44 |                |
| Subject Sample    | # 2 0.100 | 02:45 |                |
| Breath Volume     | 1.859     | 02:45 |                |
| Air Blank         | 0.000     | 02:45 |                |
| Air Blank         | 0.000     | 02:47 |                |
| Subject Sample    | # 3 0.098 | 02:48 |                |
| Breath Volume     | 1.296     | 02:48 |                |
| Air Blank         | 0.000     | 02:48 |                |
| Control Test      | 0.080     | 02:49 |                |
| Air Blank         | 0.000     | 02:49 |                |
| Diagnostics Check | OK        | 02:49 |                |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CAMPBELL,ROBERT  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: REYES FIDEL

DATE/TIME OF TEST: 06/24/2022 02:50:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:41 |
| Air Blank          | 0.000 | 02:42 |
| Control Test       | 0.082 | 02:42 |
| Air Blank          | 0.000 | 02:43 |
| Subject Sample # 1 | 0.216 | 02:45 |
| Breath Volume      | 2.265 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Air Blank          | 0.000 | 02:47 |
| Subject Sample # 2 | 0.209 | 02:48 |
| Breath Volume      | 1.390 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.080 | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Diagnostics Check  | OK    | 02:50 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SCHOEN,ELLIS  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: PADILLA MARTINEZ

DATE/TIME OF TEST: 06/24/2022 21:36:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:30 |
| Air Blank          | 0.000 | 21:30 |
| Control Test       | 0.082 | 21:31 |
| Air Blank          | 0.000 | 21:31 |
| Subject Sample # 1 | 0.219 | 21:32 |
| Breath Volume      | 1.218 | 21:32 |
| Air Blank          | 0.000 | 21:32 |
| Air Blank          | 0.000 | 21:34 |
| Subject Sample # 2 | 0.220 | 21:35 |
| Breath Volume      | 2.371 | 21:35 |
| Air Blank          | 0.000 | 21:35 |
| Control Test       | 0.081 | 21:36 |
| Air Blank          | 0.000 | 21:36 |
| Diagnostics Check  | OK    | 21:36 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SCHOEN,ELLIS  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 06/25/2022 01:12:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Control Test       | 0.081 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Subject Sample # 1 | 0.144 | 01:07 |
| Breath Volume      | 2.128 | 01:07 |
| Air Blank          | 0.000 | 01:08 |
| Air Blank          | 0.000 | 01:10 |
| Subject Sample # 2 | 0.148 | 01:10 |
| Breath Volume      | 1.996 | 01:10 |
| Air Blank          | 0.000 | 01:11 |
| Control Test       | 0.081 | 01:11 |
| Air Blank          | 0.000 | 01:12 |
| Diagnostics Check  | OK    | 01:12 |

---

Data Download Date/Time: 07/05/2022 14:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BRICKEY, ALEX  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: VALENTIN

DATE/TIME OF TEST: 06/25/2022 01:32:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:26 |
| Air Blank          | 0.000 | 01:27 |
| Control Test       | 0.082 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Subject Sample # 1 | 0.110 | 01:28 |
| Breath Volume      | 2.582 | 01:28 |
| Air Blank          | 0.000 | 01:28 |
| Air Blank          | 0.000 | 01:30 |
| Subject Sample # 2 | 0.118 | 01:31 |
| Breath Volume      | 1.902 | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.082 | 01:32 |
| Air Blank          | 0.000 | 01:32 |
| Diagnostics Check  | OK    | 01:32 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GARARD,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: FERRER RODRIGUEZ

DATE/TIME OF TEST: 06/25/2022 02:14:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:09 |
| Air Blank          | 0.000 | 02:09 |
| Control Test       | 0.082 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Subject Sample # 1 | 0.000 | 02:10 |
| Breath Volume      | 2.847 | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Air Blank          | 0.000 | 02:13 |
| Subject Sample # 2 | 0.000 | 02:13 |
| Breath Volume      | 2.914 | 02:13 |
| Air Blank          | 0.000 | 02:14 |
| Control Test       | 0.082 | 02:14 |
| Air Blank          | 0.000 | 02:14 |
| Diagnostics Check  | OK    | 02:14 |

---

Data Download Date/Time: 07/05/2022 14:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DANCEL,KAREN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: ESTELUS

DATE/TIME OF TEST: 06/25/2022 05:38:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:31 |
| Air Blank          | 0.000 | 05:31 |
| Control Test       | 0.083 | 05:32 |
| Air Blank          | 0.000 | 05:32 |
| Subject Sample # 1 | 0.119 | 05:33 |
| Breath Volume      | 1.339 | 05:33 |
| Air Blank          | 0.000 | 05:34 |
| Air Blank          | 0.000 | 05:35 |
| Subject Sample # 2 | 0.120 | 05:37 |
| Breath Volume      | 1.597 | 05:37 |
| Air Blank          | 0.000 | 05:37 |
| Control Test       | 0.081 | 05:38 |
| Air Blank          | 0.000 | 05:38 |
| Diagnostics Check  | OK    | 05:38 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KILLON,JONATHAN  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 06/26/2022 00:50:08

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 00:38 |                |
| Air Blank         | 0.000     | 00:38 |                |
| Control Test      | 0.082     | 00:39 |                |
| Air Blank         | 0.000     | 00:39 |                |
| Subject Sample    | # 1 0.091 | 00:42 | Volume Not Met |
| Breath Volume     | 0.597     | 00:42 |                |
| Air Blank         | 0.000     | 00:43 |                |
| Air Blank         | 0.000     | 00:45 |                |
| Subject Sample    | # 2 0.082 | 00:48 | Volume Not Met |
| Breath Volume     | 0.093     | 00:48 |                |
| Air Blank         | 0.000     | 00:49 |                |
| Control Test      | 0.081     | 00:49 |                |
| Air Blank         | 0.000     | 00:50 |                |
| Diagnostics Check | OK        | 00:50 |                |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ELLISON,LINDSEY  
ARRESTING OFFICER AGENCY: Winter Park PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/26/2022 01:43:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:37 |
| Air Blank          | 0.000 | 01:37 |
| Control Test       | 0.082 | 01:38 |
| Air Blank          | 0.000 | 01:38 |
| Subject Sample # 1 | 0.000 | 01:39 |
| Breath Volume      | 1.414 | 01:39 |
| Air Blank          | 0.000 | 01:39 |
| Air Blank          | 0.000 | 01:41 |
| Subject Sample # 2 | 0.000 | 01:42 |
| Breath Volume      | 1.402 | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.082 | 01:43 |
| Air Blank          | 0.000 | 01:43 |
| Diagnostics Check  | OK    | 01:43 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ESPAILLAT,LIANA  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 06/26/2022 07:58:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:51 |
| Air Blank          | 0.000 | 07:52 |
| Control Test       | 0.082 | 07:52 |
| Air Blank          | 0.000 | 07:53 |
| Subject Sample # 1 | 0.096 | 07:53 |
| Breath Volume      | 2.019 | 07:53 |
| Air Blank          | 0.000 | 07:54 |
| Air Blank          | 0.000 | 07:56 |
| Subject Sample # 2 | 0.095 | 07:56 |
| Breath Volume      | 2.613 | 07:56 |
| Air Blank          | 0.000 | 07:57 |
| Control Test       | 0.082 | 07:57 |
| Air Blank          | 0.000 | 07:58 |
| Diagnostics Check  | OK    | 07:58 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DEVINE,JASWANTIE  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 12:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: CLOUDEN

DATE/TIME OF TEST: 06/26/2022 13:53:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:47 |
| Air Blank          | 0.000 | 13:48 |
| Control Test       | 0.082 | 13:48 |
| Air Blank          | 0.000 | 13:48 |
| Subject Sample # 1 | 0.000 | 13:49 |
| Breath Volume      | 2.128 | 13:49 |
| Air Blank          | 0.000 | 13:49 |
| Air Blank          | 0.000 | 13:51 |
| Subject Sample # 2 | 0.000 | 13:52 |
| Breath Volume      | 1.226 | 13:52 |
| Air Blank          | 0.000 | 13:52 |
| Control Test       | 0.082 | 13:52 |
| Air Blank          | 0.000 | 13:53 |
| Diagnostics Check  | OK    | 13:53 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GOSS,AARON  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/27/2022 03:08:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Control Test       | 0.083 | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Subject Sample # 1 | 0.202 | 03:03 |
| Breath Volume      | 2.417 | 03:03 |
| Air Blank          | 0.000 | 03:04 |
| Air Blank          | 0.000 | 03:06 |
| Subject Sample # 2 | 0.196 | 03:06 |
| Breath Volume      | 1.578 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Control Test       | 0.082 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Diagnostics Check  | OK    | 03:08 |

---

Data Download Date/Time: 07/05/2022 14:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PETERSON,THOMAS  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 06/28/2022 01:23:07

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 01:11 |                |
| Air Blank         | 0.000     | 01:12 |                |
| Control Test      | 0.083     | 01:12 |                |
| Air Blank         | 0.000     | 01:12 |                |
| Subject Sample    | # 1 0.000 | 01:16 | Volume Not Met |
| Breath Volume     | 0.519     | 01:16 |                |
| Air Blank         | 0.000     | 01:16 |                |
| Air Blank         | 0.000     | 01:18 |                |
| Subject Sample    | # 2 0.000 | 01:21 | Volume Not Met |
| Breath Volume     | 1.003     | 01:21 |                |
| Air Blank         | 0.000     | 01:22 |                |
| Control Test      | 0.082     | 01:22 |                |
| Air Blank         | 0.000     | 01:23 |                |
| Diagnostics Check | OK        | 01:23 |                |

---

Data Download Date/Time: 07/05/2022 14:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SOSA, EMMANUEL  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: SANTIAGO ROSADO

DATE/TIME OF TEST: 06/28/2022 02:07:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Control Test       | 0.081 | 01:58 |
| Air Blank          | 0.000 | 01:59 |
| Subject Sample # 1 | 0.151 | 02:02 |
| Breath Volume      | 1.503 | 02:02 |
| Air Blank          | 0.000 | 02:02 |
| Air Blank          | 0.000 | 02:04 |
| Subject Sample # 2 | 0.146 | 02:06 |
| Breath Volume      | 1.570 | 02:06 |
| Air Blank          | 0.000 | 02:06 |
| Control Test       | 0.080 | 02:07 |
| Air Blank          | 0.000 | 02:07 |
| Diagnostics Check  | OK    | 02:07 |

---

Data Download Date/Time: 07/05/2022 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: IRELAND,DAVID  
ARRESTING OFFICER AGENCY: Edgewood PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 17:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/29/2022 18:57:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:51 |
| Air Blank          | 0.000 | 18:52 |
| Control Test       | 0.081 | 18:52 |
| Air Blank          | 0.000 | 18:52 |
| Subject Sample # 1 | 0.292 | 18:53 |
| Breath Volume      | 2.367 | 18:53 |
| Air Blank          | 0.000 | 18:53 |
| Air Blank          | 0.000 | 18:55 |
| Subject Sample # 2 | 0.274 | 18:56 |
| Breath Volume      | 1.515 | 18:56 |
| Air Blank          | 0.000 | 18:56 |
| Control Test       | 0.080 | 18:57 |
| Air Blank          | 0.000 | 18:57 |
| Diagnostics Check  | OK    | 18:57 |

---

Data Download Date/Time: 07/05/2022 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CURRAN,JASMINE  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 06/29/2022 23:09:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:03 |
| Air Blank          | 0.000 | 23:03 |
| Control Test       | 0.082 | 23:03 |
| Air Blank          | 0.000 | 23:04 |
| Subject Sample # 1 | 0.000 | 23:04 |
| Breath Volume      | 1.839 | 23:04 |
| Air Blank          | 0.000 | 23:05 |
| Air Blank          | 0.000 | 23:07 |
| Subject Sample # 2 | 0.000 | 23:07 |
| Breath Volume      | 2.257 | 23:07 |
| Air Blank          | 0.000 | 23:08 |
| Control Test       | 0.082 | 23:08 |
| Air Blank          | 0.000 | 23:08 |
| Diagnostics Check  | OK    | 23:09 |

---

Data Download Date/Time: 07/05/2022 14:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPENCER, JOHN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 06/01/2022 21:48:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:42 |
| Air Blank          | 0.000 | 21:43 |
| Control Test       | 0.080 | 21:43 |
| Air Blank          | 0.000 | 21:44 |
| Subject Sample # 1 | 0.000 | 21:44 |
| Breath Volume      | 1.464 | 21:44 |
| Air Blank          | 0.000 | 21:44 |
| Air Blank          | 0.000 | 21:46 |
| Subject Sample # 2 | 0.000 | 21:47 |
| Breath Volume      | 1.519 | 21:47 |
| Air Blank          | 0.000 | 21:47 |
| Control Test       | 0.080 | 21:48 |
| Air Blank          | 0.000 | 21:48 |
| Diagnostics Check  | OK    | 21:48 |

---

Data Download Date/Time: 06/13/2022 13:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RIVERA, HERIBERTO  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: ALGIN

DATE/TIME OF TEST: 06/02/2022 03:20:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:13 |
| Air Blank          | 0.000 | 03:13 |
| Control Test       | 0.080 | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Subject Sample # 1 | 0.154 | 03:15 |
| Breath Volume      | 1.664 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Subject Sample # 2 | 0.158 | 03:18 |
| Breath Volume      | 2.207 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Control Test       | 0.080 | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Diagnostics Check  | OK    | 03:20 |

---

Data Download Date/Time: 06/13/2022 13:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CAMPBELL,ROBERT  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: GUEHL

DATE/TIME OF TEST: 06/04/2022 01:32:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:26 |
| Air Blank          | 0.000 | 01:27 |
| Control Test       | 0.081 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Subject Sample # 1 | 0.000 | 01:28 |
| Breath Volume      | 1.476 | 01:28 |
| Air Blank          | 0.000 | 01:29 |
| Air Blank          | 0.000 | 01:31 |
| Subject Sample # 2 | 0.000 | 01:31 |
| Breath Volume      | 2.457 | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.080 | 01:32 |
| Air Blank          | 0.000 | 01:32 |
| Diagnostics Check  | OK    | 01:32 |

---

Data Download Date/Time: 06/13/2022 13:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SOSA, EMMANUEL  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/04/2022 04:28:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:22 |
| Air Blank          | 0.000 | 04:22 |
| Control Test       | 0.081 | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Subject Sample # 1 | 0.136 | 04:24 |
| Breath Volume      | 1.886 | 04:24 |
| Air Blank          | 0.000 | 04:24 |
| Air Blank          | 0.000 | 04:26 |
| Subject Sample # 2 | 0.139 | 04:27 |
| Breath Volume      | 2.265 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Control Test       | 0.081 | 04:28 |
| Air Blank          | 0.000 | 04:28 |
| Diagnostics Check  | OK    | 04:28 |

---

Data Download Date/Time: 06/13/2022 13:13



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ESPAILLAT,LIANA  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:59

SUBJECT'S LAST NAME: GUEHL

DATE/TIME OF TEST: 06/04/2022 12:30:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:24 |
| Air Blank          | 0.000 | 12:24 |
| Control Test       | 0.081 | 12:25 |
| Air Blank          | 0.000 | 12:25 |
| Subject Sample # 1 | 0.000 | 12:26 |
| Breath Volume      | 3.214 | 12:26 |
| Air Blank          | 0.000 | 12:26 |
| Air Blank          | 0.000 | 12:28 |
| Subject Sample # 2 | 0.000 | 12:28 |
| Breath Volume      | 3.425 | 12:28 |
| Air Blank          | 0.000 | 12:29 |
| Control Test       | 0.081 | 12:29 |
| Air Blank          | 0.000 | 12:30 |
| Diagnostics Check  | OK    | 12:30 |

---

Data Download Date/Time: 06/13/2022 13:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SOSA,EMMANUEL  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: CASTANEDRA IGLESIAS

DATE/TIME OF TEST: 06/04/2022 23:59:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:52 |
| Air Blank          | 0.000 | 23:53 |
| Control Test       | 0.080 | 23:53 |
| Air Blank          | 0.000 | 23:54 |
| Subject Sample # 1 | 0.094 | 23:54 |
| Breath Volume      | 2.410 | 23:54 |
| Air Blank          | 0.000 | 23:55 |
| Air Blank          | 0.000 | 23:57 |
| Subject Sample # 2 | 0.096 | 23:57 |
| Breath Volume      | 2.257 | 23:57 |
| Air Blank          | 0.000 | 23:58 |
| Control Test       | 0.079 | 23:58 |
| Air Blank          | 0.000 | 23:59 |
| Diagnostics Check  | OK    | 23:59 |

---

Data Download Date/Time: 06/13/2022 13:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CEDENO GARCIA,BRIAN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: FRANQUIZ OTERO

DATE/TIME OF TEST: 06/05/2022 04:27:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Control Test       | 0.080 | 04:21 |
| Air Blank          | 0.000 | 04:22 |
| Subject Sample # 1 | 0.129 | 04:22 |
| Breath Volume      | 4.191 | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Air Blank          | 0.000 | 04:25 |
| Subject Sample # 2 | 0.132 | 04:25 |
| Breath Volume      | 3.730 | 04:25 |
| Air Blank          | 0.000 | 04:26 |
| Control Test       | 0.079 | 04:26 |
| Air Blank          | 0.000 | 04:27 |
| Diagnostics Check  | OK    | 04:27 |

---

Data Download Date/Time: 06/13/2022 13:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ESPAILLAT,LIANA  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 06:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: MARAGH

DATE/TIME OF TEST: 06/05/2022 07:41:09

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 07:31 |                      |
| Air Blank         | 0.000 | 07:31 |                      |
| Control Test      | 0.080 | 07:32 |                      |
| Air Blank         | 0.000 | 07:32 |                      |
| Subject Sample    | # 1   | 07:35 | No Sample Provided   |
| Breath Volume     | 0.000 | 07:35 |                      |
| Air Blank         | 0.000 | 07:36 |                      |
| Air Blank         | 0.000 | 07:38 |                      |
| Subject Sample    | # 2   | 07:39 | Subject Test Refused |
| Breath Volume     | 0.000 | 07:39 |                      |
| Air Blank         | 0.000 | 07:40 |                      |
| Control Test      | 0.081 | 07:40 |                      |
| Air Blank         | 0.000 | 07:41 |                      |
| Diagnostics Check | OK    | 07:41 |                      |

---

Data Download Date/Time: 06/13/2022 13:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WOLFE, KENNETH  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 15:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 06/05/2022 16:42:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:36 |
| Air Blank          | 0.000 | 16:36 |
| Control Test       | 0.080 | 16:37 |
| Air Blank          | 0.000 | 16:37 |
| Subject Sample # 1 | 0.036 | 16:38 |
| Breath Volume      | 2.859 | 16:38 |
| Air Blank          | 0.000 | 16:38 |
| Air Blank          | 0.000 | 16:40 |
| Subject Sample # 2 | 0.035 | 16:41 |
| Breath Volume      | 2.898 | 16:41 |
| Air Blank          | 0.000 | 16:42 |
| Control Test       | 0.080 | 16:42 |
| Air Blank          | 0.000 | 16:42 |
| Diagnostics Check  | OK    | 16:42 |

---

Data Download Date/Time: 06/13/2022 13:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VINER, DENNIS  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GARCIA MEJIAS

DATE/TIME OF TEST: 06/08/2022 02:25:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:18 |
| Air Blank          | 0.000 | 02:19 |
| Control Test       | 0.081 | 02:19 |
| Air Blank          | 0.000 | 02:20 |
| Subject Sample # 1 | 0.083 | 02:20 |
| Breath Volume      | 2.015 | 02:20 |
| Air Blank          | 0.000 | 02:21 |
| Air Blank          | 0.000 | 02:23 |
| Subject Sample # 2 | 0.080 | 02:24 |
| Breath Volume      | 1.574 | 02:24 |
| Air Blank          | 0.000 | 02:24 |
| Control Test       | 0.080 | 02:25 |
| Air Blank          | 0.000 | 02:25 |
| Diagnostics Check  | OK    | 02:25 |

---

Data Download Date/Time: 06/13/2022 13:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KILLON, JONATHAN  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DEMREW

DATE/TIME OF TEST: 06/08/2022 03:41:49

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:25 |                |
| Air Blank         | 0.000     | 03:26 |                |
| Control Test      | 0.081     | 03:26 |                |
| Air Blank         | 0.000     | 03:27 |                |
| Subject Sample    | # 1 0.168 | 03:29 |                |
| Breath Volume     | 1.175     | 03:29 |                |
| Air Blank         | 0.000     | 03:30 |                |
| Air Blank         | 0.000     | 03:31 |                |
| Subject Sample    | # 2 0.144 | 03:35 | Volume Not Met |
| Breath Volume     | 0.738     | 03:35 |                |
| Air Blank         | 0.000     | 03:35 |                |
| Air Blank         | 0.000     | 03:37 |                |
| Subject Sample    | # 3 0.168 | 03:40 |                |
| Breath Volume     | 1.527     | 03:40 |                |
| Air Blank         | 0.000     | 03:40 |                |
| Control Test      | 0.077     | 03:41 |                |
| Air Blank         | 0.000     | 03:41 |                |
| Diagnostics Check | OK        | 03:41 |                |

---

Data Download Date/Time: 06/13/2022 13:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DIDOMENICO,GINO  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: WILLMAN

DATE/TIME OF TEST: 06/11/2022 03:16:12

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:13 |                      |
| Air Blank         | 0.000 | 03:13 |                      |
| Control Test      | 0.081 | 03:14 |                      |
| Air Blank         | 0.000 | 03:14 |                      |
| Subject Sample    | # 1   | 03:14 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:14 |                      |
| Air Blank         | 0.000 | 03:15 |                      |
| Control Test      | 0.079 | 03:15 |                      |
| Air Blank         | 0.000 | 03:16 |                      |
| Diagnostics Check | OK    | 03:16 |                      |

---

Data Download Date/Time: 06/13/2022 13:14



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SHIPLEY,DANIEL  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: OTENBERGER

DATE/TIME OF TEST: 06/12/2022 01:11:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Control Test       | 0.080 | 01:05 |
| Air Blank          | 0.000 | 01:06 |
| Subject Sample # 1 | 0.156 | 01:06 |
| Breath Volume      | 2.882 | 01:06 |
| Air Blank          | 0.000 | 01:07 |
| Air Blank          | 0.000 | 01:09 |
| Subject Sample # 2 | 0.155 | 01:09 |
| Breath Volume      | 2.164 | 01:09 |
| Air Blank          | 0.000 | 01:10 |
| Control Test       | 0.080 | 01:10 |
| Air Blank          | 0.000 | 01:11 |
| Diagnostics Check  | OK    | 01:11 |

---

Data Download Date/Time: 06/13/2022 13:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAYS,GARION  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MUNOZ ZUNIGA

DATE/TIME OF TEST: 06/12/2022 05:10:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:03 |
| Air Blank          | 0.000 | 05:03 |
| Control Test       | 0.080 | 05:04 |
| Air Blank          | 0.000 | 05:04 |
| Subject Sample # 1 | 0.137 | 05:05 |
| Breath Volume      | 1.496 | 05:05 |
| Air Blank          | 0.000 | 05:06 |
| Air Blank          | 0.000 | 05:08 |
| Subject Sample # 2 | 0.131 | 05:09 |
| Breath Volume      | 1.449 | 05:09 |
| Air Blank          | 0.000 | 05:09 |
| Control Test       | 0.080 | 05:10 |
| Air Blank          | 0.000 | 05:10 |
| Diagnostics Check  | OK    | 05:10 |

---

Data Download Date/Time: 06/13/2022 13:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VICTOR,NATHALIE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: PELAGIO PALACIOS

DATE/TIME OF TEST: 06/12/2022 06:07:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:00 |
| Air Blank          | 0.000 | 06:01 |
| Control Test       | 0.080 | 06:01 |
| Air Blank          | 0.000 | 06:02 |
| Subject Sample # 1 | 0.123 | 06:02 |
| Breath Volume      | 2.136 | 06:02 |
| Air Blank          | 0.000 | 06:03 |
| Air Blank          | 0.000 | 06:05 |
| Subject Sample # 2 | 0.130 | 06:05 |
| Breath Volume      | 1.304 | 06:05 |
| Air Blank          | 0.000 | 06:06 |
| Control Test       | 0.080 | 06:06 |
| Air Blank          | 0.000 | 06:07 |
| Diagnostics Check  | OK    | 06:07 |

---

Data Download Date/Time: 06/13/2022 13:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ESPINAL, BRIAN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: MEAURYCACERES

DATE/TIME OF TEST: 06/12/2022 08:20:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:11 |
| Air Blank          | 0.000 | 08:12 |
| Control Test       | 0.080 | 08:12 |
| Air Blank          | 0.000 | 08:12 |
| Subject Sample # 1 | 0.199 | 08:13 |
| Breath Volume      | 3.945 | 08:13 |
| Air Blank          | 0.000 | 08:14 |
| Air Blank          | 0.000 | 08:15 |
| Subject Sample # 2 | 0.160 | 08:16 |
| Breath Volume      | 1.171 | 08:16 |
| Air Blank          | 0.000 | 08:16 |
| Air Blank          | 0.000 | 08:18 |
| Subject Sample # 3 | 0.181 | 08:19 |
| Breath Volume      | 2.191 | 08:19 |
| Air Blank          | 0.000 | 08:19 |
| Control Test       | 0.080 | 08:20 |
| Air Blank          | 0.000 | 08:20 |
| Diagnostics Check  | OK    | 08:20 |

---

Data Download Date/Time: 06/13/2022 13:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DIAZ,JASON  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/12/2022 22:11:48

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:00 |                |
| Air Blank         | 0.000     | 22:00 |                |
| Control Test      | 0.080     | 22:00 |                |
| Air Blank         | 0.000     | 22:01 |                |
| Subject Sample    | # 1 0.096 | 22:04 | Volume Not Met |
| Breath Volume     | 0.179     | 22:04 |                |
| Air Blank         | 0.000     | 22:05 |                |
| Air Blank         | 0.000     | 22:07 |                |
| Subject Sample    | # 2 0.097 | 22:10 | Volume Not Met |
| Breath Volume     | 0.035     | 22:10 |                |
| Air Blank         | 0.000     | 22:10 |                |
| Control Test      | 0.080     | 22:11 |                |
| Air Blank         | 0.000     | 22:11 |                |
| Diagnostics Check | OK        | 22:11 |                |

---

Data Download Date/Time: 06/13/2022 13:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SCHOEN,ELLIS  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: GUILLENTRIGUERA

DATE/TIME OF TEST: 06/13/2022 00:33:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Control Test       | 0.080 | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Subject Sample # 1 | 0.166 | 00:28 |
| Breath Volume      | 1.683 | 00:28 |
| Air Blank          | 0.000 | 00:29 |
| Air Blank          | 0.000 | 00:31 |
| Subject Sample # 2 | 0.162 | 00:31 |
| Breath Volume      | 1.171 | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Control Test       | 0.079 | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Diagnostics Check  | OK    | 00:33 |

---

Data Download Date/Time: 06/13/2022 13:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CLABAUGH,ARIEL  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: BEASLEY

DATE/TIME OF TEST: 06/13/2022 02:12:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Control Test       | 0.081 | 02:06 |
| Air Blank          | 0.000 | 02:06 |
| Subject Sample # 1 | 0.280 | 02:08 |
| Breath Volume      | 2.199 | 02:08 |
| Air Blank          | 0.000 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Subject Sample # 2 | 0.283 | 02:11 |
| Breath Volume      | 1.441 | 02:11 |
| Air Blank          | 0.000 | 02:11 |
| Control Test       | 0.078 | 02:12 |
| Air Blank          | 0.000 | 02:12 |
| Diagnostics Check  | OK    | 02:12 |

---

Data Download Date/Time: 06/13/2022 13:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SCALES,BRANDON  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 06/13/2022 05:11:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:04 |
| Air Blank          | 0.000 | 05:05 |
| Control Test       | 0.080 | 05:05 |
| Air Blank          | 0.000 | 05:06 |
| Subject Sample # 1 | 0.213 | 05:06 |
| Breath Volume      | 1.406 | 05:06 |
| Air Blank          | 0.000 | 05:07 |
| Air Blank          | 0.000 | 05:09 |
| Subject Sample # 2 | 0.212 | 05:09 |
| Breath Volume      | 1.503 | 05:09 |
| Air Blank          | 0.000 | 05:10 |
| Control Test       | 0.080 | 05:10 |
| Air Blank          | 0.000 | 05:11 |
| Diagnostics Check  | OK    | 05:11 |

---

Data Download Date/Time: 06/13/2022 13:15



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RIDEAUX, ADAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A3  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: FUNES RAMOS

DATE/TIME OF TEST: 06/13/2022 05:53:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:47 |
| Air Blank          | 0.000 | 05:47 |
| Control Test       | 0.081 | 05:48 |
| Air Blank          | 0.000 | 05:48 |
| Subject Sample # 1 | 0.200 | 05:49 |
| Breath Volume      | 3.410 | 05:49 |
| Air Blank          | 0.000 | 05:49 |
| Air Blank          | 0.000 | 05:51 |
| Subject Sample # 2 | 0.201 | 05:52 |
| Breath Volume      | 2.871 | 05:52 |
| Air Blank          | 0.000 | 05:52 |
| Control Test       | 0.080 | 05:52 |
| Air Blank          | 0.000 | 05:53 |
| Diagnostics Check  | OK    | 05:53 |

---

Data Download Date/Time: 06/13/2022 13:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAN DYKE,T,M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 06/01/2022 01:16:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Control Test       | 0.079 | 01:10 |
| Air Blank          | 0.000 | 01:11 |
| Subject Sample # 1 | 0.142 | 01:11 |
| Breath Volume      | 3.886 | 01:11 |
| Air Blank          | 0.000 | 01:12 |
| Air Blank          | 0.000 | 01:14 |
| Subject Sample # 2 | 0.152 | 01:14 |
| Breath Volume      | 3.902 | 01:14 |
| Air Blank          | 0.000 | 01:15 |
| Control Test       | 0.077 | 01:15 |
| Air Blank          | 0.000 | 01:16 |
| Diagnostics Check  | OK    | 01:16 |

---

Data Download Date/Time: 06/03/2022 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLARIN,MISTY,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: MORANDO,B  
ARRESTING OFFICER AGENCY: Green Cove Springs  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 06/01/2022 01:54:15

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:45 |                      |
| Air Blank         | 0.000 | 01:46 |                      |
| Control Test      | 0.079 | 01:46 |                      |
| Air Blank         | 0.000 | 01:47 |                      |
| Subject Sample    | # 1   | 01:50 | No Sample Provided   |
| Breath Volume     | 0.000 | 01:50 |                      |
| Air Blank         | 0.000 | 01:50 |                      |
| Air Blank         | 0.000 | 01:52 |                      |
| Subject Sample    | # 2   | 01:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:52 |                      |
| Air Blank         | 0.000 | 01:53 |                      |
| Control Test      | 0.078 | 01:53 |                      |
| Air Blank         | 0.000 | 01:54 |                      |
| Diagnostics Check | OK    | 01:54 |                      |

---

Data Download Date/Time: 06/03/2022 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ADAMS,S  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: HENSE

DATE/TIME OF TEST: 06/01/2022 23:00:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:53 |
| Air Blank          | 0.000 | 22:54 |
| Control Test       | 0.079 | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Subject Sample # 1 | 0.123 | 22:55 |
| Breath Volume      | 1.792 | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Air Blank          | 0.000 | 22:58 |
| Subject Sample # 2 | 0.122 | 22:58 |
| Breath Volume      | 1.734 | 22:58 |
| Air Blank          | 0.000 | 22:59 |
| Control Test       | 0.079 | 22:59 |
| Air Blank          | 0.000 | 22:59 |
| Diagnostics Check  | OK    | 23:00 |

---

Data Download Date/Time: 06/03/2022 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCKENZIE,A,E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: MORANDO,B  
ARRESTING OFFICER AGENCY: Green Cove Springs  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 10:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/03/2022 23:25:50

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 23:20 |            |
| Air Blank         | 0.000 | 23:20 |            |
| Control Test      | 0.079 | 23:20 |            |
| Air Blank         | 0.000 | 23:21 |            |
| Subject Sample    | # 1   | 23:24 | RFI Detect |
| Breath Volume     | 0.000 | 23:24 |            |
| Air Blank         | 0.000 | 23:24 |            |
| Control Test      | 0.079 | 23:25 |            |
| Air Blank         | 0.000 | 23:25 |            |
| Diagnostics Check | OK    | 23:25 |            |

---

Data Download Date/Time: 06/27/2022 09:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCKENZIE,A,E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: MORANDO,B  
ARRESTING OFFICER AGENCY: Orange Park PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 10:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/03/2022 23:38:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:28 |                      |
| Air Blank         | 0.000 | 23:29 |                      |
| Control Test      | 0.078 | 23:29 |                      |
| Air Blank         | 0.000 | 23:29 |                      |
| Subject Sample    | # 1   | 23:33 | No Sample Provided   |
| Breath Volume     | 0.000 | 23:33 |                      |
| Air Blank         | 0.000 | 23:33 |                      |
| Air Blank         | 0.000 | 23:35 |                      |
| Subject Sample    | # 2   | 23:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:37 |                      |
| Air Blank         | 0.000 | 23:37 |                      |
| Control Test      | 0.078 | 23:38 |                      |
| Air Blank         | 0.000 | 23:38 |                      |
| Diagnostics Check | OK    | 23:38 |                      |

---

Data Download Date/Time: 06/27/2022 09:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLARIN,MISTY,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: HENDERSON,B  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/04/2022 01:29:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Control Test       | 0.079 | 01:23 |
| Air Blank          | 0.000 | 01:23 |
| Subject Sample # 1 | 0.188 | 01:24 |
| Breath Volume      | 3.085 | 01:24 |
| Air Blank          | 0.000 | 01:25 |
| Air Blank          | 0.000 | 01:27 |
| Subject Sample # 2 | 0.195 | 01:27 |
| Breath Volume      | 2.488 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.077 | 01:28 |
| Air Blank          | 0.000 | 01:29 |
| Diagnostics Check  | OK    | 01:29 |

---

Data Download Date/Time: 06/27/2022 09:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLARIN,MISTY,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: CRUZ,C  
ARRESTING OFFICER AGENCY: Orange Park PD  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 06/04/2022 20:29:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:23 |
| Air Blank          | 0.000 | 20:23 |
| Control Test       | 0.079 | 20:23 |
| Air Blank          | 0.000 | 20:24 |
| Subject Sample # 1 | 0.238 | 20:25 |
| Breath Volume      | 1.890 | 20:25 |
| Air Blank          | 0.000 | 20:25 |
| Air Blank          | 0.000 | 20:27 |
| Subject Sample # 2 | 0.233 | 20:28 |
| Breath Volume      | 2.343 | 20:28 |
| Air Blank          | 0.000 | 20:28 |
| Control Test       | 0.079 | 20:28 |
| Air Blank          | 0.000 | 20:29 |
| Diagnostics Check  | OK    | 20:29 |

---

Data Download Date/Time: 06/27/2022 09:44



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLARIN,MISTY,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: PARCHER,J  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HACHEZ

DATE/TIME OF TEST: 06/04/2022 23:32:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:23 |                      |
| Air Blank         | 0.000 | 23:24 |                      |
| Control Test      | 0.078 | 23:24 |                      |
| Air Blank         | 0.000 | 23:25 |                      |
| Subject Sample    | # 1   | 23:28 | No Sample Provided   |
| Breath Volume     | 0.000 | 23:28 |                      |
| Air Blank         | 0.000 | 23:28 |                      |
| Air Blank         | 0.000 | 23:30 |                      |
| Subject Sample    | # 2   | 23:30 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:30 |                      |
| Air Blank         | 0.000 | 23:31 |                      |
| Control Test      | 0.078 | 23:31 |                      |
| Air Blank         | 0.000 | 23:32 |                      |
| Diagnostics Check | OK    | 23:32 |                      |

---

Data Download Date/Time: 06/27/2022 09:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLARIN,MISY,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: HENDERSON,B  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: LAMMERS

DATE/TIME OF TEST: 06/05/2022 03:50:48

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:42 |                      |
| Air Blank         | 0.000 | 03:42 |                      |
| Control Test      | 0.078 | 03:43 |                      |
| Air Blank         | 0.000 | 03:43 |                      |
| Subject Sample    | # 1   | 03:46 | No Sample Provided   |
| Breath Volume     | 0.000 | 03:46 |                      |
| Air Blank         | 0.000 | 03:47 |                      |
| Air Blank         | 0.000 | 03:49 |                      |
| Subject Sample    | # 2   | 03:49 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:49 |                      |
| Air Blank         | 0.000 | 03:49 |                      |
| Control Test      | 0.079 | 03:50 |                      |
| Air Blank         | 0.000 | 03:50 |                      |
| Diagnostics Check | OK    | 03:50 |                      |

---

Data Download Date/Time: 06/27/2022 09:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLARIN,MISTY,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BYRNE,J  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: TYMONIEWICZ

DATE/TIME OF TEST: 06/05/2022 18:17:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:09 |
| Air Blank          | 0.000 | 18:10 |
| Control Test       | 0.079 | 18:10 |
| Air Blank          | 0.000 | 18:11 |
| Subject Sample # 1 | 0.071 | 18:12 |
| Breath Volume      | 1.253 | 18:12 |
| Air Blank          | 0.000 | 18:13 |
| Air Blank          | 0.000 | 18:14 |
| Subject Sample # 2 | 0.065 | 18:15 |
| Breath Volume      | 1.523 | 18:15 |
| Air Blank          | 0.000 | 18:16 |
| Control Test       | 0.077 | 18:16 |
| Air Blank          | 0.000 | 18:17 |
| Diagnostics Check  | OK    | 18:17 |

---

Data Download Date/Time: 06/27/2022 09:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCKENZIE,A,E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: RAWLES

DATE/TIME OF TEST: 06/09/2022 03:20:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:13 |
| Air Blank          | 0.000 | 03:13 |
| Control Test       | 0.079 | 03:14 |
| Air Blank          | 0.000 | 03:14 |
| Subject Sample # 1 | 0.081 | 03:15 |
| Breath Volume      | 1.375 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Subject Sample # 2 | 0.078 | 03:18 |
| Breath Volume      | 1.273 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Control Test       | 0.077 | 03:19 |
| Air Blank          | 0.000 | 03:20 |
| Diagnostics Check  | OK    | 03:20 |

---

Data Download Date/Time: 06/27/2022 09:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLBERT,J,T  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: CASPER,K  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: ESCHEN

DATE/TIME OF TEST: 06/10/2022 15:53:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 15:44 |                      |
| Air Blank         | 0.000 | 15:45 |                      |
| Control Test      | 0.079 | 15:45 |                      |
| Air Blank         | 0.000 | 15:45 |                      |
| Subject Sample    | # 1   | 15:49 | No Sample Provided   |
| Breath Volume     | 0.000 | 15:49 |                      |
| Air Blank         | 0.000 | 15:49 |                      |
| Air Blank         | 0.000 | 15:51 |                      |
| Subject Sample    | # 2   | 15:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 15:51 |                      |
| Air Blank         | 0.000 | 15:52 |                      |
| Control Test      | 0.079 | 15:52 |                      |
| Air Blank         | 0.000 | 15:53 |                      |
| Diagnostics Check | OK    | 15:53 |                      |

---

Data Download Date/Time: 06/27/2022 09:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAERTZ, MARK, J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ADAMS, S  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 06/11/2022 01:42:00

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:33 |                      |
| Air Blank         | 0.000 | 01:34 |                      |
| Control Test      | 0.079 | 01:34 |                      |
| Air Blank         | 0.000 | 01:34 |                      |
| Subject Sample    | # 1   | 01:37 | No Sample Provided   |
| Breath Volume     | 0.000 | 01:37 |                      |
| Air Blank         | 0.000 | 01:38 |                      |
| Air Blank         | 0.000 | 01:40 |                      |
| Subject Sample    | # 2   | 01:40 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:40 |                      |
| Air Blank         | 0.000 | 01:41 |                      |
| Control Test      | 0.078 | 01:41 |                      |
| Air Blank         | 0.000 | 01:41 |                      |
| Diagnostics Check | OK    | 01:42 |                      |

Data Download Date/Time: 06/27/2022 09:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BAUMGARTNER,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: LUYK

DATE/TIME OF TEST: 06/12/2022 18:35:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:29 |
| Air Blank          | 0.000 | 18:29 |
| Control Test       | 0.080 | 18:29 |
| Air Blank          | 0.000 | 18:30 |
| Subject Sample # 1 | 0.256 | 18:30 |
| Breath Volume      | 3.406 | 18:30 |
| Air Blank          | 0.000 | 18:31 |
| Air Blank          | 0.000 | 18:33 |
| Subject Sample # 2 | 0.253 | 18:33 |
| Breath Volume      | 2.191 | 18:33 |
| Air Blank          | 0.000 | 18:34 |
| Control Test       | 0.079 | 18:34 |
| Air Blank          | 0.000 | 18:35 |
| Diagnostics Check  | OK    | 18:35 |

---

Data Download Date/Time: 06/27/2022 09:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIDDERS,R,M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GELSOMINO,M  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: CAICEDO

DATE/TIME OF TEST: 06/17/2022 07:00:02

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 06:51 |                      |
| Air Blank         | 0.000 | 06:51 |                      |
| Control Test      | 0.079 | 06:52 |                      |
| Air Blank         | 0.000 | 06:52 |                      |
| Subject Sample    | # 1   | 06:55 | No Sample Provided   |
| Breath Volume     | 0.000 | 06:55 |                      |
| Air Blank         | 0.000 | 06:56 |                      |
| Air Blank         | 0.000 | 06:58 |                      |
| Subject Sample    | # 2   | 06:58 | Subject Test Refused |
| Breath Volume     | 0.000 | 06:58 |                      |
| Air Blank         | 0.000 | 06:59 |                      |
| Control Test      | 0.078 | 06:59 |                      |
| Air Blank         | 0.000 | 06:59 |                      |
| Diagnostics Check | OK    | 07:00 |                      |

Data Download Date/Time: 06/27/2022 09:44



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAN DYKE,T,M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: BRESSLER

DATE/TIME OF TEST: 06/18/2022 03:12:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:05 |
| Air Blank          | 0.000 | 03:05 |
| Control Test       | 0.080 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Subject Sample # 1 | 0.192 | 03:07 |
| Breath Volume      | 2.617 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Air Blank          | 0.000 | 03:10 |
| Subject Sample # 2 | 0.202 | 03:11 |
| Breath Volume      | 2.171 | 03:11 |
| Air Blank          | 0.000 | 03:11 |
| Control Test       | 0.079 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Diagnostics Check  | OK    | 03:12 |

---

Data Download Date/Time: 06/27/2022 09:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAN DYKE,T,M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: FLYNN,H  
ARRESTING OFFICER AGENCY: Green Cove Springs  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/18/2022 20:55:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:45 |
| Air Blank          | 0.000 | 20:45 |
| Control Test       | 0.079 | 20:46 |
| Air Blank          | 0.000 | 20:46 |
| Subject Sample # 1 | 0.139 | 20:50 |
| Breath Volume      | 1.195 | 20:50 |
| Air Blank          | 0.000 | 20:50 |
| Air Blank          | 0.000 | 20:52 |
| Subject Sample # 2 | 0.136 | 20:53 |
| Breath Volume      | 1.375 | 20:53 |
| Air Blank          | 0.000 | 20:54 |
| Control Test       | 0.078 | 20:54 |
| Air Blank          | 0.000 | 20:54 |
| Diagnostics Check  | OK    | 20:55 |

---

Data Download Date/Time: 06/27/2022 09:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCKENZIE,A,E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: MCKEE

DATE/TIME OF TEST: 06/19/2022 03:24:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Control Test       | 0.078 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Subject Sample # 1 | 0.187 | 03:19 |
| Breath Volume      | 2.675 | 03:19 |
| Air Blank          | 0.000 | 03:20 |
| Air Blank          | 0.000 | 03:22 |
| Subject Sample # 2 | 0.184 | 03:22 |
| Breath Volume      | 2.910 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Control Test       | 0.078 | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Diagnostics Check  | OK    | 03:24 |

---

Data Download Date/Time: 06/27/2022 09:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCKENZIE,A,E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/30/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: DARREN

DATE/TIME OF TEST: 06/19/2022 18:24:30

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:15 |                      |
| Air Blank         | 0.000 | 18:16 |                      |
| Control Test      | 0.078 | 18:16 |                      |
| Air Blank         | 0.000 | 18:17 |                      |
| Subject Sample    | # 1   | 18:20 | No Sample Provided   |
| Breath Volume     | 0.000 | 18:20 |                      |
| Air Blank         | 0.000 | 18:20 |                      |
| Air Blank         | 0.000 | 18:22 |                      |
| Subject Sample    | # 2   | 18:23 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:23 |                      |
| Air Blank         | 0.000 | 18:23 |                      |
| Control Test      | 0.079 | 18:23 |                      |
| Air Blank         | 0.000 | 18:24 |                      |
| Diagnostics Check | OK    | 18:24 |                      |

---

Data Download Date/Time: 06/27/2022 09:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ADAMS,S  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: LANE JR

DATE/TIME OF TEST: 06/24/2022 23:10:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:04 |
| Air Blank          | 0.000 | 23:04 |
| Control Test       | 0.079 | 23:05 |
| Air Blank          | 0.000 | 23:05 |
| Subject Sample # 1 | 0.129 | 23:06 |
| Breath Volume      | 1.644 | 23:06 |
| Air Blank          | 0.000 | 23:06 |
| Air Blank          | 0.000 | 23:08 |
| Subject Sample # 2 | 0.120 | 23:09 |
| Breath Volume      | 1.519 | 23:09 |
| Air Blank          | 0.000 | 23:09 |
| Control Test       | 0.078 | 23:10 |
| Air Blank          | 0.000 | 23:10 |
| Diagnostics Check  | OK    | 23:10 |

---

Data Download Date/Time: 06/27/2022 09:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ADAMS,S  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/25/2022 03:23:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:17 |
| Air Blank          | 0.000 | 03:17 |
| Control Test       | 0.079 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Subject Sample # 1 | 0.198 | 03:19 |
| Breath Volume      | 2.066 | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 2 | 0.184 | 03:22 |
| Breath Volume      | 1.714 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Control Test       | 0.078 | 03:23 |
| Air Blank          | 0.000 | 03:23 |
| Diagnostics Check  | OK    | 03:23 |

---

Data Download Date/Time: 06/27/2022 09:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ADAMS,S  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: POLAND

DATE/TIME OF TEST: 06/25/2022 20:49:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:43 |
| Air Blank          | 0.000 | 20:43 |
| Control Test       | 0.078 | 20:43 |
| Air Blank          | 0.000 | 20:44 |
| Subject Sample # 1 | 0.000 | 20:45 |
| Breath Volume      | 2.265 | 20:45 |
| Air Blank          | 0.000 | 20:45 |
| Air Blank          | 0.000 | 20:47 |
| Subject Sample # 2 | 0.000 | 20:48 |
| Breath Volume      | 1.742 | 20:48 |
| Air Blank          | 0.000 | 20:48 |
| Control Test       | 0.078 | 20:49 |
| Air Blank          | 0.000 | 20:49 |
| Diagnostics Check  | OK    | 20:49 |

---

Data Download Date/Time: 06/27/2022 09:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ADAMS,S  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: HAGELE

DATE/TIME OF TEST: 06/27/2022 00:54:01

---

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 00:41 |                    |
| Air Blank         | 0.000     | 00:41 |                    |
| Control Test      | 0.079     | 00:41 |                    |
| Air Blank         | 0.000     | 00:42 |                    |
| Subject Sample    | # 1 0.124 | 00:43 |                    |
| Breath Volume     | 2.546     | 00:43 |                    |
| Air Blank         | 0.000     | 00:44 |                    |
| Air Blank         | 0.000     | 00:46 |                    |
| Subject Sample    | # 2       | 00:49 | No Sample Provided |
| Breath Volume     | 0.000     | 00:49 |                    |
| Air Blank         | 0.000     | 00:50 |                    |
| Air Blank         | 0.000     | 00:52 |                    |
| Subject Sample    | # 3 0.124 | 00:52 |                    |
| Breath Volume     | 2.312     | 00:52 |                    |
| Air Blank         | 0.000     | 00:53 |                    |
| Control Test      | 0.078     | 00:53 |                    |
| Air Blank         | 0.000     | 00:53 |                    |
| Diagnostics Check | OK        | 00:54 |                    |

---

Data Download Date/Time: 06/27/2022 09:45



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MARTINI,RYAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/09/2022 21:01:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:55 |
| Air Blank          | 0.000 | 20:56 |
| Control Test       | 0.081 | 20:56 |
| Air Blank          | 0.000 | 20:57 |
| Subject Sample # 1 | 0.000 | 20:57 |
| Breath Volume      | 2.132 | 20:57 |
| Air Blank          | 0.000 | 20:57 |
| Air Blank          | 0.000 | 20:59 |
| Subject Sample # 2 | 0.000 | 21:00 |
| Breath Volume      | 2.027 | 21:00 |
| Air Blank          | 0.000 | 21:00 |
| Control Test       | 0.080 | 21:01 |
| Air Blank          | 0.000 | 21:01 |
| Diagnostics Check  | OK    | 21:01 |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOTSON, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: DOTSON, WILLIAM  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: MILTHALER

DATE/TIME OF TEST: 06/09/2022 21:14:43

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 21:03 |                |
| Air Blank         | 0.000     | 21:03 |                |
| Control Test      | 0.080     | 21:03 |                |
| Air Blank         | 0.000     | 21:04 |                |
| Subject Sample    | # 1 0.072 | 21:04 |                |
| Breath Volume     | 1.164     | 21:04 |                |
| Air Blank         | 0.000     | 21:05 |                |
| Air Blank         | 0.000     | 21:07 |                |
| Subject Sample    | # 2 0.083 | 21:10 | Volume Not Met |
| Breath Volume     | 0.148     | 21:10 |                |
| Air Blank         | 0.000     | 21:10 |                |
| Air Blank         | 0.000     | 21:12 |                |
| Subject Sample    | # 3 0.079 | 21:13 |                |
| Breath Volume     | 1.289     | 21:13 |                |
| Air Blank         | 0.000     | 21:13 |                |
| Control Test      | 0.078     | 21:14 |                |
| Air Blank         | 0.000     | 21:14 |                |
| Diagnostics Check | OK        | 21:14 |                |

---

Data Download Date/Time: 07/05/2022 17:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MARTINI,RYAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 06/10/2022 02:12:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Control Test       | 0.082 | 02:07 |
| Air Blank          | 0.000 | 02:07 |
| Subject Sample # 1 | 0.084 | 02:08 |
| Breath Volume      | 1.820 | 02:08 |
| Air Blank          | 0.000 | 02:08 |
| Air Blank          | 0.000 | 02:10 |
| Subject Sample # 2 | 0.083 | 02:11 |
| Breath Volume      | 1.562 | 02:11 |
| Air Blank          | 0.000 | 02:11 |
| Control Test       | 0.081 | 02:12 |
| Air Blank          | 0.000 | 02:12 |
| Diagnostics Check  | OK    | 02:12 |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: VULICEVIC,RADOSAV  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/10/2022 22:55:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:48 |
| Air Blank          | 0.000 | 22:48 |
| Control Test       | 0.081 | 22:48 |
| Air Blank          | 0.000 | 22:49 |
| Subject Sample # 1 | 0.304 | 22:50 |
| Breath Volume      | 1.312 | 22:50 |
| Air Blank          | 0.000 | 22:50 |
| Air Blank          | 0.000 | 22:52 |
| Subject Sample # 2 | 0.311 | 22:53 |
| Breath Volume      | 2.183 | 22:53 |
| Air Blank          | 0.000 | 22:54 |
| Control Test       | 0.080 | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Diagnostics Check  | OK    | 22:55 |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,SAMANTHA,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MARTINS,SAMANTHA  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: MARQUES

DATE/TIME OF TEST: 06/12/2022 02:29:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:22 |
| Air Blank          | 0.000 | 02:22 |
| Control Test       | 0.081 | 02:23 |
| Air Blank          | 0.000 | 02:23 |
| Subject Sample # 1 | 0.207 | 02:24 |
| Breath Volume      | 1.628 | 02:24 |
| Air Blank          | 0.000 | 02:25 |
| Air Blank          | 0.000 | 02:27 |
| Subject Sample # 2 | 0.210 | 02:28 |
| Breath Volume      | 2.066 | 02:28 |
| Air Blank          | 0.000 | 02:28 |
| Control Test       | 0.078 | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Diagnostics Check  | OK    | 02:29 |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,SAMANTHA,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MARTIN,PAUL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/12/2022 23:51:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:45 |
| Air Blank          | 0.000 | 23:45 |
| Control Test       | 0.081 | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Subject Sample # 1 | 0.188 | 23:47 |
| Breath Volume      | 2.683 | 23:47 |
| Air Blank          | 0.000 | 23:47 |
| Air Blank          | 0.000 | 23:49 |
| Subject Sample # 2 | 0.190 | 23:49 |
| Breath Volume      | 2.734 | 23:49 |
| Air Blank          | 0.000 | 23:50 |
| Control Test       | 0.080 | 23:50 |
| Air Blank          | 0.000 | 23:51 |
| Diagnostics Check  | OK    | 23:51 |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANNOVA,CASSANDRA,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: GOMEZ,JESSE  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 09:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/15/2022 10:49:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:43 |
| Air Blank          | 0.000 | 10:43 |
| Control Test       | 0.082 | 10:43 |
| Air Blank          | 0.000 | 10:44 |
| Subject Sample # 1 | 0.000 | 10:44 |
| Breath Volume      | 1.734 | 10:44 |
| Air Blank          | 0.000 | 10:45 |
| Air Blank          | 0.000 | 10:47 |
| Subject Sample # 2 | 0.000 | 10:47 |
| Breath Volume      | 1.714 | 10:47 |
| Air Blank          | 0.000 | 10:48 |
| Control Test       | 0.082 | 10:48 |
| Air Blank          | 0.000 | 10:48 |
| Diagnostics Check  | OK    | 10:49 |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREENUP, RONNY,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: JOHNSON, KWESI  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: ROMAN PEREZ

DATE/TIME OF TEST: 06/18/2022 01:03:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Control Test       | 0.082 | 00:57 |
| Air Blank          | 0.000 | 00:58 |
| Subject Sample # 1 | 0.069 | 00:59 |
| Breath Volume      | 2.656 | 00:59 |
| Air Blank          | 0.000 | 00:59 |
| Air Blank          | 0.000 | 01:01 |
| Subject Sample # 2 | 0.066 | 01:02 |
| Breath Volume      | 2.359 | 01:02 |
| Air Blank          | 0.000 | 01:02 |
| Control Test       | 0.081 | 01:02 |
| Air Blank          | 0.000 | 01:03 |
| Diagnostics Check  | OK    | 01:03 |

---

Data Download Date/Time: 07/05/2022 17:05



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWTON,JENNA,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: NEWTON,JENNA  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CUEVAS

DATE/TIME OF TEST: 06/19/2022 03:47:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:39 |
| Air Blank          | 0.000 | 03:39 |
| Control Test       | 0.081 | 03:40 |
| Air Blank          | 0.000 | 03:40 |
| Subject Sample # 1 | 0.172 | 03:42 |
| Breath Volume      | 1.316 | 03:42 |
| Air Blank          | 0.000 | 03:42 |
| Air Blank          | 0.000 | 03:44 |
| Subject Sample # 2 | 0.173 | 03:46 |
| Breath Volume      | 2.484 | 03:46 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.079 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Diagnostics Check  | OK    | 03:47 |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLWELL,CHASE,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: COLWELL,CHASE  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 06/21/2022 01:17:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:11 |
| Air Blank          | 0.000 | 01:11 |
| Control Test       | 0.082 | 01:12 |
| Air Blank          | 0.000 | 01:12 |
| Subject Sample # 1 | 0.000 | 01:13 |
| Breath Volume      | 4.644 | 01:13 |
| Air Blank          | 0.000 | 01:13 |
| Air Blank          | 0.000 | 01:15 |
| Subject Sample # 2 | 0.000 | 01:16 |
| Breath Volume      | 4.507 | 01:16 |
| Air Blank          | 0.000 | 01:16 |
| Control Test       | 0.081 | 01:17 |
| Air Blank          | 0.000 | 01:17 |
| Diagnostics Check  | OK    | 01:17 |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOTSON, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: DOTSON, WILLIAM  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ROSSI

DATE/TIME OF TEST: 06/22/2022 22:32:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:25 |
| Air Blank          | 0.000 | 22:26 |
| Control Test       | 0.082 | 22:26 |
| Air Blank          | 0.000 | 22:26 |
| Subject Sample # 1 | 0.266 | 22:28 |
| Breath Volume      | 3.007 | 22:28 |
| Air Blank          | 0.000 | 22:29 |
| Air Blank          | 0.000 | 22:30 |
| Subject Sample # 2 | 0.252 | 22:31 |
| Breath Volume      | 2.214 | 22:31 |
| Air Blank          | 0.000 | 22:31 |
| Control Test       | 0.079 | 22:32 |
| Air Blank          | 0.000 | 22:32 |
| Diagnostics Check  | OK    | 22:32 |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: BELANGER

DATE/TIME OF TEST: 06/24/2022 22:42:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Control Test       | 0.082 | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Subject Sample # 1 | 0.302 | 22:37 |
| Breath Volume      | 3.320 | 22:37 |
| Air Blank          | 0.000 | 22:38 |
| Air Blank          | 0.000 | 22:40 |
| Subject Sample # 2 | 0.294 | 22:40 |
| Breath Volume      | 1.804 | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Control Test       | 0.081 | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Diagnostics Check  | OK    | 22:42 |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 06/25/2022 03:00:39

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 02:59 |            |
| Air Blank         |       | 03:00 | RFI Detect |
| Air Blank         | 0.000 | 03:00 |            |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 06/25/2022 03:06:30

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:01 |                      |
| Air Blank         | 0.000 | 03:02 |                      |
| Control Test      | 0.082 | 03:02 |                      |
| Air Blank         | 0.000 | 03:03 |                      |
| Subject Sample    | # 1   | 03:05 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:05 |                      |
| Air Blank         | 0.000 | 03:05 |                      |
| Control Test      | 0.082 | 03:05 |                      |
| Air Blank         | 0.000 | 03:06 |                      |
| Diagnostics Check | OK    | 03:06 |                      |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOWE,ALEXANDER  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 06/28/2022 02:12:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Control Test       | 0.081 | 02:06 |
| Air Blank          | 0.000 | 02:06 |
| Subject Sample # 1 | 0.015 | 02:07 |
| Breath Volume      | 1.500 | 02:07 |
| Air Blank          | 0.000 | 02:08 |
| Air Blank          | 0.000 | 02:10 |
| Subject Sample # 2 | 0.013 | 02:11 |
| Breath Volume      | 1.160 | 02:11 |
| Air Blank          | 0.000 | 02:11 |
| Control Test       | 0.080 | 02:11 |
| Air Blank          | 0.000 | 02:12 |
| Diagnostics Check  | OK    | 02:12 |

---

Data Download Date/Time: 07/05/2022 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: DURHAM,KARI  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 14:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: ALBANITO

DATE/TIME OF TEST: 06/28/2022 15:17:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:11 |
| Air Blank          | 0.000 | 15:11 |
| Control Test       | 0.081 | 15:11 |
| Air Blank          | 0.000 | 15:12 |
| Subject Sample # 1 | 0.136 | 15:12 |
| Breath Volume      | 2.582 | 15:12 |
| Air Blank          | 0.000 | 15:13 |
| Air Blank          | 0.000 | 15:15 |
| Subject Sample # 2 | 0.133 | 15:15 |
| Breath Volume      | 2.445 | 15:15 |
| Air Blank          | 0.000 | 15:16 |
| Control Test       | 0.081 | 15:16 |
| Air Blank          | 0.000 | 15:17 |
| Diagnostics Check  | OK    | 15:17 |

---

Data Download Date/Time: 07/05/2022 17:05



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BISHOP,TREVOR,C  
BREATH TEST OPERATOR AGENCY: Putnam County SO  
ARRESTING OFFICER: PLEIMAN,ERIC  
ARRESTING OFFICER AGENCY: Palatka PD  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28021080A2  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 06/04/2022 04:17:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Control Test       | 0.080 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Subject Sample # 1 | 0.183 | 04:12 |
| Breath Volume      | 2.234 | 04:12 |
| Air Blank          | 0.000 | 04:13 |
| Air Blank          | 0.000 | 04:15 |
| Subject Sample # 2 | 0.172 | 04:15 |
| Breath Volume      | 1.699 | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Control Test       | 0.078 | 04:16 |
| Air Blank          | 0.000 | 04:17 |
| Diagnostics Check  | OK    | 04:17 |

---

Data Download Date/Time: 06/30/2022 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPELA,JORDAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: CAPELA,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28021080A2  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: SONCRANT

DATE/TIME OF TEST: 06/10/2022 21:23:41

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 21:14 |                      |
| Air Blank         | 0.000     | 21:15 |                      |
| Control Test      | 0.079     | 21:15 |                      |
| Air Blank         | 0.000     | 21:16 |                      |
| Subject Sample    | # 1 0.068 | 21:16 |                      |
| Breath Volume     | 2.125     | 21:16 |                      |
| Air Blank         | 0.000     | 21:17 |                      |
| Air Blank         | 0.000     | 21:19 |                      |
| Subject Sample    | # 2       | 21:22 | Subject Test Refused |
| Breath Volume     | 0.000     | 21:22 |                      |
| Air Blank         | 0.000     | 21:22 |                      |
| Control Test      | 0.079     | 21:23 |                      |
| Air Blank         | 0.000     | 21:23 |                      |
| Diagnostics Check | OK        | 21:23 |                      |

---

Data Download Date/Time: 06/30/2022 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,RICHARD,S  
BREATH TEST OPERATOR AGENCY: Putnam County SO  
ARRESTING OFFICER: TISDALE,RICHARD  
ARRESTING OFFICER AGENCY: Putnam County SO  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28021080A2  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MEZA GUERRERO

DATE/TIME OF TEST: 06/11/2022 05:14:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:11 |                      |
| Air Blank         | 0.000 | 05:12 |                      |
| Control Test      | 0.079 | 05:12 |                      |
| Air Blank         | 0.000 | 05:12 |                      |
| Subject Sample    | # 1   | 05:13 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:13 |                      |
| Air Blank         | 0.000 | 05:13 |                      |
| Control Test      | 0.079 | 05:13 |                      |
| Air Blank         | 0.000 | 05:14 |                      |
| Diagnostics Check | OK    | 05:14 |                      |

---

Data Download Date/Time: 06/30/2022 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAUENHOFER,TIMOTHY,H  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FERGUSON,SHANNON  
ARRESTING OFFICER AGENCY: Putnam County SO  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28021080A2  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/20/2022 03:31:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Control Test       | 0.079 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Subject Sample # 1 | 0.111 | 03:27 |
| Breath Volume      | 2.132 | 03:27 |
| Air Blank          | 0.000 | 03:27 |
| Air Blank          | 0.000 | 03:29 |
| Subject Sample # 2 | 0.112 | 03:30 |
| Breath Volume      | 2.644 | 03:30 |
| Air Blank          | 0.000 | 03:30 |
| Control Test       | 0.078 | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Diagnostics Check  | OK    | 03:31 |

---

Data Download Date/Time: 06/30/2022 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPELA,JORDAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: CAPELA,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/20/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28021080A2  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BOCANEGRA MALAGON

DATE/TIME OF TEST: 06/20/2022 22:18:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:12 |
| Air Blank          | 0.000 | 22:12 |
| Control Test       | 0.079 | 22:12 |
| Air Blank          | 0.000 | 22:13 |
| Subject Sample # 1 | 0.088 | 22:13 |
| Breath Volume      | 1.304 | 22:13 |
| Air Blank          | 0.000 | 22:14 |
| Air Blank          | 0.000 | 22:16 |
| Subject Sample # 2 | 0.097 | 22:16 |
| Breath Volume      | 2.390 | 22:16 |
| Air Blank          | 0.000 | 22:17 |
| Control Test       | 0.079 | 22:17 |
| Air Blank          | 0.000 | 22:18 |
| Diagnostics Check  | OK    | 22:18 |

---

Data Download Date/Time: 06/30/2022 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: JONES,DAKOTA  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 06/05/2022 16:31:52

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 16:26 |                |
| Air Blank         | 0.000     | 16:26 |                |
| Control Test      | 0.082     | 16:26 |                |
| Air Blank         | 0.000     | 16:27 |                |
| Subject Sample    | # 1 0.243 | 16:30 | Volume Not Met |
| Breath Volume     | 0.335     | 16:30 |                |
| Air Blank         |           | 16:31 | Purge Fail     |
| Air Blank         | 0.000     | 16:31 |                |

---

Data Download Date/Time: 06/20/2022 18:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: JONES,DAKOTA  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 05/29/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 06/05/2022 16:42:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:34 |
| Air Blank          | 0.000 | 16:34 |
| Control Test       | 0.076 | 16:35 |
| Air Blank          | 0.000 | 16:35 |
| Subject Sample # 1 | 0.262 | 16:36 |
| Breath Volume      | 2.121 | 16:36 |
| Air Blank          | 0.000 | 16:37 |
| Air Blank          | 0.000 | 16:39 |
| Subject Sample # 2 | 0.255 | 16:40 |
| Breath Volume      | 1.375 | 16:40 |
| Air Blank          | 0.000 | 16:41 |
| Control Test       | 0.078 | 16:41 |
| Air Blank          | 0.000 | 16:42 |
| Diagnostics Check  | OK    | 16:42 |

---

Data Download Date/Time: 06/20/2022 18:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: NUSCIS,NICHOLAS  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 06/25/2022 03:54:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Control Test       | 0.082 | 03:48 |
| Air Blank          | 0.000 | 03:49 |
| Subject Sample # 1 | 0.145 | 03:49 |
| Breath Volume      | 1.316 | 03:49 |
| Air Blank          | 0.000 | 03:50 |
| Air Blank          | 0.000 | 03:52 |
| Subject Sample # 2 | 0.147 | 03:52 |
| Breath Volume      | 2.324 | 03:52 |
| Air Blank          | 0.000 | 03:53 |
| Control Test       | 0.082 | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Diagnostics Check  | OK    | 03:54 |

---

Data Download Date/Time: 07/12/2022 10:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,BRIAN,G  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: GARCIA,CORINA  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PIMENTEL RIVERA

DATE/TIME OF TEST: 06/27/2022 01:03:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Control Test       | 0.080 | 00:57 |
| Air Blank          | 0.000 | 00:58 |
| Subject Sample # 1 | 0.174 | 00:59 |
| Breath Volume      | 3.160 | 00:59 |
| Air Blank          | 0.000 | 00:59 |
| Air Blank          | 0.000 | 01:01 |
| Subject Sample # 2 | 0.176 | 01:02 |
| Breath Volume      | 3.574 | 01:02 |
| Air Blank          | 0.000 | 01:02 |
| Control Test       | 0.080 | 01:03 |
| Air Blank          | 0.000 | 01:03 |
| Diagnostics Check  | OK    | 01:03 |

---

Data Download Date/Time: 07/12/2022 10:58