

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DORTON

DATE/TIME OF TEST: 07/18/2022 00:09:53

Diagnostics Check	OK	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.000	00:05
Breath Volume	1.941	00:05
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.000	00:08
Breath Volume	1.921	00:08
Air Blank	0.000	00:08
Control Test	0.079	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 08/30/2022 21:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: NORWICK

DATE/TIME OF TEST: 07/29/2022 17:33:56

Diagnostics Check	OK	17:32	
Air Blank	0.000	17:33	
Control Test	0.074	17:33	RFI Detect
Air Blank	0.000	17:33	

Data Download Date/Time: 08/30/2022 21:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: NOWRICK

DATE/TIME OF TEST: 07/29/2022 17:42:43

Diagnostics Check	OK	17:36
Air Blank	0.000	17:37
Control Test	0.078	17:37
Air Blank	0.000	17:37
Subject Sample # 1	0.000	17:38
Breath Volume	1.914	17:38
Air Blank	0.000	17:38
Air Blank	0.000	17:40
Subject Sample # 2	0.000	17:41
Breath Volume	1.812	17:41
Air Blank	0.000	17:41
Control Test	0.079	17:42
Air Blank	0.000	17:42
Diagnostics Check	OK	17:42

Data Download Date/Time: 08/30/2022 21:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 07/29/2022 18:59:05

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.078	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.009	18:54
Breath Volume	2.621	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.007	18:57
Breath Volume	2.625	18:57
Air Blank	0.000	18:58
Control Test	0.079	18:58
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 08/30/2022 21:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: WHITTAKER

DATE/TIME OF TEST: 07/01/2022 00:40:07

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.078	00:35	
Air Blank	0.000	00:35	
Subject Sample # 1	0.158	00:36	
Breath Volume	2.226	00:36	
Air Blank	0.000	00:36	
Air Blank	0.000	00:38	
Subject Sample # 2	0.163	00:39	
Breath Volume	3.992	00:39	
Air Blank		00:39	RFI Detect
Air Blank	0.000	00:40	

Data Download Date/Time: 07/07/2022 19:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PHILIPP

DATE/TIME OF TEST: 07/01/2022 01:21:17

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.115	01:16
Breath Volume	1.925	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.115	01:19
Breath Volume	2.085	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/01/2022 03:06:26

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.084	03:02
Breath Volume	3.488	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.083	03:04
Breath Volume	3.363	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: YOKOTA

DATE/TIME OF TEST: 07/01/2022 23:20:49

Diagnostics Check	OK	23:19	
Air Blank		23:20	Ambient Fail
Air Blank		23:20	Purge Fail

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: YOKOTA

DATE/TIME OF TEST: 07/01/2022 23:28:57

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.079	23:23
Air Blank	0.000	23:23
Subject Sample # 1	0.300	23:24
Breath Volume	1.742	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:26
Subject Sample # 2	0.320	23:27
Breath Volume	2.066	23:27
Air Blank	0.000	23:28
Control Test	0.081	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: YOKOTA

DATE/TIME OF TEST: 07/02/2022 00:00:22

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.079	23:55	
Air Blank	0.000	23:55	
Subject Sample # 1	0.298	23:56	
Breath Volume	1.898	23:56	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample # 2	0.300	23:58	
Breath Volume	1.359	23:58	
Air Blank	0.000	23:59	
Control Test	0.075	23:59	Interferent Detect
Air Blank	0.000	00:00	

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: DEAVER

DATE/TIME OF TEST: 07/02/2022 01:29:48

Diagnostics Check	OK	01:28	
Air Blank		01:29	Ambient Fail
Air Blank	0.000	01:29	

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: DEAVER

DATE/TIME OF TEST: 07/02/2022 01:51:23

Diagnostics Check	OK	01:50	
Air Blank		01:50	Ambient Fail
Air Blank	0.000	01:51	

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: DEEVER

DATE/TIME OF TEST: 07/02/2022 01:54:01

Diagnostics Check	OK	01:53	
Air Blank		01:53	Ambient Fail
Air Blank	0.000	01:54	

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: DEAVER

DATE/TIME OF TEST: 07/02/2022 01:58:23

Diagnostics Check	OK	01:57	
Air Blank		01:57	Ambient Fail
Air Blank	0.000	01:58	

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: PARENT

DATE/TIME OF TEST: 07/03/2022 01:31:31

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.175	01:26
Breath Volume	1.226	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.173	01:29
Breath Volume	1.515	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 07/07/2022 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/03/2022 03:05:15

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.067	03:00
Breath Volume	2.734	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.068	03:03
Breath Volume	2.976	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/07/2022 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,JASON,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 07/09/2022 01:13:26

Diagnostics Check	OK	01:12	
Air Blank		01:12	Ambient Fail
Air Blank	0.000	01:13	

Data Download Date/Time: 09/30/2022 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DRIVER

DATE/TIME OF TEST: 07/15/2022 02:18:01

Diagnostics Check	OK	02:16	
Air Blank		02:17	Ambient Fail
Air Blank	0.000	02:18	

Data Download Date/Time: 09/30/2022 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSSI,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SAULT,NICHOLAUS
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 12:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:11

SUBJECT'S LAST NAME: FLANAGAN

DATE/TIME OF TEST: 07/01/2022 12:41:45

Diagnostics Check	OK	12:35
Air Blank	0.000	12:35
Control Test	0.080	12:36
Air Blank	0.000	12:36
Subject Sample # 1	0.226	12:37
Breath Volume	1.597	12:37
Air Blank	0.000	12:37
Air Blank	0.000	12:39
Subject Sample # 2	0.243	12:40
Breath Volume	2.437	12:40
Air Blank	0.000	12:40
Control Test	0.079	12:41
Air Blank	0.000	12:41
Diagnostics Check	OK	12:41

Data Download Date/Time: 07/27/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KARL,C
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SMITH,KARL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: GIRON PERZ

DATE/TIME OF TEST: 07/03/2022 20:55:09

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.282	20:49
Breath Volume	1.648	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:51
Subject Sample # 2	0.277	20:53
Breath Volume	1.988	20:53
Air Blank	0.000	20:54
Control Test	0.079	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 07/27/2022 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GROSSE,WILLIAM
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CONROW

DATE/TIME OF TEST: 07/05/2022 01:21:22

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.094	01:16
Breath Volume	2.964	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.094	01:19
Breath Volume	3.117	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 07/27/2022 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KARL,C
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: INDAHL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 07/07/2022 04:00:45

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:56	No Sample Provided
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Air Blank	0.000	03:58	
Subject Sample	# 2	04:00	Sequence Aborted
Breath Volume	0.000	04:00	
Air Blank	0.000	04:00	

Data Download Date/Time: 07/27/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CORBITT,JAMEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/14/2022 03:25:55

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.234	03:21
Breath Volume	1.781	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample # 2	0.250	03:24
Breath Volume	2.625	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 07/27/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MEZINORD,VILINOR
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SPERRY

DATE/TIME OF TEST: 07/15/2022 01:51:36

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.266	01:47
Breath Volume	2.121	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.268	01:50
Breath Volume	2.832	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 07/27/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SOTO,EDWIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: HORROCKS

DATE/TIME OF TEST: 07/18/2022 19:26:18

Diagnostics Check	OK	19:19
Air Blank	0.000	19:20
Control Test	0.080	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.161	19:21
Breath Volume	2.468	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.158	19:24
Breath Volume	1.886	19:24
Air Blank	0.000	19:25
Control Test	0.080	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 07/27/2022 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSSI,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ROSSI,ASHLEY
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: VALLANCE

DATE/TIME OF TEST: 07/20/2022 15:42:10

Diagnostics Check	OK	15:35
Air Blank	0.000	15:36
Control Test	0.080	15:36
Air Blank	0.000	15:37
Subject Sample # 1	0.000	15:37
Breath Volume	2.546	15:37
Air Blank	0.000	15:38
Air Blank	0.000	15:40
Subject Sample # 2	0.000	15:40
Breath Volume	2.253	15:40
Air Blank	0.000	15:41
Control Test	0.079	15:41
Air Blank	0.000	15:42
Diagnostics Check	OK	15:42

Data Download Date/Time: 07/27/2022 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GRABILL,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 07/22/2022 22:29:56

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.234	22:25
Breath Volume	1.582	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:27
Subject Sample # 2	0.217	22:28
Breath Volume	1.328	22:28
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 07/27/2022 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ALTIDOR,JUDE
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: QUIJADA

DATE/TIME OF TEST: 07/23/2022 21:30:40

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:19	
Control Test	0.080	21:19	
Air Blank	0.000	21:20	
Subject Sample	# 1 0.232	21:23	Volume Not Met
Breath Volume	0.296	21:23	
Air Blank	0.000	21:23	
Air Blank	0.000	21:25	
Subject Sample	# 2 0.263	21:25	
Breath Volume	1.304	21:25	
Air Blank	0.000	21:26	
Air Blank	0.000	21:28	
Subject Sample	# 3 0.264	21:29	
Breath Volume	1.242	21:29	
Air Blank	0.000	21:29	
Control Test	0.079	21:30	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

Data Download Date/Time: 07/27/2022 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DEMNER,AUTUMN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: LAWYER

DATE/TIME OF TEST: 07/23/2022 21:58:46

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.231	21:53
Breath Volume	1.199	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.247	21:57
Breath Volume	1.449	21:57
Air Blank	0.000	21:57
Control Test	0.080	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 07/27/2022 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 07/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CRUZ-VAZQUEZ

DATE/TIME OF TEST: 07/31/2022 01:09:37

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.160	01:04
Breath Volume	2.460	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.153	01:08
Breath Volume	2.187	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 08/31/2022 08:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KARL,C
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WAGNER,SEAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 07/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: BRIDEAU

DATE/TIME OF TEST: 07/31/2022 05:31:54

Diagnostics Check	OK	05:27	
Air Blank	0.000	05:27	
Control Test	0.081	05:28	
Air Blank	0.000	05:28	
Subject Sample	# 1	05:31	Sequence Aborted
Breath Volume	0.000	05:31	
Air Blank	0.000	05:31	

Data Download Date/Time: 08/31/2022 08:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PREVATT,JENNIFER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BUBNIS

DATE/TIME OF TEST: 07/02/2022 01:47:51

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.081	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.000	01:43
Breath Volume	3.695	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.000	01:46
Breath Volume	3.531	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DORSEY,JAXON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/03/2022 22:57:17

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:54	
Control Test	0.080	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:56	
Control Test	0.081	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARTLEY,BRADEN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: WOOLISON

DATE/TIME OF TEST: 07/03/2022 23:42:41

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.081	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.075	23:38
Breath Volume	3.554	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.074	23:41
Breath Volume	3.316	23:41
Air Blank	0.000	23:41
Control Test	0.081	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/04/2022 02:17:43

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.081	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1 0.104	02:10	Volume Not Met
Breath Volume	0.472	02:10	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2 0.106	02:16	Volume Not Met
Breath Volume	0.066	02:16	
Air Blank	0.000	02:16	
Control Test	0.079	02:17	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: DITTMER

DATE/TIME OF TEST: 07/04/2022 19:31:12

Diagnostics Check	OK	19:21
Air Blank	0.000	19:22
Control Test	0.081	19:22
Air Blank	0.000	19:23
Subject Sample # 1	0.328	19:26
Breath Volume	1.218	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:28
Subject Sample # 2	0.323	19:29
Breath Volume	1.546	19:29
Air Blank	0.000	19:30
Control Test	0.076	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: JECKOVICH,JONATHON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: STUMPER

DATE/TIME OF TEST: 07/07/2022 01:45:05

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 07/08/2022 19:48:22

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.080	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.170	19:43
Breath Volume	1.558	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.165	19:46
Breath Volume	1.191	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAMPS,BENJAMIN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SANDERS,JAKE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:58

SUBJECT'S LAST NAME: LLESHAJ

DATE/TIME OF TEST: 07/13/2022 11:24:57

Diagnostics Check	OK	11:19	
Air Blank	0.000	11:19	
Control Test	0.081	11:20	
Air Blank	0.000	11:20	
Subject Sample # 1	0.000	11:21	
Breath Volume	2.320	11:21	
Air Blank	0.000	11:21	
Air Blank	0.000	11:23	
Subject Sample # 2	0.000	11:24	
Breath Volume	1.507	11:24	
Air Blank		11:24	RFI Detect
Air Blank	0.000	11:24	

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAMPS,BENJAMIN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SANDERS,JAKE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:26

SUBJECT'S LAST NAME: NUE

DATE/TIME OF TEST: 07/13/2022 11:53:26

Diagnostics Check	OK	11:47
Air Blank	0.000	11:47
Control Test	0.080	11:48
Air Blank	0.000	11:48
Subject Sample # 1	0.000	11:49
Breath Volume	2.425	11:49
Air Blank	0.000	11:49
Air Blank	0.000	11:51
Subject Sample # 2	0.000	11:52
Breath Volume	1.746	11:52
Air Blank	0.000	11:52
Control Test	0.080	11:52
Air Blank	0.000	11:53
Diagnostics Check	OK	11:53

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: RAY, KEATON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: BUBNIS

DATE/TIME OF TEST: 07/18/2022 15:36:08

Diagnostics Check	OK	15:30
Air Blank	0.000	15:30
Control Test	0.081	15:30
Air Blank	0.000	15:31
Subject Sample # 1	0.000	15:31
Breath Volume	2.152	15:31
Air Blank	0.000	15:32
Air Blank	0.000	15:34
Subject Sample # 2	0.000	15:34
Breath Volume	1.695	15:34
Air Blank	0.000	15:35
Control Test	0.080	15:35
Air Blank	0.000	15:36
Diagnostics Check	OK	15:36

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: HALES

DATE/TIME OF TEST: 07/20/2022 00:19:29

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.189	00:11
Breath Volume	1.429	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.220	00:14
Breath Volume	2.992	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 3	0.194	00:17
Breath Volume	1.835	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAUN,ROBERT,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SLAGA,JASMINE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: CLEMENT

DATE/TIME OF TEST: 07/20/2022 23:59:39

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.081	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.000	23:55
Breath Volume	4.093	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.000	23:58
Breath Volume	3.968	23:58
Air Blank	0.000	23:58
Control Test	0.081	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: LITTLEJOHN

DATE/TIME OF TEST: 07/22/2022 01:10:50

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.095	01:06
Breath Volume	1.394	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:08
Subject Sample # 2	0.097	01:09
Breath Volume	1.625	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: FERRARI

DATE/TIME OF TEST: 07/24/2022 01:15:34

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.114	01:11
Breath Volume	3.269	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.114	01:14
Breath Volume	3.093	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 07/25/2022 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: TRIPP, LYLE
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:32

SUBJECT'S LAST NAME: SEJOUR

DATE/TIME OF TEST: 07/24/2022 09:07:13

Diagnostics Check	OK	08:54	
Air Blank	0.000	08:55	
Control Test	0.080	08:55	
Air Blank	0.000	08:55	
Subject Sample	# 1 0.214	08:59	Volume Not Met
Breath Volume	0.601	08:59	
Air Blank	0.000	08:59	
Air Blank	0.000	09:01	
Subject Sample	# 2 0.239	09:02	
Breath Volume	1.335	09:02	
Air Blank	0.000	09:03	
Air Blank	0.000	09:05	
Subject Sample	# 3 0.240	09:05	
Breath Volume	1.425	09:05	
Air Blank	0.000	09:06	
Control Test	0.079	09:06	
Air Blank	0.000	09:07	
Diagnostics Check	OK	09:07	

Data Download Date/Time: 07/25/2022 09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MCLENDON

DATE/TIME OF TEST: 07/28/2022 00:27:32

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.082	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.331	00:23
Breath Volume	2.351	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.323	00:26
Breath Volume	2.355	00:26
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 08/16/2022 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 07/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: RENDE

DATE/TIME OF TEST: 07/31/2022 01:00:10

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.081	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.157	00:55
Breath Volume	2.156	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.156	00:58
Breath Volume	1.742	00:58
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 08/16/2022 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: CHEW

DATE/TIME OF TEST: 07/01/2022 01:35:45

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.079	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.000	01:31
Breath Volume	2.054	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.000	01:34
Breath Volume	1.214	01:34
Air Blank	0.000	01:34
Control Test	0.079	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 07/06/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: BREWER ANGARITA

DATE/TIME OF TEST: 07/02/2022 04:56:08

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.117	04:51
Breath Volume	1.816	04:51
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.121	04:54
Breath Volume	1.675	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 07/06/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: NAVARRETE

DATE/TIME OF TEST: 07/07/2022 23:45:54

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.141	23:40
Breath Volume	1.464	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.139	23:44
Breath Volume	2.023	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 08/10/2022 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: KUHLMAN

DATE/TIME OF TEST: 07/21/2022 01:10:47

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.078	01:07	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:09	
Control Test	0.078	01:10	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 08/10/2022 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

INSTRUMENT REGISTERED TO: LAKE HELEN P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,TREVOR,W
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GRAY,TREVOR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: HEARTZ

DATE/TIME OF TEST: 07/13/2022 17:12:17

Diagnostics Check	OK	17:05
Air Blank	0.000	17:06
Control Test	0.079	17:06
Air Blank	0.000	17:06
Subject Sample # 1	0.182	17:07
Breath Volume	1.472	17:07
Air Blank	0.000	17:08
Air Blank	0.000	17:10
Subject Sample # 2	0.190	17:10
Breath Volume	1.121	17:10
Air Blank	0.000	17:11
Control Test	0.079	17:11
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

Data Download Date/Time: 08/11/2022 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

INSTRUMENT REGISTERED TO: LAKE HELEN P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,TREVOR,W
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GRAY,TREVOR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: KELLOWAN

DATE/TIME OF TEST: 07/19/2022 00:56:48

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.000	00:52
Breath Volume	2.421	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.000	00:55
Breath Volume	1.398	00:55
Air Blank	0.000	00:55
Control Test	0.081	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 08/11/2022 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

INSTRUMENT REGISTERED TO: LAKE HELEN P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,TREVOR,W
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: MERHIB,MICHAEL
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: LINDSEY

DATE/TIME OF TEST: 07/20/2022 01:40:51

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1	0.209	01:30	Volume Not Met
Breath Volume		0.269	01:30	
Air Blank		0.000	01:30	
Air Blank		0.000	01:32	
Subject Sample	# 2	0.236	01:33	
Breath Volume		2.339	01:33	
Air Blank		0.000	01:34	
Air Blank		0.000	01:36	
Subject Sample	# 3	0.201	01:39	Volume Not Met
Breath Volume		0.972	01:39	
Air Blank		0.000	01:39	
Control Test		0.079	01:40	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 08/11/2022 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIERA,SCOTT,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: RIERA,SCOTT
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 06/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: TOLEDO

DATE/TIME OF TEST: 07/01/2022 15:33:15

Diagnostics Check	OK	15:26
Air Blank	0.000	15:26
Control Test	0.081	15:26
Air Blank	0.000	15:27
Subject Sample # 1	0.185	15:28
Breath Volume	2.402	15:28
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.188	15:31
Breath Volume	1.945	15:31
Air Blank	0.000	15:32
Control Test	0.079	15:32
Air Blank	0.000	15:33
Diagnostics Check	OK	15:33

Data Download Date/Time: 07/30/2022 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: DIAMOND,SCOTT
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 06/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: BIDUK

DATE/TIME OF TEST: 07/04/2022 04:03:43

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.224	03:58
Breath Volume	1.355	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.213	04:01
Breath Volume	1.648	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 07/30/2022 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIERA,SCOTT,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: STOKER,RONALD
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: BUI
ARREST TIME: 16:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/04/2022 17:47:29

Diagnostics Check	OK	17:40
Air Blank	0.000	17:40
Control Test	0.080	17:41
Air Blank	0.000	17:41
Subject Sample # 1	0.194	17:42
Breath Volume	2.066	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.194	17:45
Breath Volume	1.914	17:45
Air Blank	0.000	17:46
Control Test	0.078	17:46
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 07/30/2022 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: HILTON,CRAIG
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 06/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HARKINS

DATE/TIME OF TEST: 07/07/2022 02:40:03

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.000	02:35
Breath Volume	2.042	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.000	02:38
Breath Volume	2.171	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:40

Data Download Date/Time: 07/30/2022 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANDRE,
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: PRICE,DYSHEKA
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 06/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/10/2022 02:42:56

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.122	02:38
Breath Volume	4.179	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample # 2	0.116	02:41
Breath Volume	3.433	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 07/30/2022 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS, ANDRE,
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: SNOW, TANNER
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 06/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/14/2022 02:56:21

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.184	02:50
Breath Volume	1.656	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.178	02:54
Breath Volume	2.187	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 07/30/2022 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANDRE,
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: SMITH,DESTINY
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 06/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/28/2022 02:33:04

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.138	02:28
Breath Volume	1.937	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.149	02:31
Breath Volume	2.558	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:33

Data Download Date/Time: 07/30/2022 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANDRE,
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: JOHN DOE,JOHN
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 06/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/28/2022 18:03:44

Diagnostics Check	OK	17:52	
Air Blank	0.000	17:52	
Control Test	0.081	17:53	
Air Blank	0.000	17:53	
Subject Sample	# 1	17:56	No Sample Provided
Breath Volume	0.000	17:56	
Air Blank	0.000	17:57	
Air Blank	0.000	17:59	
Subject Sample	# 2	18:02	No Sample Provided
Breath Volume	0.000	18:02	
Air Blank	0.000	18:02	
Control Test	0.080	18:03	
Air Blank	0.000	18:03	
Diagnostics Check	OK	18:03	

Data Download Date/Time: 07/30/2022 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: GALT,BRETT
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 07/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/31/2022 01:08:43

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.081	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.173	01:04
Breath Volume	1.914	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.182	01:07
Breath Volume	2.406	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 08/14/2022 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: PRICE,DYSHEKA
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 07/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 07/31/2022 01:41:41

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.162	01:36
Breath Volume	2.523	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.157	01:39
Breath Volume	2.410	01:39
Air Blank	0.000	01:40
Control Test	0.078	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 08/14/2022 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINDS,L,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HINDS,L
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: PETTIWAY

DATE/TIME OF TEST: 07/04/2022 04:28:08

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.119	04:23
Breath Volume	2.050	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.121	04:26
Breath Volume	2.269	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 07/19/2022 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD, N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FALCONER

DATE/TIME OF TEST: 07/04/2022 06:52:41

Diagnostics Check	OK	06:45
Air Blank	0.000	06:46
Control Test	0.080	06:46
Air Blank	0.000	06:47
Subject Sample # 1	0.204	06:48
Breath Volume	1.628	06:48
Air Blank	0.000	06:48
Air Blank	0.000	06:50
Subject Sample # 2	0.199	06:51
Breath Volume	1.898	06:51
Air Blank	0.000	06:51
Control Test	0.080	06:52
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

Data Download Date/Time: 07/19/2022 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WONSCH,J,W
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROBERTS,J
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: DENSON

DATE/TIME OF TEST: 07/08/2022 19:32:43

Diagnostics Check	OK	19:17	
Air Blank	0.000	19:18	
Control Test	0.080	19:18	
Air Blank	0.000	19:18	
Subject Sample	# 1 0.084	19:21	
Breath Volume	1.625	19:21	
Air Blank	0.000	19:22	
Air Blank	0.000	19:23	
Subject Sample	# 2 0.075	19:27	Volume Not Met
Breath Volume	0.019	19:27	
Air Blank	0.000	19:27	
Air Blank	0.000	19:29	
Subject Sample	# 3 0.079	19:31	
Breath Volume	1.273	19:31	
Air Blank	0.000	19:31	
Control Test	0.080	19:32	
Air Blank	0.000	19:32	
Diagnostics Check	OK	19:32	

Data Download Date/Time: 07/19/2022 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD, N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: KIWALA

DATE/TIME OF TEST: 07/10/2022 00:54:23

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.081	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.266	00:48
Breath Volume	1.296	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.266	00:52
Breath Volume	1.433	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 07/19/2022 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REAGLE,ASHLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DUMONT,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: CRIM

DATE/TIME OF TEST: 07/10/2022 01:17:50

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:14	
Control Test	0.080	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:16	RFI Detect
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.080	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 07/19/2022 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: VISCONTI,ANTHONY
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: LEIJA

DATE/TIME OF TEST: 07/10/2022 02:01:43

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.081	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.169	01:57
Breath Volume	3.277	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.154	02:00
Breath Volume	2.472	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 07/19/2022 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WONSCH,JESSE,W
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DIAZ,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 07/10/2022 18:58:09

Diagnostics Check	OK	18:51
Air Blank	0.000	18:52
Control Test	0.079	18:52
Air Blank	0.000	18:52
Subject Sample # 1	0.241	18:53
Breath Volume	2.015	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.236	18:56
Breath Volume	1.195	18:56
Air Blank	0.000	18:57
Control Test	0.079	18:57
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 07/19/2022 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/15/2022 00:45:05

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.081	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.110	00:40
Breath Volume	2.300	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.114	00:43
Breath Volume	1.742	00:43
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 07/19/2022 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUIS,BOBBY,G
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: RUIS,BOBBY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:18

SUBJECT'S LAST NAME: WHITEHEAD

DATE/TIME OF TEST: 07/19/2022 12:44:42

Diagnostics Check	OK	12:38
Air Blank	0.000	12:38
Control Test	0.081	12:39
Air Blank	0.000	12:39
Subject Sample # 1	0.086	12:40
Breath Volume	2.261	12:40
Air Blank	0.000	12:41
Air Blank	0.000	12:42
Subject Sample # 2	0.085	12:43
Breath Volume	2.121	12:43
Air Blank	0.000	12:43
Control Test	0.080	12:44
Air Blank	0.000	12:44
Diagnostics Check	OK	12:44

Data Download Date/Time: 08/04/2022 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: VISCONTI,ANTHONY
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: PAPPAS

DATE/TIME OF TEST: 07/20/2022 03:33:33

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.082	03:28
Breath Volume	1.609	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.090	03:32
Breath Volume	1.671	03:32
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 08/04/2022 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATLOCK,LOGAN,R
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: MATLOCK,LOGAN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: WESTER

DATE/TIME OF TEST: 07/23/2022 02:30:30

Diagnostics Check	OK	02:29	
Air Blank		02:30	RFI Detect
Air Blank	0.000	02:30	

Data Download Date/Time: 08/04/2022 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD, N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: EAST

DATE/TIME OF TEST: 07/26/2022 04:36:17

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.261	04:31
Breath Volume	2.207	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.254	04:34
Breath Volume	2.437	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 08/04/2022 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD, N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: SENDERLING

DATE/TIME OF TEST: 07/29/2022 01:56:08

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.169	01:50
Breath Volume	1.542	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.154	01:54
Breath Volume	2.191	01:54
Air Blank	0.000	01:55
Control Test	0.081	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 08/04/2022 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 07/29/2022 23:54:03

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.224	23:49
Breath Volume	1.445	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.218	23:52
Breath Volume	1.269	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:54

Data Download Date/Time: 08/04/2022 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: SANTSPREE

DATE/TIME OF TEST: 07/10/2022 00:10:29

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.000	00:06
Breath Volume	2.628	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.000	00:09
Breath Volume	2.308	00:09
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 07/14/2022 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: HYATT,JUSTIN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: SOHL

DATE/TIME OF TEST: 07/03/2022 20:32:04

Diagnostics Check	OK	20:28	
Air Blank	0.000	20:28	
Control Test	0.080	20:28	
Air Blank	0.000	20:29	
Subject Sample	# 1	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:31	
Control Test	0.080	20:31	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:32	

Data Download Date/Time: 07/28/2022 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, MICHAEL, D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: HYATT, JUSTIN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: SOHL

DATE/TIME OF TEST: 07/03/2022 21:45:05

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.166	21:40
Breath Volume	2.535	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:42
Subject Sample # 2	0.175	21:43
Breath Volume	3.839	21:43
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:45

Data Download Date/Time: 07/28/2022 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JOSE,L
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28420080A1
DRY GAS EXPIRY DATE: 12/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 07/20/2022 16:59:37

Diagnostics Check	OK	16:53
Air Blank	0.000	16:53
Control Test	0.081	16:54
Air Blank	0.000	16:54
Subject Sample # 1	0.233	16:55
Breath Volume	2.968	16:55
Air Blank	0.000	16:55
Air Blank	0.000	16:57
Subject Sample # 2	0.227	16:58
Breath Volume	2.062	16:58
Air Blank	0.000	16:58
Control Test	0.080	16:59
Air Blank	0.000	16:59
Diagnostics Check	OK	16:59

Data Download Date/Time: 08/17/2022 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: YANICK, M
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: VALDEZ-RODRIGUEZ

DATE/TIME OF TEST: 07/04/2022 21:55:03

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.081	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.248	21:50
Breath Volume	1.750	21:50
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.241	21:53
Breath Volume	1.582	21:53
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:54
Diagnostics Check	OK	21:55

Data Download Date/Time: 07/26/2022 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: EDGAR, J
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: TOMS

DATE/TIME OF TEST: 07/10/2022 03:47:33

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.081	03:44	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 07/26/2022 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: ROSA,EMILIO
ARRESTING OFFICER AGENCY: COCOA BEACH PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/17/2022 00:58:52

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.230	00:54
Breath Volume	1.671	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:56
Subject Sample # 2	0.239	00:57
Breath Volume	1.957	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 07/26/2022 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: PRICE, C
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: BICHANICH

DATE/TIME OF TEST: 07/25/2022 00:01:56

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.081	23:54	
Air Blank	0.000	23:54	
Subject Sample	# 2 0.183	23:57	Volume Not Met
Breath Volume	0.101	23:57	
Air Blank	0.000	23:58	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:01	
Control Test	0.079	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 07/26/2022 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: CHAMBERLAIN, J
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/25/2022 03:36:09

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.082	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.177	03:31
Breath Volume	1.996	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.171	03:34
Breath Volume	1.703	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 07/26/2022 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: BROOKS,NEILL
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: CANADA

DATE/TIME OF TEST: 07/25/2022 20:39:25

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.083	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.000	20:35
Breath Volume	2.812	20:35
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.000	20:38
Breath Volume	3.917	20:38
Air Blank	0.000	20:38
Control Test	0.080	20:38
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

Data Download Date/Time: 07/26/2022 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: RICO

DATE/TIME OF TEST: 07/06/2022 04:01:19

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.079	03:49	
Air Blank	0.000	03:50	
Subject Sample	# 1 0.291	03:53	Volume Not Met
Breath Volume	0.070	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample	# 2 0.339	03:56	
Breath Volume	2.062	03:56	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 3 0.357	03:59	
Breath Volume	1.882	03:59	
Air Blank	0.000	04:00	
Control Test	0.078	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 07/19/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: ROLLINS, SEAN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MIX

DATE/TIME OF TEST: 07/09/2022 02:58:46

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.170	02:53
Breath Volume	2.296	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.167	02:57
Breath Volume	2.812	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 07/19/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: CAMACHO, NICOLE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HAARZ

DATE/TIME OF TEST: 07/09/2022 03:37:17

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.080	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:36	
Control Test	0.079	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 07/19/2022 16:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: CAMACHO,NICOLE

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 03:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: TEAL

DATE/TIME OF TEST: 07/10/2022 04:20:06

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.179	04:15
Breath Volume	2.542	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.168	04:18
Breath Volume	2.593	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 07/19/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REASONER,DEBBIE,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BURGOS,ANTHONY
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HO YOW

DATE/TIME OF TEST: 07/18/2022 01:10:24

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.214	01:05
Breath Volume	1.187	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.222	01:08
Breath Volume	1.476	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 07/19/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KEITH,L
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BURGOS,A
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212993
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 07/18/2022 21:16:53

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:11	
Control Test	0.080	21:11	
Air Blank	0.000	21:12	
Subject Sample # 1	0.000	21:13	
Breath Volume	3.832	21:13	
Air Blank	0.000	21:13	
Air Blank		21:15	RFI Detect
Air Blank	0.000	21:15	
Control Test	0.081	21:16	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 07/19/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REED,KEITH,L

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BURGOS,A

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 212993

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 20:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 07/18/2022 21:26:31

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.000	21:22
Breath Volume	3.753	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.000	21:25
Breath Volume	3.023	21:25
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 07/19/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER, CHAD, T
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: REBANDO, BRIAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 07/25/2022 14:52:25

Diagnostics Check	OK	14:46
Air Blank	0.000	14:46
Control Test	0.080	14:46
Air Blank	0.000	14:47
Subject Sample # 1	0.248	14:48
Breath Volume	2.031	14:48
Air Blank	0.000	14:48
Air Blank	0.000	14:50
Subject Sample # 2	0.250	14:50
Breath Volume	1.667	14:50
Air Blank	0.000	14:51
Control Test	0.079	14:51
Air Blank	0.000	14:52
Diagnostics Check	OK	14:52

Data Download Date/Time: 08/23/2022 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: CAMACHO,NICOLE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 07/27/2022 01:27:13

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.259	01:22
Breath Volume	1.527	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.258	01:25
Breath Volume	1.933	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 08/23/2022 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PABON,G,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: LAWLESS,B

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 19:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/27/2022 20:23:45

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:21	
Control Test	0.080	20:21	
Air Blank	0.000	20:22	
Subject Sample	# 1	20:22	RFI Detect
Breath Volume	0.000	20:22	
Air Blank	0.000	20:22	
Control Test	0.080	20:23	
Air Blank	0.000	20:23	
Diagnostics Check	OK	20:23	

Data Download Date/Time: 08/23/2022 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PABON,G,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: LAWLESS,B

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 19:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/27/2022 20:34:52

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.081	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.000	20:28
Breath Volume	2.230	20:28
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.000	20:33
Breath Volume	2.714	20:33
Air Blank	0.000	20:33
Control Test	0.080	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 08/23/2022 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ROLLINS,SEAN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 07/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/23/2022 04:07:39

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.079	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:07	Sequence Aborted
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	

Data Download Date/Time: 08/19/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ROLLINS,SEAN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 07/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/23/2022 04:17:39

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.167	04:12
Breath Volume	1.710	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.164	04:16
Breath Volume	3.562	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 08/19/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: SOTO,ADRIAN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 07/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: LOPEZ MATIAS

DATE/TIME OF TEST: 07/27/2022 21:28:45

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.104	21:24
Breath Volume	3.808	21:24
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.100	21:27
Breath Volume	3.847	21:27
Air Blank	0.000	21:27
Control Test	0.079	21:28
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 08/19/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DACEY, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: DUNLOP, MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: GRIMES

DATE/TIME OF TEST: 07/01/2022 01:40:12

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.082	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.000	01:34
Breath Volume	1.406	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.000	01:38
Breath Volume	1.660	01:38
Air Blank	0.000	01:39
Control Test	0.082	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 07/19/2022 23:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HUY,H
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/02/2022 02:30:39

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
Control Test	0.082	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1 0.184	02:22	Volume Not Met
Breath Volume	0.578	02:22	
Air Blank	0.000	02:23	
Air Blank	0.000	02:25	
Subject Sample	# 2 0.213	02:25	
Breath Volume	2.210	02:25	
Air Blank	0.000	02:26	
Air Blank	0.000	02:28	
Subject Sample	# 3 0.215	02:29	
Breath Volume	1.566	02:29	
Air Blank	0.000	02:29	
Control Test	0.081	02:30	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 07/19/2022 23:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,ELISHA,D
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CERNUTO,ELISHA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: BUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: ALVAREZ MIJARES

DATE/TIME OF TEST: 07/02/2022 21:01:26

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.082	20:55
Air Blank	0.000	20:56
Subject Sample # 1	0.051	20:56
Breath Volume	3.007	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.052	20:59
Breath Volume	3.519	20:59
Air Blank	0.000	21:00
Control Test	0.081	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 07/19/2022 23:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: MATEO

DATE/TIME OF TEST: 07/02/2022 21:26:14

Diagnostics Check	OK	21:18
Air Blank	0.000	21:19
Control Test	0.081	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.109	21:20
Breath Volume	1.136	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.116	21:24
Breath Volume	1.898	21:24
Air Blank	0.000	21:25
Control Test	0.082	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 07/19/2022 23:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: CASAS

DATE/TIME OF TEST: 07/03/2022 03:46:05

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.082	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.167	03:41
Breath Volume	2.882	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.168	03:44
Breath Volume	3.250	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 07/19/2022 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: COLGAN,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 07/03/2022 05:36:52

Diagnostics Check	OK	05:28
Air Blank	0.000	05:29
Control Test	0.082	05:29
Air Blank	0.000	05:29
Subject Sample # 1	0.136	05:32
Breath Volume	2.335	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:34
Subject Sample # 2	0.138	05:35
Breath Volume	2.367	05:35
Air Blank	0.000	05:35
Control Test	0.081	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 07/19/2022 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: BUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/03/2022 18:46:04

Diagnostics Check	OK	18:39
Air Blank	0.000	18:40
Control Test	0.082	18:40
Air Blank	0.000	18:40
Subject Sample # 1	0.128	18:41
Breath Volume	4.300	18:41
Air Blank	0.000	18:42
Air Blank	0.000	18:43
Subject Sample # 2	0.128	18:44
Breath Volume	4.410	18:44
Air Blank	0.000	18:45
Control Test	0.081	18:45
Air Blank	0.000	18:45
Diagnostics Check	OK	18:46

Data Download Date/Time: 07/19/2022 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: O'CONNER

DATE/TIME OF TEST: 07/03/2022 21:58:51

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.082	21:53
Air Blank	0.000	21:53
Subject Sample # 1	0.225	21:54
Breath Volume	4.167	21:54
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.221	21:57
Breath Volume	3.875	21:57
Air Blank	0.000	21:57
Control Test	0.081	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 07/19/2022 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: OTZOY-TZUN

DATE/TIME OF TEST: 07/04/2022 02:07:47

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:55	
Control Test	0.082	01:55	
Air Blank	0.000	01:56	
Subject Sample	# 1 0.213	01:59	Volume Not Met
Breath Volume	0.152	01:59	
Air Blank	0.000	02:00	
Air Blank	0.000	02:02	
Subject Sample	# 2 0.224	02:02	
Breath Volume	1.429	02:02	
Air Blank	0.000	02:03	
Air Blank	0.000	02:05	
Subject Sample	# 3 0.209	02:06	
Breath Volume	1.140	02:06	
Air Blank	0.000	02:06	
Control Test	0.081	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 07/19/2022 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: COSENTINO

DATE/TIME OF TEST: 07/04/2022 02:58:31

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.082	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1 0.216	02:48	Volume Not Met
Breath Volume	0.531	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 2 0.293	02:51	
Breath Volume	1.328	02:51	
Air Blank	0.000	02:52	
Air Blank	0.000	02:54	
Subject Sample	# 3 0.281	02:56	
Breath Volume	1.527	02:56	
Air Blank	0.000	02:57	
Control Test	0.080	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 07/19/2022 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: FREDLUND

DATE/TIME OF TEST: 07/04/2022 18:44:56

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.081	18:38
Air Blank	0.000	18:39
Subject Sample # 1	0.243	18:40
Breath Volume	2.882	18:40
Air Blank	0.000	18:40
Air Blank	0.000	18:42
Subject Sample # 2	0.247	18:43
Breath Volume	3.066	18:43
Air Blank	0.000	18:44
Control Test	0.081	18:44
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

Data Download Date/Time: 07/19/2022 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERIC,JEAN,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FREDERIC,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: IDROVO

DATE/TIME OF TEST: 07/05/2022 23:56:53

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.082	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.230	23:52
Breath Volume	1.449	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.237	23:55
Breath Volume	2.265	23:55
Air Blank	0.000	23:55
Control Test	0.080	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 07/19/2022 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDARELLI,CARA,G
BREATH TEST OPERATOR AGENCY: Sewall's Point PD
ARRESTING OFFICER: CARDARELLI,CARA
ARRESTING OFFICER AGENCY: Sewall's Point PD
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: TETREAULT

DATE/TIME OF TEST: 07/06/2022 21:13:55

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:11	
Control Test	0.081	21:11	
Air Blank	0.000	21:12	
Subject Sample	# 1	21:12	Subject Test Refused
Breath Volume	0.000	21:12	
Air Blank	0.000	21:12	
Control Test	0.081	21:13	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 07/19/2022 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/08/2022 20:33:56

Diagnostics Check	OK	20:27
Air Blank	0.000	20:27
Control Test	0.082	20:28
Air Blank	0.000	20:28
Subject Sample # 1	0.206	20:29
Breath Volume	2.210	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:31
Subject Sample # 2	0.216	20:32
Breath Volume	3.625	20:32
Air Blank	0.000	20:33
Control Test	0.080	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 07/19/2022 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: VELASCO

DATE/TIME OF TEST: 07/09/2022 03:24:18

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.082	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.169	03:19
Breath Volume	3.621	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.170	03:22
Breath Volume	3.210	03:22
Air Blank	0.000	03:23
Control Test	0.082	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 07/19/2022 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,CAITLYN,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SPIRITO,GEORGE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: BUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 07/09/2022 19:00:39

Diagnostics Check	OK	18:53
Air Blank	0.000	18:54
Control Test	0.082	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.167	18:55
Breath Volume	1.574	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.167	18:59
Breath Volume	1.609	18:59
Air Blank	0.000	18:59
Control Test	0.081	19:00
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 07/19/2022 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEDER,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: REEDER,MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: LOCASCIO

DATE/TIME OF TEST: 07/12/2022 15:33:05

Diagnostics Check	OK	15:25
Air Blank	0.000	15:25
Control Test	0.081	15:25
Air Blank	0.000	15:26
Subject Sample # 1	0.177	15:28
Breath Volume	1.960	15:28
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.181	15:31
Breath Volume	1.578	15:31
Air Blank	0.000	15:32
Control Test	0.081	15:32
Air Blank	0.000	15:32
Diagnostics Check	OK	15:33

Data Download Date/Time: 07/19/2022 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: NAULTY

DATE/TIME OF TEST: 07/15/2022 03:09:29

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.082	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.041	03:04
Breath Volume	4.046	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.040	03:07
Breath Volume	3.527	03:07
Air Blank	0.000	03:08
Control Test	0.082	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 07/19/2022 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 07/16/2022 03:56:48

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.082	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1 0.097	03:48	Volume Not Met
Breath Volume	0.832	03:48	
Air Blank	0.000	03:49	
Air Blank	0.000	03:51	
Subject Sample	# 2 0.096	03:52	
Breath Volume	1.371	03:52	
Air Blank	0.000	03:52	
Air Blank	0.000	03:54	
Subject Sample	# 3 0.096	03:55	
Breath Volume	1.429	03:55	
Air Blank	0.000	03:55	
Control Test	0.082	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 07/19/2022 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BARTLETT

DATE/TIME OF TEST: 07/17/2022 03:16:56

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.246	03:12
Breath Volume	3.781	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.242	03:15
Breath Volume	3.609	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 07/19/2022 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/22/2022 02:26:15

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.082	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.310	02:18
Breath Volume	1.171	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.287	02:21
Breath Volume	1.457	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 3	0.288	02:24
Breath Volume	1.769	02:24
Air Blank	0.000	02:25
Control Test	0.081	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 08/25/2022 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SULUGUI-PAR

DATE/TIME OF TEST: 07/22/2022 23:22:13

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.081	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.149	23:17
Breath Volume	2.363	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.141	23:20
Breath Volume	1.585	23:20
Air Blank	0.000	23:21
Control Test	0.082	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 08/25/2022 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: ESTEBAN GASPAR

DATE/TIME OF TEST: 07/23/2022 07:04:54

Diagnostics Check	OK	06:58
Air Blank	0.000	06:58
Control Test	0.082	06:59
Air Blank	0.000	06:59
Subject Sample # 1	0.111	07:00
Breath Volume	2.316	07:00
Air Blank	0.000	07:01
Air Blank	0.000	07:02
Subject Sample # 2	0.110	07:03
Breath Volume	2.136	07:03
Air Blank	0.000	07:03
Control Test	0.081	07:04
Air Blank	0.000	07:04
Diagnostics Check	OK	07:04

Data Download Date/Time: 08/25/2022 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: FRANKOWSKI

DATE/TIME OF TEST: 07/23/2022 20:44:00

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.082	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.156	20:39
Breath Volume	1.832	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.164	20:42
Breath Volume	1.468	20:42
Air Blank	0.000	20:43
Control Test	0.079	20:43
Air Blank	0.000	20:43
Diagnostics Check	OK	20:44

Data Download Date/Time: 08/25/2022 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,DILLAN,R
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SPIRITO,GEORGE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: EPLING

DATE/TIME OF TEST: 07/28/2022 22:45:41

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.082	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.159	22:40
Breath Volume	1.113	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.149	22:44
Breath Volume	1.425	22:44
Air Blank	0.000	22:44
Control Test	0.081	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 08/25/2022 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: MENDEZ-TAMAS

DATE/TIME OF TEST: 07/30/2022 00:06:07

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.082	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.156	00:00
Breath Volume	1.414	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.148	00:04
Breath Volume	1.453	00:04
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 08/25/2022 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: JOAQUIN-MATEO

DATE/TIME OF TEST: 07/30/2022 00:39:38

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.081	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.243	00:33
Breath Volume	2.769	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.244	00:38
Breath Volume	1.906	00:38
Air Blank	0.000	00:38
Control Test	0.079	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 08/25/2022 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: ZAPETA-GUICOL

DATE/TIME OF TEST: 07/30/2022 04:40:18

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.248	04:35
Breath Volume	1.242	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.242	04:38
Breath Volume	1.601	04:38
Air Blank	0.000	04:39
Control Test	0.082	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 08/25/2022 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: NIKOLLI,ALEKSANDER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 07/31/2022 02:17:24

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.081	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.094	02:12
Breath Volume	4.894	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.094	02:15
Breath Volume	4.789	02:15
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 08/25/2022 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 07/31/2022 03:42:46

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.081	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.171	03:38
Breath Volume	2.253	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.175	03:41
Breath Volume	1.968	03:41
Air Blank	0.000	03:41
Control Test	0.081	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 08/25/2022 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 07/01/2022 05:24:43

Diagnostics Check	OK	05:18
Air Blank	0.000	05:18
Control Test	0.082	05:19
Air Blank	0.000	05:19
Subject Sample # 1	0.111	05:20
Breath Volume	2.210	05:20
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.108	05:23
Breath Volume	2.609	05:23
Air Blank	0.000	05:23
Control Test	0.081	05:24
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: CHILDERS

DATE/TIME OF TEST: 07/02/2022 03:52:03

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.082	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.170	03:46
Breath Volume	1.132	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.161	03:50
Breath Volume	1.457	03:50
Air Blank	0.000	03:51
Control Test	0.082	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,WILLIAM,B
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: HATLEY,KENDALL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: WINDHAM

DATE/TIME OF TEST: 07/02/2022 19:39:42

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.081	19:34
Air Blank	0.000	19:34
Subject Sample # 1	0.120	19:35
Breath Volume	2.050	19:35
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.125	19:38
Breath Volume	2.289	19:38
Air Blank	0.000	19:38
Control Test	0.081	19:39
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/03/2022 21:58:44

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.081	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.085	21:53
Breath Volume	1.402	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.083	21:57
Breath Volume	2.089	21:57
Air Blank	0.000	21:57
Control Test	0.082	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 07/05/2022 02:56:02

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.171	02:51
Breath Volume	3.113	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 2	0.171	02:54
Breath Volume	3.457	02:54
Air Blank	0.000	02:55
Control Test	0.082	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:56

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: SMITH III

DATE/TIME OF TEST: 07/06/2022 05:40:30

Diagnostics Check	OK	05:33
Air Blank	0.000	05:33
Control Test	0.082	05:33
Air Blank	0.000	05:34
Subject Sample # 1	0.167	05:34
Breath Volume	2.300	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 2	0.147	05:39
Breath Volume	1.675	05:39
Air Blank	0.000	05:39
Control Test	0.081	05:39
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 14:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:59

SUBJECT'S LAST NAME: CLEMENTI

DATE/TIME OF TEST: 07/13/2022 15:35:40

Diagnostics Check	OK	15:27
Air Blank	0.000	15:27
Control Test	0.082	15:27
Air Blank	0.000	15:28
Subject Sample # 1	0.250	15:31
Breath Volume	2.097	15:31
Air Blank	0.000	15:31
Air Blank	0.000	15:33
Subject Sample # 2	0.234	15:34
Breath Volume	1.167	15:34
Air Blank	0.000	15:34
Control Test	0.080	15:35
Air Blank	0.000	15:35
Diagnostics Check	OK	15:35

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAFA,YUSUF,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ARAFA,YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: BARBIERI

DATE/TIME OF TEST: 07/13/2022 18:57:18

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.081	18:51
Air Blank	0.000	18:51
Subject Sample # 1	0.188	18:52
Breath Volume	2.855	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.186	18:55
Breath Volume	3.640	18:55
Air Blank	0.000	18:56
Control Test	0.081	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: MUNSON

DATE/TIME OF TEST: 07/14/2022 14:04:46

Diagnostics Check	OK	13:58
Air Blank	0.000	13:58
Control Test	0.082	13:59
Air Blank	0.000	13:59
Subject Sample # 1	0.223	14:00
Breath Volume	3.070	14:00
Air Blank	0.000	14:00
Air Blank	0.000	14:02
Subject Sample # 2	0.220	14:03
Breath Volume	3.113	14:03
Air Blank	0.000	14:03
Control Test	0.081	14:04
Air Blank	0.000	14:04
Diagnostics Check	OK	14:04

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 07/15/2022 22:36:54

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.081	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.325	22:28
Breath Volume	1.679	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.297	22:32
Breath Volume	2.417	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 3	0.332	22:35
Breath Volume	1.808	22:35
Air Blank	0.000	22:36
Control Test	0.081	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHNATHAN,F
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: HATLEY,KENDALL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ANGELES

DATE/TIME OF TEST: 07/17/2022 23:41:38

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.081	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.162	23:37
Breath Volume	2.285	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.166	23:40
Breath Volume	4.007	23:40
Air Blank	0.000	23:40
Control Test	0.081	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 07/21/2022 18:49:15

Diagnostics Check	OK	18:43
Air Blank	0.000	18:43
Control Test	0.081	18:43
Air Blank	0.000	18:44
Subject Sample # 1	0.000	18:44
Breath Volume	2.382	18:44
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample # 2	0.000	18:47
Breath Volume	2.707	18:47
Air Blank	0.000	18:48
Control Test	0.082	18:48
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHNATHAN,F
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: HATLEY,KENDALL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: TROTTER

DATE/TIME OF TEST: 07/22/2022 03:14:22

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.081	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:09	No Sample Provided
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 2	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.081	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 07/22/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: BIGHAM

DATE/TIME OF TEST: 07/30/2022 00:43:29

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.082	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.099	00:38
Breath Volume	1.199	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample # 2	0.103	00:42
Breath Volume	1.679	00:42
Air Blank	0.000	00:42
Control Test	0.082	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 08/23/2022 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHNATHAN,F
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: YOUNG,JOHNATHAN
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ELDER

DATE/TIME OF TEST: 07/30/2022 03:53:35

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.082	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.186	03:49
Breath Volume	2.523	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.188	03:52
Breath Volume	2.253	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 08/23/2022 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 07/30/2022 18:30:10

Diagnostics Check	OK	18:24
Air Blank	0.000	18:24
Control Test	0.081	18:24
Air Blank	0.000	18:25
Subject Sample # 1	0.110	18:25
Breath Volume	2.308	18:25
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample # 2	0.105	18:28
Breath Volume	1.437	18:28
Air Blank	0.000	18:29
Control Test	0.082	18:29
Air Blank	0.000	18:30
Diagnostics Check	OK	18:30

Data Download Date/Time: 08/23/2022 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: THOMPSON,LATHAM
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: PARKSONS

DATE/TIME OF TEST: 07/31/2022 01:03:01

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.081	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.130	00:58
Breath Volume	1.601	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.126	01:01
Breath Volume	2.148	01:01
Air Blank	0.000	01:02
Control Test	0.081	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:03

Data Download Date/Time: 08/23/2022 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/03/2022 22:30:05

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.078	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.146	22:25
Breath Volume	1.613	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.135	22:28
Breath Volume	1.671	22:28
Air Blank	0.000	22:29
Control Test	0.077	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:30

Data Download Date/Time: 07/25/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: LEIN

DATE/TIME OF TEST: 07/04/2022 21:37:10

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.078	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.360	21:32
Breath Volume	2.296	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.375	21:35
Breath Volume	2.265	21:35
Air Blank	0.000	21:36
Control Test	0.077	21:36
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 07/25/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SULLIVAN, ROBERT
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 07/06/2022 04:30:14

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.102	04:25
Breath Volume	3.664	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.099	04:28
Breath Volume	2.269	04:28
Air Blank	0.000	04:29
Control Test	0.078	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 07/25/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SITZLER, ANTHONY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/07/2022 05:21:44

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:08	
Control Test	0.078	05:08	
Air Blank	0.000	05:08	
Subject Sample	# 1 0.362	05:11	No .020 agreement
Breath Volume	1.109	05:11	
Air Blank	0.000	05:12	
Air Blank	0.000	05:13	
Subject Sample	# 2 0.319	05:17	Volume Not Met
Breath Volume	0.230	05:17	
Air Blank	0.000	05:17	
Air Blank	0.000	05:19	
Subject Sample	# 3 0.393	05:20	No .020 agreement
Breath Volume	2.011	05:20	
Air Blank	0.000	05:20	
Control Test	0.075	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 07/25/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: CORONADO

DATE/TIME OF TEST: 07/16/2022 01:43:54

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.077	01:33	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:36	No Sample Provided
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Air Blank	0.000	01:39	
Subject Sample	# 2	01:42	No Sample Provided
Breath Volume	0.000	01:42	
Air Blank	0.000	01:43	
Control Test	0.077	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 07/25/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: ERWIN

DATE/TIME OF TEST: 07/23/2022 02:27:26

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.078	02:24	
Air Blank	0.000	02:25	
Subject Sample # 1	0.192	02:25	
Breath Volume	1.421	02:25	
Air Blank		02:26	Improper Sample
Air Blank	0.000	02:26	
Control Test	0.078	02:26	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 07/25/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: ERWIN

DATE/TIME OF TEST: 07/23/2022 02:36:10

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.077	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.199	02:31
Breath Volume	2.214	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.191	02:34
Breath Volume	2.960	02:34
Air Blank	0.000	02:35
Control Test	0.078	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 07/25/2022 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MCGLAUFILIN

DATE/TIME OF TEST: 07/25/2022 00:02:29

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.223	23:57
Breath Volume	3.710	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.235	00:01
Breath Volume	3.808	00:01
Air Blank	0.000	00:01
Control Test	0.078	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 07/25/2022 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 07/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CHURCH

DATE/TIME OF TEST: 07/25/2022 03:01:37

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.143	02:57
Breath Volume	4.789	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.132	03:00
Breath Volume	2.484	03:00
Air Blank	0.000	03:00
Control Test	0.077	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 07/25/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SULLIVAN,ROBERT
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 07/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ROBY

DATE/TIME OF TEST: 07/30/2022 23:17:29

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.137	23:12
Breath Volume	2.101	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.141	23:16
Breath Volume	2.261	23:16
Air Blank	0.000	23:16
Control Test	0.078	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 08/22/2022 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 07/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 07/31/2022 03:32:32

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.247	03:27
Breath Volume	1.890	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.243	03:31
Breath Volume	1.789	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 08/22/2022 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLARD,DANIEL,C
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DILLARD,DANIEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/02/2022 13:35:59

Diagnostics Check	OK	13:29
Air Blank	0.000	13:30
Control Test	0.081	13:30
Air Blank	0.000	13:31
Subject Sample # 1	0.000	13:31
Breath Volume	3.687	13:31
Air Blank	0.000	13:31
Air Blank	0.000	13:33
Subject Sample # 2	0.000	13:34
Breath Volume	3.652	13:34
Air Blank	0.000	13:35
Control Test	0.081	13:35
Air Blank	0.000	13:35
Diagnostics Check	OK	13:35

Data Download Date/Time: 07/22/2022 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: AIRD

DATE/TIME OF TEST: 07/09/2022 00:28:54

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:28	
Control Test	0.082	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 07/22/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STREIGOLD,DAVID,M
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: CHARLES,HERARD
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 13:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: VIVIERS

DATE/TIME OF TEST: 07/09/2022 15:28:37

Diagnostics Check	OK	15:21
Air Blank	0.000	15:21
Control Test	0.082	15:21
Air Blank	0.000	15:22
Subject Sample # 1	0.000	15:24
Breath Volume	4.011	15:24
Air Blank	0.000	15:24
Air Blank	0.000	15:26
Subject Sample # 2	0.000	15:27
Breath Volume	1.710	15:27
Air Blank	0.000	15:27
Control Test	0.081	15:28
Air Blank	0.000	15:28
Diagnostics Check	OK	15:28

Data Download Date/Time: 07/22/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: VITONE

DATE/TIME OF TEST: 07/09/2022 21:09:05

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.081	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.126	21:02
Breath Volume	1.378	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:04
Subject Sample # 2	0.133	21:07
Breath Volume	1.195	21:07
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 07/22/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/17/2022 01:13:07

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.082	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.234	01:08
Breath Volume	4.160	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.230	01:11
Breath Volume	3.167	01:11
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 07/22/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLARD,DANIEL,C
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DILLARD,DANIEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 07/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 07/24/2022 02:44:00

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.081	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.328	02:39
Breath Volume	1.835	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample # 2	0.319	02:42
Breath Volume	1.945	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:44

Data Download Date/Time: 07/25/2022 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 07/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 07/24/2022 03:12:27

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.160	03:07
Breath Volume	3.242	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.166	03:10
Breath Volume	4.179	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 07/25/2022 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 07/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: JIMENEZ JOSE

DATE/TIME OF TEST: 07/26/2022 00:58:12

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.182	00:53
Breath Volume	3.050	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.176	00:56
Breath Volume	2.882	00:56
Air Blank	0.000	00:57
Control Test	0.081	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 07/28/2022 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 07/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 07/31/2022 23:59:05

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.081	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.142	23:54
Breath Volume	2.152	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.143	23:57
Breath Volume	1.957	23:57
Air Blank	0.000	23:58
Control Test	0.081	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 08/09/2022 07:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDERS,MARC,G
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: MANDERS,MARC
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: GONZALEZ SILVA

DATE/TIME OF TEST: 07/01/2022 13:53:46

Diagnostics Check	OK	13:50	
Air Blank	0.000	13:51	
Control Test	0.080	13:51	
Air Blank	0.000	13:52	
Subject Sample	# 1	13:52	Subject Test Refused
Breath Volume	0.000	13:52	
Air Blank	0.000	13:52	
Control Test	0.081	13:53	
Air Blank	0.000	13:53	
Diagnostics Check	OK	13:53	

Data Download Date/Time: 07/24/2022 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWNSBERGER, BLAINE, T
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: WILSON, TEONDRE
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 07/02/2022 05:36:53

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.081	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.130	05:32
Breath Volume	3.378	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:34
Subject Sample # 2	0.133	05:35
Breath Volume	3.410	05:35
Air Blank	0.000	05:35
Control Test	0.081	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 07/24/2022 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRONK,A,
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: ELLIS,C
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: RHYNE JR

DATE/TIME OF TEST: 07/03/2022 00:01:49

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.081	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.196	23:56
Breath Volume	1.996	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.206	00:00
Breath Volume	2.671	00:00
Air Blank	0.000	00:00
Control Test	0.080	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 07/24/2022 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAUDERS,S,A
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/03/2022 21:48:16

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.081	21:42
Air Blank	0.000	21:43
Subject Sample # 1	0.192	21:43
Breath Volume	2.890	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.182	21:46
Breath Volume	2.031	21:46
Air Blank	0.000	21:47
Control Test	0.080	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 07/24/2022 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: WISDOM,CHRISTOPHER
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 07/05/2022 22:10:17

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:09	
Control Test	0.058	22:09	Control Outside Tolerance
Air Blank	0.000	22:10	

Data Download Date/Time: 07/24/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: WISDOM,CHRISTOPHER
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 07/05/2022 22:20:33

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.075	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.058	22:16
Breath Volume	1.296	22:16
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.062	22:19
Breath Volume	2.484	22:19
Air Blank	0.000	22:19
Control Test	0.076	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 07/24/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUEGER, ELIZABETH, A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: LINK, BRADLEY
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MOOSEN

DATE/TIME OF TEST: 07/10/2022 01:17:33

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.210	01:12
Breath Volume	1.507	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.220	01:15
Breath Volume	3.949	01:15
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 07/24/2022 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARKLUND,KARL,H
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: MARKLUND,KARL
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BOURGEOIS

DATE/TIME OF TEST: 07/13/2022 02:24:09

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 07/24/2022 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LA JEUNESSE,J,S
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: SULLIVAN,Z
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: MALDONADO RAMIREZ

DATE/TIME OF TEST: 07/17/2022 23:33:35

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:32	
		23:33	RFI Detect
Air Blank	0.000	23:33	

Data Download Date/Time: 07/24/2022 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LA JEUNESSE,J,S
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: SULLIVAN,Z
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: MALDONADO RAMIREZ

DATE/TIME OF TEST: 07/17/2022 23:38:35

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:36	
Control Test	0.082	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:37	Subject Test Refused
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Control Test	0.080	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 07/24/2022 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANT,DANIEL,F
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: GRANT,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/19/2022 11:11:43

Diagnostics Check	OK	11:05
Air Blank	0.000	11:05
Control Test	0.081	11:06
Air Blank	0.000	11:06
Subject Sample # 1	0.159	11:07
Breath Volume	4.023	11:07
Air Blank	0.000	11:07
Air Blank	0.000	11:09
Subject Sample # 2	0.162	11:10
Breath Volume	4.753	11:10
Air Blank	0.000	11:10
Control Test	0.080	11:11
Air Blank	0.000	11:11
Diagnostics Check	OK	11:11

Data Download Date/Time: 07/24/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARKLUND,KARL,H
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: FAUST,JOSHUA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/22/2022 04:39:11

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:30	
Control Test	0.079	04:30	
Air Blank	0.000	04:31	
Subject Sample	# 1 0.134	04:34	Volume Not Met
Breath Volume	0.515	04:34	
Air Blank	0.000	04:34	
Air Blank	0.000	04:36	
Subject Sample	# 2	04:37	Subject Test Refused
Breath Volume	1.078	04:37	
Air Blank	0.000	04:38	
Control Test	0.079	04:38	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 07/24/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN, GERALD, E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARRY, R
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: PERKIN

DATE/TIME OF TEST: 07/23/2022 01:46:58

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.081	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.126	01:42
Breath Volume	2.421	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:44
Subject Sample # 2	0.122	01:45
Breath Volume	1.550	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 07/24/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DOUG,J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: ROBERTS,DOUG
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 07/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: DILLION

DATE/TIME OF TEST: 07/24/2022 15:38:06

Diagnostics Check	OK	15:27
Air Blank	0.000	15:27
Control Test	0.080	15:27
Air Blank	0.000	15:28
Subject Sample # 1	0.276	15:31
Breath Volume	1.402	15:31
Air Blank	0.000	15:32
Air Blank	0.000	15:33
Subject Sample # 2	0.264	15:36
Breath Volume	1.988	15:36
Air Blank	0.000	15:37
Control Test	0.079	15:37
Air Blank	0.000	15:38
Diagnostics Check	OK	15:38

Data Download Date/Time: 08/29/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLDHAM,TAYLOR,D
BREATH TEST OPERATOR AGENCY: Mascotte PD
ARRESTING OFFICER: OLDHAM,TAYLOR
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 07/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: DIGGS

DATE/TIME OF TEST: 07/25/2022 16:54:48

Diagnostics Check	OK	16:43	
Air Blank	0.000	16:43	
Control Test	0.081	16:44	
Air Blank	0.000	16:44	
Subject Sample	# 1	16:47	No Sample Provided
Breath Volume	0.000	16:47	
Air Blank	0.000	16:48	
Air Blank	0.000	16:50	
Subject Sample	# 2	16:53	No Sample Provided
Breath Volume	0.000	16:53	
Air Blank	0.000	16:53	
Control Test	0.081	16:54	
Air Blank	0.000	16:54	
Diagnostics Check	OK	16:54	

Data Download Date/Time: 08/29/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LA JEUNESSE,J,S
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: WHITAKER,M
ARRESTING OFFICER AGENCY: Howey-In-The-Hills
LAST AGENCY INSPECTION: 07/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/25/2022 23:26:13

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:23	
Control Test	0.081	23:24	
Air Blank	0.000	23:24	
Subject Sample	# 1	23:24	Subject Test Refused
Breath Volume	0.000	23:24	
Air Blank	0.000	23:25	
Control Test	0.080	23:25	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 08/29/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: KESSINGER,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/28/2022 06:41:20

Diagnostics Check	OK	06:34
Air Blank	0.000	06:35
Control Test	0.081	06:35
Air Blank	0.000	06:36
Subject Sample # 1	0.000	06:37
Breath Volume	1.585	06:37
Air Blank	0.000	06:37
Air Blank	0.000	06:39
Subject Sample # 2	0.000	06:39
Breath Volume	2.695	06:39
Air Blank	0.000	06:40
Control Test	0.080	06:40
Air Blank	0.000	06:41
Diagnostics Check	OK	06:41

Data Download Date/Time: 08/29/2022 15:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUCZKA,JACOB,C
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: KUCZKA,JACOB
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/28/2022 23:26:33

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:15	
Control Test	0.081	23:15	
Air Blank	0.000	23:16	
Subject Sample	# 1 0.116	23:19	Volume Not Met
Breath Volume	0.363	23:19	
Air Blank	0.000	23:20	
Air Blank	0.000	23:21	
Subject Sample	# 2 0.107	23:24	Volume Not Met
Breath Volume	0.148	23:24	
Air Blank	0.000	23:25	
Control Test	0.080	23:25	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 08/29/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONDAHL,CARL,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ALLRED,JAMES
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: BOWENS

DATE/TIME OF TEST: 07/29/2022 17:59:36

Diagnostics Check	OK	17:47	
Air Blank	0.000	17:48	
Control Test	0.081	17:48	
Air Blank	0.000	17:49	
Subject Sample	# 1 0.149	17:52	Volume Not Met
Breath Volume	0.304	17:52	
Air Blank	0.000	17:52	
Air Blank	0.000	17:54	
Subject Sample	# 2 0.129	17:57	Volume Not Met
Breath Volume	0.183	17:57	
Air Blank	0.000	17:58	
Control Test	0.078	17:59	
Air Blank	0.000	17:59	
Diagnostics Check	OK	17:59	

Data Download Date/Time: 08/29/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONDAHL,CARL,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ALLRED,JAMES
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: BOWENS

DATE/TIME OF TEST: 07/29/2022 18:10:02

Diagnostics Check	OK	18:02
Air Blank	0.000	18:03
Control Test	0.078	18:03
Air Blank	0.000	18:03
Subject Sample # 1	0.146	18:04
Breath Volume	1.183	18:04
Air Blank	0.000	18:05
Air Blank	0.000	18:07
Subject Sample # 2	0.147	18:08
Breath Volume	1.304	18:08
Air Blank	0.000	18:09
Control Test	0.078	18:09
Air Blank	0.000	18:09
Diagnostics Check	OK	18:10

Data Download Date/Time: 08/29/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANSOM,JACK,A
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BILLUE,MOLLY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: RUIZ DE LEON

DATE/TIME OF TEST: 07/02/2022 03:10:57

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.172	03:06
Breath Volume	1.718	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:08
Subject Sample # 2	0.158	03:09
Breath Volume	1.468	03:09
Air Blank	0.000	03:10
Control Test	0.078	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 07/25/2022 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANSOM,JACK,A
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: RANSOM,JACK
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: GARRIDO

DATE/TIME OF TEST: 07/04/2022 19:45:30

Diagnostics Check	OK	19:30	
Air Blank	0.000	19:31	
Control Test	0.079	19:31	
Air Blank	0.000	19:31	
Subject Sample	# 1 0.109	19:35	Volume Not Met
Breath Volume	0.835	19:35	
Air Blank	0.000	19:35	
Air Blank	0.000	19:37	
Subject Sample	# 2 0.124	19:37	
Breath Volume	1.527	19:37	
Air Blank	0.000	19:38	
Air Blank	0.000	19:40	
Subject Sample	# 3 0.102	19:43	Volume Not Met
Breath Volume	0.167	19:43	
Air Blank	0.000	19:44	
Control Test	0.077	19:44	
Air Blank	0.000	19:45	
Diagnostics Check	OK	19:45	

Data Download Date/Time: 07/25/2022 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILLUE,MOLLY,S
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: WOOLEYHAN,HEATHER
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 07/10/2022 03:17:09

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.175	03:08
Breath Volume	2.242	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.150	03:12
Breath Volume	1.609	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 3	0.170	03:15
Breath Volume	1.804	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 07/25/2022 18:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILLUE,MOLLY,S
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BILLUE,MOLLY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/16/2022 03:14:01

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.209	03:09
Breath Volume	3.859	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.205	03:12
Breath Volume	3.792	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:14

Data Download Date/Time: 07/25/2022 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILLUE,MOLLY,S
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BILLUE,MOLLY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: CLEMONS

DATE/TIME OF TEST: 07/17/2022 03:54:31

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.179	03:49
Breath Volume	2.527	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.168	03:52
Breath Volume	1.492	03:52
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 07/25/2022 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JACOB,M
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: WALCUTT,BRIAN
ARRESTING OFFICER AGENCY: Volusia County Beach
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MOSIER

DATE/TIME OF TEST: 07/22/2022 23:41:42

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.080	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.007	23:37
Breath Volume	3.273	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.007	23:40
Breath Volume	2.023	23:40
Air Blank	0.000	23:40
Control Test	0.079	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 07/25/2022 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILLUE,MOLLY,S
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: SWEENEY,PARKER
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 07/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GRAZIANO

DATE/TIME OF TEST: 07/31/2022 01:37:09

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.220	01:32
Breath Volume	1.949	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.214	01:35
Breath Volume	1.296	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 08/15/2022 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEGER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SCHWALL,MATTHEW
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 06/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 16:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 07/06/2022 18:03:52

Diagnostics Check	OK	17:57
Air Blank	0.000	17:58
Control Test	0.077	17:58
Air Blank	0.000	17:59
Subject Sample # 1	0.149	17:59
Breath Volume	1.679	17:59
Air Blank	0.000	18:00
Air Blank	0.000	18:01
Subject Sample # 2	0.148	18:02
Breath Volume	1.621	18:02
Air Blank	0.000	18:02
Control Test	0.076	18:03
Air Blank	0.000	18:03
Diagnostics Check	OK	18:03

Data Download Date/Time: 07/31/2022 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVILA,NICOLE,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MILLS,JACOB
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/01/2022 02:48:40

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.082	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.200	02:44
Breath Volume	2.277	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.184	02:47
Breath Volume	1.750	02:47
Air Blank	0.000	02:47
Control Test	0.081	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 07/05/2022 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCBRIDE,CRYSTA,C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MCBRIDE,CRYSA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/02/2022 04:06:50

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.081	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1 0.178	03:55	
Breath Volume	1.570	03:55	
Air Blank	0.000	03:56	
Air Blank	0.000	03:58	
Subject Sample	# 2 0.185	04:01	Volume Not Met
Breath Volume	0.695	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 3 0.190	04:05	
Breath Volume	1.843	04:05	
Air Blank	0.000	04:05	
Control Test	0.078	04:06	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 07/05/2022 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHNIBBEN,BRIAN,K
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: ALEVISATOS,JARRAD
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 07/06/2022 00:48:41

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.160	00:43
Breath Volume	1.410	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.152	00:47
Breath Volume	1.257	00:47
Air Blank	0.000	00:47
Control Test	0.081	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 07/29/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOVANOVICH, THOMAS, E
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MEISLAHN, E
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: STITH

DATE/TIME OF TEST: 07/07/2022 20:18:48

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.258	20:13
Breath Volume	1.402	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.267	20:17
Breath Volume	1.171	20:17
Air Blank	0.000	20:17
Control Test	0.081	20:18
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 07/29/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVILA,NICOLE,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MILLS,JACOB
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: STRUBBE

DATE/TIME OF TEST: 07/09/2022 17:55:14

Diagnostics Check	OK	17:49
Air Blank	0.000	17:49
Control Test	0.081	17:49
Air Blank	0.000	17:50
Subject Sample # 1	0.209	17:50
Breath Volume	2.281	17:50
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 2	0.217	17:53
Breath Volume	2.339	17:53
Air Blank	0.000	17:54
Control Test	0.081	17:54
Air Blank	0.000	17:55
Diagnostics Check	OK	17:55

Data Download Date/Time: 07/29/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCBRIDE,CRYSTA,C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: REAHL,TIMOTHY
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 07/20/2022 15:08:11

Diagnostics Check	OK	15:00
Air Blank	0.000	15:01
Control Test	0.081	15:01
Air Blank	0.000	15:01
Subject Sample # 1	0.051	15:02
Breath Volume	1.207	15:02
Air Blank	0.000	15:03
Air Blank	0.000	15:05
Subject Sample # 2	0.052	15:06
Breath Volume	2.113	15:06
Air Blank	0.000	15:07
Control Test	0.082	15:07
Air Blank	0.000	15:08
Diagnostics Check	OK	15:08

Data Download Date/Time: 07/29/2022 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOVANOVICH, THOMAS, E
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STOKER, R
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: MEALOR

DATE/TIME OF TEST: 07/21/2022 19:54:23

Diagnostics Check	OK	19:48	
Air Blank	0.000	19:48	
Control Test	0.081	19:49	
Air Blank	0.000	19:49	
Subject Sample # 1	0.203	19:50	
Breath Volume	2.656	19:50	
Air Blank	0.000	19:51	
Air Blank		19:53	RFI Detect
Air Blank	0.000	19:53	
Control Test	0.081	19:53	
Air Blank	0.000	19:54	
Diagnostics Check	OK	19:54	

Data Download Date/Time: 08/04/2022 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOVANOVICH, THOMAS, E
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STOKER, R
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: BUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: MEALOR

DATE/TIME OF TEST: 07/21/2022 20:04:32

Diagnostics Check	OK	19:58
Air Blank	0.000	19:58
Control Test	0.080	19:58
Air Blank	0.000	19:59
Subject Sample # 1	0.196	20:00
Breath Volume	2.187	20:00
Air Blank	0.000	20:00
Air Blank	0.000	20:02
Subject Sample # 2	0.200	20:03
Breath Volume	2.027	20:03
Air Blank	0.000	20:03
Control Test	0.080	20:03
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 08/04/2022 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHNIBBEN,BRIAN,K
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MCDOWELL,ADRIAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LEGAULT

DATE/TIME OF TEST: 07/24/2022 02:45:38

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.092	02:41
Breath Volume	1.898	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.093	02:44
Breath Volume	2.125	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 08/04/2022 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHNIBBEN,BRIAN,K
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MCDOWELL,ADRIAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24220808A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: MASSARO

DATE/TIME OF TEST: 07/27/2022 22:07:22

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.081	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.071	22:02
Breath Volume	1.769	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.074	22:05
Breath Volume	2.253	22:05
Air Blank	0.000	22:06
Control Test	0.081	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 08/04/2022 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CLABAUGH,ARIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: CALLAM

DATE/TIME OF TEST: 07/01/2022 03:37:39

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.184	03:32
Breath Volume	2.554	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.178	03:35
Breath Volume	1.976	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/05/2022 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: FONSECA ROONEY

DATE/TIME OF TEST: 07/01/2022 05:34:14

Diagnostics Check	OK	05:28
Air Blank	0.000	05:29
Control Test	0.082	05:29
Air Blank	0.000	05:29
Subject Sample # 1	0.000	05:30
Breath Volume	1.488	05:30
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.000	05:32
Breath Volume	1.261	05:32
Air Blank	0.000	05:33
Control Test	0.081	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 07/05/2022 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KIPP,ALEXANDER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SOTO TRUJILLO

DATE/TIME OF TEST: 07/01/2022 22:43:55

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.081	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.134	22:37
Breath Volume	1.468	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.126	22:42
Breath Volume	1.519	22:42
Air Blank	0.000	22:42
Control Test	0.081	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 07/05/2022 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOREL,ABIGAIL
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/02/2022 00:52:34

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.082	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.083	00:48
Breath Volume	2.152	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.083	00:50
Breath Volume	1.878	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 07/05/2022 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PETERSON,THOMAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: VASS

DATE/TIME OF TEST: 07/02/2022 02:58:26

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.081	02:42	
Air Blank	0.000	02:43	
Subject Sample	# 1 0.092	02:46	Volume Not Met
Breath Volume	0.464	02:46	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 2 0.118	02:50	
Breath Volume	1.464	02:50	
Air Blank	0.000	02:51	
Air Blank	0.000	02:53	
Subject Sample	# 3 0.105	02:56	Volume Not Met
Breath Volume	0.269	02:56	
Air Blank	0.000	02:57	
Control Test	0.078	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 07/05/2022 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIESENFELT,JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: GOMEZ ALVIZURES

DATE/TIME OF TEST: 07/02/2022 04:48:20

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.163	04:43
Breath Volume	2.078	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.165	04:46
Breath Volume	2.113	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 07/05/2022 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: BENITEZ-SIFONTES,RAFAEL
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 07/23/2022 02:16:23

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.079	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.079	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 08/04/2022 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: DENHAM,TREVOR
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MATEO

DATE/TIME OF TEST: 07/23/2022 03:20:57

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.271	03:15
Breath Volume	1.933	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.284	03:19
Breath Volume	2.601	03:19
Air Blank	0.000	03:20
Control Test	0.078	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 08/04/2022 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: BENITEZ-SIFONTES,RAFAEL
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RAMIREZ-PEREZ

DATE/TIME OF TEST: 07/24/2022 03:27:10

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.143	03:22
Breath Volume	2.480	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.152	03:25
Breath Volume	1.554	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 08/04/2022 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAKEMAN,TAYLOR,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LEITCH,RICKFORD
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: OTHER
ARREST TIME: 10:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:48

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 07/26/2022 11:20:26

Diagnostics Check	OK	11:14
Air Blank	0.000	11:14
Control Test	0.079	11:14
Air Blank	0.000	11:15
Subject Sample # 1	0.000	11:16
Breath Volume	2.515	11:16
Air Blank	0.000	11:16
Air Blank	0.000	11:18
Subject Sample # 2	0.000	11:19
Breath Volume	2.851	11:19
Air Blank	0.000	11:19
Control Test	0.080	11:19
Air Blank	0.000	11:20
Diagnostics Check	OK	11:20

Data Download Date/Time: 08/04/2022 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SCHALLON,CHRISTIAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:41

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 07/23/2022 16:09:07

Diagnostics Check	OK	16:02
Air Blank	0.000	16:02
Control Test	0.080	16:03
Air Blank	0.000	16:03
Subject Sample # 1	0.083	16:04
Breath Volume	1.964	16:04
Air Blank	0.000	16:05
Air Blank	0.000	16:07
Subject Sample # 2	0.084	16:07
Breath Volume	2.152	16:07
Air Blank	0.000	16:08
Control Test	0.079	16:08
Air Blank	0.000	16:09
Diagnostics Check	OK	16:09

Data Download Date/Time: 08/16/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANG,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DUFT,KOYIE
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220193
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/28/2022 23:51:20

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.212	23:45
Breath Volume	1.289	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.197	23:49
Breath Volume	1.664	23:49
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 08/16/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTTO,ALISON,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: OTTO,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 07/03/2022 18:11:30

Diagnostics Check	OK	18:04
Air Blank	0.000	18:04
Control Test	0.081	18:05
Air Blank	0.000	18:05
Subject Sample # 1	0.237	18:06
Breath Volume	2.140	18:06
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.234	18:09
Breath Volume	1.882	18:09
Air Blank	0.000	18:10
Control Test	0.078	18:10
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 07/28/2022 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,H
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DURGIN,S
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: GIAMBACO

DATE/TIME OF TEST: 07/04/2022 14:51:14

Diagnostics Check	OK	14:43
Air Blank	0.000	14:44
Control Test	0.080	14:44
Air Blank	0.000	14:44
Subject Sample # 1	0.000	14:46
Breath Volume	3.109	14:46
Air Blank	0.000	14:46
Air Blank	0.000	14:48
Subject Sample # 2	0.000	14:49
Breath Volume	1.554	14:49
Air Blank	0.000	14:50
Control Test	0.080	14:50
Air Blank	0.000	14:51
Diagnostics Check	OK	14:51

Data Download Date/Time: 07/28/2022 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,FELICIA,N
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MORRIS,COURTNEY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: BRADSHAW

DATE/TIME OF TEST: 07/05/2022 17:59:04

Diagnostics Check	OK	17:50
Air Blank	0.000	17:51
Control Test	0.080	17:51
Air Blank	0.000	17:52
Subject Sample # 1	0.352	17:54
Breath Volume	2.355	17:54
Air Blank	0.000	17:55
Air Blank	0.000	17:56
Subject Sample # 2	0.350	17:57
Breath Volume	1.636	17:57
Air Blank	0.000	17:58
Control Test	0.077	17:58
Air Blank	0.000	17:58
Diagnostics Check	OK	17:59

Data Download Date/Time: 07/28/2022 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUBEK,STEPHEN,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: OTTO,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: CROZIER

DATE/TIME OF TEST: 07/12/2022 10:57:07

Diagnostics Check	OK	10:50
Air Blank	0.000	10:50
Control Test	0.080	10:51
Air Blank	0.000	10:51
Subject Sample # 1	0.099	10:52
Breath Volume	2.140	10:52
Air Blank	0.000	10:53
Air Blank	0.000	10:54
Subject Sample # 2	0.107	10:55
Breath Volume	4.093	10:55
Air Blank	0.000	10:56
Control Test	0.080	10:56
Air Blank	0.000	10:57
Diagnostics Check	OK	10:57

Data Download Date/Time: 07/28/2022 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,BRIAN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: LEONARD,BRIAN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: ORSINI

DATE/TIME OF TEST: 07/12/2022 17:25:51

Diagnostics Check	OK	17:18
Air Blank	0.000	17:19
Control Test	0.080	17:19
Air Blank	0.000	17:20
Subject Sample # 1	0.209	17:21
Breath Volume	1.945	17:21
Air Blank	0.000	17:21
Air Blank	0.000	17:23
Subject Sample # 2	0.210	17:24
Breath Volume	2.558	17:24
Air Blank	0.000	17:24
Control Test	0.078	17:25
Air Blank	0.000	17:25
Diagnostics Check	OK	17:25

Data Download Date/Time: 07/28/2022 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ, ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: FERINO, WESLEY
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: TAVERAS

DATE/TIME OF TEST: 07/17/2022 01:51:20

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.160	01:46
Breath Volume	1.570	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.162	01:49
Breath Volume	1.535	01:49
Air Blank	0.000	01:50
Control Test	0.078	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 07/28/2022 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDMAN,B,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SAVAGE,J
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/24/2022 03:57:53

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.169	03:53
Breath Volume	1.316	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample # 2	0.167	03:56
Breath Volume	1.726	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 07/28/2022 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,H
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DURGIN,SHAWN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 07/24/2022 08:10:45

Diagnostics Check	OK	08:04
Air Blank	0.000	08:04
Control Test	0.081	08:05
Air Blank	0.000	08:05
Subject Sample # 1	0.138	08:06
Breath Volume	1.613	08:06
Air Blank	0.000	08:06
Air Blank	0.000	08:08
Subject Sample # 2	0.150	08:09
Breath Volume	4.015	08:09
Air Blank	0.000	08:09
Control Test	0.079	08:10
Air Blank	0.000	08:10
Diagnostics Check	OK	08:10

Data Download Date/Time: 07/28/2022 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 07/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME:]

DATE/TIME OF TEST: 07/29/2022 00:45:43

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.081	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.235	00:41
Breath Volume	2.261	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.238	00:44
Breath Volume	2.144	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 08/12/2022 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTTO,ALISON,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: OTTO,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 07/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 12:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/31/2022 13:25:39

Diagnostics Check	OK	13:14	
Air Blank	0.000	13:15	
Control Test	0.081	13:15	
Air Blank	0.000	13:15	
Subject Sample	# 1 0.000	13:19	Volume Not Met
Breath Volume	0.488	13:19	
Air Blank	0.000	13:19	
Air Blank	0.000	13:21	
Subject Sample	# 2	13:24	Subject Test Refused
Breath Volume	0.000	13:24	
Air Blank	0.000	13:24	
Control Test	0.080	13:25	
Air Blank	0.000	13:25	
Diagnostics Check	OK	13:25	

Data Download Date/Time: 08/12/2022 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTTO,ALISON,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: OTTO,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 07/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 12:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 07/31/2022 13:38:41

Diagnostics Check	OK	13:32
Air Blank	0.000	13:32
Control Test	0.080	13:32
Air Blank	0.000	13:33
Subject Sample # 1	0.000	13:34
Breath Volume	1.195	13:34
Air Blank	0.000	13:35
Air Blank	0.000	13:37
Subject Sample # 2	0.000	13:37
Breath Volume	1.347	13:37
Air Blank	0.000	13:37
Control Test	0.079	13:38
Air Blank	0.000	13:38
Diagnostics Check	OK	13:38

Data Download Date/Time: 08/12/2022 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WRIGHT,TIMOTHY
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 07/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211583
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BELLANOVA

DATE/TIME OF TEST: 07/17/2022 02:26:01

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.135	02:21
Breath Volume	2.601	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.129	02:24
Breath Volume	2.167	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:26

Data Download Date/Time: 08/21/2022 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 07/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211583
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 07/22/2022 03:36:10

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.233	03:31
Breath Volume	1.800	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.220	03:34
Breath Volume	1.945	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 08/21/2022 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW,C,C
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: HELBIG,D
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211583
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SZCZEPANSKI

DATE/TIME OF TEST: 07/30/2022 04:03:52

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Subject Sample # 1	0.214	03:53	
Breath Volume	3.945	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample # 2		03:59	No Sample Provided
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Air Blank	0.000	04:01	
Subject Sample # 3		04:02	Slope Not Met
Breath Volume	2.781	04:02	
Air Blank	0.000	04:02	
Control Test	0.079	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 08/21/2022 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW,C,C
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: HELBIG,D
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211583
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SZCZEPANSKI

DATE/TIME OF TEST: 07/30/2022 04:11:10

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.080	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Control Test	0.079	04:10	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 08/21/2022 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HACKETT,BRADLEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: HEINEN

DATE/TIME OF TEST: 07/01/2022 16:54:15

Diagnostics Check	OK	16:48
Air Blank	0.000	16:48
Control Test	0.079	16:48
Air Blank	0.000	16:49
Subject Sample # 1	0.000	16:50
Breath Volume	2.390	16:50
Air Blank	0.000	16:50
Air Blank	0.000	16:52
Subject Sample # 2	0.000	16:52
Breath Volume	1.441	16:52
Air Blank	0.000	16:53
Control Test	0.079	16:53
Air Blank	0.000	16:54
Diagnostics Check	OK	16:54

Data Download Date/Time: 07/26/2022 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/04/2022 05:21:26

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.081	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.206	05:16
Breath Volume	3.179	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.192	05:19
Breath Volume	4.332	05:19
Air Blank	0.000	05:20
Control Test	0.075	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 07/26/2022 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: KING, BRIAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 07/08/2022 03:17:02

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:13	
Control Test	0.079	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:15	RFI Detect
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 07/26/2022 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACHELLER,ALEXANDRIA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BACHELLER,ALEXANDRIA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MORALES FRANCO

DATE/TIME OF TEST: 07/17/2022 04:00:01

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.168	03:55
Breath Volume	3.394	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.170	03:58
Breath Volume	2.824	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	04:00

Data Download Date/Time: 07/26/2022 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: STARNER

DATE/TIME OF TEST: 07/21/2022 04:00:37

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1 0.195	03:47	
Breath Volume	1.144	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:50	
Subject Sample	# 2 0.178	03:53	Volume Not Met
Breath Volume	0.019	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 3 0.182	03:59	Volume Not Met
Breath Volume	0.132	03:59	
Air Blank	0.000	03:59	
Control Test	0.078	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 07/26/2022 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: STARNER

DATE/TIME OF TEST: 07/21/2022 04:19:01

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.078	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1 0.189	04:06	
Breath Volume	1.246	04:06	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 2 0.161	04:11	Volume Not Met
Breath Volume	0.207	04:11	
Air Blank	0.000	04:12	
Air Blank	0.000	04:14	
Subject Sample	# 3 0.153	04:17	Volume Not Met
Breath Volume	0.222	04:17	
Air Blank	0.000	04:18	
Control Test	0.076	04:18	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 07/26/2022 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, TYLER, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BURKETT, RICHARD
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 07/24/2022 22:22:35

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.080	22:09	
Air Blank	0.000	22:09	
Subject Sample	# 1 0.058	22:12	Volume Not Met
Breath Volume	0.914	22:12	
Air Blank	0.000	22:13	
Air Blank	0.000	22:15	
Subject Sample	# 2 0.067	22:17	
Breath Volume	2.195	22:17	
Air Blank	0.000	22:18	
Air Blank	0.000	22:20	
Subject Sample	# 3 0.066	22:21	
Breath Volume	2.601	22:21	
Air Blank	0.000	22:21	
Control Test	0.078	22:21	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 07/26/2022 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, TYLER, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BURKETT, RICHARD
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: SHEPARD

DATE/TIME OF TEST: 07/29/2022 00:19:30

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.080	00:04	
Air Blank	0.000	00:04	
Subject Sample	# 1 0.143	00:07	Volume Not Met
Breath Volume	0.015	00:07	
Air Blank	0.000	00:08	
Air Blank	0.000	00:10	
Subject Sample	# 2 0.156	00:12	
Breath Volume	1.210	00:12	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample	# 3 0.149	00:17	Volume Not Met
Breath Volume	0.339	00:17	
Air Blank	0.000	00:18	
Control Test	0.077	00:18	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:19	

Data Download Date/Time: 08/25/2022 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 07/30/2022 03:06:13

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.105	03:01
Breath Volume	2.558	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.105	03:04
Breath Volume	2.535	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 08/25/2022 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: CAMPAU

DATE/TIME OF TEST: 07/31/2022 01:53:12

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.081	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.249	01:48
Breath Volume	4.914	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.249	01:51
Breath Volume	4.390	01:51
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 08/25/2022 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HERRON

DATE/TIME OF TEST: 07/31/2022 04:32:15

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.000	04:28
Breath Volume	2.230	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.000	04:30
Breath Volume	1.878	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 08/25/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BIGWOOD

DATE/TIME OF TEST: 07/01/2022 00:26:37

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.079	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.179	00:18
Breath Volume	2.351	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.149	00:22
Breath Volume	1.562	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 3	0.171	00:25
Breath Volume	2.097	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 07/31/2022 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 21:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: BURSEY

DATE/TIME OF TEST: 07/01/2022 22:00:32

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:48	
Control Test	0.079	21:49	
Air Blank	0.000	21:49	
Subject Sample	# 1 0.141	21:52	Volume Not Met
Breath Volume	0.144	21:52	
Air Blank	0.000	21:53	
Air Blank	0.000	21:55	
Subject Sample	# 2 0.136	21:55	
Breath Volume	2.386	21:55	
Air Blank	0.000	21:56	
Air Blank	0.000	21:58	
Subject Sample	# 3 0.127	21:59	
Breath Volume	1.320	21:59	
Air Blank	0.000	21:59	
Control Test	0.079	21:59	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

Data Download Date/Time: 07/31/2022 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:05

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 07/08/2022 02:20:28

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.171	02:16
Breath Volume	2.101	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.158	02:18
Breath Volume	1.625	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 07/31/2022 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BERUBE

DATE/TIME OF TEST: 07/20/2022 02:01:21

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.078	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.093	01:56
Breath Volume	1.339	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.099	01:59
Breath Volume	1.976	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 07/31/2022 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GREENLEE

DATE/TIME OF TEST: 07/21/2022 02:26:19

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.230	02:21
Breath Volume	1.511	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.225	02:24
Breath Volume	1.925	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 07/31/2022 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 07/22/2022 01:30:19

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.143	01:24
Breath Volume	1.792	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.127	01:28
Breath Volume	2.277	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 07/31/2022 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 07/27/2022 00:23:42

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.095	00:19
Breath Volume	2.207	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.090	00:22
Breath Volume	2.109	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 07/31/2022 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERVAIS, ROSEMARY, C
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANCHEZ, J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MORALES RIVERA

DATE/TIME OF TEST: 07/02/2022 04:14:17

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.082	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.171	04:09
Breath Volume	1.769	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.179	04:12
Breath Volume	2.585	04:12
Air Blank	0.000	04:13
Control Test	0.081	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 07/16/2022 23:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLAZINA,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: LUMENE

DATE/TIME OF TEST: 07/02/2022 23:00:25

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.123	22:55
Breath Volume	1.691	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.119	22:58
Breath Volume	1.589	22:58
Air Blank	0.000	22:59
Control Test	0.080	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 07/16/2022 23:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: JANKAY,ASHLEY
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: PROB/PAROL
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: TORO

DATE/TIME OF TEST: 07/03/2022 02:35:59

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.081	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.297	02:29
Breath Volume	1.523	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.292	02:34
Breath Volume	1.277	02:34
Air Blank	0.000	02:34
Control Test	0.077	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 07/16/2022 23:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PFANNERSTILL,JACOB
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: DELAROSA JIMENEZ

DATE/TIME OF TEST: 07/03/2022 05:06:43

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:05	
Control Test	0.024	05:06	Control Outside Tolerance
Air Blank	0.000	05:06	

Data Download Date/Time: 07/16/2022 23:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PFANNERSTILL,JACOB
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: DELAROSA JIMENEZ

DATE/TIME OF TEST: 07/03/2022 05:17:18

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.082	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.147	05:12
Breath Volume	1.617	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.148	05:15
Breath Volume	2.140	05:15
Air Blank	0.000	05:16
Control Test	0.082	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 07/16/2022 23:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FORTENBERRY,BRANDON
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/08/2022 23:15:50

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.081	23:04	
Air Blank	0.000	23:05	
Subject Sample	# 1 0.120	23:08	Volume Not Met
Breath Volume	0.085	23:08	
Air Blank	0.000	23:09	
Air Blank	0.000	23:11	
Subject Sample	# 2 0.109	23:14	Volume Not Met
Breath Volume	0.085	23:14	
Air Blank	0.000	23:14	
Control Test	0.081	23:15	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 07/16/2022 23:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GANT,JARRET
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: WEISZ

DATE/TIME OF TEST: 07/09/2022 00:38:35

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.195	00:34
Breath Volume	1.203	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.202	00:37
Breath Volume	1.187	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 07/16/2022 23:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RODRIGUEZ,EMMANUEL
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: DI LELLA

DATE/TIME OF TEST: 07/09/2022 04:04:03

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.154	03:59
Breath Volume	4.035	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample # 2	0.143	04:02
Breath Volume	2.335	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:04

Data Download Date/Time: 07/16/2022 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FORTENBERRY,BRADON
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: AVALOS LEYVA

DATE/TIME OF TEST: 07/09/2022 04:22:57

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.151	04:16
Breath Volume	1.199	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.152	04:21
Breath Volume	1.519	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 07/16/2022 23:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Winter Springs PD
ARRESTING OFFICER: CASTRO,JONATHAN
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: KIRWAN

DATE/TIME OF TEST: 07/11/2022 19:37:45

Diagnostics Check	OK	19:29
Air Blank	0.000	19:29
Control Test	0.081	19:30
Air Blank	0.000	19:30
Subject Sample # 1	0.141	19:32
Breath Volume	1.207	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.136	19:36
Breath Volume	1.328	19:36
Air Blank	0.000	19:36
Control Test	0.080	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 07/16/2022 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANCHEZ,JUBILEE
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 07/11/2022 23:15:24

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:12	
Control Test	0.081	23:13	
Air Blank	0.000	23:13	
Subject Sample	# 1	23:14	Subject Test Refused
Breath Volume	0.000	23:14	
Air Blank	0.000	23:14	
Control Test	0.081	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 07/16/2022 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PONDER,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: RADI

DATE/TIME OF TEST: 07/16/2022 00:44:36

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.082	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.227	00:39
Breath Volume	1.191	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.231	00:42
Breath Volume	2.945	00:42
Air Blank	0.000	00:43
Control Test	0.081	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 07/16/2022 23:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLAZINA,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MCKESEY

DATE/TIME OF TEST: 07/16/2022 02:50:27

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.082	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.196	02:45
Breath Volume	1.226	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.195	02:48
Breath Volume	1.253	02:48
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 07/16/2022 23:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GIOIELLI,ANTHONY
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: MCLARRY

DATE/TIME OF TEST: 07/17/2022 00:02:46

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.290	23:58
Breath Volume	3.691	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.290	00:01
Breath Volume	3.500	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 08/29/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UCELO,BRYAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: VENTURA VELASCO

DATE/TIME OF TEST: 07/18/2022 01:01:43

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.082	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.267	00:57
Breath Volume	1.738	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.273	01:00
Breath Volume	2.257	01:00
Air Blank	0.000	01:00
Control Test	0.082	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 08/29/2022 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RITTER,ROBERT
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KEATING

DATE/TIME OF TEST: 07/19/2022 00:37:29

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.082	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.257	00:32
Breath Volume	2.199	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.262	00:35
Breath Volume	2.292	00:35
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 08/29/2022 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: KENT

DATE/TIME OF TEST: 07/19/2022 03:26:54

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.083	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.585	03:25	
Air Blank	0.000	03:25	
Control Test	0.083	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 08/29/2022 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLAZINA,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FRIAS

DATE/TIME OF TEST: 07/21/2022 01:18:10

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.082	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.051	01:13
Breath Volume	2.007	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample # 2	0.051	01:16
Breath Volume	2.289	01:16
Air Blank	0.000	01:17
Control Test	0.082	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 08/29/2022 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RAMOS,MARK
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 07/23/2022 04:01:47

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.082	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.163	03:57
Breath Volume	1.957	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.179	04:00
Breath Volume	2.605	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 08/29/2022 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CANADY,CHRISTINA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/23/2022 04:48:26

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.082	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.117	04:44
Breath Volume	2.496	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.127	04:46
Breath Volume	1.851	04:46
Air Blank	0.000	04:47
Control Test	0.081	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 08/29/2022 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOFFETT,HALEY
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ESCALANTE

DATE/TIME OF TEST: 07/25/2022 02:08:47

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.081	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.168	02:04
Breath Volume	1.820	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.175	02:07
Breath Volume	2.429	02:07
Air Blank	0.000	02:07
Control Test	0.081	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 08/29/2022 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOFFETT,HALEY
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ESCALANTE

DATE/TIME OF TEST: 07/25/2022 03:04:13

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.081	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.140	02:59
Breath Volume	2.593	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.139	03:02
Breath Volume	2.503	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 08/29/2022 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SJOBERG

DATE/TIME OF TEST: 07/01/2022 01:51:39

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.000	01:47
Breath Volume	1.343	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.000	01:50
Breath Volume	1.375	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 07/01/2022 03:10:12

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.079	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 07/01/2022 03:18:47

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.078	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.147	03:14
Breath Volume	1.839	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.144	03:17
Breath Volume	1.457	03:17
Air Blank	0.000	03:17
Control Test	0.078	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: LABBE

DATE/TIME OF TEST: 07/01/2022 05:22:39

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.079	05:15
Air Blank	0.000	05:16
Subject Sample # 1	0.178	05:17
Breath Volume	1.105	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.194	05:20
Breath Volume	1.710	05:20
Air Blank	0.000	05:21
Control Test	0.078	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 07/14/2022 19:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EALY, STEVEN, T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORGAN, WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: ESCULANO

DATE/TIME OF TEST: 07/01/2022 17:03:50

Diagnostics Check	OK	16:50	
Air Blank	0.000	16:51	
Control Test	0.079	16:51	
Air Blank	0.000	16:51	
Subject Sample	# 1 0.210	16:54	Volume Not Met
Breath Volume	0.453	16:54	
Air Blank	0.000	16:55	
Air Blank	0.000	16:57	
Subject Sample	# 2 0.200	16:58	
Breath Volume	1.695	16:58	
Air Blank	0.000	16:59	
Air Blank	0.000	17:01	
Subject Sample	# 3 0.211	17:02	
Breath Volume	1.667	17:02	
Air Blank	0.000	17:02	
Control Test	0.075	17:03	
Air Blank	0.000	17:03	
Diagnostics Check	OK	17:03	

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/01/2022 22:13:41

Diagnostics Check	OK	22:07
Air Blank	0.000	22:08
Control Test	0.079	22:08
Air Blank	0.000	22:08
Subject Sample # 1	0.000	22:09
Breath Volume	2.652	22:09
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.000	22:12
Breath Volume	1.863	22:12
Air Blank	0.000	22:12
Control Test	0.078	22:13
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: JORDAN III

DATE/TIME OF TEST: 07/01/2022 23:24:05

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.079	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.152	23:19
Breath Volume	2.933	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.150	23:22
Breath Volume	1.835	23:22
Air Blank	0.000	23:23
Control Test	0.078	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ONEIL

DATE/TIME OF TEST: 07/02/2022 00:47:31

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.017	00:42
Breath Volume	2.398	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.018	00:45
Breath Volume	2.472	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HAZELLIEF

DATE/TIME OF TEST: 07/02/2022 01:45:39

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.079	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Control Test	0.079	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PACKARD

DATE/TIME OF TEST: 07/02/2022 04:03:04

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.078	03:59	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.079	04:02	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EALY,STEVEN,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DISHMEY,OBED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: SANDBERG

DATE/TIME OF TEST: 07/02/2022 17:40:54

Diagnostics Check	OK	17:32
Air Blank	0.000	17:33
Control Test	0.079	17:33
Air Blank	0.000	17:34
Subject Sample # 1	0.141	17:36
Breath Volume	2.082	17:36
Air Blank	0.000	17:36
Air Blank	0.000	17:38
Subject Sample # 2	0.141	17:39
Breath Volume	1.171	17:39
Air Blank	0.000	17:39
Control Test	0.078	17:40
Air Blank	0.000	17:40
Diagnostics Check	OK	17:40

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: HERRMANN

DATE/TIME OF TEST: 07/02/2022 23:35:17

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:33	Interferent Detect
Breath Volume	2.648	23:33	
Air Blank	0.000	23:34	
Control Test	0.078	23:34	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 12:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: BLAIR CLAROS

DATE/TIME OF TEST: 07/03/2022 01:36:17

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.000	01:32
Breath Volume	2.800	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.000	01:34
Breath Volume	1.964	01:34
Air Blank	0.000	01:35
Control Test	0.078	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: BENNER

DATE/TIME OF TEST: 07/03/2022 23:18:34

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.117	23:13
Breath Volume	2.097	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.117	23:16
Breath Volume	2.601	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BRYSON

DATE/TIME OF TEST: 07/04/2022 03:21:57

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.000	03:17
Breath Volume	2.570	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.000	03:20
Breath Volume	2.734	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/07/2022 22:24:07

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.079	22:21	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:23	
Control Test	0.079	22:23	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 07/08/2022 04:57:53

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.000	04:53
Breath Volume	3.015	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.000	04:56
Breath Volume	2.191	04:56
Air Blank	0.000	04:56
Control Test	0.078	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: WINKELMAN

DATE/TIME OF TEST: 07/09/2022 04:35:22

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.079	04:24	
Air Blank	0.000	04:24	
Subject Sample	# 1 0.000	04:27	Volume Not Met
Breath Volume	0.417	04:27	
Air Blank	0.000	04:28	
Air Blank	0.000	04:30	
Subject Sample	# 2 0.000	04:30	
Breath Volume	1.847	04:30	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 3 0.000	04:33	
Breath Volume	1.554	04:33	
Air Blank	0.000	04:34	
Control Test	0.078	04:34	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 07/09/2022 22:05:40

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:03	
Control Test	0.079	22:03	
Air Blank	0.000	22:04	
Subject Sample	# 1	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:04	
Control Test	0.079	22:05	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

Data Download Date/Time: 07/14/2022 19:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/10/2022 02:31:12

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.296	02:25
Breath Volume	2.019	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.310	02:29
Breath Volume	2.058	02:29
Air Blank	0.000	02:30
Control Test	0.078	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 07/14/2022 19:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: SCHENCK

DATE/TIME OF TEST: 07/10/2022 04:05:55

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.079	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.080	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 07/14/2022 19:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: LUNER

DATE/TIME OF TEST: 07/14/2022 23:02:27

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.000	22:57
Breath Volume	2.175	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.000	23:01
Breath Volume	2.335	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 08/12/2022 20:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,CHRISTOPHER,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BENTLEY,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: MONTAGUE

DATE/TIME OF TEST: 07/15/2022 16:52:31

Diagnostics Check	OK	16:48	
Air Blank	0.000	16:48	
Control Test	0.079	16:49	
Air Blank	0.000	16:49	
Subject Sample	# 1	16:51	Subject Test Refused
Breath Volume	0.000	16:51	
Air Blank	0.000	16:51	
Control Test	0.079	16:51	
Air Blank	0.000	16:52	
Diagnostics Check	OK	16:52	

Data Download Date/Time: 08/12/2022 20:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 07/16/2022 03:31:29

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.050	03:26
Breath Volume	1.238	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.048	03:29
Breath Volume	1.421	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 08/12/2022 20:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MONROY

DATE/TIME OF TEST: 07/16/2022 03:39:52

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.078	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.000	03:35
Breath Volume	2.191	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.000	03:38
Breath Volume	1.542	03:38
Air Blank	0.000	03:38
Control Test	0.078	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 08/12/2022 20:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 07/01/2022 00:24:56

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.081	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.162	00:19
Breath Volume	1.929	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.153	00:23
Breath Volume	1.828	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOPPE,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: BUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: TOMAINE

DATE/TIME OF TEST: 07/01/2022 20:12:19

Diagnostics Check	OK	20:05
Air Blank	0.000	20:06
Control Test	0.080	20:06
Air Blank	0.000	20:07
Subject Sample # 1	0.000	20:07
Breath Volume	1.292	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.000	20:10
Breath Volume	1.226	20:10
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PEREZ PENA

DATE/TIME OF TEST: 07/01/2022 22:11:02

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:01	
Control Test	0.081	22:01	
Air Blank	0.000	22:02	
Subject Sample	# 1 0.153	22:05	Volume Not Met
Breath Volume	0.328	22:05	
Air Blank	0.000	22:05	
Air Blank	0.000	22:07	
Subject Sample	# 2	22:09	Subject Test Refused
Breath Volume	0.410	22:09	
Air Blank	0.000	22:10	
Control Test	0.079	22:10	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:11	

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BELMONTE

DATE/TIME OF TEST: 07/01/2022 23:32:45

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.087	23:27
Breath Volume	1.308	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.089	23:31
Breath Volume	1.164	23:31
Air Blank	0.000	23:31
Control Test	0.079	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA'
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 07/02/2022 01:09:29

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.079	01:07	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:08	
Control Test	0.080	01:08	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/02/2022 04:21:54

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.211	04:17
Breath Volume	3.511	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:19
Subject Sample # 2	0.217	04:20
Breath Volume	2.855	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 07/02/2022 06:52:17

Diagnostics Check	OK	06:45
Air Blank	0.000	06:45
Control Test	0.080	06:45
Air Blank	0.000	06:46
Subject Sample # 1	0.175	06:47
Breath Volume	2.843	06:47
Air Blank	0.000	06:48
Air Blank	0.000	06:50
Subject Sample # 2	0.169	06:50
Breath Volume	2.320	06:50
Air Blank	0.000	06:51
Control Test	0.079	06:51
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: CHATMAN

DATE/TIME OF TEST: 07/03/2022 00:06:42

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:05	Subject Test Refused
Breath Volume	0.000	00:05	
Air Blank	0.000	00:05	
Control Test	0.080	00:06	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: MALECKAS

DATE/TIME OF TEST: 07/03/2022 01:39:32

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.173	01:35
Breath Volume	2.269	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.168	01:38
Breath Volume	2.332	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SHOULTA,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: VANNORMAN

DATE/TIME OF TEST: 07/03/2022 02:08:44

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.140	02:04
Breath Volume	2.226	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.140	02:07
Breath Volume	2.406	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 07/03/2022 04:09:20

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.115	04:04
Breath Volume	3.027	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.114	04:07
Breath Volume	2.917	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 07/14/2022 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/03/2022 04:38:57

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.079	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:34	No Sample Provided
Breath Volume	0.000	04:34	
Air Blank	0.000	04:35	
Air Blank	0.000	04:37	
Subject Sample	# 2	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:38	
Control Test	0.080	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ G, YOEL-ALEXIZ,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI, JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 07/03/2022 18:10:09

Diagnostics Check	OK	18:04
Air Blank	0.000	18:04
Control Test	0.079	18:04
Air Blank	0.000	18:05
Subject Sample # 1	0.084	18:05
Breath Volume	3.195	18:05
Air Blank	0.000	18:06
Air Blank	0.000	18:08
Subject Sample # 2	0.083	18:08
Breath Volume	3.222	18:08
Air Blank	0.000	18:09
Control Test	0.080	18:09
Air Blank	0.000	18:10
Diagnostics Check	OK	18:10

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VILLA,ALAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 07/04/2022 06:33:02

Diagnostics Check	OK	06:26
Air Blank	0.000	06:27
Control Test	0.080	06:27
Air Blank	0.000	06:28
Subject Sample # 1	0.018	06:28
Breath Volume	2.722	06:28
Air Blank	0.000	06:29
Air Blank	0.000	06:31
Subject Sample # 2	0.017	06:31
Breath Volume	2.957	06:31
Air Blank	0.000	06:32
Control Test	0.080	06:32
Air Blank	0.000	06:32
Diagnostics Check	OK	06:33

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 07/04/2022 20:08:48

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.080	20:02
Air Blank	0.000	20:03
Subject Sample # 1	0.215	20:04
Breath Volume	1.113	20:04
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.224	20:07
Breath Volume	1.894	20:07
Air Blank	0.000	20:07
Control Test	0.079	20:08
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PENNAMON
MUHAMMAD,DANDRE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male

OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: STRUBE

DATE/TIME OF TEST: 07/04/2022 21:31:31

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.079	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.180	21:27
Breath Volume	2.230	21:27
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.181	21:30
Breath Volume	2.417	21:30
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MENDEZARRIAGADA

DATE/TIME OF TEST: 07/05/2022 04:57:27

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.080	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.148	04:52
Breath Volume	2.472	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.150	04:55
Breath Volume	3.816	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MEZIERE

DATE/TIME OF TEST: 07/05/2022 22:04:20

Diagnostics Check	OK	21:59	
Air Blank	0.000	21:59	
Control Test	0.079	21:59	
Air Blank	0.000	22:00	
Subject Sample	# 1	22:02	Subject Test Refused
Breath Volume	0.000	22:02	
Air Blank	0.000	22:03	
Control Test	0.080	22:03	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:04	

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/06/2022 23:26:32

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.214	23:21
Breath Volume	2.644	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.197	23:25
Breath Volume	2.199	23:25
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: DEMOS

DATE/TIME OF TEST: 07/07/2022 01:52:20

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.175	01:47
Breath Volume	3.546	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.155	01:50
Breath Volume	1.785	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BYNUM IV

DATE/TIME OF TEST: 07/07/2022 03:52:02

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.145	03:47
Breath Volume	1.347	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.151	03:50
Breath Volume	1.175	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 07/08/2022 03:13:31

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.115	03:08
Breath Volume	1.351	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.118	03:12
Breath Volume	1.750	03:12
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOODY,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SOMROLL

DATE/TIME OF TEST: 07/08/2022 22:15:31

Diagnostics Check	OK	22:07
Air Blank	0.000	22:07
Control Test	0.080	22:08
Air Blank	0.000	22:08
Subject Sample # 1	0.247	22:10
Breath Volume	1.820	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.257	22:13
Breath Volume	1.523	22:13
Air Blank	0.000	22:14
Control Test	0.079	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 07/09/2022 04:07:49

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.163	04:03
Breath Volume	2.210	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.153	04:06
Breath Volume	1.261	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CAUDILL

DATE/TIME OF TEST: 07/09/2022 04:42:53

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.148	04:38
Breath Volume	1.414	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.151	04:41
Breath Volume	1.359	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: SIURU

DATE/TIME OF TEST: 07/09/2022 05:15:51

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:12	
Control Test	0.080	05:13	
Air Blank	0.000	05:13	
Subject Sample	# 1	05:14	Subject Test Refused
Breath Volume	0.000	05:14	
Air Blank	0.000	05:14	
Control Test	0.079	05:15	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/09/2022 05:36:23

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:26	
Control Test	0.080	05:26	
Air Blank	0.000	05:27	
Subject Sample	# 1 0.079	05:30	Volume Not Met
Breath Volume	0.035	05:30	
Air Blank	0.000	05:30	
Air Blank	0.000	05:32	
Subject Sample	# 2	05:34	Subject Test Refused
Breath Volume	0.000	05:34	
Air Blank	0.000	05:35	
Control Test	0.079	05:35	
Air Blank	0.000	05:36	
Diagnostics Check	OK	05:36	

Data Download Date/Time: 07/14/2022 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGSTROM,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STAPLES,JANNIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 12:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: SHOCKLEY

DATE/TIME OF TEST: 07/09/2022 13:25:22

Diagnostics Check	OK	13:21	
Air Blank	0.000	13:22	
Control Test	0.080	13:22	
Air Blank	0.000	13:23	
Subject Sample	# 1	13:24	Subject Test Refused
Breath Volume	0.000	13:24	
Air Blank	0.000	13:24	
Control Test	0.080	13:24	
Air Blank	0.000	13:25	
Diagnostics Check	OK	13:25	

Data Download Date/Time: 07/14/2022 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUGHES,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: LUBECKI

DATE/TIME OF TEST: 07/09/2022 20:01:17

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:58	
Control Test	0.079	19:59	
Air Blank	0.000	19:59	
Subject Sample	# 1	19:59	Subject Test Refused
Breath Volume	0.000	19:59	
Air Blank	0.000	20:00	
Control Test	0.080	20:00	
Air Blank	0.000	20:01	
Diagnostics Check	OK	20:01	

Data Download Date/Time: 07/14/2022 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 07/10/2022 01:58:18

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.152	01:53
Breath Volume	2.382	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.146	01:56
Breath Volume	1.464	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 07/14/2022 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: VASSALLO

DATE/TIME OF TEST: 07/10/2022 05:53:51

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.080	05:48
Air Blank	0.000	05:48
Subject Sample # 1	0.176	05:49
Breath Volume	1.453	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.176	05:52
Breath Volume	1.875	05:52
Air Blank	0.000	05:52
Control Test	0.080	05:53
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 07/14/2022 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROBB,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: BOBER

DATE/TIME OF TEST: 07/11/2022 01:50:44

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.000	01:46
Breath Volume	1.863	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.000	01:49
Breath Volume	1.843	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/14/2022 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FARRINGTON,C
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/11/2022 03:12:11

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.079	03:10	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.079	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 07/14/2022 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: WOLLAM

DATE/TIME OF TEST: 07/11/2022 20:15:21

Diagnostics Check	OK	20:12	
Air Blank	0.000	20:12	
Control Test	0.079	20:13	
Air Blank	0.000	20:13	
Subject Sample	# 1	20:14	Subject Test Refused
Breath Volume	0.000	20:14	
Air Blank	0.000	20:14	
Control Test	0.079	20:14	
Air Blank	0.000	20:15	
Diagnostics Check	OK	20:15	

Data Download Date/Time: 07/14/2022 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: K524072936030

DATE/TIME OF TEST: 07/12/2022 03:16:53

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.080	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.080	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 07/14/2022 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMIN,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CANSLER

DATE/TIME OF TEST: 07/12/2022 03:39:10

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.078	03:36	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.079	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 07/14/2022 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: WINDSOR

DATE/TIME OF TEST: 07/12/2022 05:04:23

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:00	
Control Test	0.080	05:01	
Air Blank	0.000	05:01	
Subject Sample	# 1	05:03	Subject Test Refused
Breath Volume	0.000	05:03	
Air Blank	0.000	05:03	
Control Test	0.080	05:03	
Air Blank	0.000	05:04	
Diagnostics Check	OK	05:04	

Data Download Date/Time: 07/14/2022 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 07/13/2022 01:41:26

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.195	01:37
Breath Volume	2.968	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.183	01:39
Breath Volume	1.746	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 07/14/2022 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: GARCIA-WANSLEY

DATE/TIME OF TEST: 07/13/2022 05:05:35

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:02	
Control Test	0.080	05:03	
Air Blank	0.000	05:03	
Subject Sample	# 1	05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:04	
Control Test	0.080	05:05	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 07/14/2022 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 07/16/2022 01:15:26

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Control Test	0.080	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 08/12/2022 20:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/16/2022 03:31:11

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.080	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	
Control Test	0.080	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 08/12/2022 20:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORRIS,BRIDGETT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 07/16/2022 05:53:07

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.080	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.149	05:48
Breath Volume	1.507	05:48
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.148	05:51
Breath Volume	1.757	05:51
Air Blank	0.000	05:52
Control Test	0.080	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 08/12/2022 20:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GUSTAFSON

DATE/TIME OF TEST: 07/17/2022 04:46:09

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.203	04:41
Breath Volume	1.410	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.190	04:44
Breath Volume	2.113	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 08/12/2022 20:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPP,ANDREW,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NURAJ,MARION
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 11:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: HOJAK

DATE/TIME OF TEST: 07/25/2022 13:01:01

Diagnostics Check	OK	12:52
Air Blank	0.000	12:53
Control Test	0.080	12:53
Air Blank	0.000	12:53
Subject Sample # 1	0.000	12:56
Breath Volume	1.656	12:56
Air Blank	0.000	12:57
Air Blank	0.000	12:59
Subject Sample # 2	0.000	12:59
Breath Volume	1.101	12:59
Air Blank	0.000	13:00
Control Test	0.079	13:00
Air Blank	0.000	13:00
Diagnostics Check	OK	13:01

Data Download Date/Time: 08/12/2022 20:19