

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN,JONATHAN,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LOGAN,JONATHAN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 18:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 07/03/2022 19:35:57

Diagnostics Check	OK	19:29
Air Blank	0.000	19:29
Control Test	0.081	19:30
Air Blank	0.000	19:30
Subject Sample # 1	0.164	19:31
Breath Volume	2.023	19:31
Air Blank	0.000	19:32
Air Blank	0.000	19:33
Subject Sample # 2	0.164	19:34
Breath Volume	3.023	19:34
Air Blank	0.000	19:35
Control Test	0.080	19:35
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 08/04/2022 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HELBIG,D
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/04/2022 01:26:33

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.202	01:21
Breath Volume	1.523	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:23
Subject Sample # 2	0.196	01:24
Breath Volume	1.519	01:24
Air Blank	0.000	01:25
Control Test	0.078	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 08/04/2022 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/10/2022 01:20:55

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.081	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.151	01:16
Breath Volume	2.523	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:18
Subject Sample # 2	0.149	01:19
Breath Volume	2.585	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 08/04/2022 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOORE,GREG,C

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MORRIS,CHRIS

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 12:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 07/11/2022 13:20:57

Diagnostics Check	OK	13:14
Air Blank	0.000	13:15
Control Test	0.081	13:15
Air Blank	0.000	13:15
Subject Sample # 1	0.000	13:16
Breath Volume	3.863	13:16
Air Blank	0.000	13:17
Air Blank	0.000	13:19
Subject Sample # 2	0.000	13:19
Breath Volume	3.472	13:19
Air Blank	0.000	13:20
Control Test	0.079	13:20
Air Blank	0.000	13:20
Diagnostics Check	OK	13:20

Data Download Date/Time: 08/04/2022 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN,JONATHAN,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: FORBES,O

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: DACKÉ

DATE/TIME OF TEST: 07/16/2022 01:29:39

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.000	01:25
Breath Volume	1.558	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.000	01:28
Breath Volume	1.535	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 08/04/2022 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MOHR,J
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: DOWDELL JR

DATE/TIME OF TEST: 07/23/2022 00:50:21

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.217	00:45
Breath Volume	1.691	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.220	00:48
Breath Volume	1.257	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 08/04/2022 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MIRELES,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: YAT REYES

DATE/TIME OF TEST: 07/01/2022 00:53:18

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.137	00:48
Breath Volume	1.535	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.143	00:51
Breath Volume	1.625	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 07/12/2022 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCHAEFER,ALLISON
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: ITZEP ORDONEZ

DATE/TIME OF TEST: 07/02/2022 03:05:45

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.228	03:01
Breath Volume	1.785	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.238	03:04
Breath Volume	1.968	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/12/2022 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHAEFER,A
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: IXMATA GUARCHAJ

DATE/TIME OF TEST: 07/03/2022 01:16:31

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.184	01:11
Breath Volume	1.136	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.200	01:14
Breath Volume	1.914	01:14
Air Blank	0.000	01:15
Control Test	0.079	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 07/12/2022 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWTON,JENNA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CLEARY,A
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 07/03/2022 04:32:55

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.107	04:28
Breath Volume	1.468	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.103	04:31
Breath Volume	1.714	04:31
Air Blank	0.000	04:31
Control Test	0.078	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 07/12/2022 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: GESSE

DATE/TIME OF TEST: 07/05/2022 02:12:16

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.194	02:07
Breath Volume	2.058	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.181	02:10
Breath Volume	1.355	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 07/12/2022 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SOMMERS,E
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BOSTON

DATE/TIME OF TEST: 07/08/2022 02:08:09

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.105	02:03
Breath Volume	1.101	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:05
Subject Sample # 2	0.105	02:06
Breath Volume	1.164	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 07/12/2022 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING, JONATHAN, L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TARMAN, CODY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: COMSTOCK

DATE/TIME OF TEST: 07/08/2022 23:09:08

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.080	23:03
Air Blank	0.000	23:04
Subject Sample # 1	0.000	23:04
Breath Volume	1.652	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.000	23:07
Breath Volume	1.550	23:07
Air Blank	0.000	23:08
Control Test	0.079	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 07/12/2022 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BIONDO,NICOLAS
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: RODRIGUEZ SILVA

DATE/TIME OF TEST: 07/09/2022 00:23:25

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.153	00:18
Breath Volume	2.570	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.155	00:21
Breath Volume	2.453	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 07/12/2022 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRAS,ANASTASIA,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MARRAS,ANASTASIA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:06

SUBJECT'S LAST NAME: PEDRO

DATE/TIME OF TEST: 07/09/2022 13:46:53

Diagnostics Check	OK	13:37
Air Blank	0.000	13:37
Control Test	0.081	13:38
Air Blank	0.000	13:38
Subject Sample # 1	0.225	13:39
Breath Volume	1.332	13:39
Air Blank	0.000	13:40
Air Blank	0.000	13:41
Subject Sample # 2	0.204	13:42
Breath Volume	2.769	13:42
Air Blank	0.000	13:43
Air Blank	0.000	13:44
Subject Sample # 3	0.191	13:45
Breath Volume	2.683	13:45
Air Blank	0.000	13:45
Control Test	0.076	13:46
Air Blank	0.000	13:46
Diagnostics Check	OK	13:46

Data Download Date/Time: 07/12/2022 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURST,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: SELF,GRANT
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: BUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ZIMMERMAN

DATE/TIME OF TEST: 07/09/2022 22:40:57

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.228	22:36
Breath Volume	3.105	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.226	22:39
Breath Volume	3.472	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 07/12/2022 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WRIGHT,CODY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 07/10/2022 01:33:46

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.078	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.096	01:28
Breath Volume	2.214	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.093	01:32
Breath Volume	1.648	01:32
Air Blank	0.000	01:32
Control Test	0.077	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 07/12/2022 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KOLLER,KEVIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: OTHER
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: KLEINHEINZ

DATE/TIME OF TEST: 07/13/2022 17:27:04

Diagnostics Check	OK	17:20
Air Blank	0.000	17:20
Control Test	0.080	17:21
Air Blank	0.000	17:21
Subject Sample # 1	0.142	17:22
Breath Volume	1.214	17:22
Air Blank	0.000	17:22
Air Blank	0.000	17:24
Subject Sample # 2	0.141	17:25
Breath Volume	1.132	17:25
Air Blank	0.000	17:26
Control Test	0.081	17:26
Air Blank	0.000	17:26
Diagnostics Check	OK	17:27

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAEFER,K,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: LOPEZ,V
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: POPOLI

DATE/TIME OF TEST: 07/14/2022 15:53:28

Diagnostics Check	OK	15:44
Air Blank	0.000	15:45
Control Test	0.081	15:45
Air Blank	0.000	15:46
Subject Sample # 1	0.138	15:47
Breath Volume	1.503	15:47
Air Blank	0.000	15:48
Air Blank	0.000	15:49
Subject Sample # 2	0.134	15:51
Breath Volume	1.566	15:51
Air Blank	0.000	15:52
Control Test	0.080	15:52
Air Blank	0.000	15:53
Diagnostics Check	OK	15:53

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/15/2022 00:00:02

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.200	23:55
Breath Volume	1.199	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.180	23:58
Breath Volume	1.312	23:58
Air Blank	0.000	23:59
Control Test	0.078	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	00:00

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ECKERT,KYLE
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: ANTOCH

DATE/TIME OF TEST: 07/15/2022 02:26:19

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.176	02:21
Breath Volume	2.761	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.180	02:24
Breath Volume	3.527	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANPELT,J,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WILLIAMS,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MEZZIO

DATE/TIME OF TEST: 07/16/2022 21:09:12

Diagnostics Check	OK	21:02
Air Blank	0.000	21:02
Control Test	0.081	21:02
Air Blank	0.000	21:03
Subject Sample # 1	0.032	21:04
Breath Volume	1.292	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:06
Subject Sample # 2	0.031	21:07
Breath Volume	1.515	21:07
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BIANDO,N
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: CHAVEZ-LOPEZ

DATE/TIME OF TEST: 07/17/2022 17:21:22

Diagnostics Check	OK	17:13
Air Blank	0.000	17:13
Control Test	0.080	17:13
Air Blank	0.000	17:14
Subject Sample # 1	0.300	17:16
Breath Volume	1.777	17:16
Air Blank	0.000	17:17
Air Blank	0.000	17:19
Subject Sample # 2	0.302	17:19
Breath Volume	3.000	17:19
Air Blank	0.000	17:20
Control Test	0.077	17:20
Air Blank	0.000	17:21
Diagnostics Check	OK	17:21

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENNIFER,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TORRES,JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/19/2022 05:29:39

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.059	05:24
Breath Volume	2.101	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.062	05:28
Breath Volume	3.207	05:28
Air Blank	0.000	05:28
Control Test	0.080	05:29
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: BAREIS

DATE/TIME OF TEST: 07/19/2022 22:01:48

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.224	21:57
Breath Volume	2.250	21:57
Air Blank	0.000	21:58
Air Blank	0.000	21:59
Subject Sample # 2	0.221	22:00
Breath Volume	1.664	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: JIMENEZ CEDENO

DATE/TIME OF TEST: 07/24/2022 02:45:09

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.081	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMEN,LOGAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEMEN,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HOKE-BEVERLY

DATE/TIME OF TEST: 07/24/2022 03:49:17

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.081	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.163	03:44
Breath Volume	1.324	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.167	03:47
Breath Volume	1.503	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: SANDS

DATE/TIME OF TEST: 07/24/2022 20:16:58

Diagnostics Check	OK	20:10
Air Blank	0.000	20:10
Control Test	0.080	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.157	20:12
Breath Volume	1.804	20:12
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.155	20:15
Breath Volume	2.304	20:15
Air Blank	0.000	20:16
Control Test	0.079	20:16
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BEITZ,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BIALOCK

DATE/TIME OF TEST: 07/25/2022 02:25:23

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.220	02:20
Breath Volume	2.140	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.212	02:23
Breath Volume	1.765	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LOPEZ Y LOPEZ

DATE/TIME OF TEST: 07/25/2022 02:31:50

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.079	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:30	RFI Detect
Breath Volume	0.433	02:30	
Air Blank	0.000	02:30	
Control Test	0.080	02:31	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LOPEZ Y LOPEZ

DATE/TIME OF TEST: 07/25/2022 02:41:52

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.078	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.266	02:35
Breath Volume	1.753	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.266	02:39
Breath Volume	1.449	02:39
Air Blank	0.000	02:40
Control Test	0.075	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: BLANTON

DATE/TIME OF TEST: 07/31/2022 18:50:43

Diagnostics Check	OK	18:44
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:45
Subject Sample # 1	0.182	18:46
Breath Volume	2.281	18:46
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample # 2	0.177	18:49
Breath Volume	1.597	18:49
Air Blank	0.000	18:49
Control Test	0.079	18:50
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

Data Download Date/Time: 08/17/2022 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PORTO HERNANDEZ

DATE/TIME OF TEST: 07/31/2022 04:18:32

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1 0.139	04:10	Volume Not Met
Breath Volume	0.105	04:10	
Air Blank	0.000	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 2 0.192	04:13	
Breath Volume	3.582	04:13	
Air Blank	0.000	04:14	
Air Blank	0.000	04:15	
Subject Sample	# 3 0.194	04:16	
Breath Volume	3.449	04:16	
Air Blank	0.000	04:17	
Control Test	0.077	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 08/14/2022 16:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,CDFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 07/10/2022 08:51:44

Diagnostics Check	OK	08:45
Air Blank	0.000	08:45
Control Test	0.080	08:45
Air Blank	0.000	08:46
Subject Sample # 1	0.094	08:47
Breath Volume	3.156	08:47
Air Blank	0.000	08:47
Air Blank	0.000	08:49
Subject Sample # 2	0.091	08:50
Breath Volume	2.058	08:50
Air Blank	0.000	08:50
Control Test	0.080	08:51
Air Blank	0.000	08:51
Diagnostics Check	OK	08:51

Data Download Date/Time: 08/04/2022 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEBO,ANDRES,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GUEVARA,W
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: BARNOFF

DATE/TIME OF TEST: 07/31/2022 20:12:21

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.079	20:05
Air Blank	0.000	20:05
Subject Sample # 1	0.199	20:07
Breath Volume	2.238	20:07
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.196	20:10
Breath Volume	1.558	20:10
Air Blank	0.000	20:11
Control Test	0.079	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 08/04/2022 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: DOMENECH MOLIERE

DATE/TIME OF TEST: 07/03/2022 21:02:28

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:56	
Control Test	0.080	20:56	
Air Blank	0.000	20:57	
Subject Sample	# 1 0.007	20:57	
Breath Volume	1.382	20:57	
Air Blank	0.000	20:58	
Air Blank	0.000	21:00	
Subject Sample	# 2 0.007	21:00	
Breath Volume	3.085	21:00	
Air Blank	0.000	21:01	
Control Test	0.079	21:01	
Air Blank		21:02	RFI Detect
Air Blank	0.000	21:02	

Data Download Date/Time: 07/18/2022 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: COURAN

DATE/TIME OF TEST: 07/03/2022 21:11:47

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.080	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.015	21:07
Breath Volume	2.570	21:07
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.015	21:10
Breath Volume	2.531	21:10
Air Blank	0.000	21:10
Control Test	0.080	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 07/18/2022 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GUTIERRES

DATE/TIME OF TEST: 07/08/2022 01:52:06

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.081	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.212	01:47
Breath Volume	2.574	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:49
Subject Sample # 2	0.205	01:50
Breath Volume	1.957	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 07/18/2022 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 07/10/2022 01:23:40

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:20	
Control Test	0.080	01:20	
Air Blank	0.000	01:21	
Subject Sample # 1	0.171	01:22	
Breath Volume	2.308	01:22	
Air Blank		01:23	Purge Fail
Air Blank	0.000	01:23	

Data Download Date/Time: 07/18/2022 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBOZA,LEONEL,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARBOZA,LEONEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: MAIDEN

DATE/TIME OF TEST: 07/12/2022 16:45:09

Diagnostics Check	OK	16:38
Air Blank	0.000	16:38
Control Test	0.079	16:38
Air Blank	0.000	16:39
Subject Sample # 1	0.234	16:40
Breath Volume	1.171	16:40
Air Blank	0.000	16:41
Air Blank	0.000	16:42
Subject Sample # 2	0.233	16:43
Breath Volume	1.574	16:43
Air Blank	0.000	16:44
Control Test	0.078	16:44
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 07/18/2022 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: SELF

DATE/TIME OF TEST: 07/13/2022 22:00:38

Diagnostics Check	OK	22:00	
Air Blank		22:00	RFI Detect
Air Blank		22:00	RFI Detect

Data Download Date/Time: 07/18/2022 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: SELF

DATE/TIME OF TEST: 07/13/2022 22:10:16

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.081	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.153	22:05
Breath Volume	1.574	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.155	22:08
Breath Volume	1.976	22:08
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 07/18/2022 10:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KOSAKOWSKI,PETER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: WILKENS-PLUMLEY

DATE/TIME OF TEST: 07/14/2022 00:20:44

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:18	
Control Test	0.080	00:18	
Air Blank	0.000	00:18	
Subject Sample	# 1 0.273	00:19	
Breath Volume	2.785	00:19	
Air Blank		00:20	Purge Fail
Air Blank	0.000	00:20	

Data Download Date/Time: 07/18/2022 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KOSAKOWSKI,PETER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: WILKENS-PLUMLEY

DATE/TIME OF TEST: 07/14/2022 00:30:16

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:23	
Control Test	0.079	00:24	
Air Blank	0.000	00:24	
Subject Sample # 1	0.269	00:25	
Breath Volume	2.058	00:25	
Air Blank	0.000	00:26	
Air Blank	0.000	00:27	
Subject Sample # 2	0.266	00:28	
Breath Volume	2.593	00:28	
Air Blank	0.000	00:29	
Control Test	0.071	00:29	Control Outside Tolerance
Air Blank	0.000	00:30	

Data Download Date/Time: 07/18/2022 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: DEMMLER

DATE/TIME OF TEST: 07/20/2022 02:00:07

Diagnostics Check	OK	01:58	
Air Blank		01:59	Ambient Fail
Air Blank		02:00	Purge Fail

Data Download Date/Time: 08/15/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JIMENEZ, WILLIAM,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JIMENEZ, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 07:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: GUILFORD

DATE/TIME OF TEST: 07/21/2022 07:47:36

Diagnostics Check	OK	07:41
Air Blank	0.000	07:41
Control Test	0.080	07:42
Air Blank	0.000	07:42
Subject Sample # 1	0.222	07:43
Breath Volume	2.105	07:43
Air Blank	0.000	07:43
Air Blank	0.000	07:45
Subject Sample # 2	0.217	07:45
Breath Volume	1.339	07:45
Air Blank	0.000	07:46
Control Test	0.078	07:46
Air Blank	0.000	07:47
Diagnostics Check	OK	07:47

Data Download Date/Time: 08/15/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,SHELBY,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: BROWN,TERRELL
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: EWERT

DATE/TIME OF TEST: 07/23/2022 03:53:53

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.197	03:48
Breath Volume	3.734	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:50
Subject Sample # 2	0.189	03:52
Breath Volume	4.582	03:52
Air Blank	0.000	03:52
Control Test	0.077	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 08/15/2022 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAEUSSLER,BRETT,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HAEUSSLER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: JOYAS RODRIGUEZ

DATE/TIME OF TEST: 07/23/2022 16:54:20

Diagnostics Check	OK	16:43	
Air Blank	0.000	16:44	
Control Test	0.079	16:44	
Air Blank	0.000	16:44	
Subject Sample	# 1 0.216	16:45	No .020 agreement
Breath Volume	1.855	16:45	
Air Blank	0.000	16:46	
Air Blank	0.000	16:48	
Subject Sample	# 2 0.182	16:49	No .020 agreement
Breath Volume	1.441	16:49	
Air Blank	0.000	16:50	
Air Blank	0.000	16:52	
Subject Sample	# 3 0.248	16:52	No .020 agreement
Breath Volume	2.113	16:52	
Air Blank	0.000	16:53	
Control Test	0.077	16:53	
Air Blank	0.000	16:54	
Diagnostics Check	OK	16:54	

Data Download Date/Time: 08/15/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAEUSSLER,BRETT,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HAEUSSLER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: JOYAS RODRIGUEZ

DATE/TIME OF TEST: 07/23/2022 17:35:39

Diagnostics Check	OK	17:29
Air Blank	0.000	17:29
Control Test	0.079	17:30
Air Blank	0.000	17:30
Subject Sample # 1	0.234	17:31
Breath Volume	3.648	17:31
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample # 2	0.229	17:33
Breath Volume	3.976	17:33
Air Blank	0.000	17:34
Control Test	0.078	17:35
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 08/15/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI, MARK, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JABLONSKI, MARK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: MARTINEZ TOVAR

DATE/TIME OF TEST: 07/24/2022 06:24:32

Diagnostics Check	OK	06:18
Air Blank	0.000	06:18
Control Test	0.078	06:19
Air Blank	0.000	06:19
Subject Sample # 1	0.115	06:20
Breath Volume	2.976	06:20
Air Blank	0.000	06:20
Air Blank	0.000	06:22
Subject Sample # 2	0.114	06:22
Breath Volume	3.398	06:22
Air Blank	0.000	06:23
Control Test	0.079	06:23
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

Data Download Date/Time: 08/15/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: DAVIDSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: BUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: BROWER

DATE/TIME OF TEST: 07/24/2022 22:40:25

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.015	22:35
Breath Volume	2.503	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.015	22:38
Breath Volume	2.382	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 08/15/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: PEREZ ANGEL

DATE/TIME OF TEST: 07/25/2022 04:05:51

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.078	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.196	04:00
Breath Volume	1.550	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.194	04:04
Breath Volume	2.132	04:04
Air Blank	0.000	04:04
Control Test	0.077	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 08/15/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENTLAND,ALEX,N
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: HUDDELSON,CHASE
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 07/26/2022 22:22:57

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.078	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.221	22:17
Breath Volume	1.324	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.205	22:21
Breath Volume	1.398	22:21
Air Blank	0.000	22:22
Control Test	0.077	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 08/15/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 07/28/2022 03:47:12

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.140	03:42
Breath Volume	2.871	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.138	03:45
Breath Volume	2.683	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 08/15/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SANSON

DATE/TIME OF TEST: 07/28/2022 04:02:39

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.163	03:57
Breath Volume	3.425	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.159	04:00
Breath Volume	2.765	04:00
Air Blank	0.000	04:01
Control Test	0.078	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 08/15/2022 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:35

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BOURNE

DATE/TIME OF TEST: 07/07/2022 04:06:07

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.078	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.000	04:00
Breath Volume	1.906	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.000	04:04
Breath Volume	1.433	04:04
Air Blank	0.000	04:05
Control Test	0.078	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 07/26/2022 19:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:22

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: DOBBS

DATE/TIME OF TEST: 07/08/2022 02:59:18

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.078	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.000	02:54
Breath Volume	1.789	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.000	02:57
Breath Volume	2.328	02:57
Air Blank	0.000	02:58
Control Test	0.077	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 07/26/2022 19:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: BONEY

DATE/TIME OF TEST: 07/08/2022 23:26:01

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.135	23:21
Breath Volume	1.996	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.144	23:24
Breath Volume	1.894	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:26

Data Download Date/Time: 07/26/2022 19:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MADDOX

DATE/TIME OF TEST: 07/10/2022 02:48:56

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.279	02:41
Breath Volume	3.660	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.257	02:44
Breath Volume	1.753	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 3	0.249	02:47
Breath Volume	1.886	02:47
Air Blank	0.000	02:48
Control Test	0.077	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 07/26/2022 19:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 16:51

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:56

SUBJECT'S LAST NAME: FUSSELL

DATE/TIME OF TEST: 07/14/2022 17:31:57

Diagnostics Check	OK	17:25
Air Blank	0.000	17:25
Control Test	0.079	17:25
Air Blank	0.000	17:26
Subject Sample # 1	0.145	17:27
Breath Volume	2.355	17:27
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.146	17:30
Breath Volume	1.710	17:30
Air Blank	0.000	17:31
Control Test	0.079	17:31
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 07/26/2022 19:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: BUTTREY, NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TREMBLE

DATE/TIME OF TEST: 07/27/2022 23:59:18

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:48	
Control Test	0.079	23:48	
Air Blank	0.000	23:49	
Subject Sample	# 1	23:52	No Sample Provided
Breath Volume	0.000	23:52	
Air Blank	0.000	23:52	
Air Blank	0.000	23:54	
Subject Sample	# 2	23:57	No Sample Provided
Breath Volume	0.000	23:57	
Air Blank	0.000	23:58	
Control Test	0.079	23:58	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 08/30/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON,IVETTE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: RICHARDSON,IVETTE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: VIRGILE

DATE/TIME OF TEST: 07/10/2022 02:55:55

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.081	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.220	02:48
Breath Volume	2.375	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.250	02:51
Breath Volume	2.242	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 3	0.228	02:54
Breath Volume	3.175	02:54
Air Blank	0.000	02:55
Control Test	0.077	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 08/10/2022 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAMAYO,FRANK,E
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CONSUEGRA,ALEJANDRO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 11:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:28

SUBJECT'S LAST NAME: DELVA

DATE/TIME OF TEST: 07/10/2022 11:53:23

Diagnostics Check	OK	11:49	
Air Blank	0.000	11:49	
Control Test	0.080	11:49	
Air Blank	0.000	11:50	
Subject Sample	# 1	11:52	Subject Test Refused
Breath Volume	0.000	11:52	
Air Blank	0.000	11:52	
Control Test	0.080	11:52	
Air Blank	0.000	11:53	
Diagnostics Check	OK	11:53	

Data Download Date/Time: 08/10/2022 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAMAYO,KEVIN,X
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TAMAYO,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: JEAN-JACQUES

DATE/TIME OF TEST: 07/15/2022 01:30:28

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.110	01:25
Breath Volume	1.867	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.111	01:28
Breath Volume	2.105	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 08/10/2022 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ROYER

DATE/TIME OF TEST: 07/16/2022 01:43:38

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.203	01:39
Breath Volume	2.988	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.196	01:42
Breath Volume	2.722	01:42
Air Blank	0.000	01:42
Control Test	0.077	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 07/28/2022 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARROQUIN,RICHARD,J

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: MARROQUIN,RICHARD

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 07/11/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG128602

DRY GAS EXPIRY DATE: 10/13/2023

VIOLATION CODE: BUI

ARREST TIME: 16:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 07/16/2022 16:52:13

Diagnostics Check	OK	16:51	
Air Blank		16:51	RFI Detect
Air Blank	0.000	16:52	

Data Download Date/Time: 08/30/2022 19:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARROQUIN,RICHARD,J

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: MARROQUIN,RICHARD

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 07/11/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG128602

DRY GAS EXPIRY DATE: 10/13/2023

VIOLATION CODE: BUI

ARREST TIME: 16:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 07/16/2022 17:00:39

Diagnostics Check	OK	16:54
Air Blank	0.000	16:54
Control Test	0.076	16:54
Air Blank	0.000	16:55
Subject Sample # 1	0.171	16:55
Breath Volume	2.292	16:55
Air Blank	0.000	16:56
Air Blank	0.000	16:58
Subject Sample # 2	0.178	16:59
Breath Volume	2.406	16:59
Air Blank	0.000	16:59
Control Test	0.077	17:00
Air Blank	0.000	17:00
Diagnostics Check	OK	17:00

Data Download Date/Time: 08/30/2022 19:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NORTH,DUANE,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: WILD,JACOB

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 06/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG129201

DRY GAS EXPIRY DATE: 10/19/2023

VIOLATION CODE: BUI

ARREST TIME: 21:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: TRAGER

DATE/TIME OF TEST: 07/04/2022 22:24:59

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:22	
Control Test	0.078	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1	22:23	Subject Test Refused
Breath Volume	0.000	22:23	
Air Blank	0.000	22:24	
Control Test	0.078	22:24	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 07/17/2022 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COGBURN, HEIDI, E

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: QUICK, JUSTIN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 07/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG129201

DRY GAS EXPIRY DATE: 10/19/2023

VIOLATION CODE: BUI

ARREST TIME: 16:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: OTOOLE

DATE/TIME OF TEST: 07/23/2022 17:41:16

Diagnostics Check	OK	17:34
Air Blank	0.000	17:35
Control Test	0.078	17:35
Air Blank	0.000	17:36
Subject Sample # 1	0.043	17:36
Breath Volume	1.238	17:36
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 2	0.043	17:39
Breath Volume	1.812	17:39
Air Blank	0.000	17:40
Control Test	0.078	17:40
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

Data Download Date/Time: 08/23/2022 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COGBURN, HEIDI, E

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: WILD, JACOB

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 07/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG129201

DRY GAS EXPIRY DATE: 10/19/2023

VIOLATION CODE: BUI

ARREST TIME: 17:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 07/23/2022 18:58:36

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.077	18:53
Air Blank	0.000	18:53
Subject Sample # 1	0.058	18:54
Breath Volume	4.480	18:54
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.057	18:57
Breath Volume	4.429	18:57
Air Blank	0.000	18:57
Control Test	0.077	18:58
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 08/23/2022 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: Yes
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: WEBB,ANDREW,H	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: WEBB,ANDREW	ARREST TIME: 16:30
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2022	OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: EVERHART	DATE/TIME OF TEST: 07/08/2022	19:18:08
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	19:12
Air Blank	0.000	19:12
Control Test	0.079	19:12
Air Blank	0.000	19:13
Subject Sample # 1	0.101	19:13
Breath Volume	3.074	19:13
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample # 2	0.091	19:16
Breath Volume	1.746	19:16
Air Blank	0.000	19:17
Control Test	0.079	19:17
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time: 07/29/2022 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRODBECK,JAMES,J

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: BRODBECK,JAMES

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: BUI

ARREST TIME: 19:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 07/08/2022 19:53:47

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:51	
Control Test	0.080	19:51	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:52	Subject Test Refused
Breath Volume	0.000	19:52	
Air Blank	0.000	19:52	
Control Test	0.080	19:53	
Air Blank	0.000	19:53	
Diagnostics Check	OK	19:53	

Data Download Date/Time: 07/29/2022 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: LAMBERT,BRIAN,K	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: LAMBERT,BRIAN	ARREST TIME: 15:42
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2022	OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: S530782972460	DATE/TIME OF TEST: 07/09/2022	16:35:06
------------------------------------	-------------------------------	----------

Diagnostics Check	OK	16:28
Air Blank	0.000	16:28
Control Test	0.079	16:29
Air Blank	0.000	16:29
Subject Sample # 1	0.185	16:30
Breath Volume	2.062	16:30
Air Blank	0.000	16:31
Air Blank	0.000	16:32
Subject Sample # 2	0.183	16:33
Breath Volume	1.425	16:33
Air Blank	0.000	16:34
Control Test	0.078	16:34
Air Blank	0.000	16:35
Diagnostics Check	OK	16:35

Data Download Date/Time: 07/29/2022 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: GRAY,SETH,M	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: GLOVER,LAURANCE	ARREST TIME: 16:15
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2022	OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: SMITH	DATE/TIME OF TEST: 07/09/2022	17:05:06
----------------------------	-------------------------------	----------

Diagnostics Check	OK	16:55
Air Blank	0.000	16:56
Control Test	0.078	16:56
Air Blank	0.000	16:56
Subject Sample # 1	0.143	16:59
Breath Volume	1.363	16:59
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.144	17:03
Breath Volume	2.156	17:03
Air Blank	0.000	17:04
Control Test	0.078	17:04
Air Blank	0.000	17:05
Diagnostics Check	OK	17:05

Data Download Date/Time: 07/29/2022 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: RICE,MICHAEL,	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)	VIOLATION CODE: BUI
ARRESTING OFFICER: RICE,MICHAEL	ARREST TIME: 16:04
ARRESTING OFFICER AGENCY: FFWCC (Lake City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2022	OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: STUART DATE/TIME OF TEST: 07/09/2022 17:29:53

Diagnostics Check	OK	17:22
Air Blank	0.000	17:23
Control Test	0.078	17:23
Air Blank	0.000	17:24
Subject Sample # 1	0.215	17:25
Breath Volume	2.136	17:25
Air Blank	0.000	17:26
Air Blank	0.000	17:27
Subject Sample # 2	0.219	17:28
Breath Volume	1.589	17:28
Air Blank	0.000	17:28
Control Test	0.078	17:29
Air Blank	0.000	17:29
Diagnostics Check	OK	17:29

Data Download Date/Time: 07/29/2022 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PRITCHARD,THOMAS,D

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: PRITCHARD,THOMAS

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: BUI

ARREST TIME: 15:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MCALEER

DATE/TIME OF TEST: 07/09/2022 17:38:20

Diagnostics Check	OK	17:35	
Air Blank	0.000	17:35	
Control Test	0.079	17:36	
Air Blank	0.000	17:36	
Subject Sample	# 1	17:36	Subject Test Refused
Breath Volume	0.000	17:36	
Air Blank	0.000	17:37	
Control Test	0.079	17:37	
Air Blank	0.000	17:38	
Diagnostics Check	OK	17:38	

Data Download Date/Time: 07/29/2022 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: BASFORD,NICOLE,M	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BASFORD,NICOLE	ARREST TIME: 16:39
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2022	OBSERVATION PERIOD BEGAN: 16:39

SUBJECT'S LAST NAME: CASEY DATE/TIME OF TEST: 07/09/2022 17:52:38

Diagnostics Check	OK	17:45
Air Blank	0.000	17:46
Control Test	0.079	17:46
Air Blank	0.000	17:47
Subject Sample # 1	0.137	17:48
Breath Volume	1.683	17:48
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.126	17:51
Breath Volume	1.449	17:51
Air Blank	0.000	17:51
Control Test	0.079	17:52
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 07/29/2022 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAND,MATTHEW,A

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: LAND,MATTHEW

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: BUI

ARREST TIME: 17:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: FULFORD

DATE/TIME OF TEST: 07/09/2022 18:08:28

Diagnostics Check	OK	18:05	
Air Blank	0.000	18:06	
Control Test	0.079	18:06	
Air Blank	0.000	18:06	
Subject Sample	# 1	18:07	Subject Test Refused
Breath Volume	0.000	18:07	
Air Blank	0.000	18:07	
Control Test	0.079	18:07	
Air Blank	0.000	18:08	
Diagnostics Check	OK	18:08	

Data Download Date/Time: 07/29/2022 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JARRELL,CHRISTIAN,P

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: JARRELL,CHRISTIAN

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: BUI

ARREST TIME: 18:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 07/09/2022 18:51:42

Diagnostics Check	OK	18:45
Air Blank	0.000	18:46
Control Test	0.080	18:46
Air Blank	0.000	18:47
Subject Sample # 1	0.152	18:47
Breath Volume	2.910	18:47
Air Blank	0.000	18:48
Air Blank	0.000	18:49
Subject Sample # 2	0.133	18:50
Breath Volume	2.652	18:50
Air Blank	0.000	18:50
Control Test	0.079	18:51
Air Blank	0.000	18:51
Diagnostics Check	OK	18:51

Data Download Date/Time: 07/29/2022 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: CORBIN,BRITTON,L	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: CORBIN,BRITTON	ARREST TIME: 18:27
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2022	OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: WHISTLER DATE/TIME OF TEST: 07/09/2022 19:14:06

Diagnostics Check	OK	19:05	
Air Blank	0.000	19:05	
Control Test	0.079	19:06	
Air Blank	0.000	19:06	
Subject Sample	# 1	19:09	No Sample Provided
Breath Volume	0.000	19:09	
Air Blank	0.000	19:10	
Air Blank	0.000	19:12	
Subject Sample	# 2	19:12	Subject Test Refused
Breath Volume	0.000	19:12	
Air Blank	0.000	19:13	
Control Test	0.080	19:13	
Air Blank	0.000	19:14	
Diagnostics Check	OK	19:14	

Data Download Date/Time: 07/29/2022 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOLNAR,JARROD,T

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: PIFER,JUSTIN

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: BUI

ARREST TIME: 18:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/09/2022 19:54:30

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:52	
Control Test	0.080	19:52	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:53	Subject Test Refused
Breath Volume	0.000	19:53	
Air Blank	0.000	19:53	
Control Test	0.079	19:53	
Air Blank	0.000	19:54	
Diagnostics Check	OK	19:54	

Data Download Date/Time: 07/29/2022 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: BRODBECK,JAMES,J	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BRODBECK,JAMES	ARREST TIME: 19:22
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2022	OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: WHITEHEAD	DATE/TIME OF TEST: 07/09/2022 20:01:27
--------------------------------	--

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:59	
Control Test	0.080	19:59	
Air Blank	0.000	19:59	
Subject Sample	# 1	20:00	Subject Test Refused
Breath Volume	0.000	20:00	
Air Blank	0.000	20:00	
Control Test	0.080	20:00	
Air Blank	0.000	20:01	
Diagnostics Check	OK	20:01	

Data Download Date/Time: 07/29/2022 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: BASFORD,NICOLE,M	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BOWER,RACHEL	ARREST TIME: 21:30
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2022	OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MESEROLL	DATE/TIME OF TEST: 07/09/2022 22:12:40
-------------------------------	--

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:10	
Control Test	0.079	22:10	
Air Blank	0.000	22:11	
Subject Sample	# 1	22:11	Subject Test Refused
Breath Volume	0.000	22:11	
Air Blank	0.000	22:11	
Control Test	0.079	22:12	
Air Blank	0.000	22:12	
Diagnostics Check	OK	22:12	

Data Download Date/Time: 07/29/2022 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: BASFORD,NICOLE,M	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: WEBB,ANDREW	ARREST TIME: 17:15
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2022	OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: CLACKLER DATE/TIME OF TEST: 07/17/2022 17:46:52

Diagnostics Check	OK	17:40
Air Blank	0.000	17:40
Control Test	0.079	17:41
Air Blank	0.000	17:41
Subject Sample # 1	0.039	17:42
Breath Volume	2.414	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:44
Subject Sample # 2	0.038	17:45
Breath Volume	2.480	17:45
Air Blank	0.000	17:45
Control Test	0.078	17:46
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

Data Download Date/Time: 07/29/2022 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUNEZ,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MONTENEGRO,ANDREA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 06/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/03/2022 02:09:29

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.080	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:08	
Control Test	0.079	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 07/28/2022 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: ALLEYNE, DAVANO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 06/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: LAUREANO

DATE/TIME OF TEST: 07/03/2022 02:25:19

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.080	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:23	Subject Test Refused
Breath Volume	0.000	02:23	
Air Blank	0.000	02:24	
Control Test	0.079	02:24	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 07/28/2022 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUNEZ,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: RODRIGUEZ,AGUSTIN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 06/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: RADOMSKI

DATE/TIME OF TEST: 07/05/2022 21:50:13

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:47	
Control Test	0.080	21:47	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:49	
Control Test	0.079	21:49	
Air Blank	0.000	21:50	
Diagnostics Check	OK	21:50	

Data Download Date/Time: 07/28/2022 15:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE,ALEXANDRA,P
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: ALLEYNE,ALEXANDRA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 06/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 07/09/2022 17:36:43

Diagnostics Check	OK	17:29
Air Blank	0.000	17:29
Control Test	0.080	17:29
Air Blank	0.000	17:30
Subject Sample # 1	0.217	17:31
Breath Volume	1.562	17:31
Air Blank	0.000	17:32
Air Blank	0.000	17:33
Subject Sample # 2	0.207	17:34
Breath Volume	1.191	17:34
Air Blank	0.000	17:35
Control Test	0.078	17:35
Air Blank	0.000	17:36
Diagnostics Check	OK	17:36

Data Download Date/Time: 07/28/2022 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: ALVARENGA,RICARDO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 06/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: PETIT

DATE/TIME OF TEST: 07/16/2022 10:27:39

Diagnostics Check	OK	10:22	
Air Blank	0.000	10:23	
Control Test	0.079	10:23	
Air Blank	0.000	10:24	
Subject Sample	# 1	10:26	Subject Test Refused
Breath Volume	0.000	10:26	
Air Blank	0.000	10:26	
Control Test	0.080	10:26	
Air Blank	0.000	10:27	
Diagnostics Check	OK	10:27	

Data Download Date/Time: 07/28/2022 15:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MILIAN,KEVIN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 06/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 07/23/2022 01:45:34

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.080	01:36	
Air Blank	0.000	01:36	
Subject Sample	# 1 0.119	01:40	Volume Not Met
Breath Volume	0.640	01:40	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 2	01:43	Subject Test Refused
Breath Volume	0.613	01:43	
Air Blank	0.000	01:44	
Control Test	0.077	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 07/28/2022 15:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,MELISSA,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: NUNEZ,BENJAMIN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 06/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 07/28/2022 03:28:19

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.224	03:23
Breath Volume	2.800	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:25
Subject Sample # 2	0.213	03:26
Breath Volume	4.433	03:26
Air Blank	0.000	03:27
Control Test	0.077	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 07/28/2022 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,MELISSA,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: OTERO,MICHAEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 06/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: RODRIGUEZ GRANADO

DATE/TIME OF TEST: 07/28/2022 05:16:36

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.212	05:11
Breath Volume	2.476	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.213	05:14
Breath Volume	1.566	05:14
Air Blank	0.000	05:15
Control Test	0.077	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 07/28/2022 15:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ, ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: CABELAS, C
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CIRIACOCABRERA

DATE/TIME OF TEST: 07/07/2022 01:21:42

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:12	
Subject Sample	# 1 0.179	01:15	Volume Not Met
Breath Volume	0.074	01:15	
Air Blank	0.000	01:16	
Air Blank	0.000	01:18	
Subject Sample	# 2	01:20	Subject Test Refused
Breath Volume	0.398	01:20	
Air Blank	0.000	01:20	
Control Test	0.078	01:21	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 07/31/2022 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: ROMERO,J
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HERNANDEZMARTINEZ

DATE/TIME OF TEST: 07/10/2022 04:02:01

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.079	03:59	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Control Test	0.079	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 07/31/2022 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 06/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: MARTINEZVALERIO

DATE/TIME OF TEST: 07/16/2022 22:56:11

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.079	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.148	22:51
Breath Volume	2.941	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.153	22:54
Breath Volume	3.109	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 07/31/2022 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JOHNSON, B
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BRANNER MOSLEY

DATE/TIME OF TEST: 07/01/2022 02:24:59

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.079	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.164	02:18
Breath Volume	1.679	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.169	02:23
Breath Volume	1.902	02:23
Air Blank	0.000	02:24
Control Test	0.077	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: DE LAS NUECES

DATE/TIME OF TEST: 07/01/2022 04:23:00

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.136	04:18
Breath Volume	1.964	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.134	04:21
Breath Volume	1.316	04:21
Air Blank	0.000	04:22
Control Test	0.077	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:23

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: BURDETT

DATE/TIME OF TEST: 07/02/2022 03:24:21

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.078	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.206	03:18
Breath Volume	1.449	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.220	03:22
Breath Volume	1.542	03:22
Air Blank	0.000	03:23
Control Test	0.078	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BROCATO,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: LU

DATE/TIME OF TEST: 07/02/2022 04:21:24

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.078	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.149	04:16
Breath Volume	2.089	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.154	04:19
Breath Volume	1.683	04:19
Air Blank	0.000	04:20
Control Test	0.078	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: CASTANON BADILLO

DATE/TIME OF TEST: 07/02/2022 05:14:36

Diagnostics Check	OK	05:07
Air Blank	0.000	05:08
Control Test	0.077	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.137	05:09
Breath Volume	1.703	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.143	05:12
Breath Volume	2.097	05:12
Air Blank	0.000	05:13
Control Test	0.078	05:13
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WATSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: RUSSO

DATE/TIME OF TEST: 07/03/2022 01:11:31

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.078	01:00	
Air Blank	0.000	01:00	
Subject Sample	# 1 0.142	01:03	Volume Not Met
Breath Volume	0.613	01:03	
Air Blank	0.000	01:04	
Air Blank	0.000	01:06	
Subject Sample	# 2 0.178	01:06	
Breath Volume	2.332	01:06	
Air Blank	0.000	01:07	
Air Blank	0.000	01:09	
Subject Sample	# 3 0.177	01:09	
Breath Volume	1.816	01:09	
Air Blank	0.000	01:10	
Control Test	0.077	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BRADWELL

DATE/TIME OF TEST: 07/03/2022 02:27:25

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.078	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:26	Subject Test Refused
Breath Volume	0.000	02:26	
Air Blank	0.000	02:26	
Control Test	0.078	02:26	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEWIS,I
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: ROSE JR

DATE/TIME OF TEST: 07/03/2022 02:53:33

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.077	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.102	02:48
Breath Volume	2.218	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.100	02:51
Breath Volume	1.699	02:51
Air Blank	0.000	02:52
Control Test	0.078	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: YARDE

DATE/TIME OF TEST: 07/03/2022 04:44:09

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.078	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.180	04:39
Breath Volume	2.019	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.174	04:42
Breath Volume	1.609	04:42
Air Blank	0.000	04:43
Control Test	0.078	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/04/2022 00:24:36

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.078	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.126	00:20
Breath Volume	1.593	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.121	00:23
Breath Volume	1.492	00:23
Air Blank	0.000	00:23
Control Test	0.078	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/04/2022 03:23:56

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.078	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.178	03:19
Breath Volume	1.324	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.168	03:22
Breath Volume	1.554	03:22
Air Blank	0.000	03:23
Control Test	0.076	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: ALINDOGAN

DATE/TIME OF TEST: 07/05/2022 01:54:26

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.078	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.168	01:49
Breath Volume	1.269	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.159	01:52
Breath Volume	3.210	01:52
Air Blank	0.000	01:53
Control Test	0.077	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: TWINING

DATE/TIME OF TEST: 07/05/2022 02:19:28

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.155	02:14
Breath Volume	2.699	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.145	02:17
Breath Volume	2.437	02:17
Air Blank	0.000	02:18
Control Test	0.077	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 07/13/2022 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WINTER,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/08/2022 06:11:13

Diagnostics Check	OK	06:04
Air Blank	0.000	06:05
Control Test	0.079	06:05
Air Blank	0.000	06:06
Subject Sample # 1	0.145	06:06
Breath Volume	2.371	06:06
Air Blank	0.000	06:07
Air Blank	0.000	06:09
Subject Sample # 2	0.134	06:09
Breath Volume	1.250	06:09
Air Blank	0.000	06:10
Control Test	0.078	06:10
Air Blank	0.000	06:11
Diagnostics Check	OK	06:11

Data Download Date/Time: 07/13/2022 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: HERNANDEZ CRUZ

DATE/TIME OF TEST: 07/09/2022 01:09:35

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.181	01:04
Breath Volume	2.269	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.179	01:07
Breath Volume	1.929	01:07
Air Blank	0.000	01:08
Control Test	0.078	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 07/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SALTI

DATE/TIME OF TEST: 07/09/2022 01:59:55

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.112	01:55
Breath Volume	1.871	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.122	01:58
Breath Volume	1.792	01:58
Air Blank	0.000	01:58
Control Test	0.078	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 07/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 07/10/2022 00:07:03

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.078	23:57
Air Blank	0.000	23:57
Subject Sample # 3	0.157	23:59
Breath Volume	1.671	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.183	00:02
Breath Volume	2.132	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:04
Subject Sample # 2	0.151	00:05
Breath Volume	1.417	00:05
Air Blank	0.000	00:06
Control Test	0.077	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:07

Data Download Date/Time: 07/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 07/10/2022 01:23:11

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.078	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:22	Sequence Aborted
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	

Data Download Date/Time: 07/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 07/10/2022 01:58:17

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:45	
Control Test	0.077	01:45	
Air Blank	0.000	01:46	
Subject Sample	# 1 0.144	01:49	Volume Not Met
Breath Volume	0.117	01:49	
Air Blank	0.000	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 2 0.172	01:52	
Breath Volume	1.980	01:52	
Air Blank	0.000	01:53	
Air Blank	0.000	01:55	
Subject Sample	# 3 0.167	01:56	
Breath Volume	1.378	01:56	
Air Blank	0.000	01:57	
Control Test	0.077	01:57	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 07/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SOUVE

DATE/TIME OF TEST: 07/10/2022 03:37:54

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.078	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.135	03:33
Breath Volume	1.710	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.133	03:36
Breath Volume	2.218	03:36
Air Blank	0.000	03:36
Control Test	0.077	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBERTS,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: OXLAJ GONZALEZ

DATE/TIME OF TEST: 07/10/2022 21:26:06

Diagnostics Check	OK	21:18
Air Blank	0.000	21:19
Control Test	0.078	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.158	21:21
Breath Volume	2.480	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:23
Subject Sample # 2	0.166	21:24
Breath Volume	2.789	21:24
Air Blank	0.000	21:25
Control Test	0.077	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 07/13/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI, H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: GUTSCHMIDT

DATE/TIME OF TEST: 07/11/2022 04:48:19

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.078	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.000	04:44
Breath Volume	3.300	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.000	04:46
Breath Volume	3.113	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 07/13/2022 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 07/11/2022 05:15:17

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.079	05:00	
Air Blank	0.000	05:00	
Subject Sample	# 1 0.252	05:03	No .020 agreement
Breath Volume	2.257	05:03	
Air Blank	0.000	05:03	
Air Blank	0.000	05:05	
Subject Sample	# 2	05:08	No Sample Provided
Breath Volume	0.000	05:08	
Air Blank	0.000	05:09	
Air Blank	0.000	05:11	
Subject Sample	# 3 0.226	05:13	No .020 agreement
Breath Volume	1.417	05:13	
Air Blank	0.000	05:14	
Control Test	0.078	05:14	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 07/13/2022 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 07/11/2022 05:33:23

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:21	
Control Test	0.078	05:21	
Air Blank	0.000	05:22	
Subject Sample	# 1	05:25	No Sample Provided
Breath Volume	0.000	05:25	
Air Blank	0.000	05:25	
Air Blank	0.000	05:27	
Subject Sample	# 2	0.234	05:28
Breath Volume	2.695	05:28	
Air Blank	0.000	05:28	
Air Blank	0.000	05:30	
Subject Sample	# 3	0.239	05:31
Breath Volume	2.710	05:31	
Air Blank	0.000	05:32	
Control Test	0.078	05:32	
Air Blank	0.000	05:33	
Diagnostics Check	OK	05:33	

Data Download Date/Time: 07/13/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: GONZALEZ GOMEZ

DATE/TIME OF TEST: 07/11/2022 22:42:36

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.078	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.194	22:36
Breath Volume	1.105	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.202	22:40
Breath Volume	1.281	22:40
Air Blank	0.000	22:41
Control Test	0.077	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 07/13/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: SANCHEZ SANCHEZ

DATE/TIME OF TEST: 07/13/2022 00:02:54

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.078	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.233	23:58
Breath Volume	1.445	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.233	00:01
Breath Volume	1.589	00:01
Air Blank	0.000	00:01
Control Test	0.078	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 07/13/2022 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: FELIX HENRIQUEZ

DATE/TIME OF TEST: 07/13/2022 04:31:17

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.078	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.062	04:26
Breath Volume	1.906	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.061	04:29
Breath Volume	1.523	04:29
Air Blank	0.000	04:30
Control Test	0.078	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 07/13/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 07/24/2022 04:20:01

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.116	04:14
Breath Volume	1.445	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.115	04:18
Breath Volume	1.566	04:18
Air Blank	0.000	04:19
Control Test	0.076	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:20

Data Download Date/Time: 08/04/2022 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: HERNANDEZ JR

DATE/TIME OF TEST: 07/24/2022 05:26:14

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.078	05:20
Air Blank	0.000	05:20
Subject Sample # 1	0.103	05:21
Breath Volume	1.843	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.102	05:24
Breath Volume	1.292	05:24
Air Blank	0.000	05:25
Control Test	0.078	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 08/04/2022 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: PRECIADO

DATE/TIME OF TEST: 07/25/2022 21:38:08

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.079	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.262	21:32
Breath Volume	1.800	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.261	21:36
Breath Volume	1.445	21:36
Air Blank	0.000	21:37
Control Test	0.076	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 08/04/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: BEHRENDT

DATE/TIME OF TEST: 07/30/2022 04:03:59

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.231	03:58
Breath Volume	1.761	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.237	04:02
Breath Volume	2.144	04:02
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 08/04/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PIERRE,C
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/31/2022 02:03:01

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.079	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1 0.211	01:53	Volume Not Met
Breath Volume	0.871	01:53	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample	# 2 0.217	01:56	
Breath Volume	1.183	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 3 0.207	02:01	
Breath Volume	1.390	02:01	
Air Blank	0.000	02:02	
Control Test	0.076	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 08/04/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: URRUTIA

DATE/TIME OF TEST: 07/01/2022 03:41:20

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.166	03:36
Breath Volume	2.050	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.168	03:39
Breath Volume	1.996	03:39
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 07/13/2022 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/02/2022 03:06:38

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.163	03:02
Breath Volume	2.546	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.166	03:05
Breath Volume	2.687	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY, JANICE, F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: CABRERA BURDIER

DATE/TIME OF TEST: 07/02/2022 05:05:47

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:53	
Control Test	0.080	04:53	
Air Blank	0.000	04:54	
Subject Sample	# 1 0.169	04:57	Volume Not Met
Breath Volume	0.257	04:57	
Air Blank	0.000	04:58	
Air Blank	0.000	04:59	
Subject Sample	# 2 0.174	05:00	
Breath Volume	1.105	05:00	
Air Blank	0.000	05:01	
Air Blank	0.000	05:02	
Subject Sample	# 3 0.172	05:04	
Breath Volume	2.019	05:04	
Air Blank	0.000	05:04	
Control Test	0.077	05:05	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: HOPPS

DATE/TIME OF TEST: 07/03/2022 01:51:58

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.079	01:39	
Air Blank	0.000	01:39	
Subject Sample	# 1 0.153	01:42	Volume Not Met
Breath Volume	0.156	01:42	
Air Blank	0.000	01:43	
Air Blank	0.000	01:45	
Subject Sample	# 2 0.171	01:46	
Breath Volume	1.226	01:46	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 3 0.170	01:50	
Breath Volume	1.457	01:50	
Air Blank	0.000	01:51	
Control Test	0.077	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: AMAYO

DATE/TIME OF TEST: 07/03/2022 03:31:51

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.148	03:27
Breath Volume	1.917	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.142	03:30
Breath Volume	2.910	03:30
Air Blank	0.000	03:30
Control Test	0.077	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: RIRIE

DATE/TIME OF TEST: 07/03/2022 04:01:02

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.144	03:56
Breath Volume	2.460	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.141	03:59
Breath Volume	3.253	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:01

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: NAVA NAVARRETTE

DATE/TIME OF TEST: 07/03/2022 05:11:19

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.079	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.208	05:06
Breath Volume	2.019	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.199	05:09
Breath Volume	2.199	05:09
Air Blank	0.000	05:10
Control Test	0.079	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: RIVERO FERNANDEZ

DATE/TIME OF TEST: 07/03/2022 06:20:19

Diagnostics Check	OK	06:12
Air Blank	0.000	06:12
Control Test	0.078	06:13
Air Blank	0.000	06:13
Subject Sample # 1	0.201	06:15
Breath Volume	1.640	06:15
Air Blank	0.000	06:15
Air Blank	0.000	06:17
Subject Sample # 2	0.194	06:18
Breath Volume	1.671	06:18
Air Blank	0.000	06:19
Control Test	0.078	06:19
Air Blank	0.000	06:20
Diagnostics Check	OK	06:20

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: MORILLO MORENO

DATE/TIME OF TEST: 07/03/2022 07:08:02

Diagnostics Check	OK	07:02
Air Blank	0.000	07:02
Control Test	0.078	07:02
Air Blank	0.000	07:03
Subject Sample # 1	0.056	07:03
Breath Volume	1.468	07:03
Air Blank	0.000	07:04
Air Blank	0.000	07:06
Subject Sample # 2	0.055	07:06
Breath Volume	1.101	07:06
Air Blank	0.000	07:07
Control Test	0.079	07:07
Air Blank	0.000	07:07
Diagnostics Check	OK	07:08

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: GALLEGOS HERNANDEZ

DATE/TIME OF TEST: 07/04/2022 01:08:45

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.115	01:04
Breath Volume	1.468	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.118	01:07
Breath Volume	1.167	01:07
Air Blank	0.000	01:07
Control Test	0.078	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ARJOMAND

DATE/TIME OF TEST: 07/04/2022 03:41:24

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.000	03:37
Breath Volume	1.890	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.000	03:40
Breath Volume	1.835	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: LINDOR

DATE/TIME OF TEST: 07/04/2022 04:33:14

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.063	04:28
Breath Volume	2.296	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.060	04:31
Breath Volume	2.828	04:31
Air Blank	0.000	04:32
Control Test	0.078	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:44

SUBJECT'S LAST NAME: BRITO-PARDUE

DATE/TIME OF TEST: 07/04/2022 09:11:02

Diagnostics Check	OK	09:04
Air Blank	0.000	09:05
Control Test	0.079	09:05
Air Blank	0.000	09:06
Subject Sample # 1	0.000	09:06
Breath Volume	3.261	09:06
Air Blank	0.000	09:07
Air Blank	0.000	09:09
Subject Sample # 2	0.000	09:09
Breath Volume	1.652	09:09
Air Blank	0.000	09:10
Control Test	0.078	09:10
Air Blank	0.000	09:10
Diagnostics Check	OK	09:11

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: ARROYO HERNANDEZ

DATE/TIME OF TEST: 07/04/2022 20:59:26

Diagnostics Check	OK	20:52
Air Blank	0.000	20:52
Control Test	0.079	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.253	20:54
Breath Volume	2.156	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.253	20:57
Breath Volume	2.042	20:57
Air Blank	0.000	20:58
Control Test	0.078	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: HACKETT

DATE/TIME OF TEST: 07/05/2022 01:31:05

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.078	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.229	01:26
Breath Volume	1.507	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:28
Subject Sample # 2	0.232	01:29
Breath Volume	1.816	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: WITHNER

DATE/TIME OF TEST: 07/06/2022 03:12:46

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.179	03:08
Breath Volume	2.144	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.173	03:11
Breath Volume	1.589	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMSON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: GALLICHAN

DATE/TIME OF TEST: 07/06/2022 15:57:18

Diagnostics Check	OK	15:50
Air Blank	0.000	15:51
Control Test	0.079	15:51
Air Blank	0.000	15:51
Subject Sample # 1	0.265	15:52
Breath Volume	2.265	15:52
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.280	15:55
Breath Volume	2.601	15:55
Air Blank	0.000	15:56
Control Test	0.078	15:56
Air Blank	0.000	15:57
Diagnostics Check	OK	15:57

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: LIMEHOUSE

DATE/TIME OF TEST: 07/07/2022 03:24:41

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.141	03:19
Breath Volume	1.789	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.139	03:23
Breath Volume	1.859	03:23
Air Blank	0.000	03:23
Control Test	0.078	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:58

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 07/07/2022 12:25:08

Diagnostics Check	OK	12:18
Air Blank	0.000	12:18
Control Test	0.078	12:19
Air Blank	0.000	12:19
Subject Sample # 1	0.310	12:20
Breath Volume	1.683	12:20
Air Blank	0.000	12:21
Air Blank	0.000	12:23
Subject Sample # 2	0.308	12:23
Breath Volume	2.339	12:23
Air Blank	0.000	12:24
Control Test	0.079	12:24
Air Blank	0.000	12:25
Diagnostics Check	OK	12:25

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HADDOCK

DATE/TIME OF TEST: 07/07/2022 21:41:40

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.079	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.171	21:36
Breath Volume	2.175	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.168	21:40
Breath Volume	1.667	21:40
Air Blank	0.000	21:40
Control Test	0.078	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 07/08/2022 05:21:16

Diagnostics Check	OK	05:14
Air Blank	0.000	05:15
Control Test	0.081	05:15
Air Blank	0.000	05:16
Subject Sample # 1	0.201	05:16
Breath Volume	1.976	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.192	05:19
Breath Volume	2.929	05:19
Air Blank	0.000	05:20
Control Test	0.079	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 07/13/2022 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: Y240117051880

DATE/TIME OF TEST: 07/09/2022 00:34:48

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.079	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.172	00:30
Breath Volume	2.101	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.174	00:33
Breath Volume	2.312	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 07/13/2022 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GRECO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: HAUG

DATE/TIME OF TEST: 07/09/2022 04:50:16

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.177	04:45
Breath Volume	1.488	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:47
Subject Sample # 2	0.171	04:48
Breath Volume	1.792	04:48
Air Blank	0.000	04:49
Control Test	0.078	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 07/13/2022 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: LINDSEY JR

DATE/TIME OF TEST: 07/09/2022 23:55:23

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.155	23:50
Breath Volume	2.992	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.153	23:53
Breath Volume	1.746	23:53
Air Blank	0.000	23:54
Control Test	0.078	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 07/13/2022 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: WINSOR

DATE/TIME OF TEST: 07/10/2022 00:21:24

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.078	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.207	00:16
Breath Volume	3.082	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.202	00:19
Breath Volume	2.933	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 07/13/2022 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: DE LA VEGA VERDECIA

DATE/TIME OF TEST: 07/10/2022 05:17:23

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.078	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.072	05:12
Breath Volume	1.882	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.069	05:15
Breath Volume	2.921	05:15
Air Blank	0.000	05:16
Control Test	0.078	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 07/13/2022 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 07/10/2022 21:26:43

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.078	21:15
Air Blank	0.000	21:15
Subject Sample # 1	0.198	21:18
Breath Volume	1.894	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:20
Subject Sample # 2	0.220	21:22
Breath Volume	1.812	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 3	0.213	21:25
Breath Volume	2.617	21:25
Air Blank	0.000	21:25
Control Test	0.077	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 07/13/2022 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MEEHAN

DATE/TIME OF TEST: 07/10/2022 23:16:52

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.223	23:12
Breath Volume	1.941	23:12
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.224	23:15
Breath Volume	2.250	23:15
Air Blank	0.000	23:15
Control Test	0.078	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 07/13/2022 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: JOHN

DATE/TIME OF TEST: 07/11/2022 04:54:44

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:43	
Control Test	0.079	04:43	
Air Blank	0.000	04:44	
Subject Sample	# 1 0.154	04:47	Volume Not Met
Breath Volume	0.210	04:47	
Air Blank	0.000	04:47	
Air Blank	0.000	04:49	
Subject Sample	# 2 0.154	04:50	
Breath Volume	2.476	04:50	
Air Blank	0.000	04:50	
Air Blank	0.000	04:52	
Subject Sample	# 3 0.157	04:53	
Breath Volume	2.539	04:53	
Air Blank	0.000	04:53	
Control Test	0.077	04:54	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 07/13/2022 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JAROSZEWSKI PEREIRA,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 07/11/2022 15:15:30

Diagnostics Check	OK	15:09	
Air Blank	0.000	15:10	
Control Test	0.079	15:10	
Air Blank	0.000	15:11	
Subject Sample	# 1 0.254	15:13	
Breath Volume	1.851	15:13	
Air Blank		15:14	RFI Detect
Air Blank	0.000	15:14	
Control Test	0.080	15:14	
Air Blank	0.000	15:15	
Diagnostics Check	OK	15:15	

Data Download Date/Time: 07/13/2022 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JAROSZEWSKI PEREIRA,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 07/11/2022 15:23:25

Diagnostics Check	OK	15:16
Air Blank	0.000	15:17
Control Test	0.078	15:17
Air Blank	0.000	15:17
Subject Sample # 1	0.254	15:18
Breath Volume	1.605	15:18
Air Blank	0.000	15:19
Air Blank	0.000	15:21
Subject Sample # 2	0.246	15:21
Breath Volume	3.167	15:21
Air Blank	0.000	15:22
Control Test	0.078	15:22
Air Blank	0.000	15:23
Diagnostics Check	OK	15:23

Data Download Date/Time: 07/13/2022 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FORSBERG,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/11/2022 22:39:51

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.079	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.173	22:35
Breath Volume	1.757	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.168	22:38
Breath Volume	2.234	22:38
Air Blank	0.000	22:38
Control Test	0.079	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 07/13/2022 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: ODERINDE

DATE/TIME OF TEST: 07/12/2022 22:39:23

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:34	
Control Test	0.081	22:35	
Air Blank	0.000	22:35	
Subject Sample	# 1	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Control Test	0.080	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 07/13/2022 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPIASI,JAKE,L
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: EKWEREKWU,CHUKWUEMEK
A
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male

OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MESCADO

DATE/TIME OF TEST: 07/24/2022 00:00:13

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.196	23:55
Breath Volume	1.847	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.194	23:58
Breath Volume	1.589	23:58
Air Blank	0.000	23:59
Control Test	0.081	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 08/25/2022 15:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STRACK,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: KLIMKOWSKI

DATE/TIME OF TEST: 07/30/2022 19:49:30

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:43	
Control Test	0.080	19:44	
Air Blank	0.000	19:44	
Subject Sample	# 1 0.194	19:47	Volume Not Met
Breath Volume	0.015	19:47	
Air Blank		19:48	Purge Fail
Air Blank	0.000	19:49	

Data Download Date/Time: 08/31/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STRACK,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: KLIMKOWSKI

DATE/TIME OF TEST: 07/30/2022 19:56:57

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:51	
Control Test	0.080	19:51	
Air Blank	0.000	19:52	
Subject Sample	# 1 0.253	19:55	Volume Not Met
Breath Volume	0.046	19:55	
Air Blank		19:56	Purge Fail
Air Blank	0.000	19:56	

Data Download Date/Time: 08/31/2022 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STRACK,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: KLIMKOWSKI

DATE/TIME OF TEST: 07/30/2022 20:32:10

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:26	
Control Test	0.081	20:26	
Air Blank	0.000	20:27	
Subject Sample	# 1 0.251	20:30	
Breath Volume	1.199	20:30	
Air Blank		20:31	Purge Fail
Air Blank		20:32	Purge Fail

Data Download Date/Time: 08/31/2022 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STRACK,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: KLIMKOWSKI

DATE/TIME OF TEST: 07/30/2022 20:39:51

Diagnostics Check	OK	20:33	
Air Blank	0.000	20:33	
Control Test	0.080	20:34	
Air Blank	0.000	20:34	
Subject Sample	# 1 0.237	20:37	Volume Not Met
Breath Volume	0.136	20:37	
Air Blank		20:38	Improper Sample
Air Blank	0.000	20:38	
Control Test	0.081	20:39	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:39	

Data Download Date/Time: 08/31/2022 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: HECKMAN JR.

DATE/TIME OF TEST: 07/03/2022 05:10:05

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.000	05:05
Breath Volume	2.308	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.000	05:08
Breath Volume	3.316	05:08
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:10

Data Download Date/Time: 08/30/2022 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: TRONE

DATE/TIME OF TEST: 07/02/2022 04:45:53

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.223	04:41
Breath Volume	1.546	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.208	04:44
Breath Volume	1.148	04:44
Air Blank	0.000	04:44
Control Test	0.079	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 07/26/2022 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: LARKINS

DATE/TIME OF TEST: 07/09/2022 04:07:16

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.065	04:02
Breath Volume	1.949	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.070	04:05
Breath Volume	1.992	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 07/26/2022 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: BALENTINE

DATE/TIME OF TEST: 07/14/2022 03:30:26

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.079	03:25	
Air Blank	0.000	03:25	
Subject Sample	# 1 0.000	03:26	
Breath Volume	1.562	03:26	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample	# 2 0.000	03:29	
Breath Volume	2.597	03:29	
Air Blank	0.000	03:29	
Control Test	0.080	03:29	
Air Blank		03:29	RFI Detect
Air Blank	0.000	03:30	

Data Download Date/Time: 07/26/2022 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: BALENTINE

DATE/TIME OF TEST: 07/14/2022 03:38:20

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.000	03:33
Breath Volume	2.417	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.000	03:36
Breath Volume	1.917	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 07/26/2022 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/17/2022 02:11:36

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.217	02:07
Breath Volume	2.886	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.221	02:10
Breath Volume	2.296	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 07/26/2022 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDD,RACHEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 07/28/2022 03:04:36

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.082	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.150	02:59
Breath Volume	2.003	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample # 2	0.141	03:03
Breath Volume	2.718	03:03
Air Blank	0.000	03:03
Control Test	0.081	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 08/26/2022 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: LARGEN,JACK
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MUSGROVE

DATE/TIME OF TEST: 07/02/2022 00:34:31

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1 0.235	00:30	Volume Not Met
Breath Volume	0.015	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 2	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:33	
Control Test	0.079	00:33	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 08/04/2022 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: VILLASENOR RODRIGUEZ

DATE/TIME OF TEST: 07/04/2022 04:53:23

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.080	04:51	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:52	
Control Test	0.081	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 08/04/2022 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,DONALD,H
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WILLIAMS,SETH
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:26

SUBJECT'S LAST NAME: SABO

DATE/TIME OF TEST: 07/04/2022 07:52:27

Diagnostics Check	OK	07:48	
Air Blank	0.000	07:49	
Control Test	0.080	07:49	
Air Blank	0.000	07:50	
Subject Sample	# 1	07:50	RFI Detect
Breath Volume	0.433	07:50	
Air Blank	0.000	07:51	
Control Test	0.079	07:51	
Air Blank	0.000	07:52	
Diagnostics Check	OK	07:52	

Data Download Date/Time: 08/04/2022 21:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,DONALD,H
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WILLIAMS,SETH
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:26

SUBJECT'S LAST NAME: SABIO

DATE/TIME OF TEST: 07/04/2022 08:01:30

Diagnostics Check	OK	07:54
Air Blank	0.000	07:55
Control Test	0.079	07:55
Air Blank	0.000	07:56
Subject Sample # 1	0.275	07:56
Breath Volume	4.652	07:56
Air Blank	0.000	07:57
Air Blank	0.000	07:59
Subject Sample # 2	0.268	07:59
Breath Volume	2.871	07:59
Air Blank	0.000	08:00
Control Test	0.079	08:00
Air Blank	0.000	08:01
Diagnostics Check	OK	08:01

Data Download Date/Time: 08/04/2022 21:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIECK, DAVID, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD, CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 07/07/2022 03:29:00

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.081	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1 0.093	03:24	
Breath Volume	1.609	03:24	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2 0.098	03:28	
Breath Volume	2.238	03:28	
Air Blank		03:28	RFI Detect
Air Blank	0.000	03:29	

Data Download Date/Time: 08/04/2022 21:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRICK,JASMIN,H
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MYRICK,JASMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: ROTHRAUFF

DATE/TIME OF TEST: 07/09/2022 00:30:38

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.081	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.195	00:26
Breath Volume	1.984	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.189	00:29
Breath Volume	2.644	00:29
Air Blank	0.000	00:29
Control Test	0.081	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 08/04/2022 21:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ALBRECHT,SYDNEY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 07/13/2022 01:16:18

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.081	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.127	01:11
Breath Volume	3.605	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.129	01:14
Breath Volume	2.375	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 08/04/2022 21:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROGAN,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WILLIAMS,TYLER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: BROGAN

DATE/TIME OF TEST: 07/20/2022 22:26:43

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:18	
Control Test	0.080	22:18	
Air Blank	0.000	22:19	
Subject Sample	# 1	22:22	No Sample Provided
Breath Volume	0.000	22:22	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample	# 2	22:25	Subject Test Refused
Breath Volume	0.000	22:25	
Air Blank	0.000	22:25	
Control Test	0.080	22:26	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:26	

Data Download Date/Time: 08/04/2022 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/22/2022 23:20:17

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:17	
Control Test	0.080	23:18	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:18	Subject Test Refused
Breath Volume	0.000	23:18	
Air Blank	0.000	23:19	
Control Test	0.081	23:19	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 08/04/2022 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ALBRECHT,SYDNEY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MCCUNE

DATE/TIME OF TEST: 07/24/2022 04:34:43

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:30	
Control Test	0.081	04:30	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:33	
Control Test	0.081	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 08/04/2022 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO, EDGARDO,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: DIXON, EZEKIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/28/2022 05:31:12

Diagnostics Check	OK	05:28	
Air Blank	0.000	05:28	
Control Test	0.080	05:29	
Air Blank	0.000	05:29	
Subject Sample	# 1	05:29	Subject Test Refused
Breath Volume	0.000	05:29	
Air Blank	0.000	05:30	
Control Test	0.080	05:30	
Air Blank	0.000	05:31	
Diagnostics Check	OK	05:31	

Data Download Date/Time: 08/04/2022 21:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRICK,JASMIN,H
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: DIXON,EZEKIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: ROBLEDO GONZALEZ

DATE/TIME OF TEST: 07/29/2022 05:12:33

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.081	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.172	05:08
Breath Volume	1.394	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.173	05:11
Breath Volume	1.445	05:11
Air Blank	0.000	05:11
Control Test	0.080	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 08/04/2022 21:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIECK, DAVID, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOLL, MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 07/29/2022 06:42:28

Diagnostics Check	OK	06:36
Air Blank	0.000	06:36
Control Test	0.080	06:37
Air Blank	0.000	06:37
Subject Sample # 1	0.000	06:38
Breath Volume	2.785	06:38
Air Blank	0.000	06:38
Air Blank	0.000	06:40
Subject Sample # 2	0.000	06:41
Breath Volume	1.929	06:41
Air Blank	0.000	06:41
Control Test	0.081	06:41
Air Blank	0.000	06:42
Diagnostics Check	OK	06:42

Data Download Date/Time: 08/04/2022 21:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TWIDDY,LINDSEY,R
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: TWIDDY,LINDSEY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/30/2022 13:12:51

Diagnostics Check	OK	13:05
Air Blank	0.000	13:05
Control Test	0.081	13:06
Air Blank	0.000	13:06
Subject Sample # 1	0.000	13:08
Breath Volume	1.160	13:08
Air Blank	0.000	13:08
Air Blank	0.000	13:10
Subject Sample # 2	0.000	13:11
Breath Volume	1.339	13:11
Air Blank	0.000	13:11
Control Test	0.080	13:12
Air Blank	0.000	13:12
Diagnostics Check	OK	13:12

Data Download Date/Time: 08/04/2022 21:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: SCHROCK

DATE/TIME OF TEST: 07/02/2022 05:30:01

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.078	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.000	05:25
Breath Volume	2.007	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.000	05:28
Breath Volume	1.320	05:28
Air Blank	0.000	05:29
Control Test	0.078	05:29
Air Blank	0.000	05:29
Diagnostics Check	OK	05:30

Data Download Date/Time: 07/21/2022 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: FREZZA

DATE/TIME OF TEST: 07/03/2022 04:14:36

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.237	04:09
Breath Volume	1.863	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.240	04:13
Breath Volume	2.101	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 07/21/2022 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 07/04/2022 01:26:42

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.110	01:21
Breath Volume	1.476	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.105	01:25
Breath Volume	1.964	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 07/21/2022 01:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: URFER

DATE/TIME OF TEST: 07/04/2022 16:00:27

Diagnostics Check	OK	15:58	
Air Blank	0.000	15:59	
Control Test	0.079	15:59	
Air Blank		15:59	RFI Detect
Air Blank	0.000	16:00	

Data Download Date/Time: 07/21/2022 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: URFER

DATE/TIME OF TEST: 07/04/2022 16:08:21

Diagnostics Check	OK	16:02
Air Blank	0.000	16:02
Control Test	0.078	16:02
Air Blank	0.000	16:03
Subject Sample # 1	0.178	16:03
Breath Volume	2.078	16:03
Air Blank	0.000	16:04
Air Blank	0.000	16:06
Subject Sample # 2	0.171	16:06
Breath Volume	2.070	16:06
Air Blank	0.000	16:07
Control Test	0.079	16:07
Air Blank	0.000	16:08
Diagnostics Check	OK	16:08

Data Download Date/Time: 07/21/2022 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 07/11/2022 18:27:21

Diagnostics Check	OK	18:21
Air Blank	0.000	18:21
Control Test	0.078	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.090	18:22
Breath Volume	2.511	18:22
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.094	18:25
Breath Volume	2.070	18:25
Air Blank	0.000	18:26
Control Test	0.077	18:26
Air Blank	0.000	18:27
Diagnostics Check	OK	18:27

Data Download Date/Time: 07/21/2022 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 07/16/2022 21:12:34

Diagnostics Check	OK	21:06
Air Blank	0.000	21:06
Control Test	0.079	21:06
Air Blank	0.000	21:07
Subject Sample # 1	0.233	21:08
Breath Volume	2.781	21:08
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.238	21:11
Breath Volume	2.128	21:11
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 07/21/2022 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 07/14/2022 00:51:52

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.079	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.164	00:47
Breath Volume	2.515	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:49
Subject Sample # 2	0.170	00:50
Breath Volume	2.609	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 08/04/2022 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 07/15/2022 04:38:30

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.220	04:33
Breath Volume	2.500	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.222	04:36
Breath Volume	2.605	04:36
Air Blank	0.000	04:37
Control Test	0.078	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 08/04/2022 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PRINCE,L
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: ERIKS

DATE/TIME OF TEST: 07/15/2022 08:33:33

Diagnostics Check	OK	08:25
Air Blank	0.000	08:26
Control Test	0.079	08:26
Air Blank	0.000	08:27
Subject Sample # 1	0.183	08:28
Breath Volume	2.750	08:28
Air Blank	0.000	08:28
Air Blank	0.000	08:30
Subject Sample # 2	0.180	08:32
Breath Volume	2.820	08:32
Air Blank	0.000	08:32
Control Test	0.078	08:32
Air Blank	0.000	08:33
Diagnostics Check	OK	08:33

Data Download Date/Time: 08/04/2022 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: GARCIA ACEVES

DATE/TIME OF TEST: 07/15/2022 23:51:30

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.099	23:47
Breath Volume	1.652	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.094	23:49
Breath Volume	1.347	23:49
Air Blank	0.000	23:50
Control Test	0.078	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 08/04/2022 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL, K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 07/16/2022 02:07:13

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.000	02:02
Breath Volume	1.484	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.000	02:05
Breath Volume	1.613	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 08/04/2022 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FENDT JR

DATE/TIME OF TEST: 07/16/2022 04:48:09

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.237	04:43
Breath Volume	1.398	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.221	04:46
Breath Volume	2.140	04:46
Air Blank	0.000	04:47
Control Test	0.078	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 08/04/2022 13:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER, T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 07/16/2022 05:22:55

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.170	05:18
Breath Volume	1.765	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.170	05:21
Breath Volume	1.457	05:21
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 08/04/2022 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 07/16/2022 06:38:36

Diagnostics Check	OK	06:31
Air Blank	0.000	06:31
Control Test	0.078	06:32
Air Blank	0.000	06:32
Subject Sample # 1	0.183	06:34
Breath Volume	2.027	06:34
Air Blank	0.000	06:34
Air Blank	0.000	06:36
Subject Sample # 2	0.174	06:37
Breath Volume	1.671	06:37
Air Blank	0.000	06:37
Control Test	0.079	06:38
Air Blank	0.000	06:38
Diagnostics Check	OK	06:38

Data Download Date/Time: 08/04/2022 13:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LOPEZ,C
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 07/16/2022 07:02:53

Diagnostics Check	OK	06:56
Air Blank	0.000	06:56
Control Test	0.078	06:57
Air Blank	0.000	06:57
Subject Sample # 1	0.150	06:58
Breath Volume	2.589	06:58
Air Blank	0.000	06:58
Air Blank	0.000	07:00
Subject Sample # 2	0.142	07:01
Breath Volume	2.160	07:01
Air Blank	0.000	07:01
Control Test	0.078	07:02
Air Blank	0.000	07:02
Diagnostics Check	OK	07:02

Data Download Date/Time: 08/04/2022 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WINTER,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:26

SUBJECT'S LAST NAME: YBARRA

DATE/TIME OF TEST: 07/16/2022 07:52:40

Diagnostics Check	OK	07:46
Air Blank	0.000	07:47
Control Test	0.079	07:47
Air Blank	0.000	07:47
Subject Sample # 1	0.154	07:48
Breath Volume	2.835	07:48
Air Blank	0.000	07:48
Air Blank	0.000	07:50
Subject Sample # 2	0.145	07:51
Breath Volume	2.574	07:51
Air Blank	0.000	07:51
Control Test	0.079	07:52
Air Blank	0.000	07:52
Diagnostics Check	OK	07:52

Data Download Date/Time: 08/04/2022 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHANNON,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: SCHULTZ

DATE/TIME OF TEST: 07/16/2022 14:18:27

Diagnostics Check	OK	14:10
Air Blank	0.000	14:10
Control Test	0.079	14:11
Air Blank	0.000	14:11
Subject Sample # 1	0.266	14:14
Breath Volume	1.296	14:14
Air Blank	0.000	14:14
Air Blank	0.000	14:16
Subject Sample # 2	0.274	14:16
Breath Volume	1.332	14:16
Air Blank	0.000	14:17
Control Test	0.079	14:17
Air Blank	0.000	14:18
Diagnostics Check	OK	14:18

Data Download Date/Time: 08/04/2022 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: COMM VEHIC
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: CALZADO NOA

DATE/TIME OF TEST: 07/16/2022 22:26:58

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.079	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.168	22:22
Breath Volume	1.273	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.169	22:25
Breath Volume	1.308	22:25
Air Blank	0.000	22:26
Control Test	0.078	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 08/04/2022 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COOPER,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 07/17/2022 03:04:54

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.201	02:59
Breath Volume	2.062	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.203	03:03
Breath Volume	2.171	03:03
Air Blank	0.000	03:03
Control Test	0.078	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 08/04/2022 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WINTER,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/17/2022 04:04:00

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.090	03:59
Breath Volume	2.164	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample # 2	0.093	04:02
Breath Volume	2.207	04:02
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:04

Data Download Date/Time: 08/04/2022 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: TROUTMAN

DATE/TIME OF TEST: 07/17/2022 21:29:00

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.079	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.174	21:24
Breath Volume	1.453	21:24
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.164	21:27
Breath Volume	2.019	21:27
Air Blank	0.000	21:28
Control Test	0.077	21:28
Air Blank	0.000	21:28
Diagnostics Check	OK	21:29

Data Download Date/Time: 08/04/2022 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: LOPEZ GONZALEZ

DATE/TIME OF TEST: 07/17/2022 23:20:49

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.079	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.148	23:16
Breath Volume	1.707	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.144	23:19
Breath Volume	2.035	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 08/04/2022 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: RAMIREZ GALLEGOS

DATE/TIME OF TEST: 07/18/2022 23:14:45

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.078	23:03	
Air Blank	0.000	23:04	
Subject Sample	# 1 0.250	23:07	Volume Not Met
Breath Volume	0.511	23:07	
Air Blank	0.000	23:08	
Air Blank	0.000	23:10	
Subject Sample	# 2 0.252	23:13	Volume Not Met
Breath Volume	0.171	23:13	
Air Blank	0.000	23:13	
Control Test	0.078	23:14	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 08/04/2022 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: RAMIREZ GALLEGOS

DATE/TIME OF TEST: 07/18/2022 23:31:00

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:20	
Control Test	0.078	23:20	
Air Blank	0.000	23:20	
Subject Sample	# 1	23:24	No Sample Provided
Breath Volume	0.000	23:24	
Air Blank	0.000	23:24	
Air Blank	0.000	23:26	
Subject Sample	# 2	23:29	No Sample Provided
Breath Volume	0.000	23:29	
Air Blank	0.000	23:30	
Control Test	0.078	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 08/04/2022 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: ZALDIVAR

DATE/TIME OF TEST: 07/20/2022 01:10:49

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.220	01:06
Breath Volume	1.898	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.220	01:09
Breath Volume	1.703	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 08/04/2022 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/20/2022 01:49:06

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:37	
Control Test	0.079	01:37	
Air Blank	0.000	01:38	
Subject Sample	# 1 0.158	01:41	Volume Not Met
Breath Volume	0.066	01:41	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample	# 2 0.156	01:44	
Breath Volume	1.484	01:44	
Air Blank	0.000	01:45	
Air Blank	0.000	01:46	
Subject Sample	# 3 0.163	01:47	
Breath Volume	1.714	01:47	
Air Blank	0.000	01:48	
Control Test	0.078	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 08/04/2022 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KOEHL,P
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BIRRA

DATE/TIME OF TEST: 07/20/2022 03:10:14

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.128	03:04
Breath Volume	1.121	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.125	03:08
Breath Volume	1.226	03:08
Air Blank	0.000	03:09
Control Test	0.078	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 08/04/2022 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: CLARRETT

DATE/TIME OF TEST: 07/20/2022 04:14:32

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.274	04:10
Breath Volume	1.750	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.258	04:12
Breath Volume	1.515	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 08/04/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: DOMINGO GARCIA

DATE/TIME OF TEST: 07/20/2022 23:37:15

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:24	
Control Test	0.079	23:25	
Air Blank	0.000	23:25	
Subject Sample	# 1 0.223	23:28	Slope Not Level
Breath Volume	1.125	23:28	
Air Blank	0.000	23:29	
Air Blank	0.000	23:31	
Subject Sample	# 2 0.237	23:32	
Breath Volume	1.515	23:32	
Air Blank	0.000	23:32	
Air Blank	0.000	23:34	
Subject Sample	# 3 0.247	23:35	
Breath Volume	1.164	23:35	
Air Blank	0.000	23:36	
Control Test	0.077	23:36	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 08/04/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WELLS, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MCKINZIE

DATE/TIME OF TEST: 07/21/2022 04:50:41

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.138	04:46
Breath Volume	2.070	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.139	04:49
Breath Volume	2.074	04:49
Air Blank	0.000	04:49
Control Test	0.078	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 08/04/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: ANASOULIS

DATE/TIME OF TEST: 07/22/2022 00:46:35

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.079	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.177	00:42
Breath Volume	2.492	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.158	00:45
Breath Volume	1.191	00:45
Air Blank	0.000	00:45
Control Test	0.079	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 08/04/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EDGE,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: TORRALBA

DATE/TIME OF TEST: 07/22/2022 22:48:38

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.079	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 1 0.175	22:41	Volume Not Met
Breath Volume	0.207	22:41	
Air Blank	0.000	22:42	
Air Blank	0.000	22:43	
Subject Sample	# 2 0.183	22:47	Volume Not Met
Breath Volume	0.230	22:47	
Air Blank	0.000	22:47	
Control Test	0.078	22:48	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 08/04/2022 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: WAJGEL

DATE/TIME OF TEST: 07/22/2022 23:31:23

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.000	23:27
Breath Volume	3.835	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.000	23:30
Breath Volume	3.222	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 08/04/2022 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 07/23/2022 02:48:40

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.203	02:44
Breath Volume	2.445	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.201	02:47
Breath Volume	2.464	02:47
Air Blank	0.000	02:47
Control Test	0.078	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 08/04/2022 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: ZUNIGA VINDAS

DATE/TIME OF TEST: 07/23/2022 03:19:40

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.133	03:14
Breath Volume	2.101	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.140	03:18
Breath Volume	3.160	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 08/04/2022 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIERMEK,T
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/23/2022 04:41:49

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.000	04:37
Breath Volume	2.121	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.000	04:40
Breath Volume	1.492	04:40
Air Blank	0.000	04:40
Control Test	0.079	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 08/04/2022 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/24/2022 03:14:13

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.207	03:05
Breath Volume	1.230	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.186	03:09
Breath Volume	1.500	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 3	0.217	03:12
Breath Volume	1.347	03:12
Air Blank	0.000	03:13
Control Test	0.078	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 08/04/2022 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: TABOADA,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GOSTON

DATE/TIME OF TEST: 07/24/2022 04:01:23

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.020	03:56
Breath Volume	1.695	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.019	03:59
Breath Volume	1.835	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 08/04/2022 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WHITLOCK, R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: GLASS

DATE/TIME OF TEST: 07/24/2022 04:40:48

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.178	04:34
Breath Volume	2.398	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.181	04:39
Breath Volume	1.320	04:39
Air Blank	0.000	04:39
Control Test	0.078	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 08/04/2022 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 07/24/2022 05:51:08

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.079	05:44
Air Blank	0.000	05:45
Subject Sample # 1	0.116	05:46
Breath Volume	1.441	05:46
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample # 2	0.114	05:49
Breath Volume	1.410	05:49
Air Blank	0.000	05:50
Control Test	0.079	05:50
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 08/04/2022 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/24/2022 07:51:00

Diagnostics Check	OK	07:43
Air Blank	0.000	07:44
Control Test	0.079	07:44
Air Blank	0.000	07:45
Subject Sample # 1	0.059	07:45
Breath Volume	1.207	07:45
Air Blank	0.000	07:46
Air Blank	0.000	07:48
Subject Sample # 2	0.055	07:49
Breath Volume	1.792	07:49
Air Blank	0.000	07:50
Control Test	0.079	07:50
Air Blank	0.000	07:50
Diagnostics Check	OK	07:51

Data Download Date/Time: 08/04/2022 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 07/25/2022 03:41:25

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.038	03:36
Breath Volume	1.578	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.036	03:39
Breath Volume	1.472	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 08/04/2022 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: WIX

DATE/TIME OF TEST: 07/26/2022 04:10:52

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.146	04:06
Breath Volume	1.777	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:08
Subject Sample # 2	0.142	04:09
Breath Volume	1.453	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 08/04/2022 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: GARCIA MORALES

DATE/TIME OF TEST: 07/27/2022 01:14:20

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.188	01:09
Breath Volume	1.406	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.201	01:12
Breath Volume	1.992	01:12
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 08/04/2022 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TABOADA,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:16

SUBJECT'S LAST NAME: HUMPHREY

DATE/TIME OF TEST: 07/27/2022 16:43:10

Diagnostics Check	OK	16:36
Air Blank	0.000	16:36
Control Test	0.080	16:37
Air Blank	0.000	16:37
Subject Sample # 1	0.298	16:38
Breath Volume	3.546	16:38
Air Blank	0.000	16:39
Air Blank	0.000	16:40
Subject Sample # 2	0.297	16:41
Breath Volume	2.902	16:41
Air Blank	0.000	16:42
Control Test	0.080	16:42
Air Blank	0.000	16:43
Diagnostics Check	OK	16:43

Data Download Date/Time: 08/04/2022 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURACEAU, R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: COSTELLO

DATE/TIME OF TEST: 07/29/2022 21:56:35

Diagnostics Check	OK	21:49
Air Blank	0.000	21:50
Control Test	0.080	21:50
Air Blank	0.000	21:50
Subject Sample # 1	0.288	21:52
Breath Volume	1.761	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.304	21:54
Breath Volume	1.644	21:54
Air Blank	0.000	21:55
Control Test	0.079	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 08/04/2022 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: GHERARDI

DATE/TIME OF TEST: 07/29/2022 23:05:39

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:57	
Control Test	0.079	22:58	
Air Blank	0.000	22:58	
Subject Sample	# 1	23:01	No Sample Provided
Breath Volume	0.000	23:01	
Air Blank	0.000	23:02	
Air Blank	0.000	23:04	
Subject Sample	# 2	23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:04	
Control Test	0.080	23:05	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

Data Download Date/Time: 08/04/2022 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHIRELING,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 07/30/2022 04:24:45

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.080	04:13	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:16	No Sample Provided
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample	# 2	0.128	04:19
Breath Volume	4.683	04:19	
Air Blank	0.000	04:20	
Air Blank	0.000	04:22	
Subject Sample	# 3	0.139	04:23
Breath Volume	4.292	04:23	
Air Blank	0.000	04:23	
Control Test	0.079	04:24	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 08/04/2022 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: LUNA MILLAN

DATE/TIME OF TEST: 07/30/2022 04:57:14

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.079	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.158	04:52
Breath Volume	2.246	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:54
Subject Sample # 2	0.164	04:55
Breath Volume	1.835	04:55
Air Blank	0.000	04:56
Control Test	0.079	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 08/04/2022 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: PAUMIER RODRIGUEZ

DATE/TIME OF TEST: 07/31/2022 02:41:22

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.223	02:36
Breath Volume	1.671	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.230	02:39
Breath Volume	2.671	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 08/04/2022 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: ESPINOZA DANIEL

DATE/TIME OF TEST: 07/31/2022 23:28:16

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.080	23:22	
Air Blank	0.000	23:23	
Subject Sample # 1	0.197	23:23	
Breath Volume	2.199	23:23	
Air Blank	0.000	23:24	
Air Blank	0.000	23:26	
Subject Sample # 2	0.203	23:26	
Breath Volume	2.789	23:26	
Air Blank	0.000	23:27	
Control Test	0.080	23:27	Interferent Detect
Air Blank	0.000	23:28	

Data Download Date/Time: 08/04/2022 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: ESPINOZA DANIEL

DATE/TIME OF TEST: 07/31/2022 23:36:42

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.199	23:32
Breath Volume	2.167	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.200	23:35
Breath Volume	1.996	23:35
Air Blank	0.000	23:35
Control Test	0.079	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 08/04/2022 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESPIAN,MIGUEL,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: DESPIAN,MIGUEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: GINDIN

DATE/TIME OF TEST: 07/03/2022 23:13:43

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.082	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.056	23:09
Breath Volume	3.488	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.056	23:12
Breath Volume	2.445	23:12
Air Blank	0.000	23:12
Control Test	0.083	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 07/27/2022 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIF,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MELILLO,FRANK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: LETSINGER

DATE/TIME OF TEST: 07/01/2022 02:52:10

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.080	02:50	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 07/16/2022 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ARIZZA,DANIEL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: GLORIOSO

DATE/TIME OF TEST: 07/01/2022 03:41:01

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.080	03:37	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Control Test	0.081	03:40	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 07/16/2022 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ARIZZA,DANIEL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: GLORIOSO

DATE/TIME OF TEST: 07/01/2022 04:01:10

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:49	
Control Test	0.080	03:49	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:53	No Sample Provided
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample	# 2	0.269	03:56
Breath Volume	1.425	03:56	
Air Blank	0.000	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 3	0.263	03:59
Breath Volume	1.781	03:59	
Air Blank	0.000	04:00	
Control Test	0.079	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 07/16/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOAGLAN,BRANDON,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HOAGLAN,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: HEBERT

DATE/TIME OF TEST: 07/01/2022 16:44:44

Diagnostics Check	OK	16:38	
Air Blank	0.000	16:38	
Control Test	0.080	16:39	
Air Blank	0.000	16:39	
Subject Sample	# 1 0.174	16:40	
Breath Volume	1.554	16:40	
Air Blank	0.000	16:41	
Air Blank	0.000	16:42	
Subject Sample	# 2 0.167	16:43	
Breath Volume	1.238	16:43	
Air Blank		16:44	RFI Detect
Air Blank	0.000	16:44	

Data Download Date/Time: 07/16/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PETERS,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CHI RAMIREZ

DATE/TIME OF TEST: 07/02/2022 02:02:23

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.157	01:57
Breath Volume	2.300	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.154	02:00
Breath Volume	2.039	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 07/16/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/02/2022 02:40:33

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.167	02:36
Breath Volume	1.871	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.152	02:38
Breath Volume	1.726	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 07/16/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,P,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LAPALME,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HOERLE

DATE/TIME OF TEST: 07/02/2022 19:46:42

Diagnostics Check	OK	19:40
Air Blank	0.000	19:40
Control Test	0.079	19:41
Air Blank	0.000	19:41
Subject Sample # 1	0.066	19:42
Breath Volume	1.683	19:42
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.068	19:45
Breath Volume	2.117	19:45
Air Blank	0.000	19:45
Control Test	0.080	19:46
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 07/16/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PODRAZA,BRIAN,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LAPALME,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 07/03/2022 16:49:09

Diagnostics Check	OK	16:45	
Air Blank	0.000	16:45	
Control Test	0.080	16:46	
Air Blank	0.000	16:46	
Subject Sample	# 1	16:47	Subject Test Refused
Breath Volume	0.000	16:47	
Air Blank	0.000	16:48	
Control Test	0.080	16:48	
Air Blank	0.000	16:49	
Diagnostics Check	OK	16:49	

Data Download Date/Time: 07/16/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PETERS,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 07/03/2022 19:53:25

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.080	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.154	19:48
Breath Volume	2.523	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.144	19:51
Breath Volume	2.355	19:51
Air Blank	0.000	19:52
Control Test	0.080	19:52
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 07/16/2022 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: ALBRIGHT

DATE/TIME OF TEST: 07/06/2022 20:06:08

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:51	
Control Test	0.080	19:51	
Air Blank	0.000	19:51	
Subject Sample	# 1 0.000	19:55	Volume Not Met
Breath Volume	0.000	19:55	
Air Blank	0.000	19:55	
Air Blank	0.000	19:57	
Subject Sample	# 2 0.000	20:00	
Breath Volume	1.585	20:00	
Air Blank	0.000	20:01	
Air Blank	0.000	20:03	
Subject Sample	# 3 0.000	20:04	
Breath Volume	1.199	20:04	
Air Blank	0.000	20:05	
Control Test	0.079	20:05	
Air Blank	0.000	20:06	
Diagnostics Check	OK	20:06	

Data Download Date/Time: 07/16/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RODRIGUEZ,DAVID
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: SHARLOW

DATE/TIME OF TEST: 07/09/2022 01:13:54

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.112	01:09
Breath Volume	1.519	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.113	01:12
Breath Volume	1.445	01:12
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 07/16/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COURINGTON,RONALD
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: STEEN

DATE/TIME OF TEST: 07/10/2022 02:41:34

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.243	02:36
Breath Volume	1.437	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.240	02:39
Breath Volume	1.320	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 07/16/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BARNLEY,KAY
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: DARLIN

DATE/TIME OF TEST: 07/10/2022 03:20:19

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.193	03:15
Breath Volume	2.312	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.191	03:18
Breath Volume	1.128	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 07/16/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAMLETT,MEGAN,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: BRAMLETT

DATE/TIME OF TEST: 07/11/2022 16:30:30

Diagnostics Check	OK	16:27	
Air Blank	0.000	16:27	
Control Test	0.080	16:28	
Air Blank	0.000	16:28	
Subject Sample	# 1	16:29	Subject Test Refused
Breath Volume	0.000	16:29	
Air Blank	0.000	16:29	
Control Test	0.079	16:29	
Air Blank	0.000	16:30	
Diagnostics Check	OK	16:30	

Data Download Date/Time: 07/16/2022 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 07/13/2022 00:38:56

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.080	00:27	
Air Blank	0.000	00:28	
Subject Sample	# 1 0.140	00:31	Volume Not Met
Breath Volume	0.058	00:31	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 2 0.140	00:37	Volume Not Met
Breath Volume	0.714	00:37	
Air Blank	0.000	00:37	
Control Test	0.080	00:38	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 07/16/2022 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 07/23/2022 18:43:32

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:41	
Control Test	0.080	18:41	
Air Blank	0.000	18:41	
Subject Sample	# 1	18:42	Subject Test Refused
Breath Volume	0.000	18:42	
Air Blank	0.000	18:42	
Control Test	0.079	18:42	
Air Blank	0.000	18:43	
Diagnostics Check	OK	18:43	

Data Download Date/Time: 08/11/2022 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COURINGTON,R
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: STOUT

DATE/TIME OF TEST: 07/24/2022 02:06:33

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.170	02:02
Breath Volume	1.886	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.161	02:05
Breath Volume	1.882	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 08/11/2022 19:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: DELPHEY

DATE/TIME OF TEST: 07/26/2022 14:35:26

Diagnostics Check	OK	14:28
Air Blank	0.000	14:28
Control Test	0.080	14:28
Air Blank	0.000	14:29
Subject Sample # 1	0.026	14:31
Breath Volume	3.332	14:31
Air Blank	0.000	14:31
Air Blank	0.000	14:33
Subject Sample # 2	0.027	14:33
Breath Volume	3.628	14:33
Air Blank	0.000	14:34
Control Test	0.081	14:34
Air Blank	0.000	14:35
Diagnostics Check	OK	14:35

Data Download Date/Time: 08/11/2022 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,DAVID,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAMOS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: BECKSTRAND

DATE/TIME OF TEST: 07/31/2022 00:00:41

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:54	
Control Test	0.080	23:54	
Air Blank	0.000	23:54	
Subject Sample # 1	0.147	23:55	
Breath Volume	5.664	23:55	
Air Blank	0.000	23:56	
Air Blank	0.000	23:57	
Subject Sample # 2	0.148	23:58	
Breath Volume	4.617	23:58	
Air Blank	0.000	23:59	
Control Test	0.080	23:59	
Air Blank	0.000	00:00	
Diagnostics Check	Fail	00:00	Temperature Regulation Test
Air Blank	0.000	00:00	

Data Download Date/Time: 08/11/2022 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,DAVID,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAMOS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: BECKSTRAND

DATE/TIME OF TEST: 07/31/2022 00:09:27

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.145	00:04
Breath Volume	4.191	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.142	00:07
Breath Volume	4.035	00:07
Air Blank	0.000	00:08
Control Test	0.078	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 08/11/2022 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TAYLOR II

DATE/TIME OF TEST: 07/01/2022 03:24:50

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:17	No Sample Provided
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample	# 2	03:23	No Sample Provided
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Control Test	0.082	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 07/26/2022 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: HO

DATE/TIME OF TEST: 07/02/2022 00:35:04

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.081	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.183	00:30
Breath Volume	2.843	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:32
Subject Sample # 2	0.193	00:33
Breath Volume	2.414	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:35

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BURRELL

DATE/TIME OF TEST: 07/02/2022 04:11:07

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.081	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.158	04:06
Breath Volume	1.562	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.172	04:09
Breath Volume	1.253	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MONDT

DATE/TIME OF TEST: 07/03/2022 00:21:48

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.144	00:17
Breath Volume	2.878	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.139	00:20
Breath Volume	3.082	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: PIAZZA

DATE/TIME OF TEST: 07/03/2022 03:21:12

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:14	No Sample Provided
Breath Volume	0.000	03:14	
Air Blank	0.000	03:14	
Air Blank	0.000	03:16	
Subject Sample	# 2	03:19	No Sample Provided
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Control Test	0.081	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ORTEGA,R
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: BUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SCHERER

DATE/TIME OF TEST: 07/04/2022 23:38:29

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:36	
Control Test	0.081	23:36	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:37	Subject Test Refused
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Control Test	0.081	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: SCHEFFLER

DATE/TIME OF TEST: 07/06/2022 00:31:46

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.082	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:24	No Sample Provided
Breath Volume	0.000	00:24	
Air Blank	0.000	00:25	
Air Blank	0.000	00:27	
Subject Sample	# 2	00:30	No Sample Provided
Breath Volume	0.000	00:30	
Air Blank	0.000	00:30	
Control Test	0.081	00:31	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: DOBANI

DATE/TIME OF TEST: 07/06/2022 03:45:28

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.239	03:37
Breath Volume	1.332	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.217	03:40
Breath Volume	1.835	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 3	0.244	03:43
Breath Volume	2.195	03:43
Air Blank	0.000	03:44
Control Test	0.076	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PALMATEER

DATE/TIME OF TEST: 07/09/2022 00:39:04

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.079	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:32	No Sample Provided
Breath Volume	0.000	00:32	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 2	00:37	No Sample Provided
Breath Volume	0.000	00:37	
Air Blank	0.000	00:38	
Control Test	0.080	00:38	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: TAMAN,KELLEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 07/09/2022 21:50:16

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:42	
Control Test	0.080	21:42	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:46	No Sample Provided
Breath Volume	0.000	21:46	
Air Blank	0.000	21:46	
Air Blank	0.000	21:48	
Subject Sample	# 2	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:49	
Control Test	0.081	21:49	
Air Blank	0.000	21:50	
Diagnostics Check	OK	21:50	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/11/2022 08:53:21

Diagnostics Check	OK	08:50	
Air Blank	0.000	08:50	
Control Test	0.081	08:51	
Air Blank	0.000	08:51	
Subject Sample	# 1	08:51	Subject Test Refused
Breath Volume	0.000	08:51	
Air Blank	0.000	08:52	
Control Test	0.081	08:52	
Air Blank	0.000	08:53	
Diagnostics Check	OK	08:53	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MANCUSO

DATE/TIME OF TEST: 07/13/2022 02:19:06

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.081	02:15	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Control Test	0.080	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE,R
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:29

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 07/13/2022 08:17:27

Diagnostics Check	OK	08:08	
Air Blank	0.000	08:09	
Control Test	0.080	08:09	
Air Blank	0.000	08:10	
Subject Sample	# 1	08:13	No Sample Provided
Breath Volume	0.000	08:13	
Air Blank	0.000	08:13	
Air Blank	0.000	08:15	
Subject Sample	# 2	08:16	Subject Test Refused
Breath Volume	0.000	08:16	
Air Blank	0.000	08:16	
Control Test	0.081	08:16	
Air Blank	0.000	08:17	
Diagnostics Check	OK	08:17	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: WHALEY

DATE/TIME OF TEST: 07/14/2022 02:33:44

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.172	02:29
Breath Volume	1.394	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.172	02:32
Breath Volume	1.710	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWEIGER,JONATHAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOPKINS,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: LAMAR

DATE/TIME OF TEST: 07/15/2022 21:59:32

Diagnostics Check	OK	21:58	
Air Blank		21:59	RFI Detect
Air Blank	0.000	21:59	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWEIGER,J,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOPKINS,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: LAMAR

DATE/TIME OF TEST: 07/15/2022 22:04:43

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:02	
Control Test	0.080	22:02	
Air Blank	0.000	22:03	
Subject Sample	# 1	22:03	Subject Test Refused
Breath Volume	0.000	22:03	
Air Blank	0.000	22:03	
Control Test	0.080	22:04	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:04	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: TAMAN,KELLEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: CULPEPPER

DATE/TIME OF TEST: 07/16/2022 22:18:47

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.081	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.000	22:14
Breath Volume	1.464	22:14
Air Blank	0.000	22:14
Air Blank	0.000	22:17
Subject Sample # 2	0.000	22:17
Breath Volume	1.312	22:17
Air Blank	0.000	22:17
Control Test	0.079	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: DE LA TORRE

DATE/TIME OF TEST: 07/16/2022 23:05:00

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:01	
Control Test	0.080	23:02	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:04	
Control Test	0.079	23:04	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:05	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: BURG

DATE/TIME OF TEST: 07/18/2022 04:12:56

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.081	04:00	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:04	No Sample Provided
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:07	
Subject Sample	# 2	0.055	04:08
Breath Volume	1.695	04:08	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 3	0.049	04:11
Breath Volume	1.437	04:11	
Air Blank	0.000	04:12	
Control Test	0.078	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS, K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 07/18/2022 05:15:27

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.178	05:11
Breath Volume	3.476	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.163	05:13
Breath Volume	2.433	05:13
Air Blank	0.000	05:14
Control Test	0.078	05:14
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:52

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 07/19/2022 08:19:33

Diagnostics Check	OK	08:12
Air Blank	0.000	08:13
Control Test	0.081	08:13
Air Blank	0.000	08:14
Subject Sample # 1	0.000	08:15
Breath Volume	1.273	08:15
Air Blank	0.000	08:15
Air Blank	0.000	08:17
Subject Sample # 2	0.000	08:18
Breath Volume	1.296	08:18
Air Blank	0.000	08:18
Control Test	0.079	08:18
Air Blank	0.000	08:19
Diagnostics Check	OK	08:19

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MEDINIS

DATE/TIME OF TEST: 07/20/2022 01:52:46

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.081	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.000	01:48
Breath Volume	1.179	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.000	01:51
Breath Volume	1.699	01:51
Air Blank	0.000	01:51
Control Test	0.081	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ, SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 07/21/2022 02:38:33

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.082	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.182	02:34
Breath Volume	1.476	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.193	02:36
Breath Volume	1.542	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SESSIONS,C,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,S
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: CRUZ MORALES

DATE/TIME OF TEST: 07/22/2022 02:03:12

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:55	
Control Test	0.081	01:55	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:59	No Sample Provided
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:02	
Control Test	0.081	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 07/23/2022 01:14:56

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.081	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.205	01:10
Breath Volume	2.203	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.208	01:13
Breath Volume	1.750	01:13
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: W340421835130

DATE/TIME OF TEST: 07/23/2022 01:38:37

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.258	01:33
Breath Volume	1.812	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.267	01:36
Breath Volume	2.332	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 07/26/2022 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKASAL,TYVON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GANNAWAY,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: BUI
ARREST TIME: 10:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 07/24/2022 00:29:28

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.081	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.000	00:25
Breath Volume	1.601	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.000	00:28
Breath Volume	2.304	00:28
Air Blank	0.000	00:28
Control Test	0.082	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 07/26/2022 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERRY,TABATHA,B
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SCOTT,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GRANDE

DATE/TIME OF TEST: 07/24/2022 02:42:06

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.000	02:37
Breath Volume	1.992	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.000	02:40
Breath Volume	1.433	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 07/26/2022 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS, K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: RESTREPO

DATE/TIME OF TEST: 07/25/2022 02:18:45

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.000	02:14
Breath Volume	3.015	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.000	02:17
Breath Volume	2.039	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 07/26/2022 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:56

SUBJECT'S LAST NAME: BAJRIC

DATE/TIME OF TEST: 07/26/2022 09:23:04

Diagnostics Check	OK	09:16
Air Blank	0.000	09:17
Control Test	0.081	09:17
Air Blank	0.000	09:18
Subject Sample # 1	0.000	09:18
Breath Volume	2.765	09:18
Air Blank	0.000	09:19
Air Blank	0.000	09:21
Subject Sample # 2	0.000	09:21
Breath Volume	2.468	09:21
Air Blank	0.000	09:22
Control Test	0.081	09:22
Air Blank	0.000	09:22
Diagnostics Check	OK	09:23

Data Download Date/Time: 07/26/2022 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTENSEN,S,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 07/29/2022 02:10:39

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.230	02:05
Breath Volume	1.425	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.239	02:09
Breath Volume	1.449	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 08/29/2022 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTENSEN,S,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: RODRIGUEZ,B
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: BUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/29/2022 23:44:49

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:37	
Control Test	0.081	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1 0.076	23:40	
Breath Volume	1.976	23:40	
Air Blank	0.000	23:40	
Air Blank	0.000	23:42	
Subject Sample	# 2 0.076	23:43	
Breath Volume	1.585	23:43	
Air Blank	0.000	23:43	
Control Test	0.080	23:44	
Air Blank		23:44	RFI Detect
Air Blank		23:44	RFI Detect

Data Download Date/Time: 08/29/2022 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTENSEN,S,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SCOTT,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: GOODRICH

DATE/TIME OF TEST: 07/30/2022 00:21:40

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.081	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.163	00:17
Breath Volume	2.460	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.159	00:20
Breath Volume	2.621	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 08/29/2022 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 07/30/2022 23:53:59

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:42	
Control Test	0.081	23:43	
Air Blank	0.000	23:43	
Subject Sample	# 1 0.170	23:46	Volume Not Met
Breath Volume	0.460	23:46	
Air Blank	0.000	23:47	
Air Blank	0.000	23:49	
Subject Sample	# 2 0.176	23:52	Volume Not Met
Breath Volume	0.148	23:52	
Air Blank	0.000	23:53	
Control Test	0.080	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 08/29/2022 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/29/2022 23:22:55

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.081	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.118	23:18
Breath Volume	2.628	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.116	23:21
Breath Volume	1.687	23:21
Air Blank	0.000	23:21
Control Test	0.079	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 08/30/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BARNETT, CODY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: DUSHANE

DATE/TIME OF TEST: 07/30/2022 21:11:53

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:08	
Control Test	0.081	21:08	
Air Blank	0.000	21:09	
Subject Sample	# 1	21:10	Subject Test Refused
Breath Volume	0.000	21:10	
Air Blank	0.000	21:10	
Control Test	0.082	21:11	
Air Blank	0.000	21:11	
Diagnostics Check	OK	21:11	

Data Download Date/Time: 08/30/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BARNETT, CODY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: DUSHANE JR.

DATE/TIME OF TEST: 07/30/2022 21:22:59

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.082	21:16
Air Blank	0.000	21:16
Subject Sample # 1	0.219	21:17
Breath Volume	1.480	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.223	21:21
Breath Volume	1.210	21:21
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 08/30/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HERNANDEZ,ALAIN,

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: BLACKARD,RAYMOND

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 25721080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: SARGENTO-PEREZ

DATE/TIME OF TEST: 07/24/2022 01:04:16

Diagnostics Check	OK	01:03	
Air Blank		01:03	RFI Detect
Air Blank	0.000	01:04	

Data Download Date/Time: 08/23/2022 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HERNANDEZ,ALAIN,

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: BLACKARD,RAYMOND

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 25721080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: SARGENTO-PEREZ

DATE/TIME OF TEST: 07/24/2022 01:13:48

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.079	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.182	01:09
Breath Volume	2.453	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.181	01:12
Breath Volume	2.621	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 08/23/2022 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/27/2022 00:45:44

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.143	00:41
Breath Volume	3.218	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.141	00:44
Breath Volume	3.085	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 08/23/2022 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: KACHMAN

DATE/TIME OF TEST: 07/02/2022 02:04:24

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.245	01:59
Breath Volume	1.449	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.234	02:02
Breath Volume	1.585	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 07/18/2022 19:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH, JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: BARLEY

DATE/TIME OF TEST: 07/04/2022 02:51:59

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.080	02:40	
Air Blank	0.000	02:40	
Subject Sample	# 1 0.149	02:43	Volume Not Met
Breath Volume	0.144	02:43	
Air Blank	0.000	02:44	
Air Blank	0.000	02:46	
Subject Sample	# 2 0.180	02:47	
Breath Volume	1.582	02:47	
Air Blank	0.000	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 3 0.189	02:50	
Breath Volume	1.789	02:50	
Air Blank	0.000	02:51	
Control Test	0.079	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 07/18/2022 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH, JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WITHERS

DATE/TIME OF TEST: 07/05/2022 03:09:39

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.080	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 07/18/2022 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: GOODIN JR

DATE/TIME OF TEST: 07/10/2022 02:26:12

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.247	02:21
Breath Volume	2.671	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:23
Subject Sample # 2	0.231	02:24
Breath Volume	2.453	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 07/18/2022 19:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,C,D
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,J
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: GREESON

DATE/TIME OF TEST: 07/13/2022 03:34:04

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.080	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1 0.174	03:32	Volume Not Met
Breath Volume	0.550	03:32	
Air Blank		03:33	Purge Fail
Air Blank	0.000	03:34	

Data Download Date/Time: 07/18/2022 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,C,D
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,J
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: GREESON

DATE/TIME OF TEST: 07/13/2022 03:36:53

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.073	03:36	Control Outside Tolerance
Air Blank	0.000	03:36	

Data Download Date/Time: 07/18/2022 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,C,D
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,J
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: GREESON

DATE/TIME OF TEST: 07/13/2022 03:46:27

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.077	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.224	03:41
Breath Volume	1.871	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.224	03:44
Breath Volume	1.507	03:44
Air Blank	0.000	03:45
Control Test	0.076	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 07/18/2022 19:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SILKOU

DATE/TIME OF TEST: 07/21/2022 02:30:39

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.186	02:25
Breath Volume	1.265	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:27
Subject Sample # 2	0.177	02:28
Breath Volume	1.578	02:28
Air Blank	0.000	02:29
Control Test	0.076	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 08/11/2022 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/24/2022 03:15:24

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.081	03:13	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:14	
Control Test	0.080	03:14	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 08/11/2022 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: GAGNON

DATE/TIME OF TEST: 07/25/2022 03:45:32

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.080	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.079	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 08/11/2022 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: OLAFSSON

DATE/TIME OF TEST: 07/29/2022 00:43:45

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.152	00:38
Breath Volume	1.187	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.155	00:41
Breath Volume	1.359	00:41
Air Blank	0.000	00:42
Control Test	0.078	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 08/11/2022 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 07/31/2022 03:42:21

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.187	03:37
Breath Volume	1.714	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:39
Subject Sample # 2	0.175	03:40
Breath Volume	1.398	03:40
Air Blank	0.000	03:41
Control Test	0.077	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 08/11/2022 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HART,JEFFREY
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 06/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/02/2022 19:30:00

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:27	
Control Test	0.077	19:27	
Air Blank	0.000	19:28	
Subject Sample	# 1	19:28	Subject Test Refused
Breath Volume	0.000	19:28	
Air Blank	0.000	19:29	
Control Test	0.075	19:29	
Air Blank	0.000	19:29	
Diagnostics Check	OK	19:30	

Data Download Date/Time: 07/06/2022 05:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SMITH,CORIN
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 06/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: BUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: GERDTS

DATE/TIME OF TEST: 07/02/2022 20:05:45

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:03	
Control Test	0.076	20:03	
Air Blank	0.000	20:04	
Subject Sample	# 1	20:04	Subject Test Refused
Breath Volume	0.000	20:04	
Air Blank	0.000	20:04	
Control Test	0.076	20:05	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 07/06/2022 05:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: UZCATEGUI-ARISMENDI

DATE/TIME OF TEST: 07/16/2022 01:42:12

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.077	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.161	01:37
Breath Volume	2.425	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.175	01:40
Breath Volume	1.343	01:40
Air Blank	0.000	01:41
Control Test	0.077	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 08/10/2022 05:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: PEREZ-ECARRI

DATE/TIME OF TEST: 07/16/2022 23:05:49

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:03	
Control Test	0.077	23:03	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:04	
Control Test	0.075	23:05	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

Data Download Date/Time: 08/10/2022 05:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: TALABA

DATE/TIME OF TEST: 07/17/2022 00:27:49

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.076	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.158	00:23
Breath Volume	2.992	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.163	00:26
Breath Volume	3.011	00:26
Air Blank	0.000	00:26
Control Test	0.077	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 08/18/2022 06:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAAVEDRA, OSCAR,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SAAVEDRA, OSCAR
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: BUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: BADAWI VALENCIA

DATE/TIME OF TEST: 07/02/2022 18:06:26

Diagnostics Check	Fail	18:05	Voltage/Current Test
Air Blank	0.000	18:06	

Data Download Date/Time: 07/06/2022 05:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAAVEDRA, OSCAR,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SAAVEDRA, OSCAR
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: BUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: BADAWI VALENCIA

DATE/TIME OF TEST: 07/02/2022 18:16:47

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.079	18:10
Air Blank	0.000	18:11
Subject Sample # 1	0.000	18:12
Breath Volume	2.007	18:12
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.000	18:15
Breath Volume	2.312	18:15
Air Blank	0.000	18:15
Control Test	0.079	18:16
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 07/06/2022 05:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,JEFFERY,C
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HART,JEFFERY
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: BUI
ARREST TIME: 17:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/02/2022 18:30:38

Diagnostics Check	Fail	18:30	Voltage/Current Test
Air Blank	0.000	18:30	

Data Download Date/Time: 07/06/2022 05:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: LEAL PEREZ

DATE/TIME OF TEST: 07/16/2022 05:50:42

Diagnostics Check	OK	05:49	
Air Blank		05:50	RFI Detect
Air Blank	0.000	05:50	

Data Download Date/Time: 08/10/2022 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: LEAL PEREZ

DATE/TIME OF TEST: 07/16/2022 05:58:47

Diagnostics Check	OK	05:52
Air Blank	0.000	05:52
Control Test	0.081	05:53
Air Blank	0.000	05:53
Subject Sample # 1	0.159	05:54
Breath Volume	1.339	05:54
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.151	05:57
Breath Volume	2.691	05:57
Air Blank	0.000	05:57
Control Test	0.080	05:58
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 08/10/2022 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: P0ACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: AGUIRRE ZEVALLOS

DATE/TIME OF TEST: 07/17/2022 04:44:07

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.000	04:39
Breath Volume	2.628	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.000	04:42
Breath Volume	1.500	04:42
Air Blank	0.000	04:43
Control Test	0.079	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 08/10/2022 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: JIMENEZ SORIA

DATE/TIME OF TEST: 07/23/2022 02:11:23

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:09	
Control Test	0.078	02:09	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:10	
Control Test	0.078	02:10	Interferent Detect
Air Blank	0.000	02:11	

Data Download Date/Time: 08/10/2022 05:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: JIMENEZ SORIA

DATE/TIME OF TEST: 07/23/2022 02:44:48

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:42	
Control Test	0.078	02:42	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:43	
Control Test	0.078	02:44	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 08/10/2022 05:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: BUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 07/28/2022 22:55:55

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:52	
Control Test	0.079	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:54	Subject Test Refused
Breath Volume	0.000	22:54	
Air Blank	0.000	22:54	
Control Test	0.079	22:55	
Air Blank	0.000	22:55	
Diagnostics Check	OK	22:55	

Data Download Date/Time: 08/10/2022 05:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: CASTILLA-GUTIERREZ

DATE/TIME OF TEST: 07/07/2022 02:53:28

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.079	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.078	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 07/12/2022 04:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCDONALD,KIRK
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GUAMAN ALVAREZ

DATE/TIME OF TEST: 07/02/2022 03:52:12

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.082	03:47
Breath Volume	1.296	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.085	03:50
Breath Volume	1.597	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/12/2022 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MALONE,KRISTINA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PEREZ LAUREANO

DATE/TIME OF TEST: 07/03/2022 03:02:52

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.081	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1 0.207	02:54	Volume Not Met
Breath Volume	0.371	02:54	
Air Blank	0.000	02:55	
Air Blank	0.000	02:57	
Subject Sample	# 2 0.193	02:58	
Breath Volume	1.117	02:58	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample	# 3 0.213	03:01	
Breath Volume	1.324	03:01	
Air Blank		03:02	Purge Fail
Air Blank	0.000	03:02	

Data Download Date/Time: 07/12/2022 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 07:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: GIRON

DATE/TIME OF TEST: 07/03/2022 08:29:03

Diagnostics Check	OK	08:22
Air Blank	0.000	08:23
Control Test	0.079	08:23
Air Blank	0.000	08:24
Subject Sample # 1	0.196	08:24
Breath Volume	1.320	08:24
Air Blank	0.000	08:25
Air Blank	0.000	08:27
Subject Sample # 2	0.205	08:27
Breath Volume	2.046	08:27
Air Blank	0.000	08:28
Control Test	0.079	08:28
Air Blank	0.000	08:28
Diagnostics Check	OK	08:29

Data Download Date/Time: 07/12/2022 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: GABRIEL GABRIEL

DATE/TIME OF TEST: 07/16/2022 21:47:28

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.079	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.187	21:42
Breath Volume	3.671	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.186	21:45
Breath Volume	3.152	21:45
Air Blank	0.000	21:46
Control Test	0.078	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 08/19/2022 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 03521080A3
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: WELLS,AUSTIN	ARREST TIME: 23:57
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2022	OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: ESTRADA	DATE/TIME OF TEST: 07/27/2022	00:30:50
------------------------------	-------------------------------	----------

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.077	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.162	00:26
Breath Volume	3.265	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.158	00:29
Breath Volume	2.550	00:29
Air Blank	0.000	00:29
Control Test	0.076	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 08/30/2022 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 03521080A3
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: WELLS,AUSTIN	ARREST TIME: 03:59
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Female
LAST AGENCY INSPECTION: 07/20/2022	OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: PITTMAN	DATE/TIME OF TEST: 07/30/2022 04:19:38
------------------------------	--

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.077	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.181	04:14
Breath Volume	1.953	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.178	04:18
Breath Volume	2.503	04:18
Air Blank	0.000	04:18
Control Test	0.077	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 08/30/2022 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WELLS,AUSTIN,D

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: EARLYWINE,GARRETT

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 07/20/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 03521080A3

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: RAMOS GUADARRAMA

DATE/TIME OF TEST: 07/30/2022 23:45:18

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:42	
Control Test	0.079	23:42	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:43	Subject Test Refused
Breath Volume	0.000	23:43	
Air Blank	0.000	23:44	
Control Test	0.079	23:44	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 08/30/2022 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAAVEDRA, OSCAR,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: SAAVEDRA, OSCAR
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 06/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 07/17/2022 20:09:01

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:06	
Control Test	0.080	20:06	
Air Blank	0.000	20:07	
Subject Sample	# 1	20:07	Subject Test Refused
Breath Volume	0.000	20:07	
Air Blank	0.000	20:08	
Control Test	0.079	20:08	
Air Blank	0.000	20:08	
Diagnostics Check	OK	20:09	

Data Download Date/Time: 07/28/2022 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/20/2022 22:21:17

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.078	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.298	22:16
Breath Volume	2.445	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:18
Subject Sample # 2	0.305	22:19
Breath Volume	2.640	22:19
Air Blank	0.000	22:20
Control Test	0.078	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 08/25/2022 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/22/2022 01:43:15

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.078	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1 0.302	01:41	
Breath Volume	1.269	01:41	
Air Blank		01:42	Purge Fail
Air Blank	0.000	01:43	

Data Download Date/Time: 08/25/2022 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 07/26/2022 03:35:25

Diagnostics Check	OK	03:34	
Air Blank		03:34	Ambient Fail
Air Blank		03:35	Purge Fail

Data Download Date/Time: 08/25/2022 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 07/26/2022 03:44:46

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.078	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.129	03:39
Breath Volume	1.304	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.125	03:43
Breath Volume	1.390	03:43
Air Blank	0.000	03:43
Control Test	0.078	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 08/25/2022 12:49