

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MODERSON,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: GUASCH,AMANDA
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 08/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: CORDOVA

DATE/TIME OF TEST: 08/28/2022 22:06:21

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.077	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.168	22:01
Breath Volume	1.316	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.180	22:04
Breath Volume	1.902	22:04
Air Blank	0.000	22:05
Control Test	0.077	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 09/26/2022 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 08/19/2022 03:13:13

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.080	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 09/22/2022 18:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: PEREZ JAHUEY

DATE/TIME OF TEST: 08/20/2022 02:00:28

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.179	01:55
Breath Volume	1.359	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.175	01:58
Breath Volume	1.519	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 09/22/2022 18:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSHANS,DEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUGHES,ASHLEY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 08/20/2022 16:59:48

Diagnostics Check	OK	16:43	
Air Blank	0.000	16:43	
Control Test	0.080	16:44	
Air Blank	0.000	16:44	
Subject Sample	# 1 0.205	16:47	
Breath Volume	1.152	16:47	
Air Blank	0.000	16:48	
Air Blank	0.000	16:50	
Subject Sample	# 2 0.220	16:53	Volume Not Met
Breath Volume	0.746	16:53	
Air Blank	0.000	16:54	
Air Blank	0.000	16:56	
Subject Sample	# 3 0.206	16:58	
Breath Volume	1.230	16:58	
Air Blank	0.000	16:58	
Control Test	0.079	16:59	
Air Blank	0.000	16:59	
Diagnostics Check	OK	16:59	

Data Download Date/Time: 09/22/2022 18:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BULIZZANO,ALYSSA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: BOLIN

DATE/TIME OF TEST: 08/20/2022 18:52:13

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:41	
Control Test	0.080	18:41	
Air Blank	0.000	18:42	
Subject Sample	# 1	18:45	No Sample Provided
Breath Volume	0.000	18:45	
Air Blank	0.000	18:45	
Air Blank	0.000	18:47	
Subject Sample	# 2	18:50	No Sample Provided
Breath Volume	0.000	18:50	
Air Blank	0.000	18:51	
Control Test	0.080	18:51	
Air Blank	0.000	18:52	
Diagnostics Check	OK	18:52	

Data Download Date/Time: 09/22/2022 18:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PEREZ,JERMAINE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 08/20/2022 20:19:12

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:16	
Control Test	0.080	20:16	
Air Blank	0.000	20:17	
Subject Sample	# 1	20:17	Subject Test Refused
Breath Volume	0.000	20:17	
Air Blank	0.000	20:18	
Control Test	0.080	20:18	
Air Blank	0.000	20:19	
Diagnostics Check	OK	20:19	

Data Download Date/Time: 09/22/2022 18:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIS,BRITT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: BROOME JR

DATE/TIME OF TEST: 08/20/2022 23:36:24

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.080	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:35	
Control Test	0.080	23:35	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 09/22/2022 18:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: JAMES,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: UCHENDU

DATE/TIME OF TEST: 08/22/2022 02:57:37

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.157	02:53
Breath Volume	1.183	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.164	02:56
Breath Volume	1.710	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 09/22/2022 18:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 08/22/2022 22:58:25

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.403	22:53
Breath Volume	2.199	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.398	22:56
Breath Volume	2.031	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 09/22/2022 18:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVANOVIC,M,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IVANOVIC,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/25/2022 22:02:14

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.177	21:57
Breath Volume	1.410	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.183	22:00
Breath Volume	2.496	22:00
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 09/22/2022 18:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/25/2022 23:28:13

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.080	23:26	
Air Blank	0.000	23:26	
Subject Sample	# 1	23:26	Subject Test Refused
Breath Volume	0.000	23:26	
Air Blank	0.000	23:27	
Control Test	0.081	23:27	
Air Blank	0.000	23:28	
Diagnostics Check	OK	23:28	

Data Download Date/Time: 09/22/2022 18:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: GENZEL

DATE/TIME OF TEST: 08/27/2022 01:02:24

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.000	00:58
Breath Volume	3.515	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.000	01:01
Breath Volume	2.644	01:01
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 09/22/2022 18:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,KAYLA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 08/27/2022 04:01:02

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.272	03:56
Breath Volume	3.515	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.282	03:59
Breath Volume	3.789	03:59
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:01

Data Download Date/Time: 09/22/2022 18:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 08/28/2022 03:59:07

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.081	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.000	03:54
Breath Volume	2.558	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.000	03:57
Breath Volume	2.667	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 09/22/2022 18:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,BRADLEY,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCLAUGHLIN,BRADLEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 08/05/2022 03:15:06

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:04	
Control Test	0.078	03:04	
Air Blank	0.000	03:04	
Subject Sample	# 1 0.181	03:08	Volume Not Met
Breath Volume	0.050	03:08	
Air Blank	0.000	03:08	
Air Blank	0.000	03:10	
Subject Sample	# 2 0.151	03:13	Volume Not Met
Breath Volume	0.050	03:13	
Air Blank	0.000	03:14	
Control Test	0.077	03:14	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 08/25/2022 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CHAMBERLAIN, J
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 08/06/2022 04:08:55

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.078	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1 0.249	04:03	
Breath Volume	1.109	04:03	
Air Blank	0.000	04:04	
Air Blank	0.000	04:05	
Subject Sample	# 2	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Control Test	0.079	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 08/25/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GIVANS,GARRETH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: KESSLER

DATE/TIME OF TEST: 08/12/2022 04:03:26

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.215	03:58
Breath Volume	1.656	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.197	04:01
Breath Volume	1.898	04:01
Air Blank	0.000	04:02
Control Test	0.076	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 08/25/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CHAMBERLAIN, J
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: EWALD

DATE/TIME OF TEST: 08/20/2022 23:25:44

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.079	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.176	23:18
Breath Volume	2.386	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.183	23:24
Breath Volume	1.472	23:24
Air Blank	0.000	23:24
Control Test	0.078	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 08/25/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/03/2022 23:51:11

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.185	23:43
Breath Volume	1.199	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.208	23:46
Breath Volume	1.347	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 3	0.196	23:49
Breath Volume	1.546	23:49
Air Blank	0.000	23:50
Control Test	0.077	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 08/10/2022 18:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: BLISS

DATE/TIME OF TEST: 08/04/2022 22:25:53

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.078	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.216	22:20
Breath Volume	1.476	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.220	22:24
Breath Volume	1.460	22:24
Air Blank	0.000	22:24
Control Test	0.078	22:25
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 08/10/2022 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/05/2022 03:32:31

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.078	03:30	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Control Test	0.076	03:31	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 08/10/2022 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 08/13/2022 22:56:44

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.078	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.109	22:51
Breath Volume	1.300	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.110	22:55
Breath Volume	1.394	22:55
Air Blank	0.000	22:55
Control Test	0.077	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 09/22/2022 18:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 08/14/2022 02:04:53

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.079	02:02	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:03	Subject Test Refused
Breath Volume	0.000	02:03	
Air Blank	0.000	02:03	
Control Test	0.078	02:04	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 09/22/2022 18:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE, CORPORAL W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE, CORPORAL W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/14/2022 03:29:08

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.079	03:27	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 09/22/2022 18:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIGH, ADAM, R
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: OTTO, ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: REEDER

DATE/TIME OF TEST: 08/13/2022 07:53:49

Diagnostics Check	OK	07:47
Air Blank	0.000	07:48
Control Test	0.078	07:48
Air Blank	0.000	07:48
Subject Sample # 1	0.239	07:49
Breath Volume	4.812	07:49
Air Blank	0.000	07:50
Air Blank	0.000	07:51
Subject Sample # 2	0.238	07:52
Breath Volume	4.148	07:52
Air Blank	0.000	07:52
Control Test	0.078	07:53
Air Blank	0.000	07:53
Diagnostics Check	OK	07:53

Data Download Date/Time: 09/28/2022 16:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: C0R0NADO

DATE/TIME OF TEST: 08/12/2022 23:22:54

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.081	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.175	23:18
Breath Volume	1.957	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:20
Subject Sample # 2	0.186	23:21
Breath Volume	1.777	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 09/14/2022 21:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BROWN,LORENZO
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SUPASWUD

DATE/TIME OF TEST: 08/14/2022 01:43:14

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.161	01:37
Breath Volume	2.011	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.162	01:41
Breath Volume	1.804	01:41
Air Blank	0.000	01:42
Control Test	0.082	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 09/14/2022 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: ESTRADA GARCIA

DATE/TIME OF TEST: 08/19/2022 02:58:16

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.183	02:53
Breath Volume	2.390	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.169	02:56
Breath Volume	1.863	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 09/14/2022 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MARTINEZ QUINONES

DATE/TIME OF TEST: 08/23/2022 04:36:30

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.081	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.000	04:32
Breath Volume	2.371	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.000	04:35
Breath Volume	1.406	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 09/14/2022 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILMAN,DUSTIN,D
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/28/2022 04:42:40

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:40	
Control Test	0.080	04:40	
Air Blank	0.000	04:41	
Subject Sample	# 1	04:41	Subject Test Refused
Breath Volume	0.000	04:41	
Air Blank	0.000	04:41	
Control Test	0.080	04:42	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 09/14/2022 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN,JONATHAN,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MOFFTT,DANIEL

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: OLEARY

DATE/TIME OF TEST: 08/02/2022 03:31:25

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.137	03:25
Breath Volume	1.996	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.129	03:29
Breath Volume	1.929	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 09/18/2022 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: OTTO,ALISON,

BREATH TEST OPERATOR AGENCY: Melbourne PD

ARRESTING OFFICER: OTTO,ALISON

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 08:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: POTEAT

DATE/TIME OF TEST: 08/09/2022 10:39:24

Diagnostics Check	OK	10:33
Air Blank	0.000	10:33
Control Test	0.082	10:33
Air Blank	0.000	10:34
Subject Sample # 1	0.000	10:35
Breath Volume	1.554	10:35
Air Blank	0.000	10:35
Air Blank	0.000	10:37
Subject Sample # 2	0.000	10:38
Breath Volume	1.203	10:38
Air Blank	0.000	10:38
Control Test	0.079	10:38
Air Blank	0.000	10:39
Diagnostics Check	OK	10:39

Data Download Date/Time: 09/18/2022 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN,JONATHAN,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: SALAZAR,V

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: HERNANDEZ-CUELLAR

DATE/TIME OF TEST: 08/10/2022 01:37:40

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.085	01:33
Breath Volume	1.347	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.083	01:36
Breath Volume	1.453	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 09/18/2022 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCDONALD, COLE, M

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MCDONALD, COLE

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 20:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 08/11/2022 20:32:28

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.080	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.086	20:27
Breath Volume	2.082	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.087	20:30
Breath Volume	1.792	20:30
Air Blank	0.000	20:31
Control Test	0.079	20:31
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 09/18/2022 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCDONALD,COLE,M

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: JUAN,CASTRO

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/11/2022 22:06:09

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.056	22:01
Breath Volume	1.765	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.056	22:04
Breath Volume	2.250	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 09/18/2022 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIDGEON,ANDREW,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: REARDON,JEFF
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 08/16/2022 09:05:34

Diagnostics Check	OK	08:58
Air Blank	0.000	08:59
Control Test	0.079	08:59
Air Blank	0.000	09:00
Subject Sample # 1	0.072	09:01
Breath Volume	1.652	09:01
Air Blank	0.000	09:01
Air Blank	0.000	09:03
Subject Sample # 2	0.066	09:04
Breath Volume	1.542	09:04
Air Blank	0.000	09:04
Control Test	0.079	09:04
Air Blank	0.000	09:05
Diagnostics Check	OK	09:05

Data Download Date/Time: 09/18/2022 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: STICKNEY

DATE/TIME OF TEST: 08/20/2022 01:16:08

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.080	01:15	
Air Blank		01:15	RFI Detect
Air Blank	0.000	01:16	

Data Download Date/Time: 09/18/2022 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: STICKNEY

DATE/TIME OF TEST: 08/20/2022 01:18:00

Diagnostics Check	OK	01:17	
Air Blank		01:17	RFI Detect
Air Blank	0.000	01:18	

Data Download Date/Time: 09/18/2022 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: STICKNEY

DATE/TIME OF TEST: 08/20/2022 01:25:11

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.080	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.000	01:20
Breath Volume	3.066	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.000	01:23
Breath Volume	3.566	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 09/18/2022 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: TRIMBOLI,MELANIE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 08/21/2022 08:57:12

Diagnostics Check	OK	08:51
Air Blank	0.000	08:51
Control Test	0.080	08:51
Air Blank	0.000	08:52
Subject Sample # 1	0.000	08:52
Breath Volume	1.250	08:52
Air Blank	0.000	08:53
Air Blank	0.000	08:55
Subject Sample # 2	0.000	08:55
Breath Volume	1.105	08:55
Air Blank	0.000	08:56
Control Test	0.080	08:56
Air Blank	0.000	08:57
Diagnostics Check	OK	08:57

Data Download Date/Time: 09/18/2022 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOOREG,GREG,C

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MEEHAN,ROBERT

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 17:07

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: DITMARS

DATE/TIME OF TEST: 08/23/2022 17:53:15

Diagnostics Check	OK	17:46
Air Blank	0.000	17:46
Control Test	0.079	17:47
Air Blank	0.000	17:47
Subject Sample # 1	0.000	17:48
Breath Volume	1.554	17:48
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample # 2	0.000	17:51
Breath Volume	1.550	17:51
Air Blank	0.000	17:52
Control Test	0.078	17:52
Air Blank	0.000	17:53
Diagnostics Check	OK	17:53

Data Download Date/Time: 09/18/2022 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN,JONATHAN,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: TRAMMEL,M

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/24/2022 00:41:55

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.108	00:35
Breath Volume	1.171	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.110	00:40
Breath Volume	1.238	00:40
Air Blank	0.000	00:41
Control Test	0.078	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 09/18/2022 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: RADIG,T
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GRUMBLES

DATE/TIME OF TEST: 08/24/2022 23:33:22

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.079	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.196	23:26
Breath Volume	1.101	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:28
Subject Sample # 2	0.182	23:31
Breath Volume	1.593	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 09/18/2022 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCDONALD,COLE,M

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MCDONALD,COLE

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: GUERNSEY

DATE/TIME OF TEST: 08/26/2022 22:12:42

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:09	
Control Test	0.080	22:10	
Air Blank	0.000	22:10	
Subject Sample	# 1	22:11	Subject Test Refused
Breath Volume	0.000	22:11	
Air Blank	0.000	22:11	
Control Test	0.079	22:12	
Air Blank	0.000	22:12	
Diagnostics Check	OK	22:12	

Data Download Date/Time: 09/18/2022 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCDONALD, COLE, M

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MCDONALD, COLE

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: GUERNSEY

DATE/TIME OF TEST: 08/26/2022 22:20:56

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.079	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.194	22:16
Breath Volume	1.917	22:16
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.187	22:19
Breath Volume	1.953	22:19
Air Blank	0.000	22:20
Control Test	0.079	22:20
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 09/18/2022 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: ARUNDEL

DATE/TIME OF TEST: 08/03/2022 03:19:24

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:17	
Control Test	0.077	03:17	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.077	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 08/05/2022 02:39:24

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.077	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.145	02:35
Breath Volume	2.039	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.137	02:37
Breath Volume	1.261	02:37
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 08/07/2022 19:31:34

Diagnostics Check	OK	19:25
Air Blank	0.000	19:26
Control Test	0.078	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.074	19:27
Breath Volume	1.718	19:27
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.071	19:30
Breath Volume	1.253	19:30
Air Blank	0.000	19:30
Control Test	0.079	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: MATHOSIAN,M
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: PANZO

DATE/TIME OF TEST: 08/08/2022 22:37:21

Diagnostics Check	OK	22:29
Air Blank	0.000	22:30
Control Test	0.078	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.196	22:32
Breath Volume	1.550	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.211	22:35
Breath Volume	1.324	22:35
Air Blank	0.000	22:36
Control Test	0.078	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: UDDIN

DATE/TIME OF TEST: 08/11/2022 03:33:17

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.148	03:28
Breath Volume	1.292	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.150	03:31
Breath Volume	1.859	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 08/12/2022 01:49:29

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.078	01:47	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Control Test	0.077	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GARCIA LIBRADO

DATE/TIME OF TEST: 08/13/2022 03:04:51

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.078	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Control Test	0.078	03:04	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: DOUD

DATE/TIME OF TEST: 08/14/2022 03:10:18

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.205	03:05
Breath Volume	1.371	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.206	03:08
Breath Volume	2.148	03:08
Air Blank	0.000	03:09
Control Test	0.078	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: MARTIN MARTINEZ

DATE/TIME OF TEST: 08/19/2022 23:26:02

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.077	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.134	23:21
Breath Volume	1.269	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.135	23:24
Breath Volume	1.191	23:24
Air Blank	0.000	23:25
Control Test	0.077	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:26

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: DETRANO

DATE/TIME OF TEST: 08/20/2022 02:05:32

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.077	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.143	02:01
Breath Volume	1.957	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.139	02:04
Breath Volume	1.718	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/20/2022 03:32:08

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.185	03:27
Breath Volume	1.386	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.191	03:30
Breath Volume	1.257	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: SEALY

DATE/TIME OF TEST: 08/20/2022 05:40:15

Diagnostics Check	OK	05:33
Air Blank	0.000	05:34
Control Test	0.077	05:34
Air Blank	0.000	05:35
Subject Sample # 1	0.167	05:35
Breath Volume	1.578	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.161	05:38
Breath Volume	1.382	05:38
Air Blank	0.000	05:39
Control Test	0.078	05:39
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DUCHENY,WES
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MOHAN

DATE/TIME OF TEST: 08/23/2022 00:14:41

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:11	
Control Test	0.078	00:11	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:13	Subject Test Refused
Breath Volume	0.000	00:13	
Air Blank	0.000	00:13	
Control Test	0.078	00:14	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 08/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: NUCCI

DATE/TIME OF TEST: 08/29/2022 22:04:20

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.078	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.213	22:00
Breath Volume	1.640	22:00
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.212	22:02
Breath Volume	1.425	22:02
Air Blank	0.000	22:03
Control Test	0.078	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 09/13/2022 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 08/30/2022 03:03:04

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:52	
Control Test	0.077	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:56	No Sample Provided
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 2	03:01	No Sample Provided
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Control Test	0.076	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 09/13/2022 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: DE LA CAMPA,BRYANT
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SCOPPETTONE

DATE/TIME OF TEST: 08/03/2022 22:10:12

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:07	
Control Test	0.079	22:08	
Air Blank	0.000	22:08	
Subject Sample	# 1	22:08	Subject Test Refused
Breath Volume	0.000	22:08	
Air Blank	0.000	22:09	
Control Test	0.080	22:09	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

Data Download Date/Time: 08/30/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: HETZEL

DATE/TIME OF TEST: 08/05/2022 00:04:55

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:00
Breath Volume	1.988	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.000	00:03
Breath Volume	2.375	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 08/30/2022 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: VELASCO

DATE/TIME OF TEST: 08/08/2022 02:48:14

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.079	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:47	
Control Test	0.079	02:47	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 08/30/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DIEZ

DATE/TIME OF TEST: 08/13/2022 23:53:09

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.136	23:48
Breath Volume	1.484	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.139	23:51
Breath Volume	2.000	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 08/30/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: REZNIK

DATE/TIME OF TEST: 08/16/2022 03:41:12

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.070	03:36
Breath Volume	3.414	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.070	03:39
Breath Volume	3.730	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 08/30/2022 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 08/18/2022 01:17:40

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:15	
Control Test	0.078	01:15	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.079	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 08/30/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: MENA

DATE/TIME OF TEST: 08/18/2022 23:59:04

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.123	23:54
Breath Volume	1.386	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.133	23:57
Breath Volume	1.445	23:57
Air Blank	0.000	23:58
Control Test	0.080	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:59

Data Download Date/Time: 08/30/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: LAUR

DATE/TIME OF TEST: 08/20/2022 04:00:55

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.147	03:56
Breath Volume	1.328	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.143	03:59
Breath Volume	1.363	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 08/30/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 08/23/2022 00:03:35

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.079	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.079	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 08/30/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/24/2022 02:31:12

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.091	02:25
Breath Volume	1.273	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.091	02:29
Breath Volume	1.750	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 08/30/2022 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: BIZZELL

DATE/TIME OF TEST: 08/28/2022 20:19:30

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.079	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.027	20:15
Breath Volume	2.203	20:15
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.026	20:18
Breath Volume	2.277	20:18
Air Blank	0.000	20:18
Control Test	0.079	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 08/30/2022 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,DAVID,J
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: MORALES,DAVID
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MERLIN

DATE/TIME OF TEST: 08/31/2022 02:37:29

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:35	
Control Test	0.079	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:36	
Control Test	0.079	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 09/27/2022 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LYON,L
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SLADE

DATE/TIME OF TEST: 08/06/2022 00:32:02

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.177	00:27
Breath Volume	3.230	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.177	00:30
Breath Volume	3.792	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:32

Data Download Date/Time: 09/16/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: WOOTEN

DATE/TIME OF TEST: 08/06/2022 05:22:39

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:20	
Control Test	0.080	05:20	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:21	Subject Test Refused
Breath Volume	0.000	05:21	
Air Blank	0.000	05:21	
Control Test	0.080	05:22	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

Data Download Date/Time: 09/16/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 08/11/2022 01:45:52

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Control Test	0.080	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 09/16/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: VOLLENWEIDER

DATE/TIME OF TEST: 08/12/2022 00:42:09

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:39	
Control Test	0.080	00:40	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:40	Subject Test Refused
Breath Volume	0.000	00:40	
Air Blank	0.000	00:41	
Control Test	0.080	00:41	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 09/16/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: FELDER

DATE/TIME OF TEST: 08/12/2022 04:17:50

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.000	04:12
Breath Volume	1.214	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.000	04:16
Breath Volume	1.980	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 09/16/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: SOLANO SILVA

DATE/TIME OF TEST: 08/21/2022 04:05:00

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.155	04:00
Breath Volume	2.082	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.158	04:03
Breath Volume	2.039	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:05

Data Download Date/Time: 09/16/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: LAMBRIGHT

DATE/TIME OF TEST: 08/27/2022 20:31:52

Diagnostics Check	OK	20:23
Air Blank	0.000	20:24
Control Test	0.079	20:24
Air Blank	0.000	20:24
Subject Sample # 1	0.128	20:26
Breath Volume	2.074	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:28
Subject Sample # 2	0.131	20:30
Breath Volume	2.156	20:30
Air Blank	0.000	20:30
Control Test	0.078	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 09/16/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 08/27/2022 23:23:55

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.080	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:22	
Control Test	0.079	23:23	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 09/16/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN, C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: FERRONE

DATE/TIME OF TEST: 08/28/2022 02:48:31

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.247	02:43
Breath Volume	2.070	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.249	02:46
Breath Volume	1.621	02:46
Air Blank	0.000	02:47
Control Test	0.078	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 09/16/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GOLDMAN

DATE/TIME OF TEST: 08/28/2022 04:56:50

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.139	04:52
Breath Volume	3.082	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.140	04:55
Breath Volume	3.406	04:55
Air Blank	0.000	04:55
Control Test	0.080	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 09/16/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 08/31/2022 20:52:25

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:50	
Control Test	0.079	20:50	
Air Blank	0.000	20:50	
Subject Sample	# 1	20:51	Subject Test Refused
Breath Volume	0.000	20:51	
Air Blank	0.000	20:51	
Control Test	0.079	20:51	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 09/16/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: LANDRY

DATE/TIME OF TEST: 08/27/2022 20:49:51

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:35	
Control Test	0.078	20:35	
Air Blank	0.000	20:35	
Subject Sample	# 1 0.253	20:37	
Breath Volume	2.324	20:37	
Air Blank	0.000	20:38	
Air Blank	0.000	20:40	
Subject Sample	# 2 0.201	20:43	Volume Not Met
Breath Volume	0.136	20:43	
Air Blank	0.000	20:44	
Air Blank	0.000	20:45	
Subject Sample	# 3 0.241	20:48	
Breath Volume	2.117	20:48	
Air Blank	0.000	20:48	
Control Test	0.077	20:49	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 08/30/2022 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMELMAN, TYLER,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BROWN, JOHN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: DUKES

DATE/TIME OF TEST: 08/19/2022 03:09:10

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.143	03:04
Breath Volume	1.386	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.137	03:07
Breath Volume	1.574	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 08/22/2022 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CONNELL,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: RICK

DATE/TIME OF TEST: 08/20/2022 13:38:23

Diagnostics Check	OK	13:31
Air Blank	0.000	13:32
Control Test	0.081	13:32
Air Blank	0.000	13:32
Subject Sample # 1	0.267	13:33
Breath Volume	3.207	13:33
Air Blank	0.000	13:34
Air Blank	0.000	13:36
Subject Sample # 2	0.273	13:36
Breath Volume	3.031	13:36
Air Blank	0.000	13:37
Control Test	0.080	13:37
Air Blank	0.000	13:38
Diagnostics Check	OK	13:38

Data Download Date/Time: 08/22/2022 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMELMAN, TYLER,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAMMELMAN, TYLER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: STINEHELPER

DATE/TIME OF TEST: 08/27/2022 01:25:28

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.146	01:20
Breath Volume	1.175	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.147	01:23
Breath Volume	1.777	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 09/27/2022 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 08/27/2022 05:23:58

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.080	05:18
Air Blank	0.000	05:19
Subject Sample # 1	0.102	05:19
Breath Volume	3.605	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.105	05:22
Breath Volume	3.660	05:22
Air Blank	0.000	05:23
Control Test	0.081	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 09/27/2022 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZIZ,DAVID,I
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: AZIZ,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 08/29/2022 05:44:30

Diagnostics Check	OK	05:37
Air Blank	0.000	05:38
Control Test	0.081	05:38
Air Blank	0.000	05:39
Subject Sample # 1	0.146	05:40
Breath Volume	1.304	05:40
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.155	05:42
Breath Volume	2.074	05:42
Air Blank	0.000	05:43
Control Test	0.081	05:43
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 09/27/2022 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS, VIOLA, L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FERNANDEZ
VILLANUEVA, AMADEO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Female

OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 08/29/2022 06:13:51

Diagnostics Check	OK	06:04	
Air Blank	0.000	06:04	
Control Test	0.081	06:04	
Air Blank	0.000	06:05	
Subject Sample	# 1 0.000	06:08	Volume Not Met
Breath Volume	0.761	06:08	
Air Blank	0.000	06:08	
Air Blank	0.000	06:10	
Subject Sample	# 2	06:13	Sequence Aborted
Breath Volume	0.777	06:13	
Air Blank	0.000	06:13	

Data Download Date/Time: 09/27/2022 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNYDER,ADRIANNE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 08/30/2022 21:36:28

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.081	21:28
Air Blank	0.000	21:28
Subject Sample # 1	0.179	21:30
Breath Volume	2.488	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.159	21:34
Breath Volume	2.347	21:34
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 09/27/2022 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRODY,A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: SCOTT,BRODY
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: ORLANDO

DATE/TIME OF TEST: 08/08/2022 15:08:20

Diagnostics Check	OK	15:00
Air Blank	0.000	15:00
Control Test	0.079	15:01
Air Blank	0.000	15:01
Subject Sample # 1	0.141	15:02
Breath Volume	1.882	15:02
Air Blank	0.000	15:02
Air Blank	0.000	15:04
Subject Sample # 2	0.134	15:06
Breath Volume	2.191	15:06
Air Blank	0.000	15:07
Control Test	0.079	15:07
Air Blank	0.000	15:08
Diagnostics Check	OK	15:08

Data Download Date/Time: 08/29/2022 14:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: CAJUSTE,ALISHIA
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: ARCHE

DATE/TIME OF TEST: 08/28/2022 03:28:24

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1 0.283	03:23	
Breath Volume	2.359	03:23	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample	# 2	03:26	Slope Not Met
Breath Volume	1.570	03:26	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 08/29/2022 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: CAJUSTE,ALISHIA
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ARCHE

DATE/TIME OF TEST: 08/28/2022 04:03:44

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.257	03:59
Breath Volume	2.187	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.250	04:02
Breath Volume	2.585	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 08/29/2022 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 08/02/2022 03:11:58

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:04	No Sample Provided
Breath Volume	0.000	03:04	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 2	03:10	No Sample Provided
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.079	03:11	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 08/08/2022 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: LEAKE

DATE/TIME OF TEST: 08/05/2022 23:11:32

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.078	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.275	23:07
Breath Volume	2.464	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.280	23:09
Breath Volume	1.828	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 08/08/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: REGALADO

DATE/TIME OF TEST: 08/06/2022 02:59:22

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.078	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:48	No Sample Provided
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:51	
Subject Sample	# 2	0.224	02:52
Breath Volume	1.867	02:52	
Air Blank	0.000	02:52	
Air Blank	0.000	02:54	
Subject Sample	# 3	0.234	02:57
Breath Volume	1.910	02:57	
Air Blank	0.000	02:58	
Control Test	0.079	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 08/08/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: QUATTROCHI

DATE/TIME OF TEST: 08/07/2022 00:41:46

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.000	00:37
Breath Volume	3.625	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.000	00:40
Breath Volume	3.335	00:40
Air Blank	0.000	00:40
Control Test	0.079	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 08/08/2022 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: BARBOSA

DATE/TIME OF TEST: 08/07/2022 04:55:49

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.080	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.117	04:51
Breath Volume	3.992	04:51
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.119	04:54
Breath Volume	3.644	04:54
Air Blank	0.000	04:54
Control Test	0.080	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 08/08/2022 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: CALVILLO

DATE/TIME OF TEST: 08/07/2022 23:08:43

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.000	23:04
Breath Volume	1.371	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.000	23:07
Breath Volume	1.812	23:07
Air Blank	0.000	23:07
Control Test	0.079	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 08/24/2022 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HARGER

DATE/TIME OF TEST: 08/08/2022 03:06:08

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.219	03:01
Breath Volume	2.511	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.205	03:04
Breath Volume	1.707	03:04
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 08/24/2022 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 08/16/2022 02:38:58

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.088	02:34
Breath Volume	2.003	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.088	02:37
Breath Volume	2.078	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 08/24/2022 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON, K
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 08/17/2022 00:47:54

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.078	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.271	00:40
Breath Volume	1.671	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.301	00:42
Breath Volume	1.277	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 3	0.304	00:46
Breath Volume	1.457	00:46
Air Blank	0.000	00:46
Control Test	0.077	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 08/24/2022 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: VANATTA

DATE/TIME OF TEST: 08/17/2022 01:19:42

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:08	
Control Test	0.079	01:09	
Air Blank	0.000	01:09	
Subject Sample	# 1	01:12	No Sample Provided
Breath Volume	0.000	01:12	
Air Blank	0.000	01:13	
Air Blank	0.000	01:15	
Subject Sample	# 2	01:18	No Sample Provided
Breath Volume	0.000	01:18	
Air Blank	0.000	01:18	
Control Test	0.079	01:19	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 08/24/2022 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STEPHENS, T
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LUCIANO

DATE/TIME OF TEST: 08/20/2022 04:26:36

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:24	
Control Test	0.078	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Control Test	0.079	04:26	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 08/24/2022 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON, K
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: EVANGELISTA

DATE/TIME OF TEST: 08/20/2022 23:17:08

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.079	23:14	
Air Blank	0.000	23:15	
Subject Sample	# 1	23:15	Subject Test Refused
Breath Volume	0.000	23:15	
Air Blank	0.000	23:16	
Control Test	0.080	23:16	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 08/24/2022 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TONY, T
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STEPHENS, T
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BRETON

DATE/TIME OF TEST: 08/21/2022 03:48:20

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.099	03:42
Breath Volume	1.554	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.097	03:46
Breath Volume	1.492	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 08/24/2022 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: RIGGALL,NICHOLAS

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 07/26/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 04:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: COSTE

DATE/TIME OF TEST: 08/21/2022 05:17:11

Diagnostics Check	OK	05:10
Air Blank	0.000	05:11
Control Test	0.078	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.016	05:12
Breath Volume	4.101	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.016	05:15
Breath Volume	4.152	05:15
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 08/24/2022 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:58

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: ZENKER

DATE/TIME OF TEST: 08/25/2022 23:39:43

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:37	
Control Test	0.078	23:37	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:39	RFI Detect
Breath Volume	0.093	23:39	
Air Blank		23:39	RFI Detect
Air Blank	0.000	23:39	

Data Download Date/Time: 09/15/2022 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:58

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: ZENKER

DATE/TIME OF TEST: 08/25/2022 23:45:08

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:42	
Control Test	0.078	23:43	
Air Blank	0.000	23:43	
Subject Sample	# 1	23:43	RFI Detect
Breath Volume	0.000	23:43	
Air Blank	0.000	23:44	
Control Test	0.078	23:44	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 09/15/2022 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:58

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: ZENKER

DATE/TIME OF TEST: 08/25/2022 23:55:43

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.077	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.262	23:50
Breath Volume	2.128	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:52
Subject Sample # 2	0.255	23:54
Breath Volume	1.585	23:54
Air Blank	0.000	23:54
Control Test	0.077	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 09/15/2022 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:31

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: HERNANDEZ-CRUZ

DATE/TIME OF TEST: 08/27/2022 03:00:22

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.183	02:54
Breath Volume	1.839	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.174	02:58
Breath Volume	1.496	02:58
Air Blank	0.000	02:59
Control Test	0.078	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 09/15/2022 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: WILSON,KALEB

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: CROW

DATE/TIME OF TEST: 08/27/2022 03:44:20

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.142	03:39
Breath Volume	3.457	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.146	03:42
Breath Volume	4.875	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 09/15/2022 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOMACK,ERIC,V
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: WOMACK,ERIC
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:52

SUBJECT'S LAST NAME: PUPO

DATE/TIME OF TEST: 08/12/2022 14:21:48

Diagnostics Check	OK	14:15
Air Blank	0.000	14:16
Control Test	0.080	14:16
Air Blank	0.000	14:17
Subject Sample # 1	0.000	14:17
Breath Volume	2.054	14:17
Air Blank	0.000	14:17
Air Blank	0.000	14:19
Subject Sample # 2	0.000	14:20
Breath Volume	3.578	14:20
Air Blank	0.000	14:20
Control Test	0.080	14:21
Air Blank	0.000	14:21
Diagnostics Check	OK	14:21

Data Download Date/Time: 09/30/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALINAS,JOSE,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: SALINAS,JOSE
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 08/13/2022 19:41:35

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.080	19:34
Air Blank	0.000	19:35
Subject Sample # 1	0.348	19:36
Breath Volume	1.261	19:36
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.353	19:39
Breath Volume	1.875	19:39
Air Blank	0.000	19:40
Control Test	0.079	19:40
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 09/30/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANDENBERG,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: NUNEZ,GEORGE
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 08/13/2022 06:33:51

Diagnostics Check	OK	06:20	
Air Blank	0.000	06:21	
Control Test	0.080	06:21	
Air Blank	0.000	06:22	
Subject Sample	# 1	06:25	No Sample Provided
Breath Volume	0.000	06:25	
Air Blank	0.000	06:25	
Air Blank	0.000	06:27	
Subject Sample	# 2	0.309	06:28
Breath Volume	2.058	06:28	
Air Blank	0.000	06:29	
Air Blank	0.000	06:31	
Subject Sample	# 3	0.315	06:32
Breath Volume	1.187	06:32	
Air Blank	0.000	06:32	
Control Test	0.078	06:33	
Air Blank	0.000	06:33	
Diagnostics Check	OK	06:33	

Data Download Date/Time: 09/16/2022 21:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENNELL,RUSSELL,M
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PENNELL,RUSSELL
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: BATTLES

DATE/TIME OF TEST: 08/01/2022 02:16:38

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.079	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.113	02:15	
Air Blank	0.000	02:16	
		02:16	RFI Detect
Air Blank	0.000	02:16	

Data Download Date/Time: 09/01/2022 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGEL,JASON,S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: BERRIOS,JOSE
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: ODEH

DATE/TIME OF TEST: 08/06/2022 03:56:03

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.141	03:51
Breath Volume	2.808	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.151	03:54
Breath Volume	2.757	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:56

Data Download Date/Time: 09/01/2022 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGEL,JASON,S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: BERRIOS,JOSE
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 08/19/2022 21:07:06

Diagnostics Check	OK	20:59
Air Blank	0.000	20:59
Control Test	0.080	21:00
Air Blank	0.000	21:00
Subject Sample # 1	0.185	21:02
Breath Volume	1.878	21:02
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.192	21:05
Breath Volume	2.003	21:05
Air Blank	0.000	21:06
Control Test	0.080	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 09/01/2022 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,S,
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: BROWN,J
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: HIGGINBOTHAM

DATE/TIME OF TEST: 08/20/2022 01:28:12

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.078	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.194	01:23
Breath Volume	2.605	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.186	01:26
Breath Volume	1.906	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 09/30/2022 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: PENA GRANADINO

DATE/TIME OF TEST: 08/13/2022 01:20:03

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.215	01:14
Breath Volume	1.144	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.214	01:18
Breath Volume	1.218	01:18
Air Blank	0.000	01:19
Control Test	0.078	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:20

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: VILLEGAS HERRERA

DATE/TIME OF TEST: 08/13/2022 01:29:37

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.078	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.161	01:25
Breath Volume	2.746	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.149	01:28
Breath Volume	1.648	01:28
Air Blank	0.000	01:28
Control Test	0.078	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 08/13/2022 02:09:45

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.077	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.135	02:04
Breath Volume	1.769	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.138	02:08
Breath Volume	2.667	02:08
Air Blank	0.000	02:08
Control Test	0.076	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: SPROULL

DATE/TIME OF TEST: 08/14/2022 01:03:09

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.078	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:56	No Sample Provided
Breath Volume	0.000	00:56	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample	# 2	01:01	No Sample Provided
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Control Test	0.078	01:02	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,HENRY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: JOVANOVIC

DATE/TIME OF TEST: 08/14/2022 02:09:00

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.184	02:04
Breath Volume	1.289	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.188	02:07
Breath Volume	1.828	02:07
Air Blank	0.000	02:08
Control Test	0.078	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:09

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 08/14/2022 02:30:49

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.077	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.238	02:23
Breath Volume	3.234	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.217	02:26
Breath Volume	3.031	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 3	0.233	02:29
Breath Volume	3.046	02:29
Air Blank	0.000	02:29
Control Test	0.076	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: PANTOJAS-RAMIREZ

DATE/TIME OF TEST: 08/14/2022 19:07:46

Diagnostics Check	OK	19:00
Air Blank	0.000	19:00
Control Test	0.078	19:00
Air Blank	0.000	19:01
Subject Sample # 1	0.175	19:02
Breath Volume	1.148	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.177	19:06
Breath Volume	2.644	19:06
Air Blank	0.000	19:06
Control Test	0.078	19:07
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,S,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,S
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: KERN

DATE/TIME OF TEST: 08/15/2022 15:39:50

Diagnostics Check	OK	15:24	
Air Blank	0.000	15:24	
Control Test	0.078	15:25	
Air Blank	0.000	15:25	
Subject Sample # 1	0.151	15:27	
Breath Volume	1.468	15:27	
Air Blank	0.000	15:28	
Air Blank	0.000	15:29	
Subject Sample # 2	0.111	15:32	Volume Not Met
Breath Volume	0.785	15:32	
Air Blank	0.000	15:33	
Air Blank	0.000	15:35	
Subject Sample # 3		15:38	No Sample Provided
Breath Volume	0.000	15:38	
Air Blank	0.000	15:38	
Control Test	0.075	15:39	
Air Blank	0.000	15:39	
Diagnostics Check	OK	15:39	

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OBRIEN,DANIEL,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: OBRIEN,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/16/2022 02:02:19

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.215	01:55
Breath Volume	1.625	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:57
Subject Sample # 2	0.195	02:00
Breath Volume	1.101	02:00
Air Blank	0.000	02:01
Control Test	0.076	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ZILBA

DATE/TIME OF TEST: 08/18/2022 01:13:04

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.078	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.171	01:08
Breath Volume	2.292	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.167	01:11
Breath Volume	1.234	01:11
Air Blank	0.000	01:12
Control Test	0.077	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:13

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 08/18/2022 22:27:16

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.078	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.230	22:22
Breath Volume	2.222	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.234	22:25
Breath Volume	1.988	22:25
Air Blank	0.000	22:26
Control Test	0.078	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: FREEDOM

DATE/TIME OF TEST: 08/20/2022 03:57:06

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.079	03:52
Breath Volume	2.382	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.079	03:55
Breath Volume	2.703	03:55
Air Blank	0.000	03:56
Control Test	0.077	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: PINOS

DATE/TIME OF TEST: 08/21/2022 03:44:09

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.078	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample	# 2	03:42	No Sample Provided
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.078	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: PINOS

DATE/TIME OF TEST: 08/21/2022 03:59:52

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.078	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1 0.168	03:50	Volume Not Met
Breath Volume	0.066	03:50	
Air Blank	0.000	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 2 0.181	03:53	
Breath Volume	1.199	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample	# 3 0.167	03:58	
Breath Volume	1.636	03:58	
Air Blank	0.000	03:58	
Control Test	0.074	03:59	Control Outside Tolerance
Air Blank	0.000	03:59	

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: PINOS

DATE/TIME OF TEST: 08/21/2022 04:09:13

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.076	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.171	04:04
Breath Volume	1.578	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:06
Subject Sample # 2	0.174	04:07
Breath Volume	1.246	04:07
Air Blank	0.000	04:08
Control Test	0.076	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,S,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,S
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 08/24/2022 09:53:57

Diagnostics Check	OK	09:46
Air Blank	0.000	09:47
Control Test	0.078	09:47
Air Blank	0.000	09:48
Subject Sample # 1	0.000	09:49
Breath Volume	2.222	09:49
Air Blank	0.000	09:49
Air Blank	0.000	09:51
Subject Sample # 2	0.000	09:52
Breath Volume	2.812	09:52
Air Blank	0.000	09:53
Control Test	0.077	09:53
Air Blank	0.000	09:53
Diagnostics Check	OK	09:53

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SAYLES

DATE/TIME OF TEST: 08/26/2022 01:18:04

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.079	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:16	No Sample Provided
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Air Blank		01:18	RFI Detect

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HERNADEZ-PEREZ,IVAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HERNANDEZ DELCID

DATE/TIME OF TEST: 08/29/2022 03:32:57

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.077	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.183	03:28
Breath Volume	2.328	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.173	03:31
Breath Volume	1.480	03:31
Air Blank	0.000	03:32
Control Test	0.077	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: BUTSON

DATE/TIME OF TEST: 08/30/2022 22:09:15

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:06	
Control Test	0.077	22:07	
Air Blank	0.000	22:07	
Subject Sample	# 1	22:07	Subject Test Refused
Breath Volume	0.000	22:07	
Air Blank	0.000	22:08	
Control Test	0.078	22:08	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BIEBER,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: SOHAIL

DATE/TIME OF TEST: 08/31/2022 04:30:32

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.181	04:25
Breath Volume	2.312	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.179	04:28
Breath Volume	1.644	04:28
Air Blank	0.000	04:29
Control Test	0.077	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 09/30/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: GUTHAS,SEAN
ARRESTING OFFICER AGENCY: Bowling Green PD
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: CRUZ-VALDOVINOS

DATE/TIME OF TEST: 08/27/2022 18:38:37

Diagnostics Check	OK	18:27	
Air Blank	0.000	18:27	
Control Test	0.080	18:27	
Air Blank	0.000	18:28	
Subject Sample	# 1 0.229	18:31	Volume Not Met
Breath Volume	0.109	18:31	
Air Blank	0.000	18:32	
Air Blank	0.000	18:33	
Subject Sample	# 2 0.209	18:37	Volume Not Met
Breath Volume	0.050	18:37	
Air Blank	0.000	18:37	
Control Test	0.078	18:38	
Air Blank	0.000	18:38	
Diagnostics Check	OK	18:38	

Data Download Date/Time: 09/24/2022 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: EVERSOLE,DONNIE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: FINK

DATE/TIME OF TEST: 08/30/2022 22:39:56

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:38	
Control Test	0.079	22:38	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:39	
		22:39	RFI Detect
Air Blank		22:39	RFI Detect

Data Download Date/Time: 09/24/2022 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: EVERSOLE,DONNIE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: FINK

DATE/TIME OF TEST: 08/30/2022 22:46:16

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:43	
Control Test	0.080	22:44	
Air Blank	0.000	22:44	
Subject Sample	# 1	22:44	Subject Test Refused
Breath Volume	0.000	22:44	
Air Blank	0.000	22:45	
Control Test	0.080	22:45	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 09/24/2022 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: EDMONDS,RICHARD,J

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: HERNADEZ,IVAN

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: DOOLEY

DATE/TIME OF TEST: 08/02/2022 01:55:27

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.207	01:51
Breath Volume	2.687	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.223	01:53
Breath Volume	2.804	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 08/17/2022 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,SHANE,C
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: NESTOR,A
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: FORTUNATO

DATE/TIME OF TEST: 08/02/2022 16:08:56

Diagnostics Check	OK	16:05	
Air Blank	0.000	16:06	
Control Test	0.079	16:06	
Air Blank	0.000	16:07	
Subject Sample	# 1	16:07	Subject Test Refused
Breath Volume	0.000	16:07	
Air Blank	0.000	16:08	
Control Test	0.079	16:08	
Air Blank	0.000	16:08	
Diagnostics Check	OK	16:08	

Data Download Date/Time: 08/17/2022 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 18:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: HAEHNEL

DATE/TIME OF TEST: 08/04/2022 19:56:11

Diagnostics Check	OK	19:47
Air Blank	0.000	19:48
Control Test	0.080	19:48
Air Blank	0.000	19:48
Subject Sample # 1	0.000	19:51
Breath Volume	2.339	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.000	19:54
Breath Volume	1.925	19:54
Air Blank	0.000	19:55
Control Test	0.080	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 08/17/2022 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: BAUERLE

DATE/TIME OF TEST: 08/11/2022 01:54:19

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.175	01:49
Breath Volume	2.425	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.166	01:52
Breath Volume	1.953	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 08/17/2022 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 19:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: KERN

DATE/TIME OF TEST: 08/11/2022 19:47:53

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:45	
Control Test	0.080	19:45	
Air Blank	0.000	19:46	
Subject Sample	# 1	19:46	Subject Test Refused
Breath Volume	0.000	19:46	
Air Blank	0.000	19:46	
Control Test	0.080	19:47	
Air Blank	0.000	19:47	
Diagnostics Check	OK	19:47	

Data Download Date/Time: 08/17/2022 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 08/12/2022 22:01:10

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:58	
Control Test	0.080	21:59	
Air Blank	0.000	21:59	
Subject Sample	# 1	21:59	Subject Test Refused
Breath Volume	0.000	21:59	
Air Blank	0.000	22:00	
Control Test	0.079	22:00	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 08/17/2022 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/06/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: PENA BOLTEADA

DATE/TIME OF TEST: 08/13/2022 02:06:53

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:04	
Control Test	0.080	02:04	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:05	
Control Test	0.080	02:06	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 08/17/2022 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: MOHAMMED

DATE/TIME OF TEST: 08/19/2022 00:50:30

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:49	Subject Test Refused
Breath Volume	0.000	00:49	
Air Blank	0.000	00:49	
Control Test	0.079	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 09/15/2022 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: DEDHO JACINTO

DATE/TIME OF TEST: 08/20/2022 03:37:12

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.128	03:32
Breath Volume	2.128	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.125	03:35
Breath Volume	1.359	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 09/15/2022 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: HOWEY

DATE/TIME OF TEST: 08/25/2022 23:19:11

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.080	23:16	
Air Blank	0.000	23:17	
Subject Sample	# 1	23:17	Subject Test Refused
Breath Volume	0.000	23:17	
Air Blank	0.000	23:18	
Control Test	0.080	23:18	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 09/15/2022 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 08/26/2022 02:57:01

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.118	02:52
Breath Volume	3.507	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.119	02:55
Breath Volume	3.929	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:57

Data Download Date/Time: 09/15/2022 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:16

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: NOVAK

DATE/TIME OF TEST: 08/27/2022 00:03:47

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.225	23:59
Breath Volume	1.441	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.232	00:02
Breath Volume	2.062	00:02
Air Blank	0.000	00:02
Control Test	0.080	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 09/15/2022 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MCMILLER

DATE/TIME OF TEST: 08/27/2022 22:52:51

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.052	22:48
Breath Volume	2.140	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.050	22:51
Breath Volume	2.761	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 09/15/2022 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 08/08/2022 03:31:06

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.182	03:26
Breath Volume	2.535	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.188	03:29
Breath Volume	2.453	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 08/22/2022 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUES, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: RODRIGUES, MICHAEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: ANDREU

DATE/TIME OF TEST: 08/13/2022 18:17:50

Diagnostics Check	OK	18:11
Air Blank	0.000	18:11
Control Test	0.081	18:12
Air Blank	0.000	18:12
Subject Sample # 1	0.144	18:13
Breath Volume	2.179	18:13
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.137	18:16
Breath Volume	1.710	18:16
Air Blank	0.000	18:16
Control Test	0.080	18:17
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

Data Download Date/Time: 08/22/2022 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/14/2022 03:12:08

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.188	03:06
Breath Volume	1.917	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.174	03:10
Breath Volume	1.917	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 08/22/2022 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,W,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:11

SUBJECT'S LAST NAME: RHODE

DATE/TIME OF TEST: 08/14/2022 09:38:04

Diagnostics Check	OK	09:31
Air Blank	0.000	09:32
Control Test	0.080	09:32
Air Blank	0.000	09:32
Subject Sample # 1	0.248	09:33
Breath Volume	1.160	09:33
Air Blank	0.000	09:34
Air Blank	0.000	09:35
Subject Sample # 2	0.260	09:36
Breath Volume	2.046	09:36
Air Blank	0.000	09:37
Control Test	0.080	09:37
Air Blank	0.000	09:37
Diagnostics Check	OK	09:38

Data Download Date/Time: 08/22/2022 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 08/15/2022 00:45:19

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.081	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Control Test	0.081	00:44	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

Data Download Date/Time: 08/22/2022 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 08/21/2022 22:32:24

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.081	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.157	22:27
Breath Volume	2.136	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.149	22:30
Breath Volume	1.394	22:30
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 08/22/2022 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 08/31/2022 03:53:11

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:45	
Control Test	0.081	03:45	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:49	No Sample Provided
Breath Volume	0.000	03:49	
Air Blank	0.000	03:49	
Air Blank	0.000	03:51	
Subject Sample	# 2	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 09/27/2022 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LITTLE,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: SPAIN

DATE/TIME OF TEST: 08/05/2022 16:07:12

Diagnostics Check	OK	16:06	
Air Blank	0.000	16:06	
Control Test	0.080	16:06	
Air Blank		16:07	RFI Detect
Air Blank		16:07	RFI Detect

Data Download Date/Time: 08/09/2022 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LITTLE,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: SPAIN

DATE/TIME OF TEST: 08/05/2022 16:15:20

Diagnostics Check	OK	16:09
Air Blank	0.000	16:09
Control Test	0.080	16:09
Air Blank	0.000	16:10
Subject Sample # 1	0.208	16:10
Breath Volume	2.210	16:10
Air Blank	0.000	16:11
Air Blank	0.000	16:13
Subject Sample # 2	0.202	16:13
Breath Volume	2.273	16:13
Air Blank	0.000	16:14
Control Test	0.079	16:14
Air Blank	0.000	16:15
Diagnostics Check	OK	16:15

Data Download Date/Time: 08/09/2022 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,HECTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LOPEZ,HECTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MCELROY

DATE/TIME OF TEST: 08/07/2022 20:01:12

Diagnostics Check	OK	19:57	
Air Blank	0.000	19:57	
Control Test	0.080	19:58	
Air Blank	0.000	19:58	
Subject Sample	# 1	20:00	Sequence Aborted
Breath Volume	0.000	20:00	
Air Blank	0.000	20:01	

Data Download Date/Time: 08/12/2022 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS, TYLER, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: RANDOR

DATE/TIME OF TEST: 08/10/2022 21:29:14

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.163	21:24
Breath Volume	1.910	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.164	21:27
Breath Volume	1.765	21:27
Air Blank	0.000	21:28
Control Test	0.079	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 08/12/2022 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,HECTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LOPEZ,HECTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 15:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 08/13/2022 17:35:38

Diagnostics Check	OK	17:29
Air Blank	0.000	17:29
Control Test	0.081	17:30
Air Blank	0.000	17:30
Subject Sample # 1	0.258	17:31
Breath Volume	2.070	17:31
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample # 2	0.244	17:34
Breath Volume	2.234	17:34
Air Blank	0.000	17:34
Control Test	0.080	17:35
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 09/27/2022 01:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 08/13/2022 22:51:39

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.146	22:47
Breath Volume	1.800	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.140	22:50
Breath Volume	1.859	22:50
Air Blank	0.000	22:50
Control Test	0.079	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 09/27/2022 01:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,HECTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LOPEZ,HECTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 08/14/2022 20:57:27

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.080	20:50
Air Blank	0.000	20:50
Subject Sample # 1	0.243	20:52
Breath Volume	2.226	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:54
Subject Sample # 2	0.231	20:55
Breath Volume	1.664	20:55
Air Blank	0.000	20:56
Control Test	0.079	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 09/27/2022 01:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JAMIE,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WHITFIELD,CIARA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 08:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: OVENS

DATE/TIME OF TEST: 08/22/2022 08:45:16

Diagnostics Check	OK	08:38
Air Blank	0.000	08:38
Control Test	0.080	08:38
Air Blank	0.000	08:39
Subject Sample # 1	0.000	08:40
Breath Volume	1.132	08:40
Air Blank	0.000	08:41
Air Blank	0.000	08:43
Subject Sample # 2	0.000	08:43
Breath Volume	1.425	08:43
Air Blank	0.000	08:44
Control Test	0.080	08:44
Air Blank	0.000	08:45
Diagnostics Check	OK	08:45

Data Download Date/Time: 09/27/2022 01:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONANNO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SMITH,JAMES
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 07:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:49

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 08/25/2022 08:18:22

Diagnostics Check	OK	08:12
Air Blank	0.000	08:12
Control Test	0.081	08:13
Air Blank	0.000	08:13
Subject Sample # 1	0.216	08:14
Breath Volume	2.488	08:14
Air Blank	0.000	08:14
Air Blank	0.000	08:16
Subject Sample # 2	0.214	08:16
Breath Volume	2.128	08:16
Air Blank	0.000	08:17
Control Test	0.079	08:17
Air Blank	0.000	08:18
Diagnostics Check	OK	08:18

Data Download Date/Time: 09/27/2022 01:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,GEORGE,O
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LOPEZ,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 08/26/2022 19:44:45

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.080	19:38
Air Blank	0.000	19:38
Subject Sample # 1	0.118	19:40
Breath Volume	1.492	19:40
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.114	19:43
Breath Volume	1.449	19:43
Air Blank	0.000	19:43
Control Test	0.080	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 09/27/2022 01:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: REINHART

DATE/TIME OF TEST: 08/04/2022 02:23:47

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.082	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.176	02:19
Breath Volume	1.867	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.181	02:22
Breath Volume	2.464	02:22
Air Blank	0.000	02:22
Control Test	0.082	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 09/14/2022 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SCONIERS

DATE/TIME OF TEST: 08/19/2022 01:46:38

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.082	01:41	
Air Blank	0.000	01:41	
Subject Sample # 1	0.000	01:42	
Breath Volume	2.722	01:42	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample # 2	0.000	01:45	
Breath Volume	2.523	01:45	
Air Blank	0.000	01:45	
Control Test	0.087	01:46	Control Outside Tolerance
Air Blank	0.000	01:46	

Data Download Date/Time: 09/14/2022 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SCONIERS

DATE/TIME OF TEST: 08/19/2022 01:57:33

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.083	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.000	01:53
Breath Volume	3.183	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.000	01:56
Breath Volume	3.527	01:56
Air Blank	0.000	01:56
Control Test	0.084	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 09/14/2022 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 08/21/2022 03:25:02

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.082	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.210	03:20
Breath Volume	2.734	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:22
Subject Sample # 2	0.221	03:23
Breath Volume	2.875	03:23
Air Blank	0.000	03:24
Control Test	0.082	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:25

Data Download Date/Time: 09/14/2022 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SANTES

DATE/TIME OF TEST: 08/26/2022 02:42:33

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.082	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.237	02:37
Breath Volume	1.593	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.244	02:41
Breath Volume	3.050	02:41
Air Blank	0.000	02:41
Control Test	0.084	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 09/14/2022 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: RIDEOUT, TIMOTHY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BROWER

DATE/TIME OF TEST: 08/04/2022 04:20:49

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.080	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:19	
Control Test	0.080	04:20	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 08/31/2022 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEARY,DANTE,J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: DUMONT,TRAVIS
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: LUCK

DATE/TIME OF TEST: 08/05/2022 00:10:08

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:58	
Control Test	0.080	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.053	00:02	Volume Not Met
Breath Volume	0.355	00:02	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample	# 2 0.059	00:05	
Breath Volume	3.039	00:05	
Air Blank	0.000	00:06	
Air Blank	0.000	00:08	
Subject Sample	# 3 0.062	00:08	
Breath Volume	2.601	00:08	
Air Blank	0.000	00:09	
Control Test	0.079	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 08/31/2022 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: LOWERY,DAISY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/31/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: RICKERT

DATE/TIME OF TEST: 08/10/2022 22:21:54

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.254	22:17
Breath Volume	1.144	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.252	22:20
Breath Volume	1.398	22:20
Air Blank	0.000	22:20
Control Test	0.078	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 08/31/2022 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: COOPER, WILLIAM
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SHOUSE

DATE/TIME OF TEST: 08/13/2022 03:30:22

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:29	
Control Test	0.079	03:29	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 08/31/2022 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COOPER, WILLIAM, J

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: RIDEOUT, TIMOTHY

ARRESTING OFFICER AGENCY: Deland PD

LAST AGENCY INSPECTION: 07/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 01:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 08/14/2022 02:13:34

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.234	02:09
Breath Volume	1.132	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.227	02:12
Breath Volume	1.238	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 08/31/2022 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: DUMONT, TRAVIS
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: BLANDINA

DATE/TIME OF TEST: 08/14/2022 20:53:32

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.080	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.129	20:48
Breath Volume	1.703	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.132	20:52
Breath Volume	2.082	20:52
Air Blank	0.000	20:52
Control Test	0.079	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 08/31/2022 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WATTS,SHANNON,N

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: STURDIVANT,TERRENCE

ARRESTING OFFICER AGENCY: Deland PD

LAST AGENCY INSPECTION: 07/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 18:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: GEAR

DATE/TIME OF TEST: 08/17/2022 19:37:53

Diagnostics Check	OK	19:28
Air Blank	0.000	19:29
Control Test	0.080	19:29
Air Blank	0.000	19:30
Subject Sample # 1	0.162	19:30
Breath Volume	3.062	19:30
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.185	19:33
Breath Volume	3.621	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:35
Subject Sample # 3	0.176	19:36
Breath Volume	2.117	19:36
Air Blank	0.000	19:36
Control Test	0.079	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 08/31/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: STURDIVANT, TERRENCE
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/18/2022 22:43:34

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.252	22:38
Breath Volume	1.320	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.249	22:42
Breath Volume	1.355	22:42
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 08/31/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEARY,DANTE,J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: CUSTODE,VITTORIO
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 11:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:07

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 08/19/2022 12:41:51

Diagnostics Check	OK	12:30	
Air Blank	0.000	12:30	
Control Test	0.080	12:31	
Air Blank	0.000	12:31	
Subject Sample	# 1 0.000	12:34	Volume Not Met
Breath Volume	0.566	12:34	
Air Blank	0.000	12:35	
Air Blank	0.000	12:37	
Subject Sample	# 2 0.000	12:40	Volume Not Met
Breath Volume	0.941	12:40	
Air Blank	0.000	12:40	
Control Test	0.080	12:41	
Air Blank	0.000	12:41	
Diagnostics Check	OK	12:41	

Data Download Date/Time: 08/31/2022 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEARY,DANTE,J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: CUSTODE,VITTORIO
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 11:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:07

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 08/19/2022 13:06:48

Diagnostics Check	OK	12:54	
Air Blank	0.000	12:54	
Control Test	0.079	12:55	
Air Blank	0.000	12:55	
Subject Sample	# 1 0.000	12:58	Volume Not Met
Breath Volume	0.886	12:58	
Air Blank	0.000	12:59	
Air Blank	0.000	13:01	
Subject Sample	# 2 0.000	13:02	
Breath Volume	1.125	13:02	
Air Blank	0.000	13:03	
Air Blank	0.000	13:05	
Subject Sample	# 3 0.000	13:05	
Breath Volume	1.113	13:05	
Air Blank	0.000	13:05	
Control Test	0.078	13:06	
Air Blank	0.000	13:06	
Diagnostics Check	OK	13:06	

Data Download Date/Time: 08/31/2022 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCKHALT,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Liberty County SO
ARRESTING OFFICER: CROSBY,M
ARRESTING OFFICER AGENCY: Liberty County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 08/24/2022 19:25:27

Diagnostics Check	OK	19:19
Air Blank	0.000	19:19
Control Test	0.078	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.000	19:21
Breath Volume	2.101	19:21
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.000	19:24
Breath Volume	1.812	19:24
Air Blank	0.000	19:24
Control Test	0.079	19:24
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 09/08/2022 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORAN,ETHAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FORAN,ETHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: PABLO

DATE/TIME OF TEST: 08/27/2022 07:37:38

Diagnostics Check	OK	07:30
Air Blank	0.000	07:31
Control Test	0.079	07:31
Air Blank	0.000	07:32
Subject Sample # 1	0.175	07:33
Breath Volume	3.027	07:33
Air Blank	0.000	07:33
Air Blank	0.000	07:35
Subject Sample # 2	0.173	07:36
Breath Volume	3.000	07:36
Air Blank	0.000	07:36
Control Test	0.078	07:37
Air Blank	0.000	07:37
Diagnostics Check	OK	07:37

Data Download Date/Time: 09/08/2022 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,JOLNETH,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON,JOLNETH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 08/04/2022 01:06:17

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.080	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:59	No Sample Provided
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Air Blank	0.000	01:01	
Subject Sample	# 2	01:04	No Sample Provided
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Control Test	0.081	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 08/08/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,JOLNETH,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: THEODORE

DATE/TIME OF TEST: 08/04/2022 01:52:52

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.219	01:48
Breath Volume	1.359	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.231	01:51
Breath Volume	1.421	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 08/08/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROPST,JOHN,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: FELICIANO

DATE/TIME OF TEST: 08/04/2022 08:04:48

Diagnostics Check	OK	07:58
Air Blank	0.000	07:58
Control Test	0.081	07:59
Air Blank	0.000	07:59
Subject Sample # 1	0.254	08:00
Breath Volume	1.660	08:00
Air Blank	0.000	08:00
Air Blank	0.000	08:02
Subject Sample # 2	0.237	08:03
Breath Volume	1.414	08:03
Air Blank	0.000	08:03
Control Test	0.080	08:04
Air Blank	0.000	08:04
Diagnostics Check	OK	08:04

Data Download Date/Time: 08/08/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 08/05/2022 02:26:52

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.146	02:21
Breath Volume	1.398	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.150	02:25
Breath Volume	1.527	02:25
Air Blank	0.000	02:25
Control Test	0.078	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 08/08/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 08/05/2022 16:25:13

Diagnostics Check	OK	16:18
Air Blank	0.000	16:19
Control Test	0.081	16:19
Air Blank	0.000	16:19
Subject Sample # 1	0.000	16:20
Breath Volume	1.632	16:20
Air Blank	0.000	16:21
Air Blank	0.000	16:23
Subject Sample # 2	0.000	16:23
Breath Volume	1.640	16:23
Air Blank	0.000	16:24
Control Test	0.081	16:24
Air Blank	0.000	16:25
Diagnostics Check	OK	16:25

Data Download Date/Time: 08/08/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKMAN,GERIG,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BLACKMAN,GERIG
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 08/05/2022 22:09:30

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	21:59
Subject Sample # 1	0.209	22:00
Breath Volume	1.363	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.186	22:04
Breath Volume	1.273	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 3	0.196	22:08
Breath Volume	1.589	22:08
Air Blank	0.000	22:08
Control Test	0.079	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 08/08/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GRAZIANO

DATE/TIME OF TEST: 08/06/2022 00:43:11

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.228	00:38
Breath Volume	1.839	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.226	00:41
Breath Volume	2.355	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 08/08/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 08/06/2022 05:11:48

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:00	
Control Test	0.083	05:01	
Air Blank	0.000	05:01	
Subject Sample	# 1	05:04	No Sample Provided
Breath Volume	0.000	05:04	
Air Blank	0.000	05:05	
Air Blank	0.000	05:07	
Subject Sample	# 2	05:10	No Sample Provided
Breath Volume	0.000	05:10	
Air Blank	0.000	05:10	
Control Test	0.083	05:11	
Air Blank	0.000	05:11	
Diagnostics Check	OK	05:11	

Data Download Date/Time: 08/08/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: LARREA

DATE/TIME OF TEST: 08/06/2022 21:35:21

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.293	21:30
Breath Volume	1.808	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.300	21:33
Breath Volume	2.425	21:33
Air Blank	0.000	21:34
Control Test	0.081	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 08/08/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 08/06/2022 22:45:23

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.079	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.322	22:40
Breath Volume	1.355	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.334	22:43
Breath Volume	2.320	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 08/08/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKMAN,GERIG,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUDLOW,JUSTIN
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 07/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/07/2022 06:40:34

Diagnostics Check	OK	06:33
Air Blank	0.000	06:33
Control Test	0.081	06:34
Air Blank	0.000	06:34
Subject Sample # 1	0.174	06:35
Breath Volume	1.925	06:35
Air Blank	0.000	06:35
Air Blank	0.000	06:37
Subject Sample # 2	0.177	06:39
Breath Volume	1.433	06:39
Air Blank	0.000	06:39
Control Test	0.080	06:39
Air Blank	0.000	06:40
Diagnostics Check	OK	06:40

Data Download Date/Time: 08/08/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,MARCO,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: OUKAL

DATE/TIME OF TEST: 08/09/2022 00:04:36

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.079	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 2 0.118	23:57	Volume Not Met
Breath Volume	0.015	23:57	
Air Blank	0.000	23:58	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.152	00:03	Volume Not Met
Breath Volume	0.015	00:03	
Air Blank	0.000	00:03	
Control Test	0.080	00:04	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKMAN,GERIG,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BLACKMAN,GERIG
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: YERBY

DATE/TIME OF TEST: 08/10/2022 23:32:11

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.080	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1 0.276	23:25	Volume Not Met
Breath Volume	0.257	23:25	
Air Blank	0.000	23:25	
Air Blank	0.000	23:27	
Subject Sample	# 2	23:30	No Sample Provided
Breath Volume	0.000	23:30	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREWS, TYREE, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BULLARD, CHRISTINA
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: FONSECA

DATE/TIME OF TEST: 08/12/2022 01:43:56

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.168	01:39
Breath Volume	1.324	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.165	01:42
Breath Volume	1.410	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKMAN,GERIG,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BLACKMAN,GERIG
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: PERDUE

DATE/TIME OF TEST: 08/12/2022 04:25:15

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:14	
Control Test	0.080	04:14	
Air Blank	0.000	04:15	
Subject Sample	# 1	04:18	No Sample Provided
Breath Volume	0.000	04:18	
Air Blank	0.000	04:18	
Air Blank	0.000	04:20	
Subject Sample	# 2	04:23	No Sample Provided
Breath Volume	0.000	04:23	
Air Blank	0.000	04:24	
Control Test	0.080	04:24	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALINAS,JOSE,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: CRANE

DATE/TIME OF TEST: 08/12/2022 23:37:28

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.175	23:33
Breath Volume	1.753	23:33
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.183	23:36
Breath Volume	1.117	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,MAVERICK,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVIS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/13/2022 00:51:28

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:43	
Control Test	0.080	00:43	
Air Blank	0.000	00:44	
Subject Sample	# 1	00:47	No Sample Provided
Breath Volume	0.000	00:47	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 2	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:50	
Control Test	0.079	00:50	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFONTANT, SEDONA, S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: LAFONTANT, SEDONA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: BLOUIN

DATE/TIME OF TEST: 08/13/2022 02:51:22

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.103	02:45
Breath Volume	2.453	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.103	02:49
Breath Volume	1.843	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,MARCO,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BOLES

DATE/TIME OF TEST: 08/13/2022 03:38:57

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:27	
Control Test	0.081	03:28	
Air Blank	0.000	03:28	
Subject Sample	# 1 0.149	03:31	Volume Not Met
Breath Volume	0.257	03:31	
Air Blank	0.000	03:32	
Air Blank	0.000	03:34	
Subject Sample	# 2 0.150	03:37	Volume Not Met
Breath Volume	0.332	03:37	
Air Blank	0.000	03:38	
Control Test	0.080	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BAMACA

DATE/TIME OF TEST: 08/13/2022 04:18:37

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.149	04:13
Breath Volume	2.367	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.142	04:17
Breath Volume	1.300	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,JOLNETH,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: SIKORA

DATE/TIME OF TEST: 08/13/2022 23:42:15

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:31	
Control Test	0.081	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:35	No Sample Provided
Breath Volume	0.000	23:35	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample	# 2	23:40	No Sample Provided
Breath Volume	0.000	23:40	
Air Blank	0.000	23:41	
Control Test	0.081	23:41	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: MATEO FELIPE

DATE/TIME OF TEST: 08/14/2022 02:35:20

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.239	02:30
Breath Volume	3.488	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.231	02:33
Breath Volume	3.015	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDRON, DAVID, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PALAU, MARIO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:48

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 08/14/2022 11:24:59

Diagnostics Check	OK	11:18
Air Blank	0.000	11:18
Control Test	0.081	11:18
Air Blank	0.000	11:19
Subject Sample # 1	0.296	11:20
Breath Volume	1.152	11:20
Air Blank	0.000	11:21
Air Blank	0.000	11:22
Subject Sample # 2	0.285	11:23
Breath Volume	2.531	11:23
Air Blank	0.000	11:24
Control Test	0.080	11:24
Air Blank	0.000	11:24
Diagnostics Check	OK	11:24

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKMAN,GERIG,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BLACKMAN,GERIG
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MCCANTS

DATE/TIME OF TEST: 08/17/2022 00:33:14

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.081	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.161	00:28
Breath Volume	1.539	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.155	00:31
Breath Volume	1.535	00:31
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE, ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 08/17/2022 04:14:43

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.080	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:07	No Sample Provided
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 2	04:13	No Sample Provided
Breath Volume	0.000	04:13	
Air Blank	0.000	04:13	
Control Test	0.081	04:14	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKMAN,GERIG,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MANCILLA RAMOS

DATE/TIME OF TEST: 08/17/2022 04:30:53

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.176	04:26
Breath Volume	1.703	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:28
Subject Sample # 2	0.169	04:29
Breath Volume	1.359	04:29
Air Blank	0.000	04:29
Control Test	0.080	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALINAS,JOSE,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ALVAREZ RIOS

DATE/TIME OF TEST: 08/18/2022 00:46:37

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.259	00:42
Breath Volume	2.738	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.249	00:45
Breath Volume	2.277	00:45
Air Blank	0.000	00:45
Control Test	0.078	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTS,JAMES,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 08/20/2022 07:52:56

Diagnostics Check	OK	07:46
Air Blank	0.000	07:47
Control Test	0.080	07:47
Air Blank	0.000	07:47
Subject Sample # 1	0.089	07:48
Breath Volume	3.226	07:48
Air Blank	0.000	07:49
Air Blank	0.000	07:50
Subject Sample # 2	0.089	07:51
Breath Volume	3.082	07:51
Air Blank	0.000	07:52
Control Test	0.080	07:52
Air Blank	0.000	07:52
Diagnostics Check	OK	07:52

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,MAVERICK,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SOTO,MAVERICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: WEISS

DATE/TIME OF TEST: 08/21/2022 02:07:18

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.080	02:05	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Control Test	0.081	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 09/23/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE, ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 08/21/2022 05:25:44

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:14	
Control Test	0.081	05:15	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:18	No Sample Provided
Breath Volume	0.000	05:18	
Air Blank	0.000	05:19	
Air Blank	0.000	05:21	
Subject Sample	# 2	05:24	No Sample Provided
Breath Volume	0.000	05:24	
Air Blank	0.000	05:24	
Control Test	0.081	05:25	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISPO,STEPHANIE,
BREATH TEST OPERATOR AGENCY: Davenport PD
ARRESTING OFFICER: BISPO,STEPHANIE
ARRESTING OFFICER AGENCY: Davenport PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: REYES BARTOLO

DATE/TIME OF TEST: 08/21/2022 06:15:09

Diagnostics Check	OK	06:07
Air Blank	0.000	06:08
Control Test	0.080	06:08
Air Blank	0.000	06:08
Subject Sample # 1	0.175	06:10
Breath Volume	1.519	06:10
Air Blank	0.000	06:11
Air Blank	0.000	06:12
Subject Sample # 2	0.171	06:13
Breath Volume	1.687	06:13
Air Blank	0.000	06:14
Control Test	0.078	06:14
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE, ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 08/22/2022 00:26:22

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.153	00:21
Breath Volume	1.789	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.151	00:24
Breath Volume	1.910	00:24
Air Blank	0.000	00:25
Control Test	0.080	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELONY,ANATUS,
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: BELONY,ANATUS
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: TATHAM

DATE/TIME OF TEST: 08/23/2022 17:49:24

Diagnostics Check	OK	17:42
Air Blank	0.000	17:43
Control Test	0.081	17:43
Air Blank	0.000	17:44
Subject Sample # 1	0.000	17:44
Breath Volume	1.824	17:44
Air Blank	0.000	17:45
Air Blank	0.000	17:47
Subject Sample # 2	0.000	17:48
Breath Volume	1.113	17:48
Air Blank	0.000	17:48
Control Test	0.080	17:48
Air Blank	0.000	17:49
Diagnostics Check	OK	17:49

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTOSH,HALLIE,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FAISON,MEGAN
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: RIVERA MARRERO

DATE/TIME OF TEST: 08/25/2022 03:03:27

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.282	02:57
Breath Volume	1.371	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.296	03:01
Breath Volume	1.617	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKMAN,GERIG,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BLACKMAN,GERIG
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 08/26/2022 01:49:17

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.081	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.136	01:44
Breath Volume	1.378	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.138	01:47
Breath Volume	1.363	01:47
Air Blank	0.000	01:48
Control Test	0.082	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,JOLNETH,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON,JOLNETH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/26/2022 22:02:41

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.081	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.234	21:58
Breath Volume	2.296	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.231	22:01
Breath Volume	2.218	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,JOLNETH,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SAINT JUSTE,YVENS
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: RAMKISHUN

DATE/TIME OF TEST: 08/26/2022 22:21:06

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.148	22:16
Breath Volume	1.781	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.152	22:19
Breath Volume	1.734	22:19
Air Blank	0.000	22:20
Control Test	0.080	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCH,JOHN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHURCH,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/27/2022 08:07:02

Diagnostics Check	OK	08:00
Air Blank	0.000	08:01
Control Test	0.081	08:01
Air Blank	0.000	08:01
Subject Sample # 1	0.166	08:02
Breath Volume	1.390	08:02
Air Blank	0.000	08:02
Air Blank	0.000	08:04
Subject Sample # 2	0.156	08:05
Breath Volume	1.726	08:05
Air Blank	0.000	08:06
Control Test	0.081	08:06
Air Blank	0.000	08:06
Diagnostics Check	OK	08:07

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCH,JOHN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 08/27/2022 23:11:49

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.000	23:07
Breath Volume	1.460	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.000	23:10
Breath Volume	1.308	23:10
Air Blank	0.000	23:10
Control Test	0.081	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PHILLIPS,JORDAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 08/28/2022 03:29:18

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.081	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1 0.000	03:22	Volume Not Met
Breath Volume	0.132	03:22	
Air Blank	0.000	03:22	
Air Blank	0.000	03:24	
Subject Sample	# 2 0.000	03:27	Volume Not Met
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCH,JOHN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHURCH,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: NEWCOMB

DATE/TIME OF TEST: 08/28/2022 04:25:57

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.143	04:21
Breath Volume	1.671	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:23
Subject Sample # 2	0.147	04:24
Breath Volume	1.746	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCH,JOHN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: NEUBERGER

DATE/TIME OF TEST: 08/28/2022 05:02:36

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.081	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.079	04:58
Breath Volume	1.902	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.072	05:01
Breath Volume	1.191	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROPST,JOHN,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: PASCAL

DATE/TIME OF TEST: 08/28/2022 12:23:57

Diagnostics Check	OK	12:12	
Air Blank	0.000	12:13	
Control Test	0.081	12:13	
Air Blank	0.000	12:13	
Subject Sample	# 1	12:17	No Sample Provided
Breath Volume	0.000	12:17	
Air Blank	0.000	12:17	
Air Blank	0.000	12:19	
Subject Sample	# 2	12:22	No Sample Provided
Breath Volume	0.000	12:22	
Air Blank	0.000	12:23	
Control Test	0.081	12:23	
Air Blank	0.000	12:23	
Diagnostics Check	OK	12:23	

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,JOLNETH,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON,JOLNETH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: LAURSEN

DATE/TIME OF TEST: 08/28/2022 21:25:08

Diagnostics Check	OK	21:18
Air Blank	0.000	21:19
Control Test	0.080	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.198	21:20
Breath Volume	1.957	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.190	21:23
Breath Volume	1.734	21:23
Air Blank	0.000	21:24
Control Test	0.080	21:24
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,JOLNETH,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON,JOLNETH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 08/29/2022 01:29:44

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.080	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.188	01:24
Breath Volume	1.660	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.194	01:28
Breath Volume	1.468	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 09/23/2022 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BRYAN,STEVEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/14/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 08:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:13

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 08/22/2022 09:54:36

Diagnostics Check	OK	09:48
Air Blank	0.000	09:48
Control Test	0.075	09:49
Air Blank	0.000	09:49
Subject Sample # 1	0.067	09:50
Breath Volume	2.789	09:50
Air Blank	0.000	09:50
Air Blank	0.000	09:52
Subject Sample # 2	0.065	09:53
Breath Volume	2.218	09:53
Air Blank	0.000	09:53
Control Test	0.076	09:54
Air Blank	0.000	09:54
Diagnostics Check	OK	09:54

Data Download Date/Time: 08/30/2022 21:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BRYAN,STEVEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 08/29/2022 22:47:29

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.076	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.024	22:43
Breath Volume	2.062	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.022	22:46
Breath Volume	2.417	22:46
Air Blank	0.000	22:46
Control Test	0.076	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 08/30/2022 21:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CROFUTT,BRITTANY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: HEARTZ

DATE/TIME OF TEST: 08/07/2022 21:58:02

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.081	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.056	21:53
Breath Volume	1.851	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.058	21:56
Breath Volume	1.964	21:56
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:57
Diagnostics Check	OK	21:58

Data Download Date/Time: 08/30/2022 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CROFUTT,BRITTANY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: ELLIOT

DATE/TIME OF TEST: 08/11/2022 02:32:13

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.190	02:27
Breath Volume	1.558	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.190	02:30
Breath Volume	1.582	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 08/30/2022 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MULLER,COREY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 08/13/2022 15:10:04

Diagnostics Check	OK	15:01
Air Blank	0.000	15:01
Control Test	0.080	15:02
Air Blank	0.000	15:02
Subject Sample # 1	0.181	15:05
Breath Volume	1.832	15:05
Air Blank	0.000	15:06
Air Blank	0.000	15:07
Subject Sample # 2	0.179	15:08
Breath Volume	1.703	15:08
Air Blank	0.000	15:09
Control Test	0.078	15:09
Air Blank	0.000	15:09
Diagnostics Check	OK	15:10

Data Download Date/Time: 08/30/2022 22:29