

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 0325341-5
BREATH TEST OPERATOR: CARTER,LARRY,J	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: CARTER,L	ARREST TIME: 06:05
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Female
LAST AGENCY INSPECTION: 08/31/2022	OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: ARMSTED	DATE/TIME OF TEST: 09/01/2022 07:09:23
------------------------------	--

Diagnostics Check	OK	06:57	
Air Blank	0.000	06:58	
Control Test	0.081	06:58	
Air Blank	0.000	06:59	
Subject Sample	# 1	07:02	No Sample Provided
Breath Volume	0.000	07:02	
Air Blank	0.000	07:02	
Air Blank	0.000	07:04	
Subject Sample	# 2	07:08	No Sample Provided
Breath Volume	0.000	07:08	
Air Blank	0.000	07:08	
Control Test	0.081	07:08	
Air Blank	0.000	07:09	
Diagnostics Check	OK	07:09	

Data Download Date/Time: 09/23/2022 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 0325341-5
BREATH TEST OPERATOR: FITCHETT, ADAM, R	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SEEBECK, JENNA	ARREST TIME: 04:03
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 08/31/2022	OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: MENDEZ-ALBAN	DATE/TIME OF TEST: 09/03/2022 04:52:21
-----------------------------------	--

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:49	
Control Test	0.080	04:50	
Air Blank	0.000	04:50	
Subject Sample	# 1	04:50	Subject Test Refused
Breath Volume	0.000	04:50	
Air Blank	0.000	04:51	
Control Test	0.080	04:51	
Air Blank	0.000	04:52	
Diagnostics Check	OK	04:52	

Data Download Date/Time: 09/23/2022 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GLENECK,JORDAN,A

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: RILEY,DEJA

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 0325341-5

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 10:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: BRAZZELL

DATE/TIME OF TEST: 09/03/2022 11:44:56

Diagnostics Check	OK	11:40	
Air Blank	0.000	11:41	
Control Test	0.081	11:41	
Air Blank	0.000	11:42	
Subject Sample	# 1	11:43	Subject Test Refused
Breath Volume	0.000	11:43	
Air Blank	0.000	11:44	
Control Test	0.080	11:44	
Air Blank	0.000	11:44	
Diagnostics Check	OK	11:44	

Data Download Date/Time: 09/23/2022 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GLENECK,JORDAN,A

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: GLENECK,JORDAN

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 0325341-5

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 14:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:13

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/03/2022 15:40:31

Diagnostics Check	OK	15:34
Air Blank	0.000	15:34
Control Test	0.080	15:34
Air Blank	0.000	15:35
Subject Sample # 1	0.260	15:35
Breath Volume	3.468	15:35
Air Blank	0.000	15:36
Air Blank	0.000	15:38
Subject Sample # 2	0.244	15:38
Breath Volume	2.792	15:38
Air Blank	0.000	15:39
Control Test	0.079	15:39
Air Blank	0.000	15:40
Diagnostics Check	OK	15:40

Data Download Date/Time: 09/23/2022 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 0325341-5
BREATH TEST OPERATOR: LAPEE,WYATT,T	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SEEBECK,JENNA	ARREST TIME: 23:36
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 08/31/2022	OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: VASQUEZ ARITA	DATE/TIME OF TEST: 09/04/2022	00:45:28
------------------------------------	-------------------------------	----------

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.181	00:40
Breath Volume	1.867	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.180	00:43
Breath Volume	1.445	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 09/23/2022 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 0325341-5
BREATH TEST OPERATOR: LAPEE,WYATT,T	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: LAPEE,WYATT	ARREST TIME: 04:21
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 08/31/2022	OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: JOHNSON DATE/TIME OF TEST: 09/04/2022 05:45:52

Diagnostics Check	OK	05:39
Air Blank	0.000	05:40
Control Test	0.081	05:40
Air Blank	0.000	05:40
Subject Sample # 1	0.085	05:41
Breath Volume	2.476	05:41
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.087	05:44
Breath Volume	3.542	05:44
Air Blank	0.000	05:44
Control Test	0.081	05:45
Air Blank	0.000	05:45
Diagnostics Check	OK	05:45

Data Download Date/Time: 09/23/2022 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 0325341-5
BREATH TEST OPERATOR: STINSON,LACY,R	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: STINSON,LACY	ARREST TIME: 21:15
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 08/31/2022	OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: JONES	DATE/TIME OF TEST: 09/04/2022	22:17:04
----------------------------	-------------------------------	----------

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.080	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.306	22:11
Breath Volume	1.617	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:13
Subject Sample # 2	0.294	22:15
Breath Volume	1.742	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:17

Data Download Date/Time: 09/23/2022 08:17

Subject Test Electronic Data

DATE/TIME OF TEST: 09/04/2022 23:45:51

Printed: 01 December 2022
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 0325341-5
BREATH TEST OPERATOR: FITCHETT, ADAM, R	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SEEBECK, JENNA	ARREST TIME: 01:55
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 08/31/2022	OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: CONNER	DATE/TIME OF TEST: 09/08/2022 02:44:24
-----------------------------	--

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.082	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:43	
Control Test	0.080	02:43	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 09/23/2022 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAPEE,WYATT,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NASH,DAVID
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LANDMESSER

DATE/TIME OF TEST: 09/09/2022 00:11:28

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.083	00:07
Breath Volume	1.617	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.083	00:09
Breath Volume	1.183	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 09/23/2022 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 0325341-5
BREATH TEST OPERATOR: SHELTON,ERIC,L	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: CALDERARO,DANA	ARREST TIME: 09:38
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 08/31/2022	OBSERVATION PERIOD BEGAN: 10:24

SUBJECT'S LAST NAME: SULLIVAN	DATE/TIME OF TEST: 09/09/2022	10:50:56
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	10:44
Air Blank	0.000	10:45
Control Test	0.081	10:45
Air Blank	0.000	10:46
Subject Sample # 1	0.000	10:46
Breath Volume	3.617	10:46
Air Blank	0.000	10:46
Air Blank	0.000	10:48
Subject Sample # 2	0.000	10:49
Breath Volume	1.894	10:49
Air Blank	0.000	10:50
Control Test	0.081	10:50
Air Blank	0.000	10:50
Diagnostics Check	OK	10:50

Data Download Date/Time: 09/23/2022 08:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VASQUEZ,GEORGE,D

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: CARROLL,SHELBY

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 0325341-5

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 21:56

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: SEWELL

DATE/TIME OF TEST: 09/09/2022 22:52:29

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:49	
Control Test	0.081	22:50	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.079	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 09/23/2022 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VASQUEZ,GEORGE,D

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: VASQUEZ,GEORGE

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 0325341-5

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 04:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/10/2022 05:11:14

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:08	
Control Test	0.081	05:09	
Air Blank	0.000	05:09	
Subject Sample	# 1	05:09	Subject Test Refused
Breath Volume	0.000	05:09	
Air Blank	0.000	05:10	
Control Test	0.081	05:10	
Air Blank	0.000	05:11	
Diagnostics Check	OK	05:11	

Data Download Date/Time: 09/23/2022 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VASQUEZ,GEORGE,D

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: DANIELS,LYLE

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 0325341-5

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 04:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: COLLAZO

DATE/TIME OF TEST: 09/10/2022 05:39:37

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.081	05:25	
Air Blank	0.000	05:26	
Subject Sample	# 1 0.174	05:29	Volume Not Met
Breath Volume	0.035	05:29	
Air Blank	0.000	05:30	
Air Blank	0.000	05:31	
Subject Sample	# 2 0.177	05:34	
Breath Volume	1.667	05:34	
Air Blank	0.000	05:35	
Air Blank	0.000	05:37	
Subject Sample	# 3 0.183	05:38	
Breath Volume	1.300	05:38	
Air Blank	0.000	05:38	
Control Test	0.078	05:39	
Air Blank	0.000	05:39	
Diagnostics Check	OK	05:39	

Data Download Date/Time: 09/23/2022 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 0325341-5
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D	DRY GAS EXPIRY DATE: 11/18/2024
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: BIERNACKI,CHAD	ARREST TIME: 03:00
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 08/31/2022	OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MEDINA	DATE/TIME OF TEST: 09/11/2022 04:04:38
-----------------------------	--

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.148	03:59
Breath Volume	1.882	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.148	04:03
Breath Volume	2.261	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 09/23/2022 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VASQUEZ,GEORGE,D

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: CARROLL,SHELBY

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 0325341-5

DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI

ARREST TIME: 04:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FLORES NUNEZ

DATE/TIME OF TEST: 09/11/2022 04:45:15

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.080	04:43	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:44	
Control Test	0.080	04:44	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:45	

Data Download Date/Time: 09/23/2022 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILES,HENRY,G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: GILES,HENRY

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: MORO MORO

DATE/TIME OF TEST: 09/01/2022 01:04:17

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.160	00:59
Breath Volume	2.644	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.154	01:02
Breath Volume	1.851	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 09/30/2022 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILES,HENRY,G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: GILES,HENRY

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 23:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: WINN GORDILLO

DATE/TIME OF TEST: 09/01/2022 23:54:21

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.000	23:49
Breath Volume	2.746	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.000	23:52
Breath Volume	1.597	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 09/30/2022 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILES,HENRY,G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: GILES,HENRY

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 03:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: STANSBURY

DATE/TIME OF TEST: 09/02/2022 03:58:14

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.117	03:53
Breath Volume	3.175	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.118	03:56
Breath Volume	3.382	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 09/30/2022 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILES,OFFICER,G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: GILES,OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 21:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: UPCHURCH

DATE/TIME OF TEST: 09/02/2022 21:54:07

Diagnostics Check	OK	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.154	21:49
Breath Volume	3.046	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:51
Subject Sample # 2	0.154	21:52
Breath Volume	3.062	21:52
Air Blank	0.000	21:53
Control Test	0.079	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 09/30/2022 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YEATES,SCOTT,W

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: YEATES,SCOTT

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 23:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 09/03/2022 00:11:38

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.079	00:01	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:04	No Sample Provided
Breath Volume	0.000	00:04	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample	# 2	00:10	No Sample Provided
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Control Test	0.079	00:11	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 09/30/2022 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILES,OFFICER,G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: GILES,OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 09/04/2022 01:04:59

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.220	01:00
Breath Volume	2.089	01:00
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.210	01:03
Breath Volume	2.085	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 09/30/2022 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRUNEAU, OFFICER,

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: BRUNEAU, OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:28

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: HERRERA-ESCAMILLA

DATE/TIME OF TEST: 09/04/2022 01:26:42

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.088	01:22
Breath Volume	2.492	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.089	01:25
Breath Volume	3.179	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 09/30/2022 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KROLICK,ALEX,M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: KROLICK,ALEX

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MAIDEN

DATE/TIME OF TEST: 09/10/2022 01:14:20

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.088	01:09
Breath Volume	2.152	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.091	01:12
Breath Volume	3.433	01:12
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 09/30/2022 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KROLICK,ALEX,M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: KROLICK,ALEX

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 03:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: LEVELLEN

DATE/TIME OF TEST: 09/10/2022 04:16:16

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.206	04:11
Breath Volume	3.160	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.208	04:14
Breath Volume	3.562	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 09/30/2022 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILES,OFFICER,G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: GILES,OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 23:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: NETTLETON

DATE/TIME OF TEST: 09/10/2022 23:28:35

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.090	23:23
Breath Volume	2.105	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:25
Subject Sample # 2	0.087	23:27
Breath Volume	1.687	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 09/30/2022 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILES,OFFICER,G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: GILES,OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 09/11/2022 02:50:07

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.101	02:45
Breath Volume	2.070	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample # 2	0.101	02:48
Breath Volume	1.886	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 09/30/2022 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STIERS,S,T

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: STIERS,S

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 17:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 09/15/2022 18:45:40

Diagnostics Check	OK	18:34	
Air Blank	0.000	18:34	
Control Test	0.079	18:35	
Air Blank	0.000	18:35	
Subject Sample	# 1	18:38	No Sample Provided
Breath Volume	0.000	18:38	
Air Blank	0.000	18:39	
Air Blank	0.000	18:41	
Subject Sample	# 2	18:44	No Sample Provided
Breath Volume	0.000	18:44	
Air Blank	0.000	18:44	
Control Test	0.079	18:45	
Air Blank	0.000	18:45	
Diagnostics Check	OK	18:45	

Data Download Date/Time: 09/30/2022 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRUNEAU, OFFICER,

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: CURCIO, OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: KEICHER

DATE/TIME OF TEST: 09/15/2022 23:19:23

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.219	23:12
Breath Volume	1.542	23:12
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.210	23:17
Breath Volume	1.421	23:17
Air Blank	0.000	23:18
Control Test	0.078	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 09/30/2022 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRUNEAU, OFFICER,

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: VANGELI, OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 09/15/2022 23:49:19

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.078	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.249	23:44
Breath Volume	3.050	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.248	23:47
Breath Volume	2.976	23:47
Air Blank	0.000	23:48
Control Test	0.078	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 09/30/2022 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VENEZIA',STEVEN,L

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: VENEZIA,STEVEN

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: HAMMOND

DATE/TIME OF TEST: 09/17/2022 02:28:28

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.249	02:23
Breath Volume	1.828	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.244	02:26
Breath Volume	2.257	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 09/30/2022 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRUNEAU,OFFICER,

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: BRUNEAU,OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 21:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: CYRUS

DATE/TIME OF TEST: 09/17/2022 21:54:39

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.077	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.189	21:49
Breath Volume	2.296	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.176	21:53
Breath Volume	1.203	21:53
Air Blank	0.000	21:53
Control Test	0.076	21:54
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 09/30/2022 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YEATES,SCOTT,W

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: YEATES,SCOTT

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 21:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: KYLE

DATE/TIME OF TEST: 09/17/2022 22:25:17

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.078	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.000	22:20
Breath Volume	2.777	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.000	22:23
Breath Volume	2.339	22:23
Air Blank	0.000	22:24
Control Test	0.078	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 09/30/2022 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KROLICK,ALEX,M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: KROLICK,ALEX

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: SCHUMANN

DATE/TIME OF TEST: 09/18/2022 02:08:50

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.200	02:04
Breath Volume	4.417	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.185	02:07
Breath Volume	2.027	02:07
Air Blank	0.000	02:07
Control Test	0.076	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 09/30/2022 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KROLICK,ALEX,M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: KROLICK,ALEX

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 21:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: MONTEIRO

DATE/TIME OF TEST: 09/18/2022 21:49:15

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.079	21:43
Air Blank	0.000	21:44
Subject Sample # 1	0.172	21:44
Breath Volume	1.523	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.175	21:47
Breath Volume	2.121	21:47
Air Blank	0.000	21:48
Control Test	0.079	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 09/30/2022 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STIERS,S,T

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: PALOMA,N

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 17:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: LEDESMA

DATE/TIME OF TEST: 09/20/2022 17:48:29

Diagnostics Check	OK	17:41
Air Blank	0.000	17:42
Control Test	0.080	17:42
Air Blank	0.000	17:43
Subject Sample # 1	0.041	17:44
Breath Volume	1.394	17:44
Air Blank	0.000	17:44
Air Blank	0.000	17:46
Subject Sample # 2	0.041	17:46
Breath Volume	1.753	17:46
Air Blank	0.000	17:47
Control Test	0.079	17:47
Air Blank	0.000	17:48
Diagnostics Check	OK	17:48

Data Download Date/Time: 09/30/2022 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STIERS,S,T

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: STIERS,S

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: WAGONER

DATE/TIME OF TEST: 09/20/2022 23:22:59

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.143	23:17
Breath Volume	2.031	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.146	23:21
Breath Volume	1.117	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 09/30/2022 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YEATES,SCOTT,W

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: YEATES,SCOTT

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: VOLKWEIN

DATE/TIME OF TEST: 09/22/2022 23:31:58

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.055	23:27
Breath Volume	1.871	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.052	23:30
Breath Volume	1.757	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 09/30/2022 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KROLICK,ALEX,M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: PALOMA,N

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 21:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: MARTINEZ-PEDRAZA

DATE/TIME OF TEST: 09/23/2022 22:17:21

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.198	22:12
Breath Volume	1.832	22:12
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.205	22:15
Breath Volume	1.300	22:15
Air Blank	0.000	22:16
Control Test	0.077	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 09/30/2022 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WOODIE,OFC,L

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: WOODIE,L

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BURNHAM

DATE/TIME OF TEST: 09/24/2022 01:19:47

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.000	01:15
Breath Volume	2.660	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.000	01:18
Breath Volume	2.769	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 09/30/2022 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KROLICK,ALEX,M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: PALOMA,N

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 20:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/24/2022 21:41:29

Diagnostics Check	OK	21:34
Air Blank	0.000	21:35
Control Test	0.079	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.048	21:36
Breath Volume	1.527	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.047	21:39
Breath Volume	2.371	21:39
Air Blank	0.000	21:40
Control Test	0.077	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 09/30/2022 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KROLICK,ALEX,M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: KROLICK,ALEX

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 20:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 09/24/2022 22:01:11

Diagnostics Check	OK	21:49	
Air Blank	0.000	21:50	
Control Test	0.078	21:50	
Air Blank	0.000	21:51	
Subject Sample	# 1	21:54	No Sample Provided
Breath Volume	0.000	21:54	
Air Blank	0.000	21:54	
Air Blank	0.000	21:56	
Subject Sample	# 2	21:59	No Sample Provided
Breath Volume	0.000	21:59	
Air Blank	0.000	22:00	
Control Test	0.076	22:00	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 09/30/2022 21:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRUNEAU, OFFICER,

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: BRUNEAU, OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 09/24/2022 23:35:49

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.078	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.051	23:31
Breath Volume	1.101	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.053	23:34
Breath Volume	1.980	23:34
Air Blank	0.000	23:34
Control Test	0.078	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 09/30/2022 21:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILES,OFFICER,G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: GILES,OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:53

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/28/2022 03:26:50

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.078	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.111	03:22
Breath Volume	3.375	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.106	03:25
Breath Volume	3.175	03:25
Air Blank	0.000	03:25
Control Test	0.078	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 09/30/2022 21:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KROLICK,ALEX,M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: KROLICK,ALEX

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35521080A2

DRY GAS EXPIRY DATE: 03/05/2024

VIOLATION CODE: DUI

ARREST TIME: 04:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MAMEDOV

DATE/TIME OF TEST: 09/30/2022 05:11:33

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.129	05:06
Breath Volume	3.468	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.128	05:09
Breath Volume	3.304	05:09
Air Blank	0.000	05:10
Control Test	0.078	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 09/30/2022 21:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVY,JAYSON,I
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LAVENTURE,KATE
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 09/07/2022 18:21:42

Diagnostics Check	OK	18:15
Air Blank	0.000	18:15
Control Test	0.079	18:16
Air Blank	0.000	18:16
Subject Sample # 1	0.251	18:17
Breath Volume	2.085	18:17
Air Blank	0.000	18:17
Air Blank	0.000	18:19
Subject Sample # 2	0.247	18:20
Breath Volume	2.710	18:20
Air Blank	0.000	18:20
Control Test	0.079	18:21
Air Blank	0.000	18:21
Diagnostics Check	OK	18:21

Data Download Date/Time: 09/13/2022 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVY,JAYSON,I
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: VALDEZ,ALEX
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: MATHERS

DATE/TIME OF TEST: 09/30/2022 23:55:00

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.079	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 1	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:54	
Control Test	0.080	23:54	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 10/17/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BARNETT,JAMES
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: MUNSON

DATE/TIME OF TEST: 09/18/2022 22:40:31

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.255	22:36
Breath Volume	2.191	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.257	22:39
Breath Volume	2.640	22:39
Air Blank	0.000	22:39
Control Test	0.078	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 10/21/2022 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BRITT

DATE/TIME OF TEST: 09/17/2022 00:11:50

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.000	00:06
Breath Volume	1.644	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.000	00:10
Breath Volume	2.406	00:10
Air Blank	0.000	00:10
Control Test	0.083	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 10/21/2022 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLSOM,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FOLSOM,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 09/20/2022 00:21:33

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.081	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.152	00:16
Breath Volume	2.867	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.149	00:19
Breath Volume	1.945	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 10/21/2022 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: RAYMOND

DATE/TIME OF TEST: 09/25/2022 23:42:00

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:39	
Control Test	0.081	23:39	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:40	Subject Test Refused
Breath Volume	0.000	23:40	
Air Blank	0.000	23:41	
Control Test	0.081	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 10/21/2022 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/26/2022 03:44:35

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.080	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.107	03:40
Breath Volume	1.605	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.099	03:42
Breath Volume	1.441	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 10/21/2022 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRABHAM,CHARLES,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BRABHAM,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CLEMONS

DATE/TIME OF TEST: 09/04/2022 03:22:23

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.206	03:17
Breath Volume	1.894	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.203	03:20
Breath Volume	3.824	03:20
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 09/13/2022 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,GARRETT,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WORTH,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BRANCH

DATE/TIME OF TEST: 09/09/2022 00:10:40

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.079	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1 0.190	00:05	
Breath Volume	1.714	00:05	
Air Blank	0.000	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 2 0.186	00:09	
Breath Volume	2.015	00:09	
Air Blank		00:10	RFI Detect
Air Blank		00:10	RFI Detect

Data Download Date/Time: 09/13/2022 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,JORDI,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILSON,MALCOM
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: HUMPHREY

DATE/TIME OF TEST: 09/09/2022 21:14:00

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.080	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.107	21:09
Breath Volume	1.929	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:11
Subject Sample # 2	0.106	21:12
Breath Volume	2.062	21:12
Air Blank	0.000	21:13
Control Test	0.080	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:14

Data Download Date/Time: 09/13/2022 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WENK,ANTHONY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: BOYKIN

DATE/TIME OF TEST: 09/18/2022 01:42:28

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.283	01:37
Breath Volume	1.925	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.268	01:40
Breath Volume	2.457	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 10/17/2022 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WENK,ANTHONY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: MCDANIEL-HINES

DATE/TIME OF TEST: 09/22/2022 21:27:00

Diagnostics Check	OK	21:19
Air Blank	0.000	21:20
Control Test	0.079	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.217	21:22
Breath Volume	2.203	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:24
Subject Sample # 2	0.224	21:25
Breath Volume	1.605	21:25
Air Blank	0.000	21:26
Control Test	0.078	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:27

Data Download Date/Time: 10/17/2022 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO,KENNETH,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PRIETO,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 09/24/2022 20:53:20

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:50	
Control Test	0.079	20:50	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:51	Subject Test Refused
Breath Volume	0.000	20:51	
Air Blank	0.000	20:52	
Control Test	0.080	20:52	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

Data Download Date/Time: 10/17/2022 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNHILL, WESLEY, D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RANDOLPH, BRENT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CARMICAL

DATE/TIME OF TEST: 09/01/2022 00:58:02

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:54	
Control Test	0.082	00:55	
Air Blank	0.000	00:55	
Subject Sample	# 1	00:56	Subject Test Refused
Breath Volume	0.000	00:56	
Air Blank	0.000	00:57	
Control Test	0.081	00:57	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 09/29/2022 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: BULLOCK

DATE/TIME OF TEST: 09/05/2022 23:31:05

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:28	
Control Test	0.081	23:28	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:30	
Control Test	0.081	23:30	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 09/29/2022 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUS,CHRISTINA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 09/10/2022 06:32:49

Diagnostics Check	OK	06:26
Air Blank	0.000	06:27
Control Test	0.082	06:27
Air Blank	0.000	06:28
Subject Sample # 1	0.105	06:28
Breath Volume	2.660	06:28
Air Blank	0.000	06:29
Air Blank	0.000	06:30
Subject Sample # 2	0.102	06:31
Breath Volume	2.789	06:31
Air Blank	0.000	06:31
Control Test	0.081	06:32
Air Blank	0.000	06:32
Diagnostics Check	OK	06:32

Data Download Date/Time: 09/29/2022 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON, RAY, C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VALENTINE, TREVOR
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: NAJERA MEZA

DATE/TIME OF TEST: 09/18/2022 04:38:12

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.082	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.243	04:33
Breath Volume	3.003	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.242	04:36
Breath Volume	2.703	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 09/29/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLONNA, TYLER, S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CAPITAN, NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: CURLEE

DATE/TIME OF TEST: 09/18/2022 23:49:50

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.082	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.067	23:45
Breath Volume	2.957	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.069	23:48
Breath Volume	2.515	23:48
Air Blank	0.000	23:48
Control Test	0.082	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 09/29/2022 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: DUNHAM

DATE/TIME OF TEST: 09/24/2022 04:37:17

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.082	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.233	04:32
Breath Volume	2.000	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.228	04:35
Breath Volume	1.781	04:35
Air Blank	0.000	04:36
Control Test	0.082	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 09/29/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNING,CADEN,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DEWIT,HARALD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: OGLESBY

DATE/TIME OF TEST: 09/25/2022 14:40:06

Diagnostics Check	OK	14:36	
Air Blank	0.000	14:36	
Control Test	0.082	14:36	
Air Blank	0.000	14:37	
Subject Sample	# 1	14:38	Subject Test Refused
Breath Volume	0.000	14:38	
Air Blank	0.000	14:39	
Control Test	0.081	14:39	
Air Blank	0.000	14:40	
Diagnostics Check	OK	14:40	

Data Download Date/Time: 09/29/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTENS, WILLIAM,
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SCHMIDT, MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 09/26/2022 00:25:22

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.083	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.000	00:21
Breath Volume	2.886	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.000	00:24
Breath Volume	3.164	00:24
Air Blank	0.000	00:24
Control Test	0.082	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 09/29/2022 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: DUNVILLE

DATE/TIME OF TEST: 09/26/2022 20:04:52

Diagnostics Check	OK	19:58
Air Blank	0.000	19:58
Control Test	0.082	19:59
Air Blank	0.000	19:59
Subject Sample # 1	0.243	20:00
Breath Volume	1.527	20:00
Air Blank	0.000	20:00
Air Blank	0.000	20:02
Subject Sample # 2	0.231	20:03
Breath Volume	1.363	20:03
Air Blank	0.000	20:03
Control Test	0.081	20:04
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 09/29/2022 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON, TYLER, L
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: MEROLA, PETER
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/04/2022 02:17:34

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.186	02:12
Breath Volume	3.964	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.181	02:15
Breath Volume	4.144	02:15
Air Blank	0.000	02:16
Control Test	0.077	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 09/21/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/07/2022 23:57:01

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.128	23:52
Breath Volume	1.128	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.144	23:55
Breath Volume	1.757	23:55
Air Blank	0.000	23:56
Control Test	0.078	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:57

Data Download Date/Time: 09/21/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERKLIN, LEYON, G
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: BYRD, LONDON
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: LANCASTER

DATE/TIME OF TEST: 09/08/2022 22:27:06

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.078	22:22	
Air Blank	0.000	22:22	
Subject Sample	# 1 0.174	22:25	Volume Not Met
Breath Volume	0.238	22:25	
Air Blank		22:26	Purge Fail
Air Blank	0.000	22:27	

Data Download Date/Time: 09/21/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERKLIN, LEYON, G
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: BYRD, LONDON
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: LANCASTER

DATE/TIME OF TEST: 09/08/2022 22:45:13

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:29	
Control Test	0.077	22:30	
Air Blank	0.000	22:30	
Subject Sample	# 1 0.155	22:32	
Breath Volume	1.812	22:32	
Air Blank	0.000	22:33	
Air Blank	0.000	22:34	
Subject Sample	# 2 0.166	22:37	Volume Not Met
Breath Volume	0.285	22:37	
Air Blank	0.000	22:38	
Air Blank	0.000	22:40	
Subject Sample	# 3 0.113	22:43	Volume Not Met
Breath Volume	0.566	22:43	
Air Blank	0.000	22:44	
Control Test	0.074	22:44	Control Outside Tolerance
Air Blank	0.000	22:45	

Data Download Date/Time: 09/21/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY, MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 09/10/2022 00:47:09

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.123	00:42
Breath Volume	1.839	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.131	00:45
Breath Volume	1.566	00:45
Air Blank	0.000	00:46
Control Test	0.076	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 09/21/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERKLIN, LEYON, G
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: KERKLIN, LEYON
ARRESTING OFFICER AGENCY: Hamilton County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SERGIO

DATE/TIME OF TEST: 09/11/2022 20:55:54

Diagnostics Check	OK	20:42	
Air Blank	0.000	20:43	
Control Test	0.075	20:43	
Air Blank	0.000	20:44	
Subject Sample	# 1 0.189	20:47	Volume Not Met
Breath Volume	0.136	20:47	
Air Blank	0.000	20:48	
Air Blank	0.000	20:49	
Subject Sample	# 2 0.196	20:52	
Breath Volume	1.195	20:52	
Air Blank	0.000	20:53	
Air Blank	0.000	20:54	
Subject Sample	# 3 0.235	20:55	
Breath Volume	2.367	20:55	
Air Blank		20:55	RFI Detect
Air Blank		20:55	RFI Detect

Data Download Date/Time: 09/21/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HERRERA,JASON
ARRESTING OFFICER AGENCY: Hamilton County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: RAMIREZ PEREZ

DATE/TIME OF TEST: 09/18/2022 00:29:52

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.078	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.168	00:25
Breath Volume	1.855	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:27
Subject Sample # 2	0.167	00:28
Breath Volume	2.496	00:28
Air Blank	0.000	00:28
Control Test	0.077	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 09/21/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSKO, KRISTOPHER, R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO, KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: CORNE

DATE/TIME OF TEST: 09/19/2022 01:32:23

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.078	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.212	01:26
Breath Volume	1.988	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.211	01:30
Breath Volume	1.542	01:30
Air Blank	0.000	01:31
Control Test	0.077	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 09/21/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEDJESKE,JAMES,R
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: TEDJESKE,JAMES
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: BRATTON

DATE/TIME OF TEST: 09/20/2022 02:09:58

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.000	02:05
Breath Volume	3.253	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.000	02:08
Breath Volume	3.359	02:08
Air Blank	0.000	02:09
Control Test	0.077	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 09/21/2022 12:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/21/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/22/2022 22:15:20

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:10	
Control Test	0.078	22:10	
Air Blank	0.000	22:10	
Subject Sample	# 1 0.175	22:11	
Breath Volume	2.906	22:11	
Air Blank	0.000	22:12	
Air Blank	0.000	22:13	
Subject Sample	# 2	22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:14	
Control Test	0.077	22:14	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:15	

Data Download Date/Time: 10/19/2022 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLESBY, MICHAEL, L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY, MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 09/24/2022 06:26:56

Diagnostics Check	OK	06:20
Air Blank	0.000	06:21
Control Test	0.079	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.000	06:22
Breath Volume	2.378	06:22
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.000	06:25
Breath Volume	2.222	06:25
Air Blank	0.000	06:26
Control Test	0.078	06:26
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

Data Download Date/Time: 10/19/2022 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DELGADO,DAVID
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: NOBLES

DATE/TIME OF TEST: 09/27/2022 19:28:02

Diagnostics Check	OK	19:19
Air Blank	0.000	19:19
Control Test	0.079	19:20
Air Blank	0.000	19:20
Subject Sample # 1	0.099	19:20
Breath Volume	1.199	19:20
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.093	19:26
Breath Volume	1.988	19:26
Air Blank	0.000	19:27
Control Test	0.081	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:28

Data Download Date/Time: 10/19/2022 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SPRING

DATE/TIME OF TEST: 09/03/2022 21:22:50

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:16
Subject Sample # 1	0.146	21:18
Breath Volume	1.695	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:20
Subject Sample # 2	0.139	21:21
Breath Volume	1.500	21:21
Air Blank	0.000	21:21
Control Test	0.075	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 09/22/2022 08:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: MILLIGAN,CONNIE
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: BRADDOCK

DATE/TIME OF TEST: 09/04/2022 00:07:36

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.061	00:03
Breath Volume	2.632	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.062	00:05
Breath Volume	2.355	00:05
Air Blank	0.000	00:06
Control Test	0.078	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 09/22/2022 08:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: TEPOLE

DATE/TIME OF TEST: 09/04/2022 07:30:30

Diagnostics Check	OK	07:22
Air Blank	0.000	07:23
Control Test	0.079	07:23
Air Blank	0.000	07:24
Subject Sample # 1	0.140	07:25
Breath Volume	2.316	07:25
Air Blank	0.000	07:26
Air Blank	0.000	07:28
Subject Sample # 2	0.134	07:28
Breath Volume	2.339	07:28
Air Blank	0.000	07:29
Control Test	0.075	07:29
Air Blank	0.000	07:30
Diagnostics Check	OK	07:30

Data Download Date/Time: 09/22/2022 08:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRBY,CHAD,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KIRBY,CHAD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: GULLEY

DATE/TIME OF TEST: 09/09/2022 19:45:45

Diagnostics Check	OK	19:38
Air Blank	0.000	19:39
Control Test	0.079	19:39
Air Blank	0.000	19:40
Subject Sample # 1	0.225	19:40
Breath Volume	1.687	19:40
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.218	19:43
Breath Volume	2.625	19:43
Air Blank	0.000	19:44
Control Test	0.079	19:45
Air Blank	0.000	19:45
Diagnostics Check	OK	19:45

Data Download Date/Time: 09/22/2022 08:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,JAMES,R
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: GUERRY,CHAD
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:22

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 09/12/2022 13:01:04

Diagnostics Check	OK	12:53
Air Blank	0.000	12:53
Control Test	0.080	12:54
Air Blank	0.000	12:54
Subject Sample # 1	0.190	12:55
Breath Volume	1.957	12:55
Air Blank	0.000	12:56
Air Blank	0.000	12:57
Subject Sample # 2	0.202	12:59
Breath Volume	2.207	12:59
Air Blank	0.000	13:00
Control Test	0.079	13:00
Air Blank	0.000	13:00
Diagnostics Check	OK	13:01

Data Download Date/Time: 09/22/2022 08:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: SCOTT,BRIAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 09/13/2022 00:01:43

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:57	
Control Test	0.080	23:57	
Air Blank	0.000	23:57	
Subject Sample	# 1	00:00	RFI Detect
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.080	00:01	
Air Blank		00:01	RFI Detect
Air Blank	0.000	00:01	

Data Download Date/Time: 09/22/2022 08:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: SCOTT,BRIAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 09/13/2022 00:09:49

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:08	
Control Test	0.079	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:09	RFI Detect
Breath Volume	0.000	00:09	
Air Blank		00:09	RFI Detect
Air Blank		00:09	RFI Detect

Data Download Date/Time: 09/22/2022 08:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: SCOTT,BRIAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 09/13/2022 00:24:36

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.079	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:17	No Sample Provided
Breath Volume	0.000	00:17	
Air Blank	0.000	00:18	
Air Blank	0.000	00:20	
Subject Sample	# 2	00:23	No Sample Provided
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.079	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 09/22/2022 08:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURR,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DURR,KYLE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: VASQUEZ GUTIERREZ

DATE/TIME OF TEST: 09/16/2022 19:29:16

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.080	19:23
Air Blank	0.000	19:24
Subject Sample # 1	0.127	19:24
Breath Volume	1.980	19:24
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.120	19:27
Breath Volume	2.097	19:27
Air Blank	0.000	19:28
Control Test	0.078	19:28
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 09/22/2022 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: RIX,ARIANA
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: KEEL

DATE/TIME OF TEST: 09/17/2022 02:48:55

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.000	02:44
Breath Volume	2.300	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.000	02:47
Breath Volume	3.570	02:47
Air Blank	0.000	02:47
Control Test	0.078	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 09/22/2022 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: SHUBERT

DATE/TIME OF TEST: 09/18/2022 00:38:56

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:33	
Control Test	0.080	00:33	
Air Blank	0.000	00:34	
Subject Sample	# 1 0.251	00:37	Volume Not Met
Breath Volume	0.246	00:37	
Air Blank		00:38	Purge Fail
Air Blank	0.000	00:38	

Data Download Date/Time: 09/22/2022 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: SHUBERT

DATE/TIME OF TEST: 09/18/2022 00:41:57

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:41	
Control Test	0.072	00:41	Control Outside Tolerance
Air Blank	0.000	00:41	

Data Download Date/Time: 09/22/2022 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: SHUBERT

DATE/TIME OF TEST: 09/18/2022 00:50:02

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.077	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1 0.269	00:48	Volume Not Met
Breath Volume	0.640	00:48	
Air Blank		00:49	Purge Fail
Air Blank	0.000	00:50	

Data Download Date/Time: 09/22/2022 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SHUBERT

DATE/TIME OF TEST: 09/18/2022 01:36:03

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.077	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.253	01:30
Breath Volume	1.148	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.245	01:34
Breath Volume	1.644	01:34
Air Blank	0.000	01:35
Control Test	0.077	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:36

Data Download Date/Time: 09/22/2022 08:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: SAGER

DATE/TIME OF TEST: 09/18/2022 18:33:23

Diagnostics Check	OK	18:30	
Air Blank	0.000	18:30	
Control Test	0.079	18:31	
Air Blank	0.000	18:31	
Subject Sample	# 1	18:31	Subject Test Refused
Breath Volume	0.000	18:31	
Air Blank	0.000	18:32	
Control Test	0.079	18:32	
Air Blank	0.000	18:33	
Diagnostics Check	OK	18:33	

Data Download Date/Time: 09/22/2022 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ETHRIDGE,ALISA,B
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: MINCKS,ROBERT
ARRESTING OFFICER AGENCY: Madison PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:44

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/17/2022 14:17:36

Diagnostics Check	OK	14:10	
Air Blank	0.000	14:11	
Control Test	0.075	14:11	
Air Blank	0.000	14:12	
Subject Sample # 1	0.311	14:13	
Breath Volume	2.613	14:13	
Air Blank	0.000	14:14	
Air Blank	0.000	14:15	
Subject Sample # 2	0.310	14:15	
Breath Volume	2.343	14:15	
Air Blank	0.000	14:16	
Control Test	0.071	14:17	Control Outside Tolerance
Air Blank	0.000	14:17	

Data Download Date/Time: 09/20/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NASH,JARRED,D
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: NASH,JARRED
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 09/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: COMFORT

DATE/TIME OF TEST: 09/25/2022 02:39:57

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.020	02:35
Breath Volume	1.781	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 2	0.019	02:38
Breath Volume	1.164	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 10/18/2022 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDY,JOE,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PASLEY,JACOB
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024023281
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/18/2022 05:17:17

Diagnostics Check	OK	05:10
Air Blank	0.000	05:11
Control Test	0.082	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.169	05:12
Breath Volume	1.433	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:14
Subject Sample # 2	0.173	05:15
Breath Volume	2.332	05:15
Air Blank	0.000	05:16
Control Test	0.081	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 10/03/2022 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: ROCHEFORT

DATE/TIME OF TEST: 09/20/2022 18:06:47

Diagnostics Check	OK	18:00
Air Blank	0.000	18:00
Control Test	0.082	18:00
Air Blank	0.000	18:01
Subject Sample # 1	0.130	18:01
Breath Volume	2.414	18:01
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 2	0.126	18:05
Breath Volume	2.757	18:05
Air Blank	0.000	18:05
Control Test	0.081	18:06
Air Blank	0.000	18:06
Diagnostics Check	OK	18:06

Data Download Date/Time: 09/21/2022 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BESETT

DATE/TIME OF TEST: 09/30/2022 23:22:45

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.082	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1 0.154	23:15	Volume Not Met
Breath Volume	0.343	23:15	
Air Blank	0.000	23:16	
Air Blank	0.000	23:17	
Subject Sample	# 2 0.167	23:20	Volume Not Met
Breath Volume	0.957	23:20	
Air Blank	0.000	23:21	
Control Test	0.075	23:22	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

Data Download Date/Time: 10/03/2022 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BESETT

DATE/TIME OF TEST: 09/30/2022 23:30:48

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.076	23:26	
Air Blank	0.000	23:26	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.015	23:28	
Air Blank	0.000	23:29	
Control Test	0.080	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 10/03/2022 15:15

Subject Test Electronic Data

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,THOMAS,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SHAFFER,ROBIN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: POLNITZ

DATE/TIME OF TEST: 09/04/2022 00:48:51

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.078	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.000	00:44
Breath Volume	1.601	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.000	00:47
Breath Volume	2.703	00:47
Air Blank	0.000	00:47
Control Test	0.075	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 10/23/2022 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMS,KYLE,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SIMMS,KYLE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 16:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:16

SUBJECT'S LAST NAME: SANCHEZ GARCIA

DATE/TIME OF TEST: 09/18/2022 16:45:19

Diagnostics Check	OK	16:36	
Air Blank	0.000	16:36	
Control Test	0.077	16:37	
Air Blank	0.000	16:37	
Subject Sample # 1	0.267	16:40	
Breath Volume	2.523	16:40	
Air Blank	0.000	16:41	
Air Blank	0.000	16:42	
Subject Sample # 2		16:43	RFI Detect
Breath Volume	0.000	16:43	
Air Blank	0.000	16:44	
Control Test	0.074	16:44	Control Outside Tolerance
Air Blank	0.000	16:45	

Data Download Date/Time: 10/23/2022 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMS,KYLE,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SIMMS,KYLE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 16:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:16

SUBJECT'S LAST NAME: SANCHEZ GARCIA

DATE/TIME OF TEST: 09/18/2022 16:55:10

Diagnostics Check	OK	16:47
Air Blank	0.000	16:48
Control Test	0.076	16:48
Air Blank	0.000	16:49
Subject Sample # 1	0.251	16:50
Breath Volume	2.652	16:50
Air Blank	0.000	16:51
Air Blank	0.000	16:53
Subject Sample # 2	0.254	16:53
Breath Volume	2.054	16:53
Air Blank	0.000	16:54
Control Test	0.075	16:54
Air Blank	0.000	16:55
Diagnostics Check	OK	16:55

Data Download Date/Time: 10/23/2022 00:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/21/2022 01:13:13

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.077	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.034	01:08
Breath Volume	1.570	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.032	01:11
Breath Volume	1.480	01:11
Air Blank	0.000	01:12
Control Test	0.076	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 10/23/2022 00:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: CASCIANO

DATE/TIME OF TEST: 09/23/2022 19:37:07

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.077	19:31
Air Blank	0.000	19:32
Subject Sample # 1	0.172	19:32
Breath Volume	1.687	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.175	19:35
Breath Volume	1.515	19:35
Air Blank	0.000	19:36
Control Test	0.077	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 10/23/2022 00:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 18:37

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 09/02/2022 19:28:58

Diagnostics Check	OK	19:17	
Air Blank	0.000	19:17	
Control Test	0.079	19:18	
Air Blank	0.000	19:18	
Subject Sample	# 1 0.239	19:21	Volume Not Met
Breath Volume	0.066	19:21	
Air Blank	0.000	19:22	
Air Blank	0.000	19:24	
Subject Sample	# 2	19:27	No Sample Provided
Breath Volume	0.000	19:27	
Air Blank	0.000	19:28	
Control Test	0.078	19:28	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:28	

Data Download Date/Time: 09/15/2022 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: KEPLINGER

DATE/TIME OF TEST: 09/02/2022 23:48:38

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.078	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1 0.123	23:41	Volume Not Met
Breath Volume	0.031	23:41	
Air Blank	0.000	23:42	
Air Blank	0.000	23:43	
Subject Sample	# 2 0.120	23:47	Slope Not Level
Breath Volume	1.171	23:47	
Air Blank	0.000	23:47	
Control Test	0.078	23:48	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 09/15/2022 08:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:59

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: KEPLINGER

DATE/TIME OF TEST: 09/03/2022 00:01:26

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.078	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.114	23:56
Breath Volume	1.410	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 2	0.114	23:59
Breath Volume	1.628	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 09/15/2022 08:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER, MATTHEW, D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER, MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 09/04/2022 02:44:09

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.079	02:40	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.079	02:43	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 09/15/2022 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: WELLS,RYAN

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 15:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: SILLIVAN

DATE/TIME OF TEST: 09/08/2022 17:03:00

Diagnostics Check	OK	16:59	
Air Blank	0.000	16:59	
Control Test	0.080	16:59	
Air Blank	0.000	17:00	
Subject Sample	# 1	17:01	Subject Test Refused
Breath Volume	0.000	17:01	
Air Blank	0.000	17:02	
Control Test	0.079	17:02	
Air Blank	0.000	17:02	
Diagnostics Check	OK	17:03	

Data Download Date/Time: 09/15/2022 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HILL,CORNELIUS

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 20:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: ETSCORN

DATE/TIME OF TEST: 09/09/2022 21:38:07

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.219	21:33
Breath Volume	2.347	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:35
Subject Sample # 2	0.221	21:36
Breath Volume	1.890	21:36
Air Blank	0.000	21:37
Control Test	0.079	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 09/15/2022 08:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALAMEDA,JORGE,L
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LONGBERG,GRANT
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: DRUMHELLER

DATE/TIME OF TEST: 09/12/2022 20:10:53

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:08	
Control Test	0.079	20:08	
Air Blank	0.000	20:09	
Subject Sample	# 1	20:09	Subject Test Refused
Breath Volume	0.000	20:09	
Air Blank	0.000	20:09	
Control Test	0.078	20:10	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 09/15/2022 08:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 09/17/2022 02:32:02

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.081	02:29	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 10/12/2022 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRNKA,LAWRENCE,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LIGHTFOOT,LANE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:24

SUBJECT'S LAST NAME: HLAVATI

DATE/TIME OF TEST: 09/17/2022 11:55:57

Diagnostics Check	OK	11:44	
Air Blank	0.000	11:45	
Control Test	0.079	11:45	
Air Blank	0.000	11:45	
Subject Sample	# 1	11:49	No Sample Provided
Breath Volume	0.000	11:49	
Air Blank	0.000	11:49	
Air Blank	0.000	11:51	
Subject Sample	# 2	11:54	Subject Test Refused
Breath Volume	0.000	11:54	
Air Blank	0.000	11:54	
Control Test	0.081	11:55	
Air Blank	0.000	11:55	
Diagnostics Check	OK	11:55	

Data Download Date/Time: 10/12/2022 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER, MATTHEW, D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER, MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HOBDAV

DATE/TIME OF TEST: 09/18/2022 23:32:55

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.079	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:25	No Sample Provided
Breath Volume	0.000	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2	23:31	No Sample Provided
Breath Volume	0.000	23:31	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 10/12/2022 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HICKS,JASON

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: BOYLAN

DATE/TIME OF TEST: 09/26/2022 01:50:18

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:47	
Control Test	0.080	01:48	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:49	
Control Test	0.080	01:49	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 10/12/2022 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CORDOVA

DATE/TIME OF TEST: 09/02/2022 02:48:38

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.189	02:43
Breath Volume	1.433	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.174	02:46
Breath Volume	1.179	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 10/05/2022 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CORTLAND,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: COLANGELO,NICHOLAS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: BLEMKER

DATE/TIME OF TEST: 09/02/2022 18:52:08

Diagnostics Check	OK	18:46
Air Blank	0.000	18:46
Control Test	0.079	18:46
Air Blank	0.000	18:47
Subject Sample # 1	0.280	18:47
Breath Volume	2.714	18:47
Air Blank	0.000	18:48
Air Blank	0.000	18:50
Subject Sample # 2	0.274	18:50
Breath Volume	2.859	18:50
Air Blank	0.000	18:51
Control Test	0.079	18:51
Air Blank	0.000	18:52
Diagnostics Check	OK	18:52

Data Download Date/Time: 10/05/2022 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,J,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ARNOLD,J
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WETZEL

DATE/TIME OF TEST: 09/03/2022 20:31:51

Diagnostics Check	OK	20:29	
Air Blank	0.000	20:29	
Control Test	0.080	20:29	
Air Blank	0.000	20:30	
Subject Sample	# 1	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:30	
Control Test	0.080	20:31	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 10/05/2022 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,J,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: KOGER,A
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: TRICE

DATE/TIME OF TEST: 09/03/2022 20:47:01

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.080	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.000	20:42
Breath Volume	3.789	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.000	20:45
Breath Volume	4.234	20:45
Air Blank	0.000	20:46
Control Test	0.079	20:46
Air Blank	0.000	20:46
Diagnostics Check	OK	20:47

Data Download Date/Time: 10/05/2022 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA, ADAEL,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ACOSTA, ADAEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 09/04/2022 01:26:33

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.149	01:22
Breath Volume	1.695	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.146	01:24
Breath Volume	2.027	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 10/05/2022 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: QUINTERO GALLARDO

DATE/TIME OF TEST: 09/04/2022 02:10:10

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.062	02:05
Breath Volume	2.292	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.058	02:08
Breath Volume	1.707	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ARELLANO,ERIC
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: OROZCO VASQUEZ

DATE/TIME OF TEST: 09/04/2022 04:47:04

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.239	04:42
Breath Volume	2.054	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 2	0.231	04:45
Breath Volume	1.308	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERIC,JEAN,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FREDERIC,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: RODRIGUES

DATE/TIME OF TEST: 09/04/2022 07:52:44

Diagnostics Check	OK	07:46
Air Blank	0.000	07:46
Control Test	0.079	07:47
Air Blank	0.000	07:47
Subject Sample # 1	0.127	07:48
Breath Volume	2.511	07:48
Air Blank	0.000	07:48
Air Blank	0.000	07:50
Subject Sample # 2	0.133	07:51
Breath Volume	3.027	07:51
Air Blank	0.000	07:51
Control Test	0.079	07:52
Air Blank	0.000	07:52
Diagnostics Check	OK	07:52

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: VELASQUEZ AMAYA

DATE/TIME OF TEST: 09/04/2022 23:22:41

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.090	23:18
Breath Volume	3.632	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.086	23:21
Breath Volume	3.824	23:21
Air Blank	0.000	23:21
Control Test	0.079	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/05/2022 00:05:48

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.079	23:58	
Air Blank	0.000	23:58	
Subject Sample	# 1	00:01	No Sample Provided
Breath Volume	0.000	00:01	
Air Blank	0.000	00:02	
Air Blank	0.000	00:04	
Subject Sample	# 2	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Control Test	0.080	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BOYCE,TREVOR
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: HORNUNG

DATE/TIME OF TEST: 09/06/2022 00:10:57

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.095	00:06
Breath Volume	1.523	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.082	00:09
Breath Volume	1.503	00:09
Air Blank	0.000	00:10
Control Test	0.077	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAZIANO,STEVEN,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GRAZIANO,STEVEN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: KAMARER

DATE/TIME OF TEST: 09/08/2022 19:22:41

Diagnostics Check	OK	19:16
Air Blank	0.000	19:16
Control Test	0.079	19:17
Air Blank	0.000	19:17
Subject Sample # 1	0.158	19:18
Breath Volume	2.277	19:18
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.162	19:21
Breath Volume	2.175	19:21
Air Blank	0.000	19:21
Control Test	0.078	19:22
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: NOELSAINT

DATE/TIME OF TEST: 09/11/2022 01:34:15

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.080	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.142	01:29
Breath Volume	2.027	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:31
Subject Sample # 2	0.142	01:32
Breath Volume	2.570	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 09/11/2022 03:14:03

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.110	03:09
Breath Volume	1.238	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.109	03:12
Breath Volume	2.285	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:14

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERIC,JEAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FREDERIC,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: MCCLAIN SALTER

DATE/TIME OF TEST: 09/13/2022 03:24:53

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1 0.177	03:17	Volume Not Met
Breath Volume	0.472	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 2 0.161	03:20	
Breath Volume	1.691	03:20	
Air Blank	0.000	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 3 0.168	03:23	
Breath Volume	2.062	03:23	
Air Blank	0.000	03:23	
Control Test	0.076	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HALL,RICHARD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: STAPLES

DATE/TIME OF TEST: 09/15/2022 23:10:22

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.208	23:02
Breath Volume	1.601	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.235	23:05
Breath Volume	3.054	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:07
Subject Sample # 3	0.221	23:08
Breath Volume	2.109	23:08
Air Blank	0.000	23:09
Control Test	0.077	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,SAMANTHA,J
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: REID,SAMANTHA
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: LANGIN

DATE/TIME OF TEST: 09/17/2022 17:35:34

Diagnostics Check	OK	17:34	
Air Blank	0.000	17:34	
Control Test	0.080	17:34	
Air Blank		17:35	RFI Detect
Air Blank	0.000	17:35	

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,SAMANTHA,J
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: REID,SAMANTHA
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: LANGIN

DATE/TIME OF TEST: 09/17/2022 17:40:12

Diagnostics Check	OK	17:37	
Air Blank	0.000	17:37	
Control Test	0.079	17:37	
Air Blank	0.000	17:38	
Subject Sample	# 1 0.000	17:38	
Breath Volume	1.437	17:38	
Air Blank		17:38	Improper Sample
Air Blank	0.000	17:39	
Control Test	0.080	17:39	
Air Blank	0.000	17:40	
Diagnostics Check	OK	17:40	

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,SAMANTHA,J
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: REID,SAMANTHA
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: LANGIN

DATE/TIME OF TEST: 09/17/2022 17:47:37

Diagnostics Check	OK	17:41
Air Blank	0.000	17:42
Control Test	0.080	17:42
Air Blank	0.000	17:43
Subject Sample # 1	0.000	17:43
Breath Volume	2.929	17:43
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.000	17:46
Breath Volume	3.277	17:46
Air Blank	0.000	17:46
Control Test	0.080	17:47
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA, ADAEL,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ACOSTA, ADAEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: SILVESTRE MATIAS

DATE/TIME OF TEST: 09/19/2022 02:09:29

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.207	02:04
Breath Volume	1.585	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.197	02:07
Breath Volume	2.003	02:07
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA, ADAEL,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ACOSTA, ADAEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: NOREN

DATE/TIME OF TEST: 09/23/2022 02:46:44

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.128	02:40
Breath Volume	1.554	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.110	02:45
Breath Volume	1.433	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: BURFORD

DATE/TIME OF TEST: 09/24/2022 00:16:48

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.233	00:12
Breath Volume	2.316	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:14
Subject Sample # 2	0.227	00:15
Breath Volume	2.187	00:15
Air Blank	0.000	00:15
Control Test	0.076	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ARELLANO,ERIK
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 08/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: MENDOZA HERNANDEZ

DATE/TIME OF TEST: 09/24/2022 07:03:59

Diagnostics Check	OK	06:57
Air Blank	0.000	06:57
Control Test	0.080	06:58
Air Blank	0.000	06:58
Subject Sample # 1	0.192	06:59
Breath Volume	3.003	06:59
Air Blank	0.000	07:00
Air Blank	0.000	07:02
Subject Sample # 2	0.191	07:02
Breath Volume	2.027	07:02
Air Blank	0.000	07:03
Control Test	0.079	07:03
Air Blank	0.000	07:03
Diagnostics Check	OK	07:03

Data Download Date/Time: 10/05/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE, CHARLES, M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 09/03/2022 14:14:16

Diagnostics Check	OK	14:00	
Air Blank	0.000	14:00	
Control Test	0.077	14:00	
Air Blank	0.000	14:01	
Subject Sample	# 1 0.013	14:01	
Breath Volume	1.753	14:01	
Air Blank	0.000	14:02	
Air Blank	0.000	14:04	
Subject Sample	# 2 0.011	14:07	Volume Not Met
Breath Volume	0.699	14:07	
Air Blank	0.000	14:07	
Air Blank	0.000	14:09	
Subject Sample	# 3 0.010	14:12	
Breath Volume	1.773	14:12	
Air Blank	0.000	14:13	
Control Test	0.078	14:13	
Air Blank	0.000	14:14	
Diagnostics Check	OK	14:14	

Data Download Date/Time: 09/21/2022 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNSED, DAVID, J
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: BURNSED, DAVID
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 09/03/2022 19:24:09

Diagnostics Check	OK	19:16
Air Blank	0.000	19:16
Control Test	0.077	19:17
Air Blank	0.000	19:17
Subject Sample # 1	0.111	19:18
Breath Volume	2.855	19:18
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.105	19:22
Breath Volume	2.250	19:22
Air Blank	0.000	19:23
Control Test	0.078	19:23
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

Data Download Date/Time: 09/21/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNSED, DAVID, J
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: EDWARDS, COURTNEY
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: TAUNTON

DATE/TIME OF TEST: 09/04/2022 20:00:11

Diagnostics Check	OK	19:53
Air Blank	0.000	19:54
Control Test	0.077	19:54
Air Blank	0.000	19:55
Subject Sample # 1	0.089	19:55
Breath Volume	2.046	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.097	19:58
Breath Volume	2.316	19:58
Air Blank	0.000	19:59
Control Test	0.078	19:59
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 09/21/2022 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNSED,DAVID,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: BURNSED,DAVID
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: STECKLEY

DATE/TIME OF TEST: 09/05/2022 18:32:44

Diagnostics Check	OK	18:29	
Air Blank	0.000	18:30	
Control Test	0.077	18:30	
Air Blank	0.000	18:31	
Subject Sample	# 1	18:31	Subject Test Refused
Breath Volume	0.000	18:31	
Air Blank	0.000	18:31	
Control Test	0.077	18:32	
Air Blank	0.000	18:32	
Diagnostics Check	OK	18:32	

Data Download Date/Time: 09/21/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEYER, CHARLES, L
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: GEYER, CHARLES
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: ADMIN
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 09/13/2022 14:25:08

Diagnostics Check	OK	14:18
Air Blank	0.000	14:18
Control Test	0.078	14:18
Air Blank	0.000	14:19
Subject Sample # 1	0.000	14:20
Breath Volume	2.210	14:20
Air Blank	0.000	14:21
Air Blank	0.000	14:23
Subject Sample # 2	0.000	14:23
Breath Volume	2.070	14:23
Air Blank	0.000	14:24
Control Test	0.077	14:24
Air Blank	0.000	14:25
Diagnostics Check	OK	14:25

Data Download Date/Time: 09/21/2022 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, MATTHEW, M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: GIPSON, R
ARRESTING OFFICER AGENCY: Taylor County SO
LAST AGENCY INSPECTION: 09/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/23/2022 02:31:30

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.077	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.216	02:26
Breath Volume	1.539	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.216	02:30
Breath Volume	3.195	02:30
Air Blank	0.000	02:30
Control Test	0.077	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 10/19/2022 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD, HEATHER, M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LAPALME, MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: CONTINO

DATE/TIME OF TEST: 09/03/2022 17:31:43

Diagnostics Check	OK	17:25
Air Blank	0.000	17:26
Control Test	0.081	17:26
Air Blank	0.000	17:26
Subject Sample # 1	0.108	17:27
Breath Volume	2.027	17:27
Air Blank	0.000	17:28
Air Blank	0.000	17:29
Subject Sample # 2	0.099	17:30
Breath Volume	1.734	17:30
Air Blank	0.000	17:30
Control Test	0.080	17:31
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 09/21/2022 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,E
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 09/04/2022 18:38:24

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.082	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.112	18:33
Breath Volume	3.375	18:33
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.111	18:36
Breath Volume	3.527	18:36
Air Blank	0.000	18:37
Control Test	0.079	18:37
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 09/21/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FRANKLIN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 09/04/2022 18:59:21

Diagnostics Check	OK	18:52
Air Blank	0.000	18:53
Control Test	0.082	18:53
Air Blank	0.000	18:54
Subject Sample # 1	0.000	18:54
Breath Volume	2.253	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.000	18:57
Breath Volume	1.808	18:57
Air Blank	0.000	18:58
Control Test	0.081	18:58
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 09/21/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: FISKE

DATE/TIME OF TEST: 09/06/2022 22:16:33

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.082	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.000	22:11
Breath Volume	1.570	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.000	22:15
Breath Volume	2.097	22:15
Air Blank	0.000	22:15
Control Test	0.081	22:15
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 09/21/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/10/2022 21:18:34

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.260	21:12
Breath Volume	1.429	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.258	21:16
Breath Volume	1.945	21:16
Air Blank	0.000	21:17
Control Test	0.080	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 09/21/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: HAGER

DATE/TIME OF TEST: 09/11/2022 02:05:25

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.171	01:59
Breath Volume	1.226	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.159	02:03
Breath Volume	1.210	02:03
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 09/21/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNNE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BURKHART,Z
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 09/13/2022 17:18:43

Diagnostics Check	OK	17:07	
Air Blank	0.000	17:07	
Control Test	0.080	17:08	
Air Blank	0.000	17:08	
Subject Sample	# 1 0.230	17:11	Volume Not Met
Breath Volume	0.066	17:11	
Air Blank	0.000	17:12	
Air Blank	0.000	17:14	
Subject Sample	# 2	17:17	No Sample Provided
Breath Volume	0.000	17:17	
Air Blank	0.000	17:17	
Control Test	0.077	17:18	
Air Blank	0.000	17:18	
Diagnostics Check	OK	17:18	

Data Download Date/Time: 09/21/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CASTILLO,DENNIS
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 09/14/2022 20:43:31

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:40	
Control Test	0.081	20:41	
Air Blank	0.000	20:41	
Subject Sample	# 1	20:42	Subject Test Refused
Breath Volume	0.000	20:42	
Air Blank	0.000	20:42	
Control Test	0.079	20:42	
Air Blank	0.000	20:43	
Diagnostics Check	OK	20:43	

Data Download Date/Time: 09/21/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 09/16/2022 02:19:54

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.195	02:15
Breath Volume	1.886	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.185	02:18
Breath Volume	1.992	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 09/21/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: BORGEN

DATE/TIME OF TEST: 09/17/2022 21:17:08

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.080	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.127	21:12
Breath Volume	1.894	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:14
Subject Sample # 2	0.121	21:15
Breath Volume	1.839	21:15
Air Blank	0.000	21:16
Control Test	0.080	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 09/21/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: SCHUMACHER

DATE/TIME OF TEST: 09/20/2022 02:39:17

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.000	02:34
Breath Volume	1.898	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.000	02:37
Breath Volume	2.605	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 09/21/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SIUZDAK,COLTIN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 09/20/2022 04:05:44

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:53	
Control Test	0.081	03:53	
Air Blank	0.000	03:54	
Subject Sample	# 1 0.206	03:55	No .020 agreement
Breath Volume	2.328	03:55	
Air Blank	0.000	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 2 0.229	03:58	No .020 agreement
Breath Volume	2.167	03:58	
Air Blank	0.000	03:59	
Air Blank	0.000	04:01	
Subject Sample	# 3	04:04	No Sample Provided
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.081	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 09/21/2022 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: BARFIELD

DATE/TIME OF TEST: 09/20/2022 22:52:51

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.080	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 1 0.000	22:41	Volume Not Met
Breath Volume	0.667	22:41	
Air Blank	0.000	22:41	
Air Blank	0.000	22:43	
Subject Sample	# 2 0.000	22:45	
Breath Volume	1.441	22:45	
Air Blank	0.000	22:46	
Air Blank	0.000	22:48	
Subject Sample	# 3 0.000	22:51	Volume Not Met
Breath Volume	0.726	22:51	
Air Blank	0.000	22:51	
Control Test	0.079	22:52	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 09/21/2022 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RICH,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: KORN

DATE/TIME OF TEST: 09/22/2022 02:34:17

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.079	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.081	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 10/19/2022 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 09/24/2022 01:34:34

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.000	01:28
Breath Volume	1.148	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.000	01:33
Breath Volume	1.632	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 10/19/2022 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/24/2022 02:46:25

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.079	02:29	
Air Blank	0.000	02:30	
Subject Sample	# 1 0.162	02:33	
Breath Volume	1.343	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 2 0.155	02:38	Volume Not Met
Breath Volume	0.433	02:38	
Air Blank	0.000	02:39	
Air Blank	0.000	02:41	
Subject Sample	# 3 0.129	02:44	Volume Not Met
Breath Volume	0.589	02:44	
Air Blank	0.000	02:45	
Control Test	0.077	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 10/19/2022 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PETERS,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: VACULA

DATE/TIME OF TEST: 09/24/2022 18:15:22

Diagnostics Check	OK	18:08
Air Blank	0.000	18:08
Control Test	0.079	18:09
Air Blank	0.000	18:09
Subject Sample # 1	0.256	18:10
Breath Volume	2.046	18:10
Air Blank	0.000	18:11
Air Blank	0.000	18:13
Subject Sample # 2	0.249	18:13
Breath Volume	1.820	18:13
Air Blank	0.000	18:14
Control Test	0.080	18:14
Air Blank	0.000	18:15
Diagnostics Check	OK	18:15

Data Download Date/Time: 10/19/2022 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICH,JOSEPH,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RICH,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/26/2022 23:04:24

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.143	22:55
Breath Volume	1.636	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.168	22:59
Breath Volume	2.472	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 3	0.153	23:02
Breath Volume	1.683	23:02
Air Blank	0.000	23:03
Control Test	0.080	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 10/19/2022 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: W534544864670

DATE/TIME OF TEST: 09/03/2022 00:25:22

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.111	00:20
Breath Volume	1.324	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:22
Subject Sample # 2	0.120	00:23
Breath Volume	1.656	00:23
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 09/22/2022 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOSS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GOETZ

DATE/TIME OF TEST: 09/02/2022 21:30:05

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:27	
Control Test	0.080	21:27	
Air Blank	0.000	21:28	
Subject Sample	# 1	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:29	
Control Test	0.079	21:29	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: CRAMER,KAYLA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: PINNEY

DATE/TIME OF TEST: 09/07/2022 22:56:58

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:54	
Control Test	0.080	22:54	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:56	
Control Test	0.078	22:56	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: CRAMER,KAYLA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: PINNEY

DATE/TIME OF TEST: 09/07/2022 23:08:20

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.119	23:03
Breath Volume	1.789	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.125	23:06
Breath Volume	2.609	23:06
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALE,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCCONNELL,AUSTIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: HENRETTA

DATE/TIME OF TEST: 09/08/2022 19:02:06

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.078	18:56
Air Blank	0.000	18:56
Subject Sample # 1	0.046	18:57
Breath Volume	1.398	18:57
Air Blank	0.000	18:58
Air Blank	0.000	18:59
Subject Sample # 2	0.045	19:00
Breath Volume	1.917	19:00
Air Blank	0.000	19:01
Control Test	0.077	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALE,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 09/09/2022 22:13:02

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.079	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.268	22:08
Breath Volume	1.875	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:10
Subject Sample # 2	0.254	22:11
Breath Volume	1.402	22:11
Air Blank	0.000	22:12
Control Test	0.079	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:13

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALE,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BRAVE,S
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 09/09/2022 22:57:48

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:47	
Control Test	0.080	22:48	
Air Blank	0.000	22:48	
Subject Sample	# 1 0.118	22:51	Volume Not Met
Breath Volume	0.050	22:51	
Air Blank	0.000	22:52	
Air Blank	0.000	22:54	
Subject Sample	# 2	22:56	Subject Test Refused
Breath Volume	0.019	22:56	
Air Blank	0.000	22:56	
Control Test	0.078	22:57	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALE,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ACEVEDO,CRYSTAL
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: TOMKIN

DATE/TIME OF TEST: 09/11/2022 01:57:04

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.197	01:52
Breath Volume	2.058	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:54
Subject Sample # 2	0.192	01:55
Breath Volume	1.656	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:57

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/11/2022 02:54:52

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.086	02:49
Breath Volume	2.042	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.089	02:53
Breath Volume	1.796	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BRAVE,S
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 09/12/2022 02:20:57

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.078	02:11	
Air Blank	0.000	02:11	
Subject Sample	# 1 0.149	02:14	Volume Not Met
Breath Volume	0.015	02:14	
Air Blank	0.000	02:15	
Air Blank	0.000	02:17	
Subject Sample	# 2	02:19	Subject Test Refused
Breath Volume	0.292	02:19	
Air Blank	0.000	02:20	
Control Test	0.078	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DEMARCUS,J
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: THOMPSON,H
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: SEELY

DATE/TIME OF TEST: 09/12/2022 21:15:44

Diagnostics Check	OK	21:08
Air Blank	0.000	21:09
Control Test	0.079	21:09
Air Blank	0.000	21:09
Subject Sample # 1	0.065	21:11
Breath Volume	1.941	21:11
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.077	21:14
Breath Volume	1.417	21:14
Air Blank	0.000	21:14
Control Test	0.078	21:15
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHALL,JALEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: JACKSON,SHAUN
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: HOUCK

DATE/TIME OF TEST: 09/13/2022 16:35:29

Diagnostics Check	OK	16:23	
Air Blank	0.000	16:23	
Control Test	0.080	16:24	
Air Blank	0.000	16:24	
Subject Sample	# 1	16:27	No Sample Provided
Breath Volume	0.000	16:27	
Air Blank	0.000	16:28	
Air Blank	0.000	16:30	
Subject Sample	# 2	16:33	Volume Not Met
Breath Volume	0.734	16:33	
Air Blank	0.000	16:34	
Control Test	0.076	16:34	
Air Blank	0.000	16:35	
Diagnostics Check	OK	16:35	

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHALL,JALEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: JACKSON,SHAUN
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: HOUCK

DATE/TIME OF TEST: 09/13/2022 16:53:45

Diagnostics Check	OK	16:41	
Air Blank	0.000	16:42	
Control Test	0.078	16:42	
Air Blank	0.000	16:43	
Subject Sample	# 1 0.216	16:46	Volume Not Met
Breath Volume	0.167	16:46	
Air Blank	0.000	16:47	
Air Blank	0.000	16:48	
Subject Sample	# 2 0.244	16:51	Volume Not Met
Breath Volume	0.097	16:51	
Air Blank	0.000	16:52	
Control Test	0.076	16:53	
Air Blank	0.000	16:53	
Diagnostics Check	OK	16:53	

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMAN,HANNS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 09/14/2022 01:48:11

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.078	01:43
Breath Volume	2.816	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:45
Subject Sample # 2	0.076	01:46
Breath Volume	2.523	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 09/14/2022 23:37:33

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.000	23:32
Breath Volume	1.980	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.000	23:36
Breath Volume	3.343	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMAN,H
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: WINDHAM

DATE/TIME OF TEST: 09/17/2022 01:12:23

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.157	01:06
Breath Volume	1.171	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:08
Subject Sample # 2	0.155	01:10
Breath Volume	1.644	01:10
Air Blank	0.000	01:11
Control Test	0.079	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: FLORENCE,C
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: AXELSON

DATE/TIME OF TEST: 09/24/2022 21:41:33

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:39	
Control Test	0.079	21:39	
Air Blank	0.000	21:39	
Subject Sample	# 1	21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:40	
Control Test	0.078	21:40	
Air Blank	0.000	21:41	
Diagnostics Check	OK	21:41	

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DEMARCUS,J
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: LOPEZ CHAVEZ

DATE/TIME OF TEST: 09/25/2022 23:10:22

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:59	
Control Test	0.080	22:59	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:03	No Sample Provided
Breath Volume	0.000	23:03	
Air Blank	0.000	23:03	
Air Blank	0.000	23:05	
Subject Sample	# 2	23:08	No Sample Provided
Breath Volume	0.000	23:08	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DEMARCUS,J
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAMOS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 09/26/2022 02:49:06

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:46	
Control Test	0.080	02:46	
Air Blank	0.000	02:47	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:48	
Control Test	0.080	02:48	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: CRAMER,KAYLA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:27

SUBJECT'S LAST NAME: CUSACK

DATE/TIME OF TEST: 09/27/2022 13:10:33

Diagnostics Check	OK	13:07	
Air Blank	0.000	13:07	
Control Test	0.080	13:08	
Air Blank	0.000	13:08	
Subject Sample	# 1	13:09	Subject Test Refused
Breath Volume	0.000	13:09	
Air Blank	0.000	13:09	
Control Test	0.081	13:09	
Air Blank	0.000	13:10	
Diagnostics Check	OK	13:10	

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DEMARCUS,J
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAMOS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/28/2022 01:19:25

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:05	
Subject Sample	# 1 0.118	01:09	Volume Not Met
Breath Volume	0.000	01:09	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample	# 2 0.147	01:13	
Breath Volume	3.730	01:13	
Air Blank	0.000	01:14	
Air Blank	0.000	01:16	
Subject Sample	# 3 0.154	01:17	
Breath Volume	4.910	01:17	
Air Blank	0.000	01:18	
Control Test	0.078	01:18	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHALL,JALEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NORDLE,LACEY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:58

SUBJECT'S LAST NAME: VAICEKAUSKAS

DATE/TIME OF TEST: 09/28/2022 10:25:40

Diagnostics Check	OK	10:18
Air Blank	0.000	10:19
Control Test	0.079	10:19
Air Blank	0.000	10:20
Subject Sample # 1	0.220	10:21
Breath Volume	2.792	10:21
Air Blank	0.000	10:21
Air Blank	0.000	10:23
Subject Sample # 2	0.221	10:23
Breath Volume	2.718	10:23
Air Blank	0.000	10:24
Control Test	0.077	10:25
Air Blank	0.000	10:25
Diagnostics Check	OK	10:25

Data Download Date/Time: 09/28/2022 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: INGLIS

DATE/TIME OF TEST: 09/29/2022 19:25:00

Diagnostics Check	OK	19:18
Air Blank	0.000	19:18
Control Test	0.078	19:19
Air Blank	0.000	19:19
Subject Sample # 1	0.170	19:20
Breath Volume	1.324	19:20
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.163	19:23
Breath Volume	2.734	19:23
Air Blank	0.000	19:24
Control Test	0.079	19:24
Air Blank	0.000	19:24
Diagnostics Check	OK	19:25

Data Download Date/Time: 10/31/2022 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ,G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 09/01/2022 20:02:03

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:59	
Control Test	0.079	19:59	
Air Blank	0.000	20:00	
Subject Sample	# 1	20:00	Subject Test Refused
Breath Volume	0.000	20:00	
Air Blank	0.000	20:01	
Control Test	0.079	20:01	
Air Blank	0.000	20:01	
Diagnostics Check	OK	20:02	

Data Download Date/Time: 09/07/2022 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MIRECKI

DATE/TIME OF TEST: 09/03/2022 00:35:10

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.079	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.167	00:29
Breath Volume	2.429	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.181	00:33
Breath Volume	2.636	00:33
Air Blank	0.000	00:34
Control Test	0.078	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 09/07/2022 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: BEFUS

DATE/TIME OF TEST: 09/03/2022 18:55:49

Diagnostics Check	OK	18:52	
Air Blank	0.000	18:53	
Control Test	0.079	18:53	
Air Blank	0.000	18:54	
Subject Sample	# 1	18:54	Subject Test Refused
Breath Volume	0.000	18:54	
Air Blank	0.000	18:54	
Control Test	0.078	18:55	
Air Blank	0.000	18:55	
Diagnostics Check	OK	18:55	

Data Download Date/Time: 09/07/2022 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANSOM,AUBREY,G
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: PIETROBURGO,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 09/03/2022 23:24:21

Diagnostics Check	OK	23:23	
Air Blank		23:23	RFI Detect
Air Blank	0.000	23:24	

Data Download Date/Time: 09/07/2022 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANSOM,AUBREY,G
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: PIETROBURGO,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 09/03/2022 23:32:01

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.171	23:27
Breath Volume	2.019	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.170	23:30
Breath Volume	1.929	23:30
Air Blank	0.000	23:31
Control Test	0.077	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:32

Data Download Date/Time: 09/07/2022 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: SASSO

DATE/TIME OF TEST: 09/04/2022 02:38:07

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.079	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:30	No Sample Provided
Breath Volume	0.000	02:30	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 2	02:36	Volume Not Met
Breath Volume	0.199	02:36	
Air Blank	0.000	02:37	
Control Test	0.078	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 09/07/2022 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: SASSO

DATE/TIME OF TEST: 09/04/2022 02:52:06

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.077	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.138	02:46
Breath Volume	1.375	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.138	02:50
Breath Volume	1.433	02:50
Air Blank	0.000	02:51
Control Test	0.077	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 09/07/2022 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 09/05/2022 03:32:16

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:29	
Control Test	0.078	03:30	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 09/07/2022 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: SARMIENTO,M
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: BOMBLY

DATE/TIME OF TEST: 09/05/2022 20:46:37

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.079	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.123	20:40
Breath Volume	2.429	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.122	20:45
Breath Volume	2.371	20:45
Air Blank	0.000	20:45
Control Test	0.078	20:46
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 09/07/2022 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIGLEY, DENISE, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ, G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: VARVEL

DATE/TIME OF TEST: 09/06/2022 20:20:32

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.078	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.000	20:15
Breath Volume	2.804	20:15
Air Blank	0.000	20:16
Air Blank	0.000	20:18
Subject Sample # 2	0.000	20:19
Breath Volume	1.902	20:19
Air Blank	0.000	20:19
Control Test	0.079	20:19
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 09/07/2022 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIGLEY,DENISE,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HALVERSON,B
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: KRUNCHKOW

DATE/TIME OF TEST: 09/07/2022 04:17:07

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.079	04:13	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Control Test	0.078	04:16	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

Data Download Date/Time: 09/07/2022 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SHOLLEY

DATE/TIME OF TEST: 09/08/2022 02:04:59

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.187	01:58
Breath Volume	1.382	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:00
Subject Sample # 2	0.176	02:03
Breath Volume	1.546	02:03
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 09/09/2022 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: STANDARD

DATE/TIME OF TEST: 09/08/2022 23:01:06

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.079	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.000	22:56
Breath Volume	3.406	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:59
Subject Sample # 2	0.000	22:59
Breath Volume	1.523	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 09/09/2022 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 09/09/2022 15:50:58

Diagnostics Check	OK	15:48	
Air Blank	0.000	15:48	
Control Test	0.079	15:48	
Air Blank	0.000	15:49	
Subject Sample	# 1	15:49	Subject Test Refused
Breath Volume	0.000	15:49	
Air Blank	0.000	15:50	
Control Test	0.078	15:50	
Air Blank	0.000	15:50	
Diagnostics Check	OK	15:50	

Data Download Date/Time: 09/12/2022 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSOB, ANTHONY, T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRSION, T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BARTELS

DATE/TIME OF TEST: 09/14/2022 04:50:33

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.079	04:39	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:43	No Sample Provided
Breath Volume	0.000	04:43	
Air Blank	0.000	04:44	
Air Blank	0.000	04:46	
Subject Sample	# 2	04:49	No Sample Provided
Breath Volume	0.000	04:49	
Air Blank	0.000	04:49	
Control Test	0.079	04:49	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 09/23/2022 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, ANTHONY, T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON, T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BARTEL

DATE/TIME OF TEST: 09/14/2022 04:57:56

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:54	
Control Test	0.079	04:55	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:56	Subject Test Refused
Breath Volume	0.000	04:56	
Air Blank	0.000	04:57	
Control Test	0.079	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 09/23/2022 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: D'ADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: DEVAULT

DATE/TIME OF TEST: 09/16/2022 05:06:15

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.167	05:01
Breath Volume	1.152	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:03
Subject Sample # 2	0.170	05:04
Breath Volume	1.335	05:04
Air Blank	0.000	05:05
Control Test	0.077	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 09/23/2022 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 09/17/2022 02:15:46

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.165	02:10
Breath Volume	1.378	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.168	02:14
Breath Volume	1.746	02:14
Air Blank	0.000	02:14
Control Test	0.078	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 09/23/2022 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/18/2022 03:20:24

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.079	03:18	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Control Test	0.079	03:19	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 09/23/2022 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIGLEY, DENISE, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ, G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: HALVERSON

DATE/TIME OF TEST: 09/21/2022 00:51:11

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.129	00:46
Breath Volume	1.582	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.122	00:49
Breath Volume	1.136	00:49
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 09/23/2022 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, ANTHONY,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DEP, DULL
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: KIERNAN

DATE/TIME OF TEST: 09/23/2022 00:36:45

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.079	00:34	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.079	00:36	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 09/23/2022 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVIS,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BUCKO

DATE/TIME OF TEST: 09/23/2022 04:39:40

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:37	
Control Test	0.079	04:37	
Air Blank	0.000	04:38	
Subject Sample	# 1	04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Control Test	0.080	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 09/23/2022 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIGLEY,DENISE,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: RIVERA,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: HERSAM

DATE/TIME OF TEST: 09/23/2022 22:00:07

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:57	
Control Test	0.079	21:58	
Air Blank	0.000	21:58	
Subject Sample	# 1	21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:59	
Control Test	0.079	21:59	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

Data Download Date/Time: 11/04/2022 07:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIGLEY, DENISE, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ, G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 09/26/2022 02:22:25

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.153	02:17
Breath Volume	1.500	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.139	02:20
Breath Volume	1.117	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 11/04/2022 07:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HORTON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 09/26/2022 23:42:01

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:28	
Control Test	0.080	23:29	
Air Blank	0.000	23:29	
Subject Sample	# 1 0.157	23:32	Volume Not Met
Breath Volume	0.140	23:32	
Air Blank	0.000	23:33	
Air Blank	0.000	23:35	
Subject Sample	# 2 0.178	23:37	
Breath Volume	1.792	23:37	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample	# 3 0.188	23:40	
Breath Volume	1.183	23:40	
Air Blank	0.000	23:41	
Control Test	0.078	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 11/04/2022 07:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: D'ADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:42

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 09/28/2022 12:17:22

Diagnostics Check	OK	12:02	
Air Blank	0.000	12:03	
Control Test	0.079	12:03	
Air Blank	0.000	12:04	
Subject Sample # 1	0.000	12:06	
Breath Volume	1.562	12:06	
Air Blank	0.000	12:06	
Air Blank	0.000	12:08	
Subject Sample # 2	0.000	12:12	Volume Not Met
Breath Volume	0.117	12:12	
Air Blank	0.000	12:12	
Air Blank	0.000	12:14	
Subject Sample # 3	0.000	12:16	
Breath Volume	1.406	12:16	
Air Blank	0.000	12:16	
Control Test	0.078	12:16	
Air Blank	0.000	12:17	
Diagnostics Check	OK	12:17	

Data Download Date/Time: 11/04/2022 07:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIGLEY, DENISE, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON, T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: KOFLER

DATE/TIME OF TEST: 09/28/2022 21:18:34

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:10	
Control Test	0.079	21:10	
Air Blank	0.000	21:11	
Subject Sample	# 1	21:14	No Sample Provided
Breath Volume	0.000	21:14	
Air Blank	0.000	21:15	
Air Blank	0.000	21:17	
Subject Sample	# 2	21:17	Subject Test Refused
Breath Volume	0.000	21:17	
Air Blank	0.000	21:17	
Control Test	0.079	21:17	
Air Blank	0.000	21:18	
Diagnostics Check	OK	21:18	

Data Download Date/Time: 11/04/2022 07:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLCHARRAN,DILON,C
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: GARGANO,STEFANO
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: DUNTON

DATE/TIME OF TEST: 09/04/2022 16:12:44

Diagnostics Check	OK	16:05
Air Blank	0.000	16:06
Control Test	0.080	16:06
Air Blank	0.000	16:07
Subject Sample # 1	0.110	16:08
Breath Volume	5.269	16:08
Air Blank	0.000	16:08
Air Blank	0.000	16:10
Subject Sample # 2	0.109	16:11
Breath Volume	4.792	16:11
Air Blank	0.000	16:11
Control Test	0.080	16:12
Air Blank	0.000	16:12
Diagnostics Check	OK	16:12

Data Download Date/Time: 10/11/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ANDERTON

DATE/TIME OF TEST: 09/07/2022 01:10:45

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.000	01:06
Breath Volume	1.410	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.000	01:09
Breath Volume	1.765	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 10/11/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 09/24/2022 19:26:09

Diagnostics Check	OK	19:14
Air Blank	0.000	19:15
Control Test	0.079	19:15
Air Blank	0.000	19:16
Subject Sample # 1	0.389	19:17
Breath Volume	1.550	19:17
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample # 2	0.362	19:21
Breath Volume	2.039	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 3	0.368	19:24
Breath Volume	1.750	19:24
Air Blank	0.000	19:25
Control Test	0.078	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 10/11/2022 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: D'AGOSTINO,STEFANO,V
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: D'AGOSTINO,STEFANO
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:19

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 09/28/2022 15:50:08

Diagnostics Check	OK	15:47	
Air Blank	0.000	15:47	
Control Test	0.079	15:48	
Air Blank	0.000	15:48	
Subject Sample	# 1	15:48	Subject Test Refused
Breath Volume	0.000	15:48	
Air Blank	0.000	15:49	
Control Test	0.079	15:49	
Air Blank	0.000	15:50	
Diagnostics Check	OK	15:50	

Data Download Date/Time: 10/11/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 09/02/2022 23:28:46

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.167	23:24
Breath Volume	1.148	23:24
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.170	23:27
Breath Volume	1.113	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 09/22/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 09/03/2022 22:05:54

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.061	22:01
Breath Volume	2.386	22:01
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.059	22:04
Breath Volume	1.421	22:04
Air Blank	0.000	22:04
Control Test	0.078	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 09/22/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 09/04/2022 01:52:14

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.198	01:46
Breath Volume	1.218	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.196	01:50
Breath Volume	1.417	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 09/22/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/04/2022 19:39:57

Diagnostics Check	OK	19:39	
Air Blank		19:39	RFI Detect
Air Blank	0.000	19:39	

Data Download Date/Time: 09/22/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/04/2022 19:53:01

Diagnostics Check	OK	19:41	
Air Blank	0.000	19:41	
Control Test	0.079	19:42	
Air Blank	0.000	19:42	
Subject Sample	# 1 0.209	19:45	Volume Not Met
Breath Volume	0.355	19:45	
Air Blank	0.000	19:46	
Air Blank	0.000	19:48	
Subject Sample	# 2 0.194	19:51	Volume Not Met
Breath Volume	0.347	19:51	
Air Blank	0.000	19:52	
Control Test	0.080	19:52	
Air Blank	0.000	19:52	
Diagnostics Check	OK	19:53	

Data Download Date/Time: 09/22/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RICKETTS,JEREMIAH
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/16/2022 04:24:48

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:22	
Control Test	0.079	04:22	
Air Blank	0.000	04:23	
Subject Sample	# 1	04:23	RFI Detect
Breath Volume	0.000	04:23	
Air Blank	0.000	04:23	
Control Test	0.080	04:24	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 09/22/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RICKETTS,JEREMIAH
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/16/2022 04:29:51

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.079	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1 0.173	04:28	
Breath Volume	3.121	04:28	
Air Blank		04:28	RFI Detect
Air Blank	0.000	04:28	
Control Test	0.080	04:29	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 09/22/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RICKETTS,JEREMIAH
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/16/2022 04:34:11

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:32	
Control Test	0.080	04:32	
Air Blank	0.000	04:33	
Subject Sample	# 1 0.161	04:33	
Breath Volume	2.339	04:33	
Air Blank		04:33	RFI Detect
Air Blank		04:33	RFI Detect
Air Blank		04:34	RFI Detect

Data Download Date/Time: 09/22/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: KOVERMAN

DATE/TIME OF TEST: 09/20/2022 05:54:53

Diagnostics Check	OK	05:48
Air Blank	0.000	05:49
Control Test	0.080	05:49
Air Blank	0.000	05:50
Subject Sample # 1	0.000	05:50
Breath Volume	3.070	05:50
Air Blank	0.000	05:51
Air Blank	0.000	05:53
Subject Sample # 2	0.000	05:53
Breath Volume	2.894	05:53
Air Blank	0.000	05:53
Control Test	0.081	05:54
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

Data Download Date/Time: 09/22/2022 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/21/2022 03:13:41

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.080	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:12	RFI Detect
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Control Test	0.081	03:13	
Air Blank		03:13	RFI Detect
Air Blank	0.000	03:13	

Data Download Date/Time: 09/22/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/21/2022 03:18:26

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:16	
Subject Sample # 1	0.142	03:17	
Breath Volume	1.785	03:17	
Air Blank		03:17	RFI Detect
Air Blank		03:17	RFI Detect
Air Blank	0.000	03:18	

Data Download Date/Time: 09/22/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: DABROWSKI

DATE/TIME OF TEST: 09/01/2022 20:12:24

Diagnostics Check	OK	20:05
Air Blank	0.000	20:05
Control Test	0.080	20:05
Air Blank	0.000	20:06
Subject Sample # 1	0.249	20:07
Breath Volume	1.890	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:09
Subject Sample # 2	0.259	20:10
Breath Volume	2.093	20:10
Air Blank	0.000	20:11
Control Test	0.079	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 09/01/2022 22:56:21

Diagnostics Check	OK	22:44	
Air Blank	0.000	22:44	
Control Test	0.081	22:45	
Air Blank	0.000	22:45	
Subject Sample	# 1 0.105	22:49	Volume Not Met
Breath Volume	0.488	22:49	
Air Blank	0.000	22:49	
Air Blank	0.000	22:51	
Subject Sample	# 2 0.100	22:54	Volume Not Met
Breath Volume	0.660	22:54	
Air Blank	0.000	22:55	
Control Test	0.080	22:55	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 09/01/2022 23:06:21

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.111	23:00
Breath Volume	1.472	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample # 2	0.113	23:04
Breath Volume	2.425	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: LORD

DATE/TIME OF TEST: 09/03/2022 03:20:59

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.081	03:17	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Control Test	0.080	03:20	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: HAGLER

DATE/TIME OF TEST: 09/03/2022 03:40:53

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.104	03:35
Breath Volume	1.828	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.092	03:39
Breath Volume	1.207	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/03/2022 05:11:28

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.139	05:06
Breath Volume	1.328	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.136	05:09
Breath Volume	1.429	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DANSER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:24

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/03/2022 09:54:25

Diagnostics Check	OK	09:45
Air Blank	0.000	09:45
Control Test	0.081	09:45
Air Blank	0.000	09:46
Subject Sample # 1	0.179	09:48
Breath Volume	1.488	09:48
Air Blank	0.000	09:49
Air Blank	0.000	09:51
Subject Sample # 2	0.178	09:52
Breath Volume	1.894	09:52
Air Blank	0.000	09:53
Control Test	0.079	09:53
Air Blank	0.000	09:54
Diagnostics Check	OK	09:54

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DESIATO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: FLORES NAJERA

DATE/TIME OF TEST: 09/03/2022 13:37:44

Diagnostics Check	OK	13:25	
Air Blank	0.000	13:25	
Control Test	0.081	13:25	
Air Blank	0.000	13:26	
Subject Sample	# 1 0.201	13:29	Slope Not Level
Breath Volume	1.734	13:29	
Air Blank	0.000	13:30	
Air Blank	0.000	13:32	
Subject Sample	# 2 0.210	13:32	No .020 agreement
Breath Volume	1.332	13:32	
Air Blank	0.000	13:33	
Air Blank	0.000	13:35	
Subject Sample	# 3 0.187	13:36	No .020 agreement
Breath Volume	1.417	13:36	
Air Blank	0.000	13:36	
Control Test	0.080	13:37	
Air Blank	0.000	13:37	
Diagnostics Check	OK	13:37	

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DESIATO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: FLORES NAJERA

DATE/TIME OF TEST: 09/03/2022 13:49:57

Diagnostics Check	OK	13:43
Air Blank	0.000	13:43
Control Test	0.080	13:43
Air Blank	0.000	13:44
Subject Sample # 1	0.210	13:45
Breath Volume	2.433	13:45
Air Blank	0.000	13:46
Air Blank	0.000	13:47
Subject Sample # 2	0.199	13:48
Breath Volume	1.964	13:48
Air Blank	0.000	13:49
Control Test	0.080	13:49
Air Blank	0.000	13:49
Diagnostics Check	OK	13:49

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KROUSE,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/03/2022 22:28:23

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.101	22:23
Breath Volume	1.925	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample # 2	0.095	22:26
Breath Volume	1.968	22:26
Air Blank	0.000	22:27
Control Test	0.080	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/03/2022 23:49:36

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.081	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.147	23:44
Breath Volume	1.761	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.143	23:47
Breath Volume	1.414	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: CALZADO DIAZ

DATE/TIME OF TEST: 09/04/2022 05:11:36

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.149	05:06
Breath Volume	1.585	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.144	05:09
Breath Volume	1.300	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 09/04/2022 20:50:56

Diagnostics Check	OK	20:43
Air Blank	0.000	20:44
Control Test	0.081	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.141	20:46
Breath Volume	1.519	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:48
Subject Sample # 2	0.140	20:49
Breath Volume	1.289	20:49
Air Blank	0.000	20:49
Control Test	0.082	20:50
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 10/13/2022 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: JULIANO

DATE/TIME OF TEST: 09/05/2022 04:12:26

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.267	04:07
Breath Volume	3.230	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.256	04:10
Breath Volume	2.785	04:10
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 10/13/2022 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: PROVEZANO

DATE/TIME OF TEST: 09/05/2022 05:07:18

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.081	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.106	05:01
Breath Volume	2.035	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:03
Subject Sample # 2	0.112	05:05
Breath Volume	1.789	05:05
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 10/13/2022 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/06/2022 02:48:56

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.178	02:43
Breath Volume	2.273	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.167	02:47
Breath Volume	1.601	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 10/13/2022 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: CIRINO JR

DATE/TIME OF TEST: 09/06/2022 04:31:07

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.135	04:26
Breath Volume	1.628	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.131	04:29
Breath Volume	1.972	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 10/13/2022 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TABOADA,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: LUCERO-ROMANO

DATE/TIME OF TEST: 09/06/2022 17:01:43

Diagnostics Check	OK	16:52
Air Blank	0.000	16:52
Control Test	0.081	16:53
Air Blank	0.000	16:53
Subject Sample # 1	0.210	16:56
Breath Volume	1.609	16:56
Air Blank	0.000	16:57
Air Blank	0.000	16:58
Subject Sample # 2	0.220	17:00
Breath Volume	1.449	17:00
Air Blank	0.000	17:00
Control Test	0.080	17:01
Air Blank	0.000	17:01
Diagnostics Check	OK	17:01

Data Download Date/Time: 10/13/2022 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 09/07/2022 04:06:11

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.134	04:01
Breath Volume	1.343	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.132	04:04
Breath Volume	1.660	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 10/13/2022 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: SEALEY

DATE/TIME OF TEST: 09/07/2022 05:10:05

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.082	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.229	05:04
Breath Volume	3.941	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.219	05:08
Breath Volume	3.863	05:08
Air Blank	0.000	05:09
Control Test	0.081	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 10/13/2022 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AMOLE,J
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: GORMAN

DATE/TIME OF TEST: 09/08/2022 00:42:06

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.183	00:37
Breath Volume	2.949	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:39
Subject Sample # 2	0.185	00:40
Breath Volume	3.007	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 10/13/2022 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: FORT FERNANDEZ

DATE/TIME OF TEST: 09/08/2022 04:01:43

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.177	03:56
Breath Volume	2.785	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.176	04:00
Breath Volume	2.371	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 10/13/2022 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: SCHEMITSCH

DATE/TIME OF TEST: 09/08/2022 20:18:29

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.080	20:12
Air Blank	0.000	20:13
Subject Sample # 1	0.209	20:13
Breath Volume	3.671	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.209	20:16
Breath Volume	3.031	20:16
Air Blank	0.000	20:17
Control Test	0.081	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 10/13/2022 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: RODRIGUE

DATE/TIME OF TEST: 09/10/2022 03:43:36

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.082	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.147	03:37
Breath Volume	1.597	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.146	03:41
Breath Volume	1.539	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 10/13/2022 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: TO

DATE/TIME OF TEST: 09/10/2022 04:22:49

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.149	04:17
Breath Volume	1.644	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:19
Subject Sample # 2	0.146	04:21
Breath Volume	1.933	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 10/13/2022 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/10/2022 05:48:29

Diagnostics Check	OK	05:43	
Air Blank	0.000	05:44	
Control Test	0.082	05:44	
Air Blank	0.000	05:45	
Subject Sample	# 1	05:46	Slope Not Met
Breath Volume	0.843	05:46	
Air Blank	0.000	05:47	
Control Test	0.081	05:47	
Air Blank	0.000	05:48	
Diagnostics Check	OK	05:48	

Data Download Date/Time: 10/13/2022 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/10/2022 06:18:58

Diagnostics Check	OK	06:11
Air Blank	0.000	06:11
Control Test	0.081	06:12
Air Blank	0.000	06:12
Subject Sample # 1	0.157	06:13
Breath Volume	1.437	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:15
Subject Sample # 2	0.153	06:17
Breath Volume	1.539	06:17
Air Blank	0.000	06:18
Control Test	0.081	06:18
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 10/13/2022 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMSON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:24

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/10/2022 14:51:38

Diagnostics Check	OK	14:45
Air Blank	0.000	14:45
Control Test	0.081	14:45
Air Blank	0.000	14:46
Subject Sample # 1	0.000	14:46
Breath Volume	1.175	14:46
Air Blank	0.000	14:47
Air Blank	0.000	14:49
Subject Sample # 2	0.000	14:50
Breath Volume	2.007	14:50
Air Blank	0.000	14:50
Control Test	0.081	14:51
Air Blank	0.000	14:51
Diagnostics Check	OK	14:51

Data Download Date/Time: 10/13/2022 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 09/11/2022 03:12:44

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.082	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.158	03:08
Breath Volume	2.511	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.158	03:11
Breath Volume	2.265	03:11
Air Blank	0.000	03:11
Control Test	0.081	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 10/13/2022 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: LOPEZ TREJO

DATE/TIME OF TEST: 09/11/2022 04:21:24

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.081	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.216	04:16
Breath Volume	4.339	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.211	04:19
Breath Volume	3.855	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 10/13/2022 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 09/11/2022 05:10:40

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.081	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.096	05:05
Breath Volume	1.328	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.106	05:09
Breath Volume	1.597	05:09
Air Blank	0.000	05:09
Control Test	0.081	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 10/13/2022 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: VAZQUEZ GARCIA

DATE/TIME OF TEST: 09/11/2022 06:47:22

Diagnostics Check	OK	06:40
Air Blank	0.000	06:41
Control Test	0.081	06:41
Air Blank	0.000	06:42
Subject Sample # 1	0.191	06:42
Breath Volume	1.199	06:42
Air Blank	0.000	06:43
Air Blank	0.000	06:44
Subject Sample # 2	0.199	06:45
Breath Volume	1.757	06:45
Air Blank	0.000	06:46
Control Test	0.081	06:46
Air Blank	0.000	06:47
Diagnostics Check	OK	06:47

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 09/11/2022 22:10:31

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.081	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.130	22:05
Breath Volume	2.046	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.131	22:08
Breath Volume	1.832	22:08
Air Blank	0.000	22:09
Control Test	0.081	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: RODRIGUEZ CASTELLON

DATE/TIME OF TEST: 09/13/2022 02:11:12

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.081	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.181	02:06
Breath Volume	2.808	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.170	02:09
Breath Volume	2.253	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DUENAS TOBAR

DATE/TIME OF TEST: 09/14/2022 03:29:03

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.082	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1 0.095	03:20	Volume Not Met
Breath Volume	0.210	03:20	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample	# 2 0.139	03:24	
Breath Volume	1.585	03:24	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample	# 3 0.122	03:27	
Breath Volume	1.300	03:27	
Air Blank	0.000	03:28	
Control Test	0.082	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: AGUIRRES-CAMPOS

DATE/TIME OF TEST: 09/15/2022 04:07:34

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.160	04:01
Breath Volume	1.402	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.160	04:05
Breath Volume	1.773	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EARLYWINE, G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: BROGIE

DATE/TIME OF TEST: 09/15/2022 05:41:13

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.191	05:35
Breath Volume	1.347	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:37
Subject Sample # 2	0.193	05:39
Breath Volume	1.695	05:39
Air Blank	0.000	05:40
Control Test	0.081	05:40
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ANDERSEN

DATE/TIME OF TEST: 09/16/2022 00:21:46

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.081	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.144	00:17
Breath Volume	2.363	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.133	00:20
Breath Volume	1.878	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/16/2022 02:20:32

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.082	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.205	02:15
Breath Volume	2.304	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.199	02:18
Breath Volume	2.527	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HARDIN

DATE/TIME OF TEST: 09/16/2022 03:13:20

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.107	03:08
Breath Volume	2.066	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.102	03:11
Breath Volume	1.667	03:11
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: ANTAYA

DATE/TIME OF TEST: 09/16/2022 04:00:00

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.060	03:55
Breath Volume	2.992	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:57
Subject Sample # 2	0.058	03:58
Breath Volume	1.960	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	04:00

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEYERS,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 09/17/2022 03:25:06

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.167	03:20
Breath Volume	1.980	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.163	03:23
Breath Volume	1.199	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ANDINO

DATE/TIME OF TEST: 09/17/2022 04:11:55

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.146	04:07
Breath Volume	1.750	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.160	04:10
Breath Volume	2.925	04:10
Air Blank	0.000	04:10
Control Test	0.081	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BOUSQUET,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: BYERS

DATE/TIME OF TEST: 09/17/2022 19:44:04

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.081	19:37
Air Blank	0.000	19:37
Subject Sample # 1	0.210	19:39
Breath Volume	1.285	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:41
Subject Sample # 2	0.212	19:42
Breath Volume	1.351	19:42
Air Blank	0.000	19:43
Control Test	0.080	19:43
Air Blank	0.000	19:43
Diagnostics Check	OK	19:44

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: MESSICK

DATE/TIME OF TEST: 09/17/2022 23:00:51

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.082	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.168	22:56
Breath Volume	1.972	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.165	22:59
Breath Volume	2.046	22:59
Air Blank	0.000	22:59
Control Test	0.081	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK, I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: CARMONA

DATE/TIME OF TEST: 09/18/2022 04:43:07

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.082	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.087	04:37
Breath Volume	1.570	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.085	04:41
Breath Volume	1.343	04:41
Air Blank	0.000	04:42
Control Test	0.081	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 10/13/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: GARCIA DIAZ

DATE/TIME OF TEST: 09/18/2022 06:49:20

Diagnostics Check	OK	06:42
Air Blank	0.000	06:42
Control Test	0.081	06:42
Air Blank	0.000	06:43
Subject Sample # 1	0.197	06:44
Breath Volume	1.542	06:44
Air Blank	0.000	06:45
Air Blank	0.000	06:46
Subject Sample # 2	0.187	06:47
Breath Volume	1.992	06:47
Air Blank	0.000	06:48
Control Test	0.080	06:48
Air Blank	0.000	06:49
Diagnostics Check	OK	06:49

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RUIZ,JOHATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:24

SUBJECT'S LAST NAME: PARSLEY

DATE/TIME OF TEST: 09/18/2022 15:52:53

Diagnostics Check	OK	15:45
Air Blank	0.000	15:45
Control Test	0.082	15:45
Air Blank	0.000	15:46
Subject Sample # 1	0.263	15:47
Breath Volume	5.742	15:47
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 2	0.253	15:51
Breath Volume	3.617	15:51
Air Blank	0.000	15:51
Control Test	0.080	15:52
Air Blank	0.000	15:52
Diagnostics Check	OK	15:52

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/19/2022 04:46:34

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.081	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.099	04:42
Breath Volume	1.398	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.096	04:44
Breath Volume	1.156	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: MOHAMAD

DATE/TIME OF TEST: 09/19/2022 06:53:18

Diagnostics Check	OK	06:46
Air Blank	0.000	06:47
Control Test	0.081	06:47
Air Blank	0.000	06:48
Subject Sample # 1	0.182	06:48
Breath Volume	1.433	06:48
Air Blank	0.000	06:49
Air Blank	0.000	06:51
Subject Sample # 2	0.180	06:51
Breath Volume	1.617	06:51
Air Blank	0.000	06:52
Control Test	0.080	06:52
Air Blank	0.000	06:53
Diagnostics Check	OK	06:53

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: BUDD

DATE/TIME OF TEST: 09/21/2022 00:12:35

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.082	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.217	00:08
Breath Volume	1.824	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.235	00:10
Breath Volume	1.585	00:10
Air Blank	0.000	00:11
Control Test	0.081	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: PITA DE ALMEIDA

DATE/TIME OF TEST: 09/21/2022 03:37:55

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.136	03:33
Breath Volume	1.648	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.138	03:36
Breath Volume	1.585	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/21/2022 21:21:00

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.083	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.000	21:16
Breath Volume	1.449	21:16
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.000	21:19
Breath Volume	2.429	21:19
Air Blank	0.000	21:20
Control Test	0.081	21:20
Air Blank	0.000	21:20
Diagnostics Check	OK	21:21

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: FERRIS

DATE/TIME OF TEST: 09/22/2022 00:05:41

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.082	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.104	00:01
Breath Volume	1.984	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.112	00:04
Breath Volume	2.113	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/22/2022 02:45:17

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.082	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.135	02:40
Breath Volume	1.773	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.129	02:43
Breath Volume	1.328	02:43
Air Blank	0.000	02:44
Control Test	0.081	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: SEPPY

DATE/TIME OF TEST: 09/22/2022 22:21:41

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:07	
Control Test	0.082	22:08	
Air Blank	0.000	22:08	
Subject Sample	# 1 0.264	22:10	
Breath Volume	1.468	22:10	
Air Blank	0.000	22:10	
Air Blank	0.000	22:12	
Subject Sample	# 2 0.260	22:15	Slope Not Level
Breath Volume	1.140	22:15	
Air Blank	0.000	22:16	
Air Blank	0.000	22:18	
Subject Sample	# 3 0.280	22:19	
Breath Volume	1.734	22:19	
Air Blank	0.000	22:20	
Control Test	0.079	22:21	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 09/23/2022 02:48:09

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.148	02:43
Breath Volume	2.332	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.139	02:46
Breath Volume	1.617	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: DE PINTO

DATE/TIME OF TEST: 09/23/2022 03:43:53

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.082	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.171	03:39
Breath Volume	1.835	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.164	03:42
Breath Volume	1.230	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: HAROLD

DATE/TIME OF TEST: 09/24/2022 03:13:32

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.141	03:08
Breath Volume	3.062	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.140	03:11
Breath Volume	2.277	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: PEREZ HERRERA

DATE/TIME OF TEST: 09/24/2022 23:31:11

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.081	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.216	23:25
Breath Volume	2.046	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.232	23:29
Breath Volume	2.796	23:29
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ALMALLOHI

DATE/TIME OF TEST: 09/25/2022 04:13:28

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.082	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.119	04:08
Breath Volume	3.132	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.118	04:11
Breath Volume	3.027	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: PABON

DATE/TIME OF TEST: 09/25/2022 05:54:59

Diagnostics Check	OK	05:47
Air Blank	0.000	05:48
Control Test	0.082	05:48
Air Blank	0.000	05:49
Subject Sample # 1	0.028	05:50
Breath Volume	2.414	05:50
Air Blank	0.000	05:50
Air Blank	0.000	05:52
Subject Sample # 2	0.025	05:53
Breath Volume	1.460	05:53
Air Blank	0.000	05:54
Control Test	0.081	05:54
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

Data Download Date/Time: 10/13/2022 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: COMM VEHIC
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 09/25/2022 22:57:47

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.082	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.171	22:53
Breath Volume	1.464	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.184	22:56
Breath Volume	1.675	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 10/13/2022 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANDCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: LEGRIER

DATE/TIME OF TEST: 09/26/2022 03:26:25

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.082	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.231	03:20
Breath Volume	1.937	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.219	03:24
Breath Volume	1.457	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 10/13/2022 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 09/29/2022 22:47:59

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.082	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.165	22:42
Breath Volume	1.175	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.158	22:46
Breath Volume	1.156	22:46
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 10/13/2022 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: AMBRUSKO

DATE/TIME OF TEST: 09/24/2022 02:25:19

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.173	02:20
Breath Volume	1.425	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.187	02:23
Breath Volume	1.808	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 10/15/2022 17:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: JANIAK

DATE/TIME OF TEST: 09/26/2022 03:30:38

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.246	03:25
Breath Volume	1.871	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.228	03:28
Breath Volume	1.312	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 10/15/2022 17:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: TORRES-TAMANIZ

DATE/TIME OF TEST: 09/29/2022 21:02:34

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.078	20:56
Air Blank	0.000	20:56
Subject Sample # 1	0.133	20:57
Breath Volume	1.234	20:57
Air Blank	0.000	20:58
Air Blank	0.000	20:59
Subject Sample # 2	0.140	21:00
Breath Volume	1.644	21:00
Air Blank	0.000	21:01
Control Test	0.079	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 10/15/2022 17:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: GOEHRING,JONATHAN
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PENTKA

DATE/TIME OF TEST: 09/03/2022 03:58:30

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.000	03:53
Breath Volume	1.609	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.000	03:57
Breath Volume	1.777	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 09/24/2022 00:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PARRY,ROBERT
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: VIDAL-SILVAN

DATE/TIME OF TEST: 09/03/2022 05:22:24

Diagnostics Check	OK	05:21	
Air Blank		05:21	RFI Detect
Air Blank	0.000	05:22	

Data Download Date/Time: 09/24/2022 00:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PARRY,ROBERT
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: VIDAL-SILVAN

DATE/TIME OF TEST: 09/03/2022 05:30:15

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.079	05:24
Air Blank	0.000	05:24
Subject Sample # 1	0.177	05:25
Breath Volume	2.960	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.167	05:28
Breath Volume	2.539	05:28
Air Blank	0.000	05:29
Control Test	0.079	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 09/24/2022 00:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBBS,K,
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: HOBBS,K
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: FINNEY

DATE/TIME OF TEST: 09/03/2022 23:07:52

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.079	23:04	
Air Blank	0.000	23:05	
Subject Sample	# 1 0.120	23:06	
Breath Volume	2.910	23:06	
Air Blank		23:06	RFI Detect
Air Blank	0.000	23:06	
Control Test	0.080	23:07	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 09/24/2022 00:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBBS,KYLE,
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: HOBBS,K
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: FINNEY

DATE/TIME OF TEST: 09/03/2022 23:17:18

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.114	23:12
Breath Volume	2.554	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.102	23:15
Breath Volume	2.093	23:15
Air Blank	0.000	23:16
Control Test	0.078	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 09/24/2022 00:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: COHEN,D
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:37

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/04/2022 08:03:08

Diagnostics Check	OK	07:58	
Air Blank	0.000	07:59	
Control Test	0.081	07:59	
Air Blank	0.000	07:59	
Subject Sample	# 1	08:02	Subject Test Refused
Breath Volume	0.000	08:02	
Air Blank	0.000	08:02	
Control Test	0.078	08:02	RFI Detect
Air Blank		08:03	RFI Detect

Data Download Date/Time: 09/24/2022 00:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: COHEN,D
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:37

SUBJECT'S LAST NAME: W452804893830

DATE/TIME OF TEST: 09/04/2022 08:11:07

Diagnostics Check	OK	08:07	
Air Blank	0.000	08:07	
Control Test	0.080	08:08	
Air Blank	0.000	08:08	
Subject Sample	# 1	08:09	Subject Test Refused
Breath Volume	0.000	08:09	
Air Blank	0.000	08:10	
Control Test	0.080	08:10	
Air Blank	0.000	08:11	
Diagnostics Check	OK	08:11	

Data Download Date/Time: 09/24/2022 00:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: COHEN,D
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:37

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/04/2022 08:19:00

Diagnostics Check	OK	08:16	
Air Blank	0.000	08:16	
Control Test	0.080	08:16	
Air Blank	0.000	08:17	
Subject Sample	# 1	08:17	Subject Test Refused
Breath Volume	0.000	08:17	
Air Blank	0.000	08:18	
Control Test	0.080	08:18	
Air Blank	0.000	08:18	
Diagnostics Check	OK	08:19	

Data Download Date/Time: 09/24/2022 00:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ROUBICEK,JORDAN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: MARCONI

DATE/TIME OF TEST: 09/05/2022 00:02:43

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.203	23:58
Breath Volume	1.484	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.213	00:01
Breath Volume	1.382	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 09/24/2022 00:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: SOTO,ADRIAN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: HACKNEY

DATE/TIME OF TEST: 09/24/2022 05:58:55

Diagnostics Check	OK	05:51
Air Blank	0.000	05:52
Control Test	0.080	05:52
Air Blank	0.000	05:53
Subject Sample # 1	0.148	05:53
Breath Volume	1.593	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.146	05:57
Breath Volume	1.199	05:57
Air Blank	0.000	05:57
Control Test	0.080	05:58
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 10/26/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: WHITEHEAD,K
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: KILBRIDE

DATE/TIME OF TEST: 09/26/2022 17:50:33

Diagnostics Check	OK	17:42	
Air Blank	0.000	17:43	
Control Test	0.079	17:43	
Air Blank	0.000	17:44	
Subject Sample # 1	0.189	17:45	
Breath Volume	1.386	17:45	
Air Blank	0.000	17:46	
Air Blank	0.000	17:48	
Subject Sample # 2	0.189	17:49	
Breath Volume	2.242	17:49	
Air Blank		17:50	RFI Detect
Air Blank	0.000	17:50	

Data Download Date/Time: 10/26/2022 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: WHITEHEAD,K
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: KILBRIDE

DATE/TIME OF TEST: 09/26/2022 17:59:49

Diagnostics Check	OK	17:59	
Air Blank		17:59	RFI Detect
Air Blank	0.000	17:59	

Data Download Date/Time: 10/26/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: WHITEHEAD,K
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: KILBRIDE

DATE/TIME OF TEST: 09/26/2022 18:06:18

Diagnostics Check	OK	18:02	
Air Blank	0.000	18:02	
Control Test	0.079	18:03	
Air Blank	0.000	18:03	
Subject Sample	# 1 0.187	18:05	
Breath Volume	1.726	18:05	
Air Blank		18:05	RFI Detect
Air Blank		18:05	RFI Detect
Air Blank	0.000	18:06	

Data Download Date/Time: 10/26/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: WHITEHEAD,K
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: KILBRIDE

DATE/TIME OF TEST: 09/26/2022 18:16:55

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.078	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.184	18:12
Breath Volume	1.753	18:12
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.183	18:15
Breath Volume	1.773	18:15
Air Blank	0.000	18:15
Control Test	0.078	18:16
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 10/26/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DOUG,J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: COHEN,DUSTIN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: EUSEBIO

DATE/TIME OF TEST: 09/27/2022 22:01:38

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:58	
Control Test	0.080	21:58	
Air Blank	0.000	21:59	
Subject Sample	# 1	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.079	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 10/26/2022 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLO,D,J
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: ROLLINS,S
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: MACIAS

DATE/TIME OF TEST: 09/28/2022 04:55:11

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:52	
Control Test	0.079	04:52	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:53	Subject Test Refused
Breath Volume	0.000	04:53	
Air Blank	0.000	04:54	
Control Test	0.080	04:54	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 10/26/2022 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONDAHL,CARL,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: PEEK

DATE/TIME OF TEST: 09/28/2022 17:34:29

Diagnostics Check	OK	17:27
Air Blank	0.000	17:27
Control Test	0.080	17:28
Air Blank	0.000	17:28
Subject Sample # 1	0.162	17:30
Breath Volume	1.292	17:30
Air Blank	0.000	17:30
Air Blank	0.000	17:32
Subject Sample # 2	0.153	17:32
Breath Volume	2.019	17:32
Air Blank	0.000	17:33
Control Test	0.078	17:33
Air Blank	0.000	17:34
Diagnostics Check	OK	17:34

Data Download Date/Time: 10/26/2022 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,STEPHEN,V
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: SHIELDS,SABASTIAN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: RAWLS

DATE/TIME OF TEST: 09/28/2022 21:43:21

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.080	21:41	
Air Blank	0.000	21:42	
Subject Sample	# 1	21:42	RFI Detect
Breath Volume	0.000	21:42	
Air Blank		21:42	RFI Detect
Air Blank	0.000	21:43	

Data Download Date/Time: 10/26/2022 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNOVA,CASSANDRA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CANNOVA,CASSANDRA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: MEOLA

DATE/TIME OF TEST: 09/02/2022 18:06:38

Diagnostics Check	OK	18:00
Air Blank	0.000	18:00
Control Test	0.079	18:01
Air Blank	0.000	18:01
Subject Sample # 1	0.213	18:02
Breath Volume	1.433	18:02
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 2	0.213	18:05
Breath Volume	2.148	18:05
Air Blank	0.000	18:05
Control Test	0.079	18:06
Air Blank	0.000	18:06
Diagnostics Check	OK	18:06

Data Download Date/Time: 09/06/2022 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL, CHASE,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL, CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SCHURR

DATE/TIME OF TEST: 09/03/2022 03:15:54

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.079	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.196	03:11
Breath Volume	2.218	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.190	03:14
Breath Volume	2.421	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 09/06/2022 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: SPATARO III

DATE/TIME OF TEST: 09/06/2022 21:32:30

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.080	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.112	21:28
Breath Volume	3.273	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.110	21:31
Breath Volume	2.550	21:31
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,PATRICIA,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LINDEMANN,ANTHONY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: HANDLIN

DATE/TIME OF TEST: 09/10/2022 07:19:07

Diagnostics Check	OK	07:10	
Air Blank	0.000	07:10	
Control Test	0.080	07:11	
Air Blank	0.000	07:11	
Subject Sample	# 1	07:14	No Sample Provided
Breath Volume	0.000	07:14	
Air Blank	0.000	07:15	
Air Blank	0.000	07:17	
Subject Sample	# 2	07:17	Subject Test Refused
Breath Volume	0.000	07:17	
Air Blank	0.000	07:18	
Control Test	0.080	07:18	
Air Blank	0.000	07:19	
Diagnostics Check	OK	07:19	

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/11/2022 02:06:20

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.085	02:01
Breath Volume	3.003	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.079	02:04
Breath Volume	2.265	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,JILLIAN,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CARROLL,JILLIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/11/2022 05:33:58

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.080	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.117	05:29
Breath Volume	2.710	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.125	05:32
Breath Volume	1.636	05:32
Air Blank	0.000	05:33
Control Test	0.079	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MAMMON

DATE/TIME OF TEST: 09/15/2022 01:59:15

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.118	01:54
Breath Volume	2.304	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:56
Subject Sample # 2	0.124	01:57
Breath Volume	3.324	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: SALDIVAR

DATE/TIME OF TEST: 09/15/2022 02:50:23

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.105	02:45
Breath Volume	1.851	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.107	02:48
Breath Volume	1.906	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DURHAM, KARI
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: HERNANDEZLOPEZ

DATE/TIME OF TEST: 09/15/2022 22:41:41

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.122	22:37
Breath Volume	3.277	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.122	22:40
Breath Volume	3.046	22:40
Air Blank	0.000	22:40
Control Test	0.080	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 09/17/2022 01:22:32

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.140	01:18
Breath Volume	2.308	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.141	01:21
Breath Volume	2.734	01:21
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO PEREZ,CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 09/17/2022 04:49:43

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.079	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.186	04:45
Breath Volume	1.101	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.203	04:48
Breath Volume	2.289	04:48
Air Blank	0.000	04:48
Control Test	0.079	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GUEVARA,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 09/17/2022 14:26:48

Diagnostics Check	OK	14:19
Air Blank	0.000	14:20
Control Test	0.079	14:20
Air Blank	0.000	14:21
Subject Sample # 1	0.334	14:21
Breath Volume	1.125	14:21
Air Blank	0.000	14:22
Air Blank	0.000	14:24
Subject Sample # 2	0.351	14:25
Breath Volume	1.730	14:25
Air Blank	0.000	14:25
Control Test	0.079	14:26
Air Blank	0.000	14:26
Diagnostics Check	OK	14:26

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DAPADREA,JOSE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: TURSI

DATE/TIME OF TEST: 09/20/2022 01:51:44

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.056	01:47
Breath Volume	2.484	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.055	01:50
Breath Volume	2.914	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FRITZ,TYLER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: STAMEN

DATE/TIME OF TEST: 09/20/2022 03:49:41

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.000	03:45
Breath Volume	1.917	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.000	03:48
Breath Volume	1.500	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BEZANSON,NICHOLAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:08

SUBJECT'S LAST NAME: RISUCCI

DATE/TIME OF TEST: 09/20/2022 11:34:44

Diagnostics Check	OK	11:28
Air Blank	0.000	11:29
Control Test	0.080	11:29
Air Blank	0.000	11:30
Subject Sample # 1	0.000	11:30
Breath Volume	2.578	11:30
Air Blank	0.000	11:30
Air Blank	0.000	11:32
Subject Sample # 2	0.000	11:33
Breath Volume	2.679	11:33
Air Blank	0.000	11:33
Control Test	0.080	11:34
Air Blank	0.000	11:34
Diagnostics Check	OK	11:34

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,JILLIAN,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FRAZZETTO,ASHLEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: YESAN

DATE/TIME OF TEST: 09/20/2022 23:22:45

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.080	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.182	23:18
Breath Volume	2.289	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.196	23:21
Breath Volume	2.585	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO PEREZ,CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: LOZADA GARCIA

DATE/TIME OF TEST: 09/23/2022 00:46:37

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.165	00:41
Breath Volume	1.585	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.166	00:45
Breath Volume	1.425	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO PEREZ,CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: MARTINEZ GOMEZ

DATE/TIME OF TEST: 09/24/2022 21:30:38

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.080	21:20
Air Blank	0.000	21:20
Subject Sample # 1	0.102	21:23
Breath Volume	1.105	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.123	21:26
Breath Volume	2.890	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 3	0.121	21:29
Breath Volume	3.507	21:29
Air Blank	0.000	21:29
Control Test	0.080	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: TOMASINI

DATE/TIME OF TEST: 09/24/2022 22:04:58

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	21:59
Subject Sample # 1	0.000	22:00
Breath Volume	1.406	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.000	22:03
Breath Volume	1.433	22:03
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 10/11/2022 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: MAHONEY

DATE/TIME OF TEST: 09/11/2022 22:18:43

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:07	
Control Test	0.079	22:07	
Air Blank	0.000	22:08	
Subject Sample	# 1 0.186	22:11	Volume Not Met
Breath Volume	0.113	22:11	
Air Blank	0.000	22:12	
Air Blank	0.000	22:14	
Subject Sample	# 2 0.174	22:17	Volume Not Met
Breath Volume	0.187	22:17	
Air Blank	0.000	22:17	
Control Test	0.077	22:18	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 09/24/2022 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/14/2022 22:15:21

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:11	
Control Test	0.078	22:12	
Air Blank	0.000	22:12	
Subject Sample	# 1 0.268	22:13	
Breath Volume	1.394	22:13	
Air Blank		22:13	RFI Detect
Air Blank	0.000	22:14	
Control Test	0.080	22:14	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:15	

Data Download Date/Time: 09/24/2022 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/14/2022 22:29:35

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:17	
Control Test	0.078	22:17	
Air Blank	0.000	22:18	
Subject Sample	# 1 0.262	22:21	Volume Not Met
Breath Volume	0.835	22:21	
Air Blank	0.000	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 2 0.275	22:24	
Breath Volume	1.460	22:24	
Air Blank	0.000	22:25	
Air Blank	0.000	22:26	
Subject Sample	# 3 0.268	22:28	
Breath Volume	2.023	22:28	
Air Blank	0.000	22:28	
Control Test	0.076	22:28	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

Data Download Date/Time: 09/24/2022 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: NOTTINGHAM

DATE/TIME OF TEST: 09/16/2022 00:32:53

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.000	00:28
Breath Volume	2.609	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.000	00:31
Breath Volume	2.824	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 09/24/2022 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHWARTZ,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 09/23/2022 23:59:39

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.079	23:54	
Air Blank	0.000	23:54	
Subject Sample	# 1 0.195	23:55	
Breath Volume	3.574	23:55	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample	# 2 0.198	23:58	
Breath Volume	4.148	23:58	
Air Blank	0.000	23:58	
Control Test	0.078	23:58	
Air Blank		23:59	RFI Detect
Air Blank	0.000	23:59	

Data Download Date/Time: 09/24/2022 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: ESPARZA BELTRAN

DATE/TIME OF TEST: 09/25/2022 02:08:03

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.192	02:03
Breath Volume	2.171	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.203	02:06
Breath Volume	1.875	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:08

Data Download Date/Time: 10/23/2022 09:30