

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: LORENZO,FRANCISCO
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 09/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: CASTELLANOS

DATE/TIME OF TEST: 09/17/2022 21:03:15

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:00	
Control Test	0.078	21:00	
Air Blank	0.000	21:01	
Subject Sample	# 1	21:01	Subject Test Refused
Breath Volume	0.000	21:01	
Air Blank	0.000	21:02	
Control Test	0.078	21:02	
Air Blank	0.000	21:03	
Diagnostics Check	OK	21:03	

Data Download Date/Time: 10/03/2022 00:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURBELO,CLAUDIA,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: CURBELO,CLAUDIA
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 09/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: SOMARRIBA

DATE/TIME OF TEST: 09/24/2022 02:18:21

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.078	02:16	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:16	Subject Test Refused
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Control Test	0.078	02:17	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 10/03/2022 00:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MINA

DATE/TIME OF TEST: 09/08/2022 21:10:55

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.077	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.000	21:06
Breath Volume	2.519	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.000	21:09
Breath Volume	2.753	21:09
Air Blank	0.000	21:10
Control Test	0.075	21:10
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 10/07/2022 01:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEN,BOB,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAINTILIEN,BOB
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/09/2022 08:54:24

Diagnostics Check	OK	08:54	
Air Blank		08:54	RFI Detect
Air Blank		08:54	RFI Detect

Data Download Date/Time: 10/07/2022 01:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEEN,BOB,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAINTILIEEN,BOB
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/09/2022 09:04:31

Diagnostics Check	OK	08:57
Air Blank	0.000	08:57
Control Test	0.077	08:58
Air Blank	0.000	08:58
Subject Sample # 1	0.145	08:59
Breath Volume	1.167	08:59
Air Blank	0.000	08:59
Air Blank	0.000	09:01
Subject Sample # 2	0.142	09:02
Breath Volume	1.992	09:02
Air Blank	0.000	09:03
Control Test	0.077	09:03
Air Blank	0.000	09:04
Diagnostics Check	OK	09:04

Data Download Date/Time: 10/07/2022 01:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEEZ,HAKEEM,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LALANNE,JERLINE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 09/17/2022 05:51:27

Diagnostics Check	OK	05:51	
Air Blank		05:51	RFI Detect
Air Blank		05:51	RFI Detect

Data Download Date/Time: 10/07/2022 01:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEEZ,HAKEEM,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LALANNE,JERLINE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 09/17/2022 05:55:27

Diagnostics Check	OK	05:54	
Air Blank		05:54	RFI Detect
Air Blank	0.000	05:55	

Data Download Date/Time: 10/07/2022 01:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEEZ,HAKEEM,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LALANNE,JERLINE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 09/17/2022 06:12:18

Diagnostics Check	OK	05:58	
Air Blank	0.000	05:59	
Control Test	0.076	05:59	
Air Blank	0.000	06:00	
Subject Sample	# 1 0.160	06:03	Volume Not Met
Breath Volume	0.421	06:03	
Air Blank	0.000	06:03	
Air Blank	0.000	06:05	
Subject Sample	# 2 0.185	06:07	
Breath Volume	2.421	06:07	
Air Blank	0.000	06:08	
Air Blank	0.000	06:10	
Subject Sample	# 3 0.189	06:10	
Breath Volume	2.226	06:10	
Air Blank	0.000	06:11	
Control Test	0.073	06:11	Control Outside Tolerance
Air Blank	0.000	06:12	

Data Download Date/Time: 10/07/2022 01:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEEZ,HAKEEM,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LALANNE,JERLINE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 09/17/2022 07:13:26

Diagnostics Check	OK	07:05
Air Blank	0.000	07:06
Control Test	0.076	07:06
Air Blank	0.000	07:06
Subject Sample # 1	0.176	07:08
Breath Volume	1.980	07:08
Air Blank	0.000	07:09
Air Blank	0.000	07:11
Subject Sample # 2	0.167	07:11
Breath Volume	1.234	07:11
Air Blank	0.000	07:12
Control Test	0.075	07:12
Air Blank	0.000	07:13
Diagnostics Check	OK	07:13

Data Download Date/Time: 10/07/2022 01:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLSTON,ANDREW,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HOLSTON,ANDREW
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 09/18/2022 01:20:51

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.077	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.078	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 10/07/2022 01:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SELENT,SCOTT
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: ARTEAGA

DATE/TIME OF TEST: 09/18/2022 04:04:19

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.077	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.000	04:00
Breath Volume	2.851	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.000	04:02
Breath Volume	2.750	04:02
Air Blank	0.000	04:03
Control Test	0.077	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 10/07/2022 01:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS, WALEX,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOUIS, WALEX
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 09/18/2022 20:36:20

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:26	
Control Test	0.077	20:27	
Air Blank	0.000	20:27	
Subject Sample	# 1 0.366	20:28	
Breath Volume	2.609	20:28	
Air Blank	0.000	20:28	
Air Blank	0.000	20:30	
Subject Sample	# 2 0.335	20:31	
Breath Volume	1.335	20:31	
Air Blank	0.000	20:32	
Air Blank	0.000	20:34	
Subject Sample	# 3 0.347	20:34	
Breath Volume	2.000	20:34	
Air Blank	0.000	20:35	
Control Test	0.074	20:35	Control Outside Tolerance
Air Blank	0.000	20:36	

Data Download Date/Time: 10/07/2022 01:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLSTON,ANDREW,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CARSON,MARQUIS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 09/25/2022 00:53:22

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.077	00:44	
Air Blank	0.000	00:44	
Subject Sample # 1	0.303	00:47	
Breath Volume	2.164	00:47	
Air Blank	0.000	00:48	
Air Blank	0.000	00:50	
Subject Sample # 2	0.299	00:51	
Breath Volume	2.050	00:51	
Air Blank	0.000	00:52	
Control Test	0.073	00:52	Control Outside Tolerance
Air Blank	0.000	00:53	

Data Download Date/Time: 10/07/2022 01:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLSTON,ANDREW,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CARSON,MARQUIS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 09/25/2022 01:03:37

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.075	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.298	00:58
Breath Volume	1.300	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.306	01:01
Breath Volume	1.332	01:01
Air Blank	0.000	01:02
Control Test	0.077	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 10/07/2022 01:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ALMONTE

DATE/TIME OF TEST: 09/18/2022 01:59:14

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.078	01:47	
Air Blank	0.000	01:47	
Subject Sample	# 1 0.143	01:50	Volume Not Met
Breath Volume	0.015	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:53	
Subject Sample	# 2 0.171	01:54	
Breath Volume	1.898	01:54	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample	# 3 0.180	01:57	
Breath Volume	2.480	01:57	
Air Blank	0.000	01:58	
Control Test	0.079	01:58	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 10/06/2022 05:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MILLIEN

DATE/TIME OF TEST: 09/25/2022 01:47:36

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.183	01:42
Breath Volume	1.449	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.192	01:45
Breath Volume	1.375	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 10/06/2022 05:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 09/02/2022 03:08:59

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.078	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:08	
Control Test	0.078	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 09/16/2022 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PAUL,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: FJERAN

DATE/TIME OF TEST: 09/03/2022 07:37:22

Diagnostics Check	OK	07:30
Air Blank	0.000	07:31
Control Test	0.077	07:31
Air Blank	0.000	07:32
Subject Sample # 1	0.123	07:32
Breath Volume	2.562	07:32
Air Blank	0.000	07:33
Air Blank	0.000	07:35
Subject Sample # 2	0.108	07:35
Breath Volume	1.578	07:35
Air Blank	0.000	07:36
Control Test	0.077	07:36
Air Blank	0.000	07:37
Diagnostics Check	OK	07:37

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 09/03/2022 21:15:08

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.077	21:09
Air Blank	0.000	21:10
Subject Sample # 1	0.214	21:10
Breath Volume	1.101	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.209	21:13
Breath Volume	1.109	21:13
Air Blank	0.000	21:14
Control Test	0.075	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ELMORE

DATE/TIME OF TEST: 09/04/2022 03:10:04

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.078	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.077	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: TARRANCE

DATE/TIME OF TEST: 09/04/2022 04:31:20

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.077	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.153	04:26
Breath Volume	1.328	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.148	04:29
Breath Volume	1.332	04:29
Air Blank	0.000	04:30
Control Test	0.077	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: QUARLES

DATE/TIME OF TEST: 09/05/2022 03:21:29

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.077	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.043	03:16
Breath Volume	1.140	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.042	03:20
Breath Volume	1.281	03:20
Air Blank	0.000	03:20
Control Test	0.078	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOCKERS,B
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: KOHN

DATE/TIME OF TEST: 09/05/2022 22:47:59

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.077	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1 0.243	22:40	Volume Not Met
Breath Volume	0.707	22:40	
Air Blank	0.000	22:41	
Air Blank	0.000	22:43	
Subject Sample	# 2 0.213	22:46	Volume Not Met
Breath Volume	0.992	22:46	
Air Blank	0.000	22:47	
Control Test	0.076	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Neptune Beach PD
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 09/05/2022 23:52:34

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.077	23:50	
Air Blank	0.000	23:51	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:51	
Control Test	0.077	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 09/07/2022 03:24:40

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.078	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Control Test	0.077	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: EUTIMIO LOPEZ

DATE/TIME OF TEST: 09/08/2022 02:50:26

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.077	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.078	02:49	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 09/09/2022 00:24:51

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.077	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.254	00:20
Breath Volume	1.703	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.251	00:23
Breath Volume	1.558	00:23
Air Blank	0.000	00:23
Control Test	0.077	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: HUDGINS

DATE/TIME OF TEST: 09/09/2022 04:40:06

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.077	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.075	04:35
Breath Volume	1.308	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.079	04:38
Breath Volume	1.710	04:38
Air Blank	0.000	04:39
Control Test	0.077	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: AVILA MARTINEZ

DATE/TIME OF TEST: 09/10/2022 02:22:35

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.078	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1 0.144	02:12	
Breath Volume	1.179	02:12	
Air Blank	0.000	02:12	
Air Blank	0.000	02:14	
Subject Sample	# 2 0.126	02:17	Volume Not Met
Breath Volume	0.390	02:17	
Air Blank	0.000	02:18	
Air Blank	0.000	02:20	
Subject Sample	# 3 0.143	02:21	
Breath Volume	1.828	02:21	
Air Blank	0.000	02:21	
Control Test	0.077	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LYON,L
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: SAUSVILLE

DATE/TIME OF TEST: 09/10/2022 21:07:47

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:05	
Control Test	0.077	21:05	
Air Blank	0.000	21:06	
Subject Sample	# 1	21:06	Subject Test Refused
Breath Volume	0.000	21:06	
Air Blank	0.000	21:06	
Control Test	0.077	21:07	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/11/2022 04:48:22

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.078	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.101	04:43
Breath Volume	1.601	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.097	04:46
Breath Volume	1.535	04:46
Air Blank	0.000	04:47
Control Test	0.077	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: TOLMACHEV

DATE/TIME OF TEST: 09/11/2022 19:59:07

Diagnostics Check	OK	19:56	
Air Blank	0.000	19:56	
Control Test	0.077	19:57	
Air Blank	0.000	19:57	
Subject Sample	# 1	19:57	Subject Test Refused
Breath Volume	0.000	19:57	
Air Blank	0.000	19:58	
Control Test	0.077	19:58	
Air Blank	0.000	19:59	
Diagnostics Check	OK	19:59	

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/12/2022 03:52:21

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.078	03:50	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.078	03:51	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/14/2022 03:52:01

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.078	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.089	03:46
Breath Volume	1.699	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.090	03:50
Breath Volume	1.535	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:52

Data Download Date/Time: 09/16/2022 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: WELTY

DATE/TIME OF TEST: 09/02/2022 00:23:50

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.077	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.000	00:19
Breath Volume	3.695	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.000	00:22
Breath Volume	2.351	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 09/16/2022 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: HAKIMZODA

DATE/TIME OF TEST: 09/02/2022 02:42:39

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.077	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.206	02:38
Breath Volume	1.640	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.198	02:41
Breath Volume	1.218	02:41
Air Blank	0.000	02:41
Control Test	0.077	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 09/16/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: DIAZ RIVERA

DATE/TIME OF TEST: 09/02/2022 20:08:24

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.077	20:03
Air Blank	0.000	20:03
Subject Sample # 1	0.000	20:04
Breath Volume	2.937	20:04
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.000	20:07
Breath Volume	2.816	20:07
Air Blank	0.000	20:07
Control Test	0.077	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 09/16/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FANFAN,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: TYO

DATE/TIME OF TEST: 09/03/2022 05:37:01

Diagnostics Check	OK	05:34	
Air Blank	0.000	05:34	
Control Test	0.078	05:34	
Air Blank	0.000	05:35	
Subject Sample	# 1	05:35	Subject Test Refused
Breath Volume	0.000	05:35	
Air Blank	0.000	05:36	
Control Test	0.077	05:36	
Air Blank	0.000	05:36	
Diagnostics Check	OK	05:37	

Data Download Date/Time: 09/16/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TYLER,M
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: PRESTON

DATE/TIME OF TEST: 09/03/2022 21:39:03

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.077	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.018	21:34
Breath Volume	2.886	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.018	21:37
Breath Volume	2.242	21:37
Air Blank	0.000	21:38
Control Test	0.078	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:39

Data Download Date/Time: 09/16/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 09/04/2022 04:09:09

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.094	04:03
Breath Volume	1.656	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.096	04:07
Breath Volume	2.156	04:07
Air Blank	0.000	04:08
Control Test	0.077	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 09/16/2022 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES, ALFONSO,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CASEY

DATE/TIME OF TEST: 09/04/2022 23:36:09

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.078	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.000	23:31
Breath Volume	1.859	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.000	23:34
Breath Volume	1.933	23:34
Air Blank	0.000	23:35
Control Test	0.077	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 09/16/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: TORREGANO

DATE/TIME OF TEST: 09/06/2022 01:34:05

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.078	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:33	
Control Test	0.077	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 09/16/2022 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: ZHAO

DATE/TIME OF TEST: 09/08/2022 04:10:05

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.078	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.175	04:05
Breath Volume	2.792	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.169	04:08
Breath Volume	2.171	04:08
Air Blank	0.000	04:09
Control Test	0.077	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:10

Data Download Date/Time: 09/16/2022 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: FLENER

DATE/TIME OF TEST: 09/10/2022 16:50:35

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.078	16:41
Air Blank	0.000	16:42
Subject Sample # 1	0.164	16:43
Breath Volume	1.218	16:43
Air Blank	0.000	16:44
Air Blank	0.000	16:46
Subject Sample # 2	0.168	16:49
Breath Volume	1.824	16:49
Air Blank	0.000	16:49
Control Test	0.077	16:50
Air Blank	0.000	16:50
Diagnostics Check	OK	16:50

Data Download Date/Time: 09/16/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: O'CONNELL,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/11/2022 04:39:19

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.078	04:37	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:38	
Control Test	0.079	04:38	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 09/16/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: ANGE

DATE/TIME OF TEST: 09/12/2022 01:31:28

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.077	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.226	01:26
Breath Volume	3.773	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.226	01:29
Breath Volume	3.964	01:29
Air Blank	0.000	01:30
Control Test	0.077	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 09/16/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES, ALFONSO,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE, S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 09/15/2022 00:10:34

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.077	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.128	00:06
Breath Volume	1.195	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.120	00:09
Breath Volume	1.914	00:09
Air Blank	0.000	00:09
Control Test	0.077	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 09/16/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/16/2022 02:16:56

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.077	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.125	02:12
Breath Volume	2.058	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.123	02:15
Breath Volume	1.750	02:15
Air Blank	0.000	02:16
Control Test	0.077	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 09/16/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 09/16/2022 03:30:32

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.077	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.128	03:25
Breath Volume	1.757	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.125	03:29
Breath Volume	1.652	03:29
Air Blank	0.000	03:29
Control Test	0.077	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 09/16/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/16/2022 04:33:41

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:31	
Control Test	0.078	04:31	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Control Test	0.078	04:33	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 09/16/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: BRESSLER

DATE/TIME OF TEST: 09/16/2022 22:55:08

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.078	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.014	22:50
Breath Volume	2.757	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.011	22:53
Breath Volume	2.800	22:53
Air Blank	0.000	22:54
Control Test	0.077	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 10/13/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 09/17/2022 01:10:41

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.077	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.242	01:06
Breath Volume	1.855	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.242	01:09
Breath Volume	1.785	01:09
Air Blank	0.000	01:09
Control Test	0.077	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 10/13/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PAILLANT,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: COPPOCK

DATE/TIME OF TEST: 09/17/2022 02:40:58

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.078	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.218	02:36
Breath Volume	1.886	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:38
Subject Sample # 2	0.217	02:39
Breath Volume	2.144	02:39
Air Blank	0.000	02:40
Control Test	0.077	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 10/13/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 09/17/2022 03:37:58

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.142	03:33
Breath Volume	1.621	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.132	03:36
Breath Volume	1.585	03:36
Air Blank	0.000	03:37
Control Test	0.077	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 10/13/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 09/17/2022 04:57:16

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:54	
Control Test	0.078	04:55	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:55	Subject Test Refused
Breath Volume	0.000	04:55	
Air Blank	0.000	04:56	
Control Test	0.078	04:56	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 10/13/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRY,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:04

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 09/17/2022 12:30:31

Diagnostics Check	OK	12:24
Air Blank	0.000	12:24
Control Test	0.078	12:25
Air Blank	0.000	12:25
Subject Sample # 1	0.158	12:26
Breath Volume	3.730	12:26
Air Blank	0.000	12:26
Air Blank	0.000	12:28
Subject Sample # 2	0.157	12:29
Breath Volume	2.839	12:29
Air Blank	0.000	12:29
Control Test	0.078	12:29
Air Blank	0.000	12:30
Diagnostics Check	OK	12:30

Data Download Date/Time: 10/13/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ALADICS

DATE/TIME OF TEST: 09/17/2022 22:33:46

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
Control Test	0.078	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:32	
Control Test	0.077	22:33	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 10/13/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: HERNANDEZ GARCIA

DATE/TIME OF TEST: 09/18/2022 01:26:34

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:24	
Control Test	0.078	01:24	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:25	Subject Test Refused
Breath Volume	0.000	01:25	
Air Blank	0.000	01:25	
Control Test	0.077	01:25	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DOTA ESPINOZA

DATE/TIME OF TEST: 09/18/2022 03:32:12

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.170	03:27
Breath Volume	2.437	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.172	03:30
Breath Volume	2.136	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: DIMOPULOS

DATE/TIME OF TEST: 09/18/2022 04:40:33

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.078	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.124	04:36
Breath Volume	1.937	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.128	04:39
Breath Volume	1.714	04:39
Air Blank	0.000	04:39
Control Test	0.077	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 09/18/2022 20:05:45

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:03	
Control Test	0.079	20:03	
Air Blank	0.000	20:04	
Subject Sample	# 1	20:04	Subject Test Refused
Breath Volume	0.000	20:04	
Air Blank	0.000	20:04	
Control Test	0.078	20:05	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: SIBILA

DATE/TIME OF TEST: 09/19/2022 02:14:15

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.078	02:12	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Control Test	0.078	02:13	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GRAHEK

DATE/TIME OF TEST: 09/19/2022 21:24:25

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:20	
Control Test	0.078	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:23	Subject Test Refused
Breath Volume	0.000	21:23	
Air Blank	0.000	21:23	
Control Test	0.078	21:23	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:24	

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/20/2022 03:53:35

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.059	03:49
Breath Volume	1.769	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.062	03:52
Breath Volume	1.304	03:52
Air Blank	0.000	03:52
Control Test	0.077	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,P,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: HEWITT

DATE/TIME OF TEST: 09/20/2022 17:16:58

Diagnostics Check	OK	17:09
Air Blank	0.000	17:10
Control Test	0.078	17:10
Air Blank	0.000	17:11
Subject Sample # 1	0.075	17:12
Breath Volume	3.246	17:12
Air Blank	0.000	17:13
Air Blank	0.000	17:15
Subject Sample # 2	0.078	17:15
Breath Volume	3.160	17:15
Air Blank	0.000	17:16
Control Test	0.078	17:16
Air Blank	0.000	17:16
Diagnostics Check	OK	17:16

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 09/20/2022 19:38:37

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:36	
Control Test	0.078	19:36	
Air Blank	0.000	19:37	
Subject Sample	# 1	19:37	Subject Test Refused
Breath Volume	0.000	19:37	
Air Blank	0.000	19:37	
Control Test	0.079	19:38	
Air Blank	0.000	19:38	
Diagnostics Check	OK	19:38	

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 09/20/2022 23:01:47

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.175	22:57
Breath Volume	1.328	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.166	23:00
Breath Volume	1.964	23:00
Air Blank	0.000	23:00
Control Test	0.078	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PINCKNEY,R
ARRESTING OFFICER AGENCY: Jax Airport Auth
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 09/21/2022 03:35:58

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.233	03:31
Breath Volume	2.679	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.237	03:34
Breath Volume	2.859	03:34
Air Blank	0.000	03:35
Control Test	0.077	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 09/21/2022 20:41:52

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.078	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.000	20:37
Breath Volume	1.394	20:37
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.000	20:40
Breath Volume	1.691	20:40
Air Blank	0.000	20:40
Control Test	0.077	20:41
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DUNNEBACK,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: SCHOVERLING

DATE/TIME OF TEST: 09/22/2022 05:21:10

Diagnostics Check	OK	05:14
Air Blank	0.000	05:15
Control Test	0.078	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.172	05:16
Breath Volume	3.105	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.174	05:19
Breath Volume	2.152	05:19
Air Blank	0.000	05:20
Control Test	0.078	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOROZOV,I
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BURKHARD

DATE/TIME OF TEST: 09/23/2022 23:42:15

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.078	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.104	23:37
Breath Volume	1.949	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.100	23:40
Breath Volume	2.132	23:40
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 09/24/2022 03:02:01

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.078	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.078	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: VAN NETZEL

DATE/TIME OF TEST: 09/24/2022 03:33:32

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.078	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.077	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 09/24/2022 20:47:23

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.078	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.224	20:42
Breath Volume	2.507	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.231	20:45
Breath Volume	2.062	20:45
Air Blank	0.000	20:46
Control Test	0.077	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: STIPE

DATE/TIME OF TEST: 09/24/2022 22:00:17

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:57	
Control Test	0.078	21:58	
Air Blank	0.000	21:58	
Subject Sample	# 1	21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:59	
Control Test	0.078	21:59	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: MARTIN-BROWN

DATE/TIME OF TEST: 09/24/2022 23:09:38

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.078	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.079	23:05
Breath Volume	1.496	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.090	23:08
Breath Volume	1.382	23:08
Air Blank	0.000	23:08
Control Test	0.076	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNES,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MATUTE HERNANDEZ

DATE/TIME OF TEST: 09/25/2022 00:29:00

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.078	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.240	00:24
Breath Volume	2.871	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.227	00:27
Breath Volume	2.125	00:27
Air Blank	0.000	00:28
Control Test	0.077	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:29

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTELT,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 09/25/2022 03:04:37

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.078	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Control Test	0.078	03:04	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 09/25/2022 04:25:37

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.079	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Control Test	0.078	04:25	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 09/25/2022 05:04:13

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.078	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.172	04:59
Breath Volume	3.160	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.166	05:02
Breath Volume	2.453	05:02
Air Blank	0.000	05:03
Control Test	0.078	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SYSWERDA,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:39

SUBJECT'S LAST NAME: BOLDEN-LOPEZ

DATE/TIME OF TEST: 09/25/2022 12:05:54

Diagnostics Check	OK	11:59
Air Blank	0.000	12:00
Control Test	0.078	12:00
Air Blank	0.000	12:01
Subject Sample # 1	0.146	12:01
Breath Volume	2.406	12:01
Air Blank	0.000	12:02
Air Blank	0.000	12:03
Subject Sample # 2	0.137	12:04
Breath Volume	2.007	12:04
Air Blank	0.000	12:04
Control Test	0.078	12:05
Air Blank	0.000	12:05
Diagnostics Check	OK	12:05

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/26/2022 01:02:12

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.077	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.226	00:52
Breath Volume	2.488	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.189	00:57
Breath Volume	1.628	00:57
Air Blank	0.000	00:58
Air Blank	0.000	00:59
Subject Sample # 3	0.196	01:00
Breath Volume	1.855	01:00
Air Blank	0.000	01:01
Control Test	0.077	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 09/26/2022 03:46:16

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.077	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.169	03:41
Breath Volume	3.554	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.166	03:44
Breath Volume	2.023	03:44
Air Blank	0.000	03:45
Control Test	0.077	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 10/13/2022 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: SANPEI

DATE/TIME OF TEST: 09/27/2022 00:37:32

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.077	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.044	00:33
Breath Volume	2.335	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.045	00:36
Breath Volume	2.214	00:36
Air Blank	0.000	00:36
Control Test	0.077	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 10/13/2022 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/27/2022 01:20:20

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.077	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:18	RFI Detect
Breath Volume	0.000	01:18	
Air Blank	0.000	01:19	
Control Test	0.078	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 10/13/2022 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/27/2022 01:28:00

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.078	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.019	01:23
Breath Volume	1.824	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:25
Subject Sample # 2	0.018	01:26
Breath Volume	1.597	01:26
Air Blank	0.000	01:27
Control Test	0.077	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:28

Data Download Date/Time: 10/13/2022 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SAMUELS

DATE/TIME OF TEST: 09/28/2022 02:52:30

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.079	02:47
Breath Volume	1.269	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.075	02:51
Breath Volume	1.523	02:51
Air Blank	0.000	02:51
Control Test	0.077	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 10/13/2022 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: KILLEBREW

DATE/TIME OF TEST: 09/28/2022 04:57:01

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:54	
Control Test	0.078	04:54	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:55	Subject Test Refused
Breath Volume	0.000	04:55	
Air Blank	0.000	04:56	
Control Test	0.078	04:56	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 10/13/2022 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 09/29/2022 04:38:41

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.077	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.157	04:34
Breath Volume	1.585	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.155	04:37
Breath Volume	1.828	04:37
Air Blank	0.000	04:37
Control Test	0.078	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 10/13/2022 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HICKEY,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 09/29/2022 20:05:18

Diagnostics Check	OK	19:58
Air Blank	0.000	19:59
Control Test	0.078	19:59
Air Blank	0.000	20:00
Subject Sample # 1	0.174	20:00
Breath Volume	4.640	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.177	20:03
Breath Volume	4.503	20:03
Air Blank	0.000	20:04
Control Test	0.078	20:04
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 10/13/2022 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOROZOV,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: SAPP

DATE/TIME OF TEST: 09/29/2022 23:08:54

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.079	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.149	23:04
Breath Volume	1.585	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.156	23:07
Breath Volume	1.816	23:07
Air Blank	0.000	23:08
Control Test	0.077	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 10/13/2022 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: HOOKS

DATE/TIME OF TEST: 09/30/2022 20:55:00

Diagnostics Check	OK	20:48
Air Blank	0.000	20:49
Control Test	0.078	20:49
Air Blank	0.000	20:49
Subject Sample # 1	0.040	20:50
Breath Volume	2.019	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.041	20:53
Breath Volume	1.968	20:53
Air Blank	0.000	20:54
Control Test	0.078	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:55

Data Download Date/Time: 10/13/2022 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 09/30/2022 23:06:50

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.077	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.142	23:02
Breath Volume	2.066	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:04
Subject Sample # 2	0.146	23:05
Breath Volume	1.945	23:05
Air Blank	0.000	23:05
Control Test	0.078	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 10/13/2022 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: FILLINGIM

DATE/TIME OF TEST: 09/01/2022 21:18:24

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.129	21:13
Breath Volume	1.796	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.134	21:16
Breath Volume	1.843	21:16
Air Blank	0.000	21:17
Control Test	0.078	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 09/16/2022 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/01/2022 23:28:13

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.079	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.106	23:23
Breath Volume	2.964	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.107	23:26
Breath Volume	2.777	23:26
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 09/16/2022 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SWOPE

DATE/TIME OF TEST: 09/02/2022 01:23:25

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.080	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.079	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 09/16/2022 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: FREDRICKSON

DATE/TIME OF TEST: 09/02/2022 02:24:49

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.090	02:20
Breath Volume	2.261	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.089	02:23
Breath Volume	2.972	02:23
Air Blank	0.000	02:23
Control Test	0.079	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 09/16/2022 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HEATH,M
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: BUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: LISECKI

DATE/TIME OF TEST: 09/02/2022 19:26:39

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:24	
Control Test	0.079	19:24	
Air Blank	0.000	19:24	
Subject Sample	# 1	19:25	Subject Test Refused
Breath Volume	0.000	19:25	
Air Blank	0.000	19:25	
Control Test	0.079	19:26	
Air Blank	0.000	19:26	
Diagnostics Check	OK	19:26	

Data Download Date/Time: 09/16/2022 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: MANAYAN

DATE/TIME OF TEST: 09/03/2022 00:34:52

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.079	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.021	00:30
Breath Volume	2.855	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.021	00:33
Breath Volume	2.468	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 09/16/2022 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 09/03/2022 04:32:03

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.228	04:26
Breath Volume	2.097	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.220	04:30
Breath Volume	1.593	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:32

Data Download Date/Time: 09/16/2022 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: COATES JR

DATE/TIME OF TEST: 09/03/2022 05:48:00

Diagnostics Check	OK	05:41
Air Blank	0.000	05:41
Control Test	0.079	05:42
Air Blank	0.000	05:42
Subject Sample # 1	0.172	05:43
Breath Volume	2.320	05:43
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample # 2	0.165	05:46
Breath Volume	1.613	05:46
Air Blank	0.000	05:47
Control Test	0.079	05:47
Air Blank	0.000	05:47
Diagnostics Check	OK	05:48

Data Download Date/Time: 09/16/2022 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRY,H
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 09/03/2022 20:48:38

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.185	20:44
Breath Volume	2.527	20:44
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.187	20:47
Breath Volume	2.476	20:47
Air Blank	0.000	20:47
Control Test	0.078	20:48
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 09/16/2022 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 09/03/2022 23:16:27

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:14	
Control Test	0.080	23:14	
Air Blank	0.000	23:14	
Subject Sample	# 1	23:15	Subject Test Refused
Breath Volume	0.000	23:15	
Air Blank	0.000	23:15	
Control Test	0.079	23:15	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 09/16/2022 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEAL IV,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 09/04/2022 00:08:53

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.079	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:07	
Control Test	0.079	00:08	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: GRIER

DATE/TIME OF TEST: 09/04/2022 02:38:44

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.152	02:34
Breath Volume	1.910	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.148	02:37
Breath Volume	2.468	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: LANDERS

DATE/TIME OF TEST: 09/04/2022 03:45:17

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.172	03:40
Breath Volume	1.511	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:42
Subject Sample # 2	0.172	03:43
Breath Volume	1.695	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 09/04/2022 19:53:44

Diagnostics Check	OK	19:47
Air Blank	0.000	19:47
Control Test	0.079	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.082	19:48
Breath Volume	1.113	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.081	19:52
Breath Volume	1.855	19:52
Air Blank	0.000	19:52
Control Test	0.079	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOCKERS,B
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: LAMBETH

DATE/TIME OF TEST: 09/04/2022 23:13:28

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.079	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:12	
Control Test	0.079	23:12	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: IBO

DATE/TIME OF TEST: 09/04/2022 23:58:39

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.079	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.238	23:51
Breath Volume	1.527	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.239	23:56
Breath Volume	1.433	23:56
Air Blank	0.000	23:57
Control Test	0.078	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEMPF,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: TISDALE

DATE/TIME OF TEST: 09/05/2022 01:40:36

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.079	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:39	RFI Detect
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.079	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEMPF,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: TISDALE

DATE/TIME OF TEST: 09/05/2022 01:48:17

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:45	
Control Test	0.080	01:46	
Air Blank	0.000	01:46	
Subject Sample	# 1	01:46	Subject Test Refused
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Control Test	0.079	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DE'SOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: KIGER

DATE/TIME OF TEST: 09/05/2022 02:39:10

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.171	02:34
Breath Volume	2.695	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.170	02:37
Breath Volume	2.738	02:37
Air Blank	0.000	02:38
Control Test	0.077	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/05/2022 03:26:36

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.079	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.079	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/05/2022 22:50:34

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.079	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.121	22:46
Breath Volume	1.570	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.116	22:49
Breath Volume	1.398	22:49
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: CURRAN

DATE/TIME OF TEST: 09/06/2022 00:35:02

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.234	00:30
Breath Volume	1.800	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.234	00:33
Breath Volume	1.867	00:33
Air Blank	0.000	00:34
Control Test	0.078	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:35

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 09/06/2022 04:49:34

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.079	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:48	
Control Test	0.079	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 09/08/2022 00:56:56

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.079	00:56	
Air Blank		00:56	Sequence Aborted
Air Blank	0.000	00:56	

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 09/08/2022 01:29:16

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.251	01:24
Breath Volume	1.414	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.247	01:27
Breath Volume	1.878	01:27
Air Blank	0.000	01:28
Control Test	0.077	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 09/08/2022 03:31:35

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:29	
Control Test	0.079	03:29	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/08/2022 18:45:35

Diagnostics Check	OK	18:38
Air Blank	0.000	18:39
Control Test	0.079	18:39
Air Blank	0.000	18:40
Subject Sample # 1	0.199	18:41
Breath Volume	2.312	18:41
Air Blank	0.000	18:41
Air Blank	0.000	18:43
Subject Sample # 2	0.203	18:44
Breath Volume	2.335	18:44
Air Blank	0.000	18:44
Control Test	0.078	18:44
Air Blank	0.000	18:45
Diagnostics Check	OK	18:45

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/10/2022 04:33:03

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.115	04:28
Breath Volume	1.496	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.117	04:31
Breath Volume	2.195	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:33

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/10/2022 23:17:08

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.080	23:14	
Air Blank	0.000	23:15	
Subject Sample	# 1	23:15	Subject Test Refused
Breath Volume	0.000	23:15	
Air Blank	0.000	23:16	
Control Test	0.081	23:16	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 09/16/2022 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: PUNCH

DATE/TIME OF TEST: 09/11/2022 01:03:02

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.080	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Control Test	0.080	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VO,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/11/2022 04:21:53

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Control Test	0.079	04:21	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,P,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FERNANDEZ,W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/11/2022 11:41:48

Diagnostics Check	OK	11:35
Air Blank	0.000	11:36
Control Test	0.080	11:36
Air Blank	0.000	11:36
Subject Sample # 1	0.169	11:37
Breath Volume	2.242	11:37
Air Blank	0.000	11:37
Air Blank	0.000	11:39
Subject Sample # 2	0.173	11:40
Breath Volume	2.296	11:40
Air Blank	0.000	11:40
Control Test	0.079	11:41
Air Blank	0.000	11:41
Diagnostics Check	OK	11:41

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,P,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/11/2022 15:06:34

Diagnostics Check	OK	15:00
Air Blank	0.000	15:00
Control Test	0.079	15:01
Air Blank	0.000	15:01
Subject Sample # 1	0.201	15:02
Breath Volume	1.511	15:02
Air Blank	0.000	15:02
Air Blank	0.000	15:04
Subject Sample # 2	0.186	15:05
Breath Volume	1.464	15:05
Air Blank	0.000	15:05
Control Test	0.079	15:05
Air Blank	0.000	15:06
Diagnostics Check	OK	15:06

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: CASTRO-TORRES

DATE/TIME OF TEST: 09/12/2022 00:55:25

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:53	
Control Test	0.079	00:53	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:54	Subject Test Refused
Breath Volume	0.000	00:54	
Air Blank	0.000	00:54	
Control Test	0.080	00:54	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/12/2022 01:56:38

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:54	
Control Test	0.079	01:54	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:55	
Control Test	0.080	01:56	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 09/12/2022 19:50:17

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:47	
Control Test	0.080	19:48	
Air Blank	0.000	19:48	
Subject Sample	# 1	19:48	Subject Test Refused
Breath Volume	0.000	19:48	
Air Blank	0.000	19:49	
Control Test	0.080	19:49	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 09/13/2022 01:31:38

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.152	01:27
Breath Volume	2.250	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.150	01:30
Breath Volume	1.664	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/13/2022 19:49:24

Diagnostics Check	OK	19:46	
Air Blank	0.000	19:47	
Control Test	0.080	19:47	
Air Blank	0.000	19:47	
Subject Sample	# 1	19:48	Subject Test Refused
Breath Volume	0.000	19:48	
Air Blank	0.000	19:48	
Control Test	0.080	19:48	
Air Blank	0.000	19:49	
Diagnostics Check	OK	19:49	

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: VANKOUGHNETT

DATE/TIME OF TEST: 09/13/2022 23:11:58

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.000	23:07
Breath Volume	2.531	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.000	23:10
Breath Volume	2.902	23:10
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/15/2022 03:50:37

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.039	03:46
Breath Volume	2.210	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.042	03:49
Breath Volume	3.089	03:49
Air Blank	0.000	03:49
Control Test	0.079	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 09/16/2022 02:05:27

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.080	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.079	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: HOLZENDORF

DATE/TIME OF TEST: 09/16/2022 03:58:18

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.158	03:53
Breath Volume	2.140	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.163	03:56
Breath Volume	1.691	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 09/16/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/17/2022 01:06:58

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.080	01:04	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 09/19/2022 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MCCAUL

DATE/TIME OF TEST: 09/17/2022 04:09:05

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.094	04:04
Breath Volume	1.414	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.094	04:07
Breath Volume	1.394	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 09/19/2022 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBER,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HICKS,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:33

SUBJECT'S LAST NAME: TORRES LUGO

DATE/TIME OF TEST: 09/17/2022 09:00:28

Diagnostics Check	OK	08:54
Air Blank	0.000	08:54
Control Test	0.079	08:54
Air Blank	0.000	08:55
Subject Sample # 1	0.000	08:55
Breath Volume	2.464	08:55
Air Blank	0.000	08:56
Air Blank	0.000	08:58
Subject Sample # 2	0.000	08:59
Breath Volume	1.375	08:59
Air Blank	0.000	08:59
Control Test	0.079	08:59
Air Blank	0.000	09:00
Diagnostics Check	OK	09:00

Data Download Date/Time: 09/19/2022 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: MARTINEZ PERDOMO

DATE/TIME OF TEST: 09/17/2022 20:21:20

Diagnostics Check	OK	20:18	
Air Blank	0.000	20:18	
Control Test	0.080	20:19	
Air Blank	0.000	20:19	
Subject Sample	# 1	20:19	Subject Test Refused
Breath Volume	0.000	20:19	
Air Blank	0.000	20:20	
Control Test	0.079	20:20	
Air Blank	0.000	20:21	
Diagnostics Check	OK	20:21	

Data Download Date/Time: 09/19/2022 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: GOU

DATE/TIME OF TEST: 09/18/2022 02:26:27

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.249	02:20
Breath Volume	1.226	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.250	02:24
Breath Volume	1.410	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 09/19/2022 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 09/18/2022 03:36:26

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.226	03:31
Breath Volume	1.402	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.217	03:34
Breath Volume	1.128	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 09/19/2022 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBER,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 09/18/2022 07:22:13

Diagnostics Check	OK	07:15
Air Blank	0.000	07:16
Control Test	0.079	07:16
Air Blank	0.000	07:17
Subject Sample # 1	0.196	07:17
Breath Volume	2.703	07:17
Air Blank	0.000	07:18
Air Blank	0.000	07:19
Subject Sample # 2	0.197	07:20
Breath Volume	2.816	07:20
Air Blank	0.000	07:21
Control Test	0.078	07:21
Air Blank	0.000	07:22
Diagnostics Check	OK	07:22

Data Download Date/Time: 09/19/2022 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GLASPIE

DATE/TIME OF TEST: 09/19/2022 03:55:25

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:53	
Control Test	0.079	03:53	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:54	Subject Test Refused
Breath Volume	0.000	03:54	
Air Blank	0.000	03:54	
Control Test	0.080	03:54	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 09/19/2022 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Atlantic Beach PD
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 09/20/2022 00:07:24

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.200	00:02
Breath Volume	1.609	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:04
Subject Sample # 2	0.201	00:05
Breath Volume	1.382	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 10/13/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: NEIL

DATE/TIME OF TEST: 09/22/2022 03:37:51

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.190	03:32
Breath Volume	1.515	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.202	03:36
Breath Volume	1.960	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 10/13/2022 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Atlantic Beach PD
ARRESTING OFFICER: MCLENDON,R
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/24/2022 03:25:05

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.080	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 10/13/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: JIMENEZ GOMEZ

DATE/TIME OF TEST: 09/25/2022 22:37:41

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.081	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.136	22:32
Breath Volume	1.835	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.130	22:36
Breath Volume	2.117	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 10/13/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: PRIDE

DATE/TIME OF TEST: 09/25/2022 23:47:35

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.200	23:42
Breath Volume	1.789	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.199	23:46
Breath Volume	1.777	23:46
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 10/13/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: DRAGOMIR,I
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GOMEZ ALVAREZ

DATE/TIME OF TEST: 09/26/2022 02:29:47

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.161	02:23
Breath Volume	1.109	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.166	02:28
Breath Volume	1.421	02:28
Air Blank	0.000	02:28
Control Test	0.077	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 10/13/2022 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FANFAN,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/27/2022 02:11:19

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.080	02:09	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:10	
Control Test	0.080	02:10	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 10/13/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GOLDEN

DATE/TIME OF TEST: 09/28/2022 01:17:47

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.110	01:13
Breath Volume	1.136	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.117	01:16
Breath Volume	1.304	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 10/13/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,C,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BARNES,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: HOLLINGSWORTH

DATE/TIME OF TEST: 09/01/2022 01:30:44

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.135	01:26
Breath Volume	2.519	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.135	01:29
Breath Volume	2.304	01:29
Air Blank	0.000	01:29
Control Test	0.077	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 09/13/2022 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,C,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JACKSON,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/02/2022 01:57:18

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.078	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.193	01:52
Breath Volume	1.242	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.195	01:55
Breath Volume	2.285	01:55
Air Blank	0.000	01:56
Control Test	0.075	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 09/13/2022 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PICKETT,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BARNES,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 09/02/2022 02:44:46

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.077	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.120	02:39
Breath Volume	1.238	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample # 2	0.121	02:43
Breath Volume	1.203	02:43
Air Blank	0.000	02:43
Control Test	0.075	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 09/13/2022 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PICKETT,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SELLERS,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CONOVER

DATE/TIME OF TEST: 09/02/2022 04:19:33

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:17	
Control Test	0.076	04:17	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:18	
Control Test	0.077	04:18	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 09/13/2022 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,C,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DILL,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MERCHANT

DATE/TIME OF TEST: 09/10/2022 22:25:16

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.078	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.200	22:20
Breath Volume	2.472	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.199	22:23
Breath Volume	2.027	22:23
Air Blank	0.000	22:24
Control Test	0.075	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 09/13/2022 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES, BRANDON, T
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GORDON, G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: BLACKWELL

DATE/TIME OF TEST: 09/13/2022 01:13:45

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.077	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.162	01:08
Breath Volume	1.613	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.173	01:12
Breath Volume	1.542	01:12
Air Blank	0.000	01:12
Control Test	0.077	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 09/13/2022 16:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,BRANDON,T
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GORDON,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/18/2022 00:30:46

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.078	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.203	00:26
Breath Volume	2.015	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.190	00:29
Breath Volume	1.492	00:29
Air Blank	0.000	00:29
Control Test	0.076	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 10/24/2022 18:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: PENZA,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 09/19/2022 14:38:38

Diagnostics Check	OK	14:35	
Air Blank	0.000	14:36	
Control Test	0.078	14:36	
Air Blank	0.000	14:37	
Subject Sample	# 1	14:37	Subject Test Refused
Breath Volume	0.000	14:37	
Air Blank	0.000	14:37	
Control Test	0.078	14:38	
Air Blank	0.000	14:38	
Diagnostics Check	OK	14:38	

Data Download Date/Time: 10/24/2022 18:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JACKSON,ROBERT
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: GUARDADO

DATE/TIME OF TEST: 09/24/2022 21:46:45

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.078	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.173	21:42
Breath Volume	3.320	21:42
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.163	21:45
Breath Volume	3.046	21:45
Air Blank	0.000	21:45
Control Test	0.076	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 10/24/2022 18:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JACKSON,ROBERT
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 09/25/2022 01:59:21

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.078	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.163	01:54
Breath Volume	3.437	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.151	01:57
Breath Volume	2.156	01:57
Air Blank	0.000	01:58
Control Test	0.075	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 10/24/2022 18:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: TERRY,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: HEMPSTEAD

DATE/TIME OF TEST: 09/25/2022 03:40:36

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.078	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.147	03:35
Breath Volume	1.304	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.156	03:39
Breath Volume	1.203	03:39
Air Blank	0.000	03:39
Control Test	0.077	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 10/24/2022 18:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOBLEY, JONATHAN, A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MOBLEY, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0602108082
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RUISE

DATE/TIME OF TEST: 09/02/2022 01:56:57

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.077	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.230	01:52
Breath Volume	1.363	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.227	01:55
Breath Volume	1.257	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 09/25/2022 21:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CALEB,A
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: BODKIN,MICHAEL
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 08/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: DURHAM

DATE/TIME OF TEST: 09/03/2022 04:52:34

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.078	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.157	04:46
Breath Volume	1.515	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.154	04:50
Breath Volume	1.484	04:50
Air Blank	0.000	04:51
Control Test	0.076	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 09/25/2022 21:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,GLENN,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,GLENN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: MCRAE

DATE/TIME OF TEST: 09/17/2022 22:07:42

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.078	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.184	22:02
Breath Volume	1.421	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:04
Subject Sample # 2	0.201	22:05
Breath Volume	2.964	22:05
Air Blank	0.000	22:06
Control Test	0.077	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 10/03/2022 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,L,
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: LEWIS,L
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: TROUG

DATE/TIME OF TEST: 09/22/2022 13:58:06

Diagnostics Check	OK	13:55	
Air Blank	0.000	13:55	
Control Test	0.079	13:56	
Air Blank	0.000	13:56	
Subject Sample	# 1	13:56	Subject Test Refused
Breath Volume	0.000	13:56	
Air Blank	0.000	13:57	
Control Test	0.075	13:57	
Air Blank	0.000	13:58	
Diagnostics Check	OK	13:58	

Data Download Date/Time: 10/03/2022 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEHEAD,BRANDON,A
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: BELLAMY,MOSES
ARRESTING OFFICER AGENCY: Starke PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 09/29/2022 14:39:13

Diagnostics Check	OK	14:24	
Air Blank	0.000	14:25	
Control Test	0.078	14:25	
Air Blank	0.000	14:25	
Subject Sample	# 1 0.276	14:26	
Breath Volume	2.382	14:26	
Air Blank	0.000	14:27	
Air Blank	0.000	14:29	
Subject Sample	# 2 0.239	14:32	Volume Not Met
Breath Volume	0.089	14:32	
Air Blank	0.000	14:32	
Air Blank	0.000	14:34	
Subject Sample	# 3	14:37	No Sample Provided
Breath Volume	0.000	14:37	
Air Blank	0.000	14:38	
Control Test	0.076	14:38	
Air Blank	0.000	14:39	
Diagnostics Check	OK	14:39	

Data Download Date/Time: 10/03/2022 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: HIGHTOWER

DATE/TIME OF TEST: 09/05/2022 19:29:58

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:27	
Control Test	0.080	19:28	
Air Blank	0.000	19:28	
Subject Sample	# 1	19:28	RFI Detect
Breath Volume	0.000	19:28	
Air Blank	0.000	19:29	
Control Test	0.008	19:29	RFI Detect
Air Blank		19:29	RFI Detect

Data Download Date/Time: 09/16/2022 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: CHERWAK

DATE/TIME OF TEST: 09/08/2022 19:08:21

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:07	
Control Test	0.080	19:07	
Air Blank	0.000	19:07	
		19:08	RFI Detect
Air Blank		19:08	RFI Detect

Data Download Date/Time: 09/16/2022 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: CHERWAK

DATE/TIME OF TEST: 09/08/2022 19:26:05

Diagnostics Check	OK	19:12	
Air Blank	0.000	19:12	
Control Test	0.080	19:13	
Air Blank	0.000	19:13	
Subject Sample	# 1 0.187	19:16	Volume Not Met
Breath Volume	0.343	19:16	
Air Blank	0.000	19:17	
Air Blank	0.000	19:19	
Subject Sample	# 2 0.215	19:21	
Breath Volume	2.050	19:21	
Air Blank	0.000	19:21	
Air Blank	0.000	19:23	
Subject Sample	# 3 0.216	19:24	
Breath Volume	1.664	19:24	
Air Blank	0.000	19:25	
Control Test	0.076	19:25	
Air Blank	0.000	19:26	
Diagnostics Check	OK	19:26	

Data Download Date/Time: 09/16/2022 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHNER,KYLER,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TERHUNE,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: SEMISA

DATE/TIME OF TEST: 09/17/2022 00:23:30

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.081	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.000	00:19
Breath Volume	3.191	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.000	00:22
Breath Volume	2.929	00:22
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 10/11/2022 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, TYLER, M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JONES, TYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: OGUNJI

DATE/TIME OF TEST: 09/17/2022 04:18:31

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.081	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.154	04:13
Breath Volume	1.378	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.152	04:16
Breath Volume	2.101	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 10/11/2022 18:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: HILTON

DATE/TIME OF TEST: 09/17/2022 23:49:09

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.000	23:44
Breath Volume	3.894	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.000	23:47
Breath Volume	1.464	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 10/11/2022 18:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMPSON,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAMPSON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: HOSKINS

DATE/TIME OF TEST: 09/19/2022 20:55:54

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:51	
Control Test	0.080	20:51	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:54	Subject Test Refused
Breath Volume	0.000	20:54	
Air Blank	0.000	20:55	
Control Test	0.081	20:55	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

Data Download Date/Time: 10/11/2022 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: LEBAKKEN

DATE/TIME OF TEST: 09/20/2022 18:56:05

Diagnostics Check	OK	18:44	
Air Blank	0.000	18:45	
Control Test	0.080	18:45	
Air Blank	0.000	18:45	
Subject Sample	# 1	18:49	No Sample Provided
Breath Volume	0.000	18:49	
Air Blank	0.000	18:49	
Air Blank	0.000	18:51	
Subject Sample	# 2	18:54	No Sample Provided
Breath Volume	0.000	18:54	
Air Blank	0.000	18:55	
Control Test	0.080	18:55	
Air Blank	0.000	18:55	
Diagnostics Check	OK	18:56	

Data Download Date/Time: 10/11/2022 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TIEDEMAN,NICHOLAS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/21/2022 02:33:53

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.081	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:26	No Sample Provided
Breath Volume	0.000	02:26	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample	# 2	02:32	No Sample Provided
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Control Test	0.081	02:33	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 10/11/2022 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 09/21/2022 23:36:42

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.000	23:32
Breath Volume	3.488	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.000	23:35
Breath Volume	3.730	23:35
Air Blank	0.000	23:35
Control Test	0.081	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 10/11/2022 18:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAHL,RANDY,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAHL,RANDY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: BROXSON

DATE/TIME OF TEST: 09/22/2022 06:07:55

Diagnostics Check	OK	05:58	
Air Blank	0.000	05:59	
Control Test	0.080	05:59	
Air Blank	0.000	06:00	
Subject Sample	# 1	06:03	No Sample Provided
Breath Volume	0.000	06:03	
Air Blank	0.000	06:03	
Air Blank	0.000	06:05	
Subject Sample	# 2	06:06	Subject Test Refused
Breath Volume	0.000	06:06	
Air Blank	0.000	06:07	
Control Test	0.081	06:07	
Air Blank	0.000	06:07	
Diagnostics Check	OK	06:07	

Data Download Date/Time: 10/11/2022 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/23/2022 00:50:31

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:39	
Control Test	0.081	00:39	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:43	No Sample Provided
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Air Blank	0.000	00:46	
Subject Sample	# 2	00:49	No Sample Provided
Breath Volume	0.000	00:49	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 10/11/2022 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: SUAREZ SILVA

DATE/TIME OF TEST: 09/24/2022 03:26:15

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.154	03:21
Breath Volume	2.035	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.157	03:24
Breath Volume	1.800	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 10/11/2022 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: MEEKS-BRUNACHE

DATE/TIME OF TEST: 09/24/2022 23:21:10

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:18	
Control Test	0.080	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 1	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:20	
Control Test	0.081	23:20	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 10/11/2022 18:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: SORRELL

DATE/TIME OF TEST: 09/25/2022 05:06:03

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.081	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.000	05:01
Breath Volume	1.523	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.000	05:04
Breath Volume	1.460	05:04
Air Blank	0.000	05:05
Control Test	0.081	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:06

Data Download Date/Time: 10/11/2022 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TERHUNE,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 09/25/2022 21:51:25

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:43	
Control Test	0.079	21:43	
Air Blank	0.000	21:44	
Subject Sample	# 1	21:47	No Sample Provided
Breath Volume	0.000	21:47	
Air Blank	0.000	21:47	
Air Blank	0.000	21:49	
Subject Sample	# 2	21:50	Subject Test Refused
Breath Volume	0.000	21:50	
Air Blank	0.000	21:50	
Control Test	0.079	21:50	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 10/11/2022 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: TROTTER

DATE/TIME OF TEST: 09/29/2022 00:06:30

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.081	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.000	00:02
Breath Volume	1.894	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.000	00:05
Breath Volume	1.617	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 10/11/2022 18:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: BRINSON

DATE/TIME OF TEST: 09/30/2022 20:16:44

Diagnostics Check	OK	20:13	
Air Blank	0.000	20:14	
Control Test	0.080	20:14	
Air Blank	0.000	20:15	
Subject Sample	# 1	20:15	Subject Test Refused
Breath Volume	0.000	20:15	
Air Blank	0.000	20:15	
Control Test	0.080	20:16	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:16	

Data Download Date/Time: 10/11/2022 18:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TIEDEMANN,NICHOLAS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 09/01/2022 03:28:31

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.000	03:23
Breath Volume	3.425	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.000	03:27
Breath Volume	2.273	03:27
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 09/16/2022 20:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROBINS II,JERRY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: BUTLER-ANDERSON

DATE/TIME OF TEST: 09/01/2022 05:57:56

Diagnostics Check	OK	05:46	
Air Blank	0.000	05:47	
Control Test	0.082	05:47	
Air Blank	0.000	05:47	
Subject Sample	# 1	05:50	No Sample Provided
Breath Volume	0.000	05:50	
Air Blank	0.000	05:51	
Air Blank	0.000	05:53	
Subject Sample	# 2	05:56	No Sample Provided
Breath Volume	0.000	05:56	
Air Blank	0.000	05:57	
Control Test	0.082	05:57	
Air Blank	0.000	05:57	
Diagnostics Check	OK	05:57	

Data Download Date/Time: 09/16/2022 20:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 09/02/2022 02:22:09

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:19	
Control Test	0.082	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 09/16/2022 20:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 09/03/2022 05:22:34

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.082	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.132	05:18
Breath Volume	3.667	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.129	05:21
Breath Volume	3.921	05:21
Air Blank	0.000	05:21
Control Test	0.082	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/03/2022 05:37:59

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.082	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.142	05:33
Breath Volume	3.675	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:35
Subject Sample # 2	0.145	05:36
Breath Volume	1.949	05:36
Air Blank	0.000	05:37
Control Test	0.082	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/03/2022 22:53:54

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.082	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.079	22:49
Breath Volume	2.339	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.066	22:52
Breath Volume	2.203	22:52
Air Blank	0.000	22:53
Control Test	0.082	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 12:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 09/04/2022 00:28:27

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.082	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.000	00:23
Breath Volume	3.574	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.000	00:27
Breath Volume	4.488	00:27
Air Blank	0.000	00:27
Control Test	0.082	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CUEVAS

DATE/TIME OF TEST: 09/04/2022 01:42:12

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.082	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.154	01:37
Breath Volume	1.742	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.161	01:40
Breath Volume	1.183	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BICKNELL,BRANDON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 09/05/2022 00:35:30

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:33	
Control Test	0.082	00:33	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:34	Subject Test Refused
Breath Volume	0.000	00:34	
Air Blank	0.000	00:34	
Control Test	0.081	00:34	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: HIGHTOWER

DATE/TIME OF TEST: 09/05/2022 19:19:18

Diagnostics Check	OK	19:16	
Air Blank	0.000	19:17	
Control Test	0.082	19:17	
Air Blank	0.000	19:18	
Subject Sample	# 1	19:18	RFI Detect
Breath Volume	0.000	19:18	
Air Blank		19:18	RFI Detect
Air Blank	0.000	19:19	

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: HIGHTOWER

DATE/TIME OF TEST: 09/05/2022 19:23:01

Diagnostics Check	OK	19:21	
Air Blank	0.000	19:21	
Control Test	0.082	19:21	
Air Blank	0.000	19:22	
Subject Sample	# 1	19:22	RFI Detect
Breath Volume	0.000	19:22	
Air Blank		19:22	RFI Detect
Air Blank		19:23	RFI Detect

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: HIGHTOWER

DATE/TIME OF TEST: 09/05/2022 19:42:03

Diagnostics Check	OK	19:32
Air Blank	0.000	19:33
Control Test	0.082	19:33
Air Blank	0.000	19:34
Subject Sample # 1	0.281	19:34
Breath Volume	1.917	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.251	19:37
Breath Volume	1.433	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 3	0.283	19:40
Breath Volume	1.664	19:40
Air Blank	0.000	19:41
Control Test	0.081	19:41
Air Blank	0.000	19:41
Diagnostics Check	OK	19:42

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KELLY,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MATEIKAT

DATE/TIME OF TEST: 09/05/2022 19:59:46

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.081	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.000	19:55
Breath Volume	2.480	19:55
Air Blank	0.000	19:55
Air Blank	0.000	19:57
Subject Sample # 2	0.000	19:58
Breath Volume	1.804	19:58
Air Blank	0.000	19:58
Control Test	0.082	19:59
Air Blank	0.000	19:59
Diagnostics Check	OK	19:59

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DUFRENE,DUSTIN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: FOUST

DATE/TIME OF TEST: 09/05/2022 21:01:45

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:50	
Control Test	0.082	20:51	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:54	No Sample Provided
Breath Volume	0.000	20:54	
Air Blank	0.000	20:55	
Air Blank	0.000	20:57	
Subject Sample	# 2	21:00	No Sample Provided
Breath Volume	0.000	21:00	
Air Blank	0.000	21:00	
Control Test	0.082	21:01	
Air Blank	0.000	21:01	
Diagnostics Check	OK	21:01	

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARTY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: VAN SICKLE

DATE/TIME OF TEST: 09/08/2022 03:40:18

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.082	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.195	03:35
Breath Volume	5.246	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.193	03:38
Breath Volume	4.464	03:38
Air Blank	0.000	03:39
Control Test	0.081	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: CHERWAK

DATE/TIME OF TEST: 09/08/2022 18:48:24

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:46	
Control Test	0.082	18:47	
Air Blank	0.000	18:47	
Subject Sample	# 1	18:48	RFI Detect
Breath Volume	0.000	18:48	
Air Blank		18:48	RFI Detect
Air Blank		18:48	RFI Detect

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: CHERWAK

DATE/TIME OF TEST: 09/08/2022 18:52:31

Diagnostics Check	OK	18:50	
Air Blank	0.000	18:50	
Control Test	0.082	18:51	
Air Blank	0.000	18:51	
Subject Sample	# 1	18:51	RFI Detect
Breath Volume	0.000	18:51	
Air Blank		18:52	RFI Detect
Air Blank	0.000	18:52	

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: CHERWAK

DATE/TIME OF TEST: 09/08/2022 18:55:42

Diagnostics Check	OK	18:55	
Air Blank	0.000	18:55	
		18:55	RFI Detect
Air Blank		18:55	RFI Detect

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: CHERWAK

DATE/TIME OF TEST: 09/08/2022 18:58:00

Diagnostics Check	OK	18:57	
Air Blank		18:57	RFI Detect
Air Blank	0.000	18:58	

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: CHERWAK

DATE/TIME OF TEST: 09/08/2022 19:04:07

Diagnostics Check	OK	19:03	
Air Blank		19:03	RFI Detect
Air Blank	0.000	19:04	

Data Download Date/Time: 09/16/2022 20:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TIEDEMANN,NICHOLAS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 09/09/2022 00:47:08

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.082	00:43	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:46	
Control Test	0.082	00:46	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: ZINK

DATE/TIME OF TEST: 09/09/2022 04:33:34

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:30	
Control Test	0.083	04:30	
Air Blank	0.000	04:30	
Subject Sample	# 1	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Control Test	0.082	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELENZ,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ELENZ,TIMOTHY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/10/2022 01:08:45

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.210	01:04
Breath Volume	2.964	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 2	0.210	01:07
Breath Volume	2.675	01:07
Air Blank	0.000	01:07
Control Test	0.081	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: GARLETTS

DATE/TIME OF TEST: 09/10/2022 01:24:51

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.074	01:20
Breath Volume	2.933	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample # 2	0.073	01:23
Breath Volume	3.003	01:23
Air Blank	0.000	01:23
Control Test	0.081	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 09/10/2022 02:45:19

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.081	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.114	02:40
Breath Volume	1.816	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.102	02:43
Breath Volume	1.332	02:43
Air Blank	0.000	02:44
Control Test	0.081	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BANKICH

DATE/TIME OF TEST: 09/11/2022 00:27:14

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.082	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.000	00:22
Breath Volume	4.183	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.000	00:25
Breath Volume	3.753	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: GRIMES

DATE/TIME OF TEST: 09/11/2022 05:35:12

Diagnostics Check	OK	05:30	
Air Blank	0.000	05:31	
Control Test	0.082	05:31	
Air Blank	0.000	05:31	
Subject Sample	# 1	05:33	Subject Test Refused
Breath Volume	0.000	05:33	
Air Blank	0.000	05:34	
Control Test	0.082	05:34	
Air Blank	0.000	05:35	
Diagnostics Check	OK	05:35	

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ADAMS,BRIANA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: INESON

DATE/TIME OF TEST: 09/11/2022 21:53:46

Diagnostics Check	OK	21:47
Air Blank	0.000	21:48
Control Test	0.081	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.309	21:49
Breath Volume	1.308	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:51
Subject Sample # 2	0.290	21:52
Breath Volume	1.640	21:52
Air Blank	0.000	21:52
Control Test	0.081	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: PACKER

DATE/TIME OF TEST: 09/12/2022 03:39:25

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.034	03:34
Breath Volume	1.410	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.036	03:37
Breath Volume	1.972	03:37
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHNER,KYLER,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CASTELLANETA,MARC
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 09/13/2022 01:01:49

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.083	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Control Test	0.082	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JONES,TYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: REA

DATE/TIME OF TEST: 09/13/2022 03:22:16

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.082	03:20	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.082	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 09/15/2022 04:09:13

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.083	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Control Test	0.082	04:08	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MCKEEN

DATE/TIME OF TEST: 09/15/2022 05:08:34

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:05	
Control Test	0.083	05:05	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:07	Subject Test Refused
Breath Volume	0.000	05:07	
Air Blank	0.000	05:07	
Control Test	0.080	05:07	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MCKEEN

DATE/TIME OF TEST: 09/15/2022 05:17:39

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.082	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.220	05:12
Breath Volume	2.636	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.219	05:16
Breath Volume	1.738	05:16
Air Blank	0.000	05:16
Control Test	0.082	05:17
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 09/16/2022 20:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/15/2022 23:57:45

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.082	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.215	23:53
Breath Volume	2.171	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.207	23:56
Breath Volume	1.527	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 09/16/2022 20:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TIEDEMANN,NICHOLAS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: KNEEN

DATE/TIME OF TEST: 09/16/2022 04:52:30

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.081	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.187	04:47
Breath Volume	3.355	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.185	04:51
Breath Volume	2.964	04:51
Air Blank	0.000	04:51
Control Test	0.081	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 09/16/2022 20:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MILLER,TIM
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 09/02/2022 03:15:25

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.082	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.180	03:10
Breath Volume	3.097	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.175	03:13
Breath Volume	3.101	03:13
Air Blank	0.000	03:14
Control Test	0.082	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 09/28/2022 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,BRANDON,L
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: ARROYO,BRANDON
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GILLEY

DATE/TIME OF TEST: 09/17/2022 20:41:47

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.082	20:31
Air Blank	0.000	20:31
Subject Sample # 1	0.221	20:32
Breath Volume	2.640	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.194	20:37
Breath Volume	1.261	20:37
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 3	0.190	20:40
Breath Volume	1.679	20:40
Air Blank	0.000	20:40
Control Test	0.081	20:41
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 09/28/2022 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN, WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LARA OSORIO

DATE/TIME OF TEST: 09/04/2022 03:35:33

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.081	03:31	
Air Blank	0.000	03:31	
Subject Sample # 1	0.123	03:31	
Breath Volume	1.417	03:31	
Air Blank	0.000	03:32	
Air Blank		03:34	RFI Detect
Air Blank	0.000	03:34	
Control Test	0.082	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 09/28/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN, WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LARA OSORIO

DATE/TIME OF TEST: 09/04/2022 03:46:08

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.120	03:41
Breath Volume	1.355	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.126	03:44
Breath Volume	1.734	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 09/28/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,FORREST,H
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALLEN,FORREST
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: BUNDY 3

DATE/TIME OF TEST: 09/11/2022 21:14:14

Diagnostics Check	OK	21:07
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.303	21:09
Breath Volume	3.492	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:11
Subject Sample # 2	0.305	21:12
Breath Volume	2.484	21:12
Air Blank	0.000	21:13
Control Test	0.080	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 09/28/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: AGUILERA

DATE/TIME OF TEST: 09/17/2022 22:24:42

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.195	22:19
Breath Volume	2.117	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.210	22:23
Breath Volume	2.066	22:23
Air Blank	0.000	22:23
Control Test	0.080	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 09/28/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: KIGHT,BRANDON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 09/18/2022 05:42:09

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.081	05:35
Air Blank	0.000	05:36
Subject Sample # 1	0.165	05:37
Breath Volume	1.648	05:37
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.160	05:40
Breath Volume	1.523	05:40
Air Blank	0.000	05:41
Control Test	0.079	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 09/28/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, CALE, V
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: WALSH, DAKOTA
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 09/20/2022 01:56:06

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.082	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.000	01:51
Breath Volume	1.691	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.000	01:54
Breath Volume	2.222	01:54
Air Blank	0.000	01:55
Control Test	0.082	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 09/28/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FURMAN,FREDDIE,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: YOUNG,THOMAS
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: SALAS

DATE/TIME OF TEST: 09/20/2022 17:16:53

Diagnostics Check	OK	17:01	
Air Blank	0.000	17:01	
Control Test	0.081	17:01	
Air Blank	0.000	17:02	
Subject Sample	# 1 0.203	17:05	Volume Not Met
Breath Volume	0.667	17:05	
Air Blank	0.000	17:06	
Air Blank	0.000	17:07	
Subject Sample	# 2 0.216	17:11	
Breath Volume	1.816	17:11	
Air Blank	0.000	17:12	
Air Blank	0.000	17:13	
Subject Sample	# 3 0.214	17:14	
Breath Volume	1.359	17:14	
Air Blank	0.000	17:15	
Control Test	0.077	17:16	
Air Blank	0.000	17:16	
Diagnostics Check	OK	17:16	

Data Download Date/Time: 09/28/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: RAHN,BRENNEEN
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 09/24/2022 02:24:54

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:24	
Control Test	0.079	02:24	RFI Detect
Air Blank		02:24	RFI Detect

Data Download Date/Time: 09/28/2022 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BRENNEN,RAHN
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 09/24/2022 02:29:55

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.081	02:27	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:29	
Control Test	0.080	02:29	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 09/28/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, CALE, V
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SMITH, CALE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: LOTT

DATE/TIME OF TEST: 09/29/2022 02:12:17

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.082	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.154	02:07
Breath Volume	3.050	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.154	02:10
Breath Volume	1.914	02:10
Air Blank	0.000	02:11
Control Test	0.082	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 10/24/2022 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,ERIN,M
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MARTIN,ERIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: MALTAIS

DATE/TIME OF TEST: 09/03/2022 22:54:24

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.135	22:49
Breath Volume	1.433	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:51
Subject Sample # 2	0.127	22:52
Breath Volume	1.105	22:52
Air Blank	0.000	22:53
Control Test	0.078	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 09/28/2022 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: GREER

DATE/TIME OF TEST: 09/18/2022 21:18:45

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:13
Subject Sample # 1	0.161	21:14
Breath Volume	1.574	21:14
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.159	21:17
Breath Volume	1.996	21:17
Air Blank	0.000	21:17
Control Test	0.079	21:18
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 09/28/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAENZ,MELISSA,P
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SMITH,BRENDEN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 12:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ADMAS

DATE/TIME OF TEST: 09/06/2022 00:52:49

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.077	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.184	00:48
Breath Volume	2.984	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.177	00:51
Breath Volume	3.398	00:51
Air Blank	0.000	00:51
Control Test	0.078	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 09/29/2022 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAENZ,MELISSA,P
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: FULFORD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: HOWEN

DATE/TIME OF TEST: 09/23/2022 22:16:10

Diagnostics Check	OK	22:13	
Air Blank	0.000	22:13	
Control Test	0.077	22:13	
Air Blank	0.000	22:14	
Subject Sample	# 1	22:14	Subject Test Refused
Breath Volume	0.000	22:14	
Air Blank	0.000	22:15	
Control Test	0.077	22:15	
Air Blank	0.000	22:16	
Diagnostics Check	OK	22:16	

Data Download Date/Time: 09/29/2022 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: CLINE

DATE/TIME OF TEST: 09/03/2022 01:00:10

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.081	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	00:58	Subject Test Refused
Breath Volume	0.000	00:58	
Air Blank	0.000	00:59	
Control Test	0.081	00:59	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 09/30/2022 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARZA,JESSE,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: LYNN,KELVIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 09/04/2022 08:56:55

Diagnostics Check	OK	08:50
Air Blank	0.000	08:50
Control Test	0.081	08:51
Air Blank	0.000	08:51
Subject Sample # 1	0.130	08:52
Breath Volume	3.335	08:52
Air Blank	0.000	08:52
Air Blank	0.000	08:54
Subject Sample # 2	0.130	08:55
Breath Volume	3.855	08:55
Air Blank	0.000	08:56
Control Test	0.080	08:56
Air Blank	0.000	08:56
Diagnostics Check	OK	08:56

Data Download Date/Time: 09/30/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,JANISIA,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: STAPLETON,JACOB
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 09/08/2022 20:07:35

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.080	20:02
Air Blank	0.000	20:02
Subject Sample # 1	0.257	20:03
Breath Volume	2.832	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.256	20:06
Breath Volume	3.527	20:06
Air Blank	0.000	20:06
Control Test	0.079	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 09/30/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAFFA,AMELIA,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BAFFA,AMELIA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 09/21/2022 19:34:46

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.080	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.156	19:30
Breath Volume	1.976	19:30
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.161	19:33
Breath Volume	2.214	19:33
Air Blank	0.000	19:33
Control Test	0.079	19:34
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 09/30/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEFFERSON,CALIDIA,
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: KAISER,TONY
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 09/25/2022 02:43:32

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.081	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:42	RFI Detect
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 09/30/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: CHUN CUZ

DATE/TIME OF TEST: 09/25/2022 14:51:42

Diagnostics Check	OK	14:45
Air Blank	0.000	14:46
Control Test	0.080	14:46
Air Blank	0.000	14:46
Subject Sample # 1	0.258	14:47
Breath Volume	2.480	14:47
Air Blank	0.000	14:47
Air Blank	0.000	14:49
Subject Sample # 2	0.250	14:50
Breath Volume	2.507	14:50
Air Blank	0.000	14:50
Control Test	0.080	14:51
Air Blank	0.000	14:51
Diagnostics Check	OK	14:51

Data Download Date/Time: 09/30/2022 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: STEIN

DATE/TIME OF TEST: 09/01/2022 16:39:23

Diagnostics Check	OK	16:31
Air Blank	0.000	16:32
Control Test	0.080	16:32
Air Blank	0.000	16:33
Subject Sample # 1	0.000	16:34
Breath Volume	1.707	16:34
Air Blank	0.000	16:35
Air Blank	0.000	16:37
Subject Sample # 2	0.000	16:38
Breath Volume	1.355	16:38
Air Blank	0.000	16:38
Control Test	0.080	16:38
Air Blank	0.000	16:39
Diagnostics Check	OK	16:39

Data Download Date/Time: 09/27/2022 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: CERRATO

DATE/TIME OF TEST: 09/03/2022 00:34:45

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:32	
Control Test	0.080	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:33	
Control Test	0.080	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 09/27/2022 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 13:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: FLORES BLANCO

DATE/TIME OF TEST: 09/03/2022 14:19:46

Diagnostics Check	OK	14:16	
Air Blank	0.000	14:16	
Control Test	0.079	14:16	
Air Blank	0.000	14:17	
Subject Sample	# 1	14:18	RFI Detect
Breath Volume	0.054	14:18	
Air Blank	0.000	14:18	
Control Test	0.079	14:19	
Air Blank	0.000	14:19	
Diagnostics Check	OK	14:19	

Data Download Date/Time: 09/27/2022 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 13:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: FLORES BLANCO

DATE/TIME OF TEST: 09/03/2022 14:30:43

Diagnostics Check	OK	14:22
Air Blank	0.000	14:22
Control Test	0.079	14:23
Air Blank	0.000	14:23
Subject Sample # 1	0.238	14:24
Breath Volume	1.917	14:24
Air Blank	0.000	14:25
Air Blank	0.000	14:27
Subject Sample # 2	0.227	14:29
Breath Volume	1.351	14:29
Air Blank	0.000	14:29
Control Test	0.079	14:30
Air Blank	0.000	14:30
Diagnostics Check	OK	14:30

Data Download Date/Time: 09/27/2022 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: DILL

DATE/TIME OF TEST: 09/03/2022 22:35:51

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:33	
Control Test	0.079	22:33	
Air Blank	0.000	22:34	
Subject Sample	# 1	22:34	Subject Test Refused
Breath Volume	0.000	22:34	
Air Blank	0.000	22:34	
Control Test	0.079	22:35	
Air Blank	0.000	22:35	
Diagnostics Check	OK	22:35	

Data Download Date/Time: 09/27/2022 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS, MARK, K
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL, MISSY
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 09/04/2022 04:17:07

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.082	04:12
Breath Volume	1.652	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.082	04:15
Breath Volume	2.914	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 09/27/2022 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS, MARK, K
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: CASIAN, S
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MCCORD

DATE/TIME OF TEST: 09/08/2022 22:03:13

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.121	21:58
Breath Volume	2.527	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.118	22:01
Breath Volume	1.550	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 09/27/2022 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESCHAMPS,KATELYN,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: CASIAN,SKYLAR
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: WANDER

DATE/TIME OF TEST: 09/08/2022 23:46:00

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.079	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:45	
Control Test	0.079	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 09/27/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 09:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:29

SUBJECT'S LAST NAME: ABRUZZI

DATE/TIME OF TEST: 09/09/2022 10:55:25

Diagnostics Check	OK	10:49
Air Blank	0.000	10:49
Control Test	0.080	10:50
Air Blank	0.000	10:50
Subject Sample # 1	0.000	10:51
Breath Volume	2.953	10:51
Air Blank	0.000	10:51
Air Blank	0.000	10:53
Subject Sample # 2	0.000	10:54
Breath Volume	3.179	10:54
Air Blank	0.000	10:54
Control Test	0.079	10:54
Air Blank	0.000	10:55
Diagnostics Check	OK	10:55

Data Download Date/Time: 09/27/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 09/10/2022 01:19:43

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.080	01:17	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:18	Subject Test Refused
Breath Volume	0.000	01:18	
Air Blank	0.000	01:18	
Control Test	0.080	01:19	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 09/27/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 14:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: ERWIN

DATE/TIME OF TEST: 09/10/2022 15:05:21

Diagnostics Check	OK	14:56
Air Blank	0.000	14:56
Control Test	0.080	14:57
Air Blank	0.000	14:57
Subject Sample # 1	0.143	15:00
Breath Volume	2.125	15:00
Air Blank	0.000	15:00
Air Blank	0.000	15:02
Subject Sample # 2	0.124	15:03
Breath Volume	1.453	15:03
Air Blank	0.000	15:04
Control Test	0.080	15:04
Air Blank	0.000	15:05
Diagnostics Check	OK	15:05

Data Download Date/Time: 09/27/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/10/2022 22:52:22

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:48	
Control Test	0.079	22:49	
Air Blank	0.000	22:49	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.079	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 09/27/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: GUTIERREZ PINEDA

DATE/TIME OF TEST: 09/11/2022 00:11:37

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.173	00:07
Breath Volume	1.902	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.191	00:10
Breath Volume	1.640	00:10
Air Blank	0.000	00:10
Control Test	0.079	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 09/27/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 12:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:48

SUBJECT'S LAST NAME: LUNSFORD

DATE/TIME OF TEST: 09/11/2022 13:15:20

Diagnostics Check	OK	13:08
Air Blank	0.000	13:09
Control Test	0.080	13:09
Air Blank	0.000	13:10
Subject Sample # 1	0.130	13:10
Breath Volume	1.222	13:10
Air Blank	0.000	13:11
Air Blank	0.000	13:13
Subject Sample # 2	0.136	13:13
Breath Volume	2.101	13:13
Air Blank	0.000	13:14
Control Test	0.079	13:14
Air Blank	0.000	13:15
Diagnostics Check	OK	13:15

Data Download Date/Time: 09/27/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: GWINN

DATE/TIME OF TEST: 09/15/2022 00:54:28

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.080	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.000	00:49
Breath Volume	2.152	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.000	00:53
Breath Volume	2.472	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 09/27/2022 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DAVID,E
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: CASIAN,SKYLAR
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: KLUEGEL

DATE/TIME OF TEST: 09/16/2022 04:31:19

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:28	
Control Test	0.079	04:28	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:30	
Control Test	0.080	04:30	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 09/27/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 09/19/2022 21:49:06

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.080	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.000	21:44
Breath Volume	3.546	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.000	21:47
Breath Volume	3.953	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 09/27/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 09/23/2022 23:33:17

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.202	23:28
Breath Volume	4.035	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.202	23:31
Breath Volume	2.707	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 09/27/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: SANTIBANEZ GONZALEZ

DATE/TIME OF TEST: 09/24/2022 02:55:10

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.080	02:44	
Air Blank	0.000	02:44	
Subject Sample	# 1 0.256	02:47	Volume Not Met
Breath Volume	0.226	02:47	
Air Blank	0.000	02:48	
Air Blank	0.000	02:50	
Subject Sample	# 2 0.251	02:53	Volume Not Met
Breath Volume	0.640	02:53	
Air Blank	0.000	02:54	
Control Test	0.079	02:54	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 09/27/2022 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: ANDRES

DATE/TIME OF TEST: 09/25/2022 00:17:34

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.000	00:13
Breath Volume	1.976	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.000	00:16
Breath Volume	2.476	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 09/27/2022 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: DOVAL,RAMIRO
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/25/2022 01:51:33

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.252	01:47
Breath Volume	2.644	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.245	01:50
Breath Volume	2.277	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 09/27/2022 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SANTIBANEZ

DATE/TIME OF TEST: 09/25/2022 02:05:22

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.157	02:00
Breath Volume	1.824	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.157	02:03
Breath Volume	1.914	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 09/27/2022 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BOBGNA,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 12:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:04

SUBJECT'S LAST NAME: OOMS

DATE/TIME OF TEST: 09/01/2022 13:31:43

Diagnostics Check	OK	13:24
Air Blank	0.000	13:24
Control Test	0.079	13:25
Air Blank	0.000	13:25
Subject Sample # 1	0.000	13:27
Breath Volume	1.382	13:27
Air Blank	0.000	13:27
Air Blank	0.000	13:29
Subject Sample # 2	0.000	13:30
Breath Volume	2.714	13:30
Air Blank	0.000	13:30
Control Test	0.079	13:31
Air Blank	0.000	13:31
Diagnostics Check	OK	13:31

Data Download Date/Time: 09/23/2022 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCROGGIN,JOHN,C
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SCROGGIN,JOHN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: NOAKES

DATE/TIME OF TEST: 09/02/2022 00:00:19

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.138	23:55
Breath Volume	1.777	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.142	23:58
Breath Volume	2.609	23:58
Air Blank	0.000	23:59
Control Test	0.077	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 09/23/2022 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE, B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: GENNETT

DATE/TIME OF TEST: 09/02/2022 02:04:50

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.078	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.155	02:00
Breath Volume	1.992	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.158	02:03
Breath Volume	2.015	02:03
Air Blank	0.000	02:03
Control Test	0.079	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 09/23/2022 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCROGGIN,JOHN,C
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SCROGGIN,JOHN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CONCEPCION

DATE/TIME OF TEST: 09/02/2022 22:39:25

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
Control Test	0.079	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:35	No Sample Provided
Breath Volume	0.000	22:35	
Air Blank	0.000	22:35	
Air Blank	0.000	22:37	
Subject Sample	# 2	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Control Test	0.079	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 09/23/2022 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CPL.,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROLLANDO,OFC
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/03/2022 01:08:15

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:06	Subject Test Refused
Breath Volume	0.000	01:06	
Air Blank	0.000	01:07	
Control Test	0.078	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 09/23/2022 08:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SZOTT,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: JACOVIELLO

DATE/TIME OF TEST: 09/03/2022 01:45:50

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.078	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Control Test	0.078	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 09/23/2022 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUCIER,M,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SAUCIER,M
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 09/03/2022 02:18:25

Diagnostics Check	OK	02:17	
Air Blank		02:17	RFI Detect
Air Blank	0.000	02:18	

Data Download Date/Time: 09/23/2022 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUCIER,M,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SAUCIER,M
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 09/03/2022 02:28:31

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.172	02:24
Breath Volume	2.320	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.182	02:26
Breath Volume	1.996	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCROGGIN,JOHN,C
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SCROGGIN,JOHN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RUSS COMEGYS

DATE/TIME OF TEST: 09/03/2022 03:37:33

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.160	03:32
Breath Volume	2.929	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.155	03:35
Breath Volume	3.160	03:35
Air Blank	0.000	03:36
Control Test	0.077	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCROGGIN,JOHN,C
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: WICKERT,PAUL
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 09/03/2022 04:43:17

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:40	
Control Test	0.078	04:40	
Air Blank	0.000	04:41	
Subject Sample	# 1	04:41	RFI Detect
Breath Volume	0.000	04:41	
Air Blank	0.000	04:42	
Control Test	0.078	04:42	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCROGGIN,JOHN,C
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: WICKET,PAUL
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 09/03/2022 04:51:54

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.078	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.055	04:47
Breath Volume	1.433	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.052	04:50
Breath Volume	1.128	04:50
Air Blank	0.000	04:51
Control Test	0.078	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAFF,P,S
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CORTES-GUZMAN,JESUS
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: RAMIREZ-GONZALEZ

DATE/TIME OF TEST: 09/03/2022 22:27:47

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.079	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.130	22:23
Breath Volume	1.324	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample # 2	0.137	22:26
Breath Volume	1.843	22:26
Air Blank	0.000	22:26
Control Test	0.078	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAFF,P,S
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CAUDILL

DATE/TIME OF TEST: 09/03/2022 23:20:30

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:17	
Control Test	0.079	23:17	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.078	23:19	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAFF,P,S
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROLLANDO,T
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: LARA HERNANDEZ

DATE/TIME OF TEST: 09/04/2022 00:43:48

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.145	00:39
Breath Volume	2.242	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.140	00:42
Breath Volume	1.820	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FINNEGAN,Z
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 09/04/2022 01:30:20

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.000	01:26
Breath Volume	2.855	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.000	01:28
Breath Volume	2.957	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CPL.,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEACH,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CUDILO

DATE/TIME OF TEST: 09/04/2022 01:46:32

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.017	01:41
Breath Volume	1.226	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.019	01:45
Breath Volume	2.984	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCROGGIN,JOHN,C
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SCROGGIN,JOHN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 09/04/2022 02:15:12

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.273	02:10
Breath Volume	2.222	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 2	0.266	02:13
Breath Volume	2.035	02:13
Air Blank	0.000	02:14
Control Test	0.076	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAFF,P,S
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ALFORD

DATE/TIME OF TEST: 09/04/2022 03:54:25

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.079	03:52	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.079	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LANG,JEFF
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 11:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: FORTE

DATE/TIME OF TEST: 09/04/2022 12:13:38

Diagnostics Check	OK	12:07
Air Blank	0.000	12:07
Control Test	0.080	12:08
Air Blank	0.000	12:08
Subject Sample # 1	0.000	12:09
Breath Volume	1.988	12:09
Air Blank	0.000	12:09
Air Blank	0.000	12:11
Subject Sample # 2	0.000	12:12
Breath Volume	2.480	12:12
Air Blank	0.000	12:12
Control Test	0.079	12:13
Air Blank	0.000	12:13
Diagnostics Check	OK	12:13

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,NICOLE,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SHEPPARD,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/04/2022 21:13:25

Diagnostics Check	OK	21:01	
Air Blank	0.000	21:02	
Control Test	0.080	21:02	
Air Blank	0.000	21:03	
Subject Sample	# 1 0.197	21:06	Volume Not Met
Breath Volume	0.199	21:06	
Air Blank	0.000	21:06	
Air Blank	0.000	21:08	
Subject Sample	# 2	21:11	Subject Test Refused
Breath Volume	0.035	21:11	
Air Blank	0.000	21:12	
Control Test	0.077	21:12	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYALA,KASSANDRA,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: WICKERT,PAUL
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: CARMAN

DATE/TIME OF TEST: 09/05/2022 00:53:24

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.276	00:48
Breath Volume	2.750	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.274	00:51
Breath Volume	3.070	00:51
Air Blank	0.000	00:52
Control Test	0.078	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROLLANDO, T
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: STRUNK

DATE/TIME OF TEST: 09/07/2022 03:57:57

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.113	03:53
Breath Volume	1.695	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.112	03:56
Breath Volume	1.285	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: AYERS,P
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 09/07/2022 22:48:14

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.079	22:46	
Air Blank	0.000	22:46	
Subject Sample	# 1	22:46	Subject Test Refused
Breath Volume	0.000	22:46	
Air Blank	0.000	22:47	
Control Test	0.079	22:47	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA, ADAEL,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ACOSTA, ADAEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/08/2022 05:26:35

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.079	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.000	05:22
Breath Volume	1.492	05:22
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.000	05:25
Breath Volume	1.894	05:25
Air Blank	0.000	05:25
Control Test	0.078	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEACH,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: RABBIT

DATE/TIME OF TEST: 09/09/2022 00:22:40

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:20	
Control Test	0.078	00:20	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.078	00:22	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BOLOGNA,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 09/09/2022 14:50:26

Diagnostics Check	OK	14:42
Air Blank	0.000	14:42
Control Test	0.078	14:43
Air Blank	0.000	14:43
Subject Sample # 1	0.334	14:44
Breath Volume	1.554	14:44
Air Blank	0.000	14:44
Air Blank	0.000	14:46
Subject Sample # 2	0.333	14:48
Breath Volume	2.003	14:48
Air Blank	0.000	14:49
Control Test	0.078	14:49
Air Blank	0.000	14:50
Diagnostics Check	OK	14:50

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RRODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EVANS,HUNTER
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/09/2022 20:30:08

Diagnostics Check	OK	20:23
Air Blank	0.000	20:24
Control Test	0.078	20:24
Air Blank	0.000	20:24
Subject Sample # 1	0.196	20:25
Breath Volume	1.738	20:25
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.189	20:28
Breath Volume	1.554	20:28
Air Blank	0.000	20:29
Control Test	0.077	20:29
Air Blank	0.000	20:30
Diagnostics Check	OK	20:30

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,DUSTIN,B
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: LIGHTSEY,DUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: KLUCINEC

DATE/TIME OF TEST: 09/09/2022 21:22:22

Diagnostics Check	OK	21:21	
Air Blank		21:21	RFI Detect
Air Blank	0.000	21:22	

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,DUSTIN,B
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: LIGHTSEY,DUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: K425499503910

DATE/TIME OF TEST: 09/09/2022 21:30:48

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.079	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.028	21:26
Breath Volume	1.210	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.025	21:29
Breath Volume	1.117	21:29
Air Blank	0.000	21:29
Control Test	0.078	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HENGST,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 09/10/2022 16:09:23

Diagnostics Check	OK	16:05	
Air Blank	0.000	16:06	
Control Test	0.080	16:06	
Air Blank	0.000	16:07	
Subject Sample	# 1	16:08	Subject Test Refused
Breath Volume	0.000	16:08	
Air Blank	0.000	16:08	
Control Test	0.079	16:08	
Air Blank	0.000	16:09	
Diagnostics Check	OK	16:09	

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HERNANDEZ, S
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: GOLDEN

DATE/TIME OF TEST: 09/11/2022 02:17:43

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.102	02:13
Breath Volume	2.261	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.099	02:16
Breath Volume	1.425	02:16
Air Blank	0.000	02:16
Control Test	0.078	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SELDES, Z
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: ALVES

DATE/TIME OF TEST: 09/11/2022 21:00:15

Diagnostics Check	OK	20:57	
Air Blank	0.000	20:57	
Control Test	0.079	20:58	
Air Blank	0.000	20:58	
Subject Sample	# 1	20:58	Subject Test Refused
Breath Volume	0.000	20:58	
Air Blank	0.000	20:59	
Control Test	0.078	20:59	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:00	

Data Download Date/Time: 09/23/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUCIER,M,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SAUCIER,M
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: CUEVAS

DATE/TIME OF TEST: 09/12/2022 02:35:39

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.079	02:28	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:31	No Sample Provided
Breath Volume	0.000	02:31	
Air Blank	0.000	02:32	
Air Blank	0.000	02:34	
Subject Sample	# 2	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.079	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 09/23/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE, D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: DABLOW

DATE/TIME OF TEST: 09/15/2022 23:13:03

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.326	23:03
Breath Volume	1.273	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.305	23:07
Breath Volume	1.207	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 3	0.315	23:11
Breath Volume	1.152	23:11
Air Blank	0.000	23:12
Control Test	0.076	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:13

Data Download Date/Time: 09/23/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTMAN,KEVIN,S
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JONES,IAN
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: ANTANAVICH

DATE/TIME OF TEST: 09/16/2022 13:53:49

Diagnostics Check	OK	13:51	
Air Blank	0.000	13:51	
Control Test	0.078	13:51	
Air Blank	0.000	13:52	
Subject Sample	# 1	13:52	Subject Test Refused
Breath Volume	0.000	13:52	
Air Blank	0.000	13:52	
Control Test	0.078	13:53	
Air Blank	0.000	13:53	
Diagnostics Check	OK	13:53	

Data Download Date/Time: 09/23/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LANG,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: BESSANT

DATE/TIME OF TEST: 09/16/2022 23:03:18

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.078	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.120	22:58
Breath Volume	2.207	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.118	23:01
Breath Volume	2.496	23:01
Air Blank	0.000	23:02
Control Test	0.078	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 09/23/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WICKERT,P
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: RYERSON

DATE/TIME OF TEST: 09/18/2022 21:25:19

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:22	
Control Test	0.079	21:23	
Air Blank	0.000	21:23	
Subject Sample	# 1	21:23	Subject Test Refused
Breath Volume	0.000	21:23	
Air Blank	0.000	21:24	
Control Test	0.079	21:24	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 09/23/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SINGH,T
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BAUMER

DATE/TIME OF TEST: 09/18/2022 23:07:46

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.079	23:05	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:06	Subject Test Refused
Breath Volume	0.000	23:06	
Air Blank	0.000	23:06	
Control Test	0.079	23:07	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 09/23/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUCIER,M,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SAUCIER,M
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 09/20/2022 04:43:37

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:41	
Control Test	0.079	04:41	
Air Blank	0.000	04:41	
Subject Sample	# 1	04:42	Subject Test Refused
Breath Volume	0.000	04:42	
Air Blank	0.000	04:42	
Control Test	0.079	04:43	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 09/23/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCELVY,J
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 09/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MCISAAC

DATE/TIME OF TEST: 09/21/2022 00:19:02

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.080	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.182	00:14
Breath Volume	2.007	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.179	00:17
Breath Volume	1.839	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:19

Data Download Date/Time: 09/23/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SELDES,Z
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SLAVEN

DATE/TIME OF TEST: 09/21/2022 03:07:26

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.112	03:02
Breath Volume	1.343	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.105	03:05
Breath Volume	1.664	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 09/23/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: REDDEN,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 09/22/2022 21:32:10

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.000	21:27
Breath Volume	2.085	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.000	21:30
Breath Volume	2.125	21:30
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 09/23/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NEWMAN,P
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 09/23/2022 17:31:49

Diagnostics Check	OK	17:28	
Air Blank	0.000	17:29	
Control Test	0.079	17:29	
Air Blank	0.000	17:30	
Subject Sample	# 1	17:30	Subject Test Refused
Breath Volume	0.000	17:30	
Air Blank	0.000	17:30	
Control Test	0.079	17:31	
Air Blank	0.000	17:31	
Diagnostics Check	OK	17:31	

Data Download Date/Time: 10/26/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: EXATUS

DATE/TIME OF TEST: 09/25/2022 00:15:10

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.079	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.167	00:10
Breath Volume	1.398	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.175	00:13
Breath Volume	1.394	00:13
Air Blank	0.000	00:14
Control Test	0.078	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 10/26/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FRAKER,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 09/25/2022 04:35:28

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.212	04:30
Breath Volume	1.535	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.199	04:33
Breath Volume	1.140	04:33
Air Blank	0.000	04:34
Control Test	0.077	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 10/26/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE, D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: GAW

DATE/TIME OF TEST: 09/25/2022 18:58:11

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.079	18:51
Air Blank	0.000	18:51
Subject Sample # 1	0.116	18:53
Breath Volume	1.406	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:55
Subject Sample # 2	0.114	18:56
Breath Volume	1.425	18:56
Air Blank	0.000	18:57
Control Test	0.078	18:57
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 10/26/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HERNANDEZ, S
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 09/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: MICHAELSON

DATE/TIME OF TEST: 09/25/2022 19:32:26

Diagnostics Check	OK	19:29	
Air Blank	0.000	19:29	
Control Test	0.079	19:30	
Air Blank	0.000	19:30	
Subject Sample	# 1	19:31	Subject Test Refused
Breath Volume	0.000	19:31	
Air Blank	0.000	19:31	
Control Test	0.079	19:31	
Air Blank	0.000	19:32	
Diagnostics Check	OK	19:32	

Data Download Date/Time: 10/26/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAFF,P,S
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOLLEY,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: SAVINO DONLEY

DATE/TIME OF TEST: 09/30/2022 23:57:38

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.163	23:53
Breath Volume	1.550	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.173	23:56
Breath Volume	1.433	23:56
Air Blank	0.000	23:56
Control Test	0.078	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 10/26/2022 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV, OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRUNO, OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: FITZPATRICK

DATE/TIME OF TEST: 09/07/2022 04:11:46

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.066	04:06
Breath Volume	2.375	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:08
Subject Sample # 2	0.059	04:10
Breath Volume	2.078	04:10
Air Blank	0.000	04:10
Control Test	0.078	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COLE,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 09/09/2022 04:06:37

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.000	04:02
Breath Volume	3.800	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.000	04:05
Breath Volume	4.078	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SMYTH,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 09/10/2022 21:02:27

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.000	20:58
Breath Volume	1.304	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.000	21:01
Breath Volume	1.652	21:01
Air Blank	0.000	21:01
Control Test	0.079	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GRAMMATICA

DATE/TIME OF TEST: 09/14/2022 02:01:42

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.129	01:57
Breath Volume	1.250	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.132	02:00
Breath Volume	1.289	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 09/15/2022 00:43:19

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.148	00:38
Breath Volume	2.476	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.140	00:41
Breath Volume	1.777	00:41
Air Blank	0.000	00:42
Control Test	0.077	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI, OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI, OFC. M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/15/2022 03:19:02

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.000	03:14
Breath Volume	2.578	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.000	03:17
Breath Volume	3.464	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:19

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHWAB,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: TOVILA-PERNIA

DATE/TIME OF TEST: 09/16/2022 05:15:53

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:13	
Control Test	0.079	05:13	
Air Blank	0.000	05:14	
Subject Sample	# 1	05:15	Sequence Aborted
Breath Volume	0.000	05:15	
Air Blank	0.000	05:15	

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO, OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHWAB, OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: TOVILA-PERNIA

DATE/TIME OF TEST: 09/16/2022 05:24:39

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:18	
Control Test	0.077	05:19	
Air Blank	0.000	05:19	
Subject Sample # 1	0.241	05:20	
Breath Volume	1.312	05:20	
Air Blank	0.000	05:21	
Air Blank	0.000	05:23	
Subject Sample # 2		05:23	Sequence Aborted
Breath Volume	0.156	05:23	
Air Blank	0.000	05:24	

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHWAB,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: TOVILA-PERNIA

DATE/TIME OF TEST: 09/16/2022 05:42:12

Diagnostics Check	OK	05:36
Air Blank	0.000	05:36
Control Test	0.079	05:36
Air Blank	0.000	05:37
Subject Sample # 1	0.244	05:37
Breath Volume	2.007	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.247	05:40
Breath Volume	1.621	05:40
Air Blank	0.000	05:41
Control Test	0.078	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: MILARDO

DATE/TIME OF TEST: 09/16/2022 23:29:15

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:22
Subject Sample # 1	0.102	23:24
Breath Volume	3.011	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:26
Subject Sample # 2	0.099	23:27
Breath Volume	1.210	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: SARVER

DATE/TIME OF TEST: 09/17/2022 04:28:20

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.079	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1 0.207	04:17	
Breath Volume	1.203	04:17	
Air Blank	0.000	04:18	
Air Blank	0.000	04:19	
Subject Sample	# 2 0.181	04:23	Volume Not Met
Breath Volume	0.734	04:23	
Air Blank	0.000	04:23	
Air Blank	0.000	04:25	
Subject Sample	# 3 0.208	04:26	
Breath Volume	2.621	04:26	
Air Blank	0.000	04:27	
Control Test	0.077	04:27	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BLANCO,J
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: RIVERA RODRIGUEZ

DATE/TIME OF TEST: 09/17/2022 05:11:54

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.000	05:07
Breath Volume	1.605	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.000	05:10
Breath Volume	1.519	05:10
Air Blank	0.000	05:10
Control Test	0.079	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MIKLUSCAK,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:56

SUBJECT'S LAST NAME: CARRERABARRIOS

DATE/TIME OF TEST: 09/17/2022 11:27:02

Diagnostics Check	OK	11:20
Air Blank	0.000	11:21
Control Test	0.079	11:21
Air Blank	0.000	11:22
Subject Sample # 1	0.164	11:22
Breath Volume	3.593	11:22
Air Blank	0.000	11:23
Air Blank	0.000	11:25
Subject Sample # 2	0.182	11:25
Breath Volume	3.152	11:25
Air Blank	0.000	11:26
Control Test	0.078	11:26
Air Blank	0.000	11:26
Diagnostics Check	OK	11:27

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SURRAN,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KENNY,S
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: FEINS

DATE/TIME OF TEST: 09/18/2022 21:09:43

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.080	21:02
Air Blank	0.000	21:02
Subject Sample # 1	0.000	21:05
Breath Volume	1.617	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.000	21:08
Breath Volume	2.117	21:08
Air Blank	0.000	21:08
Control Test	0.079	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SKINNER,OFFICER
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: DULL

DATE/TIME OF TEST: 09/18/2022 22:49:43

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.165	22:45
Breath Volume	2.015	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:47
Subject Sample # 2	0.157	22:48
Breath Volume	1.554	22:48
Air Blank	0.000	22:48
Control Test	0.076	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: BROOME

DATE/TIME OF TEST: 09/19/2022 23:22:45

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.205	23:18
Breath Volume	1.722	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.197	23:21
Breath Volume	1.664	23:21
Air Blank	0.000	23:21
Control Test	0.078	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SOLOVJEV

DATE/TIME OF TEST: 09/22/2022 03:20:18

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.070	03:15
Breath Volume	2.972	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.071	03:18
Breath Volume	2.562	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: URFER

DATE/TIME OF TEST: 09/22/2022 21:16:03

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.080	21:10
Air Blank	0.000	21:10
Subject Sample # 1	0.076	21:11
Breath Volume	2.687	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.076	21:14
Breath Volume	1.332	21:14
Air Blank	0.000	21:15
Control Test	0.079	21:15
Air Blank	0.000	21:15
Diagnostics Check	OK	21:16

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI,OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: PACHECO,OFC. E.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 09/22/2022 23:59:34

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.078	23:56	
Air Blank	0.000	23:56	
Subject Sample	# 1	23:58	Sequence Aborted
Breath Volume	0.242	23:58	
Air Blank	0.000	23:59	

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI,OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI,OFC. M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 09/23/2022 00:14:21

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.077	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1 0.064	00:06	Volume Not Met
Breath Volume	1.042	00:06	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample	# 2 0.084	00:09	
Breath Volume	1.289	00:09	
Air Blank	0.000	00:10	
Air Blank	0.000	00:12	
Subject Sample	# 3 0.087	00:12	
Breath Volume	1.320	00:12	
Air Blank	0.000	00:13	
Control Test	0.078	00:13	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: CESAR

DATE/TIME OF TEST: 09/23/2022 02:43:51

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.078	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.173	02:39
Breath Volume	3.156	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.175	02:42
Breath Volume	1.917	02:42
Air Blank	0.000	02:42
Control Test	0.078	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: HERNANDEZ,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: ALBRECHT

DATE/TIME OF TEST: 09/23/2022 22:00:32

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.081	21:53
Air Blank	0.000	21:53
Subject Sample # 1	0.106	21:55
Breath Volume	1.480	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.099	21:58
Breath Volume	1.289	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: SAY

DATE/TIME OF TEST: 09/24/2022 02:30:48

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.218	02:25
Breath Volume	1.898	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:27
Subject Sample # 2	0.215	02:29
Breath Volume	1.558	02:29
Air Blank	0.000	02:29
Control Test	0.076	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SMYTH,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: EPSTEIN

DATE/TIME OF TEST: 09/24/2022 04:21:00

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.193	04:16
Breath Volume	2.601	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.184	04:19
Breath Volume	1.949	04:19
Air Blank	0.000	04:20
Control Test	0.076	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:21

Data Download Date/Time: 09/24/2022 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NELSON,OFC A
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SAINT-CYR

DATE/TIME OF TEST: 09/25/2022 00:10:10

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.133	00:05
Breath Volume	1.500	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.134	00:08
Breath Volume	1.511	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 10/23/2022 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LUNDSTROM,OFC B
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: HERNANDEZ-ALVAREZ

DATE/TIME OF TEST: 09/25/2022 23:56:21

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.079	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.186	23:51
Breath Volume	2.828	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.166	23:54
Breath Volume	1.257	23:54
Air Blank	0.000	23:55
Control Test	0.075	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 10/23/2022 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TINFINA,LOGAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TINFINA,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: CASTILLO RESENDIZ

DATE/TIME OF TEST: 09/26/2022 03:35:48

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.080	03:34	
Air Blank	0.000	03:35	
		03:35	RFI Detect
Air Blank	0.000	03:35	

Data Download Date/Time: 10/23/2022 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TINFINA,LOGAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TINFINA,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: CASTILLO RESENDIZ

DATE/TIME OF TEST: 09/26/2022 03:44:03

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.124	03:39
Breath Volume	2.378	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.129	03:42
Breath Volume	2.136	03:42
Air Blank	0.000	03:43
Control Test	0.078	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:44

Data Download Date/Time: 10/23/2022 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLOW,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:58

SUBJECT'S LAST NAME: TORNELLO

DATE/TIME OF TEST: 09/27/2022 15:25:59

Diagnostics Check	OK	15:19
Air Blank	0.000	15:20
Control Test	0.079	15:20
Air Blank	0.000	15:20
Subject Sample # 1	0.267	15:21
Breath Volume	1.988	15:21
Air Blank	0.000	15:22
Air Blank	0.000	15:24
Subject Sample # 2	0.270	15:24
Breath Volume	1.996	15:24
Air Blank	0.000	15:25
Control Test	0.078	15:25
Air Blank	0.000	15:25
Diagnostics Check	OK	15:25

Data Download Date/Time: 10/23/2022 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: CARPENTIER

DATE/TIME OF TEST: 09/29/2022 21:17:06

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.080	21:07
Air Blank	0.000	21:08
Subject Sample # 1	0.147	21:08
Breath Volume	2.464	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.126	21:12
Breath Volume	1.660	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:14
Subject Sample # 3	0.134	21:15
Breath Volume	1.585	21:15
Air Blank	0.000	21:16
Control Test	0.078	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 10/23/2022 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: MACKINNON

DATE/TIME OF TEST: 09/02/2022 17:45:50

Diagnostics Check	OK	17:39
Air Blank	0.000	17:39
Control Test	0.081	17:39
Air Blank	0.000	17:40
Subject Sample # 1	0.202	17:41
Breath Volume	3.140	17:41
Air Blank	0.000	17:41
Air Blank	0.000	17:43
Subject Sample # 2	0.200	17:44
Breath Volume	2.964	17:44
Air Blank	0.000	17:44
Control Test	0.079	17:45
Air Blank	0.000	17:45
Diagnostics Check	OK	17:45

Data Download Date/Time: 09/24/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: OSTRIKOV

DATE/TIME OF TEST: 09/03/2022 01:10:07

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.081	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.153	01:05
Breath Volume	2.214	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.160	01:08
Breath Volume	1.722	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 09/24/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: KRAVCHENKO

DATE/TIME OF TEST: 09/03/2022 01:19:35

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.068	01:14
Breath Volume	2.582	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.064	01:18
Breath Volume	2.226	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 09/24/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PELOT

DATE/TIME OF TEST: 09/03/2022 02:05:43

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.081	01:55	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:58	No Sample Provided
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2	02:04	No Sample Provided
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.080	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 09/24/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,DEP G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: COLLAZO-RUVALCABA

DATE/TIME OF TEST: 09/03/2022 03:54:32

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.081	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.149	03:50
Breath Volume	2.507	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.145	03:52
Breath Volume	1.285	03:52
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 09/24/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: RODRIGUEZ,E
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 09/03/2022 14:22:09

Diagnostics Check	OK	14:18	
Air Blank	0.000	14:19	
Control Test	0.081	14:19	
Air Blank	0.000	14:19	
Subject Sample	# 1	14:20	Slope Not Met
Breath Volume	1.128	14:20	
Air Blank	0.000	14:21	
Control Test	0.082	14:21	
Air Blank	0.000	14:22	
Diagnostics Check	OK	14:22	

Data Download Date/Time: 09/24/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: RODRIGUEZ,E
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:23

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 09/03/2022 14:52:00

Diagnostics Check	OK	14:45
Air Blank	0.000	14:45
Control Test	0.080	14:46
Air Blank	0.000	14:46
Subject Sample # 1	0.271	14:47
Breath Volume	1.996	14:47
Air Blank	0.000	14:48
Air Blank	0.000	14:49
Subject Sample # 2	0.252	14:50
Breath Volume	1.617	14:50
Air Blank	0.000	14:51
Control Test	0.077	14:51
Air Blank	0.000	14:51
Diagnostics Check	OK	14:52

Data Download Date/Time: 09/24/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: BAURIEDL

DATE/TIME OF TEST: 09/03/2022 18:44:12

Diagnostics Check	OK	18:37
Air Blank	0.000	18:37
Control Test	0.081	18:37
Air Blank	0.000	18:38
Subject Sample # 1	0.176	18:39
Breath Volume	1.914	18:39
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.174	18:42
Breath Volume	1.832	18:42
Air Blank	0.000	18:43
Control Test	0.080	18:43
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

Data Download Date/Time: 09/24/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: SELLECK

DATE/TIME OF TEST: 09/03/2022 20:31:51

Diagnostics Check	OK	20:29	
Air Blank	0.000	20:29	
Control Test	0.081	20:29	
Air Blank	0.000	20:30	
Subject Sample	# 1	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:30	
Control Test	0.081	20:31	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 09/24/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: GOMEZ SANCHEZ

DATE/TIME OF TEST: 09/04/2022 00:57:43

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.242	00:53
Breath Volume	1.531	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.253	00:56
Breath Volume	1.644	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 09/24/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: ROUANET

DATE/TIME OF TEST: 09/04/2022 01:41:51

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.143	01:37
Breath Volume	2.589	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.146	01:40
Breath Volume	2.960	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 09/24/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: BOURDEAU

DATE/TIME OF TEST: 09/05/2022 18:54:45

Diagnostics Check	OK	18:48
Air Blank	0.000	18:48
Control Test	0.081	18:48
Air Blank	0.000	18:49
Subject Sample # 1	0.172	18:50
Breath Volume	2.351	18:50
Air Blank	0.000	18:50
Air Blank	0.000	18:52
Subject Sample # 2	0.172	18:53
Breath Volume	1.660	18:53
Air Blank	0.000	18:53
Control Test	0.081	18:54
Air Blank	0.000	18:54
Diagnostics Check	OK	18:54

Data Download Date/Time: 09/24/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, JOSHUA, H
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HARRISON, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 10:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:09

SUBJECT'S LAST NAME: BOLINGER

DATE/TIME OF TEST: 09/02/2022 11:35:20

Diagnostics Check	OK	11:32	
Air Blank	0.000	11:32	
Control Test	0.081	11:33	
Air Blank	0.000	11:33	
Subject Sample	# 1	11:33	Subject Test Refused
Breath Volume	0.000	11:33	
Air Blank	0.000	11:34	
Control Test	0.080	11:34	
Air Blank	0.000	11:35	
Diagnostics Check	OK	11:35	

Data Download Date/Time: 09/29/2022 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,AUSTIN,B
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: MORRIS,JACK
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: RIVAS-DOMINGO

DATE/TIME OF TEST: 09/04/2022 06:42:03

Diagnostics Check	OK	06:35
Air Blank	0.000	06:36
Control Test	0.081	06:36
Air Blank	0.000	06:37
Subject Sample # 1	0.098	06:37
Breath Volume	2.488	06:37
Air Blank	0.000	06:38
Air Blank	0.000	06:39
Subject Sample # 2	0.097	06:40
Breath Volume	2.328	06:40
Air Blank	0.000	06:41
Control Test	0.080	06:41
Air Blank	0.000	06:41
Diagnostics Check	OK	06:42

Data Download Date/Time: 09/29/2022 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS, JAKE, M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SWIGGETT, TIMOTHY
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/07/2022 00:14:06

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.081	00:11	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:13	
Control Test	0.080	00:13	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 09/29/2022 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSHAUGHNESSY,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Chiefland PD
ARRESTING OFFICER: OSHAUGHNESSY,TIMOTHY
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: NEESE

DATE/TIME OF TEST: 09/10/2022 04:17:38

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.081	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.139	04:13
Breath Volume	1.871	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.140	04:15
Breath Volume	1.398	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 09/29/2022 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS, JAKE, M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ADKINS, JAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: BABCOCK

DATE/TIME OF TEST: 09/10/2022 23:23:57

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.080	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:23	
Control Test	0.080	23:23	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 09/29/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,AUSTIN,B
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: POLK,AUSTIN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 09/16/2022 19:20:15

Diagnostics Check	OK	19:13
Air Blank	0.000	19:13
Control Test	0.080	19:14
Air Blank	0.000	19:14
Subject Sample # 1	0.114	19:15
Breath Volume	1.187	19:15
Air Blank	0.000	19:16
Air Blank	0.000	19:17
Subject Sample # 2	0.111	19:18
Breath Volume	1.699	19:18
Air Blank	0.000	19:19
Control Test	0.078	19:19
Air Blank	0.000	19:20
Diagnostics Check	OK	19:20

Data Download Date/Time: 09/29/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL, JAIRUS, S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCCONNELL, JAIRUS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/18/2022 12:06:53

Diagnostics Check	OK	12:00
Air Blank	0.000	12:00
Control Test	0.081	12:01
Air Blank	0.000	12:01
Subject Sample # 1	0.171	12:02
Breath Volume	1.820	12:02
Air Blank	0.000	12:02
Air Blank	0.000	12:04
Subject Sample # 2	0.171	12:05
Breath Volume	1.308	12:05
Air Blank	0.000	12:05
Control Test	0.081	12:06
Air Blank	0.000	12:06
Diagnostics Check	OK	12:06

Data Download Date/Time: 09/29/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: PEPPER

DATE/TIME OF TEST: 09/21/2022 21:30:08

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.080	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.036	21:25
Breath Volume	1.781	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.032	21:28
Breath Volume	1.160	21:28
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 09/29/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,AUSTIN,B
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: AUSTIN,POLK
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 09/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 09/30/2022 18:48:11

Diagnostics Check	OK	18:41
Air Blank	0.000	18:41
Control Test	0.080	18:42
Air Blank	0.000	18:42
Subject Sample # 1	0.015	18:43
Breath Volume	1.613	18:43
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 2	0.013	18:46
Breath Volume	2.007	18:46
Air Blank	0.000	18:47
Control Test	0.079	18:47
Air Blank	0.000	18:48
Diagnostics Check	OK	18:48

Data Download Date/Time: 10/03/2022 17:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: DARE

DATE/TIME OF TEST: 09/03/2022 01:04:05

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.081	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.030	00:59
Breath Volume	1.628	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:01
Subject Sample # 2	0.028	01:02
Breath Volume	1.613	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:04

Data Download Date/Time: 09/22/2022 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KING,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MOLENDIA

DATE/TIME OF TEST: 09/03/2022 02:06:17

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.080	02:04	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Control Test	0.080	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 09/22/2022 21:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: VIZCARRA

DATE/TIME OF TEST: 09/03/2022 03:35:20

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.081	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.234	03:30
Breath Volume	1.476	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample # 2	0.241	03:33
Breath Volume	1.714	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 09/22/2022 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BIGLIN,AMANDA
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 09/04/2022 17:21:16

Diagnostics Check	OK	17:18	
Air Blank	0.000	17:18	
Control Test	0.080	17:18	
Air Blank	0.000	17:19	
Subject Sample	# 1	17:19	Subject Test Refused
Breath Volume	0.000	17:19	
Air Blank	0.000	17:20	
Control Test	0.080	17:20	
Air Blank	0.000	17:21	
Diagnostics Check	OK	17:21	

Data Download Date/Time: 09/22/2022 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: GEISEN

DATE/TIME OF TEST: 09/04/2022 19:30:28

Diagnostics Check	OK	19:29	
Air Blank		19:30	RFI Detect
Air Blank		19:30	RFI Detect

Data Download Date/Time: 09/22/2022 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: TORRES BERRIOS

DATE/TIME OF TEST: 09/04/2022 21:53:49

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:45	
Control Test	0.080	21:45	
Air Blank	0.000	21:45	
Subject Sample	# 1 0.135	21:49	Volume Not Met
Breath Volume	0.367	21:49	
Air Blank	0.000	21:49	
Air Blank	0.000	21:51	
Subject Sample	# 2	21:52	Subject Test Refused
Breath Volume	0.015	21:52	
Air Blank	0.000	21:52	
Control Test	0.079	21:53	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 09/22/2022 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: HASSAN

DATE/TIME OF TEST: 09/08/2022 03:44:49

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:38	
Control Test	0.080	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1 0.237	03:39	
Breath Volume	1.523	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 2 0.238	03:43	
Breath Volume	1.429	03:43	
Air Blank	0.000	03:43	
Control Test	0.079	03:44	
Air Blank		03:44	RFI Detect
Air Blank	0.000	03:44	

Data Download Date/Time: 09/22/2022 21:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/08/2022 04:56:14

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:53	
Control Test	0.079	04:54	
Air Blank	0.000	04:54	
Subject Sample	# 1	04:54	Subject Test Refused
Breath Volume	0.000	04:54	
Air Blank	0.000	04:55	
Control Test	0.080	04:55	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 09/22/2022 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: GIRARDI

DATE/TIME OF TEST: 09/09/2022 02:36:32

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.000	02:31
Breath Volume	2.210	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.000	02:35
Breath Volume	2.066	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 09/22/2022 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: COWLES

DATE/TIME OF TEST: 09/11/2022 01:29:01

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.000	01:24
Breath Volume	2.015	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.000	01:27
Breath Volume	1.683	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:29

Data Download Date/Time: 09/22/2022 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: PARA

DATE/TIME OF TEST: 09/16/2022 02:14:31

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 09/22/2022 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/17/2022 01:58:54

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.029	01:54
Breath Volume	1.500	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.027	01:57
Breath Volume	1.718	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 09/22/2022 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: VERNI

DATE/TIME OF TEST: 09/17/2022 23:56:42

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.081	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.000	23:52
Breath Volume	2.750	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.000	23:55
Breath Volume	2.589	23:55
Air Blank	0.000	23:55
Control Test	0.080	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 09/22/2022 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ESPINOSA, ANGEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 09/11/2022 04:14:08

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.082	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.274	04:09
Breath Volume	1.992	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.283	04:12
Breath Volume	2.140	04:12
Air Blank	0.000	04:13
Control Test	0.082	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 09/14/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOSA, EMMANUEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 09/11/2022 04:37:10

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.082	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.175	04:32
Breath Volume	1.855	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.169	04:35
Breath Volume	1.867	04:35
Air Blank	0.000	04:36
Control Test	0.081	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 09/14/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 09/11/2022 05:10:16

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.082	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.000	05:05
Breath Volume	4.148	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.000	05:08
Breath Volume	3.781	05:08
Air Blank	0.000	05:09
Control Test	0.082	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 09/14/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TAPIA, JACOB
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: FLASHMAN

DATE/TIME OF TEST: 09/11/2022 22:41:50

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.082	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.000	22:37
Breath Volume	3.234	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.000	22:40
Breath Volume	5.183	22:40
Air Blank	0.000	22:40
Control Test	0.082	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 09/14/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOFFMAN, KYLE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 09/01/2022 02:18:26

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.203	02:14
Breath Volume	3.453	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.201	02:16
Breath Volume	3.691	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 09/14/2022 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD,ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ARCHILA VELA

DATE/TIME OF TEST: 09/01/2022 03:00:58

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.223	02:56
Breath Volume	2.835	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:58
Subject Sample # 2	0.222	02:59
Breath Volume	1.214	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 09/14/2022 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA, VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HUGO VAQUER

DATE/TIME OF TEST: 09/01/2022 04:15:59

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.129	04:11
Breath Volume	3.644	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.134	04:14
Breath Volume	3.371	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 09/14/2022 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GOLDENBERG, EVA
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: SHAH

DATE/TIME OF TEST: 09/01/2022 05:16:20

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.080	05:09
Air Blank	0.000	05:09
Subject Sample # 1	0.133	05:11
Breath Volume	1.773	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.138	05:14
Breath Volume	1.898	05:14
Air Blank	0.000	05:15
Control Test	0.079	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 09/14/2022 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ORTIZ,PAUL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 09/02/2022 01:51:23

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.190	01:46
Breath Volume	3.761	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.182	01:49
Breath Volume	3.609	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 09/14/2022 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER, DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: CALIX SORTO

DATE/TIME OF TEST: 09/02/2022 02:09:35

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.080	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.198	02:05
Breath Volume	2.375	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.189	02:08
Breath Volume	2.617	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 09/14/2022 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: TLAPIZACO IBARRA

DATE/TIME OF TEST: 09/02/2022 03:12:49

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.143	03:08
Breath Volume	1.554	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.145	03:11
Breath Volume	1.562	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 09/14/2022 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI, ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BARLATIER

DATE/TIME OF TEST: 09/02/2022 04:03:08

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.144	03:58
Breath Volume	1.453	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.151	04:01
Breath Volume	1.468	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 09/14/2022 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 09/02/2022 23:55:12

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.233	23:50
Breath Volume	5.117	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:52
Subject Sample # 2	0.237	23:53
Breath Volume	5.304	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 09/14/2022 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOSA,EMMANUEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MEJIADELEON

DATE/TIME OF TEST: 09/03/2022 01:23:11

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.266	01:17
Breath Volume	2.898	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.257	01:21
Breath Volume	1.289	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 09/14/2022 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALLE, OMAR
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ESCOBEDO ROSARIO

DATE/TIME OF TEST: 09/03/2022 03:38:22

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.173	03:33
Breath Volume	1.878	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.174	03:36
Breath Volume	2.285	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUSSAK, KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SANTAS

DATE/TIME OF TEST: 09/03/2022 04:58:07

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.150	04:53
Breath Volume	1.734	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.139	04:56
Breath Volume	1.285	04:56
Air Blank	0.000	04:57
Control Test	0.081	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KRUPINSKI,LUKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: MARZULLOALAMO

DATE/TIME OF TEST: 09/03/2022 06:56:55

Diagnostics Check	OK	06:50
Air Blank	0.000	06:51
Control Test	0.080	06:51
Air Blank	0.000	06:51
Subject Sample # 1	0.129	06:52
Breath Volume	3.296	06:52
Air Blank	0.000	06:52
Air Blank	0.000	06:54
Subject Sample # 2	0.129	06:55
Breath Volume	3.054	06:55
Air Blank	0.000	06:56
Control Test	0.080	06:56
Air Blank	0.000	06:56
Diagnostics Check	OK	06:56

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 09/03/2022 21:25:36

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:14	
Control Test	0.080	21:14	
Air Blank	0.000	21:15	
Subject Sample	# 1 0.296	21:16	
Breath Volume	1.281	21:16	
Air Blank	0.000	21:16	
Air Blank	0.000	21:18	
Subject Sample	# 2 0.260	21:20	
Breath Volume	1.593	21:20	
Air Blank	0.000	21:21	
Air Blank	0.000	21:23	
Subject Sample	# 3	21:24	Subject Test Refused
Breath Volume	0.000	21:24	
Air Blank	0.000	21:24	
Control Test	0.076	21:25	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: AMERIS

DATE/TIME OF TEST: 09/04/2022 01:26:18

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.199	01:20
Breath Volume	2.097	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.188	01:24
Breath Volume	1.847	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAL,RAJIV
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: ZAW

DATE/TIME OF TEST: 09/04/2022 03:47:18

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.101	03:42
Breath Volume	1.562	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.101	03:45
Breath Volume	2.390	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHIPLEY,DANIEL
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: FLORES MIRANDA

DATE/TIME OF TEST: 09/04/2022 04:31:55

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.162	04:27
Breath Volume	1.527	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.162	04:30
Breath Volume	1.312	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STOWE,NICOLAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 09/04/2022 05:21:47

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.164	05:17
Breath Volume	1.574	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:19
Subject Sample # 2	0.151	05:20
Breath Volume	1.109	05:20
Air Blank	0.000	05:20
Control Test	0.079	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IRELAND,DAVID
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: HERNANDEAAGRON

DATE/TIME OF TEST: 09/04/2022 06:22:06

Diagnostics Check	OK	06:15
Air Blank	0.000	06:16
Control Test	0.080	06:16
Air Blank	0.000	06:16
Subject Sample # 1	0.194	06:17
Breath Volume	1.140	06:17
Air Blank	0.000	06:17
Air Blank	0.000	06:19
Subject Sample # 2	0.205	06:20
Breath Volume	2.269	06:20
Air Blank	0.000	06:21
Control Test	0.080	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TEJERA,LUIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: GODOY

DATE/TIME OF TEST: 09/04/2022 12:41:18

Diagnostics Check	OK	12:29	
Air Blank	0.000	12:30	
Control Test	0.080	12:30	
Air Blank	0.000	12:30	
Subject Sample	# 1	12:34	No Sample Provided
Breath Volume	0.000	12:34	
Air Blank	0.000	12:34	
Air Blank	0.000	12:36	
Subject Sample	# 2	0.175	12:39
Breath Volume	0.132	12:39	Volume Not Met
Air Blank	0.000	12:40	
Control Test	0.081	12:40	
Air Blank	0.000	12:41	
Diagnostics Check	OK	12:41	

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIEBKNECHT,KEVIN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HEABERLIN

DATE/TIME OF TEST: 09/05/2022 02:53:28

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.081	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.195	02:48
Breath Volume	1.160	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.190	02:51
Breath Volume	1.175	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCCARTHY, MICHAEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 09/05/2022 04:13:08

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:10	
Control Test	0.080	04:11	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Control Test	0.080	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AMARAL,MARCELO
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: VASQUEZ CHAY

DATE/TIME OF TEST: 09/05/2022 23:55:36

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.081	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.157	23:51
Breath Volume	2.281	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.153	23:54
Breath Volume	2.085	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER, OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: POWELL, KALEY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:38

SUBJECT'S LAST NAME: JEFFRIES

DATE/TIME OF TEST: 09/06/2022 10:10:21

Diagnostics Check	OK	10:03
Air Blank	0.000	10:03
Control Test	0.079	10:04
Air Blank	0.000	10:04
Subject Sample # 1	0.252	10:05
Breath Volume	1.308	10:05
Air Blank	0.000	10:06
Air Blank	0.000	10:07
Subject Sample # 2	0.261	10:08
Breath Volume	1.792	10:08
Air Blank	0.000	10:09
Control Test	0.079	10:09
Air Blank	0.000	10:10
Diagnostics Check	OK	10:10

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DALLAGNOL

DATE/TIME OF TEST: 09/07/2022 01:17:36

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:15	
Control Test	0.081	01:15	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.080	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KRUPINSKI,LUKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: METELUS

DATE/TIME OF TEST: 09/07/2022 05:07:38

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.081	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.082	05:02
Breath Volume	2.460	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.076	05:06
Breath Volume	1.925	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: NICHOLAS

DATE/TIME OF TEST: 09/08/2022 00:42:04

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.267	00:37
Breath Volume	2.089	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.256	00:40
Breath Volume	1.730	00:40
Air Blank	0.000	00:41
Control Test	0.078	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:42

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMMONS,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DESIR

DATE/TIME OF TEST: 09/08/2022 02:14:27

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.000	02:10
Breath Volume	1.277	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.000	02:13
Breath Volume	1.472	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SEAY,ERICA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: BOLANOS CAMPOS

DATE/TIME OF TEST: 09/09/2022 00:55:41

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.175	00:47
Breath Volume	2.878	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.152	00:50
Breath Volume	2.042	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 3	0.178	00:54
Breath Volume	2.199	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STOWE,NICOLAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 09/09/2022 01:13:21

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.081	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.187	01:08
Breath Volume	1.617	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.188	01:11
Breath Volume	1.609	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 09/09/2022 02:22:59

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.204	02:18
Breath Volume	2.589	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.208	02:21
Breath Volume	2.773	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: SNELLING

DATE/TIME OF TEST: 09/09/2022 04:41:15

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:30	
Control Test	0.081	04:30	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:34	No Sample Provided
Breath Volume	0.000	04:34	
Air Blank	0.000	04:34	
Air Blank	0.000	04:36	
Subject Sample	# 2	04:39	No Sample Provided
Breath Volume	0.000	04:39	
Air Blank	0.000	04:40	
Control Test	0.081	04:40	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BAKSH,SAEED
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: KARLOWICZ

DATE/TIME OF TEST: 09/09/2022 20:29:31

Diagnostics Check	OK	20:23
Air Blank	0.000	20:23
Control Test	0.080	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.306	20:25
Breath Volume	2.144	20:25
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.299	20:27
Breath Volume	1.832	20:27
Air Blank	0.000	20:28
Control Test	0.078	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: SANTOS ORTEGA

DATE/TIME OF TEST: 09/10/2022 02:44:10

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.084	02:39
Breath Volume	2.281	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.079	02:42
Breath Volume	2.234	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GRISALES,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SANCHEZ FERMIN

DATE/TIME OF TEST: 09/10/2022 04:09:26

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.113	04:05
Breath Volume	1.207	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.115	04:07
Breath Volume	1.667	04:07
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 09/14/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PREDELUS, SMITH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 09/10/2022 04:54:19

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.196	04:49
Breath Volume	2.843	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.200	04:52
Breath Volume	2.753	04:52
Air Blank	0.000	04:53
Control Test	0.081	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 09/14/2022 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BARS LAND, BRENT
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: SOOKRAM

DATE/TIME OF TEST: 09/10/2022 07:55:51

Diagnostics Check	OK	07:49
Air Blank	0.000	07:50
Control Test	0.080	07:50
Air Blank	0.000	07:50
Subject Sample # 1	0.185	07:51
Breath Volume	1.597	07:51
Air Blank	0.000	07:51
Air Blank	0.000	07:53
Subject Sample # 2	0.186	07:54
Breath Volume	1.796	07:54
Air Blank	0.000	07:54
Control Test	0.080	07:55
Air Blank	0.000	07:55
Diagnostics Check	OK	07:55

Data Download Date/Time: 09/14/2022 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: UMANA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: CHINORODRIGUEZ

DATE/TIME OF TEST: 09/10/2022 08:29:21

Diagnostics Check	OK	08:22
Air Blank	0.000	08:23
Control Test	0.080	08:23
Air Blank	0.000	08:24
Subject Sample # 1	0.121	08:24
Breath Volume	2.574	08:24
Air Blank	0.000	08:25
Air Blank	0.000	08:27
Subject Sample # 2	0.120	08:27
Breath Volume	1.707	08:27
Air Blank	0.000	08:28
Control Test	0.080	08:28
Air Blank	0.000	08:29
Diagnostics Check	OK	08:29

Data Download Date/Time: 09/14/2022 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KRUPINSKI,LUKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: THOMAS ABRAHAM

DATE/TIME OF TEST: 09/14/2022 04:10:07

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.114	04:05
Breath Volume	3.828	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.118	04:08
Breath Volume	3.878	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 09/14/2022 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 09/18/2022 02:01:07

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.207	01:56
Breath Volume	2.703	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.206	01:59
Breath Volume	2.468	01:59
Air Blank	0.000	02:00
Control Test	0.077	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 10/26/2022 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: RUIZGORDILLO

DATE/TIME OF TEST: 09/19/2022 00:09:13

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.081	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.220	00:04
Breath Volume	1.359	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.219	00:07
Breath Volume	3.363	00:07
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 10/26/2022 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FLYNN,H
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: LAXTON

DATE/TIME OF TEST: 09/02/2022 23:49:49

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.184	23:44
Breath Volume	1.214	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.184	23:48
Breath Volume	2.367	23:48
Air Blank	0.000	23:48
Control Test	0.078	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FUENTEZ,J
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CRABTREE

DATE/TIME OF TEST: 09/03/2022 03:37:07

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.199	03:32
Breath Volume	2.187	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.199	03:35
Breath Volume	2.617	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GARLOW,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/03/2022 05:14:28

Diagnostics Check	OK	05:07
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.189	05:09
Breath Volume	3.082	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.183	05:12
Breath Volume	1.781	05:12
Air Blank	0.000	05:13
Control Test	0.078	05:13
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FLYNN,H
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 09/07/2022 01:58:36

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:56	
Control Test	0.079	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1	01:57	Subject Test Refused
Breath Volume	0.000	01:57	
Air Blank	0.000	01:57	
Control Test	0.079	01:58	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELSOMINO, MATTHEW, J
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: GELSOMINO, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 09/09/2022 22:36:19

Diagnostics Check	OK	22:29
Air Blank	0.000	22:30
Control Test	0.079	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.000	22:32
Breath Volume	2.398	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.000	22:34
Breath Volume	1.683	22:34
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MACON

DATE/TIME OF TEST: 09/09/2022 23:32:18

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:23	
Control Test	0.079	23:23	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:27	No Sample Provided
Breath Volume	0.000	23:27	
Air Blank	0.000	23:27	
Air Blank	0.000	23:29	
Subject Sample	# 2	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGBLOOD,Y,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: WHITING

DATE/TIME OF TEST: 09/12/2022 02:34:06

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.116	02:29
Breath Volume	2.269	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.112	02:32
Breath Volume	1.699	02:32
Air Blank	0.000	02:33
Control Test	0.078	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 09/18/2022 03:41:05

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.080	03:27	
Air Blank	0.000	03:27	
Subject Sample	# 1 0.171	03:28	
Breath Volume	1.238	03:28	
Air Blank	0.000	03:29	
Air Blank	0.000	03:30	
Subject Sample	# 2 0.000	03:34	Volume Not Met
Breath Volume	0.093	03:34	
Air Blank	0.000	03:34	
Air Blank	0.000	03:36	
Subject Sample	# 3	03:39	No Sample Provided
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Control Test	0.079	03:40	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MCFETRIDGE,P
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 09/19/2022 22:31:39

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
		22:31	RFI Detect
Air Blank		22:31	RFI Detect

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MCFETRIDGE,P
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 09/19/2022 22:40:23

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.337	22:35
Breath Volume	2.835	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.340	22:38
Breath Volume	4.226	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 09/20/2022 18:22:42

Diagnostics Check	OK	18:15
Air Blank	0.000	18:16
Control Test	0.079	18:16
Air Blank	0.000	18:16
Subject Sample # 1	0.000	18:18
Breath Volume	2.511	18:18
Air Blank	0.000	18:18
Air Blank	0.000	18:20
Subject Sample # 2	0.000	18:21
Breath Volume	2.437	18:21
Air Blank	0.000	18:21
Control Test	0.078	18:22
Air Blank	0.000	18:22
Diagnostics Check	OK	18:22

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CHURCH,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: DIANE

DATE/TIME OF TEST: 09/21/2022 23:02:08

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.223	22:56
Breath Volume	1.527	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.218	23:00
Breath Volume	1.746	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 09/25/2022 03:09:10

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.079	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Air Blank	0.000	03:06	
Subject Sample	# 2	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FLYNN,H
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 09/25/2022 03:18:36

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.141	03:14
Breath Volume	2.039	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.142	03:17
Breath Volume	2.179	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,S,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILLIAMS,T.R.
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: PACK

DATE/TIME OF TEST: 09/25/2022 07:40:16

Diagnostics Check	OK	07:31	
Air Blank	0.000	07:32	
Control Test	0.080	07:32	
Air Blank	0.000	07:33	
Subject Sample	# 1	07:36	No Sample Provided
Breath Volume	0.000	07:36	
Air Blank	0.000	07:36	
Air Blank	0.000	07:38	
Subject Sample	# 2	07:38	Subject Test Refused
Breath Volume	0.000	07:38	
Air Blank	0.000	07:39	
Control Test	0.080	07:39	
Air Blank	0.000	07:40	
Diagnostics Check	OK	07:40	

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDERS,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FUTCH,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:23

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 09/25/2022 07:58:21

Diagnostics Check	OK	07:45	
Air Blank	0.000	07:45	
Control Test	0.080	07:46	
Air Blank	0.000	07:46	
Subject Sample	# 1 0.115	07:47	
Breath Volume	1.222	07:47	
Air Blank	0.000	07:47	
Air Blank	0.000	07:49	
Subject Sample	# 2 0.098	07:52	Volume Not Met
Breath Volume	1.195	07:52	
Air Blank	0.000	07:53	
Air Blank	0.000	07:55	
Subject Sample	# 3 0.114	07:56	
Breath Volume	1.449	07:56	
Air Blank	0.000	07:57	
Control Test	0.079	07:57	
Air Blank	0.000	07:58	
Diagnostics Check	OK	07:58	

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGBLOOD,Y,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MUSTAFA,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: COPE

DATE/TIME OF TEST: 09/25/2022 22:50:19

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.358	22:45
Breath Volume	2.726	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.356	22:48
Breath Volume	3.312	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 10/25/2022 07:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIXON,QUADRY,D
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: GARZA,IVAN
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 09/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: CLINE

DATE/TIME OF TEST: 09/10/2022 21:36:52

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.081	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.085	21:32
Breath Volume	1.394	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:34
Subject Sample # 2	0.084	21:35
Breath Volume	1.824	21:35
Air Blank	0.000	21:35
Control Test	0.081	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 10/26/2022 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRABER,KIMBERLY,L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: DAME,TOMMY
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 09/06/2022 22:39:48

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
Control Test	0.079	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:35	No Sample Provided
Breath Volume	0.000	22:35	
Air Blank	0.000	22:35	
Air Blank	0.000	22:37	
Subject Sample	# 2	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Control Test	0.079	22:39	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 09/27/2022 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,DUSTIN,A
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: FERGUSON,KYRSTEN
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: LOPEZ-ESCALANTE

DATE/TIME OF TEST: 09/11/2022 23:38:49

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.078	23:26	
Air Blank	0.000	23:26	
Subject Sample	# 1 0.159	23:29	Volume Not Met
Breath Volume	0.085	23:29	
Air Blank	0.000	23:30	
Air Blank	0.000	23:32	
Subject Sample	# 2 0.170	23:33	
Breath Volume	3.437	23:33	
Air Blank	0.000	23:34	
Air Blank	0.000	23:36	
Subject Sample	# 3 0.159	23:37	
Breath Volume	1.597	23:37	
Air Blank	0.000	23:37	
Control Test	0.072	23:38	Control Outside Tolerance
Air Blank	0.000	23:38	

Data Download Date/Time: 09/27/2022 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,TREVOR,C
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: DRUMMOND,ALEX
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: BRENNEMAN

DATE/TIME OF TEST: 09/13/2022 21:33:50

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.079	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.000	21:29
Breath Volume	2.511	21:29
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.000	21:32
Breath Volume	1.976	21:32
Air Blank	0.000	21:32
Control Test	0.080	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 09/27/2022 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYANT,BRANDEN,J
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: CASSIDY,A
ARRESTING OFFICER AGENCY: Palatka PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 09/17/2022 17:47:11

Diagnostics Check	OK	17:40
Air Blank	0.000	17:41
Control Test	0.079	17:41
Air Blank	0.000	17:41
Subject Sample # 1	0.237	17:42
Breath Volume	1.902	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:44
Subject Sample # 2	0.232	17:45
Breath Volume	2.414	17:45
Air Blank	0.000	17:46
Control Test	0.079	17:46
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 09/27/2022 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,TREVOR,C
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: PLEIMAN,ERIC
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 09/18/2022 23:41:02

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.140	23:36
Breath Volume	1.406	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:38
Subject Sample # 2	0.150	23:39
Breath Volume	1.738	23:39
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:41

Data Download Date/Time: 09/27/2022 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAUENHOFER,TIMOTHY,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FRAUENHOFER,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 09/30/2022 18:52:26

Diagnostics Check	OK	18:46
Air Blank	0.000	18:46
Control Test	0.079	18:46
Air Blank	0.000	18:47
Subject Sample # 1	0.035	18:47
Breath Volume	3.792	18:47
Air Blank	0.000	18:48
Air Blank	0.000	18:50
Subject Sample # 2	0.031	18:50
Breath Volume	3.343	18:50
Air Blank	0.000	18:51
Control Test	0.078	18:51
Air Blank	0.000	18:52
Diagnostics Check	OK	18:52

Data Download Date/Time: 10/25/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:28

SUBJECT'S LAST NAME: RASADO

DATE/TIME OF TEST: 09/06/2022 12:57:19

Diagnostics Check	OK	12:48
Air Blank	0.000	12:48
Control Test	0.081	12:49
Air Blank	0.000	12:49
Subject Sample # 1	0.015	12:52
Breath Volume	1.886	12:52
Air Blank	0.000	12:52
Air Blank	0.000	12:54
Subject Sample # 2	0.014	12:55
Breath Volume	2.527	12:55
Air Blank	0.000	12:56
Control Test	0.080	12:56
Air Blank	0.000	12:57
Diagnostics Check	OK	12:57

Data Download Date/Time: 09/18/2022 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HANSELL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: ALNAFFAIE

DATE/TIME OF TEST: 09/01/2022 00:38:27

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:24	
Control Test	0.081	00:25	
Air Blank	0.000	00:25	
Subject Sample # 1	0.289	00:26	
Breath Volume	1.714	00:26	
Air Blank	0.000	00:27	
Air Blank	0.000	00:28	
Subject Sample # 2		00:31	No Sample Provided
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample # 3		00:37	Subject Test Refused
Breath Volume	0.000	00:37	
Air Blank	0.000	00:37	
Control Test	0.079	00:37	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 09/30/2022 07:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDERO,JAVIER,O
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CORDERO,JAVIER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RIVERA-ORTIZ

DATE/TIME OF TEST: 09/03/2022 03:15:47

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.195	03:11
Breath Volume	2.757	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.207	03:14
Breath Volume	2.660	03:14
Air Blank	0.000	03:14
Control Test	0.081	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 09/30/2022 07:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: PIETRILABOY

DATE/TIME OF TEST: 09/03/2022 04:33:23

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.131	04:28
Breath Volume	1.582	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.143	04:31
Breath Volume	1.222	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 09/30/2022 07:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: MALAVE ARRIAGA

DATE/TIME OF TEST: 09/09/2022 22:01:22

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.082	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.168	21:56
Breath Volume	2.480	21:56
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.159	21:59
Breath Volume	1.207	21:59
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 09/30/2022 07:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONTANEZ,YOMARI,I
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GRIEVE,ALEX
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/10/2022 00:34:39

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.082	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.196	00:29
Breath Volume	2.273	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.190	00:32
Breath Volume	2.632	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 09/30/2022 07:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: HAMILTON-AMBURGEY

DATE/TIME OF TEST: 09/10/2022 22:28:07

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.081	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.015	22:23
Breath Volume	1.347	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample # 2	0.013	22:26
Breath Volume	1.492	22:26
Air Blank	0.000	22:27
Control Test	0.081	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HANSELL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: BELEN-CARBONELL

DATE/TIME OF TEST: 09/11/2022 01:25:01

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.208	01:20
Breath Volume	1.343	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample # 2	0.208	01:23
Breath Volume	1.824	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:25

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEREZ,HECTOR
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: SEQUEIRA DE ZELEDON

DATE/TIME OF TEST: 09/11/2022 02:43:06

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.080	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:32	No Sample Provided
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Air Blank	0.000	02:34	
Subject Sample	# 2	0.274	02:37
Breath Volume	1.449	02:37	
Air Blank	0.000	02:38	
Air Blank	0.000	02:40	
Subject Sample	# 3	0.275	02:41
Breath Volume	1.320	02:41	
Air Blank		02:42	Purge Fail
Air Blank	0.000	02:43	

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENWALL,ROBERT,W
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GREENWALL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 14:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:24

SUBJECT'S LAST NAME: WAKWFIELD

DATE/TIME OF TEST: 09/14/2022 14:51:29

Diagnostics Check	OK	14:45
Air Blank	0.000	14:45
Control Test	0.081	14:46
Air Blank	0.000	14:46
Subject Sample # 1	0.248	14:47
Breath Volume	1.574	14:47
Air Blank	0.000	14:47
Air Blank	0.000	14:49
Subject Sample # 2	0.248	14:49
Breath Volume	1.675	14:49
Air Blank	0.000	14:50
Control Test	0.080	14:50
Air Blank	0.000	14:51
Diagnostics Check	OK	14:51

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CARTAGENA

DATE/TIME OF TEST: 09/16/2022 04:24:57

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.081	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.019	04:20
Breath Volume	1.207	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.017	04:23
Breath Volume	2.269	04:23
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AKINS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: AKINS,JUSTIN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ORAMAS BASULTO

DATE/TIME OF TEST: 09/16/2022 23:54:41

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.081	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.120	23:50
Breath Volume	2.390	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.119	23:53
Breath Volume	1.449	23:53
Air Blank	0.000	23:53
Control Test	0.081	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 07:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:46

SUBJECT'S LAST NAME: RODRIGUEZ GODOY

DATE/TIME OF TEST: 09/21/2022 08:14:33

Diagnostics Check	OK	08:06
Air Blank	0.000	08:07
Control Test	0.081	08:07
Air Blank	0.000	08:08
Subject Sample # 1	0.082	08:09
Breath Volume	1.410	08:09
Air Blank	0.000	08:10
Air Blank	0.000	08:11
Subject Sample # 2	0.076	08:12
Breath Volume	1.281	08:12
Air Blank	0.000	08:13
Control Test	0.079	08:13
Air Blank	0.000	08:14
Diagnostics Check	OK	08:14

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 07:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:46

SUBJECT'S LAST NAME: RODRIGUEZ GODOY

DATE/TIME OF TEST: 09/21/2022 08:27:37

Diagnostics Check	OK	08:20
Air Blank	0.000	08:20
Control Test	0.079	08:20
Air Blank	0.000	08:21
Subject Sample # 1	0.070	08:22
Breath Volume	1.492	08:22
Air Blank	0.000	08:23
Air Blank	0.000	08:25
Subject Sample # 2	0.070	08:26
Breath Volume	1.410	08:26
Air Blank	0.000	08:26
Control Test	0.079	08:27
Air Blank	0.000	08:27
Diagnostics Check	OK	08:27

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBERI, FRANCIS, J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: PARTANEN, TYLER
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 09/22/2022 03:01:20

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.115	02:56
Breath Volume	2.207	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.106	02:59
Breath Volume	2.214	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ESCALANTEALVARENGA

DATE/TIME OF TEST: 09/24/2022 01:47:12

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.081	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.155	01:42
Breath Volume	4.097	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.147	01:45
Breath Volume	3.855	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,CARLOS,I
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TORRES,CARLOS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: FABIUS

DATE/TIME OF TEST: 09/24/2022 05:09:36

Diagnostics Check	OK	05:08	
Air Blank		05:09	Ambient Fail
Air Blank	0.000	05:09	

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,CARLOS,I
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TORRES,CARLOS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: FABIUS

DATE/TIME OF TEST: 09/24/2022 05:21:07

Diagnostics Check	OK	05:12
Air Blank	0.000	05:13
Control Test	0.081	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.085	05:15
Breath Volume	1.332	05:15
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.082	05:19
Breath Volume	1.332	05:19
Air Blank	0.000	05:20
Control Test	0.080	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROSZEWSKI PEREIRA, THIAGO,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JAROSZEWSKI
PEREIRA, THIAGO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male

OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: BUBON

DATE/TIME OF TEST: 09/25/2022 00:34:38

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.081	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.122	00:29
Breath Volume	4.203	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.117	00:32
Breath Volume	4.054	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 09/25/2022 00:50:54

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:48	
Control Test	0.081	00:48	
Air Blank	0.000	00:49	
Subject Sample	# 1	00:49	Subject Test Refused
Breath Volume	0.000	00:49	
Air Blank	0.000	00:49	
Control Test	0.081	00:50	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONTANEZ,YOMARI,I
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: FONTANEZ,YOMARI
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 09/25/2022 02:31:01

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.081	02:28	
Air Blank	0.000	02:28	
Subject Sample # 1	0.186	02:29	
Breath Volume	1.453	02:29	
Air Blank		02:29	RFI Detect
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONTANEZ,YOMARI,I
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: FONTANEZ,YOMARI
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 09/25/2022 02:38:50

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.081	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1 0.169	02:37	Volume Not Met
Breath Volume	0.226	02:37	
Air Blank		02:38	Purge Fail
Air Blank	0.000	02:38	

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONTANEZ,YOMARI,I
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: FONTANEZ,YOMARI
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 09/25/2022 02:49:07

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.179	02:44
Breath Volume	2.027	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.175	02:47
Breath Volume	2.007	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 09/25/2022 03:57:09

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.161	03:52
Breath Volume	1.976	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:54
Subject Sample # 2	0.158	03:55
Breath Volume	1.316	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: QUIROSVALENTIN

DATE/TIME OF TEST: 09/25/2022 06:07:48

Diagnostics Check	OK	06:04	
Air Blank	0.000	06:04	
Control Test	0.081	06:05	
Air Blank	0.000	06:05	
Subject Sample	# 1	06:06	Subject Test Refused
Breath Volume	0.000	06:06	
Air Blank	0.000	06:06	
Control Test	0.081	06:07	
Air Blank	0.000	06:07	
Diagnostics Check	OK	06:07	

Data Download Date/Time: 09/30/2022 07:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWEENEY,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SWEENEY,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 07:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: MEDINA VELEZ

DATE/TIME OF TEST: 09/27/2022 08:55:40

Diagnostics Check	OK	08:47
Air Blank	0.000	08:47
Control Test	0.081	08:48
Air Blank	0.000	08:48
Subject Sample # 1	0.000	08:50
Breath Volume	2.734	08:50
Air Blank	0.000	08:51
Air Blank	0.000	08:53
Subject Sample # 2	0.000	08:54
Breath Volume	2.539	08:54
Air Blank	0.000	08:54
Control Test	0.079	08:55
Air Blank	0.000	08:55
Diagnostics Check	OK	08:55

Data Download Date/Time: 09/30/2022 07:13