

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRITO,XAVIER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: CULBERT

DATE/TIME OF TEST: 09/03/2022 23:02:10

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.025	22:57
Breath Volume	3.062	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.024	23:00
Breath Volume	3.023	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 09/26/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: CALDERON-MENDEZ

DATE/TIME OF TEST: 09/04/2022 21:59:46

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.080	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.149	21:55
Breath Volume	3.265	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.148	21:58
Breath Volume	2.761	21:58
Air Blank	0.000	21:58
Control Test	0.080	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 09/26/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: YOC PEREZ

DATE/TIME OF TEST: 09/05/2022 01:22:54

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.233	01:18
Breath Volume	2.261	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:20
Subject Sample # 2	0.222	01:21
Breath Volume	1.687	01:21
Air Blank	0.000	01:21
Control Test	0.079	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 09/26/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/09/2022 02:02:04

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.109	01:57
Breath Volume	1.320	01:57
Air Blank	0.000	01:58
Air Blank	0.000	01:59
Subject Sample # 2	0.109	02:00
Breath Volume	1.437	02:00
Air Blank	0.000	02:01
Control Test	0.077	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:02

Data Download Date/Time: 09/26/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: DE LA CRUZ BATEN

DATE/TIME OF TEST: 09/03/2022 05:00:05

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.080	04:47	
Air Blank	0.000	04:47	
Subject Sample # 1	0.174	04:49	
Breath Volume	1.535	04:49	
Air Blank	0.000	04:50	
Air Blank	0.000	04:51	
Subject Sample # 2	0.142	04:54	Volume Not Met
Breath Volume	0.164	04:54	
Air Blank	0.000	04:56	
Air Blank	0.000	04:57	
Subject Sample # 3	0.159	04:58	
Breath Volume	1.933	04:58	
Air Blank		04:59	Purge Fail
Air Blank	0.000	05:00	

Data Download Date/Time: 09/26/2022 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: BORJAS GAITAN

DATE/TIME OF TEST: 09/18/2022 00:09:16

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.173	00:04
Breath Volume	2.273	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.171	00:07
Breath Volume	2.390	00:07
Air Blank	0.000	00:08
Control Test	0.080	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 09/26/2022 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KOSAKOWSKI,PETTER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: NUEVA

DATE/TIME OF TEST: 09/20/2022 05:30:06

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.081	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.155	05:24
Breath Volume	2.378	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.144	05:28
Breath Volume	1.210	05:28
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 09/26/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: KEITA

DATE/TIME OF TEST: 09/22/2022 19:56:25

Diagnostics Check	OK	19:48
Air Blank	0.000	19:49
Control Test	0.080	19:49
Air Blank	0.000	19:50
Subject Sample # 1	0.381	19:51
Breath Volume	2.289	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.379	19:54
Breath Volume	1.906	19:54
Air Blank	0.000	19:55
Control Test	0.078	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 09/26/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL, REAGHN, S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIRIGNANO, JESSICA
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LOPEZ RAMIREZ

DATE/TIME OF TEST: 09/23/2022 03:23:28

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.165	03:18
Breath Volume	1.578	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample # 2	0.176	03:21
Breath Volume	2.562	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 09/26/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DORSEY,M
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: DIFRANCESCO

DATE/TIME OF TEST: 09/30/2022 22:55:13

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.000	22:50
Breath Volume	2.640	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.000	22:53
Breath Volume	2.648	22:53
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 10/25/2022 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: TIMONEY,RYAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 09/02/2022 21:35:53

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:30
Subject Sample # 1	0.311	21:31
Breath Volume	3.308	21:31
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.299	21:34
Breath Volume	2.988	21:34
Air Blank	0.000	21:34
Control Test	0.078	21:35
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 09/19/2022 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WALTER,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: BLACKSHAW

DATE/TIME OF TEST: 09/03/2022 03:41:28

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.270	03:36
Breath Volume	2.585	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.259	03:39
Breath Volume	2.867	03:39
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 09/19/2022 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA RUA,BRYAN,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HORNE,ASHTON
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 09/03/2022 22:58:49

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.182	22:50
Breath Volume	3.515	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:52
Subject Sample # 2	0.146	22:54
Breath Volume	2.007	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 3	0.164	22:57
Breath Volume	3.750	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 09/19/2022 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: DE LA RUA,BRYAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 09/04/2022 02:57:05

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.158	02:52
Breath Volume	1.562	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.166	02:55
Breath Volume	1.675	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 09/19/2022 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA RUA,BRYAN,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 09/05/2022 01:54:55

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.071	01:50
Breath Volume	2.957	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.074	01:53
Breath Volume	2.179	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 09/19/2022 06:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 09/05/2022 05:10:29

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:08	
Control Test	0.079	05:08	
Air Blank	0.000	05:08	
Subject Sample	# 1	05:09	Subject Test Refused
Breath Volume	0.000	05:09	
Air Blank	0.000	05:09	
Control Test	0.080	05:09	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

Data Download Date/Time: 09/19/2022 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: METOYER

DATE/TIME OF TEST: 09/05/2022 21:25:26

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:23	
Control Test	0.080	21:23	
Air Blank	0.000	21:23	
Subject Sample	# 1	21:24	Subject Test Refused
Breath Volume	0.000	21:24	
Air Blank	0.000	21:24	
Control Test	0.079	21:24	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 09/19/2022 06:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CALHOUN,KEVIN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 09/10/2022 16:03:42

Diagnostics Check	OK	15:59	
Air Blank	0.000	15:59	
Control Test	0.080	15:59	
Air Blank	0.000	16:00	
Subject Sample	# 1	16:02	Subject Test Refused
Breath Volume	0.226	16:02	
Air Blank	0.000	16:02	
Control Test	0.081	16:03	
Air Blank	0.000	16:03	
Diagnostics Check	OK	16:03	

Data Download Date/Time: 09/19/2022 06:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: BRANDT

DATE/TIME OF TEST: 09/11/2022 19:03:16

Diagnostics Check	OK	18:56
Air Blank	0.000	18:56
Control Test	0.079	18:57
Air Blank	0.000	18:57
Subject Sample # 1	0.319	18:58
Breath Volume	1.289	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:00
Subject Sample # 2	0.318	19:01
Breath Volume	2.246	19:01
Air Blank	0.000	19:02
Control Test	0.078	19:02
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 09/19/2022 06:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARECCIA,GREGORY,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DE CASTRO SILVA

DATE/TIME OF TEST: 09/12/2022 02:52:15

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:48	No Sample Provided
Breath Volume	0.000	02:48	
Air Blank	0.000	02:48	
Air Blank	0.000	02:50	
Subject Sample	# 2	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 09/19/2022 06:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: BONNER

DATE/TIME OF TEST: 09/12/2022 04:25:55

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:22	
Control Test	0.080	04:22	
Air Blank	0.000	04:23	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Control Test	0.080	04:25	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 09/19/2022 06:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: IGLESIAS,JORGE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CHESTER

DATE/TIME OF TEST: 09/03/2022 23:31:54

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.081	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1 0.070	23:24	Volume Not Met
Breath Volume	0.308	23:24	
Air Blank	0.000	23:25	
Air Blank	0.000	23:27	
Subject Sample	# 2 0.092	23:30	Volume Not Met
Breath Volume	0.773	23:30	
Air Blank	0.000	23:30	
Control Test	0.080	23:31	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 10/26/2022 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: IGLESIAS,JORGE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CHESTER

DATE/TIME OF TEST: 09/03/2022 23:46:47

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.079	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 1 0.083	23:39	Volume Not Met
Breath Volume	0.464	23:39	
Air Blank	0.000	23:40	
Air Blank	0.000	23:42	
Subject Sample	# 2 0.080	23:45	Volume Not Met
Breath Volume	0.417	23:45	
Air Blank	0.000	23:45	
Control Test	0.080	23:46	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 10/26/2022 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: IGLESIAS,JORGE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CHESTER

DATE/TIME OF TEST: 09/03/2022 23:59:47

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.103	23:55
Breath Volume	1.417	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.104	23:58
Breath Volume	1.160	23:58
Air Blank	0.000	23:58
Control Test	0.081	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 10/26/2022 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIERRA,ALIEK,O
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SIERRA,ALIEK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ESTRELLA MEJIA

DATE/TIME OF TEST: 09/04/2022 00:28:16

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.080	00:26	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:27	
Control Test	0.081	00:27	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 10/26/2022 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: ESCANDELL SILVA

DATE/TIME OF TEST: 09/05/2022 03:51:51

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.206	03:47
Breath Volume	1.675	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.210	03:50
Breath Volume	1.207	03:50
Air Blank	0.000	03:50
Control Test	0.081	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 10/26/2022 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIERRA,ALIEK,O
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SIERRA,ALIEK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MORETA CORNIELLE

DATE/TIME OF TEST: 09/10/2022 04:50:14

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:47	
Control Test	0.081	04:48	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:49	
Control Test	0.082	04:49	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 10/26/2022 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MONTE DE OCA
VIERA,ALEXANDER
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female

OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 09/11/2022 04:02:19

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.079	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.189	03:57
Breath Volume	1.820	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.194	04:00
Breath Volume	2.441	04:00
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 10/26/2022 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES, AMAURY, A
BREATH TEST OPERATOR AGENCY: FHP TROOP K
ARRESTING OFFICER: QUINONES, AMAURY
ARRESTING OFFICER AGENCY: FHP TROOP K
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 07:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:24

SUBJECT'S LAST NAME: OLIVA

DATE/TIME OF TEST: 09/11/2022 09:15:10

Diagnostics Check	OK	09:08
Air Blank	0.000	09:08
Control Test	0.081	09:08
Air Blank	0.000	09:09
Subject Sample # 1	0.112	09:10
Breath Volume	2.363	09:10
Air Blank	0.000	09:11
Air Blank	0.000	09:13
Subject Sample # 2	0.121	09:13
Breath Volume	2.796	09:13
Air Blank	0.000	09:14
Control Test	0.081	09:14
Air Blank	0.000	09:15
Diagnostics Check	OK	09:15

Data Download Date/Time: 10/26/2022 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAXIMILIEN, RONALD,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MAXIMILIEN, RONALD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 08:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:19

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/11/2022 10:01:17

Diagnostics Check	OK	09:54
Air Blank	0.000	09:55
Control Test	0.081	09:55
Air Blank	0.000	09:56
Subject Sample # 1	0.199	09:56
Breath Volume	1.890	09:56
Air Blank	0.000	09:57
Air Blank	0.000	09:58
Subject Sample # 2	0.200	09:59
Breath Volume	1.925	09:59
Air Blank	0.000	10:00
Control Test	0.080	10:00
Air Blank	0.000	10:01
Diagnostics Check	OK	10:01

Data Download Date/Time: 10/26/2022 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODDEN,DANIEL,R
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BODDEN,DANIEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/14/2022 18:47:20

Diagnostics Check	OK	18:40
Air Blank	0.000	18:40
Control Test	0.081	18:41
Air Blank	0.000	18:41
Subject Sample # 1	0.085	18:42
Breath Volume	1.921	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.082	18:45
Breath Volume	2.722	18:45
Air Blank	0.000	18:46
Control Test	0.081	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 10/26/2022 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIERRA,ALIEK,O
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SIERRA,ALIEK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 12:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/18/2022 01:17:49

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.000	01:13
Breath Volume	2.621	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.000	01:16
Breath Volume	2.960	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 10/26/2022 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIERRA,ALIEK,O
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SIERRA',ALIEK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: SANTES

DATE/TIME OF TEST: 09/18/2022 05:10:04

Diagnostics Check	OK	04:59	
Air Blank	0.000	05:00	
Control Test	0.081	05:00	
Air Blank	0.000	05:00	
Subject Sample	# 1 0.148	05:04	Volume Not Met
Breath Volume	0.960	05:04	
Air Blank	0.000	05:04	
Air Blank	0.000	05:06	
Subject Sample	# 2	05:08	Subject Test Refused
Breath Volume	1.101	05:08	
Air Blank	0.000	05:09	
Control Test	0.079	05:09	
Air Blank	0.000	05:09	
Diagnostics Check	OK	05:10	

Data Download Date/Time: 10/26/2022 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIERRA,ALIEK,O
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SIERRA,ALIEK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: SANTES

DATE/TIME OF TEST: 09/18/2022 05:46:55

Diagnostics Check	OK	05:39
Air Blank	0.000	05:39
Control Test	0.080	05:39
Air Blank	0.000	05:40
Subject Sample # 1	0.147	05:41
Breath Volume	2.054	05:41
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.149	05:45
Breath Volume	2.031	05:45
Air Blank	0.000	05:46
Control Test	0.077	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 10/26/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: BOSQUE,JOEL
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: OTHER
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: GONZALEZ HERNANDEZ

DATE/TIME OF TEST: 09/18/2022 08:08:08

Diagnostics Check	OK	08:05	
Air Blank	0.000	08:05	
Control Test	0.081	08:06	
Air Blank	0.000	08:06	
Subject Sample	# 1	08:06	Subject Test Refused
Breath Volume	0.000	08:06	
Air Blank	0.000	08:07	
Control Test	0.082	08:07	
Air Blank	0.000	08:08	
Diagnostics Check	OK	08:08	

Data Download Date/Time: 10/26/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JAIME,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: VEGA,JAIME
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/18/2022 17:27:16

Diagnostics Check	OK	17:20
Air Blank	0.000	17:21
Control Test	0.082	17:21
Air Blank	0.000	17:21
Subject Sample # 1	0.192	17:22
Breath Volume	1.453	17:22
Air Blank	0.000	17:23
Air Blank	0.000	17:25
Subject Sample # 2	0.189	17:25
Breath Volume	1.308	17:25
Air Blank	0.000	17:26
Control Test	0.081	17:26
Air Blank	0.000	17:27
Diagnostics Check	OK	17:27

Data Download Date/Time: 10/26/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ
AGUILERA,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SIERRA,ALIEK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023

VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RODRIGUEZ AGUILERA

DATE/TIME OF TEST: 09/24/2022 01:40:18

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.080	01:38	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:38	Subject Test Refused
Breath Volume	0.000	01:38	
Air Blank	0.000	01:39	
Control Test	0.080	01:39	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 10/26/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIERRA,ALIEK,O
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SIERRA,ALIEK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RODRIGUEZ AGUILERA

DATE/TIME OF TEST: 09/24/2022 01:47:56

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:45	
Control Test	0.078	01:45	
Air Blank	0.000	01:46	
Subject Sample	# 1	01:46	Subject Test Refused
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Control Test	0.080	01:47	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 10/26/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,YUDIEL,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SANTANA,YUDIEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: PEREZ MACHADO

DATE/TIME OF TEST: 09/24/2022 01:58:26

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:55	
Control Test	0.081	01:56	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:57	Subject Test Refused
Breath Volume	0.000	01:57	
Air Blank	0.000	01:57	
Control Test	0.080	01:57	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 10/26/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSQUERA,ODEL,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MOSQUERA,ODEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PADRINO SOBRINO

DATE/TIME OF TEST: 09/24/2022 02:31:58

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.082	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.189	02:27
Breath Volume	1.484	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:29
Subject Sample # 2	0.197	02:30
Breath Volume	2.304	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 10/26/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SIERRA,ALIEK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/25/2022 00:15:05

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.126	00:10
Breath Volume	1.355	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.130	00:13
Breath Volume	1.164	00:13
Air Blank	0.000	00:14
Control Test	0.081	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 10/26/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: ROMERO,J
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: GUTIERREZ FUNDORA

DATE/TIME OF TEST: 09/25/2022 10:42:24

Diagnostics Check	OK	10:35
Air Blank	0.000	10:36
Control Test	0.081	10:36
Air Blank	0.000	10:37
Subject Sample # 1	0.220	10:37
Breath Volume	1.363	10:37
Air Blank	0.000	10:38
Air Blank	0.000	10:40
Subject Sample # 2	0.209	10:40
Breath Volume	1.453	10:40
Air Blank	0.000	10:41
Control Test	0.078	10:41
Air Blank	0.000	10:42
Diagnostics Check	OK	10:42

Data Download Date/Time: 10/26/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: ALAVREZ

DATE/TIME OF TEST: 09/18/2022 03:20:06

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.078	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.196	03:15
Breath Volume	2.078	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.196	03:18
Breath Volume	2.781	03:18
Air Blank	0.000	03:19
Control Test	0.077	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:20

Data Download Date/Time: 09/23/2022 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: XORXE PABLO

DATE/TIME OF TEST: 09/17/2022 00:24:35

Diagnostics Check	OK	00:24	
Air Blank		00:24	RFI Detect
Air Blank	0.000	00:24	

Data Download Date/Time: 09/26/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: XORXE PABLO

DATE/TIME OF TEST: 09/17/2022 00:30:31

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.075	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:28	RFI Detect
Breath Volume	0.621	00:28	
Air Blank	0.000	00:29	
Control Test	0.076	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 09/26/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: XORXE PABLO

DATE/TIME OF TEST: 09/17/2022 00:35:25

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.075	00:33	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:34	RFI Detect
Breath Volume	0.000	00:34	
Air Blank	0.000	00:34	
Control Test	0.076	00:34	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 09/26/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: XORXE PABLO

DATE/TIME OF TEST: 09/17/2022 00:38:40

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.036	00:38	RFI Detect
Air Blank	0.000	00:38	

Data Download Date/Time: 09/26/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: XORXE PABLO

DATE/TIME OF TEST: 09/17/2022 00:47:22

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.078	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.188	00:42
Breath Volume	3.343	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.182	00:45
Breath Volume	3.457	00:45
Air Blank	0.000	00:46
Control Test	0.076	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 09/26/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: KNOWLES

DATE/TIME OF TEST: 09/17/2022 01:16:33

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.188	01:11
Breath Volume	3.140	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.185	01:14
Breath Volume	2.296	01:14
Air Blank	0.000	01:15
Control Test	0.077	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 09/26/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLAZOS JR,DIEGO,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COLLAZOS JR,DIEGO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: CASTILLO PAZ

DATE/TIME OF TEST: 09/18/2022 01:31:19

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.213	01:26
Breath Volume	1.441	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.225	01:29
Breath Volume	2.589	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 09/26/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: GADEA PEREIRA

DATE/TIME OF TEST: 09/18/2022 02:44:47

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.144	02:40
Breath Volume	1.511	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.153	02:43
Breath Volume	1.476	02:43
Air Blank	0.000	02:43
Control Test	0.077	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 09/26/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: RICH, JOSEPH	ARREST TIME: 22:02
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 08/29/2022	OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: PATEL	DATE/TIME OF TEST: 09/02/2022 22:44:25
----------------------------	--

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.080	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.156	22:39
Breath Volume	1.886	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.153	22:42
Breath Volume	1.394	22:42
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 09/26/2022 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: GALLOWAY,GRANT	ARREST TIME: 23:59
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 08/29/2022	OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: RATLIFF	DATE/TIME OF TEST: 09/03/2022	00:26:23
------------------------------	-------------------------------	----------

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.233	00:21
Breath Volume	2.273	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.233	00:24
Breath Volume	2.500	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 09/26/2022 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: RICH,JOSEPH	ARREST TIME: 23:32
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 08/29/2022	OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: ASHRIF	DATE/TIME OF TEST: 09/04/2022	00:15:51
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.107	00:11
Breath Volume	2.027	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:13
Subject Sample # 2	0.109	00:14
Breath Volume	1.140	00:14
Air Blank	0.000	00:14
Control Test	0.078	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 09/26/2022 09:12

Subject Test Electronic Data

DATE/TIME OF TEST: 09/04/2022 22:01:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: GALLOWAY, GRANT	ARREST TIME: 00:51
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 08/29/2022	OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: WINTERS JR	DATE/TIME OF TEST: 09/17/2022 01:23:22
---------------------------------	--

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.231	01:18
Breath Volume	1.722	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:20
Subject Sample # 2	0.236	01:21
Breath Volume	1.484	01:21
Air Blank	0.000	01:22
Control Test	0.076	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 09/26/2022 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: DE PAZ RIZO

DATE/TIME OF TEST: 09/01/2022 20:38:44

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:32	
Control Test	0.081	20:32	
Air Blank	0.000	20:33	
Subject Sample # 1	0.193	20:34	
Breath Volume	1.792	20:34	
Air Blank	0.000	20:35	
Air Blank	0.000	20:36	
Subject Sample # 2	0.194	20:37	
Breath Volume	1.921	20:37	
Air Blank		20:38	Purge Fail
Air Blank	0.000	20:38	

Data Download Date/Time: 09/26/2022 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: DE PAZ RIZO

DATE/TIME OF TEST: 09/01/2022 20:47:21

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.080	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.186	20:42
Breath Volume	1.859	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.191	20:45
Breath Volume	1.878	20:45
Air Blank	0.000	20:46
Control Test	0.083	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 09/26/2022 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ORELLANA LICONA

DATE/TIME OF TEST: 09/02/2022 01:31:08

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.081	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.181	01:26
Breath Volume	2.867	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:28
Subject Sample # 2	0.191	01:29
Breath Volume	1.941	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 09/26/2022 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 09/17/2022 03:14:34

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.077	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.105	03:09
Breath Volume	2.167	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.106	03:13
Breath Volume	1.335	03:13
Air Blank	0.000	03:13
Control Test	0.076	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 09/22/2022 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 09/18/2022 00:35:22

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:32	
Control Test	0.077	00:32	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:34	Subject Test Refused
Breath Volume	0.000	00:34	
Air Blank	0.000	00:34	
Control Test	0.077	00:34	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 09/22/2022 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 09/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 09/30/2022 21:59:28

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:57	
Control Test	0.078	21:57	
Air Blank	0.000	21:57	
Subject Sample	# 1	21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:58	
Control Test	0.078	21:58	
Air Blank	0.000	21:59	
Diagnostics Check	OK	21:59	

Data Download Date/Time: 10/19/2022 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022365
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 09/03/2022 03:32:42

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.170	03:27
Breath Volume	1.804	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.177	03:31
Breath Volume	3.593	03:31
Air Blank	0.000	03:31
Control Test	0.077	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 09/26/2022 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022365
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: JUAREZ MARTINEZ

DATE/TIME OF TEST: 09/05/2022 04:14:55

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.222	04:10
Breath Volume	2.691	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.218	04:13
Breath Volume	2.792	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 09/26/2022 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022365
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: CRISP

DATE/TIME OF TEST: 09/08/2022 04:35:56

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.156	04:30
Breath Volume	1.929	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:32
Subject Sample # 2	0.147	04:34
Breath Volume	1.414	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 09/26/2022 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JIMENEZ, WILLIAM,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JIMENEZ, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022365
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: KAY

DATE/TIME OF TEST: 09/09/2022 15:09:44

Diagnostics Check	OK	15:02
Air Blank	0.000	15:03
Control Test	0.080	15:03
Air Blank	0.000	15:04
Subject Sample # 1	0.000	15:05
Breath Volume	2.722	15:05
Air Blank	0.000	15:05
Air Blank	0.000	15:07
Subject Sample # 2	0.000	15:08
Breath Volume	2.183	15:08
Air Blank	0.000	15:08
Control Test	0.078	15:09
Air Blank	0.000	15:09
Diagnostics Check	OK	15:09

Data Download Date/Time: 09/26/2022 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATES,BRIAN,A
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GATES,BRIAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022365
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/24/2022 23:31:06

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:25	
Control Test	0.080	23:25	
Air Blank	0.000	23:26	
Subject Sample	# 1 0.196	23:26	
Breath Volume	3.160	23:26	
Air Blank	0.000	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 2 0.181	23:29	
Breath Volume	3.390	23:29	
Air Blank	0.000	23:30	
Control Test	0.069	23:30	Control Outside Tolerance
Air Blank	0.000	23:31	

Data Download Date/Time: 09/26/2022 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATES,BRIAN,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: WILSON,MATT
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022365
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/25/2022 04:44:18

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.175	04:39
Breath Volume	1.285	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:41
Subject Sample # 2	0.178	04:42
Breath Volume	1.683	04:42
Air Blank	0.000	04:43
Control Test	0.077	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 09/26/2022 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CENTENO COLORADO

DATE/TIME OF TEST: 09/04/2022 01:24:11

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.083	01:21	
Air Blank	0.000	01:21	
Subject Sample # 1	0.205	01:22	
Breath Volume	2.570	01:22	
Air Blank		01:22	RFI Detect
Air Blank	0.000	01:23	
Control Test	0.081	01:23	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 09/26/2022 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CENTENO COLORADO

DATE/TIME OF TEST: 09/04/2022 01:29:31

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:28	
Control Test	0.081	01:28	
Air Blank		01:28	Ambient Fail
Air Blank	0.000	01:29	

Data Download Date/Time: 09/26/2022 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CENTENO COLORADO

DATE/TIME OF TEST: 09/04/2022 01:38:47

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.080	01:33	
Air Blank	0.000	01:33	
Subject Sample # 1	0.200	01:34	
Breath Volume	1.699	01:34	
Air Blank	0.000	01:35	
Air Blank	0.000	01:36	
Subject Sample # 2	0.209	01:37	
Breath Volume	1.812	01:37	
Air Blank	0.000	01:37	
Control Test	0.080	01:38	Interferent Detect
Air Blank	0.000	01:38	

Data Download Date/Time: 09/26/2022 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RUBEL,NEIL,F

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: RUBEL,NEIL

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1327124

DRY GAS EXPIRY DATE: 05/14/2023

VIOLATION CODE: DUI

ARREST TIME: 00:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DININNO

DATE/TIME OF TEST: 09/15/2022 01:24:20

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.078	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Control Test	0.077	01:23	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 09/22/2022 18:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1327124
BREATH TEST OPERATOR: DEW,REECE,K	DRY GAS EXPIRY DATE: 05/14/2023
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: DEW,REECE	ARREST TIME: 23:08
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Male
LAST AGENCY INSPECTION: 08/09/2022	OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: KINSEY DATE/TIME OF TEST: 09/16/2022 02:17:30

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.000	02:13
Breath Volume	1.382	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.000	02:16
Breath Volume	1.558	02:16
Air Blank	0.000	02:16
Control Test	0.078	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 09/22/2022 18:03

Subject Test Electronic Data

DATE/TIME OF TEST: 09/16/2022 23:54:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 08/28/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 202303

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:04

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: PUTMAN

DATE/TIME OF TEST: 09/10/2022 02:17:48

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.170	02:13
Breath Volume	1.429	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:15
Subject Sample # 2	0.164	02:16
Breath Volume	1.570	02:16
Air Blank	0.000	02:16
Control Test	0.077	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 10/05/2022 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TERUZ,GABRIEL,S

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: TERUZ,GABRIEL

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 3024023571

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 02:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GERBIER

DATE/TIME OF TEST: 09/03/2022 03:12:17

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.082	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.175	03:07
Breath Volume	1.703	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.159	03:10
Breath Volume	2.714	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 09/26/2022 14:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 09/03/2022 04:44:24

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.080	04:42	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:43	
Control Test	0.078	04:43	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 10/06/2022 05:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 09/03/2022 04:51:05

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.078	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:49	Subject Test Refused
Breath Volume	0.000	04:49	
Air Blank	0.000	04:50	
Control Test	0.078	04:50	
Air Blank	0.000	04:51	
Diagnostics Check	OK	04:51	

Data Download Date/Time: 10/06/2022 05:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: VALENZUELA

DATE/TIME OF TEST: 09/04/2022 20:49:53

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.078	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.135	20:45
Breath Volume	2.035	20:45
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.137	20:48
Breath Volume	1.796	20:48
Air Blank	0.000	20:48
Control Test	0.078	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 10/06/2022 05:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LENBIEN,L
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: CANALAS ROMEO

DATE/TIME OF TEST: 09/05/2022 23:04:54

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.175	22:59
Breath Volume	2.132	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.183	23:03
Breath Volume	2.703	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 10/06/2022 05:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 09/11/2022 03:24:10

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.078	03:22	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Control Test	0.079	03:23	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 10/06/2022 05:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTAGLIA,P
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: ZELEDON

DATE/TIME OF TEST: 09/11/2022 06:16:52

Diagnostics Check	OK	06:10	
Air Blank	0.000	06:10	
Control Test	0.078	06:11	
Air Blank	0.000	06:11	
Subject Sample # 1	0.140	06:12	
Breath Volume	1.476	06:12	
Air Blank	0.000	06:12	
Air Blank	0.000	06:14	
Subject Sample # 2	0.142	06:15	
Breath Volume	1.203	06:15	
Air Blank	0.000	06:15	
Control Test	0.078	06:16	
Air Blank		06:16	RFI Detect
Air Blank	0.000	06:16	

Data Download Date/Time: 10/06/2022 05:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTAGLIA,P
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: ZELEDON

DATE/TIME OF TEST: 09/11/2022 06:26:14

Diagnostics Check	OK	06:19
Air Blank	0.000	06:20
Control Test	0.078	06:20
Air Blank	0.000	06:21
Subject Sample # 1	0.126	06:21
Breath Volume	1.519	06:21
Air Blank	0.000	06:22
Air Blank	0.000	06:24
Subject Sample # 2	0.136	06:24
Breath Volume	1.476	06:24
Air Blank	0.000	06:25
Control Test	0.079	06:25
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

Data Download Date/Time: 10/06/2022 05:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/15/2022 05:39:02

Diagnostics Check	OK	05:32
Air Blank	0.000	05:33
Control Test	0.079	05:33
Air Blank	0.000	05:34
Subject Sample # 1	0.000	05:34
Breath Volume	3.261	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 2	0.000	05:37
Breath Volume	3.375	05:37
Air Blank	0.000	05:38
Control Test	0.079	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:39

Data Download Date/Time: 10/06/2022 05:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MEDINA PENA

DATE/TIME OF TEST: 09/25/2022 00:23:54

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.078	00:12	
Air Blank	0.000	00:12	
Subject Sample	# 1 0.196	00:15	Volume Not Met
Breath Volume	0.476	00:15	
Air Blank	0.000	00:16	
Air Blank	0.000	00:18	
Subject Sample	# 2 0.201	00:19	No .020 agreement
Breath Volume	2.460	00:19	
Air Blank	0.000	00:19	
Air Blank	0.000	00:21	
Subject Sample	# 3 0.160	00:22	No .020 agreement
Breath Volume	1.207	00:22	
Air Blank	0.000	00:22	
Control Test	0.076	00:23	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 10/06/2022 05:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MEDINA PENA

DATE/TIME OF TEST: 09/25/2022 00:34:57

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.078	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.197	00:30
Breath Volume	2.714	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.215	00:33
Breath Volume	1.578	00:33
Air Blank	0.000	00:34
Control Test	0.077	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 10/06/2022 05:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HILDALGO,K
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/25/2022 04:44:05

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.078	04:41	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:42	Subject Test Refused
Breath Volume	0.000	04:42	
Air Blank	0.000	04:43	
Control Test	0.078	04:43	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 10/06/2022 05:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MULLEN,STEVE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: ROBSON

DATE/TIME OF TEST: 09/01/2022 20:06:47

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.079	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.240	20:02
Breath Volume	1.765	20:02
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.236	20:05
Breath Volume	1.914	20:05
Air Blank	0.000	20:05
Control Test	0.079	20:06
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 09/04/2022 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: MCNAMARA

DATE/TIME OF TEST: 09/03/2022 05:44:05

Diagnostics Check	OK	05:39	
Air Blank	0.000	05:39	
Control Test	0.081	05:39	
Air Blank	0.000	05:40	
Subject Sample	# 1	05:42	RFI Detect
Breath Volume	0.015	05:42	
Air Blank	0.000	05:43	
Control Test	0.079	05:43	
Air Blank	0.000	05:44	
Diagnostics Check	OK	05:44	

Data Download Date/Time: 09/04/2022 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: MCNAMARA

DATE/TIME OF TEST: 09/03/2022 05:54:21

Diagnostics Check	OK	05:46
Air Blank	0.000	05:46
Control Test	0.080	05:46
Air Blank	0.000	05:47
Subject Sample # 1	0.117	05:48
Breath Volume	1.984	05:48
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.115	05:52
Breath Volume	1.957	05:52
Air Blank	0.000	05:53
Control Test	0.080	05:53
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

Data Download Date/Time: 09/04/2022 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUC,NORBENS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 09/03/2022 06:31:49

Diagnostics Check	OK	06:25
Air Blank	0.000	06:25
Control Test	0.077	06:26
Air Blank	0.000	06:26
Subject Sample # 1	0.190	06:27
Breath Volume	1.839	06:27
Air Blank	0.000	06:27
Air Blank	0.000	06:29
Subject Sample # 2	0.183	06:30
Breath Volume	1.371	06:30
Air Blank	0.000	06:30
Control Test	0.076	06:31
Air Blank	0.000	06:31
Diagnostics Check	OK	06:31

Data Download Date/Time: 09/04/2022 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALSHAIER,ELIAS
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/04/2022 03:55:31

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.228	03:50
Breath Volume	2.234	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.225	03:53
Breath Volume	2.484	03:53
Air Blank	0.000	03:54
Control Test	0.078	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 10/05/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: BONILLA PAZ

DATE/TIME OF TEST: 09/04/2022 18:30:10

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.080	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.177	18:25
Breath Volume	2.062	18:25
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample # 2	0.164	18:28
Breath Volume	2.082	18:28
Air Blank	0.000	18:29
Control Test	0.080	18:29
Air Blank	0.000	18:30
Diagnostics Check	OK	18:30

Data Download Date/Time: 10/05/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAMMAD,MARINA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: BUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: TOSTI

DATE/TIME OF TEST: 09/04/2022 19:53:14

Diagnostics Check	OK	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.069	19:48
Breath Volume	3.718	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.070	19:51
Breath Volume	2.851	19:51
Air Blank	0.000	19:52
Control Test	0.081	19:52
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 10/05/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: FERNANDEZ MONTERREY

DATE/TIME OF TEST: 09/04/2022 22:54:11

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.079	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.000	22:49
Breath Volume	2.613	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.000	22:52
Breath Volume	2.664	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 10/05/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ENGEL,ANTHONY
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 09/05/2022 07:45:02

Diagnostics Check	OK	07:38
Air Blank	0.000	07:39
Control Test	0.080	07:39
Air Blank	0.000	07:39
Subject Sample # 1	0.145	07:40
Breath Volume	3.187	07:40
Air Blank	0.000	07:41
Air Blank	0.000	07:42
Subject Sample # 2	0.145	07:43
Breath Volume	2.105	07:43
Air Blank	0.000	07:44
Control Test	0.081	07:44
Air Blank	0.000	07:44
Diagnostics Check	OK	07:45

Data Download Date/Time: 10/05/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HALES,ANDREW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: KINCHELOE

DATE/TIME OF TEST: 09/09/2022 22:59:09

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.115	22:54
Breath Volume	1.359	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.116	22:57
Breath Volume	2.292	22:57
Air Blank	0.000	22:58
Control Test	0.078	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 10/05/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCCOLLIN-DAVIS,DEVON
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 11:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:53

SUBJECT'S LAST NAME: ROCHA

DATE/TIME OF TEST: 09/10/2022 13:23:48

Diagnostics Check	OK	13:17
Air Blank	0.000	13:18
Control Test	0.080	13:18
Air Blank	0.000	13:18
Subject Sample # 1	0.000	13:19
Breath Volume	2.847	13:19
Air Blank	0.000	13:19
Air Blank	0.000	13:21
Subject Sample # 2	0.000	13:22
Breath Volume	2.847	13:22
Air Blank	0.000	13:22
Control Test	0.081	13:23
Air Blank	0.000	13:23
Diagnostics Check	OK	13:23

Data Download Date/Time: 10/05/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GREEN,SAMANTHA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: MONSEAU

DATE/TIME OF TEST: 09/11/2022 18:34:29

Diagnostics Check	OK	18:28
Air Blank	0.000	18:28
Control Test	0.081	18:28
Air Blank	0.000	18:29
Subject Sample # 1	0.232	18:29
Breath Volume	3.066	18:29
Air Blank	0.000	18:30
Air Blank	0.000	18:32
Subject Sample # 2	0.217	18:32
Breath Volume	2.359	18:32
Air Blank	0.000	18:33
Control Test	0.078	18:33
Air Blank	0.000	18:34
Diagnostics Check	OK	18:34

Data Download Date/Time: 10/05/2022 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MAJOR,BERNARD
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: BALLESTEROS DOMINGUE

DATE/TIME OF TEST: 09/11/2022 22:56:44

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.087	22:52
Breath Volume	1.566	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:54
Subject Sample # 2	0.078	22:55
Breath Volume	2.066	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 10/05/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALSHAIER,ELIAS
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MELVILLE

DATE/TIME OF TEST: 09/12/2022 00:45:37

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.078	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.165	00:40
Breath Volume	1.761	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:42
Subject Sample # 2	0.151	00:44
Breath Volume	1.164	00:44
Air Blank	0.000	00:44
Control Test	0.077	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 10/05/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KATES PAULUS,DEANO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ALBERT

DATE/TIME OF TEST: 09/21/2022 03:22:51

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.147	03:18
Breath Volume	1.714	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample # 2	0.144	03:21
Breath Volume	1.503	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 10/05/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 09/21/2022 04:11:50

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.166	04:07
Breath Volume	1.468	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.169	04:10
Breath Volume	1.652	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 10/05/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: REBOYAL

DATE/TIME OF TEST: 09/23/2022 23:17:11

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.078	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.191	23:12
Breath Volume	2.171	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.183	23:15
Breath Volume	2.574	23:15
Air Blank	0.000	23:16
Control Test	0.077	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 10/05/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DOVI,FRANK
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MORALES TORRES

DATE/TIME OF TEST: 09/24/2022 01:18:31

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.078	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.275	01:13
Breath Volume	2.863	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.268	01:16
Breath Volume	2.875	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 10/05/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WALKER,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MENDOZA GUILLEN

DATE/TIME OF TEST: 09/24/2022 02:03:56

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.079	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.242	01:59
Breath Volume	1.347	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.237	02:02
Breath Volume	2.066	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 10/05/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: GEORGES

DATE/TIME OF TEST: 09/24/2022 20:13:41

Diagnostics Check	OK	20:04
Air Blank	0.000	20:05
Control Test	0.079	20:05
Air Blank	0.000	20:06
Subject Sample # 1	0.188	20:06
Breath Volume	1.285	20:06
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.179	20:11
Breath Volume	2.406	20:11
Air Blank	0.000	20:12
Control Test	0.081	20:13
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 10/05/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 09/25/2022 03:34:39

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.254	03:29
Breath Volume	1.449	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.246	03:32
Breath Volume	2.042	03:32
Air Blank	0.000	03:33
Control Test	0.077	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 10/05/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHANIEL
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: NEPTUNE

DATE/TIME OF TEST: 09/25/2022 04:40:14

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:28	
Control Test	0.079	04:29	
Air Blank	0.000	04:29	
Subject Sample	# 1 0.153	04:32	Volume Not Met
Breath Volume	0.156	04:32	
Air Blank	0.000	04:33	
Air Blank	0.000	04:35	
Subject Sample	# 2 0.190	04:38	Volume Not Met
Breath Volume	0.066	04:38	
Air Blank	0.000	04:39	
Control Test	0.083	04:39	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 10/05/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RUSSELL,AMON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: DIAZ GONZALEZ

DATE/TIME OF TEST: 09/25/2022 06:18:16

Diagnostics Check	OK	06:11
Air Blank	0.000	06:12
Control Test	0.079	06:12
Air Blank	0.000	06:13
Subject Sample # 1	0.271	06:13
Breath Volume	2.289	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:16
Subject Sample # 2	0.267	06:16
Breath Volume	2.578	06:16
Air Blank	0.000	06:17
Control Test	0.078	06:17
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 10/05/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: YOUNG,J
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: LLOVET

DATE/TIME OF TEST: 09/30/2022 23:10:39

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:03
Air Blank	0.000	23:04
Subject Sample # 1	0.165	23:05
Breath Volume	2.109	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.161	23:09
Breath Volume	1.222	23:09
Air Blank	0.000	23:09
Control Test	0.079	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 10/05/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ALEX,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: YOUNG,ALEX
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 09/17/2022 05:43:05

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:40	
Control Test	0.080	05:41	
Air Blank	0.000	05:41	
Subject Sample	# 1	05:41	Subject Test Refused
Breath Volume	0.000	05:41	
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 09/28/2022 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: GLADIN

DATE/TIME OF TEST: 09/29/2022 23:51:56

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.081	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.150	23:47
Breath Volume	3.226	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.142	23:50
Breath Volume	3.328	23:50
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 10/24/2022 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: TOLLEY

DATE/TIME OF TEST: 09/30/2022 03:10:44

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.081	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 2	03:08	Volume Not Met
Breath Volume	0.019	03:08	
Air Blank	0.000	03:09	
Control Test	0.080	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 10/24/2022 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: TOLLEY

DATE/TIME OF TEST: 09/30/2022 03:20:21

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.140	03:14
Breath Volume	1.671	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.135	03:18
Breath Volume	1.496	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 10/24/2022 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIKES,ROBERT
ARRESTING OFFICER AGENCY: Parker PD
LAST AGENCY INSPECTION: 09/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: SIMONTON

DATE/TIME OF TEST: 09/30/2022 05:09:29

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.113	05:04
Breath Volume	2.695	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.096	05:07
Breath Volume	1.769	05:07
Air Blank	0.000	05:08
Control Test	0.082	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 10/24/2022 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 09/30/2022 06:15:19

Diagnostics Check	OK	06:08
Air Blank	0.000	06:08
Control Test	0.082	06:09
Air Blank	0.000	06:09
Subject Sample # 1	0.173	06:10
Breath Volume	2.394	06:10
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample # 2	0.176	06:13
Breath Volume	2.332	06:13
Air Blank	0.000	06:14
Control Test	0.081	06:14
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

Data Download Date/Time: 10/24/2022 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHM,JONATHAN,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: BOHM,BOHM
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MONBERG

DATE/TIME OF TEST: 09/02/2022 01:46:55

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.203	01:42
Breath Volume	2.777	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.198	01:45
Breath Volume	2.605	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 09/21/2022 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: PECHOVA

DATE/TIME OF TEST: 09/14/2022 00:21:10

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.077	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.191	00:16
Breath Volume	1.750	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.183	00:19
Breath Volume	1.542	00:19
Air Blank	0.000	00:20
Control Test	0.077	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 09/26/2022 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 09/02/2022 01:26:37

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.154	01:21
Breath Volume	2.363	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.158	01:24
Breath Volume	2.542	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 09/21/2022 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: SATTERTHWAITE

DATE/TIME OF TEST: 09/08/2022 19:55:41

Diagnostics Check	OK	19:38	
Air Blank	0.000	19:39	
Control Test	0.079	19:39	
Air Blank	0.000	19:40	
Subject Sample	# 1	19:43	No Sample Provided
Breath Volume	0.000	19:43	
Air Blank	0.000	19:44	
Air Blank	0.000	19:45	
Subject Sample	# 2	0.232	19:48
Breath Volume	1.167	19:48	
Air Blank	0.000	19:48	
Air Blank	0.000	19:50	
Subject Sample	# 3	0.223	19:53
Breath Volume	0.015	19:53	Volume Not Met
Air Blank	0.000	19:54	
Control Test	0.077	19:54	
Air Blank	0.000	19:55	
Diagnostics Check	OK	19:55	

Data Download Date/Time: 09/21/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER\JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 09/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: M216010640195

DATE/TIME OF TEST: 09/22/2022 23:45:44

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:30	
Control Test	0.079	23:31	
Air Blank	0.000	23:31	
Subject Sample	# 1	23:34	No Sample Provided
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample	# 2	23:38	
Breath Volume	1.628	23:38	
Air Blank	0.000	23:39	
Air Blank	0.000	23:41	
Subject Sample	# 3	23:44	Volume Not Met
Breath Volume	0.328	23:44	
Air Blank	0.000	23:44	
Control Test	0.081	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 10/25/2022 01:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLIARD,M,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SHERAN,,
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 09/04/2022 07:05:55

Diagnostics Check	Fail	07:05	Internal Printer Test
Air Blank	0.000	07:05	

Data Download Date/Time: 09/05/2022 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLIARD,M,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SHERAN,,
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 09/04/2022 07:09:35

Diagnostics Check	Fail	07:09	Internal Printer Test
Air Blank	0.000	07:09	

Data Download Date/Time: 09/05/2022 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,E,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BAXTER,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/06/2022 17:55:18

Diagnostics Check	OK	17:48
Air Blank	0.000	17:49
Control Test	0.077	17:49
Air Blank	0.000	17:50
Subject Sample # 1	0.270	17:50
Breath Volume	3.093	17:50
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 2	0.279	17:53
Breath Volume	1.988	17:53
Air Blank	0.000	17:54
Control Test	0.077	17:54
Air Blank	0.000	17:55
Diagnostics Check	OK	17:55

Data Download Date/Time: 10/05/2022 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/09/2022 02:37:22

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.027	02:32
Breath Volume	1.605	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.027	02:35
Breath Volume	1.316	02:35
Air Blank	0.000	02:36
Control Test	0.077	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 10/05/2022 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,S,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BENITEZ,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 09/10/2022 00:26:06

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.077	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.200	00:20
Breath Volume	1.640	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.201	00:24
Breath Volume	1.695	00:24
Air Blank	0.000	00:25
Control Test	0.077	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 10/05/2022 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDS,FELICIA,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: JOHNSON,TONY
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SNEDEKER

DATE/TIME OF TEST: 09/10/2022 01:50:39

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.126	01:44
Breath Volume	2.488	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.111	01:48
Breath Volume	1.500	01:48
Air Blank	0.000	01:49
Control Test	0.076	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 10/05/2022 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL, ADAM, J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PURSLEY, B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SHERIDAN

DATE/TIME OF TEST: 09/10/2022 03:11:50

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.078	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.160	03:07
Breath Volume	2.746	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.155	03:10
Breath Volume	1.503	03:10
Air Blank	0.000	03:10
Control Test	0.077	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 10/05/2022 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSIL,LUCAS,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SUEROLANTIGUA,J
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:29

SUBJECT'S LAST NAME: YUMANCULEX

DATE/TIME OF TEST: 09/11/2022 13:04:48

Diagnostics Check	OK	12:58
Air Blank	0.000	12:58
Control Test	0.079	12:58
Air Blank	0.000	12:59
Subject Sample # 1	0.187	12:59
Breath Volume	2.394	12:59
Air Blank	0.000	13:00
Air Blank	0.000	13:02
Subject Sample # 2	0.184	13:03
Breath Volume	3.296	13:03
Air Blank	0.000	13:03
Control Test	0.078	13:04
Air Blank	0.000	13:04
Diagnostics Check	OK	13:04

Data Download Date/Time: 10/05/2022 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,AUSTIN,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: NUNO

DATE/TIME OF TEST: 09/13/2022 03:41:27

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.077	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.056	03:36
Breath Volume	2.472	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.055	03:39
Breath Volume	4.015	03:39
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASALA,J,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: VALENTIN,E
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ALCIME

DATE/TIME OF TEST: 09/13/2022 23:57:21

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.078	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.204	23:52
Breath Volume	2.609	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.203	23:55
Breath Volume	2.113	23:55
Air Blank	0.000	23:56
Control Test	0.076	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,T,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BURGER,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: WELDON

DATE/TIME OF TEST: 09/17/2022 04:32:12

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.189	04:27
Breath Volume	2.738	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.194	04:30
Breath Volume	2.421	04:30
Air Blank	0.000	04:31
Control Test	0.077	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,DANIEL,S
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GRAY,J
ARRESTING OFFICER AGENCY: MARCO ISLAND PD
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/18/2022 14:55:59

Diagnostics Check	OK	14:49
Air Blank	0.000	14:49
Control Test	0.078	14:50
Air Blank	0.000	14:50
Subject Sample # 1	0.000	14:51
Breath Volume	2.332	14:51
Air Blank	0.000	14:52
Air Blank	0.000	14:54
Subject Sample # 2	0.000	14:54
Breath Volume	2.023	14:54
Air Blank	0.000	14:55
Control Test	0.078	14:55
Air Blank	0.000	14:55
Diagnostics Check	OK	14:55

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GOODMAN,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: GODINEZ-SANCHEZ

DATE/TIME OF TEST: 09/18/2022 20:38:08

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:26	
Control Test	0.079	20:26	
Air Blank	0.000	20:27	
Subject Sample	# 1 0.217	20:30	Volume Not Met
Breath Volume	0.070	20:30	
Air Blank	0.000	20:31	
Air Blank	0.000	20:32	
Subject Sample	# 2 0.213	20:33	
Breath Volume	1.707	20:33	
Air Blank	0.000	20:34	
Air Blank	0.000	20:36	
Subject Sample	# 3 0.214	20:36	
Breath Volume	1.699	20:36	
Air Blank	0.000	20:37	
Control Test	0.076	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: SCHUETT

DATE/TIME OF TEST: 09/18/2022 20:49:10

Diagnostics Check	OK	20:42
Air Blank	0.000	20:42
Control Test	0.078	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.159	20:44
Breath Volume	2.480	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:46
Subject Sample # 2	0.161	20:47
Breath Volume	2.917	20:47
Air Blank	0.000	20:48
Control Test	0.077	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,A,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MERCADO,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: SHEAR

DATE/TIME OF TEST: 09/20/2022 09:22:10

Diagnostics Check	OK	09:15
Air Blank	0.000	09:16
Control Test	0.078	09:16
Air Blank	0.000	09:17
Subject Sample # 1	0.000	09:17
Breath Volume	1.757	09:17
Air Blank	0.000	09:18
Air Blank	0.000	09:20
Subject Sample # 2	0.000	09:20
Breath Volume	1.515	09:20
Air Blank	0.000	09:21
Control Test	0.078	09:21
Air Blank	0.000	09:22
Diagnostics Check	OK	09:22

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LUKASZ,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: WEISENBACH

DATE/TIME OF TEST: 09/20/2022 21:50:02

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.078	21:44
Air Blank	0.000	21:44
Subject Sample # 1	0.297	21:45
Breath Volume	3.707	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:47
Subject Sample # 2	0.295	21:48
Breath Volume	3.855	21:48
Air Blank	0.000	21:49
Control Test	0.076	21:49
Air Blank	0.000	21:49
Diagnostics Check	OK	21:50

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,DANIEL,E
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: VILLANUEVA,H
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: GLOD

DATE/TIME OF TEST: 09/20/2022 23:15:11

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.078	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.077	23:10
Breath Volume	3.824	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:12
Subject Sample # 2	0.071	23:13
Breath Volume	1.804	23:13
Air Blank	0.000	23:14
Control Test	0.078	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: RAMOS,L
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MATAMOROS

DATE/TIME OF TEST: 09/21/2022 00:03:37

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.078	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.167	23:57
Breath Volume	1.871	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.155	00:01
Breath Volume	1.839	00:01
Air Blank	0.000	00:02
Control Test	0.075	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LESNIEWSKI,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: ROSALES TURCIOS

DATE/TIME OF TEST: 09/23/2022 01:57:19

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.230	01:52
Breath Volume	2.800	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:54
Subject Sample # 2	0.226	01:55
Breath Volume	2.808	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,J.,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GASH,C.
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: OLVERA TREJO

DATE/TIME OF TEST: 09/24/2022 22:53:10

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.078	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.158	22:48
Breath Volume	1.187	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:50
Subject Sample # 2	0.150	22:51
Breath Volume	1.242	22:51
Air Blank	0.000	22:52
Control Test	0.077	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA,R,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SHERAN,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: ROMERO HERRERA

DATE/TIME OF TEST: 09/27/2022 01:00:18

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.079	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.198	00:55
Breath Volume	1.570	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample # 2	0.216	00:58
Breath Volume	3.347	00:58
Air Blank	0.000	00:59
Control Test	0.078	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 10/05/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,ERIC,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: KORPOLINSKI

DATE/TIME OF TEST: 09/30/2022 21:27:45

Diagnostics Check	OK	21:21
Air Blank	0.000	21:22
Control Test	0.078	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.125	21:23
Breath Volume	1.746	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.130	21:26
Breath Volume	2.167	21:26
Air Blank	0.000	21:26
Control Test	0.078	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 10/05/2022 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGGINS,M,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HUGGINS,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/15/2022 00:05:20

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.148	23:59
Breath Volume	1.175	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.148	00:03
Breath Volume	1.593	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 10/06/2022 04:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,C,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINION,C
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: GARCES COLINA

DATE/TIME OF TEST: 09/20/2022 08:21:26

Diagnostics Check	OK	08:13
Air Blank	0.000	08:14
Control Test	0.077	08:14
Air Blank	0.000	08:15
Subject Sample # 1	0.163	08:16
Breath Volume	2.242	08:16
Air Blank	0.000	08:17
Air Blank	0.000	08:18
Subject Sample # 2	0.153	08:19
Breath Volume	1.937	08:19
Air Blank	0.000	08:20
Control Test	0.077	08:20
Air Blank	0.000	08:21
Diagnostics Check	OK	08:21

Data Download Date/Time: 10/06/2022 04:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AGUILAR,A
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SALES ORDONEZ

DATE/TIME OF TEST: 09/22/2022 01:52:02

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.082	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Control Test	0.080	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 10/06/2022 04:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMONTE,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: SALAS RABANALES

DATE/TIME OF TEST: 09/01/2022 16:42:54

Diagnostics Check	OK	16:35
Air Blank	0.000	16:35
Control Test	0.077	16:35
Air Blank	0.000	16:36
Subject Sample # 1	0.200	16:38
Breath Volume	1.132	16:38
Air Blank	0.000	16:39
Air Blank	0.000	16:40
Subject Sample # 2	0.181	16:41
Breath Volume	1.117	16:41
Air Blank	0.000	16:41
Control Test	0.076	16:42
Air Blank	0.000	16:42
Diagnostics Check	OK	16:42

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: GOMEZ HERNANDEZ

DATE/TIME OF TEST: 09/01/2022 22:57:40

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:49	
Control Test	0.077	22:50	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:53	No Sample Provided
Breath Volume	0.000	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 2	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:56	
Control Test	0.077	22:57	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: CASTRO-RIVERA

DATE/TIME OF TEST: 09/02/2022 03:40:04

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.077	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.189	03:35
Breath Volume	2.156	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.197	03:38
Breath Volume	2.406	03:38
Air Blank	0.000	03:39
Control Test	0.076	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:40

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIDSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 09/02/2022 15:54:36

Diagnostics Check	OK	15:48
Air Blank	0.000	15:48
Control Test	0.078	15:49
Air Blank	0.000	15:49
Subject Sample # 1	0.000	15:50
Breath Volume	4.593	15:50
Air Blank	0.000	15:50
Air Blank	0.000	15:52
Subject Sample # 2	0.000	15:53
Breath Volume	3.089	15:53
Air Blank	0.000	15:53
Control Test	0.078	15:54
Air Blank	0.000	15:54
Diagnostics Check	OK	15:54

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PASS,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: BUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: GETZ

DATE/TIME OF TEST: 09/02/2022 17:19:35

Diagnostics Check	OK	17:13
Air Blank	0.000	17:13
Control Test	0.078	17:14
Air Blank	0.000	17:14
Subject Sample # 1	0.065	17:15
Breath Volume	2.429	17:15
Air Blank	0.000	17:15
Air Blank	0.000	17:17
Subject Sample # 2	0.063	17:18
Breath Volume	2.554	17:18
Air Blank	0.000	17:18
Control Test	0.078	17:18
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 09/02/2022 21:37:52

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.077	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.202	21:33
Breath Volume	1.578	21:33
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.197	21:36
Breath Volume	1.632	21:36
Air Blank	0.000	21:36
Control Test	0.079	21:37
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SQUIBBS

DATE/TIME OF TEST: 09/03/2022 02:07:58

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.077	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.206	02:03
Breath Volume	2.105	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:05
Subject Sample # 2	0.205	02:06
Breath Volume	2.027	02:06
Air Blank	0.000	02:07
Control Test	0.077	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GRECO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 09/03/2022 05:01:19

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.078	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.132	04:56
Breath Volume	1.675	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 2	0.124	04:59
Breath Volume	1.726	04:59
Air Blank	0.000	05:00
Control Test	0.077	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 09/03/2022 05:40:53

Diagnostics Check	OK	05:33
Air Blank	0.000	05:34
Control Test	0.077	05:34
Air Blank	0.000	05:34
Subject Sample # 1	0.111	05:36
Breath Volume	2.929	05:36
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.112	05:39
Breath Volume	3.101	05:39
Air Blank	0.000	05:39
Control Test	0.077	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: ANSELM

DATE/TIME OF TEST: 09/03/2022 06:43:40

Diagnostics Check	OK	06:36
Air Blank	0.000	06:36
Control Test	0.077	06:37
Air Blank	0.000	06:37
Subject Sample # 1	0.226	06:38
Breath Volume	1.894	06:38
Air Blank	0.000	06:39
Air Blank	0.000	06:41
Subject Sample # 2	0.218	06:42
Breath Volume	1.648	06:42
Air Blank	0.000	06:42
Control Test	0.076	06:43
Air Blank	0.000	06:43
Diagnostics Check	OK	06:43

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: ELKINS SR

DATE/TIME OF TEST: 09/03/2022 22:21:31

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.077	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.209	22:16
Breath Volume	2.351	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.204	22:19
Breath Volume	2.027	22:19
Air Blank	0.000	22:20
Control Test	0.077	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY, JANICE, F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GEERTSEMA

DATE/TIME OF TEST: 09/04/2022 00:19:30

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.078	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.126	00:13
Breath Volume	1.488	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.122	00:17
Breath Volume	1.421	00:17
Air Blank	0.000	00:18
Control Test	0.077	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/04/2022 00:59:10

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.077	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.207	00:54
Breath Volume	2.488	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:56
Subject Sample # 2	0.205	00:57
Breath Volume	2.261	00:57
Air Blank	0.000	00:58
Control Test	0.077	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/04/2022 05:27:09

Diagnostics Check	OK	05:18
Air Blank	0.000	05:19
Control Test	0.077	05:19
Air Blank	0.000	05:20
Subject Sample # 1	0.182	05:22
Breath Volume	1.949	05:22
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.189	05:25
Breath Volume	1.769	05:25
Air Blank	0.000	05:26
Control Test	0.077	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: REYES CORTES

DATE/TIME OF TEST: 09/04/2022 19:04:41

Diagnostics Check	OK	19:01	
Air Blank	0.000	19:01	
Control Test	0.077	19:01	
Air Blank	0.000	19:02	
Subject Sample	# 1	19:03	Subject Test Refused
Breath Volume	0.000	19:03	
Air Blank	0.000	19:03	
Control Test	0.077	19:04	
Air Blank	0.000	19:04	
Diagnostics Check	OK	19:04	

Data Download Date/Time: 10/13/2022 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: SAMERI-TALEGHANI

DATE/TIME OF TEST: 09/04/2022 20:19:24

Diagnostics Check	OK	20:12
Air Blank	0.000	20:12
Control Test	0.078	20:12
Air Blank	0.000	20:13
Subject Sample # 1	0.046	20:14
Breath Volume	2.593	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:16
Subject Sample # 2	0.045	20:17
Breath Volume	2.453	20:17
Air Blank	0.000	20:18
Control Test	0.077	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 10/13/2022 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: RIVAS SALAZAR

DATE/TIME OF TEST: 09/05/2022 01:24:31

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.078	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.134	01:19
Breath Volume	2.332	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.135	01:22
Breath Volume	2.781	01:22
Air Blank	0.000	01:23
Control Test	0.077	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 10/13/2022 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: RODRIGUEZ HERNANDEZ

DATE/TIME OF TEST: 09/05/2022 04:01:01

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.077	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.200	03:56
Breath Volume	1.796	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.183	03:59
Breath Volume	1.148	03:59
Air Blank	0.000	04:00
Control Test	0.076	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:01

Data Download Date/Time: 10/13/2022 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 09/05/2022 22:16:38

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.078	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.173	22:11
Breath Volume	2.343	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.177	22:15
Breath Volume	3.011	22:15
Air Blank	0.000	22:15
Control Test	0.078	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 10/13/2022 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY, JANICE, F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: FERNANDEZ FUENTES

DATE/TIME OF TEST: 09/06/2022 02:23:50

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.078	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.184	02:18
Breath Volume	1.742	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.179	02:22
Breath Volume	1.832	02:22
Air Blank	0.000	02:22
Control Test	0.077	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: SANFILIPO

DATE/TIME OF TEST: 09/06/2022 22:01:35

Diagnostics Check	OK	21:54
Air Blank	0.000	21:55
Control Test	0.078	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.145	21:56
Breath Volume	1.910	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.156	22:00
Breath Volume	2.257	22:00
Air Blank	0.000	22:00
Control Test	0.078	22:00
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: SIEMERING

DATE/TIME OF TEST: 09/07/2022 04:02:05

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.078	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.153	03:57
Breath Volume	2.234	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 2	0.152	04:00
Breath Volume	2.441	04:00
Air Blank	0.000	04:01
Control Test	0.077	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 09/08/2022 00:44:25

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.197	00:39
Breath Volume	1.347	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.193	00:42
Breath Volume	2.082	00:42
Air Blank	0.000	00:43
Control Test	0.078	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: ADMIN
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 09/08/2022 11:18:46

Diagnostics Check	OK	11:04	
Air Blank	0.000	11:04	
Control Test	0.078	11:05	
Air Blank	0.000	11:05	
Subject Sample	# 1 0.103	11:06	
Breath Volume	3.246	11:06	
Air Blank	0.000	11:06	
Air Blank	0.000	11:08	
Subject Sample	# 2	11:11	No Sample Provided
Breath Volume	0.000	11:11	
Air Blank	0.000	11:12	
Air Blank	0.000	11:14	
Subject Sample	# 3	11:17	No Sample Provided
Breath Volume	0.000	11:17	
Air Blank	0.000	11:17	
Control Test	0.079	11:18	
Air Blank	0.000	11:18	
Diagnostics Check	OK	11:18	

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY, JANICE, F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK, I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 09/09/2022 04:13:43

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:10	
Control Test	0.078	04:11	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Control Test	0.078	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 09/09/2022 23:03:59

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.078	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.163	22:59
Breath Volume	1.652	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:01
Subject Sample # 2	0.161	23:02
Breath Volume	2.019	23:02
Air Blank	0.000	23:03
Control Test	0.077	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAZQUEZ,V
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: VACA

DATE/TIME OF TEST: 09/10/2022 07:24:43

Diagnostics Check	OK	07:18
Air Blank	0.000	07:18
Control Test	0.078	07:19
Air Blank	0.000	07:19
Subject Sample # 1	0.135	07:20
Breath Volume	3.027	07:20
Air Blank	0.000	07:20
Air Blank	0.000	07:22
Subject Sample # 2	0.127	07:23
Breath Volume	2.796	07:23
Air Blank	0.000	07:23
Control Test	0.077	07:24
Air Blank	0.000	07:24
Diagnostics Check	OK	07:24

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: KRAUS PADILLA

DATE/TIME OF TEST: 09/11/2022 02:17:53

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:05	
Control Test	0.078	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1 0.127	02:09	Volume Not Met
Breath Volume	0.378	02:09	
Air Blank	0.000	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 2 0.124	02:12	
Breath Volume	1.871	02:12	
Air Blank	0.000	02:13	
Air Blank	0.000	02:15	
Subject Sample	# 3 0.133	02:16	
Breath Volume	2.093	02:16	
Air Blank	0.000	02:16	
Control Test	0.077	02:17	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: GIDDENS

DATE/TIME OF TEST: 09/11/2022 05:30:45

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.079	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.214	05:25
Breath Volume	1.367	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.221	05:29
Breath Volume	2.511	05:29
Air Blank	0.000	05:29
Control Test	0.077	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: DABROWSKI

DATE/TIME OF TEST: 09/11/2022 20:18:09

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.079	20:12
Air Blank	0.000	20:12
Subject Sample # 1	0.281	20:13
Breath Volume	1.855	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:15
Subject Sample # 2	0.274	20:16
Breath Volume	2.117	20:16
Air Blank	0.000	20:17
Control Test	0.078	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: PAREDES

DATE/TIME OF TEST: 09/12/2022 02:17:44

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.078	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.189	02:13
Breath Volume	3.167	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.182	02:16
Breath Volume	2.226	02:16
Air Blank	0.000	02:16
Control Test	0.077	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: LIPTAK

DATE/TIME OF TEST: 09/12/2022 03:52:52

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.077	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.221	03:48
Breath Volume	1.792	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.210	03:51
Breath Volume	2.253	03:51
Air Blank	0.000	03:51
Control Test	0.078	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: SINKLER

DATE/TIME OF TEST: 09/12/2022 05:40:10

Diagnostics Check	OK	05:33
Air Blank	0.000	05:34
Control Test	0.078	05:34
Air Blank	0.000	05:35
Subject Sample # 1	0.187	05:35
Breath Volume	1.800	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.185	05:38
Breath Volume	1.222	05:38
Air Blank	0.000	05:39
Control Test	0.078	05:39
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CROSBY,J
ARRESTING OFFICER AGENCY: Seminole Res (Hills)
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: JAGERSKY

DATE/TIME OF TEST: 09/12/2022 18:19:50

Diagnostics Check	OK	18:13
Air Blank	0.000	18:13
Control Test	0.077	18:14
Air Blank	0.000	18:14
Subject Sample # 1	0.158	18:15
Breath Volume	2.812	18:15
Air Blank	0.000	18:15
Air Blank	0.000	18:17
Subject Sample # 2	0.161	18:18
Breath Volume	1.847	18:18
Air Blank	0.000	18:18
Control Test	0.078	18:19
Air Blank	0.000	18:19
Diagnostics Check	OK	18:19

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 09/14/2022 03:12:49

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.078	03:02	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:05	No Sample Provided
Breath Volume	0.000	03:05	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample	# 2	03:11	No Sample Provided
Breath Volume	0.000	03:11	
Air Blank	0.000	03:11	
Control Test	0.078	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: PEREZ GOMEZ

DATE/TIME OF TEST: 09/14/2022 05:09:42

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.078	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.227	05:04
Breath Volume	1.398	05:04
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.216	05:08
Breath Volume	1.937	05:08
Air Blank	0.000	05:08
Control Test	0.078	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 09/15/2022 00:36:50

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.078	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.171	00:32
Breath Volume	2.300	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.164	00:35
Breath Volume	1.675	00:35
Air Blank	0.000	00:35
Control Test	0.078	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MERAZ JUAREZ

DATE/TIME OF TEST: 09/15/2022 01:37:03

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.078	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.159	01:32
Breath Volume	1.625	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.158	01:35
Breath Volume	2.449	01:35
Air Blank	0.000	01:36
Control Test	0.076	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:37

Data Download Date/Time: 10/13/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: MERINO MORENO

DATE/TIME OF TEST: 09/15/2022 03:18:55

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.107	03:14
Breath Volume	1.347	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.107	03:17
Breath Volume	1.742	03:17
Air Blank	0.000	03:17
Control Test	0.078	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CASTELLI,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:08

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 09/15/2022 11:35:36

Diagnostics Check	OK	11:28
Air Blank	0.000	11:29
Control Test	0.078	11:29
Air Blank	0.000	11:30
Subject Sample # 1	0.173	11:30
Breath Volume	1.261	11:30
Air Blank	0.000	11:31
Air Blank	0.000	11:33
Subject Sample # 2	0.175	11:34
Breath Volume	2.417	11:34
Air Blank	0.000	11:34
Control Test	0.078	11:34
Air Blank	0.000	11:35
Diagnostics Check	OK	11:35

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: EASTWOOD

DATE/TIME OF TEST: 09/15/2022 22:26:52

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.078	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.071	22:22
Breath Volume	2.250	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:24
Subject Sample # 2	0.065	22:25
Breath Volume	1.179	22:25
Air Blank	0.000	22:25
Control Test	0.077	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 09/16/2022 02:33:56

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.078	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.167	02:29
Breath Volume	2.144	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.174	02:32
Breath Volume	1.429	02:32
Air Blank	0.000	02:33
Control Test	0.077	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RYAN, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: VINCENTE

DATE/TIME OF TEST: 09/17/2022 19:15:41

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.078	19:09
Air Blank	0.000	19:09
Subject Sample # 1	0.183	19:10
Breath Volume	1.781	19:10
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.193	19:13
Breath Volume	1.886	19:13
Air Blank	0.000	19:14
Control Test	0.078	19:15
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 09/17/2022 22:53:57

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.078	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.290	22:49
Breath Volume	2.484	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:51
Subject Sample # 2	0.287	22:52
Breath Volume	2.597	22:52
Air Blank	0.000	22:53
Control Test	0.077	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MISHRA

DATE/TIME OF TEST: 09/18/2022 02:55:50

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.176	02:51
Breath Volume	3.031	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.176	02:54
Breath Volume	2.269	02:54
Air Blank	0.000	02:54
Control Test	0.078	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: GONZALEZ LOPEZ

DATE/TIME OF TEST: 09/18/2022 05:09:12

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.078	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.160	05:03
Breath Volume	3.460	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.145	05:07
Breath Volume	1.593	05:07
Air Blank	0.000	05:08
Control Test	0.077	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: GARCIA AGUILA

DATE/TIME OF TEST: 09/18/2022 05:40:49

Diagnostics Check	OK	05:32	
Air Blank	0.000	05:32	
Control Test	0.078	05:33	
Air Blank	0.000	05:33	
Subject Sample	# 1	05:36	No Sample Provided
Breath Volume	0.000	05:36	
Air Blank	0.000	05:37	
Air Blank	0.000	05:39	
Subject Sample	# 2	05:39	Subject Test Refused
Breath Volume	0.000	05:39	
Air Blank	0.000	05:39	
Control Test	0.077	05:40	
Air Blank	0.000	05:40	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BOUSQUET,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: HOWARD WILSON

DATE/TIME OF TEST: 09/18/2022 19:58:46

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.077	19:51
Air Blank	0.000	19:52
Subject Sample # 1	0.166	19:54
Breath Volume	2.285	19:54
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.164	19:57
Breath Volume	1.945	19:57
Air Blank	0.000	19:57
Control Test	0.077	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: LOPEZ MARTINEZ

DATE/TIME OF TEST: 09/18/2022 20:55:45

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.078	20:48
Air Blank	0.000	20:48
Subject Sample # 1	0.241	20:51
Breath Volume	1.546	20:51
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.227	20:54
Breath Volume	1.835	20:54
Air Blank	0.000	20:54
Control Test	0.076	20:55
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/20/2022 01:55:35

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:47	
Control Test	0.078	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:51	No Sample Provided
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Air Blank	0.000	01:53	
Subject Sample	# 2	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Control Test	0.077	01:54	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: HANDY

DATE/TIME OF TEST: 09/20/2022 21:42:03

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.079	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.151	21:37
Breath Volume	1.738	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.140	21:40
Breath Volume	1.156	21:40
Air Blank	0.000	21:41
Control Test	0.078	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:42

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MANCHEGO PINZON

DATE/TIME OF TEST: 09/21/2022 02:41:28

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.077	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.087	02:34
Breath Volume	2.273	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.086	02:39
Breath Volume	2.027	02:39
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/21/2022 03:36:59

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.159	03:32
Breath Volume	1.546	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.171	03:35
Breath Volume	1.378	03:35
Air Blank	0.000	03:36
Control Test	0.077	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBERTS,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 09/22/2022 03:19:48

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.174	03:15
Breath Volume	2.617	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.160	03:18
Breath Volume	1.492	03:18
Air Blank	0.000	03:18
Control Test	0.078	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 09/22/2022 21:37:14

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.079	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.112	21:32
Breath Volume	1.617	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.116	21:35
Breath Volume	1.546	21:35
Air Blank	0.000	21:36
Control Test	0.078	21:36
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/22/2022 22:46:25

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.078	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.197	22:39
Breath Volume	1.375	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.193	22:44
Breath Volume	1.316	22:44
Air Blank	0.000	22:45
Control Test	0.077	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 10/13/2022 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: PEGUERO LOPEZ

DATE/TIME OF TEST: 09/23/2022 04:47:28

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.078	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.171	04:42
Breath Volume	2.429	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.177	04:45
Breath Volume	2.226	04:45
Air Blank	0.000	04:46
Control Test	0.078	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: POLLOCK,D
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:27

SUBJECT'S LAST NAME: DUGAN

DATE/TIME OF TEST: 09/23/2022 11:54:40

Diagnostics Check	OK	11:48
Air Blank	0.000	11:48
Control Test	0.078	11:48
Air Blank	0.000	11:49
Subject Sample # 1	0.141	11:49
Breath Volume	1.347	11:49
Air Blank	0.000	11:50
Air Blank	0.000	11:52
Subject Sample # 2	0.133	11:53
Breath Volume	1.246	11:53
Air Blank	0.000	11:53
Control Test	0.078	11:54
Air Blank	0.000	11:54
Diagnostics Check	OK	11:54

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: HENRIQUEZ

DATE/TIME OF TEST: 09/24/2022 01:06:23

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.154	01:01
Breath Volume	1.628	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.157	01:04
Breath Volume	1.894	01:04
Air Blank	0.000	01:05
Control Test	0.077	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MYLES III

DATE/TIME OF TEST: 09/24/2022 01:53:09

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.077	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.215	01:48
Breath Volume	1.613	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.215	01:51
Breath Volume	1.800	01:51
Air Blank	0.000	01:52
Control Test	0.078	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: ZAPONI

DATE/TIME OF TEST: 09/24/2022 04:15:44

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.077	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.238	04:09
Breath Volume	1.136	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.254	04:14
Breath Volume	1.953	04:14
Air Blank	0.000	04:14
Control Test	0.078	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FORSBERG, B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 09/24/2022 04:41:08

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:37	
Control Test	0.078	04:38	
Air Blank	0.000	04:38	
Subject Sample	# 1	04:39	Subject Test Refused
Breath Volume	0.000	04:39	
Air Blank	0.000	04:40	
Control Test	0.077	04:40	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 09/24/2022 07:09:49

Diagnostics Check	OK	07:03
Air Blank	0.000	07:03
Control Test	0.078	07:04
Air Blank	0.000	07:04
Subject Sample # 1	0.175	07:05
Breath Volume	2.714	07:05
Air Blank	0.000	07:05
Air Blank	0.000	07:07
Subject Sample # 2	0.170	07:08
Breath Volume	2.910	07:08
Air Blank	0.000	07:08
Control Test	0.078	07:09
Air Blank	0.000	07:09
Diagnostics Check	OK	07:09

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 09/24/2022 22:52:17

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.210	22:47
Breath Volume	3.597	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.207	22:50
Breath Volume	3.183	22:50
Air Blank	0.000	22:51
Control Test	0.077	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: VIERA

DATE/TIME OF TEST: 09/25/2022 00:36:23

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.078	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1 0.170	00:29	Volume Not Met
Breath Volume	0.070	00:29	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 2 0.167	00:34	Volume Not Met
Breath Volume	0.035	00:34	
Air Blank	0.000	00:35	
Control Test	0.077	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMSON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 09/25/2022 16:18:57

Diagnostics Check	OK	16:10
Air Blank	0.000	16:11
Control Test	0.079	16:11
Air Blank	0.000	16:11
Subject Sample # 1	0.279	16:14
Breath Volume	3.601	16:14
Air Blank	0.000	16:14
Air Blank	0.000	16:16
Subject Sample # 2	0.296	16:17
Breath Volume	3.046	16:17
Air Blank	0.000	16:18
Control Test	0.076	16:18
Air Blank	0.000	16:18
Diagnostics Check	OK	16:18

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: VASQUEZ MARTINEZ

DATE/TIME OF TEST: 09/25/2022 23:32:24

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.078	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.182	23:27
Breath Volume	1.890	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.194	23:30
Breath Volume	1.519	23:30
Air Blank	0.000	23:31
Control Test	0.078	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: SOLEIMANI

DATE/TIME OF TEST: 09/26/2022 02:05:31

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.078	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.000	02:01
Breath Volume	1.156	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.000	02:04
Breath Volume	1.222	02:04
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEYERS,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: RO

DATE/TIME OF TEST: 09/26/2022 22:34:11

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.078	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.061	22:29
Breath Volume	2.585	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.059	22:32
Breath Volume	2.253	22:32
Air Blank	0.000	22:33
Control Test	0.077	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STILES,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:56

SUBJECT'S LAST NAME: AGOSTINI

DATE/TIME OF TEST: 09/27/2022 10:23:44

Diagnostics Check	OK	10:17
Air Blank	0.000	10:17
Control Test	0.078	10:17
Air Blank	0.000	10:18
Subject Sample # 1	0.000	10:19
Breath Volume	2.859	10:19
Air Blank	0.000	10:19
Air Blank	0.000	10:21
Subject Sample # 2	0.000	10:22
Breath Volume	2.585	10:22
Air Blank	0.000	10:22
Control Test	0.078	10:23
Air Blank	0.000	10:23
Diagnostics Check	OK	10:23

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: PHELPS

DATE/TIME OF TEST: 09/29/2022 23:59:48

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.078	23:45	
Air Blank	0.000	23:45	
Subject Sample	# 1 0.147	23:48	Volume Not Met
Breath Volume	0.023	23:48	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample	# 2 0.247	23:52	
Breath Volume	1.136	23:52	
Air Blank	0.000	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 3 0.224	23:58	Volume Not Met
Breath Volume	0.210	23:58	
Air Blank	0.000	23:58	
Control Test	0.077	23:59	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: PHELPS

DATE/TIME OF TEST: 09/30/2022 00:18:39

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.077	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1 0.196	00:07	Volume Not Met
Breath Volume	0.585	00:07	
Air Blank	0.000	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 2 0.224	00:11	
Breath Volume	1.437	00:11	
Air Blank	0.000	00:12	
Air Blank	0.000	00:13	
Subject Sample	# 3 0.226	00:16	Volume Not Met
Breath Volume	0.562	00:16	
Air Blank	0.000	00:17	
Control Test	0.077	00:18	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: CUSTODIO

DATE/TIME OF TEST: 09/30/2022 20:17:51

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.079	20:12
Air Blank	0.000	20:12
Subject Sample # 1	0.165	20:13
Breath Volume	1.433	20:13
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.150	20:16
Breath Volume	1.554	20:16
Air Blank	0.000	20:16
Control Test	0.079	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 10/13/2022 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BUARQUE DE LIMA

DATE/TIME OF TEST: 09/04/2022 04:03:48

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 09/29/2022 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ANDRADE,ESTEBAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 09/04/2022 05:18:43

Diagnostics Check	OK	05:15	
Air Blank	0.000	05:16	
Control Test	0.080	05:16	
Air Blank	0.000	05:17	
Subject Sample	# 1	05:17	Subject Test Refused
Breath Volume	0.000	05:17	
Air Blank	0.000	05:17	
Control Test	0.081	05:18	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:18	

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: RODRIGUEZ PINEDA

DATE/TIME OF TEST: 09/08/2022 19:42:20

Diagnostics Check	OK	19:39	
Air Blank	0.000	19:39	
Control Test	0.080	19:40	
Air Blank	0.000	19:40	
Subject Sample	# 1	19:40	Subject Test Refused
Breath Volume	0.000	19:40	
Air Blank	0.000	19:41	
Control Test	0.081	19:41	
Air Blank	0.000	19:42	
Diagnostics Check	OK	19:42	

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,ELAM,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: THORNBRUGH,ELAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 09/09/2022 04:38:21

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.029	04:33
Breath Volume	1.285	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.030	04:36
Breath Volume	4.132	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: BUTISTE

DATE/TIME OF TEST: 09/12/2022 20:37:24

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:35	
Control Test	0.081	20:35	
Air Blank	0.000	20:35	
Subject Sample	# 1	20:36	Subject Test Refused
Breath Volume	0.000	20:36	
Air Blank	0.000	20:36	
Control Test	0.081	20:36	
Air Blank	0.000	20:37	
Diagnostics Check	OK	20:37	

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALBO,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BUSCEMI,ANTHONY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SEELEY

DATE/TIME OF TEST: 09/16/2022 01:36:18

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:28	
Control Test	0.081	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:32	No Sample Provided
Breath Volume	0.000	01:32	
Air Blank	0.000	01:32	
Air Blank	0.000	01:34	
Subject Sample	# 2	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:35	
Control Test	0.080	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI, RYAN, P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CARPENTER, BENJAMIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SLY

DATE/TIME OF TEST: 09/16/2022 23:44:09

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.185	23:39
Breath Volume	1.363	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.169	23:42
Breath Volume	1.242	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,ELAM,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: THORNBRUGH,ELAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: CASTRILLO

DATE/TIME OF TEST: 09/18/2022 04:58:09

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.081	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1 0.097	04:50	Volume Not Met
Breath Volume	0.460	04:50	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample	# 2 0.105	04:56	Volume Not Met
Breath Volume	0.531	04:56	
Air Blank	0.000	04:57	
Control Test	0.077	04:57	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,ELAM,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: THORNBRUGH,ELAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: CASTRILLO

DATE/TIME OF TEST: 09/18/2022 05:08:15

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.078	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.133	05:02
Breath Volume	2.179	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.124	05:06
Breath Volume	1.796	05:06
Air Blank	0.000	05:07
Control Test	0.076	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: DIONISE

DATE/TIME OF TEST: 09/18/2022 23:37:53

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.080	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:36	
Control Test	0.081	23:37	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BETZ,FRANK
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 09/19/2022 09:41:31

Diagnostics Check	OK	09:38	
Air Blank	0.000	09:39	
Control Test	0.080	09:39	
Air Blank	0.000	09:39	
Subject Sample	# 1	09:40	Subject Test Refused
Breath Volume	0.000	09:40	
Air Blank	0.000	09:40	
Control Test	0.080	09:40	
Air Blank	0.000	09:41	
Diagnostics Check	OK	09:41	

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BETZ,FRANK
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 09/20/2022 18:04:01

Diagnostics Check	OK	17:57
Air Blank	0.000	17:58
Control Test	0.081	17:58
Air Blank	0.000	17:58
Subject Sample # 1	0.243	17:59
Breath Volume	2.144	17:59
Air Blank	0.000	18:00
Air Blank	0.000	18:01
Subject Sample # 2	0.241	18:02
Breath Volume	2.515	18:02
Air Blank	0.000	18:03
Control Test	0.079	18:03
Air Blank	0.000	18:03
Diagnostics Check	OK	18:04

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 09/21/2022 03:31:42

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.119	03:26
Breath Volume	1.410	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.120	03:30
Breath Volume	2.179	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 09/22/2022 01:55:30

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.081	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1 0.169	01:46	Volume Not Met
Breath Volume	0.062	01:46	
Air Blank	0.000	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 2 0.189	01:50	
Breath Volume	2.562	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 3 0.203	01:53	
Breath Volume	2.503	01:53	
Air Blank	0.000	01:54	
Control Test	0.077	01:54	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 09/29/2022 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 09/24/2022 23:02:40

Diagnostics Check	OK	22:59	
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:01	Subject Test Refused
Breath Volume	0.000	23:01	
Air Blank	0.000	23:01	
Control Test	0.080	23:02	
Air Blank	0.000	23:02	
Diagnostics Check	OK	23:02	

Data Download Date/Time: 09/29/2022 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIERRO,ALEXANDER,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FIERRO,ALEX
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 09/25/2022 05:34:43

Diagnostics Check	OK	05:22	
Air Blank	0.000	05:23	
Control Test	0.080	05:23	
Air Blank	0.000	05:24	
Subject Sample	# 1	05:27	No Sample Provided
Breath Volume	0.000	05:27	
Air Blank	0.000	05:27	
Air Blank	0.000	05:29	
Subject Sample	# 2	05:32	Volume Not Met
Breath Volume	0.105	05:32	
Air Blank	0.000	05:33	
Control Test	0.080	05:34	
Air Blank	0.000	05:34	
Diagnostics Check	OK	05:34	

Data Download Date/Time: 09/29/2022 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIERRO,ALEX,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FIERRO,ALEX
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 09/25/2022 05:50:48

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:39	
Control Test	0.079	05:39	
Air Blank	0.000	05:40	
Subject Sample	# 1 0.156	05:43	Volume Not Met
Breath Volume	0.167	05:43	
Air Blank	0.000	05:44	
Air Blank	0.000	05:45	
Subject Sample	# 2 0.157	05:49	Volume Not Met
Breath Volume	0.296	05:49	
Air Blank	0.000	05:49	
Control Test	0.077	05:50	
Air Blank	0.000	05:50	
Diagnostics Check	OK	05:50	

Data Download Date/Time: 09/29/2022 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIERRO,ALEX,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FIERRO,ALEX
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 09/25/2022 05:55:44

Diagnostics Check	OK	05:52	
Air Blank	0.000	05:53	
Control Test	0.078	05:53	
Air Blank	0.000	05:54	
Subject Sample	# 1	05:54	Subject Test Refused
Breath Volume	0.000	05:54	
Air Blank	0.000	05:54	
Control Test	0.079	05:55	
Air Blank	0.000	05:55	
Diagnostics Check	OK	05:55	

Data Download Date/Time: 09/29/2022 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: MENDOZA OLIVARES

DATE/TIME OF TEST: 09/25/2022 22:05:08

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.180	21:59
Breath Volume	1.769	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.179	22:03
Breath Volume	2.246	22:03
Air Blank	0.000	22:04
Control Test	0.078	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 09/29/2022 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALBO,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GALBO,NICHOLAS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: WILLS

DATE/TIME OF TEST: 09/26/2022 01:00:25

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.080	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:56	No Sample Provided
Breath Volume	0.000	00:56	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample	# 2	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Control Test	0.080	00:59	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 09/29/2022 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 09:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:47

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 09/26/2022 10:24:43

Diagnostics Check	OK	10:12	
Air Blank	0.000	10:13	
Control Test	0.080	10:13	
Air Blank	0.000	10:14	
Subject Sample	# 1 0.277	10:17	Volume Not Met
Breath Volume	0.121	10:17	
Air Blank	0.000	10:18	
Air Blank	0.000	10:19	
Subject Sample	# 2 0.226	10:22	Volume Not Met
Breath Volume	0.546	10:22	
Air Blank	0.000	10:23	
Control Test	0.076	10:24	
Air Blank	0.000	10:24	
Diagnostics Check	OK	10:24	

Data Download Date/Time: 09/29/2022 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 09:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:47

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 09/26/2022 10:40:04

Diagnostics Check	OK	10:26	
Air Blank	0.000	10:26	
Control Test	0.077	10:27	
Air Blank	0.000	10:27	
Subject Sample	# 1 0.266	10:30	Volume Not Met
Breath Volume	0.812	10:30	
Air Blank	0.000	10:31	
Air Blank	0.000	10:33	
Subject Sample	# 2 0.294	10:34	
Breath Volume	1.726	10:34	
Air Blank	0.000	10:35	
Air Blank	0.000	10:36	
Subject Sample	# 3 0.291	10:38	
Breath Volume	2.203	10:38	
Air Blank	0.000	10:39	
Control Test	0.076	10:39	
Air Blank	0.000	10:39	
Diagnostics Check	OK	10:40	

Data Download Date/Time: 09/29/2022 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,LEONARDO,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HERNANDEZ,LEONARDO
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 09/28/2022 19:20:52

Diagnostics Check	OK	19:17	
Air Blank	0.000	19:18	
Control Test	0.080	19:18	
Air Blank	0.000	19:19	
Subject Sample	# 1	19:19	Subject Test Refused
Breath Volume	0.000	19:19	
Air Blank	0.000	19:19	
Control Test	0.081	19:20	
Air Blank	0.000	19:20	
Diagnostics Check	OK	19:20	

Data Download Date/Time: 09/29/2022 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: BERG,TIM

ARRESTING OFFICER AGENCY: Hillsborough Co SO

LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/15/2023

VIOLATION CODE: BUI

ARREST TIME: 13:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 13:22

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 09/03/2022 13:50:06

Diagnostics Check	OK	13:43
Air Blank	0.000	13:43
Control Test	0.078	13:43
Air Blank	0.000	13:44
Subject Sample # 1	0.035	13:45
Breath Volume	2.449	13:45
Air Blank	0.000	13:45
Air Blank	0.000	13:47
Subject Sample # 2	0.034	13:48
Breath Volume	2.632	13:48
Air Blank	0.000	13:49
Control Test	0.077	13:49
Air Blank	0.000	13:50
Diagnostics Check	OK	13:50

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASS,JUSTIN,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILKINS,M
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/15/2023
VIOLATION CODE: BUI
ARREST TIME: 15:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: HINTON

DATE/TIME OF TEST: 09/03/2022 15:48:07

Diagnostics Check	OK	15:41
Air Blank	0.000	15:41
Control Test	0.077	15:42
Air Blank	0.000	15:42
Subject Sample # 1	0.067	15:43
Breath Volume	3.375	15:43
Air Blank	0.000	15:44
Air Blank	0.000	15:46
Subject Sample # 2	0.058	15:46
Breath Volume	1.597	15:46
Air Blank	0.000	15:47
Control Test	0.077	15:47
Air Blank	0.000	15:48
Diagnostics Check	OK	15:48

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: MCCORMACK,EAMON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/15/2023
VIOLATION CODE: BUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: CALLICOAT

DATE/TIME OF TEST: 09/03/2022 20:43:13

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:32	
Control Test	0.077	20:32	
Air Blank	0.000	20:33	
Subject Sample	# 1	20:36	No Sample Provided
Breath Volume	0.000	20:36	
Air Blank	0.000	20:36	
Air Blank	0.000	20:38	
Subject Sample	# 2	20:41	No Sample Provided
Breath Volume	0.000	20:41	
Air Blank	0.000	20:42	
Control Test	0.078	20:42	
Air Blank	0.000	20:43	
Diagnostics Check	OK	20:43	

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: PASS,JUSTIN

ARRESTING OFFICER AGENCY: Hillsborough Co SO

LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/15/2023

VIOLATION CODE: BUI

ARREST TIME: 21:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 09/03/2022 22:23:48

Diagnostics Check	OK	22:12	
Air Blank	0.000	22:12	
Control Test	0.077	22:13	
Air Blank	0.000	22:13	
Subject Sample	# 1	22:16	No Sample Provided
Breath Volume	0.000	22:16	
Air Blank	0.000	22:17	
Air Blank	0.000	22:19	
Subject Sample	# 2	22:22	No Sample Provided
Breath Volume	0.000	22:22	
Air Blank	0.000	22:22	
Control Test	0.078	22:23	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASS,JUSTIN,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILKINS,M
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/15/2023
VIOLATION CODE: BUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: RIX

DATE/TIME OF TEST: 09/04/2022 14:25:13

Diagnostics Check	OK	14:13	
Air Blank	0.000	14:14	
Control Test	0.077	14:14	
Air Blank	0.000	14:15	
Subject Sample	# 1	14:18	No Sample Provided
Breath Volume	0.000	14:18	
Air Blank	0.000	14:18	
Air Blank	0.000	14:20	
Subject Sample	# 2	14:23	No Sample Provided
Breath Volume	0.000	14:23	
Air Blank	0.000	14:24	
Control Test	0.078	14:24	
Air Blank	0.000	14:25	
Diagnostics Check	OK	14:25	

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ,BENNETT,L

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: RODRIGUEZ,BENNETT

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/15/2023

VIOLATION CODE: BUI

ARREST TIME: 15:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: ANTONUCCI

DATE/TIME OF TEST: 09/04/2022 16:11:26

Diagnostics Check	OK	16:08	
Air Blank	0.000	16:08	
Control Test	0.078	16:09	
Air Blank	0.000	16:09	
Subject Sample	# 1	16:10	Subject Test Refused
Breath Volume	0.000	16:10	
Air Blank	0.000	16:10	
Control Test	0.078	16:10	
Air Blank	0.000	16:11	
Diagnostics Check	OK	16:11	

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, OFC B,
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PASS, OFC J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/15/2023
VIOLATION CODE: BUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: RUCABADO FUXENCH

DATE/TIME OF TEST: 09/04/2022 16:20:43

Diagnostics Check	OK	16:14	
Air Blank	0.000	16:14	
Control Test	0.077	16:15	
Air Blank	0.000	16:15	
Subject Sample # 1	0.051	16:16	
Breath Volume	2.136	16:16	
Air Blank	0.000	16:16	
Air Blank	0.000	16:18	
Subject Sample # 2	0.051	16:19	
Breath Volume	2.253	16:19	
Air Blank	0.000	16:19	
Control Test	0.078	16:20	
Air Blank		16:20	RFI Detect
Air Blank	0.000	16:20	

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ, OFC B,

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILKINS, OFC M

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/15/2023

VIOLATION CODE: BUI

ARREST TIME: 17:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: KOLLERR

DATE/TIME OF TEST: 09/04/2022 18:09:56

Diagnostics Check	OK	18:03
Air Blank	0.000	18:04
Control Test	0.078	18:04
Air Blank	0.000	18:04
Subject Sample # 1	0.078	18:05
Breath Volume	2.371	18:05
Air Blank	0.000	18:05
Air Blank	0.000	18:07
Subject Sample # 2	0.075	18:08
Breath Volume	2.628	18:08
Air Blank	0.000	18:09
Control Test	0.078	18:09
Air Blank	0.000	18:09
Diagnostics Check	OK	18:09

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ, OFC B,

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: MCCORMACK, OFC E

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/15/2023

VIOLATION CODE: BUI

ARREST TIME: 17:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: KOELKER

DATE/TIME OF TEST: 09/04/2022 18:18:49

Diagnostics Check	OK	18:12
Air Blank	0.000	18:12
Control Test	0.077	18:13
Air Blank	0.000	18:13
Subject Sample # 1	0.042	18:14
Breath Volume	2.183	18:14
Air Blank	0.000	18:14
Air Blank	0.000	18:16
Subject Sample # 2	0.041	18:17
Breath Volume	2.011	18:17
Air Blank	0.000	18:17
Control Test	0.077	18:18
Air Blank	0.000	18:18
Diagnostics Check	OK	18:18

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PASS,JUSTIN,E

BREATH TEST OPERATOR AGENCY: Hillsborough Co SO

ARRESTING OFFICER: BERG,TIM

ARRESTING OFFICER AGENCY: Hillsborough Co SO

LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/15/2023

VIOLATION CODE: BUI

ARREST TIME: 19:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 09/04/2022 19:49:08

Diagnostics Check	OK	19:42
Air Blank	0.000	19:42
Control Test	0.079	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.169	19:44
Breath Volume	2.472	19:44
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.159	19:47
Breath Volume	2.332	19:47
Air Blank	0.000	19:48
Control Test	0.078	19:48
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ, OFC B,

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: HOPPE, OFC A

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/15/2023

VIOLATION CODE: BUI

ARREST TIME: 14:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: MAYE NOPAL

DATE/TIME OF TEST: 09/05/2022 15:14:47

Diagnostics Check	OK	15:08
Air Blank	0.000	15:08
Control Test	0.077	15:09
Air Blank	0.000	15:09
Subject Sample # 1	0.013	15:10
Breath Volume	1.656	15:10
Air Blank	0.000	15:10
Air Blank	0.000	15:12
Subject Sample # 2	0.013	15:13
Breath Volume	2.925	15:13
Air Blank	0.000	15:13
Control Test	0.078	15:14
Air Blank	0.000	15:14
Diagnostics Check	OK	15:14

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVIS,REECE,H
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: ALVIS,REECE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/15/2023
VIOLATION CODE: BUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 09/05/2022 18:42:30

Diagnostics Check	OK	18:31	
Air Blank	0.000	18:31	
Control Test	0.078	18:31	
Air Blank	0.000	18:32	
Subject Sample	# 1	18:35	No Sample Provided
Breath Volume	0.000	18:35	
Air Blank	0.000	18:35	
Air Blank	0.000	18:37	
Subject Sample	# 2	18:41	No Sample Provided
Breath Volume	0.000	18:41	
Air Blank	0.000	18:41	
Control Test	0.078	18:41	
Air Blank	0.000	18:42	
Diagnostics Check	OK	18:42	

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: LAMOUREUX,MATTHEW

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/15/2023

VIOLATION CODE: BUI

ARREST TIME: 19:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: RALHAN

DATE/TIME OF TEST: 09/05/2022 20:22:43

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:11	
Control Test	0.078	20:12	
Air Blank	0.000	20:12	
Subject Sample	# 1	20:15	No Sample Provided
Breath Volume	0.000	20:15	
Air Blank	0.000	20:16	
Air Blank	0.000	20:18	
Subject Sample	# 2	20:21	No Sample Provided
Breath Volume	0.000	20:21	
Air Blank	0.000	20:21	
Control Test	0.078	20:22	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:22	

Data Download Date/Time: 10/24/2022 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DANIELS,WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/03/2022 03:35:26

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:34	
Control Test	0.080	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 09/29/2022 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO, JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CRISTEA, VICTOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MCNALLY

DATE/TIME OF TEST: 09/09/2022 01:44:23

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Control Test	0.079	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 09/29/2022 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BERARDI III

DATE/TIME OF TEST: 09/14/2022 01:35:58

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:33	
Control Test	0.080	01:33	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:35	
Control Test	0.080	01:35	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 09/29/2022 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/18/2022 02:38:23

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.195	02:33
Breath Volume	1.691	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.189	02:36
Breath Volume	1.144	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 09/29/2022 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEMIEC,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Key Colony Beach PD
ARRESTING OFFICER: FOSTER,MEGHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HENDRYX

DATE/TIME OF TEST: 09/24/2022 02:58:17

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.177	02:53
Breath Volume	1.480	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:55
Subject Sample # 2	0.185	02:56
Breath Volume	2.507	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 09/29/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: AGUANNO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 11:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:58

SUBJECT'S LAST NAME: LOPEZ GONZALEZ

DATE/TIME OF TEST: 09/24/2022 12:34:15

Diagnostics Check	OK	12:24
Air Blank	0.000	12:24
Control Test	0.079	12:25
Air Blank	0.000	12:25
Subject Sample # 1	0.030	12:27
Breath Volume	1.414	12:27
Air Blank	0.000	12:27
Air Blank	0.000	12:29
Subject Sample # 2	0.027	12:32
Breath Volume	1.457	12:32
Air Blank	0.000	12:33
Control Test	0.080	12:33
Air Blank	0.000	12:34
Diagnostics Check	OK	12:34

Data Download Date/Time: 09/29/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: VONDERSAAR

DATE/TIME OF TEST: 09/28/2022 03:24:52

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.079	03:17	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:20	No Sample Provided
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Air Blank	0.000	03:23	
Subject Sample	# 2	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Control Test	0.080	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 09/29/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANZA,CHRISTIAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LANZA,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MCNEIL

DATE/TIME OF TEST: 09/02/2022 23:03:01

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.200	22:58
Breath Volume	1.343	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.213	23:01
Breath Volume	2.347	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:03

Data Download Date/Time: 09/22/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARR,LUCAS,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KARR,LUCAS
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: VAVREK

DATE/TIME OF TEST: 09/03/2022 22:47:44

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.080	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:40	No Sample Provided
Breath Volume	0.000	22:40	
Air Blank	0.000	22:41	
Air Blank	0.000	22:43	
Subject Sample	# 2	22:46	No Sample Provided
Breath Volume	0.000	22:46	
Air Blank	0.000	22:46	
Control Test	0.081	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 09/22/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,M,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MOORE,M
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/04/2022 03:04:33

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.164	02:59
Breath Volume	1.199	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.166	03:02
Breath Volume	1.589	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 09/22/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRINGER,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: STRINGER,MICHAEL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/07/2022 19:28:32

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.081	19:21
Air Blank	0.000	19:22
Subject Sample # 1	0.179	19:22
Breath Volume	1.886	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.169	19:26
Breath Volume	1.363	19:26
Air Blank	0.000	19:27
Control Test	0.081	19:27
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 09/22/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MARIO,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RODRIGUEZ,MARIO
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 09/13/2022 15:44:44

Diagnostics Check	OK	15:33	
Air Blank	0.000	15:33	
Control Test	0.081	15:34	
Air Blank	0.000	15:34	
Subject Sample	# 1	15:37	No Sample Provided
Breath Volume	0.000	15:37	
Air Blank	0.000	15:38	
Air Blank	0.000	15:40	
Subject Sample	# 2	15:43	No Sample Provided
Breath Volume	0.000	15:43	
Air Blank	0.000	15:43	
Control Test	0.081	15:44	
Air Blank	0.000	15:44	
Diagnostics Check	OK	15:44	

Data Download Date/Time: 09/22/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSBORNE,HUNTER,C
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: OSBORNE,HUNTER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 09/15/2022 01:38:09

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.081	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:37	Sequence Aborted
Breath Volume	0.000	01:37	
Air Blank	0.000	01:38	

Data Download Date/Time: 09/22/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CARNECCHIA

DATE/TIME OF TEST: 09/15/2022 01:47:44

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.000	01:43
Breath Volume	1.941	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.000	01:46
Breath Volume	2.109	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 09/22/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RUSSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 09/16/2022 02:34:00

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.225	02:28
Breath Volume	1.285	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample # 2	0.209	02:32
Breath Volume	1.257	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:34

Data Download Date/Time: 09/22/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETTS,J,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/16/2022 05:35:27

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.155	05:30
Breath Volume	1.593	05:30
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.155	05:33
Breath Volume	2.617	05:33
Air Blank	0.000	05:34
Control Test	0.080	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 09/22/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: RAMIREZ HERNANDEZ

DATE/TIME OF TEST: 09/17/2022 05:56:38

Diagnostics Check	OK	05:50
Air Blank	0.000	05:50
Control Test	0.081	05:51
Air Blank	0.000	05:51
Subject Sample # 1	0.015	05:52
Breath Volume	1.597	05:52
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample # 2	0.013	05:55
Breath Volume	1.351	05:55
Air Blank	0.000	05:55
Control Test	0.079	05:56
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

Data Download Date/Time: 09/22/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MUNDORFF,TERRY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 14:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 09/20/2022 15:26:57

Diagnostics Check	OK	15:20
Air Blank	0.000	15:21
Control Test	0.082	15:21
Air Blank	0.000	15:22
Subject Sample # 1	0.000	15:22
Breath Volume	2.843	15:22
Air Blank	0.000	15:23
Air Blank	0.000	15:25
Subject Sample # 2	0.000	15:25
Breath Volume	2.566	15:25
Air Blank	0.000	15:26
Control Test	0.081	15:26
Air Blank	0.000	15:26
Diagnostics Check	OK	15:26

Data Download Date/Time: 09/22/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/21/2022 04:13:33

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.135	04:08
Breath Volume	3.980	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.137	04:11
Breath Volume	2.667	04:11
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 09/22/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETTS,J,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 09/24/2022 03:10:17

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.080	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.080	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 10/27/2022 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTIS,LAYNE,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: CURTIS,LAYNE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: STENBACK

DATE/TIME OF TEST: 09/24/2022 04:03:28

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.080	03:52	
Air Blank	0.000	03:52	
Subject Sample	# 1 0.232	03:55	Volume Not Met
Breath Volume	0.000	03:55	
Air Blank	0.000	03:56	
Air Blank	0.000	03:58	
Subject Sample	# 2 0.252	04:01	Volume Not Met
Breath Volume	0.023	04:01	
Air Blank	0.000	04:02	
Control Test	0.077	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 10/27/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTIS,LAYNE,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: CURTIS,LAYNE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: STENBACK

DATE/TIME OF TEST: 09/24/2022 04:17:18

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.076	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.275	04:11
Breath Volume	1.277	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.269	04:15
Breath Volume	1.261	04:15
Air Blank	0.000	04:16
Control Test	0.077	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 10/27/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WELFENBERG,ASHTON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370810
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/25/2022 21:56:35

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.081	21:50
Air Blank	0.000	21:51
Subject Sample # 1	0.095	21:51
Breath Volume	2.371	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.093	21:54
Breath Volume	2.285	21:54
Air Blank	0.000	21:55
Control Test	0.080	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 10/27/2022 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ELLIOTT,ZACHARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: MAZZUCA

DATE/TIME OF TEST: 09/26/2022 18:19:39

Diagnostics Check	OK	18:15	
Air Blank	0.000	18:15	
Control Test	0.080	18:16	
Air Blank	0.000	18:16	
Subject Sample	# 1	18:18	Subject Test Refused
Breath Volume	0.000	18:18	
Air Blank	0.000	18:18	
Control Test	0.080	18:19	
Air Blank	0.000	18:19	
Diagnostics Check	OK	18:19	

Data Download Date/Time: 10/27/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REUL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: REUL,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: STEINER

DATE/TIME OF TEST: 09/27/2022 19:03:34

Diagnostics Check	OK	18:56	
Air Blank	0.000	18:57	
Control Test	0.080	18:57	
Air Blank	0.000	18:58	
Subject Sample	# 1 0.115	18:58	
Breath Volume	1.292	18:58	
Air Blank	0.000	18:59	
Air Blank	0.000	19:01	
Subject Sample	# 2 0.108	19:01	
Breath Volume	1.335	19:01	
Air Blank	0.000	19:02	
Control Test	0.076	19:03	RFI Detect
Air Blank	0.000	19:03	

Data Download Date/Time: 10/27/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REUL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: REUL,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: STEINER

DATE/TIME OF TEST: 09/27/2022 19:14:41

Diagnostics Check	OK	19:07
Air Blank	0.000	19:07
Control Test	0.079	19:08
Air Blank	0.000	19:08
Subject Sample # 1	0.107	19:09
Breath Volume	1.269	19:09
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample # 2	0.109	19:12
Breath Volume	1.429	19:12
Air Blank	0.000	19:13
Control Test	0.078	19:14
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

Data Download Date/Time: 10/27/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWLING,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOWLING,MICHAEL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: KUNDERT

DATE/TIME OF TEST: 09/28/2022 04:07:22

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.124	04:02
Breath Volume	1.863	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.123	04:05
Breath Volume	1.695	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 10/27/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MARIO,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RODRIGUEZ,MARIO
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: GOLA

DATE/TIME OF TEST: 09/28/2022 18:03:46

Diagnostics Check	OK	17:52	
Air Blank	0.000	17:52	
Control Test	0.080	17:53	
Air Blank	0.000	17:53	
Subject Sample	# 1	17:56	No Sample Provided
Breath Volume	0.000	17:56	
Air Blank	0.000	17:57	
Air Blank	0.000	17:59	
Subject Sample	# 2	18:02	No Sample Provided
Breath Volume	0.000	18:02	
Air Blank	0.000	18:02	
Control Test	0.080	18:03	
Air Blank	0.000	18:03	
Diagnostics Check	OK	18:03	

Data Download Date/Time: 10/27/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MAY,MIYA,K

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: MAY,MIYA

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: DUI

ARREST TIME: 02:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 09/03/2022 02:42:19

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.081	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.127	02:34
Breath Volume	1.457	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.148	02:37
Breath Volume	1.339	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 3	0.144	02:40
Breath Volume	3.746	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 09/27/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MARTINEZ,HENDRICK
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: HATHERLY

DATE/TIME OF TEST: 09/03/2022 06:14:24

Diagnostics Check	OK	06:08
Air Blank	0.000	06:09
Control Test	0.081	06:09
Air Blank	0.000	06:09
Subject Sample # 1	0.000	06:10
Breath Volume	3.535	06:10
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.000	06:13
Breath Volume	3.636	06:13
Air Blank	0.000	06:13
Control Test	0.081	06:13
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

Data Download Date/Time: 09/27/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WOODS,LINDSEY,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: HUDSON,JOSHUA

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: DUI

ARREST TIME: 01:37

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 09/10/2022 02:27:20

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.081	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.207	02:22
Breath Volume	1.472	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 2	0.206	02:25
Breath Volume	2.148	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 09/27/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASWELL,AUSTIN,J

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: HASWELL,AUSTIN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: DUI

ARREST TIME: 18:05

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: NICHOLAS

DATE/TIME OF TEST: 09/17/2022 18:32:00

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.081	18:26
Air Blank	0.000	18:26
Subject Sample # 1	0.231	18:27
Breath Volume	1.542	18:27
Air Blank	0.000	18:28
Air Blank	0.000	18:29
Subject Sample # 2	0.226	18:30
Breath Volume	2.570	18:30
Air Blank	0.000	18:31
Control Test	0.081	18:31
Air Blank	0.000	18:31
Diagnostics Check	OK	18:32

Data Download Date/Time: 09/27/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROPER, WILLIAM, D

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: ROPER, WILLIAM

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: DUI

ARREST TIME: 22:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 09/19/2022 23:06:09

Diagnostics Check	OK	22:59	
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:00	
Subject Sample	# 1 0.239	23:01	
Breath Volume	1.628	23:01	
Air Blank	0.000	23:02	
Air Blank	0.000	23:04	
Subject Sample	# 2 0.242	23:04	
Breath Volume	1.960	23:04	
Air Blank	0.000	23:05	
Control Test	0.080	23:05	
Air Blank		23:05	RFI Detect
Air Blank	0.000	23:06	

Data Download Date/Time: 09/27/2022 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HUDSON,JOSHUA,K

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: HUDSON,JOSHUA

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: DUI

ARREST TIME: 01:17

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: CULVER

DATE/TIME OF TEST: 09/24/2022 02:14:54

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.082	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.000	02:10
Breath Volume	3.082	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.000	02:13
Breath Volume	3.792	02:13
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 09/27/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HUDSON,JOSHUA,K

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: HUDSON,JOSHUA

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 08/16/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: DUI

ARREST TIME: 04:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: ALSHEHRI

DATE/TIME OF TEST: 09/25/2022 04:37:53

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.082	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1 0.225	04:30	Volume Not Met
Breath Volume	0.101	04:30	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2 0.174	04:36	Volume Not Met
Breath Volume	0.117	04:36	
Air Blank	0.000	04:36	
Control Test	0.080	04:37	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 09/27/2022 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,KARISSA,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MORALES,KARISSA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/28/2022 01:41:02

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.081	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.175	01:36
Breath Volume	1.679	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.170	01:39
Breath Volume	1.738	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:41

Data Download Date/Time: 10/31/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN,BRENDAN,D

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: BUI,RICHARD

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 09/26/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1349789

DRY GAS EXPIRY DATE: 07/15/2023

VIOLATION CODE: DUI

ARREST TIME: 01:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: GHAIMIMI

DATE/TIME OF TEST: 09/29/2022 01:45:30

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.082	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.217	01:40
Breath Volume	2.828	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample # 2	0.208	01:43
Breath Volume	3.578	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 10/31/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: BEAECH

DATE/TIME OF TEST: 09/02/2022 01:08:20

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.152	01:03
Breath Volume	1.730	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.145	01:06
Breath Volume	1.398	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 09/02/2022 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PHILHOWER

DATE/TIME OF TEST: 09/03/2022 20:44:28

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.078	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.179	20:39
Breath Volume	2.605	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.180	20:42
Breath Volume	1.824	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 09/30/2022 15:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02021080A1

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:16

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: VENEZIO

DATE/TIME OF TEST: 09/07/2022 23:22:04

Diagnostics Check

Fail

23:21

Internal Printer Test

Air Blank

0.000

23:22

Data Download Date/Time: 09/30/2022 15:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02021080A1

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:16

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: VENEZIO

DATE/TIME OF TEST: 09/07/2022 23:30:10

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.079	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.204	23:25
Breath Volume	1.140	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:27
Subject Sample # 2	0.202	23:28
Breath Volume	1.523	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 09/30/2022 15:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ETTEL,M
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: JAILLET-ZARZYCKI

DATE/TIME OF TEST: 09/08/2022 01:05:32

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.000	01:01
Breath Volume	1.511	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.000	01:04
Breath Volume	1.238	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 09/30/2022 15:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CARREGAL

DATE/TIME OF TEST: 09/08/2022 01:52:41

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.125	01:48
Breath Volume	2.753	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.120	01:51
Breath Volume	2.554	01:51
Air Blank	0.000	01:51
Control Test	0.080	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 09/30/2022 15:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 09/09/2022 23:11:28

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.080	23:05	
Air Blank	0.000	23:05	
Subject Sample # 1	0.252	23:06	
Breath Volume	2.054	23:06	
Air Blank	0.000	23:06	
Air Blank	0.000	23:08	
Subject Sample # 2	0.246	23:09	
Breath Volume	1.582	23:09	
Air Blank	0.000	23:10	
Control Test	0.080	23:10	
Air Blank	0.000	23:10	
Diagnostics Check	Fail	23:11	Internal Printer Test
Air Blank	0.000	23:11	

Data Download Date/Time: 09/30/2022 15:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: HOLDERFIELD

DATE/TIME OF TEST: 09/15/2022 22:51:20

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.171	22:46
Breath Volume	1.593	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:48
Subject Sample # 2	0.163	22:49
Breath Volume	1.285	22:49
Air Blank	0.000	22:50
Control Test	0.079	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 09/30/2022 15:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: PATRICK

DATE/TIME OF TEST: 09/23/2022 00:55:35

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.081	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.202	00:51
Breath Volume	1.519	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.197	00:53
Breath Volume	1.792	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 09/30/2022 15:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYA,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAYA,ZACHARY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346773
DRY GAS EXPIRY DATE: 07/08/2023
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CUC CUZ

DATE/TIME OF TEST: 09/04/2022 00:24:53

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.077	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.225	00:19
Breath Volume	1.671	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.217	00:23
Breath Volume	2.394	00:23
Air Blank	0.000	00:23
Control Test	0.076	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 09/30/2022 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: VENTRY, WILLIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346773
DRY GAS EXPIRY DATE: 07/08/2023
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: DRAIN

DATE/TIME OF TEST: 09/05/2022 22:40:10

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.078	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.197	22:35
Breath Volume	3.824	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.204	22:38
Breath Volume	4.074	22:38
Air Blank	0.000	22:39
Control Test	0.078	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 09/30/2022 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYA,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAYA,ZACHARY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346773
DRY GAS EXPIRY DATE: 07/08/2023
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: HENNESSEY

DATE/TIME OF TEST: 09/06/2022 16:07:31

Diagnostics Check	OK	16:04	
Air Blank	0.000	16:05	
Control Test	0.077	16:05	
Air Blank	0.000	16:05	
Subject Sample	# 1	16:06	Subject Test Refused
Breath Volume	0.000	16:06	
Air Blank	0.000	16:06	
Control Test	0.077	16:06	
Air Blank	0.000	16:07	
Diagnostics Check	OK	16:07	

Data Download Date/Time: 09/30/2022 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEPREZ,JACOB,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DEPREZ,JACOB
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346773
DRY GAS EXPIRY DATE: 07/08/2023
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MONGOGNA

DATE/TIME OF TEST: 09/13/2022 23:11:34

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.078	23:09	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Control Test	0.078	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 09/30/2022 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DEAN, AUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346773
DRY GAS EXPIRY DATE: 07/08/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: LUDWELL

DATE/TIME OF TEST: 09/24/2022 02:40:57

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.077	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.011	02:36
Breath Volume	2.246	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.010	02:39
Breath Volume	2.804	02:39
Air Blank	0.000	02:40
Control Test	0.077	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 09/30/2022 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DEAN, AUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346773
DRY GAS EXPIRY DATE: 07/08/2023
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 09/30/2022 01:12:58

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:02	
Control Test	0.077	01:02	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:05	No Sample Provided
Breath Volume	0.000	01:05	
Air Blank	0.000	01:06	
Air Blank	0.000	01:08	
Subject Sample	# 2	01:11	No Sample Provided
Breath Volume	0.000	01:11	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 09/30/2022 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ADAMES, KRISTAL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346773
DRY GAS EXPIRY DATE: 07/08/2023
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PERDUE

DATE/TIME OF TEST: 09/30/2022 03:25:10

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.078	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:18	No Sample Provided
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample	# 2	03:23	No Sample Provided
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Control Test	0.079	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 09/30/2022 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: LEATIOTA,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: TRASTER

DATE/TIME OF TEST: 09/01/2022 21:42:20

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.080	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.024	21:37
Breath Volume	1.781	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.023	21:40
Breath Volume	2.070	21:40
Air Blank	0.000	21:41
Control Test	0.079	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 09/02/2022 17:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYERS,SPENCER,O
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: HAND,BRANDON
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:27

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 09/02/2022 10:14:27

Diagnostics Check	OK	10:08
Air Blank	0.000	10:08
Control Test	0.080	10:08
Air Blank	0.000	10:09
Subject Sample # 1	0.175	10:09
Breath Volume	3.425	10:09
Air Blank	0.000	10:10
Air Blank	0.000	10:12
Subject Sample # 2	0.163	10:12
Breath Volume	2.890	10:12
Air Blank	0.000	10:13
Control Test	0.076	10:13
Air Blank	0.000	10:14
Diagnostics Check	OK	10:14

Data Download Date/Time: 09/02/2022 17:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAPELA,JORDAN,A

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: CAPELA,JORDAN

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 202303

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:53

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: HERTER

DATE/TIME OF TEST: 09/05/2022 04:35:13

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.130	04:30
Breath Volume	2.136	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.121	04:33
Breath Volume	1.738	04:33
Air Blank	0.000	04:34
Control Test	0.078	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: REECE,STEPHEN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 09/06/2022 02:57:43

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.079	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.071	02:53
Breath Volume	2.078	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.070	02:56
Breath Volume	2.687	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: JAGIELSKI

DATE/TIME OF TEST: 09/09/2022 16:24:59

Diagnostics Check	OK	16:18
Air Blank	0.000	16:18
Control Test	0.079	16:19
Air Blank	0.000	16:19
Subject Sample # 1	0.100	16:20
Breath Volume	1.765	16:20
Air Blank	0.000	16:20
Air Blank	0.000	16:22
Subject Sample # 2	0.099	16:23
Breath Volume	1.500	16:23
Air Blank	0.000	16:24
Control Test	0.079	16:24
Air Blank	0.000	16:24
Diagnostics Check	OK	16:24

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: WHITEHEAD

DATE/TIME OF TEST: 09/11/2022 01:06:23

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.080	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.134	01:00
Breath Volume	1.726	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.137	01:04
Breath Volume	1.230	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: SALATA

DATE/TIME OF TEST: 09/11/2022 03:19:19

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.064	03:15
Breath Volume	3.203	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.067	03:17
Breath Volume	3.316	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JOHNATHAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BROWN,JOHNATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: DUFF

DATE/TIME OF TEST: 09/12/2022 01:47:09

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.126	01:42
Breath Volume	2.242	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:44
Subject Sample # 2	0.127	01:45
Breath Volume	3.242	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIRSHIKOVA,ALINA,A
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: NOBORE,OJHOSA
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: COMINSKY

DATE/TIME OF TEST: 09/14/2022 02:24:55

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.180	02:19
Breath Volume	1.617	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.168	02:23
Breath Volume	1.687	02:23
Air Blank	0.000	02:24
Control Test	0.077	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALLOUN,SAMUEL,C
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: KINNEY,CARRIANNE
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: RUDEEN

DATE/TIME OF TEST: 09/14/2022 15:22:45

Diagnostics Check	OK	15:16
Air Blank	0.000	15:16
Control Test	0.079	15:16
Air Blank	0.000	15:17
Subject Sample # 1	0.165	15:18
Breath Volume	1.441	15:18
Air Blank	0.000	15:18
Air Blank	0.000	15:20
Subject Sample # 2	0.163	15:21
Breath Volume	1.125	15:21
Air Blank	0.000	15:21
Control Test	0.079	15:22
Air Blank	0.000	15:22
Diagnostics Check	OK	15:22

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARRILL,JOSHUA,D

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: BAKER,LOGAN

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 202303

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 06:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: FALLMANN

DATE/TIME OF TEST: 09/15/2022 07:21:03

Diagnostics Check	OK	07:14
Air Blank	0.000	07:15
Control Test	0.080	07:15
Air Blank	0.000	07:16
Subject Sample # 1	0.000	07:16
Breath Volume	2.125	07:16
Air Blank	0.000	07:17
Air Blank	0.000	07:19
Subject Sample # 2	0.000	07:19
Breath Volume	3.132	07:19
Air Blank	0.000	07:20
Control Test	0.079	07:20
Air Blank	0.000	07:20
Diagnostics Check	OK	07:21

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: MCCracken

DATE/TIME OF TEST: 09/17/2022 00:43:20

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.215	00:38
Breath Volume	3.292	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.222	00:41
Breath Volume	4.226	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SALIFU,EZUL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: KLOECKNER

DATE/TIME OF TEST: 09/17/2022 02:29:05

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.304	02:21
Breath Volume	1.179	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:23
Subject Sample # 2	0.347	02:24
Breath Volume	2.027	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:26
Subject Sample # 3	0.333	02:27
Breath Volume	1.437	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 09/17/2022 23:12:17

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.199	23:06
Breath Volume	1.968	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.191	23:10
Breath Volume	1.347	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ALMAGUER,BRANDON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 09/19/2022 14:19:10

Diagnostics Check	OK	14:10	
Air Blank	0.000	14:11	
Control Test	0.079	14:11	
Air Blank	0.000	14:12	
Subject Sample	# 1 0.292	14:12	
Breath Volume	1.300	14:12	
Air Blank	0.000	14:13	
Air Blank	0.000	14:15	
Subject Sample	# 2	14:17	Subject Test Refused
Breath Volume	0.000	14:17	
Air Blank	0.000	14:18	
Control Test	0.078	14:18	
Air Blank	0.000	14:19	
Diagnostics Check	OK	14:19	

Data Download Date/Time: 09/29/2022 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUI,DANIEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: PARKER,TAYLOR
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GALJOUR

DATE/TIME OF TEST: 09/20/2022 00:02:08

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.080	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.154	23:57
Breath Volume	1.234	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.162	00:00
Breath Volume	1.519	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 09/29/2022 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 202303
BREATH TEST OPERATOR: HEYMANS,MORNE,	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HEYMANS,MORNE	ARREST TIME: 17:28
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 08/31/2022	OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: FRANCISCO	DATE/TIME OF TEST: 09/23/2022	19:18:17
--------------------------------	-------------------------------	----------

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:15	
Control Test	0.079	19:16	
Air Blank	0.000	19:16	
Subject Sample	# 1	19:16	Subject Test Refused
Breath Volume	0.000	19:16	
Air Blank	0.000	19:17	
Control Test	0.078	19:17	
Air Blank	0.000	19:18	
Diagnostics Check	OK	19:18	

Data Download Date/Time: 09/29/2022 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUI,DANIEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: NORCROSS,AUSTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 09/24/2022 00:07:22

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:55	
Control Test	0.079	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 3 0.170	23:59	Volume Not Met
Breath Volume	0.507	23:59	
Air Blank	0.000	23:59	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.185	00:02	
Breath Volume	2.648	00:02	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample	# 2 0.172	00:05	
Breath Volume	1.351	00:05	
Air Blank	0.000	00:06	
Control Test	0.078	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 09/29/2022 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEYMANS,MORNE,

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: HEYMANS,MORNE

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 202303

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:11

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 09/24/2022 04:39:42

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:37	
Control Test	0.080	04:37	
Air Blank	0.000	04:38	
Subject Sample	# 1	04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Control Test	0.080	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 09/29/2022 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYMANS,MORNE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: LEAR,TREY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 09/24/2022 22:07:41

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.203	22:02
Breath Volume	2.144	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.204	22:06
Breath Volume	2.429	22:06
Air Blank	0.000	22:06
Control Test	0.080	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 09/29/2022 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YARBOROUGH,ERIC,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: DELAY,DEVIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SCHRAMM

DATE/TIME OF TEST: 09/24/2022 23:51:49

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.183	23:47
Breath Volume	1.632	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.184	23:50
Breath Volume	2.519	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 09/29/2022 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARPER,DARIN,W

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: HARPER,DARIN

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 202303

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ORTI JR

DATE/TIME OF TEST: 09/25/2022 00:54:35

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.163	00:49
Breath Volume	2.980	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.168	00:53
Breath Volume	2.367	00:53
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 09/29/2022 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRILL,JOSHUA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCGUIRE,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:59

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 09/25/2022 17:26:42

Diagnostics Check	OK	17:20
Air Blank	0.000	17:20
Control Test	0.080	17:20
Air Blank	0.000	17:21
Subject Sample # 1	0.198	17:22
Breath Volume	1.632	17:22
Air Blank	0.000	17:22
Air Blank	0.000	17:24
Subject Sample # 2	0.218	17:25
Breath Volume	2.859	17:25
Air Blank	0.000	17:25
Control Test	0.079	17:26
Air Blank	0.000	17:26
Diagnostics Check	OK	17:26

Data Download Date/Time: 09/29/2022 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: FIGUEROA,KAI
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: KOCSIS

DATE/TIME OF TEST: 09/25/2022 23:44:10

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.079	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.292	23:39
Breath Volume	1.976	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:41
Subject Sample # 2	0.306	23:42
Breath Volume	2.648	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 09/29/2022 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:43

SUBJECT'S LAST NAME: DEANE

DATE/TIME OF TEST: 09/26/2022 16:18:57

Diagnostics Check	OK	16:12
Air Blank	0.000	16:13
Control Test	0.080	16:13
Air Blank	0.000	16:13
Subject Sample # 1	0.149	16:14
Breath Volume	2.355	16:14
Air Blank	0.000	16:15
Air Blank	0.000	16:16
Subject Sample # 2	0.145	16:17
Breath Volume	3.277	16:17
Air Blank	0.000	16:18
Control Test	0.080	16:18
Air Blank	0.000	16:18
Diagnostics Check	OK	16:18

Data Download Date/Time: 09/29/2022 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALLOUN,SAMUEL,C
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MANGOLD,KENNETH
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: PAGUE

DATE/TIME OF TEST: 09/27/2022 16:53:11

Diagnostics Check	OK	16:47
Air Blank	0.000	16:47
Control Test	0.080	16:47
Air Blank	0.000	16:48
Subject Sample # 1	0.000	16:48
Breath Volume	2.070	16:48
Air Blank	0.000	16:49
Air Blank	0.000	16:51
Subject Sample # 2	0.000	16:51
Breath Volume	1.808	16:51
Air Blank	0.000	16:52
Control Test	0.078	16:52
Air Blank	0.000	16:53
Diagnostics Check	OK	16:53

Data Download Date/Time: 09/29/2022 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAHNSEN,A,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: ARAN,H
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: HOLFEDER

DATE/TIME OF TEST: 09/27/2022 20:05:00

Diagnostics Check	OK	19:58
Air Blank	0.000	19:59
Control Test	0.080	19:59
Air Blank	0.000	20:00
Subject Sample # 1	0.196	20:00
Breath Volume	1.406	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.204	20:03
Breath Volume	1.304	20:03
Air Blank	0.000	20:04
Control Test	0.077	20:04
Air Blank	0.000	20:04
Diagnostics Check	OK	20:05

Data Download Date/Time: 09/29/2022 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOLDBERG,JOSH
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/02/2022 03:39:02

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.212	03:34
Breath Volume	2.203	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample # 2	0.219	03:37
Breath Volume	2.664	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:39

Data Download Date/Time: 09/04/2022 00:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: AJTUM

DATE/TIME OF TEST: 09/01/2022 22:07:34

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.081	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.114	22:02
Breath Volume	1.277	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:04
Subject Sample # 2	0.131	22:05
Breath Volume	1.656	22:05
Air Blank	0.000	22:06
Control Test	0.081	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,CARLOS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: MENDEZ MENDEZ

DATE/TIME OF TEST: 09/01/2022 22:41:57

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.081	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:40	
Control Test	0.081	22:41	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GARCIACARILLO

DATE/TIME OF TEST: 09/01/2022 23:39:38

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:36	
Control Test	0.081	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:38	Subject Test Refused
Breath Volume	0.000	23:38	
Air Blank	0.000	23:38	
Control Test	0.081	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: VELASCO VERNAL

DATE/TIME OF TEST: 09/02/2022 21:17:13

Diagnostics Check	OK	21:14	
Air Blank	0.000	21:14	
Control Test	0.081	21:14	
Air Blank	0.000	21:15	
Subject Sample # 1	0.256	21:15	
Breath Volume	2.691	21:15	
Air Blank		21:16	Purge Fail
Air Blank	0.000	21:17	

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: VELASCO VERNAL

DATE/TIME OF TEST: 09/02/2022 21:26:12

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.080	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.247	21:20
Breath Volume	2.343	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.244	21:24
Breath Volume	2.507	21:24
Air Blank	0.000	21:25
Control Test	0.077	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: SANTIAGO FLORES

DATE/TIME OF TEST: 09/03/2022 00:33:56

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.081	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.102	00:29
Breath Volume	1.179	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.096	00:32
Breath Volume	2.617	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,VICTOR,R
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: LUGO,VICTOR
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: GILLESPIE

DATE/TIME OF TEST: 09/04/2022 13:52:35

Diagnostics Check	OK	13:52	
Air Blank		13:52	RFI Detect
Air Blank		13:52	RFI Detect

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,VICTOR,R
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: LUGO,VICTOR
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: GILLESPIE

DATE/TIME OF TEST: 09/04/2022 14:01:58

Diagnostics Check	OK	13:55
Air Blank	0.000	13:55
Control Test	0.081	13:56
Air Blank	0.000	13:56
Subject Sample # 1	0.000	13:57
Breath Volume	2.480	13:57
Air Blank	0.000	13:58
Air Blank	0.000	14:00
Subject Sample # 2	0.000	14:00
Breath Volume	3.167	14:00
Air Blank	0.000	14:01
Control Test	0.080	14:01
Air Blank	0.000	14:01
Diagnostics Check	OK	14:01

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/09/2022 22:05:17

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:02	
Control Test	0.082	22:03	
Air Blank	0.000	22:03	
Subject Sample	# 1	22:03	Subject Test Refused
Breath Volume	0.000	22:03	
Air Blank	0.000	22:04	
Control Test	0.082	22:04	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/10/2022 02:38:05

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:30	
Control Test	0.081	02:30	
Air Blank	0.000	02:31	
Subject Sample # 1	0.238	02:32	
Breath Volume	1.480	02:32	
Air Blank	0.000	02:33	
Air Blank	0.000	02:35	
Subject Sample # 2	0.222	02:36	
Breath Volume	2.070	02:36	
Air Blank		02:37	Purge Fail
Air Blank	0.000	02:38	

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,GARCIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: CABRERA MARROQUIN

DATE/TIME OF TEST: 09/10/2022 21:19:23

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.081	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.000	21:14
Breath Volume	2.578	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.000	21:17
Breath Volume	1.699	21:17
Air Blank	0.000	21:18
Control Test	0.080	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: CRUZ,JORGE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 09/10/2022 21:33:39

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.081	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.122	21:28
Breath Volume	2.832	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:30
Subject Sample # 2	0.127	21:31
Breath Volume	1.281	21:31
Air Blank	0.000	21:32
Control Test	0.080	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: JOSE PASCUAL

DATE/TIME OF TEST: 09/10/2022 23:52:14

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:42	
Control Test	0.081	23:43	
Air Blank	0.000	23:43	
Subject Sample	# 1	23:46	No Sample Provided
Breath Volume	0.000	23:46	
Air Blank	0.000	23:47	
Air Blank	0.000	23:49	
Subject Sample	# 2	23:50	
Breath Volume	2.019	23:50	
Air Blank		23:51	Purge Fail
Air Blank	0.000	23:52	

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: JOSE PASCUAL

DATE/TIME OF TEST: 09/10/2022 23:59:47

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.076	23:56	
Air Blank	0.000	23:56	
Subject Sample	# 1 0.192	23:58	
Breath Volume	1.332	23:58	
Air Blank		23:59	Purge Fail
Air Blank	0.000	23:59	

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: JOSE PASCUAL

DATE/TIME OF TEST: 09/11/2022 00:10:21

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.075	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.207	00:05
Breath Volume	1.625	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:07
Subject Sample # 2	0.191	00:08
Breath Volume	3.152	00:08
Air Blank	0.000	00:09
Control Test	0.077	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCORE,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: SINCORE,MICHAEL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 09/13/2022 14:54:28

Diagnostics Check	OK	14:51	
Air Blank	0.000	14:51	
Control Test	0.081	14:52	
Air Blank	0.000	14:52	
Subject Sample	# 1	14:53	Subject Test Refused
Breath Volume	0.000	14:53	
Air Blank	0.000	14:53	
Control Test	0.080	14:53	
Air Blank	0.000	14:54	
Diagnostics Check	OK	14:54	

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 09:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: STROEMER

DATE/TIME OF TEST: 09/15/2022 10:38:15

Diagnostics Check	OK	10:30
Air Blank	0.000	10:31
Control Test	0.081	10:31
Air Blank	0.000	10:32
Subject Sample # 1	0.000	10:33
Breath Volume	1.804	10:33
Air Blank	0.000	10:33
Air Blank	0.000	10:35
Subject Sample # 2	0.000	10:36
Breath Volume	2.214	10:36
Air Blank	0.000	10:37
Control Test	0.080	10:37
Air Blank	0.000	10:38
Diagnostics Check	OK	10:38

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIROGA,DIEGO,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: QUIROGA,DIEGO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 09/16/2022 00:39:50

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.082	00:32	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:35	No Sample Provided
Breath Volume	0.000	00:35	
Air Blank	0.000	00:36	
Air Blank	0.000	00:38	
Subject Sample	# 2	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:38	
Control Test	0.082	00:39	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/16/2022 23:03:46

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.299	22:54
Breath Volume	2.968	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:56
Subject Sample # 2	0.259	22:58
Breath Volume	1.464	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:00
Subject Sample # 3	0.267	23:01
Breath Volume	3.148	23:01
Air Blank	0.000	23:02
Control Test	0.077	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 09/19/2022 19:37:01

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:28	
Control Test	0.081	19:29	
Air Blank	0.000	19:29	
Subject Sample	# 1	19:32	No Sample Provided
Breath Volume	0.000	19:32	
Air Blank	0.000	19:33	
Air Blank	0.000	19:35	
Subject Sample	# 2	19:35	Subject Test Refused
Breath Volume	0.000	19:35	
Air Blank	0.000	19:36	
Control Test	0.082	19:36	
Air Blank	0.000	19:36	
Diagnostics Check	OK	19:37	

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINNOCENT,L,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: INNOCENT,L
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PEDROMATEO

DATE/TIME OF TEST: 09/20/2022 01:34:39

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.081	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Control Test	0.082	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,CARLOS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 09/22/2022 20:33:47

Diagnostics Check	OK	20:27
Air Blank	0.000	20:27
Control Test	0.081	20:28
Air Blank	0.000	20:28
Subject Sample # 1	0.000	20:29
Breath Volume	1.921	20:29
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.000	20:32
Breath Volume	2.269	20:32
Air Blank	0.000	20:32
Control Test	0.081	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 09/30/2022 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: STEPHENSON,JENNIFER
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 09/22/2022 23:30:58

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:28	
Control Test	0.082	23:28	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:30	
Control Test	0.082	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 09/30/2022 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: JIMENEZ MARTIN

DATE/TIME OF TEST: 09/23/2022 22:19:46

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.081	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.159	22:14
Breath Volume	1.152	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.147	22:17
Breath Volume	1.757	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 09/30/2022 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: BONILLA MORAN

DATE/TIME OF TEST: 09/24/2022 21:50:05

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.082	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.183	21:44
Breath Volume	1.496	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.175	21:48
Breath Volume	2.371	21:48
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 09/30/2022 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: SHEFFIELD JR

DATE/TIME OF TEST: 09/06/2022 01:48:07

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.167	01:43
Breath Volume	1.640	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:45
Subject Sample # 2	0.161	01:46
Breath Volume	1.648	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 10/05/2022 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 09/06/2022 02:54:45

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.081	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.140	02:50
Breath Volume	1.273	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.143	02:53
Breath Volume	1.535	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 10/05/2022 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEAN,ALEXIA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 06:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:14

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 09/08/2022 08:46:51

Diagnostics Check	OK	08:38
Air Blank	0.000	08:38
Control Test	0.081	08:39
Air Blank	0.000	08:39
Subject Sample # 1	0.127	08:40
Breath Volume	1.808	08:40
Air Blank	0.000	08:41
Air Blank	0.000	08:43
Subject Sample # 2	0.121	08:45
Breath Volume	2.085	08:45
Air Blank	0.000	08:45
Control Test	0.080	08:46
Air Blank	0.000	08:46
Diagnostics Check	OK	08:46

Data Download Date/Time: 10/05/2022 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RUSSELL,A
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: WOODARD JR

DATE/TIME OF TEST: 09/10/2022 01:54:30

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.081	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.128	01:49
Breath Volume	2.023	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.137	01:52
Breath Volume	2.445	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 10/05/2022 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LIPTOK,LAUREN
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 09/10/2022 03:14:25

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.081	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.158	03:09
Breath Volume	1.363	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.161	03:12
Breath Volume	1.769	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 10/05/2022 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RYAN,JASON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: RODRIGUEZ LEMUS

DATE/TIME OF TEST: 09/10/2022 23:11:05

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.176	23:06
Breath Volume	3.187	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:08
Subject Sample # 2	0.175	23:09
Breath Volume	2.812	23:09
Air Blank	0.000	23:10
Control Test	0.078	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:11

Data Download Date/Time: 10/05/2022 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCKENNA,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: FARIAS VELAZQUEZ

DATE/TIME OF TEST: 09/14/2022 23:43:03

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:30	
Control Test	0.081	23:30	
Air Blank	0.000	23:31	
Subject Sample	# 1 0.126	23:34	Volume Not Met
Breath Volume	0.398	23:34	
Air Blank	0.000	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 2 0.129	23:38	
Breath Volume	1.226	23:38	
Air Blank	0.000	23:39	
Air Blank	0.000	23:40	
Subject Sample	# 3 0.131	23:41	
Breath Volume	1.316	23:41	
Air Blank	0.000	23:42	
Control Test	0.077	23:42	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 10/05/2022 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 09/16/2022 21:43:08

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.313	21:36
Breath Volume	1.773	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.303	21:41
Breath Volume	1.648	21:41
Air Blank	0.000	21:42
Control Test	0.078	21:42
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 10/05/2022 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GALLAGHER,CHRIS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: DELVA

DATE/TIME OF TEST: 09/17/2022 05:54:21

Diagnostics Check	OK	05:47
Air Blank	0.000	05:48
Control Test	0.081	05:48
Air Blank	0.000	05:49
Subject Sample # 1	0.162	05:49
Breath Volume	3.253	05:49
Air Blank	0.000	05:50
Air Blank	0.000	05:52
Subject Sample # 2	0.157	05:52
Breath Volume	2.351	05:52
Air Blank	0.000	05:53
Control Test	0.079	05:53
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

Data Download Date/Time: 10/05/2022 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: DA SILVA ALFONCO

DATE/TIME OF TEST: 09/17/2022 06:20:11

Diagnostics Check	OK	06:13
Air Blank	0.000	06:13
Control Test	0.079	06:13
Air Blank	0.000	06:14
Subject Sample # 1	0.121	06:15
Breath Volume	2.402	06:15
Air Blank	0.000	06:16
Air Blank	0.000	06:17
Subject Sample # 2	0.120	06:18
Breath Volume	1.531	06:18
Air Blank	0.000	06:19
Control Test	0.079	06:19
Air Blank	0.000	06:20
Diagnostics Check	OK	06:20

Data Download Date/Time: 10/05/2022 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 09/18/2022 05:09:52

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:58	
Control Test	0.082	04:58	
Air Blank	0.000	04:59	
Subject Sample	# 1 0.157	05:02	Volume Not Met
Breath Volume	0.101	05:02	
Air Blank	0.000	05:03	
Air Blank	0.000	05:05	
Subject Sample	# 2 0.014	05:08	Volume Not Met
Breath Volume	0.019	05:08	
Air Blank	0.000	05:08	
Control Test	0.079	05:09	
Air Blank	0.000	05:09	
Diagnostics Check	OK	05:09	

Data Download Date/Time: 10/05/2022 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/18/2022 20:35:43

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.081	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.165	20:31
Breath Volume	2.191	20:31
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.163	20:34
Breath Volume	1.394	20:34
Air Blank	0.000	20:34
Control Test	0.079	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 10/05/2022 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/18/2022 22:42:58

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.224	22:38
Breath Volume	1.773	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:40
Subject Sample # 2	0.227	22:41
Breath Volume	2.011	22:41
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 10/05/2022 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: GRANADOS

DATE/TIME OF TEST: 09/19/2022 04:01:21

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.200	03:56
Breath Volume	1.691	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.200	03:59
Breath Volume	1.523	03:59
Air Blank	0.000	04:00
Control Test	0.077	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 10/05/2022 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LIPTAK,LAUREN
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: THEAS

DATE/TIME OF TEST: 09/26/2022 01:35:09

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.082	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.175	01:30
Breath Volume	1.511	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.176	01:33
Breath Volume	1.773	01:33
Air Blank	0.000	01:34
Control Test	0.079	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 10/05/2022 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEONE,DANIEL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/26/2022 02:10:46

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.184	02:05
Breath Volume	3.992	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.177	02:08
Breath Volume	4.234	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 10/05/2022 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: GARRIQUES

DATE/TIME OF TEST: 09/26/2022 03:08:41

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.147	03:04
Breath Volume	2.019	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.147	03:06
Breath Volume	2.929	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 10/05/2022 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MENDOZA,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: RUBIO MAR

DATE/TIME OF TEST: 09/29/2022 20:27:52

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.081	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.298	20:23
Breath Volume	1.742	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.280	20:26
Breath Volume	2.281	20:26
Air Blank	0.000	20:26
Control Test	0.080	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 10/05/2022 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 09/30/2022 03:52:32

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.082	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.149	03:47
Breath Volume	1.218	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.141	03:50
Breath Volume	1.894	03:50
Air Blank	0.000	03:51
Control Test	0.083	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 11/03/2022 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WELLS,AUSTIN,D

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: WIBERLY,RYAN

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 3521080A3

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: DILL

DATE/TIME OF TEST: 09/03/2022 03:05:39

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.128	03:00
Breath Volume	3.585	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.120	03:03
Breath Volume	2.433	03:03
Air Blank	0.000	03:04
Control Test	0.077	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 09/26/2022 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WELLS,AUSTIN,D

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: EARLYWINE,GARETT

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 3521080A3

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/20/2022 08:24:13

Diagnostics Check	OK	08:17
Air Blank	0.000	08:17
Control Test	0.078	08:18
Air Blank	0.000	08:18
Subject Sample # 1	0.224	08:19
Breath Volume	1.156	08:19
Air Blank	0.000	08:20
Air Blank	0.000	08:22
Subject Sample # 2	0.208	08:22
Breath Volume	1.187	08:22
Air Blank	0.000	08:23
Control Test	0.076	08:23
Air Blank	0.000	08:24
Diagnostics Check	OK	08:24

Data Download Date/Time: 09/26/2022 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: Yes
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 3521080A3
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: GONZALEZ,CHRISTOPHER	ARREST TIME: 00:18
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 08/29/2022	OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MULLIN	DATE/TIME OF TEST: 09/21/2022	00:47:02
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.075	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.145	00:42
Breath Volume	1.546	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:44
Subject Sample # 2	0.161	00:45
Breath Volume	1.132	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:47

Data Download Date/Time: 09/26/2022 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 3521080A3
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: EARLYWINE,GARETT	ARREST TIME: 02:00
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Female
LAST AGENCY INSPECTION: 08/29/2022	OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WALKER DATE/TIME OF TEST: 09/21/2022 02:21:52

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.083	02:21	Interferent Detect
Air Blank	0.000	02:21	

Data Download Date/Time: 09/26/2022 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WELLS,AUSTIN,D

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: EARLYWINE,GARETT

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 3521080A3

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/21/2022 02:28:04

Diagnostics Check

Fail

02:27

Internal Printer Test

Air Blank

0.000

02:28

Data Download Date/Time: 09/26/2022 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WELLS,AUSTIN,D

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: EARLYWINE,GARETT

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 3521080A3

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/21/2022 02:31:14

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	RFI Detect
Air Blank	0.000	02:31	

Data Download Date/Time: 09/26/2022 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 3521080A3
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: EARLYWINE,GARETT	ARREST TIME: 02:00
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Female
LAST AGENCY INSPECTION: 08/29/2022	OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WALKER DATE/TIME OF TEST: 09/21/2022 02:46:22

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.076	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1 0.173	02:37	Volume Not Met
Breath Volume	0.972	02:37	
Air Blank	0.000	02:38	
Air Blank	0.000	02:40	
Subject Sample	# 2 0.177	02:41	
Breath Volume	1.570	02:41	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample	# 3 0.172	02:44	
Breath Volume	1.644	02:44	
Air Blank	0.000	02:45	
Control Test	0.073	02:45	Control Outside Tolerance
Air Blank	0.000	02:46	

Data Download Date/Time: 09/26/2022 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 3521080A3
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: PALFY,RACHEL	ARREST TIME: 02:44
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 08/29/2022	OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MARCELO	DATE/TIME OF TEST: 09/25/2022	03:02:10
------------------------------	-------------------------------	----------

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.084	02:57
Breath Volume	2.585	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.082	03:00
Breath Volume	2.800	03:00
Air Blank	0.000	03:01
Control Test	0.077	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 09/26/2022 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 03521080A3
BREATH TEST OPERATOR: NAEYAERT,JEFF,H	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: NAEYAERT,JEFF	ARREST TIME: 00:09
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 09/02/2022	OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RHODEN DATE/TIME OF TEST: 09/04/2022 00:56:22

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.082	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.214	00:51
Breath Volume	2.531	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.211	00:54
Breath Volume	2.539	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 10/28/2022 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0352108A3
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/02/2022 21:35:21

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.080	21:29
Air Blank	0.000	21:30
Subject Sample # 1	0.074	21:30
Breath Volume	2.023	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.082	21:33
Breath Volume	1.980	21:33
Air Blank	0.000	21:34
Control Test	0.079	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 09/27/2022 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0352108A3
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: FRYSON

DATE/TIME OF TEST: 09/03/2022 02:24:56

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.196	02:20
Breath Volume	1.636	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.206	02:23
Breath Volume	1.679	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 09/27/2022 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0352108A3
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: GUTIERREZ MARTINEZ

DATE/TIME OF TEST: 09/04/2022 00:29:37

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.079	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.162	00:24
Breath Volume	2.285	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.168	00:27
Breath Volume	3.597	00:27
Air Blank	0.000	00:28
Control Test	0.076	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 09/27/2022 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0352108A3
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 09/04/2022 17:22:18

Diagnostics Check	OK	17:15
Air Blank	0.000	17:16
Control Test	0.080	17:16
Air Blank	0.000	17:17
Subject Sample # 1	0.279	17:17
Breath Volume	2.695	17:17
Air Blank	0.000	17:18
Air Blank	0.000	17:20
Subject Sample # 2	0.265	17:20
Breath Volume	2.253	17:20
Air Blank	0.000	17:21
Control Test	0.075	17:21
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

Data Download Date/Time: 09/27/2022 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 0352108A3
BREATH TEST OPERATOR: ALLEN,JAMES,	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop H	VIOLATION CODE: DUI
ARRESTING OFFICER: ALLEN,JAMES	ARREST TIME: 01:24
ARRESTING OFFICER AGENCY: FHP Troop H	GENDER: Male
LAST AGENCY INSPECTION: 08/18/2022	OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: OGLETREE DATE/TIME OF TEST: 09/05/2022 02:03:19

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.078	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.074	01:58
Breath Volume	3.093	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:00
Subject Sample # 2	0.080	02:01
Breath Volume	4.187	02:01
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 09/27/2022 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOSCHETTO, ANTHONY,

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: MOSCHETTO, ANTHONY

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 0352108A3

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/06/2022 03:13:08

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.191	03:08
Breath Volume	2.851	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.188	03:11
Breath Volume	2.964	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 09/27/2022 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0352108A3
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 09/06/2022 21:42:00

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:41	
Control Test	0.078	21:41	
Air Blank		21:41	RFI Detect
Air Blank	0.000	21:42	

Data Download Date/Time: 09/27/2022 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0352108A3
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 09/06/2022 21:49:00

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:44	
Control Test	0.078	21:44	
Air Blank	0.000	21:45	
Subject Sample	# 1	21:47	Subject Test Refused
Breath Volume	0.000	21:47	
Air Blank	0.000	21:48	
Control Test	0.077	21:48	
Air Blank	0.000	21:48	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 09/27/2022 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0352108A3
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: HARLEY

DATE/TIME OF TEST: 09/24/2022 19:42:25

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.098	19:37
Breath Volume	1.730	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.094	19:40
Breath Volume	1.425	19:40
Air Blank	0.000	19:41
Control Test	0.080	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 09/27/2022 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0352108A3
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: BULLARD

DATE/TIME OF TEST: 09/26/2022 21:39:54

Diagnostics Check	OK	21:31	
Air Blank	0.000	21:32	
Control Test	0.080	21:32	
Air Blank	0.000	21:32	
Subject Sample # 1	0.000	21:33	
Breath Volume	1.480	21:33	
Air Blank	0.000	21:33	
Air Blank	0.000	21:35	
Subject Sample # 2		21:38	Subject Test Refused
Breath Volume	0.632	21:38	
Air Blank	0.000	21:38	
Control Test	0.078	21:39	
Air Blank	0.000	21:39	
Diagnostics Check	OK	21:39	

Data Download Date/Time: 09/27/2022 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 03521080A3
BREATH TEST OPERATOR: RIVERA,VICTOR,J	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop D	VIOLATION CODE: DUI
ARRESTING OFFICER: RIVERA,VICTOR	ARREST TIME: 23:15
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Male
LAST AGENCY INSPECTION: 08/10/2022	OBSERVATION PERIOD BEGAN: 23:31
SUBJECT'S LAST NAME: DIAZ	DATE/TIME OF TEST: 09/16/2022 23:52:22

Diagnostics Check	Fail	23:51	Internal Printer Test
Air Blank	0.000	23:52	

Data Download Date/Time: 09/23/2022 08:49

Subject Test Electronic Data

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIVERA,VICTOR,J

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: RIVERA,VICTOR

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 08/10/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 03521080A3

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/17/2022 00:11:21

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.078	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.150	00:06
Breath Volume	1.792	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.144	00:09
Breath Volume	1.640	00:09
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 09/23/2022 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 03521080A3
BREATH TEST OPERATOR: RIVERA,VICTOR,J	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop D	VIOLATION CODE: DUI
ARRESTING OFFICER: MARTINEZ,SERGIO	ARREST TIME: 22:07
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Male
LAST AGENCY INSPECTION: 08/10/2022	OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: IGNOT ROSARIO	DATE/TIME OF TEST: 09/17/2022 23:24:31
------------------------------------	--

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.079	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.195	23:19
Breath Volume	1.562	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.186	23:22
Breath Volume	2.339	23:22
Air Blank	0.000	23:23
Control Test	0.081	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 09/23/2022 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MAYHAND

DATE/TIME OF TEST: 09/03/2022 01:30:23

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.081	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.078	01:25
Breath Volume	1.828	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.084	01:28
Breath Volume	1.472	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 09/28/2022 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 09/03/2022 21:58:59

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.080	21:53
Air Blank	0.000	21:53
Subject Sample # 1	0.233	21:54
Breath Volume	1.917	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:56
Subject Sample # 2	0.231	21:57
Breath Volume	1.769	21:57
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 09/28/2022 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MCCLELLAND

DATE/TIME OF TEST: 09/03/2022 22:03:09

Diagnostics Check	Fail	22:02	Internal Printer Test
Air Blank	0.000	22:03	

Data Download Date/Time: 09/28/2022 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MCCLELLAND

DATE/TIME OF TEST: 09/03/2022 22:18:29

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:05	
Control Test	0.080	22:05	
Air Blank	0.000	22:06	
Subject Sample	# 1 0.215	22:08	
Breath Volume	1.164	22:08	
Air Blank	0.000	22:09	
Air Blank	0.000	22:10	
Subject Sample	# 2 0.195	22:13	Volume Not Met
Breath Volume	0.226	22:13	
Air Blank	0.000	22:14	
Air Blank	0.000	22:16	
Subject Sample	# 3 0.219	22:16	
Breath Volume	1.347	22:16	
Air Blank	0.000	22:17	
Control Test	0.079	22:17	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 09/28/2022 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MORRIS,CHASE
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: IACOVINO

DATE/TIME OF TEST: 09/09/2022 22:08:07

Diagnostics Check	OK	21:59	
Air Blank	0.000	22:00	
Control Test	0.082	22:00	
Air Blank	0.000	22:00	
Subject Sample	# 1	22:04	No Sample Provided
Breath Volume	0.000	22:04	
Air Blank	0.000	22:04	
Air Blank	0.000	22:06	
Subject Sample	# 2	22:06	Subject Test Refused
Breath Volume	0.000	22:06	
Air Blank	0.000	22:07	
Control Test	0.081	22:07	
Air Blank	0.000	22:08	
Diagnostics Check	OK	22:08	

Data Download Date/Time: 09/28/2022 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: DELANEY,TYELER
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/09/2022 23:08:47

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.098	23:03
Breath Volume	1.632	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:05
Subject Sample # 2	0.102	23:07
Breath Volume	1.503	23:07
Air Blank	0.000	23:07
Control Test	0.080	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 09/28/2022 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HOGAN,JEREMY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 09/10/2022 00:20:06

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.198	00:15
Breath Volume	1.945	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:17
Subject Sample # 2	0.187	00:18
Breath Volume	1.660	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 09/28/2022 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 09/11/2022 02:17:32

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.252	02:12
Breath Volume	1.722	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.250	02:15
Breath Volume	1.863	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 09/28/2022 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PATTISON

DATE/TIME OF TEST: 09/15/2022 02:31:27

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.227	02:26
Breath Volume	1.343	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.229	02:29
Breath Volume	2.386	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 09/28/2022 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 11:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 09/16/2022 23:40:29

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.082	23:26	
Air Blank	0.000	23:26	
Subject Sample	# 1 0.178	23:29	
Breath Volume	1.273	23:29	
Air Blank	0.000	23:30	
Air Blank	0.000	23:32	
Subject Sample	# 2 0.169	23:35	Volume Not Met
Breath Volume	0.507	23:35	
Air Blank	0.000	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 3	23:38	Subject Test Refused
Breath Volume	0.445	23:38	
Air Blank	0.000	23:39	
Control Test	0.080	23:39	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:40	

Data Download Date/Time: 09/28/2022 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SMITH,CALE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: VOGLESSONG

DATE/TIME OF TEST: 09/17/2022 01:02:25

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.184	00:57
Breath Volume	1.406	00:57
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.189	01:00
Breath Volume	1.457	01:00
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 09/28/2022 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/17/2022 04:25:16

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:22	
Control Test	0.081	04:22	
Air Blank	0.000	04:23	
Subject Sample	# 1	04:23	Subject Test Refused
Breath Volume	0.000	04:23	
Air Blank	0.000	04:24	
Control Test	0.081	04:24	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 09/28/2022 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JOSE,L
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: LOWERY,JOHN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:07

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 09/03/2022 08:43:54

Diagnostics Check	OK	08:32	
Air Blank	0.000	08:32	
Control Test	0.080	08:33	
Air Blank	0.000	08:33	
Subject Sample	# 1 0.052	08:36	Volume Not Met
Breath Volume	0.000	08:36	
Air Blank	0.000	08:37	
Air Blank	0.000	08:39	
Subject Sample	# 2 0.055	08:42	Volume Not Met
Breath Volume	0.031	08:42	
Air Blank	0.000	08:42	
Control Test	0.077	08:43	
Air Blank	0.000	08:43	
Diagnostics Check	OK	08:43	

Data Download Date/Time: 09/18/2022 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JOSE,L
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: LOWERY,JOHN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:07

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 09/03/2022 09:02:54

Diagnostics Check	OK	08:51
Air Blank	0.000	08:52
Control Test	0.080	08:52
Air Blank	0.000	08:53
Subject Sample # 1	0.053	08:55
Breath Volume	1.820	08:55
Air Blank	0.000	08:56
Air Blank	0.000	08:58
Subject Sample # 2	0.050	09:01
Breath Volume	1.539	09:01
Air Blank	0.000	09:01
Control Test	0.078	09:02
Air Blank	0.000	09:02
Diagnostics Check	OK	09:02

Data Download Date/Time: 09/18/2022 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAWRENCE,JOSHUA,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: THOMAS,MEGAN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG129201

DRY GAS EXPIRY DATE: 10/19/2023

VIOLATION CODE: BUI

ARREST TIME: 17:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: HOUGH

DATE/TIME OF TEST: 09/02/2022 18:09:43

Diagnostics Check

Fail

18:09

Internal Printer Test

Air Blank

0.000

18:09

Data Download Date/Time: 09/24/2022 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAWRENCE,JOSHUA,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: THOMAS,MEGAN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG129201

DRY GAS EXPIRY DATE: 10/19/2023

VIOLATION CODE: BUI

ARREST TIME: 17:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: HOUGH

DATE/TIME OF TEST: 09/02/2022 18:18:50

Diagnostics Check	OK	18:12
Air Blank	0.000	18:12
Control Test	0.076	18:13
Air Blank	0.000	18:13
Subject Sample # 1	0.136	18:14
Breath Volume	1.457	18:14
Air Blank	0.000	18:15
Air Blank	0.000	18:16
Subject Sample # 2	0.150	18:17
Breath Volume	1.785	18:17
Air Blank	0.000	18:17
Control Test	0.076	18:18
Air Blank	0.000	18:18
Diagnostics Check	OK	18:18

Data Download Date/Time: 09/24/2022 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILMORE,RICHARD,L

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: HANEY,MIKE

ARRESTING OFFICER AGENCY: FFWCC (Lake City)

LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: HINSON

DATE/TIME OF TEST: 09/05/2022 20:13:30

Diagnostics Check	OK	20:07
Air Blank	0.000	20:07
Control Test	0.078	20:08
Air Blank	0.000	20:08
Subject Sample # 1	0.079	20:08
Breath Volume	2.933	20:08
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.075	20:11
Breath Volume	2.312	20:11
Air Blank	0.000	20:12
Control Test	0.078	20:12
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 10/03/2022 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 15421080A1
BREATH TEST OPERATOR: GILMORE,RICHARD,L	DRY GAS EXPIRY DATE: 08/05/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)	VIOLATION CODE: BUI
ARRESTING OFFICER: HAUSLER,TREVOR	ARREST TIME: 21:32
ARRESTING OFFICER AGENCY: FFWCC (Lake City)	GENDER: Male
LAST AGENCY INSPECTION: 08/17/2022	OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: WHITE DATE/TIME OF TEST: 09/05/2022 22:34:43

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.078	22:29
Air Blank	0.000	22:29
Subject Sample # 1	0.000	22:30
Breath Volume	3.726	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.000	22:33
Breath Volume	2.804	22:33
Air Blank	0.000	22:33
Control Test	0.076	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 10/03/2022 14:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007503

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 09/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 19021080A2

DRY GAS EXPIRY DATE: 09/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 09/24/2022 01:30:12

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.167	01:25
Breath Volume	2.722	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.163	01:28
Breath Volume	2.906	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 10/24/2022 12:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007503

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 09/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 19021080A2

DRY GAS EXPIRY DATE: 09/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:24

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: QUINTANILLA

DATE/TIME OF TEST: 09/25/2022 02:26:34

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.132	02:22
Breath Volume	2.531	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.121	02:25
Breath Volume	1.476	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 10/24/2022 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007503

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGES,MARCO,A

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: HODGES,MARCO

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 09/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 19021080A2

DRY GAS EXPIRY DATE: 09/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 09/26/2022 04:19:28

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.079	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1 0.196	04:12	Volume Not Met
Breath Volume	0.105	04:12	
Air Blank	0.000	04:12	
Air Blank	0.000	04:14	
Subject Sample	# 2	04:17	No Sample Provided
Breath Volume	0.000	04:17	
Air Blank	0.000	04:18	
Control Test	0.078	04:18	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 10/24/2022 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,ALBERT,J

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: WILLIAMS,ALBERT

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 19021080A2

DRY GAS EXPIRY DATE: 09/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: LEFFEL

DATE/TIME OF TEST: 09/02/2022 22:30:01

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:19	
Control Test	0.081	22:19	
Air Blank	0.000	22:19	
Subject Sample	# 1	22:23	No Sample Provided
Breath Volume	0.000	22:23	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample	# 2	22:28	No Sample Provided
Breath Volume	0.000	22:28	
Air Blank	0.000	22:29	
Control Test	0.082	22:29	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:30	

Data Download Date/Time: 09/23/2022 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BLACKMAN,GERIG,A

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: BLACKMAN,GERIG

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 08/08/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 19021080A2

DRY GAS EXPIRY DATE: 09/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 09/13/2022 00:57:33

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.000	00:53
Breath Volume	1.546	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.000	00:56
Breath Volume	1.464	00:56
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 09/23/2022 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,E
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/03/2022 01:04:25

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.079	01:01	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:03	
Control Test	0.079	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 09/26/2022 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTE,KYLE,M
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BEUFORD,S
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WERT

DATE/TIME OF TEST: 09/10/2022 02:13:30

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.079	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1 0.204	02:02	
Breath Volume	1.160	02:02	
Air Blank	0.000	02:03	
Air Blank	0.000	02:05	
Subject Sample	# 2 0.176	02:08	Volume Not Met
Breath Volume	0.996	02:08	
Air Blank	0.000	02:09	
Air Blank	0.000	02:10	
Subject Sample	# 3 0.190	02:11	
Breath Volume	1.566	02:11	
Air Blank	0.000	02:12	
Control Test	0.077	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 09/26/2022 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SVENDSEN,C,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SVENDSEN,C
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 09/13/2022 13:14:33

Diagnostics Check	OK	13:08
Air Blank	0.000	13:08
Control Test	0.079	13:09
Air Blank	0.000	13:09
Subject Sample # 1	0.000	13:10
Breath Volume	3.156	13:10
Air Blank	0.000	13:10
Air Blank	0.000	13:12
Subject Sample # 2	0.000	13:13
Breath Volume	2.273	13:13
Air Blank	0.000	13:13
Control Test	0.079	13:13
Air Blank	0.000	13:14
Diagnostics Check	OK	13:14

Data Download Date/Time: 09/26/2022 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRETT,RYAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/13/2022 22:07:19

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.303	22:02
Breath Volume	2.761	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:04
Subject Sample # 2	0.285	22:05
Breath Volume	1.996	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 09/26/2022 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,CALEB,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LOTT,CALEB
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: COURT ORD
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: PEREZ ALVAREZ

DATE/TIME OF TEST: 09/15/2022 12:17:24

Diagnostics Check	OK	12:11
Air Blank	0.000	12:11
Control Test	0.079	12:11
Air Blank	0.000	12:12
Subject Sample # 1	0.000	12:12
Breath Volume	2.890	12:12
Air Blank	0.000	12:13
Air Blank	0.000	12:15
Subject Sample # 2	0.000	12:16
Breath Volume	2.531	12:16
Air Blank	0.000	12:16
Control Test	0.080	12:16
Air Blank	0.000	12:17
Diagnostics Check	OK	12:17

Data Download Date/Time: 09/26/2022 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: FORBES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: CARRIGAN

DATE/TIME OF TEST: 09/15/2022 19:23:40

Diagnostics Check	OK	19:20	
Air Blank	0.000	19:20	
Control Test	0.079	19:21	
Air Blank	0.000	19:21	
Subject Sample	# 1	19:22	Subject Test Refused
Breath Volume	0.000	19:22	
Air Blank	0.000	19:22	
Control Test	0.080	19:22	
Air Blank	0.000	19:23	
Diagnostics Check	OK	19:23	

Data Download Date/Time: 09/26/2022 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GILLIARD,STEVEN
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: KAHL

DATE/TIME OF TEST: 09/16/2022 19:48:16

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.079	19:42
Air Blank	0.000	19:42
Subject Sample # 1	0.085	19:43
Breath Volume	2.734	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:45
Subject Sample # 2	0.077	19:46
Breath Volume	2.132	19:46
Air Blank	0.000	19:47
Control Test	0.079	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 09/26/2022 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: OMALLEY,MICHAEL
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/18/2022 03:52:50

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.197	03:47
Breath Volume	1.589	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.192	03:51
Breath Volume	2.960	03:51
Air Blank	0.000	03:51
Control Test	0.077	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 09/26/2022 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,R,C
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GILLIARD,S
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 08/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/19/2022 22:44:02

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:41	
Control Test	0.080	22:41	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:42	
Control Test	0.080	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:44	

Data Download Date/Time: 09/26/2022 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,CALEB,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: JONES,WESLEY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GOYETTE

DATE/TIME OF TEST: 09/25/2022 03:33:40

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.080	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 09/26/2022 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIEL,KYLE,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BAKER,DUSTIN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/03/2022 03:39:55

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.082	03:35
Breath Volume	1.488	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.082	03:38
Breath Volume	1.851	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIEL,KYLE,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: FORNAL,DOMINC
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: MADISON

DATE/TIME OF TEST: 09/04/2022 06:10:32

Diagnostics Check	OK	06:02
Air Blank	0.000	06:03
Control Test	0.080	06:03
Air Blank	0.000	06:04
Subject Sample # 1	0.226	06:05
Breath Volume	1.914	06:05
Air Blank	0.000	06:06
Air Blank	0.000	06:08
Subject Sample # 2	0.224	06:08
Breath Volume	1.542	06:08
Air Blank	0.000	06:09
Control Test	0.078	06:09
Air Blank	0.000	06:10
Diagnostics Check	OK	06:10

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: KARR,SEAN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: BUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: HERSHBERGER

DATE/TIME OF TEST: 09/04/2022 17:25:49

Diagnostics Check	OK	17:12	
Air Blank	0.000	17:13	
Control Test	0.079	17:13	
Air Blank	0.000	17:14	
Subject Sample	# 1 0.140	17:17	Volume Not Met
Breath Volume	0.144	17:17	
Air Blank	0.000	17:18	
Air Blank	0.000	17:19	
Subject Sample	# 2 0.164	17:20	
Breath Volume	1.503	17:20	
Air Blank	0.000	17:21	
Air Blank	0.000	17:23	
Subject Sample	# 3 0.153	17:24	
Breath Volume	1.230	17:24	
Air Blank	0.000	17:24	
Control Test	0.075	17:25	
Air Blank	0.000	17:25	
Diagnostics Check	OK	17:25	

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DITTMAN, LAVERNE, T
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: FORNAL, DOMINIC
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 09/04/2022 23:34:30

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.172	23:29
Breath Volume	1.683	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.163	23:32
Breath Volume	2.285	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIEL,KYLE,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BAKER,DUSTIN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: QUEZADA

DATE/TIME OF TEST: 09/05/2022 01:30:38

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:28	
Control Test	0.079	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:29	Subject Test Refused
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Control Test	0.079	01:30	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: FORNAL,DOMINIC
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: PABON

DATE/TIME OF TEST: 09/08/2022 22:40:44

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.078	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.116	22:35
Breath Volume	2.828	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.111	22:39
Breath Volume	1.886	22:39
Air Blank	0.000	22:39
Control Test	0.078	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: FORNAL,DOMINIC
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 09/17/2022 03:30:12

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.138	03:25
Breath Volume	2.007	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.140	03:28
Breath Volume	1.886	03:28
Air Blank	0.000	03:29
Control Test	0.075	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIEL,KYLE,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: POVEDA,CARLOS
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: SANTIZ

DATE/TIME OF TEST: 09/18/2022 02:43:48

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:37	
Control Test	0.080	02:37	
Air Blank	0.000	02:38	
Subject Sample	# 1 0.141	02:38	
Breath Volume	1.699	02:38	
Air Blank	0.000	02:39	
Air Blank	0.000	02:41	
Subject Sample	# 2 0.145	02:42	
Breath Volume	2.105	02:42	
Air Blank	0.000	02:42	
Control Test	0.079	02:43	
Air Blank		02:43	RFI Detect
Air Blank	0.000	02:43	

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIEL,KYLE,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BAKER,DUSTIN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 09/18/2022 02:54:43

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.135	02:50
Breath Volume	2.023	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.127	02:53
Breath Volume	2.039	02:53
Air Blank	0.000	02:53
Control Test	0.076	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIEL,KYLE,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: POVEDA,CARLOS
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: VILLEGAS HERNANDEZ

DATE/TIME OF TEST: 09/19/2022 04:06:39

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.258	03:58
Breath Volume	2.269	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.236	04:01
Breath Volume	1.878	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 3	0.229	04:04
Breath Volume	1.410	04:04
Air Blank	0.000	04:05
Control Test	0.077	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BAKER,DUSTIN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: POWELL JR

DATE/TIME OF TEST: 09/23/2022 00:33:24

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.241	00:28
Breath Volume	1.460	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.246	00:31
Breath Volume	1.742	00:31
Air Blank	0.000	00:32
Control Test	0.078	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DITTMAN, LAVERNE, T
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: FORNAL, DOMINIC
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: JIMENEZ GARCIA

DATE/TIME OF TEST: 09/23/2022 04:37:06

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:34	
Control Test	0.079	04:34	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:36	
Control Test	0.079	04:36	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIEL,KYLE,J
BREATH TEST OPERATOR AGENCY: Arcadia PD
ARRESTING OFFICER: RIOS,DANIELLE
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/27/2022 00:24:03

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.080	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:22	Subject Test Refused
Breath Volume	0.000	00:22	
Air Blank	0.000	00:23	
Control Test	0.080	00:23	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 10/24/2022 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: KLAPATCH

DATE/TIME OF TEST: 09/08/2022 00:59:01

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.078	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:58	
Control Test	0.076	00:58	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:59	

Data Download Date/Time: 09/30/2022 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SIDES

DATE/TIME OF TEST: 09/10/2022 03:24:43

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.059	03:20
Breath Volume	1.382	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.065	03:23
Breath Volume	2.187	03:23
Air Blank	0.000	03:23
Control Test	0.078	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 09/30/2022 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: COWART

DATE/TIME OF TEST: 09/12/2022 01:26:40

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.077	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.134	01:21
Breath Volume	2.300	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.129	01:25
Breath Volume	1.707	01:25
Air Blank	0.000	01:25
Control Test	0.076	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 09/30/2022 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 09/15/2022 01:55:26

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:45	
Control Test	0.078	01:46	
Air Blank	0.000	01:46	
Subject Sample	# 1 0.214	01:47	
Breath Volume	1.648	01:47	
Air Blank	0.000	01:48	
Air Blank	0.000	01:50	
Subject Sample	# 2 0.218	01:53	
Breath Volume	2.519	01:53	
Air Blank	0.000	01:53	
Control Test	0.076	01:54	
Air Blank	0.000	01:54	
Diagnostics Check	Fail	01:54	Internal Printer Test
Air Blank	0.000	01:55	

Data Download Date/Time: 09/30/2022 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 09/15/2022 02:02:30

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.077	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:01	
Control Test	0.076	02:01	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 09/30/2022 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: STEINER

DATE/TIME OF TEST: 09/15/2022 23:43:00

Diagnostics Check	OK	23:42	
Air Blank		23:42	RFI Detect
Air Blank	0.000	23:43	

Data Download Date/Time: 09/30/2022 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: STEINER

DATE/TIME OF TEST: 09/15/2022 23:49:08

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:45	
Control Test	0.079	23:45	
Air Blank	0.000	23:46	
Subject Sample # 1	0.069	23:47	
Breath Volume	2.281	23:47	
Air Blank		23:47	RFI Detect
Air Blank	0.000	23:48	
Control Test	0.078	23:48	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time: 09/30/2022 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: STEINER

DATE/TIME OF TEST: 09/15/2022 23:57:45

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.078	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.066	23:52
Breath Volume	1.816	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.064	23:56
Breath Volume	1.558	23:56
Air Blank	0.000	23:56
Control Test	0.077	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 09/30/2022 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: PRAKASH

DATE/TIME OF TEST: 09/20/2022 19:47:44

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.077	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.132	19:43
Breath Volume	2.355	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.129	19:46
Breath Volume	1.972	19:46
Air Blank	0.000	19:46
Control Test	0.077	19:47
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 09/30/2022 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: KIELESZEWSKI

DATE/TIME OF TEST: 09/22/2022 16:29:30

Diagnostics Check	OK	16:22	
Air Blank	0.000	16:23	
Control Test	0.078	16:23	
Air Blank	0.000	16:23	
Subject Sample # 1	0.131	16:24	
Breath Volume	2.843	16:24	
Air Blank	0.000	16:25	
Air Blank	0.000	16:26	
Subject Sample # 2	0.127	16:27	
Breath Volume	3.132	16:27	
Air Blank	0.000	16:28	
Control Test	0.078	16:28	
Air Blank	0.000	16:28	
Diagnostics Check	Fail	16:29	Internal Printer Test
Air Blank	0.000	16:29	

Data Download Date/Time: 09/30/2022 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: CYBULSKI

DATE/TIME OF TEST: 09/23/2022 02:37:30

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.154	02:32
Breath Volume	1.597	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.156	02:35
Breath Volume	2.023	02:35
Air Blank	0.000	02:36
Control Test	0.076	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 09/30/2022 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CRAWFORD,JUDSON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HANCOCK

DATE/TIME OF TEST: 09/24/2022 03:10:24

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.077	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.174	03:05
Breath Volume	2.683	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.168	03:08
Breath Volume	3.394	03:08
Air Blank	0.000	03:09
Control Test	0.077	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 09/30/2022 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILES,STEPHANIE,N
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SILES,STEPHANIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: FROMETA

DATE/TIME OF TEST: 09/03/2022 01:36:09

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.083	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.194	01:31
Breath Volume	1.460	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:33
Subject Sample # 2	0.175	01:34
Breath Volume	1.144	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 09/21/2022 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: GALINDO GOMEZ,MARIA,A	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: LEWIS,SHAREKA	ARREST TIME: 08:46
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Female
LAST AGENCY INSPECTION: 08/25/2022	OBSERVATION PERIOD BEGAN: 09:29

SUBJECT'S LAST NAME: GALINDO GOME	DATE/TIME OF TEST: 09/03/2022	09:59:05
-----------------------------------	-------------------------------	----------

Diagnostics Check	OK	09:51
Air Blank	0.000	09:51
Control Test	0.082	09:51
Air Blank	0.000	09:52
Subject Sample # 1	0.222	09:54
Breath Volume	1.781	09:54
Air Blank	0.000	09:55
Air Blank	0.000	09:56
Subject Sample # 2	0.218	09:57
Breath Volume	2.000	09:57
Air Blank	0.000	09:58
Control Test	0.080	09:58
Air Blank	0.000	09:59
Diagnostics Check	OK	09:59

Data Download Date/Time: 09/21/2022 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,CAMILA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: GONZALEZ,CAMILA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 22:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/04/2022 23:32:14

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.082	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.203	23:27
Breath Volume	1.769	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:29
Subject Sample # 2	0.214	23:30
Breath Volume	2.074	23:30
Air Blank	0.000	23:31
Control Test	0.081	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 09/21/2022 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,JASMIN,I
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,JASMIN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: SENATUS

DATE/TIME OF TEST: 09/09/2022 22:01:10

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:50	
Control Test	0.083	21:51	
Air Blank	0.000	21:51	
Subject Sample	# 1 0.193	21:54	Volume Not Met
Breath Volume	0.093	21:54	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample	# 2 0.213	21:59	
Breath Volume	1.996	21:59	
Air Blank		22:00	Purge Fail
Air Blank	0.000	22:01	

Data Download Date/Time: 09/21/2022 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: UBEDA,D
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 09/10/2022 01:57:34

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.081	01:55	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:56	
Control Test	0.082	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 09/21/2022 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHO,HARRY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VILLEGAS,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 09/10/2022 20:57:08

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:55	
Control Test	0.082	20:56	
Air Blank		20:56	RFI Detect
Air Blank	0.000	20:57	

Data Download Date/Time: 09/21/2022 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHO,HARRY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VILLEGAS,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 09/10/2022 21:04:22

Diagnostics Check	OK	20:58	
Air Blank	0.000	20:59	
Control Test	0.082	20:59	
Air Blank	0.000	21:00	
Subject Sample	# 1 0.185	21:00	
Breath Volume	1.300	21:00	
Air Blank	0.000	21:01	
Air Blank	0.000	21:03	
Subject Sample	# 2	21:03	RFI Detect
Breath Volume	0.671	21:03	
Air Blank		21:04	RFI Detect
Air Blank		21:04	RFI Detect

Data Download Date/Time: 09/21/2022 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHO,HARRY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VILLEGA,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 09/10/2022 21:07:52

Diagnostics Check	OK	21:07	
Air Blank		21:07	RFI Detect
Air Blank		21:07	RFI Detect

Data Download Date/Time: 09/21/2022 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEALY,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WATTLEY,E
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: KAREN MARIELA

DATE/TIME OF TEST: 09/10/2022 22:42:04

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:40	
Control Test	0.083	22:41	
Air Blank		22:41	Ambient Fail
Air Blank	0.000	22:42	

Data Download Date/Time: 09/21/2022 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEALY,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WATTLEY,E
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: KAREN MARIELA

DATE/TIME OF TEST: 09/10/2022 22:52:58

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.081	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.235	22:48
Breath Volume	1.812	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.241	22:51
Breath Volume	1.910	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 09/21/2022 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 04:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: DE LA ROSA

DATE/TIME OF TEST: 09/17/2022 05:42:23

Diagnostics Check	OK	05:36
Air Blank	0.000	05:36
Control Test	0.081	05:36
Air Blank	0.000	05:37
Subject Sample # 1	0.177	05:37
Breath Volume	3.296	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.162	05:40
Breath Volume	3.871	05:40
Air Blank	0.000	05:41
Control Test	0.081	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 09/21/2022 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MONDESIR,JOHNNY,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: DIAZ,ROBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 22:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: BARCELO ECHEVARRIA

DATE/TIME OF TEST: 09/17/2022 23:56:00

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.081	23:48	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:51	No Sample Provided
Breath Volume	0.000	23:51	
Air Blank	0.000	23:52	
Air Blank	0.000	23:54	
Subject Sample	# 2	23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:55	
Control Test	0.082	23:55	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 09/21/2022 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BERMUDEZ,FERNANDO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: BERMUDEZ,FERNANDO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 14:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/20/2022 16:40:54

Diagnostics Check	OK	16:37	
Air Blank	0.000	16:38	
Control Test	0.081	16:38	
Air Blank	0.000	16:39	
Subject Sample	# 1	16:39	Subject Test Refused
Breath Volume	0.000	16:39	
Air Blank	0.000	16:39	
Control Test	0.082	16:40	
Air Blank	0.000	16:40	
Diagnostics Check	OK	16:40	

Data Download Date/Time: 09/21/2022 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,CAMILA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: GONZALEZ,CAMILA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 09/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 01:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: DIAZ MARTIN

DATE/TIME OF TEST: 09/25/2022 02:24:02

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.083	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.082	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 10/21/2022 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLMEDO,REBECCA,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OLMEDO,REBECCA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 13:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: FLORES RODRIGUEZ

DATE/TIME OF TEST: 09/26/2022 14:47:55

Diagnostics Check	OK	14:45	
Air Blank	0.000	14:45	
Control Test	0.083	14:45	
Air Blank	0.000	14:46	
Subject Sample	# 1	14:46	Subject Test Refused
Breath Volume	0.000	14:46	
Air Blank	0.000	14:46	
Control Test	0.082	14:47	
Air Blank	0.000	14:47	
Diagnostics Check	OK	14:47	

Data Download Date/Time: 10/21/2022 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CHITTENDEN
RODRIGUEZ, MICHELLE,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CARRERO, BRANDON

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 22:51

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: CHITTENDEN RODRIGUEZ

DATE/TIME OF TEST: 09/10/2022 00:17:45

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:15	
Control Test	0.079	00:15	
Air Blank	0.000	00:16	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:16	
Control Test	0.080	00:17	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 09/22/2022 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARRERO,BRANDON,R

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CARRERO,BRANDON

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 22:51

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: CHITTENDEN RODRIGUEZ

DATE/TIME OF TEST: 09/10/2022 00:31:17

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.079	00:28	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Control Test	0.079	00:30	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 09/22/2022 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHO,HARRY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VILLEGAS,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 09/10/2022 21:18:32

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.079	21:12
Air Blank	0.000	21:12
Subject Sample # 1	0.187	21:13
Breath Volume	1.324	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.183	21:16
Breath Volume	1.542	21:16
Air Blank	0.000	21:17
Control Test	0.077	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 09/22/2022 12:35