

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: LORENZO,FRANCISCO
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 01/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: CRUZ MERCADO

DATE/TIME OF TEST: 01/08/2023 20:23:18

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:20	
Control Test	0.080	20:20	
Air Blank	0.000	20:21	
Subject Sample	# 1	20:21	Subject Test Refused
Breath Volume	0.000	20:21	
Air Blank	0.000	20:22	
Control Test	0.081	20:22	
Air Blank	0.000	20:23	
Diagnostics Check	OK	20:23	

Data Download Date/Time: 02/08/2023 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULLES,ALBERTO,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: PULLES,ALBERTO
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 01/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: BISCHOFF YAMIN

DATE/TIME OF TEST: 01/15/2023 00:30:22

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.080	00:27	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 02/08/2023 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURBELO MESA,CLAUDIA,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: CURBELO MESA,CLAUDIA
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 01/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/15/2023 03:09:30

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 02/08/2023 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULLES,ALBERTO,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: PULLES,ALBERTO
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 01/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: CARBONELL-ARGUELLES

DATE/TIME OF TEST: 01/15/2023 06:36:10

Diagnostics Check	OK	06:29
Air Blank	0.000	06:30
Control Test	0.081	06:30
Air Blank	0.000	06:30
Subject Sample # 1	0.215	06:31
Breath Volume	1.781	06:31
Air Blank	0.000	06:32
Air Blank	0.000	06:33
Subject Sample # 2	0.195	06:34
Breath Volume	2.296	06:34
Air Blank	0.000	06:35
Control Test	0.079	06:35
Air Blank	0.000	06:36
Diagnostics Check	OK	06:36

Data Download Date/Time: 02/08/2023 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULLES,ALBERTO,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: PULLES,ALBERTO
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 01/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MARRERO GONZALEZ

DATE/TIME OF TEST: 01/20/2023 04:44:30

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.077	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.215	04:39
Breath Volume	2.691	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.201	04:42
Breath Volume	1.605	04:42
Air Blank	0.000	04:43
Control Test	0.075	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 02/08/2023 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: LORENZO,FRANCISCO
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 01/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MORALES VALENCIA

DATE/TIME OF TEST: 01/21/2023 04:10:41

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.276	04:06
Breath Volume	2.746	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:08
Subject Sample # 2	0.278	04:08
Breath Volume	3.027	04:08
Air Blank	0.000	04:09
Control Test	0.077	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 02/08/2023 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASQUERI,JESUS,N
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: ROBBINS,JUSTIN
ARRESTING OFFICER AGENCY: MIAMI SPRINGS PD
LAST AGENCY INSPECTION: 01/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:03

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 01/26/2023 09:29:25

Diagnostics Check	OK	09:25	
Air Blank	0.000	09:25	
Control Test	0.078	09:26	
Air Blank	0.000	09:26	
Subject Sample	# 1	09:28	Subject Test Refused
Breath Volume	0.000	09:28	
Air Blank	0.000	09:28	
Control Test	0.079	09:28	
Air Blank	0.000	09:29	
Diagnostics Check	OK	09:29	

Data Download Date/Time: 02/08/2023 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALDANA,ERICK,N
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: CARRERO,EMMANUEL
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 01/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: VILLAMUES

DATE/TIME OF TEST: 01/28/2023 01:14:15

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.080	01:12	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:13	
Control Test	0.080	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 02/08/2023 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURBELO,CLAUDIA,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: CURBELO,CLAUDIA
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 01/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: LOPEZ MURIEL

DATE/TIME OF TEST: 01/29/2023 06:18:02

Diagnostics Check	OK	06:15	
Air Blank	0.000	06:15	
Control Test	0.079	06:15	
Air Blank	0.000	06:16	
Subject Sample	# 1	06:16	Subject Test Refused
Breath Volume	0.000	06:16	
Air Blank	0.000	06:17	
Control Test	0.078	06:17	
Air Blank	0.000	06:17	
Diagnostics Check	OK	06:18	

Data Download Date/Time: 02/08/2023 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BOZA HERRERA

DATE/TIME OF TEST: 01/02/2023 04:23:15

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.078	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.149	04:18
Breath Volume	3.144	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:20
Subject Sample # 2	0.153	04:21
Breath Volume	3.140	04:21
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 01/05/2023 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CADE JR

DATE/TIME OF TEST: 01/11/2023 04:10:00

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.170	04:04
Breath Volume	1.921	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.165	04:08
Breath Volume	1.234	04:08
Air Blank	0.000	04:09
Control Test	0.077	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:10

Data Download Date/Time: 02/16/2023 07:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: LEON CABRERA

DATE/TIME OF TEST: 01/13/2023 01:39:26

Diagnostics Check	OK	01:38	
Air Blank		01:38	RFI Detect
Air Blank	0.000	01:39	

Data Download Date/Time: 02/16/2023 07:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: LEON CABRERA

DATE/TIME OF TEST: 01/13/2023 01:44:20

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.078	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:43	
Control Test	0.078	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 02/16/2023 07:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/14/2023 05:37:10

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.078	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.000	05:32
Breath Volume	2.843	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.000	05:35
Breath Volume	3.546	05:35
Air Blank	0.000	05:36
Control Test	0.079	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 02/16/2023 07:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 01/14/2023 21:38:18

Diagnostics Check	OK	21:31
Air Blank	0.000	21:32
Control Test	0.079	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.013	21:34
Breath Volume	1.128	21:34
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.010	21:36
Breath Volume	1.320	21:36
Air Blank	0.000	21:37
Control Test	0.077	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 02/16/2023 07:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,NATALIE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: RIVERO,NATALIE
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/24/2023 23:21:48

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.078	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.060	23:15
Breath Volume	1.503	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:17
Subject Sample # 2	0.058	23:20
Breath Volume	2.328	23:20
Air Blank	0.000	23:20
Control Test	0.078	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 02/16/2023 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 01/25/2023 01:33:17

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.077	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.215	01:27
Breath Volume	1.324	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.215	01:31
Breath Volume	1.503	01:31
Air Blank	0.000	01:32
Control Test	0.078	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 02/16/2023 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIE,BOB,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: 1102,SAINTILIE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: AUGUSTE

DATE/TIME OF TEST: 01/29/2023 12:14:32

Diagnostics Check	OK	12:14	
Air Blank		12:14	RFI Detect
Air Blank		12:14	RFI Detect

Data Download Date/Time: 02/16/2023 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEEN,BOB,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAINTILIEEN,BOB
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: AUGUSTE

DATE/TIME OF TEST: 01/29/2023 12:23:33

Diagnostics Check	OK	12:16
Air Blank	0.000	12:17
Control Test	0.078	12:17
Air Blank	0.000	12:18
Subject Sample # 1	0.088	12:19
Breath Volume	2.152	12:19
Air Blank	0.000	12:19
Air Blank	0.000	12:21
Subject Sample # 2	0.083	12:22
Breath Volume	2.183	12:22
Air Blank	0.000	12:22
Control Test	0.078	12:22
Air Blank	0.000	12:23
Diagnostics Check	OK	12:23

Data Download Date/Time: 02/16/2023 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEEN,BOB,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAINTILIEEN,BOB
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/30/2023 08:02:58

Diagnostics Check	OK	07:56
Air Blank	0.000	07:57
Control Test	0.078	07:57
Air Blank	0.000	07:57
Subject Sample # 1	0.161	07:58
Breath Volume	2.066	07:58
Air Blank	0.000	07:59
Air Blank	0.000	08:00
Subject Sample # 2	0.163	08:01
Breath Volume	1.640	08:01
Air Blank	0.000	08:02
Control Test	0.078	08:02
Air Blank	0.000	08:02
Diagnostics Check	OK	08:02

Data Download Date/Time: 02/16/2023 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEEN,BOB,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAINTILIEEN,BOB
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: BERRIOS

DATE/TIME OF TEST: 01/30/2023 13:32:03

Diagnostics Check	OK	13:20	
Air Blank	0.000	13:21	
Control Test	0.077	13:21	
Air Blank	0.000	13:21	
Subject Sample	# 1	13:25	No Sample Provided
Breath Volume	0.000	13:25	
Air Blank	0.000	13:25	
Air Blank	0.000	13:27	
Subject Sample	# 2	13:30	No Sample Provided
Breath Volume	0.000	13:30	
Air Blank	0.000	13:31	
Control Test	0.077	13:31	
Air Blank	0.000	13:31	
Diagnostics Check	OK	13:32	

Data Download Date/Time: 02/16/2023 07:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: RAMOS DE ALMONTE

DATE/TIME OF TEST: 01/01/2023 01:38:21

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.203	01:33
Breath Volume	2.949	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.212	01:36
Breath Volume	2.179	01:36
Air Blank	0.000	01:37
Control Test	0.078	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 01/12/2023 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FLANDERS,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MCEACHERN

DATE/TIME OF TEST: 01/01/2023 02:52:59

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.079	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:52	
Control Test	0.079	02:52	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 01/12/2023 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: MCCLAREN

DATE/TIME OF TEST: 01/01/2023 03:58:59

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.079	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:58	
Control Test	0.079	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 01/12/2023 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 01/01/2023 06:12:29

Diagnostics Check	OK	06:05
Air Blank	0.000	06:06
Control Test	0.079	06:06
Air Blank	0.000	06:07
Subject Sample # 1	0.066	06:07
Breath Volume	1.992	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 2	0.065	06:11
Breath Volume	1.410	06:11
Air Blank	0.000	06:11
Control Test	0.080	06:11
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

Data Download Date/Time: 01/12/2023 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 01/02/2023 00:46:32

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.080	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.161	00:42
Breath Volume	2.785	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.160	00:45
Breath Volume	2.402	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 01/12/2023 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: PITCHFORD

DATE/TIME OF TEST: 01/02/2023 02:50:41

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.080	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.080	02:50	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 01/12/2023 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: LEHMAN

DATE/TIME OF TEST: 01/02/2023 03:17:39

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.000	03:11
Breath Volume	1.218	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.000	03:16
Breath Volume	1.226	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 01/12/2023 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HUMES,U
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: PIKE

DATE/TIME OF TEST: 01/02/2023 20:05:53

Diagnostics Check	OK	20:03	
Air Blank	0.000	20:03	
Control Test	0.080	20:03	
Air Blank	0.000	20:04	
Subject Sample	# 1	20:04	Subject Test Refused
Breath Volume	0.000	20:04	
Air Blank	0.000	20:04	
Control Test	0.079	20:05	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 01/04/2023 01:03:36

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.030	00:59
Breath Volume	2.777	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.029	01:02
Breath Volume	2.250	01:02
Air Blank	0.000	01:02
Control Test	0.079	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: YISRAEL-LAGUER

DATE/TIME OF TEST: 01/04/2023 03:52:55

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.000	03:48
Breath Volume	1.550	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.000	03:51
Breath Volume	1.347	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: BEAMAN

DATE/TIME OF TEST: 01/04/2023 05:10:23

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:08	
Control Test	0.079	05:08	
Air Blank	0.000	05:08	
Subject Sample	# 1	05:09	Subject Test Refused
Breath Volume	0.000	05:09	
Air Blank	0.000	05:09	
Control Test	0.079	05:09	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 01/04/2023 20:16:36

Diagnostics Check	OK	20:13	
Air Blank	0.000	20:14	
Control Test	0.080	20:14	
Air Blank	0.000	20:15	
Subject Sample	# 1	20:15	Subject Test Refused
Breath Volume	0.000	20:15	
Air Blank	0.000	20:15	
Control Test	0.080	20:16	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:16	

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 01/04/2023 22:38:28

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.080	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.106	22:33
Breath Volume	1.664	22:33
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.098	22:36
Breath Volume	1.117	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEMPF,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/05/2023 02:14:17

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.147	02:09
Breath Volume	1.535	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.143	02:12
Breath Volume	1.695	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MASOOD,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MICHAELS

DATE/TIME OF TEST: 01/05/2023 03:18:23

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.000	03:13
Breath Volume	1.960	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.000	03:17
Breath Volume	1.949	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: SHAY

DATE/TIME OF TEST: 01/06/2023 00:25:12

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.120	00:20
Breath Volume	2.687	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.121	00:23
Breath Volume	2.132	00:23
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: NOVOTNY

DATE/TIME OF TEST: 01/06/2023 03:12:42

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.137	03:08
Breath Volume	1.277	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.137	03:11
Breath Volume	1.144	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,S,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BATEH,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:52

SUBJECT'S LAST NAME: KREISMAN

DATE/TIME OF TEST: 01/06/2023 13:21:59

Diagnostics Check	OK	13:18	
Air Blank	0.000	13:18	
Control Test	0.082	13:19	
Air Blank	0.000	13:19	
Subject Sample	# 1	13:20	Subject Test Refused
Breath Volume	0.000	13:20	
Air Blank	0.000	13:21	
Control Test	0.081	13:21	
Air Blank	0.000	13:21	
Diagnostics Check	OK	13:21	

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: DAGLEY

DATE/TIME OF TEST: 01/06/2023 23:18:47

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.082	23:16	
Air Blank	0.000	23:17	
Subject Sample	# 1	23:17	Subject Test Refused
Breath Volume	0.000	23:17	
Air Blank	0.000	23:17	
Control Test	0.081	23:18	
Air Blank	0.000	23:18	
Diagnostics Check	OK	23:18	

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: PANZO QUIAHUA

DATE/TIME OF TEST: 01/07/2023 00:11:01

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.082	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:10	
Control Test	0.081	00:10	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 01/07/2023 01:31:22

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.081	01:29	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Control Test	0.081	01:30	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/07/2023 02:11:21

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.081	02:09	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:10	
Control Test	0.081	02:10	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/07/2023 04:26:54

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.174	04:22
Breath Volume	1.542	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.167	04:25
Breath Volume	1.941	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/07/2023 23:58:47

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.080	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:57	
Control Test	0.080	23:58	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: SALMERON FLORES

DATE/TIME OF TEST: 01/08/2023 01:06:33

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.058	01:02
Breath Volume	3.351	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.060	01:05
Breath Volume	3.140	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: BEASLEY

DATE/TIME OF TEST: 01/08/2023 02:19:49

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.118	02:15
Breath Volume	1.488	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.113	02:18
Breath Volume	1.339	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: ROBERGE

DATE/TIME OF TEST: 01/08/2023 04:39:20

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.202	04:35
Breath Volume	2.363	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.200	04:37
Breath Volume	1.269	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: VEGA-GUEVARA

DATE/TIME OF TEST: 01/08/2023 05:41:27

Diagnostics Check	OK	05:34
Air Blank	0.000	05:35
Control Test	0.081	05:35
Air Blank	0.000	05:36
Subject Sample # 1	0.141	05:36
Breath Volume	1.394	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.142	05:39
Breath Volume	1.550	05:39
Air Blank	0.000	05:40
Control Test	0.080	05:40
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBER,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 01/08/2023 18:32:48

Diagnostics Check	OK	18:26
Air Blank	0.000	18:26
Control Test	0.080	18:27
Air Blank	0.000	18:27
Subject Sample # 1	0.000	18:28
Breath Volume	3.410	18:28
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample # 2	0.000	18:31
Breath Volume	3.527	18:31
Air Blank	0.000	18:31
Control Test	0.080	18:32
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/08/2023 19:49:33

Diagnostics Check	OK	19:39
Air Blank	0.000	19:39
Control Test	0.080	19:40
Air Blank	0.000	19:40
Subject Sample # 1	0.230	19:41
Breath Volume	1.773	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.251	19:44
Breath Volume	1.125	19:44
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 3	0.240	19:48
Breath Volume	1.765	19:48
Air Blank	0.000	19:48
Control Test	0.080	19:48
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: SCEE

DATE/TIME OF TEST: 01/09/2023 01:13:50

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.163	01:09
Breath Volume	1.945	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:11
Subject Sample # 2	0.156	01:12
Breath Volume	2.371	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 01/09/2023 02:26:58

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.205	02:22
Breath Volume	1.949	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.206	02:25
Breath Volume	2.597	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: FEDERER

DATE/TIME OF TEST: 01/09/2023 05:48:55

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.080	05:43
Air Blank	0.000	05:44
Subject Sample # 1	0.155	05:44
Breath Volume	1.468	05:44
Air Blank	0.000	05:45
Air Blank	0.000	05:46
Subject Sample # 2	0.151	05:47
Breath Volume	1.277	05:47
Air Blank	0.000	05:48
Control Test	0.080	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SIMKANICH

DATE/TIME OF TEST: 01/09/2023 23:47:16

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.151	23:42
Breath Volume	3.019	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.150	23:45
Breath Volume	2.953	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: KOSTENSKI

DATE/TIME OF TEST: 01/10/2023 00:48:46

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.081	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.000	00:44
Breath Volume	2.019	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.000	00:47
Breath Volume	1.410	00:47
Air Blank	0.000	00:47
Control Test	0.080	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 01/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: MENDOZA PADILLA

DATE/TIME OF TEST: 01/10/2023 22:53:25

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.093	22:48
Breath Volume	3.144	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.092	22:51
Breath Volume	3.582	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 01/12/2023 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 01/11/2023 02:29:39

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.081	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.057	02:24
Breath Volume	1.855	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.059	02:28
Breath Volume	1.742	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 01/12/2023 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: KINDER

DATE/TIME OF TEST: 01/12/2023 02:56:22

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.062	02:51
Breath Volume	1.550	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.059	02:54
Breath Volume	2.050	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 01/12/2023 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 01/12/2023 04:55:06

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:43	
Control Test	0.081	04:43	
Air Blank	0.000	04:44	
Subject Sample	# 1 0.053	04:47	Volume Not Met
Breath Volume	0.523	04:47	
Air Blank	0.000	04:47	
Air Blank	0.000	04:49	
Subject Sample	# 2 0.067	04:50	
Breath Volume	1.984	04:50	
Air Blank	0.000	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 3 0.065	04:53	
Breath Volume	1.984	04:53	
Air Blank	0.000	04:54	
Control Test	0.081	04:54	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 01/12/2023 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: ABRASHI

DATE/TIME OF TEST: 01/08/2023 00:25:59

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.163	00:21
Breath Volume	1.843	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.153	00:24
Breath Volume	1.250	00:24
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 01/12/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: HASANOVIC

DATE/TIME OF TEST: 01/09/2023 01:35:25

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.078	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.114	01:31
Breath Volume	2.414	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.115	01:33
Breath Volume	3.328	01:33
Air Blank	0.000	01:34
Control Test	0.078	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 01/12/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBER,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: ST MARTIN

DATE/TIME OF TEST: 01/09/2023 16:02:16

Diagnostics Check	OK	15:55
Air Blank	0.000	15:56
Control Test	0.078	15:56
Air Blank	0.000	15:57
Subject Sample # 1	0.313	15:57
Breath Volume	4.320	15:57
Air Blank	0.000	15:58
Air Blank	0.000	16:00
Subject Sample # 2	0.326	16:00
Breath Volume	3.996	16:00
Air Blank	0.000	16:01
Control Test	0.077	16:01
Air Blank	0.000	16:02
Diagnostics Check	OK	16:02

Data Download Date/Time: 01/12/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES, W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WELLONS

DATE/TIME OF TEST: 01/10/2023 03:46:25

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.190	03:42
Breath Volume	2.839	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.192	03:44
Breath Volume	2.660	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 01/12/2023 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 01/15/2023 03:09:17

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.081	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:08	
Control Test	0.081	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 01/31/2023 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: MARTIN-DOMINGO

DATE/TIME OF TEST: 01/15/2023 23:26:19

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.201	23:21
Breath Volume	2.507	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:23
Subject Sample # 2	0.200	23:24
Breath Volume	2.546	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 01/31/2023 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: FERRIERE

DATE/TIME OF TEST: 01/19/2023 23:55:45

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:53	
Control Test	0.079	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:54	
Control Test	0.079	23:55	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 01/31/2023 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: LOMBARD

DATE/TIME OF TEST: 01/27/2023 03:52:45

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.082	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.082	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 01/31/2023 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOUNT,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SHARKEY,E
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: BACKSTRAND

DATE/TIME OF TEST: 01/02/2023 20:48:11

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.078	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.093	20:43
Breath Volume	1.660	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.096	20:46
Breath Volume	2.050	20:46
Air Blank	0.000	20:47
Control Test	0.079	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 01/24/2023 23:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES,JOSEAS,I
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BURNETTE,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: GREANEY

DATE/TIME OF TEST: 01/04/2023 18:26:53

Diagnostics Check	OK	18:24	
Air Blank	0.000	18:24	
Control Test	0.077	18:24	
Air Blank	0.000	18:25	
Subject Sample	# 1	18:25	Subject Test Refused
Breath Volume	0.000	18:25	
Air Blank	0.000	18:25	
Control Test	0.077	18:26	
Air Blank	0.000	18:26	
Diagnostics Check	OK	18:26	

Data Download Date/Time: 01/24/2023 23:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOELZER,K,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MORRIS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:03

SUBJECT'S LAST NAME: EPPERSON

DATE/TIME OF TEST: 01/05/2023 12:37:34

Diagnostics Check	OK	12:31
Air Blank	0.000	12:31
Control Test	0.078	12:31
Air Blank	0.000	12:32
Subject Sample # 1	0.019	12:33
Breath Volume	1.152	12:33
Air Blank	0.000	12:33
Air Blank	0.000	12:35
Subject Sample # 2	0.016	12:36
Breath Volume	3.363	12:36
Air Blank	0.000	12:36
Control Test	0.077	12:36
Air Blank	0.000	12:37
Diagnostics Check	OK	12:37

Data Download Date/Time: 01/24/2023 23:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,BRADEN,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ANSTETT,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: FELDMAN

DATE/TIME OF TEST: 01/06/2023 23:25:52

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.042	23:21
Breath Volume	2.812	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:23
Subject Sample # 2	0.040	23:24
Breath Volume	2.597	23:24
Air Blank	0.000	23:24
Control Test	0.078	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 01/24/2023 23:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEFFERSON,A,S
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: EAST,Z
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/15/2023 04:17:19

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:14	
Control Test	0.080	04:14	
Air Blank	0.000	04:15	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Control Test	0.079	04:16	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

Data Download Date/Time: 01/24/2023 23:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BRANDON,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 01/15/2023 16:26:45

Diagnostics Check	OK	16:20
Air Blank	0.000	16:21
Control Test	0.079	16:21
Air Blank	0.000	16:21
Subject Sample # 1	0.000	16:22
Breath Volume	2.660	16:22
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.000	16:25
Breath Volume	2.328	16:25
Air Blank	0.000	16:25
Control Test	0.079	16:26
Air Blank	0.000	16:26
Diagnostics Check	OK	16:26

Data Download Date/Time: 01/24/2023 23:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSIDY,BRADLEY,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DILL,TODD
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: GRODIVANT

DATE/TIME OF TEST: 01/15/2023 22:29:39

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.079	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.203	22:24
Breath Volume	2.621	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.210	22:28
Breath Volume	4.472	22:28
Air Blank	0.000	22:28
Control Test	0.078	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 01/24/2023 23:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,B,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: WOODRUFF

DATE/TIME OF TEST: 01/23/2023 00:19:32

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.078	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.172	00:13
Breath Volume	2.277	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.181	00:17
Breath Volume	2.406	00:17
Air Blank	0.000	00:18
Control Test	0.075	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 01/24/2023 23:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ADAMS,MATTHEW
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: GARCIE

DATE/TIME OF TEST: 01/08/2023 22:40:16

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:29	
Control Test	0.078	22:30	
Air Blank	0.000	22:30	
Subject Sample	# 1	22:33	No Sample Provided
Breath Volume	0.000	22:33	
Air Blank	0.000	22:34	
Air Blank	0.000	22:36	
Subject Sample	# 2	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:39	
Control Test	0.079	22:39	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 02/01/2023 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,THOMAS,H
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: REEVES,DAVID
ARRESTING OFFICER AGENCY: Trenton PD
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: SPRAWLING

DATE/TIME OF TEST: 01/14/2023 00:17:37

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:15	
Control Test	0.080	00:15	
Air Blank	0.000	00:16	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:16	
Control Test	0.079	00:17	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 02/01/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MARTINEZ GOMEZ

DATE/TIME OF TEST: 01/16/2023 23:31:32

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.078	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.139	23:27
Breath Volume	2.042	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.130	23:30
Breath Volume	2.484	23:30
Air Blank	0.000	23:30
Control Test	0.078	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 02/01/2023 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROIANO,JOSHUA,R
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: SHEFFIELD,AUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 01/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BRASWELL

DATE/TIME OF TEST: 01/22/2023 21:28:26

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:26	
Control Test	0.080	21:26	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:27	RFI Detect
Breath Volume	0.000	21:27	
Air Blank	0.000	21:27	
Control Test	0.079	21:28	
Air Blank		21:28	RFI Detect
Air Blank		21:28	RFI Detect

Data Download Date/Time: 02/15/2023 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROIANO,JOSHUA,R
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: SHEFFIELD,AUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 01/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BRASWELL

DATE/TIME OF TEST: 01/22/2023 21:37:41

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.080	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.167	21:32
Breath Volume	2.074	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.163	21:36
Breath Volume	1.945	21:36
Air Blank	0.000	21:36
Control Test	0.078	21:37
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 02/15/2023 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: 2S01,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: AGUILAR REYES

DATE/TIME OF TEST: 01/28/2023 20:56:09

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:53	
Control Test	0.081	20:54	
Air Blank	0.000	20:54	
Subject Sample	# 1	20:54	Subject Test Refused
Breath Volume	0.000	20:54	
Air Blank	0.000	20:55	
Control Test	0.080	20:55	
Air Blank	0.000	20:56	
Diagnostics Check	OK	20:56	

Data Download Date/Time: 02/15/2023 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,GLENN,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,GLENN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CARTWRIGHT

DATE/TIME OF TEST: 01/01/2023 00:17:44

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.076	00:16	
Air Blank		00:17	RFI Detect
Air Blank	0.000	00:17	

Data Download Date/Time: 01/16/2023 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,GLENN,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,GLENN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CARTWRIGHT

DATE/TIME OF TEST: 01/01/2023 00:26:20

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.076	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.073	00:21
Breath Volume	3.511	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.072	00:24
Breath Volume	3.558	00:24
Air Blank	0.000	00:25
Control Test	0.076	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 01/16/2023 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,GLENN,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,GLENN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: COUNCIL

DATE/TIME OF TEST: 01/01/2023 04:13:38

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:11	
Control Test	0.076	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Control Test	0.076	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 01/16/2023 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUDSWORTH,SGT,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: DREW,DEP.
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/03/2023 00:03:45

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.075	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 2	23:59	No Sample Provided
Breath Volume	0.000	23:59	
Air Blank	0.000	00:00	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.076	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 01/16/2023 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUDSWORTH,SGT,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: MALLEY,DEP
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 01/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: OSTEN

DATE/TIME OF TEST: 01/16/2023 19:40:37

Diagnostics Check	OK	19:33	
Air Blank	0.000	19:34	
Control Test	0.077	19:34	
Air Blank	0.000	19:35	
Subject Sample # 1	0.179	19:35	
Breath Volume	2.113	19:35	
Air Blank	0.000	19:36	
Air Blank	0.000	19:38	
Subject Sample # 2	0.170	19:39	
Breath Volume	1.804	19:39	
Air Blank	0.000	19:39	
Control Test	0.074	19:40	Control Outside Tolerance
Air Blank	0.000	19:40	

Data Download Date/Time: 02/28/2023 17:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: KONKEL,WYATT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 01/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CRENSHAW

DATE/TIME OF TEST: 01/24/2023 03:28:58

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.078	03:21	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:24	No Sample Provided
Breath Volume	0.000	03:24	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	
Control Test	0.078	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 02/28/2023 17:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEBE,CAMERON,C
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BEEBE,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PALLA

DATE/TIME OF TEST: 01/25/2023 21:55:02

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:54	
Control Test	0.073	21:54	Control Outside Tolerance
Air Blank	0.000	21:55	

Data Download Date/Time: 02/28/2023 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEBE,CAMERON,C
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BEEBE,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PALLA

DATE/TIME OF TEST: 01/25/2023 22:06:17

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.077	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.165	22:01
Breath Volume	2.320	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.169	22:04
Breath Volume	1.843	22:04
Air Blank	0.000	22:05
Control Test	0.077	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 02/28/2023 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ORANTES QUEVEDO

DATE/TIME OF TEST: 01/08/2023 02:19:02

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.161	02:14
Breath Volume	1.398	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.160	02:17
Breath Volume	1.394	02:17
Air Blank	0.000	02:18
Control Test	0.077	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:19

Data Download Date/Time: 02/01/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,DANIEL,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: JUEN,MARK
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: LOUX

DATE/TIME OF TEST: 01/12/2023 04:12:34

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.104	04:07
Breath Volume	1.457	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.106	04:11
Breath Volume	2.386	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 02/01/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASOLO,DOREEN,M
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: HAMILTON,LUCAS
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 12/30/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 133593
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 01/28/2023 01:36:17

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.077	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.201	01:31
Breath Volume	1.402	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.209	01:34
Breath Volume	2.082	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 03/06/2023 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 01/01/2023 01:22:55

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.081	01:20	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:21	Subject Test Refused
Breath Volume	0.000	01:21	
Air Blank	0.000	01:22	
Control Test	0.081	01:22	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 01/07/2023 21:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEAVER,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: APOLONIO

DATE/TIME OF TEST: 01/01/2023 03:13:51

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.082	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.102	03:07
Breath Volume	1.144	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.102	03:12
Breath Volume	2.351	03:12
Air Blank	0.000	03:12
Control Test	0.080	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 01/07/2023 21:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,DAVID,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MUNDT,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 01/01/2023 06:41:00

Diagnostics Check	OK	06:34
Air Blank	0.000	06:35
Control Test	0.080	06:35
Air Blank	0.000	06:36
Subject Sample # 1	0.137	06:36
Breath Volume	2.011	06:36
Air Blank	0.000	06:37
Air Blank	0.000	06:39
Subject Sample # 2	0.131	06:39
Breath Volume	1.925	06:39
Air Blank	0.000	06:40
Control Test	0.079	06:40
Air Blank	0.000	06:40
Diagnostics Check	OK	06:41

Data Download Date/Time: 01/07/2023 21:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONNELLY,DARYA,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: PEA,CLAIRSE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: DONNELLY

DATE/TIME OF TEST: 01/02/2023 15:29:49

Diagnostics Check	OK	15:25	
Air Blank	0.000	15:26	
Control Test	0.081	15:26	
Air Blank	0.000	15:27	
Subject Sample	# 1	15:28	Subject Test Refused
Breath Volume	0.000	15:28	
Air Blank	0.000	15:28	
Control Test	0.081	15:29	
Air Blank	0.000	15:29	
Diagnostics Check	OK	15:29	

Data Download Date/Time: 01/07/2023 21:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: ENFINGER

DATE/TIME OF TEST: 01/04/2023 01:32:01

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.081	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.111	01:27
Breath Volume	1.269	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.105	01:30
Breath Volume	2.113	01:30
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:32

Data Download Date/Time: 01/07/2023 21:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 01/06/2023 02:04:17

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.083	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.159	01:59
Breath Volume	1.808	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.162	02:02
Breath Volume	1.968	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 01/07/2023 21:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS, BRIDGID, E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM, JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: T520065806630

DATE/TIME OF TEST: 01/06/2023 20:35:31

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:34	
Control Test	0.082	20:34	
Air Blank		20:35	RFI Detect
Air Blank	0.000	20:35	

Data Download Date/Time: 01/07/2023 21:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS, BRIDGID, E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM, JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/06/2023 20:43:45

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.083	20:38
Air Blank	0.000	20:38
Subject Sample # 1	0.000	20:39
Breath Volume	1.375	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.000	20:42
Breath Volume	2.191	20:42
Air Blank	0.000	20:42
Control Test	0.081	20:43
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 01/07/2023 21:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/06/2023 23:22:14

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.082	23:20	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:21	RFI Detect
Breath Volume	0.000	23:21	
Air Blank		23:21	RFI Detect
Air Blank	0.000	23:22	

Data Download Date/Time: 01/07/2023 21:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/06/2023 23:32:40

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:24	
Control Test	0.082	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 1	23:28	No Sample Provided
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Air Blank	0.000	23:30	
Subject Sample	# 2	23:31	Subject Test Refused
Breath Volume	0.000	23:31	
Air Blank	0.000	23:31	
Control Test	0.083	23:32	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 01/07/2023 21:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, TYLER, M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JONES, TYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 01/07/2023 03:14:35

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.082	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Control Test	0.082	03:14	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 01/07/2023 21:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CRAFT

DATE/TIME OF TEST: 01/07/2023 03:40:55

Diagnostics Check	OK	03:40	
Air Blank		03:40	RFI Detect
Air Blank	0.000	03:40	

Data Download Date/Time: 01/07/2023 21:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CRAFT

DATE/TIME OF TEST: 01/07/2023 03:49:49

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.082	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.000	03:45
Breath Volume	1.359	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.000	03:48
Breath Volume	1.476	03:48
Air Blank	0.000	03:48
Control Test	0.082	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 01/07/2023 21:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JONES,DALLAS,L

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: JONES,DALLAS

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1346146

DRY GAS EXPIRY DATE: 07/07/2023

VIOLATION CODE: DUI

ARREST TIME: 21:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: MCGILL

DATE/TIME OF TEST: 01/23/2023 22:23:16

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.082	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.000	22:18
Breath Volume	2.265	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.000	22:21
Breath Volume	3.871	22:21
Air Blank	0.000	22:22
Control Test	0.081	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 02/28/2023 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSICK,ROBERT,C
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: CORBIN,TRUETT
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/29/2023 09:33:23

Diagnostics Check	OK	09:30	
Air Blank	0.000	09:30	
Control Test	0.081	09:31	
Air Blank	0.000	09:31	
Subject Sample	# 1	09:32	Subject Test Refused
Breath Volume	0.000	09:32	
Air Blank	0.000	09:32	
Control Test	0.082	09:32	
Air Blank	0.000	09:33	
Diagnostics Check	OK	09:33	

Data Download Date/Time: 02/28/2023 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSICK,ROBERT,C
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: CORBIN,TRUETT
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/29/2023 09:38:00

Diagnostics Check	OK	09:35	
Air Blank	0.000	09:35	
Control Test	0.081	09:35	
Air Blank	0.000	09:36	
Subject Sample	# 1	09:36	Subject Test Refused
Breath Volume	0.000	09:36	
Air Blank	0.000	09:37	
Control Test	0.082	09:37	
Air Blank	0.000	09:37	
Diagnostics Check	OK	09:38	

Data Download Date/Time: 02/28/2023 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,ERIN,M
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MARTIN,ERIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:06

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 01/05/2023 10:33:53

Diagnostics Check	OK	10:26
Air Blank	0.000	10:27
Control Test	0.081	10:27
Air Blank	0.000	10:28
Subject Sample # 1	0.000	10:29
Breath Volume	1.457	10:29
Air Blank	0.000	10:29
Air Blank	0.000	10:31
Subject Sample # 2	0.000	10:32
Breath Volume	1.421	10:32
Air Blank	0.000	10:32
Control Test	0.081	10:33
Air Blank	0.000	10:33
Diagnostics Check	OK	10:33

Data Download Date/Time: 01/23/2023 13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HALE,CALEB
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 01/06/2023 07:24:05

Diagnostics Check	OK	07:15
Air Blank	0.000	07:16
Control Test	0.081	07:16
Air Blank	0.000	07:17
Subject Sample # 1	0.092	07:19
Breath Volume	1.156	07:19
Air Blank	0.000	07:19
Air Blank	0.000	07:21
Subject Sample # 2	0.088	07:22
Breath Volume	1.503	07:22
Air Blank	0.000	07:23
Control Test	0.080	07:23
Air Blank	0.000	07:24
Diagnostics Check	OK	07:24

Data Download Date/Time: 01/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DEMARCO, MICHAEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 01/06/2023 23:10:56

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.082	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.065	23:06
Breath Volume	1.578	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.074	23:09
Breath Volume	2.574	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 01/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MARTIN, COLBY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: PACE

DATE/TIME OF TEST: 01/08/2023 01:39:59

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.082	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.239	01:35
Breath Volume	1.500	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:37
Subject Sample # 2	0.238	01:38
Breath Volume	2.609	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 01/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 01/12/2023 23:01:44

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.080	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.197	22:57
Breath Volume	2.441	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.201	23:00
Breath Volume	4.476	23:00
Air Blank	0.000	23:00
Control Test	0.080	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 01/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ALEX,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: YOUNG,THOMAS
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: GLENN

DATE/TIME OF TEST: 01/15/2023 06:12:07

Diagnostics Check	OK	06:04
Air Blank	0.000	06:05
Control Test	0.080	06:05
Air Blank	0.000	06:06
Subject Sample # 1	0.125	06:07
Breath Volume	1.703	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:09
Subject Sample # 2	0.128	06:10
Breath Volume	1.421	06:10
Air Blank	0.000	06:11
Control Test	0.081	06:11
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

Data Download Date/Time: 01/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHROW,TRACY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BARRAS,HOPE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: KORHNAK

DATE/TIME OF TEST: 01/15/2023 06:31:18

Diagnostics Check	OK	06:25
Air Blank	0.000	06:25
Control Test	0.080	06:25
Air Blank	0.000	06:26
Subject Sample # 1	0.178	06:26
Breath Volume	1.710	06:26
Air Blank	0.000	06:27
Air Blank	0.000	06:29
Subject Sample # 2	0.180	06:29
Breath Volume	1.312	06:29
Air Blank	0.000	06:30
Control Test	0.081	06:30
Air Blank	0.000	06:31
Diagnostics Check	OK	06:31

Data Download Date/Time: 01/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSE,STEVEN,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: JONES,KAVIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: YORK

DATE/TIME OF TEST: 01/17/2023 23:49:35

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.081	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.117	23:44
Breath Volume	1.460	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.123	23:48
Breath Volume	1.601	23:48
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 01/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/20/2023 20:44:29

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.081	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.101	20:39
Breath Volume	1.460	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.101	20:43
Breath Volume	2.035	20:43
Air Blank	0.000	20:43
Control Test	0.081	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 01/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BRUECKNER,APRIL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: PADELLA

DATE/TIME OF TEST: 01/20/2023 21:15:12

Diagnostics Check	OK	21:06
Air Blank	0.000	21:06
Control Test	0.081	21:07
Air Blank	0.000	21:07
Subject Sample # 1	0.188	21:09
Breath Volume	1.339	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:11
Subject Sample # 2	0.199	21:13
Breath Volume	2.093	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 01/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSE,STEVEN,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 01/22/2023 03:02:59

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.150	02:58
Breath Volume	2.191	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.154	03:01
Breath Volume	1.875	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 01/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN, WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 01/22/2023 04:07:41

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.252	04:02
Breath Volume	2.371	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.240	04:06
Breath Volume	2.628	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 01/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNABB,KENNETH,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FOSTER,ALEX
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/27/2023 23:29:24

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.081	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.000	23:25
Breath Volume	1.464	23:25
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.000	23:28
Breath Volume	3.304	23:28
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 02/27/2023 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONAHAN,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MONAHAN,TIMOTHY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 01/29/2023 04:50:30

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:45	
Control Test	0.080	04:46	
Air Blank	0.000	04:46	
Subject Sample	# 1	04:49	Subject Test Refused
Breath Volume	0.050	04:49	
Air Blank	0.000	04:49	
Control Test	0.080	04:49	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 02/27/2023 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHASON,HUNTER,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HOGAN,JEREMY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SHIVER

DATE/TIME OF TEST: 01/28/2023 03:52:17

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 1 0.150	03:48	
Breath Volume	1.152	03:48	
Air Blank	0.000	03:49	
Air Blank	0.000	03:51	
Subject Sample	# 2 0.158	03:51	
Breath Volume	1.910	03:51	
Air Blank		03:51	RFI Detect
Air Blank	0.000	03:52	

Data Download Date/Time: 02/27/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: PERDUE

DATE/TIME OF TEST: 01/07/2023 00:18:34

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.080	00:17	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:17	Subject Test Refused
Breath Volume	0.000	00:17	
Air Blank		00:18	RFI Detect
Air Blank	0.000	00:18	

Data Download Date/Time: 02/23/2023 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: KERR,ANDREW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: AMBERSON CROWTHER

DATE/TIME OF TEST: 01/14/2023 19:57:19

Diagnostics Check	OK	19:50
Air Blank	0.000	19:50
Control Test	0.079	19:50
Air Blank	0.000	19:51
Subject Sample # 1	0.222	19:52
Breath Volume	2.007	19:52
Air Blank	0.000	19:53
Air Blank	0.000	19:55
Subject Sample # 2	0.225	19:55
Breath Volume	1.988	19:55
Air Blank	0.000	19:56
Control Test	0.078	19:56
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 02/23/2023 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMPSON,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAMPSON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: DOUGLASS

DATE/TIME OF TEST: 01/01/2023 05:37:52

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.183	05:33
Breath Volume	1.980	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:35
Subject Sample # 2	0.174	05:36
Breath Volume	4.550	05:36
Air Blank	0.000	05:36
Control Test	0.080	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 01/31/2023 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALAS,LUIS,E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCMILLIAN,RUSSELL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 11:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:08

SUBJECT'S LAST NAME: PERDOMO ROMERO

DATE/TIME OF TEST: 01/01/2023 12:48:40

Diagnostics Check	OK	12:37	
Air Blank	0.000	12:37	
Control Test	0.080	12:38	
Air Blank	0.000	12:38	
Subject Sample	# 1	12:41	No Sample Provided
Breath Volume	0.000	12:41	
Air Blank	0.000	12:42	
Air Blank	0.000	12:44	
Subject Sample	# 2	12:47	No Sample Provided
Breath Volume	0.000	12:47	
Air Blank	0.000	12:47	
Control Test	0.081	12:48	
Air Blank	0.000	12:48	
Diagnostics Check	OK	12:48	

Data Download Date/Time: 01/31/2023 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMATEER,TRACY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PALMATEER,TRACY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LENIART

DATE/TIME OF TEST: 01/06/2023 20:38:25

Diagnostics Check	OK	20:31
Air Blank	0.000	20:32
Control Test	0.080	20:32
Air Blank	0.000	20:33
Subject Sample # 1	0.193	20:33
Breath Volume	1.480	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.197	20:36
Breath Volume	1.738	20:36
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 01/31/2023 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALAS,LUIS,E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: FARWICK,SARAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: HOOSE

DATE/TIME OF TEST: 01/13/2023 18:01:48

Diagnostics Check	OK	17:50	
Air Blank	0.000	17:50	
Control Test	0.082	17:51	
Air Blank	0.000	17:51	
Subject Sample	# 1	17:54	No Sample Provided
Breath Volume	0.000	17:54	
Air Blank	0.000	17:55	
Air Blank	0.000	17:57	
Subject Sample	# 2	18:00	No Sample Provided
Breath Volume	0.000	18:00	
Air Blank	0.000	18:00	
Control Test	0.083	18:01	
Air Blank	0.000	18:01	
Diagnostics Check	OK	18:01	

Data Download Date/Time: 01/31/2023 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: ERWIN

DATE/TIME OF TEST: 01/19/2023 21:24:06

Diagnostics Check	OK	21:12	
Air Blank	0.000	21:13	
Control Test	0.080	21:13	
Air Blank	0.000	21:13	
Subject Sample	# 1 0.065	21:17	Volume Not Met
Breath Volume	0.222	21:17	
Air Blank	0.000	21:17	
Air Blank	0.000	21:19	
Subject Sample	# 2 0.058	21:22	Volume Not Met
Breath Volume	0.558	21:22	
Air Blank	0.000	21:23	
Control Test	0.081	21:23	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:24	

Data Download Date/Time: 01/31/2023 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: ERWIN

DATE/TIME OF TEST: 01/19/2023 21:31:29

Diagnostics Check	OK	21:28	
Air Blank	0.000	21:29	
Control Test	0.081	21:29	
Air Blank	0.000	21:29	
Subject Sample	# 1	21:30	Subject Test Refused
Breath Volume	0.000	21:30	
Air Blank	0.000	21:30	
Control Test	0.082	21:30	
Air Blank	0.000	21:31	
Diagnostics Check	OK	21:31	

Data Download Date/Time: 01/31/2023 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DIAZ, RAMON
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: CABALLERO

DATE/TIME OF TEST: 01/21/2023 02:20:08

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.082	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.177	02:15
Breath Volume	2.195	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.168	02:18
Breath Volume	1.824	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 01/31/2023 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ, JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: YEAGER

DATE/TIME OF TEST: 01/25/2023 01:08:48

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.081	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	01:01	No Sample Provided
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample	# 2	01:07	No Sample Provided
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.082	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 01/31/2023 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 01/25/2023 20:12:49

Diagnostics Check	OK	20:04
Air Blank	0.000	20:05
Control Test	0.081	20:05
Air Blank	0.000	20:05
Subject Sample # 1	0.093	20:07
Breath Volume	1.230	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:09
Subject Sample # 2	0.089	20:11
Breath Volume	1.750	20:11
Air Blank	0.000	20:11
Control Test	0.080	20:12
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 01/31/2023 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOST,CRAIG,R
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DIAZ, RAMON
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: AYALA AREVALO

DATE/TIME OF TEST: 01/27/2023 02:29:25

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.080	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.081	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 01/31/2023 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PARSONS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MALL

DATE/TIME OF TEST: 01/29/2023 00:12:45

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:10	
Control Test	0.082	00:10	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:11	
Control Test	0.083	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 01/31/2023 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: SPIRES

DATE/TIME OF TEST: 01/09/2023 14:04:18

Diagnostics Check	OK	13:56
Air Blank	0.000	13:56
Control Test	0.080	13:56
Air Blank	0.000	13:57
Subject Sample # 1	0.000	13:58
Breath Volume	1.144	13:58
Air Blank	0.000	13:58
Air Blank	0.000	14:00
Subject Sample # 2	0.000	14:02
Breath Volume	1.109	14:02
Air Blank	0.000	14:03
Control Test	0.081	14:03
Air Blank	0.000	14:04
Diagnostics Check	OK	14:04

Data Download Date/Time: 01/23/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 01/15/2023 22:27:57

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:24	
Control Test	0.080	22:25	
Air Blank	0.000	22:25	
Subject Sample	# 1	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:27	
Control Test	0.080	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 01/23/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 01/19/2023 20:17:47

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.000	20:13
Breath Volume	1.234	20:13
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.000	20:16
Breath Volume	3.382	20:16
Air Blank	0.000	20:16
Control Test	0.080	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 01/23/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, MARK, A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: GARCIA, K
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: BALTAZAR

DATE/TIME OF TEST: 01/20/2023 18:37:16

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.080	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.145	18:32
Breath Volume	1.660	18:32
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.146	18:35
Breath Volume	2.203	18:35
Air Blank	0.000	18:36
Control Test	0.079	18:36
Air Blank	0.000	18:37
Diagnostics Check	OK	18:37

Data Download Date/Time: 01/23/2023 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: OLVERA JIMENEZ

DATE/TIME OF TEST: 01/21/2023 00:44:55

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Control Test	0.080	00:44	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 01/23/2023 14:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK, STEVEN, D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK, STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: YE

DATE/TIME OF TEST: 01/21/2023 22:23:58

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.079	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.158	22:19
Breath Volume	2.597	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.155	22:22
Breath Volume	1.542	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 01/23/2023 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: KELLY,GARRETT
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 01/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 01/23/2023 19:04:24

Diagnostics Check	OK	19:00	
Air Blank	0.000	19:00	
Control Test	0.080	19:01	
Air Blank	0.000	19:01	
Subject Sample	# 1	19:03	Subject Test Refused
Breath Volume	0.000	19:03	
Air Blank	0.000	19:03	
Control Test	0.081	19:03	
Air Blank	0.000	19:04	
Diagnostics Check	OK	19:04	

Data Download Date/Time: 02/23/2023 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTO,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: RAVIROSA,ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402345249
DRY GAS EXPIRY DATE: 02/15/2025
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: FALKENBACH

DATE/TIME OF TEST: 01/28/2023 22:00:24

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.080	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.175	21:55
Breath Volume	1.917	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.174	21:58
Breath Volume	1.796	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 02/23/2023 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: KUBITSCHK,LEONARD
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 01/06/2023 10:08:35

Diagnostics Check	OK	10:02
Air Blank	0.000	10:03
Control Test	0.080	10:03
Air Blank	0.000	10:04
Subject Sample # 1	0.000	10:04
Breath Volume	2.089	10:04
Air Blank	0.000	10:04
Air Blank	0.000	10:06
Subject Sample # 2	0.000	10:07
Breath Volume	1.656	10:07
Air Blank	0.000	10:07
Control Test	0.079	10:07
Air Blank	0.000	10:08
Diagnostics Check	OK	10:08

Data Download Date/Time: 01/25/2023 07:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: ESPRIELLA

DATE/TIME OF TEST: 01/08/2023 02:08:59

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.174	02:04
Breath Volume	2.367	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:06
Subject Sample # 2	0.164	02:07
Breath Volume	2.644	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 01/25/2023 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHUNN,PETER,E
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LABARBERA,SUMMER
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: KAKELA

DATE/TIME OF TEST: 01/09/2023 18:57:39

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:46	
Control Test	0.079	18:47	
Air Blank	0.000	18:47	
Subject Sample	# 1	18:50	No Sample Provided
Breath Volume	0.000	18:50	
Air Blank	0.000	18:51	
Air Blank	0.000	18:53	
Subject Sample	# 2	18:56	No Sample Provided
Breath Volume	0.000	18:56	
Air Blank	0.000	18:56	
Control Test	0.080	18:57	
Air Blank	0.000	18:57	
Diagnostics Check	OK	18:57	

Data Download Date/Time: 01/25/2023 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHUNN,PETER,E
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: CHUNN,PETER
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:24

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 01/16/2023 11:07:50

Diagnostics Check	OK	11:01
Air Blank	0.000	11:02
Control Test	0.080	11:02
Air Blank	0.000	11:02
Subject Sample # 1	0.000	11:03
Breath Volume	1.816	11:03
Air Blank	0.000	11:03
Air Blank	0.000	11:05
Subject Sample # 2	0.000	11:06
Breath Volume	1.363	11:06
Air Blank	0.000	11:06
Control Test	0.081	11:07
Air Blank	0.000	11:07
Diagnostics Check	OK	11:07

Data Download Date/Time: 01/25/2023 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: BRINK,R
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 01/19/2023 01:23:31

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.179	01:18
Breath Volume	2.175	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:20
Subject Sample # 2	0.171	01:21
Breath Volume	1.609	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 01/25/2023 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF, MARK, D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LAWLESS, A
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ORDYNSKI

DATE/TIME OF TEST: 01/20/2023 23:52:19

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:47	No Sample Provided
Breath Volume	0.000	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:50	
Subject Sample	# 2	23:50	Subject Test Refused
Breath Volume	0.000	23:50	
Air Blank	0.000	23:51	
Control Test	0.079	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 01/25/2023 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUBITSCHK,LEONARD,H
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: VARGAS,ANGEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SAPIENZA

DATE/TIME OF TEST: 01/21/2023 00:50:37

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.079	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.232	00:46
Breath Volume	1.519	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.230	00:49
Breath Volume	1.574	00:49
Air Blank	0.000	00:49
Control Test	0.078	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 01/25/2023 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MAURE,R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: FLORES VELAZQUEZ

DATE/TIME OF TEST: 01/28/2023 17:18:12

Diagnostics Check	OK	17:05	
Air Blank	0.000	17:05	
Control Test	0.079	17:06	
Air Blank	0.000	17:06	
Subject Sample	# 1 0.149	17:09	Volume Not Met
Breath Volume	0.796	17:09	
Air Blank	0.000	17:10	
Air Blank	0.000	17:12	
Subject Sample	# 2 0.156	17:12	
Breath Volume	1.257	17:12	
Air Blank	0.000	17:13	
Air Blank	0.000	17:15	
Subject Sample	# 3 0.154	17:16	
Breath Volume	1.140	17:16	
Air Blank	0.000	17:17	
Control Test	0.078	17:17	
Air Blank	0.000	17:18	
Diagnostics Check	OK	17:18	

Data Download Date/Time: 02/01/2023 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 01/28/2023 19:53:27

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:44	
Control Test	0.079	19:44	
Air Blank	0.000	19:44	
Subject Sample	# 1 0.204	19:48	Volume Not Met
Breath Volume	0.195	19:48	
Air Blank	0.000	19:49	
Air Blank	0.000	19:50	
Subject Sample	# 2	19:52	Subject Test Refused
Breath Volume	0.000	19:52	
Air Blank	0.000	19:52	
Control Test	0.077	19:53	
Air Blank		19:53	RFI Detect
Air Blank		19:53	RFI Detect

Data Download Date/Time: 02/01/2023 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 01/28/2023 19:58:24

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:56	
Control Test	0.079	19:56	
Air Blank	0.000	19:56	
Subject Sample	# 1	19:57	Subject Test Refused
Breath Volume	0.000	19:57	
Air Blank	0.000	19:57	
Control Test	0.079	19:57	
Air Blank	0.000	19:58	
Diagnostics Check	OK	19:58	

Data Download Date/Time: 02/01/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOON,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 01/28/2023 23:24:16

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.208	23:19
Breath Volume	1.730	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.196	23:22
Breath Volume	1.546	23:22
Air Blank	0.000	23:23
Control Test	0.078	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 02/01/2023 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: HOOPER

DATE/TIME OF TEST: 01/29/2023 02:53:52

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.079	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.079	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 02/01/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOON, C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: REIO

DATE/TIME OF TEST: 01/29/2023 23:35:20

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.078	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.191	23:29
Breath Volume	1.246	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.209	23:33
Breath Volume	1.867	23:33
Air Blank	0.000	23:34
Control Test	0.076	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 02/01/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SZOTT,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: FOWKES

DATE/TIME OF TEST: 01/30/2023 23:12:11

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.078	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.089	23:07
Breath Volume	1.984	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.088	23:10
Breath Volume	1.941	23:10
Air Blank	0.000	23:11
Control Test	0.078	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 02/01/2023 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CPL.,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BORJAS,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LEISENRING

DATE/TIME OF TEST: 01/31/2023 01:35:32

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.079	01:33	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.078	01:34	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 02/01/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,CONCETTA,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 01/01/2023 00:48:58

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.177	00:44
Breath Volume	1.738	00:44
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.180	00:47
Breath Volume	1.761	00:47
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 02/01/2023 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,CONCETTA,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SINGH,T
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 01/01/2023 01:32:28

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:24	
Control Test	0.080	01:24	
Air Blank	0.000	01:25	
Subject Sample # 1	0.196	01:27	
Breath Volume	2.035	01:27	
Air Blank	0.000	01:28	
Air Blank	0.000	01:30	
Subject Sample # 2		01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:31	
Control Test	0.078	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,CONCETTA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: RODRIGUES

DATE/TIME OF TEST: 01/01/2023 02:05:09

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.143	02:00
Breath Volume	1.660	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.156	02:03
Breath Volume	1.851	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEACH,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 01/07/2023 04:40:02

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.082	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.154	04:35
Breath Volume	1.851	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:37
Subject Sample # 2	0.138	04:38
Breath Volume	1.664	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:40

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KESKINEN,D
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/07/2023 23:45:39

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:43	
Control Test	0.082	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:44	
Control Test	0.082	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEACH,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 01/08/2023 00:55:36

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.081	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.000	00:51
Breath Volume	2.269	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.000	00:54
Breath Volume	1.378	00:54
Air Blank	0.000	00:54
Control Test	0.081	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOLLEY,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HOYT-SANCHEZ

DATE/TIME OF TEST: 01/08/2023 03:28:03

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.081	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.145	03:23
Breath Volume	2.207	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:25
Subject Sample # 2	0.149	03:26
Breath Volume	2.136	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:28

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON, M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: CEPEDA

DATE/TIME OF TEST: 01/10/2023 01:52:27

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.081	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: OOLEY, J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: SUMMERFORD

DATE/TIME OF TEST: 01/10/2023 02:30:03

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.082	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.000	02:25
Breath Volume	2.773	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.000	02:28
Breath Volume	2.546	02:28
Air Blank	0.000	02:29
Control Test	0.082	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:30

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SHEPPARD,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:08

SUBJECT'S LAST NAME: ABRAM

DATE/TIME OF TEST: 01/10/2023 13:41:35

Diagnostics Check	OK	13:36	
Air Blank	0.000	13:36	
Control Test	0.082	13:37	
Air Blank	0.000	13:37	
Subject Sample	# 1	13:40	Subject Test Refused
Breath Volume	0.000	13:40	
Air Blank	0.000	13:40	
Control Test	0.082	13:40	
Air Blank	0.000	13:41	
Diagnostics Check	OK	13:41	

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: FAUSTINO

DATE/TIME OF TEST: 01/11/2023 03:13:44

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.081	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.000	03:09
Breath Volume	1.765	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.000	03:12
Breath Volume	1.300	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Fellsmere PD
ARRESTING OFFICER: ROLLANDO, T
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 01/14/2023 01:28:18

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.082	01:25	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:27	
Control Test	0.082	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NEWBERRY, L
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BISSONNETTE

DATE/TIME OF TEST: 01/14/2023 02:36:40

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.083	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.259	02:31
Breath Volume	1.984	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.259	02:35
Breath Volume	1.210	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CONNOLLY, B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/14/2023 03:47:20

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.082	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.070	03:42
Breath Volume	1.289	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.072	03:45
Breath Volume	1.605	03:45
Air Blank	0.000	03:46
Control Test	0.083	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LYNCH, D
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/14/2023 05:01:46

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.082	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.100	04:56
Breath Volume	2.054	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.110	05:00
Breath Volume	1.636	05:00
Air Blank	0.000	05:00
Control Test	0.082	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MAURE,R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 01/14/2023 16:38:17

Diagnostics Check	OK	16:36	
Air Blank	0.000	16:37	
Control Test	0.083	16:37	
Air Blank		16:37	RFI Detect
Air Blank	0.000	16:38	

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MAURE,R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:13

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 01/14/2023 16:45:08

Diagnostics Check	OK	16:39	
Air Blank	0.000	16:40	
Control Test	0.083	16:40	
Air Blank	0.000	16:41	
Subject Sample	# 1	16:43	Subject Test Refused
Breath Volume	0.000	16:43	
Air Blank	0.000	16:44	
Control Test	0.083	16:44	
Air Blank	0.000	16:45	
Diagnostics Check	OK	16:45	

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOON,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: RAPP

DATE/TIME OF TEST: 01/15/2023 00:02:27

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.081	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:01	
Control Test	0.082	00:01	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: HERNANDEZ,C
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 01/15/2023 02:36:43

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:34	
Control Test	0.082	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Control Test	0.082	02:36	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 02/01/2023 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 01/15/2023 05:06:53

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:04	
Control Test	0.083	05:04	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:05	Subject Test Refused
Breath Volume	0.000	05:05	
Air Blank	0.000	05:05	
Control Test	0.082	05:06	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEIGHAN,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 01/16/2023 21:55:43

Diagnostics Check	OK	21:52	
Air Blank	0.000	21:53	
Control Test	0.081	21:53	
Air Blank	0.000	21:54	
Subject Sample	# 1	21:54	Subject Test Refused
Breath Volume	0.000	21:54	
Air Blank	0.000	21:54	
Control Test	0.082	21:55	
Air Blank	0.000	21:55	
Diagnostics Check	OK	21:55	

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CONNOLLY,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: DUNFORD

DATE/TIME OF TEST: 01/19/2023 20:16:59

Diagnostics Check	OK	20:08
Air Blank	0.000	20:09
Control Test	0.081	20:09
Air Blank	0.000	20:10
Subject Sample # 1	0.079	20:10
Breath Volume	1.675	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.085	20:15
Breath Volume	1.398	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: OOLEY,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: BOLLINGER

DATE/TIME OF TEST: 01/19/2023 20:46:08

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:43	
Control Test	0.080	20:44	
Air Blank	0.000	20:44	
Subject Sample	# 1	20:44	Subject Test Refused
Breath Volume	0.000	20:44	
Air Blank	0.000	20:45	
Control Test	0.081	20:45	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:46	

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEACH,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: DENNARD

DATE/TIME OF TEST: 01/21/2023 04:41:18

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.000	04:36
Breath Volume	1.988	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.000	04:39
Breath Volume	1.589	04:39
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPUTO,TAYLOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: CAPUTO,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: OLEGARIO

DATE/TIME OF TEST: 01/21/2023 20:07:00

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.081	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.138	20:02
Breath Volume	2.546	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:04
Subject Sample # 2	0.135	20:05
Breath Volume	2.976	20:05
Air Blank	0.000	20:06
Control Test	0.080	20:06
Air Blank	0.000	20:06
Diagnostics Check	OK	20:07

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILAR, RODRIGO,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: AGUILAR, RODRIGO
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 01/21/2023 21:37:08

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:34	
Control Test	0.081	21:35	
Air Blank	0.000	21:35	
Subject Sample	# 1	21:35	Subject Test Refused
Breath Volume	0.000	21:35	
Air Blank	0.000	21:36	
Control Test	0.081	21:36	
Air Blank	0.000	21:37	
Diagnostics Check	OK	21:37	

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BORJAS,N
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 01/22/2023 04:47:43

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:45	
Control Test	0.082	04:45	
Air Blank	0.000	04:46	
Subject Sample	# 1	04:46	Subject Test Refused
Breath Volume	0.000	04:46	
Air Blank	0.000	04:46	
Control Test	0.081	04:47	
Air Blank	0.000	04:47	
Diagnostics Check	OK	04:47	

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EVANS, H
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: BAYONNE

DATE/TIME OF TEST: 01/24/2023 00:52:47

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.082	00:50	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:51	
Control Test	0.082	00:52	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: OOLEY, J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 01/24/2023 22:11:23

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.081	21:55	
Air Blank	0.000	21:56	
Subject Sample # 1	0.193	21:58	
Breath Volume	2.050	21:58	
Air Blank	0.000	21:59	
Air Blank	0.000	22:01	
Subject Sample # 2		22:04	No Sample Provided
Breath Volume	0.000	22:04	
Air Blank	0.000	22:04	
Air Blank	0.000	22:06	
Subject Sample # 3		22:10	No Sample Provided
Breath Volume	0.000	22:10	
Air Blank	0.000	22:10	
Control Test	0.081	22:10	
Air Blank	0.000	22:11	
Diagnostics Check	OK	22:11	

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CPL.,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEACH,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CLAUSS

DATE/TIME OF TEST: 01/26/2023 22:42:36

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:40	
Control Test	0.082	22:40	
Air Blank	0.000	22:41	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:41	
Control Test	0.081	22:42	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CPL.,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PIZZUTI,OFC.
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: DHEEDENE

DATE/TIME OF TEST: 01/27/2023 02:50:09

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.082	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:48	Subject Test Refused
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	
Control Test	0.082	02:49	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROTH, C
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: STARR

DATE/TIME OF TEST: 01/28/2023 00:05:26

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.081	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Control Test	0.082	00:04	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CONSALO, A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: NAPOLITANO

DATE/TIME OF TEST: 01/28/2023 01:37:00

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.082	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.202	01:32
Breath Volume	2.136	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.194	01:35
Breath Volume	2.167	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:37

Data Download Date/Time: 02/01/2023 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SMYTH,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: PERES BHATT

DATE/TIME OF TEST: 01/14/2023 02:43:46

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.190	02:39
Breath Volume	1.367	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.182	02:42
Breath Volume	1.246	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 02/07/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ALAMPI

DATE/TIME OF TEST: 01/15/2023 19:56:12

Diagnostics Check	OK	19:41	
Air Blank	0.000	19:41	
Control Test	0.081	19:42	
Air Blank	0.000	19:42	
Subject Sample	# 1 0.261	19:45	Volume Not Met
Breath Volume	0.835	19:45	
Air Blank	0.000	19:46	
Air Blank	0.000	19:48	
Subject Sample	# 2 0.254	19:48	
Breath Volume	1.246	19:48	
Air Blank	0.000	19:49	
Air Blank	0.000	19:51	
Subject Sample	# 3 0.242	19:54	Volume Not Met
Breath Volume	0.496	19:54	
Air Blank	0.000	19:55	
Control Test	0.079	19:55	
Air Blank	0.000	19:56	
Diagnostics Check	OK	19:56	

Data Download Date/Time: 02/07/2023 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ALAMPI

DATE/TIME OF TEST: 01/15/2023 20:02:25

Diagnostics Check	OK	19:57	
Air Blank	0.000	19:58	
Control Test	0.080	19:58	
Air Blank	0.000	19:59	
Subject Sample	# 1	20:00	Subject Test Refused
Breath Volume	0.152	20:00	
Air Blank	0.000	20:01	
Control Test	0.080	20:01	
Air Blank	0.000	20:02	
Diagnostics Check	OK	20:02	

Data Download Date/Time: 02/07/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI, OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI, OFC. M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CEGLIA

DATE/TIME OF TEST: 01/01/2023 01:43:29

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.082	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.091	01:38
Breath Volume	4.398	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.091	01:41
Breath Volume	4.027	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: MILSOM

DATE/TIME OF TEST: 01/01/2023 20:37:33

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.082	20:31
Air Blank	0.000	20:32
Subject Sample # 1	0.395	20:32
Breath Volume	3.046	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.400	20:35
Breath Volume	2.269	20:35
Air Blank	0.000	20:36
Control Test	0.078	20:36
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIRLEY,THOMAS,W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SHIRLEY,THOMAS
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:08

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 01/02/2023 14:18:20

Diagnostics Check	OK	14:15	
Air Blank	0.000	14:15	
Control Test	0.081	14:16	
Air Blank	0.000	14:16	
Subject Sample	# 1	14:16	Subject Test Refused
Breath Volume	0.000	14:16	
Air Blank	0.000	14:17	
Control Test	0.082	14:17	
Air Blank	0.000	14:18	
Diagnostics Check	OK	14:18	

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 01/02/2023 23:59:32

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.078	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.149	23:55
Breath Volume	1.402	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.155	23:58
Breath Volume	1.265	23:58
Air Blank	0.000	23:58
Control Test	0.080	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 01/04/2023 20:26:59

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.081	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.000	20:22
Breath Volume	2.105	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.000	20:25
Breath Volume	1.980	20:25
Air Blank	0.000	20:26
Control Test	0.080	20:26
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/06/2023 03:46:49

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.142	03:42
Breath Volume	3.183	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.138	03:45
Breath Volume	3.261	03:45
Air Blank	0.000	03:45
Control Test	0.080	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MEDICO

DATE/TIME OF TEST: 01/07/2023 01:34:29

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.081	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Control Test	0.082	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: BIRMINGHAM

DATE/TIME OF TEST: 01/07/2023 04:50:59

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.082	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.203	04:45
Breath Volume	1.644	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.192	04:49
Breath Volume	1.867	04:49
Air Blank	0.000	04:50
Control Test	0.082	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 01/07/2023 22:07:02

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.083	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.000	22:02
Breath Volume	2.878	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.000	22:05
Breath Volume	2.089	22:05
Air Blank	0.000	22:06
Control Test	0.082	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:07

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: ALVARADO ANTUNEZ

DATE/TIME OF TEST: 01/08/2023 19:39:39

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.081	19:34
Air Blank	0.000	19:34
Subject Sample # 1	0.248	19:35
Breath Volume	2.601	19:35
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.244	19:38
Breath Volume	2.414	19:38
Air Blank	0.000	19:38
Control Test	0.080	19:39
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTA,R
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: COMYNS-SUITTO

DATE/TIME OF TEST: 01/09/2023 02:51:51

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.082	02:38	
Air Blank	0.000	02:38	
Subject Sample	# 1 0.104	02:41	
Breath Volume	1.816	02:41	
Air Blank	0.000	02:42	
Air Blank	0.000	02:43	
Subject Sample	# 2 0.085	02:46	Volume Not Met
Breath Volume	0.601	02:46	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 3 0.093	02:50	
Breath Volume	1.410	02:50	
Air Blank	0.000	02:50	
Control Test	0.076	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV, OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STEVENS, OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/09/2023 19:36:49

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.081	19:31
Air Blank	0.000	19:32
Subject Sample # 1	0.229	19:32
Breath Volume	1.894	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:34
Subject Sample # 2	0.227	19:35
Breath Volume	1.781	19:35
Air Blank	0.000	19:35
Control Test	0.078	19:36
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ROBERTS,OFC D
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MARTINCEK

DATE/TIME OF TEST: 01/10/2023 02:57:34

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.198	02:53
Breath Volume	3.937	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.192	02:55
Breath Volume	2.480	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: BRUMBACH

DATE/TIME OF TEST: 01/11/2023 22:33:00

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:17	
Control Test	0.080	22:17	
Air Blank	0.000	22:17	
Subject Sample	# 1 0.202	22:20	
Breath Volume	1.402	22:20	
Air Blank	0.000	22:21	
Air Blank	0.000	22:23	
Subject Sample	# 2 0.204	22:26	Volume Not Met
Breath Volume	0.460	22:26	
Air Blank	0.000	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 3 0.184	22:31	
Breath Volume	1.125	22:31	
Air Blank	0.000	22:32	
Control Test	0.076	22:32	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYTH,GARY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SMYTH,GARY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: DYAYEV

DATE/TIME OF TEST: 01/15/2023 00:18:15

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.082	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.165	00:13
Breath Volume	2.164	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.158	00:16
Breath Volume	1.644	00:16
Air Blank	0.000	00:17
Control Test	0.081	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYTH,GARY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRUNO,NICHOLAS
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MAYO

DATE/TIME OF TEST: 01/15/2023 00:41:58

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.083	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1 0.067	00:30	Volume Not Met
Breath Volume	0.347	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:33	
Subject Sample	# 2 0.080	00:34	
Breath Volume	1.433	00:34	
Air Blank	0.000	00:35	
Air Blank	0.000	00:37	
Subject Sample	# 3 0.059	00:40	Volume Not Met
Breath Volume	0.035	00:40	
Air Blank	0.000	00:41	
Control Test	0.082	00:41	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: JAIME

DATE/TIME OF TEST: 01/15/2023 02:33:30

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.084	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.137	02:29
Breath Volume	1.851	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.126	02:31
Breath Volume	1.648	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYTH,GARY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SIPE,RAYCHEL
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: GERVASI

DATE/TIME OF TEST: 01/15/2023 02:57:44

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.082	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.184	02:53
Breath Volume	2.457	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.187	02:56
Breath Volume	2.453	02:56
Air Blank	0.000	02:56
Control Test	0.082	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI, OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI, OFC. M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: OSTRICKI

DATE/TIME OF TEST: 01/17/2023 23:25:16

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.172	23:20
Breath Volume	1.109	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.176	23:23
Breath Volume	1.484	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KENNY,S
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: LAFOY

DATE/TIME OF TEST: 01/20/2023 20:48:53

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.083	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.000	20:44
Breath Volume	2.839	20:44
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.000	20:47
Breath Volume	3.792	20:47
Air Blank	0.000	20:47
Control Test	0.082	20:48
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: RHOADES

DATE/TIME OF TEST: 01/20/2023 23:53:35

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.082	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.000	23:49
Breath Volume	1.984	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.000	23:52
Breath Volume	1.503	23:52
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: VILLASANCHEZ

DATE/TIME OF TEST: 01/22/2023 02:54:20

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.083	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1 0.000	02:49	Volume Not Met
Breath Volume	0.125	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 2	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: VILLASANCHEZ

DATE/TIME OF TEST: 01/22/2023 03:03:23

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:58	
Control Test	0.082	02:58	
Air Blank	0.000	02:58	
Subject Sample	# 1 0.159	02:59	
Breath Volume	2.734	02:59	
Air Blank	0.000	03:00	
Air Blank	0.000	03:01	
Subject Sample	# 2 0.146	03:02	
Breath Volume	1.117	03:02	
Air Blank		03:02	RFI Detect
Air Blank	0.000	03:03	

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/22/2023 21:36:49

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.082	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.079	21:32
Breath Volume	1.351	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:34
Subject Sample # 2	0.083	21:35
Breath Volume	1.257	21:35
Air Blank	0.000	21:35
Control Test	0.082	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KLEIN,B
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MCCracken

DATE/TIME OF TEST: 01/23/2023 01:20:57

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.082	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.230	01:16
Breath Volume	1.175	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.219	01:19
Breath Volume	2.128	01:19
Air Blank	0.000	01:20
Control Test	0.081	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV, OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: JONES, OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: LAVERTY

DATE/TIME OF TEST: 01/24/2023 21:00:39

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.082	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.142	20:56
Breath Volume	1.765	20:56
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.147	20:59
Breath Volume	1.421	20:59
Air Blank	0.000	20:59
Control Test	0.079	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: KISS

DATE/TIME OF TEST: 01/24/2023 21:39:42

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.082	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.000	21:35
Breath Volume	2.363	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.000	21:38
Breath Volume	2.250	21:38
Air Blank	0.000	21:38
Control Test	0.080	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 01/25/2023 02:37:09

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.082	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.000	02:32
Breath Volume	1.613	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.000	02:35
Breath Volume	2.042	02:35
Air Blank	0.000	02:36
Control Test	0.082	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 02/01/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI, OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI, OFC. M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 01/26/2023 03:27:09

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.140	03:22
Breath Volume	2.277	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.134	03:25
Breath Volume	3.121	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 02/01/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: CREADON

DATE/TIME OF TEST: 01/26/2023 03:40:26

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.081	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:39	
Control Test	0.079	03:39	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 02/01/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIETRON,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/27/2023 02:43:36

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.082	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.170	02:39
Breath Volume	3.781	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.156	02:42
Breath Volume	2.207	02:42
Air Blank	0.000	02:42
Control Test	0.082	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 02/01/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI, OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI, OFC. M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MAHONEY JR

DATE/TIME OF TEST: 01/27/2023 03:17:41

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.082	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.192	03:13
Breath Volume	3.785	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.193	03:16
Breath Volume	4.460	03:16
Air Blank	0.000	03:16
Control Test	0.081	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 02/01/2023 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: KARLOFF

DATE/TIME OF TEST: 01/27/2023 19:55:15

Diagnostics Check	OK	19:48
Air Blank	0.000	19:49
Control Test	0.081	19:49
Air Blank	0.000	19:50
Subject Sample # 1	0.215	19:50
Breath Volume	1.171	19:50
Air Blank	0.000	19:51
Air Blank	0.000	19:52
Subject Sample # 2	0.224	19:53
Breath Volume	2.121	19:53
Air Blank	0.000	19:54
Control Test	0.080	19:54
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 02/01/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SALEM

DATE/TIME OF TEST: 01/28/2023 02:00:08

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.067	01:55
Breath Volume	1.636	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.065	01:58
Breath Volume	1.550	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 02/01/2023 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: TIMA

DATE/TIME OF TEST: 01/28/2023 03:42:48

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.082	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.130	03:38
Breath Volume	1.984	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.131	03:41
Breath Volume	1.730	03:41
Air Blank	0.000	03:41
Control Test	0.082	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 02/01/2023 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KEOGH,OFC R
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MALCOLM

DATE/TIME OF TEST: 01/28/2023 04:41:01

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.081	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:30	No Sample Provided
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2	0.169	04:34
Breath Volume	1.488	04:34	
Air Blank	0.000	04:35	
Air Blank	0.000	04:37	
Subject Sample	# 3	0.162	04:39
Breath Volume	1.996	04:39	
Air Blank	0.000	04:40	
Control Test	0.077	04:40	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 02/01/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV, OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SUNDERLAND, OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 01/29/2023 00:44:32

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.081	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.168	00:38
Breath Volume	2.093	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.162	00:42
Breath Volume	1.671	00:42
Air Blank	0.000	00:43
Control Test	0.078	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 02/01/2023 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BATTAGLIA

DATE/TIME OF TEST: 01/29/2023 03:51:54

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Subject Sample # 1	0.052	03:39	
Breath Volume	1.128	03:39	
Air Blank	0.000	03:39	
Air Blank	0.000	03:41	
Subject Sample # 2	0.059	03:44	Volume Not Met
Breath Volume	0.035	03:44	
Air Blank	0.000	03:45	
Air Blank	0.000	03:47	
Subject Sample # 3	0.051	03:50	Volume Not Met
Breath Volume	0.304	03:50	
Air Blank	0.000	03:50	
Control Test	0.080	03:51	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 02/01/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BATTAGLIA

DATE/TIME OF TEST: 01/29/2023 04:02:58

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.054	03:57
Breath Volume	1.183	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.047	04:01
Breath Volume	1.660	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 02/01/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: DEL CASTILLO

DATE/TIME OF TEST: 01/20/2023 20:32:17

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:31	
Control Test	0.000	20:31	Control Outside Tolerance
Air Blank	0.000	20:32	

Data Download Date/Time: 02/09/2023 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: DEL CASTILLO

DATE/TIME OF TEST: 01/20/2023 20:42:01

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.080	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.248	20:37
Breath Volume	2.339	20:37
Air Blank	0.000	20:38
Air Blank	0.000	20:39
Subject Sample # 2	0.243	20:40
Breath Volume	1.941	20:40
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:41
Diagnostics Check	OK	20:42

Data Download Date/Time: 02/09/2023 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: JOYA

DATE/TIME OF TEST: 01/22/2023 04:03:48

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.081	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.081	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 01/26/2023 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUGO, OSCAR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: LUGO PAGAN

DATE/TIME OF TEST: 01/21/2023 03:57:04

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.145	03:52
Breath Volume	1.535	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.146	03:55
Breath Volume	2.074	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:57

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GOLDENBERG,EVA
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CHISOLM

DATE/TIME OF TEST: 01/21/2023 04:17:44

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.123	04:12
Breath Volume	1.742	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample # 2	0.120	04:16
Breath Volume	1.718	04:16
Air Blank	0.000	04:16
Control Test	0.078	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: SABOYA DE MELLO

DATE/TIME OF TEST: 01/21/2023 05:36:27

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.079	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.161	05:32
Breath Volume	1.804	05:32
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.161	05:34
Breath Volume	1.531	05:34
Air Blank	0.000	05:35
Control Test	0.079	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEWIS,BEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/21/2023 15:57:39

Diagnostics Check	OK	15:51
Air Blank	0.000	15:51
Control Test	0.079	15:52
Air Blank	0.000	15:52
Subject Sample # 1	0.000	15:53
Breath Volume	1.386	15:53
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.000	15:56
Breath Volume	1.945	15:56
Air Blank	0.000	15:56
Control Test	0.080	15:57
Air Blank	0.000	15:57
Diagnostics Check	OK	15:57

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO, WILFRED
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: ZAROMBA

DATE/TIME OF TEST: 01/21/2023 17:06:22

Diagnostics Check	OK	16:55
Air Blank	0.000	16:55
Control Test	0.079	16:56
Air Blank	0.000	16:56
Subject Sample # 1	0.186	16:59
Breath Volume	1.238	16:59
Air Blank	0.000	17:00
Air Blank	0.000	17:01
Subject Sample # 2	0.187	17:04
Breath Volume	1.957	17:04
Air Blank	0.000	17:05
Control Test	0.079	17:05
Air Blank	0.000	17:06
Diagnostics Check	OK	17:06

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ, JOEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: RAMIREZ LEON

DATE/TIME OF TEST: 01/22/2023 04:55:28

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.249	04:50
Breath Volume	1.472	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.232	04:53
Breath Volume	2.253	04:53
Air Blank	0.000	04:54
Control Test	0.077	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RINCON,RENZO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: PHILEMON

DATE/TIME OF TEST: 01/22/2023 07:57:02

Diagnostics Check	OK	07:51
Air Blank	0.000	07:51
Control Test	0.080	07:51
Air Blank	0.000	07:52
Subject Sample # 1	0.000	07:52
Breath Volume	2.449	07:52
Air Blank	0.000	07:53
Air Blank	0.000	07:55
Subject Sample # 2	0.000	07:55
Breath Volume	3.296	07:55
Air Blank	0.000	07:56
Control Test	0.079	07:56
Air Blank	0.000	07:56
Diagnostics Check	OK	07:57

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCALES, BRANDON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: FRAME

DATE/TIME OF TEST: 01/22/2023 23:35:17

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.259	23:30
Breath Volume	4.003	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.259	23:33
Breath Volume	3.199	23:33
Air Blank	0.000	23:34
Control Test	0.078	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIEBKNECHT, KEVIN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: DIOR

DATE/TIME OF TEST: 01/24/2023 03:37:08

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.179	03:32
Breath Volume	1.906	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.179	03:35
Breath Volume	2.285	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TERRETT,NICHOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BLAIR JR

DATE/TIME OF TEST: 01/24/2023 22:20:17

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.079	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.095	22:15
Breath Volume	3.425	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.092	22:18
Breath Volume	3.011	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: GUTIERREZ CONTRERAS

DATE/TIME OF TEST: 01/25/2023 00:43:38

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.169	00:39
Breath Volume	1.359	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.170	00:42
Breath Volume	1.695	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILLS,DARREN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: SPIVY

DATE/TIME OF TEST: 01/25/2023 01:34:31

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.080	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.267	01:29
Breath Volume	1.394	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.285	01:32
Breath Volume	1.644	01:32
Air Blank	0.000	01:33
Control Test	0.078	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: BIBLER

DATE/TIME OF TEST: 01/25/2023 03:50:35

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.181	03:46
Breath Volume	3.527	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.190	03:49
Breath Volume	3.863	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KING,ARTHUR
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: BURITICA

DATE/TIME OF TEST: 01/26/2023 19:07:24

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:04	
Control Test	0.080	19:04	
Air Blank	0.000	19:05	
Subject Sample	# 1	19:06	Subject Test Refused
Breath Volume	0.000	19:06	
Air Blank	0.000	19:06	
Control Test	0.080	19:06	
Air Blank	0.000	19:07	
Diagnostics Check	OK	19:07	

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: BA YAT

DATE/TIME OF TEST: 01/26/2023 23:45:05

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1 0.175	23:34	
Breath Volume	1.269	23:34	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample	# 2 0.158	23:40	Volume Not Met
Breath Volume	0.289	23:40	
Air Blank	0.000	23:40	
Air Blank	0.000	23:42	
Subject Sample	# 3 0.163	23:43	
Breath Volume	1.605	23:43	
Air Blank	0.000	23:44	
Control Test	0.077	23:44	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 02/06/2023 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CRESPO,BRYAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 01/27/2023 00:49:07

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.080	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1 0.026	00:41	Volume Not Met
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 2 0.024	00:44	
Breath Volume	1.457	00:44	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample	# 3 0.025	00:47	
Breath Volume	1.234	00:47	
Air Blank	0.000	00:48	
Control Test	0.079	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 01/27/2023 02:48:12

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:37	
Control Test	0.080	02:37	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:41	No Sample Provided
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Air Blank	0.000	02:43	
Subject Sample	# 2	02:46	No Sample Provided
Breath Volume	0.000	02:46	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUNT,FRANCIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: CHARRIEZ

DATE/TIME OF TEST: 01/27/2023 03:47:57

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.080	03:37	
Air Blank	0.000	03:37	
Subject Sample	# 1 0.143	03:40	Volume Not Met
Breath Volume	0.152	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 2 0.131	03:46	Volume Not Met
Breath Volume	0.195	03:46	
Air Blank	0.000	03:47	
Control Test	0.079	03:47	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUNT,FRANCIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: CHARRIEZ

DATE/TIME OF TEST: 01/27/2023 04:00:01

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.078	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.148	03:53
Breath Volume	1.203	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.149	03:58
Breath Volume	1.265	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	04:00

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUGO, OSCAR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 01/27/2023 04:16:29

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.078	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.245	04:10
Breath Volume	1.519	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.238	04:14
Breath Volume	1.343	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT, RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 01/27/2023 04:26:17

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.079	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:25	
Control Test	0.079	04:25	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHMIEDER, NICOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/27/2023 22:33:04

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.252	22:28
Breath Volume	2.625	22:28
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.251	22:31
Breath Volume	1.468	22:31
Air Blank	0.000	22:32
Control Test	0.078	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:33

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI, ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 01/29/2023 02:14:35

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.184	02:10
Breath Volume	1.703	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.192	02:13
Breath Volume	1.152	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIESENFELT,JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: LAURISTON

DATE/TIME OF TEST: 01/29/2023 05:20:53

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:09	
Control Test	0.079	05:09	
Air Blank	0.000	05:09	
Subject Sample	# 1 0.099	05:13	Volume Not Met
Breath Volume	0.976	05:13	
Air Blank	0.000	05:13	
Air Blank	0.000	05:15	
Subject Sample	# 2 0.115	05:16	
Breath Volume	2.398	05:16	
Air Blank	0.000	05:16	
Air Blank	0.000	05:18	
Subject Sample	# 3 0.113	05:19	
Breath Volume	1.328	05:19	
Air Blank	0.000	05:19	
Control Test	0.078	05:20	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ, SERGIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/29/2023 19:41:37

Diagnostics Check	OK	19:35
Air Blank	0.000	19:35
Control Test	0.079	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.268	19:37
Breath Volume	1.718	19:37
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.267	19:40
Breath Volume	2.125	19:40
Air Blank	0.000	19:40
Control Test	0.078	19:41
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ, JEFFRIE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: CARDONA MARTINEZ

DATE/TIME OF TEST: 01/29/2023 21:48:35

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.079	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.236	21:44
Breath Volume	2.109	21:44
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.241	21:47
Breath Volume	1.968	21:47
Air Blank	0.000	21:47
Control Test	0.079	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS, CHRISTOPHER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: GIBNEY

DATE/TIME OF TEST: 01/30/2023 01:38:21

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.206	01:33
Breath Volume	2.457	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.190	01:36
Breath Volume	1.812	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES, JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: BOVAIRD

DATE/TIME OF TEST: 01/30/2023 02:34:09

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.139	02:29
Breath Volume	2.433	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.139	02:32
Breath Volume	2.722	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: UMANA, RAUL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: DE LOS REYES ARROYO

DATE/TIME OF TEST: 01/30/2023 03:29:29

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.037	03:25
Breath Volume	3.457	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.037	03:27
Breath Volume	2.207	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CROCK, NICOLLE
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: GILCHRIST

DATE/TIME OF TEST: 01/30/2023 04:42:14

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.173	04:37
Breath Volume	3.214	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.172	04:40
Breath Volume	3.410	04:40
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL, KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BRADFORD

DATE/TIME OF TEST: 01/31/2023 02:02:08

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.093	01:57
Breath Volume	1.929	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.093	02:00
Breath Volume	3.054	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COMBAS,JOHN
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:24

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 01/31/2023 15:57:16

Diagnostics Check	OK	15:50
Air Blank	0.000	15:51
Control Test	0.080	15:51
Air Blank	0.000	15:51
Subject Sample # 1	0.252	15:52
Breath Volume	1.535	15:52
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.265	15:55
Breath Volume	1.304	15:55
Air Blank	0.000	15:56
Control Test	0.078	15:56
Air Blank	0.000	15:57
Diagnostics Check	OK	15:57

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPENCER,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: GALLORUIZ

DATE/TIME OF TEST: 01/31/2023 17:11:42

Diagnostics Check	OK	17:05
Air Blank	0.000	17:05
Control Test	0.079	17:06
Air Blank	0.000	17:06
Subject Sample # 1	0.000	17:07
Breath Volume	1.105	17:07
Air Blank	0.000	17:07
Air Blank	0.000	17:09
Subject Sample # 2	0.000	17:10
Breath Volume	2.054	17:10
Air Blank	0.000	17:10
Control Test	0.079	17:11
Air Blank	0.000	17:11
Diagnostics Check	OK	17:11

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MELLA,MELISSA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MAUGER

DATE/TIME OF TEST: 01/31/2023 23:26:19

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.196	23:21
Breath Volume	2.167	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.184	23:24
Breath Volume	1.199	23:24
Air Blank	0.000	23:25
Control Test	0.078	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 02/06/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 01/09/2023 19:33:31

Diagnostics Check	OK	19:26
Air Blank	0.000	19:27
Control Test	0.082	19:27
Air Blank	0.000	19:27
Subject Sample # 1	0.090	19:28
Breath Volume	1.300	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.090	19:32
Breath Volume	1.449	19:32
Air Blank	0.000	19:32
Control Test	0.081	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 01/30/2023 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDZINSKI,BRANDON,E
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ROSA,CHANCE
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/12/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: DISTEFANO

DATE/TIME OF TEST: 01/22/2023 00:32:30

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.078	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.000	00:28
Breath Volume	2.062	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.000	00:31
Breath Volume	2.328	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 01/30/2023 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDZINSKI,BRANDON,E
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMANN,HANS
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 12/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SIEMBIDA

DATE/TIME OF TEST: 01/22/2023 22:43:12

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.078	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.130	22:38
Breath Volume	3.679	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.132	22:41
Breath Volume	3.093	22:41
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 01/30/2023 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,REBECCA,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: FENNELLY

DATE/TIME OF TEST: 01/24/2023 17:46:22

Diagnostics Check	OK	17:40
Air Blank	0.000	17:40
Control Test	0.080	17:40
Air Blank	0.000	17:41
Subject Sample # 1	0.111	17:42
Breath Volume	2.546	17:42
Air Blank	0.000	17:42
Air Blank	0.000	17:44
Subject Sample # 2	0.101	17:44
Breath Volume	2.867	17:44
Air Blank	0.000	17:45
Control Test	0.077	17:45
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

Data Download Date/Time: 01/30/2023 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: DAVENPORT,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 01/28/2023 04:55:01

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:50	
Control Test	0.079	04:51	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:53	Subject Test Refused
Breath Volume	0.000	04:53	
Air Blank	0.000	04:54	
Control Test	0.079	04:54	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 01/30/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/29/2023 01:11:08

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.147	01:05
Breath Volume	1.476	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.148	01:09
Breath Volume	1.531	01:09
Air Blank	0.000	01:10
Control Test	0.082	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 01/30/2023 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PHILLIPS,MERCEDES
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: SAMMON

DATE/TIME OF TEST: 01/15/2023 04:04:37

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.082	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.167	03:59
Breath Volume	1.621	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.155	04:03
Breath Volume	2.355	04:03
Air Blank	0.000	04:03
Control Test	0.082	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 02/14/2023 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,KWESI
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: CLIFFORD

DATE/TIME OF TEST: 01/15/2023 04:18:46

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.082	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.091	04:14
Breath Volume	1.121	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.088	04:17
Breath Volume	1.324	04:17
Air Blank	0.000	04:17
Control Test	0.082	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 02/14/2023 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DRAGOSLAV,MAKSIMOVIC
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: AZOFEIFA

DATE/TIME OF TEST: 01/15/2023 05:10:49

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.082	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.168	05:06
Breath Volume	1.324	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.174	05:09
Breath Volume	1.136	05:09
Air Blank	0.000	05:09
Control Test	0.082	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 02/14/2023 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,JILLIAN,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,KWESI
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 01/16/2023 00:48:59

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.081	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.259	00:44
Breath Volume	2.054	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:46
Subject Sample # 2	0.253	00:47
Breath Volume	1.640	00:47
Air Blank	0.000	00:48
Control Test	0.082	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 02/14/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: LOONEY ARCHER

DATE/TIME OF TEST: 01/18/2023 19:32:59

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.081	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.161	19:27
Breath Volume	1.261	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.153	19:31
Breath Volume	1.394	19:31
Air Blank	0.000	19:32
Control Test	0.081	19:32
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 02/14/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILZER,WOLFGANG,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CABRAL,MATTHEW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: DILZER

DATE/TIME OF TEST: 01/19/2023 19:07:38

Diagnostics Check	OK	18:57
Air Blank	0.000	18:58
Control Test	0.082	18:58
Air Blank	0.000	18:59
Subject Sample # 1	0.234	18:59
Breath Volume	1.589	18:59
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample # 2	0.274	19:02
Breath Volume	1.441	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 3	0.270	19:06
Breath Volume	2.531	19:06
Air Blank	0.000	19:06
Control Test	0.080	19:07
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 02/14/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: HECHLER

DATE/TIME OF TEST: 01/19/2023 20:49:49

Diagnostics Check	OK	20:43
Air Blank	0.000	20:44
Control Test	0.081	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.360	20:45
Breath Volume	1.425	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:47
Subject Sample # 2	0.347	20:48
Breath Volume	1.714	20:48
Air Blank	0.000	20:48
Control Test	0.080	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 02/14/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: NOFI

DATE/TIME OF TEST: 01/20/2023 22:35:09

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.081	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.154	22:30
Breath Volume	2.480	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.147	22:33
Breath Volume	1.914	22:33
Air Blank	0.000	22:34
Control Test	0.081	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 02/14/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: LUTZ

DATE/TIME OF TEST: 01/25/2023 21:35:24

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.082	21:29
Air Blank	0.000	21:30
Subject Sample # 1	0.070	21:31
Breath Volume	3.070	21:31
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.070	21:33
Breath Volume	2.363	21:33
Air Blank	0.000	21:34
Control Test	0.082	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 02/14/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARCHIORI,JONATHAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: KICHNER

DATE/TIME OF TEST: 01/27/2023 00:20:06

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.083	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:12	No Sample Provided
Breath Volume	0.000	00:12	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample	# 2	0.084	00:15
Breath Volume	1.804	00:15	
Air Blank	0.000	00:16	
Air Blank	0.000	00:18	
Subject Sample	# 3	0.079	00:18
Breath Volume	1.410	00:18	
Air Blank	0.000	00:19	
Control Test	0.082	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 02/14/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SERRES, DEVAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: LOISELLE

DATE/TIME OF TEST: 01/29/2023 00:29:45

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.083	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.117	00:25
Breath Volume	1.257	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.120	00:28
Breath Volume	1.492	00:28
Air Blank	0.000	00:28
Control Test	0.083	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 02/14/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 01/01/2023 00:26:41

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.000	00:22
Breath Volume	3.777	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.000	00:25
Breath Volume	3.546	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 01/17/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DEMARIA,,JAMES
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 01/01/2023 08:04:32

Diagnostics Check	OK	07:58
Air Blank	0.000	07:59
Control Test	0.080	07:59
Air Blank	0.000	07:59
Subject Sample # 1	0.180	08:00
Breath Volume	2.800	08:00
Air Blank	0.000	08:00
Air Blank	0.000	08:02
Subject Sample # 2	0.183	08:03
Breath Volume	2.675	08:03
Air Blank	0.000	08:03
Control Test	0.080	08:03
Air Blank	0.000	08:04
Diagnostics Check	OK	08:04

Data Download Date/Time: 01/17/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,CLAYTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:44

SUBJECT'S LAST NAME: BAROCH

DATE/TIME OF TEST: 01/01/2023 15:10:59

Diagnostics Check	OK	15:04
Air Blank	0.000	15:04
Control Test	0.080	15:05
Air Blank	0.000	15:05
Subject Sample # 1	0.000	15:06
Breath Volume	3.093	15:06
Air Blank	0.000	15:07
Air Blank	0.000	15:09
Subject Sample # 2	0.000	15:09
Breath Volume	4.039	15:09
Air Blank	0.000	15:10
Control Test	0.080	15:10
Air Blank	0.000	15:10
Diagnostics Check	OK	15:10

Data Download Date/Time: 01/17/2023 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ, S
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: BACHELER

DATE/TIME OF TEST: 01/03/2023 22:02:57

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:52	
Control Test	0.080	21:52	
Air Blank	0.000	21:52	
Subject Sample	# 1	21:55	No Sample Provided
Breath Volume	0.000	21:55	
Air Blank	0.000	21:56	
Air Blank	0.000	21:58	
Subject Sample	# 2	22:01	No Sample Provided
Breath Volume	0.000	22:01	
Air Blank	0.000	22:02	
Control Test	0.080	22:02	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: GARLIK

DATE/TIME OF TEST: 01/04/2023 18:48:29

Diagnostics Check	OK	18:42
Air Blank	0.000	18:42
Control Test	0.080	18:43
Air Blank	0.000	18:43
Subject Sample # 1	0.000	18:44
Breath Volume	1.523	18:44
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 2	0.000	18:47
Breath Volume	2.109	18:47
Air Blank	0.000	18:47
Control Test	0.079	18:47
Air Blank	0.000	18:48
Diagnostics Check	OK	18:48

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 01/05/2023 00:11:02

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.000	00:06
Breath Volume	1.339	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.000	00:09
Breath Volume	2.492	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:11

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 01/05/2023 00:47:35

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.128	00:43
Breath Volume	1.390	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.127	00:46
Breath Volume	1.781	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 01/05/2023 02:18:22

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.080	02:16	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:17	
Control Test	0.080	02:17	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,CARLOS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: OCONNELL,BRIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 01/05/2023 03:30:52

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:26	No Sample Provided
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample	# 2	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:29	
Control Test	0.080	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 01/05/2023 19:06:07

Diagnostics Check	OK	18:59
Air Blank	0.000	19:00
Control Test	0.081	19:00
Air Blank	0.000	19:01
Subject Sample # 1	0.000	19:01
Breath Volume	2.035	19:01
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.000	19:04
Breath Volume	2.218	19:04
Air Blank	0.000	19:05
Control Test	0.080	19:05
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/06/2023 01:23:03

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.080	01:12	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:16	No Sample Provided
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Air Blank	0.000	01:18	
Subject Sample	# 2	01:21	No Sample Provided
Breath Volume	0.000	01:21	
Air Blank	0.000	01:22	
Control Test	0.081	01:22	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SCHAEFFER

DATE/TIME OF TEST: 01/07/2023 01:58:08

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.139	01:53
Breath Volume	1.597	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.137	01:56
Breath Volume	1.546	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: O'CONNELL,BRIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: RUIZ-MENDOZA

DATE/TIME OF TEST: 01/08/2023 05:54:23

Diagnostics Check	OK	05:48
Air Blank	0.000	05:48
Control Test	0.081	05:49
Air Blank	0.000	05:49
Subject Sample # 1	0.170	05:50
Breath Volume	3.121	05:50
Air Blank	0.000	05:50
Air Blank	0.000	05:52
Subject Sample # 2	0.183	05:52
Breath Volume	3.250	05:52
Air Blank	0.000	05:53
Control Test	0.080	05:53
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: FALCON

DATE/TIME OF TEST: 01/09/2023 02:37:15

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.081	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.081	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: RIETMAN

DATE/TIME OF TEST: 01/10/2023 00:09:06

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.080	00:07	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:08	
Control Test	0.080	00:08	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,CARLOS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DEMARIA,JAMES
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MCKEE

DATE/TIME OF TEST: 01/10/2023 03:31:34

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.145	03:23
Breath Volume	4.507	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.121	03:26
Breath Volume	1.140	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 3	0.144	03:30
Breath Volume	4.835	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:56

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 01/10/2023 08:22:49

Diagnostics Check	OK	08:16
Air Blank	0.000	08:17
Control Test	0.081	08:17
Air Blank	0.000	08:18
Subject Sample # 1	0.047	08:18
Breath Volume	1.566	08:18
Air Blank	0.000	08:19
Air Blank	0.000	08:21
Subject Sample # 2	0.044	08:21
Breath Volume	2.117	08:21
Air Blank	0.000	08:21
Control Test	0.079	08:22
Air Blank	0.000	08:22
Diagnostics Check	OK	08:22

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 01/11/2023 11:46:24

Diagnostics Check	OK	11:40
Air Blank	0.000	11:40
Control Test	0.080	11:41
Air Blank	0.000	11:41
Subject Sample # 1	0.000	11:42
Breath Volume	2.847	11:42
Air Blank	0.000	11:42
Air Blank	0.000	11:44
Subject Sample # 2	0.000	11:45
Breath Volume	2.632	11:45
Air Blank	0.000	11:45
Control Test	0.080	11:45
Air Blank	0.000	11:46
Diagnostics Check	OK	11:46

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BOCKMAN-PEDERSON

DATE/TIME OF TEST: 01/11/2023 23:33:19

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.158	23:27
Breath Volume	1.382	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:29
Subject Sample # 2	0.174	23:31
Breath Volume	1.890	23:31
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 01/12/2023 00:37:27

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.204	00:32
Breath Volume	2.964	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.205	00:35
Breath Volume	2.621	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 01/17/2023 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: SEBKARSHAD

DATE/TIME OF TEST: 01/12/2023 01:35:37

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.081	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.000	01:31
Breath Volume	1.757	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.000	01:34
Breath Volume	1.363	01:34
Air Blank	0.000	01:34
Control Test	0.079	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/12/2023 04:02:58

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.113	03:58
Breath Volume	2.101	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.111	04:01
Breath Volume	2.222	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 01/12/2023 15:24:19

Diagnostics Check	OK	15:18	
Air Blank	0.000	15:18	
Control Test	0.080	15:18	
Air Blank	0.000	15:19	
Subject Sample # 1	0.000	15:20	
Breath Volume	3.007	15:20	
Air Blank	0.000	15:20	
Air Blank	0.000	15:22	
Subject Sample # 2		15:22	RFI Detect
Breath Volume	0.910	15:22	
Air Blank	0.000	15:23	
Control Test	0.080	15:23	
Air Blank	0.000	15:24	
Diagnostics Check	OK	15:24	

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 01/12/2023 15:31:46

Diagnostics Check	OK	15:26
Air Blank	0.000	15:26
Control Test	0.080	15:26
Air Blank	0.000	15:27
Subject Sample # 1	0.000	15:27
Breath Volume	2.101	15:27
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.000	15:30
Breath Volume	1.824	15:30
Air Blank	0.000	15:30
Control Test	0.080	15:31
Air Blank	0.000	15:31
Diagnostics Check	OK	15:31

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: DOAN

DATE/TIME OF TEST: 01/12/2023 22:02:52

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.174	21:58
Breath Volume	1.617	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.168	22:01
Breath Volume	1.265	22:01
Air Blank	0.000	22:01
Control Test	0.079	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HOWSER

DATE/TIME OF TEST: 01/12/2023 23:31:46

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.049	23:27
Breath Volume	2.441	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.049	23:30
Breath Volume	2.777	23:30
Air Blank	0.000	23:30
Control Test	0.080	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 01/12/2023 23:59:14

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.177	23:54
Breath Volume	1.941	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.176	23:57
Breath Volume	2.832	23:57
Air Blank	0.000	23:58
Control Test	0.080	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 01/13/2023 20:19:37

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:11	
Control Test	0.081	20:12	
Air Blank	0.000	20:12	
Subject Sample	# 1	20:15	No Sample Provided
Breath Volume	0.000	20:15	
Air Blank	0.000	20:16	
Air Blank	0.000	20:18	
Subject Sample	# 2	20:18	Subject Test Refused
Breath Volume	0.000	20:18	
Air Blank	0.000	20:18	
Control Test	0.080	20:19	
Air Blank	0.000	20:19	
Diagnostics Check	OK	20:19	

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: GOLA

DATE/TIME OF TEST: 01/13/2023 23:45:02

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.081	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.191	23:40
Breath Volume	1.175	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.180	23:43
Breath Volume	1.648	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:45

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: DORNS

DATE/TIME OF TEST: 01/14/2023 02:09:15

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.135	02:04
Breath Volume	1.593	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.133	02:07
Breath Volume	1.523	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: JODAH

DATE/TIME OF TEST: 01/14/2023 03:28:56

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.172	03:24
Breath Volume	1.132	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.173	03:27
Breath Volume	1.335	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: VALDEZ II

DATE/TIME OF TEST: 01/14/2023 20:11:03

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:02	
Control Test	0.081	20:02	
Air Blank	0.000	20:03	
Subject Sample	# 1	20:06	No Sample Provided
Breath Volume	0.000	20:06	
Air Blank	0.000	20:07	
Air Blank	0.000	20:09	
Subject Sample	# 2	20:09	Subject Test Refused
Breath Volume	0.000	20:09	
Air Blank	0.000	20:10	
Control Test	0.081	20:10	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:11	

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/14/2023 22:27:23

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.081	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.145	22:23
Breath Volume	2.214	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.141	22:25
Breath Volume	1.789	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: KHAMVONGSA

DATE/TIME OF TEST: 01/15/2023 00:00:32

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.081	23:55
Air Blank	0.000	23:56
Subject Sample # 1	0.150	23:56
Breath Volume	1.359	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.153	23:59
Breath Volume	1.777	23:59
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: PIEDRA

DATE/TIME OF TEST: 01/15/2023 02:35:59

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.081	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:29	No Sample Provided
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Air Blank	0.000	02:31	
Subject Sample	# 2	02:34	No Sample Provided
Breath Volume	0.000	02:34	
Air Blank	0.000	02:35	
Control Test	0.081	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: RUBIO GONZALEZ

DATE/TIME OF TEST: 01/15/2023 19:39:53

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:28	
Control Test	0.081	19:29	
Air Blank	0.000	19:29	
Subject Sample	# 1	19:32	No Sample Provided
Breath Volume	0.000	19:32	
Air Blank	0.000	19:33	
Air Blank	0.000	19:35	
Subject Sample	# 2	19:38	No Sample Provided
Breath Volume	0.000	19:38	
Air Blank	0.000	19:38	
Control Test	0.081	19:39	
Air Blank	0.000	19:39	
Diagnostics Check	OK	19:39	

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 01/16/2023 00:14:02

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.081	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.184	00:09
Breath Volume	1.179	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.186	00:12
Breath Volume	1.265	00:12
Air Blank	0.000	00:13
Control Test	0.081	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:14

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: CHRISTOPHER

DATE/TIME OF TEST: 01/17/2023 03:59:28

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:45	
Control Test		0.080	03:45	
Air Blank		0.000	03:46	
Subject Sample	# 1	0.274	03:49	Volume Not Met
Breath Volume		0.273	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:51	
Subject Sample	# 2	0.294	03:52	
Breath Volume		1.359	03:52	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 3	0.286	03:57	Volume Not Met
Breath Volume		0.753	03:57	
Air Blank		0.000	03:58	
Control Test		0.079	03:58	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 01/17/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSELMAN,CLINT,A
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CASSELMAN,CLINT
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LUNA APONTE

DATE/TIME OF TEST: 01/19/2023 21:29:46

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.081	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.000	21:25
Breath Volume	2.320	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.000	21:28
Breath Volume	3.281	21:28
Air Blank	0.000	21:28
Control Test	0.081	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 01/27/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/22/2023 04:06:03

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.081	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.116	04:01
Breath Volume	2.441	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.116	04:04
Breath Volume	2.007	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:06

Data Download Date/Time: 01/27/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBERI, FRANCIS, J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MONCAYO, MASON
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MORENO-MORALES

DATE/TIME OF TEST: 01/22/2023 06:15:06

Diagnostics Check	OK	06:08
Air Blank	0.000	06:08
Control Test	0.081	06:08
Air Blank	0.000	06:09
Subject Sample # 1	0.124	06:09
Breath Volume	1.343	06:09
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.112	06:13
Breath Volume	1.285	06:13
Air Blank	0.000	06:14
Control Test	0.079	06:14
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

Data Download Date/Time: 01/27/2023 14:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDERO,JAVIER,O
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PADILLA,HENRY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: BOWENS

DATE/TIME OF TEST: 01/22/2023 06:46:03

Diagnostics Check	OK	06:36
Air Blank	0.000	06:36
Control Test	0.081	06:37
Air Blank	0.000	06:37
Subject Sample # 1	0.210	06:40
Breath Volume	1.640	06:40
Air Blank	0.000	06:41
Air Blank	0.000	06:43
Subject Sample # 2	0.220	06:44
Breath Volume	1.925	06:44
Air Blank	0.000	06:45
Control Test	0.078	06:45
Air Blank	0.000	06:45
Diagnostics Check	OK	06:46

Data Download Date/Time: 01/27/2023 14:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: LANZO, MIGUEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: HERNANDEZ GUTIERREZ

DATE/TIME OF TEST: 01/22/2023 21:58:42

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.081	21:53
Air Blank	0.000	21:53
Subject Sample # 1	0.104	21:54
Breath Volume	3.042	21:54
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.105	21:57
Breath Volume	3.179	21:57
Air Blank	0.000	21:57
Control Test	0.080	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 01/27/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: DEVILLE

DATE/TIME OF TEST: 01/26/2023 21:36:52

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.081	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.205	21:32
Breath Volume	1.992	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.200	21:35
Breath Volume	1.796	21:35
Air Blank	0.000	21:35
Control Test	0.082	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 01/27/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRON,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MONCAYO,MASON
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: HENAO-CAMARGO

DATE/TIME OF TEST: 01/27/2023 03:35:37

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.082	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.101	03:31
Breath Volume	3.902	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.088	03:33
Breath Volume	2.406	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 01/27/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: GOMEZ HERNANDEZ

DATE/TIME OF TEST: 01/27/2023 23:36:05

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.081	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.148	23:31
Breath Volume	2.289	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.156	23:34
Breath Volume	1.253	23:34
Air Blank	0.000	23:35
Control Test	0.081	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:36

Data Download Date/Time: 02/24/2023 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 01/29/2023 00:53:16

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.082	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.172	00:48
Breath Volume	2.617	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.153	00:51
Breath Volume	1.820	00:51
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 02/24/2023 14:12