

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTZER,ETHAN,E
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BARTZER,ETHAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: RIPPEY

DATE/TIME OF TEST: 02/02/2023 01:02:33

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.078	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.210	00:58
Breath Volume	4.027	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.215	01:00
Breath Volume	3.054	01:00
Air Blank	0.000	01:01
Control Test	0.077	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 02/02/2023 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: OTHER
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/02/2023 16:39:54

Diagnostics Check	OK	16:38	
Air Blank	0.000	16:38	
Control Test	0.073	16:39	Control Outside Tolerance
Air Blank	0.000	16:39	

Data Download Date/Time: 02/25/2023 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: OTHER
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/02/2023 16:50:22

Diagnostics Check	OK	16:41
Air Blank	0.000	16:41
Control Test	0.080	16:42
Air Blank	0.000	16:42
Subject Sample # 1	0.103	16:43
Breath Volume	3.871	16:43
Air Blank	0.000	16:43
Air Blank	0.000	16:45
Subject Sample # 2	0.082	16:46
Breath Volume	3.589	16:46
Air Blank	0.000	16:46
Air Blank	0.000	16:48
Subject Sample # 3	0.087	16:48
Breath Volume	1.847	16:48
Air Blank	0.000	16:49
Control Test	0.079	16:49
Air Blank	0.000	16:50
Diagnostics Check	OK	16:50

Data Download Date/Time: 02/25/2023 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: OTHER
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/02/2023 16:57:28

Diagnostics Check	OK	16:51
Air Blank	0.000	16:51
Control Test	0.079	16:52
Air Blank	0.000	16:52
Subject Sample # 1	0.093	16:52
Breath Volume	3.195	16:52
Air Blank	0.000	16:53
Air Blank	0.000	16:55
Subject Sample # 2	0.094	16:55
Breath Volume	4.148	16:55
Air Blank	0.000	16:56
Control Test	0.079	16:56
Air Blank	0.000	16:57
Diagnostics Check	OK	16:57

Data Download Date/Time: 02/25/2023 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/02/2023 18:29:24

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.080	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.137	18:25
Breath Volume	4.191	18:25
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.125	18:27
Breath Volume	2.597	18:27
Air Blank	0.000	18:28
Control Test	0.079	18:28
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 02/25/2023 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: OTHER
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/02/2023 18:42:05

Diagnostics Check	OK	18:30	
Air Blank	0.000	18:30	
Control Test	0.079	18:31	
Air Blank	0.000	18:31	
Subject Sample	# 1 0.147	18:32	No .020 agreement
Breath Volume	3.796	18:32	
Air Blank	0.000	18:32	
Air Blank	0.000	18:34	
Subject Sample	# 2 0.100	18:35	No .020 agreement
Breath Volume	2.363	18:35	
Air Blank	0.000	18:35	
Air Blank	0.000	18:37	
Subject Sample	# 3	18:40	No Sample Provided
Breath Volume	0.000	18:40	
Air Blank	0.000	18:41	
Control Test	0.079	18:41	
Air Blank	0.000	18:41	
Diagnostics Check	OK	18:42	

Data Download Date/Time: 02/25/2023 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SALJANIN,C
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 02/12/2023 22:40:55

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.166	22:35
Breath Volume	2.207	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.158	22:39
Breath Volume	2.074	22:39
Air Blank	0.000	22:39
Control Test	0.079	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 02/25/2023 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,LANDON,L
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: EDWARDS,LANDON
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: ELIASSEN

DATE/TIME OF TEST: 02/14/2023 15:33:54

Diagnostics Check	OK	15:25
Air Blank	0.000	15:26
Control Test	0.080	15:26
Air Blank	0.000	15:27
Subject Sample # 1	0.000	15:28
Breath Volume	1.156	15:28
Air Blank	0.000	15:29
Air Blank	0.000	15:31
Subject Sample # 2	0.000	15:32
Breath Volume	1.171	15:32
Air Blank	0.000	15:32
Control Test	0.080	15:33
Air Blank	0.000	15:33
Diagnostics Check	OK	15:33

Data Download Date/Time: 02/25/2023 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSSI,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: EDWARDS,LANDON
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: CUSHING

DATE/TIME OF TEST: 02/15/2023 17:12:28

Diagnostics Check	OK	17:06
Air Blank	0.000	17:06
Control Test	0.079	17:06
Air Blank	0.000	17:07
Subject Sample # 1	0.149	17:07
Breath Volume	1.445	17:07
Air Blank	0.000	17:08
Air Blank	0.000	17:10
Subject Sample # 2	0.156	17:10
Breath Volume	1.277	17:10
Air Blank	0.000	17:11
Control Test	0.079	17:11
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

Data Download Date/Time: 02/25/2023 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DAMON,RUSSELL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 02/17/2023 03:44:53

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.143	03:40
Breath Volume	1.125	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.143	03:43
Breath Volume	1.523	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 02/25/2023 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GRABILL,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 02/18/2023 03:13:28

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.190	03:08
Breath Volume	2.617	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.198	03:11
Breath Volume	2.773	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 02/25/2023 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BROOKS,JACOB
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 02/19/2023 03:10:59

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1 0.133	03:03	Volume Not Met
Breath Volume	0.363	03:03	
Air Blank	0.000	03:04	
Air Blank	0.000	03:06	
Subject Sample	# 2 0.131	03:09	Volume Not Met
Breath Volume	0.050	03:09	
Air Blank	0.000	03:10	
Control Test	0.076	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 02/25/2023 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SPELLER,BENJAMIN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: CONCEPCION

DATE/TIME OF TEST: 02/24/2023 21:53:14

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.079	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.380	21:48
Breath Volume	3.085	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:50
Subject Sample # 2	0.370	21:51
Breath Volume	2.484	21:51
Air Blank	0.000	21:52
Control Test	0.078	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 02/25/2023 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: FLETCHER,A
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 02/25/2023 02:23:00

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.104	02:18
Breath Volume	2.062	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.105	02:21
Breath Volume	1.679	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:23

Data Download Date/Time: 02/25/2023 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: HOWELL,J
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 02/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: KEENAN

DATE/TIME OF TEST: 02/26/2023 03:29:05

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.119	03:23
Breath Volume	2.402	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.118	03:27
Breath Volume	1.714	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 02/28/2023 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,D,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WEAVER,D
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:43

SUBJECT'S LAST NAME: AMACHER

DATE/TIME OF TEST: 02/01/2023 12:07:34

Diagnostics Check	OK	12:03	
Air Blank	0.000	12:04	
Control Test	0.081	12:04	
Air Blank	0.000	12:05	
Subject Sample	# 1	12:06	Subject Test Refused
Breath Volume	0.000	12:06	
Air Blank	0.000	12:06	
Control Test	0.081	12:06	
Air Blank	0.000	12:07	
Diagnostics Check	OK	12:07	

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: WEESNER

DATE/TIME OF TEST: 02/02/2023 18:45:23

Diagnostics Check	OK	18:41	
Air Blank	0.000	18:42	
Control Test	0.080	18:42	
Air Blank	0.000	18:43	
Subject Sample	# 1	18:44	Subject Test Refused
Breath Volume	0.000	18:44	
Air Blank	0.000	18:44	
Control Test	0.081	18:44	
Air Blank	0.000	18:45	
Diagnostics Check	OK	18:45	

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BARRETT,JOSHUA
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: CRISTE

DATE/TIME OF TEST: 02/02/2023 20:22:51

Diagnostics Check	OK	20:19	
Air Blank	0.000	20:20	
Control Test	0.081	20:20	
Air Blank	0.000	20:21	
Subject Sample	# 1	20:21	Subject Test Refused
Breath Volume	0.000	20:21	
Air Blank	0.000	20:21	
Control Test	0.081	20:22	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:22	

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SANDERS,JAKE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: DEATON

DATE/TIME OF TEST: 02/05/2023 03:31:14

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.188	03:26
Breath Volume	1.246	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.182	03:29
Breath Volume	1.300	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HUZIOR,N
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: HOUSER

DATE/TIME OF TEST: 02/05/2023 19:56:09

Diagnostics Check	OK	19:47
Air Blank	0.000	19:48
Control Test	0.081	19:48
Air Blank	0.000	19:49
Subject Sample # 1	0.055	19:50
Breath Volume	1.343	19:50
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.051	19:54
Breath Volume	1.378	19:54
Air Blank	0.000	19:55
Control Test	0.081	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SANDERS,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: TRESS

DATE/TIME OF TEST: 02/08/2023 22:24:36

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.150	22:19
Breath Volume	1.445	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:21
Subject Sample # 2	0.149	22:23
Breath Volume	1.257	22:23
Air Blank	0.000	22:23
Control Test	0.081	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 02/10/2023 02:30:58

Diagnostics Check	OK	02:30	
Air Blank		02:30	RFI Detect
Air Blank		02:30	RFI Detect

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 02/10/2023 02:47:03

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.082	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1 0.159	02:38	Volume Not Met
Breath Volume	0.152	02:38	
Air Blank	0.000	02:39	
Air Blank	0.000	02:41	
Subject Sample	# 2 0.163	02:42	
Breath Volume	1.593	02:42	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample	# 3 0.164	02:45	
Breath Volume	1.191	02:45	
Air Blank	0.000	02:46	
Control Test	0.081	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BRYAN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CARTER,BRYAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/13/2023 01:01:32

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.081	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.184	00:56
Breath Volume	1.500	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.176	01:00
Breath Volume	1.593	01:00
Air Blank	0.000	01:00
Control Test	0.081	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOFFORD,JOSEPH,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WOFFORD,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: SANDY

DATE/TIME OF TEST: 02/13/2023 10:30:19

Diagnostics Check	OK	10:23
Air Blank	0.000	10:24
Control Test	0.082	10:24
Air Blank	0.000	10:25
Subject Sample # 1	0.050	10:25
Breath Volume	1.667	10:25
Air Blank	0.000	10:26
Air Blank	0.000	10:28
Subject Sample # 2	0.048	10:28
Breath Volume	2.835	10:28
Air Blank	0.000	10:29
Control Test	0.082	10:29
Air Blank	0.000	10:30
Diagnostics Check	OK	10:30

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRAY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WRAY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 02/14/2023 23:58:07

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.082	23:56	
Air Blank	0.000	23:56	
Subject Sample	# 1	23:56	Subject Test Refused
Breath Volume	0.000	23:56	
Air Blank	0.000	23:57	
Control Test	0.082	23:57	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDEN,A,G
BREATH TEST OPERATOR AGENCY: Bunnell PD
ARRESTING OFFICER: OGDEN,A
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 02/18/2023 20:32:09

Diagnostics Check	OK	20:31	
Air Blank		20:31	RFI Detect
Air Blank	0.000	20:32	

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDEN,ALTON,
BREATH TEST OPERATOR AGENCY: Bunnell PD
ARRESTING OFFICER: OGDEN,ALTON
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 02/18/2023 20:45:23

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:34	
Control Test	0.082	20:34	
Air Blank	0.000	20:35	
Subject Sample	# 1	20:38	No Sample Provided
Breath Volume	0.000	20:38	
Air Blank	0.000	20:38	
Air Blank	0.000	20:40	
Subject Sample	# 2	20:44	No Sample Provided
Breath Volume	0.000	20:44	
Air Blank	0.000	20:44	
Control Test	0.082	20:44	
Air Blank	0.000	20:45	
Diagnostics Check	OK	20:45	

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MOORHEAD,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/30/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: KACZMARSKI

DATE/TIME OF TEST: 02/20/2023 00:05:09

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.082	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.185	00:00
Breath Volume	1.429	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.180	00:03
Breath Volume	1.488	00:03
Air Blank	0.000	00:04
Control Test	0.082	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 02/20/2023 16:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ROSSI,C
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: ANDRADE

DATE/TIME OF TEST: 02/26/2023 02:25:47

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.083	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.113	02:21
Breath Volume	1.410	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.103	02:24
Breath Volume	1.382	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 03/16/2023 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:42

SUBJECT'S LAST NAME: HERNANDEZ-RODRIGUEZ

DATE/TIME OF TEST: 02/26/2023 09:10:54

Diagnostics Check	OK	09:02
Air Blank	0.000	09:03
Control Test	0.082	09:03
Air Blank	0.000	09:04
Subject Sample # 1	0.169	09:04
Breath Volume	1.628	09:04
Air Blank	0.000	09:05
Air Blank	0.000	09:07
Subject Sample # 2	0.165	09:09
Breath Volume	1.597	09:09
Air Blank	0.000	09:10
Control Test	0.082	09:10
Air Blank	0.000	09:10
Diagnostics Check	OK	09:10

Data Download Date/Time: 03/16/2023 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: MIRAMAR PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 02/04/2023 01:05:29

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.021	01:00
Breath Volume	2.351	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:02
Subject Sample # 2	0.018	01:04
Breath Volume	1.765	01:04
Air Blank	0.000	01:04
Control Test	0.081	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 03/13/2023 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: GOLDBERG,J
ARRESTING OFFICER AGENCY: OTHER STATE
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: RAMIREZ HERMIDA

DATE/TIME OF TEST: 02/04/2023 01:24:50

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.082	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.066	01:20
Breath Volume	1.601	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.063	01:23
Breath Volume	1.453	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 03/13/2023 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: MIRAMAR PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: LAVELANET

DATE/TIME OF TEST: 02/04/2023 01:44:07

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.081	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.027	01:39
Breath Volume	1.531	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.028	01:42
Breath Volume	1.250	01:42
Air Blank	0.000	01:43
Control Test	0.081	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 03/13/2023 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: AREVALO

DATE/TIME OF TEST: 02/10/2023 03:00:39

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.190	02:55
Breath Volume	2.445	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.183	02:59
Breath Volume	2.203	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 03/13/2023 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MARCIER

DATE/TIME OF TEST: 02/11/2023 02:49:36

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.081	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.000	02:45
Breath Volume	1.253	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.000	02:48
Breath Volume	1.746	02:48
Air Blank	0.000	02:48
Control Test	0.081	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 03/13/2023 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: PAREDES MAYEN

DATE/TIME OF TEST: 02/12/2023 05:29:04

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.081	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.193	05:24
Breath Volume	2.328	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.190	05:27
Breath Volume	1.824	05:27
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:29

Data Download Date/Time: 03/13/2023 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: GOMEZ PEREZ

DATE/TIME OF TEST: 02/19/2023 05:28:55

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.081	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.207	05:24
Breath Volume	2.593	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:26
Subject Sample # 2	0.194	05:27
Breath Volume	2.007	05:27
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 03/13/2023 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

INSTRUMENT REGISTERED TO: LAKE HELEN P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOUGERE, MARCEL, A
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: WOODIN, STEPHEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 02/09/2023 01:33:59

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.130	01:29
Breath Volume	1.761	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.124	01:32
Breath Volume	2.289	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 02/28/2023 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

INSTRUMENT REGISTERED TO: LAKE HELEN P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,TREVOR,W
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GRAY,TREVOR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/26/2023 03:45:35

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1 0.088	03:37	Volume Not Met
Breath Volume	0.804	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:39	
Subject Sample	# 2 0.120	03:40	
Breath Volume	1.820	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 3 0.115	03:44	
Breath Volume	1.167	03:44	
Air Blank	0.000	03:44	
Control Test	0.078	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 02/28/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIK
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: SAYSAVANG

DATE/TIME OF TEST: 02/03/2023 01:51:57

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:40	
Control Test	0.080	01:41	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:44	No Sample Provided
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample	# 2	01:50	No Sample Provided
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 02/23/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/04/2023 01:20:09

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.207	01:15
Breath Volume	2.128	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.216	01:18
Breath Volume	1.808	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 02/23/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: WELTMAN

DATE/TIME OF TEST: 02/11/2023 02:03:12

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.080	01:55	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:58	No Sample Provided
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:02	
Control Test	0.081	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 02/23/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 02/24/2023 19:57:52

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.080	19:52
Air Blank	0.000	19:52
Subject Sample # 1	0.260	19:53
Breath Volume	2.238	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:55
Subject Sample # 2	0.245	19:56
Breath Volume	2.445	19:56
Air Blank	0.000	19:56
Control Test	0.079	19:57
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 02/25/2023 23:48:17

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.081	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.220	23:43
Breath Volume	1.328	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.215	23:46
Breath Volume	1.523	23:46
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANDRE,
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: MECHTEL,BRANDON
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 01/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 02/09/2023 04:06:53

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.145	04:02
Breath Volume	3.421	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.152	04:05
Breath Volume	3.699	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 02/21/2023 18:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS, ANDRE,
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: PRISE, DYSHEKA
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 01/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ROLLEFSON

DATE/TIME OF TEST: 02/10/2023 22:32:26

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.081	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 1 0.167	22:25	Volume Not Met
Breath Volume	0.714	22:25	
Air Blank	0.000	22:25	
Air Blank	0.000	22:27	
Subject Sample	# 2 0.170	22:30	Volume Not Met
Breath Volume	0.605	22:30	
Air Blank	0.000	22:31	
Control Test	0.079	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 02/21/2023 18:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPELAND,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: COPELAND,MICHAEL
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 02/16/2023 22:22:44

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.082	22:20	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:21	
Control Test	0.081	22:22	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 03/03/2023 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEARSON,ROBERT,B
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: CARDONE,DOUGLAS
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 02/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:36

SUBJECT'S LAST NAME: KAUFMAN

DATE/TIME OF TEST: 02/12/2023 13:05:09

Diagnostics Check	OK	12:58
Air Blank	0.000	12:59
Control Test	0.081	12:59
Air Blank	0.000	13:00
Subject Sample # 1	0.234	13:00
Breath Volume	2.308	13:00
Air Blank	0.000	13:01
Air Blank	0.000	13:03
Subject Sample # 2	0.226	13:03
Breath Volume	2.582	13:03
Air Blank	0.000	13:04
Control Test	0.079	13:04
Air Blank	0.000	13:05
Diagnostics Check	OK	13:05

Data Download Date/Time: 03/02/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: MORRIS,JOSHUA
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 12:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 02/22/2023 13:37:13

Diagnostics Check	OK	13:30
Air Blank	0.000	13:31
Control Test	0.080	13:31
Air Blank	0.000	13:32
Subject Sample # 1	0.210	13:32
Breath Volume	1.562	13:32
Air Blank	0.000	13:33
Air Blank	0.000	13:35
Subject Sample # 2	0.205	13:35
Breath Volume	1.609	13:35
Air Blank	0.000	13:36
Control Test	0.080	13:36
Air Blank	0.000	13:37
Diagnostics Check	OK	13:37

Data Download Date/Time: 03/31/2023 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEARSON,ROBERT,B
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: OVERTON,SCOTT
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 02/25/2023 19:46:13

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.080	19:40
Air Blank	0.000	19:41
Subject Sample # 1	0.173	19:41
Breath Volume	1.753	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:43
Subject Sample # 2	0.171	19:44
Breath Volume	1.804	19:44
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 03/31/2023 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: APONTE,ALEX
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 02/08/2023 02:57:23

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.080	02:54	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:56	
Control Test	0.080	02:56	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 02/24/2023 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PABON,G,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: CORBETT,R

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:27

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: QUILAS

DATE/TIME OF TEST: 02/13/2023 23:14:34

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.080	23:07
Air Blank	0.000	23:07
Subject Sample # 1	0.191	23:08
Breath Volume	1.503	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.173	23:12
Breath Volume	1.222	23:12
Air Blank	0.000	23:13
Control Test	0.078	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 02/24/2023 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LOWE,MAYCEE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: RAINES

DATE/TIME OF TEST: 02/14/2023 20:07:42

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.081	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.000	20:02
Breath Volume	2.761	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.000	20:06
Breath Volume	3.246	20:06
Air Blank	0.000	20:06
Control Test	0.081	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 02/24/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: CHAN, KAM-WING

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: LOMAX

DATE/TIME OF TEST: 02/15/2023 03:01:24

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.157	02:56
Breath Volume	1.703	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.156	02:59
Breath Volume	2.453	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 02/24/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PABON,G,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LOWE,M
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SOCUALAYA

DATE/TIME OF TEST: 02/20/2023 04:48:20

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.127	04:43
Breath Volume	4.097	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.121	04:46
Breath Volume	3.730	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 02/24/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,E
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 02/09/2023 21:53:27

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.080	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.051	21:48
Breath Volume	2.156	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.050	21:51
Breath Volume	3.417	21:51
Air Blank	0.000	21:52
Control Test	0.079	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 02/22/2023 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, RONALD,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: DEATON, GARRETT
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: MILLETTE

DATE/TIME OF TEST: 02/11/2023 07:34:53

Diagnostics Check	OK	07:30	
Air Blank	0.000	07:30	
Control Test	0.079	07:31	
Air Blank	0.000	07:31	
Subject Sample	# 1	07:33	Interferent Detect
Breath Volume	3.484	07:33	
Air Blank	0.000	07:34	
Control Test	0.076	07:34	Interferent Detect
Air Blank	0.000	07:34	

Data Download Date/Time: 02/22/2023 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, RONALD,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: DEATON, GARRETT
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: MILLETTE

DATE/TIME OF TEST: 02/11/2023 07:45:16

Diagnostics Check	OK	07:37
Air Blank	0.000	07:37
Control Test	0.079	07:37
Air Blank	0.000	07:38
Subject Sample # 1	0.133	07:39
Breath Volume	2.183	07:39
Air Blank	0.000	07:39
Air Blank	0.000	07:41
Subject Sample # 2	0.135	07:43
Breath Volume	2.957	07:43
Air Blank	0.000	07:44
Control Test	0.079	07:44
Air Blank	0.000	07:45
Diagnostics Check	OK	07:45

Data Download Date/Time: 02/22/2023 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ROLLINS,SEAN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: HAINRICK

DATE/TIME OF TEST: 02/13/2023 19:49:38

Diagnostics Check	OK	19:48	
Air Blank		19:49	RFI Detect
Air Blank	0.000	19:49	

Data Download Date/Time: 02/22/2023 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ROLLINS,SEAN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: HAINRICK

DATE/TIME OF TEST: 02/13/2023 19:58:49

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.081	19:53
Air Blank	0.000	19:53
Subject Sample # 1	0.207	19:54
Breath Volume	1.867	19:54
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.203	19:57
Breath Volume	2.070	19:57
Air Blank	0.000	19:57
Control Test	0.082	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 02/22/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,E
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: EASTRIDGE,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: COUCH

DATE/TIME OF TEST: 02/23/2023 00:48:55

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.184	00:44
Breath Volume	1.238	00:44
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.185	00:47
Breath Volume	1.894	00:47
Air Blank	0.000	00:47
Control Test	0.080	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 03/22/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,R,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: PENA,R
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: BUI

DATE/TIME OF TEST: 02/28/2023 18:58:25

Diagnostics Check	OK	18:47	
Air Blank	0.000	18:47	
Control Test	0.080	18:47	
Air Blank	0.000	18:48	
Subject Sample	# 1	18:51	No Sample Provided
Breath Volume	0.000	18:51	
Air Blank	0.000	18:51	
Air Blank	0.000	18:53	
Subject Sample	# 2	18:57	No Sample Provided
Breath Volume	0.000	18:57	
Air Blank	0.000	18:57	
Control Test	0.080	18:57	
Air Blank	0.000	18:58	
Diagnostics Check	OK	18:58	

Data Download Date/Time: 03/22/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: TRISTER

DATE/TIME OF TEST: 02/14/2023 23:04:24

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.195	22:59
Breath Volume	2.265	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.191	23:02
Breath Volume	2.375	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 03/09/2023 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: SHAW,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: LAUDERDALE

DATE/TIME OF TEST: 02/14/2023 23:38:05

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.079	23:36	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 03/09/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARROLL,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: GABRICK

DATE/TIME OF TEST: 02/15/2023 19:17:11

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.080	19:11
Air Blank	0.000	19:11
Subject Sample # 1	0.150	19:12
Breath Volume	2.875	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.143	19:15
Breath Volume	2.441	19:15
Air Blank	0.000	19:16
Control Test	0.079	19:16
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: HOEFLING

DATE/TIME OF TEST: 02/15/2023 21:54:29

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:49
Subject Sample # 1	0.223	21:49
Breath Volume	3.019	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.220	21:52
Breath Volume	3.734	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: HAYDEN

DATE/TIME OF TEST: 02/16/2023 04:25:28

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.206	04:17
Breath Volume	1.929	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:19
Subject Sample # 2	0.184	04:20
Breath Volume	1.878	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:22
Subject Sample # 3	0.195	04:23
Breath Volume	2.589	04:23
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERIC,JEAN,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FREDERIC,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 02/18/2023 22:39:46

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.081	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.102	22:34
Breath Volume	1.222	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:36
Subject Sample # 2	0.097	22:38
Breath Volume	1.308	22:38
Air Blank	0.000	22:38
Control Test	0.078	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: LANGLEY

DATE/TIME OF TEST: 02/19/2023 00:16:37

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.159	00:12
Breath Volume	1.125	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.163	00:15
Breath Volume	1.738	00:15
Air Blank	0.000	00:15
Control Test	0.079	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: KUNTZ

DATE/TIME OF TEST: 02/19/2023 20:31:38

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.080	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.218	20:27
Breath Volume	1.984	20:27
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.224	20:30
Breath Volume	1.421	20:30
Air Blank	0.000	20:30
Control Test	0.080	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERIC,JEAN,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FREDERIC,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 02/19/2023 23:13:42

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.000	23:09
Breath Volume	2.273	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.000	23:12
Breath Volume	2.261	23:12
Air Blank	0.000	23:12
Control Test	0.080	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY, CHRISTOPHER, L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: KOGER, AARON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: BUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: GRADY

DATE/TIME OF TEST: 02/20/2023 00:05:41

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:05	Sequence Aborted
Breath Volume	0.000	00:05	
Air Blank	0.000	00:05	

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 02/20/2023 17:55:49

Diagnostics Check	OK	17:49
Air Blank	0.000	17:49
Control Test	0.080	17:50
Air Blank	0.000	17:50
Subject Sample # 1	0.218	17:51
Breath Volume	1.843	17:51
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 2	0.222	17:54
Breath Volume	2.281	17:54
Air Blank	0.000	17:54
Control Test	0.079	17:55
Air Blank	0.000	17:55
Diagnostics Check	OK	17:55

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: DESROSIER

DATE/TIME OF TEST: 02/21/2023 03:45:08

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.080	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.000	03:40
Breath Volume	1.640	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.000	03:43
Breath Volume	1.660	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: GUSSE

DATE/TIME OF TEST: 02/21/2023 16:30:18

Diagnostics Check	OK	16:23
Air Blank	0.000	16:23
Control Test	0.079	16:24
Air Blank	0.000	16:24
Subject Sample # 1	0.295	16:25
Breath Volume	2.425	16:25
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.282	16:28
Breath Volume	3.406	16:28
Air Blank	0.000	16:29
Control Test	0.079	16:29
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: KYLE

DATE/TIME OF TEST: 02/23/2023 02:37:27

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.245	02:32
Breath Volume	2.585	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.239	02:35
Breath Volume	2.406	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: STERLE

DATE/TIME OF TEST: 02/23/2023 04:14:17

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.229	04:09
Breath Volume	2.964	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.218	04:12
Breath Volume	1.550	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOGER,AARON,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: KOGER,AARON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/23/2023 16:42:37

Diagnostics Check	OK	16:36
Air Blank	0.000	16:37
Control Test	0.079	16:37
Air Blank	0.000	16:37
Subject Sample # 1	0.212	16:38
Breath Volume	3.519	16:38
Air Blank	0.000	16:38
Air Blank	0.000	16:40
Subject Sample # 2	0.193	16:41
Breath Volume	3.777	16:41
Air Blank	0.000	16:41
Control Test	0.077	16:42
Air Blank	0.000	16:42
Diagnostics Check	OK	16:42

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LANDERS,KRISTOFER
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/23/2023 20:34:18

Diagnostics Check	OK	20:28
Air Blank	0.000	20:28
Control Test	0.080	20:28
Air Blank	0.000	20:29
Subject Sample # 1	0.000	20:29
Breath Volume	3.324	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.000	20:32
Breath Volume	1.929	20:32
Air Blank	0.000	20:33
Control Test	0.080	20:33
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: SANTOS OLIVERA

DATE/TIME OF TEST: 02/25/2023 01:43:37

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.080	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1 0.113	01:32	
Breath Volume	1.281	01:32	
Air Blank	0.000	01:32	
Air Blank	0.000	01:34	
Subject Sample	# 2 0.111	01:37	Volume Not Met
Breath Volume	0.777	01:37	
Air Blank	0.000	01:38	
Air Blank	0.000	01:40	
Subject Sample	# 3 0.111	01:42	
Breath Volume	2.085	01:42	
Air Blank	0.000	01:42	
Control Test	0.078	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: PENA SUAREZ

DATE/TIME OF TEST: 02/27/2023 01:01:24

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.124	00:56
Breath Volume	1.281	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.125	00:59
Breath Volume	1.179	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 02/11/2023 20:41:09

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.082	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.201	20:35
Breath Volume	3.261	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 2	0.210	20:39
Breath Volume	2.441	20:39
Air Blank	0.000	20:40
Control Test	0.082	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 03/10/2023 18:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NEGRON,JOHN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: BARTOLON LUCAS

DATE/TIME OF TEST: 02/18/2023 05:05:13

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.080	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.177	05:00
Breath Volume	1.257	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:02
Subject Sample # 2	0.168	05:03
Breath Volume	1.230	05:03
Air Blank	0.000	05:04
Control Test	0.077	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 03/07/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: DE LA CRUZ,JUAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: CORTES VASQUEZ

DATE/TIME OF TEST: 02/05/2023 18:14:41

Diagnostics Check	OK	18:04	
Air Blank	0.000	18:05	
Control Test	0.077	18:05	
Air Blank	0.000	18:05	
Subject Sample	# 1 0.257	18:07	
Breath Volume	1.914	18:07	
Air Blank	0.000	18:08	
Air Blank	0.000	18:10	
Subject Sample	# 2 0.245	18:13	
Breath Volume	1.187	18:13	
Air Blank	0.000	18:13	
Control Test	0.074	18:14	Control Outside Tolerance
Air Blank	0.000	18:14	

Data Download Date/Time: 02/24/2023 17:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,CALE,V
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: WILLIAMS,AARON
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 02/11/2023 01:02:19

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.077	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.197	00:57
Breath Volume	1.113	00:57
Air Blank	0.000	00:58
Air Blank	0.000	00:59
Subject Sample # 2	0.194	01:00
Breath Volume	1.187	01:00
Air Blank	0.000	01:01
Control Test	0.077	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 02/24/2023 17:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COSME,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 02/11/2023 22:20:58

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.078	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.088	22:16
Breath Volume	2.707	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:18
Subject Sample # 2	0.085	22:19
Breath Volume	2.281	22:19
Air Blank	0.000	22:20
Control Test	0.077	22:20
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 02/24/2023 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HALE,CALEB
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/12/2023 02:33:06

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.077	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.151	02:28
Breath Volume	2.089	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.155	02:31
Breath Volume	1.898	02:31
Air Blank	0.000	02:32
Control Test	0.077	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 02/24/2023 17:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CASEY,TYLER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 02/12/2023 03:22:22

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.077	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.181	03:17
Breath Volume	1.367	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.165	03:20
Breath Volume	1.531	03:20
Air Blank	0.000	03:21
Control Test	0.076	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 02/24/2023 17:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SILVOY,SHONA
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: BOND

DATE/TIME OF TEST: 02/12/2023 22:36:52

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
Control Test	0.077	22:31	
Air Blank	0.000	22:32	
Subject Sample # 1	0.130	22:32	
Breath Volume	3.339	22:32	
Air Blank	0.000	22:33	
Air Blank	0.000	22:34	
Subject Sample # 2	0.123	22:35	
Breath Volume	1.460	22:35	
Air Blank	0.000	22:35	
Control Test	0.074	22:36	Control Outside Tolerance
Air Blank	0.000	22:36	

Data Download Date/Time: 02/24/2023 17:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 02/13/2023 21:44:30

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.076	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.201	21:39
Breath Volume	2.484	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.204	21:43
Breath Volume	2.535	21:43
Air Blank	0.000	21:43
Control Test	0.076	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 02/24/2023 17:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CASEY,TYLER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 02/14/2023 00:00:17

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.077	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.184	23:55
Breath Volume	2.515	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.182	23:58
Breath Volume	1.167	23:58
Air Blank	0.000	23:59
Control Test	0.076	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 02/24/2023 17:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SILVOY,SHONA
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: CARCAMO

DATE/TIME OF TEST: 02/14/2023 20:08:16

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.077	20:00
Air Blank	0.000	20:01
Subject Sample # 1	0.218	20:03
Breath Volume	2.078	20:03
Air Blank	0.000	20:04
Air Blank	0.000	20:05
Subject Sample # 2	0.214	20:06
Breath Volume	2.140	20:06
Air Blank	0.000	20:07
Control Test	0.076	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 02/24/2023 17:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ZACCARO

DATE/TIME OF TEST: 02/15/2023 01:45:44

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.077	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1 0.185	01:35	Volume Not Met
Breath Volume	0.386	01:35	
Air Blank	0.000	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 2 0.202	01:38	
Breath Volume	2.394	01:38	
Air Blank	0.000	01:39	
Air Blank	0.000	01:41	
Subject Sample	# 3 0.192	01:44	Volume Not Met
Breath Volume	0.062	01:44	
Air Blank	0.000	01:44	
Control Test	0.076	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 02/24/2023 17:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 02/16/2023 02:33:50

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.076	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.263	02:29
Breath Volume	1.597	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.273	02:32
Breath Volume	1.746	02:32
Air Blank	0.000	02:32
Control Test	0.075	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 02/24/2023 17:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINGRY,HOBIE,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: KINGRY,HOBIE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 02/18/2023 02:29:45

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.078	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.000	02:25
Breath Volume	1.875	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.000	02:28
Breath Volume	2.082	02:28
Air Blank	0.000	02:28
Control Test	0.078	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 02/24/2023 17:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: LANZA VILLALOBO

DATE/TIME OF TEST: 02/19/2023 03:31:51

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.078	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1 0.212	03:22	Volume Not Met
Breath Volume	0.148	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 2 0.198	03:26	
Breath Volume	1.316	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample	# 3 0.214	03:30	
Breath Volume	1.941	03:30	
Air Blank	0.000	03:30	
Control Test	0.076	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 02/24/2023 17:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: DE LA CRUZ,JUAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 02/21/2023 05:11:46

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.078	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.144	05:07
Breath Volume	1.640	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.151	05:10
Breath Volume	1.777	05:10
Air Blank	0.000	05:10
Control Test	0.076	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 02/24/2023 17:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CASEY,TYLER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/22/2023 04:23:31

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.077	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.172	04:18
Breath Volume	2.074	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.168	04:22
Breath Volume	2.019	04:22
Air Blank	0.000	04:22
Control Test	0.076	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 02/24/2023 17:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 02/23/2023 03:59:10

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.077	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.169	03:54
Breath Volume	1.449	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.175	03:57
Breath Volume	2.285	03:57
Air Blank	0.000	03:58
Control Test	0.076	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 02/24/2023 17:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORE,BECKY,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CASEY,TYLER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HILBURN

DATE/TIME OF TEST: 02/26/2023 02:03:22

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.075	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1 0.246	02:00	
Breath Volume	1.492	02:00	
Air Blank	0.000	02:01	
Air Blank		02:03	Improper Sample
Air Blank		02:03	Improper Sample
Air Blank		02:03	Improper Sample

Data Download Date/Time: 04/10/2023 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORE,BECKY,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CASEY,TYLER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HILBURN

DATE/TIME OF TEST: 02/26/2023 02:09:33

Diagnostics Check	OK	02:08	
Air Blank		02:09	Ambient Fail
Air Blank	0.000	02:09	

Data Download Date/Time: 04/10/2023 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORE,BECKY,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CASEY,TYLER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HILBURN

DATE/TIME OF TEST: 02/26/2023 02:13:00

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:12	
Control Test	0.074	02:12	Control Outside Tolerance
Air Blank	0.000	02:13	

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DENNIS,JONATHAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: TULLY

DATE/TIME OF TEST: 02/02/2023 04:07:57

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.076	04:03
Breath Volume	2.226	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.071	04:06
Breath Volume	1.625	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MENDEZ,VIOLET
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: DRURY

DATE/TIME OF TEST: 02/02/2023 23:26:54

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.190	23:22
Breath Volume	3.523	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:24
Subject Sample # 2	0.181	23:25
Breath Volume	3.664	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JACKSON,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CUSTODIO DE OLIVEIRA

DATE/TIME OF TEST: 02/05/2023 04:37:41

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.081	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.113	04:33
Breath Volume	2.562	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.110	04:36
Breath Volume	1.531	04:36
Air Blank	0.000	04:36
Control Test	0.081	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,ROHAN,G
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: LEON,BRIAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: BAMACA LOPEZ

DATE/TIME OF TEST: 02/05/2023 21:39:20

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.081	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.223	21:34
Breath Volume	2.660	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.207	21:37
Breath Volume	2.886	21:37
Air Blank	0.000	21:38
Control Test	0.081	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MATERA,DERICK
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/06/2023 05:05:27

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:01	
Control Test	0.081	05:01	
Air Blank	0.000	05:01	
Subject Sample	# 1	05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:04	
Control Test	0.081	05:04	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: BEYRENT

DATE/TIME OF TEST: 02/06/2023 23:14:26

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.082	23:09	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:13	
Control Test	0.081	23:13	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DENNIS, JONATHAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: PATITUCCI

DATE/TIME OF TEST: 02/07/2023 20:38:36

Diagnostics Check	OK	20:32
Air Blank	0.000	20:32
Control Test	0.081	20:33
Air Blank	0.000	20:33
Subject Sample # 1	0.050	20:34
Breath Volume	2.429	20:34
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.048	20:37
Breath Volume	2.875	20:37
Air Blank	0.000	20:37
Control Test	0.080	20:38
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HOYOS,.
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: RAMIREZ-ANGULO

DATE/TIME OF TEST: 02/08/2023 01:10:38

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.081	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.188	01:05
Breath Volume	1.277	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:07
Subject Sample # 2	0.206	01:09
Breath Volume	1.933	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DENNIS, JONATHAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: RODRIGUEZ RODRIGUEZ

DATE/TIME OF TEST: 02/08/2023 04:00:08

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.231	03:55
Breath Volume	1.535	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.227	03:58
Breath Volume	1.429	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ARMSTRONG,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: TOMAS MATEO

DATE/TIME OF TEST: 02/09/2023 00:51:34

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.085	00:46
Breath Volume	3.171	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.075	00:50
Breath Volume	2.617	00:50
Air Blank	0.000	00:50
Control Test	0.081	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HAACK PEREZ

DATE/TIME OF TEST: 02/09/2023 03:57:46

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.081	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.147	03:53
Breath Volume	1.734	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample # 2	0.153	03:56
Breath Volume	2.222	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: MORALES DOMINGUEZ

DATE/TIME OF TEST: 02/11/2023 00:58:29

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.112	00:54
Breath Volume	1.453	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.107	00:57
Breath Volume	2.105	00:57
Air Blank	0.000	00:57
Control Test	0.081	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PHILLIPS, MERCEDES
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 02/11/2023 02:55:26

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.198	02:50
Breath Volume	1.988	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.185	02:53
Breath Volume	1.320	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SMUCKER

DATE/TIME OF TEST: 02/12/2023 02:02:01

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.054	01:57
Breath Volume	2.835	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.055	02:00
Breath Volume	2.632	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:02

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRUM,JACOB,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NEWBERN,JAMES
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: LUIS-DIEGO

DATE/TIME OF TEST: 02/12/2023 11:36:55

Diagnostics Check	OK	11:30
Air Blank	0.000	11:30
Control Test	0.080	11:31
Air Blank	0.000	11:31
Subject Sample # 1	0.205	11:32
Breath Volume	1.214	11:32
Air Blank	0.000	11:32
Air Blank	0.000	11:34
Subject Sample # 2	0.192	11:35
Breath Volume	1.820	11:35
Air Blank	0.000	11:36
Control Test	0.080	11:36
Air Blank	0.000	11:36
Diagnostics Check	OK	11:36

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SERRES, DEVAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: BOYCE

DATE/TIME OF TEST: 02/13/2023 00:53:14

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.177	00:48
Breath Volume	1.476	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.170	00:51
Breath Volume	2.023	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MAKSIMOVIC, DRAGOSLAV
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 02/13/2023 01:03:49

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.081	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.105	00:59
Breath Volume	2.121	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.103	01:02
Breath Volume	1.937	01:02
Air Blank	0.000	01:02
Control Test	0.081	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DELGATTI, BRADLEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 02/13/2023 01:14:50

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.143	01:10
Breath Volume	1.906	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.147	01:13
Breath Volume	2.277	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: GERRISH

DATE/TIME OF TEST: 02/13/2023 03:49:59

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.166	03:44
Breath Volume	2.019	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.157	03:48
Breath Volume	1.308	03:48
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: JESUS PEDRO

DATE/TIME OF TEST: 02/13/2023 04:26:57

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.182	04:22
Breath Volume	3.613	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.176	04:25
Breath Volume	3.718	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHIERA, FERNANDO,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CORNIER, GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: CONDRY

DATE/TIME OF TEST: 02/15/2023 16:35:35

Diagnostics Check	OK	16:34	
Air Blank	0.000	16:34	
Control Test	0.079	16:35	RFI Detect
Air Blank	0.000	16:35	

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHIERA,FERNANDO,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CORNIER,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: CONDRY

DATE/TIME OF TEST: 02/15/2023 16:44:50

Diagnostics Check	OK	16:38
Air Blank	0.000	16:39
Control Test	0.081	16:39
Air Blank	0.000	16:40
Subject Sample # 1	0.161	16:40
Breath Volume	2.417	16:40
Air Blank	0.000	16:41
Air Blank	0.000	16:42
Subject Sample # 2	0.154	16:43
Breath Volume	2.265	16:43
Air Blank	0.000	16:43
Control Test	0.081	16:44
Air Blank	0.000	16:44
Diagnostics Check	OK	16:44

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: FEWSTER

DATE/TIME OF TEST: 02/16/2023 23:47:08

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.081	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.235	23:41
Breath Volume	1.582	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.226	23:45
Breath Volume	1.710	23:45
Air Blank	0.000	23:46
Control Test	0.081	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 02/21/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KARSNICK,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: OXLAJ-SARAT

DATE/TIME OF TEST: 02/17/2023 01:41:42

Diagnostics Check	OK	01:41	
Air Blank		01:41	RFI Detect
Air Blank	0.000	01:41	

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KARSNICK,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: OXLAJ-SARAT

DATE/TIME OF TEST: 02/17/2023 01:50:06

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.150	01:45
Breath Volume	3.675	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.142	01:48
Breath Volume	2.113	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: COKER,JUSTIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: NEVEAU

DATE/TIME OF TEST: 02/18/2023 01:00:25

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.271	00:52
Breath Volume	2.167	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.249	00:55
Breath Volume	1.839	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 3	0.256	00:58
Breath Volume	1.769	00:58
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JACKSON,TYRESE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: CALMO GOMEZ

DATE/TIME OF TEST: 02/18/2023 02:57:53

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.081	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.118	02:53
Breath Volume	2.250	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.119	02:56
Breath Volume	2.089	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/18/2023 04:38:14

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.058	04:33
Breath Volume	1.550	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.059	04:36
Breath Volume	1.343	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CAPURSI,R
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: DALMAN

DATE/TIME OF TEST: 02/18/2023 05:49:06

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.081	05:43
Air Blank	0.000	05:44
Subject Sample # 1	0.203	05:44
Breath Volume	2.980	05:44
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.203	05:47
Breath Volume	2.683	05:47
Air Blank	0.000	05:48
Control Test	0.080	05:48
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/18/2023 23:09:11

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.081	23:06	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:07	Subject Test Refused
Breath Volume	0.000	23:07	
Air Blank	0.000	23:08	
Control Test	0.081	23:08	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING, JONATHAN, L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: KIRCIKYAN, RAFFI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PEREZ SANTOS

DATE/TIME OF TEST: 02/19/2023 03:39:40

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.186	03:35
Breath Volume	2.550	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.185	03:38
Breath Volume	3.601	03:38
Air Blank	0.000	03:38
Control Test	0.078	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: KERKMAN

DATE/TIME OF TEST: 02/19/2023 07:34:27

Diagnostics Check	OK	07:28
Air Blank	0.000	07:28
Control Test	0.081	07:29
Air Blank	0.000	07:29
Subject Sample # 1	0.099	07:30
Breath Volume	3.199	07:30
Air Blank	0.000	07:30
Air Blank	0.000	07:32
Subject Sample # 2	0.094	07:32
Breath Volume	2.171	07:32
Air Blank	0.000	07:33
Control Test	0.081	07:33
Air Blank	0.000	07:34
Diagnostics Check	OK	07:34

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,CLAYTON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 02/19/2023 22:18:56

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.081	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.270	22:14
Breath Volume	3.238	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:16
Subject Sample # 2	0.257	22:17
Breath Volume	3.136	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 02/20/2023 01:17:21

Diagnostics Check	OK	01:16	
Air Blank		01:16	RFI Detect
Air Blank	0.000	01:17	

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 02/20/2023 01:25:49

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.158	01:21
Breath Volume	2.171	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:23
Subject Sample # 2	0.165	01:24
Breath Volume	2.195	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRUM,JACOB,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NEWBERN,JAMES
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:11

SUBJECT'S LAST NAME: DAIGLE

DATE/TIME OF TEST: 02/20/2023 13:39:56

Diagnostics Check	OK	13:33
Air Blank	0.000	13:34
Control Test	0.081	13:34
Air Blank	0.000	13:34
Subject Sample # 1	0.000	13:35
Breath Volume	1.671	13:35
Air Blank	0.000	13:35
Air Blank	0.000	13:37
Subject Sample # 2	0.000	13:38
Breath Volume	1.953	13:38
Air Blank	0.000	13:39
Control Test	0.081	13:39
Air Blank	0.000	13:39
Diagnostics Check	OK	13:39

Data Download Date/Time: 02/21/2023 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: KARDUM

DATE/TIME OF TEST: 02/23/2023 02:33:54

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.167	02:28
Breath Volume	1.839	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample # 2	0.163	02:32
Breath Volume	3.476	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MELLO JR

DATE/TIME OF TEST: 02/24/2023 02:42:00

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.081	02:39	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:41	
Control Test	0.081	02:41	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BATISTA,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: NERRY

DATE/TIME OF TEST: 02/24/2023 20:41:09

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.081	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.298	20:36
Breath Volume	2.246	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.306	20:39
Breath Volume	3.160	20:39
Air Blank	0.000	20:40
Control Test	0.080	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MENDEZ, VIOLET
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: OBANDO

DATE/TIME OF TEST: 02/25/2023 03:17:14

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.196	03:12
Breath Volume	3.132	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.185	03:15
Breath Volume	3.500	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARNETT, CODY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/25/2023 03:34:18

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.080	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BAIR,J
ARRESTING OFFICER AGENCY: FL Gulf Coast UPD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: VENTRE

DATE/TIME OF TEST: 02/25/2023 03:55:20

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.175	03:50
Breath Volume	2.339	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.175	03:53
Breath Volume	2.179	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/26/2023 03:08:24

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:03	
Control Test	0.081	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.081	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: PIGOTT

DATE/TIME OF TEST: 02/27/2023 21:20:32

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:17	
Control Test	0.080	21:18	
Air Blank	0.000	21:18	
Subject Sample	# 1	21:19	Subject Test Refused
Breath Volume	0.000	21:19	
Air Blank	0.000	21:19	
Control Test	0.081	21:19	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DEFILLIPIS,MICHAEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: JEWETT

DATE/TIME OF TEST: 02/03/2023 02:39:28

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.164	02:34
Breath Volume	1.460	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.173	02:37
Breath Volume	2.500	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 02/24/2023 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: CHARLES,JAMESLOO
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: BAIRD

DATE/TIME OF TEST: 02/06/2023 03:32:35

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.081	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Control Test	0.081	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 02/24/2023 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: CHARLES,JAMESLOO
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 02/12/2023 04:38:10

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.131	04:33
Breath Volume	2.796	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:35
Subject Sample # 2	0.131	04:36
Breath Volume	2.449	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 02/24/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: ROBERTS,NATALIE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: SANCHEZ LOPEZ

DATE/TIME OF TEST: 02/12/2023 21:08:17

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.082	21:02
Air Blank	0.000	21:02
Subject Sample # 1	0.205	21:03
Breath Volume	2.914	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample # 2	0.207	21:06
Breath Volume	3.218	21:06
Air Blank	0.000	21:07
Control Test	0.079	21:07
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

Data Download Date/Time: 02/24/2023 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: ROBERTS,NATALIE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/13/2023 02:21:06

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.081	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Control Test	0.081	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 02/24/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: ZACARIAS PEREZ

DATE/TIME OF TEST: 02/17/2023 18:46:34

Diagnostics Check	OK	18:39
Air Blank	0.000	18:40
Control Test	0.080	18:40
Air Blank	0.000	18:41
Subject Sample # 1	0.081	18:42
Breath Volume	2.410	18:42
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.082	18:45
Breath Volume	2.746	18:45
Air Blank	0.000	18:45
Control Test	0.080	18:45
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 02/24/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,EDWARD,E
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MENDEZ,ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: KUHLMANN

DATE/TIME OF TEST: 02/23/2023 21:56:34

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.081	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.128	21:52
Breath Volume	3.992	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.126	21:55
Breath Volume	3.714	21:55
Air Blank	0.000	21:55
Control Test	0.081	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 02/24/2023 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO, ENRIQUE, A
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: WOOLEYHAN, HEATHER
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: WHELAN

DATE/TIME OF TEST: 02/01/2023 06:08:45

Diagnostics Check	OK	05:58
Air Blank	0.000	05:58
Control Test	0.080	05:59
Air Blank	0.000	05:59
Subject Sample # 1	0.090	06:02
Breath Volume	1.351	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.089	06:07
Breath Volume	1.246	06:07
Air Blank	0.000	06:07
Control Test	0.078	06:08
Air Blank	0.000	06:08
Diagnostics Check	OK	06:08

Data Download Date/Time: 02/16/2023 23:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRAZANA,DANIEL,R
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CARRAZANA,DANIEL
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/13/2023 15:32:35

Diagnostics Check	OK	15:26
Air Blank	0.000	15:26
Control Test	0.081	15:27
Air Blank	0.000	15:27
Subject Sample # 1	0.000	15:28
Breath Volume	2.796	15:28
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.000	15:31
Breath Volume	2.257	15:31
Air Blank	0.000	15:31
Control Test	0.079	15:31
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

Data Download Date/Time: 02/16/2023 23:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEGER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: FEGER,JOSHUA
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 02/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: ARMITAGE

DATE/TIME OF TEST: 02/06/2023 17:42:09

Diagnostics Check	OK	17:36
Air Blank	0.000	17:36
Control Test	0.078	17:37
Air Blank	0.000	17:37
Subject Sample # 1	0.000	17:38
Breath Volume	1.734	17:38
Air Blank	0.000	17:38
Air Blank	0.000	17:40
Subject Sample # 2	0.000	17:40
Breath Volume	1.859	17:40
Air Blank	0.000	17:41
Control Test	0.078	17:41
Air Blank	0.000	17:42
Diagnostics Check	OK	17:42

Data Download Date/Time: 03/24/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIPPEON,DANIEL,P
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: FEGER,JOSHUA
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 02/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:22

SUBJECT'S LAST NAME: KODIS

DATE/TIME OF TEST: 02/21/2023 10:31:49

Diagnostics Check	OK	10:25
Air Blank	0.000	10:26
Control Test	0.078	10:26
Air Blank	0.000	10:27
Subject Sample # 1	0.000	10:27
Breath Volume	2.832	10:27
Air Blank	0.000	10:28
Air Blank	0.000	10:30
Subject Sample # 2	0.000	10:30
Breath Volume	3.011	10:30
Air Blank	0.000	10:30
Control Test	0.076	10:31
Air Blank	0.000	10:31
Diagnostics Check	OK	10:31

Data Download Date/Time: 03/24/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEGER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: HAGAN,KASEY
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 02/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 11:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:56

SUBJECT'S LAST NAME: HOLCOMB

DATE/TIME OF TEST: 02/23/2023 12:23:47

Diagnostics Check	OK	12:18
Air Blank	0.000	12:18
Control Test	0.078	12:18
Air Blank	0.000	12:19
Subject Sample # 1	0.000	12:19
Breath Volume	2.167	12:19
Air Blank	0.000	12:20
Air Blank	0.000	12:22
Subject Sample # 2	0.000	12:22
Breath Volume	2.074	12:22
Air Blank	0.000	12:22
Control Test	0.078	12:23
Air Blank	0.000	12:23
Diagnostics Check	OK	12:23

Data Download Date/Time: 03/24/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFFNEY,CASEY,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: MENDOZA,JONOTHAN
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MATUSEK

DATE/TIME OF TEST: 02/08/2023 02:32:56

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.082	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.246	02:24
Breath Volume	2.593	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.223	02:27
Breath Volume	1.898	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 3	0.229	02:31
Breath Volume	1.714	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 03/25/2023 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFFNEY,CASEY,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: MEANY,TREVOR
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: DREVER

DATE/TIME OF TEST: 02/09/2023 00:30:12

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.082	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.000	00:25
Breath Volume	2.644	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.000	00:28
Breath Volume	2.746	00:28
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 03/25/2023 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS,TERRY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 02/27/2023 23:10:23

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.220	23:05
Breath Volume	1.656	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:07
Subject Sample # 2	0.220	23:08
Breath Volume	1.992	23:08
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 03/30/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JASON,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FITZGERALD,KYLE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:06

SUBJECT'S LAST NAME: LAPIERRE

DATE/TIME OF TEST: 02/28/2023 11:01:07

Diagnostics Check	OK	10:54
Air Blank	0.000	10:54
Control Test	0.080	10:54
Air Blank	0.000	10:55
Subject Sample # 1	0.000	10:56
Breath Volume	3.105	10:56
Air Blank	0.000	10:56
Air Blank	0.000	10:58
Subject Sample # 2	0.000	10:59
Breath Volume	1.382	10:59
Air Blank	0.000	11:00
Control Test	0.081	11:00
Air Blank	0.000	11:01
Diagnostics Check	OK	11:01

Data Download Date/Time: 03/30/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAENGCHOTE,PHANUDEJ
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 02/04/2023 02:25:47

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.177	02:21
Breath Volume	2.292	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.175	02:24
Breath Volume	1.347	02:24
Air Blank	0.000	02:24
Control Test	0.079	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 02/14/2023 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FLORES,MATHEUS
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PETTAY

DATE/TIME OF TEST: 02/05/2023 03:15:23

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.081	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.213	03:10
Breath Volume	1.656	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.224	03:13
Breath Volume	1.628	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 02/14/2023 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MALDONADO,ADAM
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: TOLLIVER

DATE/TIME OF TEST: 02/05/2023 05:25:24

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.080	05:20
Air Blank	0.000	05:20
Subject Sample # 1	0.063	05:21
Breath Volume	2.773	05:21
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.062	05:23
Breath Volume	2.593	05:23
Air Blank	0.000	05:24
Control Test	0.079	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 02/14/2023 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KOPINSKI,PATRYK
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/11/2023 05:18:59

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.080	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.173	05:14
Breath Volume	2.394	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.172	05:17
Breath Volume	2.250	05:17
Air Blank	0.000	05:18
Control Test	0.079	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 02/14/2023 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 02/18/2023 03:32:00

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.110	03:27
Breath Volume	2.390	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.114	03:30
Breath Volume	2.437	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:32

Data Download Date/Time: 03/14/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUTFL, RAED
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: TUR ESLES

DATE/TIME OF TEST: 02/19/2023 02:32:43

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.104	02:28
Breath Volume	1.550	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample # 2	0.100	02:31
Breath Volume	1.375	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 03/14/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI, ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: NAVARRO CARDOZO

DATE/TIME OF TEST: 02/20/2023 04:27:10

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.220	04:22
Breath Volume	1.281	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.231	04:25
Breath Volume	1.308	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 03/14/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SMITH, JAMES
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/23/2023 21:13:04

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.080	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.017	21:06
Breath Volume	1.304	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.016	21:11
Breath Volume	1.335	21:11
Air Blank	0.000	21:12
Control Test	0.081	21:12
Air Blank	0.000	21:12
Diagnostics Check	OK	21:13

Data Download Date/Time: 03/14/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: UMANA, RAUL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: SUAREZ VILLEGAS

DATE/TIME OF TEST: 02/24/2023 21:49:18

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.080	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.113	21:45
Breath Volume	1.382	21:45
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.120	21:47
Breath Volume	1.664	21:47
Air Blank	0.000	21:48
Control Test	0.078	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 03/14/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: SANTIAGO SANTOS

DATE/TIME OF TEST: 02/25/2023 01:18:47

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.081	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.014	01:14
Breath Volume	2.589	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.013	01:17
Breath Volume	2.851	01:17
Air Blank	0.000	01:17
Control Test	0.079	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 03/14/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER, DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: ROSA TOLBERT

DATE/TIME OF TEST: 02/25/2023 02:14:00

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.080	02:01	
Air Blank	0.000	02:02	
Subject Sample # 1	0.058	02:04	
Breath Volume	1.675	02:04	
Air Blank	0.000	02:04	
Air Blank	0.000	02:06	
Subject Sample # 2	0.044	02:09	Volume Not Met
Breath Volume	0.562	02:09	
Air Blank	0.000	02:10	
Air Blank	0.000	02:12	
Subject Sample # 3		02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Control Test	0.079	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 03/14/2023 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: SHEREL

DATE/TIME OF TEST: 02/25/2023 03:35:05

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.149	03:30
Breath Volume	2.253	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.148	03:33
Breath Volume	1.421	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 03/14/2023 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SALES MARTIN

DATE/TIME OF TEST: 02/26/2023 04:13:37

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.237	04:09
Breath Volume	2.640	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.234	04:12
Breath Volume	1.703	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 03/14/2023 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: LOAN

DATE/TIME OF TEST: 02/09/2023 23:11:56

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.081	23:03
Air Blank	0.000	23:04
Subject Sample # 1	0.263	23:06
Breath Volume	2.136	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.260	23:10
Breath Volume	2.582	23:10
Air Blank	0.000	23:11
Control Test	0.078	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 02/14/2023 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: KRAMER,KAYLEE
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 01/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: GIANCATARINO

DATE/TIME OF TEST: 02/06/2023 02:11:27

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.151	02:06
Breath Volume	2.828	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.155	02:09
Breath Volume	4.398	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 02/20/2023 07:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK, MITCHELL, A
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SCHALLON, CHRISTIAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 01/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 02/17/2023 00:44:25

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:40	
Control Test	0.080	00:40	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.179	00:42	
Air Blank	0.000	00:43	
Control Test	0.080	00:43	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 02/20/2023 07:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: KRAMER,TREVOR
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 02/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: ROSENQUIST

DATE/TIME OF TEST: 02/17/2023 23:38:55

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:34	
Control Test	0.080	23:34	
Air Blank	0.000	23:35	
Subject Sample	# 1	23:37	Subject Test Refused
Breath Volume	0.000	23:37	
Air Blank	0.000	23:38	
Control Test	0.080	23:38	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 02/20/2023 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RODRIGUEZ,EMMANUEL
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/01/2023 03:08:44

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.152	03:04
Breath Volume	1.164	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.145	03:07
Breath Volume	2.289	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 02/08/2023 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 13:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 02/01/2023 13:57:17

Diagnostics Check	OK	13:51
Air Blank	0.000	13:51
Control Test	0.080	13:51
Air Blank	0.000	13:52
Subject Sample # 1	0.202	13:52
Breath Volume	2.468	13:52
Air Blank	0.000	13:53
Air Blank	0.000	13:55
Subject Sample # 2	0.198	13:55
Breath Volume	2.449	13:55
Air Blank	0.000	13:56
Control Test	0.080	13:56
Air Blank	0.000	13:57
Diagnostics Check	OK	13:57

Data Download Date/Time: 02/08/2023 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PONDER,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: PONDER,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: SHADD

DATE/TIME OF TEST: 02/02/2023 16:30:41

Diagnostics Check	OK	16:24
Air Blank	0.000	16:25
Control Test	0.080	16:25
Air Blank	0.000	16:25
Subject Sample # 1	0.000	16:26
Breath Volume	1.960	16:26
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.000	16:29
Breath Volume	2.199	16:29
Air Blank	0.000	16:29
Control Test	0.079	16:30
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 02/08/2023 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DIAL,V
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 02/03/2023 03:28:24

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.226	03:23
Breath Volume	1.781	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.229	03:26
Breath Volume	1.863	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 02/08/2023 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: IRIZARRY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: CHESTNUT

DATE/TIME OF TEST: 02/03/2023 04:44:48

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.146	04:40
Breath Volume	1.324	04:40
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.147	04:43
Breath Volume	2.355	04:43
Air Blank	0.000	04:43
Control Test	0.080	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 02/08/2023 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUMSTEIN,REBEKAH,L
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,C
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: CORNIEL

DATE/TIME OF TEST: 02/04/2023 17:41:38

Diagnostics Check	OK	17:40	
Air Blank	0.000	17:40	
Control Test	0.080	17:40	
Air Blank		17:41	RFI Detect
Air Blank	0.000	17:41	

Data Download Date/Time: 02/08/2023 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUMSTEIN,REBEKAH,L
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,C
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: CORNIEL

DATE/TIME OF TEST: 02/04/2023 17:55:58

Diagnostics Check	OK	17:44
Air Blank	0.000	17:44
Control Test	0.079	17:44
Air Blank	0.000	17:45
Subject Sample # 1	0.232	17:47
Breath Volume	2.023	17:47
Air Blank	0.000	17:48
Air Blank	0.000	17:49
Subject Sample # 2	0.211	17:50
Breath Volume	1.960	17:50
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 3	0.225	17:54
Breath Volume	1.425	17:54
Air Blank	0.000	17:55
Control Test	0.077	17:55
Air Blank	0.000	17:55
Diagnostics Check	OK	17:55

Data Download Date/Time: 02/08/2023 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GANT,JARRET
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: VELAND

DATE/TIME OF TEST: 02/05/2023 02:48:43

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.144	02:44
Breath Volume	1.199	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.140	02:47
Breath Volume	1.363	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 02/08/2023 04:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,JERRY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 02/05/2023 04:10:28

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.116	04:05
Breath Volume	1.687	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.115	04:08
Breath Volume	1.640	04:08
Air Blank	0.000	04:09
Control Test	0.080	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 02/08/2023 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: TEEHAN,M
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: AHLQUIST

DATE/TIME OF TEST: 02/05/2023 08:29:19

Diagnostics Check	OK	08:21
Air Blank	0.000	08:22
Control Test	0.080	08:22
Air Blank	0.000	08:23
Subject Sample # 1	0.157	08:24
Breath Volume	3.031	08:24
Air Blank	0.000	08:24
Air Blank	0.000	08:26
Subject Sample # 2	0.153	08:27
Breath Volume	2.433	08:27
Air Blank	0.000	08:28
Control Test	0.080	08:28
Air Blank	0.000	08:29
Diagnostics Check	OK	08:29

Data Download Date/Time: 02/08/2023 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FORTENBERRY,B
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/05/2023 21:05:30

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:54	
Control Test	0.079	20:54	
Air Blank	0.000	20:55	
Subject Sample	# 1 0.217	20:58	Volume Not Met
Breath Volume	0.269	20:58	
Air Blank	0.000	20:58	
Air Blank	0.000	21:00	
Subject Sample	# 2 0.205	21:03	Volume Not Met
Breath Volume	0.164	21:03	
Air Blank	0.000	21:04	
Control Test	0.079	21:04	
Air Blank	0.000	21:05	
Diagnostics Check	OK	21:05	

Data Download Date/Time: 02/08/2023 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FORTENBERRY,B
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/05/2023 21:27:41

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:12	
Control Test	0.079	21:12	
Air Blank	0.000	21:12	
Subject Sample	# 1 0.226	21:16	
Breath Volume	1.851	21:16	
Air Blank	0.000	21:16	
Air Blank	0.000	21:18	
Subject Sample	# 2 0.196	21:21	Volume Not Met
Breath Volume	0.101	21:21	
Air Blank	0.000	21:22	
Air Blank	0.000	21:24	
Subject Sample	# 3 0.212	21:25	
Breath Volume	1.460	21:25	
Air Blank		21:26	Purge Fail
Air Blank	0.000	21:27	

Data Download Date/Time: 02/08/2023 04:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REHDER,PATRICK,W
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: JOHNSON,ANDREW
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: COOLEY

DATE/TIME OF TEST: 02/06/2023 19:29:19

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:25	
Control Test	0.077	19:26	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:28	Sequence Aborted
Breath Volume	0.000	19:28	
Air Blank	0.000	19:29	

Data Download Date/Time: 02/08/2023 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REHDER,PATRICK,W
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GIOIELLI,ANTHONY
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: SALAS LOPEZ

DATE/TIME OF TEST: 02/06/2023 20:06:46

Diagnostics Check	OK	19:52	
Air Blank	0.000	19:53	
Control Test	0.078	19:53	
Air Blank	0.000	19:54	
Subject Sample	# 1 0.234	19:57	Volume Not Met
Breath Volume	0.523	19:57	
Air Blank	0.000	19:58	
Air Blank	0.000	19:59	
Subject Sample	# 2 0.244	20:02	
Breath Volume	1.242	20:02	
Air Blank	0.000	20:03	
Air Blank	0.000	20:04	
Subject Sample	# 3	20:05	RFI Detect
Breath Volume	0.972	20:05	
Air Blank	0.000	20:05	
Control Test	0.076	20:06	
Air Blank	0.000	20:06	
Diagnostics Check	OK	20:06	

Data Download Date/Time: 02/08/2023 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REHDER,PATRICK,W
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GIOIELLI,ANTHONY
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: SALAS LOPEZ

DATE/TIME OF TEST: 02/06/2023 20:11:43

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.078	20:08
Air Blank	0.000	20:09
Subject Sample # 1	0.249	20:10
Breath Volume	1.105	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:11

Data Download Date/Time: 02/08/2023 04:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,S,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,S
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 10:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:29

SUBJECT'S LAST NAME: OROZCO AGUILAR

DATE/TIME OF TEST: 02/09/2023 12:13:08

Diagnostics Check	OK	12:05
Air Blank	0.000	12:06
Control Test	0.081	12:06
Air Blank	0.000	12:06
Subject Sample # 1	0.197	12:07
Breath Volume	2.730	12:07
Air Blank	0.000	12:08
Air Blank	0.000	12:10
Subject Sample # 2	0.181	12:11
Breath Volume	2.433	12:11
Air Blank	0.000	12:12
Control Test	0.080	12:12
Air Blank	0.000	12:13
Diagnostics Check	OK	12:13

Data Download Date/Time: 02/22/2023 21:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HAMILTON,DWAYNE
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/10/2023 22:11:34

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.080	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.136	22:07
Breath Volume	1.765	22:07
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.139	22:10
Breath Volume	2.218	22:10
Air Blank	0.000	22:10
Control Test	0.080	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 02/23/2023 04:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANCHEZ,JUBILEE
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 02/10/2023 23:26:13

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.092	23:21
Breath Volume	2.214	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:23
Subject Sample # 2	0.090	23:24
Breath Volume	1.503	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 02/23/2023 04:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIDEROA,KEVIN,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: HIDEROA,KEVIN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: COTTO-MORENO

DATE/TIME OF TEST: 02/11/2023 17:48:31

Diagnostics Check	OK	17:45	
Air Blank	0.000	17:46	
Control Test	0.080	17:46	
Air Blank	0.000	17:46	
Subject Sample	# 1	17:47	Subject Test Refused
Breath Volume	0.000	17:47	
Air Blank	0.000	17:47	
Control Test	0.080	17:47	
Air Blank	0.000	17:48	
Diagnostics Check	OK	17:48	

Data Download Date/Time: 02/23/2023 04:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: JOHNSON,A
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 02/12/2023 02:41:07

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.176	02:36
Breath Volume	1.574	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:38
Subject Sample # 2	0.170	02:39
Breath Volume	1.667	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 02/23/2023 04:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UCELO,B
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: REAMER

DATE/TIME OF TEST: 02/12/2023 04:16:58

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.189	04:11
Breath Volume	1.992	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.174	04:15
Breath Volume	1.878	04:15
Air Blank	0.000	04:16
Control Test	0.078	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 02/23/2023 04:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KAISER,DAVID
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 02/12/2023 19:28:31

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:26	
Control Test	0.078	19:26	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:27	Subject Test Refused
Breath Volume	0.000	19:27	
Air Blank	0.000	19:27	
Control Test	0.077	19:27	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:28	

Data Download Date/Time: 02/23/2023 04:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KAISER,DAVID
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 02/12/2023 19:47:22

Diagnostics Check	OK	19:40	
Air Blank	0.000	19:41	
Control Test	0.075	19:41	
Air Blank	0.000	19:42	
Subject Sample	# 1 0.236	19:42	
Breath Volume	2.195	19:42	
Air Blank	0.000	19:43	
Air Blank	0.000	19:45	
Subject Sample	# 2 0.240	19:45	
Breath Volume	2.425	19:45	
Air Blank	0.000	19:46	
Control Test	0.071	19:46	Control Outside Tolerance
Air Blank	0.000	19:47	

Data Download Date/Time: 02/23/2023 04:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PENA,U
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 02/12/2023 23:42:41

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.000	23:42	Control Outside Tolerance
Air Blank	0.000	23:42	

Data Download Date/Time: 02/23/2023 04:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PENA,U
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 02/12/2023 23:46:27

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:45	
Control Test	0.000	23:45	Control Outside Tolerance
Air Blank	0.000	23:46	

Data Download Date/Time: 02/23/2023 04:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PENA,U
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 02/13/2023 00:00:58

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.048	23:56
Breath Volume	3.132	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.046	23:59
Breath Volume	3.394	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 02/23/2023 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROSASCO,PIERO
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 02/16/2023 22:58:11

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.000	22:53
Breath Volume	4.652	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.000	22:56
Breath Volume	4.328	22:56
Air Blank	0.000	22:57
Control Test	0.078	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 02/23/2023 04:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: POZADA,ANDREW
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: PICKETT

DATE/TIME OF TEST: 02/17/2023 03:50:47

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.215	03:46
Breath Volume	2.976	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.212	03:49
Breath Volume	2.855	03:49
Air Blank	0.000	03:49
Control Test	0.077	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 02/23/2023 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VANROMPA,VITO
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: BALBINO FIGUEIRA

DATE/TIME OF TEST: 02/18/2023 01:18:02

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.119	01:13
Breath Volume	1.472	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample # 2	0.112	01:16
Breath Volume	2.339	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:18

Data Download Date/Time: 02/23/2023 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RAMOS,MARK
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: KINGSLAND

DATE/TIME OF TEST: 02/18/2023 05:46:53

Diagnostics Check	OK	05:40
Air Blank	0.000	05:41
Control Test	0.080	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.000	05:42
Breath Volume	3.070	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample # 2	0.000	05:45
Breath Volume	2.527	05:45
Air Blank	0.000	05:45
Control Test	0.079	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 02/23/2023 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANTIAGO,LUMANS
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: STEFAN

DATE/TIME OF TEST: 02/19/2023 00:40:20

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.278	00:35
Breath Volume	2.617	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.269	00:38
Breath Volume	3.289	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 02/23/2023 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RAINSFORD,BRITTANY
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: DASILVA

DATE/TIME OF TEST: 02/19/2023 01:40:10

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.174	01:35
Breath Volume	1.996	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:37
Subject Sample # 2	0.174	01:38
Breath Volume	2.925	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 02/23/2023 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RODRIGUEZ,EMMANUEL
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: GLASS

DATE/TIME OF TEST: 02/19/2023 01:54:34

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.171	01:49
Breath Volume	2.210	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.167	01:53
Breath Volume	3.058	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 02/23/2023 04:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RILEY,B
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: AREVALO

DATE/TIME OF TEST: 02/21/2023 02:52:50

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.238	02:48
Breath Volume	2.257	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.230	02:51
Breath Volume	1.921	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 02/23/2023 04:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: WILLIAM

DATE/TIME OF TEST: 02/21/2023 23:52:28

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:49	
Control Test	0.079	23:49	
Air Blank	0.000	23:50	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 02/23/2023 04:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: WILLIAM

DATE/TIME OF TEST: 02/22/2023 00:39:49

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.178	00:34
Breath Volume	1.414	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.160	00:38
Breath Volume	1.199	00:38
Air Blank	0.000	00:38
Control Test	0.079	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 02/23/2023 04:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DIAL,V
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: DEVERNEY

DATE/TIME OF TEST: 02/22/2023 02:24:28

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.113	02:19
Breath Volume	1.601	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.098	02:22
Breath Volume	1.199	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 02/23/2023 04:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BRYANT,STEVEN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ALDRICH

DATE/TIME OF TEST: 02/24/2023 01:48:15

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.217	01:43
Breath Volume	3.945	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.212	01:46
Breath Volume	4.035	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 02/28/2023 01:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLAZINA,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BECKSTROM

DATE/TIME OF TEST: 02/26/2023 01:01:21

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.270	00:56
Breath Volume	4.082	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.276	00:59
Breath Volume	4.121	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 02/28/2023 01:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RILEY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: NOBOA

DATE/TIME OF TEST: 02/26/2023 03:43:07

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.179	03:38
Breath Volume	3.417	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.181	03:41
Breath Volume	3.460	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 02/28/2023 01:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: CREEMER

DATE/TIME OF TEST: 02/26/2023 04:00:47

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.000	03:56
Breath Volume	3.097	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.000	03:59
Breath Volume	2.875	03:59
Air Blank	0.000	03:59
Control Test	0.078	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 02/28/2023 01:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS, ADAM, J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MCGAVERN, S
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 02/26/2023 07:14:12

Diagnostics Check	OK	07:07
Air Blank	0.000	07:07
Control Test	0.079	07:08
Air Blank	0.000	07:08
Subject Sample # 1	0.122	07:09
Breath Volume	3.855	07:09
Air Blank	0.000	07:10
Air Blank	0.000	07:12
Subject Sample # 2	0.111	07:12
Breath Volume	2.867	07:12
Air Blank	0.000	07:13
Control Test	0.078	07:13
Air Blank	0.000	07:14
Diagnostics Check	OK	07:14

Data Download Date/Time: 02/28/2023 01:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOHRI,BRENDAN,R
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LOHRI,BRENDAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: LAPPAS

DATE/TIME OF TEST: 02/28/2023 18:43:27

Diagnostics Check	OK	18:37
Air Blank	0.000	18:37
Control Test	0.080	18:37
Air Blank	0.000	18:38
Subject Sample # 1	0.158	18:38
Breath Volume	2.972	18:38
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.165	18:41
Breath Volume	2.753	18:41
Air Blank	0.000	18:42
Control Test	0.080	18:42
Air Blank	0.000	18:43
Diagnostics Check	OK	18:43

Data Download Date/Time: 03/14/2023 03:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HALL,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 02/28/2023 21:38:50

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.211	21:33
Breath Volume	2.535	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.209	21:37
Breath Volume	2.500	21:37
Air Blank	0.000	21:37
Control Test	0.079	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 03/14/2023 03:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRITT,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BADILLO

DATE/TIME OF TEST: 02/04/2023 04:12:01

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.079	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 1	04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:11	
Control Test	0.080	04:11	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KING,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 02/04/2023 04:33:36

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:31	
Control Test	0.080	04:31	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Control Test	0.079	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 02/05/2023 04:45:04

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.080	04:42	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:44	
Control Test	0.079	04:44	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:45	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIDY,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: GLEIS

DATE/TIME OF TEST: 02/05/2023 19:45:20

Diagnostics Check	OK	19:33	
Air Blank	0.000	19:34	
Control Test	0.079	19:34	
Air Blank	0.000	19:35	
Subject Sample	# 1	19:38	No Sample Provided
Breath Volume	0.000	19:38	
Air Blank	0.000	19:38	
Air Blank	0.000	19:40	
Subject Sample	# 2	19:43	No Sample Provided
Breath Volume	0.000	19:43	
Air Blank	0.000	19:44	
Control Test	0.079	19:44	
Air Blank	0.000	19:45	
Diagnostics Check	OK	19:45	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIS,DEPUTY B.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/05/2023 21:58:22

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.079	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.171	21:53
Breath Volume	2.562	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.164	21:56
Breath Volume	1.683	21:56
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: DORSO

DATE/TIME OF TEST: 02/06/2023 04:35:39

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:32	
Control Test	0.079	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:34	Subject Test Refused
Breath Volume	0.000	04:34	
Air Blank	0.000	04:34	
Control Test	0.079	04:35	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGANARO,JESSICA,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MANGANARO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: CATARELLI

DATE/TIME OF TEST: 02/06/2023 17:54:45

Diagnostics Check	OK	17:47
Air Blank	0.000	17:47
Control Test	0.079	17:48
Air Blank	0.000	17:48
Subject Sample # 1	0.235	17:49
Breath Volume	2.332	17:49
Air Blank	0.000	17:50
Air Blank	0.000	17:52
Subject Sample # 2	0.233	17:52
Breath Volume	2.636	17:52
Air Blank	0.000	17:53
Control Test	0.078	17:54
Air Blank	0.000	17:54
Diagnostics Check	OK	17:54

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSHANS,DEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HERY,DIEDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: PETALAS

DATE/TIME OF TEST: 02/07/2023 03:29:35

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.135	03:24
Breath Volume	2.484	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.144	03:27
Breath Volume	3.015	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: PHIVILAY

DATE/TIME OF TEST: 02/07/2023 19:52:29

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.079	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.000	19:48
Breath Volume	1.316	19:48
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.000	19:51
Breath Volume	1.207	19:51
Air Blank	0.000	19:51
Control Test	0.080	19:51
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOFMAN,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: SPEAKE

DATE/TIME OF TEST: 02/10/2023 21:08:40

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:06	
Control Test	0.079	21:06	
Air Blank	0.000	21:07	
Subject Sample	# 1	21:07	Subject Test Refused
Breath Volume	0.000	21:07	
Air Blank	0.000	21:07	
Control Test	0.078	21:08	
Air Blank	0.000	21:08	
Diagnostics Check	OK	21:08	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: NOLTON

DATE/TIME OF TEST: 02/10/2023 22:42:28

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.079	22:40	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:41	
Control Test	0.080	22:41	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: HANLEY

DATE/TIME OF TEST: 02/10/2023 23:34:54

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.000	23:30
Breath Volume	1.492	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.000	23:33
Breath Volume	1.582	23:33
Air Blank	0.000	23:33
Control Test	0.079	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 02/11/2023 01:11:38

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:09	
Control Test	0.079	01:09	
Air Blank	0.000	01:09	
Subject Sample	# 1	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Control Test	0.079	01:11	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: SKUKALEK

DATE/TIME OF TEST: 02/12/2023 04:56:29

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.034	04:51
Breath Volume	2.992	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.032	04:54
Breath Volume	2.859	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HERY,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: LIAN

DATE/TIME OF TEST: 02/12/2023 05:09:56

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.230	05:04
Breath Volume	1.460	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:06
Subject Sample # 2	0.232	05:07
Breath Volume	1.468	05:07
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 02/12/2023 23:43:26

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.079	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.022	23:38
Breath Volume	1.617	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.021	23:41
Breath Volume	1.558	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: GRIFFIS

DATE/TIME OF TEST: 02/13/2023 02:24:42

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.174	02:19
Breath Volume	2.429	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.170	02:22
Breath Volume	1.742	02:22
Air Blank	0.000	02:23
Control Test	0.078	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: MACKER

DATE/TIME OF TEST: 02/13/2023 02:52:49

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.078	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.078	02:52	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SARKHOSH,AZEEN
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 02/13/2023 04:49:33

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.079	04:36	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:40	No Sample Provided
Breath Volume	0.000	04:40	
Air Blank	0.000	04:41	
Air Blank	0.000	04:43	
Subject Sample	# 2	0.169	04:44
Breath Volume	1.492	04:44	
Air Blank	0.000	04:45	
Air Blank	0.000	04:46	
Subject Sample	# 3	0.155	04:47
Breath Volume	1.332	04:47	
Air Blank	0.000	04:48	
Control Test	0.079	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 02/13/2023 23:57:19

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.079	23:54	
Air Blank	0.000	23:55	
Subject Sample	# 1	23:55	Subject Test Refused
Breath Volume	0.000	23:55	
Air Blank	0.000	23:56	
Control Test	0.079	23:56	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:57	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP G,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: SOKOL

DATE/TIME OF TEST: 02/14/2023 05:33:01

Diagnostics Check	OK	05:25
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:26
Subject Sample # 1	0.000	05:28
Breath Volume	1.492	05:28
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.000	05:31
Breath Volume	1.816	05:31
Air Blank	0.000	05:32
Control Test	0.079	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:33

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKEIRIK,DEPUTY D.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: KISSINGER

DATE/TIME OF TEST: 02/14/2023 20:17:12

Diagnostics Check	OK	20:10
Air Blank	0.000	20:10
Control Test	0.079	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.229	20:12
Breath Volume	2.710	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:14
Subject Sample # 2	0.225	20:15
Breath Volume	2.312	20:15
Air Blank	0.000	20:16
Control Test	0.078	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIDY,DEP,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SEVOSTJANOV,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: GAUD SR

DATE/TIME OF TEST: 02/15/2023 02:14:01

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.079	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:07	No Sample Provided
Breath Volume	0.000	02:07	
Air Blank	0.000	02:07	
Air Blank	0.000	02:09	
Subject Sample	# 2	02:12	No Sample Provided
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Control Test	0.079	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIBLEY, TRAVIS, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORGAN, WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 13:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:52

SUBJECT'S LAST NAME: BURRELL

DATE/TIME OF TEST: 02/16/2023 14:25:25

Diagnostics Check	OK	14:13	
Air Blank	0.000	14:13	
Control Test	0.079	14:14	
Air Blank	0.000	14:14	
Subject Sample	# 1 0.173	14:17	Volume Not Met
Breath Volume	0.082	14:17	
Air Blank	0.000	14:18	
Air Blank	0.000	14:20	
Subject Sample	# 2 0.163	14:23	Volume Not Met
Breath Volume	1.066	14:23	
Air Blank	0.000	14:24	
Control Test	0.076	14:24	
Air Blank	0.000	14:25	
Diagnostics Check	OK	14:25	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,K,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 02/17/2023 23:42:05

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.133	23:37
Breath Volume	2.292	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:39
Subject Sample # 2	0.133	23:40
Breath Volume	3.007	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REVILLA,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/18/2023 02:31:11

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
Control Test	0.079	02:18	
Air Blank	0.000	02:19	
Subject Sample	# 1 0.151	02:22	Volume Not Met
Breath Volume	0.207	02:22	
Air Blank	0.000	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 2 0.153	02:25	
Breath Volume	1.417	02:25	
Air Blank	0.000	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 3 0.140	02:29	
Breath Volume	1.691	02:29	
Air Blank	0.000	02:30	
Control Test	0.078	02:30	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: DU HART

DATE/TIME OF TEST: 02/18/2023 04:58:20

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.079	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.177	04:52
Breath Volume	1.371	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:54
Subject Sample # 2	0.171	04:56
Breath Volume	1.390	04:56
Air Blank	0.000	04:57
Control Test	0.077	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 09:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:31

SUBJECT'S LAST NAME: RAMEY

DATE/TIME OF TEST: 02/18/2023 09:58:13

Diagnostics Check	OK	09:51
Air Blank	0.000	09:52
Control Test	0.079	09:52
Air Blank	0.000	09:53
Subject Sample # 1	0.000	09:53
Breath Volume	1.753	09:53
Air Blank	0.000	09:54
Air Blank	0.000	09:56
Subject Sample # 2	0.000	09:56
Breath Volume	1.691	09:56
Air Blank	0.000	09:57
Control Test	0.080	09:57
Air Blank	0.000	09:58
Diagnostics Check	OK	09:58

Data Download Date/Time: 02/23/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: LINGERFELT

DATE/TIME OF TEST: 02/04/2023 04:58:33

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:52
Subject Sample # 1	0.000	04:53
Breath Volume	2.292	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.000	04:57
Breath Volume	2.281	04:57
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 02/23/2023 09:28