

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 02/09/2023 04:09:31

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.190	04:03
Breath Volume	2.308	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:05
Subject Sample # 2	0.185	04:07
Breath Volume	1.851	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 03/13/2023 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GAPPA

DATE/TIME OF TEST: 02/15/2023 04:10:57

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.081	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Control Test	0.080	04:10	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 03/13/2023 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYES JR,DARRON,O
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: HUDDELSON,CHASE
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BAESSLER

DATE/TIME OF TEST: 02/16/2023 23:07:02

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.080	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.328	23:02
Breath Volume	1.601	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:04
Subject Sample # 2	0.326	23:05
Breath Volume	1.781	23:05
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:07

Data Download Date/Time: 03/13/2023 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERYCKEN,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/18/2023 04:50:03

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:40	
Control Test	0.080	04:41	
Air Blank	0.000	04:41	
Subject Sample	# 1 0.281	04:42	
Breath Volume	1.332	04:42	
Air Blank	0.000	04:43	
Air Blank	0.000	04:45	
Subject Sample	# 2	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:49	
Control Test	0.078	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 03/13/2023 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: FERNANDEZ PERDOMO

DATE/TIME OF TEST: 02/19/2023 04:03:33

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.192	03:58
Breath Volume	1.894	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.193	04:01
Breath Volume	1.515	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 03/13/2023 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: HERNANDEZ LOPEZ

DATE/TIME OF TEST: 02/20/2023 03:51:57

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.078	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.151	03:47
Breath Volume	1.683	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.140	03:50
Breath Volume	1.179	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 03/13/2023 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,SHELBY,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GARDNER,SHELBY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/20/2023 20:22:04

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.081	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.000	20:17
Breath Volume	2.410	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.000	20:20
Breath Volume	1.957	20:20
Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:22

Data Download Date/Time: 03/13/2023 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TINFINA,LOGAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TINFINA,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: STEMPEL

DATE/TIME OF TEST: 02/21/2023 03:01:04

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.114	02:56
Breath Volume	1.519	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.122	02:59
Breath Volume	1.226	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:01

Data Download Date/Time: 03/13/2023 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: DE LEON-DIAZ

DATE/TIME OF TEST: 02/23/2023 00:35:43

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.165	00:30
Breath Volume	1.214	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.166	00:34
Breath Volume	1.777	00:34
Air Blank	0.000	00:34
Control Test	0.079	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 03/13/2023 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,SHELBY,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GARDNER,SHELBY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/26/2023 03:45:43

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.136	03:41
Breath Volume	1.218	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.139	03:44
Breath Volume	1.125	03:44
Air Blank	0.000	03:44
Control Test	0.078	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 03/13/2023 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATES,BRIAN,A
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GATES,BRIAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 02/27/2023 02:49:09

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.198	02:44
Breath Volume	2.390	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.209	02:47
Breath Volume	2.250	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 03/13/2023 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 02/28/2023 21:09:29

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:05	
Control Test	0.075	21:06	
Air Blank	0.000	21:06	
Subject Sample	# 1	21:08	Subject Test Refused
Breath Volume	0.000	21:08	
Air Blank	0.000	21:08	
Control Test	0.075	21:08	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

Data Download Date/Time: 03/13/2023 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: BURROUGHS

DATE/TIME OF TEST: 02/02/2023 17:58:06

Diagnostics Check	OK	17:51
Air Blank	0.000	17:52
Control Test	0.082	17:52
Air Blank	0.000	17:53
Subject Sample # 1	0.000	17:53
Breath Volume	2.062	17:53
Air Blank	0.000	17:54
Air Blank	0.000	17:56
Subject Sample # 2	0.000	17:56
Breath Volume	1.503	17:56
Air Blank	0.000	17:57
Control Test	0.082	17:57
Air Blank	0.000	17:58
Diagnostics Check	OK	17:58

Data Download Date/Time: 02/06/2023 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: FAY

DATE/TIME OF TEST: 02/04/2023 16:41:08

Diagnostics Check	OK	16:34	
Air Blank	0.000	16:35	
Control Test	0.081	16:35	
Air Blank	0.000	16:35	
Subject Sample	# 1 0.205	16:36	
Breath Volume	2.628	16:36	
Air Blank	0.000	16:37	
Air Blank	0.000	16:39	
Subject Sample	# 2 0.202	16:39	
Breath Volume	2.191	16:39	
Air Blank		16:40	RFI Detect
Air Blank		16:41	RFI Detect

Data Download Date/Time: 02/06/2023 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: FAY

DATE/TIME OF TEST: 02/04/2023 16:53:03

Diagnostics Check	OK	16:44	
Air Blank	0.000	16:45	
Control Test	0.080	16:45	
Air Blank	0.000	16:46	
Subject Sample	# 1 0.187	16:47	
Breath Volume	4.320	16:47	
Air Blank	0.000	16:48	
Air Blank	0.000	16:49	
Subject Sample	# 2 0.189	16:50	
Breath Volume	1.605	16:50	
Air Blank	0.000	16:51	
Control Test	0.077	16:52	
Air Blank		16:52	RFI Detect
Air Blank	0.000	16:53	

Data Download Date/Time: 02/06/2023 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROESCH,RYAN,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: FAY

DATE/TIME OF TEST: 02/04/2023 17:04:22

Diagnostics Check	OK	16:57
Air Blank	0.000	16:57
Control Test	0.080	16:58
Air Blank	0.000	16:58
Subject Sample # 1	0.198	16:59
Breath Volume	1.816	16:59
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.203	17:02
Breath Volume	1.785	17:02
Air Blank	0.000	17:03
Control Test	0.079	17:03
Air Blank	0.000	17:04
Diagnostics Check	OK	17:04

Data Download Date/Time: 02/06/2023 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: OLVERA HURTADO

DATE/TIME OF TEST: 02/17/2023 22:50:04

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.217	22:42
Breath Volume	1.484	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.211	22:48
Breath Volume	2.281	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:50

Data Download Date/Time: 03/13/2023 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: VASQUEZ-OXLAJ

DATE/TIME OF TEST: 02/18/2023 18:58:47

Diagnostics Check	OK	18:51
Air Blank	0.000	18:51
Control Test	0.081	18:52
Air Blank	0.000	18:52
Subject Sample # 1	0.145	18:53
Breath Volume	2.367	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.143	18:57
Breath Volume	2.503	18:57
Air Blank	0.000	18:57
Control Test	0.079	18:58
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 03/13/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI, MARK, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JABLONSKI, MARK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: RAMIREZ BARRERA

DATE/TIME OF TEST: 02/19/2023 01:16:56

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.216	01:12
Breath Volume	2.996	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:14
Subject Sample # 2	0.221	01:15
Breath Volume	3.332	01:15
Air Blank	0.000	01:16
Control Test	0.077	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 03/13/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 02/21/2023 17:29:30

Diagnostics Check	OK	17:22
Air Blank	0.000	17:22
Control Test	0.080	17:23
Air Blank	0.000	17:23
Subject Sample # 1	0.000	17:24
Breath Volume	2.257	17:24
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 2	0.000	17:28
Breath Volume	2.558	17:28
Air Blank	0.000	17:28
Control Test	0.081	17:28
Air Blank	0.000	17:29
Diagnostics Check	OK	17:29

Data Download Date/Time: 03/13/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCI,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: RICCI,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:42

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 02/24/2023 13:14:29

Diagnostics Check	OK	13:08
Air Blank	0.000	13:08
Control Test	0.081	13:09
Air Blank	0.000	13:09
Subject Sample # 1	0.000	13:10
Breath Volume	2.460	13:10
Air Blank	0.000	13:10
Air Blank	0.000	13:12
Subject Sample # 2	0.000	13:13
Breath Volume	1.910	13:13
Air Blank	0.000	13:13
Control Test	0.079	13:13
Air Blank	0.000	13:14
Diagnostics Check	OK	13:14

Data Download Date/Time: 03/13/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BATEN-GARCIA

DATE/TIME OF TEST: 02/26/2023 02:27:57

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.210	02:23
Breath Volume	2.695	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.204	02:26
Breath Volume	2.308	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 03/13/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,ELIER,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SANTANA,ELIER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:43

SUBJECT'S LAST NAME: GURDIAN GURDIAN

DATE/TIME OF TEST: 02/27/2023 08:14:23

Diagnostics Check	OK	08:04
Air Blank	0.000	08:05
Control Test	0.080	08:05
Air Blank	0.000	08:05
Subject Sample # 1	0.139	08:07
Breath Volume	1.761	08:07
Air Blank	0.000	08:08
Air Blank	0.000	08:10
Subject Sample # 2	0.130	08:12
Breath Volume	2.335	08:12
Air Blank	0.000	08:13
Control Test	0.077	08:13
Air Blank	0.000	08:14
Diagnostics Check	OK	08:14

Data Download Date/Time: 03/13/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: RANDO

DATE/TIME OF TEST: 02/01/2023 01:41:43

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.081	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.151	01:37
Breath Volume	2.847	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.148	01:40
Breath Volume	2.035	01:40
Air Blank	0.000	01:40
Control Test	0.078	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 02/07/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 02/05/2023 04:01:59

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.176	03:54
Breath Volume	2.394	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.154	03:57
Breath Volume	1.925	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 3	0.159	04:00
Breath Volume	2.246	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 02/07/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: EARGOOD

DATE/TIME OF TEST: 02/07/2023 20:45:17

Diagnostics Check	OK	20:35
Air Blank	0.000	20:36
Control Test	0.079	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.065	20:40
Breath Volume	1.335	20:40
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.071	20:43
Breath Volume	1.593	20:43
Air Blank	0.000	20:44
Control Test	0.077	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: COUTURE

DATE/TIME OF TEST: 02/09/2023 01:24:26

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.078	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.156	01:19
Breath Volume	2.953	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:21
Subject Sample # 2	0.151	01:22
Breath Volume	2.632	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: COMM VEHIC
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 02/09/2023 02:20:40

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.105	02:16
Breath Volume	1.539	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.113	02:19
Breath Volume	2.234	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GACHEL

DATE/TIME OF TEST: 02/09/2023 04:26:56

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.168	04:22
Breath Volume	2.035	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.171	04:25
Breath Volume	1.863	04:25
Air Blank	0.000	04:26
Control Test	0.078	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: TAPOLYAI

DATE/TIME OF TEST: 02/10/2023 01:28:58

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.177	01:24
Breath Volume	1.140	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.173	01:27
Breath Volume	1.640	01:27
Air Blank	0.000	01:28
Control Test	0.077	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 02/10/2023 02:01:54

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.078	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.189	01:57
Breath Volume	2.183	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.199	02:00
Breath Volume	2.417	02:00
Air Blank	0.000	02:00
Control Test	0.078	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: RHETT

DATE/TIME OF TEST: 02/11/2023 03:36:54

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:35	
Control Test	0.079	03:36	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: RHETT

DATE/TIME OF TEST: 02/11/2023 03:53:15

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.079	03:42	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:46	No Sample Provided
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Air Blank	0.000	03:48	
Subject Sample	# 2	03:51	No Sample Provided
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.078	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: GOOLSBY III

DATE/TIME OF TEST: 02/11/2023 20:01:20

Diagnostics Check	OK	19:54
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:55
Subject Sample # 1	0.097	19:56
Breath Volume	2.015	19:56
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.099	19:59
Breath Volume	2.160	19:59
Air Blank	0.000	20:00
Control Test	0.079	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: LABARTA

DATE/TIME OF TEST: 02/11/2023 22:15:56

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.078	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.197	22:11
Breath Volume	2.128	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.209	22:14
Breath Volume	1.683	22:14
Air Blank	0.000	22:15
Control Test	0.078	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/11/2023 22:43:45

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.079	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.266	22:39
Breath Volume	3.042	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.265	22:42
Breath Volume	2.398	22:42
Air Blank	0.000	22:42
Control Test	0.078	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PINA PINA

DATE/TIME OF TEST: 02/12/2023 03:42:57

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.154	03:37
Breath Volume	1.511	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.146	03:41
Breath Volume	1.273	03:41
Air Blank	0.000	03:42
Control Test	0.077	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: KOTH

DATE/TIME OF TEST: 02/12/2023 04:03:14

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.168	03:58
Breath Volume	1.343	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.164	04:01
Breath Volume	1.382	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: WADLEY III

DATE/TIME OF TEST: 02/12/2023 04:21:58

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.078	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.198	04:17
Breath Volume	2.570	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:19
Subject Sample # 2	0.198	04:20
Breath Volume	1.566	04:20
Air Blank	0.000	04:21
Control Test	0.079	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 02/12/2023 04:41:44

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.078	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.038	04:36
Breath Volume	1.980	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:38
Subject Sample # 2	0.037	04:40
Breath Volume	1.972	04:40
Air Blank	0.000	04:40
Control Test	0.079	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: IRONS

DATE/TIME OF TEST: 02/13/2023 00:11:54

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.175	00:07
Breath Volume	1.531	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:09
Subject Sample # 2	0.181	00:10
Breath Volume	2.988	00:10
Air Blank	0.000	00:10
Control Test	0.078	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 03/02/2023 09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: LIPRANDO

DATE/TIME OF TEST: 02/13/2023 00:29:19

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.149	00:24
Breath Volume	1.640	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.150	00:27
Breath Volume	1.464	00:27
Air Blank	0.000	00:28
Control Test	0.079	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: GOODFELLOW

DATE/TIME OF TEST: 02/13/2023 05:27:47

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.117	05:23
Breath Volume	1.742	05:23
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.127	05:26
Breath Volume	1.742	05:26
Air Blank	0.000	05:26
Control Test	0.079	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: PALACIO

DATE/TIME OF TEST: 02/13/2023 23:32:41

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.081	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.140	23:28
Breath Volume	1.929	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.147	23:31
Breath Volume	1.335	23:31
Air Blank	0.000	23:31
Control Test	0.080	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 02/14/2023 03:08:24

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.246	03:03
Breath Volume	2.214	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.244	03:06
Breath Volume	1.781	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: TAGGART

DATE/TIME OF TEST: 02/14/2023 05:03:35

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.234	04:58
Breath Volume	3.347	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.239	05:01
Breath Volume	3.449	05:01
Air Blank	0.000	05:02
Control Test	0.079	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: STEPHAN

DATE/TIME OF TEST: 02/15/2023 01:53:50

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.083	01:48
Breath Volume	1.328	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.080	01:52
Breath Volume	1.707	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 02/16/2023 22:52:20

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:40	
Control Test	0.080	22:40	
Air Blank	0.000	22:41	
Subject Sample	# 1 0.261	22:41	
Breath Volume	1.781	22:41	
Air Blank	0.000	22:42	
Air Blank	0.000	22:44	
Subject Sample	# 2 0.224	22:47	Volume Not Met
Breath Volume	0.714	22:47	
Air Blank	0.000	22:48	
Air Blank	0.000	22:49	
Subject Sample	# 3 0.263	22:50	
Breath Volume	1.898	22:50	
Air Blank	0.000	22:51	
Control Test	0.076	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MOCKAPETRIS

DATE/TIME OF TEST: 02/18/2023 01:53:02

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.151	01:48
Breath Volume	1.406	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:50
Subject Sample # 2	0.151	01:51
Breath Volume	1.210	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:53

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 02/18/2023 04:03:36

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.078	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1 0.128	03:56	Volume Not Met
Breath Volume	0.574	03:56	
Air Blank	0.000	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 2 0.128	04:01	Volume Not Met
Breath Volume	0.523	04:01	
Air Blank	0.000	04:02	
Control Test	0.077	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 02/18/2023 04:14:49

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.076	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.124	04:06
Breath Volume	2.433	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.102	04:09
Breath Volume	1.316	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 3	0.125	04:13
Breath Volume	2.511	04:13
Air Blank	0.000	04:13
Control Test	0.078	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: LOPEZ ALAMILLA

DATE/TIME OF TEST: 02/19/2023 02:29:29

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.156	02:24
Breath Volume	1.750	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.164	02:27
Breath Volume	1.160	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: EPPS

DATE/TIME OF TEST: 02/19/2023 05:01:42

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.228	04:56
Breath Volume	1.457	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 2	0.234	05:00
Breath Volume	1.679	05:00
Air Blank	0.000	05:00
Control Test	0.078	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: SIERPA AVILAN

DATE/TIME OF TEST: 02/19/2023 05:17:57

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.079	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.185	05:12
Breath Volume	1.367	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.198	05:16
Breath Volume	1.328	05:16
Air Blank	0.000	05:17
Control Test	0.078	05:17
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK, I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: BORGES VILARINO

DATE/TIME OF TEST: 02/19/2023 05:40:31

Diagnostics Check	OK	05:32
Air Blank	0.000	05:33
Control Test	0.079	05:33
Air Blank	0.000	05:34
Subject Sample # 1	0.179	05:35
Breath Volume	1.531	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.189	05:38
Breath Volume	1.675	05:38
Air Blank	0.000	05:39
Control Test	0.078	05:39
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: BOULES

DATE/TIME OF TEST: 02/22/2023 01:33:47

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.110	01:28
Breath Volume	2.574	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.117	01:32
Breath Volume	1.355	01:32
Air Blank	0.000	01:32
Control Test	0.079	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: SIMON LOPEZ

DATE/TIME OF TEST: 02/25/2023 02:08:44

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.079	02:03
Breath Volume	1.300	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.079	02:07
Breath Volume	2.019	02:07
Air Blank	0.000	02:07
Control Test	0.078	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: BAHAM III

DATE/TIME OF TEST: 02/25/2023 05:43:46

Diagnostics Check	OK	05:36
Air Blank	0.000	05:36
Control Test	0.079	05:37
Air Blank	0.000	05:37
Subject Sample # 1	0.240	05:38
Breath Volume	2.203	05:38
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample # 2	0.257	05:42
Breath Volume	2.562	05:42
Air Blank	0.000	05:42
Control Test	0.078	05:43
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: GAUDETTE

DATE/TIME OF TEST: 02/26/2023 00:55:33

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.161	00:48
Breath Volume	1.214	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.156	00:53
Breath Volume	1.101	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: LAGAR HERNANDEZ

DATE/TIME OF TEST: 02/26/2023 01:17:43

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.176	01:12
Breath Volume	1.734	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.176	01:16
Breath Volume	1.335	01:16
Air Blank	0.000	01:16
Control Test	0.077	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: GILLESPIE

DATE/TIME OF TEST: 02/26/2023 05:17:21

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.072	05:10
Breath Volume	1.531	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:12
Subject Sample # 2	0.061	05:15
Breath Volume	1.320	05:15
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/27/2023 02:16:03

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.167	02:11
Breath Volume	3.066	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample # 2	0.166	02:14
Breath Volume	1.789	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:16

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: TAMARGO

DATE/TIME OF TEST: 02/28/2023 03:17:21

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.185	03:07
Breath Volume	1.253	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.214	03:12
Breath Volume	2.265	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 3	0.197	03:15
Breath Volume	1.511	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 03/02/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: COX,GABRIEL
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 01/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: GORAKHNAUTH

DATE/TIME OF TEST: 02/02/2023 01:19:37

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.202	01:14
Breath Volume	1.449	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.196	01:17
Breath Volume	1.988	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 02/21/2023 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KIANA,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 01/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: STRANTE MAGNUSSON

DATE/TIME OF TEST: 02/04/2023 23:05:55

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.080	22:59
Air Blank	0.000	23:00
Subject Sample # 1	0.253	23:01
Breath Volume	2.542	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:03
Subject Sample # 2	0.244	23:04
Breath Volume	2.789	23:04
Air Blank	0.000	23:04
Control Test	0.078	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 02/21/2023 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNE,ASHTON,D
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 01/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/05/2023 02:52:46

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.080	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.080	02:52	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 02/21/2023 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BUGALLO,GABRIEL
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 01/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/10/2023 01:09:29

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.000	01:04
Breath Volume	1.554	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.000	01:08
Breath Volume	3.191	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 02/21/2023 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA RUA,BRYAN,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 01/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/13/2023 02:53:19

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.109	02:48
Breath Volume	3.929	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.105	02:51
Breath Volume	2.027	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 02/21/2023 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA RUA,BRYAN,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 01/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ABECASIS

DATE/TIME OF TEST: 02/15/2023 03:03:29

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.123	02:58
Breath Volume	4.039	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.121	03:01
Breath Volume	4.097	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 02/21/2023 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: STEPHENSON,MATTHEW
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 01/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: NIEVES SANTIAGO

DATE/TIME OF TEST: 02/17/2023 01:11:06

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.183	01:05
Breath Volume	3.453	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:07
Subject Sample # 2	0.179	01:09
Breath Volume	3.644	01:09
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 02/21/2023 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 02/21/2023 02:07:05

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.155	02:02
Breath Volume	1.414	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.145	02:05
Breath Volume	1.113	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 02/21/2023 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KIANA,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MCCOSH JR

DATE/TIME OF TEST: 02/22/2023 02:24:18

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.079	02:22	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.079	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 03/15/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: FELIX,REYSON
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: GINTER

DATE/TIME OF TEST: 02/24/2023 16:06:44

Diagnostics Check	OK	15:59
Air Blank	0.000	15:59
Control Test	0.079	16:00
Air Blank	0.000	16:00
Subject Sample # 1	0.248	16:01
Breath Volume	1.546	16:01
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.254	16:04
Breath Volume	2.898	16:04
Air Blank	0.000	16:05
Control Test	0.079	16:06
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

Data Download Date/Time: 03/15/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BAUMGARTEN,CHAYA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: ORSINI

DATE/TIME OF TEST: 02/25/2023 00:38:43

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.079	00:31	
Air Blank	0.000	00:31	
Subject Sample	# 1 0.048	00:34	
Breath Volume	1.511	00:34	
Air Blank	0.000	00:35	
Air Blank	0.000	00:37	
Subject Sample	# 2	00:37	RFI Detect
Breath Volume	0.000	00:37	
Air Blank	0.000	00:37	
Control Test	0.078	00:38	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 03/15/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BAUMGARTEN,CHAYA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: ORSINI

DATE/TIME OF TEST: 02/25/2023 00:53:25

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.079	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1 0.035	00:46	Volume Not Met
Breath Volume	0.398	00:46	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 2 0.029	00:51	Volume Not Met
Breath Volume	0.152	00:51	
Air Blank	0.000	00:52	
Control Test	0.078	00:52	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 03/15/2023 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,HALEY,M
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,D
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: BRONSTEIN

DATE/TIME OF TEST: 02/25/2023 01:05:38

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:03	
Control Test	0.079	01:03	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:04	
Control Test	0.079	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 03/15/2023 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BAUMGARTEN,CHAYA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: ORSINI

DATE/TIME OF TEST: 02/25/2023 01:43:52

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.079	01:36	
Air Blank	0.000	01:36	
Subject Sample	# 1 0.019	01:39	Volume Not Met
Breath Volume	0.332	01:39	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 2	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.079	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 03/15/2023 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: CHAFIN BARROSO

DATE/TIME OF TEST: 02/27/2023 03:31:08

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.079	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 03/15/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KIANA,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: HEDRICK

DATE/TIME OF TEST: 02/28/2023 03:11:59

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.238	03:07
Breath Volume	3.789	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample # 2	0.240	03:10
Breath Volume	3.386	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 03/15/2023 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ARIEL,N
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: HERNANDEDZ,ARIEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: VILLGA ZAYAS

DATE/TIME OF TEST: 02/04/2023 16:51:21

Diagnostics Check	OK	16:45	
Air Blank	0.000	16:46	
Control Test	0.082	16:46	
Air Blank	0.000	16:47	
Subject Sample	# 1	16:49	Subject Test Refused
Breath Volume	0.000	16:49	
Air Blank	0.000	16:50	
Control Test	0.082	16:50	
Air Blank	0.000	16:51	
Diagnostics Check	OK	16:51	

Data Download Date/Time: 04/06/2023 04:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIDRON,JULIO,C
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CIDRON,JULIO
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MORENO DORTA

DATE/TIME OF TEST: 02/14/2023 02:29:22

Diagnostics Check	OK	02:28	
Air Blank		02:28	Ambient Fail
Air Blank	0.000	02:29	

Data Download Date/Time: 04/06/2023 04:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIDRON,JULIO,C
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CIDRON,JULIO
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MORENO DORTA

DATE/TIME OF TEST: 02/14/2023 02:41:29

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.146	02:34
Breath Volume	1.402	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.135	02:39
Breath Volume	1.156	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 04/06/2023 04:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 07:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GONZALEZ DOMINGUEZ

DATE/TIME OF TEST: 02/21/2023 20:37:38

Diagnostics Check	OK	20:31
Air Blank	0.000	20:32
Control Test	0.082	20:32
Air Blank	0.000	20:33
Subject Sample # 1	0.000	20:33
Breath Volume	2.523	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.000	20:36
Breath Volume	2.691	20:36
Air Blank	0.000	20:36
Control Test	0.082	20:37
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 04/06/2023 04:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ-BONILLA, RONALD,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MELENDEZ-BONILLA, RONALD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: MARIN GARCIA

DATE/TIME OF TEST: 02/24/2023 06:01:37

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.082	05:54
Air Blank	0.000	05:55
Subject Sample # 1	0.100	05:56
Breath Volume	2.359	05:56
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.109	06:00
Breath Volume	2.800	06:00
Air Blank	0.000	06:00
Control Test	0.082	06:01
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

Data Download Date/Time: 04/06/2023 04:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WROY,WALNER,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: WROY,WALNER
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: GARAY

DATE/TIME OF TEST: 02/24/2023 21:42:02

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.082	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.173	21:37
Breath Volume	1.519	21:37
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.173	21:40
Breath Volume	3.175	21:40
Air Blank	0.000	21:41
Control Test	0.078	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:42

Data Download Date/Time: 04/06/2023 04:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEALEGRE, YESSICA, A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DELGADO, GIANCARLO
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MONTEALEGRE

DATE/TIME OF TEST: 02/24/2023 23:18:57

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:06	
Control Test	0.082	23:06	
Air Blank	0.000	23:07	
Subject Sample	# 1 0.185	23:10	Slope Not Level
Breath Volume	1.226	23:10	
Air Blank	0.000	23:11	
Air Blank	0.000	23:12	
Subject Sample	# 2 0.204	23:13	
Breath Volume	2.066	23:13	
Air Blank	0.000	23:14	
Air Blank	0.000	23:15	
Subject Sample	# 3 0.194	23:17	
Breath Volume	2.433	23:17	
Air Blank	0.000	23:18	
Control Test	0.079	23:18	
Air Blank	0.000	23:18	
Diagnostics Check	OK	23:18	

Data Download Date/Time: 04/06/2023 04:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402231696
BREATH TEST OPERATOR: RIVERA,VICTOR,J	DRY GAS EXPIRY DATE: 10/07/2024
BREATH TEST OPERATOR AGENCY: FHP Troop D	VIOLATION CODE: DUI
ARRESTING OFFICER: RIVERA,VICTOR	ARREST TIME: 00:12
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Male
LAST AGENCY INSPECTION: 02/18/2023	OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: RICARDO	DATE/TIME OF TEST: 02/26/2023 00:48:10
------------------------------	--

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.081	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.178	00:43
Breath Volume	1.839	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.176	00:46
Breath Volume	1.335	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 03/16/2023 07:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 02/03/2023 01:32:46

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.162	01:24
Breath Volume	1.964	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.138	01:28
Breath Volume	3.343	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 3	0.142	01:31
Breath Volume	4.046	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 03/02/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,KRISTOPHER,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HALL,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 02/11/2023 03:30:25

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.078	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:22	No Sample Provided
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Air Blank	0.000	03:24	
Subject Sample	# 2	0.201	03:25
Breath Volume	2.828	03:25	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample	# 3	0.190	03:28
Breath Volume	2.343	03:28	
Air Blank	0.000	03:29	
Control Test	0.077	03:29	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 03/02/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MUKTAR

DATE/TIME OF TEST: 02/17/2023 03:44:56

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.078	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.133	03:40
Breath Volume	4.156	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.126	03:43
Breath Volume	4.421	03:43
Air Blank	0.000	03:44
Control Test	0.077	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 03/02/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: SWEET

DATE/TIME OF TEST: 02/07/2023 20:42:01

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:39	
Control Test	0.079	20:39	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:40	Subject Test Refused
Breath Volume	0.000	20:40	
Air Blank	0.000	20:41	
Control Test	0.078	20:41	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:42	

Data Download Date/Time: 03/13/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/08/2023 23:27:45

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.079	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.219	23:23
Breath Volume	1.933	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.218	23:26
Breath Volume	1.382	23:26
Air Blank	0.000	23:26
Control Test	0.078	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 03/13/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: AGUILERA

DATE/TIME OF TEST: 02/08/2023 23:36:41

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.078	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.397	23:32
Breath Volume	1.562	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.393	23:34
Breath Volume	1.390	23:34
Air Blank	0.000	23:35
Control Test	0.078	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 03/13/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRIGNANO,JESSICA,L
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: SIRIGNANO,JESSICA
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: VIPARINA

DATE/TIME OF TEST: 02/09/2023 04:03:40

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.245	03:57
Breath Volume	1.503	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.247	04:01
Breath Volume	2.433	04:01
Air Blank	0.000	04:02
Control Test	0.075	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 03/13/2023 14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,TY,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WILSON,MATTHEW
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: CATALAM

DATE/TIME OF TEST: 02/11/2023 04:04:55

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.078	04:03	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:04	Sequence Aborted
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	

Data Download Date/Time: 03/13/2023 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,TY,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WILSON,MATTHEW
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: CATALAM

DATE/TIME OF TEST: 02/11/2023 04:13:57

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.101	04:09
Breath Volume	1.882	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:11
Subject Sample # 2	0.107	04:12
Breath Volume	1.664	04:12
Air Blank	0.000	04:13
Control Test	0.076	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 03/13/2023 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SAVAGE

DATE/TIME OF TEST: 02/12/2023 03:58:49

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.121	03:54
Breath Volume	1.195	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.122	03:57
Breath Volume	1.292	03:57
Air Blank	0.000	03:57
Control Test	0.078	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 03/13/2023 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: HERNANDEZ SANTIAGO

DATE/TIME OF TEST: 02/12/2023 08:01:43

Diagnostics Check	OK	07:54
Air Blank	0.000	07:54
Control Test	0.078	07:55
Air Blank	0.000	07:55
Subject Sample # 1	0.017	07:56
Breath Volume	1.234	07:56
Air Blank	0.000	07:57
Air Blank	0.000	07:58
Subject Sample # 2	0.016	08:00
Breath Volume	1.164	08:00
Air Blank	0.000	08:00
Control Test	0.078	08:01
Air Blank	0.000	08:01
Diagnostics Check	OK	08:01

Data Download Date/Time: 03/13/2023 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: JARAMILLO VALEZ

DATE/TIME OF TEST: 02/17/2023 22:13:36

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:11	
Control Test	0.079	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:12	Subject Test Refused
Breath Volume	0.000	22:12	
Air Blank	0.000	22:12	
Control Test	0.079	22:12	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

Data Download Date/Time: 03/13/2023 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERYCKEN,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VERYCKEN,MATTHEW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: REYESGUZMAN

DATE/TIME OF TEST: 02/18/2023 04:41:22

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.079	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.140	04:36
Breath Volume	1.679	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.128	04:39
Breath Volume	1.460	04:39
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 03/13/2023 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRIGNANO,JESSICA,L
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: CASTRO-AGIS,JORGE
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: DELGADO PENA

DATE/TIME OF TEST: 02/19/2023 21:17:34

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.079	21:09
Air Blank	0.000	21:10
Subject Sample # 1	0.208	21:11
Breath Volume	3.652	21:11
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.189	21:15
Breath Volume	3.261	21:15
Air Blank	0.000	21:16
Control Test	0.076	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 03/13/2023 14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 02/21/2023 03:43:53

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.078	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.078	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 03/13/2023 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BALDE

DATE/TIME OF TEST: 02/25/2023 21:42:22

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.079	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.236	21:37
Breath Volume	3.242	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.229	21:40
Breath Volume	2.171	21:40
Air Blank	0.000	21:41
Control Test	0.078	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 03/13/2023 14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: FORMAN

DATE/TIME OF TEST: 02/27/2023 04:53:29

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.078	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.000	04:47
Breath Volume	1.210	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.000	04:52
Breath Volume	1.554	04:52
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 03/13/2023 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: GONZALEZ,CHRISTOPHER,M	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: EARLYWINE,G	ARREST TIME: 03:41
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 01/31/2023	OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: ADAMES	DATE/TIME OF TEST: 02/25/2023 04:26:39
-----------------------------	--

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.145	04:21
Breath Volume	2.085	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.148	04:24
Breath Volume	2.332	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 02/28/2023 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: DENNY

DATE/TIME OF TEST: 02/02/2023 20:08:11

Diagnostics Check	OK	20:01
Air Blank	0.000	20:02
Control Test	0.081	20:02
Air Blank	0.000	20:02
Subject Sample # 1	0.188	20:03
Breath Volume	2.941	20:03
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.183	20:06
Breath Volume	2.000	20:06
Air Blank	0.000	20:07
Control Test	0.081	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 02/06/2023 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 02/03/2023 20:16:14

Diagnostics Check	OK	20:09
Air Blank	0.000	20:10
Control Test	0.081	20:10
Air Blank	0.000	20:11
Subject Sample # 1	0.230	20:11
Breath Volume	1.636	20:11
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.229	20:14
Breath Volume	1.601	20:14
Air Blank	0.000	20:15
Control Test	0.080	20:15
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 02/06/2023 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 02/05/2023 01:32:12

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.082	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.169	01:27
Breath Volume	3.425	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.170	01:30
Breath Volume	1.492	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 02/06/2023 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PAQUIN

DATE/TIME OF TEST: 02/15/2023 00:25:06

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.082	00:19	
Air Blank	0.000	00:20	
Subject Sample # 1	0.076	00:21	
Breath Volume	2.546	00:21	
Air Blank	0.000	00:22	
Air Blank		00:23	RFI Detect
Air Blank	0.000	00:24	
Control Test	0.082	00:24	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 03/13/2023 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PAQUIN

DATE/TIME OF TEST: 02/15/2023 00:34:21

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.081	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.076	00:29
Breath Volume	2.492	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.073	00:32
Breath Volume	2.117	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 03/13/2023 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 02/18/2023 00:58:23

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.082	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.024	00:53
Breath Volume	2.687	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.021	00:56
Breath Volume	4.062	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 03/13/2023 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: KREILING

DATE/TIME OF TEST: 02/23/2023 00:21:04

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.160	00:16
Breath Volume	1.957	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:18
Subject Sample # 2	0.156	00:19
Breath Volume	3.179	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:21

Data Download Date/Time: 03/13/2023 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARARD,WILLIAM,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: GARARD,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402231696
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: WAITERS

DATE/TIME OF TEST: 02/26/2023 04:22:41

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.082	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.147	04:17
Breath Volume	1.710	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.138	04:21
Breath Volume	1.406	04:21
Air Blank	0.000	04:21
Control Test	0.082	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 03/16/2023 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DALTON, CHEYNE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 02/03/2023 01:06:10

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.139	01:01
Breath Volume	1.582	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:03
Subject Sample # 2	0.139	01:04
Breath Volume	1.566	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 02/09/2023 03:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: SEAY

DATE/TIME OF TEST: 02/20/2023 22:02:51

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.120	21:58
Breath Volume	2.390	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.119	22:01
Breath Volume	2.148	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 02/28/2023 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: ADAMES

DATE/TIME OF TEST: 02/25/2023 04:08:16

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.081	04:02	
Air Blank	0.000	04:02	
Subject Sample	# 1 0.153	04:03	
Breath Volume	3.605	04:03	
Air Blank	0.000	04:04	
Air Blank	0.000	04:05	
Subject Sample	# 2	04:06	Interferent Detect
Breath Volume	2.328	04:06	
Air Blank	0.000	04:07	
Control Test	0.079	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 02/28/2023 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 01/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 02/09/2023 03:50:47

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:48	
Control Test	0.079	03:48	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:49	Subject Test Refused
Breath Volume	0.000	03:49	
Air Blank	0.000	03:49	
Control Test	0.079	03:50	
Air Blank	0.000	03:50	
Diagnostics Check	OK	03:50	

Data Download Date/Time: 02/14/2023 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: STAPLETON

DATE/TIME OF TEST: 02/02/2023 00:14:35

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:12	
Control Test	0.080	00:12	
Air Blank	0.000	00:13	
Subject Sample	# 1	00:13	Subject Test Refused
Breath Volume	0.000	00:13	
Air Blank	0.000	00:13	
Control Test	0.079	00:13	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 02/06/2023 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: DE LEON MARCOS

DATE/TIME OF TEST: 02/04/2023 05:43:24

Diagnostics Check	OK	05:39	
Air Blank	0.000	05:40	
Control Test	0.081	05:40	
Air Blank	0.000	05:41	
Subject Sample	# 1	05:42	Subject Test Refused
Breath Volume	0.000	05:42	
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 02/06/2023 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ROMERO SORTO

DATE/TIME OF TEST: 02/05/2023 02:48:26

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.161	02:43
Breath Volume	2.863	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.166	02:46
Breath Volume	2.480	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 02/06/2023 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BRUNGER

DATE/TIME OF TEST: 02/06/2023 03:20:03

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:17	
Control Test	0.080	03:17	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 02/06/2023 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 02/13/2023 00:02:55

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:02	
Control Test	0.000	00:02	Control Outside Tolerance
Air Blank	0.000	00:02	

Data Download Date/Time: 03/13/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 02/13/2023 00:12:45

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.145	00:07
Breath Volume	1.394	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.138	00:11
Breath Volume	1.746	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 03/13/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: CARREON DOMINGUEZ

DATE/TIME OF TEST: 02/25/2023 01:36:22

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.026	01:31
Breath Volume	1.781	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.024	01:34
Breath Volume	2.343	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 03/13/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ZETINO LOPEZ

DATE/TIME OF TEST: 02/26/2023 02:06:51

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.080	02:00	
Air Blank	0.000	02:01	
Subject Sample # 1	0.233	02:02	
Breath Volume	2.164	02:02	
Air Blank	0.000	02:02	
Air Blank	0.000	02:04	
Subject Sample # 2	0.234	02:05	
Breath Volume	2.324	02:05	
Air Blank	0.000	02:05	
Control Test	0.000	02:06	Control Outside Tolerance
Air Blank	0.000	02:06	

Data Download Date/Time: 03/13/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ZETINO LOPEZ

DATE/TIME OF TEST: 02/26/2023 02:15:04

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.230	02:10
Breath Volume	1.949	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 2	0.220	02:13
Breath Volume	2.355	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:15

Data Download Date/Time: 03/13/2023 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: MORENO ESPINOZA

DATE/TIME OF TEST: 02/05/2023 20:38:41

Diagnostics Check	OK	20:30
Air Blank	0.000	20:30
Control Test	0.083	20:30
Air Blank	0.000	20:31
Subject Sample # 1	0.250	20:33
Breath Volume	1.750	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.255	20:37
Breath Volume	1.562	20:37
Air Blank	0.000	20:37
Control Test	0.081	20:38
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 02/06/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOYETT,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/24/2023 18:48:58

Diagnostics Check	OK	18:38
Air Blank	0.000	18:39
Control Test	0.078	18:39
Air Blank	0.000	18:40
Subject Sample # 1	0.251	18:42
Breath Volume	1.535	18:42
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.232	18:47
Breath Volume	1.257	18:47
Air Blank	0.000	18:47
Control Test	0.077	18:48
Air Blank	0.000	18:48
Diagnostics Check	OK	18:48

Data Download Date/Time: 03/06/2023 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 01:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: FROESE

DATE/TIME OF TEST: 02/18/2023 01:24:22

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.081	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.103	01:19
Breath Volume	1.132	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.110	01:22
Breath Volume	1.371	01:22
Air Blank	0.000	01:23
Control Test	0.081	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 03/28/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,MATTHEW,M

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: SMITH,MATTHEW

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:21

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BOYLES

DATE/TIME OF TEST: 02/18/2023 04:09:45

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.083	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.178	04:04
Breath Volume	1.937	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.175	04:08
Breath Volume	1.335	04:08
Air Blank	0.000	04:08
Control Test	0.081	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 03/28/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: FARLEY,JOSEPH,E	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: FARLEY,JOSEPH	ARREST TIME: 01:46
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 02/15/2023	OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SCHUUR 2	DATE/TIME OF TEST: 02/19/2023	01:56:44
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.147	01:52
Breath Volume	1.578	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.149	01:55
Breath Volume	1.542	01:55
Air Blank	0.000	01:55
Control Test	0.081	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 03/28/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: FARLEY,JOSEPH,E	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: FARLEY,JOSEPH	ARREST TIME: 00:47
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Female
LAST AGENCY INSPECTION: 02/15/2023	OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: ANDERSON	DATE/TIME OF TEST: 02/24/2023	01:00:58
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.153	00:56
Breath Volume	1.808	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:58
Subject Sample # 2	0.150	00:59
Breath Volume	1.800	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 03/28/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,MATTHEW,M

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: SMITH,MATTHEW

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PRADO

DATE/TIME OF TEST: 02/24/2023 03:52:35

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.126	03:47
Breath Volume	2.515	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.123	03:50
Breath Volume	2.144	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 03/28/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FORMAS,Z
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: COLLADO GUERRERO

DATE/TIME OF TEST: 02/10/2023 01:44:59

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.130	01:40
Breath Volume	2.613	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.129	01:43
Breath Volume	2.703	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 02/16/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: DIAZ MONSERRATE

DATE/TIME OF TEST: 02/11/2023 04:08:36

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.066	04:01
Breath Volume	1.718	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.068	04:07
Breath Volume	1.992	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 02/16/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MAYORGA RIVAS

DATE/TIME OF TEST: 02/18/2023 05:39:00

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.078	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.146	05:34
Breath Volume	2.980	05:34
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.143	05:37
Breath Volume	1.855	05:37
Air Blank	0.000	05:38
Control Test	0.080	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:39

Data Download Date/Time: 03/02/2023 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HOWARD,A
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: EUSEDA

DATE/TIME OF TEST: 02/19/2023 03:55:05

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:54	
Control Test	0.079	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 03/02/2023 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: ROLLERSON

DATE/TIME OF TEST: 02/23/2023 05:17:05

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.221	05:12
Breath Volume	1.621	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:14
Subject Sample # 2	0.224	05:15
Breath Volume	1.863	05:15
Air Blank	0.000	05:16
Control Test	0.078	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:17

Data Download Date/Time: 03/02/2023 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: NADAL,S
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 02/24/2023 04:38:40

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.080	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.152	04:34
Breath Volume	2.523	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.153	04:37
Breath Volume	1.574	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 03/02/2023 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: PEREZ BIGORRA

DATE/TIME OF TEST: 02/26/2023 03:04:43

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.136	03:00
Breath Volume	1.222	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.145	03:03
Breath Volume	1.355	03:03
Air Blank	0.000	03:03
Control Test	0.080	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 03/02/2023 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: RINCHERE-GUERRIER

DATE/TIME OF TEST: 02/11/2023 22:44:00

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:33	
Control Test	0.081	22:33	
Air Blank	0.000	22:33	
Subject Sample	# 1	22:37	No Sample Provided
Breath Volume	0.000	22:37	
Air Blank	0.000	22:37	
Air Blank	0.000	22:39	
Subject Sample	# 2	22:42	No Sample Provided
Breath Volume	0.000	22:42	
Air Blank	0.000	22:43	
Control Test	0.080	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:44	

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEONE,DANIEL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: TEIXEIRA

DATE/TIME OF TEST: 02/12/2023 00:12:03

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.098	00:07
Breath Volume	1.218	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.097	00:10
Breath Volume	3.488	00:10
Air Blank	0.000	00:11
Control Test	0.081	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:12

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REMEDIOS,ALEXANDER
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 08:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:36

SUBJECT'S LAST NAME: MARTINEZ SALAS

DATE/TIME OF TEST: 02/12/2023 11:21:51

Diagnostics Check	OK	11:15
Air Blank	0.000	11:15
Control Test	0.082	11:15
Air Blank	0.000	11:16
Subject Sample # 1	0.224	11:17
Breath Volume	2.136	11:17
Air Blank	0.000	11:17
Air Blank	0.000	11:19
Subject Sample # 2	0.218	11:20
Breath Volume	2.531	11:20
Air Blank	0.000	11:20
Control Test	0.080	11:21
Air Blank	0.000	11:21
Diagnostics Check	OK	11:21

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARTINEZ,YORDANY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/13/2023 02:30:49

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.089	02:26
Breath Volume	2.152	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.089	02:29
Breath Volume	2.484	02:29
Air Blank	0.000	02:29
Control Test	0.083	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TORRES,MELVIN
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 02/13/2023 21:34:02

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.081	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.243	21:29
Breath Volume	1.296	21:29
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.241	21:32
Breath Volume	1.183	21:32
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:34

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: FACEY

DATE/TIME OF TEST: 02/14/2023 19:02:22

Diagnostics Check	OK	18:53
Air Blank	0.000	18:54
Control Test	0.081	18:54
Air Blank	0.000	18:55
Subject Sample # 1	0.139	18:57
Breath Volume	1.292	18:57
Air Blank	0.000	18:58
Air Blank	0.000	18:59
Subject Sample # 2	0.136	19:00
Breath Volume	1.332	19:00
Air Blank	0.000	19:01
Control Test	0.079	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SCHLAUDER

DATE/TIME OF TEST: 02/15/2023 00:23:26

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.114	00:18
Breath Volume	2.792	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.110	00:21
Breath Volume	2.531	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DIVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/15/2023 02:17:55

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.102	02:12
Breath Volume	1.878	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.098	02:16
Breath Volume	1.726	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MAYES

DATE/TIME OF TEST: 02/15/2023 03:19:04

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.160	03:13
Breath Volume	1.199	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.144	03:17
Breath Volume	1.695	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:19

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILSON,LAMAR
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 08:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:18

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/15/2023 09:57:29

Diagnostics Check	OK	09:51
Air Blank	0.000	09:51
Control Test	0.080	09:51
Air Blank	0.000	09:52
Subject Sample # 1	0.120	09:52
Breath Volume	2.957	09:52
Air Blank	0.000	09:53
Air Blank	0.000	09:55
Subject Sample # 2	0.129	09:55
Breath Volume	2.921	09:55
Air Blank	0.000	09:56
Control Test	0.081	09:56
Air Blank	0.000	09:57
Diagnostics Check	OK	09:57

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MONTANO

DATE/TIME OF TEST: 02/15/2023 22:14:10

Diagnostics Check	OK	22:07
Air Blank	0.000	22:08
Control Test	0.081	22:08
Air Blank	0.000	22:09
Subject Sample # 1	0.137	22:09
Breath Volume	1.957	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.149	22:12
Breath Volume	2.054	22:12
Air Blank	0.000	22:13
Control Test	0.082	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCVEY,KRISTINA
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: LONDONO VASQUEZ

DATE/TIME OF TEST: 02/17/2023 04:20:22

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.196	04:15
Breath Volume	3.128	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.195	04:18
Breath Volume	3.367	04:18
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONOTHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/18/2023 02:25:04

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.178	02:19
Breath Volume	2.386	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.170	02:23
Breath Volume	2.308	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:25

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALSHAIER,ELIAS
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: AVSYUK

DATE/TIME OF TEST: 02/18/2023 02:50:40

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.189	02:46
Breath Volume	1.449	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.191	02:49
Breath Volume	1.757	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RUSSELL,AMON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MOCK

DATE/TIME OF TEST: 02/18/2023 03:53:07

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.081	03:42	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:46	No Sample Provided
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Air Blank	0.000	03:48	
Subject Sample	# 2	03:51	No Sample Provided
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RUSSELL,AMON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MOCK

DATE/TIME OF TEST: 02/18/2023 04:17:37

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.174	04:12
Breath Volume	2.167	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.156	04:15
Breath Volume	1.113	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAPOTA,JOSHUA
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: DAHL

DATE/TIME OF TEST: 02/18/2023 20:22:40

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.080	20:15
Air Blank	0.000	20:16
Subject Sample # 1	0.164	20:17
Breath Volume	1.121	20:17
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.161	20:20
Breath Volume	1.183	20:20
Air Blank	0.000	20:21
Control Test	0.081	20:22
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOLDBERG,JOSHUA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: IRAHETA

DATE/TIME OF TEST: 02/19/2023 01:04:51

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.241	00:59
Breath Volume	2.109	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:01
Subject Sample # 2	0.233	01:03
Breath Volume	1.949	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GIANI,N
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: BUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: WINSOR

DATE/TIME OF TEST: 02/21/2023 00:51:19

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.106	00:46
Breath Volume	2.652	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.111	00:49
Breath Volume	4.140	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BENINCASA,VINCENT
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/21/2023 02:11:22

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.209	02:06
Breath Volume	2.132	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.195	02:09
Breath Volume	1.535	02:09
Air Blank	0.000	02:10
Control Test	0.078	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 03/11/2023 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: RUSALIC

DATE/TIME OF TEST: 02/26/2023 22:56:57

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.081	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.063	22:52
Breath Volume	2.160	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.060	22:55
Breath Volume	1.570	22:55
Air Blank	0.000	22:56
Control Test	0.081	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 03/11/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MAJOR,BERNARD
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: TARKER

DATE/TIME OF TEST: 02/27/2023 07:16:02

Diagnostics Check	OK	07:07
Air Blank	0.000	07:08
Control Test	0.081	07:08
Air Blank	0.000	07:09
Subject Sample # 1	0.089	07:11
Breath Volume	1.148	07:11
Air Blank	0.000	07:12
Air Blank	0.000	07:14
Subject Sample # 2	0.098	07:14
Breath Volume	1.656	07:14
Air Blank	0.000	07:15
Control Test	0.079	07:15
Air Blank	0.000	07:15
Diagnostics Check	OK	07:16

Data Download Date/Time: 03/11/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 02/04/2023 00:51:37

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.083	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.211	00:46
Breath Volume	1.800	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.218	00:49
Breath Volume	1.187	00:49
Air Blank	0.000	00:50
Control Test	0.082	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 02/27/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 02/04/2023 02:24:40

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.083	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.116	02:20
Breath Volume	2.632	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.119	02:23
Breath Volume	2.300	02:23
Air Blank	0.000	02:23
Control Test	0.084	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 02/27/2023 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SETH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: SHARAMEK

DATE/TIME OF TEST: 02/04/2023 04:06:45

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.083	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.130	04:01
Breath Volume	1.699	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.122	04:05
Breath Volume	2.093	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 02/27/2023 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: WHILEY

DATE/TIME OF TEST: 02/12/2023 22:37:15

Diagnostics Check	OK	22:36	
Air Blank		22:36	RFI Detect
Air Blank	0.000	22:37	

Data Download Date/Time: 02/27/2023 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: WHITLEY

DATE/TIME OF TEST: 02/12/2023 22:42:57

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:39	
Control Test	0.083	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:42	
Control Test	0.082	22:42	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 02/27/2023 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/13/2023 01:11:33

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.084	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.167	01:05
Breath Volume	1.402	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.169	01:09
Breath Volume	1.496	01:09
Air Blank	0.000	01:10
Control Test	0.081	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 02/27/2023 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/13/2023 01:33:23

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.160	01:28
Breath Volume	1.625	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.157	01:31
Breath Volume	1.859	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 02/27/2023 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: BOHM,JONASTHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: FURLONG

DATE/TIME OF TEST: 02/02/2023 01:30:22

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.124	01:25
Breath Volume	1.960	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.123	01:28
Breath Volume	1.964	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 02/28/2023 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: SINISTERRA

DATE/TIME OF TEST: 02/10/2023 00:43:26

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.080	00:35	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:38	No Sample Provided
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 2	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 02/28/2023 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: BHOORASINGH

DATE/TIME OF TEST: 02/11/2023 02:44:29

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.283	02:38
Breath Volume	2.457	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.278	02:42
Breath Volume	1.656	02:42
Air Blank	0.000	02:43
Control Test	0.076	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 02/28/2023 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: SIKES

DATE/TIME OF TEST: 02/19/2023 23:37:22

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.078	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.173	23:32
Breath Volume	2.558	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.183	23:35
Breath Volume	2.328	23:35
Air Blank	0.000	23:36
Control Test	0.077	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 03/05/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: COKA

DATE/TIME OF TEST: 02/20/2023 23:56:49

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.077	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.163	23:51
Breath Volume	1.226	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.163	23:55
Breath Volume	1.105	23:55
Air Blank	0.000	23:55
Control Test	0.076	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 03/05/2023 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,DEAN,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PARKS,DEAN
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 02/02/2023 00:41:44

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:38	
Control Test	0.078	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:40	Subject Test Refused
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Control Test	0.078	00:41	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 02/07/2023 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,DEAN,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PARKS,DEAN
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 02/02/2023 23:33:25

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:28	
Control Test	0.078	23:29	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:32	
Control Test	0.078	23:32	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:33	

Data Download Date/Time: 02/07/2023 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA,R,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ZYNDA,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 02/04/2023 04:14:26

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.078	04:13	RFI Detect
Air Blank	0.000	04:14	

Data Download Date/Time: 02/07/2023 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA,R,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ZYNDA,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 02/04/2023 04:23:20

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.078	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.169	04:18
Breath Volume	1.535	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.167	04:21
Breath Volume	1.730	04:21
Air Blank	0.000	04:22
Control Test	0.077	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 02/07/2023 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,T,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PURSLEY,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: BRISTOL

DATE/TIME OF TEST: 02/04/2023 22:21:19

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.079	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.062	22:16
Breath Volume	1.742	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.058	22:19
Breath Volume	2.414	22:19
Air Blank	0.000	22:20
Control Test	0.077	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 02/07/2023 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,GATLIN,T
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WENNLAND,FREDERICK
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MELIDEO

DATE/TIME OF TEST: 02/09/2023 22:05:33

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.078	22:00
Air Blank	0.000	22:00
Subject Sample # 1	0.164	22:01
Breath Volume	2.125	22:01
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.149	22:03
Breath Volume	2.257	22:03
Air Blank	0.000	22:04
Control Test	0.078	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 03/09/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA, URIEL,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PEREZ, O
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 02/11/2023 20:45:18

Diagnostics Check	OK	20:38
Air Blank	0.000	20:39
Control Test	0.078	20:39
Air Blank	0.000	20:39
Subject Sample # 1	0.226	20:40
Breath Volume	1.253	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.236	20:43
Breath Volume	1.546	20:43
Air Blank	0.000	20:44
Control Test	0.077	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 03/09/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,MEGAN,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: EMILIUS,ETHIN
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: DOMINGORAMIREZ

DATE/TIME OF TEST: 02/18/2023 21:01:44

Diagnostics Check	OK	20:52
Air Blank	0.000	20:52
Control Test	0.078	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.243	20:54
Breath Volume	3.054	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:56
Subject Sample # 2	0.218	20:57
Breath Volume	3.511	20:57
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 3	0.228	21:00
Breath Volume	3.113	21:00
Air Blank	0.000	21:00
Control Test	0.075	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 03/09/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,URIEL,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MENDOZA,URIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/25/2023 01:06:19

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.078	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.308	01:01
Breath Volume	1.777	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.304	01:04
Breath Volume	1.636	01:04
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 03/09/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WHITE,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: RESTREPO

DATE/TIME OF TEST: 02/25/2023 04:30:33

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.160	04:25
Breath Volume	2.003	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.161	04:28
Breath Volume	2.613	04:28
Air Blank	0.000	04:29
Control Test	0.078	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 03/09/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCE,M,O
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ZERPA,E
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:38

SUBJECT'S LAST NAME: CACERES-ACOSTA

DATE/TIME OF TEST: 02/26/2023 11:08:51

Diagnostics Check	OK	11:00
Air Blank	0.000	11:01
Control Test	0.078	11:01
Air Blank	0.000	11:02
Subject Sample # 1	0.163	11:03
Breath Volume	1.531	11:03
Air Blank	0.000	11:03
Air Blank	0.000	11:05
Subject Sample # 2	0.158	11:07
Breath Volume	1.156	11:07
Air Blank	0.000	11:07
Control Test	0.076	11:08
Air Blank	0.000	11:08
Diagnostics Check	OK	11:08

Data Download Date/Time: 03/09/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASALA,J,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CONROY,J
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/08/2023 22:31:49

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.082	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.067	22:27
Breath Volume	2.601	22:27
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.061	22:30
Breath Volume	2.085	22:30
Air Blank	0.000	22:30
Control Test	0.079	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENRRIQUEZ,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOVA,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: JARQUINGURDIAN

DATE/TIME OF TEST: 02/11/2023 05:09:21

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.080	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.165	05:04
Breath Volume	1.214	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.169	05:07
Breath Volume	1.261	05:07
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENNINI,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: AMOS

DATE/TIME OF TEST: 02/11/2023 13:35:58

Diagnostics Check	OK	13:29
Air Blank	0.000	13:30
Control Test	0.080	13:30
Air Blank	0.000	13:30
Subject Sample # 1	0.182	13:31
Breath Volume	1.347	13:31
Air Blank	0.000	13:32
Air Blank	0.000	13:33
Subject Sample # 2	0.176	13:34
Breath Volume	1.218	13:34
Air Blank	0.000	13:35
Control Test	0.079	13:35
Air Blank	0.000	13:35
Diagnostics Check	OK	13:35

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,J.,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESPIN,A.
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PEREZ JIMENEZ

DATE/TIME OF TEST: 02/12/2023 03:52:34

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.258	03:44
Breath Volume	2.191	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.229	03:47
Breath Volume	1.402	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 3	0.272	03:50
Breath Volume	3.324	03:50
Air Blank	0.000	03:51
Control Test	0.077	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECORELLA,R,K
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GONZALEZ,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 02/14/2023 02:54:02

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.000	02:49
Breath Volume	2.378	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.000	02:52
Breath Volume	2.507	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:54

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,D,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOWLIN,NA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:57

SUBJECT'S LAST NAME: GILLIGAN

DATE/TIME OF TEST: 02/14/2023 15:30:17

Diagnostics Check	OK	15:23
Air Blank	0.000	15:24
Control Test	0.081	15:24
Air Blank	0.000	15:25
Subject Sample # 1	0.000	15:26
Breath Volume	1.941	15:26
Air Blank	0.000	15:26
Air Blank	0.000	15:28
Subject Sample # 2	0.000	15:28
Breath Volume	1.824	15:28
Air Blank	0.000	15:29
Control Test	0.079	15:29
Air Blank	0.000	15:30
Diagnostics Check	OK	15:30

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GRANESE,J
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: PENALOZA GUTIERREZ

DATE/TIME OF TEST: 02/14/2023 17:52:58

Diagnostics Check	OK	17:45
Air Blank	0.000	17:45
Control Test	0.080	17:46
Air Blank	0.000	17:46
Subject Sample # 1	0.227	17:48
Breath Volume	3.250	17:48
Air Blank	0.000	17:49
Air Blank	0.000	17:50
Subject Sample # 2	0.245	17:51
Breath Volume	1.800	17:51
Air Blank	0.000	17:52
Control Test	0.077	17:52
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDASH, MATTHEW, B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MACIQUES, R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: ROJAS ESPINOSA

DATE/TIME OF TEST: 02/15/2023 00:07:40

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.133	00:02
Breath Volume	2.125	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.121	00:06
Breath Volume	2.015	00:06
Air Blank	0.000	00:06
Control Test	0.079	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NEGRON,JOHN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: BARTOLON LUCAS

DATE/TIME OF TEST: 02/18/2023 04:38:17

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:33	
Control Test	0.081	04:33	
Air Blank	0.000	04:34	
Subject Sample	# 1 0.193	04:37	
Breath Volume	1.449	04:37	
Air Blank		04:37	Purge Fail
Air Blank	0.000	04:38	

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NEGRON,JOHN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: BARTOLON LUCAS

DATE/TIME OF TEST: 02/18/2023 04:40:57

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:40	
Control Test	0.074	04:40	Control Outside Tolerance
Air Blank	0.000	04:40	

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NEGRON,JOHN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: BARTOLON LUCAS

DATE/TIME OF TEST: 02/18/2023 04:49:12

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.077	04:43	
Air Blank	0.000	04:43	
Subject Sample # 1	0.181	04:44	
Breath Volume	1.238	04:44	
Air Blank	0.000	04:45	
Air Blank	0.000	04:46	
Subject Sample # 2	0.178	04:47	
Breath Volume	1.160	04:47	
Air Blank	0.000	04:48	
Control Test	0.073	04:48	Control Outside Tolerance
Air Blank	0.000	04:49	

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,T,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PRUCHNIK,A
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: MACIAS

DATE/TIME OF TEST: 02/18/2023 21:36:18

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.080	21:30
Air Blank	0.000	21:30
Subject Sample # 1	0.234	21:31
Breath Volume	1.746	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:33
Subject Sample # 2	0.231	21:34
Breath Volume	2.664	21:34
Air Blank	0.000	21:35
Control Test	0.078	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PURSLEY,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 02/19/2023 03:37:07

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.233	03:32
Breath Volume	2.304	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.219	03:35
Breath Volume	2.394	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MORRIS,H
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 02/20/2023 03:00:35

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.145	02:55
Breath Volume	2.843	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.146	02:58
Breath Volume	3.039	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WILLIAMS,ZACHARY
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: CASILLAS

DATE/TIME OF TEST: 02/20/2023 03:50:34

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.080	03:43	
Air Blank	0.000	03:43	
Subject Sample # 1	0.237	03:44	
Breath Volume	2.511	03:44	
Air Blank	0.000	03:45	
Air Blank	0.000	03:46	
Subject Sample # 2	0.257	03:49	
Breath Volume	1.386	03:49	
Air Blank		03:50	Purge Fail
Air Blank	0.000	03:50	

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLOAT,LUCAS,G
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: RAMOS,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: SUTKUS

DATE/TIME OF TEST: 02/21/2023 01:25:48

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.217	01:20
Breath Volume	1.195	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.210	01:24
Breath Volume	1.359	01:24
Air Blank	0.000	01:24
Control Test	0.078	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CLUCHEY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/22/2023 00:58:41

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.139	00:53
Breath Volume	1.496	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample # 2	0.138	00:56
Breath Volume	1.593	00:56
Air Blank	0.000	00:57
Control Test	0.075	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,J.,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SENA,M.
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: PENA PINEDA

DATE/TIME OF TEST: 02/25/2023 03:54:33

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.240	03:49
Breath Volume	1.574	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.243	03:52
Breath Volume	1.933	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CLUCHEY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 02/27/2023 05:33:10

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.000	05:28
Breath Volume	2.050	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.000	05:31
Breath Volume	2.050	05:31
Air Blank	0.000	05:32
Control Test	0.079	05:32
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 03/07/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRITES,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,JONATHAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: FARMERIE

DATE/TIME OF TEST: 02/03/2023 21:33:54

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.080	21:28
Air Blank	0.000	21:28
Subject Sample # 1	0.144	21:29
Breath Volume	1.656	21:29
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.132	21:32
Breath Volume	1.460	21:32
Air Blank	0.000	21:32
Control Test	0.080	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 02/28/2023 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GANK,T
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: ALBRIGHT

DATE/TIME OF TEST: 02/07/2023 09:07:36

Diagnostics Check	OK	09:00
Air Blank	0.000	09:01
Control Test	0.080	09:01
Air Blank	0.000	09:01
Subject Sample # 1	0.000	09:02
Breath Volume	2.226	09:02
Air Blank	0.000	09:02
Air Blank	0.000	09:04
Subject Sample # 2	0.000	09:06
Breath Volume	1.281	09:06
Air Blank	0.000	09:06
Control Test	0.079	09:06
Air Blank	0.000	09:07
Diagnostics Check	OK	09:07

Data Download Date/Time: 02/28/2023 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCWARTZ,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: KOTTHOFF

DATE/TIME OF TEST: 02/08/2023 21:26:08

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.081	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.175	21:21
Breath Volume	2.023	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:23
Subject Sample # 2	0.173	21:24
Breath Volume	2.179	21:24
Air Blank	0.000	21:25
Control Test	0.081	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 02/28/2023 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DEGRASSE,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/09/2023 20:43:17

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:35	
Control Test	0.081	20:35	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:39	No Sample Provided
Breath Volume	0.000	20:39	
Air Blank	0.000	20:39	
Air Blank	0.000	20:41	
Subject Sample	# 2	20:41	Subject Test Refused
Breath Volume	0.000	20:41	
Air Blank	0.000	20:42	
Control Test	0.079	20:42	
Air Blank	0.000	20:43	
Diagnostics Check	OK	20:43	

Data Download Date/Time: 02/28/2023 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRITES,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JANKOWSKI,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 02/13/2023 23:06:44

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.080	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.143	23:02
Breath Volume	2.339	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.137	23:05
Breath Volume	1.843	23:05
Air Blank	0.000	23:05
Control Test	0.077	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 02/28/2023 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAZZARESE,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SELLA,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: OTHER
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 02/16/2023 23:30:47

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.081	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.223	23:25
Breath Volume	1.675	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.221	23:29
Breath Volume	2.296	23:29
Air Blank	0.000	23:29
Control Test	0.080	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 02/28/2023 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAZZARESE,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: SICILIAN

DATE/TIME OF TEST: 02/17/2023 00:19:33

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.081	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.124	00:14
Breath Volume	2.167	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.117	00:17
Breath Volume	1.988	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAZZARESE,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SELLA,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: OTHER
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 02/17/2023 05:04:59

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.101	05:00
Breath Volume	3.058	05:00
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.102	05:03
Breath Volume	2.695	05:03
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 02/17/2023 07:59:00

Diagnostics Check	OK	07:52
Air Blank	0.000	07:53
Control Test	0.080	07:53
Air Blank	0.000	07:54
Subject Sample # 1	0.048	07:54
Breath Volume	2.656	07:54
Air Blank	0.000	07:55
Air Blank	0.000	07:56
Subject Sample # 2	0.047	07:57
Breath Volume	2.281	07:57
Air Blank	0.000	07:58
Control Test	0.080	07:58
Air Blank	0.000	07:58
Diagnostics Check	OK	07:59

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 02/17/2023 17:42:01

Diagnostics Check	OK	17:30	
Air Blank	0.000	17:31	
Control Test	0.080	17:31	
Air Blank	0.000	17:31	
Subject Sample	# 1 0.000	17:35	Volume Not Met
Breath Volume	0.441	17:35	
Air Blank	0.000	17:35	
Air Blank	0.000	17:37	
Subject Sample	# 2	17:40	No Sample Provided
Breath Volume	0.000	17:40	
Air Blank	0.000	17:41	
Control Test	0.079	17:41	
Air Blank	0.000	17:41	
Diagnostics Check	OK	17:42	

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WYMER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 02/17/2023 22:09:07

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.081	21:55	
Air Blank	0.000	21:56	
Subject Sample	# 1 0.000	21:59	Volume Not Met
Breath Volume	0.855	21:59	
Air Blank	0.000	22:00	
Air Blank	0.000	22:02	
Subject Sample	# 2 0.000	22:04	
Breath Volume	1.449	22:04	
Air Blank	0.000	22:05	
Air Blank	0.000	22:07	
Subject Sample	# 3 0.000	22:07	
Breath Volume	1.351	22:07	
Air Blank	0.000	22:08	
Control Test	0.080	22:08	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MCGONIGAL BUSQUE

DATE/TIME OF TEST: 02/18/2023 02:46:08

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.179	02:41
Breath Volume	2.703	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:43
Subject Sample # 2	0.181	02:44
Breath Volume	1.648	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,CDFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 13:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 02/18/2023 13:45:15

Diagnostics Check	OK	13:38
Air Blank	0.000	13:38
Control Test	0.081	13:39
Air Blank	0.000	13:39
Subject Sample # 1	0.000	13:40
Breath Volume	1.816	13:40
Air Blank	0.000	13:40
Air Blank	0.000	13:42
Subject Sample # 2	0.000	13:43
Breath Volume	1.804	13:43
Air Blank	0.000	13:44
Control Test	0.080	13:44
Air Blank	0.000	13:45
Diagnostics Check	OK	13:45

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: SELWYN

DATE/TIME OF TEST: 02/21/2023 15:52:17

Diagnostics Check	OK	15:45
Air Blank	0.000	15:46
Control Test	0.081	15:46
Air Blank	0.000	15:47
Subject Sample # 1	0.122	15:47
Breath Volume	1.257	15:47
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 2	0.121	15:50
Breath Volume	1.441	15:50
Air Blank	0.000	15:51
Control Test	0.079	15:51
Air Blank	0.000	15:52
Diagnostics Check	OK	15:52

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARCHMAN,MARCHMAN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CCSO,CCSO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 02/21/2023 20:02:31

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.081	19:56
Air Blank	0.000	19:56
Subject Sample # 1	0.126	19:57
Breath Volume	2.835	19:57
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.127	20:00
Breath Volume	3.078	20:00
Air Blank	0.000	20:01
Control Test	0.080	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONGA,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: IRSUNIT,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 02/22/2023 15:46:03

Diagnostics Check	OK	15:38
Air Blank	0.000	15:38
Control Test	0.080	15:39
Air Blank	0.000	15:39
Subject Sample # 1	0.044	15:40
Breath Volume	1.437	15:40
Air Blank	0.000	15:41
Air Blank	0.000	15:43
Subject Sample # 2	0.041	15:44
Breath Volume	1.667	15:44
Air Blank	0.000	15:45
Control Test	0.080	15:45
Air Blank	0.000	15:45
Diagnostics Check	OK	15:46

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARCHMAN,MARCHMAN,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DEPUTY,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 02/23/2023 18:15:32

Diagnostics Check	OK	18:09
Air Blank	0.000	18:09
Control Test	0.080	18:09
Air Blank	0.000	18:10
Subject Sample # 1	0.000	18:11
Breath Volume	2.109	18:11
Air Blank	0.000	18:11
Air Blank	0.000	18:13
Subject Sample # 2	0.000	18:14
Breath Volume	2.281	18:14
Air Blank	0.000	18:14
Control Test	0.080	18:14
Air Blank	0.000	18:15
Diagnostics Check	OK	18:15

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORD,JAMES,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MARCHMAN,MARCHMAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 02/24/2023 18:49:51

Diagnostics Check	OK	18:43
Air Blank	0.000	18:43
Control Test	0.078	18:43
Air Blank	0.000	18:44
Subject Sample # 1	0.000	18:45
Breath Volume	3.398	18:45
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample # 2	0.000	18:48
Breath Volume	3.625	18:48
Air Blank	0.000	18:48
Control Test	0.078	18:49
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BANNAN-MATOS

DATE/TIME OF TEST: 02/27/2023 23:45:33

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.000	23:40
Breath Volume	1.355	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.000	23:44
Breath Volume	1.925	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 02/28/2023 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SITKO,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 02/28/2023 00:14:10

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.237	00:09
Breath Volume	1.453	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.233	00:12
Breath Volume	1.429	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 02/28/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RILEY,OFC
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BELTON

DATE/TIME OF TEST: 02/28/2023 01:54:25

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.232	01:49
Breath Volume	1.773	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.230	01:52
Breath Volume	2.191	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 02/28/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILAR,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORETON,MYKELL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: FUNDORA ACOSTA

DATE/TIME OF TEST: 02/16/2023 01:39:15

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.145	01:30
Breath Volume	2.562	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.167	01:34
Breath Volume	2.007	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 3	0.158	01:37
Breath Volume	2.117	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 03/09/2023 03:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILAR,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AGUILAR,ALEXANDER
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: CABEZAS

DATE/TIME OF TEST: 02/16/2023 23:52:30

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.081	23:50	
Air Blank	0.000	23:50	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:51	
Control Test	0.081	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 03/09/2023 03:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIL,ALEJANDRO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORETON,MYKELL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CETO CHAVEZ

DATE/TIME OF TEST: 02/17/2023 00:37:25

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.080	00:26	
Air Blank	0.000	00:26	
Subject Sample	# 1 0.108	00:30	Volume Not Met
Breath Volume	0.429	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 2 0.096	00:35	Volume Not Met
Breath Volume	0.953	00:35	
Air Blank	0.000	00:36	
Control Test	0.073	00:36	Control Outside Tolerance
Air Blank	0.000	00:37	

Data Download Date/Time: 03/09/2023 03:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIL,ALEJANDRO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORETON,MYKELL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CETO CHAVEZ

DATE/TIME OF TEST: 02/17/2023 01:04:55

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.119	01:00
Breath Volume	1.984	01:00
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.108	01:03
Breath Volume	1.328	01:03
Air Blank	0.000	01:03
Control Test	0.078	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 03/09/2023 03:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLEDO,MIGUEL,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ROBLEDO,MIGUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:21

SUBJECT'S LAST NAME: PAULINO

DATE/TIME OF TEST: 02/18/2023 09:54:19

Diagnostics Check	OK	09:50	
Air Blank	0.000	09:51	
Control Test	0.080	09:51	
Air Blank	0.000	09:52	
Subject Sample	# 1	09:52	Subject Test Refused
Breath Volume	0.000	09:52	
Air Blank	0.000	09:53	
Control Test	0.081	09:53	
Air Blank	0.000	09:54	
Diagnostics Check	OK	09:54	

Data Download Date/Time: 03/09/2023 03:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:23

SUBJECT'S LAST NAME: ARAGON OLIVA

DATE/TIME OF TEST: 02/18/2023 11:54:16

Diagnostics Check	OK	11:45
Air Blank	0.000	11:45
Control Test	0.080	11:46
Air Blank	0.000	11:46
Subject Sample # 1	0.168	11:48
Breath Volume	1.656	11:48
Air Blank	0.000	11:49
Air Blank	0.000	11:51
Subject Sample # 2	0.158	11:52
Breath Volume	1.562	11:52
Air Blank	0.000	11:53
Control Test	0.078	11:53
Air Blank	0.000	11:54
Diagnostics Check	OK	11:54

Data Download Date/Time: 03/09/2023 03:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILAR,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORETON,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ABEA

DATE/TIME OF TEST: 02/22/2023 01:12:59

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:10	
Control Test	0.080	01:10	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:12	
Control Test	0.081	01:12	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 03/09/2023 03:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,VANESSA,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORETON,MYKELL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ESPINOZA GONZALEZ

DATE/TIME OF TEST: 02/25/2023 01:40:47

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.204	01:35
Breath Volume	1.402	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:37
Subject Sample # 2	0.211	01:39
Breath Volume	2.339	01:39
Air Blank	0.000	01:39
Control Test	0.078	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 03/09/2023 03:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/01/2023 23:37:34

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.083	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.179	23:32
Breath Volume	1.410	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:34
Subject Sample # 2	0.179	23:35
Breath Volume	1.593	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 03/11/2023 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: PABON

DATE/TIME OF TEST: 02/02/2023 22:09:36

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.083	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.169	22:04
Breath Volume	1.937	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.175	22:07
Breath Volume	2.171	22:07
Air Blank	0.000	22:08
Control Test	0.082	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 03/11/2023 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 02/05/2023 21:58:31

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.082	21:53
Air Blank	0.000	21:53
Subject Sample # 1	0.215	21:54
Breath Volume	3.035	21:54
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.214	21:56
Breath Volume	2.683	21:56
Air Blank	0.000	21:57
Control Test	0.082	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 03/11/2023 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/06/2023 01:45:40

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.083	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.186	01:41
Breath Volume	1.675	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.184	01:44
Breath Volume	2.089	01:44
Air Blank	0.000	01:44
Control Test	0.082	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 03/11/2023 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: OSBORNE,TIFFANI
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: ROBSON

DATE/TIME OF TEST: 02/08/2023 20:10:45

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.083	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.053	20:05
Breath Volume	2.125	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.052	20:09
Breath Volume	1.992	20:09
Air Blank	0.000	20:09
Control Test	0.083	20:10
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 02/08/2023 21:09:01

Diagnostics Check	OK	20:58	
Air Blank	0.000	20:58	
Control Test	0.082	20:59	
Air Blank	0.000	20:59	
Subject Sample	# 1 0.166	21:02	Volume Not Met
Breath Volume	0.199	21:02	
Air Blank	0.000	21:03	
Air Blank	0.000	21:05	
Subject Sample	# 2	21:07	Subject Test Refused
Breath Volume	0.152	21:07	
Air Blank	0.000	21:08	
Control Test	0.081	21:08	
Air Blank	0.000	21:08	
Diagnostics Check	OK	21:09	

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BEAUREGARD,JEFF
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: APONTE

DATE/TIME OF TEST: 02/08/2023 22:11:02

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.082	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.145	22:05
Breath Volume	1.351	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:07
Subject Sample # 2	0.139	22:09
Breath Volume	1.878	22:09
Air Blank	0.000	22:10
Control Test	0.082	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:11

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BROWN,EDDIE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: FLEURIVAL

DATE/TIME OF TEST: 02/10/2023 00:23:28

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.083	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.192	00:18
Breath Volume	2.117	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.187	00:21
Breath Volume	1.882	00:21
Air Blank	0.000	00:22
Control Test	0.082	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BEAUREGARD,JEFF
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: HELD

DATE/TIME OF TEST: 02/10/2023 00:51:43

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.083	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.188	00:47
Breath Volume	1.972	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.190	00:50
Breath Volume	1.964	00:50
Air Blank	0.000	00:50
Control Test	0.081	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALSHAIER,ELIAS
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: DYER

DATE/TIME OF TEST: 02/10/2023 03:02:52

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.082	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.157	02:58
Breath Volume	2.445	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample # 2	0.152	03:01
Breath Volume	1.988	03:01
Air Blank	0.000	03:01
Control Test	0.082	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CASTILLO,MICHAEL
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 02/10/2023 04:26:28

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.276	04:21
Breath Volume	3.851	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.279	04:24
Breath Volume	3.222	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JEREMY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: GIRARDEY

DATE/TIME OF TEST: 02/10/2023 21:29:30

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.082	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.377	21:24
Breath Volume	2.460	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.357	21:27
Breath Volume	2.234	21:27
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,MARCOS
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 09:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:41

SUBJECT'S LAST NAME: SOLIMAN

DATE/TIME OF TEST: 02/20/2023 12:16:24

Diagnostics Check	OK	12:10
Air Blank	0.000	12:10
Control Test	0.082	12:10
Air Blank	0.000	12:11
Subject Sample # 1	0.173	12:11
Breath Volume	2.808	12:11
Air Blank	0.000	12:12
Air Blank	0.000	12:14
Subject Sample # 2	0.158	12:14
Breath Volume	1.570	12:14
Air Blank	0.000	12:15
Control Test	0.083	12:15
Air Blank	0.000	12:16
Diagnostics Check	OK	12:16

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:12

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 02/21/2023 13:44:09

Diagnostics Check	OK	13:37
Air Blank	0.000	13:38
Control Test	0.083	13:38
Air Blank	0.000	13:39
Subject Sample # 1	0.000	13:39
Breath Volume	3.718	13:39
Air Blank	0.000	13:40
Air Blank	0.000	13:42
Subject Sample # 2	0.000	13:42
Breath Volume	3.671	13:42
Air Blank	0.000	13:43
Control Test	0.083	13:43
Air Blank	0.000	13:44
Diagnostics Check	OK	13:44

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LYON,ALVARO
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/22/2023 04:24:07

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.083	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.276	04:19
Breath Volume	2.136	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.286	04:22
Breath Volume	2.261	04:22
Air Blank	0.000	04:23
Control Test	0.082	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 02/22/2023 06:07:56

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.082	06:02
Air Blank	0.000	06:02
Subject Sample # 1	0.125	06:03
Breath Volume	2.468	06:03
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.125	06:06
Breath Volume	1.914	06:06
Air Blank	0.000	06:07
Control Test	0.082	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRUJILLO,ELAINE
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 02/23/2023 14:55:37

Diagnostics Check	OK	14:47	
Air Blank	0.000	14:47	
Control Test	0.083	14:47	
Air Blank	0.000	14:48	
Subject Sample	# 1	14:51	No Sample Provided
Breath Volume	0.000	14:51	
Air Blank	0.000	14:52	
Air Blank	0.000	14:54	
Subject Sample	# 2	14:54	Subject Test Refused
Breath Volume	0.000	14:54	
Air Blank	0.000	14:54	
Control Test	0.083	14:55	
Air Blank	0.000	14:55	
Diagnostics Check	OK	14:55	

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: BLUM

DATE/TIME OF TEST: 02/24/2023 21:19:45

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.082	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.208	21:15
Breath Volume	1.976	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:17
Subject Sample # 2	0.215	21:18
Breath Volume	1.515	21:18
Air Blank	0.000	21:18
Control Test	0.082	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 03/11/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROUTE,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: LANDOLT

DATE/TIME OF TEST: 02/01/2023 18:16:55

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.080	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.045	18:12
Breath Volume	2.222	18:12
Air Blank	0.000	18:13
Air Blank	0.000	18:14
Subject Sample # 2	0.045	18:15
Breath Volume	2.144	18:15
Air Blank	0.000	18:15
Control Test	0.079	18:16
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 02/07/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAASE,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 02/02/2023 03:03:22

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.150	02:58
Breath Volume	2.718	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.158	03:01
Breath Volume	1.582	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 02/07/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: PUENTES

DATE/TIME OF TEST: 02/03/2023 01:41:37

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.208	01:37
Breath Volume	1.984	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.204	01:40
Breath Volume	1.625	01:40
Air Blank	0.000	01:40
Control Test	0.078	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 02/07/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BALSON

DATE/TIME OF TEST: 02/04/2023 01:50:01

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.124	01:45
Breath Volume	2.207	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.115	01:48
Breath Volume	1.628	01:48
Air Blank	0.000	01:49
Control Test	0.078	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:50

Data Download Date/Time: 02/07/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GRILL

DATE/TIME OF TEST: 02/04/2023 04:18:05

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.215	04:13
Breath Volume	2.558	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.219	04:16
Breath Volume	2.324	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 02/07/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: BARRIENTOS

DATE/TIME OF TEST: 02/05/2023 22:09:54

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.080	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.159	22:05
Breath Volume	2.136	22:05
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.162	22:08
Breath Volume	1.546	22:08
Air Blank	0.000	22:08
Control Test	0.079	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 02/07/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: CLOUTIER

DATE/TIME OF TEST: 02/06/2023 02:35:17

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.174	02:30
Breath Volume	1.683	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:32
Subject Sample # 2	0.168	02:33
Breath Volume	2.269	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 02/07/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HARRISON,GARY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 02/06/2023 16:57:40

Diagnostics Check	OK	16:50
Air Blank	0.000	16:51
Control Test	0.080	16:51
Air Blank	0.000	16:51
Subject Sample # 1	0.271	16:53
Breath Volume	1.769	16:53
Air Blank	0.000	16:53
Air Blank	0.000	16:55
Subject Sample # 2	0.280	16:56
Breath Volume	3.070	16:56
Air Blank	0.000	16:56
Control Test	0.079	16:57
Air Blank	0.000	16:57
Diagnostics Check	OK	16:57

Data Download Date/Time: 02/07/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 02/11/2023 01:08:36

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.149	01:03
Breath Volume	1.785	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.149	01:07
Breath Volume	1.527	01:07
Air Blank	0.000	01:07
Control Test	0.078	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 03/02/2023 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCON,K
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: SHOATS

DATE/TIME OF TEST: 02/15/2023 07:38:19

Diagnostics Check	OK	07:29	
Air Blank	0.000	07:30	
Control Test	0.080	07:30	
Air Blank	0.000	07:31	
Subject Sample	# 1	07:34	No Sample Provided
Breath Volume	0.000	07:34	
Air Blank	0.000	07:34	
Air Blank	0.000	07:36	
Subject Sample	# 2	07:36	Subject Test Refused
Breath Volume	0.000	07:36	
Air Blank	0.000	07:37	
Control Test	0.080	07:37	
Air Blank	0.000	07:38	
Diagnostics Check	OK	07:38	

Data Download Date/Time: 03/02/2023 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BRANDON

DATE/TIME OF TEST: 02/18/2023 00:41:14

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.079	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.178	00:36
Breath Volume	1.445	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.188	00:39
Breath Volume	1.652	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 03/02/2023 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 02/20/2023 00:40:49

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.000	00:36
Breath Volume	3.230	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.000	00:39
Breath Volume	2.621	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 03/02/2023 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: TRAMMELL

DATE/TIME OF TEST: 02/22/2023 00:55:46

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1 0.237	00:51	Volume Not Met
Breath Volume	0.066	00:51	
Air Blank	0.000	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 2	00:54	Subject Test Refused
Breath Volume	0.000	00:54	
Air Blank	0.000	00:54	
Control Test	0.077	00:55	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 03/02/2023 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WASHINGTON,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: TORRES JR

DATE/TIME OF TEST: 02/28/2023 23:56:18

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.079	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.193	23:51
Breath Volume	1.621	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:53
Subject Sample # 2	0.198	23:54
Breath Volume	1.585	23:54
Air Blank	0.000	23:55
Control Test	0.078	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 03/02/2023 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLLINS,CHASE,M

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: COLLINS,CHASE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1346146

DRY GAS EXPIRY DATE: 07/07/2023

VIOLATION CODE: DUI

ARREST TIME: 09:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:54

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/03/2023 10:29:15

Diagnostics Check	OK	10:22
Air Blank	0.000	10:23
Control Test	0.083	10:23
Air Blank	0.000	10:23
Subject Sample # 1	0.000	10:24
Breath Volume	1.312	10:24
Air Blank	0.000	10:24
Air Blank	0.000	10:26
Subject Sample # 2	0.000	10:27
Breath Volume	1.648	10:27
Air Blank	0.000	10:28
Control Test	0.082	10:28
Air Blank	0.000	10:29
Diagnostics Check	OK	10:29

Data Download Date/Time: 02/23/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROEDEL,JUSTIN,M

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: JAMES,DEQUAN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1346146

DRY GAS EXPIRY DATE: 07/07/2023

VIOLATION CODE: DUI

ARREST TIME: 01:28

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: EDMOND

DATE/TIME OF TEST: 02/04/2023 02:39:05

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.084	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.274	02:34
Breath Volume	2.453	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.278	02:37
Breath Volume	2.152	02:37
Air Blank	0.000	02:38
Control Test	0.082	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 02/23/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY-IVERSEN,MIYA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAY-IVERSEN,MIYA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 02/14/2023 23:02:26

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.083	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.000	22:57
Breath Volume	2.179	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.000	23:01
Breath Volume	3.273	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 02/23/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CHASE,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COLLINS,CHASE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/15/2023 17:29:55

Diagnostics Check	OK	17:18	
Air Blank	0.000	17:18	
Control Test	0.082	17:18	
Air Blank	0.000	17:19	
Subject Sample	# 1 0.198	17:22	Volume Not Met
Breath Volume	0.054	17:22	
Air Blank	0.000	17:23	
Air Blank	0.000	17:25	
Subject Sample	# 2 0.202	17:28	Volume Not Met
Breath Volume	0.324	17:28	
Air Blank	0.000	17:28	
Control Test	0.080	17:29	
Air Blank	0.000	17:29	
Diagnostics Check	OK	17:29	

Data Download Date/Time: 02/23/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLLINS,CHASE,M

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: COLLINS,CHASE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1346146

DRY GAS EXPIRY DATE: 07/07/2023

VIOLATION CODE: DUI

ARREST TIME: 16:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/15/2023 18:49:35

Diagnostics Check	OK	18:42
Air Blank	0.000	18:42
Control Test	0.081	18:43
Air Blank	0.000	18:43
Subject Sample # 1	0.172	18:44
Breath Volume	1.363	18:44
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample # 2	0.177	18:47
Breath Volume	1.726	18:47
Air Blank	0.000	18:48
Control Test	0.081	18:48
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 02/23/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROEDEL,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ROEDEL,JUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: PIZANO

DATE/TIME OF TEST: 02/19/2023 03:17:45

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.082	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1 0.120	03:04	
Breath Volume	2.667	03:04	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 2 0.119	03:10	Volume Not Met
Breath Volume	0.429	03:10	
Air Blank	0.000	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 3 0.109	03:15	Volume Not Met
Breath Volume	0.316	03:15	
Air Blank	0.000	03:16	
Control Test	0.079	03:17	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 02/23/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEARN,SHELDON,J

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: HASWELL,AUSTIN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1346146

DRY GAS EXPIRY DATE: 07/07/2023

VIOLATION CODE: DUI

ARREST TIME: 04:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/19/2023 05:19:13

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.082	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.097	05:13
Breath Volume	3.894	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.101	05:17
Breath Volume	3.960	05:17
Air Blank	0.000	05:18
Control Test	0.080	05:18
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 02/23/2023 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFKJEN,THOMAS,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GAFKJEN,THOMAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: CAPUZZO

DATE/TIME OF TEST: 02/20/2023 12:25:01

Diagnostics Check	OK	12:22	
Air Blank	0.000	12:22	
Control Test	0.082	12:22	
Air Blank	0.000	12:23	
Subject Sample	# 1	12:23	Subject Test Refused
Breath Volume	0.000	12:23	
Air Blank	0.000	12:24	
Control Test	0.081	12:24	
Air Blank	0.000	12:24	
Diagnostics Check	OK	12:25	

Data Download Date/Time: 02/23/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOVELL,ADAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 02/03/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: HIGLEY

DATE/TIME OF TEST: 02/03/2023 22:22:21

Diagnostics Check	Fail	22:21	Internal Printer Test
Air Blank	0.000	22:22	

Data Download Date/Time: 02/06/2023 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOVELL,ADAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 02/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: HIGLEY

DATE/TIME OF TEST: 02/03/2023 22:32:31

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.081	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.146	22:27
Breath Volume	1.261	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.142	22:30
Breath Volume	1.375	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 02/06/2023 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOVELL,ADAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 02/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: PALACIOUS

DATE/TIME OF TEST: 02/04/2023 23:32:07

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.081	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.139	23:27
Breath Volume	1.468	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.130	23:30
Breath Volume	1.890	23:30
Air Blank	0.000	23:31
Control Test	0.078	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 02/06/2023 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY, REX, M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: TOMBERLIN, K
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: PIEDAD CAMACHO

DATE/TIME OF TEST: 02/26/2023 09:16:02

Diagnostics Check	OK	09:09
Air Blank	0.000	09:10
Control Test	0.080	09:10
Air Blank	0.000	09:11
Subject Sample # 1	0.086	09:11
Breath Volume	2.023	09:11
Air Blank	0.000	09:12
Air Blank	0.000	09:14
Subject Sample # 2	0.081	09:14
Breath Volume	3.281	09:14
Air Blank	0.000	09:15
Control Test	0.078	09:15
Air Blank	0.000	09:15
Diagnostics Check	OK	09:16

Data Download Date/Time: 04/03/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: MALDONADO CRUZ

DATE/TIME OF TEST: 02/09/2023 18:56:43

Diagnostics Check	OK	18:50
Air Blank	0.000	18:50
Control Test	0.079	18:50
Air Blank	0.000	18:51
Subject Sample # 1	0.193	18:52
Breath Volume	2.652	18:52
Air Blank	0.000	18:52
Air Blank	0.000	18:54
Subject Sample # 2	0.198	18:55
Breath Volume	2.691	18:55
Air Blank	0.000	18:55
Control Test	0.079	18:56
Air Blank	0.000	18:56
Diagnostics Check	OK	18:56

Data Download Date/Time: 03/02/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: AUGUSTINE

DATE/TIME OF TEST: 02/09/2023 23:33:47

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.066	23:29
Breath Volume	4.300	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.066	23:32
Breath Volume	4.476	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 03/02/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: BOLANOS

DATE/TIME OF TEST: 02/10/2023 00:15:12

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.079	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.161	00:10
Breath Volume	1.730	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.172	00:13
Breath Volume	1.812	00:13
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 03/02/2023 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: RODRIGUEZ VELASQUEZ

DATE/TIME OF TEST: 02/10/2023 01:15:52

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.187	01:08
Breath Volume	1.347	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.176	01:14
Breath Volume	1.675	01:14
Air Blank	0.000	01:14
Control Test	0.078	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 03/02/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: SPOLAR

DATE/TIME OF TEST: 02/10/2023 22:52:20

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.191	22:47
Breath Volume	1.480	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:49
Subject Sample # 2	0.187	22:50
Breath Volume	2.074	22:50
Air Blank	0.000	22:51
Control Test	0.078	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 03/02/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Plant City PD
ARRESTING OFFICER: DURRANCE,C
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SURIANO ANTONIO

DATE/TIME OF TEST: 02/11/2023 02:04:03

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.113	01:59
Breath Volume	1.894	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:01
Subject Sample # 2	0.114	02:02
Breath Volume	1.839	02:02
Air Blank	0.000	02:03
Control Test	0.078	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:04

Data Download Date/Time: 03/02/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: HORNER

DATE/TIME OF TEST: 02/11/2023 04:16:13

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.078	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.156	04:11
Breath Volume	3.351	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.156	04:14
Breath Volume	3.339	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 03/02/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: MBO MEDJA

DATE/TIME OF TEST: 02/11/2023 06:42:20

Diagnostics Check	OK	06:34
Air Blank	0.000	06:35
Control Test	0.078	06:35
Air Blank	0.000	06:36
Subject Sample # 1	0.116	06:37
Breath Volume	1.285	06:37
Air Blank	0.000	06:38
Air Blank	0.000	06:40
Subject Sample # 2	0.117	06:40
Breath Volume	1.328	06:40
Air Blank	0.000	06:41
Control Test	0.079	06:41
Air Blank	0.000	06:42
Diagnostics Check	OK	06:42

Data Download Date/Time: 03/02/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEWIS,I
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: RENGIFO OCHOA

DATE/TIME OF TEST: 02/12/2023 01:58:41

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.157	01:53
Breath Volume	1.578	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:55
Subject Sample # 2	0.157	01:57
Breath Volume	2.132	01:57
Air Blank	0.000	01:57
Control Test	0.079	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 03/02/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/12/2023 03:06:13

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.079	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.046	03:01
Breath Volume	1.503	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.046	03:04
Breath Volume	1.351	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 03/02/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 02/12/2023 03:26:49

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.180	03:21
Breath Volume	1.339	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.189	03:25
Breath Volume	2.105	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 03/02/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/12/2023 04:32:08

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.142	04:27
Breath Volume	1.289	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.143	04:30
Breath Volume	1.660	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 03/02/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: NIZ LOPEZ

DATE/TIME OF TEST: 02/12/2023 04:50:51

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.079	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.286	04:46
Breath Volume	2.312	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.281	04:49
Breath Volume	1.882	04:49
Air Blank	0.000	04:49
Control Test	0.078	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 03/02/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/12/2023 06:31:01

Diagnostics Check	OK	06:24
Air Blank	0.000	06:24
Control Test	0.079	06:25
Air Blank	0.000	06:25
Subject Sample # 1	0.108	06:26
Breath Volume	1.425	06:26
Air Blank	0.000	06:26
Air Blank	0.000	06:28
Subject Sample # 2	0.108	06:29
Breath Volume	2.382	06:29
Air Blank	0.000	06:30
Control Test	0.079	06:30
Air Blank	0.000	06:30
Diagnostics Check	OK	06:31

Data Download Date/Time: 03/02/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GREGORY,D
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: BAUTISTA TRINIDAD

DATE/TIME OF TEST: 02/12/2023 07:50:50

Diagnostics Check	OK	07:44
Air Blank	0.000	07:44
Control Test	0.080	07:44
Air Blank	0.000	07:45
Subject Sample # 1	0.234	07:46
Breath Volume	1.414	07:46
Air Blank	0.000	07:46
Air Blank	0.000	07:48
Subject Sample # 2	0.237	07:49
Breath Volume	2.171	07:49
Air Blank	0.000	07:49
Control Test	0.080	07:50
Air Blank	0.000	07:50
Diagnostics Check	OK	07:50

Data Download Date/Time: 03/02/2023 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 02/13/2023 00:36:36

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.124	00:31
Breath Volume	1.132	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.120	00:34
Breath Volume	1.773	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 03/02/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: CARTER JR

DATE/TIME OF TEST: 02/13/2023 01:34:48

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.137	01:29
Breath Volume	2.308	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.137	01:33
Breath Volume	3.355	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 03/02/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: TOBY

DATE/TIME OF TEST: 02/13/2023 03:37:55

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.200	03:33
Breath Volume	1.656	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.200	03:36
Breath Volume	1.886	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 03/02/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: CAMBRIDGE

DATE/TIME OF TEST: 02/13/2023 04:13:02

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.182	04:08
Breath Volume	3.425	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.179	04:11
Breath Volume	3.246	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:13

Data Download Date/Time: 03/02/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: PEREZ SALES

DATE/TIME OF TEST: 02/13/2023 06:04:44

Diagnostics Check	OK	05:56
Air Blank	0.000	05:56
Control Test	0.079	05:57
Air Blank	0.000	05:57
Subject Sample # 1	0.176	06:00
Breath Volume	2.335	06:00
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.176	06:03
Breath Volume	3.234	06:03
Air Blank	0.000	06:03
Control Test	0.080	06:04
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 03/02/2023 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SUMMIT

DATE/TIME OF TEST: 02/14/2023 02:59:12

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.079	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1 0.129	02:51	Volume Not Met
Breath Volume	0.460	02:51	
Air Blank	0.000	02:52	
Air Blank	0.000	02:54	
Subject Sample	# 2 0.128	02:57	Volume Not Met
Breath Volume	0.136	02:57	
Air Blank	0.000	02:58	
Control Test	0.079	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 03/02/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: NEW

DATE/TIME OF TEST: 02/14/2023 04:23:36

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.129	04:19
Breath Volume	2.757	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.127	04:21
Breath Volume	1.824	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 03/02/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 02/14/2023 05:22:36

Diagnostics Check	OK	05:13
Air Blank	0.000	05:14
Control Test	0.079	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.182	05:16
Breath Volume	1.167	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:18
Subject Sample # 2	0.182	05:20
Breath Volume	1.128	05:20
Air Blank	0.000	05:21
Control Test	0.079	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 03/02/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: DUONG

DATE/TIME OF TEST: 02/15/2023 03:55:40

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.184	03:50
Breath Volume	1.898	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.197	03:54
Breath Volume	2.675	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 03/02/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: GABBARD

DATE/TIME OF TEST: 02/15/2023 09:42:36

Diagnostics Check	OK	09:35
Air Blank	0.000	09:36
Control Test	0.080	09:36
Air Blank	0.000	09:36
Subject Sample # 1	0.072	09:37
Breath Volume	1.425	09:37
Air Blank	0.000	09:38
Air Blank	0.000	09:40
Subject Sample # 2	0.070	09:41
Breath Volume	1.226	09:41
Air Blank	0.000	09:41
Control Test	0.079	09:41
Air Blank	0.000	09:42
Diagnostics Check	OK	09:42

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: STEVENSON II

DATE/TIME OF TEST: 02/15/2023 22:15:29

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.192	22:10
Breath Volume	1.847	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.183	22:13
Breath Volume	1.355	22:13
Air Blank	0.000	22:14
Control Test	0.079	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: HAMMONS

DATE/TIME OF TEST: 02/16/2023 00:36:32

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:33	
Control Test	0.080	00:33	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 02/16/2023 02:26:21

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.119	02:21
Breath Volume	2.421	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.117	02:24
Breath Volume	2.210	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: GOMEZ RODRIGUEZ

DATE/TIME OF TEST: 02/16/2023 04:05:00

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.109	04:00
Breath Volume	2.199	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.110	04:03
Breath Volume	1.871	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:05

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 02/17/2023 00:12:11

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.079	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:05	No Sample Provided
Breath Volume	0.000	00:05	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample	# 2	00:10	No Sample Provided
Breath Volume	0.000	00:10	
Air Blank	0.000	00:11	
Control Test	0.079	00:11	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: NIETO SILVA

DATE/TIME OF TEST: 02/17/2023 04:03:29

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.079	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.080	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBERTS,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: BLEDSOE

DATE/TIME OF TEST: 02/17/2023 06:17:00

Diagnostics Check	OK	06:10
Air Blank	0.000	06:10
Control Test	0.080	06:10
Air Blank	0.000	06:11
Subject Sample # 1	0.063	06:11
Breath Volume	1.195	06:11
Air Blank	0.000	06:12
Air Blank	0.000	06:14
Subject Sample # 2	0.069	06:15
Breath Volume	1.753	06:15
Air Blank	0.000	06:16
Control Test	0.079	06:16
Air Blank	0.000	06:16
Diagnostics Check	OK	06:17

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOLMSTROM,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: BENITEZ-PEREZ

DATE/TIME OF TEST: 02/17/2023 23:23:28

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:19	
Control Test	0.079	23:19	
Air Blank	0.000	23:20	
Subject Sample	# 1	23:21	Subject Test Refused
Breath Volume	0.195	23:21	
Air Blank	0.000	23:22	
Control Test	0.080	23:22	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBERTS,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: TAWNEY

DATE/TIME OF TEST: 02/18/2023 02:54:17

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.184	02:49
Breath Volume	3.277	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.170	02:52
Breath Volume	3.093	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/18/2023 04:06:35

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.121	04:01
Breath Volume	1.308	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.139	04:04
Breath Volume	2.800	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/18/2023 04:48:31

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.099	04:43
Breath Volume	3.820	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.100	04:46
Breath Volume	3.046	04:46
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: MCCOMAS

DATE/TIME OF TEST: 02/19/2023 02:15:47

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.080	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:14	Subject Test Refused
Breath Volume	0.000	02:14	
Air Blank	0.000	02:14	
Control Test	0.080	02:15	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 02/19/2023 02:36:33

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.088	02:31
Breath Volume	2.843	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.084	02:34
Breath Volume	2.457	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: BARRETO

DATE/TIME OF TEST: 02/19/2023 05:00:13

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.166	04:55
Breath Volume	1.304	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.170	04:58
Breath Volume	2.011	04:58
Air Blank	0.000	04:59
Control Test	0.081	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LAMDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: OTANO STABLE

DATE/TIME OF TEST: 02/20/2023 02:19:21

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.183	02:14
Breath Volume	1.324	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.193	02:17
Breath Volume	1.566	02:17
Air Blank	0.000	02:18
Control Test	0.078	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: CELESTIN

DATE/TIME OF TEST: 02/20/2023 08:09:33

Diagnostics Check	OK	07:56	
Air Blank	0.000	07:57	
Control Test	0.080	07:57	
Air Blank	0.000	07:57	
Subject Sample	# 1 0.241	08:01	Volume Not Met
Breath Volume	0.621	08:01	
Air Blank	0.000	08:01	
Air Blank	0.000	08:03	
Subject Sample	# 2 0.259	08:04	
Breath Volume	1.308	08:04	
Air Blank	0.000	08:05	
Air Blank	0.000	08:07	
Subject Sample	# 3 0.254	08:07	
Breath Volume	2.101	08:07	
Air Blank	0.000	08:08	
Control Test	0.077	08:08	
Air Blank	0.000	08:09	
Diagnostics Check	OK	08:09	

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: SANFORD

DATE/TIME OF TEST: 02/20/2023 22:55:50

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.158	22:50
Breath Volume	2.214	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.143	22:54
Breath Volume	2.023	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 03/02/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: SULTENFUSS

DATE/TIME OF TEST: 02/21/2023 01:29:52

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.165	01:25
Breath Volume	2.328	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.169	01:28
Breath Volume	1.601	01:28
Air Blank	0.000	01:28
Control Test	0.079	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 02/22/2023 03:15:36

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:04	
Control Test	0.079	03:04	
Air Blank	0.000	03:05	
Subject Sample	# 1 0.139	03:08	Volume Not Met
Breath Volume	0.050	03:08	
Air Blank	0.000	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 2 0.162	03:13	Volume Not Met
Breath Volume	0.437	03:13	
Air Blank	0.000	03:14	
Control Test	0.077	03:14	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 02/22/2023 03:32:03

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.078	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1 0.167	03:22	Volume Not Met
Breath Volume	0.660	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 2 0.171	03:27	
Breath Volume	1.386	03:27	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample	# 3 0.169	03:30	
Breath Volume	1.492	03:30	
Air Blank	0.000	03:31	
Control Test	0.077	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: CONE

DATE/TIME OF TEST: 02/24/2023 22:46:16

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.079	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.115	22:40
Breath Volume	1.113	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:42
Subject Sample # 2	0.105	22:44
Breath Volume	1.914	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MAGAN

DATE/TIME OF TEST: 02/25/2023 04:37:49

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.080	04:26	
Air Blank	0.000	04:27	
Subject Sample	# 1 0.125	04:30	Volume Not Met
Breath Volume	0.214	04:30	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2 0.099	04:36	Volume Not Met
Breath Volume	0.113	04:36	
Air Blank	0.000	04:36	
Control Test	0.079	04:37	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MAGAN

DATE/TIME OF TEST: 02/25/2023 05:01:41

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:53	
Control Test	0.078	04:53	
Air Blank	0.000	04:54	
Subject Sample	# 1 0.103	04:57	Volume Not Met
Breath Volume	0.234	04:57	
Air Blank	0.000	04:58	
Air Blank	0.000	05:00	
Subject Sample	# 2	05:00	Subject Test Refused
Breath Volume	0.000	05:00	
Air Blank	0.000	05:00	
Control Test	0.078	05:01	
Air Blank	0.000	05:01	
Diagnostics Check	OK	05:01	

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: GAGNON

DATE/TIME OF TEST: 02/25/2023 05:50:26

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.078	05:44
Air Blank	0.000	05:45
Subject Sample # 1	0.000	05:46
Breath Volume	2.386	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.000	05:49
Breath Volume	2.320	05:49
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 02/25/2023 21:55:22

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.191	21:50
Breath Volume	1.238	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.188	21:53
Breath Volume	1.363	21:53
Air Blank	0.000	21:54
Control Test	0.080	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CANCANON DIAZ

DATE/TIME OF TEST: 02/26/2023 01:26:40

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.116	01:22
Breath Volume	3.035	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.119	01:25
Breath Volume	2.347	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBERTS,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: GIMESON

DATE/TIME OF TEST: 02/26/2023 06:25:31

Diagnostics Check	OK	06:18
Air Blank	0.000	06:19
Control Test	0.079	06:19
Air Blank	0.000	06:20
Subject Sample # 1	0.142	06:20
Breath Volume	1.554	06:20
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.140	06:23
Breath Volume	1.960	06:23
Air Blank	0.000	06:24
Control Test	0.079	06:24
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 02/26/2023 23:32:38

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.080	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1 0.205	23:25	Volume Not Met
Breath Volume	0.019	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 2 0.205	23:30	Volume Not Met
Breath Volume	0.871	23:30	
Air Blank	0.000	23:31	
Control Test	0.077	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BENAJIM

DATE/TIME OF TEST: 02/26/2023 23:46:00

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:34	
Control Test	0.077	23:34	
Air Blank	0.000	23:35	
Subject Sample	# 1 0.198	23:38	Volume Not Met
Breath Volume	0.683	23:38	
Air Blank	0.000	23:39	
Air Blank	0.000	23:41	
Subject Sample	# 2 0.195	23:44	Volume Not Met
Breath Volume	0.820	23:44	
Air Blank	0.000	23:45	
Control Test	0.077	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 02/27/2023 00:03:01

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.076	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 2 0.165	23:55	Volume Not Met
Breath Volume	0.343	23:55	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample	# 1 0.199	00:01	Volume Not Met
Breath Volume	0.113	00:01	
Air Blank	0.000	00:02	
Control Test	0.077	00:02	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: KABACK

DATE/TIME OF TEST: 02/27/2023 00:46:13

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.077	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.234	00:41
Breath Volume	3.031	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.229	00:44
Breath Volume	2.714	00:44
Air Blank	0.000	00:45
Control Test	0.077	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: WATSON JR

DATE/TIME OF TEST: 02/27/2023 22:02:44

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.079	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.164	21:58
Breath Volume	1.703	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.175	22:01
Breath Volume	1.933	22:01
Air Blank	0.000	22:01
Control Test	0.079	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: PINO

DATE/TIME OF TEST: 02/28/2023 21:38:16

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.079	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.149	21:33
Breath Volume	1.855	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.150	21:36
Breath Volume	1.847	21:36
Air Blank	0.000	21:37
Control Test	0.079	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 03/02/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 02/05/2023 06:42:57

Diagnostics Check	OK	06:36	
Air Blank	0.000	06:37	
Control Test	0.080	06:37	
Air Blank	0.000	06:38	
Subject Sample	# 1 0.095	06:38	
Breath Volume	1.199	06:38	
Air Blank	0.000	06:39	
Air Blank	0.000	06:41	
Subject Sample	# 2 0.098	06:42	
Breath Volume	1.386	06:42	
Air Blank		06:42	RFI Detect
Air Blank		06:42	RFI Detect

Data Download Date/Time: 02/16/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: DERVIL

DATE/TIME OF TEST: 02/07/2023 02:25:53

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.192	02:21
Breath Volume	2.667	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:23
Subject Sample # 2	0.196	02:24
Breath Volume	2.175	02:24
Air Blank	0.000	02:24
Control Test	0.076	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 02/16/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KETCHAM,JENNIFER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MCKEAN

DATE/TIME OF TEST: 02/08/2023 03:44:13

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:37	No Sample Provided
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample	# 2	03:42	No Sample Provided
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 02/16/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI, RYAN, P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALLEN, ALEXANDER
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: KEMMERER

DATE/TIME OF TEST: 02/10/2023 00:16:44

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.138	00:11
Breath Volume	3.343	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.136	00:15
Breath Volume	3.425	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 02/16/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: THORNBRUGH,E
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: EATMAN

DATE/TIME OF TEST: 02/11/2023 01:42:46

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:41	Subject Test Refused
Breath Volume	0.000	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:42	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 02/16/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: PREKOP

DATE/TIME OF TEST: 02/11/2023 02:32:22

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.203	02:27
Breath Volume	2.273	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:29
Subject Sample # 2	0.197	02:30
Breath Volume	1.707	02:30
Air Blank	0.000	02:31
Control Test	0.077	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 02/16/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: GRUBA,JACK
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BEREZNANIN

DATE/TIME OF TEST: 02/11/2023 03:38:16

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.079	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Control Test	0.081	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 02/16/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALBO,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TITKO,NICOLE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 02/13/2023 00:26:39

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.169	00:22
Breath Volume	1.746	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.175	00:24
Breath Volume	1.406	00:24
Air Blank	0.000	00:25
Control Test	0.077	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 02/16/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HESSE,JOSH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CARIDAD

DATE/TIME OF TEST: 02/14/2023 03:59:32

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.142	03:53
Breath Volume	1.757	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.141	03:57
Breath Volume	2.667	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 02/16/2023 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,KENNETH,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: KAADE

DATE/TIME OF TEST: 02/15/2023 17:33:54

Diagnostics Check	OK	17:22
Air Blank	0.000	17:23
Control Test	0.080	17:23
Air Blank	0.000	17:23
Subject Sample # 1	0.282	17:24
Breath Volume	2.546	17:24
Air Blank	0.000	17:25
Air Blank	0.000	17:27
Subject Sample # 2	0.256	17:27
Breath Volume	1.156	17:27
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 3	0.269	17:31
Breath Volume	2.054	17:31
Air Blank	0.000	17:32
Control Test	0.079	17:33
Air Blank	0.000	17:33
Diagnostics Check	OK	17:33

Data Download Date/Time: 02/16/2023 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: THORNBRUGH,E
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: CENTENO

DATE/TIME OF TEST: 02/16/2023 05:56:06

Diagnostics Check	OK	05:53	
Air Blank	0.000	05:53	
Control Test	0.080	05:53	
Air Blank	0.000	05:54	
Subject Sample	# 1	05:54	Subject Test Refused
Breath Volume	0.000	05:54	
Air Blank	0.000	05:55	
Control Test	0.080	05:55	
Air Blank	0.000	05:56	
Diagnostics Check	OK	05:56	

Data Download Date/Time: 02/16/2023 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI, RYAN, P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ROBERTS, ERIC
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: GRAZIANO

DATE/TIME OF TEST: 02/16/2023 19:10:41

Diagnostics Check	OK	19:07	
Air Blank	0.000	19:07	
Control Test	0.079	19:07	
Air Blank	0.000	19:08	
Subject Sample	# 1	19:09	Subject Test Refused
Breath Volume	0.000	19:09	
Air Blank	0.000	19:09	
Control Test	0.079	19:10	
Air Blank	0.000	19:10	
Diagnostics Check	OK	19:10	

Data Download Date/Time: 03/06/2023 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: QUAD,LUIS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 02/18/2023 00:56:48

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:54	
Control Test	0.080	00:54	
Air Blank	0.000	00:55	
Subject Sample	# 1	00:55	Subject Test Refused
Breath Volume	0.000	00:55	
Air Blank	0.000	00:55	
Control Test	0.080	00:56	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 03/06/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,J
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 02/18/2023 23:58:06

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.081	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.234	23:51
Breath Volume	1.882	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.246	23:55
Breath Volume	1.769	23:55
Air Blank	0.000	23:57
Control Test	0.076	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 03/06/2023 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,J
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 02/19/2023 02:45:27

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.080	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 03/06/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ZITO,ANTHONY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: VERTICHIO

DATE/TIME OF TEST: 02/21/2023 03:12:31

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.184	03:07
Breath Volume	2.695	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.177	03:10
Breath Volume	1.785	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 03/06/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: GRECU

DATE/TIME OF TEST: 02/22/2023 20:58:13

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:55	
Control Test	0.079	20:56	
Air Blank	0.000	20:56	
Subject Sample	# 1	20:56	Subject Test Refused
Breath Volume	0.000	20:56	
Air Blank	0.000	20:57	
Control Test	0.080	20:57	
Air Blank	0.000	20:58	
Diagnostics Check	OK	20:58	

Data Download Date/Time: 03/06/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PIKE,TREVOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: STADNYCKI

DATE/TIME OF TEST: 02/22/2023 21:16:35

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.081	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.018	21:12
Breath Volume	2.261	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.017	21:15
Breath Volume	2.953	21:15
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 03/06/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: KNOWLES

DATE/TIME OF TEST: 02/25/2023 02:03:50

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.080	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 03/06/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 02/25/2023 02:48:57

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.135	02:44
Breath Volume	3.269	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.131	02:47
Breath Volume	2.449	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 03/06/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALBO,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MULTY,T
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 02/25/2023 05:51:01

Diagnostics Check	OK	05:43
Air Blank	0.000	05:43
Control Test	0.080	05:43
Air Blank	0.000	05:44
Subject Sample # 1	0.176	05:45
Breath Volume	1.328	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.169	05:49
Breath Volume	1.660	05:49
Air Blank	0.000	05:50
Control Test	0.077	05:50
Air Blank	0.000	05:50
Diagnostics Check	OK	05:51

Data Download Date/Time: 03/06/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURR,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: DURR,KYLE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: HAMILL

DATE/TIME OF TEST: 02/28/2023 17:48:25

Diagnostics Check	OK	17:45	
Air Blank	0.000	17:45	
Control Test	0.080	17:46	
Air Blank	0.000	17:46	
Subject Sample	# 1	17:47	Subject Test Refused
Breath Volume	0.000	17:47	
Air Blank	0.000	17:47	
Control Test	0.080	17:47	
Air Blank	0.000	17:48	
Diagnostics Check	OK	17:48	

Data Download Date/Time: 03/06/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YETTER,JAMES,R
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: SPIRITO,GEORGE
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024022906
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: BUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: CORBETT

DATE/TIME OF TEST: 02/28/2023 18:01:21

Diagnostics Check	OK	17:54
Air Blank	0.000	17:54
Control Test	0.081	17:55
Air Blank	0.000	17:55
Subject Sample # 1	0.000	17:56
Breath Volume	1.171	17:56
Air Blank	0.000	17:57
Air Blank	0.000	17:59
Subject Sample # 2	0.000	17:59
Breath Volume	1.890	17:59
Air Blank	0.000	18:00
Control Test	0.078	18:00
Air Blank	0.000	18:01
Diagnostics Check	OK	18:01

Data Download Date/Time: 03/06/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FLANEGAN,BROOKE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 02/01/2023 02:38:04

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.000	02:33
Breath Volume	1.441	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.000	02:36
Breath Volume	1.898	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:38

Data Download Date/Time: 02/16/2023 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: POPE,NICOLAS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/04/2023 05:21:24

Diagnostics Check	OK	05:12
Air Blank	0.000	05:13
Control Test	0.080	05:13
Air Blank	0.000	05:14
Subject Sample # 1	0.232	05:15
Breath Volume	1.820	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:17
Subject Sample # 2	0.228	05:19
Breath Volume	1.136	05:19
Air Blank	0.000	05:20
Control Test	0.078	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 02/16/2023 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRERO,KEVEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 02/05/2023 22:28:30

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.081	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.159	22:23
Breath Volume	2.445	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.162	22:26
Breath Volume	2.273	22:26
Air Blank	0.000	22:27
Control Test	0.079	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 02/16/2023 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: POPE,NICHOLAS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: MARROQUIN-ESTRADA

DATE/TIME OF TEST: 02/06/2023 00:19:53

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.079	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.218	00:13
Breath Volume	1.101	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:15
Subject Sample # 2	0.207	00:17
Breath Volume	1.535	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 02/16/2023 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRERO,KEVEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: BRAVENDER

DATE/TIME OF TEST: 02/09/2023 02:52:06

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.080	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 02/16/2023 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,MEGHAN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: NIEMIEC,NICHOLAS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: SOBERG

DATE/TIME OF TEST: 02/12/2023 21:54:09

Diagnostics Check	OK	21:47
Air Blank	0.000	21:48
Control Test	0.081	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.128	21:49
Breath Volume	3.343	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.131	21:52
Breath Volume	3.558	21:52
Air Blank	0.000	21:53
Control Test	0.079	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 02/16/2023 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,THOMAS,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FOSTER,MEGHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/12/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CARUSO

DATE/TIME OF TEST: 02/13/2023 00:42:57

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.191	00:38
Breath Volume	1.519	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample # 2	0.202	00:41
Breath Volume	1.148	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 02/16/2023 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARTINEZ,KEVEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: TONSMEIRE

DATE/TIME OF TEST: 02/17/2023 20:53:31

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:51	
Control Test	0.080	20:51	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:52	Subject Test Refused
Breath Volume	0.000	20:52	
Air Blank	0.000	20:52	
Control Test	0.080	20:52	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

Data Download Date/Time: 03/07/2023 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEMIEC,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Key Colony Beach PD
ARRESTING OFFICER: FOSTER,MEGHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 02/27/2023 01:52:27

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:41	
Subject Sample	# 1 0.129	01:45	Volume Not Met
Breath Volume	0.136	01:45	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample	# 2 0.150	01:50	Volume Not Met
Breath Volume	0.398	01:50	
Air Blank	0.000	01:51	
Control Test	0.078	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 03/07/2023 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEMIEC,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Key Colony Beach PD
ARRESTING OFFICER: FOSTER,MEGHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 02/27/2023 02:11:09

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.078	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1 0.150	02:00	
Breath Volume	2.390	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:03	
Subject Sample	# 2 0.144	02:06	Volume Not Met
Breath Volume	0.000	02:06	
Air Blank	0.000	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 3 0.149	02:09	
Breath Volume	2.644	02:09	
Air Blank	0.000	02:10	
Control Test	0.077	02:10	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 03/07/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENEAU,ALISA,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WILLIAMS,RONNIE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 02/05/2023 16:16:46

Diagnostics Check	OK	16:09
Air Blank	0.000	16:10
Control Test	0.081	16:10
Air Blank	0.000	16:11
Subject Sample # 1	0.000	16:12
Breath Volume	1.457	16:12
Air Blank	0.000	16:12
Air Blank	0.000	16:14
Subject Sample # 2	0.000	16:15
Breath Volume	1.265	16:15
Air Blank	0.000	16:15
Control Test	0.080	16:16
Air Blank	0.000	16:16
Diagnostics Check	OK	16:16

Data Download Date/Time: 02/20/2023 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRUH,DYLAN,C
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: OSTHED,HENRIK
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 02/05/2023 22:37:02

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:34	
Control Test	0.081	22:34	
Air Blank	0.000	22:35	
Subject Sample	# 1	22:35	Subject Test Refused
Breath Volume	0.000	22:35	
Air Blank	0.000	22:36	
Control Test	0.080	22:36	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:37	

Data Download Date/Time: 02/20/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: BULLARD

DATE/TIME OF TEST: 02/06/2023 00:31:43

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.081	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.161	00:26
Breath Volume	1.289	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:28
Subject Sample # 2	0.141	00:29
Breath Volume	1.773	00:29
Air Blank	0.000	00:30
Control Test	0.078	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 02/20/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: CANTRELL,SEAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SHABER

DATE/TIME OF TEST: 02/07/2023 03:18:41

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.082	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.147	03:13
Breath Volume	1.835	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.149	03:16
Breath Volume	2.421	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 02/20/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETTS,J,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: MAYBERRY

DATE/TIME OF TEST: 02/09/2023 21:16:05

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.081	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.155	21:09
Breath Volume	1.242	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:11
Subject Sample # 2	0.139	21:14
Breath Volume	2.242	21:14
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 02/20/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELHO,ISAC,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SHUMAN,GARRET
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MCDUFFIE

DATE/TIME OF TEST: 02/09/2023 23:37:53

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:34	
Control Test	0.080	23:34	
Air Blank	0.000	23:35	
Subject Sample	# 1	23:37	Sequence Aborted
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	

Data Download Date/Time: 02/20/2023 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FENTON,CODY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FENTON,CODY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: MCGILL

DATE/TIME OF TEST: 02/10/2023 18:38:52

Diagnostics Check	OK	18:35	
Air Blank	0.000	18:36	
Control Test	0.080	18:36	
Air Blank	0.000	18:37	
Subject Sample	# 1	18:37	Subject Test Refused
Breath Volume	0.000	18:37	
Air Blank	0.000	18:37	
Control Test	0.080	18:38	
Air Blank	0.000	18:38	
Diagnostics Check	OK	18:38	

Data Download Date/Time: 02/20/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,F,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: DILEY,K
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: AGUILAR CASTILLO

DATE/TIME OF TEST: 02/11/2023 01:13:46

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:02	
Control Test	0.081	01:02	
Air Blank	0.000	01:02	
Subject Sample	# 1 0.136	01:03	
Breath Volume	3.171	01:03	
Air Blank	0.000	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 2	01:08	No Sample Provided
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample	# 3 0.150	01:11	
Breath Volume	5.945	01:11	
Air Blank	0.000	01:12	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 02/20/2023 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STILLER,CLINTON,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: STILLER

DATE/TIME OF TEST: 02/11/2023 17:47:10

Diagnostics Check	OK	17:44	
Air Blank	0.000	17:44	
Control Test	0.080	17:44	
Air Blank	0.000	17:45	
Subject Sample	# 1	17:45	Subject Test Refused
Breath Volume	0.000	17:45	
Air Blank	0.000	17:46	
Control Test	0.078	17:46	
Air Blank	0.000	17:47	
Diagnostics Check	OK	17:47	

Data Download Date/Time: 02/20/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEIGH,D,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WITT,W
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 02/11/2023 20:01:15

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:55	
Control Test	0.080	19:56	
Air Blank	0.000	19:56	
Subject Sample	# 1 0.226	19:59	Volume Not Met
Breath Volume	0.128	19:59	
Air Blank		20:00	Purge Fail
Air Blank	0.000	20:01	

Data Download Date/Time: 02/20/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEIGH,D,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WITT,W
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 02/11/2023 20:08:38

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:04	
Control Test	0.075	20:04	
Air Blank	0.000	20:05	
Subject Sample	# 1 0.255	20:07	
Breath Volume	1.925	20:07	
Air Blank		20:07	Purge Fail
Air Blank	0.000	20:08	

Data Download Date/Time: 02/20/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEIGH,D,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WITT,W
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 02/11/2023 20:12:51

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:11	
Control Test	0.072	20:12	Control Outside Tolerance
Air Blank	0.000	20:12	

Data Download Date/Time: 02/20/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEIGH,D,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WITT,W
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 02/11/2023 20:33:08

Diagnostics Check	OK	20:26
Air Blank	0.000	20:26
Control Test	0.077	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.253	20:28
Breath Volume	1.265	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:30
Subject Sample # 2	0.250	20:31
Breath Volume	1.804	20:31
Air Blank	0.000	20:32
Control Test	0.077	20:32
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 02/20/2023 15:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELHO,ISAC,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SHUMAN,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WILLINGHAM

DATE/TIME OF TEST: 02/14/2023 23:57:07

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.000	23:52
Breath Volume	2.285	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.000	23:55
Breath Volume	2.261	23:55
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 02/20/2023 15:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUTRAL,WILLIAM,Z
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FUTRAL,WILLIAM
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PARENT

DATE/TIME OF TEST: 02/15/2023 23:49:00

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.197	23:43
Breath Volume	1.742	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.196	23:47
Breath Volume	1.757	23:47
Air Blank	0.000	23:48
Control Test	0.078	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:49

Data Download Date/Time: 02/20/2023 15:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUTRAL,WILLIAM,Z
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FUTRAL,WILLIAM
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: THOMASON

DATE/TIME OF TEST: 02/16/2023 03:40:05

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.081	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:39	
Control Test	0.081	03:39	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 02/20/2023 15:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,ISAAC,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ECK,ISAAC
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: KELTY

DATE/TIME OF TEST: 02/17/2023 21:55:24

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.081	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.028	21:50
Breath Volume	1.632	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.028	21:53
Breath Volume	1.921	21:53
Air Blank	0.000	21:54
Control Test	0.080	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 02/20/2023 15:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: ORANJE

DATE/TIME OF TEST: 02/19/2023 07:12:59

Diagnostics Check	OK	07:05
Air Blank	0.000	07:06
Control Test	0.080	07:06
Air Blank	0.000	07:07
Subject Sample # 1	0.054	07:08
Breath Volume	1.781	07:08
Air Blank	0.000	07:08
Air Blank	0.000	07:10
Subject Sample # 2	0.052	07:11
Breath Volume	2.382	07:11
Air Blank	0.000	07:12
Control Test	0.079	07:12
Air Blank	0.000	07:12
Diagnostics Check	OK	07:12

Data Download Date/Time: 02/20/2023 15:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROPER, WILLIAM, D

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: ROPER, WILLIAM

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1346146

DRY GAS EXPIRY DATE: 07/07/2023

VIOLATION CODE: DUI

ARREST TIME: 18:49

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: CRABTREE

DATE/TIME OF TEST: 02/26/2023 19:32:29

Diagnostics Check	OK	19:25
Air Blank	0.000	19:26
Control Test	0.081	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.046	19:27
Breath Volume	2.480	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.046	19:30
Breath Volume	1.929	19:30
Air Blank	0.000	19:31
Control Test	0.081	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 03/09/2023 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: FORIS

DATE/TIME OF TEST: 02/03/2023 02:31:37

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.170	02:25
Breath Volume	1.246	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.166	02:29
Breath Volume	1.859	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 03/03/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CROOMS

DATE/TIME OF TEST: 02/04/2023 02:41:51

Diagnostics Check	Fail	02:41	Internal Printer Test
Air Blank	0.000	02:41	

Data Download Date/Time: 03/03/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CROOMS

DATE/TIME OF TEST: 02/04/2023 02:49:38

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.081	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.204	02:45
Breath Volume	2.824	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.205	02:48
Breath Volume	2.914	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 03/03/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SITARSKI

DATE/TIME OF TEST: 02/10/2023 23:57:49

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.079	23:50	
Air Blank	0.000	23:51	
Subject Sample # 1	0.192	23:52	
Breath Volume	2.050	23:52	
Air Blank	0.000	23:53	
Air Blank	0.000	23:54	
Subject Sample # 2	0.187	23:55	
Breath Volume	1.976	23:55	
Air Blank	0.000	23:56	
Control Test	0.077	23:56	
Air Blank	0.000	23:57	
Diagnostics Check	Fail	23:57	Analytical Stability Test
Air Blank	0.000	23:57	

Data Download Date/Time: 03/03/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MASTERS

DATE/TIME OF TEST: 02/12/2023 00:24:01

Diagnostics Check	Fail	00:23	Internal Printer Test
Air Blank	0.000	00:24	

Data Download Date/Time: 03/03/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 00:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MASTERS

DATE/TIME OF TEST: 02/12/2023 00:25:22

Diagnostics Check

Fail

00:24

Internal Printer Test

Air Blank

0.000

00:25

Data Download Date/Time: 03/03/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MASTERS

DATE/TIME OF TEST: 02/12/2023 00:33:56

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.113	00:29
Breath Volume	2.769	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:31
Subject Sample # 2	0.100	00:32
Breath Volume	1.714	00:32
Air Blank	0.000	00:32
Control Test	0.078	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 03/03/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/12/2023 02:21:57

Diagnostics Check	Fail	02:21	Internal Printer Test
Air Blank	0.000	02:21	

Data Download Date/Time: 03/03/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/12/2023 02:23:28

Diagnostics Check	Fail	02:23	Internal Printer Test
Air Blank	0.000	02:23	

Data Download Date/Time: 03/03/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/12/2023 02:31:13

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.158	02:26
Breath Volume	2.027	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.153	02:29
Breath Volume	1.710	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 03/03/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 23:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: STILLMUNKS

DATE/TIME OF TEST: 02/24/2023 23:17:32

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.079	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1 0.108	23:12	
Breath Volume	2.417	23:12	
Air Blank	0.000	23:13	
Air Blank	0.000	23:14	
Subject Sample	# 2 0.109	23:15	
Breath Volume	2.136	23:15	
Air Blank	0.000	23:16	
Control Test	0.080	23:16	
Air Blank	0.000	23:16	
Diagnostics Check	Fail	23:17	Internal Printer Test
Air Blank	0.000	23:17	

Data Download Date/Time: 03/03/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: STILLMUNKS

DATE/TIME OF TEST: 02/24/2023 23:19:27

Diagnostics Check	Fail	23:18	Internal Printer Test
Air Blank	0.000	23:19	

Data Download Date/Time: 03/03/2023 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 23:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: STILLMUNKS

DATE/TIME OF TEST: 02/24/2023 23:21:03

Diagnostics Check

Fail

23:20

Internal Printer Test

Air Blank

0.000

23:21

Data Download Date/Time: 03/03/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: STILLMUNKS

DATE/TIME OF TEST: 02/24/2023 23:23:07

Diagnostics Check	Fail	23:22	Internal Printer Test
Air Blank	0.000	23:23	

Data Download Date/Time: 03/03/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: STILLMUNKS

DATE/TIME OF TEST: 02/24/2023 23:30:22

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.102	23:25
Breath Volume	2.320	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.099	23:28
Breath Volume	2.015	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 03/03/2023 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 02:42

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LAVERY

DATE/TIME OF TEST: 02/26/2023 02:58:52

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:46	
Control Test	0.078	02:47	
Air Blank	0.000	02:47	
Subject Sample	# 1 0.166	02:48	
Breath Volume	2.156	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 2 0.149	02:53	Volume Not Met
Breath Volume	0.296	02:53	
Air Blank	0.000	02:54	
Air Blank	0.000	02:56	
Subject Sample	# 3 0.161	02:57	
Breath Volume	1.285	02:57	
Air Blank	0.000	02:57	
Control Test	0.079	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 03/03/2023 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: ULMER

DATE/TIME OF TEST: 02/06/2023 23:50:43

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.146	23:42
Breath Volume	1.628	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.173	23:46
Breath Volume	1.292	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 3	0.149	23:49
Breath Volume	2.175	23:49
Air Blank	0.000	23:49
Control Test	0.078	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 02/28/2023 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCALIA, ADAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: ROBLEDA VAZQUEZ

DATE/TIME OF TEST: 02/24/2023 23:54:32

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:47	No Sample Provided
Breath Volume	0.000	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:50	
Subject Sample	# 2	23:53	No Sample Provided
Breath Volume	0.000	23:53	
Air Blank	0.000	23:53	
Control Test	0.082	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 02/28/2023 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ, JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 02/25/2023 20:34:46

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.081	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.309	20:27
Breath Volume	3.632	20:27
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.277	20:29
Breath Volume	1.398	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 3	0.306	20:33
Breath Volume	3.433	20:33
Air Blank	0.000	20:33
Control Test	0.081	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 02/28/2023 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: JAROUSSE,DANIELLE	ARREST TIME: 02:56
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Female
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SZTEPA	DATE/TIME OF TEST: 02/02/2023	03:34:43
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.000	03:29
Breath Volume	1.164	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.000	03:33
Breath Volume	1.472	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 03/03/2023 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CORREA,JONATHAN,E

BREATH TEST OPERATOR AGENCY: St. Augustine PD

ARRESTING OFFICER: CORREA,JONATHAN

ARRESTING OFFICER AGENCY: St. Augustine PD

LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211543

DRY GAS EXPIRY DATE: 01/05/2024

VIOLATION CODE: DUI

ARREST TIME: 18:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: LEMIRE

DATE/TIME OF TEST: 02/06/2023 19:27:17

Diagnostics Check	OK	19:19
Air Blank	0.000	19:20
Control Test	0.078	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.105	19:22
Breath Volume	1.492	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.105	19:25
Breath Volume	1.449	19:25
Air Blank	0.000	19:26
Control Test	0.078	19:26
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 03/03/2023 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: PLATEROMALDONADO

DATE/TIME OF TEST: 02/08/2023 04:13:23

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.078	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.137	04:07
Breath Volume	1.687	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.134	04:11
Breath Volume	2.031	04:11
Air Blank	0.000	04:12
Control Test	0.078	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 03/03/2023 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RIGGINS,BRENDAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 02/09/2023 15:30:56

Diagnostics Check	OK	15:24
Air Blank	0.000	15:24
Control Test	0.079	15:24
Air Blank	0.000	15:25
Subject Sample # 1	0.275	15:26
Breath Volume	1.121	15:26
Air Blank	0.000	15:27
Air Blank	0.000	15:28
Subject Sample # 2	0.273	15:29
Breath Volume	1.121	15:29
Air Blank	0.000	15:30
Control Test	0.077	15:30
Air Blank	0.000	15:30
Diagnostics Check	OK	15:30

Data Download Date/Time: 03/03/2023 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: EMINISOR,ADAM
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 02/09/2023 18:35:12

Diagnostics Check	OK	18:23	
Air Blank	0.000	18:24	
Control Test	0.078	18:24	
Air Blank	0.000	18:24	
Subject Sample	# 1 0.083	18:27	Volume Not Met
Breath Volume	0.050	18:27	
Air Blank	0.000	18:28	
Air Blank	0.000	18:30	
Subject Sample	# 2 0.082	18:33	Volume Not Met
Breath Volume	0.015	18:33	
Air Blank	0.000	18:34	
Control Test	0.075	18:34	
Air Blank	0.000	18:35	
Diagnostics Check	OK	18:35	

Data Download Date/Time: 03/03/2023 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MCCANDLESS,MATTHEW	ARREST TIME: 15:36
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Female
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: GRAHAM	DATE/TIME OF TEST: 02/12/2023 16:27:09
-----------------------------	--

Diagnostics Check	OK	16:20
Air Blank	0.000	16:20
Control Test	0.079	16:21
Air Blank	0.000	16:21
Subject Sample # 1	0.118	16:22
Breath Volume	1.695	16:22
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.115	16:25
Breath Volume	2.089	16:25
Air Blank	0.000	16:26
Control Test	0.079	16:26
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

Data Download Date/Time: 03/03/2023 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: NORCROSS,AUSTIN	ARREST TIME: 19:38
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Female
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: VANBROCKLIN	DATE/TIME OF TEST: 02/12/2023 20:30:50
----------------------------------	--

Diagnostics Check	OK	20:23
Air Blank	0.000	20:23
Control Test	0.079	20:24
Air Blank	0.000	20:24
Subject Sample # 1	0.000	20:26
Breath Volume	1.527	20:26
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.000	20:29
Breath Volume	1.222	20:29
Air Blank	0.000	20:29
Control Test	0.079	20:30
Air Blank	0.000	20:30
Diagnostics Check	OK	20:30

Data Download Date/Time: 03/03/2023 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: JAROUSSE,DANIELLE	ARREST TIME: 23:05
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: O'DANIEL	DATE/TIME OF TEST: 02/12/2023	23:45:29
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.020	23:40
Breath Volume	2.875	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.019	23:44
Breath Volume	2.328	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 03/03/2023 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: FINNEY

DATE/TIME OF TEST: 02/16/2023 03:21:50

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.078	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.103	03:17
Breath Volume	2.437	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.111	03:20
Breath Volume	2.351	03:20
Air Blank	0.000	03:20
Control Test	0.078	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 03/03/2023 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: SCHEURICH

DATE/TIME OF TEST: 02/16/2023 22:25:02

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.078	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.083	22:20
Breath Volume	2.402	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.084	22:23
Breath Volume	1.691	22:23
Air Blank	0.000	22:24
Control Test	0.078	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:25

Data Download Date/Time: 03/03/2023 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: PARKER,D,J	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ETTEL,M	ARREST TIME: 20:41
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: MAY	DATE/TIME OF TEST: 02/18/2023	22:31:07
--------------------------	-------------------------------	----------

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.078	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.110	22:26
Breath Volume	4.425	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.112	22:29
Breath Volume	4.476	22:29
Air Blank	0.000	22:30
Control Test	0.076	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 03/03/2023 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,A
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 02/19/2023 03:18:01

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.202	03:13
Breath Volume	3.121	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.208	03:16
Breath Volume	4.527	03:16
Air Blank	0.000	03:17
Control Test	0.078	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:18

Data Download Date/Time: 03/03/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAPELA,JORDAN,A

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: CAPELA,JORDAN

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211543

DRY GAS EXPIRY DATE: 01/05/2024

VIOLATION CODE: DUI

ARREST TIME: 03:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 02/19/2023 05:01:32

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.078	04:56
Air Blank	0.000	04:56
Subject Sample # 1	0.000	04:57
Breath Volume	1.757	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.000	05:00
Breath Volume	1.789	05:00
Air Blank	0.000	05:00
Control Test	0.076	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 03/03/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNING,SEAN,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CANNING,SEAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 02/19/2023 20:53:42

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:47
Subject Sample # 1	0.404	20:48
Breath Volume	1.996	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.400	20:52
Breath Volume	1.820	20:52
Air Blank	0.000	20:52
Control Test	0.075	20:53
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 03/03/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HEARD,AALIYAH	ARREST TIME: 23:01
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: THURSTON	DATE/TIME OF TEST: 02/20/2023	23:51:26
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.079	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1 0.233	23:46	
Breath Volume	1.261	23:46	
Air Blank	0.000	23:47	
Air Blank	0.000	23:48	
Subject Sample	# 2	23:50	RFI Detect
Breath Volume	0.710	23:50	
Air Blank		23:50	Improper Sample
Air Blank	0.000	23:51	

Data Download Date/Time: 03/03/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JAROUSSE,DANIELLE,

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: HEARD,AALIYAH

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211543

DRY GAS EXPIRY DATE: 01/05/2024

VIOLATION CODE: DUI

ARREST TIME: 23:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: THURSTON

DATE/TIME OF TEST: 02/21/2023 00:03:35

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.077	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.216	23:57
Breath Volume	1.355	23:57
Air Blank	0.000	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.232	00:01
Breath Volume	1.718	00:01
Air Blank	0.000	00:02
Control Test	0.076	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 03/03/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PROUT,ANGELA	ARREST TIME: 12:12
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 12:46

SUBJECT'S LAST NAME: JONES	DATE/TIME OF TEST: 02/24/2023 13:13:13
----------------------------	--

Diagnostics Check	OK	13:06
Air Blank	0.000	13:06
Control Test	0.078	13:07
Air Blank	0.000	13:07
Subject Sample # 1	0.247	13:08
Breath Volume	1.976	13:08
Air Blank	0.000	13:08
Air Blank	0.000	13:10
Subject Sample # 2	0.247	13:11
Breath Volume	1.914	13:11
Air Blank	0.000	13:12
Control Test	0.078	13:12
Air Blank	0.000	13:13
Diagnostics Check	OK	13:13

Data Download Date/Time: 03/03/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: CANAAN,AUSTIN	ARREST TIME: 18:10
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: HEIL	DATE/TIME OF TEST: 02/24/2023	19:05:06
---------------------------	-------------------------------	----------

Diagnostics Check	OK	18:58
Air Blank	0.000	18:59
Control Test	0.078	18:59
Air Blank	0.000	18:59
Subject Sample # 1	0.000	19:00
Breath Volume	1.460	19:00
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample # 2	0.000	19:03
Breath Volume	1.214	19:03
Air Blank	0.000	19:04
Control Test	0.078	19:04
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

Data Download Date/Time: 03/03/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MELKERSON

DATE/TIME OF TEST: 02/24/2023 22:17:14

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.079	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.000	22:12
Breath Volume	2.343	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.000	22:15
Breath Volume	2.660	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 03/03/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 02/24/2023 23:50:43

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.078	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.113	23:46
Breath Volume	2.253	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.109	23:49
Breath Volume	1.859	23:49
Air Blank	0.000	23:49
Control Test	0.078	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 03/03/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARRILL,JOSHUA,D

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: THOMPSON,SEAN

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211543

DRY GAS EXPIRY DATE: 01/05/2024

VIOLATION CODE: DUI

ARREST TIME: 12:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 02/25/2023 13:56:34

Diagnostics Check	OK	13:50
Air Blank	0.000	13:50
Control Test	0.078	13:51
Air Blank	0.000	13:51
Subject Sample # 1	0.000	13:52
Breath Volume	2.960	13:52
Air Blank	0.000	13:52
Air Blank	0.000	13:54
Subject Sample # 2	0.000	13:55
Breath Volume	2.238	13:55
Air Blank	0.000	13:55
Control Test	0.079	13:55
Air Blank	0.000	13:56
Diagnostics Check	OK	13:56

Data Download Date/Time: 03/03/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MCCANDLESS,MATTHEW	ARREST TIME: 14:45
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 16:26

SUBJECT'S LAST NAME: STANLEY	DATE/TIME OF TEST: 02/26/2023	16:54:32
------------------------------	-------------------------------	----------

Diagnostics Check	OK	16:46
Air Blank	0.000	16:46
Control Test	0.078	16:47
Air Blank	0.000	16:47
Subject Sample # 1	0.000	16:48
Breath Volume	2.226	16:48
Air Blank	0.000	16:48
Air Blank	0.000	16:50
Subject Sample # 2	0.000	16:53
Breath Volume	2.250	16:53
Air Blank	0.000	16:53
Control Test	0.080	16:53
Air Blank	0.000	16:54
Diagnostics Check	OK	16:54

Data Download Date/Time: 03/03/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: AYUSO

DATE/TIME OF TEST: 02/06/2023 00:27:45

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.082	00:22	
Air Blank	0.000	00:23	
Subject Sample	# 1	00:25	Subject Test Refused
Breath Volume	0.109	00:25	
Air Blank	0.000	00:26	
Control Test	0.083	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 02/27/2023 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,W
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 02/23/2023 06:13:04

Diagnostics Check	OK	06:09	
Air Blank	0.000	06:10	
Control Test	0.081	06:10	
Air Blank	0.000	06:10	
Subject Sample	# 1	06:11	Subject Test Refused
Breath Volume	0.000	06:11	
Air Blank	0.000	06:11	
Control Test	0.081	06:11	
Air Blank	0.000	06:12	
Diagnostics Check	Fail	06:12	Internal Printer Test
Air Blank	0.000	06:13	

Data Download Date/Time: 05/12/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,ANGELA,F
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WILLIAMS,KYLE
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402469736
DRY GAS EXPIRY DATE: 06/19/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: JERNIGAN

DATE/TIME OF TEST: 02/25/2023 02:57:37

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.081	02:55	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Control Test	0.082	02:56	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 05/12/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARX,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: GRIFFIN,BRADLEY
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: VASALLO

DATE/TIME OF TEST: 02/26/2023 01:07:58

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.108	01:02
Breath Volume	1.312	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.117	01:06
Breath Volume	1.593	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 05/12/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELSOMINO,MATTHEW,J
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: GELSOMINO,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: ESTEP

DATE/TIME OF TEST: 02/18/2023 21:00:44

Diagnostics Check	Fail	21:00	Internal Printer Test
Air Blank	0.000	21:00	

Data Download Date/Time: 03/28/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELSOMINO,MATTHEW,J
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: GELSOMINO,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: ESTEP

DATE/TIME OF TEST: 02/18/2023 21:15:36

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:04	
Control Test	0.079	21:04	
Air Blank	0.000	21:04	
Subject Sample	# 1 0.104	21:08	Volume Not Met
Breath Volume	0.000	21:08	
Air Blank	0.000	21:09	
Air Blank	0.000	21:10	
Subject Sample	# 2 0.109	21:13	Volume Not Met
Breath Volume	0.035	21:13	
Air Blank	0.000	21:14	
Control Test	0.082	21:14	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:15	

Data Download Date/Time: 03/28/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: MILLS,SCOTT,T	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: MOROZOV,SERGEY	ARREST TIME: 05:03
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: CARLEY	DATE/TIME OF TEST: 02/02/2023 05:28:29
-----------------------------	--

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.082	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.000	05:23
Breath Volume	2.769	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.000	05:27
Breath Volume	2.789	05:27
Air Blank	0.000	05:27
Control Test	0.081	05:27
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 02/15/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: MILLS,SCOTT,T	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: MILLS,SCOTT	ARREST TIME: 21:20
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Female
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: ZEMANEK	DATE/TIME OF TEST: 02/03/2023 21:35:52
------------------------------	--

Diagnostics Check	Fail	21:35	Internal Printer Test
Air Blank	0.000	21:35	

Data Download Date/Time: 02/15/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: MILLS,SCOTT,T	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: MILLS,SCOTT	ARREST TIME: 21:20
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Female
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: ZEMANEK	DATE/TIME OF TEST: 02/03/2023 21:44:34
------------------------------	--

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.082	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.100	21:39
Breath Volume	3.324	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.092	21:42
Breath Volume	2.175	21:42
Air Blank	0.000	21:43
Control Test	0.081	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 02/15/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: DICRISTOFANO

DATE/TIME OF TEST: 02/07/2023 00:06:01

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.083	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.069	00:01
Breath Volume	2.550	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:03
Subject Sample # 2	0.066	00:04
Breath Volume	2.460	00:04
Air Blank	0.000	00:05
Control Test	0.082	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:06

Data Download Date/Time: 02/15/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: MILLS,SCNOTT,T	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: MILLS,SCNOTT	ARREST TIME: 01:03
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Female
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BULLARD FETRGUSOCN	DATE/TIME OF TEST: 02/12/2023 01:14:57
---	--

Diagnostics Check	Fail	01:14	Internal Printer Test
Air Blank	0.000	01:14	

Data Download Date/Time: 02/15/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: MILLS,SNCOTT,T	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: MILLS,SCOTT	ARREST TIME: 01:03
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Female
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BULLARDFERGUSON	DATE/TIME OF TEST: 02/12/2023 01:18:00
--------------------------------------	--

Diagnostics Check	Fail	01:17	Internal Printer Test
Air Blank	0.000	01:18	

Data Download Date/Time: 02/15/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: MILLS,SCOTT,T	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: MILLS,SCOTT	ARREST TIME: 01:03
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Female
LAST AGENCY INSPECTION: 01/25/2023	OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BULLARD FTERGUSOCN	DATE/TIME OF TEST: 02/12/2023 01:27:22
---	--

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.082	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.194	01:22
Breath Volume	1.296	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.182	01:25
Breath Volume	2.406	01:25
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 02/15/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402231696
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 02/18/2023 00:40:31

Diagnostics Check	Fail	00:40	Internal Printer Test
Air Blank	0.000	00:40	

Data Download Date/Time: 02/22/2023 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402231696
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 02/18/2023 00:42:25

Diagnostics Check	Fail	00:41	Internal Printer Test
Air Blank	0.000	00:42	

Data Download Date/Time: 02/22/2023 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ALEX,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: COSME,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 02/16/2023 01:32:52

Diagnostics Check	Fail	01:32	Internal Printer Test
Air Blank	0.000	01:32	

Data Download Date/Time: 02/27/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ALEX,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: COSME,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 02/16/2023 01:44:37

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.082	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:38
Breath Volume	1.351	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.000	01:43
Breath Volume	1.613	01:43
Air Blank	0.000	01:43
Control Test	0.081	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 02/27/2023 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ALEX,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HOUSE,STEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: LERMA

DATE/TIME OF TEST: 02/19/2023 04:04:54

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.083	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:03	
Control Test	0.081	04:04	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 02/27/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ALEX,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/20/2023 02:39:23

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.083	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.151	02:34
Breath Volume	2.414	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.134	02:37
Breath Volume	1.507	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 02/27/2023 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: DIVVER

DATE/TIME OF TEST: 02/15/2023 21:07:26

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:54	
Control Test	0.079	20:54	
Air Blank	0.000	20:55	
Subject Sample	# 1 0.122	20:58	Volume Not Met
Breath Volume	0.687	20:58	
Air Blank	0.000	20:58	
Air Blank	0.000	21:00	
Subject Sample	# 2 0.123	21:01	
Breath Volume	1.359	21:01	
Air Blank	0.000	21:02	
Air Blank	0.000	21:04	
Subject Sample	# 3 0.119	21:05	
Breath Volume	1.292	21:05	
Air Blank	0.000	21:06	
Control Test	0.080	21:06	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time: 03/02/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: FAHRNER

DATE/TIME OF TEST: 02/18/2023 04:14:27

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.093	04:09
Breath Volume	1.773	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.094	04:12
Breath Volume	1.691	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 03/02/2023 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: EQUILUZ ALFARO

DATE/TIME OF TEST: 02/24/2023 02:52:24

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:49	
Control Test	0.079	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.079	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 03/02/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: BARTOLO BARTOLO

DATE/TIME OF TEST: 02/18/2023 00:13:54

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.082	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.181	00:08
Breath Volume	1.640	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.174	00:12
Breath Volume	1.257	00:12
Air Blank	0.000	00:12
Control Test	0.082	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 03/02/2023 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: FIGUEROA GRACIA

DATE/TIME OF TEST: 02/18/2023 03:13:51

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.083	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.208	03:07
Breath Volume	1.519	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.207	03:12
Breath Volume	1.554	03:12
Air Blank	0.000	03:12
Control Test	0.083	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 03/02/2023 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: HEARD II

DATE/TIME OF TEST: 02/18/2023 05:16:43

Diagnostics Check	OK	05:15	
Air Blank	0.000	05:15	
Control Test	0.087	05:16	Interferent Detect
Air Blank	0.000	05:16	

Data Download Date/Time: 03/02/2023 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: HEARD II

DATE/TIME OF TEST: 02/18/2023 05:25:52

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.084	05:20
Air Blank	0.000	05:20
Subject Sample # 1	0.000	05:21
Breath Volume	3.339	05:21
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.000	05:24
Breath Volume	2.949	05:24
Air Blank	0.000	05:24
Control Test	0.082	05:25
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 03/02/2023 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: CARMICHAEL

DATE/TIME OF TEST: 02/28/2023 22:18:50

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.081	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.198	22:13
Breath Volume	1.480	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:15
Subject Sample # 2	0.191	22:17
Breath Volume	1.335	22:17
Air Blank	0.000	22:17
Control Test	0.080	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 03/02/2023 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA, MARCANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HERNANDEZ SHILON

DATE/TIME OF TEST: 02/16/2023 23:33:24

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:25	
Control Test	0.078	23:25	
Air Blank	0.000	23:26	
Subject Sample	# 1	23:29	No Sample Provided
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Air Blank	0.000	23:31	
Subject Sample	# 2	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:32	
Control Test	0.078	23:32	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:33	

Data Download Date/Time: 03/02/2023 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO, MATTHEW
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: NOSS

DATE/TIME OF TEST: 02/17/2023 00:36:00

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.077	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.152	00:29
Breath Volume	1.855	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:31
Subject Sample # 2	0.142	00:34
Breath Volume	1.648	00:34
Air Blank	0.000	00:35
Control Test	0.077	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:36

Data Download Date/Time: 03/02/2023 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: REYES CASTILLO

DATE/TIME OF TEST: 02/26/2023 03:49:01

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.077	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.106	03:44
Breath Volume	2.535	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.104	03:47
Breath Volume	2.316	03:47
Air Blank	0.000	03:48
Control Test	0.078	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:49

Data Download Date/Time: 03/02/2023 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: ROMAGUERA REYES

DATE/TIME OF TEST: 02/26/2023 22:21:53

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.078	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.111	22:17
Breath Volume	1.984	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.103	22:20
Breath Volume	1.515	22:20
Air Blank	0.000	22:20
Control Test	0.078	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 03/02/2023 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFFMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: HOFFMAN,JOSHUA
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 01/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 02/13/2023 00:24:22

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.172	00:19
Breath Volume	1.484	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.184	00:22
Breath Volume	1.941	00:22
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 03/02/2023 22:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SACKETT

DATE/TIME OF TEST: 02/18/2023 03:37:03

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.119	03:32
Breath Volume	1.460	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.116	03:35
Breath Volume	1.511	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

Data Download Date/Time: 03/02/2023 22:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HYLTON,ROBERT
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 01/26/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ADORNO CLAVIJO

DATE/TIME OF TEST: 02/18/2023 21:47:54

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.080	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.236	21:43
Breath Volume	1.933	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:45
Subject Sample # 2	0.240	21:46
Breath Volume	1.800	21:46
Air Blank	0.000	21:46
Control Test	0.078	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 03/02/2023 22:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFFMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 02/25/2023 04:03:00

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.066	03:58
Breath Volume	1.886	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.064	04:01
Breath Volume	1.953	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:03

Data Download Date/Time: 03/02/2023 22:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DETTMAN,AMANDA
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: HESSELINK

DATE/TIME OF TEST: 02/20/2023 04:06:18

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.080	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.081	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 02/24/2023 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,ANTHONY,K
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: TOMBERLIN,K
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:57

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 02/01/2023 11:26:04

Diagnostics Check	OK	11:18
Air Blank	0.000	11:18
Control Test	0.081	11:18
Air Blank	0.000	11:19
Subject Sample # 1	0.093	11:20
Breath Volume	1.656	11:20
Air Blank	0.000	11:21
Air Blank	0.000	11:23
Subject Sample # 2	0.084	11:24
Breath Volume	1.421	11:24
Air Blank	0.000	11:25
Control Test	0.080	11:25
Air Blank	0.000	11:25
Diagnostics Check	OK	11:26

Data Download Date/Time: 02/24/2023 20:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN, CHAZ,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN, CHAZ
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 14:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:29

SUBJECT'S LAST NAME: KOLEK

DATE/TIME OF TEST: 02/02/2023 14:58:22

Diagnostics Check	OK	14:51
Air Blank	0.000	14:51
Control Test	0.081	14:52
Air Blank	0.000	14:52
Subject Sample # 1	0.111	14:53
Breath Volume	3.503	14:53
Air Blank	0.000	14:54
Air Blank	0.000	14:55
Subject Sample # 2	0.106	14:56
Breath Volume	3.464	14:56
Air Blank	0.000	14:57
Control Test	0.079	14:57
Air Blank	0.000	14:58
Diagnostics Check	OK	14:58

Data Download Date/Time: 02/24/2023 20:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: TOMBERLIN, K
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:42

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 02/07/2023 11:08:47

Diagnostics Check	OK	11:02
Air Blank	0.000	11:02
Control Test	0.081	11:03
Air Blank	0.000	11:03
Subject Sample # 1	0.000	11:04
Breath Volume	1.191	11:04
Air Blank	0.000	11:04
Air Blank	0.000	11:06
Subject Sample # 2	0.000	11:07
Breath Volume	2.324	11:07
Air Blank	0.000	11:07
Control Test	0.080	11:08
Air Blank	0.000	11:08
Diagnostics Check	OK	11:08

Data Download Date/Time: 02/24/2023 20:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,U
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,A
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: GLUCK

DATE/TIME OF TEST: 02/10/2023 23:02:52

Diagnostics Check	OK	22:59	
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:01	Subject Test Refused
Breath Volume	0.000	23:01	
Air Blank	0.000	23:01	
Control Test	0.080	23:02	
Air Blank	0.000	23:02	
Diagnostics Check	OK	23:02	

Data Download Date/Time: 02/24/2023 20:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,DEPUTY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: PERALTA

DATE/TIME OF TEST: 02/11/2023 21:42:37

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:40	
Control Test	0.081	21:40	
Air Blank	0.000	21:41	
Subject Sample	# 1	21:41	Subject Test Refused
Breath Volume	0.000	21:41	
Air Blank	0.000	21:41	
Control Test	0.080	21:41	
Air Blank	0.000	21:42	
Diagnostics Check	OK	21:42	

Data Download Date/Time: 02/24/2023 20:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SERGEYEV,PS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 02/12/2023 01:10:05

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Control Test	0.080	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 02/24/2023 20:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WALKER,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 02/14/2023 04:08:21

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.082	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.000	04:03
Breath Volume	1.167	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.000	04:06
Breath Volume	1.171	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 02/24/2023 20:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,J,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GONZALEZ,SANDRA
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: KOLEK

DATE/TIME OF TEST: 02/17/2023 16:40:46

Diagnostics Check	OK	16:34
Air Blank	0.000	16:34
Control Test	0.082	16:34
Air Blank	0.000	16:35
Subject Sample # 1	0.032	16:36
Breath Volume	3.292	16:36
Air Blank	0.000	16:36
Air Blank	0.000	16:38
Subject Sample # 2	0.027	16:39
Breath Volume	2.746	16:39
Air Blank	0.000	16:39
Control Test	0.081	16:40
Air Blank	0.000	16:40
Diagnostics Check	OK	16:40

Data Download Date/Time: 02/24/2023 20:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,J,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PECK,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 12:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: GREGIS

DATE/TIME OF TEST: 02/23/2023 13:17:04

Diagnostics Check	OK	13:16	
Air Blank		13:16	RFI Detect
Air Blank	0.000	13:17	

Data Download Date/Time: 02/24/2023 20:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,J,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PECK,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 12:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: GREGIS

DATE/TIME OF TEST: 02/23/2023 13:25:26

Diagnostics Check	OK	13:18
Air Blank	0.000	13:19
Control Test	0.081	13:19
Air Blank	0.000	13:20
Subject Sample # 1	0.000	13:20
Breath Volume	1.960	13:20
Air Blank	0.000	13:21
Air Blank	0.000	13:23
Subject Sample # 2	0.000	13:24
Breath Volume	1.441	13:24
Air Blank	0.000	13:24
Control Test	0.081	13:24
Air Blank	0.000	13:25
Diagnostics Check	OK	13:25

Data Download Date/Time: 02/24/2023 20:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,DEPUTY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 02/24/2023 23:06:57

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.081	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.000	23:02
Breath Volume	3.335	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.000	23:05
Breath Volume	3.480	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 03/24/2023 01:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,CALEB,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BARR,GARRETT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BIVINS

DATE/TIME OF TEST: 02/25/2023 02:54:12

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.081	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.080	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 03/24/2023 01:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,C,H
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PANTOJA,Y
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/25/2023 21:26:53

Diagnostics Check	OK	21:14	
Air Blank	0.000	21:15	
Control Test	0.080	21:15	
Air Blank	0.000	21:15	
Subject Sample	# 1 0.249	21:17	
Breath Volume	1.386	21:17	
Air Blank	0.000	21:17	
Air Blank	0.000	21:19	
Subject Sample	# 2 0.227	21:22	Volume Not Met
Breath Volume	0.019	21:22	
Air Blank	0.000	21:23	
Air Blank	0.000	21:25	
Subject Sample	# 3	21:25	Subject Test Refused
Breath Volume	0.000	21:25	
Air Blank	0.000	21:25	
Control Test	0.076	21:26	
Air Blank	0.000	21:26	
Diagnostics Check	OK	21:26	

Data Download Date/Time: 03/24/2023 01:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: SCHUM,MELONIE
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LITCHFIELD

DATE/TIME OF TEST: 02/05/2023 03:04:43

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.052	02:59
Breath Volume	1.160	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.050	03:03
Breath Volume	1.234	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 02/09/2023 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BAKER,DUSTIN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 02/18/2023 00:11:58

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.000	00:07
Breath Volume	1.878	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.000	00:10
Breath Volume	1.890	00:10
Air Blank	0.000	00:11
Control Test	0.079	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 03/30/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAQUETTE,AARON,Y
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 02/22/2023 01:55:46

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.253	01:48
Breath Volume	2.894	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:50
Subject Sample # 2	0.229	01:51
Breath Volume	1.734	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample # 3	0.236	01:54
Breath Volume	2.574	01:54
Air Blank	0.000	01:54
Control Test	0.079	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 03/30/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,A
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/04/2023 00:32:18

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.224	00:27
Breath Volume	1.925	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.223	00:30
Breath Volume	3.769	00:30
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 03/03/2023 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 02/04/2023 19:01:32

Diagnostics Check	Fail	19:01	Internal Printer Test
Air Blank	0.000	19:01	

Data Download Date/Time: 03/03/2023 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 02/04/2023 19:13:53

Diagnostics Check	OK	19:02	
Air Blank	0.000	19:03	
Control Test	0.078	19:03	
Air Blank	0.000	19:04	
Subject Sample	# 1 0.211	19:07	Volume Not Met
Breath Volume	0.015	19:07	
Air Blank	0.000	19:08	
Air Blank	0.000	19:09	
Subject Sample	# 2	19:12	Subject Test Refused
Breath Volume	0.601	19:12	
Air Blank	0.000	19:12	
Control Test	0.075	19:13	
Air Blank	0.000	19:13	
Diagnostics Check	OK	19:13	

Data Download Date/Time: 03/03/2023 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: PARADISO

DATE/TIME OF TEST: 02/07/2023 22:53:58

Diagnostics Check	Fail	22:53	Internal Printer Test
Air Blank	0.000	22:53	

Data Download Date/Time: 03/03/2023 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: PARADISO

DATE/TIME OF TEST: 02/07/2023 23:02:57

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.078	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.126	22:58
Breath Volume	1.492	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.117	23:01
Breath Volume	1.597	23:01
Air Blank	0.000	23:02
Control Test	0.077	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 03/03/2023 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 02/09/2023 01:00:47

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.078	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.159	00:55
Breath Volume	2.714	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.158	00:59
Breath Volume	2.464	00:59
Air Blank	0.000	00:59
Control Test	0.077	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 03/03/2023 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAHNSEN,A,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: HUGHES,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: SOPOTNICK

DATE/TIME OF TEST: 02/28/2023 16:21:50

Diagnostics Check	OK	16:15
Air Blank	0.000	16:15
Control Test	0.078	16:15
Air Blank	0.000	16:16
Subject Sample # 1	0.307	16:16
Breath Volume	1.277	16:16
Air Blank	0.000	16:17
Air Blank	0.000	16:19
Subject Sample # 2	0.320	16:20
Breath Volume	2.187	16:20
Air Blank	0.000	16:20
Control Test	0.077	16:21
Air Blank	0.000	16:21
Diagnostics Check	OK	16:21

Data Download Date/Time: 03/03/2023 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ,ALBERTO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RUBIO,R

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 01/18/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 06:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: ALFARO CHAVES

DATE/TIME OF TEST: 02/16/2023 07:19:00

Diagnostics Check	OK	07:16	
Air Blank	0.000	07:16	
Control Test	0.082	07:16	
Air Blank	0.000	07:17	
Subject Sample	# 1	07:17	Subject Test Refused
Breath Volume	0.000	07:17	
Air Blank	0.000	07:18	
Control Test	0.081	07:18	
Air Blank	0.000	07:18	
Diagnostics Check	OK	07:19	

Data Download Date/Time: 02/16/2023 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 05:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: ORTEGA RODRIGUEZ

DATE/TIME OF TEST: 02/17/2023 06:57:34

Diagnostics Check	OK	06:49
Air Blank	0.000	06:49
Control Test	0.081	06:50
Air Blank	0.000	06:50
Subject Sample # 1	0.173	06:52
Breath Volume	1.316	06:52
Air Blank	0.000	06:53
Air Blank	0.000	06:55
Subject Sample # 2	0.179	06:55
Breath Volume	2.105	06:55
Air Blank	0.000	06:56
Control Test	0.079	06:56
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

Data Download Date/Time: 05/05/2023 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,GUILLERMO,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GUERRERO,RAFAEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 08:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: ANDINO

DATE/TIME OF TEST: 02/18/2023 10:07:53

Diagnostics Check	OK	10:04	
Air Blank	0.000	10:05	
Control Test	0.081	10:05	
Air Blank	0.000	10:05	
Subject Sample	# 1	10:06	Subject Test Refused
Breath Volume	0.000	10:06	
Air Blank	0.000	10:06	
Control Test	0.081	10:07	
Air Blank	0.000	10:07	
Diagnostics Check	OK	10:07	

Data Download Date/Time: 05/05/2023 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BERMUDEZ,FERNANDO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MENDEZ,M

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 20:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: MARTINEZ CAMPO

DATE/TIME OF TEST: 02/21/2023 21:09:11

Diagnostics Check	OK	21:06	
Air Blank	0.000	21:06	
Control Test	0.082	21:07	
Air Blank	0.000	21:07	
Subject Sample	# 1	21:07	Subject Test Refused
Breath Volume	0.000	21:07	
Air Blank	0.000	21:08	
Control Test	0.080	21:08	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

Data Download Date/Time: 05/05/2023 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARTIN,ALEJANDRO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: SOMARRIBA,YADER

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 12:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:47

SUBJECT'S LAST NAME: PEREZ AREAS

DATE/TIME OF TEST: 02/24/2023 13:10:28

Diagnostics Check	OK	13:07	
Air Blank	0.000	13:08	
Control Test	0.082	13:08	
Air Blank	0.000	13:08	
Subject Sample	# 1	13:09	Subject Test Refused
Breath Volume	0.000	13:09	
Air Blank	0.000	13:09	
Control Test	0.083	13:09	
Air Blank	0.000	13:10	
Diagnostics Check	OK	13:10	

Data Download Date/Time: 05/05/2023 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO	ARREST TIME: 02:23
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Female
LAST AGENCY INSPECTION: 02/16/2023	OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: VALLADARES DATE/TIME OF TEST: 02/25/2023 03:10:21

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.082	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.111	03:05
Breath Volume	1.683	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.117	03:08
Breath Volume	1.222	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 05/05/2023 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: WHITT,JESSE,T	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)	VIOLATION CODE: BUI
ARRESTING OFFICER: BEGERA,JERRY	ARREST TIME: 13:54
ARRESTING OFFICER AGENCY: FFWCC (West Palm)	GENDER: Male
LAST AGENCY INSPECTION: 02/16/2023	OBSERVATION PERIOD BEGAN: 13:54

SUBJECT'S LAST NAME: MACHADO DATE/TIME OF TEST: 02/26/2023 18:06:39

Diagnostics Check	OK	18:00	
Air Blank	0.000	18:01	
Control Test	0.081	18:01	
Air Blank	0.000	18:02	
Subject Sample	# 1 0.093	18:05	Volume Not Met
Breath Volume	0.273	18:05	
Air Blank		18:06	Purge Fail
Air Blank	0.000	18:06	

Data Download Date/Time: 05/05/2023 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WHITT,JESSE,T

BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)

ARRESTING OFFICER: WHITT,JESSE

ARRESTING OFFICER AGENCY: FFWCC (West Palm)

LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: BUI

ARREST TIME: 13:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 13:54

SUBJECT'S LAST NAME: MACHADO

DATE/TIME OF TEST: 02/26/2023 18:14:28

Diagnostics Check	OK	18:11	
Air Blank	0.000	18:12	
Control Test	0.079	18:12	
Air Blank	0.000	18:12	
Subject Sample	# 1	18:13	Subject Test Refused
Breath Volume	0.000	18:13	
Air Blank	0.000	18:13	
Control Test	0.081	18:13	
Air Blank	0.000	18:14	
Diagnostics Check	OK	18:14	

Data Download Date/Time: 05/05/2023 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GUTIERREZ,DANIEL,F

BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)

ARRESTING OFFICER: GUTIERREZ,DANIEL

ARRESTING OFFICER AGENCY: FFWCC (West Palm)

LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: BUI

ARREST TIME: 17:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/26/2023 18:56:37

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:54	
Control Test	0.081	18:54	
Air Blank	0.000	18:55	
Subject Sample	# 1	18:55	Subject Test Refused
Breath Volume	0.000	18:55	
Air Blank	0.000	18:55	
Control Test	0.080	18:55	
Air Blank	0.000	18:56	
Diagnostics Check	OK	18:56	

Data Download Date/Time: 05/05/2023 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ANTHONY,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CRUZ,ANTHONY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SALAZAR TRINIDAD

DATE/TIME OF TEST: 02/26/2023 21:59:44

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.082	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.214	21:54
Breath Volume	1.765	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:56
Subject Sample # 2	0.203	21:57
Breath Volume	1.656	21:57
Air Blank	0.000	21:58
Control Test	0.079	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 05/05/2023 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CAMILA,P
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,CAMILA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: AGUILLON

DATE/TIME OF TEST: 02/18/2023 04:27:50

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.195	04:20
Breath Volume	2.863	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.167	04:23
Breath Volume	2.101	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 3	0.170	04:26
Breath Volume	1.148	04:26
Air Blank	0.000	04:26
Control Test	0.077	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 03/14/2023 17:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO	ARREST TIME: 20:18
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Female
LAST AGENCY INSPECTION: 02/16/2023	OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: BLANCO LOPEZ	DATE/TIME OF TEST: 02/21/2023 21:09:38
-----------------------------------	--

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:00	
Control Test	0.078	21:01	
Air Blank	0.000	21:01	
Subject Sample	# 1	21:04	No Sample Provided
Breath Volume	0.000	21:04	
Air Blank	0.000	21:05	
Air Blank	0.000	21:07	
Subject Sample	# 2	21:08	Subject Test Refused
Breath Volume	0.000	21:08	
Air Blank	0.000	21:08	
Control Test	0.076	21:09	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

Data Download Date/Time: 03/14/2023 17:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,CAMILA,P

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CABRERA,ROGER

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 04:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: CHAMORRO FRANCO

DATE/TIME OF TEST: 02/25/2023 05:17:58

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.079	05:12
Air Blank	0.000	05:12
Subject Sample # 1	0.178	05:13
Breath Volume	1.710	05:13
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.183	05:16
Breath Volume	2.937	05:16
Air Blank	0.000	05:17
Control Test	0.077	05:17
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 03/14/2023 17:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ACOSTA,CAROLINA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: LEIVA

DATE/TIME OF TEST: 02/25/2023 06:48:56

Diagnostics Check	OK	06:42
Air Blank	0.000	06:42
Control Test	0.079	06:42
Air Blank	0.000	06:43
Subject Sample # 1	0.207	06:44
Breath Volume	1.601	06:44
Air Blank	0.000	06:44
Air Blank	0.000	06:46
Subject Sample # 2	0.208	06:47
Breath Volume	1.640	06:47
Air Blank	0.000	06:48
Control Test	0.078	06:48
Air Blank	0.000	06:48
Diagnostics Check	OK	06:48

Data Download Date/Time: 03/14/2023 17:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARBALLO,ANGEL,O
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MARTINEZ RAMOS,JORGE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 02/25/2023 22:13:06

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.078	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.119	22:08
Breath Volume	2.308	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:10
Subject Sample # 2	0.131	22:11
Breath Volume	2.238	22:11
Air Blank	0.000	22:12
Control Test	0.078	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 03/14/2023 17:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCHE,EDEL,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ROSADO,ANGEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: BUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: BROCHE

DATE/TIME OF TEST: 02/26/2023 22:46:12

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:43	
Control Test	0.079	22:44	
Air Blank	0.000	22:44	
Subject Sample	# 1	22:44	Subject Test Refused
Breath Volume	0.000	22:44	
Air Blank	0.000	22:45	
Control Test	0.079	22:45	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 03/14/2023 17:00