

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,ALEXANDRIA,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GONZALEZ,KAITLIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 03/02/2023 11:13:01

Diagnostics Check	OK	11:10	
Air Blank	0.000	11:10	
Control Test	0.081	11:10	
Air Blank	0.000	11:11	
Subject Sample	# 1	11:11	Subject Test Refused
Breath Volume	0.000	11:11	
Air Blank	0.000	11:12	
Control Test	0.082	11:12	
Air Blank	0.000	11:12	
Diagnostics Check	OK	11:13	

Data Download Date/Time: 03/14/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUTTERLY,BROOKE,L
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: RICO,MATEO
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 03/11/2023 02:56:26

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.082	02:54	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.082	02:55	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 03/14/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUGH,BRYAN,W
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: PONAS,CLIFFORD
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 03/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/20/2023 22:06:06

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.079	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.205	22:00
Breath Volume	2.316	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.200	22:04
Breath Volume	2.070	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 04/14/2023 01:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,ROBERT,R
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: DENNIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: THAMES

DATE/TIME OF TEST: 03/10/2023 02:05:18

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.081	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.179	02:00
Breath Volume	1.847	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.175	02:03
Breath Volume	2.070	02:03
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 03/20/2023 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,ROBERT,R
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEWIS,OLAJUWON
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: POOLE

DATE/TIME OF TEST: 03/11/2023 23:02:59

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.214	22:57
Breath Volume	1.386	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.215	23:01
Breath Volume	1.562	23:01
Air Blank	0.000	23:02
Control Test	0.081	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 03/20/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAMOS,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 03/12/2023 22:49:32

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.080	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.188	22:44
Breath Volume	2.933	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.178	22:48
Breath Volume	3.015	22:48
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 03/20/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSTICE,JOSEPH,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JUSTICE,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: MAULDEN

DATE/TIME OF TEST: 03/02/2023 13:28:10

Diagnostics Check	OK	13:21
Air Blank	0.000	13:22
Control Test	0.080	13:22
Air Blank	0.000	13:23
Subject Sample # 1	0.000	13:23
Breath Volume	4.160	13:23
Air Blank	0.000	13:24
Air Blank	0.000	13:26
Subject Sample # 2	0.000	13:26
Breath Volume	3.589	13:26
Air Blank	0.000	13:27
Control Test	0.078	13:27
Air Blank	0.000	13:28
Diagnostics Check	OK	13:28

Data Download Date/Time: 03/20/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENRIGHT,PAUL,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WINFREE,SAIGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: ENRIGHT

DATE/TIME OF TEST: 03/02/2023 21:00:33

Diagnostics Check	OK	20:53
Air Blank	0.000	20:54
Control Test	0.079	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.200	20:56
Breath Volume	2.292	20:56
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.191	20:58
Breath Volume	1.582	20:58
Air Blank	0.000	20:59
Control Test	0.078	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 03/20/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,R
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HOLLOWAY,HAGEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: WESTRICK

DATE/TIME OF TEST: 03/02/2023 23:48:24

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.204	23:43
Breath Volume	1.367	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.198	23:46
Breath Volume	1.539	23:46
Air Blank	0.000	23:47
Control Test	0.078	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 03/20/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:54

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 03/04/2023 08:21:15

Diagnostics Check	OK	08:14
Air Blank	0.000	08:14
Control Test	0.080	08:15
Air Blank	0.000	08:15
Subject Sample # 1	0.132	08:16
Breath Volume	3.007	08:16
Air Blank	0.000	08:17
Air Blank	0.000	08:19
Subject Sample # 2	0.127	08:19
Breath Volume	2.511	08:19
Air Blank	0.000	08:20
Control Test	0.077	08:20
Air Blank	0.000	08:21
Diagnostics Check	OK	08:21

Data Download Date/Time: 03/20/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 07:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: HINOTE

DATE/TIME OF TEST: 03/08/2023 08:47:13

Diagnostics Check	OK	08:40
Air Blank	0.000	08:41
Control Test	0.080	08:41
Air Blank	0.000	08:42
Subject Sample # 1	0.189	08:42
Breath Volume	2.078	08:42
Air Blank	0.000	08:43
Air Blank	0.000	08:45
Subject Sample # 2	0.188	08:45
Breath Volume	3.351	08:45
Air Blank	0.000	08:46
Control Test	0.078	08:46
Air Blank	0.000	08:47
Diagnostics Check	OK	08:47

Data Download Date/Time: 03/20/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,SEAN,C
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LONG,SEAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ILLING-VINSON

DATE/TIME OF TEST: 03/12/2023 23:17:40

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.171	23:12
Breath Volume	3.363	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.160	23:15
Breath Volume	1.820	23:15
Air Blank	0.000	23:16
Control Test	0.078	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 03/20/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: FLAHERTY,ZACHARY
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: BIDNICK

DATE/TIME OF TEST: 03/16/2023 03:47:28

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:40	No Sample Provided
Breath Volume	0.000	03:40	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 2	03:46	No Sample Provided
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Control Test	0.081	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 04/20/2023 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,JORDI,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BOHANNON,M
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MATIAS PEREZ

DATE/TIME OF TEST: 03/16/2023 23:24:59

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.079	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.211	23:20
Breath Volume	3.132	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.213	23:23
Breath Volume	2.886	23:23
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 04/20/2023 12:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,DILLON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: TILLMAN,JOE
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 03/17/2023 17:33:23

Diagnostics Check	OK	17:20	
Air Blank	0.000	17:21	
Control Test	0.080	17:21	
Air Blank	0.000	17:21	
Subject Sample	# 1 0.234	17:24	Volume Not Met
Breath Volume	0.117	17:24	
Air Blank	0.000	17:25	
Air Blank	0.000	17:27	
Subject Sample	# 2 0.300	17:28	No .020 agreement
Breath Volume	1.394	17:28	
Air Blank	0.000	17:29	
Air Blank	0.000	17:30	
Subject Sample	# 3 0.274	17:31	No .020 agreement
Breath Volume	1.660	17:31	
Air Blank	0.000	17:32	
Control Test	0.079	17:32	
Air Blank	0.000	17:33	
Diagnostics Check	OK	17:33	

Data Download Date/Time: 04/20/2023 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,DILLON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: TILLMAN,JOE
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 03/17/2023 17:44:48

Diagnostics Check	OK	17:37
Air Blank	0.000	17:37
Control Test	0.079	17:38
Air Blank	0.000	17:38
Subject Sample # 1	0.294	17:39
Breath Volume	1.363	17:39
Air Blank	0.000	17:40
Air Blank	0.000	17:42
Subject Sample # 2	0.294	17:43
Breath Volume	1.246	17:43
Air Blank	0.000	17:43
Control Test	0.078	17:44
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

Data Download Date/Time: 04/20/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DECZYNSKI,JAMES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: IVEY

DATE/TIME OF TEST: 03/18/2023 02:43:54

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.119	02:39
Breath Volume	1.644	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample # 2	0.116	02:42
Breath Volume	2.015	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 04/20/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CARLTON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CARLTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 03/20/2023 19:44:22

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.079	19:34
Air Blank	0.000	19:35
Subject Sample # 1	0.308	19:36
Breath Volume	3.015	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:38
Subject Sample # 2	0.287	19:39
Breath Volume	2.109	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:41
Subject Sample # 3	0.278	19:42
Breath Volume	2.917	19:42
Air Blank	0.000	19:43
Control Test	0.077	19:43
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 04/20/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,JOSHUA,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILSON,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: REYES RODRIGUEZ

DATE/TIME OF TEST: 03/22/2023 03:41:53

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.081	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.259	03:37
Breath Volume	3.675	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.255	03:40
Breath Volume	3.558	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 04/20/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: SNIDER,RYAN
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: ENFINGER

DATE/TIME OF TEST: 03/25/2023 03:41:08

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.080	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:34	No Sample Provided
Breath Volume	0.000	03:34	
Air Blank	0.000	03:34	
Air Blank	0.000	03:36	
Subject Sample	# 2	03:39	No Sample Provided
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Control Test	0.081	03:40	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 04/20/2023 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: CONTRERAS,JAVIER
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: AKPLOGAN FIHO

DATE/TIME OF TEST: 03/26/2023 23:54:05

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:38	
Control Test	0.080	23:38	
Air Blank	0.000	23:39	
Subject Sample	# 1 0.161	23:42	Volume Not Met
Breath Volume	0.824	23:42	
Air Blank	0.000	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 2 0.179	23:46	
Breath Volume	2.980	23:46	
Air Blank	0.000	23:47	
Air Blank	0.000	23:49	
Subject Sample	# 3 0.160	23:52	Volume Not Met
Breath Volume	0.914	23:52	
Air Blank	0.000	23:53	
Control Test	0.080	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 04/20/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: CONTRERAS,JAVIER
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: AKPLOGAN FIHO

DATE/TIME OF TEST: 03/27/2023 00:19:29

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.172	00:14
Breath Volume	1.886	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.182	00:17
Breath Volume	2.605	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 04/20/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SALTER,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: MORIN

DATE/TIME OF TEST: 03/29/2023 03:53:49

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.240	03:49
Breath Volume	4.890	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.226	03:52
Breath Volume	3.664	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 04/20/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAIG,ANYCE,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: YORK,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:24

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 03/04/2023 07:53:21

Diagnostics Check	OK	07:46
Air Blank	0.000	07:47
Control Test	0.080	07:47
Air Blank	0.000	07:47
Subject Sample # 1	0.163	07:48
Breath Volume	3.687	07:48
Air Blank	0.000	07:49
Air Blank	0.000	07:51
Subject Sample # 2	0.162	07:51
Breath Volume	3.746	07:51
Air Blank	0.000	07:52
Control Test	0.078	07:52
Air Blank	0.000	07:53
Diagnostics Check	OK	07:53

Data Download Date/Time: 03/31/2023 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 03/05/2023 02:29:22

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.126	02:24
Breath Volume	1.613	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.134	02:27
Breath Volume	1.578	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 03/31/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: ALTILIO

DATE/TIME OF TEST: 03/12/2023 21:30:17

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.080	21:24
Air Blank	0.000	21:24
Subject Sample # 1	0.092	21:25
Breath Volume	2.390	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.085	21:28
Breath Volume	1.945	21:28
Air Blank	0.000	21:29
Control Test	0.078	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 03/31/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: HORN

DATE/TIME OF TEST: 03/16/2023 05:19:56

Diagnostics Check	OK	05:13
Air Blank	0.000	05:13
Control Test	0.080	05:13
Air Blank	0.000	05:14
Subject Sample # 1	0.169	05:14
Breath Volume	1.687	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.173	05:18
Breath Volume	1.621	05:18
Air Blank	0.000	05:19
Control Test	0.080	05:19
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 03/31/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMATEER,TRACY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PALMATEER,TRACY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 03/17/2023 01:45:23

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.079	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1 0.087	01:40	
Breath Volume	3.500	01:40	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample	# 2 0.083	01:43	
Breath Volume	3.160	01:43	
Air Blank	0.000	01:44	
Control Test	0.078	01:44	
Air Blank		01:44	RFI Detect
Air Blank	0.000	01:45	

Data Download Date/Time: 03/31/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEYES,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GEORGE,AMANDA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: CRADDOCK

DATE/TIME OF TEST: 03/31/2023 00:18:12

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.079	00:12	
Air Blank	0.000	00:13	
Subject Sample	# 1 0.209	00:13	
Breath Volume	2.722	00:13	
Air Blank	0.000	00:14	
Air Blank	0.000	00:16	
Subject Sample	# 2	00:16	RFI Detect
Breath Volume	0.539	00:16	
Air Blank	0.000	00:17	
Control Test	0.077	00:17	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 03/31/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELAN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PHELAN,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 03/29/2023 02:26:12

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.169	02:21
Breath Volume	1.113	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.181	02:24
Breath Volume	1.769	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 03/30/2023 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDEN,RYAN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GOLDEN,RYAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: CRAWLEY

DATE/TIME OF TEST: 03/05/2023 18:58:49

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.083	18:53
Air Blank	0.000	18:53
Subject Sample # 1	0.000	18:54
Breath Volume	4.394	18:54
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.000	18:57
Breath Volume	4.796	18:57
Air Blank	0.000	18:57
Control Test	0.081	18:58
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 04/07/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRILL,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MERRILL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 09:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:13

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 03/08/2023 10:41:17

Diagnostics Check	OK	10:34
Air Blank	0.000	10:34
Control Test	0.082	10:35
Air Blank	0.000	10:35
Subject Sample # 1	0.000	10:36
Breath Volume	3.039	10:36
Air Blank	0.000	10:37
Air Blank	0.000	10:39
Subject Sample # 2	0.000	10:39
Breath Volume	3.464	10:39
Air Blank	0.000	10:40
Control Test	0.081	10:40
Air Blank	0.000	10:41
Diagnostics Check	OK	10:41

Data Download Date/Time: 04/07/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: WILDHABER

DATE/TIME OF TEST: 03/09/2023 02:30:29

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.082	02:27	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Control Test	0.081	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 04/07/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JAMES,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LOPEZ,BRIAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 03/10/2023 01:49:15

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.082	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:42	No Sample Provided
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample	# 2	01:47	No Sample Provided
Breath Volume	0.000	01:47	
Air Blank	0.000	01:48	
Control Test	0.082	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 04/07/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:24

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 03/10/2023 14:54:48

Diagnostics Check	OK	14:48
Air Blank	0.000	14:48
Control Test	0.082	14:49
Air Blank	0.000	14:49
Subject Sample # 1	0.000	14:50
Breath Volume	2.445	14:50
Air Blank	0.000	14:50
Air Blank	0.000	14:52
Subject Sample # 2	0.000	14:53
Breath Volume	2.484	14:53
Air Blank	0.000	14:53
Control Test	0.081	14:54
Air Blank	0.000	14:54
Diagnostics Check	OK	14:54

Data Download Date/Time: 04/07/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FOSTER,COLE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: ST CLAIR

DATE/TIME OF TEST: 03/10/2023 22:56:07

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.083	22:42	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:46	No Sample Provided
Breath Volume	0.000	22:46	
Air Blank	0.000	22:46	
Air Blank	0.000	22:48	
Subject Sample	# 2	0.164	22:51
Breath Volume	2.656	22:51	
Air Blank	0.000	22:51	
Air Blank	0.000	22:53	
Subject Sample	# 3	0.168	22:54
Breath Volume	2.601	22:54	
Air Blank	0.000	22:55	
Control Test	0.081	22:55	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 04/07/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEISANSAL,HARMON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: VARDEMAN

DATE/TIME OF TEST: 03/12/2023 00:50:03

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.159	00:45
Breath Volume	1.539	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.151	00:48
Breath Volume	1.355	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:50

Data Download Date/Time: 04/07/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: LUCIANO

DATE/TIME OF TEST: 03/12/2023 01:17:18

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.082	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.338	01:10
Breath Volume	1.261	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.349	01:15
Breath Volume	2.660	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 04/07/2023 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/12/2023 20:34:51

Diagnostics Check	OK	20:27
Air Blank	0.000	20:27
Control Test	0.082	20:28
Air Blank	0.000	20:28
Subject Sample # 1	0.174	20:30
Breath Volume	3.074	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:32
Subject Sample # 2	0.180	20:33
Breath Volume	2.117	20:33
Air Blank	0.000	20:33
Control Test	0.080	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 04/07/2023 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEISANSAL,HARMON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: TREMBLAY

DATE/TIME OF TEST: 03/12/2023 22:15:31

Diagnostics Check	OK	22:07
Air Blank	0.000	22:07
Control Test	0.081	22:08
Air Blank	0.000	22:08
Subject Sample # 1	0.257	22:10
Breath Volume	1.953	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.244	22:13
Breath Volume	1.875	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 04/07/2023 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELENZ,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ELENZ,TIMOTHY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/13/2023 01:16:09

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.081	01:05	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:09	No Sample Provided
Breath Volume	0.000	01:09	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample	# 2	01:14	No Sample Provided
Breath Volume	0.000	01:14	
Air Blank	0.000	01:15	
Control Test	0.081	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 04/07/2023 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/18/2023 02:40:55

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.083	02:37	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Control Test	0.082	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 04/07/2023 08:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIEDEMANN,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TIEDEMANN,NICHOLAS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: MOSES JR

DATE/TIME OF TEST: 03/18/2023 18:29:38

Diagnostics Check	OK	18:18	
Air Blank	0.000	18:18	
Control Test	0.082	18:19	
Air Blank	0.000	18:19	
Subject Sample	# 1	18:22	No Sample Provided
Breath Volume	0.000	18:22	
Air Blank	0.000	18:23	
Air Blank	0.000	18:25	
Subject Sample	# 2	18:28	No Sample Provided
Breath Volume	0.000	18:28	
Air Blank	0.000	18:28	
Control Test	0.081	18:29	
Air Blank	0.000	18:29	
Diagnostics Check	OK	18:29	

Data Download Date/Time: 04/07/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPLING,JEFF,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RICHMOND,DONALD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: ASHCRAFT

DATE/TIME OF TEST: 03/19/2023 19:57:58

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.083	19:52
Air Blank	0.000	19:52
Subject Sample # 1	0.000	19:53
Breath Volume	2.214	19:53
Air Blank	0.000	19:53
Air Blank	0.000	19:55
Subject Sample # 2	0.000	19:56
Breath Volume	1.292	19:56
Air Blank	0.000	19:57
Control Test	0.084	19:57
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 04/07/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,BOBBY,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARRISON,BOBBY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MINK

DATE/TIME OF TEST: 03/19/2023 22:23:40

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.084	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.272	22:19
Breath Volume	2.316	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.272	22:21
Breath Volume	2.191	22:21
Air Blank	0.000	22:22
Control Test	0.082	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 04/07/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 03/20/2023 04:01:52

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.083	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.118	03:56
Breath Volume	1.976	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.120	04:00
Breath Volume	1.667	04:00
Air Blank	0.000	04:00
Control Test	0.084	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 04/07/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: STERLING

DATE/TIME OF TEST: 03/23/2023 03:00:40

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.082	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.076	02:55
Breath Volume	1.687	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.071	02:59
Breath Volume	1.121	02:59
Air Blank	0.000	02:59
Control Test	0.081	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 04/07/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/24/2023 07:05:31

Diagnostics Check	OK	06:57
Air Blank	0.000	06:58
Control Test	0.082	06:58
Air Blank	0.000	06:58
Subject Sample # 1	0.189	07:00
Breath Volume	1.390	07:00
Air Blank	0.000	07:01
Air Blank	0.000	07:03
Subject Sample # 2	0.184	07:03
Breath Volume	1.542	07:03
Air Blank	0.000	07:04
Control Test	0.081	07:04
Air Blank	0.000	07:05
Diagnostics Check	OK	07:05

Data Download Date/Time: 04/07/2023 08:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WHITE,TRAVIS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/25/2023 03:46:12

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.080	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Control Test	0.081	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 04/07/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: GAINES

DATE/TIME OF TEST: 03/25/2023 03:54:41

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.103	03:50
Breath Volume	1.222	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.099	03:53
Breath Volume	1.488	03:53
Air Blank	0.000	03:53
Control Test	0.081	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 04/07/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: TATUM

DATE/TIME OF TEST: 03/26/2023 01:35:53

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.082	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.104	01:31
Breath Volume	1.382	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.100	01:34
Breath Volume	1.343	01:34
Air Blank	0.000	01:34
Control Test	0.081	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 04/07/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BRUCE,HEATHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 03/26/2023 21:30:41

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.083	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.000	21:26
Breath Volume	1.101	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.000	21:29
Breath Volume	1.566	21:29
Air Blank	0.000	21:29
Control Test	0.081	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 04/07/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAHL,RANDY,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAHL,RANDY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 03/27/2023 03:25:04

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.083	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.000	03:20
Breath Volume	6.386	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.000	03:23
Breath Volume	5.656	03:23
Air Blank	0.000	03:24
Control Test	0.082	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:25

Data Download Date/Time: 04/07/2023 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WHITE,TRAVIS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 03/29/2023 04:47:16

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:44	
Control Test	0.081	04:44	
Air Blank	0.000	04:45	
Subject Sample	# 1	04:45	Subject Test Refused
Breath Volume	0.000	04:45	
Air Blank	0.000	04:46	
Control Test	0.082	04:46	
Air Blank	0.000	04:47	
Diagnostics Check	OK	04:47	

Data Download Date/Time: 04/07/2023 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 03/29/2023 05:32:42

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.081	05:27
Air Blank	0.000	05:27
Subject Sample # 1	0.171	05:28
Breath Volume	2.218	05:28
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.178	05:31
Breath Volume	2.285	05:31
Air Blank	0.000	05:31
Control Test	0.082	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 04/07/2023 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: STOVER

DATE/TIME OF TEST: 03/29/2023 21:42:06

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.082	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.039	21:37
Breath Volume	1.253	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.039	21:40
Breath Volume	2.203	21:40
Air Blank	0.000	21:41
Control Test	0.082	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 04/07/2023 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODDO,ISAIAH,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COOLEY,LOGAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: MCLEOD

DATE/TIME OF TEST: 03/31/2023 19:55:45

Diagnostics Check	OK	19:49
Air Blank	0.000	19:50
Control Test	0.081	19:50
Air Blank	0.000	19:50
Subject Sample # 1	0.214	19:51
Breath Volume	1.941	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:53
Subject Sample # 2	0.199	19:54
Breath Volume	2.781	19:54
Air Blank	0.000	19:54
Control Test	0.080	19:55
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 04/07/2023 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSER,KEN,L
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: MOSER,KEN
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 03/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 007
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: 260512714561

DATE/TIME OF TEST: 03/27/2023 16:15:45

Diagnostics Check	OK	16:12	
Air Blank	0.000	16:13	
Control Test	0.079	16:13	
Air Blank	0.000	16:14	
Subject Sample	# 1	16:14	Subject Test Refused
Breath Volume	0.000	16:14	
Air Blank	0.000	16:14	
Control Test	0.079	16:15	
Air Blank	0.000	16:15	
Diagnostics Check	OK	16:15	

Data Download Date/Time: 04/11/2023 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGRETTE,JOAN,P
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: NEGRETTE,JOAN
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/11/2023 23:49:34

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:48	
Control Test	0.080	23:48	
Air Blank		23:49	RFI Detect
Air Blank	0.000	23:49	

Data Download Date/Time: 03/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGRETTE,JOAN,P
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: NEGRETTE,JOAN
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/11/2023 23:57:43

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.264	23:52
Breath Volume	1.464	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.245	23:56
Breath Volume	1.246	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 03/23/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PIND

DATE/TIME OF TEST: 03/03/2023 04:09:39

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:07	
Control Test	0.077	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:08	Subject Test Refused
Breath Volume	0.000	04:08	
Air Blank	0.000	04:08	
Control Test	0.078	04:09	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 03/27/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/04/2023 03:48:45

Diagnostics Check	OK	03:47	
Air Blank		03:48	RFI Detect
Air Blank	0.000	03:48	

Data Download Date/Time: 03/27/2023 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/04/2023 03:53:45

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.077	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:52	RFI Detect
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.078	03:53	
Air Blank		03:53	RFI Detect
Air Blank	0.000	03:53	

Data Download Date/Time: 03/27/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/04/2023 03:57:45

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.077	03:56	
Air Blank		03:57	RFI Detect
Air Blank	0.000	03:57	

Data Download Date/Time: 03/27/2023 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/04/2023 04:10:03

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.077	04:02	
Air Blank	0.000	04:03	
Subject Sample # 1	0.158	04:04	
Breath Volume	2.050	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:07	
Subject Sample # 2	0.173	04:08	
Breath Volume	1.628	04:08	
Air Blank	0.000	04:09	
Control Test	0.074	04:09	Control Outside Tolerance
Air Blank	0.000	04:10	

Data Download Date/Time: 03/27/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: ROWSEY

DATE/TIME OF TEST: 03/07/2023 06:36:15

Diagnostics Check	OK	06:33	
Air Blank	0.000	06:33	
Control Test	0.078	06:34	
Air Blank	0.000	06:34	
Subject Sample	# 1	06:34	Subject Test Refused
Breath Volume	0.000	06:34	
Air Blank	0.000	06:35	
Control Test	0.079	06:35	
Air Blank	0.000	06:36	
Diagnostics Check	OK	06:36	

Data Download Date/Time: 03/27/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: MCGREW

DATE/TIME OF TEST: 03/08/2023 20:21:36

Diagnostics Check	OK	20:18	
Air Blank	0.000	20:19	
Control Test	0.078	20:19	
Air Blank	0.000	20:19	
Subject Sample	# 1	20:20	Subject Test Refused
Breath Volume	0.000	20:20	
Air Blank	0.000	20:20	
Control Test	0.079	20:21	
Air Blank	0.000	20:21	
Diagnostics Check	OK	20:21	

Data Download Date/Time: 03/27/2023 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRITTANY,N
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: SWEARS,AARON
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 03/15/2023 06:07:58

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:05	
Control Test	0.079	06:06	
Air Blank	0.000	06:06	
Subject Sample	# 1	06:07	Subject Test Refused
Breath Volume	0.000	06:07	
Air Blank		06:07	RFI Detect
Air Blank	0.000	06:07	

Data Download Date/Time: 03/27/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/24/2023 01:50:44

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.077	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.080	01:46
Breath Volume	2.417	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.086	01:49
Breath Volume	2.312	01:49
Air Blank	0.000	01:49
Control Test	0.078	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 03/27/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: JONES,T
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: JANIGAN

DATE/TIME OF TEST: 03/25/2023 00:37:16

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.224	00:32
Breath Volume	2.257	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.214	00:35
Breath Volume	2.703	00:35
Air Blank	0.000	00:36
Control Test	0.077	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 03/27/2023 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN,JASON,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LORD,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 03/30/2023 22:10:51

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.078	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.139	22:06
Breath Volume	1.960	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:08
Subject Sample # 2	0.149	22:09
Breath Volume	3.234	22:09
Air Blank	0.000	22:09
Control Test	0.078	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 05/04/2023 08:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PASLEY,JACOB,W

BREATH TEST OPERATOR AGENCY: Washington County SO

ARRESTING OFFICER: LEE,TYLER

ARRESTING OFFICER AGENCY: Washington County SO

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 3024023281

DRY GAS EXPIRY DATE: 01/20/2025

VIOLATION CODE: DUI

ARREST TIME: 08:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/24/2023 08:44:30

Diagnostics Check	OK	08:40	
Air Blank	0.000	08:41	
Control Test	0.082	08:41	
Air Blank	0.000	08:42	
Subject Sample	# 1	08:43	Subject Test Refused
Breath Volume	0.000	08:43	
Air Blank	0.000	08:43	
Control Test	0.082	08:43	
Air Blank	0.000	08:44	
Diagnostics Check	OK	08:44	

Data Download Date/Time: 05/01/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 3024023281
BREATH TEST OPERATOR: PASLEY,JACOB,W	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: Washington County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PASLEY,JACOB	ARREST TIME: 12:56
ARRESTING OFFICER AGENCY: Washington County SO	GENDER: Female
LAST AGENCY INSPECTION: 03/14/2023	OBSERVATION PERIOD BEGAN: 12:56

SUBJECT'S LAST NAME: FRAIZER	DATE/TIME OF TEST: 03/28/2023 13:43:34
------------------------------	--

Diagnostics Check	OK	13:35
Air Blank	0.000	13:35
Control Test	0.081	13:36
Air Blank	0.000	13:36
Subject Sample # 1	0.178	13:38
Breath Volume	1.847	13:38
Air Blank	0.000	13:39
Air Blank	0.000	13:41
Subject Sample # 2	0.177	13:41
Breath Volume	1.292	13:41
Air Blank	0.000	13:42
Control Test	0.079	13:42
Air Blank	0.000	13:43
Diagnostics Check	OK	13:43

Data Download Date/Time: 05/01/2023 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: CARIGG

DATE/TIME OF TEST: 03/02/2023 22:18:01

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.082	22:15	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:16	Subject Test Refused
Breath Volume	0.000	22:16	
Air Blank	0.000	22:17	
Control Test	0.082	22:17	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 03/07/2023 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUSTER,CORY,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GUSTER,CORY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BUCKLER

DATE/TIME OF TEST: 03/08/2023 01:21:59

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.082	01:21	RFI Detect
Air Blank	0.000	01:21	

Data Download Date/Time: 04/07/2023 23:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUSTER,CORY,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GUSTER,CORY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BUCKLER

DATE/TIME OF TEST: 03/08/2023 01:24:33

Diagnostics Check	OK	01:23	
Air Blank		01:24	RFI Detect
Air Blank	0.000	01:24	

Data Download Date/Time: 04/07/2023 23:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUSTER,CORY,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GUSTER,CORY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BUCKLER

DATE/TIME OF TEST: 03/08/2023 01:37:20

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.082	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.209	01:28
Breath Volume	1.921	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.171	01:32
Breath Volume	2.300	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 3	0.192	01:35
Breath Volume	2.703	01:35
Air Blank	0.000	01:36
Control Test	0.076	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 04/07/2023 23:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 03/15/2023 22:03:54

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.083	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.139	21:58
Breath Volume	3.488	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.139	22:02
Breath Volume	3.195	22:02
Air Blank	0.000	22:02
Control Test	0.082	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 04/07/2023 23:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:13

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 03/16/2023 11:40:17

Diagnostics Check	OK	11:33
Air Blank	0.000	11:34
Control Test	0.082	11:34
Air Blank	0.000	11:35
Subject Sample # 1	0.000	11:35
Breath Volume	2.136	11:35
Air Blank	0.000	11:36
Air Blank	0.000	11:38
Subject Sample # 2	0.000	11:38
Breath Volume	2.062	11:38
Air Blank	0.000	11:39
Control Test	0.082	11:39
Air Blank	0.000	11:40
Diagnostics Check	OK	11:40

Data Download Date/Time: 04/07/2023 23:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,F,H
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: YOUNG,R
ARRESTING OFFICER AGENCY: Holmes County SO
LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 03/11/2023 21:40:13

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.082	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.000	21:35
Breath Volume	1.781	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.000	21:38
Breath Volume	2.156	21:38
Air Blank	0.000	21:39
Control Test	0.081	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 03/14/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 2402328118
BREATH TEST OPERATOR: HARDY,JOE,E	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: LONG,KOBI	ARREST TIME: 21:01
ARRESTING OFFICER AGENCY: Bonifay PD	GENDER: Male
LAST AGENCY INSPECTION: 03/14/2023	OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: HOLSTON	DATE/TIME OF TEST: 03/18/2023	23:01:35
------------------------------	-------------------------------	----------

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.082	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.111	22:56
Breath Volume	1.816	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.098	22:59
Breath Volume	1.656	22:59
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 05/01/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 2402328118
BREATH TEST OPERATOR: HARDY,JOE,E	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: HARDY,JOE	ARREST TIME: 01:24
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Male
LAST AGENCY INSPECTION: 03/14/2023	OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: EDMINSON	DATE/TIME OF TEST: 03/25/2023 02:34:42
-------------------------------	--

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.161	02:30
Breath Volume	3.085	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.159	02:33
Breath Volume	2.691	02:33
Air Blank	0.000	02:33
Control Test	0.078	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 05/01/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: UNDERHILL

DATE/TIME OF TEST: 03/25/2023 19:54:45

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.080	19:48
Air Blank	0.000	19:49
Subject Sample # 1	0.145	19:50
Breath Volume	3.175	19:50
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.141	19:53
Breath Volume	3.859	19:53
Air Blank	0.000	19:53
Control Test	0.079	19:54
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 05/01/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 2402328118
BREATH TEST OPERATOR: GUSTER,CORY,J	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: LONG,KOBI	ARREST TIME: 20:01
ARRESTING OFFICER AGENCY: Bonifay PD	GENDER: Female
LAST AGENCY INSPECTION: 03/14/2023	OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: RAVENELLE	DATE/TIME OF TEST: 03/27/2023 22:13:31
--------------------------------	--

Diagnostics Check	OK	22:07
Air Blank	0.000	22:07
Control Test	0.082	22:07
Air Blank	0.000	22:08
Subject Sample # 1	0.000	22:08
Breath Volume	2.144	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.000	22:12
Breath Volume	1.445	22:12
Air Blank	0.000	22:12
Control Test	0.082	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 05/01/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SABATINE,MATTHEW
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: CHESTNUT

DATE/TIME OF TEST: 03/03/2023 20:51:02

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.076	20:45
Air Blank	0.000	20:45
Subject Sample # 1	0.180	20:46
Breath Volume	3.539	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:48
Subject Sample # 2	0.174	20:49
Breath Volume	2.531	20:49
Air Blank	0.000	20:50
Control Test	0.077	20:50
Air Blank	0.000	20:50
Diagnostics Check	OK	20:51

Data Download Date/Time: 04/10/2023 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAFFER,ROBIN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SHAFFER,ROBIN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/11/2023 00:13:54

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.077	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.156	00:09
Breath Volume	3.753	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.143	00:12
Breath Volume	1.582	00:12
Air Blank	0.000	00:13
Control Test	0.075	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 04/10/2023 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: IGLESIAS

DATE/TIME OF TEST: 03/03/2023 21:23:31

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.082	21:18
Air Blank	0.000	21:18
Subject Sample # 1	0.000	21:19
Breath Volume	2.578	21:19
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.000	21:22
Breath Volume	2.835	21:22
Air Blank	0.000	21:22
Control Test	0.082	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 03/29/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MOXEY

DATE/TIME OF TEST: 03/04/2023 03:08:20

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.082	03:05	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:07	Sequence Aborted
Breath Volume	0.000	03:07	
Air Blank	0.000	03:08	

Data Download Date/Time: 03/29/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MOXEY

DATE/TIME OF TEST: 03/04/2023 03:32:42

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.209	03:28
Breath Volume	1.332	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.200	03:31
Breath Volume	2.023	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 03/29/2023 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA,ARMANDO,R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA,ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 03/04/2023 03:47:53

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.214	03:43
Breath Volume	4.109	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.210	03:46
Breath Volume	3.542	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 03/29/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: CHAJ CHAJ

DATE/TIME OF TEST: 03/04/2023 19:50:49

Diagnostics Check	OK	19:44
Air Blank	0.000	19:44
Control Test	0.082	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.187	19:46
Breath Volume	1.511	19:46
Air Blank	0.000	19:47
Air Blank	0.000	19:48
Subject Sample # 2	0.175	19:49
Breath Volume	1.519	19:49
Air Blank	0.000	19:49
Control Test	0.078	19:50
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 03/29/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF, MARK, D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: WADE, TYLER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GOMEZ PEREZ

DATE/TIME OF TEST: 03/11/2023 03:46:33

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.082	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.136	03:42
Breath Volume	1.121	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.133	03:45
Breath Volume	1.968	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 03/29/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: MORALES,FABIAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COVERT

DATE/TIME OF TEST: 03/13/2023 02:13:28

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.081	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.147	02:08
Breath Volume	1.382	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.155	02:11
Breath Volume	1.992	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 03/29/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 03/13/2023 04:41:38

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.082	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.058	04:37
Breath Volume	1.472	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.052	04:40
Breath Volume	1.500	04:40
Air Blank	0.000	04:40
Control Test	0.081	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 03/29/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HARDMAN

DATE/TIME OF TEST: 03/17/2023 00:47:29

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.082	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.139	00:42
Breath Volume	2.003	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.133	00:45
Breath Volume	1.484	00:45
Air Blank	0.000	00:46
Control Test	0.082	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 03/29/2023 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: VALENZUELA

DATE/TIME OF TEST: 03/19/2023 02:32:45

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.081	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.200	02:24
Breath Volume	1.175	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.179	02:28
Breath Volume	1.691	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 3	0.205	02:31
Breath Volume	1.207	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 03/29/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: FAUSTIN

DATE/TIME OF TEST: 03/20/2023 00:58:21

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.082	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.131	00:53
Breath Volume	1.792	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample # 2	0.130	00:56
Breath Volume	1.398	00:56
Air Blank	0.000	00:57
Control Test	0.081	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 03/29/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/26/2023 00:00:25

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.081	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.291	23:56
Breath Volume	3.195	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.300	23:58
Breath Volume	2.300	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 03/29/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAGLIONI,DAVID,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MENDEZ,JORGE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: ZEMAN

DATE/TIME OF TEST: 03/29/2023 15:22:26

Diagnostics Check	OK	15:15
Air Blank	0.000	15:16
Control Test	0.082	15:16
Air Blank	0.000	15:17
Subject Sample # 1	0.187	15:17
Breath Volume	2.351	15:17
Air Blank	0.000	15:18
Air Blank	0.000	15:20
Subject Sample # 2	0.187	15:20
Breath Volume	1.394	15:20
Air Blank	0.000	15:21
Control Test	0.082	15:21
Air Blank	0.000	15:22
Diagnostics Check	OK	15:22

Data Download Date/Time: 04/18/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEYER, CHARLES, L
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: YOUNG, JESSE
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: WINDHAM

DATE/TIME OF TEST: 03/13/2023 21:11:36

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.078	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.000	21:07
Breath Volume	2.324	21:07
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.000	21:10
Breath Volume	1.785	21:10
Air Blank	0.000	21:10
Control Test	0.078	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 04/03/2023 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 03/03/2023 15:53:20

Diagnostics Check	OK	15:50	
Air Blank	0.000	15:50	
Control Test	0.081	15:51	
Air Blank	0.000	15:51	
Subject Sample	# 1	15:51	Subject Test Refused
Breath Volume	0.000	15:51	
Air Blank	0.000	15:52	
Control Test	0.082	15:52	
Air Blank	0.000	15:53	
Diagnostics Check	OK	15:53	

Data Download Date/Time: 03/17/2023 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 03/04/2023 23:36:01

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.082	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Control Test	0.082	23:35	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 03/17/2023 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WHITAKER,G
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 03/05/2023 22:29:34

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.082	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.207	22:24
Breath Volume	2.125	22:24
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.200	22:27
Breath Volume	1.921	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 03/17/2023 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/10/2023 22:31:08

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:28	
Control Test	0.082	22:28	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:29	Subject Test Refused
Breath Volume	0.000	22:29	
Air Blank	0.000	22:30	
Control Test	0.082	22:30	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 03/17/2023 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: EIKENBERRY

DATE/TIME OF TEST: 03/10/2023 23:02:48

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.082	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.000	22:58
Breath Volume	1.406	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.000	23:01
Breath Volume	1.351	23:01
Air Blank	0.000	23:01
Control Test	0.082	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 03/17/2023 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VAUGHN,CURTIS,L

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: VAUGHN,CURTIS

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 02/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30822080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LAUDERMILCH

DATE/TIME OF TEST: 03/12/2023 01:54:10

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.080	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.081	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 03/17/2023 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: GALLOWAY

DATE/TIME OF TEST: 03/12/2023 20:54:40

Diagnostics Check	OK	20:47
Air Blank	0.000	20:48
Control Test	0.082	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.096	20:50
Breath Volume	1.355	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.094	20:53
Breath Volume	1.113	20:53
Air Blank	0.000	20:53
Control Test	0.082	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 03/17/2023 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREZKOP,ZACH,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: MCCARTY

DATE/TIME OF TEST: 03/14/2023 01:56:56

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.082	01:54	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:56	
Control Test	0.080	01:56	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 03/17/2023 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREZKOP,ZACH,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CABRERA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 03/14/2023 03:14:14

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.081	03:12	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.081	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 03/17/2023 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREZKOP,ZACH,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MCNEELEY,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: BICKERTON

DATE/TIME OF TEST: 03/15/2023 01:16:57

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.082	01:14	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:16	
Control Test	0.081	01:16	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 03/17/2023 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: ROQUE

DATE/TIME OF TEST: 03/17/2023 23:24:49

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.179	23:15
Breath Volume	1.843	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.158	23:20
Breath Volume	1.839	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 3	0.161	23:23
Breath Volume	1.199	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 04/11/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: DESTINO

DATE/TIME OF TEST: 03/18/2023 04:22:01

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.081	04:19	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:21	
Control Test	0.081	04:21	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 04/11/2023 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: FREDRIKSEN

DATE/TIME OF TEST: 03/19/2023 23:34:16

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.082	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:33	
Control Test	0.082	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 04/11/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: CHARLTON

DATE/TIME OF TEST: 03/24/2023 20:13:43

Diagnostics Check	OK	20:06
Air Blank	0.000	20:07
Control Test	0.082	20:07
Air Blank	0.000	20:07
Subject Sample # 1	0.157	20:08
Breath Volume	2.292	20:08
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.158	20:12
Breath Volume	1.781	20:12
Air Blank	0.000	20:12
Control Test	0.082	20:13
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 04/11/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,DAVID,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAMOS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RIOUX

DATE/TIME OF TEST: 03/26/2023 00:43:01

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.082	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.165	00:35
Breath Volume	1.496	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.158	00:41
Breath Volume	1.503	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:43

Data Download Date/Time: 04/11/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: KATZ

DATE/TIME OF TEST: 03/26/2023 19:51:37

Diagnostics Check	OK	19:44
Air Blank	0.000	19:45
Control Test	0.081	19:45
Air Blank	0.000	19:45
Subject Sample # 1	0.227	19:47
Breath Volume	1.703	19:47
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.230	19:50
Breath Volume	1.800	19:50
Air Blank	0.000	19:50
Control Test	0.081	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 04/11/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLITO,BRAD,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MORITO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: DANKER

DATE/TIME OF TEST: 03/27/2023 16:13:33

Diagnostics Check	OK	16:10	
Air Blank	0.000	16:11	
Control Test	0.082	16:11	
Air Blank	0.000	16:11	
Subject Sample	# 1	16:12	Subject Test Refused
Breath Volume	0.000	16:12	
Air Blank	0.000	16:12	
Control Test	0.081	16:12	
Air Blank	0.000	16:13	
Diagnostics Check	OK	16:13	

Data Download Date/Time: 04/11/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLITO,BRAD,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,VINCENT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: LEVOY

DATE/TIME OF TEST: 03/27/2023 16:44:49

Diagnostics Check	OK	16:38
Air Blank	0.000	16:38
Control Test	0.081	16:39
Air Blank	0.000	16:39
Subject Sample # 1	0.326	16:40
Breath Volume	3.015	16:40
Air Blank	0.000	16:40
Air Blank	0.000	16:42
Subject Sample # 2	0.315	16:43
Breath Volume	3.222	16:43
Air Blank	0.000	16:43
Control Test	0.081	16:44
Air Blank	0.000	16:44
Diagnostics Check	OK	16:44

Data Download Date/Time: 04/11/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,DAVID,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAMOS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: FINNERTY

DATE/TIME OF TEST: 03/31/2023 03:11:19

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.083	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.034	03:06
Breath Volume	2.671	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.034	03:09
Breath Volume	2.453	03:09
Air Blank	0.000	03:10
Control Test	0.082	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 04/11/2023 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROZZELL,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/31/2023 23:48:27

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.082	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.088	23:43
Breath Volume	1.859	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.083	23:46
Breath Volume	1.929	23:46
Air Blank	0.000	23:47
Control Test	0.083	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 04/11/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDZINSKI,BRANDON,E
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: CAMER,KAYLA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 03/01/2023 01:51:13

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:40	
Control Test	0.077	01:40	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:44	No Sample Provided
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Air Blank	0.000	01:46	
Subject Sample	# 2	01:49	Volume Not Met
Breath Volume	0.855	01:49	
Air Blank	0.000	01:50	
Control Test	0.079	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 03/02/2023 00:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDZINSKI,BRANDON,E
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: CRAMER,KAYLA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 03/01/2023 02:06:18

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.078	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1 0.160	01:59	
Breath Volume	1.781	01:59	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2	02:04	Subject Test Refused
Breath Volume	1.039	02:04	
Air Blank	0.000	02:05	
Control Test	0.075	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 03/02/2023 00:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL, TYLER, C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MEJIAS, JOE
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 03/03/2023 00:24:29

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.078	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.078	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 03/27/2023 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GUINANE,JOHN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: HERGERT

DATE/TIME OF TEST: 03/04/2023 20:00:06

Diagnostics Check	OK	19:53
Air Blank	0.000	19:54
Control Test	0.078	19:54
Air Blank	0.000	19:55
Subject Sample # 1	0.144	19:55
Breath Volume	3.136	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.143	19:58
Breath Volume	3.296	19:58
Air Blank	0.000	19:59
Control Test	0.078	19:59
Air Blank	0.000	19:59
Diagnostics Check	OK	20:00

Data Download Date/Time: 03/27/2023 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HINDERHOFER,EDWARD
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: FENNE

DATE/TIME OF TEST: 03/06/2023 05:08:41

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:05	
Control Test	0.078	05:05	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:07	Subject Test Refused
Breath Volume	0.148	05:07	
Air Blank	0.000	05:07	
Control Test	0.078	05:08	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 03/27/2023 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL, TYLER, C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HIGHTOWER, TUCKER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 03/10/2023 19:50:22

Diagnostics Check	OK	19:40
Air Blank	0.000	19:40
Control Test	0.078	19:41
Air Blank	0.000	19:41
Subject Sample # 1	0.000	19:44
Breath Volume	2.167	19:44
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.000	19:48
Breath Volume	1.351	19:48
Air Blank	0.000	19:49
Control Test	0.079	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 03/27/2023 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: DECOTEAU,TIM
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/17/2023 14:54:37

Diagnostics Check	OK	14:51	
Air Blank	0.000	14:52	
Control Test	0.080	14:52	
Air Blank	0.000	14:53	
Subject Sample	# 1	14:53	Subject Test Refused
Breath Volume	0.000	14:53	
Air Blank	0.000	14:53	
Control Test	0.078	14:54	
Air Blank	0.000	14:54	
Diagnostics Check	OK	14:54	

Data Download Date/Time: 03/27/2023 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDZINSKI,BRANDON,E
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: THOMPSON,TYLER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: NEVILLE

DATE/TIME OF TEST: 03/18/2023 03:21:59

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.078	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.089	03:17
Breath Volume	3.019	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.089	03:20
Breath Volume	2.898	03:20
Air Blank	0.000	03:21
Control Test	0.077	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 03/27/2023 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,DENISE,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HINDERHOFER,EDWARD
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 03/18/2023 04:29:03

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.078	04:26	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:28	
Control Test	0.078	04:28	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 03/27/2023 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,DENISE,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WHITMORE,JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: FREEBORN

DATE/TIME OF TEST: 03/19/2023 00:44:15

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.146	00:39
Breath Volume	1.750	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.149	00:42
Breath Volume	1.289	00:42
Air Blank	0.000	00:43
Control Test	0.076	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 03/27/2023 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SANTAMARIA,J
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: CRAY

DATE/TIME OF TEST: 03/20/2023 10:24:41

Diagnostics Check	OK	10:21	
Air Blank	0.000	10:22	
Control Test	0.080	10:22	
Air Blank	0.000	10:23	
Subject Sample	# 1	10:23	Subject Test Refused
Breath Volume	0.000	10:23	
Air Blank	0.000	10:23	
Control Test	0.080	10:24	
Air Blank	0.000	10:24	
Diagnostics Check	OK	10:24	

Data Download Date/Time: 03/27/2023 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ADKINS,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MOLASKY

DATE/TIME OF TEST: 03/21/2023 23:42:04

Diagnostics Check	OK	23:41	
Air Blank		23:41	Ambient Fail
Air Blank	0.000	23:42	

Data Download Date/Time: 03/27/2023 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ATKINS,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MOLASKY

DATE/TIME OF TEST: 03/22/2023 00:12:35

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.078	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1 0.136	00:01	Volume Not Met
Breath Volume	1.066	00:01	
Air Blank	0.000	00:02	
Air Blank	0.000	00:03	
Subject Sample	# 2 0.142	00:06	
Breath Volume	1.230	00:06	
Air Blank	0.000	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 3 0.152	00:10	
Breath Volume	1.902	00:10	
Air Blank	0.000	00:11	
Control Test	0.076	00:11	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 03/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NEIGER,K
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: KNIFFEN

DATE/TIME OF TEST: 03/24/2023 16:24:51

Diagnostics Check	OK	16:18	
Air Blank	0.000	16:18	
Control Test	0.078	16:18	
Air Blank	0.000	16:19	
Subject Sample	# 1 0.343	16:20	
Breath Volume	3.281	16:20	
Air Blank	0.000	16:20	
Air Blank	0.000	16:22	
Subject Sample	# 2 0.329	16:23	
Breath Volume	3.082	16:23	
Air Blank	0.000	16:23	
Control Test	0.078	16:24	Interferent Detect
Air Blank	0.000	16:24	

Data Download Date/Time: 03/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NEIGER,K
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: KNIFFEN

DATE/TIME OF TEST: 03/24/2023 16:34:18

Diagnostics Check	OK	16:30	
Air Blank	0.000	16:31	
Control Test	0.078	16:31	
Air Blank	0.000	16:32	
Subject Sample	# 1	16:32	Slope Not Met
Breath Volume	1.652	16:32	
Air Blank	0.000	16:33	
Control Test	0.079	16:33	Interferent Detect
Air Blank	0.000	16:34	

Data Download Date/Time: 03/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NEIGER,K
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: KNIFFEN

DATE/TIME OF TEST: 03/24/2023 16:48:41

Diagnostics Check	OK	16:41
Air Blank	0.000	16:42
Control Test	0.078	16:42
Air Blank	0.000	16:42
Subject Sample # 1	0.331	16:43
Breath Volume	2.140	16:43
Air Blank	0.000	16:44
Air Blank	0.000	16:46
Subject Sample # 2	0.313	16:47
Breath Volume	2.546	16:47
Air Blank	0.000	16:47
Control Test	0.078	16:48
Air Blank	0.000	16:48
Diagnostics Check	OK	16:48

Data Download Date/Time: 03/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ADKINS,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: HENGEL

DATE/TIME OF TEST: 03/26/2023 22:10:07

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.077	22:04
Air Blank	0.000	22:04
Subject Sample # 1	0.197	22:05
Breath Volume	2.281	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:07
Subject Sample # 2	0.192	22:08
Breath Volume	2.246	22:08
Air Blank	0.000	22:09
Control Test	0.077	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 03/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHALL,JALEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LAYNE,GARTH
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 03/28/2023 16:09:12

Diagnostics Check	OK	16:02
Air Blank	0.000	16:02
Control Test	0.078	16:03
Air Blank	0.000	16:03
Subject Sample # 1	0.172	16:04
Breath Volume	1.218	16:04
Air Blank	0.000	16:04
Air Blank	0.000	16:06
Subject Sample # 2	0.176	16:07
Breath Volume	1.671	16:07
Air Blank	0.000	16:08
Control Test	0.078	16:08
Air Blank	0.000	16:09
Diagnostics Check	OK	16:09

Data Download Date/Time: 04/30/2023 01:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,ERNEST,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ROSA,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 213073
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: LIVERMAN

DATE/TIME OF TEST: 03/29/2023 04:17:07

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.076	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.222	04:12
Breath Volume	5.125	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample # 2	0.216	04:15
Breath Volume	5.250	04:15
Air Blank	0.000	04:16
Control Test	0.075	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 04/30/2023 01:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BROWN,BROWN
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: WHITMAN

DATE/TIME OF TEST: 03/04/2023 16:30:52

Diagnostics Check	OK	16:25	
Air Blank	0.000	16:25	
Control Test	0.080	16:25	
Air Blank	0.000	16:26	
Subject Sample	# 1	16:29	No Sample Provided
Breath Volume	0.000	16:29	
Air Blank	0.000	16:30	
Air Blank	0.000	16:30	

Data Download Date/Time: 03/30/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATCH,VIRGINIA,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HATCH,VIRGINIA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: BOATWRIGHT

DATE/TIME OF TEST: 03/04/2023 20:13:12

Diagnostics Check	OK	20:06
Air Blank	0.000	20:07
Control Test	0.080	20:07
Air Blank	0.000	20:07
Subject Sample # 1	0.000	20:08
Breath Volume	1.992	20:08
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.000	20:11
Breath Volume	1.593	20:11
Air Blank	0.000	20:12
Control Test	0.079	20:12
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 03/30/2023 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: FUNCHES

DATE/TIME OF TEST: 03/08/2023 00:55:20

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.079	00:52	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:54	
Control Test	0.080	00:54	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 03/30/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, ANTHONY, T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: MELILLO, FHP
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: SMELTZ

DATE/TIME OF TEST: 03/08/2023 21:55:13

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.000	21:50
Breath Volume	1.699	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.000	21:53
Breath Volume	1.957	21:53
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 03/30/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ERDMAN,N
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: METCALF

DATE/TIME OF TEST: 03/13/2023 02:04:54

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.081	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.174	02:00
Breath Volume	1.488	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.185	02:03
Breath Volume	1.871	02:03
Air Blank	0.000	02:03
Control Test	0.081	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 03/30/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHOY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ADOLF,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 03/14/2023 03:16:52

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.168	03:10
Breath Volume	1.335	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.160	03:15
Breath Volume	1.351	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 03/30/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DELLEFEMINE,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:14

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 03/14/2023 15:42:06

Diagnostics Check	OK	15:34
Air Blank	0.000	15:34
Control Test	0.080	15:35
Air Blank	0.000	15:35
Subject Sample # 1	0.148	15:36
Breath Volume	1.429	15:36
Air Blank	0.000	15:37
Air Blank	0.000	15:38
Subject Sample # 2	0.150	15:40
Breath Volume	1.628	15:40
Air Blank	0.000	15:41
Control Test	0.079	15:41
Air Blank	0.000	15:42
Diagnostics Check	OK	15:42

Data Download Date/Time: 03/30/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: WELLINGTON

DATE/TIME OF TEST: 03/15/2023 00:11:01

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:58	
Control Test	0.081	23:58	
Air Blank	0.000	23:58	
Subject Sample	# 1	00:02	No Sample Provided
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Air Blank	0.000	00:04	
Subject Sample	# 2	0.260	00:06
Breath Volume	1.132	00:06	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample	# 3	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:10	
Control Test	0.080	00:10	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 03/30/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/15/2023 21:10:43

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.081	21:04
Air Blank	0.000	21:04
Subject Sample # 1	0.221	21:05
Breath Volume	1.398	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:07
Subject Sample # 2	0.212	21:09
Breath Volume	1.214	21:09
Air Blank	0.000	21:09
Control Test	0.080	21:10
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 03/30/2023 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BANNON,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/16/2023 22:33:05

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:30	
Control Test	0.081	22:31	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:32	
Control Test	0.081	22:32	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 03/30/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,RANDY,K
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: ROBINSON,RANDY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: S160557923660

DATE/TIME OF TEST: 03/18/2023 02:12:05

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.081	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.116	02:07
Breath Volume	3.375	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.113	02:10
Breath Volume	3.234	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 03/30/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,RANDY,K
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: BLOTZ,GEOFFREY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: PARREAULT

DATE/TIME OF TEST: 03/18/2023 03:38:47

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:27	
Control Test	0.080	03:28	
Air Blank	0.000	03:28	
Subject Sample	# 1 0.141	03:31	Volume Not Met
Breath Volume	0.558	03:31	
Air Blank	0.000	03:32	
Air Blank	0.000	03:34	
Subject Sample	# 2 0.160	03:37	Volume Not Met
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Control Test	0.079	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 03/30/2023 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,RANDY,K
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: ADOLF,CHARLEY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/18/2023 05:21:01

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.080	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.125	05:16
Breath Volume	3.121	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:18
Subject Sample # 2	0.123	05:19
Breath Volume	2.429	05:19
Air Blank	0.000	05:20
Control Test	0.079	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:21

Data Download Date/Time: 03/30/2023 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: BESTERFELDT

DATE/TIME OF TEST: 03/18/2023 19:47:22

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.081	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.080	19:42
Breath Volume	1.394	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.079	19:45
Breath Volume	1.265	19:45
Air Blank	0.000	19:46
Control Test	0.079	19:46
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 03/30/2023 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: LOCKLEY

DATE/TIME OF TEST: 03/24/2023 19:12:51

Diagnostics Check	OK	19:10	
Air Blank	0.000	19:10	
Control Test	0.080	19:10	
Air Blank	0.000	19:11	
Subject Sample	# 1	19:11	Subject Test Refused
Breath Volume	0.000	19:11	
Air Blank	0.000	19:11	
Control Test	0.080	19:12	
Air Blank	0.000	19:12	
Diagnostics Check	OK	19:12	

Data Download Date/Time: 03/30/2023 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: RAMOS,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: SCHUTT

DATE/TIME OF TEST: 03/29/2023 00:52:22

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.079	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.208	00:47
Breath Volume	1.800	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.224	00:50
Breath Volume	1.660	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 03/30/2023 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRANSUE,MICHAEL,C
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: TRANSUE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DELVA

DATE/TIME OF TEST: 03/05/2023 01:37:36

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.131	01:32
Breath Volume	2.703	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.130	01:36
Breath Volume	2.085	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 03/16/2023 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: RAPP,ERIC
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 03/05/2023 02:30:04

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.022	02:25
Breath Volume	1.984	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.021	02:28
Breath Volume	1.898	02:28
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:30

Data Download Date/Time: 03/16/2023 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: RAPP,ERIC
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GRASON

DATE/TIME OF TEST: 03/05/2023 04:58:31

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.194	04:53
Breath Volume	2.500	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:55
Subject Sample # 2	0.194	04:56
Breath Volume	2.558	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 03/16/2023 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: SULLIVAN, ZACH
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HESLIN

DATE/TIME OF TEST: 03/11/2023 02:51:08

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.246	02:46
Breath Volume	2.046	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.243	02:49
Breath Volume	2.023	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 03/16/2023 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODEN,ANDREW,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RODEN,ANDREW
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 03/11/2023 17:28:59

Diagnostics Check	OK	17:26	
Air Blank	0.000	17:26	
Control Test	0.079	17:26	
Air Blank	0.000	17:27	
Subject Sample	# 1	17:27	Subject Test Refused
Breath Volume	0.000	17:27	
Air Blank	0.000	17:28	
Control Test	0.079	17:28	
Air Blank	0.000	17:28	
Diagnostics Check	OK	17:28	

Data Download Date/Time: 03/16/2023 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: POSTMA,M
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 02/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: WILLEY

DATE/TIME OF TEST: 03/12/2023 00:27:02

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.180	00:22
Breath Volume	1.457	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.184	00:25
Breath Volume	1.675	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:27

Data Download Date/Time: 03/16/2023 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BANNICK,COURTNEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 02/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 03/13/2023 00:37:26

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.079	00:35	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:36	
Control Test	0.079	00:36	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 03/16/2023 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUCZKA,JACOB,C
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: KUCZKA,JACOB
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KRUPA

DATE/TIME OF TEST: 03/15/2023 00:09:44

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:58	
Control Test	0.080	23:58	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.156	00:02	Volume Not Met
Breath Volume	0.355	00:02	
Air Blank	0.000	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 2 0.171	00:07	Volume Not Met
Breath Volume	0.144	00:07	
Air Blank	0.000	00:08	
Control Test	0.080	00:09	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 03/16/2023 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: COE,BARRY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: YAZELL

DATE/TIME OF TEST: 03/19/2023 02:34:16

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.079	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 05/08/2023 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRONK,A,
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: CRONK,A
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 03/19/2023 03:19:57

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:19	
Air Blank		03:19	RFI Detect
Air Blank	0.000	03:19	

Data Download Date/Time: 05/08/2023 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRONK,A,
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: CRONK,A
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 03/19/2023 03:29:27

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.084	03:24
Breath Volume	4.148	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.081	03:27
Breath Volume	4.238	03:27
Air Blank	0.000	03:28
Control Test	0.077	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 05/08/2023 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: POSTMA,M
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: KOPP

DATE/TIME OF TEST: 03/21/2023 02:39:18

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.153	02:34
Breath Volume	1.738	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.165	02:37
Breath Volume	1.343	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 05/08/2023 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: PEREZ,FRANCISCO
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: MIERZWAK

DATE/TIME OF TEST: 03/21/2023 23:12:07

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.079	23:10	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:11	
Control Test	0.079	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 05/08/2023 13:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: RAPP,ERIC
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/23/2023 01:12:59

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:10	
Control Test	0.079	01:10	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 05/08/2023 13:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREANO,JACK,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: KRUEGER,ELIZABETH
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MILLEDGE

DATE/TIME OF TEST: 03/23/2023 22:30:24

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:28	
Control Test	0.079	22:28	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:29	Subject Test Refused
Breath Volume	0.000	22:29	
Air Blank	0.000	22:29	
Control Test	0.080	22:29	
Air Blank	0.000	22:30	
Diagnostics Check	OK	22:30	

Data Download Date/Time: 05/08/2023 13:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BANNICK,COURTNEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/26/2023 02:01:53

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.080	01:59	
Air Blank	0.000	01:59	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.079	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 05/08/2023 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONDAHL,CARL,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PERALTA,BRYAN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: POFF

DATE/TIME OF TEST: 03/29/2023 17:39:56

Diagnostics Check	OK	17:28	
Air Blank	0.000	17:28	
Control Test	0.080	17:29	
Air Blank	0.000	17:29	
Subject Sample	# 1	17:32	No Sample Provided
Breath Volume	0.000	17:32	
Air Blank	0.000	17:33	
Air Blank	0.000	17:35	
Subject Sample	# 2	17:38	No Sample Provided
Breath Volume	0.000	17:38	
Air Blank	0.000	17:38	
Control Test	0.079	17:39	
Air Blank	0.000	17:39	
Diagnostics Check	OK	17:39	

Data Download Date/Time: 05/08/2023 13:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONDAHL,CARL,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PERALTA,BRYAN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: POFF

DATE/TIME OF TEST: 03/29/2023 17:45:54

Diagnostics Check	OK	17:42	
Air Blank	0.000	17:42	
Control Test	0.079	17:43	
Air Blank	0.000	17:43	
Subject Sample	# 1	17:44	Subject Test Refused
Breath Volume	0.000	17:44	
Air Blank	0.000	17:44	
Control Test	0.079	17:45	
Air Blank	0.000	17:45	
Diagnostics Check	OK	17:45	

Data Download Date/Time: 05/08/2023 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: POSTMA,M
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: PIONTEK

DATE/TIME OF TEST: 03/30/2023 02:54:35

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.079	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.162	02:49
Breath Volume	1.199	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.151	02:53
Breath Volume	1.398	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 05/08/2023 13:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COFFEY-SHADE,W,T

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: COFFEY-SHADE,W

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 04:02

GENDER: Female

OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: IVERY

DATE/TIME OF TEST: 03/01/2023 06:35:21

Diagnostics Check	OK	06:30	
Air Blank	0.000	06:31	
Control Test	0.078	06:31	
Air Blank	0.000	06:31	
Subject Sample	# 1	06:33	Subject Test Refused
Breath Volume	0.019	06:33	
Air Blank	0.000	06:34	
Control Test	0.079	06:34	
Air Blank	0.000	06:35	
Diagnostics Check	OK	06:35	

Data Download Date/Time: 03/29/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MARSKI

DATE/TIME OF TEST: 03/11/2023 04:31:37

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.139	04:27
Breath Volume	3.703	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.135	04:30
Breath Volume	2.746	04:30
Air Blank	0.000	04:30
Control Test	0.080	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 03/29/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: MAPLE

DATE/TIME OF TEST: 03/13/2023 00:21:18

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:10	
Control Test	0.079	00:10	
Air Blank	0.000	00:11	
Subject Sample	# 1 0.110	00:14	Volume Not Met
Breath Volume	0.343	00:14	
Air Blank	0.000	00:14	
Air Blank	0.000	00:16	
Subject Sample	# 2 0.113	00:19	Volume Not Met
Breath Volume	0.339	00:19	
Air Blank	0.000	00:20	
Control Test	0.079	00:20	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 03/29/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: MAPLE

DATE/TIME OF TEST: 03/13/2023 00:31:04

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.079	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.110	00:25
Breath Volume	1.386	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:27
Subject Sample # 2	0.098	00:29
Breath Volume	1.300	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:31

Data Download Date/Time: 03/29/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: ALGARIN HERNANDEZ

DATE/TIME OF TEST: 03/15/2023 22:25:04

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.152	22:19
Breath Volume	1.117	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.148	22:23
Breath Volume	1.398	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:25

Data Download Date/Time: 03/29/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 03/16/2023 19:20:38

Diagnostics Check	OK	19:12
Air Blank	0.000	19:13
Control Test	0.079	19:13
Air Blank	0.000	19:14
Subject Sample # 1	0.156	19:15
Breath Volume	1.898	19:15
Air Blank	0.000	19:15
Air Blank	0.000	19:17
Subject Sample # 2	0.149	19:19
Breath Volume	1.300	19:19
Air Blank	0.000	19:19
Control Test	0.079	19:20
Air Blank	0.000	19:20
Diagnostics Check	OK	19:20

Data Download Date/Time: 03/29/2023 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCLAIN, DAVID, R

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: DAGASTINO, STEFANO

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: NANASCA

DATE/TIME OF TEST: 03/18/2023 03:43:38

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.152	03:39
Breath Volume	2.039	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.155	03:42
Breath Volume	1.648	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 03/29/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COFFEY-SHADE, W, T

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: COFFEY-SHADE, W

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: HEARNDON

DATE/TIME OF TEST: 03/18/2023 22:49:03

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:43	
Control Test	0.079	22:43	
Air Blank	0.000	22:44	
Subject Sample	# 1 0.302	22:44	
Breath Volume	3.093	22:44	
Air Blank	0.000	22:45	
Air Blank	0.000	22:47	
Subject Sample	# 2 0.292	22:47	
Breath Volume	2.968	22:47	
Air Blank	0.000	22:48	
Control Test	0.074	22:48	Control Outside Tolerance
Air Blank	0.000	22:49	

Data Download Date/Time: 03/29/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEY-SHADE, WILLIAM, T
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: COFFEY-SHADE, W
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: HEARN DON

DATE/TIME OF TEST: 03/18/2023 23:00:55

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.078	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.288	22:56
Breath Volume	3.062	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:58
Subject Sample # 2	0.294	22:59
Breath Volume	3.148	22:59
Air Blank	0.000	23:00
Control Test	0.078	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 03/29/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COFFEY-SHADE,W,T

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: COFFEY-SHADE,W

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 04:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: ROXBURGH

DATE/TIME OF TEST: 03/19/2023 05:12:33

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.222	05:08
Breath Volume	1.296	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.214	05:11
Breath Volume	3.855	05:11
Air Blank	0.000	05:11
Control Test	0.078	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 03/29/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,DAVID,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN,DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CORONADO

DATE/TIME OF TEST: 03/30/2023 01:25:26

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:24	
Control Test	0.079	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 04/19/2023 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELHO,ISAC,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BOTELHO,ISAC
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/01/2023 01:23:40

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.083	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.158	01:19
Breath Volume	1.503	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.166	01:22
Breath Volume	1.605	01:22
Air Blank	0.000	01:22
Control Test	0.082	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEASEL,KELSEY,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: FEASEL,KELSEY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/01/2023 02:16:11

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.082	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.060	02:11
Breath Volume	2.605	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.059	02:14
Breath Volume	2.949	02:14
Air Blank	0.000	02:15
Control Test	0.081	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALGADO,BRANDON,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: SALGADO,BRANDON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PAULINE

DATE/TIME OF TEST: 03/01/2023 02:41:33

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.082	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.000	02:37
Breath Volume	2.046	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.000	02:40
Breath Volume	1.757	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSNYAK,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOSNYAK,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 03/02/2023 22:34:22

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:25	
Control Test	0.082	22:26	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:29	No Sample Provided
Breath Volume	0.000	22:29	
Air Blank	0.000	22:30	
Air Blank	0.000	22:32	
Subject Sample	# 2	22:33	Subject Test Refused
Breath Volume	0.000	22:33	
Air Blank	0.000	22:33	
Control Test	0.083	22:33	
Air Blank	0.000	22:34	
Diagnostics Check	OK	22:34	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,MAXWELL,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COLEMAN,MAXWELL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: HENNEMAN

DATE/TIME OF TEST: 03/03/2023 19:22:23

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:06	
Control Test	0.082	19:07	
Air Blank	0.000	19:07	
Subject Sample	# 1	19:11	No Sample Provided
Breath Volume	0.000	19:11	
Air Blank	0.000	19:12	
Air Blank	0.000	19:13	
Subject Sample	# 2	0.288	19:15
Breath Volume	2.796	19:15	
Air Blank	0.000	19:16	
Air Blank	0.000	19:17	
Subject Sample	# 3	0.278	19:20
Breath Volume	2.214	19:20	
Air Blank	0.000	19:21	
Control Test	0.077	19:21	
Air Blank	0.000	19:22	
Diagnostics Check	OK	19:22	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELHO,ISAC,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BOTELHO,ISAC
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 03/03/2023 20:44:32

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.081	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.263	20:39
Breath Volume	2.894	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.271	20:42
Breath Volume	3.031	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: H155733578010

DATE/TIME OF TEST: 03/03/2023 23:03:35

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.081	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:56	No Sample Provided
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 2	23:02	No Sample Provided
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Control Test	0.082	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 03/04/2023 02:57:20

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.082	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.000	02:53
Breath Volume	2.312	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.000	02:55
Breath Volume	2.183	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANZA,CHRISTIAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LANZA,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SHORT

DATE/TIME OF TEST: 03/04/2023 05:00:18

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.082	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.218	04:55
Breath Volume	3.414	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.206	04:58
Breath Volume	3.464	04:58
Air Blank	0.000	04:59
Control Test	0.081	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTEZ,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: CAMACHO,LUIS
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 03/07/2023 08:10:45

Diagnostics Check	OK	08:04	
Air Blank	0.000	08:04	
Control Test	0.083	08:05	
Air Blank	0.000	08:06	
Subject Sample	# 1	08:09	No Sample Provided
Breath Volume	0.000	08:09	
Air Blank		08:09	Purge Fail
Air Blank		08:10	Purge Fail

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTEZ,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: CAMACHO,LUIS
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 03/07/2023 08:18:49

Diagnostics Check	OK	08:17	
Air Blank		08:18	Ambient Fail
Air Blank	0.000	08:18	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTEZ,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: CAMACHO,LUIS
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 03/07/2023 08:32:10

Diagnostics Check	OK	08:20	
Air Blank	0.000	08:21	
Control Test	0.082	08:21	
Air Blank	0.000	08:21	
Subject Sample	# 1	08:25	No Sample Provided
Breath Volume	0.000	08:25	
Air Blank	0.000	08:25	
Air Blank	0.000	08:27	
Subject Sample	# 2	08:30	No Sample Provided
Breath Volume	0.000	08:30	
Air Blank		08:31	Purge Fail
Air Blank	0.000	08:32	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAINO,JOSHUA,V
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: BRAUN,DANIELLE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BRAGMAN

DATE/TIME OF TEST: 03/07/2023 22:59:12

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.083	22:46	
Air Blank	0.000	22:47	
Subject Sample	# 1 0.142	22:50	Volume Not Met
Breath Volume	0.578	22:50	
Air Blank	0.000	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 2 0.167	22:54	
Breath Volume	1.476	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:57	
Subject Sample	# 3 0.165	22:57	
Breath Volume	1.242	22:57	
Air Blank	0.000	22:58	
Control Test	0.080	22:58	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEMLER,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SEMLER,JOSEPH
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BANNER

DATE/TIME OF TEST: 03/09/2023 04:21:56

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.082	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.182	04:17
Breath Volume	2.425	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:19
Subject Sample # 2	0.192	04:20
Breath Volume	2.605	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWELL,BARRY,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SEARS,TREVER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: FITZHERBERT

DATE/TIME OF TEST: 03/12/2023 03:17:32

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.083	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:10	No Sample Provided
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Air Blank	0.000	03:13	
Subject Sample	# 2	03:16	No Sample Provided
Breath Volume	0.000	03:16	
Air Blank	0.000	03:16	
Control Test	0.083	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,F,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LOPEZ,F
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/12/2023 04:37:16

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.083	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1 0.131	04:27	
Breath Volume	1.171	04:27	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample	# 2	04:32	No Sample Provided
Breath Volume	0.000	04:32	
Air Blank	0.000	04:33	
Air Blank	0.000	04:35	
Subject Sample	# 3 0.133	04:35	
Breath Volume	2.441	04:35	
Air Blank	0.000	04:36	
Control Test	0.082	04:36	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCHESTER,MILTON,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MANCHESTER,MILTON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MORVANT

DATE/TIME OF TEST: 03/14/2023 01:24:10

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.083	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.197	01:19
Breath Volume	2.972	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.204	01:22
Breath Volume	2.312	01:22
Air Blank	0.000	01:23
Control Test	0.083	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: PICKFORD

DATE/TIME OF TEST: 03/14/2023 17:28:09

Diagnostics Check	OK	17:21
Air Blank	0.000	17:22
Control Test	0.085	17:22
Air Blank	0.000	17:23
Subject Sample # 1	0.245	17:23
Breath Volume	1.812	17:23
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 2	0.238	17:26
Breath Volume	1.976	17:26
Air Blank	0.000	17:27
Control Test	0.083	17:27
Air Blank	0.000	17:28
Diagnostics Check	OK	17:28

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUIS,ALYSSA,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ROUIS,ALYSSA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: HEINITZ

DATE/TIME OF TEST: 03/18/2023 00:19:10

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.083	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.174	00:14
Breath Volume	1.187	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.172	00:17
Breath Volume	1.191	00:17
Air Blank	0.000	00:18
Control Test	0.082	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HERNANDEZ,IVAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: LANGLOIS

DATE/TIME OF TEST: 03/18/2023 01:55:55

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.082	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.000	01:51
Breath Volume	1.765	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.000	01:54
Breath Volume	2.117	01:54
Air Blank	0.000	01:55
Control Test	0.082	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MELCHOR PEREZ

DATE/TIME OF TEST: 03/18/2023 03:15:48

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.083	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:14	Sequence Aborted
Breath Volume	0.019	03:14	
Air Blank	0.000	03:15	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MELCHOR PEREZ

DATE/TIME OF TEST: 03/18/2023 03:31:19

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.078	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1 0.143	03:22	Volume Not Met
Breath Volume	0.078	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 2 0.147	03:25	
Breath Volume	2.261	03:25	
Air Blank	0.000	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 3 0.155	03:29	
Breath Volume	1.933	03:29	
Air Blank	0.000	03:30	
Control Test	0.075	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SIMONSEN,ERIK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: VANHOUTEN

DATE/TIME OF TEST: 03/18/2023 04:43:02

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:37	
Control Test	0.076	04:37	
Air Blank	0.000	04:38	
Subject Sample	# 1	04:41	No Sample Provided
Breath Volume	0.000	04:41	
Air Blank		04:42	Purge Fail
Air Blank		04:43	Purge Fail

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SIMONSEN,ERIK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: VANHOUTEN

DATE/TIME OF TEST: 03/18/2023 04:46:32

Diagnostics Check	OK	04:45	
Air Blank		04:46	Ambient Fail
Air Blank	0.000	04:46	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SIMONSEN,ERIK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: VANHOUTEN

DATE/TIME OF TEST: 03/18/2023 04:59:13

Diagnostics Check	OK	04:48	
Air Blank	0.000	04:48	
Control Test	0.081	04:49	
Air Blank	0.000	04:49	
Subject Sample	# 1 0.242	04:52	
Breath Volume	1.105	04:52	
Air Blank	0.000	04:53	
Air Blank	0.000	04:54	
Subject Sample	# 2 0.248	04:57	Volume Not Met
Breath Volume	0.175	04:57	
Air Blank		04:58	Purge Fail
Air Blank	0.000	04:59	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SIMONSEN,ERIK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: VANHOUTEN

DATE/TIME OF TEST: 03/18/2023 05:04:07

Diagnostics Check	OK	05:01	
Air Blank	0.000	05:01	
Control Test	0.075	05:01	
Air Blank	0.000	05:02	
Subject Sample	# 1 0.238	05:02	
Breath Volume	1.753	05:02	
Air Blank		05:03	Purge Fail
Air Blank	0.000	05:04	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SIMONSEN,ERIK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: VANHOUTEN

DATE/TIME OF TEST: 03/18/2023 05:20:53

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:06	
Control Test	0.079	05:06	
Air Blank	0.000	05:07	
Subject Sample	# 1 0.219	05:10	Volume Not Met
Breath Volume	0.035	05:10	
Air Blank	0.000	05:11	
Air Blank	0.000	05:12	
Subject Sample	# 2 0.222	05:15	
Breath Volume	1.847	05:15	
Air Blank	0.000	05:16	
Air Blank	0.000	05:17	
Subject Sample	# 3 0.233	05:19	
Breath Volume	1.585	05:19	
Air Blank		05:20	Purge Fail
Air Blank	0.000	05:20	

Data Download Date/Time: 03/29/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SOTELO

DATE/TIME OF TEST: 03/19/2023 04:52:51

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.084	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.227	04:48
Breath Volume	2.000	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.218	04:51
Breath Volume	1.167	04:51
Air Blank	0.000	04:51
Control Test	0.080	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 03/29/2023 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/20/2023 00:10:56

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.084	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.207	00:06
Breath Volume	1.351	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:08
Subject Sample # 2	0.199	00:09
Breath Volume	1.828	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 03/29/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ORR

DATE/TIME OF TEST: 03/24/2023 02:17:10

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.083	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.292	02:12
Breath Volume	2.082	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.295	02:15
Breath Volume	2.339	02:15
Air Blank	0.000	02:16
Control Test	0.082	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 03/29/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENTSCH,RICHARD,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: JENTSCH,RICHARD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MUNTUN

DATE/TIME OF TEST: 03/26/2023 00:02:23

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.084	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.232	23:56
Breath Volume	1.421	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 1	0.236	00:00
Breath Volume	1.339	00:00
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 03/29/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ RIVERA, ANDREW, J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOSNYAK, ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: FLORES HERNANDEZ

DATE/TIME OF TEST: 03/25/2023 01:42:11

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.083	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.267	01:37
Breath Volume	2.296	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.273	01:40
Breath Volume	2.160	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 03/29/2023 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEIGH,DAVID,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: BURLEIGH,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: CRUZ PEREZ

DATE/TIME OF TEST: 03/27/2023 00:14:57

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.083	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.203	00:07
Breath Volume	2.187	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.227	00:10
Breath Volume	2.160	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 3	0.233	00:13
Breath Volume	2.519	00:13
Air Blank	0.000	00:14
Control Test	0.082	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 03/29/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 03/27/2023 02:43:10

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.083	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.170	02:38
Breath Volume	3.125	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample # 2	0.173	02:41
Breath Volume	2.324	02:41
Air Blank	0.000	02:42
Control Test	0.083	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 03/29/2023 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETTS,J,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 03/27/2023 23:25:26

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.083	23:11	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:15	No Sample Provided
Breath Volume	0.000	23:15	
Air Blank	0.000	23:15	
Air Blank	0.000	23:17	
Subject Sample	# 2	0.116	23:18
Breath Volume	3.238	23:18	
Air Blank	0.000	23:19	
Air Blank	0.000	23:21	
Subject Sample	# 3	0.115	23:23
Breath Volume	4.792	23:23	
Air Blank	0.000	23:24	
Control Test	0.082	23:24	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 03/29/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETTS,J,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 03/28/2023 19:50:51

Diagnostics Check	OK	19:42
Air Blank	0.000	19:43
Control Test	0.083	19:43
Air Blank	0.000	19:43
Subject Sample # 1	0.196	19:45
Breath Volume	1.750	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.186	19:49
Breath Volume	1.210	19:49
Air Blank	0.000	19:49
Control Test	0.083	19:50
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 03/29/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MICHAELIDES,S,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MICHAELIDES,S
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: BEVILLE

DATE/TIME OF TEST: 03/28/2023 21:31:01

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.084	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.000	21:26
Breath Volume	3.011	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.000	21:29
Breath Volume	1.503	21:29
Air Blank	0.000	21:30
Control Test	0.084	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:31

Data Download Date/Time: 03/29/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: COMM VEHIC
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CORD

DATE/TIME OF TEST: 03/01/2023 01:00:49

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.082	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.113	00:56
Breath Volume	2.503	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.111	00:59
Breath Volume	2.070	00:59
Air Blank	0.000	00:59
Control Test	0.081	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 03/02/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: HOUSER

DATE/TIME OF TEST: 03/01/2023 03:49:40

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.082	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.189	03:44
Breath Volume	1.617	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.171	03:48
Breath Volume	1.378	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 03/02/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: RINCON

DATE/TIME OF TEST: 03/01/2023 05:05:02

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.081	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.149	05:00
Breath Volume	2.085	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:02
Subject Sample # 2	0.161	05:03
Breath Volume	2.082	05:03
Air Blank	0.000	05:04
Control Test	0.081	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:05

Data Download Date/Time: 03/02/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WHITLOCK,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: AMAYA

DATE/TIME OF TEST: 03/01/2023 18:15:06

Diagnostics Check	OK	18:08
Air Blank	0.000	18:08
Control Test	0.082	18:08
Air Blank	0.000	18:09
Subject Sample # 1	0.210	18:10
Breath Volume	1.507	18:10
Air Blank	0.000	18:11
Air Blank	0.000	18:12
Subject Sample # 2	0.218	18:13
Breath Volume	1.132	18:13
Air Blank	0.000	18:14
Control Test	0.081	18:14
Air Blank	0.000	18:15
Diagnostics Check	OK	18:15

Data Download Date/Time: 03/02/2023 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: TAPIA

DATE/TIME OF TEST: 03/02/2023 02:57:44

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.082	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Control Test	0.081	02:57	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 03/02/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: DELORENZO

DATE/TIME OF TEST: 03/02/2023 22:37:42

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.083	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.059	22:32
Breath Volume	1.277	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:34
Subject Sample # 2	0.066	22:36
Breath Volume	1.996	22:36
Air Blank	0.000	22:36
Control Test	0.082	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 04/10/2023 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PAGAN-OLMO

DATE/TIME OF TEST: 03/03/2023 04:08:33

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.083	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.080	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 04/10/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: TAJAH

DATE/TIME OF TEST: 03/05/2023 00:25:11

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:15	
Control Test	0.083	00:15	
Air Blank	0.000	00:16	
Subject Sample	# 1	00:19	No Sample Provided
Breath Volume	0.000	00:19	
Air Blank	0.000	00:19	
Air Blank	0.000	00:21	
Subject Sample	# 2	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:24	
Control Test	0.082	00:24	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 04/10/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: DE LOS SANTOS BAUTIS

DATE/TIME OF TEST: 03/05/2023 01:10:11

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.082	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.204	01:05
Breath Volume	1.925	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:07
Subject Sample # 2	0.204	01:08
Breath Volume	1.484	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 04/10/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WHITLOCK, R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: LAZARRE

DATE/TIME OF TEST: 03/05/2023 02:52:08

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.081	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.172	02:46
Breath Volume	1.937	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.171	02:50
Breath Volume	1.972	02:50
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 04/10/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 03/05/2023 04:40:33

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.092	04:36
Breath Volume	2.628	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.089	04:38
Breath Volume	2.132	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 04/10/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 03/05/2023 05:15:18

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.082	05:09
Air Blank	0.000	05:09
Subject Sample # 1	0.131	05:10
Breath Volume	1.660	05:10
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.135	05:13
Breath Volume	2.125	05:13
Air Blank	0.000	05:14
Control Test	0.080	05:14
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 04/10/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/06/2023 04:49:47

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.083	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.149	04:44
Breath Volume	3.144	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.143	04:48
Breath Volume	2.898	04:48
Air Blank	0.000	04:48
Control Test	0.081	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 04/10/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHANNON,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: CHERRY

DATE/TIME OF TEST: 03/06/2023 17:37:51

Diagnostics Check	OK	17:25
Air Blank	0.000	17:26
Control Test	0.082	17:26
Air Blank	0.000	17:26
Subject Sample # 1	0.238	17:27
Breath Volume	2.691	17:27
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.212	17:31
Breath Volume	1.558	17:31
Air Blank	0.000	17:32
Air Blank	0.000	17:34
Subject Sample # 3	0.215	17:36
Breath Volume	1.371	17:36
Air Blank	0.000	17:36
Control Test	0.081	17:37
Air Blank	0.000	17:37
Diagnostics Check	OK	17:37

Data Download Date/Time: 04/10/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 03/08/2023 04:01:53

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.082	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.135	03:56
Breath Volume	1.363	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.133	04:00
Breath Volume	1.558	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: RAMIREZ BASULTO

DATE/TIME OF TEST: 03/08/2023 04:53:54

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.082	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.130	04:48
Breath Volume	1.480	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.146	04:52
Breath Volume	1.472	04:52
Air Blank	0.000	04:52
Control Test	0.082	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: KELLEY JR

DATE/TIME OF TEST: 03/10/2023 03:54:23

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.083	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.176	03:45
Breath Volume	2.511	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:47
Subject Sample # 2	0.151	03:49
Breath Volume	2.207	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 3	0.159	03:52
Breath Volume	1.832	03:52
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: SIFUENTES

DATE/TIME OF TEST: 03/11/2023 06:16:19

Diagnostics Check	OK	06:09
Air Blank	0.000	06:10
Control Test	0.083	06:10
Air Blank	0.000	06:11
Subject Sample # 1	0.121	06:11
Breath Volume	1.171	06:11
Air Blank	0.000	06:12
Air Blank	0.000	06:14
Subject Sample # 2	0.133	06:14
Breath Volume	1.179	06:14
Air Blank	0.000	06:15
Control Test	0.082	06:15
Air Blank	0.000	06:16
Diagnostics Check	OK	06:16

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GARARD III

DATE/TIME OF TEST: 03/11/2023 22:23:44

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.082	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.080	22:18
Breath Volume	3.585	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.083	22:21
Breath Volume	3.257	22:21
Air Blank	0.000	22:22
Control Test	0.082	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 03/12/2023 03:46:38

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.083	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.000	03:42
Breath Volume	2.574	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.000	03:45
Breath Volume	3.269	03:45
Air Blank	0.000	03:45
Control Test	0.082	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: NAVA SANCHEZ

DATE/TIME OF TEST: 03/13/2023 00:12:35

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.084	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.110	00:07
Breath Volume	1.890	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.119	00:10
Breath Volume	1.777	00:10
Air Blank	0.000	00:11
Control Test	0.081	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: NADIMPALLI

DATE/TIME OF TEST: 03/14/2023 00:28:26

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.084	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.086	00:23
Breath Volume	1.773	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.090	00:26
Breath Volume	1.593	00:26
Air Blank	0.000	00:27
Control Test	0.082	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: BARILLAS OLIVAS

DATE/TIME OF TEST: 03/15/2023 00:33:46

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.084	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:32	
Control Test	0.082	00:33	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/16/2023 03:21:33

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.085	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.076	03:16
Breath Volume	4.097	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.077	03:19
Breath Volume	3.976	03:19
Air Blank	0.000	03:20
Control Test	0.083	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: PETRONE

DATE/TIME OF TEST: 03/16/2023 20:42:24

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:30	
Control Test	0.081	20:31	
Air Blank	0.000	20:31	
Subject Sample	# 1 0.282	20:34	Volume Not Met
Breath Volume	0.050	20:34	
Air Blank	0.000	20:35	
Air Blank	0.000	20:37	
Subject Sample	# 2 0.274	20:40	Volume Not Met
Breath Volume	0.050	20:40	
Air Blank	0.000	20:41	
Control Test	0.075	20:41	
Air Blank	0.000	20:42	
Diagnostics Check	OK	20:42	

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: HARBEY

DATE/TIME OF TEST: 03/17/2023 03:55:52

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.275	03:50
Breath Volume	3.164	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.271	03:53
Breath Volume	3.222	03:53
Air Blank	0.000	03:54
Control Test	0.083	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 03/18/2023 01:48:55

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.084	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.087	01:44
Breath Volume	2.449	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:46
Subject Sample # 2	0.088	01:47
Breath Volume	3.320	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: DE LOS RIOS

DATE/TIME OF TEST: 03/18/2023 02:37:01

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.082	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.165	02:32
Breath Volume	3.421	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.168	02:35
Breath Volume	3.242	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:37

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MATUTE GUARDADO

DATE/TIME OF TEST: 03/18/2023 04:02:18

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.082	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1 0.201	03:50	Volume Not Met
Breath Volume	0.542	03:50	
Air Blank	0.000	03:51	
Air Blank	0.000	03:53	
Subject Sample	# 2 0.237	03:54	
Breath Volume	2.054	03:54	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample	# 3	04:00	No Sample Provided
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MATUTE GUARDADO

DATE/TIME OF TEST: 03/18/2023 04:12:12

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.221	04:07
Breath Volume	1.746	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.231	04:10
Breath Volume	1.816	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK, I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: ROJAS NAVA

DATE/TIME OF TEST: 03/19/2023 05:16:04

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.082	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.140	05:11
Breath Volume	2.910	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:13
Subject Sample # 2	0.146	05:14
Breath Volume	3.156	05:14
Air Blank	0.000	05:15
Control Test	0.081	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:16

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SPENCE,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:09

SUBJECT'S LAST NAME: GOMEZ PERES

DATE/TIME OF TEST: 03/19/2023 10:39:38

Diagnostics Check	OK	10:31
Air Blank	0.000	10:31
Control Test	0.082	10:32
Air Blank	0.000	10:32
Subject Sample # 1	0.213	10:33
Breath Volume	1.476	10:33
Air Blank	0.000	10:34
Air Blank	0.000	10:36
Subject Sample # 2	0.225	10:37
Breath Volume	2.253	10:37
Air Blank	0.000	10:38
Control Test	0.083	10:38
Air Blank	0.000	10:39
Diagnostics Check	OK	10:39

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: JANES

DATE/TIME OF TEST: 03/22/2023 00:44:48

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:39	
Control Test	0.083	00:40	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:43	
Control Test	0.082	00:44	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 04/10/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 03/23/2023 01:12:30

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.084	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.084	01:07
Breath Volume	1.867	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.081	01:10
Breath Volume	1.628	01:10
Air Blank	0.000	01:11
Control Test	0.083	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: DAWSON II

DATE/TIME OF TEST: 03/23/2023 03:35:49

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.083	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.222	03:31
Breath Volume	1.160	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.227	03:34
Breath Volume	2.367	03:34
Air Blank	0.000	03:34
Control Test	0.083	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/23/2023 20:07:19

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.084	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.000	20:02
Breath Volume	1.511	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.000	20:05
Breath Volume	1.757	20:05
Air Blank	0.000	20:06
Control Test	0.083	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MATTIELLO

DATE/TIME OF TEST: 03/25/2023 02:43:27

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.083	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.230	02:35
Breath Volume	2.101	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.260	02:38
Breath Volume	2.140	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 3	0.246	02:41
Breath Volume	2.371	02:41
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 03/25/2023 04:26:25

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.082	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.103	04:20
Breath Volume	1.820	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.102	04:24
Breath Volume	1.558	04:24
Air Blank	0.000	04:25
Control Test	0.082	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: ALSTON

DATE/TIME OF TEST: 03/25/2023 05:38:31

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.082	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.172	05:33
Breath Volume	1.828	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.165	05:36
Breath Volume	1.164	05:36
Air Blank	0.000	05:37
Control Test	0.081	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: RODRIGUEZ-RAMIREZ

DATE/TIME OF TEST: 03/26/2023 00:39:53

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.082	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.221	00:35
Breath Volume	2.664	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.222	00:38
Breath Volume	2.105	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 03/26/2023 03:25:25

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.083	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.185	03:20
Breath Volume	1.878	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:22
Subject Sample # 2	0.190	03:23
Breath Volume	1.804	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: FAIN

DATE/TIME OF TEST: 03/27/2023 00:11:14

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.082	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.162	00:06
Breath Volume	1.464	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.169	00:09
Breath Volume	1.539	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEWIS,I
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: CASEY

DATE/TIME OF TEST: 03/27/2023 00:24:41

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.081	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.135	00:20
Breath Volume	2.503	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.134	00:22
Breath Volume	1.808	00:22
Air Blank	0.000	00:23
Control Test	0.081	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: QUESADA LOPEZ

DATE/TIME OF TEST: 03/27/2023 01:38:13

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.082	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.079	01:33
Breath Volume	1.859	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:35
Subject Sample # 2	0.081	01:36
Breath Volume	1.429	01:36
Air Blank	0.000	01:37
Control Test	0.082	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/27/2023 03:11:09

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.082	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.179	03:06
Breath Volume	2.152	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:08
Subject Sample # 2	0.187	03:09
Breath Volume	2.277	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: OATMAN JR

DATE/TIME OF TEST: 03/28/2023 01:37:15

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.082	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.134	01:32
Breath Volume	1.488	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.136	01:35
Breath Volume	1.765	01:35
Air Blank	0.000	01:36
Control Test	0.082	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/30/2023 01:30:15

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.083	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.207	01:25
Breath Volume	1.906	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.212	01:28
Breath Volume	1.894	01:28
Air Blank	0.000	01:29
Control Test	0.082	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: GODINEZ PEREZ

DATE/TIME OF TEST: 03/31/2023 00:59:19

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.084	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:58	
Control Test	0.081	00:58	
Air Blank	0.000	00:59	
Diagnostics Check	OK	00:59	

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HATCHETT JR

DATE/TIME OF TEST: 03/31/2023 02:49:06

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.084	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.183	02:42
Breath Volume	1.363	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.193	02:47
Breath Volume	1.148	02:47
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: TAYLOR II

DATE/TIME OF TEST: 03/31/2023 04:24:01

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.082	04:08	
Air Blank	0.000	04:09	
Subject Sample # 1	0.153	04:10	
Breath Volume	3.097	04:10	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample # 2		04:16	No Sample Provided
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample # 3	0.112	04:22	Volume Not Met
Breath Volume	1.085	04:22	
Air Blank	0.000	04:23	
Control Test	0.083	04:23	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: ROSALEZ

DATE/TIME OF TEST: 03/31/2023 22:57:52

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.083	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.150	22:53
Breath Volume	1.593	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.156	22:56
Breath Volume	1.808	22:56
Air Blank	0.000	22:56
Control Test	0.082	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 04/10/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ROBERT,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JOHNSON,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: UPTON

DATE/TIME OF TEST: 03/01/2023 21:15:02

Diagnostics Check	OK	21:08
Air Blank	0.000	21:09
Control Test	0.081	21:09
Air Blank	0.000	21:10
Subject Sample # 1	0.162	21:10
Breath Volume	1.730	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.157	21:13
Breath Volume	2.382	21:13
Air Blank	0.000	21:14
Control Test	0.078	21:14
Air Blank	0.000	21:14
Diagnostics Check	OK	21:15

Data Download Date/Time: 03/10/2023 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MUSGRAVE

DATE/TIME OF TEST: 03/05/2023 01:01:33

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.180	00:56
Breath Volume	1.644	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.179	01:00
Breath Volume	1.457	01:00
Air Blank	0.000	01:00
Control Test	0.081	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 03/10/2023 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: AN

DATE/TIME OF TEST: 03/08/2023 15:18:14

Diagnostics Check	OK	15:09
Air Blank	0.000	15:10
Control Test	0.081	15:10
Air Blank	0.000	15:10
Subject Sample # 1	0.302	15:13
Breath Volume	2.066	15:13
Air Blank	0.000	15:14
Air Blank	0.000	15:16
Subject Sample # 2	0.296	15:16
Breath Volume	1.675	15:16
Air Blank	0.000	15:17
Control Test	0.080	15:17
Air Blank	0.000	15:18
Diagnostics Check	OK	15:18

Data Download Date/Time: 03/10/2023 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: MUNOZ ORTIZ

DATE/TIME OF TEST: 03/09/2023 05:30:11

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.167	05:25
Breath Volume	1.746	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:27
Subject Sample # 2	0.166	05:28
Breath Volume	1.621	05:28
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 03/10/2023 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: MENCY

DATE/TIME OF TEST: 03/11/2023 21:03:58

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.080	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.000	20:59
Breath Volume	2.472	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.000	21:02
Breath Volume	3.187	21:02
Air Blank	0.000	21:03
Control Test	0.081	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 04/08/2023 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: BADGER

DATE/TIME OF TEST: 03/11/2023 21:45:35

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:34	
Control Test	0.081	21:34	
Air Blank	0.000	21:35	
Subject Sample	# 1 0.198	21:38	Volume Not Met
Breath Volume	0.062	21:38	
Air Blank	0.000	21:39	
Air Blank	0.000	21:40	
Subject Sample	# 2 0.177	21:44	Volume Not Met
Breath Volume	0.070	21:44	
Air Blank	0.000	21:44	
Control Test	0.080	21:45	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 04/08/2023 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: BADGER

DATE/TIME OF TEST: 03/11/2023 22:01:44

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:50	
Control Test	0.081	21:50	
Air Blank	0.000	21:51	
Subject Sample	# 1 0.208	21:54	Volume Not Met
Breath Volume	0.328	21:54	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample	# 2 0.175	22:00	Volume Not Met
Breath Volume	0.039	22:00	
Air Blank	0.000	22:00	
Control Test	0.080	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 04/08/2023 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: RACE

DATE/TIME OF TEST: 03/13/2023 20:48:42

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.081	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.052	20:44
Breath Volume	2.203	20:44
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.052	20:47
Breath Volume	1.730	20:47
Air Blank	0.000	20:47
Control Test	0.081	20:48
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 04/08/2023 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELFREY,PAUL,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: PELFREY,PAUL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: STOLLINGS

DATE/TIME OF TEST: 03/14/2023 20:49:52

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.082	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.074	20:45
Breath Volume	3.882	20:45
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.073	20:48
Breath Volume	3.773	20:48
Air Blank	0.000	20:48
Control Test	0.082	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 04/08/2023 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: HARTLEY

DATE/TIME OF TEST: 03/19/2023 15:37:00

Diagnostics Check	OK	15:30
Air Blank	0.000	15:31
Control Test	0.081	15:31
Air Blank	0.000	15:31
Subject Sample # 1	0.040	15:32
Breath Volume	3.265	15:32
Air Blank	0.000	15:33
Air Blank	0.000	15:35
Subject Sample # 2	0.036	15:35
Breath Volume	3.644	15:35
Air Blank	0.000	15:36
Control Test	0.081	15:36
Air Blank	0.000	15:36
Diagnostics Check	OK	15:37

Data Download Date/Time: 04/08/2023 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ROBERT,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JOHNSON,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: CHUMNEY

DATE/TIME OF TEST: 03/19/2023 20:13:37

Diagnostics Check	OK	20:07
Air Blank	0.000	20:07
Control Test	0.081	20:08
Air Blank	0.000	20:08
Subject Sample # 1	0.000	20:09
Breath Volume	3.070	20:09
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.000	20:12
Breath Volume	3.550	20:12
Air Blank	0.000	20:12
Control Test	0.081	20:13
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 04/08/2023 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 03/19/2023 20:40:54

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.081	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.198	20:36
Breath Volume	2.335	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:38
Subject Sample # 2	0.194	20:39
Breath Volume	2.894	20:39
Air Blank	0.000	20:40
Control Test	0.081	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 04/08/2023 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: CLEVELAND

DATE/TIME OF TEST: 03/22/2023 02:16:14

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.081	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.125	02:10
Breath Volume	1.550	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.125	02:14
Breath Volume	1.859	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 04/08/2023 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: VITELLO

DATE/TIME OF TEST: 03/23/2023 03:15:00

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.092	03:09
Breath Volume	1.589	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.074	03:13
Breath Volume	1.480	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:15

Data Download Date/Time: 04/08/2023 16:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: DUPREE

DATE/TIME OF TEST: 03/24/2023 04:15:33

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.141	04:07
Breath Volume	1.664	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.164	04:10
Breath Volume	1.992	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 3	0.172	04:14
Breath Volume	2.777	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 04/08/2023 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 03/25/2023 22:37:13

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.081	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.125	22:32
Breath Volume	1.351	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.121	22:35
Breath Volume	1.328	22:35
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 04/08/2023 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: HAMBY

DATE/TIME OF TEST: 03/26/2023 16:25:45

Diagnostics Check	OK	16:19
Air Blank	0.000	16:20
Control Test	0.080	16:20
Air Blank	0.000	16:20
Subject Sample # 1	0.000	16:21
Breath Volume	1.484	16:21
Air Blank	0.000	16:21
Air Blank	0.000	16:23
Subject Sample # 2	0.000	16:24
Breath Volume	1.230	16:24
Air Blank	0.000	16:24
Control Test	0.081	16:25
Air Blank	0.000	16:25
Diagnostics Check	OK	16:25

Data Download Date/Time: 04/08/2023 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CONSIGLIATI,EMANUELE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: OROZCO MONTES

DATE/TIME OF TEST: 03/01/2023 21:34:01

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.162	21:29
Breath Volume	1.542	21:29
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.155	21:32
Breath Volume	1.785	21:32
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:34

Data Download Date/Time: 03/08/2023 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: CHAPLIN

DATE/TIME OF TEST: 03/02/2023 01:35:52

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.080	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.197	01:28
Breath Volume	1.191	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.229	01:31
Breath Volume	2.085	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 3	0.195	01:34
Breath Volume	1.359	01:34
Air Blank	0.000	01:34
Control Test	0.079	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 03/08/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: VOELKER

DATE/TIME OF TEST: 03/02/2023 17:58:32

Diagnostics Check	OK	17:52
Air Blank	0.000	17:52
Control Test	0.081	17:53
Air Blank	0.000	17:53
Subject Sample # 1	0.000	17:54
Breath Volume	2.460	17:54
Air Blank	0.000	17:54
Air Blank	0.000	17:56
Subject Sample # 2	0.000	17:57
Breath Volume	1.515	17:57
Air Blank	0.000	17:57
Control Test	0.080	17:57
Air Blank	0.000	17:58
Diagnostics Check	OK	17:58

Data Download Date/Time: 03/08/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: JASKIER

DATE/TIME OF TEST: 03/03/2023 21:21:48

Diagnostics Check	OK	21:15
Air Blank	0.000	21:16
Control Test	0.081	21:16
Air Blank	0.000	21:17
Subject Sample # 1	0.038	21:17
Breath Volume	2.171	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:19
Subject Sample # 2	0.037	21:20
Breath Volume	1.320	21:20
Air Blank	0.000	21:20
Control Test	0.082	21:21
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 03/08/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 03/04/2023 01:56:41

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.081	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.155	01:52
Breath Volume	3.160	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.152	01:55
Breath Volume	3.117	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 03/08/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ECHEVARRIA VERA

DATE/TIME OF TEST: 03/04/2023 03:29:56

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.123	03:24
Breath Volume	3.875	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.116	03:28
Breath Volume	2.355	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 03/08/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,FRANCISCO,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 03/05/2023 00:11:44

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.081	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.253	00:04
Breath Volume	2.035	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.268	00:10
Breath Volume	1.148	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 03/08/2023 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ARNOLD,P
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 03/05/2023 01:53:28

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.164	01:49
Breath Volume	2.234	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.154	01:52
Breath Volume	1.347	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 03/08/2023 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: MERO GUADAMUD

DATE/TIME OF TEST: 03/05/2023 02:55:13

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.198	02:50
Breath Volume	1.703	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.204	02:53
Breath Volume	1.996	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 03/08/2023 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,CHEYANNE,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOPEZ,JACQIE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ZOLLNER

DATE/TIME OF TEST: 03/06/2023 04:31:12

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.096	04:26
Breath Volume	2.949	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.092	04:29
Breath Volume	3.046	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 03/08/2023 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/07/2023 22:29:08

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.081	22:27	
Air Blank	0.000	22:27	
Subject Sample	# 1	22:27	Subject Test Refused
Breath Volume	0.000	22:27	
Air Blank	0.000	22:28	
Control Test	0.080	22:28	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

Data Download Date/Time: 03/08/2023 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,JILLIAN,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: VULICEVIC,RADOSAV
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: DE OGANDO

DATE/TIME OF TEST: 03/08/2023 00:27:43

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.080	00:17	
Air Blank	0.000	00:17	
Subject Sample	# 1 0.165	00:20	Volume Not Met
Breath Volume	0.117	00:20	
Air Blank	0.000	00:21	
Air Blank	0.000	00:23	
Subject Sample	# 2 0.175	00:26	Volume Not Met
Breath Volume	0.398	00:26	
Air Blank	0.000	00:26	
Control Test	0.079	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 03/08/2023 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,JILLIAN,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: VULICEVIC,RADOSAV
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: DE OGANDO

DATE/TIME OF TEST: 03/08/2023 00:40:57

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.180	00:34
Breath Volume	1.113	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.177	00:39
Breath Volume	1.410	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 03/08/2023 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: ROSSANO

DATE/TIME OF TEST: 03/08/2023 22:41:30

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.081	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.160	22:37
Breath Volume	2.718	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.149	22:40
Breath Volume	1.925	22:40
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: KINCAID

DATE/TIME OF TEST: 03/09/2023 03:57:26

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.123	03:50
Breath Volume	1.656	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.125	03:55
Breath Volume	1.597	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: WINSLOW

DATE/TIME OF TEST: 03/09/2023 20:45:17

Diagnostics Check	OK	20:38
Air Blank	0.000	20:39
Control Test	0.081	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.325	20:40
Breath Volume	1.863	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.306	20:43
Breath Volume	1.972	20:43
Air Blank	0.000	20:44
Control Test	0.080	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,FRANCISCO,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: SMYTH

DATE/TIME OF TEST: 03/09/2023 22:41:09

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.081	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.121	22:36
Breath Volume	1.937	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.118	22:39
Breath Volume	2.250	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOPEZ,JACQIE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 03/09/2023 23:03:58

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.081	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.000	22:59
Breath Volume	2.558	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.000	23:02
Breath Volume	2.789	23:02
Air Blank	0.000	23:03
Control Test	0.081	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: LAGRAVE

DATE/TIME OF TEST: 03/10/2023 01:22:55

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.000	01:18
Breath Volume	2.769	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.000	01:21
Breath Volume	2.171	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: HATCH

DATE/TIME OF TEST: 03/11/2023 00:09:28

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.081	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.087	00:04
Breath Volume	1.730	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.084	00:08
Breath Volume	1.453	00:08
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: TRACY

DATE/TIME OF TEST: 03/11/2023 22:20:14

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.081	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.000	22:15
Breath Volume	1.769	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.000	22:18
Breath Volume	2.351	22:18
Air Blank	0.000	22:19
Control Test	0.082	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ELLIS,SYDNEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: FANES-SON

DATE/TIME OF TEST: 03/12/2023 05:38:47

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.081	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.165	05:34
Breath Volume	1.699	05:34
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.159	05:37
Breath Volume	3.421	05:37
Air Blank	0.000	05:37
Control Test	0.080	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: OUDERKIRK

DATE/TIME OF TEST: 03/14/2023 02:39:59

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.000	02:35
Breath Volume	3.652	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.000	02:38
Breath Volume	3.753	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: CHAPAS

DATE/TIME OF TEST: 03/15/2023 03:53:51

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.081	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.120	03:49
Breath Volume	2.429	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.119	03:52
Breath Volume	2.828	03:52
Air Blank	0.000	03:52
Control Test	0.081	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: FETTY

DATE/TIME OF TEST: 03/15/2023 04:07:38

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.153	04:03
Breath Volume	1.308	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.148	04:06
Breath Volume	1.343	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LINDEMANN, ANTHONY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/17/2023 00:21:44

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.000	00:17
Breath Volume	1.257	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.000	00:20
Breath Volume	1.902	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 04/07/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DELGATTI, BRAD
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: REMINICK

DATE/TIME OF TEST: 03/17/2023 01:33:18

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.091	01:28
Breath Volume	1.839	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.086	01:31
Breath Volume	1.390	01:31
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON,WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/17/2023 04:04:16

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.135	03:59
Breath Volume	1.617	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.126	04:02
Breath Volume	1.453	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: HATTON

DATE/TIME OF TEST: 03/17/2023 20:55:32

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.080	20:50
Air Blank	0.000	20:50
Subject Sample # 1	0.172	20:50
Breath Volume	1.363	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.163	20:54
Breath Volume	1.152	20:54
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 03/17/2023 23:03:25

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.000	22:59
Breath Volume	2.777	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.000	23:02
Breath Volume	1.660	23:02
Air Blank	0.000	23:02
Control Test	0.081	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MELLO

DATE/TIME OF TEST: 03/17/2023 23:57:24

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.094	23:53
Breath Volume	1.640	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.094	23:55
Breath Volume	1.757	23:55
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO PEREZ,CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 03/18/2023 01:36:13

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.081	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.111	01:31
Breath Volume	2.933	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:33
Subject Sample # 2	0.113	01:34
Breath Volume	3.269	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGIEL,ADRIAN,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARCHIORI,JONATHAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: DIAZ MARTINEZ

DATE/TIME OF TEST: 03/18/2023 04:44:22

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:41	
Control Test	0.081	04:41	
Air Blank	0.000	04:41	
Subject Sample	# 1	04:43	RFI Detect
Breath Volume	0.000	04:43	
Air Blank		04:44	RFI Detect
Air Blank		04:44	RFI Detect

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGIEL,ADRIAN,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARCHIORI,JONATHAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: DIAZ MARTINEZ

DATE/TIME OF TEST: 03/18/2023 04:57:25

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:47	
Control Test	0.081	04:48	
Air Blank	0.000	04:48	
Subject Sample	# 1 0.183	04:49	No .020 agreement
Breath Volume	2.488	04:49	
Air Blank	0.000	04:49	
Air Blank	0.000	04:51	
Subject Sample	# 2 0.160	04:52	No .020 agreement
Breath Volume	2.796	04:52	
Air Blank	0.000	04:53	
Air Blank	0.000	04:55	
Subject Sample	# 3 0.205	04:55	No .020 agreement
Breath Volume	2.457	04:55	
Air Blank	0.000	04:56	
Control Test	0.080	04:56	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGIEL,ADRIAN,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CHESNUT,JONATHAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 03/18/2023 05:53:12

Diagnostics Check	OK	05:46
Air Blank	0.000	05:46
Control Test	0.081	05:47
Air Blank	0.000	05:47
Subject Sample # 1	0.261	05:48
Breath Volume	1.523	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.269	05:51
Breath Volume	1.597	05:51
Air Blank	0.000	05:52
Control Test	0.080	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ARNOLD,PAUL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:57

SUBJECT'S LAST NAME: MONVILLE

DATE/TIME OF TEST: 03/18/2023 09:24:51

Diagnostics Check	OK	09:18
Air Blank	0.000	09:18
Control Test	0.082	09:19
Air Blank	0.000	09:19
Subject Sample # 1	0.144	09:20
Breath Volume	3.023	09:20
Air Blank	0.000	09:21
Air Blank	0.000	09:22
Subject Sample # 2	0.138	09:23
Breath Volume	2.460	09:23
Air Blank	0.000	09:23
Control Test	0.081	09:24
Air Blank	0.000	09:24
Diagnostics Check	OK	09:24

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: RIVERA RIVERA

DATE/TIME OF TEST: 03/19/2023 03:05:45

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.172	03:01
Breath Volume	2.222	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.168	03:04
Breath Volume	2.851	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANKOSKY,DONALD,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FACCIOLO

DATE/TIME OF TEST: 03/19/2023 23:27:32

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.229	23:22
Breath Volume	3.085	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.228	23:26
Breath Volume	2.820	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,FRANCISCO,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINEZ,FRANCISCO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 03/20/2023 03:24:39

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.135	03:20
Breath Volume	3.453	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.117	03:23
Breath Volume	1.320	03:23
Air Blank	0.000	03:23
Control Test	0.080	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,JILLIAN,F
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FRAZZETTO,ASHLEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 03/22/2023 05:15:04

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.082	04:59	
Air Blank	0.000	04:59	
Subject Sample	# 1 0.049	05:02	
Breath Volume	1.824	05:02	
Air Blank	0.000	05:03	
Air Blank	0.000	05:04	
Subject Sample	# 2	05:08	No Sample Provided
Breath Volume	0.000	05:08	
Air Blank	0.000	05:08	
Air Blank	0.000	05:10	
Subject Sample	# 3	05:13	No Sample Provided
Breath Volume	0.000	05:13	
Air Blank	0.000	05:14	
Control Test	0.081	05:14	
Air Blank	0.000	05:14	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: GREEN IV

DATE/TIME OF TEST: 03/23/2023 04:11:27

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.124	04:06
Breath Volume	1.640	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.124	04:09
Breath Volume	1.593	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: BAUMAN

DATE/TIME OF TEST: 03/24/2023 02:29:22

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.081	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.190	02:25
Breath Volume	1.957	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.182	02:27
Breath Volume	3.144	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 03/25/2023 00:27:33

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.081	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.087	00:23
Breath Volume	2.250	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.092	00:26
Breath Volume	1.851	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: MAYFIELD

DATE/TIME OF TEST: 03/25/2023 02:47:42

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.080	02:39	
Air Blank	0.000	02:40	
Subject Sample	# 1 0.225	02:43	Volume Not Met
Breath Volume	0.089	02:43	
Air Blank	0.000	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 2	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.079	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLAUSNER,JAMIE,N
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,KWESI
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GUSTAFSON

DATE/TIME OF TEST: 03/25/2023 03:56:57

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.211	03:52
Breath Volume	3.136	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.206	03:55
Breath Volume	2.019	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESNUT,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HOWES,JOHN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: MCGIFFEN

DATE/TIME OF TEST: 03/27/2023 21:37:28

Diagnostics Check	OK	21:36	
Air Blank		21:37	RFI Detect
Air Blank	0.000	21:37	

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GUEVARA,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:22

SUBJECT'S LAST NAME: MAHONEY

DATE/TIME OF TEST: 03/28/2023 09:51:22

Diagnostics Check	OK	09:44
Air Blank	0.000	09:45
Control Test	0.082	09:45
Air Blank	0.000	09:45
Subject Sample # 1	0.000	09:46
Breath Volume	1.277	09:46
Air Blank	0.000	09:47
Air Blank	0.000	09:49
Subject Sample # 2	0.000	09:50
Breath Volume	1.375	09:50
Air Blank	0.000	09:50
Control Test	0.081	09:50
Air Blank	0.000	09:51
Diagnostics Check	OK	09:51

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DELGATTI, BRADLEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: DIAZ CRUZ

DATE/TIME OF TEST: 03/29/2023 20:59:15

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.081	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.000	20:54
Breath Volume	3.902	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.000	20:57
Breath Volume	3.671	20:57
Air Blank	0.000	20:58
Control Test	0.081	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PANTOJA, ABRAHAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: ROSSETTE

DATE/TIME OF TEST: 03/30/2023 18:40:10

Diagnostics Check	OK	18:33
Air Blank	0.000	18:34
Control Test	0.081	18:34
Air Blank	0.000	18:35
Subject Sample # 1	0.360	18:35
Breath Volume	3.117	18:35
Air Blank	0.000	18:36
Air Blank	0.000	18:38
Subject Sample # 2	0.356	18:38
Breath Volume	2.863	18:38
Air Blank	0.000	18:39
Control Test	0.080	18:39
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

Data Download Date/Time: 04/07/2023 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 03/02/2023 21:33:21

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.079	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.349	21:28
Breath Volume	3.355	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.366	21:31
Breath Volume	1.738	21:31
Air Blank	0.000	21:32
Control Test	0.078	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 03/30/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: KITCHENS

DATE/TIME OF TEST: 03/06/2023 20:14:58

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.079	20:09
Air Blank	0.000	20:09
Subject Sample # 1	0.161	20:10
Breath Volume	3.476	20:10
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.153	20:13
Breath Volume	3.691	20:13
Air Blank	0.000	20:14
Control Test	0.078	20:14
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 03/30/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 03/08/2023 02:17:25

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.111	02:12
Breath Volume	2.042	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.117	02:15
Breath Volume	4.125	02:15
Air Blank	0.000	02:16
Control Test	0.078	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 03/30/2023 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: SAVIO

DATE/TIME OF TEST: 03/12/2023 16:55:06

Diagnostics Check	OK	16:49
Air Blank	0.000	16:49
Control Test	0.078	16:49
Air Blank	0.000	16:50
Subject Sample # 1	0.253	16:50
Breath Volume	2.253	16:50
Air Blank	0.000	16:51
Air Blank	0.000	16:53
Subject Sample # 2	0.250	16:53
Breath Volume	2.316	16:53
Air Blank	0.000	16:54
Control Test	0.077	16:54
Air Blank	0.000	16:55
Diagnostics Check	OK	16:55

Data Download Date/Time: 03/30/2023 10:24