

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: DIEDESCH,LAURA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: DRAPEAU

DATE/TIME OF TEST: 03/06/2023 04:10:21

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.078	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.229	04:02
Breath Volume	1.507	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.199	04:05
Breath Volume	1.324	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 3	0.225	04:08
Breath Volume	1.585	04:08
Air Blank	0.000	04:09
Control Test	0.077	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 04/21/2023 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: CHAPPELL

DATE/TIME OF TEST: 03/18/2023 22:19:20

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:05	
Control Test	0.078	22:05	
Air Blank	0.000	22:05	
Subject Sample	# 1 0.205	22:06	
Breath Volume	1.214	22:06	
Air Blank	0.000	22:07	
Air Blank	0.000	22:09	
Subject Sample	# 2 0.181	22:12	Volume Not Met
Breath Volume	0.023	22:12	
Air Blank	0.000	22:12	
Air Blank	0.000	22:14	
Subject Sample	# 3	22:17	No Sample Provided
Breath Volume	0.000	22:17	
Air Blank	0.000	22:18	
Control Test	0.077	22:18	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

Data Download Date/Time: 04/21/2023 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CORBITT,JAMEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 02/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MCCLLOUD

DATE/TIME OF TEST: 03/04/2023 22:53:21

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.079	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.183	22:47
Breath Volume	1.628	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.200	22:51
Breath Volume	1.367	22:51
Air Blank	0.000	22:52
Control Test	0.077	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 03/29/2023 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSSI,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TORRES,LYDIA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 02/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 03/22/2023 16:08:42

Diagnostics Check	OK	16:02
Air Blank	0.000	16:02
Control Test	0.080	16:03
Air Blank	0.000	16:03
Subject Sample # 1	0.120	16:04
Breath Volume	1.488	16:04
Air Blank	0.000	16:04
Air Blank	0.000	16:06
Subject Sample # 2	0.126	16:07
Breath Volume	1.750	16:07
Air Blank	0.000	16:07
Control Test	0.078	16:08
Air Blank	0.000	16:08
Diagnostics Check	OK	16:08

Data Download Date/Time: 03/29/2023 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WATTS,SHANNON,N

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: RIDEOUT,TIMOTHY

ARRESTING OFFICER AGENCY: Deland PD

LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 23:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HERNANDEZ CANO

DATE/TIME OF TEST: 03/25/2023 00:07:40

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:05	
Control Test	0.079	00:05	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:06	
Control Test	0.079	00:07	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 04/04/2023 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WATTS,SHANNON,N

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: RIDEOUT,TIMOTHY

ARRESTING OFFICER AGENCY: Deland PD

LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 23:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HERNANDEZ CANO

DATE/TIME OF TEST: 03/25/2023 01:31:23

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.162	01:26
Breath Volume	1.660	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.163	01:29
Breath Volume	1.218	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 04/04/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLER,COREY,C
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: QUICK,JUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 03/25/2023 18:55:43

Diagnostics Check	OK	18:54	
Air Blank		18:55	Ambient Fail
Air Blank	0.000	18:55	

Data Download Date/Time: 04/04/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLER,COREY,C
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: QUICK,JUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 03/25/2023 19:05:17

Diagnostics Check	OK	18:58
Air Blank	0.000	18:58
Control Test	0.079	18:59
Air Blank	0.000	18:59
Subject Sample # 1	0.134	19:00
Breath Volume	1.699	19:00
Air Blank	0.000	19:01
Air Blank	0.000	19:02
Subject Sample # 2	0.135	19:03
Breath Volume	1.472	19:03
Air Blank	0.000	19:04
Control Test	0.079	19:04
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

Data Download Date/Time: 04/04/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLER,COREY,C
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: RIDEOUT,TIMOTHY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 03/26/2023 02:35:11

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:34	No Sample Provided
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Air Blank	0.000	02:35	

Data Download Date/Time: 04/04/2023 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLER,COREY,C
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: RIDEOUT,TIMOTHY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 03/26/2023 02:40:22

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Control Test	0.080	02:39	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 04/04/2023 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: CLARK,DAMON
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: MATLICK

DATE/TIME OF TEST: 03/29/2023 01:55:16

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.080	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:53	Subject Test Refused
Breath Volume	0.000	01:53	
Air Blank	0.000	01:54	
Control Test	0.080	01:54	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 04/04/2023 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINEY,CRISTA,N
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: JOHNSON,DAVID
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BLACKMON

DATE/TIME OF TEST: 03/03/2023 03:08:25

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.082	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.082	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 03/16/2023 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI,PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MARTINEZ III

DATE/TIME OF TEST: 03/04/2023 03:29:48

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:27	
Control Test	0.082	03:27	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:28	Subject Test Refused
Breath Volume	0.000	03:28	
Air Blank	0.000	03:28	
Control Test	0.083	03:29	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 03/16/2023 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILL,JOHN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BUECKER

DATE/TIME OF TEST: 03/05/2023 04:26:34

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.081	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.000	04:22
Breath Volume	2.031	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.000	04:25
Breath Volume	2.531	04:25
Air Blank	0.000	04:25
Control Test	0.082	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 03/16/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,C
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BALDINO

DATE/TIME OF TEST: 03/05/2023 19:37:33

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.081	19:31
Air Blank	0.000	19:32
Subject Sample # 1	0.280	19:33
Breath Volume	1.464	19:33
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.289	19:35
Breath Volume	1.339	19:35
Air Blank	0.000	19:36
Control Test	0.081	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 03/16/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDY,K
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: MEEKINS

DATE/TIME OF TEST: 03/09/2023 21:01:53

Diagnostics Check	OK	20:59	
Air Blank	0.000	20:59	
Control Test	0.082	20:59	
Air Blank	0.000	21:00	
Subject Sample	# 1	21:00	Subject Test Refused
Breath Volume	0.000	21:00	
Air Blank	0.000	21:00	
Control Test	0.082	21:01	
Air Blank	0.000	21:01	
Diagnostics Check	OK	21:01	

Data Download Date/Time: 03/16/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 03/11/2023 03:52:07

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.082	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.046	03:47
Breath Volume	3.769	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.046	03:50
Breath Volume	3.945	03:50
Air Blank	0.000	03:51
Control Test	0.083	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 03/16/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISS,ADAM,B
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DORSEY,JAXON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 03/12/2023 05:18:44

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:14	
Control Test	0.081	05:15	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:17	Subject Test Refused
Breath Volume	0.000	05:17	
Air Blank	0.000	05:17	
Control Test	0.081	05:18	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:18	

Data Download Date/Time: 03/16/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LOGSDON,D
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 03/14/2023 01:02:08

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.217	00:57
Breath Volume	1.226	00:57
Air Blank	0.000	00:58
Air Blank	0.000	00:59
Subject Sample # 2	0.205	01:00
Breath Volume	1.691	01:00
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 03/16/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BRYAN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: COZZONE,GAETANO
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 02/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: ORDONEZ ORTIZ

DATE/TIME OF TEST: 03/15/2023 19:07:49

Diagnostics Check	OK	19:00
Air Blank	0.000	19:00
Control Test	0.081	19:01
Air Blank	0.000	19:01
Subject Sample # 1	0.275	19:02
Breath Volume	1.113	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:04
Subject Sample # 2	0.276	19:06
Breath Volume	1.824	19:06
Air Blank	0.000	19:06
Control Test	0.081	19:07
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 03/16/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: FRIEDMAN

DATE/TIME OF TEST: 03/16/2023 04:24:48

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:24	
		04:24	RFI Detect
Air Blank	0.000	04:24	

Data Download Date/Time: 04/12/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: HERNANDEZ ALMEIDA

DATE/TIME OF TEST: 03/17/2023 02:15:23

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.082	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.158	02:09
Breath Volume	1.222	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample # 2	0.166	02:13
Breath Volume	1.691	02:13
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 04/12/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 03/18/2023 02:00:31

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.187	01:54
Breath Volume	1.855	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.170	01:59
Breath Volume	1.644	01:59
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 04/12/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 03/23/2023 03:12:06

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.169	03:07
Breath Volume	4.027	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.168	03:10
Breath Volume	3.230	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:12

Data Download Date/Time: 04/12/2023 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:47

SUBJECT'S LAST NAME: ARGO

DATE/TIME OF TEST: 03/13/2023 16:17:54

Diagnostics Check	OK	16:10
Air Blank	0.000	16:10
Control Test	0.081	16:10
Air Blank	0.000	16:11
Subject Sample # 1	0.000	16:13
Breath Volume	2.335	16:13
Air Blank	0.000	16:13
Air Blank	0.000	16:15
Subject Sample # 2	0.000	16:16
Breath Volume	2.007	16:16
Air Blank	0.000	16:16
Control Test	0.081	16:17
Air Blank	0.000	16:17
Diagnostics Check	OK	16:17

Data Download Date/Time: 04/10/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 03/18/2023 01:48:50

Diagnostics Check	OK	01:48	
Air Blank		01:48	RFI Detect
Air Blank	0.000	01:48	

Data Download Date/Time: 04/10/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 03/18/2023 01:57:03

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.081	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.096	01:52
Breath Volume	2.921	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.094	01:55
Breath Volume	3.429	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:57

Data Download Date/Time: 04/10/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CASTILLO AVELINO

DATE/TIME OF TEST: 03/25/2023 01:28:08

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.081	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.177	01:23
Breath Volume	2.453	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:25
Subject Sample # 2	0.182	01:26
Breath Volume	3.246	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 04/10/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: LORENZO GREGORIO

DATE/TIME OF TEST: 03/25/2023 21:07:10

Diagnostics Check	OK	20:57	
Air Blank	0.000	20:57	
Control Test	0.081	20:57	
Air Blank	0.000	20:58	
Subject Sample	# 1	21:01	No Sample Provided
Breath Volume	0.000	21:01	
Air Blank	0.000	21:02	
Air Blank	0.000	21:04	
Subject Sample	# 2	21:05	Subject Test Refused
Breath Volume	0.000	21:05	
Air Blank	0.000	21:06	
Control Test	0.081	21:06	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time: 04/10/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: KEINZ

DATE/TIME OF TEST: 03/31/2023 11:42:40

Diagnostics Check	OK	11:36
Air Blank	0.000	11:36
Control Test	0.079	11:36
Air Blank	0.000	11:37
Subject Sample # 1	0.183	11:37
Breath Volume	1.304	11:37
Air Blank	0.000	11:38
Air Blank	0.000	11:40
Subject Sample # 2	0.198	11:40
Breath Volume	2.332	11:40
Air Blank	0.000	11:41
Control Test	0.077	11:42
Air Blank	0.000	11:42
Diagnostics Check	OK	11:42

Data Download Date/Time: 04/10/2023 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: FRIESE

DATE/TIME OF TEST: 03/31/2023 19:11:22

Diagnostics Check	OK	19:04
Air Blank	0.000	19:04
Control Test	0.081	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.192	19:06
Breath Volume	2.234	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.195	19:09
Breath Volume	2.136	19:09
Air Blank	0.000	19:10
Control Test	0.079	19:10
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 04/10/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU, ARIC, J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ, SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MATALONIS JR

DATE/TIME OF TEST: 03/03/2023 02:24:17

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.081	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 1 0.112	02:14	Volume Not Met
Breath Volume	0.203	02:14	
Air Blank	0.000	02:15	
Air Blank	0.000	02:17	
Subject Sample	# 2 0.123	02:18	
Breath Volume	1.761	02:18	
Air Blank	0.000	02:19	
Air Blank	0.000	02:21	
Subject Sample	# 3 0.128	02:22	
Breath Volume	2.394	02:22	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIK
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 03/03/2023 02:57:04

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.215	02:52
Breath Volume	1.636	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.206	02:55
Breath Volume	1.593	02:55
Air Blank	0.000	02:56
Control Test	0.078	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:57

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: RIZZO

DATE/TIME OF TEST: 03/03/2023 03:09:07

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.141	03:04
Breath Volume	1.402	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:06
Subject Sample # 2	0.135	03:07
Breath Volume	1.285	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/03/2023 03:36:22

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.173	03:31
Breath Volume	1.394	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.177	03:34
Breath Volume	1.367	03:34
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: KNAPP

DATE/TIME OF TEST: 03/04/2023 00:33:39

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.081	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.024	00:29
Breath Volume	4.539	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.023	00:32
Breath Volume	3.335	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: KUHN

DATE/TIME OF TEST: 03/04/2023 01:54:29

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.124	01:50
Breath Volume	1.734	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.114	01:52
Breath Volume	1.292	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: GOMOS JACOME

DATE/TIME OF TEST: 03/04/2023 03:12:02

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:05	No Sample Provided
Breath Volume	0.000	03:05	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 2	03:10	No Sample Provided
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.080	03:11	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 03/04/2023 05:09:21

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.185	05:04
Breath Volume	1.171	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.189	05:07
Breath Volume	1.214	05:07
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA,RICHARD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ANDREE,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: TREXLER

DATE/TIME OF TEST: 03/04/2023 19:07:29

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.080	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.220	19:02
Breath Volume	3.089	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.218	19:05
Breath Volume	4.878	19:05
Air Blank	0.000	19:06
Control Test	0.079	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BACA,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: BUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BYRNE

DATE/TIME OF TEST: 03/04/2023 23:56:42

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.061	23:52
Breath Volume	3.003	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.058	23:55
Breath Volume	2.308	23:55
Air Blank	0.000	23:55
Control Test	0.078	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ ORTEGA,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 03/05/2023 00:12:56

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.080	00:10	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Control Test	0.079	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTENSEN,S,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ ORTEGA,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 03/05/2023 05:47:05

Diagnostics Check	OK	05:39
Air Blank	0.000	05:39
Control Test	0.080	05:40
Air Blank	0.000	05:40
Subject Sample # 1	0.138	05:42
Breath Volume	1.445	05:42
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.142	05:45
Breath Volume	1.714	05:45
Air Blank	0.000	05:46
Control Test	0.080	05:46
Air Blank	0.000	05:47
Diagnostics Check	OK	05:47

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ANDREE,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: RAYMOND

DATE/TIME OF TEST: 03/05/2023 22:05:08

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.000	22:00
Breath Volume	4.074	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.000	22:03
Breath Volume	3.960	22:03
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 03/29/2023 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/07/2023 05:57:50

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.080	05:52
Air Blank	0.000	05:52
Subject Sample # 1	0.000	05:53
Breath Volume	1.976	05:53
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.000	05:56
Breath Volume	1.832	05:56
Air Blank	0.000	05:56
Control Test	0.080	05:57
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: EASON

DATE/TIME OF TEST: 03/08/2023 15:31:51

Diagnostics Check	OK	15:25
Air Blank	0.000	15:26
Control Test	0.080	15:26
Air Blank	0.000	15:27
Subject Sample # 1	0.000	15:27
Breath Volume	1.667	15:27
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.000	15:30
Breath Volume	1.636	15:30
Air Blank	0.000	15:30
Control Test	0.079	15:31
Air Blank	0.000	15:31
Diagnostics Check	OK	15:31

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 03/08/2023 19:50:27

Diagnostics Check	OK	19:39	
Air Blank	0.000	19:39	
Control Test	0.080	19:39	
Air Blank	0.000	19:40	
Subject Sample	# 1	19:43	No Sample Provided
Breath Volume	0.000	19:43	
Air Blank	0.000	19:43	
Air Blank	0.000	19:45	
Subject Sample	# 2	19:49	No Sample Provided
Breath Volume	0.000	19:49	
Air Blank	0.000	19:49	
Control Test	0.080	19:49	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: SHEELEY

DATE/TIME OF TEST: 03/08/2023 23:35:14

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.081	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.000	23:30
Breath Volume	1.464	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.000	23:33
Breath Volume	1.613	23:33
Air Blank	0.000	23:34
Control Test	0.078	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DOBRILA

DATE/TIME OF TEST: 03/09/2023 00:52:19

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:41	
Control Test	0.080	00:41	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:45	No Sample Provided
Breath Volume	0.000	00:45	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample	# 2	00:50	No Sample Provided
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOUCK, GENE, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BOUCK

DATE/TIME OF TEST: 03/09/2023 02:52:36

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.080	02:50	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/09/2023 03:28:22

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.163	03:23
Breath Volume	1.687	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.159	03:26
Breath Volume	1.843	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 03/09/2023 23:26:16

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:22	
Control Test	0.080	23:23	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:24	Subject Test Refused
Breath Volume	0.000	23:24	
Air Blank	0.000	23:25	
Control Test	0.080	23:25	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: CASTILLO-GARCIA

DATE/TIME OF TEST: 03/10/2023 00:56:13

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.222	00:51
Breath Volume	2.511	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.210	00:54
Breath Volume	2.414	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/10/2023 01:34:57

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.080	01:27	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:30	No Sample Provided
Breath Volume	0.000	01:30	
Air Blank	0.000	01:31	
Air Blank	0.000	01:33	
Subject Sample	# 2	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ANDREE,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: COATES

DATE/TIME OF TEST: 03/11/2023 21:27:45

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.080	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.350	21:19
Breath Volume	1.351	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.328	21:23
Breath Volume	1.562	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:25
Subject Sample # 3	0.316	21:26
Breath Volume	1.390	21:26
Air Blank	0.000	21:26
Control Test	0.075	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 03/18/2023 20:54:26

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.000	20:50
Breath Volume	2.136	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.000	20:53
Breath Volume	1.671	20:53
Air Blank	0.000	20:53
Control Test	0.080	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SESSIONS,C,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/22/2023 00:13:31

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.081	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.165	00:08
Breath Volume	1.164	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.172	00:11
Breath Volume	2.367	00:11
Air Blank	0.000	00:12
Control Test	0.079	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 03/29/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOKER,RONALD,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: STOKER,RONALD
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: BUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 03/25/2023 20:00:08

Diagnostics Check	OK	19:48	
Air Blank	0.000	19:48	
Control Test	0.080	19:49	
Air Blank	0.000	19:49	
Subject Sample	# 1 0.170	19:52	Volume Not Met
Breath Volume	0.144	19:52	
Air Blank	0.000	19:53	
Air Blank	0.000	19:55	
Subject Sample	# 2 0.153	19:58	Volume Not Met
Breath Volume	0.363	19:58	
Air Blank	0.000	19:59	
Control Test	0.079	19:59	
Air Blank	0.000	20:00	
Diagnostics Check	OK	20:00	

Data Download Date/Time: 03/26/2023 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DIAZ, DIANA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: CASSIDY

DATE/TIME OF TEST: 03/07/2023 23:23:50

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.082	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.225	23:19
Breath Volume	3.273	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.231	23:22
Breath Volume	1.734	23:22
Air Blank	0.000	23:22
Control Test	0.081	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 04/10/2023 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BRYAN,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: PHELPS,W
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/03/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 03/12/2023 01:00:33

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.082	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.243	00:56
Breath Volume	2.199	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.248	00:59
Breath Volume	2.628	00:59
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 04/10/2023 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 03/15/2023 23:52:43

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.082	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.055	23:48
Breath Volume	1.734	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.051	23:51
Breath Volume	1.867	23:51
Air Blank	0.000	23:51
Control Test	0.082	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 04/10/2023 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 03/16/2023 23:19:20

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.082	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.255	23:14
Breath Volume	2.101	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.271	23:17
Breath Volume	2.242	23:17
Air Blank	0.000	23:18
Control Test	0.082	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 04/10/2023 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: VASQUEZ ARREAGA

DATE/TIME OF TEST: 03/18/2023 20:35:38

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.081	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.165	20:31
Breath Volume	1.726	20:31
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.171	20:34
Breath Volume	1.367	20:34
Air Blank	0.000	20:34
Control Test	0.081	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 04/10/2023 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,GEORGE,H
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: EDWARDS,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: CHAMBERY

DATE/TIME OF TEST: 03/23/2023 00:08:10

Diagnostics Check	OK	23:59	
Air Blank	0.000	00:00	
Control Test	0.082	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:04	No Sample Provided
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Air Blank	0.000	00:06	
Subject Sample	# 2	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:07	
Control Test	0.082	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 04/10/2023 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAYNE,ANDREW,S
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BAYNE,ANDREW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: PULLIN

DATE/TIME OF TEST: 03/19/2023 04:25:08

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.265	04:20
Breath Volume	2.976	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:22
Subject Sample # 2	0.253	04:23
Breath Volume	1.957	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 04/21/2023 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISSONNETTE,JACOB,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CASON,RYAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: HEALY

DATE/TIME OF TEST: 03/20/2023 00:55:16

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.081	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.161	00:50
Breath Volume	3.136	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.161	00:53
Breath Volume	1.980	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 04/21/2023 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISSONNETTE,JACOB,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BROWN,ZECHARIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RAPALO

DATE/TIME OF TEST: 03/28/2023 03:17:33

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.080	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 1 0.181	03:12	
Breath Volume	1.695	03:12	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample	# 2 0.184	03:15	
Breath Volume	2.410	03:15	
Air Blank	0.000	03:16	
Control Test	0.081	03:16	
Air Blank		03:17	RFI Detect
Air Blank	0.000	03:17	

Data Download Date/Time: 04/21/2023 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GARNER,MICHAEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/05/2023 03:39:06

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.120	03:33
Breath Volume	1.457	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.135	03:37
Breath Volume	1.113	03:37
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 03/31/2023 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: LABRIE,DAVID
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 02/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 12:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: MOELLER

DATE/TIME OF TEST: 03/18/2023 13:33:59

Diagnostics Check	OK	13:27
Air Blank	0.000	13:28
Control Test	0.081	13:28
Air Blank	0.000	13:29
Subject Sample # 1	0.000	13:29
Breath Volume	2.222	13:29
Air Blank	0.000	13:30
Air Blank	0.000	13:32
Subject Sample # 2	0.000	13:32
Breath Volume	1.929	13:32
Air Blank	0.000	13:33
Control Test	0.080	13:33
Air Blank	0.000	13:33
Diagnostics Check	OK	13:33

Data Download Date/Time: 03/31/2023 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BERRIDGE,K
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: LONGWELL

DATE/TIME OF TEST: 03/07/2023 00:23:22

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.157	00:18
Breath Volume	1.269	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.153	00:21
Breath Volume	1.886	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 03/24/2023 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LOWE,MAYCEE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 03/18/2023 01:09:59

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.079	01:07	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Control Test	0.079	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 03/24/2023 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: TUFTS,KEVIN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/26/2023 00:12:33

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.173	00:08
Breath Volume	1.210	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.154	00:11
Breath Volume	1.589	00:11
Air Blank	0.000	00:11
Control Test	0.077	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 04/24/2023 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,J,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 03/01/2023 22:18:20

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.155	22:13
Breath Volume	3.324	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:15
Subject Sample # 2	0.148	22:16
Breath Volume	2.367	22:16
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 03/22/2023 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHRADER,E,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: RICHARDSON,VINCE
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: CORRAL

DATE/TIME OF TEST: 03/03/2023 00:46:22

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.079	00:35	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:39	No Sample Provided
Breath Volume	0.000	00:39	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 2	00:44	No Sample Provided
Breath Volume	0.000	00:44	
Air Blank	0.000	00:45	
Control Test	0.080	00:45	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 03/22/2023 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ROLLINS,SEAN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 02/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/05/2023 19:58:04

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:55	
Control Test	0.080	19:55	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:56	Subject Test Refused
Breath Volume	0.000	19:56	
Air Blank	0.000	19:57	
Control Test	0.080	19:57	
Air Blank	0.000	19:57	
Diagnostics Check	OK	19:58	

Data Download Date/Time: 03/22/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,R,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: PENA,R
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: KNEAL

DATE/TIME OF TEST: 03/18/2023 01:07:31

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:56	
Control Test	0.080	00:56	
Air Blank	0.000	00:57	
Subject Sample	# 1	01:00	No Sample Provided
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Air Blank	0.000	01:03	
Subject Sample	# 2	01:06	No Sample Provided
Breath Volume	0.000	01:06	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 03/22/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,R,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: PENA,R
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: IZQUIERDO SOTOLONGO

DATE/TIME OF TEST: 03/19/2023 20:57:26

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.081	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.219	20:52
Breath Volume	2.156	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.227	20:55
Breath Volume	2.414	20:55
Air Blank	0.000	20:56
Control Test	0.081	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 03/22/2023 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: PEREZ,FRANCISCO
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 02/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 03/20/2023 03:29:07

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.081	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 03/22/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,E
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 03/26/2023 02:43:58

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.080	02:33	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:36	No Sample Provided
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 2	02:42	No Sample Provided
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.080	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 04/19/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 03/04/2023 20:18:19

Diagnostics Check	OK	20:12
Air Blank	0.000	20:12
Control Test	0.080	20:12
Air Blank	0.000	20:13
Subject Sample # 1	0.178	20:13
Breath Volume	3.488	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.179	20:16
Breath Volume	3.027	20:16
Air Blank	0.000	20:17
Control Test	0.080	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 03/14/2023 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: CAMPO LOPEZ

DATE/TIME OF TEST: 03/04/2023 23:58:35

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:48	
Control Test	0.080	23:48	
Air Blank	0.000	23:49	
Subject Sample	# 1 0.257	23:51	
Breath Volume	1.402	23:51	
Air Blank	0.000	23:52	
Air Blank	0.000	23:54	
Subject Sample	# 2	23:57	Subject Test Refused
Breath Volume	0.289	23:57	
Air Blank	0.000	23:57	
Control Test	0.080	23:57	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 03/14/2023 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: ALVAREZ ACOSTA

DATE/TIME OF TEST: 03/05/2023 02:50:02

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.080	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank		02:49	RFI Detect
Air Blank	0.000	02:50	

Data Download Date/Time: 03/14/2023 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: ALVAREZ ACOSTA

DATE/TIME OF TEST: 03/05/2023 02:55:48

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.080	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.080	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 03/14/2023 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: CADY

DATE/TIME OF TEST: 03/03/2023 00:56:25

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.179	00:51
Breath Volume	2.152	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.181	00:54
Breath Volume	2.261	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEITFELD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: HEITFELD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: SODINI

DATE/TIME OF TEST: 03/03/2023 16:29:11

Diagnostics Check	OK	16:23
Air Blank	0.000	16:23
Control Test	0.080	16:23
Air Blank	0.000	16:24
Subject Sample # 1	0.162	16:24
Breath Volume	1.652	16:24
Air Blank	0.000	16:25
Air Blank	0.000	16:27
Subject Sample # 2	0.156	16:27
Breath Volume	1.488	16:27
Air Blank	0.000	16:28
Control Test	0.079	16:28
Air Blank	0.000	16:29
Diagnostics Check	OK	16:29

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERIC,JEAN,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FREDERIC,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 03/05/2023 05:25:47

Diagnostics Check	OK	05:22	
Air Blank	0.000	05:23	
Control Test	0.080	05:23	
Air Blank	0.000	05:24	
Subject Sample	# 1	05:24	Subject Test Refused
Breath Volume	0.000	05:24	
Air Blank	0.000	05:24	
Control Test	0.080	05:25	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 03/05/2023 19:15:13

Diagnostics Check	OK	19:08
Air Blank	0.000	19:08
Control Test	0.080	19:08
Air Blank	0.000	19:09
Subject Sample # 1	0.224	19:10
Breath Volume	1.109	19:10
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.215	19:13
Breath Volume	1.808	19:13
Air Blank	0.000	19:14
Control Test	0.078	19:14
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPUGNARDI,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SANTAMARIA,LEONARDO
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 03/06/2023 00:41:27

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.170	00:36
Breath Volume	3.457	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.170	00:39
Breath Volume	2.910	00:39
Air Blank	0.000	00:40
Control Test	0.078	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: CERNUTO,MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: LESLIE

DATE/TIME OF TEST: 03/07/2023 01:07:02

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.080	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ESCALANTE-ESCALANTE

DATE/TIME OF TEST: 03/07/2023 03:21:36

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.287	03:13
Breath Volume	2.011	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.253	03:17
Breath Volume	1.351	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 3	0.288	03:20
Breath Volume	2.175	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: FORBES,RAY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: COMO

DATE/TIME OF TEST: 03/08/2023 02:06:39

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.192	02:01
Breath Volume	2.558	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.201	02:05
Breath Volume	2.843	02:05
Air Blank	0.000	02:05
Control Test	0.079	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: CHAN

DATE/TIME OF TEST: 03/08/2023 17:23:09

Diagnostics Check	OK	17:14
Air Blank	0.000	17:14
Control Test	0.080	17:15
Air Blank	0.000	17:15
Subject Sample # 1	0.260	17:18
Breath Volume	1.101	17:18
Air Blank	0.000	17:19
Air Blank	0.000	17:20
Subject Sample # 2	0.258	17:21
Breath Volume	1.320	17:21
Air Blank	0.000	17:22
Control Test	0.078	17:22
Air Blank	0.000	17:23
Diagnostics Check	OK	17:23

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERIC,JEAN,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FREDERIC,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 03/09/2023 00:02:43

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.079	23:55
Air Blank	0.000	23:55
Subject Sample # 2	0.204	23:58
Breath Volume	2.183	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.197	00:01
Breath Volume	2.121	00:01
Air Blank	0.000	00:01
Control Test	0.076	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 03/09/2023 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PETERS,DARIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: TORREZ

DATE/TIME OF TEST: 03/19/2023 00:44:53

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.077	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.175	00:40
Breath Volume	1.800	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.176	00:43
Breath Volume	2.117	00:43
Air Blank	0.000	00:43
Control Test	0.077	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 04/07/2023 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,ELISHA,D
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CASTLE,ALICIA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 03/26/2023 23:03:11

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:55	
Control Test	0.078	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:59	No Sample Provided
Breath Volume	0.000	22:59	
Air Blank	0.000	22:59	
Air Blank	0.000	23:01	
Subject Sample	# 2	23:01	Subject Test Refused
Breath Volume	0.000	23:01	
Air Blank	0.000	23:02	
Control Test	0.078	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 04/07/2023 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,ELISHA,D
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CERNUTO,ELISHA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: D'ASCANIO

DATE/TIME OF TEST: 03/27/2023 02:57:04

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.028	02:52
Breath Volume	2.160	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.025	02:55
Breath Volume	3.019	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:57

Data Download Date/Time: 04/07/2023 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: VIDAL GONGORA

DATE/TIME OF TEST: 03/10/2023 23:52:40

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:49	
Control Test	0.081	23:49	
Air Blank	0.000	23:50	
Subject Sample	# 1	23:51	RFI Detect
Breath Volume	0.000	23:51	
Air Blank	0.000	23:52	
Control Test	0.081	23:52	
Air Blank		23:52	RFI Detect
Air Blank		23:52	RFI Detect

Data Download Date/Time: 04/11/2023 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MARTINEZ GOMEZ

DATE/TIME OF TEST: 03/12/2023 00:47:17

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:36	
Control Test	0.081	00:36	
Air Blank	0.000	00:36	
Subject Sample	# 1 0.069	00:40	Volume Not Met
Breath Volume	0.398	00:40	
Air Blank	0.000	00:40	
Air Blank	0.000	00:42	
Subject Sample	# 2 0.097	00:45	Volume Not Met
Breath Volume	0.152	00:45	
Air Blank	0.000	00:46	
Control Test	0.081	00:46	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 04/11/2023 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 03/18/2023 18:47:40

Diagnostics Check	OK	18:43	
Air Blank	0.000	18:44	
Control Test	0.083	18:44	
Air Blank	0.000	18:45	
Subject Sample	# 1	18:46	Subject Test Refused
Breath Volume	0.000	18:46	
Air Blank	0.000	18:46	
Control Test	0.082	18:47	
Air Blank	0.000	18:47	
Diagnostics Check	OK	18:47	

Data Download Date/Time: 04/11/2023 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 03/21/2023 00:20:07

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.083	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.000	00:15
Breath Volume	1.281	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.000	00:18
Breath Volume	1.460	00:18
Air Blank	0.000	00:19
Control Test	0.082	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 04/11/2023 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: WINKLE

DATE/TIME OF TEST: 03/24/2023 17:37:42

Diagnostics Check	OK	17:26	
Air Blank	0.000	17:26	
Control Test	0.081	17:26	
Air Blank	0.000	17:27	
Subject Sample	# 1 0.147	17:30	Volume Not Met
Breath Volume	0.066	17:30	
Air Blank	0.000	17:31	
Air Blank	0.000	17:33	
Subject Sample	# 2 0.161	17:36	Volume Not Met
Breath Volume	0.117	17:36	
Air Blank	0.000	17:36	
Control Test	0.081	17:37	
Air Blank	0.000	17:37	
Diagnostics Check	OK	17:37	

Data Download Date/Time: 04/11/2023 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: GOBER,JOSHUA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HIMES

DATE/TIME OF TEST: 03/31/2023 03:27:03

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.082	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.228	03:22
Breath Volume	1.117	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.228	03:25
Breath Volume	1.109	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:27

Data Download Date/Time: 04/11/2023 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 03/31/2023 22:59:45

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.102	22:55
Breath Volume	1.296	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.107	22:58
Breath Volume	1.437	22:58
Air Blank	0.000	22:58
Control Test	0.081	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 04/11/2023 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: CAZAD,AUNDREA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 03/31/2023 23:37:15

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.081	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.011	23:32
Breath Volume	1.554	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.009	23:35
Breath Volume	1.378	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 04/11/2023 14:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTCHINSON,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HUTCHINSON,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: PRESCOTT

DATE/TIME OF TEST: 03/02/2023 00:15:42

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.076	00:13	
Air Blank	0.000	00:13	
Subject Sample	# 1	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:14	
Control Test	0.077	00:15	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONAHAN,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MONAHAN,TIMOTHY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: SALZMAN

DATE/TIME OF TEST: 03/02/2023 06:11:59

Diagnostics Check	OK	06:05
Air Blank	0.000	06:06
Control Test	0.077	06:06
Air Blank	0.000	06:07
Subject Sample # 1	0.000	06:07
Breath Volume	3.589	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 2	0.000	06:10
Breath Volume	3.105	06:10
Air Blank	0.000	06:11
Control Test	0.076	06:11
Air Blank	0.000	06:11
Diagnostics Check	OK	06:11

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 13:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: BEENE

DATE/TIME OF TEST: 03/03/2023 15:05:35

Diagnostics Check	OK	14:52	
Air Blank	0.000	14:52	
Control Test	0.076	14:52	
Air Blank	0.000	14:53	
Subject Sample	# 1 0.208	14:56	Volume Not Met
Breath Volume	0.062	14:56	
Air Blank	0.000	14:57	
Air Blank	0.000	14:59	
Subject Sample	# 2 0.188	15:00	
Breath Volume	1.285	15:00	
Air Blank	0.000	15:00	
Air Blank	0.000	15:02	
Subject Sample	# 3 0.203	15:04	
Breath Volume	1.464	15:04	
Air Blank	0.000	15:04	
Control Test	0.074	15:05	Control Outside Tolerance
Air Blank	0.000	15:05	

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 13:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: BEENE

DATE/TIME OF TEST: 03/03/2023 15:08:50

Diagnostics Check	OK	15:07	
Air Blank	0.000	15:07	
Control Test	0.074	15:08	Control Outside Tolerance
Air Blank	0.000	15:08	

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MINNICK

DATE/TIME OF TEST: 03/04/2023 20:48:23

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.077	20:42
Air Blank	0.000	20:42
Subject Sample # 1	0.189	20:43
Breath Volume	1.492	20:43
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.178	20:46
Breath Volume	1.890	20:46
Air Blank	0.000	20:47
Control Test	0.076	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTCHINSON,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HUTCHINSON,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/05/2023 03:27:38

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.076	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Control Test	0.076	03:27	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: DE LA CRUZ,JUAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/06/2023 02:02:13

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.075	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1 0.116	01:55	Volume Not Met
Breath Volume	0.035	01:55	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample	# 2 0.084	02:00	Volume Not Met
Breath Volume	0.191	02:00	
Air Blank	0.000	02:01	
Control Test	0.075	02:01	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: DE LA CRUZ,JUAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: 0000000000

DATE/TIME OF TEST: 03/06/2023 02:15:56

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.075	02:09	
Air Blank	0.000	02:10	
Subject Sample # 1	0.101	02:11	
Breath Volume	2.453	02:11	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample # 2	0.101	02:14	
Breath Volume	2.414	02:14	
Air Blank	0.000	02:15	
Control Test	0.074	02:15	Control Outside Tolerance
Air Blank	0.000	02:15	

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CRUZ,ADAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 03/09/2023 02:02:48

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.076	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1 0.109	01:57	
Breath Volume	1.816	01:57	
Air Blank	0.000	01:58	
Air Blank	0.000	02:00	
Subject Sample	# 2 0.114	02:01	
Breath Volume	3.382	02:01	
Air Blank	0.000	02:01	
Control Test	0.074	02:02	Control Outside Tolerance
Air Blank	0.000	02:02	

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,CALE,V
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SMITH,CALE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: LESNER

DATE/TIME OF TEST: 03/17/2023 17:39:42

Diagnostics Check	OK	17:36	
Air Blank	0.000	17:37	
Control Test	0.075	17:37	
Air Blank	0.000	17:37	
Subject Sample	# 1	17:38	Subject Test Refused
Breath Volume	0.000	17:38	
Air Blank	0.000	17:38	
Control Test	0.076	17:39	
Air Blank	0.000	17:39	
Diagnostics Check	OK	17:39	

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,CALE,V
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SMITH,CALE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: PRUITT

DATE/TIME OF TEST: 03/18/2023 23:33:13

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.076	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1 0.234	23:25	Volume Not Met
Breath Volume	0.082	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2 0.258	23:31	Volume Not Met
Breath Volume	0.199	23:31	
Air Blank	0.000	23:32	
Control Test	0.074	23:32	Control Outside Tolerance
Air Blank	0.000	23:33	

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RYAN,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROWN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/23/2023 00:57:24

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.081	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.201	00:53
Breath Volume	2.398	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.192	00:55
Breath Volume	1.847	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,CALE,V
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SMITH,CALE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 03/24/2023 22:38:22

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.081	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.199	22:33
Breath Volume	2.734	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.201	22:36
Breath Volume	2.843	22:36
Air Blank	0.000	22:37
Control Test	0.079	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 04/10/2023 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/25/2023 04:16:58

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.180	04:12
Breath Volume	1.527	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample # 2	0.189	04:15
Breath Volume	2.375	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 04/10/2023 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PAUGA,JOHN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: ESPARZA

DATE/TIME OF TEST: 03/25/2023 18:38:56

Diagnostics Check	OK	18:30
Air Blank	0.000	18:31
Control Test	0.080	18:31
Air Blank	0.000	18:32
Subject Sample # 1	0.319	18:34
Breath Volume	1.316	18:34
Air Blank	0.000	18:35
Air Blank	0.000	18:36
Subject Sample # 2	0.339	18:37
Breath Volume	1.687	18:37
Air Blank	0.000	18:38
Control Test	0.079	18:38
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 04/10/2023 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: TRICE

DATE/TIME OF TEST: 03/26/2023 03:06:44

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.102	03:02
Breath Volume	3.652	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.098	03:05
Breath Volume	3.257	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 04/10/2023 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

INSTRUMENT REGISTERED TO: NAS PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,WILLIAM,W
BREATH TEST OPERATOR AGENCY: NAS Pensacola
ARRESTING OFFICER: MILLER,WILLIAM
ARRESTING OFFICER AGENCY: NAS Pensacola
LAST AGENCY INSPECTION: 02/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RIMER

DATE/TIME OF TEST: 03/04/2023 03:48:59

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.079	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:47	RFI Detect
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.079	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 03/05/2023 23:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

INSTRUMENT REGISTERED TO: NAS PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,WILLIAM,W
BREATH TEST OPERATOR AGENCY: NAS Pensacola
ARRESTING OFFICER: MILLER,WILLIAM
ARRESTING OFFICER AGENCY: NAS Pensacola
LAST AGENCY INSPECTION: 02/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RIMER

DATE/TIME OF TEST: 03/04/2023 04:00:38

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.076	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.164	03:55
Breath Volume	2.597	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.166	03:59
Breath Volume	3.269	03:59
Air Blank	0.000	03:59
Control Test	0.076	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 03/05/2023 23:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

INSTRUMENT REGISTERED TO: NAS PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAVARES,PETER,M
BREATH TEST OPERATOR AGENCY: NAS Pensacola
ARRESTING OFFICER: TAVARES,PETER
ARRESTING OFFICER AGENCY: NAS Pensacola
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 03/18/2023 03:39:13

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.253	03:34
Breath Volume	3.320	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.243	03:37
Breath Volume	1.933	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 04/05/2023 23:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: TZUNUX BRITO

DATE/TIME OF TEST: 03/01/2023 04:36:21

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.045	04:31
Breath Volume	1.804	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.043	04:34
Breath Volume	1.648	04:34
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: PRIEST

DATE/TIME OF TEST: 03/02/2023 02:58:35

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.147	02:53
Breath Volume	1.242	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.148	02:57
Breath Volume	2.898	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BYRD,B
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: DELANEY

DATE/TIME OF TEST: 03/03/2023 00:06:22

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.184	00:01
Breath Volume	2.316	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.176	00:04
Breath Volume	2.917	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ARNOLD,PAUL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: REALE

DATE/TIME OF TEST: 03/04/2023 03:02:25

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.081	02:50	
Air Blank	0.000	02:50	
Subject Sample	# 1 0.061	02:53	Volume Not Met
Breath Volume	0.234	02:53	
Air Blank	0.000	02:54	
Air Blank	0.000	02:56	
Subject Sample	# 2 0.080	02:57	
Breath Volume	1.335	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample	# 3 0.075	03:00	
Breath Volume	1.664	03:00	
Air Blank	0.000	03:01	
Control Test	0.079	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MATERA,DERIK
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: KIWAN

DATE/TIME OF TEST: 03/04/2023 03:22:49

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1 0.167	03:17	Volume Not Met
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample	# 2	03:21	Subject Test Refused
Breath Volume	0.105	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: PIRONTI

DATE/TIME OF TEST: 03/04/2023 03:44:15

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.066	03:39
Breath Volume	2.093	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.063	03:42
Breath Volume	2.144	03:42
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ARMSTRONG,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: IRIGOYEN IBARRA

DATE/TIME OF TEST: 03/05/2023 00:12:05

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:10	
Subject Sample	# 1	00:10	Subject Test Refused
Breath Volume	0.000	00:10	
Air Blank	0.000	00:11	
Control Test	0.081	00:11	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ARMSTRONG,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: ROCHELLE

DATE/TIME OF TEST: 03/10/2023 05:02:36

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.082	04:56
Air Blank	0.000	04:56
Subject Sample # 1	0.139	04:57
Breath Volume	4.855	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.136	05:01
Breath Volume	4.109	05:01
Air Blank	0.000	05:01
Control Test	0.081	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,ROHAN,G
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: RODRIGUEZ,EDDIE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: CHIC CASTRO

DATE/TIME OF TEST: 03/10/2023 06:48:20

Diagnostics Check	OK	06:42
Air Blank	0.000	06:42
Control Test	0.081	06:42
Air Blank	0.000	06:43
Subject Sample # 1	0.137	06:43
Breath Volume	1.980	06:43
Air Blank	0.000	06:44
Air Blank	0.000	06:46
Subject Sample # 2	0.134	06:46
Breath Volume	3.074	06:46
Air Blank	0.000	06:47
Control Test	0.081	06:47
Air Blank	0.000	06:48
Diagnostics Check	OK	06:48

Data Download Date/Time: 03/13/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SCHWARTZE

DATE/TIME OF TEST: 03/11/2023 01:10:41

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.356	01:06
Breath Volume	2.367	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.350	01:09
Breath Volume	1.816	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 03/13/2023 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,LINDSAY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMYRNIOS,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: ASHCRAFT

DATE/TIME OF TEST: 03/11/2023 13:41:45

Diagnostics Check	OK	13:34
Air Blank	0.000	13:34
Control Test	0.080	13:35
Air Blank	0.000	13:35
Subject Sample # 1	0.234	13:36
Breath Volume	2.085	13:36
Air Blank	0.000	13:36
Air Blank	0.000	13:38
Subject Sample # 2	0.233	13:40
Breath Volume	1.542	13:40
Air Blank	0.000	13:40
Control Test	0.078	13:41
Air Blank	0.000	13:41
Diagnostics Check	OK	13:41

Data Download Date/Time: 03/13/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAK,KYLE,F
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DEAK,KYLE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:41

SUBJECT'S LAST NAME: TRAUSOTT

DATE/TIME OF TEST: 03/12/2023 07:09:32

Diagnostics Check	OK	07:03
Air Blank	0.000	07:03
Control Test	0.081	07:04
Air Blank	0.000	07:04
Subject Sample # 1	0.195	07:05
Breath Volume	1.933	07:05
Air Blank	0.000	07:05
Air Blank	0.000	07:07
Subject Sample # 2	0.200	07:08
Breath Volume	1.230	07:08
Air Blank	0.000	07:08
Control Test	0.080	07:08
Air Blank	0.000	07:09
Diagnostics Check	OK	07:09

Data Download Date/Time: 03/13/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: RUDINSKY

DATE/TIME OF TEST: 03/14/2023 02:15:46

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.125	02:11
Breath Volume	2.203	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.116	02:14
Breath Volume	2.121	02:14
Air Blank	0.000	02:14
Control Test	0.080	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 04/20/2023 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMEN,LOGAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MCLAUGHLIN,TERENCE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:52

SUBJECT'S LAST NAME: CASON

DATE/TIME OF TEST: 03/18/2023 10:19:15

Diagnostics Check	OK	10:13	
Air Blank	0.000	10:13	
Control Test	0.081	10:14	
Air Blank	0.000	10:14	
Subject Sample	# 1	10:17	Subject Test Refused
Breath Volume	0.421	10:17	
Air Blank	0.000	10:18	
Control Test	0.081	10:18	
Air Blank	0.000	10:19	
Diagnostics Check	OK	10:19	

Data Download Date/Time: 04/20/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: AMENDOLA,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: GASPAR

DATE/TIME OF TEST: 03/20/2023 04:27:06

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.171	04:22
Breath Volume	2.105	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.168	04:25
Breath Volume	1.949	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 04/20/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,JILLIAN,F
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FRAZZETTO,ASHLEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 03/22/2023 06:18:50

Diagnostics Check	OK	06:10
Air Blank	0.000	06:10
Control Test	0.081	06:11
Air Blank	0.000	06:11
Subject Sample # 1	0.035	06:13
Breath Volume	1.351	06:13
Air Blank	0.000	06:13
Air Blank	0.000	06:15
Subject Sample # 2	0.033	06:17
Breath Volume	1.691	06:17
Air Blank	0.000	06:17
Control Test	0.079	06:18
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 04/20/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: KRAIZA

DATE/TIME OF TEST: 03/23/2023 06:40:21

Diagnostics Check	OK	06:33
Air Blank	0.000	06:34
Control Test	0.081	06:34
Air Blank	0.000	06:35
Subject Sample # 1	0.080	06:35
Breath Volume	1.457	06:35
Air Blank	0.000	06:36
Air Blank	0.000	06:38
Subject Sample # 2	0.072	06:38
Breath Volume	1.468	06:38
Air Blank	0.000	06:39
Control Test	0.080	06:39
Air Blank	0.000	06:40
Diagnostics Check	OK	06:40

Data Download Date/Time: 04/20/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JACKSON,TYRESE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: SANTIZO PEREZ

DATE/TIME OF TEST: 03/24/2023 04:42:23

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.081	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.200	04:37
Breath Volume	3.804	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.207	04:40
Breath Volume	3.480	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 04/20/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HERNANDEZ CASTILLO

DATE/TIME OF TEST: 03/26/2023 02:04:55

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.219	02:00
Breath Volume	3.445	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:02
Subject Sample # 2	0.210	02:03
Breath Volume	3.027	02:03
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 04/20/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,LINDSAY,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SEMICLOSE,CHARLES
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:39

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 03/26/2023 08:34:19

Diagnostics Check	OK	08:33	
Air Blank		08:33	RFI Detect
Air Blank	0.000	08:34	

Data Download Date/Time: 04/20/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,LINDSAY,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SEMICLOSE,CHARLES
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:39

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 03/26/2023 08:41:58

Diagnostics Check	OK	08:35
Air Blank	0.000	08:36
Control Test	0.081	08:36
Air Blank	0.000	08:37
Subject Sample # 1	0.067	08:37
Breath Volume	2.460	08:37
Air Blank	0.000	08:38
Air Blank	0.000	08:40
Subject Sample # 2	0.062	08:40
Breath Volume	1.488	08:40
Air Blank	0.000	08:41
Control Test	0.079	08:41
Air Blank	0.000	08:41
Diagnostics Check	OK	08:41

Data Download Date/Time: 04/20/2023 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMEN,LOGAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MAIK,CHAD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:32

SUBJECT'S LAST NAME: NOVOM

DATE/TIME OF TEST: 03/27/2023 14:54:34

Diagnostics Check	OK	14:53	
Air Blank	0.000	14:53	
Control Test	0.080	14:53	
Air Blank		14:54	RFI Detect
Air Blank	0.000	14:54	

Data Download Date/Time: 04/20/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMEN,LOGAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MAIK,CHAD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:32

SUBJECT'S LAST NAME: NOVOM

DATE/TIME OF TEST: 03/27/2023 15:03:32

Diagnostics Check	OK	14:56
Air Blank	0.000	14:57
Control Test	0.081	14:57
Air Blank	0.000	14:58
Subject Sample # 1	0.146	14:59
Breath Volume	1.445	14:59
Air Blank	0.000	14:59
Air Blank	0.000	15:01
Subject Sample # 2	0.138	15:02
Breath Volume	1.507	15:02
Air Blank	0.000	15:02
Control Test	0.080	15:02
Air Blank	0.000	15:03
Diagnostics Check	OK	15:03

Data Download Date/Time: 04/20/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ARMSTRONG,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: TARDIF

DATE/TIME OF TEST: 03/27/2023 21:27:30

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.000	21:23
Breath Volume	4.421	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.000	21:26
Breath Volume	3.496	21:26
Air Blank	0.000	21:26
Control Test	0.081	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 04/20/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCVEIGH,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: KOSTENO

DATE/TIME OF TEST: 03/27/2023 23:05:08

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:02	
Control Test	0.080	23:02	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:04	
Control Test	0.081	23:04	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

Data Download Date/Time: 04/20/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PEACOCK

DATE/TIME OF TEST: 03/28/2023 01:57:37

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.218	01:53
Breath Volume	1.187	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.216	01:56
Breath Volume	1.367	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 04/20/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: FERRER

DATE/TIME OF TEST: 03/29/2023 05:20:11

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.081	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.000	05:16
Breath Volume	3.324	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.000	05:18
Breath Volume	2.156	05:18
Air Blank	0.000	05:19
Control Test	0.081	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 04/20/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: TERNS

DATE/TIME OF TEST: 03/30/2023 02:46:31

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.000	02:42
Breath Volume	3.769	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.000	02:45
Breath Volume	3.945	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 04/20/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKPATRICK,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: KIRKPATRICK,MATTHEW
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: MOSES

DATE/TIME OF TEST: 03/24/2023 08:36:01

Diagnostics Check	OK	08:23	
Air Blank	0.000	08:23	
Control Test	0.079	08:23	
Air Blank	0.000	08:24	
Subject Sample	# 1 0.000	08:27	Volume Not Met
Breath Volume	0.660	08:27	
Air Blank	0.000	08:27	
Air Blank	0.000	08:29	
Subject Sample	# 2 0.000	08:30	
Breath Volume	2.179	08:30	
Air Blank	0.000	08:31	
Air Blank	0.000	08:33	
Subject Sample	# 3 0.000	08:34	
Breath Volume	1.804	08:34	
Air Blank	0.000	08:35	
Control Test	0.080	08:35	
Air Blank	0.000	08:35	
Diagnostics Check	OK	08:36	

Data Download Date/Time: 04/27/2023 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPLING,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: GOTZ,KATHRYN
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: TADDEO

DATE/TIME OF TEST: 03/01/2023 17:48:50

Diagnostics Check	OK	17:40
Air Blank	0.000	17:41
Control Test	0.080	17:41
Air Blank	0.000	17:42
Subject Sample # 1	0.228	17:43
Breath Volume	1.484	17:43
Air Blank	0.000	17:44
Air Blank	0.000	17:46
Subject Sample # 2	0.216	17:47
Breath Volume	1.343	17:47
Air Blank	0.000	17:47
Control Test	0.076	17:48
Air Blank	0.000	17:48
Diagnostics Check	OK	17:48

Data Download Date/Time: 03/21/2023 01:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTZ,KATHRYN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: MALPHURS,GARY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: DUKES

DATE/TIME OF TEST: 03/02/2023 13:00:37

Diagnostics Check	OK	12:52	
Air Blank	0.000	12:52	
Control Test	0.081	12:52	
Air Blank	0.000	12:53	
Subject Sample	# 1	12:56	No Sample Provided
Breath Volume	0.000	12:56	
Air Blank	0.000	12:57	
Air Blank	0.000	12:59	
Subject Sample	# 2	12:59	Subject Test Refused
Breath Volume	0.000	12:59	
Air Blank	0.000	12:59	
Control Test	0.079	13:00	
Air Blank	0.000	13:00	
Diagnostics Check	OK	13:00	

Data Download Date/Time: 03/21/2023 01:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADEMACHER,RICHARD,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BILLUE,MOLLY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 03/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 03/25/2023 03:32:56

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.093	03:28
Breath Volume	2.304	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.093	03:31
Breath Volume	1.960	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 04/14/2023 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEGER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: FEGER,JOSHUA
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 02/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 13:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: STIPP

DATE/TIME OF TEST: 03/03/2023 14:36:10

Diagnostics Check	OK	14:29
Air Blank	0.000	14:30
Control Test	0.077	14:30
Air Blank	0.000	14:31
Subject Sample # 1	0.000	14:32
Breath Volume	1.660	14:32
Air Blank	0.000	14:32
Air Blank	0.000	14:34
Subject Sample # 2	0.000	14:34
Breath Volume	1.746	14:34
Air Blank	0.000	14:35
Control Test	0.075	14:35
Air Blank	0.000	14:36
Diagnostics Check	OK	14:36

Data Download Date/Time: 03/24/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEGER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: FEGER,JOSHUA
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 02/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: COCCO

DATE/TIME OF TEST: 03/06/2023 16:52:08

Diagnostics Check	OK	16:46
Air Blank	0.000	16:46
Control Test	0.076	16:46
Air Blank	0.000	16:47
Subject Sample # 1	0.000	16:47
Breath Volume	1.492	16:47
Air Blank	0.000	16:48
Air Blank	0.000	16:50
Subject Sample # 2	0.000	16:50
Breath Volume	1.632	16:50
Air Blank	0.000	16:51
Control Test	0.077	16:51
Air Blank	0.000	16:52
Diagnostics Check	OK	16:52

Data Download Date/Time: 03/24/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEGER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: FEGER,JOSHUA
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 02/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HATHAWAY

DATE/TIME OF TEST: 03/23/2023 19:47:11

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.077	19:42
Air Blank	0.000	19:42
Subject Sample # 1	0.000	19:43
Breath Volume	1.539	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.000	19:45
Breath Volume	1.656	19:45
Air Blank	0.000	19:46
Control Test	0.077	19:46
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 03/24/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRONDIN,B,D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: MENDOZA,JONATHAN
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 02/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 09:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:47

SUBJECT'S LAST NAME: SERRANO-GONZALEZ

DATE/TIME OF TEST: 03/20/2023 10:25:32

Diagnostics Check	OK	10:14	
Air Blank	0.000	10:15	
Control Test	0.081	10:15	
Air Blank	0.000	10:16	
Subject Sample	# 1 0.155	10:19	Volume Not Met
Breath Volume	0.070	10:19	
Air Blank	0.000	10:20	
Air Blank	0.000	10:21	
Subject Sample	# 2	10:23	Subject Test Refused
Breath Volume	0.050	10:23	
Air Blank	0.000	10:24	
Control Test	0.079	10:24	
Air Blank	0.000	10:25	
Diagnostics Check	OK	10:25	

Data Download Date/Time: 03/25/2023 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFFNEY,CASEY,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: GAFFNEY,CASEY
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/25/2023 20:12:34

Diagnostics Check	OK	20:06
Air Blank	0.000	20:06
Control Test	0.080	20:07
Air Blank	0.000	20:07
Subject Sample # 1	0.225	20:08
Breath Volume	1.339	20:08
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.229	20:11
Breath Volume	1.648	20:11
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 05/01/2023 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MIRANDA,JAVIER
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: WESTNER

DATE/TIME OF TEST: 03/03/2023 00:07:19

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.211	00:02
Breath Volume	3.015	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.221	00:05
Breath Volume	3.652	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 03/30/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFLEUR, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: LAFLEUR, MATTHEW
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 03/04/2023 03:04:05

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:58	
Control Test	0.080	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1 0.133	03:00	
Breath Volume	1.929	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 2	03:03	Sequence Aborted
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	

Data Download Date/Time: 03/30/2023 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HILL,ROBERT,J

BREATH TEST OPERATOR AGENCY: Cocoa Beach PD

ARRESTING OFFICER: HILL,ROBERT

ARRESTING OFFICER AGENCY: Cocoa Beach PD

LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 25721080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ROCKEY

DATE/TIME OF TEST: 03/05/2023 01:33:17

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.272	01:28
Breath Volume	2.808	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.266	01:31
Breath Volume	3.968	01:31
Air Blank	0.000	01:32
Control Test	0.078	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 03/30/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROSA,EMILIO
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 03/07/2023 03:01:18

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.081	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 1 0.129	02:54	Volume Not Met
Breath Volume	0.324	02:54	
Air Blank	0.000	02:54	
Air Blank	0.000	02:56	
Subject Sample	# 2	02:59	No Sample Provided
Breath Volume	0.000	02:59	
Air Blank	0.000	03:00	
Control Test	0.080	03:00	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 03/30/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GIVANS, GARRETH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: MANDA

DATE/TIME OF TEST: 03/07/2023 23:49:56

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.104	23:45
Breath Volume	2.187	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.098	23:48
Breath Volume	2.640	23:48
Air Blank	0.000	23:49
Control Test	0.077	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 03/30/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,T,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS,T
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BUFORD

DATE/TIME OF TEST: 03/23/2023 04:45:32

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.095	04:41
Breath Volume	3.257	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.095	04:44
Breath Volume	3.187	04:44
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 03/30/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERLAIN, JOSHUA, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CHAMBERLAIN, JOSHUA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: EICHELBERG

DATE/TIME OF TEST: 03/28/2023 21:12:27

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.080	21:06
Air Blank	0.000	21:07
Subject Sample # 1	0.359	21:07
Breath Volume	2.582	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.339	21:10
Breath Volume	2.503	21:10
Air Blank	0.000	21:11
Control Test	0.080	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 03/30/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 03/05/2023 04:20:52

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.086	04:16
Breath Volume	1.332	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.083	04:19
Breath Volume	2.476	04:19
Air Blank	0.000	04:19
Control Test	0.078	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 03/14/2023 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WHALEN, THOMAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PEREZ DURAN

DATE/TIME OF TEST: 03/08/2023 03:54:44

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.100	03:50
Breath Volume	2.128	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:52
Subject Sample # 2	0.091	03:53
Breath Volume	1.648	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 03/14/2023 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: SEIJO

DATE/TIME OF TEST: 03/09/2023 01:57:03

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.232	01:52
Breath Volume	1.835	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:54
Subject Sample # 2	0.227	01:55
Breath Volume	2.093	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:57

Data Download Date/Time: 03/14/2023 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUNT, FRANCIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: CHAIR

DATE/TIME OF TEST: 03/09/2023 03:12:18

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.183	03:08
Breath Volume	1.886	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.192	03:10
Breath Volume	1.273	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 03/14/2023 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHOUN, ELLIS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: PEREZLOPEZ

DATE/TIME OF TEST: 03/10/2023 02:02:38

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.186	01:57
Breath Volume	1.492	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.180	02:01
Breath Volume	1.984	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 03/14/2023 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL, ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 03/10/2023 03:47:05

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.065	03:42
Breath Volume	2.609	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.059	03:45
Breath Volume	1.855	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:47

Data Download Date/Time: 03/14/2023 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIESENFELT,JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: BATISTA GRULLON

DATE/TIME OF TEST: 03/11/2023 04:16:20

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.095	04:12
Breath Volume	2.457	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.094	04:14
Breath Volume	1.765	04:14
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 03/14/2023 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAUERNFEIND,BRANDON,M
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: LIVINGSTON,JAVON
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 02/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WEHLAND

DATE/TIME OF TEST: 03/04/2023 03:06:56

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:04	
Control Test	0.080	03:04	
Air Blank	0.000	03:05	
Subject Sample	# 1	03:05	Subject Test Refused
Breath Volume	0.000	03:05	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:06	
Diagnostics Check	OK	03:06	

Data Download Date/Time: 04/06/2023 08:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENDER,DEREK,A
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: LIVINGSTON,JAVON
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 02/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: CRISSMAN

DATE/TIME OF TEST: 03/15/2023 00:30:13

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.080	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:29	
Control Test	0.078	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 04/06/2023 08:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENDER,DEREK,A
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: KRAMER,TERVOR
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 02/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: SHUTES

DATE/TIME OF TEST: 03/19/2023 03:20:53

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.080	03:18	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:20	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 04/06/2023 08:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOREHEAD,AVERY,F
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SCHALLON,CHRISTIAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 02/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 10:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 03/21/2023 11:39:11

Diagnostics Check	OK	11:36	
Air Blank	0.000	11:36	
Control Test	0.079	11:37	
Air Blank	0.000	11:37	
Subject Sample	# 1	11:37	Subject Test Refused
Breath Volume	0.000	11:37	
Air Blank	0.000	11:38	
Control Test	0.080	11:38	
Air Blank	0.000	11:39	
Diagnostics Check	OK	11:39	

Data Download Date/Time: 04/06/2023 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,JACOB,A
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: ROSS,JACOB
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 03/01/2023 18:07:49

Diagnostics Check	OK	18:00
Air Blank	0.000	18:00
Control Test	0.081	18:01
Air Blank	0.000	18:01
Subject Sample # 1	0.000	18:02
Breath Volume	1.832	18:02
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 2	0.000	18:06
Breath Volume	1.703	18:06
Air Blank	0.000	18:06
Control Test	0.078	18:07
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

Data Download Date/Time: 03/14/2023 03:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HALL,JONATHAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/02/2023 19:32:11

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:23	
Control Test	0.080	19:24	
Air Blank	0.000	19:24	
Subject Sample	# 1 0.244	19:27	Volume Not Met
Breath Volume	0.066	19:27	
Air Blank	0.000	19:28	
Air Blank	0.000	19:30	
Subject Sample	# 2	19:30	Subject Test Refused
Breath Volume	0.000	19:30	
Air Blank	0.000	19:31	
Control Test	0.078	19:31	
Air Blank	0.000	19:32	
Diagnostics Check	OK	19:32	

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RAMOS,MARK
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 03/04/2023 04:02:06

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.191	03:57
Breath Volume	2.410	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.181	04:00
Breath Volume	3.195	04:00
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: QUINTAL,VIRGEL
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: VUKOMANOVIC

DATE/TIME OF TEST: 03/04/2023 23:34:36

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.120	23:29
Breath Volume	2.035	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.123	23:33
Breath Volume	1.707	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLAZINA,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 03/05/2023 02:41:28

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.085	02:37
Breath Volume	1.667	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.085	02:39
Breath Volume	1.203	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: WITT,P
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: LUCAS MARTINEZ

DATE/TIME OF TEST: 03/06/2023 03:19:54

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.195	03:15
Breath Volume	2.503	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.195	03:18
Breath Volume	2.136	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KAREEM,D
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: BORROMEO RAIMUNDO

DATE/TIME OF TEST: 03/06/2023 05:09:28

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.080	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.276	05:04
Breath Volume	3.492	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.278	05:07
Breath Volume	3.656	05:07
Air Blank	0.000	05:08
Control Test	0.079	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS,E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: MARSHALL,COREY
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/09/2023 06:02:19

Diagnostics Check	OK	05:54
Air Blank	0.000	05:55
Control Test	0.080	05:55
Air Blank	0.000	05:56
Subject Sample # 1	0.123	05:57
Breath Volume	1.371	05:57
Air Blank	0.000	05:58
Air Blank	0.000	06:00
Subject Sample # 2	0.117	06:00
Breath Volume	1.820	06:00
Air Blank	0.000	06:01
Control Test	0.078	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOONEY,RYA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 03/10/2023 03:19:45

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.146	03:15
Breath Volume	3.851	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.148	03:18
Breath Volume	2.390	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ESTEVES,VIRGILIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 03/10/2023 03:32:04

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.132	03:27
Breath Volume	3.812	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.131	03:30
Breath Volume	3.589	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:32

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLAZINA,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DURAN-SWEATT

DATE/TIME OF TEST: 03/11/2023 03:43:37

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.127	03:38
Breath Volume	1.746	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.125	03:42
Breath Volume	1.867	03:42
Air Blank	0.000	03:42
Control Test	0.078	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VITELLARO,N
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 03/11/2023 21:46:14

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.240	21:41
Breath Volume	1.957	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.231	21:44
Breath Volume	1.871	21:44
Air Blank	0.000	21:45
Control Test	0.078	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,C
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: DELMORAL

DATE/TIME OF TEST: 03/12/2023 00:54:20

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.017	00:49
Breath Volume	2.117	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.017	00:52
Breath Volume	1.246	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: HILLS

DATE/TIME OF TEST: 03/12/2023 01:58:06

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.239	01:52
Breath Volume	1.480	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.232	01:56
Breath Volume	1.800	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 03/14/2023 03:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLYNN, COLBY, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: TUERO, ADRIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: GLYNN

DATE/TIME OF TEST: 03/12/2023 21:29:30

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.079	21:24
Air Blank	0.000	21:24
Subject Sample # 1	0.162	21:25
Breath Volume	2.269	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.159	21:27
Breath Volume	2.742	21:27
Air Blank	0.000	21:28
Control Test	0.079	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 03/14/2023 03:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS,E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: NOLAND,THOMAS
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: CLEMON

DATE/TIME OF TEST: 03/13/2023 07:38:03

Diagnostics Check	OK	07:35	
Air Blank	0.000	07:35	
Control Test	0.080	07:35	
Air Blank	0.000	07:36	
Subject Sample	# 1	07:36	RFI Detect
Breath Volume	0.000	07:36	
Air Blank	0.000	07:37	
Control Test	0.080	07:37	
Air Blank	0.000	07:37	
Diagnostics Check	OK	07:38	

Data Download Date/Time: 03/14/2023 03:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS,E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: NOLAND,THOMAS
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: CLEMON

DATE/TIME OF TEST: 03/13/2023 07:51:26

Diagnostics Check	OK	07:40	
Air Blank	0.000	07:40	
Control Test	0.080	07:40	
Air Blank	0.000	07:41	
Subject Sample	# 1	07:44	No Sample Provided
Breath Volume	0.000	07:44	
Air Blank	0.000	07:44	
Air Blank	0.000	07:46	
Subject Sample	# 2	07:50	No Sample Provided
Breath Volume	0.000	07:50	
Air Blank	0.000	07:50	
Control Test	0.080	07:50	
Air Blank	0.000	07:51	
Diagnostics Check	OK	07:51	

Data Download Date/Time: 03/14/2023 03:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS,E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: NOLAND,THOMAS
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: CLEMON

DATE/TIME OF TEST: 03/13/2023 08:06:57

Diagnostics Check	OK	07:55	
Air Blank	0.000	07:56	
Control Test	0.080	07:56	
Air Blank	0.000	07:56	
Subject Sample	# 1 0.115	08:00	Volume Not Met
Breath Volume	0.589	08:00	
Air Blank	0.000	08:00	
Air Blank	0.000	08:02	
Subject Sample	# 2 0.119	08:02	
Breath Volume	1.742	08:02	
Air Blank	0.000	08:03	
Air Blank	0.000	08:05	
		08:05	RFI Detect
Air Blank	0.000	08:06	
Control Test	0.079	08:06	
Air Blank	0.000	08:06	
Diagnostics Check	OK	08:06	

Data Download Date/Time: 03/14/2023 03:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS,E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: NOLAND,THOMAS
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: CLEMON

DATE/TIME OF TEST: 03/13/2023 08:15:09

Diagnostics Check	OK	08:08
Air Blank	0.000	08:08
Control Test	0.080	08:09
Air Blank	0.000	08:09
Subject Sample # 1	0.109	08:10
Breath Volume	1.558	08:10
Air Blank	0.000	08:10
Air Blank	0.000	08:12
Subject Sample # 2	0.109	08:13
Breath Volume	1.582	08:13
Air Blank	0.000	08:14
Control Test	0.079	08:14
Air Blank	0.000	08:15
Diagnostics Check	OK	08:15

Data Download Date/Time: 03/14/2023 03:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVEZ,CESAR,G
BREATH TEST OPERATOR AGENCY: Winter Springs PD
ARRESTING OFFICER: MOXLEY,MICHAEL
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: AGEE

DATE/TIME OF TEST: 03/15/2023 07:34:32

Diagnostics Check	OK	07:28
Air Blank	0.000	07:28
Control Test	0.080	07:28
Air Blank	0.000	07:29
Subject Sample # 1	0.167	07:29
Breath Volume	2.191	07:29
Air Blank	0.000	07:30
Air Blank	0.000	07:32
Subject Sample # 2	0.171	07:32
Breath Volume	2.925	07:32
Air Blank	0.000	07:33
Control Test	0.079	07:33
Air Blank	0.000	07:34
Diagnostics Check	OK	07:34

Data Download Date/Time: 03/27/2023 03:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VITELLARO,N
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 03/16/2023 03:40:11

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.206	03:35
Breath Volume	3.449	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.202	03:38
Breath Volume	2.703	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 03/27/2023 03:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/17/2023 00:12:46

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.275	00:07
Breath Volume	1.457	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.262	00:11
Breath Volume	2.070	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 03/27/2023 03:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REHDER,PATRICK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: CHIN-LEE

DATE/TIME OF TEST: 03/17/2023 01:17:47

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.079	01:13
Breath Volume	2.410	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.086	01:16
Breath Volume	2.562	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 03/27/2023 03:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UCELO,BRYAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: VASQUEZ VASQUEZ

DATE/TIME OF TEST: 03/18/2023 01:16:52

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.194	01:12
Breath Volume	3.707	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:14
Subject Sample # 2	0.194	01:15
Breath Volume	2.492	01:15
Air Blank	0.000	01:15
Control Test	0.079	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 03/27/2023 03:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RODRIGUEZ,EMMANUEL
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: WULFF

DATE/TIME OF TEST: 03/18/2023 02:11:04

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.078	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.223	02:06
Breath Volume	2.320	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.226	02:09
Breath Volume	2.261	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:11

Data Download Date/Time: 03/27/2023 03:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: NAUGHTON

DATE/TIME OF TEST: 03/18/2023 03:05:26

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.148	03:00
Breath Volume	2.785	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.149	03:03
Breath Volume	2.664	03:03
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 03/27/2023 03:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GANT,JERRY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/18/2023 03:28:56

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.165	03:24
Breath Volume	2.535	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.168	03:27
Breath Volume	2.808	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 03/27/2023 03:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,JERRY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/18/2023 04:45:31

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.079	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.167	04:40
Breath Volume	1.734	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.179	04:43
Breath Volume	1.230	04:43
Air Blank	0.000	04:44
Control Test	0.078	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 03/28/2023 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RILEY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: SALES LOPEZ

DATE/TIME OF TEST: 03/18/2023 22:06:43

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.161	22:02
Breath Volume	2.324	22:02
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.146	22:05
Breath Volume	1.593	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 03/28/2023 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANTIAGO,LUMANS
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 03/19/2023 00:56:02

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.108	00:51
Breath Volume	3.011	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.105	00:54
Breath Volume	3.406	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:56

Data Download Date/Time: 03/28/2023 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DASQUE,DARREN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: SPILMAN

DATE/TIME OF TEST: 03/19/2023 01:57:25

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.305	01:49
Breath Volume	3.054	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:51
Subject Sample # 2	0.272	01:52
Breath Volume	2.265	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 3	0.280	01:55
Breath Volume	1.589	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 03/28/2023 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS,E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: NOLAND,THOMAS
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: NEIDHARDT

DATE/TIME OF TEST: 03/19/2023 16:09:33

Diagnostics Check	OK	15:59	
Air Blank	0.000	16:00	
Control Test	0.079	16:00	
Air Blank	0.000	16:01	
Subject Sample	# 1	16:04	No Sample Provided
Breath Volume	0.000	16:04	
Air Blank	0.000	16:04	
Air Blank	0.000	16:06	
Subject Sample	# 2	16:08	RFI Detect
Breath Volume	0.000	16:08	
Air Blank	0.000	16:08	
Control Test	0.081	16:08	
Air Blank	0.000	16:09	
Diagnostics Check	OK	16:09	

Data Download Date/Time: 03/28/2023 02:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS,E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: NOLAND,THOMAS
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: NEIDHARDT

DATE/TIME OF TEST: 03/19/2023 16:23:18

Diagnostics Check	OK	16:11	
Air Blank	0.000	16:12	
Control Test	0.080	16:12	
Air Blank	0.000	16:13	
Subject Sample	# 1	16:16	No Sample Provided
Breath Volume	0.000	16:16	
Air Blank	0.000	16:16	
Air Blank	0.000	16:18	
Subject Sample	# 2	16:21	No Sample Provided
Breath Volume	0.000	16:21	
Air Blank	0.000	16:22	
Control Test	0.081	16:22	
Air Blank	0.000	16:23	
Diagnostics Check	OK	16:23	

Data Download Date/Time: 03/28/2023 02:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROMAN,J
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SHANAHAN

DATE/TIME OF TEST: 03/19/2023 22:17:53

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.079	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.245	22:13
Breath Volume	2.191	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.247	22:16
Breath Volume	2.640	22:16
Air Blank	0.000	22:16
Control Test	0.079	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 03/28/2023 02:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CHALK

DATE/TIME OF TEST: 03/20/2023 03:07:43

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.133	03:03
Breath Volume	1.785	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.128	03:06
Breath Volume	1.937	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 03/28/2023 04:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: AMIN,SAHIL
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 03/20/2023 04:05:13

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.157	04:00
Breath Volume	2.234	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.152	04:03
Breath Volume	2.324	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 03/28/2023 04:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: POZADA,A
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: ABUDIAR

DATE/TIME OF TEST: 03/22/2023 03:53:33

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.189	03:47
Breath Volume	2.195	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.175	03:52
Breath Volume	2.652	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 03/28/2023 04:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,S,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,S
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 03/23/2023 14:37:03

Diagnostics Check	OK	14:31
Air Blank	0.000	14:31
Control Test	0.080	14:31
Air Blank	0.000	14:32
Subject Sample # 1	0.067	14:32
Breath Volume	3.449	14:32
Air Blank	0.000	14:33
Air Blank	0.000	14:35
Subject Sample # 2	0.063	14:35
Breath Volume	3.230	14:35
Air Blank	0.000	14:36
Control Test	0.080	14:36
Air Blank	0.000	14:36
Diagnostics Check	OK	14:37

Data Download Date/Time: 03/28/2023 04:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: JOHNSON,A
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: DOWNING

DATE/TIME OF TEST: 03/25/2023 00:26:32

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.189	00:21
Breath Volume	2.035	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.184	00:25
Breath Volume	1.621	00:25
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 03/28/2023 04:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: JOHNSON,M
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 03/25/2023 22:06:34

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.079	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.000	22:00
Breath Volume	1.699	22:00
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.000	22:05
Breath Volume	1.253	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 03/28/2023 04:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DIAL,V
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/26/2023 04:02:48

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.079	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1 0.089	03:58	
Breath Volume	1.324	03:58	
Air Blank	0.000	03:58	
Air Blank	0.000	04:00	
Subject Sample	# 2 0.094	04:01	
Breath Volume	2.128	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:02	
Air Blank		04:02	RFI Detect
Air Blank	0.000	04:02	

Data Download Date/Time: 03/28/2023 04:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/26/2023 04:51:39

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.100	04:47
Breath Volume	1.621	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.098	04:50
Breath Volume	1.136	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 03/28/2023 04:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RILEY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: MIZELL

DATE/TIME OF TEST: 03/26/2023 21:50:48

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.079	21:44
Air Blank	0.000	21:45
Subject Sample # 1	0.087	21:46
Breath Volume	2.332	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.080	21:49
Breath Volume	1.246	21:49
Air Blank	0.000	21:49
Control Test	0.079	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 03/28/2023 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,LOGAN
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: LOPEZ-CHAVEZ

DATE/TIME OF TEST: 03/26/2023 22:47:29

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.046	22:42
Breath Volume	1.703	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.050	22:46
Breath Volume	1.515	22:46
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 03/28/2023 04:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: JIMENEZ RIVERO

DATE/TIME OF TEST: 03/27/2023 00:44:22

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.079	00:31	
Air Blank	0.000	00:31	
Subject Sample	# 1 0.120	00:34	Volume Not Met
Breath Volume	0.636	00:34	
Air Blank	0.000	00:35	
Air Blank	0.000	00:37	
Subject Sample	# 2 0.140	00:39	
Breath Volume	1.851	00:39	
Air Blank	0.000	00:40	
Air Blank	0.000	00:41	
Subject Sample	# 3 0.140	00:42	
Breath Volume	1.796	00:42	
Air Blank	0.000	00:43	
Control Test	0.078	00:43	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 03/28/2023 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GONZALEZ,JOAQUIN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: REYES REYES

DATE/TIME OF TEST: 03/27/2023 01:16:53

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.142	01:12
Breath Volume	1.304	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.144	01:15
Breath Volume	1.257	01:15
Air Blank	0.000	01:15
Control Test	0.079	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 04/05/2023 00:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANROMPA,V,
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: MCCRANIE,S
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 03/28/2023 20:44:04

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.079	20:38
Air Blank	0.000	20:38
Subject Sample # 1	0.195	20:39
Breath Volume	1.574	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:41
Subject Sample # 2	0.191	20:42
Breath Volume	1.460	20:42
Air Blank	0.000	20:43
Control Test	0.079	20:43
Air Blank	0.000	20:43
Diagnostics Check	OK	20:44

Data Download Date/Time: 04/05/2023 00:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PENA,U
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/28/2023 23:27:54

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.079	23:22
Air Blank	0.000	23:22
Subject Sample # 1	0.157	23:23
Breath Volume	2.156	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:25
Subject Sample # 2	0.152	23:26
Breath Volume	1.820	23:26
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 04/05/2023 00:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 03/23/2023 03:31:49

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.079	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	
Control Test	0.078	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 04/19/2023 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HICKS,J
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: RAECKERS

DATE/TIME OF TEST: 03/25/2023 21:56:27

Diagnostics Check	OK	21:47
Air Blank	0.000	21:48
Control Test	0.078	21:48
Air Blank	0.000	21:49
Subject Sample # 1	0.059	21:51
Breath Volume	1.476	21:51
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.056	21:54
Breath Volume	1.503	21:54
Air Blank	0.000	21:55
Control Test	0.077	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 04/19/2023 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGSON,ROLAND,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SINGSON,ROLAND
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: KADANE

DATE/TIME OF TEST: 03/25/2023 23:02:38

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.078	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.085	22:58
Breath Volume	1.375	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.088	23:00
Breath Volume	1.367	23:00
Air Blank	0.000	23:01
Control Test	0.076	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 04/19/2023 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LORING,BLAINE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: SIMONS

DATE/TIME OF TEST: 03/25/2023 23:35:26

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.077	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.083	23:30
Breath Volume	2.214	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.083	23:33
Breath Volume	1.757	23:33
Air Blank	0.000	23:34
Control Test	0.076	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 04/19/2023 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PAVLAT

DATE/TIME OF TEST: 03/26/2023 00:25:41

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.078	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.111	00:20
Breath Volume	2.136	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.109	00:23
Breath Volume	2.023	00:23
Air Blank	0.000	00:24
Control Test	0.076	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 04/19/2023 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SAMPAIO

DATE/TIME OF TEST: 03/26/2023 03:09:11

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.078	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:08	
Control Test	0.078	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 04/19/2023 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,B,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 03/22/2023 22:57:16

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.081	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.103	22:51
Breath Volume	1.851	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.096	22:55
Breath Volume	1.312	22:55
Air Blank	0.000	22:56
Control Test	0.081	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 03/23/2023 21:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 03/25/2023 01:24:43

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.080	01:22	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Control Test	0.081	01:24	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 04/19/2023 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,B,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 03/25/2023 04:35:03

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.080	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:34	
Control Test	0.080	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 04/19/2023 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,B,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 03/25/2023 23:15:32

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:12	
Control Test	0.079	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:14	Subject Test Refused
Breath Volume	0.000	23:14	
Air Blank	0.000	23:14	
Control Test	0.079	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 04/19/2023 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 03/26/2023 01:43:00

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.155	01:37
Breath Volume	2.183	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.156	01:41
Breath Volume	2.117	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:43

Data Download Date/Time: 04/19/2023 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,B,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: GALINDO

DATE/TIME OF TEST: 03/26/2023 03:25:47

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:23	
Control Test	0.079	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:24	
Control Test	0.079	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 04/19/2023 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,B,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: PACITTI

DATE/TIME OF TEST: 03/30/2023 23:48:53

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.081	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:47	
Control Test	0.081	23:48	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 04/19/2023 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,E,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: CABLES

DATE/TIME OF TEST: 03/31/2023 01:45:03

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.114	01:40
Breath Volume	1.707	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.115	01:43
Breath Volume	1.648	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:45

Data Download Date/Time: 04/19/2023 12:56