

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: RICO

DATE/TIME OF TEST: 03/11/2023 20:44:25

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.080	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.171	20:39
Breath Volume	2.328	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.176	20:42
Breath Volume	3.250	20:42
Air Blank	0.000	20:43
Control Test	0.079	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 04/07/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/12/2023 05:38:04

Diagnostics Check		OK	05:23	
Air Blank		0.000	05:23	
Control Test		0.079	05:24	
Air Blank		0.000	05:24	
Subject Sample	# 1	0.252	05:27	Volume Not Met
Breath Volume		0.117	05:27	
Air Blank		0.000	05:28	
Air Blank		0.000	05:30	
Subject Sample	# 2	0.259	05:30	
Breath Volume		1.730	05:30	
Air Blank		0.000	05:31	
Air Blank		0.000	05:33	
Subject Sample	# 3	0.246	05:36	Volume Not Met
Breath Volume		0.320	05:36	
Air Blank		0.000	05:37	
Control Test		0.078	05:37	
Air Blank		0.000	05:37	
Diagnostics Check		OK	05:38	

Data Download Date/Time: 04/07/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:41

SUBJECT'S LAST NAME: DIAZ PASCUAL

DATE/TIME OF TEST: 03/12/2023 11:06:41

Diagnostics Check	OK	11:01	
Air Blank	0.000	11:01	
Control Test	0.081	11:02	
Air Blank	0.000	11:02	
Subject Sample	# 1	11:05	Subject Test Refused
Breath Volume	0.000	11:05	
Air Blank	0.000	11:05	
Control Test	0.081	11:06	
Air Blank	0.000	11:06	
Diagnostics Check	OK	11:06	

Data Download Date/Time: 04/07/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: PIDCOCK

DATE/TIME OF TEST: 03/13/2023 01:28:05

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.081	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.212	01:23
Breath Volume	2.757	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:25
Subject Sample # 2	0.208	01:26
Breath Volume	1.730	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 04/07/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERIC,JEAN,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FREDERIC,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: AL BATNIJI

DATE/TIME OF TEST: 03/14/2023 01:03:56

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:01	
Control Test	0.081	01:01	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:02	
Control Test	0.081	01:03	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 04/07/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: GASPAR

DATE/TIME OF TEST: 03/15/2023 18:38:51

Diagnostics Check	OK	18:32
Air Blank	0.000	18:32
Control Test	0.080	18:33
Air Blank	0.000	18:33
Subject Sample # 1	0.075	18:34
Breath Volume	2.359	18:34
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.071	18:37
Breath Volume	1.773	18:37
Air Blank	0.000	18:37
Control Test	0.080	18:38
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 04/07/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MATEO

DATE/TIME OF TEST: 03/16/2023 03:01:41

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.256	02:57
Breath Volume	2.421	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.250	03:00
Breath Volume	1.457	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 04/07/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TOMAS-RAMIREZ

DATE/TIME OF TEST: 03/17/2023 03:02:31

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.238	02:58
Breath Volume	3.289	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.252	03:00
Breath Volume	3.250	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 04/07/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/18/2023 04:07:03

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.100	04:02
Breath Volume	3.253	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.103	04:05
Breath Volume	3.480	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:07

Data Download Date/Time: 04/07/2023 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: KERN

DATE/TIME OF TEST: 03/19/2023 00:19:52

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:17	
Control Test	0.080	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:18	
Control Test	0.080	00:19	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:19	

Data Download Date/Time: 04/07/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRADY,REBECCA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: HOSKINS

DATE/TIME OF TEST: 03/19/2023 00:33:30

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.080	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:25	No Sample Provided
Breath Volume	0.000	00:25	
Air Blank	0.000	00:25	
Air Blank	0.000	00:27	
Subject Sample	# 2	0.285	00:28
Breath Volume	2.773	00:28	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 3	0.268	00:31
Breath Volume	2.101	00:31	
Air Blank	0.000	00:32	
Control Test	0.079	00:32	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 04/07/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: CERNUTO,MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: VANHEUSEN

DATE/TIME OF TEST: 03/21/2023 01:30:35

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.076	01:25
Breath Volume	2.312	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.078	01:29
Breath Volume	1.976	01:29
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 04/07/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: DONATO

DATE/TIME OF TEST: 03/21/2023 23:34:26

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.165	23:29
Breath Volume	1.437	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.147	23:32
Breath Volume	1.183	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHAM,MATTHEW,H
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ISHAM,MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 03/23/2023 00:21:41

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.081	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.166	00:17
Breath Volume	1.593	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.164	00:20
Breath Volume	4.015	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 03/23/2023 20:01:00

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.081	19:52
Air Blank	0.000	19:52
Subject Sample # 1	0.118	19:54
Breath Volume	1.589	19:54
Air Blank	0.000	19:55
Air Blank	0.000	19:57
Subject Sample # 2	0.135	19:59
Breath Volume	2.875	19:59
Air Blank	0.000	20:00
Control Test	0.079	20:00
Air Blank	0.000	20:00
Diagnostics Check	OK	20:01

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 03/24/2023 19:33:00

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.081	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.176	19:27
Breath Volume	1.453	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.169	19:31
Breath Volume	1.750	19:31
Air Blank	0.000	19:32
Control Test	0.080	19:32
Air Blank	0.000	19:32
Diagnostics Check	OK	19:33

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: FORBES,RAY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/26/2023 02:36:10

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.081	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.208	02:31
Breath Volume	1.945	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.207	02:34
Breath Volume	1.320	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: CURTLAND

DATE/TIME OF TEST: 03/26/2023 23:15:23

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.198	23:10
Breath Volume	2.679	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.202	23:13
Breath Volume	2.765	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WADE,TYLER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: REYES GRANADOS

DATE/TIME OF TEST: 03/27/2023 01:04:57

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.205	01:00
Breath Volume	1.691	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:02
Subject Sample # 2	0.202	01:03
Breath Volume	1.539	01:03
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 03/28/2023 01:55:05

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.169	01:50
Breath Volume	3.277	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.177	01:53
Breath Volume	1.890	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 03/29/2023 22:06:27

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.194	22:02
Breath Volume	1.988	22:02
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.183	22:04
Breath Volume	1.507	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 03/31/2023 02:46:49

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.126	02:42
Breath Volume	3.738	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.126	02:45
Breath Volume	3.679	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,KEVIN,E
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MORRIS,KEVIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:58

SUBJECT'S LAST NAME: MEASMER

DATE/TIME OF TEST: 03/31/2023 12:30:06

Diagnostics Check	OK	12:24
Air Blank	0.000	12:24
Control Test	0.080	12:25
Air Blank	0.000	12:25
Subject Sample # 1	0.000	12:25
Breath Volume	2.296	12:25
Air Blank	0.000	12:26
Air Blank	0.000	12:28
Subject Sample # 2	0.000	12:28
Breath Volume	1.445	12:28
Air Blank	0.000	12:29
Control Test	0.081	12:29
Air Blank	0.000	12:30
Diagnostics Check	OK	12:30

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PEREZ,CALEB
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 02/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 03/04/2023 02:27:27

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.144	02:21
Breath Volume	1.367	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.145	02:25
Breath Volume	1.992	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 03/16/2023 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,JUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LONEY,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 03/26/2023 21:09:30

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.079	21:04
Air Blank	0.000	21:04
Subject Sample # 1	0.148	21:05
Breath Volume	2.324	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.149	21:07
Breath Volume	1.597	21:07
Air Blank	0.000	21:08
Control Test	0.077	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 04/27/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: KENINGER,DAVID
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: HUTCHESON

DATE/TIME OF TEST: 03/27/2023 15:14:11

Diagnostics Check	OK	15:02	
Air Blank	0.000	15:02	
Control Test	0.079	15:03	
Air Blank	0.000	15:03	
Subject Sample	# 1 0.177	15:06	Volume Not Met
Breath Volume	0.246	15:06	
Air Blank	0.000	15:07	
Air Blank	0.000	15:09	
Subject Sample	# 2 0.169	15:12	Volume Not Met
Breath Volume	0.351	15:12	
Air Blank	0.000	15:13	
Control Test	0.077	15:13	
Air Blank	0.000	15:14	
Diagnostics Check	OK	15:14	

Data Download Date/Time: 04/27/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: KENINGER,DAVID
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: HUTCHESON

DATE/TIME OF TEST: 03/27/2023 15:23:55

Diagnostics Check	OK	15:16
Air Blank	0.000	15:17
Control Test	0.076	15:17
Air Blank	0.000	15:18
Subject Sample # 1	0.221	15:19
Breath Volume	1.738	15:19
Air Blank	0.000	15:19
Air Blank	0.000	15:21
Subject Sample # 2	0.228	15:22
Breath Volume	1.847	15:22
Air Blank	0.000	15:22
Control Test	0.077	15:23
Air Blank	0.000	15:23
Diagnostics Check	OK	15:23

Data Download Date/Time: 04/27/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS, TERRY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 03/01/2023 03:23:59

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.219	03:19
Breath Volume	1.914	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.215	03:22
Breath Volume	2.113	03:22
Air Blank	0.000	03:23
Control Test	0.078	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 03/30/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS,TERRY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: BORSAK

DATE/TIME OF TEST: 03/03/2023 23:41:21

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.078	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.022	23:36
Breath Volume	1.539	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.019	23:39
Breath Volume	1.824	23:39
Air Blank	0.000	23:40
Control Test	0.076	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 03/30/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,GEORGE,O
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LOPEZ,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MCCABE

DATE/TIME OF TEST: 03/04/2023 22:47:48

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.079	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.076	22:43
Breath Volume	1.691	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:45
Subject Sample # 2	0.069	22:46
Breath Volume	1.554	22:46
Air Blank	0.000	22:46
Control Test	0.079	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 03/30/2023 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS,TERRY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 03/08/2023 20:37:34

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:32	
Control Test	0.079	20:33	
Air Blank	0.000	20:33	
Subject Sample	# 1	20:35	Subject Test Refused
Breath Volume	0.015	20:35	
Air Blank	0.000	20:36	
Control Test	0.076	20:36	
Air Blank	0.000	20:37	
Diagnostics Check	OK	20:37	

Data Download Date/Time: 03/30/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: YANICK,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 03/18/2023 03:57:28

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.128	03:52
Breath Volume	2.472	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.119	03:55
Breath Volume	3.277	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 03/30/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,TERRY,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS,T
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BEADNELL

DATE/TIME OF TEST: 03/23/2023 01:59:36

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.181	01:54
Breath Volume	2.125	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.171	01:57
Breath Volume	3.238	01:57
Air Blank	0.000	01:58
Control Test	0.076	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 03/30/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HANEY,TIMOTHY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: HANNAN

DATE/TIME OF TEST: 03/24/2023 05:10:00

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.000	05:05
Breath Volume	1.691	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.000	05:08
Breath Volume	1.847	05:08
Air Blank	0.000	05:09
Control Test	0.078	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:10

Data Download Date/Time: 03/30/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: TARDIFF

DATE/TIME OF TEST: 03/07/2023 03:25:34

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.142	03:20
Breath Volume	2.855	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.140	03:23
Breath Volume	2.882	03:23
Air Blank	0.000	03:24
Control Test	0.078	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,S,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MILLIKEN,C
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: MCNAUGHTON

DATE/TIME OF TEST: 03/07/2023 14:30:32

Diagnostics Check	OK	14:18	
Air Blank	0.000	14:18	
Control Test	0.079	14:19	
Air Blank	0.000	14:19	
Subject Sample	# 1	14:22	No Sample Provided
Breath Volume	0.000	14:22	
Air Blank	0.000	14:23	
Air Blank	0.000	14:25	
Subject Sample	# 2	0.265	14:25
Breath Volume	2.085	14:25	
Air Blank	0.000	14:26	
Air Blank	0.000	14:28	
Subject Sample	# 3	0.266	14:28
Breath Volume	1.828	14:28	
Air Blank	0.000	14:29	
Control Test	0.078	14:29	
Air Blank	0.000	14:30	
Diagnostics Check	OK	14:30	

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 03/07/2023 19:09:07

Diagnostics Check	OK	19:02
Air Blank	0.000	19:02
Control Test	0.080	19:03
Air Blank	0.000	19:03
Subject Sample # 1	0.000	19:04
Breath Volume	2.035	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.000	19:07
Breath Volume	1.718	19:07
Air Blank	0.000	19:08
Control Test	0.079	19:08
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: WAY

DATE/TIME OF TEST: 03/10/2023 02:35:36

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.080	02:27	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:31	No Sample Provided
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 2	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.080	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDERS,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LANIER,Z
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/12/2023 17:31:42

Diagnostics Check	OK	17:28	
Air Blank	0.000	17:28	
Control Test	0.080	17:29	
Air Blank	0.000	17:29	
Subject Sample	# 1	17:30	Interferent Detect
Breath Volume	1.652	17:30	
Air Blank	0.000	17:30	
Control Test	0.080	17:31	
Air Blank	0.000	17:31	
Diagnostics Check	OK	17:31	

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 03/12/2023 22:09:38

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:01	
Control Test	0.080	22:01	
Air Blank	0.000	22:02	
Subject Sample	# 1	22:05	No Sample Provided
Breath Volume	0.000	22:05	
Air Blank	0.000	22:05	
Air Blank	0.000	22:07	
Subject Sample	# 2	22:08	Subject Test Refused
Breath Volume	0.000	22:08	
Air Blank	0.000	22:08	
Control Test	0.079	22:09	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/14/2023 03:54:58

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.252	03:50
Breath Volume	2.289	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.245	03:53
Breath Volume	3.816	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/15/2023 00:34:28

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:30	No Sample Provided
Breath Volume	0.000	00:30	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample	# 2	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:33	
Control Test	0.081	00:33	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BURNETTE

DATE/TIME OF TEST: 03/15/2023 02:36:39

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.081	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.000	02:32
Breath Volume	1.367	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.000	02:35
Breath Volume	1.593	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TIBER,TARIQ
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: BOYETTE

DATE/TIME OF TEST: 03/16/2023 01:41:55

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.081	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:37
Breath Volume	3.472	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.000	01:40
Breath Volume	3.425	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MURPHY,JONES
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 03/17/2023 03:28:52

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.192	03:24
Breath Volume	1.820	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.207	03:27
Breath Volume	2.394	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: VAN HOF,J
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: HASTINGS

DATE/TIME OF TEST: 03/18/2023 02:28:02

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.187	02:23
Breath Volume	1.281	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.183	02:26
Breath Volume	2.835	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:28

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 03/19/2023 02:32:23

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.092	02:27
Breath Volume	2.292	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.091	02:30
Breath Volume	1.839	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: PINKERTON

DATE/TIME OF TEST: 03/19/2023 04:26:42

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.080	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 1 0.097	04:19	Volume Not Met
Breath Volume	0.250	04:19	
Air Blank	0.000	04:20	
Air Blank	0.000	04:21	
Subject Sample	# 2 0.094	04:25	Volume Not Met
Breath Volume	1.074	04:25	
Air Blank	0.000	04:25	
Control Test	0.080	04:26	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 03/24/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HEIN,P
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/24/2023 23:59:36

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.079	23:47	
Air Blank	0.000	23:47	
Subject Sample	# 1 0.000	23:48	
Breath Volume	1.171	23:48	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample	# 2 0.000	23:54	Volume Not Met
Breath Volume	0.187	23:54	
Air Blank	0.000	23:54	
Air Blank	0.000	23:56	
Subject Sample	# 3 0.000	23:58	
Breath Volume	1.160	23:58	
Air Blank	0.000	23:58	
Control Test	0.078	23:59	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 04/28/2023 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 03/25/2023 22:04:04

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.200	21:59
Breath Volume	1.960	21:59
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.209	22:02
Breath Volume	2.941	22:02
Air Blank	0.000	22:03
Control Test	0.077	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:04

Data Download Date/Time: 04/28/2023 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: NEWTON,N
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: FERNNDEZ

DATE/TIME OF TEST: 03/26/2023 19:02:25

Diagnostics Check	OK	18:55
Air Blank	0.000	18:56
Control Test	0.079	18:56
Air Blank	0.000	18:56
Subject Sample # 1	0.304	18:57
Breath Volume	2.859	18:57
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.303	19:00
Breath Volume	2.597	19:00
Air Blank	0.000	19:01
Control Test	0.078	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 04/28/2023 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: BLUMENSTEIN

DATE/TIME OF TEST: 03/26/2023 20:22:21

Diagnostics Check	OK	20:05	
Air Blank	0.000	20:06	
Control Test	0.080	20:06	
Air Blank	0.000	20:07	
Subject Sample # 1	0.160	20:09	
Breath Volume	1.304	20:09	
Air Blank	0.000	20:10	
Air Blank	0.000	20:11	
Subject Sample # 2	0.140	20:15	Volume Not Met
Breath Volume	0.648	20:15	
Air Blank	0.000	20:15	
Air Blank	0.000	20:17	
Subject Sample # 3	0.131	20:20	Slope Not Level
Breath Volume	1.484	20:20	
Air Blank	0.000	20:21	
Control Test	0.077	20:21	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:22	

Data Download Date/Time: 04/28/2023 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: BLUMENSTEIN

DATE/TIME OF TEST: 03/26/2023 20:53:23

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:45	
Control Test	0.078	20:45	
Air Blank	0.000	20:46	
Subject Sample	# 1	20:49	No Sample Provided
Breath Volume	0.000	20:49	
Air Blank	0.000	20:49	
Air Blank	0.000	20:51	
Subject Sample	# 2	20:52	Subject Test Refused
Breath Volume	0.000	20:52	
Air Blank	0.000	20:52	
Control Test	0.079	20:52	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

Data Download Date/Time: 04/28/2023 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: RIVERA CARABALLO

DATE/TIME OF TEST: 03/28/2023 19:07:59

Diagnostics Check	OK	19:01
Air Blank	0.000	19:01
Control Test	0.080	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.066	19:03
Breath Volume	2.187	19:03
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.065	19:06
Breath Volume	2.292	19:06
Air Blank	0.000	19:07
Control Test	0.079	19:07
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 04/28/2023 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TBER,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 03/29/2023 21:35:08

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.079	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.393	21:29
Breath Volume	3.597	21:29
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.379	21:33
Breath Volume	3.542	21:33
Air Blank	0.000	21:34
Control Test	0.079	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 04/28/2023 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/29/2023 23:11:37

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.080	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.187	23:06
Breath Volume	1.679	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.172	23:09
Breath Volume	1.214	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 04/28/2023 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 03/30/2023 03:26:48

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.080	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:22	No Sample Provided
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 2	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.080	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: OUTTIN

DATE/TIME OF TEST: 03/31/2023 01:41:14

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.079	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:29	No Sample Provided
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Air Blank	0.000	01:31	
Subject Sample	# 2	0.047	01:34
Breath Volume	2.414	01:34	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample	# 3	0.051	01:39
Breath Volume	3.101	01:39	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOMAS ANTONIO,ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: REINER

DATE/TIME OF TEST: 03/10/2023 16:52:30

Diagnostics Check	OK	16:46
Air Blank	0.000	16:46
Control Test	0.080	16:47
Air Blank	0.000	16:47
Subject Sample # 1	0.000	16:48
Breath Volume	1.496	16:48
Air Blank	0.000	16:48
Air Blank	0.000	16:50
Subject Sample # 2	0.000	16:51
Breath Volume	3.035	16:51
Air Blank	0.000	16:51
Control Test	0.080	16:51
Air Blank	0.000	16:52
Diagnostics Check	OK	16:52

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MASON,CHAD
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SADLER

DATE/TIME OF TEST: 03/10/2023 22:54:30

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.129	22:50
Breath Volume	1.949	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.127	22:52
Breath Volume	1.574	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: BIELERT

DATE/TIME OF TEST: 03/11/2023 01:41:19

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:28	
Control Test	0.080	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1 0.165	01:32	Volume Not Met
Breath Volume	0.199	01:32	
Air Blank	0.000	01:32	
Air Blank	0.000	01:34	
Subject Sample	# 2 0.165	01:35	
Breath Volume	2.324	01:35	
Air Blank	0.000	01:36	
Air Blank	0.000	01:38	
Subject Sample	# 3 0.175	01:39	
Breath Volume	2.398	01:39	
Air Blank	0.000	01:40	
Control Test	0.078	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 03/11/2023 03:30:53

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.116	03:26
Breath Volume	1.792	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.118	03:29
Breath Volume	3.195	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORGADO,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:38

SUBJECT'S LAST NAME: PHELPS

DATE/TIME OF TEST: 03/11/2023 08:22:50

Diagnostics Check	OK	08:16
Air Blank	0.000	08:16
Control Test	0.080	08:16
Air Blank	0.000	08:17
Subject Sample # 1	0.212	08:18
Breath Volume	2.765	08:18
Air Blank	0.000	08:18
Air Blank	0.000	08:20
Subject Sample # 2	0.211	08:21
Breath Volume	2.707	08:21
Air Blank	0.000	08:21
Control Test	0.080	08:22
Air Blank	0.000	08:22
Diagnostics Check	OK	08:22

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WILKINSON,KATELYN
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: BELU

DATE/TIME OF TEST: 03/11/2023 23:55:03

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.000	23:50
Breath Volume	2.710	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.000	23:53
Breath Volume	2.992	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:55

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: SONTAY GOMEZ

DATE/TIME OF TEST: 03/12/2023 00:51:12

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.079	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.145	00:46
Breath Volume	2.226	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.149	00:49
Breath Volume	2.109	00:49
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 03/12/2023 03:32:18

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.192	03:27
Breath Volume	3.277	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.186	03:30
Breath Volume	2.546	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAYLOR,AUSTIN
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: BOLAND

DATE/TIME OF TEST: 03/12/2023 04:40:34

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.162	04:36
Breath Volume	3.007	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.164	04:38
Breath Volume	1.953	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROCHA,LUIS
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: LEPHART

DATE/TIME OF TEST: 03/12/2023 07:35:24

Diagnostics Check	OK	07:30	
Air Blank	0.000	07:31	
Control Test	0.080	07:31	
Air Blank	0.000	07:31	
Subject Sample	# 1	07:34	Subject Test Refused
Breath Volume	0.000	07:34	
Air Blank	0.000	07:34	
Control Test	0.080	07:34	
Air Blank	0.000	07:35	
Diagnostics Check	OK	07:35	

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOWARD,EDWARD
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: BOYLAN

DATE/TIME OF TEST: 03/12/2023 22:09:04

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.081	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.000	22:04
Breath Volume	2.031	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.000	22:07
Breath Volume	1.914	22:07
Air Blank	0.000	22:08
Control Test	0.079	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:09

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: COLIN

DATE/TIME OF TEST: 03/13/2023 02:24:08

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.081	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1 0.136	02:16	Volume Not Met
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample	# 2 0.123	02:22	Volume Not Met
Breath Volume	0.812	02:22	
Air Blank	0.000	02:23	
Control Test	0.078	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBERTS,ROGER
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: KOHL

DATE/TIME OF TEST: 03/13/2023 18:20:37

Diagnostics Check	OK	18:14
Air Blank	0.000	18:14
Control Test	0.080	18:14
Air Blank	0.000	18:15
Subject Sample # 1	0.253	18:16
Breath Volume	1.410	18:16
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.244	18:19
Breath Volume	1.261	18:19
Air Blank	0.000	18:19
Control Test	0.079	18:20
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MULLEN,GREGORY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 03/13/2023 19:16:43

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.079	19:10
Air Blank	0.000	19:11
Subject Sample # 1	0.242	19:11
Breath Volume	2.453	19:11
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample # 2	0.224	19:15
Breath Volume	2.574	19:15
Air Blank	0.000	19:15
Control Test	0.080	19:16
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BROWN,LEILA
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: FOX JR

DATE/TIME OF TEST: 03/13/2023 22:25:26

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.079	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.171	22:20
Breath Volume	1.785	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.172	22:23
Breath Volume	1.519	22:23
Air Blank	0.000	22:24
Control Test	0.078	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 03/14/2023 00:19:02

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.079	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.122	00:14
Breath Volume	3.167	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:16
Subject Sample # 2	0.130	00:17
Breath Volume	3.296	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:19

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: DEMOSTHENE

DATE/TIME OF TEST: 03/14/2023 04:47:15

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.039	04:42
Breath Volume	1.218	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.036	04:45
Breath Volume	1.617	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HAMMEL,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: NEVERETT

DATE/TIME OF TEST: 03/14/2023 16:20:51

Diagnostics Check	OK	16:14
Air Blank	0.000	16:14
Control Test	0.080	16:15
Air Blank	0.000	16:15
Subject Sample # 1	0.187	16:16
Breath Volume	1.214	16:16
Air Blank	0.000	16:16
Air Blank	0.000	16:18
Subject Sample # 2	0.186	16:19
Breath Volume	1.730	16:19
Air Blank	0.000	16:19
Control Test	0.081	16:20
Air Blank	0.000	16:20
Diagnostics Check	OK	16:20

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FISHER,CHRISTOPHER
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: DANDY

DATE/TIME OF TEST: 03/14/2023 20:16:30

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.080	20:09
Air Blank	0.000	20:10
Subject Sample # 1	0.190	20:11
Breath Volume	2.796	20:11
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.179	20:14
Breath Volume	2.929	20:14
Air Blank	0.000	20:15
Control Test	0.079	20:15
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 03/15/2023 03:53:52

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.081	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.198	03:49
Breath Volume	2.320	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.207	03:52
Breath Volume	2.457	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: NORDE

DATE/TIME OF TEST: 03/15/2023 04:56:31

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:45	
Control Test	0.081	04:45	
Air Blank	0.000	04:46	
Subject Sample	# 1 0.151	04:49	Volume Not Met
Breath Volume	0.292	04:49	
Air Blank	0.000	04:49	
Air Blank	0.000	04:51	
Subject Sample	# 2 0.157	04:54	Volume Not Met
Breath Volume	0.273	04:54	
Air Blank	0.000	04:55	
Control Test	0.078	04:55	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOVETT,JAMES
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: JAUDAS

DATE/TIME OF TEST: 03/15/2023 23:59:20

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.181	23:54
Breath Volume	1.156	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.187	23:57
Breath Volume	2.050	23:57
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHARLES, JAMES LOO
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: MANTANONA

DATE/TIME OF TEST: 03/16/2023 04:35:47

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.186	04:30
Breath Volume	1.226	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.187	04:34
Breath Volume	2.429	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TISDALE, KATIE
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: BARRIOS CHAJ

DATE/TIME OF TEST: 03/18/2023 00:36:17

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.000	00:32
Breath Volume	1.964	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.000	00:34
Breath Volume	2.082	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: BARTOLOME DIEGO

DATE/TIME OF TEST: 03/18/2023 01:11:43

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.160	01:07
Breath Volume	2.640	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.156	01:10
Breath Volume	1.761	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VESCOVA,VALENTINA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: VESCOVA

DATE/TIME OF TEST: 03/18/2023 06:57:59

Diagnostics Check	OK	06:51
Air Blank	0.000	06:52
Control Test	0.080	06:52
Air Blank	0.000	06:52
Subject Sample # 1	0.134	06:53
Breath Volume	3.011	06:53
Air Blank	0.000	06:54
Air Blank	0.000	06:55
Subject Sample # 2	0.130	06:56
Breath Volume	2.371	06:56
Air Blank	0.000	06:57
Control Test	0.080	06:57
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORRISON,ERIC
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:11

SUBJECT'S LAST NAME: SERRANO-RAMIREZ

DATE/TIME OF TEST: 03/18/2023 13:25:22

Diagnostics Check	OK	13:18
Air Blank	0.000	13:18
Control Test	0.080	13:18
Air Blank	0.000	13:19
Subject Sample # 1	0.197	13:20
Breath Volume	2.164	13:20
Air Blank	0.000	13:21
Air Blank	0.000	13:23
Subject Sample # 2	0.191	13:23
Breath Volume	2.460	13:23
Air Blank	0.000	13:24
Control Test	0.079	13:24
Air Blank	0.000	13:25
Diagnostics Check	OK	13:25

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARTINEZ,MARVIN
ARRESTING OFFICER AGENCY: Atlantis PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: MEJIA GARCIA

DATE/TIME OF TEST: 03/18/2023 22:32:08

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:20	
Control Test	0.080	22:21	
Air Blank	0.000	22:21	
Subject Sample	# 1 0.157	22:24	Volume Not Met
Breath Volume	0.503	22:24	
Air Blank	0.000	22:25	
Air Blank	0.000	22:27	
Subject Sample	# 2 0.161	22:30	Volume Not Met
Breath Volume	0.414	22:30	
Air Blank	0.000	22:31	
Control Test	0.078	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: ECHEGOYEN HERNANDEZ

DATE/TIME OF TEST: 03/19/2023 01:38:47

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.133	01:34
Breath Volume	2.167	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.135	01:37
Breath Volume	1.882	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: TORRES FIGUEREDO

DATE/TIME OF TEST: 03/19/2023 04:55:47

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:44	
Control Test	0.079	04:44	
Air Blank	0.000	04:44	
Subject Sample	# 1 0.206	04:47	Volume Not Met
Breath Volume	0.460	04:47	
Air Blank	0.000	04:48	
Air Blank	0.000	04:50	
Subject Sample	# 2 0.213	04:51	
Breath Volume	1.421	04:51	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample	# 3 0.223	04:54	
Breath Volume	1.523	04:54	
Air Blank	0.000	04:54	
Control Test	0.078	04:55	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: VIVAS

DATE/TIME OF TEST: 03/19/2023 22:27:44

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.081	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.195	22:23
Breath Volume	2.285	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.197	22:26
Breath Volume	1.601	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: TEJEDA CHAVEZ

DATE/TIME OF TEST: 03/20/2023 00:15:06

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.080	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.211	00:10
Breath Volume	2.277	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.209	00:13
Breath Volume	2.167	00:13
Air Blank	0.000	00:14
Control Test	0.078	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MARQUIS

DATE/TIME OF TEST: 03/23/2023 03:33:17

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.153	03:28
Breath Volume	3.050	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.147	03:31
Breath Volume	2.347	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 03/23/2023 04:10:07

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.123	04:05
Breath Volume	1.316	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:07
Subject Sample # 2	0.120	04:08
Breath Volume	1.437	04:08
Air Blank	0.000	04:09
Control Test	0.080	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: VO

DATE/TIME OF TEST: 03/24/2023 01:48:00

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:37	
Control Test	0.081	01:37	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:41	No Sample Provided
Breath Volume	0.000	01:41	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample	# 2	01:46	No Sample Provided
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Control Test	0.080	01:47	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: FREED

DATE/TIME OF TEST: 03/24/2023 22:32:52

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.159	22:28
Breath Volume	2.378	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:30
Subject Sample # 2	0.157	22:31
Breath Volume	2.371	22:31
Air Blank	0.000	22:31
Control Test	0.079	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PEDRAZA

DATE/TIME OF TEST: 03/25/2023 01:53:33

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:46	No Sample Provided
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 2	01:52	No Sample Provided
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.080	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: BOSTON

DATE/TIME OF TEST: 03/25/2023 03:03:57

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.000	02:59
Breath Volume	2.269	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.000	03:02
Breath Volume	1.863	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOVETT,JAMES
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SALDARRIAGA

DATE/TIME OF TEST: 03/25/2023 04:30:50

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.119	04:26
Breath Volume	2.531	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:28
Subject Sample # 2	0.121	04:29
Breath Volume	1.742	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: OLIVER JR

DATE/TIME OF TEST: 03/25/2023 07:02:09

Diagnostics Check	OK	06:55
Air Blank	0.000	06:55
Control Test	0.080	06:56
Air Blank	0.000	06:56
Subject Sample # 1	0.000	06:57
Breath Volume	3.015	06:57
Air Blank	0.000	06:58
Air Blank	0.000	07:00
Subject Sample # 2	0.000	07:00
Breath Volume	2.781	07:00
Air Blank	0.000	07:01
Control Test	0.080	07:01
Air Blank	0.000	07:02
Diagnostics Check	OK	07:02

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BELMAN,JONATHAN
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: LOPEZ JUAREZ

DATE/TIME OF TEST: 03/25/2023 22:45:07

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.209	22:39
Breath Volume	1.761	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.209	22:43
Breath Volume	2.273	22:43
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: DOMINGO GARCIA

DATE/TIME OF TEST: 03/26/2023 03:13:09

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.212	03:08
Breath Volume	2.210	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.215	03:11
Breath Volume	1.792	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BYRNE,CHRIS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:28

SUBJECT'S LAST NAME: CHEVERRER OROZCO

DATE/TIME OF TEST: 03/26/2023 13:19:55

Diagnostics Check	OK	13:13
Air Blank	0.000	13:13
Control Test	0.081	13:14
Air Blank	0.000	13:14
Subject Sample # 1	0.237	13:15
Breath Volume	1.609	13:15
Air Blank	0.000	13:16
Air Blank	0.000	13:17
Subject Sample # 2	0.243	13:18
Breath Volume	1.808	13:18
Air Blank	0.000	13:18
Control Test	0.080	13:19
Air Blank	0.000	13:19
Diagnostics Check	OK	13:19

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SEIER,BRADLEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: MONTEJO DOMINGO

DATE/TIME OF TEST: 03/26/2023 19:59:31

Diagnostics Check	OK	19:53
Air Blank	0.000	19:53
Control Test	0.080	19:54
Air Blank	0.000	19:54
Subject Sample # 1	0.253	19:55
Breath Volume	1.566	19:55
Air Blank	0.000	19:55
Air Blank	0.000	19:57
Subject Sample # 2	0.249	19:57
Breath Volume	1.890	19:57
Air Blank	0.000	19:58
Control Test	0.080	19:58
Air Blank	0.000	19:59
Diagnostics Check	OK	19:59

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ, ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 03/29/2023 01:13:09

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.119	01:08
Breath Volume	2.488	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.106	01:11
Breath Volume	1.191	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GITTENS, THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DIAZ PEREZ

DATE/TIME OF TEST: 03/29/2023 22:34:02

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.081	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.142	22:29
Breath Volume	1.156	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.139	22:32
Breath Volume	1.332	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:34

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: NASSIR-HESSARAKI

DATE/TIME OF TEST: 03/31/2023 00:36:42

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.081	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.131	00:32
Breath Volume	1.890	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.128	00:35
Breath Volume	2.171	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 03/31/2023 23:16:47

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.081	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.000	23:12
Breath Volume	2.140	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.000	23:15
Breath Volume	2.652	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 03/18/2023 01:32:45

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.080	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.113	01:28
Breath Volume	2.527	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.114	01:31
Breath Volume	3.285	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 03/28/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,CAMRYN,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HILL,CAMRYN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 03/04/2023 22:52:23

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:49	
Control Test	0.082	22:49	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:51	RFI Detect
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.081	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 04/21/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,CAMRYN,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HILL,CAMRYN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 03/04/2023 23:02:34

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.081	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.095	22:57
Breath Volume	1.312	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.103	23:00
Breath Volume	1.179	23:00
Air Blank	0.000	23:01
Control Test	0.083	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 04/21/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTZER,ETHAN,E
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SMITH,NICKOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: DIX

DATE/TIME OF TEST: 03/08/2023 01:04:49

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.083	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.070	00:58
Breath Volume	2.226	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.066	01:02
Breath Volume	2.277	01:02
Air Blank	0.000	01:03
Control Test	0.077	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 04/21/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADER,JAMES,D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: HANDY

DATE/TIME OF TEST: 03/02/2023 04:42:27

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:31	
Control Test	0.078	04:31	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:35	No Sample Provided
Breath Volume	0.000	04:35	
Air Blank	0.000	04:35	
Air Blank	0.000	04:37	
Subject Sample	# 2	04:41	No Sample Provided
Breath Volume	0.000	04:41	
Air Blank	0.000	04:41	
Control Test	0.078	04:41	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 03/29/2023 20:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,MAURQUICE,D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MILLER,MAURQUICE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: ROQUE

DATE/TIME OF TEST: 03/03/2023 06:14:19

Diagnostics Check	OK	06:05
Air Blank	0.000	06:05
Control Test	0.078	06:06
Air Blank	0.000	06:06
Subject Sample # 1	0.238	06:07
Breath Volume	1.304	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 2	0.225	06:12
Breath Volume	1.140	06:12
Air Blank	0.000	06:13
Control Test	0.076	06:13
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

Data Download Date/Time: 03/29/2023 20:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN, TYLER, B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: ANDA, JOHN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 03/14/2023 19:58:40

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.077	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.265	19:53
Breath Volume	3.550	19:53
Air Blank	0.000	19:55
Air Blank	0.000	19:56
Subject Sample # 2	0.261	19:56
Breath Volume	3.492	19:56
Air Blank	0.000	19:57
Control Test	0.075	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 03/29/2023 20:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STREETER,RAFAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/28/2023 20:43:23

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.078	20:36
Air Blank	0.000	20:36
Subject Sample # 1	0.182	20:37
Breath Volume	2.187	20:37
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.164	20:41
Breath Volume	2.273	20:41
Air Blank	0.000	20:42
Control Test	0.075	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 04/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CAMILA,P
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PADILLA,ARNOLD
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: RAMOS RAMIREZ

DATE/TIME OF TEST: 03/19/2023 21:52:09

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.163	21:47
Breath Volume	1.843	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.171	21:50
Breath Volume	1.171	21:50
Air Blank	0.000	21:51
Control Test	0.078	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 04/13/2023 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROJAS CORTES,EUGENIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,Y
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: ROJAS CORTES

DATE/TIME OF TEST: 03/20/2023 02:16:36

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:13	
Control Test	0.079	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:15	RFI Detect
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.079	02:16	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 04/13/2023 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROJAS CORTES,EUGENIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,Y
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: ROJAS CORTES

DATE/TIME OF TEST: 03/20/2023 02:45:58

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.079	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:45	
Control Test	0.080	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 04/13/2023 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOPEZ,ANTHONY,M

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CAPAZ,TAMMY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1497672

DRY GAS EXPIRY DATE: 05/20/2024

VIOLATION CODE: DUI

ARREST TIME: 00:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: OLIVA

DATE/TIME OF TEST: 03/22/2023 01:15:45

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:13	
Control Test	0.079	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Control Test	0.079	01:15	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 04/13/2023 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1497672
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A	DRY GAS EXPIRY DATE: 05/20/2024
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO	ARREST TIME: 23:20
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Male
LAST AGENCY INSPECTION: 03/14/2023	OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HERNANDEZ MACHADO	DATE/TIME OF TEST: 03/24/2023 00:10:15
--	--

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:04	
Control Test	0.079	00:05	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:08	No Sample Provided
Breath Volume	0.000	00:08	
Air Blank		00:09	Purge Fail
Air Blank	0.000	00:10	

Data Download Date/Time: 04/13/2023 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1497672

DRY GAS EXPIRY DATE: 05/20/2024

VIOLATION CODE: DUI

ARREST TIME: 23:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HERNANDEZ MACHADO

DATE/TIME OF TEST: 03/24/2023 00:18:54

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:16	
Control Test	0.079	00:16	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:17	Subject Test Refused
Breath Volume	0.000	00:17	
Air Blank	0.000	00:17	
Control Test	0.079	00:18	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 04/13/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ,ALBERTO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: COBO,J

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1497672

DRY GAS EXPIRY DATE: 05/20/2024

VIOLATION CODE: DUI

ARREST TIME: 06:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: PADRON PALMERO

DATE/TIME OF TEST: 03/25/2023 07:24:02

Diagnostics Check	OK	07:15
Air Blank	0.000	07:16
Control Test	0.079	07:16
Air Blank	0.000	07:17
Subject Sample # 1	0.196	07:19
Breath Volume	3.121	07:19
Air Blank	0.000	07:20
Air Blank	0.000	07:21
Subject Sample # 2	0.200	07:22
Breath Volume	2.164	07:22
Air Blank	0.000	07:23
Control Test	0.076	07:23
Air Blank	0.000	07:23
Diagnostics Check	OK	07:24

Data Download Date/Time: 04/13/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,R
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: RIVER MEDRANO

DATE/TIME OF TEST: 03/25/2023 21:16:27

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:05	
Control Test	0.080	21:05	
Air Blank	0.000	21:05	
Subject Sample	# 1 0.165	21:09	Volume Not Met
Breath Volume	0.437	21:09	
Air Blank	0.000	21:09	
Air Blank	0.000	21:11	
Subject Sample	# 2 0.161	21:14	Volume Not Met
Breath Volume	0.160	21:14	
Air Blank	0.000	21:15	
Control Test	0.078	21:15	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 04/13/2023 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: RIVERA MEDRANO

DATE/TIME OF TEST: 03/25/2023 21:27:22

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.079	21:21
Air Blank	0.000	21:21
Subject Sample # 1	0.218	21:22
Breath Volume	2.062	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.223	21:25
Breath Volume	2.179	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 04/13/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: RODRIGUEZ ROJAS

DATE/TIME OF TEST: 03/26/2023 00:14:30

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.079	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.130	00:09
Breath Volume	2.531	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.129	00:12
Breath Volume	2.894	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 04/13/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,R,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RAMOS,R
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ANDURAY LOPEZ

DATE/TIME OF TEST: 03/26/2023 04:46:30

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.079	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.218	04:41
Breath Volume	1.871	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:43
Subject Sample # 2	0.203	04:44
Breath Volume	2.910	04:44
Air Blank	0.000	04:45
Control Test	0.077	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 04/13/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVARRIA,DORA,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARAUZ,A
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: MORALESSANCHEZ

DATE/TIME OF TEST: 03/28/2023 00:04:03

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 04/13/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO,KENNETH,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: PRIETO,KENNETH
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: LINDBACK

DATE/TIME OF TEST: 03/01/2023 00:47:40

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.079	00:41	
Air Blank	0.000	00:42	
Subject Sample # 1	0.142	00:43	
Breath Volume	2.265	00:43	
Air Blank	0.000	00:44	
Air Blank	0.000	00:45	
Subject Sample # 2	0.145	00:46	
Breath Volume	1.832	00:46	
Air Blank		00:47	RFI Detect
Air Blank	0.000	00:47	

Data Download Date/Time: 03/07/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: ROTA

DATE/TIME OF TEST: 03/03/2023 19:32:50

Diagnostics Check	OK	19:20	
Air Blank	0.000	19:20	
Control Test	0.080	19:20	
Air Blank	0.000	19:21	
Subject Sample	# 1 0.086	19:24	Volume Not Met
Breath Volume	0.933	19:24	
Air Blank	0.000	19:25	
Air Blank	0.000	19:27	
Subject Sample	# 2 0.092	19:27	
Breath Volume	1.210	19:27	
Air Blank	0.000	19:28	
Air Blank	0.000	19:30	
Subject Sample	# 3 0.100	19:31	
Breath Volume	1.714	19:31	
Air Blank	0.000	19:31	
Control Test	0.078	19:32	
Air Blank	0.000	19:32	
Diagnostics Check	OK	19:32	

Data Download Date/Time: 03/07/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONACO,ASHLEY,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HARRILL,ANDERSEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 02/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 03/05/2023 03:43:15

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.167	03:38
Breath Volume	1.414	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.175	03:41
Breath Volume	1.273	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 03/07/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GIL,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MCCARTHY,MARGARET
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 03/09/2023 06:30:04

Diagnostics Check	OK	06:26	
Air Blank	0.000	06:27	
Control Test	0.078	06:27	
Air Blank	0.000	06:28	
Subject Sample	# 1	06:28	Subject Test Refused
Breath Volume	0.000	06:28	
Air Blank	0.000	06:29	
Control Test	0.077	06:29	
Air Blank	0.000	06:29	
Diagnostics Check	OK	06:30	

Data Download Date/Time: 04/05/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ESPINOSA, OSCAR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: STILSON

DATE/TIME OF TEST: 03/10/2023 22:14:28

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:12	
Control Test	0.079	22:12	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:13	
Control Test	0.079	22:13	
Air Blank	0.000	22:14	
Diagnostics Check	OK	22:14	

Data Download Date/Time: 04/05/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,ANFERNEE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 03/12/2023 01:41:20

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.078	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank		01:40	RFI Detect
Air Blank	0.000	01:41	

Data Download Date/Time: 04/05/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CORTES, CHRISTIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: DORSEY

DATE/TIME OF TEST: 03/13/2023 00:37:14

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.078	00:35	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:36	
Control Test	0.078	00:36	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 04/05/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONACO,ASHLEY,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MONACO,ASHLEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/15/2023 01:32:22

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.078	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.204	01:27
Breath Volume	2.976	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.201	01:30
Breath Volume	1.996	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 04/05/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GARCIA,JULIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: RACHOWICZ

DATE/TIME OF TEST: 03/15/2023 02:58:23

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.077	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.247	02:53
Breath Volume	3.027	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.255	02:56
Breath Volume	2.632	02:56
Air Blank	0.000	02:57
Control Test	0.076	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 04/05/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONACO,ASHLEY,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MONACO,ASHLEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: CHAVEZ-NUNEZ

DATE/TIME OF TEST: 03/18/2023 02:22:38

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.071	02:18
Breath Volume	2.054	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.074	02:21
Breath Volume	1.871	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 04/05/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HARRILL,ANDERSEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: TAMARGO

DATE/TIME OF TEST: 03/19/2023 02:19:55

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.078	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:18	
Control Test	0.078	02:19	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 04/05/2023 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,JORGE,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: DOMINGUEZ,OSCAR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: BLUCHER

DATE/TIME OF TEST: 03/22/2023 00:20:52

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:18	
Control Test	0.080	00:18	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:19	Subject Test Refused
Breath Volume	0.000	00:19	
Air Blank	0.000	00:19	
Control Test	0.078	00:20	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 04/05/2023 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 10:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: BRALEY

DATE/TIME OF TEST: 03/23/2023 23:30:55

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.080	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Control Test	0.079	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 04/05/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,JORGE,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ESPINOZA,OSCAR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 03/24/2023 23:10:06

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Subject Sample # 1	0.048	23:05
Breath Volume	1.574	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.051	23:08
Breath Volume	1.359	23:08
Air Blank	0.000	23:09
Control Test	0.078	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 04/05/2023 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,JORGE,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 03/25/2023 03:38:15

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.078	03:36	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Control Test	0.080	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 04/05/2023 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STURGEON,JOHN,S
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: STURGEON,JOHN
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: BUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: SANCHEZ HERNANDEZ

DATE/TIME OF TEST: 03/25/2023 23:10:10

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:55	
Control Test	0.079	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1 0.176	22:59	Volume Not Met
Breath Volume	0.203	22:59	
Air Blank	0.000	22:59	
Air Blank	0.000	23:01	
Subject Sample	# 2 0.192	23:02	
Breath Volume	1.468	23:02	
Air Blank	0.000	23:03	
Air Blank	0.000	23:05	
Subject Sample	# 3 0.189	23:08	Volume Not Met
Breath Volume	0.121	23:08	
Air Blank	0.000	23:09	
Control Test	0.077	23:09	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

Data Download Date/Time: 04/05/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLEY,KYLE,C
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: POLEY,KYLE
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: BUI
ARREST TIME: 13:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: BALASQUIDE

DATE/TIME OF TEST: 03/26/2023 14:37:15

Diagnostics Check	OK	14:30
Air Blank	0.000	14:31
Control Test	0.078	14:31
Air Blank	0.000	14:31
Subject Sample # 1	0.047	14:32
Breath Volume	2.972	14:32
Air Blank	0.000	14:33
Air Blank	0.000	14:35
Subject Sample # 2	0.045	14:35
Breath Volume	3.066	14:35
Air Blank	0.000	14:36
Control Test	0.076	14:36
Air Blank	0.000	14:37
Diagnostics Check	OK	14:37

Data Download Date/Time: 04/05/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALAZAR,JUSTIN,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: ESPINOSA,MIGUEL
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: BUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: JASPER

DATE/TIME OF TEST: 03/26/2023 20:31:21

Diagnostics Check	OK	20:19	
Air Blank	0.000	20:20	
Control Test	0.078	20:20	
Air Blank	0.000	20:21	
Subject Sample	# 1	20:24	No Sample Provided
Breath Volume	0.000	20:24	
Air Blank	0.000	20:24	
Air Blank	0.000	20:26	
Subject Sample	# 2	20:29	No Sample Provided
Breath Volume	0.000	20:29	
Air Blank	0.000	20:30	
Control Test	0.079	20:30	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 04/05/2023 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,DANIEL,W
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: MARSHALL,DANIEL
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/28/2023 21:28:00

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.000	21:23
Breath Volume	1.554	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.000	21:26
Breath Volume	3.816	21:26
Air Blank	0.000	21:27
Control Test	0.077	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:28

Data Download Date/Time: 04/05/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROITH,GREGORY,S
BREATH TEST OPERATOR AGENCY: FHP TROOP A EAST
ARRESTING OFFICER: GONZALEZ SANDAR,JESUS
ARRESTING OFFICER AGENCY: FHP TROOP A EAST
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: ROITH

DATE/TIME OF TEST: 03/29/2023 03:02:57

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.078	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.048	02:58
Breath Volume	2.179	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample # 2	0.050	03:01
Breath Volume	2.058	03:01
Air Blank	0.000	03:02
Control Test	0.076	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 04/05/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASIMIRO
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SANTIAGO MARTEL

DATE/TIME OF TEST: 03/12/2023 01:57:18

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.025	01:52
Breath Volume	2.453	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.026	01:55
Breath Volume	1.992	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 03/13/2023 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 03/12/2023 08:52:40

Diagnostics Check	OK	08:45
Air Blank	0.000	08:46
Control Test	0.079	08:46
Air Blank	0.000	08:46
Subject Sample # 1	0.180	08:47
Breath Volume	1.488	08:47
Air Blank	0.000	08:48
Air Blank	0.000	08:49
Subject Sample # 2	0.180	08:50
Breath Volume	2.371	08:50
Air Blank	0.000	08:51
Control Test	0.080	08:52
Air Blank	0.000	08:52
Diagnostics Check	OK	08:52

Data Download Date/Time: 03/13/2023 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,MARCOS
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 06:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:04

SUBJECT'S LAST NAME: BARRETO ALFARO

DATE/TIME OF TEST: 03/12/2023 09:36:57

Diagnostics Check	OK	09:30
Air Blank	0.000	09:30
Control Test	0.079	09:30
Air Blank	0.000	09:31
Subject Sample # 1	0.158	09:32
Breath Volume	2.050	09:32
Air Blank	0.000	09:32
Air Blank	0.000	09:34
Subject Sample # 2	0.149	09:35
Breath Volume	1.785	09:35
Air Blank	0.000	09:36
Control Test	0.078	09:36
Air Blank	0.000	09:36
Diagnostics Check	OK	09:36

Data Download Date/Time: 03/13/2023 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 03/13/2023 05:19:28

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.079	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.189	05:14
Breath Volume	1.468	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:16
Subject Sample # 2	0.190	05:17
Breath Volume	2.246	05:17
Air Blank	0.000	05:18
Control Test	0.077	05:18
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 03/13/2023 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 03/15/2023 05:09:29

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.077	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.067	05:04
Breath Volume	2.039	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.065	05:07
Breath Volume	2.035	05:07
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MARTINEZ URQUIZA

DATE/TIME OF TEST: 03/16/2023 02:04:51

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.079	01:53	
Air Blank	0.000	01:54	
Subject Sample	# 1 0.263	01:57	Volume Not Met
Breath Volume	0.945	01:57	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 2 0.268	02:00	
Breath Volume	1.804	02:00	
Air Blank	0.000	02:00	
Air Blank	0.000	02:02	
Subject Sample	# 3 0.286	02:03	
Breath Volume	1.394	02:03	
Air Blank	0.000	02:03	
Control Test	0.077	02:04	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 03/16/2023 02:21:43

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.078	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.167	02:17
Breath Volume	3.042	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.149	02:20
Breath Volume	1.179	02:20
Air Blank	0.000	02:20
Control Test	0.077	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: FRIEDMAN

DATE/TIME OF TEST: 03/16/2023 05:19:29

Diagnostics Check	OK	05:13
Air Blank	0.000	05:13
Control Test	0.078	05:13
Air Blank	0.000	05:14
Subject Sample # 1	0.144	05:14
Breath Volume	3.046	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.144	05:17
Breath Volume	2.878	05:17
Air Blank	0.000	05:18
Control Test	0.078	05:18
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: STETSON

DATE/TIME OF TEST: 03/16/2023 23:23:16

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.079	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.055	23:17
Breath Volume	1.199	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.047	23:21
Breath Volume	1.406	23:21
Air Blank	0.000	23:22
Control Test	0.078	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: METHELUS

DATE/TIME OF TEST: 03/17/2023 00:19:18

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.078	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.247	00:14
Breath Volume	1.835	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.246	00:17
Breath Volume	1.937	00:17
Air Blank	0.000	00:18
Control Test	0.078	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CASTELLANO,SALVATORE
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: GATAULLIN

DATE/TIME OF TEST: 03/18/2023 04:50:25

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.200	04:44
Breath Volume	1.187	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.190	04:48
Breath Volume	2.753	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HEPBURN,ANDREW
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: LIVIDINI

DATE/TIME OF TEST: 03/18/2023 18:18:00

Diagnostics Check	OK	18:11
Air Blank	0.000	18:11
Control Test	0.079	18:12
Air Blank	0.000	18:12
Subject Sample # 1	0.123	18:13
Breath Volume	2.582	18:13
Air Blank	0.000	18:14
Air Blank	0.000	18:15
Subject Sample # 2	0.121	18:16
Breath Volume	2.714	18:16
Air Blank	0.000	18:17
Control Test	0.078	18:17
Air Blank	0.000	18:17
Diagnostics Check	OK	18:18

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BROWN,JEREMIAH
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: MATUL GONZALEZ

DATE/TIME OF TEST: 03/18/2023 19:30:37

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.078	19:23
Air Blank	0.000	19:24
Subject Sample # 1	0.165	19:25
Breath Volume	2.515	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 2	0.159	19:28
Breath Volume	1.488	19:28
Air Blank	0.000	19:29
Control Test	0.077	19:29
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 03/20/2023 04:18:52

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.182	04:14
Breath Volume	1.921	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.178	04:17
Breath Volume	1.746	04:17
Air Blank	0.000	04:17
Control Test	0.077	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CHARNECHO,TAYLOR
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: HINOJOSA

DATE/TIME OF TEST: 03/21/2023 02:51:18

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.000	02:46
Breath Volume	1.867	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.000	02:49
Breath Volume	2.042	02:49
Air Blank	0.000	02:50
Control Test	0.078	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ,FRANK
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: VISNIC

DATE/TIME OF TEST: 03/21/2023 06:21:45

Diagnostics Check	OK	06:15
Air Blank	0.000	06:15
Control Test	0.079	06:16
Air Blank	0.000	06:16
Subject Sample # 1	0.160	06:17
Breath Volume	2.710	06:17
Air Blank	0.000	06:17
Air Blank	0.000	06:19
Subject Sample # 2	0.153	06:20
Breath Volume	1.863	06:20
Air Blank	0.000	06:20
Control Test	0.079	06:21
Air Blank	0.000	06:21
Diagnostics Check	OK	06:21

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOLDBERG,JOSHUA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/21/2023 21:25:00

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:21	
Control Test	0.078	21:22	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:23	Slope Not Met
Breath Volume	1.167	21:23	
Air Blank	0.000	21:24	
Control Test	0.078	21:24	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOLDBERG,JOSHUA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/21/2023 22:09:49

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.079	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.272	22:04
Breath Volume	2.304	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.280	22:08
Breath Volume	2.644	22:08
Air Blank	0.000	22:08
Control Test	0.078	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/21/2023 23:47:01

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.249	23:42
Breath Volume	2.781	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.245	23:45
Breath Volume	1.742	23:45
Air Blank	0.000	23:46
Control Test	0.077	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:47

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/26/2023 03:37:21

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.078	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.172	03:31
Breath Volume	1.148	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.163	03:35
Breath Volume	1.585	03:35
Air Blank	0.000	03:36
Control Test	0.077	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRUJILLO,ELAINE
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 08:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:41

SUBJECT'S LAST NAME: CALONA ESCOBAR

DATE/TIME OF TEST: 03/26/2023 10:12:40

Diagnostics Check	OK	10:06
Air Blank	0.000	10:06
Control Test	0.078	10:06
Air Blank	0.000	10:07
Subject Sample # 1	0.150	10:07
Breath Volume	2.910	10:07
Air Blank	0.000	10:08
Air Blank	0.000	10:10
Subject Sample # 2	0.148	10:11
Breath Volume	2.917	10:11
Air Blank	0.000	10:11
Control Test	0.078	10:12
Air Blank	0.000	10:12
Diagnostics Check	OK	10:12

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BENINCASA,VINCENT
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: COLASANTE SUAREZ

DATE/TIME OF TEST: 03/29/2023 05:27:33

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.079	05:20
Air Blank	0.000	05:20
Subject Sample # 1	0.197	05:22
Breath Volume	2.750	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.207	05:25
Breath Volume	2.804	05:25
Air Blank	0.000	05:26
Control Test	0.077	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FONSECA,LAVINCH
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 03/30/2023 06:38:39

Diagnostics Check	OK	06:35	
Air Blank	0.000	06:36	
Control Test	0.078	06:36	
Air Blank	0.000	06:37	
Subject Sample	# 1	06:37	Subject Test Refused
Breath Volume	0.000	06:37	
Air Blank	0.000	06:37	
Control Test	0.078	06:38	
Air Blank	0.000	06:38	
Diagnostics Check	OK	06:38	

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 03/03/2023 22:36:42

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.079	22:30
Air Blank	0.000	22:30
Subject Sample # 1	0.186	22:31
Breath Volume	1.683	22:31
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.188	22:34
Breath Volume	2.179	22:34
Air Blank	0.000	22:35
Control Test	0.079	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 03/10/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ,ISIAH
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/05/2023 02:10:16

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.244	02:05
Breath Volume	2.683	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.242	02:08
Breath Volume	2.046	02:08
Air Blank	0.000	02:09
Control Test	0.077	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 03/10/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MANAK,EDWARD
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/05/2023 03:05:35

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.078	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.346	03:00
Breath Volume	1.191	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.342	03:03
Breath Volume	1.460	03:03
Air Blank	0.000	03:04
Control Test	0.077	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 03/10/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: RAMIREZ-VELASQUEZ

DATE/TIME OF TEST: 03/05/2023 04:00:20

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.142	03:55
Breath Volume	1.769	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.129	03:58
Breath Volume	1.160	03:58
Air Blank	0.000	03:59
Control Test	0.077	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 03/10/2023 17:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: LOPEZ AJPUAC

DATE/TIME OF TEST: 03/05/2023 23:26:44

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.079	23:22	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:24	Subject Test Refused
Breath Volume	0.246	23:24	
Air Blank	0.000	23:25	
Control Test	0.080	23:26	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 03/10/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: JACOBSEN

DATE/TIME OF TEST: 03/06/2023 03:31:41

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.000	03:27
Breath Volume	1.589	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.000	03:30
Breath Volume	1.558	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 03/10/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEACH, ZACHARY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 02/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: MALOW

DATE/TIME OF TEST: 03/08/2023 19:32:50

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.079	19:27
Air Blank	0.000	19:27
Subject Sample # 1	0.267	19:28
Breath Volume	3.351	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:30
Subject Sample # 2	0.270	19:31
Breath Volume	2.816	19:31
Air Blank	0.000	19:31
Control Test	0.079	19:32
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 03/10/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: CINTO MACARIO

DATE/TIME OF TEST: 03/09/2023 04:38:36

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.138	04:33
Breath Volume	1.246	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.137	04:36
Breath Volume	2.457	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 03/10/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: JIMENEZ MARCOS

DATE/TIME OF TEST: 03/10/2023 00:15:19

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.080	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.260	00:10
Breath Volume	2.761	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:12
Subject Sample # 2	0.255	00:13
Breath Volume	1.960	00:13
Air Blank	0.000	00:14
Control Test	0.076	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 03/10/2023 17:03