

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BIONDO,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: LIU

DATE/TIME OF TEST: 04/01/2023 03:42:00

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.000	03:37
Breath Volume	2.183	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.000	03:40
Breath Volume	2.046	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:42

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEDHAM,K,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BRYANT,S
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: FIGUEROA POUCHIE

DATE/TIME OF TEST: 04/01/2023 20:36:22

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.080	20:29
Air Blank	0.000	20:29
Subject Sample # 1	0.141	20:30
Breath Volume	1.160	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.150	20:34
Breath Volume	1.988	20:34
Air Blank	0.000	20:35
Control Test	0.080	20:35
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PETERS,DARIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: HERNANDEZ LOPEZ

DATE/TIME OF TEST: 04/02/2023 03:09:18

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.186	03:04
Breath Volume	1.828	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.196	03:07
Breath Volume	3.468	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LOWER, TIFFANY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: PACIFICI

DATE/TIME OF TEST: 04/04/2023 22:19:30

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.132	22:15
Breath Volume	3.722	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.147	22:17
Breath Volume	3.953	22:17
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PETERS,DARIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: DAVENPORT

DATE/TIME OF TEST: 04/06/2023 03:23:19

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.081	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.156	03:18
Breath Volume	1.500	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample # 2	0.158	03:21
Breath Volume	2.089	03:21
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 04/07/2023 00:29:31

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.081	00:27	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.081	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 04/07/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/08/2023 23:02:02

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.202	22:57
Breath Volume	3.500	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.197	23:00
Breath Volume	3.730	23:00
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:02

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:29

SUBJECT'S LAST NAME: BAILON PEREZ

DATE/TIME OF TEST: 04/09/2023 09:02:41

Diagnostics Check	OK	08:55
Air Blank	0.000	08:56
Control Test	0.081	08:56
Air Blank	0.000	08:57
Subject Sample # 1	0.196	08:58
Breath Volume	1.769	08:58
Air Blank	0.000	08:58
Air Blank	0.000	09:00
Subject Sample # 2	0.204	09:01
Breath Volume	2.125	09:01
Air Blank	0.000	09:01
Control Test	0.080	09:02
Air Blank	0.000	09:02
Diagnostics Check	OK	09:02

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: CHAMBON

DATE/TIME OF TEST: 04/10/2023 02:43:42

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.239	02:38
Breath Volume	1.644	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample # 2	0.253	02:42
Breath Volume	1.257	02:42
Air Blank	0.000	02:42
Control Test	0.078	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: WOOSTER

DATE/TIME OF TEST: 04/10/2023 22:29:49

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.193	22:25
Breath Volume	3.406	22:25
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.193	22:28
Breath Volume	2.878	22:28
Air Blank	0.000	22:28
Control Test	0.080	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: MAGRO

DATE/TIME OF TEST: 04/11/2023 09:49:10

Diagnostics Check	OK	09:42
Air Blank	0.000	09:42
Control Test	0.081	09:43
Air Blank	0.000	09:43
Subject Sample # 1	0.018	09:44
Breath Volume	1.843	09:44
Air Blank	0.000	09:45
Air Blank	0.000	09:47
Subject Sample # 2	0.017	09:47
Breath Volume	2.828	09:47
Air Blank	0.000	09:48
Control Test	0.080	09:48
Air Blank	0.000	09:49
Diagnostics Check	OK	09:49

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SAMSELSKI

DATE/TIME OF TEST: 04/11/2023 20:06:21

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.081	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.202	20:01
Breath Volume	2.519	20:01
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.210	20:04
Breath Volume	2.410	20:04
Air Blank	0.000	20:05
Control Test	0.080	20:05
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PETERS,DARIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 04/12/2023 04:04:13

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.126	03:59
Breath Volume	4.167	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.113	04:02
Breath Volume	2.023	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: LUJAN

DATE/TIME OF TEST: 04/12/2023 22:08:07

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.081	22:00
Air Blank	0.000	22:00
Subject Sample # 1	0.147	22:02
Breath Volume	1.449	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.145	22:06
Breath Volume	1.421	22:06
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: JACINTO

DATE/TIME OF TEST: 04/13/2023 01:20:12

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.081	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.289	01:15
Breath Volume	1.363	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:17
Subject Sample # 2	0.269	01:18
Breath Volume	2.597	01:18
Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDARELLI,CARA,G
BREATH TEST OPERATOR AGENCY: Sewall's Point PD
ARRESTING OFFICER: CARDARELLI,CARA
ARRESTING OFFICER AGENCY: Sewall's Point PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: IRIZARRY

DATE/TIME OF TEST: 04/14/2023 01:40:40

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.081	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.264	01:36
Breath Volume	4.015	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.256	01:39
Breath Volume	3.242	01:39
Air Blank	0.000	01:39
Control Test	0.078	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: WEIK

DATE/TIME OF TEST: 04/14/2023 21:20:59

Diagnostics Check	OK	21:06	
Air Blank	0.000	21:07	
Control Test	0.081	21:07	
Air Blank	0.000	21:07	
Subject Sample	# 1 0.154	21:08	
Breath Volume	1.316	21:08	
Air Blank	0.000	21:08	
Air Blank	0.000	21:10	
Subject Sample	# 2 0.158	21:13	Volume Not Met
Breath Volume	0.031	21:13	
Air Blank	0.000	21:14	
Air Blank	0.000	21:16	
Subject Sample	# 3 0.121	21:19	Volume Not Met
Breath Volume	0.050	21:19	
Air Blank	0.000	21:20	
Control Test	0.076	21:20	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 04/16/2023 01:31:56

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:16	
Control Test	0.080	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1 0.232	01:20	Volume Not Met
Breath Volume	1.035	01:20	
Air Blank	0.000	01:20	
Air Blank	0.000	01:22	
Subject Sample	# 2 0.218	01:24	
Breath Volume	1.250	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 3 0.231	01:30	Volume Not Met
Breath Volume	0.503	01:30	
Air Blank	0.000	01:30	
Control Test	0.079	01:31	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TOUSSAINT,WESNES
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: ROMERO HERNANDEZ

DATE/TIME OF TEST: 04/17/2023 00:48:24

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.255	00:43
Breath Volume	2.921	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.258	00:46
Breath Volume	2.761	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HATTON,NICKOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MURILLO GOMEZ

DATE/TIME OF TEST: 04/17/2023 02:54:48

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.081	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.095	02:50
Breath Volume	2.089	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.098	02:53
Breath Volume	2.566	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/20/2023 22:57:47

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:55	
Control Test	0.081	22:55	
Air Blank	0.000	22:56	
Subject Sample	# 1	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:56	
Control Test	0.080	22:57	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MULROONEY

DATE/TIME OF TEST: 04/21/2023 01:55:46

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.081	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Control Test	0.080	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: FORBES,RAY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/23/2023 00:05:01

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.081	00:02	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:04	
Control Test	0.080	00:04	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ALVAREZ-ORANTES

DATE/TIME OF TEST: 04/23/2023 03:18:49

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.000	03:14
Breath Volume	3.214	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.000	03:17
Breath Volume	1.550	03:17
Air Blank	0.000	03:17
Control Test	0.081	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WADE,TYLER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: BLANC

DATE/TIME OF TEST: 04/23/2023 06:54:31

Diagnostics Check	OK	06:48
Air Blank	0.000	06:48
Control Test	0.081	06:49
Air Blank	0.000	06:49
Subject Sample # 1	0.130	06:50
Breath Volume	2.250	06:50
Air Blank	0.000	06:50
Air Blank	0.000	06:52
Subject Sample # 2	0.130	06:53
Breath Volume	1.527	06:53
Air Blank	0.000	06:53
Control Test	0.080	06:53
Air Blank	0.000	06:54
Diagnostics Check	OK	06:54

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: RODNEY

DATE/TIME OF TEST: 04/24/2023 00:45:35

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.081	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.156	00:41
Breath Volume	1.808	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.159	00:44
Breath Volume	2.878	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: DALEY

DATE/TIME OF TEST: 04/25/2023 02:28:54

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.154	02:24
Breath Volume	2.214	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.144	02:27
Breath Volume	1.910	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: HASLOM

DATE/TIME OF TEST: 04/25/2023 21:45:09

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.081	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.076	21:40
Breath Volume	1.800	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:42
Subject Sample # 2	0.075	21:43
Breath Volume	2.371	21:43
Air Blank	0.000	21:44
Control Test	0.081	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: LAMARRE

DATE/TIME OF TEST: 04/26/2023 00:02:48

Diagnostics Check	OK	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.214	23:58
Breath Volume	1.199	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.223	00:01
Breath Volume	1.671	00:01
Air Blank	0.000	00:01
Control Test	0.079	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTHOLOMEW, MARK, A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BARTHOLOMEW, MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: HARWOOD

DATE/TIME OF TEST: 04/28/2023 17:28:01

Diagnostics Check	OK	17:22
Air Blank	0.000	17:22
Control Test	0.081	17:22
Air Blank	0.000	17:23
Subject Sample # 1	0.000	17:23
Breath Volume	1.507	17:23
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 2	0.000	17:26
Breath Volume	4.136	17:26
Air Blank	0.000	17:27
Control Test	0.080	17:27
Air Blank	0.000	17:27
Diagnostics Check	OK	17:28

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: TOMAS-FRANCISCO

DATE/TIME OF TEST: 04/29/2023 01:32:29

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.081	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.163	01:27
Breath Volume	2.933	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.159	01:30
Breath Volume	2.000	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 04/29/2023 17:33:21

Diagnostics Check	OK	17:23	
Air Blank	0.000	17:24	
Control Test	0.081	17:24	
Air Blank	0.000	17:25	
Subject Sample	# 1 0.359	17:25	No .020 agreement
Breath Volume	1.738	17:25	
Air Blank	0.000	17:26	
Air Blank	0.000	17:28	
Subject Sample	# 2 0.385	17:28	No .020 agreement
Breath Volume	3.292	17:28	
Air Blank	0.000	17:29	
Air Blank	0.000	17:31	
Subject Sample	# 3 0.303	17:31	No .020 agreement
Breath Volume	1.316	17:31	
Air Blank	0.000	17:32	
Control Test	0.080	17:32	
Air Blank	0.000	17:33	
Diagnostics Check	OK	17:33	

Data Download Date/Time: 05/04/2023 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: ISAAC

DATE/TIME OF TEST: 04/30/2023 00:18:59

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.081	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.146	00:14
Breath Volume	1.386	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:16
Subject Sample # 2	0.143	00:17
Breath Volume	2.792	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 05/04/2023 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: BACKUS

DATE/TIME OF TEST: 04/30/2023 02:29:37

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.208	02:25
Breath Volume	2.542	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.197	02:28
Breath Volume	3.738	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 05/04/2023 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: JOSE JIMENEZ

DATE/TIME OF TEST: 04/30/2023 21:44:12

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.081	21:37
Air Blank	0.000	21:37
Subject Sample # 1	0.240	21:39
Breath Volume	3.378	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:41
Subject Sample # 2	0.238	21:42
Breath Volume	3.535	21:42
Air Blank	0.000	21:43
Control Test	0.080	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 05/04/2023 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 04/30/2023 22:15:52

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.081	22:14	
Air Blank		22:15	RFI Detect
Air Blank	0.000	22:15	

Data Download Date/Time: 05/04/2023 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 04/30/2023 22:20:52

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:18	
Control Test	0.080	22:18	
Air Blank	0.000	22:19	
Subject Sample	# 1	22:19	Subject Test Refused
Breath Volume	0.000	22:19	
Air Blank	0.000	22:19	
Control Test	0.080	22:20	
Air Blank	0.000	22:20	
Diagnostics Check	OK	22:20	

Data Download Date/Time: 05/04/2023 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLO,DEREK,J
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARISH,CLINT
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: HEREDIA

DATE/TIME OF TEST: 04/06/2023 06:36:37

Diagnostics Check	OK	06:28
Air Blank	0.000	06:28
Control Test	0.079	06:28
Air Blank	0.000	06:29
Subject Sample # 1	0.045	06:31
Breath Volume	2.203	06:31
Air Blank	0.000	06:31
Air Blank	0.000	06:33
Subject Sample # 2	0.039	06:34
Breath Volume	1.562	06:34
Air Blank	0.000	06:35
Control Test	0.075	06:36
Air Blank	0.000	06:36
Diagnostics Check	OK	06:36

Data Download Date/Time: 04/27/2023 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: POTH,LAUREN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: BURNETTE

DATE/TIME OF TEST: 04/29/2023 00:41:16

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:38	
Control Test	0.079	00:38	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:39	Subject Test Refused
Breath Volume	0.000	00:39	
Air Blank	0.000	00:40	
Control Test	0.077	00:40	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 05/26/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY, MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 04/11/2023 02:16:06

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:02	
Control Test	0.080	02:02	
Air Blank	0.000	02:03	
Subject Sample # 1	0.000	02:03	
Breath Volume	1.937	02:03	
Air Blank	0.000	02:03	
Air Blank	0.000	02:05	
Subject Sample # 2		02:09	No Sample Provided
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Air Blank	0.000	02:11	
Subject Sample # 3		02:14	No Sample Provided
Breath Volume	0.000	02:14	
Air Blank	0.000	02:15	
Control Test	0.080	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 04/28/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY, MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 04/11/2023 03:04:21

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.000	03:00
Breath Volume	2.339	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.000	03:02
Breath Volume	2.375	03:02
Air Blank	0.000	03:03
Control Test	0.077	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 04/28/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,T,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS,T
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: ESSIG

DATE/TIME OF TEST: 04/14/2023 23:06:16

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.079	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.045	23:01
Breath Volume	2.652	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:03
Subject Sample # 2	0.041	23:04
Breath Volume	2.664	23:04
Air Blank	0.000	23:05
Control Test	0.077	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 04/28/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: YANICK, MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: FLORES GALINDO

DATE/TIME OF TEST: 04/16/2023 00:18:00

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.245	00:13
Breath Volume	1.925	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:15
Subject Sample # 2	0.231	00:16
Breath Volume	1.804	00:16
Air Blank	0.000	00:17
Control Test	0.078	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:18

Data Download Date/Time: 04/28/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: KING,BRIAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MAXFIELD

DATE/TIME OF TEST: 04/18/2023 01:22:01

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.079	01:17	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:20	Subject Test Refused
Breath Volume	0.000	01:20	
Air Blank	0.000	01:21	
Control Test	0.080	01:21	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 04/28/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BATCH,JOSEPH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MENDOZA CRUZ

DATE/TIME OF TEST: 04/22/2023 03:32:26

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.176	03:27
Breath Volume	2.402	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.172	03:30
Breath Volume	1.445	03:30
Air Blank	0.000	03:31
Control Test	0.075	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 04/28/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: TARR,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: HERNANDEZ DE LEON

DATE/TIME OF TEST: 04/23/2023 05:08:27

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.204	05:03
Breath Volume	2.804	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.201	05:06
Breath Volume	3.171	05:06
Air Blank	0.000	05:07
Control Test	0.077	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 04/28/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERLAIN,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CHAMBERLAIN,JOSHUA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: BELANGER

DATE/TIME OF TEST: 04/24/2023 23:25:05

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.079	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.145	23:20
Breath Volume	2.726	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:22
Subject Sample # 2	0.132	23:23
Breath Volume	1.648	23:23
Air Blank	0.000	23:24
Control Test	0.076	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 04/28/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MATTHEWS, HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 04/25/2023 17:43:21

Diagnostics Check	OK	17:36
Air Blank	0.000	17:37
Control Test	0.080	17:37
Air Blank	0.000	17:38
Subject Sample # 1	0.325	17:38
Breath Volume	1.464	17:38
Air Blank	0.000	17:39
Air Blank	0.000	17:41
Subject Sample # 2	0.313	17:41
Breath Volume	1.363	17:41
Air Blank	0.000	17:42
Control Test	0.077	17:42
Air Blank	0.000	17:43
Diagnostics Check	OK	17:43

Data Download Date/Time: 04/28/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY, MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: WADDELL

DATE/TIME OF TEST: 04/30/2023 21:55:01

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:45	
Control Test	0.080	21:46	
Air Blank	0.000	21:46	
Subject Sample # 1	0.162	21:49	
Breath Volume	2.152	21:49	
Air Blank	0.000	21:50	
Air Blank	0.000	21:51	
Subject Sample # 2	0.153	21:53	
Breath Volume	2.128	21:53	
Air Blank	0.000	21:54	
Control Test	0.073	21:54	Control Outside Tolerance
Air Blank	0.000	21:55	

Data Download Date/Time: 05/25/2023 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY, MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: WADDELL

DATE/TIME OF TEST: 04/30/2023 22:11:25

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:58	
Control Test	0.078	21:59	
Air Blank	0.000	21:59	
Subject Sample	# 1 0.147	22:00	
Breath Volume	2.015	22:00	
Air Blank	0.000	22:01	
Air Blank	0.000	22:03	
Subject Sample	# 2	22:06	No Sample Provided
Breath Volume	0.000	22:06	
Air Blank	0.000	22:06	
Air Blank	0.000	22:08	
Subject Sample	# 3 0.166	22:09	
Breath Volume	1.164	22:09	
Air Blank	0.000	22:10	
Control Test	0.078	22:10	
Air Blank	0.000	22:11	
Diagnostics Check	OK	22:11	

Data Download Date/Time: 05/25/2023 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY, MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: WADDELL

DATE/TIME OF TEST: 04/30/2023 22:31:03

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.077	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.146	22:25
Breath Volume	1.363	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.147	22:29
Breath Volume	2.070	22:29
Air Blank	0.000	22:30
Control Test	0.076	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:31

Data Download Date/Time: 05/25/2023 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: FARLIN

DATE/TIME OF TEST: 04/10/2023 18:47:48

Diagnostics Check	OK	18:41
Air Blank	0.000	18:41
Control Test	0.078	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.000	18:43
Breath Volume	2.476	18:43
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.000	18:46
Breath Volume	3.179	18:46
Air Blank	0.000	18:46
Control Test	0.078	18:47
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 05/15/2023 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: ELSBERRY

DATE/TIME OF TEST: 04/13/2023 06:14:56

Diagnostics Check	OK	06:07
Air Blank	0.000	06:08
Control Test	0.079	06:08
Air Blank	0.000	06:08
Subject Sample # 1	0.251	06:09
Breath Volume	2.410	06:09
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.239	06:13
Breath Volume	1.800	06:13
Air Blank	0.000	06:14
Control Test	0.077	06:14
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

Data Download Date/Time: 05/15/2023 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: BURCH

DATE/TIME OF TEST: 04/13/2023 20:13:53

Diagnostics Check	OK	20:06
Air Blank	0.000	20:06
Control Test	0.078	20:07
Air Blank	0.000	20:07
Subject Sample # 1	0.118	20:08
Breath Volume	1.816	20:08
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.115	20:12
Breath Volume	1.753	20:12
Air Blank	0.000	20:12
Control Test	0.077	20:13
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 05/15/2023 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: SORIANO

DATE/TIME OF TEST: 04/15/2023 19:29:38

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:28	
Control Test	0.000	19:29	RFI Detect
Air Blank	0.000	19:29	

Data Download Date/Time: 05/15/2023 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: SORIANO

DATE/TIME OF TEST: 04/15/2023 19:38:48

Diagnostics Check	OK	19:31
Air Blank	0.000	19:31
Control Test	0.078	19:32
Air Blank	0.000	19:32
Subject Sample # 1	0.175	19:33
Breath Volume	1.839	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.168	19:37
Breath Volume	1.558	19:37
Air Blank	0.000	19:37
Control Test	0.076	19:38
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 05/15/2023 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: VALDEZ ESQUIVEL

DATE/TIME OF TEST: 04/15/2023 22:51:59

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.078	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.010	22:47
Breath Volume	2.457	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:50
Subject Sample # 2	0.010	22:50
Breath Volume	2.785	22:50
Air Blank	0.000	22:51
Control Test	0.079	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 05/15/2023 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRIGNANO,JESSICA,L
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: RIOS,MARCOS
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:16

SUBJECT'S LAST NAME: ANDRADE

DATE/TIME OF TEST: 04/16/2023 08:49:13

Diagnostics Check	OK	08:41
Air Blank	0.000	08:42
Control Test	0.079	08:42
Air Blank	0.000	08:43
Subject Sample # 1	0.027	08:44
Breath Volume	1.972	08:44
Air Blank	0.000	08:44
Air Blank	0.000	08:46
Subject Sample # 2	0.025	08:47
Breath Volume	1.699	08:47
Air Blank	0.000	08:48
Control Test	0.079	08:48
Air Blank	0.000	08:49
Diagnostics Check	OK	08:49

Data Download Date/Time: 05/15/2023 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 04/20/2023 01:13:00

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.151	01:08
Breath Volume	1.675	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.158	01:11
Breath Volume	2.996	01:11
Air Blank	0.000	01:12
Control Test	0.077	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:13

Data Download Date/Time: 05/15/2023 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: DE PAOLA

DATE/TIME OF TEST: 04/21/2023 00:52:45

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.170	00:48
Breath Volume	1.578	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:50
Subject Sample # 2	0.169	00:51
Breath Volume	1.765	00:51
Air Blank	0.000	00:51
Control Test	0.076	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 05/15/2023 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: DINH

DATE/TIME OF TEST: 04/23/2023 00:14:33

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.048	00:09
Breath Volume	1.378	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.047	00:12
Breath Volume	2.460	00:12
Air Blank	0.000	00:13
Control Test	0.078	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 05/15/2023 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TINFINA,LOGAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TINFINA,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ZACKERY

DATE/TIME OF TEST: 04/24/2023 00:11:09

Diagnostics Check	OK	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.000	00:06
Breath Volume	1.824	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.000	00:09
Breath Volume	1.203	00:09
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 05/15/2023 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: CENTOLANZA

DATE/TIME OF TEST: 04/25/2023 13:55:19

Diagnostics Check	OK	13:48
Air Blank	0.000	13:49
Control Test	0.079	13:49
Air Blank	0.000	13:50
Subject Sample # 1	0.324	13:50
Breath Volume	2.609	13:50
Air Blank	0.000	13:51
Air Blank	0.000	13:53
Subject Sample # 2	0.321	13:53
Breath Volume	2.390	13:53
Air Blank	0.000	13:54
Control Test	0.078	13:54
Air Blank	0.000	13:55
Diagnostics Check	OK	13:55

Data Download Date/Time: 05/15/2023 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/27/2023 03:57:43

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.067	03:53
Breath Volume	3.992	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.068	03:56
Breath Volume	3.753	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 05/15/2023 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: AMAYA-NUNEZ

DATE/TIME OF TEST: 04/28/2023 16:34:44

Diagnostics Check	OK	16:28
Air Blank	0.000	16:28
Control Test	0.079	16:28
Air Blank	0.000	16:29
Subject Sample # 1	0.288	16:30
Breath Volume	2.269	16:30
Air Blank	0.000	16:30
Air Blank	0.000	16:32
Subject Sample # 2	0.307	16:33
Breath Volume	1.656	16:33
Air Blank	0.000	16:33
Control Test	0.077	16:34
Air Blank	0.000	16:34
Diagnostics Check	OK	16:34

Data Download Date/Time: 05/15/2023 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 08:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: GARCIA CASIQUE

DATE/TIME OF TEST: 04/30/2023 09:40:38

Diagnostics Check	OK	09:33
Air Blank	0.000	09:33
Control Test	0.079	09:33
Air Blank	0.000	09:34
Subject Sample # 1	0.208	09:35
Breath Volume	2.867	09:35
Air Blank	0.000	09:35
Air Blank	0.000	09:37
Subject Sample # 2	0.210	09:38
Breath Volume	1.902	09:38
Air Blank	0.000	09:39
Control Test	0.078	09:40
Air Blank	0.000	09:40
Diagnostics Check	OK	09:40

Data Download Date/Time: 05/15/2023 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: LIZENBEE

DATE/TIME OF TEST: 04/01/2023 00:41:03

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.080	00:33	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:36	No Sample Provided
Breath Volume	0.000	00:36	
Air Blank	0.000	00:37	
Air Blank	0.000	00:39	
Subject Sample	# 2	00:39	Subject Test Refused
Breath Volume	0.000	00:39	
Air Blank	0.000	00:40	
Control Test	0.080	00:40	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/02/2023 01:43:43

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.266	01:38
Breath Volume	2.074	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.264	01:42
Breath Volume	1.300	01:42
Air Blank	0.000	01:42
Control Test	0.078	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CROSBY,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: CASILLAS

DATE/TIME OF TEST: 04/02/2023 19:54:20

Diagnostics Check	OK	19:47
Air Blank	0.000	19:48
Control Test	0.080	19:48
Air Blank	0.000	19:49
Subject Sample # 1	0.226	19:49
Breath Volume	2.902	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.225	19:52
Breath Volume	2.289	19:52
Air Blank	0.000	19:53
Control Test	0.079	19:53
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: ST.ONGE

DATE/TIME OF TEST: 04/03/2023 19:28:21

Diagnostics Check	OK	19:18	
Air Blank	0.000	19:19	
Control Test	0.080	19:19	
Air Blank	0.000	19:20	
Subject Sample	# 1	19:23	No Sample Provided
Breath Volume	0.000	19:23	
Air Blank	0.000	19:23	
Air Blank	0.000	19:25	
Subject Sample	# 2	19:27	RFI Detect
Breath Volume	0.000	19:27	
Air Blank		19:27	RFI Detect
Air Blank	0.000	19:28	

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: ST.ONGE

DATE/TIME OF TEST: 04/03/2023 20:01:23

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.080	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.261	19:56
Breath Volume	2.687	19:56
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.265	19:59
Breath Volume	3.132	19:59
Air Blank	0.000	20:00
Control Test	0.078	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 04/06/2023 03:33:20

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:25	No Sample Provided
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2	0.191	03:28
Breath Volume	1.414	03:28	
Air Blank	0.000	03:29	
Air Blank	0.000	03:31	
Subject Sample	# 3	0.193	03:31
Breath Volume	2.738	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/06/2023 21:03:28

Diagnostics Check	OK	20:57
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.280	20:58
Breath Volume	2.027	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.272	21:01
Breath Volume	1.816	21:01
Air Blank	0.000	21:02
Control Test	0.079	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SKEEN,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: BARRON

DATE/TIME OF TEST: 04/07/2023 08:21:27

Diagnostics Check	OK	08:12
Air Blank	0.000	08:13
Control Test	0.080	08:13
Air Blank	0.000	08:14
Subject Sample # 1	0.369	08:16
Breath Volume	2.582	08:16
Air Blank	0.000	08:17
Air Blank	0.000	08:19
Subject Sample # 2	0.367	08:19
Breath Volume	2.300	08:19
Air Blank	0.000	08:20
Control Test	0.076	08:20
Air Blank	0.000	08:21
Diagnostics Check	OK	08:21

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BOLGER

DATE/TIME OF TEST: 04/08/2023 02:14:49

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.134	02:09
Breath Volume	3.109	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.135	02:13
Breath Volume	3.292	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HELWIG

DATE/TIME OF TEST: 04/08/2023 23:47:56

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.192	23:43
Breath Volume	2.195	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.195	23:46
Breath Volume	1.144	23:46
Air Blank	0.000	23:47
Control Test	0.078	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 04/09/2023 03:55:54

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.079	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1 0.176	03:42	
Breath Volume	1.582	03:42	
Air Blank	0.000	03:43	
Air Blank	0.000	03:45	
Subject Sample	# 2	03:48	No Sample Provided
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Air Blank	0.000	03:51	
Subject Sample	# 3 0.162	03:54	Volume Not Met
Breath Volume	0.195	03:54	
Air Blank	0.000	03:54	
Control Test	0.080	03:55	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/11/2023 02:51:17

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.155	02:46
Breath Volume	2.117	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.140	02:49
Breath Volume	2.371	02:49
Air Blank	0.000	02:50
Control Test	0.078	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: LARY

DATE/TIME OF TEST: 04/12/2023 03:28:18

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.000	03:23
Breath Volume	3.597	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.000	03:26
Breath Volume	3.617	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/15/2023 23:00:46

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.079	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:56	No Sample Provided
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 2	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	22:59	
Control Test	0.080	23:00	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 04/28/2023 08:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: SKELTON JR

DATE/TIME OF TEST: 04/16/2023 00:23:43

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:15	
Control Test	0.080	00:15	
Air Blank	0.000	00:16	
Subject Sample	# 1	00:19	No Sample Provided
Breath Volume	0.000	00:19	
Air Blank	0.000	00:20	
Air Blank	0.000	00:22	
Subject Sample	# 2	00:22	Subject Test Refused
Breath Volume	0.000	00:22	
Air Blank	0.000	00:22	
Control Test	0.080	00:23	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 04/28/2023 08:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CROSBY,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: KARDON

DATE/TIME OF TEST: 04/16/2023 19:51:49

Diagnostics Check	OK	19:45
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:46
Subject Sample # 1	0.202	19:47
Breath Volume	2.113	19:47
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.206	19:50
Breath Volume	2.632	19:50
Air Blank	0.000	19:50
Control Test	0.079	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 04/28/2023 08:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: DOAN

DATE/TIME OF TEST: 04/20/2023 03:27:33

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.181	03:22
Breath Volume	2.328	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.164	03:25
Breath Volume	2.429	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 04/28/2023 08:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SHARQAWI

DATE/TIME OF TEST: 04/21/2023 00:05:23

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:57	
Control Test	0.080	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1	00:01	No Sample Provided
Breath Volume	0.000	00:01	
Air Blank	0.000	00:01	
Air Blank	0.000	00:03	
Subject Sample	# 2	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Control Test	0.080	00:04	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 04/28/2023 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 04/25/2023 22:54:54

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.079	22:47	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:50	No Sample Provided
Breath Volume	0.000	22:50	
Air Blank	0.000	22:51	
Air Blank	0.000	22:53	
Subject Sample	# 2	22:53	Subject Test Refused
Breath Volume	0.000	22:53	
Air Blank	0.000	22:53	
Control Test	0.080	22:54	
Air Blank	0.000	22:54	
Diagnostics Check	OK	22:54	

Data Download Date/Time: 04/28/2023 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 04/28/2023 01:20:35

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.212	01:15
Breath Volume	1.246	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.230	01:18
Breath Volume	1.601	01:18
Air Blank	0.000	01:19
Control Test	0.078	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 04/28/2023 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/28/2023 03:53:08

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.079	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1 0.130	03:43	Volume Not Met
Breath Volume	1.378	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2 0.163	03:47	
Breath Volume	2.000	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:50	
Subject Sample	# 3 0.161	03:51	
Breath Volume	1.957	03:51	
Air Blank	0.000	03:52	
Control Test	0.077	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 04/28/2023 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: WILKEN II

DATE/TIME OF TEST: 04/28/2023 21:29:31

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.079	21:21	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:25	No Sample Provided
Breath Volume	0.000	21:25	
Air Blank	0.000	21:25	
Air Blank	0.000	21:27	
Subject Sample	# 2	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:28	
Control Test	0.079	21:28	
Air Blank	0.000	21:29	
Diagnostics Check	OK	21:29	

Data Download Date/Time: 05/26/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: WINSLOW

DATE/TIME OF TEST: 04/29/2023 02:24:54

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:16	
Control Test	0.080	02:16	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:20	No Sample Provided
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample	# 2	02:23	Subject Test Refused
Breath Volume	0.000	02:23	
Air Blank	0.000	02:23	
Control Test	0.080	02:24	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 05/26/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CROSBY,H
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 04/30/2023 01:44:28

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.150	01:39
Breath Volume	3.335	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.147	01:42
Breath Volume	3.023	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 05/26/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HERANDEZ,E
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: RHODEN

DATE/TIME OF TEST: 04/30/2023 02:43:35

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.159	02:38
Breath Volume	1.214	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.170	02:41
Breath Volume	2.183	02:41
Air Blank	0.000	02:42
Control Test	0.077	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 05/26/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/30/2023 02:59:38

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.154	02:55
Breath Volume	1.660	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.159	02:58
Breath Volume	1.394	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 05/26/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CROSBY,H
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 04/30/2023 20:07:40

Diagnostics Check	OK	19:53	
Air Blank	0.000	19:53	
Control Test	0.079	19:54	
Air Blank	0.000	19:54	
Subject Sample	# 1 0.196	19:57	Volume Not Met
Breath Volume	0.039	19:57	
Air Blank	0.000	19:58	
Air Blank	0.000	20:00	
Subject Sample	# 2 0.217	20:01	
Breath Volume	1.761	20:01	
Air Blank	0.000	20:02	
Air Blank	0.000	20:04	
Subject Sample	# 3 0.211	20:05	
Breath Volume	2.175	20:05	
Air Blank	0.000	20:06	
Control Test	0.079	20:07	
Air Blank	0.000	20:07	
Diagnostics Check	OK	20:07	

Data Download Date/Time: 05/26/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PICARD,RONNIE
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: MACHLEID

DATE/TIME OF TEST: 04/15/2023 00:45:35

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.147	00:40
Breath Volume	1.402	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.145	00:44
Breath Volume	1.164	00:44
Air Blank	0.000	00:44
Control Test	0.079	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANEGO,DAVID
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/15/2023 01:45:33

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.156	01:41
Breath Volume	2.476	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.154	01:43
Breath Volume	1.855	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: LOPEZ ORDONEZ

DATE/TIME OF TEST: 04/15/2023 03:45:07

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.080	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.230	03:40
Breath Volume	3.027	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:42
Subject Sample # 2	0.231	03:43
Breath Volume	3.320	03:43
Air Blank	0.000	03:44
Control Test	0.078	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANEGO,DAVID
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: WOZNICK

DATE/TIME OF TEST: 04/16/2023 00:33:55

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.227	00:29
Breath Volume	1.894	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:31
Subject Sample # 2	0.228	00:32
Breath Volume	1.957	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GODINEZ SANCHEZ

DATE/TIME OF TEST: 04/16/2023 01:25:05

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.212	01:19
Breath Volume	1.234	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.217	01:23
Breath Volume	1.367	01:23
Air Blank	0.000	01:24
Control Test	0.077	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: JADOS

DATE/TIME OF TEST: 04/16/2023 06:40:16

Diagnostics Check	OK	06:34	
Air Blank	0.000	06:34	
Control Test	0.080	06:35	
Air Blank	0.000	06:35	
Subject Sample	# 1	06:38	Subject Test Refused
Breath Volume	0.179	06:38	
Air Blank	0.000	06:39	
Control Test	0.080	06:39	
Air Blank	0.000	06:40	
Diagnostics Check	OK	06:40	

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: TORRES MURILLO

DATE/TIME OF TEST: 04/16/2023 07:17:44

Diagnostics Check	OK	07:13	
Air Blank	0.000	07:14	
Control Test	0.079	07:14	
Air Blank	0.000	07:15	
Subject Sample	# 1	07:16	Subject Test Refused
Breath Volume	0.000	07:16	
Air Blank	0.000	07:16	
Control Test	0.080	07:17	
Air Blank	0.000	07:17	
Diagnostics Check	OK	07:17	

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NELER, SAM
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:21

SUBJECT'S LAST NAME: SALES

DATE/TIME OF TEST: 04/16/2023 09:58:43

Diagnostics Check	OK	09:52
Air Blank	0.000	09:53
Control Test	0.080	09:53
Air Blank	0.000	09:53
Subject Sample # 1	0.142	09:54
Breath Volume	2.132	09:54
Air Blank	0.000	09:54
Air Blank	0.000	09:56
Subject Sample # 2	0.130	09:57
Breath Volume	1.410	09:57
Air Blank	0.000	09:57
Control Test	0.080	09:58
Air Blank	0.000	09:58
Diagnostics Check	OK	09:58

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 04/16/2023 23:10:15

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.170	23:05
Breath Volume	1.261	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.182	23:08
Breath Volume	1.929	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 04/17/2023 01:05:12

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.080	01:02	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:04	
Control Test	0.080	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: GOEHRINGER

DATE/TIME OF TEST: 04/17/2023 01:19:29

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.162	01:14
Breath Volume	1.679	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.158	01:17
Breath Volume	3.074	01:17
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BRANT

DATE/TIME OF TEST: 04/17/2023 22:56:36

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.109	22:52
Breath Volume	2.050	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.099	22:55
Breath Volume	1.566	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: ISAGUIRRE

DATE/TIME OF TEST: 04/18/2023 21:01:30

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.081	20:55
Air Blank	0.000	20:56
Subject Sample # 1	0.224	20:56
Breath Volume	3.449	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.223	20:59
Breath Volume	2.929	20:59
Air Blank	0.000	21:00
Control Test	0.080	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: AUGUSTIN

DATE/TIME OF TEST: 04/18/2023 21:23:20

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.079	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.180	21:18
Breath Volume	1.246	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.180	21:21
Breath Volume	1.402	21:21
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 04/19/2023 00:49:17

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.081	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.000	00:45
Breath Volume	1.753	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.000	00:47
Breath Volume	1.839	00:47
Air Blank	0.000	00:48
Control Test	0.081	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RINI,LAWRENCE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: BODO DE SOLER

DATE/TIME OF TEST: 04/20/2023 11:59:01

Diagnostics Check	OK	11:52
Air Blank	0.000	11:52
Control Test	0.081	11:53
Air Blank	0.000	11:53
Subject Sample # 1	0.000	11:54
Breath Volume	1.132	11:54
Air Blank	0.000	11:54
Air Blank	0.000	11:56
Subject Sample # 2	0.000	11:57
Breath Volume	1.195	11:57
Air Blank	0.000	11:58
Control Test	0.081	11:58
Air Blank	0.000	11:58
Diagnostics Check	OK	11:59

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: MATIAS

DATE/TIME OF TEST: 04/20/2023 20:36:21

Diagnostics Check	OK	20:30
Air Blank	0.000	20:30
Control Test	0.081	20:30
Air Blank	0.000	20:31
Subject Sample # 1	0.165	20:31
Breath Volume	3.453	20:31
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.158	20:34
Breath Volume	3.171	20:34
Air Blank	0.000	20:35
Control Test	0.079	20:35
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: PADILLA DISCUA

DATE/TIME OF TEST: 04/21/2023 03:06:02

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.137	03:01
Breath Volume	1.707	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.149	03:04
Breath Volume	2.070	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:06

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:59

SUBJECT'S LAST NAME: PASCUAL ANDRES

DATE/TIME OF TEST: 04/21/2023 17:36:01

Diagnostics Check	OK	17:29
Air Blank	0.000	17:29
Control Test	0.080	17:29
Air Blank	0.000	17:30
Subject Sample # 1	0.249	17:31
Breath Volume	1.402	17:31
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample # 2	0.234	17:34
Breath Volume	1.613	17:34
Air Blank	0.000	17:35
Control Test	0.079	17:35
Air Blank	0.000	17:35
Diagnostics Check	OK	17:36

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: OLIVER VAZQUEZ

DATE/TIME OF TEST: 04/22/2023 05:13:53

Diagnostics Check	OK	05:07
Air Blank	0.000	05:08
Control Test	0.081	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.232	05:09
Breath Volume	1.992	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:11
Subject Sample # 2	0.225	05:12
Breath Volume	2.269	05:12
Air Blank	0.000	05:12
Control Test	0.080	05:13
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 04/23/2023 00:38:27

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.081	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.191	00:33
Breath Volume	1.394	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.193	00:36
Breath Volume	1.589	00:36
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PICARD,RONNIE
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 04/23/2023 02:12:41

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.081	02:04	
Air Blank	0.000	02:05	
Subject Sample	# 1 0.141	02:08	Volume Not Met
Breath Volume	0.019	02:08	
Air Blank	0.000	02:09	
Air Blank	0.000	02:11	
Subject Sample	# 2	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Control Test	0.079	02:12	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DIAS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 04/23/2023 05:32:10

Diagnostics Check	OK	05:25
Air Blank	0.000	05:26
Control Test	0.081	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.161	05:27
Breath Volume	2.125	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.155	05:30
Breath Volume	1.867	05:30
Air Blank	0.000	05:31
Control Test	0.080	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CLAUDIO,JOSE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:42

SUBJECT'S LAST NAME: RAMIREZ RAMIREZ

DATE/TIME OF TEST: 04/23/2023 10:26:51

Diagnostics Check	OK	10:19
Air Blank	0.000	10:19
Control Test	0.080	10:19
Air Blank	0.000	10:20
Subject Sample # 1	0.182	10:22
Breath Volume	2.117	10:22
Air Blank	0.000	10:22
Air Blank	0.000	10:24
Subject Sample # 2	0.182	10:25
Breath Volume	1.828	10:25
Air Blank	0.000	10:25
Control Test	0.078	10:26
Air Blank	0.000	10:26
Diagnostics Check	OK	10:26

Data Download Date/Time: 05/19/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BENNETT,JONATHAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: ALICEA JR

DATE/TIME OF TEST: 04/25/2023 17:55:15

Diagnostics Check	OK	17:48
Air Blank	0.000	17:48
Control Test	0.081	17:49
Air Blank	0.000	17:49
Subject Sample # 1	0.235	17:50
Breath Volume	2.710	17:50
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 2	0.230	17:53
Breath Volume	2.878	17:53
Air Blank	0.000	17:54
Control Test	0.080	17:54
Air Blank	0.000	17:55
Diagnostics Check	OK	17:55

Data Download Date/Time: 05/19/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 04/25/2023 19:29:44

Diagnostics Check	OK	19:22
Air Blank	0.000	19:22
Control Test	0.081	19:23
Air Blank	0.000	19:23
Subject Sample # 1	0.217	19:24
Breath Volume	2.261	19:24
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.211	19:28
Breath Volume	2.539	19:28
Air Blank	0.000	19:28
Control Test	0.079	19:29
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 05/19/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CARONE

DATE/TIME OF TEST: 04/29/2023 02:23:05

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.112	02:17
Breath Volume	1.546	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.109	02:21
Breath Volume	1.246	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 05/19/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BALLATORI,MARC
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: NEHER

DATE/TIME OF TEST: 04/30/2023 02:24:11

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.080	02:19
Breath Volume	2.316	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.079	02:22
Breath Volume	2.179	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 05/19/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: REYES MORALES

DATE/TIME OF TEST: 04/30/2023 20:18:35

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.082	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.193	20:13
Breath Volume	1.683	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:15
Subject Sample # 2	0.189	20:16
Breath Volume	1.968	20:16
Air Blank	0.000	20:17
Control Test	0.079	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 05/19/2023 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: WARONKER

DATE/TIME OF TEST: 04/30/2023 21:31:45

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.081	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.000	21:26
Breath Volume	1.402	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.000	21:30
Breath Volume	1.199	21:30
Air Blank	0.000	21:30
Control Test	0.080	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 05/19/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOOHEY,DOUGLAS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: GARCIA-MORENO

DATE/TIME OF TEST: 04/30/2023 22:54:42

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.081	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.194	22:49
Breath Volume	2.003	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.197	22:53
Breath Volume	2.039	22:53
Air Blank	0.000	22:53
Control Test	0.078	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 05/19/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ALVES SILVA

DATE/TIME OF TEST: 04/01/2023 03:38:49

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.271	03:34
Breath Volume	1.664	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample # 2	0.266	03:37
Breath Volume	1.640	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BRYANT,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: MALLONEE

DATE/TIME OF TEST: 04/01/2023 06:52:50

Diagnostics Check	OK	06:46
Air Blank	0.000	06:46
Control Test	0.081	06:46
Air Blank	0.000	06:47
Subject Sample # 1	0.145	06:48
Breath Volume	3.652	06:48
Air Blank	0.000	06:49
Air Blank	0.000	06:50
Subject Sample # 2	0.138	06:51
Breath Volume	2.503	06:51
Air Blank	0.000	06:51
Control Test	0.080	06:52
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: CONSOLINO

DATE/TIME OF TEST: 04/01/2023 21:13:08

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.079	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.000	21:08
Breath Volume	2.738	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.000	21:11
Breath Volume	2.886	21:11
Air Blank	0.000	21:12
Control Test	0.079	21:12
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: OVEREND

DATE/TIME OF TEST: 04/02/2023 00:14:07

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.191	00:09
Breath Volume	2.207	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.199	00:12
Breath Volume	1.800	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: MANGAN

DATE/TIME OF TEST: 04/02/2023 02:45:40

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.138	02:41
Breath Volume	1.460	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.140	02:44
Breath Volume	1.917	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GUERRIER,JEFFERSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: PAREDES PAVON

DATE/TIME OF TEST: 04/02/2023 08:26:06

Diagnostics Check	OK	08:19
Air Blank	0.000	08:19
Control Test	0.080	08:20
Air Blank	0.000	08:20
Subject Sample # 1	0.105	08:21
Breath Volume	2.613	08:21
Air Blank	0.000	08:22
Air Blank	0.000	08:23
Subject Sample # 2	0.105	08:24
Breath Volume	2.171	08:24
Air Blank	0.000	08:25
Control Test	0.080	08:25
Air Blank	0.000	08:26
Diagnostics Check	OK	08:26

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ONG,ANDY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 04/02/2023 14:34:05

Diagnostics Check	OK	14:27
Air Blank	0.000	14:27
Control Test	0.080	14:28
Air Blank	0.000	14:28
Subject Sample # 1	0.065	14:29
Breath Volume	1.937	14:29
Air Blank	0.000	14:30
Air Blank	0.000	14:31
Subject Sample # 2	0.066	14:32
Breath Volume	2.484	14:32
Air Blank	0.000	14:33
Control Test	0.079	14:33
Air Blank	0.000	14:34
Diagnostics Check	OK	14:34

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: REA,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: BUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 04/02/2023 17:12:03

Diagnostics Check	OK	17:05
Air Blank	0.000	17:05
Control Test	0.080	17:06
Air Blank	0.000	17:06
Subject Sample # 1	0.167	17:07
Breath Volume	1.929	17:07
Air Blank	0.000	17:08
Air Blank	0.000	17:09
Subject Sample # 2	0.176	17:10
Breath Volume	1.628	17:10
Air Blank	0.000	17:11
Control Test	0.079	17:11
Air Blank	0.000	17:11
Diagnostics Check	OK	17:12

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZIMMERMAN,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: CASSANELLI

DATE/TIME OF TEST: 04/02/2023 18:52:06

Diagnostics Check	OK	18:46
Air Blank	0.000	18:46
Control Test	0.080	18:46
Air Blank	0.000	18:47
Subject Sample # 1	0.000	18:47
Breath Volume	1.515	18:47
Air Blank	0.000	18:48
Air Blank	0.000	18:50
Subject Sample # 2	0.000	18:50
Breath Volume	2.191	18:50
Air Blank	0.000	18:51
Control Test	0.080	18:51
Air Blank	0.000	18:52
Diagnostics Check	OK	18:52

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CERNUTO,ELISHA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: BUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: VILLEGAS

DATE/TIME OF TEST: 04/02/2023 19:41:33

Diagnostics Check	OK	19:34
Air Blank	0.000	19:34
Control Test	0.080	19:35
Air Blank	0.000	19:35
Subject Sample # 1	0.154	19:36
Breath Volume	2.621	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.152	19:39
Breath Volume	2.351	19:39
Air Blank	0.000	19:40
Control Test	0.079	19:40
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: OLIVA

DATE/TIME OF TEST: 04/02/2023 20:29:53

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.248	20:24
Breath Volume	1.203	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:26
Subject Sample # 2	0.253	20:28
Breath Volume	2.539	20:28
Air Blank	0.000	20:28
Control Test	0.078	20:29
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: ZELAYANDIA HERNANDEZ

DATE/TIME OF TEST: 04/02/2023 22:42:55

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.209	22:38
Breath Volume	3.304	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.209	22:41
Breath Volume	3.386	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 04/14/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HABER,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: DIAZ ESCOBAR

DATE/TIME OF TEST: 04/03/2023 01:37:32

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.193	01:33
Breath Volume	1.171	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.196	01:35
Breath Volume	1.847	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: SHANGOLD

DATE/TIME OF TEST: 04/03/2023 19:58:25

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.080	19:52
Air Blank	0.000	19:52
Subject Sample # 1	0.196	19:53
Breath Volume	2.722	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.200	19:56
Breath Volume	2.753	19:56
Air Blank	0.000	19:57
Control Test	0.080	19:57
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: ALMAZAN

DATE/TIME OF TEST: 04/03/2023 22:26:08

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.080	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.199	22:21
Breath Volume	1.863	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.203	22:24
Breath Volume	3.035	22:24
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: CERBONE

DATE/TIME OF TEST: 04/04/2023 22:45:10

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.021	22:40
Breath Volume	1.187	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.020	22:43
Breath Volume	1.250	22:43
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO,LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: MOISE

DATE/TIME OF TEST: 04/05/2023 00:47:55

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.253	00:43
Breath Volume	1.496	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.248	00:46
Breath Volume	1.316	00:46
Air Blank	0.000	00:47
Control Test	0.078	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MCLAMORE

DATE/TIME OF TEST: 04/06/2023 02:37:09

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.181	02:31
Breath Volume	2.152	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.182	02:35
Breath Volume	2.257	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TESSMER, SCOTT
ARRESTING OFFICER AGENCY: Atlantis PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: BINGHAM

DATE/TIME OF TEST: 04/06/2023 10:52:31

Diagnostics Check	OK	10:46
Air Blank	0.000	10:46
Control Test	0.081	10:47
Air Blank	0.000	10:47
Subject Sample # 1	0.000	10:48
Breath Volume	2.082	10:48
Air Blank	0.000	10:48
Air Blank	0.000	10:50
Subject Sample # 2	0.000	10:51
Breath Volume	2.101	10:51
Air Blank	0.000	10:51
Control Test	0.080	10:51
Air Blank	0.000	10:52
Diagnostics Check	OK	10:52

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIER, MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 04/07/2023 02:15:51

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.186	02:11
Breath Volume	1.109	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.183	02:14
Breath Volume	1.261	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: ISOFF

DATE/TIME OF TEST: 04/07/2023 23:00:59

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.308	22:52
Breath Volume	2.277	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.287	22:56
Breath Volume	1.578	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:58
Subject Sample # 3	0.301	22:59
Breath Volume	3.562	22:59
Air Blank	0.000	23:00
Control Test	0.078	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/08/2023 04:04:34

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.080	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.062	04:00
Breath Volume	2.449	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.061	04:03
Breath Volume	3.191	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: ROMES

DATE/TIME OF TEST: 04/08/2023 05:14:58

Diagnostics Check	OK	05:07
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.208	05:10
Breath Volume	2.503	05:10
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.207	05:13
Breath Volume	2.855	05:13
Air Blank	0.000	05:14
Control Test	0.080	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SIMMONS,TIMOTHY
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: MANASSE

DATE/TIME OF TEST: 04/08/2023 06:34:10

Diagnostics Check	OK	06:27
Air Blank	0.000	06:28
Control Test	0.080	06:28
Air Blank	0.000	06:29
Subject Sample # 1	0.000	06:29
Breath Volume	1.519	06:29
Air Blank	0.000	06:30
Air Blank	0.000	06:32
Subject Sample # 2	0.000	06:32
Breath Volume	1.183	06:32
Air Blank	0.000	06:33
Control Test	0.080	06:33
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DIEZ,DARRELL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:49

SUBJECT'S LAST NAME: TARVER

DATE/TIME OF TEST: 04/08/2023 08:30:35

Diagnostics Check	OK	08:23
Air Blank	0.000	08:24
Control Test	0.081	08:24
Air Blank	0.000	08:25
Subject Sample # 1	0.000	08:25
Breath Volume	1.789	08:25
Air Blank	0.000	08:26
Air Blank	0.000	08:28
Subject Sample # 2	0.000	08:29
Breath Volume	1.519	08:29
Air Blank	0.000	08:29
Control Test	0.081	08:29
Air Blank	0.000	08:30
Diagnostics Check	OK	08:30

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/08/2023 21:51:36

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:46
Subject Sample # 1	0.158	21:46
Breath Volume	3.367	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.159	21:50
Breath Volume	3.710	21:50
Air Blank	0.000	21:50
Control Test	0.079	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO, LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/08/2023 23:02:05

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.218	22:57
Breath Volume	2.019	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.219	23:00
Breath Volume	2.328	23:00
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO,LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: WAYT

DATE/TIME OF TEST: 04/09/2023 19:43:24

Diagnostics Check	OK	19:34	
Air Blank	0.000	19:35	
Control Test	0.080	19:35	
Air Blank	0.000	19:36	
Subject Sample # 1	0.227	19:38	
Breath Volume	1.148	19:38	
Air Blank	0.000	19:39	
Air Blank	0.000	19:40	
Subject Sample # 2		19:42	Subject Test Refused
Breath Volume	0.000	19:42	
Air Blank	0.000	19:42	
Control Test	0.077	19:42	
Air Blank	0.000	19:43	
Diagnostics Check	OK	19:43	

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: PEREZ SANCHEZ

DATE/TIME OF TEST: 04/09/2023 23:10:44

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.182	23:05
Breath Volume	1.937	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.177	23:09
Breath Volume	2.304	23:09
Air Blank	0.000	23:09
Control Test	0.079	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: VO

DATE/TIME OF TEST: 04/10/2023 00:40:17

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.148	00:35
Breath Volume	1.710	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:37
Subject Sample # 2	0.150	00:38
Breath Volume	1.917	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 04/11/2023 01:21:28

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.255	01:16
Breath Volume	2.253	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.263	01:19
Breath Volume	2.289	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACHATE,CHRIS
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 03/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: HEAVENER

DATE/TIME OF TEST: 04/14/2023 05:37:25

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.080	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.054	05:33
Breath Volume	2.175	05:33
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.052	05:35
Breath Volume	1.437	05:35
Air Blank	0.000	05:36
Control Test	0.080	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 04/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CASEY

DATE/TIME OF TEST: 04/14/2023 20:33:53

Diagnostics Check	OK	20:26
Air Blank	0.000	20:27
Control Test	0.081	20:27
Air Blank	0.000	20:28
Subject Sample # 1	0.007	20:29
Breath Volume	1.964	20:29
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.000	20:32
Breath Volume	1.472	20:32
Air Blank	0.000	20:32
Control Test	0.077	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: REMISSAINTHE

DATE/TIME OF TEST: 04/15/2023 02:21:08

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.080	02:16
Breath Volume	1.343	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.081	02:19
Breath Volume	1.785	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOMAS ANTONIO,ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: CHAJ CHAJ

DATE/TIME OF TEST: 04/15/2023 04:35:19

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.152	04:29
Breath Volume	1.683	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.140	04:33
Breath Volume	2.097	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/16/2023 01:16:35

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.080	01:05	
Air Blank	0.000	01:05	
Subject Sample	# 1 0.210	01:08	Volume Not Met
Breath Volume	0.343	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample	# 2 0.221	01:11	
Breath Volume	1.718	01:11	
Air Blank	0.000	01:12	
Air Blank	0.000	01:14	
Subject Sample	# 3 0.207	01:14	
Breath Volume	1.148	01:14	
Air Blank	0.000	01:15	
Control Test	0.078	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: VILLAREAL

DATE/TIME OF TEST: 04/17/2023 23:59:19

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:48	
Control Test	0.080	23:48	
Air Blank	0.000	23:49	
Subject Sample	# 1 0.140	23:52	Volume Not Met
Breath Volume	0.050	23:52	
Air Blank	0.000	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 2	23:57	No Sample Provided
Breath Volume	0.000	23:57	
Air Blank	0.000	23:58	
Control Test	0.078	23:58	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: VILLAREAL

DATE/TIME OF TEST: 04/18/2023 00:11:27

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.142	00:06
Breath Volume	1.167	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.153	00:09
Breath Volume	1.378	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: PEREZ MENDEZ

DATE/TIME OF TEST: 04/22/2023 00:20:45

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.081	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.221	00:13
Breath Volume	1.171	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.245	00:16
Breath Volume	1.414	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 3	0.243	00:19
Breath Volume	1.218	00:19
Air Blank	0.000	00:19
Control Test	0.079	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEPEINE,KENS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: WASSERMAN

DATE/TIME OF TEST: 04/22/2023 06:09:58

Diagnostics Check	OK	06:01
Air Blank	0.000	06:02
Control Test	0.080	06:02
Air Blank	0.000	06:03
Subject Sample # 1	0.106	06:04
Breath Volume	1.789	06:04
Air Blank	0.000	06:05
Air Blank	0.000	06:07
Subject Sample # 2	0.105	06:08
Breath Volume	1.640	06:08
Air Blank	0.000	06:09
Control Test	0.079	06:09
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,JUAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: COCHE CHOJPEN

DATE/TIME OF TEST: 04/22/2023 22:42:46

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.223	22:38
Breath Volume	2.011	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.230	22:41
Breath Volume	2.527	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEGRAFFENREID,JADA
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: HAY

DATE/TIME OF TEST: 04/23/2023 02:39:27

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:35	No Sample Provided
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Air Blank	0.000	02:37	
Subject Sample	# 2	02:38	Subject Test Refused
Breath Volume	0.000	02:38	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NEUS,KEVIN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: ELISIER

DATE/TIME OF TEST: 04/23/2023 05:34:22

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.081	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.086	05:29
Breath Volume	1.593	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.085	05:32
Breath Volume	2.597	05:32
Air Blank	0.000	05:33
Control Test	0.080	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BYRNE,CHRIS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: BYLES

DATE/TIME OF TEST: 04/24/2023 01:32:33

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.176	01:28
Breath Volume	1.593	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.186	01:30
Breath Volume	1.621	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GRAU

DATE/TIME OF TEST: 04/26/2023 01:55:04

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.243	01:46
Breath Volume	1.320	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.210	01:50
Breath Volume	1.109	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 3	0.250	01:53
Breath Volume	2.074	01:53
Air Blank	0.000	01:54
Control Test	0.078	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:55

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: GASPAR MAURICIO

DATE/TIME OF TEST: 04/26/2023 23:00:58

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.222	22:56
Breath Volume	2.597	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.223	22:59
Breath Volume	2.195	22:59
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HILL, JEREMY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:11

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 04/27/2023 09:47:58

Diagnostics Check	OK	09:41
Air Blank	0.000	09:42
Control Test	0.080	09:42
Air Blank	0.000	09:43
Subject Sample # 1	0.140	09:43
Breath Volume	2.289	09:43
Air Blank	0.000	09:44
Air Blank	0.000	09:45
Subject Sample # 2	0.138	09:46
Breath Volume	3.187	09:46
Air Blank	0.000	09:47
Control Test	0.079	09:47
Air Blank	0.000	09:47
Diagnostics Check	OK	09:47

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: AGUILAR ALARCON

DATE/TIME OF TEST: 04/28/2023 01:43:36

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.000	01:39
Breath Volume	1.800	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.000	01:42
Breath Volume	2.550	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLACKMAN,VICTOR
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LOPEZ GARCIA

DATE/TIME OF TEST: 04/28/2023 02:15:26

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.195	02:10
Breath Volume	1.300	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.199	02:13
Breath Volume	2.347	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/28/2023 21:41:16

Diagnostics Check	OK	21:34
Air Blank	0.000	21:35
Control Test	0.081	21:35
Air Blank	0.000	21:35
Subject Sample # 1	0.210	21:36
Breath Volume	1.843	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.222	21:39
Breath Volume	2.613	21:39
Air Blank	0.000	21:40
Control Test	0.080	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 04/28/2023 22:08:24

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.080	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.169	22:03
Breath Volume	2.996	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.169	22:06
Breath Volume	3.433	22:06
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: WHALEN

DATE/TIME OF TEST: 04/29/2023 03:43:44

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.292	03:36
Breath Volume	1.476	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.314	03:39
Breath Volume	1.894	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 3	0.312	03:42
Breath Volume	1.734	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: DELUNA

DATE/TIME OF TEST: 04/29/2023 20:40:29

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.080	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.101	20:36
Breath Volume	3.226	20:36
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 2	0.100	20:38
Breath Volume	2.707	20:38
Air Blank	0.000	20:39
Control Test	0.080	20:39
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BYRNE, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: MARTIN MARTIN

DATE/TIME OF TEST: 04/29/2023 23:20:43

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.156	23:16
Breath Volume	2.484	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.152	23:19
Breath Volume	1.304	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: PARK

DATE/TIME OF TEST: 04/30/2023 05:50:34

Diagnostics Check	OK	05:44	
Air Blank	0.000	05:44	
Control Test	0.080	05:44	
Air Blank	0.000	05:45	
Subject Sample	# 1 0.204	05:45	
Breath Volume	1.406	05:45	
Air Blank	0.000	05:46	
Air Blank	0.000	05:48	
Subject Sample	# 2	05:48	Slope Not Met
Breath Volume	1.703	05:48	
Air Blank	0.000	05:49	
Control Test	0.080	05:49	
Air Blank	0.000	05:50	
Diagnostics Check	OK	05:50	

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: PARK

DATE/TIME OF TEST: 04/30/2023 06:18:58

Diagnostics Check	OK	06:12
Air Blank	0.000	06:13
Control Test	0.080	06:13
Air Blank	0.000	06:13
Subject Sample # 1	0.209	06:14
Breath Volume	1.820	06:14
Air Blank	0.000	06:15
Air Blank	0.000	06:16
Subject Sample # 2	0.202	06:17
Breath Volume	1.875	06:17
Air Blank	0.000	06:18
Control Test	0.079	06:18
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 05/19/2023 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: ALLGOR

DATE/TIME OF TEST: 04/12/2023 03:21:43

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.013	03:17
Breath Volume	2.664	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.012	03:20
Breath Volume	3.289	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 04/25/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 04/21/2023 00:40:00

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:39	
Control Test	0.000	00:39	RFI Detect
Air Blank	0.000	00:40	

Data Download Date/Time: 04/25/2023 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 04/21/2023 00:47:57

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.010	00:43
Breath Volume	2.265	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.011	00:46
Breath Volume	2.175	00:46
Air Blank	0.000	00:47
Control Test	0.082	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 04/25/2023 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: DARBY

DATE/TIME OF TEST: 04/24/2023 23:11:17

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.054	23:06
Breath Volume	1.437	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.069	23:09
Breath Volume	2.019	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 04/25/2023 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: CZEBOTAR

DATE/TIME OF TEST: 04/28/2023 23:25:51

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:22	
Control Test	0.079	23:23	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:24	Subject Test Refused
Breath Volume	0.000	23:24	
Air Blank	0.000	23:24	
Control Test	0.080	23:25	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 05/17/2023 15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 04/30/2023 01:19:11

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.133	01:14
Breath Volume	1.222	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.141	01:17
Breath Volume	2.609	01:17
Air Blank	0.000	01:18
Control Test	0.077	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 05/17/2023 15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTZER,ETHAN,E
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SMITH,NICKOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: SCHAAB

DATE/TIME OF TEST: 04/09/2023 22:01:29

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:59	
Control Test	0.082	21:59	
Air Blank	0.000	21:59	
Subject Sample	# 1	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.080	22:00	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 04/21/2023 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAASTRA,WILLIAM,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KAASTRA,WILLIAM
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ROCHELLE

DATE/TIME OF TEST: 04/03/2023 01:33:21

Diagnostics Check	OK	01:32	
Air Blank		01:32	RFI Detect
Air Blank	0.000	01:33	

Data Download Date/Time: 04/24/2023 21:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAASTRA, WILLIAM, S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KAASTRA, WILLIAM
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ROCHELLE

DATE/TIME OF TEST: 04/03/2023 01:39:46

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.079	01:36	
Air Blank	0.000	01:36	
Subject Sample # 1	0.187	01:38	
Breath Volume	2.957	01:38	
Air Blank		01:39	Purge Fail
Air Blank	0.000	01:39	

Data Download Date/Time: 04/24/2023 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAASTRA, WILLIAM, S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KAASTRA, WILLIAM
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: R240013836830

DATE/TIME OF TEST: 04/03/2023 01:53:35

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.076	01:44	
Air Blank	0.000	01:44	
Subject Sample # 1	0.167	01:45	
Breath Volume	1.570	01:45	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample # 2	0.192	01:48	
Breath Volume	1.230	01:48	
Air Blank	0.000	01:49	
Air Blank	0.000	01:50	
Subject Sample # 3	0.183	01:51	
Breath Volume	2.531	01:51	
Air Blank	0.000	01:52	
Control Test	0.073	01:52	Control Outside Tolerance
Air Blank	0.000	01:53	

Data Download Date/Time: 04/24/2023 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: ARANDA MUNOZ

DATE/TIME OF TEST: 04/04/2023 00:11:41

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.078	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.229	00:06
Breath Volume	2.179	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.223	00:09
Breath Volume	1.796	00:09
Air Blank	0.000	00:10
Control Test	0.077	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 04/24/2023 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,DEBRA,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WALSH,KASEY
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: RAFFERTY

DATE/TIME OF TEST: 04/04/2023 15:18:00

Diagnostics Check	OK	15:15	
Air Blank	0.000	15:15	
Control Test	0.077	15:16	
Air Blank	0.000	15:16	
Subject Sample	# 1	15:17	Sequence Aborted
Breath Volume	0.000	15:17	
Air Blank	0.000	15:18	

Data Download Date/Time: 04/24/2023 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMSON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: GOMEZ HERNANDEZ

DATE/TIME OF TEST: 04/10/2023 15:56:31

Diagnostics Check	OK	15:49
Air Blank	0.000	15:49
Control Test	0.080	15:50
Air Blank	0.000	15:50
Subject Sample # 1	0.206	15:51
Breath Volume	1.335	15:51
Air Blank	0.000	15:52
Air Blank	0.000	15:54
Subject Sample # 2	0.204	15:54
Breath Volume	1.695	15:54
Air Blank	0.000	15:55
Control Test	0.078	15:55
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

Data Download Date/Time: 05/02/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: SOUZA

DATE/TIME OF TEST: 04/11/2023 23:14:01

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:10	
Control Test	0.080	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Control Test	0.080	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 05/02/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: PENAGOS

DATE/TIME OF TEST: 04/12/2023 04:16:51

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.214	04:12
Breath Volume	2.195	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.205	04:14
Breath Volume	1.722	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 05/02/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAMILTON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:26

SUBJECT'S LAST NAME: MCGEORGE

DATE/TIME OF TEST: 04/12/2023 10:52:50

Diagnostics Check	OK	10:46
Air Blank	0.000	10:46
Control Test	0.079	10:47
Air Blank	0.000	10:47
Subject Sample # 1	0.000	10:48
Breath Volume	2.488	10:48
Air Blank	0.000	10:48
Air Blank	0.000	10:50
Subject Sample # 2	0.000	10:51
Breath Volume	1.664	10:51
Air Blank	0.000	10:51
Control Test	0.078	10:52
Air Blank	0.000	10:52
Diagnostics Check	OK	10:52

Data Download Date/Time: 05/02/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 04/13/2023 04:26:14

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.180	04:18
Breath Volume	2.187	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.156	04:21
Breath Volume	1.769	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 3	0.157	04:24
Breath Volume	1.738	04:24
Air Blank	0.000	04:25
Control Test	0.078	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 05/02/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 04/13/2023 21:50:56

Diagnostics Check	OK	21:43
Air Blank	0.000	21:44
Control Test	0.079	21:44
Air Blank	0.000	21:44
Subject Sample # 1	0.224	21:46
Breath Volume	1.855	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:48
Subject Sample # 2	0.233	21:49
Breath Volume	1.855	21:49
Air Blank	0.000	21:49
Control Test	0.077	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 05/02/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MOGAVERO

DATE/TIME OF TEST: 04/14/2023 00:57:49

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.079	00:50	
Air Blank	0.000	00:51	
Subject Sample	# 1 0.144	00:52	
Breath Volume	1.449	00:52	
Air Blank	0.000	00:53	
Air Blank	0.000	00:54	
Subject Sample	# 2 0.161	00:55	
Breath Volume	1.777	00:55	
Air Blank	0.000	00:56	
Control Test	0.079	00:56	
Air Blank	0.000	00:57	
Diagnostics Check	Fail	00:57	Temperature Regulation Test
Air Blank	0.000	00:57	

Data Download Date/Time: 05/02/2023 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MOGAVERO

DATE/TIME OF TEST: 04/14/2023 01:09:24

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.079	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.152	01:04
Breath Volume	1.390	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 2	0.162	01:07
Breath Volume	2.164	01:07
Air Blank	0.000	01:08
Control Test	0.078	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 05/02/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 04/15/2023 04:49:24

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.079	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.184	04:42
Breath Volume	1.140	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 2	0.184	04:47
Breath Volume	1.164	04:47
Air Blank	0.000	04:48
Control Test	0.077	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 05/02/2023 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEYERS,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MANLEY

DATE/TIME OF TEST: 04/15/2023 05:26:16

Diagnostics Check	OK	05:16	
Air Blank	0.000	05:16	
Control Test	0.077	05:16	
Air Blank	0.000	05:17	
Subject Sample	# 1 0.187	05:20	Volume Not Met
Breath Volume	0.070	05:20	
Air Blank	0.000	05:21	
Air Blank	0.000	05:22	
Subject Sample	# 2	05:24	Subject Test Refused
Breath Volume	0.046	05:24	
Air Blank	0.000	05:25	
Control Test	0.073	05:25	Control Outside Tolerance
Air Blank	0.000	05:26	

Data Download Date/Time: 05/02/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BUKOVZA

DATE/TIME OF TEST: 04/18/2023 02:11:07

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.081	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.037	02:06
Breath Volume	3.648	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample # 2	0.037	02:09
Breath Volume	3.093	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 05/02/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: PETTIJOHN

DATE/TIME OF TEST: 04/18/2023 02:50:57

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.162	02:46
Breath Volume	1.246	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.168	02:49
Breath Volume	1.882	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 05/02/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: MOYE

DATE/TIME OF TEST: 04/19/2023 04:10:11

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.081	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.177	04:05
Breath Volume	2.929	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:07
Subject Sample # 2	0.179	04:08
Breath Volume	2.109	04:08
Air Blank	0.000	04:09
Control Test	0.077	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 05/02/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 04/19/2023 22:20:23

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:17	
Control Test	0.081	22:17	
Air Blank	0.000	22:18	
Subject Sample	# 1 0.333	22:19	
Breath Volume	2.593	22:19	
Air Blank		22:19	Purge Fail
Air Blank	0.000	22:20	

Data Download Date/Time: 05/02/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 04/19/2023 22:30:22

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:23	
Control Test	0.075	22:24	
Air Blank	0.000	22:24	
Subject Sample # 1	0.327	22:25	
Breath Volume	3.218	22:25	
Air Blank	0.000	22:26	
Air Blank	0.000	22:27	
Subject Sample # 2	0.323	22:28	
Breath Volume	2.253	22:28	
Air Blank	0.000	22:29	
Control Test	0.074	22:29	Control Outside Tolerance
Air Blank	0.000	22:30	

Data Download Date/Time: 05/02/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NONE,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: SYS CHECK
ARREST TIME: 07:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: NONE

DATE/TIME OF TEST: 04/26/2023 08:17:44

Diagnostics Check	OK	08:11
Air Blank	0.000	08:12
Control Test	0.080	08:12
Air Blank	0.000	08:13
Subject Sample # 1	0.000	08:13
Breath Volume	2.363	08:13
Air Blank	0.000	08:14
Air Blank	0.000	08:16
Subject Sample # 2	0.000	08:16
Breath Volume	1.914	08:16
Air Blank	0.000	08:16
Control Test	0.078	08:17
Air Blank	0.000	08:17
Diagnostics Check	OK	08:17

Data Download Date/Time: 05/02/2023 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, LAURA,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HERNANDEZ, LAURA
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 04/06/2023 17:12:25

Diagnostics Check	OK	17:08	
Air Blank	0.000	17:09	
Control Test	0.080	17:09	
Air Blank	0.000	17:09	
Subject Sample	# 1	17:11	Sequence Aborted
Breath Volume	0.000	17:11	
Air Blank	0.000	17:12	

Data Download Date/Time: 04/14/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, LAURA,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HERNANDEZ, LAURA
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 04/06/2023 17:20:43

Diagnostics Check	OK	17:15	
Air Blank	0.000	17:16	
Control Test	0.081	17:16	
Air Blank	0.000	17:17	
Subject Sample	# 1	17:19	Subject Test Refused
Breath Volume	0.000	17:19	
Air Blank	0.000	17:19	
Control Test	0.081	17:20	
Air Blank	0.000	17:20	
Diagnostics Check	OK	17:20	

Data Download Date/Time: 04/14/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AFONSO,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GOLDBERG,JUSTIN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 04/09/2023 04:38:48

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.114	04:33
Breath Volume	3.824	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.121	04:37
Breath Volume	3.835	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 04/14/2023 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,GARY,B
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GOLDBERG,JUSTIN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 04/12/2023 22:28:05

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:24	
Control Test	0.081	22:25	
Air Blank	0.000	22:25	
Subject Sample	# 1	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 04/14/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEIL,DARRIN,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: MCNEIL,DARRIN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/17/2023 01:33:58

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.080	01:20	
Air Blank	0.000	01:21	
Subject Sample # 1	0.142	01:23	
Breath Volume	3.597	01:23	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample # 2	0.103	01:28	Volume Not Met
Breath Volume	0.019	01:28	
Air Blank	0.000	01:29	
Air Blank	0.000	01:31	
Subject Sample # 3	0.137	01:32	
Breath Volume	1.988	01:32	
Air Blank	0.000	01:33	
Control Test	0.076	01:33	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 05/19/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AFONSO,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CASTRO,JUAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: ROBLERO

DATE/TIME OF TEST: 04/22/2023 19:51:47

Diagnostics Check	OK	19:44
Air Blank	0.000	19:44
Control Test	0.080	19:45
Air Blank	0.000	19:45
Subject Sample # 1	0.245	19:46
Breath Volume	2.375	19:46
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.244	19:49
Breath Volume	3.781	19:49
Air Blank	0.000	19:50
Control Test	0.080	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 05/19/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AFONSO,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: THIMOTHEE,CHARLES
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BUBAINE-TORRES

DATE/TIME OF TEST: 04/23/2023 04:44:55

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.119	04:40
Breath Volume	1.574	04:40
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.101	04:43
Breath Volume	1.652	04:43
Air Blank	0.000	04:43
Control Test	0.079	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 05/19/2023 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEIL,DARRIN,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: THIMOTHEE,CHARLES
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/30/2023 02:32:15

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank		02:31	Purge Fail
Air Blank		02:32	Purge Fail

Data Download Date/Time: 05/19/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEIL,DARRIN,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: THIMOTHEE,CHARLES
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/30/2023 02:37:45

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.069	02:37	Control Outside Tolerance
Air Blank	0.000	02:37	

Data Download Date/Time: 05/19/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLGOOD,JEFFREY,S
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ALLGOOD,JEFFREY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 03/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/06/2023 21:30:31

Diagnostics Check	OK	21:21
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.000	21:24
Breath Volume	1.269	21:24
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.000	21:29
Breath Volume	1.207	21:29
Air Blank	0.000	21:29
Control Test	0.078	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 04/12/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER, MONICA, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MILLENDER, MONICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:13

SUBJECT'S LAST NAME: CONLIN

DATE/TIME OF TEST: 04/07/2023 12:39:47

Diagnostics Check	OK	12:33
Air Blank	0.000	12:33
Control Test	0.080	12:34
Air Blank	0.000	12:34
Subject Sample # 1	0.000	12:35
Breath Volume	1.976	12:35
Air Blank	0.000	12:36
Air Blank	0.000	12:38
Subject Sample # 2	0.000	12:38
Breath Volume	2.183	12:38
Air Blank	0.000	12:38
Control Test	0.080	12:39
Air Blank	0.000	12:39
Diagnostics Check	OK	12:39

Data Download Date/Time: 04/12/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/14/2023 03:16:39

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.081	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.123	03:12
Breath Volume	2.792	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.118	03:15
Breath Volume	2.218	03:15
Air Blank	0.000	03:15
Control Test	0.078	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 05/09/2023 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: ERWIN

DATE/TIME OF TEST: 04/15/2023 17:17:21

Diagnostics Check	OK	17:14	
Air Blank	0.000	17:14	
Control Test	0.081	17:14	
Air Blank	0.000	17:15	
Subject Sample	# 1	17:15	Subject Test Refused
Breath Volume	0.000	17:15	
Air Blank	0.000	17:16	
Control Test	0.081	17:16	
Air Blank	0.000	17:17	
Diagnostics Check	OK	17:17	

Data Download Date/Time: 05/09/2023 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: ERWIN

DATE/TIME OF TEST: 04/15/2023 17:33:05

Diagnostics Check	OK	17:25
Air Blank	0.000	17:26
Control Test	0.080	17:26
Air Blank	0.000	17:27
Subject Sample # 1	0.171	17:28
Breath Volume	1.281	17:28
Air Blank	0.000	17:29
Air Blank	0.000	17:31
Subject Sample # 2	0.179	17:31
Breath Volume	1.359	17:31
Air Blank	0.000	17:32
Control Test	0.078	17:32
Air Blank	0.000	17:33
Diagnostics Check	OK	17:33

Data Download Date/Time: 05/09/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA, ADALBERTO, L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA, ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: TREVINO

DATE/TIME OF TEST: 04/17/2023 18:18:41

Diagnostics Check	OK	18:10	
Air Blank	0.000	18:11	
Control Test	0.083	18:11	
Air Blank	0.000	18:11	
Subject Sample # 1	0.000	18:12	
Breath Volume	1.875	18:12	
Air Blank	0.000	18:13	
Air Blank	0.000	18:15	
Subject Sample # 2	0.000	18:17	
Breath Volume	1.722	18:17	
Air Blank	0.000	18:17	
Control Test	0.102	18:17	Interferent Detect
Air Blank		18:18	Purge Fail

Data Download Date/Time: 05/09/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSTICE,JOSEPH,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JUSTICE,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/20/2023 01:29:01

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.082	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1 0.068	01:25	Volume Not Met
Breath Volume	0.144	01:25	
Air Blank	0.000	01:26	
Air Blank	0.000	01:28	
Subject Sample	# 2	01:28	Sequence Aborted
Breath Volume	0.000	01:28	
Air Blank	0.000	01:29	

Data Download Date/Time: 05/09/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,DAVID,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MUNDT,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: TANK

DATE/TIME OF TEST: 04/21/2023 03:52:33

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.157	03:47
Breath Volume	1.398	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.141	03:50
Breath Volume	1.402	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 05/09/2023 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSTICE,JOSEPH,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JUSTICE,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: KOHLENBERG

DATE/TIME OF TEST: 04/22/2023 02:16:28

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.073	02:11
Breath Volume	2.648	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.071	02:14
Breath Volume	3.128	02:14
Air Blank	0.000	02:15
Control Test	0.078	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 05/09/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANNING,CARL,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MANNING,CARL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 11:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: TORRES MONTANEZ

DATE/TIME OF TEST: 04/24/2023 23:34:25

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:29	
Control Test	0.081	23:29	
Air Blank	0.000	23:30	
Subject Sample	# 1 0.136	23:30	
Breath Volume	3.402	23:30	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample	# 2 0.137	23:33	
Breath Volume	2.753	23:33	
Air Blank		23:33	RFI Detect
Air Blank	0.000	23:34	

Data Download Date/Time: 05/09/2023 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 04/28/2023 05:16:36

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.081	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.156	05:07
Breath Volume	2.007	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:09
Subject Sample # 2	0.135	05:10
Breath Volume	1.121	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 3	0.142	05:14
Breath Volume	2.003	05:14
Air Blank	0.000	05:15
Control Test	0.078	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 05/09/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 09:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/28/2023 10:41:45

Diagnostics Check	OK	10:35
Air Blank	0.000	10:35
Control Test	0.081	10:36
Air Blank	0.000	10:36
Subject Sample # 1	0.000	10:37
Breath Volume	2.917	10:37
Air Blank	0.000	10:37
Air Blank	0.000	10:39
Subject Sample # 2	0.000	10:40
Breath Volume	3.140	10:40
Air Blank	0.000	10:40
Control Test	0.078	10:41
Air Blank	0.000	10:41
Diagnostics Check	OK	10:41

Data Download Date/Time: 05/09/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,S
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 04/01/2023 22:18:12

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.227	22:13
Breath Volume	3.753	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.220	22:16
Breath Volume	3.742	22:16
Air Blank	0.000	22:17
Control Test	0.078	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 04/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA, ADALBERTO, L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA, ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 13:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: MCCOLLEY

DATE/TIME OF TEST: 04/04/2023 14:22:28

Diagnostics Check	OK	14:17	
Air Blank	0.000	14:18	
Control Test	0.078	14:18	
Air Blank	0.000	14:18	
Subject Sample	# 1	14:21	Subject Test Refused
Breath Volume	0.000	14:21	
Air Blank	0.000	14:21	
Control Test	0.078	14:21	
Air Blank	0.000	14:22	
Diagnostics Check	OK	14:22	

Data Download Date/Time: 04/12/2023 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: EDGAR

DATE/TIME OF TEST: 04/10/2023 18:11:37

Diagnostics Check	OK	18:05
Air Blank	0.000	18:05
Control Test	0.078	18:05
Air Blank	0.000	18:06
Subject Sample # 1	0.218	18:06
Breath Volume	4.109	18:06
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.218	18:10
Breath Volume	2.937	18:10
Air Blank	0.000	18:10
Control Test	0.075	18:11
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 04/12/2023 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 04/01/2023 04:00:34

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.080	03:58	
Air Blank	0.000	03:58	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.080	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 04/13/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 04/03/2023 02:34:04

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.019	02:28
Breath Volume	1.230	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.018	02:32
Breath Volume	1.597	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:34

Data Download Date/Time: 04/13/2023 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BAEZ-CAMACHO

DATE/TIME OF TEST: 04/14/2023 04:30:08

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.078	04:28	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:28	Subject Test Refused
Breath Volume	0.000	04:28	
Air Blank	0.000	04:29	
Control Test	0.080	04:29	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 05/31/2023 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,HOPE,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: JONES,HOPE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 10:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: MORRA

DATE/TIME OF TEST: 04/15/2023 12:28:01

Diagnostics Check	OK	12:20
Air Blank	0.000	12:20
Control Test	0.080	12:20
Air Blank	0.000	12:21
Subject Sample # 1	0.000	12:21
Breath Volume	1.539	12:21
Air Blank	0.000	12:22
Air Blank	0.000	12:24
Subject Sample # 2	0.000	12:26
Breath Volume	1.972	12:26
Air Blank	0.000	12:27
Control Test	0.078	12:27
Air Blank	0.000	12:27
Diagnostics Check	OK	12:28

Data Download Date/Time: 05/31/2023 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1497672
BREATH TEST OPERATOR: NOGUERA,CHRISTOPHER,A	DRY GAS EXPIRY DATE: 05/20/2024
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: LOPEZ,RYAN	ARREST TIME: 02:24
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Male
LAST AGENCY INSPECTION: 04/13/2023	OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: GARCIA OSORIO	DATE/TIME OF TEST: 04/17/2023 03:29:07
------------------------------------	--

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.000	03:24
Breath Volume	2.265	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.000	03:27
Breath Volume	1.562	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 05/31/2023 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SENMARTIN

DATE/TIME OF TEST: 04/20/2023 00:34:39

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:32	
Control Test	0.079	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:33	
Control Test	0.080	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 05/31/2023 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,ELSTON,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,WILLIAMS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: PEREZ LLANO

DATE/TIME OF TEST: 04/22/2023 00:50:22

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:38	
Control Test	0.079	00:38	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:42	No Sample Provided
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Air Blank	0.000	00:44	
Subject Sample	# 2	00:49	No Sample Provided
Breath Volume	0.000	00:49	
Air Blank	0.000	00:49	
Control Test	0.079	00:49	
Air Blank		00:50	RFI Detect
Air Blank		00:50	RFI Detect

Data Download Date/Time: 05/31/2023 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DANELIA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,FERNANDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 04/22/2023 19:31:00

Diagnostics Check	OK	19:20	
Air Blank	0.000	19:21	
Control Test	0.080	19:21	
Air Blank	0.000	19:21	
Subject Sample	# 1	19:24	No Sample Provided
Breath Volume	0.000	19:24	
Air Blank	0.000	19:25	
Air Blank	0.000	19:27	
Subject Sample	# 2	19:29	Subject Test Refused
Breath Volume	0.000	19:29	
Air Blank	0.000	19:30	
Control Test	0.080	19:30	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:31	

Data Download Date/Time: 05/31/2023 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MESA,ERIC
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PLANTE

DATE/TIME OF TEST: 04/23/2023 02:44:21

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.079	02:39	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.079	02:43	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 05/31/2023 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORINSTEIN,JOSEPH
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: GREENFIELD

DATE/TIME OF TEST: 04/01/2023 18:13:29

Diagnostics Check	OK	18:07
Air Blank	0.000	18:07
Control Test	0.078	18:08
Air Blank	0.000	18:08
Subject Sample # 1	0.000	18:09
Breath Volume	1.667	18:09
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.000	18:12
Breath Volume	1.937	18:12
Air Blank	0.000	18:12
Control Test	0.079	18:12
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 04/05/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ESPINOZA,O
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: PEREZ GARCIA

DATE/TIME OF TEST: 04/08/2023 05:25:14

Diagnostics Check	OK	05:18
Air Blank	0.000	05:19
Control Test	0.079	05:19
Air Blank	0.000	05:19
Subject Sample # 1	0.000	05:20
Breath Volume	1.371	05:20
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.000	05:23
Breath Volume	1.152	05:23
Air Blank	0.000	05:24
Control Test	0.079	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: DASH

DATE/TIME OF TEST: 04/11/2023 21:38:06

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.078	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.200	21:32
Breath Volume	1.746	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.186	21:36
Breath Volume	2.058	21:36
Air Blank	0.000	21:37
Control Test	0.079	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: JONES,DALLAS
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 10:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 04/18/2023 00:33:50

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:29	
Control Test	0.078	00:30	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:32	
Control Test	0.078	00:33	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,ANDRES,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RAMIREZ,ANDRES
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 04/20/2023 22:09:17

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:06	
Control Test	0.079	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 1	22:07	Subject Test Refused
Breath Volume	0.000	22:07	
Air Blank	0.000	22:08	
Control Test	0.079	22:08	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: FEREIRA,SALISHA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: VALENTIM

DATE/TIME OF TEST: 04/21/2023 03:35:50

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.078	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1 0.103	03:28	Volume Not Met
Breath Volume	0.511	03:28	
Air Blank	0.000	03:29	
Air Blank	0.000	03:30	
Subject Sample	# 2 0.087	03:34	Volume Not Met
Breath Volume	0.312	03:34	
Air Blank	0.000	03:34	
Control Test	0.078	03:35	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: FEREIRA,SALISHA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: VALENTIM

DATE/TIME OF TEST: 04/21/2023 04:07:11

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.088	04:01
Breath Volume	2.128	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.087	04:05
Breath Volume	1.375	04:05
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ONEILL,CONNER
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 04/22/2023 20:10:21

Diagnostics Check	OK	20:03	
Air Blank	0.000	20:04	
Control Test	0.078	20:04	
Air Blank	0.000	20:05	
Subject Sample	# 1 0.123	20:06	
Breath Volume	1.769	20:06	
Air Blank	0.000	20:07	
Air Blank	0.000	20:08	
Subject Sample	# 2	20:08	Subject Test Refused
Breath Volume	0.000	20:08	
Air Blank	0.000	20:09	
Control Test	0.077	20:09	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ONEILL,CONNER
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 04/22/2023 20:15:41

Diagnostics Check	OK	20:15	
Air Blank		20:15	RFI Detect
Air Blank	0.000	20:15	

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ONEILL,CONNER
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 04/22/2023 20:25:12

Diagnostics Check	OK	20:17
Air Blank	0.000	20:17
Control Test	0.078	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.120	20:20
Breath Volume	1.722	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:22
Subject Sample # 2	0.115	20:23
Breath Volume	1.734	20:23
Air Blank	0.000	20:24
Control Test	0.078	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONACO,ASHLEY,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MONACO,ASHLEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MIGOYA

DATE/TIME OF TEST: 04/21/2023 00:49:52

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:48	
Control Test	0.078	00:48	
Air Blank	0.000	00:49	
Subject Sample	# 1	00:49	Sequence Aborted
Breath Volume	0.000	00:49	
Air Blank	0.000	00:49	

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONACO,ASHLEY,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MONACO,ASHLEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MIGOYA

DATE/TIME OF TEST: 04/21/2023 00:55:39

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:53	
Control Test	0.079	00:53	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:55	Sequence Aborted
Breath Volume	0.000	00:55	
Air Blank	0.000	00:55	

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONACO,ASHLEY,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MONACO,ASHLEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MIGOYA

DATE/TIME OF TEST: 04/21/2023 01:12:46

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.118	01:08
Breath Volume	2.257	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.115	01:11
Breath Volume	1.378	01:11
Air Blank	0.000	01:11
Control Test	0.078	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 11:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:57

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/26/2023 12:20:43

Diagnostics Check	OK	12:17	
Air Blank	0.000	12:18	
Control Test	0.079	12:18	
Air Blank	0.000	12:18	
Subject Sample	# 1	12:19	Subject Test Refused
Breath Volume	0.000	12:19	
Air Blank	0.000	12:19	
Control Test	0.079	12:20	
Air Blank	0.000	12:20	
Diagnostics Check	OK	12:20	

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: O'NEILL,CONNER
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: FARRELL

DATE/TIME OF TEST: 04/26/2023 22:49:45

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:47	
Control Test	0.078	22:47	
Air Blank	0.000	22:48	
Subject Sample	# 1	22:48	Subject Test Refused
Breath Volume	0.000	22:48	
Air Blank	0.000	22:48	
Control Test	0.079	22:49	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HARRILL,ANDERSEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 04/29/2023 02:41:18

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.144	02:35
Breath Volume	1.816	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 2	0.151	02:39
Breath Volume	1.484	02:39
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANEY,MICHAEL,P
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: BARBER,EDWARD
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: BUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: WOLFE

DATE/TIME OF TEST: 04/29/2023 18:27:06

Diagnostics Check	OK	18:20
Air Blank	0.000	18:21
Control Test	0.080	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.000	18:22
Breath Volume	1.332	18:22
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.000	18:25
Breath Volume	1.714	18:25
Air Blank	0.000	18:26
Control Test	0.078	18:26
Air Blank	0.000	18:27
Diagnostics Check	OK	18:27

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 04/30/2023 23:28:27

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.078	23:22
Air Blank	0.000	23:22
Subject Sample # 1	0.148	23:23
Breath Volume	2.089	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.147	23:26
Breath Volume	1.648	23:26
Air Blank	0.000	23:27
Control Test	0.078	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 05/04/2023 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: ZULAIMA SAINZ

DATE/TIME OF TEST: 04/23/2023 22:46:57

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:44	
Control Test	0.081	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:45	
Control Test	0.081	22:46	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 04/25/2023 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/01/2023 23:06:45

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.079	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.227	23:01
Breath Volume	2.273	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:03
Subject Sample # 2	0.218	23:05
Breath Volume	1.824	23:05
Air Blank	0.000	23:05
Control Test	0.078	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 04/01/2023 23:47:06

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.164	23:42
Breath Volume	1.601	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.168	23:45
Breath Volume	2.605	23:45
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: FERNANDES CAMPOS

DATE/TIME OF TEST: 04/05/2023 01:47:32

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.078	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1 0.225	01:40	Volume Not Met
Breath Volume	0.050	01:40	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample	# 2	01:46	No Sample Provided
Breath Volume	0.000	01:46	
Air Blank	0.000	01:46	
Control Test	0.076	01:46	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: FERNANDES CAMPOS

DATE/TIME OF TEST: 04/05/2023 02:14:42

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:05	
Control Test	0.078	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1 0.263	02:09	Volume Not Met
Breath Volume	1.054	02:09	
Air Blank	0.000	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 2	02:12	RFI Detect
Breath Volume	0.511	02:12	
Air Blank	0.000	02:13	
Control Test	0.074	02:14	Control Outside Tolerance
Air Blank	0.000	02:14	

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: FERNANDES CAMPOS

DATE/TIME OF TEST: 04/05/2023 02:31:48

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.076	02:18	
Air Blank	0.000	02:18	
Subject Sample	# 1 0.224	02:21	Volume Not Met
Breath Volume	0.160	02:21	
Air Blank	0.000	02:22	
Air Blank	0.000	02:24	
Subject Sample	# 2 0.223	02:26	
Breath Volume	1.320	02:26	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample	# 3 0.227	02:29	
Breath Volume	1.218	02:29	
Air Blank	0.000	02:30	
Control Test	0.073	02:31	Control Outside Tolerance
Air Blank	0.000	02:31	

Data Download Date/Time: 04/12/2023 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALSHAIER,ELIAS
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/15/2023 00:18:12

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:04	
Control Test	0.079	00:04	
Air Blank	0.000	00:05	
Subject Sample # 1	0.092	00:05	
Breath Volume	1.632	00:05	
Air Blank	0.000	00:06	
Air Blank	0.000	00:08	
Subject Sample # 2		00:11	No Sample Provided
Breath Volume	0.000	00:11	
Air Blank	0.000	00:11	
Air Blank	0.000	00:13	
Subject Sample # 3		00:16	No Sample Provided
Breath Volume	0.000	00:16	
Air Blank	0.000	00:17	
Control Test	0.079	00:17	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 05/03/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BENINCASA,VINCENT
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 04/15/2023 04:39:57

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.131	04:34
Breath Volume	1.628	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:36
Subject Sample # 2	0.134	04:38
Breath Volume	1.730	04:38
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 05/03/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VALDES,YANELIS
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: VILLOTA

DATE/TIME OF TEST: 04/15/2023 04:58:43

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.078	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.105	04:54
Breath Volume	2.359	04:54
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.109	04:57
Breath Volume	2.078	04:57
Air Blank	0.000	04:57
Control Test	0.078	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 05/03/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUIS,B
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:16

SUBJECT'S LAST NAME: SOUBBOTIN

DATE/TIME OF TEST: 04/16/2023 16:48:38

Diagnostics Check	OK	16:41
Air Blank	0.000	16:41
Control Test	0.079	16:42
Air Blank	0.000	16:42
Subject Sample # 1	0.150	16:43
Breath Volume	3.226	16:43
Air Blank	0.000	16:44
Air Blank	0.000	16:46
Subject Sample # 2	0.130	16:47
Breath Volume	1.617	16:47
Air Blank	0.000	16:47
Control Test	0.078	16:48
Air Blank	0.000	16:48
Diagnostics Check	OK	16:48

Data Download Date/Time: 05/03/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AMPARO,AURIS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: CASTANO PINEDA

DATE/TIME OF TEST: 04/30/2023 00:58:35

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.166	00:53
Breath Volume	2.500	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.158	00:56
Breath Volume	4.906	00:56
Air Blank	0.000	00:57
Control Test	0.078	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 05/03/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,JAVIER
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: OSCEOLA

DATE/TIME OF TEST: 04/30/2023 07:53:26

Diagnostics Check	OK	07:47
Air Blank	0.000	07:47
Control Test	0.079	07:47
Air Blank	0.000	07:48
Subject Sample # 1	0.063	07:48
Breath Volume	3.078	07:48
Air Blank	0.000	07:49
Air Blank	0.000	07:51
Subject Sample # 2	0.061	07:51
Breath Volume	3.660	07:51
Air Blank	0.000	07:52
Control Test	0.079	07:52
Air Blank	0.000	07:53
Diagnostics Check	OK	07:53

Data Download Date/Time: 05/03/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 08:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: FAVERO

DATE/TIME OF TEST: 04/30/2023 09:45:59

Diagnostics Check	OK	09:38
Air Blank	0.000	09:38
Control Test	0.079	09:39
Air Blank	0.000	09:39
Subject Sample # 1	0.109	09:41
Breath Volume	1.324	09:41
Air Blank	0.000	09:41
Air Blank	0.000	09:43
Subject Sample # 2	0.111	09:44
Breath Volume	1.535	09:44
Air Blank	0.000	09:45
Control Test	0.078	09:45
Air Blank	0.000	09:45
Diagnostics Check	OK	09:45

Data Download Date/Time: 05/03/2023 11:02