

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAPELA,JORDAN,A

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: CAPELA,JORDAN

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 18:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 05/18/2023 19:13:10

Diagnostics Check	OK	19:06
Air Blank	0.000	19:07
Control Test	0.081	19:07
Air Blank	0.000	19:07
Subject Sample # 1	0.246	19:08
Breath Volume	2.816	19:08
Air Blank	0.000	19:09
Air Blank	0.000	19:10
Subject Sample # 2	0.247	19:11
Breath Volume	1.125	19:11
Air Blank	0.000	19:12
Control Test	0.079	19:12
Air Blank	0.000	19:13
Diagnostics Check	OK	19:13

Data Download Date/Time: 06/01/2023 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAPELA,JORDAN,A

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: CAPELA,JORDAN

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 04:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/19/2023 05:07:21

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.081	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1 0.162	05:03	Volume Not Met
Breath Volume	0.085	05:03	
Air Blank	0.000	05:04	
Air Blank	0.000	05:05	
Subject Sample	# 2	05:05	Subject Test Refused
Breath Volume	0.000	05:05	
Air Blank	0.000	05:06	
Control Test	0.080	05:06	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 06/01/2023 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAPELA,JORDAN,A

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: CAPELA,JORDAN

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 01:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ROMER

DATE/TIME OF TEST: 05/28/2023 01:32:17

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.082	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.152	01:27
Breath Volume	2.496	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.150	01:30
Breath Volume	2.964	01:30
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 06/01/2023 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPPOCK,JOEL,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: OGBURN

DATE/TIME OF TEST: 05/06/2023 01:11:26

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.081	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.099	01:06
Breath Volume	1.417	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.098	01:09
Breath Volume	1.296	01:09
Air Blank	0.000	01:10
Control Test	0.081	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 05/09/2023 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLANDA, RICARDO, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HOLANDA, RICARDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: SHEARER

DATE/TIME OF TEST: 05/07/2023 19:17:31

Diagnostics Check	OK	19:10
Air Blank	0.000	19:11
Control Test	0.079	19:11
Air Blank	0.000	19:12
Subject Sample # 1	0.110	19:13
Breath Volume	3.800	19:13
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.108	19:16
Breath Volume	3.644	19:16
Air Blank	0.000	19:16
Control Test	0.078	19:16
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 05/09/2023 10:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLTON,JUSTIN,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOLTON,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 05/13/2023 16:24:50

Diagnostics Check	OK	16:18
Air Blank	0.000	16:19
Control Test	0.079	16:19
Air Blank	0.000	16:20
Subject Sample # 1	0.170	16:20
Breath Volume	2.535	16:20
Air Blank	0.000	16:21
Air Blank	0.000	16:22
Subject Sample # 2	0.161	16:23
Breath Volume	1.683	16:23
Air Blank	0.000	16:23
Control Test	0.079	16:24
Air Blank	0.000	16:24
Diagnostics Check	OK	16:24

Data Download Date/Time: 06/07/2023 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,CHAD,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LYNCH,CHAD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 05/20/2023 03:05:22

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.202	03:00
Breath Volume	1.859	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.219	03:03
Breath Volume	1.117	03:03
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 06/07/2023 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NASH,JUSTIN,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA,ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 05/21/2023 21:52:24

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.207	21:47
Breath Volume	1.664	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.203	21:50
Breath Volume	2.292	21:50
Air Blank	0.000	21:51
Control Test	0.079	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 06/07/2023 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,DAVID,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MUNDT,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: CUMMINS

DATE/TIME OF TEST: 05/26/2023 03:00:30

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.205	02:56
Breath Volume	2.605	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.209	02:59
Breath Volume	2.761	02:59
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 06/07/2023 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: SALTER

DATE/TIME OF TEST: 05/28/2023 22:03:28

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.146	21:58
Breath Volume	2.078	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.152	22:01
Breath Volume	1.664	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 06/07/2023 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,DAVID,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MUNDT,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/29/2023 01:12:10

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.262	01:07
Breath Volume	2.437	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.262	01:10
Breath Volume	1.890	01:10
Air Blank	0.000	01:11
Control Test	0.079	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 06/07/2023 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA, ADALBERTO, L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA, ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MCDUFFIE

DATE/TIME OF TEST: 05/29/2023 21:34:04

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:23	
Control Test	0.081	21:23	
Air Blank	0.000	21:23	
Subject Sample	# 1 0.119	21:26	Volume Not Met
Breath Volume	0.000	21:26	
Air Blank	0.000	21:27	
Air Blank	0.000	21:29	
Subject Sample	# 2 0.098	21:32	Volume Not Met
Breath Volume	0.312	21:32	
Air Blank	0.000	21:33	
Control Test	0.081	21:33	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:34	

Data Download Date/Time: 06/07/2023 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA, ADALBERTO, L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA, ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MCDUFFIE

DATE/TIME OF TEST: 05/29/2023 21:53:48

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:38	
Control Test	0.080	21:38	
Air Blank	0.000	21:39	
Subject Sample	# 1 0.106	21:42	Volume Not Met
Breath Volume	0.613	21:42	
Air Blank	0.000	21:43	
Air Blank	0.000	21:45	
Subject Sample	# 2 0.122	21:46	
Breath Volume	1.238	21:46	
Air Blank	0.000	21:47	
Air Blank	0.000	21:49	
Subject Sample	# 3 0.101	21:52	Volume Not Met
Breath Volume	0.015	21:52	
Air Blank	0.000	21:52	
Control Test	0.080	21:53	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 06/07/2023 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAMOS,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: QUAMME

DATE/TIME OF TEST: 05/13/2023 18:39:21

Diagnostics Check	OK	18:36	
Air Blank	0.000	18:36	
Control Test	0.081	18:37	
Air Blank	0.000	18:37	
Subject Sample	# 1	18:37	Subject Test Refused
Breath Volume	0.000	18:37	
Air Blank	0.000	18:38	
Control Test	0.080	18:38	
Air Blank	0.000	18:39	
Diagnostics Check	OK	18:39	

Data Download Date/Time: 06/07/2023 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINFREE,SAIGE,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WINFREE,SAIGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:46

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 05/19/2023 15:13:43

Diagnostics Check	OK	15:06
Air Blank	0.000	15:07
Control Test	0.081	15:07
Air Blank	0.000	15:08
Subject Sample # 1	0.000	15:09
Breath Volume	1.425	15:09
Air Blank	0.000	15:09
Air Blank	0.000	15:11
Subject Sample # 2	0.000	15:12
Breath Volume	2.253	15:12
Air Blank	0.000	15:12
Control Test	0.080	15:13
Air Blank	0.000	15:13
Diagnostics Check	OK	15:13

Data Download Date/Time: 06/07/2023 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWER,RACHEL,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BOWER,RACHEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: BUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: KITCHEN

DATE/TIME OF TEST: 05/28/2023 18:09:49

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:07	
Control Test	0.081	18:07	
Air Blank	0.000	18:08	
Subject Sample	# 1	18:08	Subject Test Refused
Breath Volume	0.000	18:08	
Air Blank	0.000	18:08	
Control Test	0.082	18:09	
Air Blank	0.000	18:09	
Diagnostics Check	OK	18:09	

Data Download Date/Time: 06/07/2023 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAHR,DANIEL,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HAHR,DANIEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: BUI
ARREST TIME: 17:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: ULLRICH

DATE/TIME OF TEST: 05/28/2023 19:50:06

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:47	
Control Test	0.082	19:47	
Air Blank	0.000	19:48	
Subject Sample	# 1	19:48	Subject Test Refused
Breath Volume	0.000	19:48	
Air Blank	0.000	19:49	
Control Test	0.082	19:49	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 06/07/2023 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: HUNTER,JOHN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 05/28/2023 21:13:34

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:11	
Control Test	0.081	21:11	
Air Blank	0.000	21:11	
Subject Sample	# 1	21:12	Subject Test Refused
Breath Volume	0.000	21:12	
Air Blank	0.000	21:12	
Control Test	0.081	21:12	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 06/07/2023 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO,KENNETH,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PRIETO,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DAVISON

DATE/TIME OF TEST: 05/02/2023 00:09:03

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.080	00:07	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:08	
Control Test	0.081	00:08	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 05/09/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,JOSHUA,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILSON,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 05/02/2023 22:49:48

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:38	
Control Test	0.081	22:39	
Air Blank	0.000	22:39	
Subject Sample	# 1 0.000	22:42	Volume Not Met
Breath Volume	0.097	22:42	
Air Blank	0.000	22:43	
Air Blank	0.000	22:45	
Subject Sample	# 2	22:48	No Sample Provided
Breath Volume	0.000	22:48	
Air Blank	0.000	22:48	
Control Test	0.080	22:49	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 05/09/2023 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,JOSHUA,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILSON,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 05/02/2023 23:11:21

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:02	
Control Test	0.081	23:02	
Air Blank	0.000	23:03	
Subject Sample	# 1 0.000	23:06	Volume Not Met
Breath Volume	0.042	23:06	
Air Blank	0.000	23:06	
Air Blank	0.000	23:08	
Subject Sample	# 2	23:09	Subject Test Refused
Breath Volume	0.000	23:09	
Air Blank	0.000	23:10	
Control Test	0.081	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 05/09/2023 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEYES,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KEYES,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: STRAUGHN

DATE/TIME OF TEST: 05/03/2023 18:53:51

Diagnostics Check	OK	18:51	
Air Blank	0.000	18:51	
Control Test	0.081	18:51	
Air Blank	0.000	18:52	
Subject Sample	# 1	18:52	Subject Test Refused
Breath Volume	0.000	18:52	
Air Blank	0.000	18:52	
Control Test	0.081	18:53	
Air Blank	0.000	18:53	
Diagnostics Check	OK	18:53	

Data Download Date/Time: 05/09/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: SAPUNDJIEFF,THOMAS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MARTONE

DATE/TIME OF TEST: 05/05/2023 03:53:01

Diagnostics Check	OK	03:52	
Air Blank		03:52	RFI Detect
Air Blank	0.000	03:53	

Data Download Date/Time: 05/09/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: SAPUNDJIEFF,THOMAS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MARTONE

DATE/TIME OF TEST: 05/05/2023 04:07:51

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1	04:00	No Sample Provided
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample	# 2	04:06	No Sample Provided
Breath Volume	0.000	04:06	
Air Blank	0.000	04:06	
Control Test	0.082	04:07	
Air Blank	0.000	04:07	
Diagnostics Check	OK	04:07	

Data Download Date/Time: 05/09/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO,KENNETH,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MURPHY,JAVON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PLEAS

DATE/TIME OF TEST: 05/06/2023 01:49:23

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.081	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.122	01:45
Breath Volume	1.625	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.122	01:47
Breath Volume	2.953	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 05/09/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO,KENNETH,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MURPHY,JAVON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PLEAS

DATE/TIME OF TEST: 05/06/2023 01:58:30

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.081	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.119	01:54
Breath Volume	2.554	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.118	01:56
Breath Volume	2.453	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 05/09/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,MAURQUICE,D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WILLIAMS,JUSTIN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/06/2023 04:49:50

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.080	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:49	Sequence Aborted
Breath Volume	0.000	04:49	
Air Blank	0.000	04:49	

Data Download Date/Time: 05/09/2023 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEPETE,RYAN,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DEPETE,RYAN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: SEGER

DATE/TIME OF TEST: 05/06/2023 23:29:40

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.081	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.191	23:25
Breath Volume	1.636	23:25
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.188	23:28
Breath Volume	2.359	23:28
Air Blank	0.000	23:28
Control Test	0.079	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 05/09/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LEIGHA,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: JONES,LEIGHA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 05/10/2023 05:24:18

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.081	05:18
Air Blank	0.000	05:19
Subject Sample # 1	0.148	05:19
Breath Volume	1.675	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.155	05:22
Breath Volume	2.007	05:22
Air Blank	0.000	05:23
Control Test	0.081	05:23
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

Data Download Date/Time: 06/06/2023 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: UPSHAW,JASON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: RANSOM

DATE/TIME OF TEST: 05/13/2023 05:02:33

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.081	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.157	04:58
Breath Volume	4.660	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.161	05:01
Breath Volume	3.078	05:01
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 06/06/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: JATHAR

DATE/TIME OF TEST: 05/14/2023 05:25:30

Diagnostics Check	OK	05:16
Air Blank	0.000	05:17
Control Test	0.080	05:17
Air Blank	0.000	05:18
Subject Sample # 1	0.117	05:20
Breath Volume	1.550	05:20
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.116	05:23
Breath Volume	2.011	05:23
Air Blank	0.000	05:24
Control Test	0.080	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 06/06/2023 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MILMAN,ANDREW
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: SAMS

DATE/TIME OF TEST: 05/15/2023 21:03:15

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:00	
Control Test	0.081	21:01	
Air Blank	0.000	21:01	
Subject Sample	# 1	21:01	Subject Test Refused
Breath Volume	0.000	21:01	
Air Blank	0.000	21:02	
Control Test	0.081	21:02	
Air Blank	0.000	21:03	
Diagnostics Check	OK	21:03	

Data Download Date/Time: 06/06/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: TILLMAN, J
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: BURCHARD

DATE/TIME OF TEST: 05/18/2023 19:26:51

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.080	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.197	19:22
Breath Volume	1.531	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.191	19:25
Breath Volume	1.367	19:25
Air Blank	0.000	19:25
Control Test	0.081	19:26
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 06/06/2023 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,JORDI,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAEZ,JORDI
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GRAYBEAL

DATE/TIME OF TEST: 05/20/2023 04:01:35

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.174	03:57
Breath Volume	1.835	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.165	04:00
Breath Volume	2.261	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 06/06/2023 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,JORDI,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILSON,MALCOLM
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ROSILES

DATE/TIME OF TEST: 05/21/2023 04:34:13

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.132	04:29
Breath Volume	1.746	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.126	04:32
Breath Volume	2.019	04:32
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 06/06/2023 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WILSON,MALCOLM
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MARION

DATE/TIME OF TEST: 05/22/2023 00:57:18

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.107	00:52
Breath Volume	2.265	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.107	00:55
Breath Volume	2.378	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 06/06/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ARREDONDO

DATE/TIME OF TEST: 05/27/2023 00:19:32

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.080	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.012	00:15
Breath Volume	3.453	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.011	00:18
Breath Volume	3.695	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 06/06/2023 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARDOIN,ALEX,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PUELLO,JAVIER
ARRESTING OFFICER AGENCY: Shalimar PD
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 05/03/2023 18:56:18

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:53	
Control Test	0.081	18:54	
Air Blank	0.000	18:54	
Subject Sample	# 1	18:54	Subject Test Refused
Breath Volume	0.000	18:54	
Air Blank	0.000	18:55	
Control Test	0.079	18:55	
Air Blank	0.000	18:56	
Diagnostics Check	OK	18:56	

Data Download Date/Time: 05/23/2023 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNICK, BIANCA, V
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KEYES, CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 06:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: ADEPOJU

DATE/TIME OF TEST: 05/07/2023 07:24:45

Diagnostics Check	OK	07:21	
Air Blank	0.000	07:22	
Control Test	0.080	07:22	
Air Blank	0.000	07:23	
Subject Sample	# 1	07:23	Subject Test Refused
Breath Volume	0.000	07:23	
Air Blank	0.000	07:23	
Control Test	0.079	07:24	
Air Blank	0.000	07:24	
Diagnostics Check	OK	07:24	

Data Download Date/Time: 05/23/2023 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNICK, BIANCA, V
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KEYES, CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 06:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: ADEPOJU

DATE/TIME OF TEST: 05/07/2023 07:34:30

Diagnostics Check	OK	07:31	
Air Blank	0.000	07:32	
Control Test	0.079	07:32	
Air Blank	0.000	07:32	
Subject Sample	# 1	07:33	Subject Test Refused
Breath Volume	0.000	07:33	
Air Blank	0.000	07:33	
Control Test	0.079	07:33	
Air Blank	0.000	07:34	
Diagnostics Check	OK	07:34	

Data Download Date/Time: 05/23/2023 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/07/2023 23:53:17

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.229	23:47
Breath Volume	1.777	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.215	23:51
Breath Volume	1.875	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 05/23/2023 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO, WESTON, C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HOWARD, ELLIOTT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 05/09/2023 17:43:09

Diagnostics Check	OK	17:35
Air Blank	0.000	17:36
Control Test	0.080	17:36
Air Blank	0.000	17:36
Subject Sample # 1	0.148	17:38
Breath Volume	1.468	17:38
Air Blank	0.000	17:39
Air Blank	0.000	17:41
Subject Sample # 2	0.138	17:41
Breath Volume	1.640	17:41
Air Blank	0.000	17:42
Control Test	0.078	17:42
Air Blank	0.000	17:43
Diagnostics Check	OK	17:43

Data Download Date/Time: 05/23/2023 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 14:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: WALDRON MARSH

DATE/TIME OF TEST: 05/23/2023 16:08:25

Diagnostics Check	OK	16:00
Air Blank	0.000	16:01
Control Test	0.080	16:01
Air Blank	0.000	16:01
Subject Sample # 1	0.267	16:03
Breath Volume	2.460	16:03
Air Blank	0.000	16:04
Air Blank	0.000	16:05
Subject Sample # 2	0.253	16:06
Breath Volume	2.125	16:06
Air Blank	0.000	16:07
Control Test	0.075	16:07
Air Blank	0.000	16:08
Diagnostics Check	OK	16:08

Data Download Date/Time: 06/29/2023 19:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 05/10/2023 00:45:00

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.073	00:39
Breath Volume	1.433	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.072	00:43
Breath Volume	2.554	00:43
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:45

Data Download Date/Time: 05/27/2023 16:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,GERALD,V
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: SMITH,LORNE
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: O'HARA

DATE/TIME OF TEST: 05/13/2023 14:42:25

Diagnostics Check	OK	14:36
Air Blank	0.000	14:36
Control Test	0.079	14:36
Air Blank	0.000	14:37
Subject Sample # 1	0.163	14:37
Breath Volume	2.964	14:37
Air Blank	0.000	14:38
Air Blank	0.000	14:40
Subject Sample # 2	0.160	14:40
Breath Volume	2.921	14:40
Air Blank	0.000	14:41
Control Test	0.079	14:41
Air Blank	0.000	14:42
Diagnostics Check	OK	14:42

Data Download Date/Time: 05/27/2023 16:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGE,ESLY,J
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: HODGE,ESLY
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: NICHOLAS

DATE/TIME OF TEST: 05/20/2023 06:51:57

Diagnostics Check	OK	06:45
Air Blank	0.000	06:45
Control Test	0.079	06:46
Air Blank	0.000	06:46
Subject Sample # 1	0.152	06:47
Breath Volume	1.203	06:47
Air Blank	0.000	06:48
Air Blank	0.000	06:49
Subject Sample # 2	0.151	06:50
Breath Volume	1.406	06:50
Air Blank	0.000	06:51
Control Test	0.077	06:51
Air Blank	0.000	06:51
Diagnostics Check	OK	06:51

Data Download Date/Time: 05/27/2023 16:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASTOR,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KASTOR,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WEATHERSPOON

DATE/TIME OF TEST: 05/21/2023 04:06:17

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1 0.139	03:53	
Breath Volume	2.511	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample	# 2 0.122	03:59	Volume Not Met
Breath Volume	0.414	03:59	
Air Blank	0.000	03:59	
Air Blank	0.000	04:01	
Subject Sample	# 3	04:04	No Sample Provided
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.077	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 05/27/2023 17:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOONE,RODNEY,E
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: BOONE,RODNEY
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: BUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: BONEY

DATE/TIME OF TEST: 05/21/2023 19:12:28

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.080	19:06
Air Blank	0.000	19:07
Subject Sample # 1	0.141	19:07
Breath Volume	3.988	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.138	19:10
Breath Volume	4.726	19:10
Air Blank	0.000	19:11
Control Test	0.079	19:11
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 05/27/2023 17:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASTOR,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KIRBY,CHAD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 05/26/2023 00:42:18

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.081	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.000	00:38
Breath Volume	1.546	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.000	00:40
Breath Volume	2.066	00:40
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 05/27/2023 17:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,ANGELA,F
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: RECCA,MICHAEL
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: MIDTGARD

DATE/TIME OF TEST: 05/26/2023 20:55:34

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.080	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.177	20:51
Breath Volume	1.707	20:51
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.162	20:53
Breath Volume	1.816	20:53
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 05/27/2023 17:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,RANDY,A
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: COVINGTON,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: VICKERY

DATE/TIME OF TEST: 05/27/2023 20:50:25

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:47	
Control Test	0.080	20:48	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:49	Subject Test Refused
Breath Volume	0.000	20:49	
Air Blank	0.000	20:49	
Control Test	0.080	20:49	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

Data Download Date/Time: 06/25/2023 00:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: POPE

DATE/TIME OF TEST: 05/29/2023 03:34:52

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.080	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:34	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 06/25/2023 00:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,JAMES,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TERRELL,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: SHUTTS

DATE/TIME OF TEST: 05/01/2023 16:19:01

Diagnostics Check	OK	16:17	
Air Blank	0.000	16:17	
Control Test	0.080	16:18	
Air Blank		16:18	RFI Detect
Air Blank	0.000	16:19	

Data Download Date/Time: 05/05/2023 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,JAMES,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TERRELL,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:56

SUBJECT'S LAST NAME: SHUTTS

DATE/TIME OF TEST: 05/01/2023 16:32:46

Diagnostics Check	OK	16:22
Air Blank	0.000	16:23
Control Test	0.079	16:23
Air Blank	0.000	16:24
Subject Sample # 1	0.172	16:24
Breath Volume	2.003	16:24
Air Blank	0.000	16:25
Air Blank	0.000	16:27
Subject Sample # 2	0.149	16:27
Breath Volume	1.632	16:27
Air Blank	0.000	16:28
Air Blank	0.000	16:30
Subject Sample # 3	0.149	16:31
Breath Volume	1.316	16:31
Air Blank	0.000	16:31
Control Test	0.079	16:32
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

Data Download Date/Time: 05/05/2023 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: KOBYLARSKI

DATE/TIME OF TEST: 05/05/2023 01:48:45

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.083	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Control Test	0.082	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 05/05/2023 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,JAMES,R
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: STEPHENS,MAURICE
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 05/27/2023 23:00:49

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.202	22:56
Breath Volume	1.128	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.199	22:59
Breath Volume	1.277	22:59
Air Blank	0.000	22:59
Control Test	0.079	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 07/11/2023 07:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AKERS,CHARLES,H
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: YATES,JERRY
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: BUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: AKERS

DATE/TIME OF TEST: 05/28/2023 20:45:22

Diagnostics Check	OK	20:41	
Air Blank	0.000	20:41	
Control Test	0.081	20:42	
Air Blank	0.000	20:42	
Subject Sample	# 1	20:43	Subject Test Refused
Breath Volume	0.000	20:43	
Air Blank	0.000	20:44	
Control Test	0.080	20:44	
Air Blank	0.000	20:45	
Diagnostics Check	OK	20:45	

Data Download Date/Time: 07/11/2023 07:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: ALFORD,EVERETT
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: IRVIN

DATE/TIME OF TEST: 05/05/2023 15:54:01

Diagnostics Check	OK	15:42	
Air Blank	0.000	15:42	
Control Test	0.079	15:43	
Air Blank	0.000	15:43	
Subject Sample	# 1 0.247	15:46	Volume Not Met
Breath Volume	0.074	15:46	
Air Blank	0.000	15:47	
Air Blank	0.000	15:49	
Subject Sample	# 2 0.243	15:52	Volume Not Met
Breath Volume	0.000	15:52	
Air Blank	0.000	15:53	
Control Test	0.077	15:53	
Air Blank	0.000	15:53	
Diagnostics Check	OK	15:54	

Data Download Date/Time: 06/01/2023 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRBY,CHAD,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KIRBY,CHAD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: LIPA

DATE/TIME OF TEST: 05/10/2023 19:28:50

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.078	19:22
Air Blank	0.000	19:22
Subject Sample # 1	0.000	19:24
Breath Volume	2.277	19:24
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.000	19:27
Breath Volume	2.527	19:27
Air Blank	0.000	19:27
Control Test	0.077	19:28
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 06/01/2023 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEMMONS,LAQUINTA,M
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PLEMMONS,LAQUINTA
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 05/17/2023 00:13:47

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:11	
Control Test	0.079	00:11	
Air Blank	0.000	00:12	
Subject Sample	# 1	00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:12	
Control Test	0.077	00:13	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 06/01/2023 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADRID,AUSTIN,L
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: BARRON,EDWARD
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 05/14/2023 06:32:42

Diagnostics Check	OK	06:26
Air Blank	0.000	06:26
Control Test	0.080	06:27
Air Blank	0.000	06:27
Subject Sample # 1	0.118	06:28
Breath Volume	1.777	06:28
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample # 2	0.114	06:31
Breath Volume	2.339	06:31
Air Blank	0.000	06:31
Control Test	0.078	06:32
Air Blank	0.000	06:32
Diagnostics Check	OK	06:32

Data Download Date/Time: 05/26/2023 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 3024023281
BREATH TEST OPERATOR: PASLEY,JACOB,W	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: Washington County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PASLEY,JACOB	ARREST TIME: 19:20
ARRESTING OFFICER AGENCY: Washington County SO	GENDER: Female
LAST AGENCY INSPECTION: 05/01/2023	OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HAMILTON	DATE/TIME OF TEST: 05/09/2023 22:25:15
-------------------------------	--

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.082	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.190	22:19
Breath Volume	1.574	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.191	22:23
Breath Volume	1.804	22:23
Air Blank	0.000	22:24
Control Test	0.081	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 06/21/2023 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 3024023281
BREATH TEST OPERATOR: GUSTER,CORY,J	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: GUSTER,CORY	ARREST TIME: 19:00
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Male
LAST AGENCY INSPECTION: 05/01/2023	OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: GLASS	DATE/TIME OF TEST: 05/27/2023 19:32:21
----------------------------	--

Diagnostics Check	OK	19:29	
Air Blank	0.000	19:29	
Control Test	0.082	19:30	
Air Blank	0.000	19:30	
Subject Sample	# 1	19:30	Subject Test Refused
Breath Volume	0.000	19:30	
Air Blank	0.000	19:31	
Control Test	0.082	19:31	
Air Blank	0.000	19:32	
Diagnostics Check	OK	19:32	

Data Download Date/Time: 06/21/2023 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEWIS,TIMOTHY,B

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: LEWIS,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 3024023281

DRY GAS EXPIRY DATE: 01/20/2025

VIOLATION CODE: DUI

ARREST TIME: 20:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:19

SUBJECT'S LAST NAME: BATSON

DATE/TIME OF TEST: 05/29/2023 21:25:41

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:22	
Control Test	0.082	21:22	
Air Blank	0.000	21:23	
Subject Sample	# 1	21:24	Subject Test Refused
Breath Volume	0.000	21:24	
Air Blank	0.000	21:24	
Control Test	0.081	21:25	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 06/21/2023 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 05/05/2023 02:25:02

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.082	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.225	02:19
Breath Volume	2.343	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.216	02:23
Breath Volume	2.945	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:25

Data Download Date/Time: 06/20/2023 15:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CALOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ABNEY

DATE/TIME OF TEST: 05/07/2023 03:23:08

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.082	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Control Test	0.081	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 06/20/2023 15:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 05/18/2023 17:10:50

Diagnostics Check	OK	17:07	
Air Blank	0.000	17:08	
Control Test	0.082	17:08	
Air Blank	0.000	17:09	
Subject Sample	# 1	17:09	Subject Test Refused
Breath Volume	0.000	17:09	
Air Blank	0.000	17:09	
Control Test	0.082	17:10	
Air Blank	0.000	17:10	
Diagnostics Check	OK	17:10	

Data Download Date/Time: 06/20/2023 15:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROWN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/19/2023 00:23:18

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.083	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.129	00:18
Breath Volume	4.812	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.126	00:21
Breath Volume	4.808	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 06/20/2023 15:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDY,JOE,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARDY,JOE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 13:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: LOMONOSOVA

DATE/TIME OF TEST: 05/16/2023 14:54:21

Diagnostics Check	OK	14:43
Air Blank	0.000	14:43
Control Test	0.081	14:43
Air Blank	0.000	14:44
Subject Sample # 1	0.209	14:46
Breath Volume	2.746	14:46
Air Blank	0.000	14:46
Air Blank	0.000	14:48
Subject Sample # 2	0.185	14:49
Breath Volume	1.347	14:49
Air Blank	0.000	14:49
Air Blank	0.000	14:51
Subject Sample # 3	0.193	14:52
Breath Volume	1.277	14:52
Air Blank	0.000	14:53
Control Test	0.080	14:53
Air Blank	0.000	14:54
Diagnostics Check	OK	14:54

Data Download Date/Time: 06/21/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: ZERO TOL
ARREST TIME: 12:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 05/22/2023 02:05:47

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.082	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.130	02:01
Breath Volume	3.671	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.127	02:04
Breath Volume	2.003	02:04
Air Blank	0.000	02:04
Control Test	0.081	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 06/21/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: HART,DYLAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SHORT

DATE/TIME OF TEST: 05/10/2023 02:27:31

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.175	02:22
Breath Volume	1.582	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.169	02:26
Breath Volume	2.546	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 06/05/2023 03:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: THOMPSON,WILLIAM
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 05/11/2023 00:00:52

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:49	
Control Test	0.081	23:49	
Air Blank	0.000	23:50	
Subject Sample	# 1 0.204	23:50	
Breath Volume	1.140	23:50	
Air Blank	0.000	23:51	
Air Blank	0.000	23:53	
Subject Sample	# 2 0.162	23:56	Volume Not Met
Breath Volume	0.167	23:56	
Air Blank	0.000	23:57	
Air Blank	0.000	23:59	
Subject Sample	# 3 0.195	23:59	
Breath Volume	2.203	23:59	
Air Blank	0.000	23:59	
Control Test	0.079	00:00	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 06/05/2023 03:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,JOHN,H
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: HUNTER,JOHN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/13/2023 22:26:37

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.080	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.133	22:22
Breath Volume	1.582	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.136	22:25
Breath Volume	2.636	22:25
Air Blank	0.000	22:25
Control Test	0.081	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 06/05/2023 03:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SHAFFER,ROBIN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: BETTS

DATE/TIME OF TEST: 05/19/2023 20:05:46

Diagnostics Check	OK	19:53	
Air Blank	0.000	19:53	
Control Test	0.080	19:54	
Air Blank	0.000	19:54	
Subject Sample	# 1	19:57	No Sample Provided
Breath Volume	0.000	19:57	
Air Blank	0.000	19:58	
Air Blank	0.000	20:00	
Subject Sample	# 2	0.029	20:00
Breath Volume	1.734	20:00	
Air Blank	0.000	20:01	
Air Blank	0.000	20:03	
Subject Sample	# 3	0.026	20:04
Breath Volume	1.128	20:04	
Air Blank	0.000	20:04	
Control Test	0.081	20:05	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 06/05/2023 03:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: HART,DYLAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/20/2023 03:40:28

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.129	03:35
Breath Volume	3.343	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.126	03:38
Breath Volume	3.105	03:38
Air Blank	0.000	03:39
Control Test	0.081	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 06/05/2023 03:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON, WESLEY, A
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: HENDERSON, WESLEY
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: PINEGAR

DATE/TIME OF TEST: 05/29/2023 00:02:54

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.139	23:58
Breath Volume	4.000	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.134	00:01
Breath Volume	2.800	00:01
Air Blank	0.000	00:01
Control Test	0.081	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 06/05/2023 03:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: GOMILLION,LANCE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: HULBURD

DATE/TIME OF TEST: 05/29/2023 21:19:19

Diagnostics Check	OK	21:12
Air Blank	0.000	21:12
Control Test	0.080	21:13
Air Blank	0.000	21:13
Subject Sample # 1	0.129	21:14
Breath Volume	2.007	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.133	21:17
Breath Volume	1.136	21:17
Air Blank	0.000	21:18
Control Test	0.080	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 06/05/2023 03:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SAMMONS

DATE/TIME OF TEST: 05/11/2023 23:30:05

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.082	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.113	23:25
Breath Volume	1.554	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:27
Subject Sample # 2	0.119	23:28
Breath Volume	1.207	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 05/30/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY, LAINE, A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MEMON, MATTHEW
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 05/14/2023 02:33:59

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.187	02:29
Breath Volume	2.863	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.182	02:32
Breath Volume	2.507	02:32
Air Blank	0.000	02:33
Control Test	0.081	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 05/30/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: INGRAM,JENNIFER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/14/2023 02:53:05

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.222	02:48
Breath Volume	2.414	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.229	02:51
Breath Volume	2.335	02:51
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 05/30/2023 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,ERIC,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ROBERTS,MICHAEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:54

SUBJECT'S LAST NAME: FARINA

DATE/TIME OF TEST: 05/14/2023 14:26:56

Diagnostics Check	OK	14:20
Air Blank	0.000	14:20
Control Test	0.081	14:21
Air Blank	0.000	14:21
Subject Sample # 1	0.308	14:22
Breath Volume	3.039	14:22
Air Blank	0.000	14:22
Air Blank	0.000	14:24
Subject Sample # 2	0.302	14:25
Breath Volume	3.437	14:25
Air Blank	0.000	14:25
Control Test	0.080	14:26
Air Blank	0.000	14:26
Diagnostics Check	OK	14:26

Data Download Date/Time: 05/30/2023 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOKES,SHAYNE,R
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: STOKES,SHAYNE
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: GREIN REYNOLDS

DATE/TIME OF TEST: 05/14/2023 14:48:23

Diagnostics Check	OK	14:41
Air Blank	0.000	14:42
Control Test	0.080	14:42
Air Blank	0.000	14:42
Subject Sample # 1	0.346	14:43
Breath Volume	1.253	14:43
Air Blank	0.000	14:44
Air Blank	0.000	14:45
Subject Sample # 2	0.352	14:46
Breath Volume	1.496	14:46
Air Blank	0.000	14:47
Control Test	0.080	14:47
Air Blank	0.000	14:48
Diagnostics Check	OK	14:48

Data Download Date/Time: 05/30/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: RUTHERFORD,ROBERT
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 05/15/2023 02:42:32

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.163	02:37
Breath Volume	2.195	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.166	02:40
Breath Volume	1.835	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 05/30/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 06/18/2023
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CERDA

DATE/TIME OF TEST: 05/20/2023 03:05:03

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.081	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.193	03:00
Breath Volume	1.566	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:02
Subject Sample # 2	0.205	03:03
Breath Volume	1.726	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:05

Data Download Date/Time: 05/30/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 05/23/2023 04:13:21

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.082	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1 0.110	04:04	Volume Not Met
Breath Volume	0.566	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:07	
Subject Sample	# 2 0.000	04:09	
Breath Volume	1.597	04:09	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample	# 3	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Control Test	0.080	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 05/30/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: DOCHERTY

DATE/TIME OF TEST: 05/23/2023 14:25:55

Diagnostics Check	OK	14:22	
Air Blank	0.000	14:23	
Control Test	0.081	14:23	
Air Blank	0.000	14:24	
Subject Sample	# 1	14:24	RFI Detect
Breath Volume	0.000	14:24	
Air Blank	0.000	14:24	
Control Test	0.081	14:25	
Air Blank	0.000	14:25	
Diagnostics Check	OK	14:25	

Data Download Date/Time: 05/30/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: DOCHERTY

DATE/TIME OF TEST: 05/23/2023 14:33:37

Diagnostics Check	OK	14:27
Air Blank	0.000	14:27
Control Test	0.082	14:28
Air Blank	0.000	14:28
Subject Sample # 1	0.024	14:29
Breath Volume	3.066	14:29
Air Blank	0.000	14:29
Air Blank	0.000	14:31
Subject Sample # 2	0.020	14:32
Breath Volume	2.804	14:32
Air Blank	0.000	14:32
Control Test	0.080	14:33
Air Blank	0.000	14:33
Diagnostics Check	OK	14:33

Data Download Date/Time: 05/30/2023 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILKEY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: RAMIREZ,STEVEN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: HATHEWAY

DATE/TIME OF TEST: 05/24/2023 11:58:42

Diagnostics Check	OK	11:55	
Air Blank	0.000	11:55	
Control Test	0.082	11:56	
Air Blank	0.000	11:56	
Subject Sample	# 1	11:57	RFI Detect
Breath Volume	0.722	11:57	
Air Blank	0.000	11:57	
Control Test	0.082	11:58	
Air Blank	0.000	11:58	
Diagnostics Check	OK	11:58	

Data Download Date/Time: 05/30/2023 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILKEY,J,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: RAMIREZ,STEVEN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: HATHEWAY

DATE/TIME OF TEST: 05/24/2023 12:06:46

Diagnostics Check	OK	12:00
Air Blank	0.000	12:01
Control Test	0.081	12:01
Air Blank	0.000	12:01
Subject Sample # 1	0.000	12:02
Breath Volume	1.179	12:02
Air Blank	0.000	12:02
Air Blank	0.000	12:04
Subject Sample # 2	0.000	12:05
Breath Volume	1.992	12:05
Air Blank	0.000	12:05
Control Test	0.081	12:06
Air Blank	0.000	12:06
Diagnostics Check	OK	12:06

Data Download Date/Time: 05/30/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SAN HERAZO,BRAYAN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 05/01/2023 00:43:42

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.081	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:35	No Sample Provided
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Air Blank	0.000	00:37	
Subject Sample	# 2	0.166	00:38
Breath Volume	1.972	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:40	
Subject Sample	# 3	0.174	00:42
Breath Volume	2.136	00:42	
Air Blank	0.000	00:42	
Control Test	0.082	00:43	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 05/30/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WADE,TYLER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: GOODWYNE

DATE/TIME OF TEST: 05/05/2023 21:35:47

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:35	
Control Test	0.081	21:35	RFI Detect
Air Blank	0.000	21:35	

Data Download Date/Time: 05/30/2023 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WADE,TYLER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: GOODWYNE

DATE/TIME OF TEST: 05/05/2023 21:46:18

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.081	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.198	21:41
Breath Volume	3.167	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.188	21:44
Breath Volume	3.003	21:44
Air Blank	0.000	21:45
Control Test	0.079	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 05/30/2023 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEROY,TAD,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WAITES,DEJAS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 05/06/2023 07:33:37

Diagnostics Check	OK	07:27
Air Blank	0.000	07:27
Control Test	0.082	07:27
Air Blank	0.000	07:28
Subject Sample # 1	0.055	07:28
Breath Volume	1.492	07:28
Air Blank	0.000	07:29
Air Blank	0.000	07:31
Subject Sample # 2	0.054	07:32
Breath Volume	1.871	07:32
Air Blank	0.000	07:32
Control Test	0.082	07:33
Air Blank	0.000	07:33
Diagnostics Check	OK	07:33

Data Download Date/Time: 05/30/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUBITSCHK,LEONARD,H
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: ROUNDTREE

DATE/TIME OF TEST: 05/07/2023 21:59:58

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.082	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.082	21:55
Breath Volume	1.703	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.080	21:58
Breath Volume	1.433	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 05/30/2023 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA,ARMANDO,R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA,ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: PETTWAY

DATE/TIME OF TEST: 05/09/2023 01:17:14

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.082	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.065	01:12
Breath Volume	1.726	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.066	01:15
Breath Volume	2.613	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 05/30/2023 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: POLING

DATE/TIME OF TEST: 05/27/2023 01:33:48

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.082	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.215	01:28
Breath Volume	3.019	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.217	01:32
Breath Volume	3.066	01:32
Air Blank	0.000	01:32
Control Test	0.081	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 05/30/2023 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CORTLAND,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: JONES,CORTLAND
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 05/28/2023 00:15:15

Diagnostics Check	OK	00:14	
Air Blank		00:15	RFI Detect
Air Blank		00:15	RFI Detect

Data Download Date/Time: 05/30/2023 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CORTLAND,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: JONES,CORTLAND
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 05/28/2023 00:32:32

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.082	00:19	
Air Blank	0.000	00:19	
Subject Sample	# 1 0.017	00:22	Volume Not Met
Breath Volume	0.144	00:22	
Air Blank	0.000	00:23	
Air Blank	0.000	00:25	
Subject Sample	# 2 0.014	00:27	
Breath Volume	1.242	00:27	
Air Blank	0.000	00:28	
Air Blank	0.000	00:30	
Subject Sample	# 3 0.013	00:31	
Breath Volume	1.156	00:31	
Air Blank	0.000	00:31	
Control Test	0.081	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 05/30/2023 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAKE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/28/2023 03:22:35

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.082	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.129	03:18
Breath Volume	2.355	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.115	03:21
Breath Volume	1.226	03:21
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 05/30/2023 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: SAE KOW

DATE/TIME OF TEST: 05/28/2023 21:07:24

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:56	
Control Test	0.083	20:56	
Air Blank	0.000	20:56	
Subject Sample	# 1	20:59	No Sample Provided
Breath Volume	0.000	20:59	
Air Blank	0.000	21:00	
Air Blank	0.000	21:02	
Subject Sample	# 2	21:05	Volume Not Met
Breath Volume	1.093	21:05	
Air Blank	0.000	21:06	
Control Test	0.080	21:06	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time: 05/30/2023 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: SAE KOW

DATE/TIME OF TEST: 05/28/2023 21:20:07

Diagnostics Check	OK	21:09	
Air Blank	0.000	21:10	
Control Test	0.078	21:10	
Air Blank	0.000	21:11	
Subject Sample	# 1 0.255	21:14	Volume Not Met
Breath Volume	0.000	21:14	
Air Blank	0.000	21:15	
Air Blank	0.000	21:16	
Subject Sample	# 2	21:18	Subject Test Refused
Breath Volume	0.089	21:18	
Air Blank	0.000	21:19	
Control Test	0.076	21:19	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 05/30/2023 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: SISCO

DATE/TIME OF TEST: 05/06/2023 18:59:40

Diagnostics Check	OK	18:53
Air Blank	0.000	18:54
Control Test	0.080	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.000	18:55
Breath Volume	2.523	18:55
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.000	18:58
Breath Volume	1.917	18:58
Air Blank	0.000	18:58
Control Test	0.080	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 05/25/2023 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: ARD

DATE/TIME OF TEST: 05/11/2023 13:51:23

Diagnostics Check	OK	13:45
Air Blank	0.000	13:45
Control Test	0.080	13:45
Air Blank	0.000	13:46
Subject Sample # 1	0.000	13:47
Breath Volume	2.765	13:47
Air Blank	0.000	13:47
Air Blank	0.000	13:49
Subject Sample # 2	0.000	13:50
Breath Volume	1.843	13:50
Air Blank	0.000	13:50
Control Test	0.080	13:50
Air Blank	0.000	13:51
Diagnostics Check	OK	13:51

Data Download Date/Time: 05/25/2023 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: VILLALON HERNANDEZ

DATE/TIME OF TEST: 05/12/2023 10:06:26

Diagnostics Check	OK	10:00
Air Blank	0.000	10:00
Control Test	0.080	10:01
Air Blank	0.000	10:01
Subject Sample # 1	0.000	10:02
Breath Volume	2.164	10:02
Air Blank	0.000	10:02
Air Blank	0.000	10:04
Subject Sample # 2	0.000	10:05
Breath Volume	2.296	10:05
Air Blank	0.000	10:05
Control Test	0.080	10:05
Air Blank	0.000	10:06
Diagnostics Check	OK	10:06

Data Download Date/Time: 05/25/2023 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBLINO,NICK,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TUCKER,WILLIAM
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/13/2023 01:01:53

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.000	00:56
Breath Volume	1.289	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.000	01:00
Breath Volume	1.109	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 05/25/2023 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 05/15/2023 12:29:44

Diagnostics Check	OK	12:23
Air Blank	0.000	12:24
Control Test	0.080	12:24
Air Blank	0.000	12:25
Subject Sample # 1	0.000	12:25
Breath Volume	2.312	12:25
Air Blank	0.000	12:25
Air Blank	0.000	12:27
Subject Sample # 2	0.000	12:28
Breath Volume	2.191	12:28
Air Blank	0.000	12:28
Control Test	0.080	12:29
Air Blank	0.000	12:29
Diagnostics Check	OK	12:29

Data Download Date/Time: 05/25/2023 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROW,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GONZALEZ,GEORGE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: CROW

DATE/TIME OF TEST: 05/16/2023 09:04:54

Diagnostics Check	OK	08:59
Air Blank	0.000	08:59
Control Test	0.080	08:59
Air Blank	0.000	09:00
Subject Sample # 1	0.000	09:00
Breath Volume	1.757	09:00
Air Blank	0.000	09:01
Air Blank	0.000	09:03
Subject Sample # 2	0.000	09:03
Breath Volume	2.437	09:03
Air Blank	0.000	09:03
Control Test	0.079	09:04
Air Blank	0.000	09:04
Diagnostics Check	OK	09:04

Data Download Date/Time: 05/25/2023 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBLINO,NICK,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: HEPTINSTALL,DAVID
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: FAVELA

DATE/TIME OF TEST: 05/17/2023 16:52:27

Diagnostics Check	OK	16:40	
Air Blank	0.000	16:40	
Control Test	0.080	16:41	
Air Blank	0.000	16:41	
Subject Sample	# 1 0.123	16:44	Volume Not Met
Breath Volume	0.406	16:44	
Air Blank	0.000	16:45	
Air Blank	0.000	16:47	
Subject Sample	# 2 0.104	16:50	Volume Not Met
Breath Volume	0.730	16:50	
Air Blank	0.000	16:51	
Control Test	0.077	16:51	
Air Blank	0.000	16:52	
Diagnostics Check	OK	16:52	

Data Download Date/Time: 05/25/2023 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: GRAZIANO

DATE/TIME OF TEST: 05/18/2023 11:44:00

Diagnostics Check	OK	11:34
Air Blank	0.000	11:34
Control Test	0.080	11:35
Air Blank	0.000	11:35
Subject Sample # 1	0.086	11:37
Breath Volume	2.296	11:37
Air Blank	0.000	11:38
Air Blank	0.000	11:40
Subject Sample # 2	0.079	11:42
Breath Volume	2.468	11:42
Air Blank	0.000	11:43
Control Test	0.079	11:43
Air Blank	0.000	11:43
Diagnostics Check	OK	11:44

Data Download Date/Time: 05/25/2023 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: ANGLIN

DATE/TIME OF TEST: 05/19/2023 09:01:54

Diagnostics Check	OK	08:55
Air Blank	0.000	08:56
Control Test	0.080	08:56
Air Blank	0.000	08:57
Subject Sample # 1	0.000	08:57
Breath Volume	1.687	08:57
Air Blank	0.000	08:58
Air Blank	0.000	09:00
Subject Sample # 2	0.000	09:00
Breath Volume	1.964	09:00
Air Blank	0.000	09:01
Control Test	0.080	09:01
Air Blank	0.000	09:01
Diagnostics Check	OK	09:01

Data Download Date/Time: 05/25/2023 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 05/19/2023 11:09:46

Diagnostics Check	OK	11:03
Air Blank	0.000	11:04
Control Test	0.080	11:04
Air Blank	0.000	11:05
Subject Sample # 1	0.000	11:05
Breath Volume	2.007	11:05
Air Blank	0.000	11:06
Air Blank	0.000	11:08
Subject Sample # 2	0.000	11:08
Breath Volume	1.742	11:08
Air Blank	0.000	11:08
Control Test	0.080	11:09
Air Blank	0.000	11:09
Diagnostics Check	OK	11:09

Data Download Date/Time: 05/25/2023 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: GRANT-PERKINS

DATE/TIME OF TEST: 05/19/2023 13:12:17

Diagnostics Check	OK	13:05
Air Blank	0.000	13:06
Control Test	0.080	13:06
Air Blank	0.000	13:06
Subject Sample # 1	0.000	13:08
Breath Volume	1.894	13:08
Air Blank	0.000	13:08
Air Blank	0.000	13:10
Subject Sample # 2	0.000	13:10
Breath Volume	1.730	13:10
Air Blank	0.000	13:11
Control Test	0.079	13:11
Air Blank	0.000	13:12
Diagnostics Check	OK	13:12

Data Download Date/Time: 05/25/2023 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/28/2023 05:58:12

Diagnostics Check	OK	05:51
Air Blank	0.000	05:52
Control Test	0.081	05:52
Air Blank	0.000	05:53
Subject Sample # 1	0.186	05:53
Breath Volume	1.687	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.175	05:56
Breath Volume	1.320	05:56
Air Blank	0.000	05:57
Control Test	0.080	05:57
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 06/16/2023 18:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: TAPIA CASIANO

DATE/TIME OF TEST: 05/30/2023 04:33:22

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.123	04:28
Breath Volume	1.757	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.116	04:31
Breath Volume	1.492	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 06/16/2023 18:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: SKYBERG

DATE/TIME OF TEST: 05/31/2023 10:59:26

Diagnostics Check	OK	10:53
Air Blank	0.000	10:53
Control Test	0.080	10:54
Air Blank	0.000	10:54
Subject Sample # 1	0.000	10:55
Breath Volume	2.382	10:55
Air Blank	0.000	10:55
Air Blank	0.000	10:57
Subject Sample # 2	0.000	10:58
Breath Volume	3.074	10:58
Air Blank	0.000	10:58
Control Test	0.080	10:58
Air Blank	0.000	10:59
Diagnostics Check	OK	10:59

Data Download Date/Time: 06/16/2023 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: DOMINGUEZ-SUAREZ

DATE/TIME OF TEST: 05/31/2023 16:41:41

Diagnostics Check	OK	16:35
Air Blank	0.000	16:36
Control Test	0.081	16:36
Air Blank	0.000	16:37
Subject Sample # 1	0.000	16:37
Breath Volume	2.113	16:37
Air Blank	0.000	16:37
Air Blank	0.000	16:39
Subject Sample # 2	0.000	16:40
Breath Volume	2.058	16:40
Air Blank	0.000	16:40
Control Test	0.080	16:41
Air Blank	0.000	16:41
Diagnostics Check	OK	16:41

Data Download Date/Time: 06/16/2023 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCKETTE,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: LUCKETTE,JUSTIN
ARRESTING OFFICER AGENCY: Hamilton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: COFFEE

DATE/TIME OF TEST: 05/04/2023 03:59:58

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.082	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1 0.203	03:58	Volume Not Met
Breath Volume	0.773	03:58	
Air Blank		03:59	Purge Fail
Air Blank	0.000	03:59	

Data Download Date/Time: 05/28/2023 18:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCKETTE,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: LUCKETTE,JUSTIN
ARRESTING OFFICER AGENCY: Hamilton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: COFFEE

DATE/TIME OF TEST: 05/04/2023 04:10:46

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.202	04:06
Breath Volume	2.300	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:08
Subject Sample # 2	0.201	04:09
Breath Volume	2.328	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 05/28/2023 18:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: DE JESUS

DATE/TIME OF TEST: 05/13/2023 16:22:46

Diagnostics Check	OK	16:16
Air Blank	0.000	16:17
Control Test	0.082	16:17
Air Blank	0.000	16:17
Subject Sample # 1	0.112	16:18
Breath Volume	3.859	16:18
Air Blank	0.000	16:18
Air Blank	0.000	16:20
Subject Sample # 2	0.112	16:21
Breath Volume	3.886	16:21
Air Blank	0.000	16:21
Control Test	0.082	16:22
Air Blank	0.000	16:22
Diagnostics Check	OK	16:22

Data Download Date/Time: 05/28/2023 18:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNSED,DAVID,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: BURNSED,DAVID
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: BUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 05/28/2023 20:39:25

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.082	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.000	20:35
Breath Volume	4.210	20:35
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.000	20:38
Breath Volume	4.671	20:38
Air Blank	0.000	20:38
Control Test	0.081	20:38
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

Data Download Date/Time: 06/29/2023 02:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 05/22/2023 17:19:12

Diagnostics Check	OK	17:12
Air Blank	0.000	17:12
Control Test	0.081	17:13
Air Blank	0.000	17:13
Subject Sample # 1	0.103	17:14
Breath Volume	1.328	17:14
Air Blank	0.000	17:15
Air Blank	0.000	17:17
Subject Sample # 2	0.110	17:17
Breath Volume	1.222	17:17
Air Blank	0.000	17:18
Control Test	0.081	17:18
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

Data Download Date/Time: 05/24/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 05/22/2023 17:27:00

Diagnostics Check	OK	17:20
Air Blank	0.000	17:21
Control Test	0.082	17:21
Air Blank	0.000	17:22
Subject Sample # 1	0.112	17:22
Breath Volume	1.589	17:22
Air Blank	0.000	17:23
Air Blank	0.000	17:25
Subject Sample # 2	0.103	17:25
Breath Volume	1.402	17:25
Air Blank	0.000	17:26
Control Test	0.082	17:26
Air Blank	0.000	17:26
Diagnostics Check	OK	17:27

Data Download Date/Time: 05/24/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,JOHN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 05/03/2023 02:45:01

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.081	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1 0.238	02:35	Volume Not Met
Breath Volume	0.667	02:35	
Air Blank	0.000	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 2 0.289	02:39	
Breath Volume	2.765	02:39	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 3 0.279	02:43	
Breath Volume	2.671	02:43	
Air Blank	0.000	02:44	
Control Test	0.079	02:44	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 05/11/2023 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SAVERINO,G
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/04/2023 21:41:28

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:39	
Control Test	0.083	21:39	
Air Blank	0.000	21:39	
Subject Sample	# 1	21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:40	
Control Test	0.081	21:40	
Air Blank	0.000	21:41	
Diagnostics Check	OK	21:41	

Data Download Date/Time: 05/11/2023 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 05/05/2023 00:51:50

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:45	
Control Test	0.082	00:46	
Air Blank	0.000	00:46	
Subject Sample	# 1 0.225	00:47	
Breath Volume	1.878	00:47	
Air Blank	0.000	00:48	
Air Blank	0.000	00:50	
Subject Sample	# 2	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:50	
Control Test	0.078	00:51	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 05/11/2023 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,K
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: NILES

DATE/TIME OF TEST: 05/06/2023 23:00:51

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:58	
Control Test	0.083	22:58	
Air Blank	0.000	22:59	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	22:59	
Control Test	0.082	23:00	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 05/11/2023 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SIUZDAK,COLTIN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: WILLIAMS-CADIGAN

DATE/TIME OF TEST: 05/08/2023 01:45:45

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.083	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.000	01:41
Breath Volume	1.511	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.000	01:44
Breath Volume	3.089	01:44
Air Blank	0.000	01:44
Control Test	0.082	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 05/11/2023 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREZKOP,ZACHARY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SAVERINO,G
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/08/2023 19:32:21

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:29	
Control Test	0.083	19:29	
Air Blank	0.000	19:30	
Subject Sample	# 1	19:30	Subject Test Refused
Breath Volume	0.000	19:30	
Air Blank	0.000	19:31	
Control Test	0.081	19:31	
Air Blank	0.000	19:32	
Diagnostics Check	OK	19:32	

Data Download Date/Time: 05/11/2023 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREZKOP,ZACHARY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 05/10/2023 00:00:15

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:57	
Control Test	0.082	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:59	
Control Test	0.082	23:59	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 05/11/2023 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SAVERINO,G
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PRICHARD

DATE/TIME OF TEST: 05/10/2023 01:25:33

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.082	01:22	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:24	
Control Test	0.082	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 05/11/2023 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:44

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/11/2023 11:11:45

Diagnostics Check	OK	11:05
Air Blank	0.000	11:05
Control Test	0.082	11:06
Air Blank	0.000	11:06
Subject Sample # 1	0.000	11:07
Breath Volume	1.343	11:07
Air Blank	0.000	11:07
Air Blank	0.000	11:09
Subject Sample # 2	0.000	11:10
Breath Volume	1.238	11:10
Air Blank	0.000	11:10
Control Test	0.083	11:11
Air Blank	0.000	11:11
Diagnostics Check	OK	11:11

Data Download Date/Time: 05/11/2023 18:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MILIAN-ECHEVARRIA

DATE/TIME OF TEST: 05/13/2023 03:04:05

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.082	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:03	
Control Test	0.082	03:03	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KIBBEY,BRADLEY,A

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: KIBBEY,BRADLEY

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30822080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 05:18

GENDER: Female

OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: CRESPO

DATE/TIME OF TEST: 05/13/2023 06:49:07

Diagnostics Check	OK	06:42
Air Blank	0.000	06:42
Control Test	0.081	06:42
Air Blank	0.000	06:43
Subject Sample # 1	0.115	06:44
Breath Volume	2.109	06:44
Air Blank	0.000	06:45
Air Blank	0.000	06:46
Subject Sample # 2	0.116	06:47
Breath Volume	2.152	06:47
Air Blank	0.000	06:48
Control Test	0.082	06:48
Air Blank	0.000	06:49
Diagnostics Check	OK	06:49

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: GILLIAM

DATE/TIME OF TEST: 05/13/2023 21:01:04

Diagnostics Check	OK	20:58	
Air Blank	0.000	20:58	
Control Test	0.082	20:59	
Air Blank	0.000	20:59	
Subject Sample	# 1	20:59	Subject Test Refused
Breath Volume	0.000	20:59	
Air Blank	0.000	21:00	
Control Test	0.081	21:00	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:01	

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 05/14/2023 01:00:47

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.082	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1 0.305	00:53	Volume Not Met
Breath Volume	0.031	00:53	
Air Blank	0.000	00:54	
Air Blank	0.000	00:55	
Subject Sample	# 2 0.244	00:59	Volume Not Met
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Control Test	0.080	01:00	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SIUZDAK,C
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: TORREGIANTE

DATE/TIME OF TEST: 05/16/2023 02:13:32

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.082	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.182	02:08
Breath Volume	1.761	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.187	02:11
Breath Volume	1.781	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: RUBBA

DATE/TIME OF TEST: 05/21/2023 01:11:59

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.083	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.000	01:07
Breath Volume	1.140	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.000	01:10
Breath Volume	2.238	01:10
Air Blank	0.000	01:11
Control Test	0.083	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MCNEELEY,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/23/2023 23:27:48

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.083	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.218	23:23
Breath Volume	1.425	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:25
Subject Sample # 2	0.221	23:26
Breath Volume	1.332	23:26
Air Blank	0.000	23:26
Control Test	0.081	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: KINCADE

DATE/TIME OF TEST: 05/24/2023 21:29:20

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.082	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.150	21:24
Breath Volume	1.570	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.141	21:27
Breath Volume	1.199	21:27
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: ROCK

DATE/TIME OF TEST: 05/25/2023 04:27:54

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.082	04:24	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Control Test	0.082	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: ROCK

DATE/TIME OF TEST: 05/25/2023 04:37:13

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.082	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.152	04:32
Breath Volume	1.300	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.163	04:35
Breath Volume	2.421	04:35
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KIBBEY,BRADLEY,A

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: KIBBEY,BRADLEY

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30822080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 19:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GUTIERREZ PEREZ

DATE/TIME OF TEST: 05/26/2023 20:06:43

Diagnostics Check	OK	20:05	
Air Blank	0.000	20:06	
Control Test	0.000	20:06	RFI Detect
Air Blank		20:06	RFI Detect

Data Download Date/Time: 06/21/2023 19:02

07 August 2023

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GUTIERREZ PEREZ

DATE/TIME OF TEST: 05/26/2023 20:21:38

Diagnostics Check	OK	20:08	
Air Blank	0.000	20:09	
Control Test	0.082	20:09	
Air Blank	0.000	20:10	
Subject Sample	# 1 0.195	20:13	Volume Not Met
Breath Volume	0.398	20:13	
Air Blank	0.000	20:14	
Air Blank	0.000	20:15	
Subject Sample	# 2 0.198	20:16	
Breath Volume	1.226	20:16	
Air Blank	0.000	20:17	
Air Blank	0.000	20:19	
Subject Sample	# 3 0.191	20:19	
Breath Volume	1.164	20:19	
Air Blank	0.000	20:20	
Control Test	0.080	20:21	
Air Blank	0.000	20:21	
Diagnostics Check	OK	20:21	

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREZKOP,ZACHARY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ABDULWAHAB,N
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: WALK

DATE/TIME OF TEST: 05/26/2023 21:38:23

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.082	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.221	21:33
Breath Volume	1.289	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.225	21:36
Breath Volume	1.218	21:36
Air Blank	0.000	21:37
Control Test	0.081	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,E
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: BUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: LAFARY

DATE/TIME OF TEST: 05/28/2023 22:27:48

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:19	
Control Test	0.083	22:19	
Air Blank	0.000	22:20	
Subject Sample	# 1	22:23	No Sample Provided
Breath Volume	0.000	22:23	
Air Blank	0.000	22:24	
Air Blank	0.000	22:26	
Subject Sample	# 2	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:26	
Control Test	0.083	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: PETRU,MIROS
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 05/29/2023 18:43:39

Diagnostics Check	OK	18:36
Air Blank	0.000	18:36
Control Test	0.083	18:36
Air Blank	0.000	18:37
Subject Sample # 1	0.075	18:38
Breath Volume	1.488	18:38
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.082	18:42
Breath Volume	1.613	18:42
Air Blank	0.000	18:42
Control Test	0.082	18:43
Air Blank	0.000	18:43
Diagnostics Check	OK	18:43

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 05/29/2023 22:32:24

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:29	
Control Test	0.083	22:30	
Air Blank	0.000	22:30	
Subject Sample	# 1	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Control Test	0.083	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/29/2023 23:08:43

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:06	
Control Test	0.082	23:06	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:07	Subject Test Refused
Breath Volume	0.000	23:07	
Air Blank	0.000	23:07	
Control Test	0.083	23:08	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIGIA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BENDER

DATE/TIME OF TEST: 05/30/2023 00:28:52

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.082	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:27	
Control Test	0.082	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSNYAK,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: CANTRELL,SHANE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: KNAPP

DATE/TIME OF TEST: 05/25/2023 02:23:12

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:13	No Sample Provided
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample	# 2	0.065	02:16
Breath Volume	2.203	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample	# 3	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:22	
Control Test	0.079	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEASEL,KELSEY,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: FEASEL,KELSEY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ZAHARIAS

DATE/TIME OF TEST: 05/27/2023 04:01:32

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.079	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1 0.181	03:54	Volume Not Met
Breath Volume	0.527	03:54	
Air Blank	0.000	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 2 0.198	03:59	Volume Not Met
Breath Volume	0.078	03:59	
Air Blank	0.000	04:00	
Control Test	0.079	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICKSON,LAUREN,C
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: DICKSON,LAUREN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: BUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/27/2023 23:38:45

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.077	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.122	23:34
Breath Volume	3.117	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.124	23:37
Breath Volume	3.593	23:37
Air Blank	0.000	23:37
Control Test	0.078	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GUTIERREZ RAMIREZ

DATE/TIME OF TEST: 05/29/2023 02:44:36

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.091	02:40
Breath Volume	1.308	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.090	02:43
Breath Volume	1.464	02:43
Air Blank	0.000	02:43
Control Test	0.078	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JESSE,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BROWN,JESSE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/30/2023 00:10:25

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.323	00:05
Breath Volume	2.402	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.319	00:08
Breath Volume	2.085	00:08
Air Blank	0.000	00:09
Control Test	0.078	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/30/2023 18:27:09

Diagnostics Check	OK	18:16	
Air Blank	0.000	18:17	
Control Test	0.080	18:17	
Air Blank	0.000	18:17	
Subject Sample	# 1 0.080	18:21	Volume Not Met
Breath Volume	0.210	18:21	
Air Blank	0.000	18:21	
Air Blank	0.000	18:23	
Subject Sample	# 2	18:25	Subject Test Refused
Breath Volume	0.113	18:25	
Air Blank	0.000	18:26	
Control Test	0.079	18:26	
Air Blank	0.000	18:27	
Diagnostics Check	OK	18:27	

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY, SPENSER, J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KENNEDY, SPENSER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MASTROSERIO

DATE/TIME OF TEST: 05/30/2023 21:01:57

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.000	20:57
Breath Volume	2.132	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.000	21:00
Breath Volume	2.253	21:00
Air Blank	0.000	21:01
Control Test	0.079	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,YOUNG
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: SPINELLA

DATE/TIME OF TEST: 05/31/2023 01:50:43

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.233	01:46
Breath Volume	1.390	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.224	01:49
Breath Volume	1.390	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: REYNOLDS,DONOVAN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: OWNBY

DATE/TIME OF TEST: 05/05/2023 01:42:13

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.080	01:40	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 05/29/2023 02:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ROSARIO,F
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: BRENT

DATE/TIME OF TEST: 05/06/2023 12:34:35

Diagnostics Check	OK	12:25	
Air Blank	0.000	12:26	
Control Test	0.080	12:26	
Air Blank	0.000	12:27	
Subject Sample	# 1	12:30	No Sample Provided
Breath Volume	0.000	12:30	
Air Blank	0.000	12:30	
Air Blank	0.000	12:32	
Subject Sample	# 2	12:33	Subject Test Refused
Breath Volume	0.000	12:33	
Air Blank	0.000	12:33	
Control Test	0.080	12:33	
Air Blank	0.000	12:34	
Diagnostics Check	OK	12:34	

Data Download Date/Time: 05/29/2023 02:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRINKLE,BRITTANY,N
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NEIGER,K
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: LINGERFELT

DATE/TIME OF TEST: 05/06/2023 22:54:19

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.120	22:48
Breath Volume	1.218	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.117	22:52
Breath Volume	1.535	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 05/29/2023 02:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRINKLE,BRITTANY,N
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WELLMAN,A
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SUDANTAS

DATE/TIME OF TEST: 05/07/2023 00:20:07

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 1 0.159	00:09	
Breath Volume	2.062	00:09	
Air Blank	0.000	00:10	
Air Blank	0.000	00:11	
Subject Sample	# 2 0.143	00:14	Volume Not Met
Breath Volume	0.699	00:14	
Air Blank	0.000	00:15	
Air Blank	0.000	00:17	
Subject Sample	# 3 0.146	00:18	
Breath Volume	1.921	00:18	
Air Blank	0.000	00:19	
Control Test	0.077	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 05/29/2023 02:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBRIGHT,BRIAN',M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 05/06/2023 00:24:32

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:22	
Control Test	0.080	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.080	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 05/31/2023 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBRIGHT,BRIAN,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KNIGHT,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: FREIER

DATE/TIME OF TEST: 05/06/2023 01:02:36

Diagnostics Check	OK	00:59	
Air Blank	0.000	01:00	
Control Test	0.080	01:00	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.079	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 05/31/2023 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANSOM,AUBREY,G
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: RANSOM,AUBREY
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: DORMAND

DATE/TIME OF TEST: 05/06/2023 23:08:15

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.081	23:06	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:06	Subject Test Refused
Breath Volume	0.000	23:06	
Air Blank	0.000	23:07	
Control Test	0.081	23:07	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

Data Download Date/Time: 05/31/2023 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBRIGHT,BRIAN,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: CAPERELL

DATE/TIME OF TEST: 05/09/2023 17:12:53

Diagnostics Check	OK	17:04
Air Blank	0.000	17:04
Control Test	0.081	17:05
Air Blank	0.000	17:05
Subject Sample # 1	0.000	17:08
Breath Volume	1.578	17:08
Air Blank	0.000	17:09
Air Blank	0.000	17:11
Subject Sample # 2	0.000	17:11
Breath Volume	1.179	17:11
Air Blank	0.000	17:11
Control Test	0.080	17:12
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

Data Download Date/Time: 05/31/2023 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: D'ADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: OCHS

DATE/TIME OF TEST: 05/11/2023 00:19:47

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.159	00:14
Breath Volume	2.230	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.166	00:18
Breath Volume	1.152	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 05/31/2023 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, ANTHONY, T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON, TIM
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 05/12/2023 23:49:08

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:46	
Control Test	0.080	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:48	
Control Test	0.080	23:48	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time: 05/31/2023 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:31

SUBJECT'S LAST NAME: HARPST

DATE/TIME OF TEST: 05/13/2023 15:55:48

Diagnostics Check	OK	15:52	
Air Blank	0.000	15:53	
Control Test	0.079	15:53	
Air Blank	0.000	15:54	
Subject Sample	# 1	15:54	Subject Test Refused
Breath Volume	0.000	15:54	
Air Blank	0.000	15:54	
Control Test	0.078	15:55	
Air Blank	0.000	15:55	
Diagnostics Check	OK	15:55	

Data Download Date/Time: 05/31/2023 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ELIZABETH,A
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: BROWN,ELIZABETH
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: RISE

DATE/TIME OF TEST: 05/17/2023 00:48:40

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.113	00:43
Breath Volume	2.085	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.116	00:47
Breath Volume	2.078	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BOTTING

DATE/TIME OF TEST: 05/19/2023 02:11:07

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.214	02:06
Breath Volume	1.519	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample # 2	0.214	02:09
Breath Volume	1.117	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBRIGHT,BRIAN,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HUNTER,D
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:04

SUBJECT'S LAST NAME: BAGLEY

DATE/TIME OF TEST: 05/19/2023 07:30:56

Diagnostics Check	OK	07:24
Air Blank	0.000	07:25
Control Test	0.080	07:25
Air Blank	0.000	07:26
Subject Sample # 1	0.000	07:26
Breath Volume	1.761	07:26
Air Blank	0.000	07:27
Air Blank	0.000	07:29
Subject Sample # 2	0.000	07:29
Breath Volume	1.734	07:29
Air Blank	0.000	07:30
Control Test	0.079	07:30
Air Blank	0.000	07:30
Diagnostics Check	OK	07:30

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: BAZATA

DATE/TIME OF TEST: 05/19/2023 19:55:39

Diagnostics Check	OK	19:52	
Air Blank	0.000	19:53	
Control Test	0.080	19:53	
Air Blank	0.000	19:54	
Subject Sample	# 1	19:54	Subject Test Refused
Breath Volume	0.000	19:54	
Air Blank	0.000	19:54	
Control Test	0.080	19:55	
Air Blank	0.000	19:55	
Diagnostics Check	OK	19:55	

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BANNON,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: COUSINET

DATE/TIME OF TEST: 05/20/2023 03:41:10

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.080	03:39	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Control Test	0.080	03:40	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BROWNING,J
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 05/20/2023 19:03:41

Diagnostics Check	OK	19:00	
Air Blank	0.000	19:00	
Control Test	0.080	19:01	
Air Blank	0.000	19:01	
Subject Sample	# 1	19:02	Subject Test Refused
Breath Volume	0.000	19:02	
Air Blank	0.000	19:02	
Control Test	0.079	19:03	
Air Blank	0.000	19:03	
Diagnostics Check	OK	19:03	

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,L
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ENSOR

DATE/TIME OF TEST: 05/21/2023 00:00:30

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.080	23:47	
Air Blank	0.000	23:48	
Subject Sample	# 1 0.091	23:51	Volume Not Met
Breath Volume	0.316	23:51	
Air Blank	0.000	23:52	
Air Blank	0.000	23:54	
Subject Sample	# 2 0.102	23:55	
Breath Volume	2.019	23:55	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample	# 3 0.099	23:58	
Breath Volume	1.144	23:58	
Air Blank	0.000	23:59	
Control Test	0.079	23:59	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURO,JAMES,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/21/2023 14:21:42

Diagnostics Check	OK	14:18	
Air Blank	0.000	14:18	
Control Test	0.080	14:18	
Air Blank	0.000	14:19	
Subject Sample	# 1	14:20	Subject Test Refused
Breath Volume	0.000	14:20	
Air Blank	0.000	14:20	
Control Test	0.080	14:21	
Air Blank	0.000	14:21	
Diagnostics Check	OK	14:21	

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ULRICH,B
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 05/21/2023 21:15:49

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:13	
Control Test	0.080	21:13	
Air Blank	0.000	21:14	
Subject Sample	# 1	21:14	Subject Test Refused
Breath Volume	0.000	21:14	
Air Blank	0.000	21:14	
Control Test	0.080	21:15	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:15	

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIGLEY, DENISE, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ADOLF, C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: PACE

DATE/TIME OF TEST: 05/26/2023 17:58:43

Diagnostics Check	OK	17:52
Air Blank	0.000	17:52
Control Test	0.081	17:53
Air Blank	0.000	17:53
Subject Sample # 1	0.000	17:54
Breath Volume	3.531	17:54
Air Blank	0.000	17:54
Air Blank	0.000	17:56
Subject Sample # 2	0.000	17:57
Breath Volume	2.335	17:57
Air Blank	0.000	17:57
Control Test	0.081	17:58
Air Blank	0.000	17:58
Diagnostics Check	OK	17:58

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, ANTHONY, T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON, TIM
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/27/2023 00:55:06

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.184	00:50
Breath Volume	2.808	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.177	00:53
Breath Volume	1.968	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FISHER,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: HUFFMAN

DATE/TIME OF TEST: 05/27/2023 19:26:50

Diagnostics Check	OK	19:22	
Air Blank	0.000	19:22	
Control Test	0.080	19:23	
Air Blank	0.000	19:23	
Subject Sample	# 1	19:25	Subject Test Refused
Breath Volume	0.000	19:25	
Air Blank	0.000	19:25	
Control Test	0.080	19:26	
Air Blank	0.000	19:26	
Diagnostics Check	OK	19:26	

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, ANTHONY, T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SARMEINTO, M
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: SCHULT

DATE/TIME OF TEST: 05/28/2023 19:39:47

Diagnostics Check	OK	19:32
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:34
Subject Sample # 1	0.040	19:35
Breath Volume	2.140	19:35
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.041	19:38
Breath Volume	1.292	19:38
Air Blank	0.000	19:38
Control Test	0.080	19:39
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: BELGICA

DATE/TIME OF TEST: 05/28/2023 23:20:57

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.119	23:16
Breath Volume	1.117	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.125	23:19
Breath Volume	2.210	23:19
Air Blank	0.000	23:20
Control Test	0.079	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: NAPOLITANO

DATE/TIME OF TEST: 05/29/2023 20:39:05

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:27	
Control Test	0.079	20:28	
Air Blank	0.000	20:28	
Subject Sample	# 1 0.253	20:31	Volume Not Met
Breath Volume	0.652	20:31	
Air Blank	0.000	20:32	
Air Blank	0.000	20:34	
Subject Sample	# 2 0.284	20:37	Volume Not Met
Breath Volume	0.066	20:37	
Air Blank	0.000	20:38	
Control Test	0.079	20:38	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:39	

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: NAPOLITANO

DATE/TIME OF TEST: 05/29/2023 20:52:20

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:41	
Control Test	0.078	20:41	
Air Blank	0.000	20:41	
Subject Sample	# 1 0.278	20:45	Volume Not Met
Breath Volume	0.035	20:45	
Air Blank	0.000	20:45	
Air Blank	0.000	20:47	
Subject Sample	# 2 0.192	20:50	Volume Not Met
Breath Volume	0.226	20:50	
Air Blank	0.000	20:51	
Control Test	0.078	20:51	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 05/31/2023 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: PRANOUSKES,DENNIS
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 05/14/2023 04:50:08

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.079	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.180	04:45
Breath Volume	1.363	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:47
Subject Sample # 2	0.174	04:48
Breath Volume	1.347	04:48
Air Blank	0.000	04:49
Control Test	0.077	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 05/26/2023 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,N,
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILDEMAN-STUART,N
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/17/2023 02:08:25

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.000	02:04
Breath Volume	3.324	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.000	02:07
Breath Volume	2.441	02:07
Air Blank	0.000	02:07
Control Test	0.078	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 05/26/2023 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: POTH,LAUREN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: NEMEC

DATE/TIME OF TEST: 05/18/2023 23:38:49

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.080	23:33	
Air Blank	0.000	23:34	
Subject Sample	# 1 0.166	23:37	Volume Not Met
Breath Volume	1.074	23:37	
Air Blank		23:38	Purge Fail
Air Blank	0.000	23:38	

Data Download Date/Time: 05/26/2023 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: POTH,LAUREN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: NEMEC

DATE/TIME OF TEST: 05/18/2023 23:44:10

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:43	
Control Test	0.073	23:43	Control Outside Tolerance
Air Blank	0.000	23:44	

Data Download Date/Time: 05/26/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: POTH,LAUREN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: NEMEC

DATE/TIME OF TEST: 05/18/2023 23:52:05

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.077	23:47	
Air Blank	0.000	23:47	
Subject Sample	# 1 0.162	23:50	Volume Not Met
Breath Volume	0.492	23:50	
Air Blank		23:51	Purge Fail
Air Blank	0.000	23:52	

Data Download Date/Time: 05/26/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: POTH,LAUREN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: N520793529411

DATE/TIME OF TEST: 05/19/2023 00:10:08

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:56	
Control Test	0.076	23:57	
Air Blank	0.000	23:57	
Subject Sample	# 3 0.174	23:58	
Breath Volume	1.148	23:58	
Air Blank	0.000	23:59	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.149	00:04	Volume Not Met
Breath Volume	1.042	00:04	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample	# 2 0.164	00:08	
Breath Volume	1.261	00:08	
Air Blank	0.000	00:09	
Control Test	0.076	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 05/26/2023 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: PUGMIRE,TRENT
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: DEBORDE

DATE/TIME OF TEST: 05/23/2023 23:31:58

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.079	23:29	
Air Blank	0.000	23:30	
Subject Sample	# 1	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:31	
Control Test	0.080	23:31	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 05/26/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH, LAUREN, A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: RICHETTI, KALEIGH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 05/24/2023 00:37:10

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.073	00:32
Breath Volume	2.296	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.072	00:35
Breath Volume	2.027	00:35
Air Blank	0.000	00:36
Control Test	0.077	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 05/26/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RICHETTI,K
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: DORSEY

DATE/TIME OF TEST: 05/27/2023 21:35:20

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.080	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.197	21:29
Breath Volume	1.316	21:29
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.202	21:33
Breath Volume	1.996	21:33
Air Blank	0.000	21:34
Control Test	0.077	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 06/21/2023 19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,D,J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: PARISH,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MC NEAL

DATE/TIME OF TEST: 05/28/2023 01:57:09

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.079	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:56	
Control Test	0.079	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 06/21/2023 19:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: POTH,LAUREN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: NESPECA

DATE/TIME OF TEST: 05/28/2023 22:59:54

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.080	22:57	
Air Blank	0.000	22:57	
Subject Sample	# 1	22:58	Subject Test Refused
Breath Volume	0.000	22:58	
Air Blank	0.000	22:58	
Control Test	0.080	22:59	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 06/21/2023 19:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: ANDREANO,JACK
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: NESPECA

DATE/TIME OF TEST: 05/28/2023 23:07:27

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.080	23:05	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:06	Subject Test Refused
Breath Volume	0.000	23:06	
Air Blank	0.000	23:06	
Control Test	0.079	23:06	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 06/21/2023 19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,N,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILDEMAN-STUART,N
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: ROE

DATE/TIME OF TEST: 05/30/2023 02:28:52

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.124	02:24
Breath Volume	1.855	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.124	02:27
Breath Volume	1.691	02:27
Air Blank	0.000	02:27
Control Test	0.078	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 06/21/2023 19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: MCCOOK

DATE/TIME OF TEST: 05/16/2023 23:40:08

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.182	23:35
Breath Volume	1.320	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:37
Subject Sample # 2	0.176	23:38
Breath Volume	1.812	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 06/28/2023 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 05/20/2023 00:58:32

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.079	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Control Test	0.080	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 06/28/2023 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: LATIMER,JUSTIN

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 05/21/2023 03:18:32

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.078	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 06/28/2023 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: LOLLEY

DATE/TIME OF TEST: 05/22/2023 07:03:19

Diagnostics Check	OK	06:57
Air Blank	0.000	06:57
Control Test	0.080	06:57
Air Blank	0.000	06:58
Subject Sample # 1	0.179	06:59
Breath Volume	1.757	06:59
Air Blank	0.000	06:59
Air Blank	0.000	07:01
Subject Sample # 2	0.188	07:01
Breath Volume	1.796	07:01
Air Blank	0.000	07:02
Control Test	0.079	07:02
Air Blank	0.000	07:03
Diagnostics Check	OK	07:03

Data Download Date/Time: 06/28/2023 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECCUE,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BENVENUTI,RICHARD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GANDIA

DATE/TIME OF TEST: 05/01/2023 03:26:21

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.000	03:22
Breath Volume	1.960	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.000	03:24
Breath Volume	1.378	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 10:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: RAMNARAIN

DATE/TIME OF TEST: 05/02/2023 23:06:50

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:55	
Control Test	0.081	22:55	
Air Blank	0.000	22:56	
Subject Sample	# 1 0.153	22:59	Volume Not Met
Breath Volume	0.417	22:59	
Air Blank	0.000	23:00	
Air Blank	0.000	23:02	
Subject Sample	# 2 0.161	23:05	Volume Not Met
Breath Volume	0.812	23:05	
Air Blank	0.000	23:05	
Control Test	0.081	23:06	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: RAMNARAIN

DATE/TIME OF TEST: 05/02/2023 23:23:13

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:09	
Subject Sample	# 1 0.192	23:10	
Breath Volume	1.433	23:10	
Air Blank	0.000	23:10	
Air Blank	0.000	23:12	
Subject Sample	# 2 0.134	23:15	Volume Not Met
Breath Volume	0.035	23:15	
Air Blank	0.000	23:16	
Air Blank	0.000	23:18	
Subject Sample	# 3 0.164	23:21	Volume Not Met
Breath Volume	0.113	23:21	
Air Blank	0.000	23:22	
Control Test	0.079	23:22	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 05/03/2023 00:32:45

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1 0.099	00:30	
Breath Volume	1.929	00:30	
Air Blank		00:31	RFI Detect
Air Blank	0.000	00:31	
Control Test	0.080	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 05/03/2023 00:41:17

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.081	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.106	00:36
Breath Volume	2.511	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.106	00:39
Breath Volume	1.593	00:39
Air Blank	0.000	00:40
Control Test	0.077	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENTSCH,RICHARD,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: CANTRELL,SEAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: TRIPP

DATE/TIME OF TEST: 05/03/2023 01:21:24

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.202	01:16
Breath Volume	1.691	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.205	01:19
Breath Volume	1.671	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEMLER,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SEMLER,JOSEPH
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: BOHLIN

DATE/TIME OF TEST: 05/04/2023 01:52:47

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.081	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:52	Sequence Aborted
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEMLER,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SEMLER,JOSEPH
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: BOHLIN

DATE/TIME OF TEST: 05/04/2023 02:03:30

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.000	01:59
Breath Volume	1.273	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.000	02:02
Breath Volume	1.367	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARIAS,W,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ARIAS,W
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 05/04/2023 18:28:01

Diagnostics Check	OK	18:24	
Air Blank	0.000	18:25	
Control Test	0.082	18:25	
Air Blank	0.000	18:26	
Subject Sample	# 1	18:26	Subject Test Refused
Breath Volume	0.000	18:26	
Air Blank	0.000	18:27	
Control Test	0.082	18:27	
Air Blank	0.000	18:27	
Diagnostics Check	OK	18:28	

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JESSICA,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: DICKEY

DATE/TIME OF TEST: 05/05/2023 01:00:34

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.082	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.215	00:56
Breath Volume	4.140	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.220	00:58
Breath Volume	1.175	00:58
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RIDDLE,BRIANNA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/05/2023 02:28:33

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.081	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.081	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REUL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: REUL,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/05/2023 21:45:18

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:42	
Control Test	0.080	21:43	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:43	Subject Test Refused
Breath Volume	0.000	21:43	
Air Blank	0.000	21:44	
Control Test	0.082	21:44	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMBERT,CODY,R
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: NORRIS,THOMAS
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 05/06/2023 21:27:53

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.081	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.198	21:23
Breath Volume	1.855	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:25
Subject Sample # 2	0.191	21:26
Breath Volume	1.773	21:26
Air Blank	0.000	21:26
Control Test	0.079	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAINO,JOSHUA,V
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: PARKER,JOSHUA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 05/08/2023 02:31:02

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.000	02:25
Breath Volume	1.605	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.000	02:29
Breath Volume	1.867	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:31

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SIMONSEN,ERIK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 05/09/2023 00:10:22

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:08	
Control Test	0.082	00:08	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:09	
Control Test	0.082	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 05/10/2023 06:12:14

Diagnostics Check	OK	06:06
Air Blank	0.000	06:06
Control Test	0.081	06:06
Air Blank	0.000	06:07
Subject Sample # 1	0.000	06:08
Breath Volume	2.183	06:08
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 2	0.000	06:10
Breath Volume	1.593	06:10
Air Blank	0.000	06:11
Control Test	0.079	06:11
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUBAISY, MUHAMMAD ALAMIN,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: QUBAISY, MUHAMMAD-
ALAMIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Male

OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: PIKE

DATE/TIME OF TEST: 05/12/2023 11:03:20

Diagnostics Check	OK	10:49	
Air Blank	0.000	10:50	
Control Test	0.081	10:50	
Air Blank	0.000	10:51	
Subject Sample	# 1 0.227	10:54	Volume Not Met
Breath Volume	0.140	10:54	
Air Blank	0.000	10:55	
Air Blank	0.000	10:56	
Subject Sample	# 2 0.228	10:57	
Breath Volume	1.742	10:57	
Air Blank	0.000	10:58	
Air Blank	0.000	11:00	
Subject Sample	# 3 0.248	11:01	
Breath Volume	1.320	11:01	
Air Blank	0.000	11:02	
Control Test	0.077	11:02	
Air Blank	0.000	11:03	
Diagnostics Check	OK	11:03	

Data Download Date/Time: 05/24/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEMLER,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SEMLER,JOSEPH
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: GOGUEN

DATE/TIME OF TEST: 05/12/2023 23:43:55

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.082	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.076	23:39
Breath Volume	1.566	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:41
Subject Sample # 2	0.058	23:42
Breath Volume	1.183	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 05/24/2023 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEASEL,KELSEY,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: FEASEL,KELSEY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: GLUHAREEF

DATE/TIME OF TEST: 05/13/2023 04:10:00

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.082	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.163	04:05
Breath Volume	2.277	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.155	04:08
Breath Volume	1.382	04:08
Air Blank	0.000	04:09
Control Test	0.077	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:10

Data Download Date/Time: 05/24/2023 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MENDEZ PEREZ

DATE/TIME OF TEST: 05/13/2023 04:57:55

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:45	
Control Test	0.081	04:45	
Air Blank	0.000	04:46	
Subject Sample	# 1 0.199	04:49	Volume Not Met
Breath Volume	0.328	04:49	
Air Blank	0.000	04:50	
Air Blank	0.000	04:52	
Subject Sample	# 2 0.211	04:53	
Breath Volume	1.785	04:53	
Air Blank	0.000	04:53	
Air Blank	0.000	04:55	
Subject Sample	# 3 0.213	04:56	
Breath Volume	1.859	04:56	
Air Blank	0.000	04:57	
Control Test	0.081	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 05/24/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKINNER,KAILEIGH,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: SKINNER,KAILEIGH
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 05/14/2023 01:36:05

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.081	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.045	01:31
Breath Volume	1.125	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.038	01:34
Breath Volume	1.597	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 05/24/2023 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEASEL,KELSEY,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: FOGARTY,ANTHONY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 05/14/2023 03:29:48

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.081	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.119	03:25
Breath Volume	1.785	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.105	03:28
Breath Volume	1.507	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 05/24/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUIS,ALYSSA,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ROUIS,ALYSSA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: SEEGOPAUL

DATE/TIME OF TEST: 05/14/2023 04:23:15

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.082	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.150	04:18
Breath Volume	2.445	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:20
Subject Sample # 2	0.156	04:21
Breath Volume	1.242	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 05/24/2023 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: ACABAL CXELEOH

DATE/TIME OF TEST: 05/14/2023 05:38:39

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.081	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.170	05:34
Breath Volume	3.648	05:34
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.164	05:37
Breath Volume	2.070	05:37
Air Blank	0.000	05:37
Control Test	0.079	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 05/24/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DILLON,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SHUMAN,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 08:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: OKER

DATE/TIME OF TEST: 05/14/2023 09:19:19

Diagnostics Check	OK	09:10
Air Blank	0.000	09:10
Control Test	0.082	09:11
Air Blank	0.000	09:11
Subject Sample # 1	0.060	09:12
Breath Volume	1.339	09:12
Air Blank	0.000	09:13
Air Blank	0.000	09:15
Subject Sample # 2	0.062	09:17
Breath Volume	3.917	09:17
Air Blank	0.000	09:18
Control Test	0.081	09:18
Air Blank	0.000	09:19
Diagnostics Check	OK	09:19

Data Download Date/Time: 05/24/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELCHER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MELCHER,ANDREW
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/04/2023 01:51:43

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.077	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.207	01:47
Breath Volume	1.515	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.203	01:50
Breath Volume	2.085	01:50
Air Blank	0.000	01:50
Control Test	0.078	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 05/10/2023 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:48

SUBJECT'S LAST NAME: KANTZ

DATE/TIME OF TEST: 05/04/2023 08:14:40

Diagnostics Check	OK	08:08
Air Blank	0.000	08:09
Control Test	0.080	08:09
Air Blank	0.000	08:09
Subject Sample # 1	0.192	08:10
Breath Volume	1.558	08:10
Air Blank	0.000	08:10
Air Blank	0.000	08:12
Subject Sample # 2	0.194	08:13
Breath Volume	2.398	08:13
Air Blank	0.000	08:13
Control Test	0.081	08:14
Air Blank	0.000	08:14
Diagnostics Check	OK	08:14

Data Download Date/Time: 05/10/2023 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BENNETT,BRANDY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: HOGLAND

DATE/TIME OF TEST: 05/04/2023 22:07:19

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.176	22:02
Breath Volume	2.390	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.166	22:05
Breath Volume	1.804	22:05
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 05/10/2023 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: LONG JR

DATE/TIME OF TEST: 05/09/2023 01:05:29

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.078	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.183	01:01
Breath Volume	2.386	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.187	01:04
Breath Volume	2.503	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 05/10/2023 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELCHER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MELCHER,ANDREW
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/09/2023 01:57:17

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.122	01:52
Breath Volume	1.847	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.118	01:55
Breath Volume	2.562	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 05/10/2023 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCASTER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: LANCASTER,JONATHAN
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 05/16/2023 04:10:41

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.143	04:06
Breath Volume	1.738	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.142	04:09
Breath Volume	3.039	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 06/08/2023 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: EMERSON

DATE/TIME OF TEST: 05/28/2023 00:43:38

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.081	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.047	00:39
Breath Volume	1.863	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.048	00:42
Breath Volume	1.593	00:42
Air Blank	0.000	00:42
Control Test	0.081	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 06/08/2023 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 05:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 05/28/2023 07:15:59

Diagnostics Check	OK	07:07	
Air Blank	0.000	07:07	
Control Test	0.081	07:07	
Air Blank	0.000	07:08	
Subject Sample	# 1	07:11	No Sample Provided
Breath Volume	0.000	07:11	
Air Blank	0.000	07:12	
Air Blank	0.000	07:14	
Subject Sample	# 2	07:14	Subject Test Refused
Breath Volume	0.000	07:14	
Air Blank	0.000	07:15	
Control Test	0.081	07:15	
Air Blank	0.000	07:15	
Diagnostics Check	OK	07:15	

Data Download Date/Time: 06/08/2023 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ROBERT,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: ANGULO

DATE/TIME OF TEST: 05/29/2023 07:11:50

Diagnostics Check	OK	07:04
Air Blank	0.000	07:05
Control Test	0.080	07:05
Air Blank	0.000	07:05
Subject Sample # 1	0.128	07:07
Breath Volume	2.683	07:07
Air Blank	0.000	07:07
Air Blank	0.000	07:09
Subject Sample # 2	0.127	07:10
Breath Volume	2.000	07:10
Air Blank	0.000	07:10
Control Test	0.080	07:11
Air Blank	0.000	07:11
Diagnostics Check	OK	07:11

Data Download Date/Time: 06/08/2023 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 08:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/31/2023 08:59:14

Diagnostics Check	OK	08:51
Air Blank	0.000	08:52
Control Test	0.080	08:52
Air Blank	0.000	08:53
Subject Sample # 1	0.000	08:54
Breath Volume	1.187	08:54
Air Blank	0.000	08:54
Air Blank	0.000	08:56
Subject Sample # 2	0.000	08:57
Breath Volume	1.167	08:57
Air Blank	0.000	08:58
Control Test	0.081	08:58
Air Blank	0.000	08:59
Diagnostics Check	OK	08:59

Data Download Date/Time: 06/08/2023 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/04/2023 02:14:02

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.081	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.164	02:09
Breath Volume	3.398	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.163	02:12
Breath Volume	2.699	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:14

Data Download Date/Time: 05/11/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOPEZ,JACQIE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: HAGAN

DATE/TIME OF TEST: 05/04/2023 02:40:42

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.081	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.128	02:36
Breath Volume	3.007	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.124	02:39
Breath Volume	3.007	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 05/11/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MONTS

DATE/TIME OF TEST: 05/04/2023 23:43:08

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.033	23:38
Breath Volume	1.828	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.032	23:41
Breath Volume	1.476	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 05/11/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: RAINERO

DATE/TIME OF TEST: 05/05/2023 20:21:37

Diagnostics Check	OK	20:14
Air Blank	0.000	20:14
Control Test	0.081	20:14
Air Blank	0.000	20:15
Subject Sample # 1	0.225	20:16
Breath Volume	1.363	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.227	20:20
Breath Volume	1.644	20:20
Air Blank	0.000	20:20
Control Test	0.080	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 05/11/2023 12:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: HRUZA

DATE/TIME OF TEST: 05/05/2023 22:29:57

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.080	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1	22:28	Subject Test Refused
Breath Volume	0.378	22:28	
Air Blank	0.000	22:29	
Control Test	0.081	22:29	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

Data Download Date/Time: 05/11/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREETHAM, CAMRON
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: ALEMAN GARCIA

DATE/TIME OF TEST: 05/06/2023 20:29:44

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.081	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.334	20:24
Breath Volume	1.269	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.339	20:28
Breath Volume	1.406	20:28
Air Blank	0.000	20:28
Control Test	0.080	20:29
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 05/11/2023 12:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,FRANCISCO,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINEZ,FRANCISCO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 05/09/2023 04:39:45

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.081	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1 0.117	04:29	Volume Not Met
Breath Volume	0.316	04:29	
Air Blank	0.000	04:29	
Air Blank	0.000	04:31	
Subject Sample	# 2 0.155	04:34	
Breath Volume	1.488	04:34	
Air Blank	0.000	04:34	
Air Blank	0.000	04:36	
Subject Sample	# 3 0.143	04:38	
Breath Volume	1.699	04:38	
Air Blank	0.000	04:38	
Control Test	0.079	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 05/11/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,STEPHEN,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: EDWARDS,ANTHONY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:27

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 05/09/2023 08:53:46

Diagnostics Check	OK	08:47
Air Blank	0.000	08:47
Control Test	0.081	08:48
Air Blank	0.000	08:48
Subject Sample # 1	0.000	08:49
Breath Volume	3.238	08:49
Air Blank	0.000	08:49
Air Blank	0.000	08:51
Subject Sample # 2	0.000	08:52
Breath Volume	3.230	08:52
Air Blank	0.000	08:52
Control Test	0.080	08:53
Air Blank	0.000	08:53
Diagnostics Check	OK	08:53

Data Download Date/Time: 05/11/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS, MERCEDES, E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PHILLIPS, MERCEDES
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: GADDIS

DATE/TIME OF TEST: 05/21/2023 04:35:28

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.081	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.079	04:30
Breath Volume	2.156	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.079	04:34
Breath Volume	2.671	04:34
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 06/07/2023 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CRUZ BALCAZAR

DATE/TIME OF TEST: 05/01/2023 03:19:20

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Control Test	0.080	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 05/17/2023 18:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/03/2023 22:30:40

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.079	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.089	22:25
Breath Volume	3.425	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.088	22:29
Breath Volume	3.718	22:29
Air Blank	0.000	22:29
Control Test	0.078	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 05/17/2023 18:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: TRIER

DATE/TIME OF TEST: 05/04/2023 17:27:35

Diagnostics Check	OK	17:20
Air Blank	0.000	17:21
Control Test	0.079	17:21
Air Blank	0.000	17:21
Subject Sample # 1	0.185	17:23
Breath Volume	1.753	17:23
Air Blank	0.000	17:23
Air Blank	0.000	17:25
Subject Sample # 2	0.180	17:26
Breath Volume	1.562	17:26
Air Blank	0.000	17:26
Control Test	0.078	17:26
Air Blank	0.000	17:27
Diagnostics Check	OK	17:27

Data Download Date/Time: 05/17/2023 18:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/04/2023 22:15:33

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.235	22:11
Breath Volume	2.921	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.219	22:14
Breath Volume	1.734	22:14
Air Blank	0.000	22:14
Control Test	0.077	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 05/04/2023 22:34:20

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.202	22:29
Breath Volume	1.277	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.217	22:32
Breath Volume	1.457	22:32
Air Blank	0.000	22:33
Control Test	0.078	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/05/2023 04:00:40

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.079	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.080	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: LUGAR

DATE/TIME OF TEST: 05/06/2023 00:45:58

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.078	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.084	00:41
Breath Volume	1.250	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:43
Subject Sample # 2	0.084	00:44
Breath Volume	1.511	00:44
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: VALENCIA-MARQUEZ

DATE/TIME OF TEST: 05/06/2023 01:31:47

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.060	01:27
Breath Volume	1.652	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.063	01:30
Breath Volume	1.867	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: OWSIANKA

DATE/TIME OF TEST: 05/06/2023 03:25:39

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.165	03:21
Breath Volume	2.851	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.161	03:24
Breath Volume	2.230	03:24
Air Blank	0.000	03:24
Control Test	0.078	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: CAIRE

DATE/TIME OF TEST: 05/06/2023 18:24:10

Diagnostics Check	OK	18:18
Air Blank	0.000	18:18
Control Test	0.079	18:18
Air Blank	0.000	18:19
Subject Sample # 1	0.219	18:19
Breath Volume	1.871	18:19
Air Blank	0.000	18:20
Air Blank	0.000	18:22
Subject Sample # 2	0.232	18:22
Breath Volume	3.078	18:22
Air Blank	0.000	18:23
Control Test	0.077	18:23
Air Blank	0.000	18:24
Diagnostics Check	OK	18:24

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SUNDERLAND,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: SCHROYER

DATE/TIME OF TEST: 05/07/2023 00:08:41

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.115	00:04
Breath Volume	1.796	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.117	00:07
Breath Volume	1.480	00:07
Air Blank	0.000	00:07
Control Test	0.077	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ROBERTS,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 05/07/2023 02:55:06

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.000	02:50
Breath Volume	2.367	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.000	02:53
Breath Volume	2.500	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: NORDQUIST

DATE/TIME OF TEST: 05/07/2023 07:45:36

Diagnostics Check	OK	07:39
Air Blank	0.000	07:40
Control Test	0.080	07:40
Air Blank	0.000	07:40
Subject Sample # 1	0.085	07:41
Breath Volume	3.875	07:41
Air Blank	0.000	07:41
Air Blank	0.000	07:43
Subject Sample # 2	0.084	07:44
Breath Volume	3.457	07:44
Air Blank	0.000	07:44
Control Test	0.078	07:45
Air Blank	0.000	07:45
Diagnostics Check	OK	07:45

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:21

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 05/07/2023 11:49:14

Diagnostics Check	OK	11:43
Air Blank	0.000	11:43
Control Test	0.080	11:43
Air Blank	0.000	11:44
Subject Sample # 1	0.000	11:44
Breath Volume	1.878	11:44
Air Blank	0.000	11:45
Air Blank	0.000	11:47
Subject Sample # 2	0.000	11:47
Breath Volume	1.839	11:47
Air Blank	0.000	11:48
Control Test	0.080	11:48
Air Blank	0.000	11:49
Diagnostics Check	OK	11:49

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:31

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 05/07/2023 16:02:55

Diagnostics Check	OK	15:56
Air Blank	0.000	15:57
Control Test	0.080	15:57
Air Blank	0.000	15:58
Subject Sample # 1	0.019	15:58
Breath Volume	2.128	15:58
Air Blank	0.000	15:59
Air Blank	0.000	16:01
Subject Sample # 2	0.017	16:01
Breath Volume	1.488	16:01
Air Blank	0.000	16:02
Control Test	0.079	16:02
Air Blank	0.000	16:02
Diagnostics Check	OK	16:02

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 05/08/2023 05:51:39

Diagnostics Check	OK	05:43
Air Blank	0.000	05:44
Control Test	0.079	05:44
Air Blank	0.000	05:44
Subject Sample # 1	0.142	05:46
Breath Volume	2.300	05:46
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample # 2	0.142	05:50
Breath Volume	1.484	05:50
Air Blank	0.000	05:50
Control Test	0.076	05:51
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: RAMIREZ,OFC.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: MENDOZA-VELASCO

DATE/TIME OF TEST: 05/10/2023 00:36:05

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.079	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.170	00:31
Breath Volume	1.769	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.166	00:34
Breath Volume	1.796	00:34
Air Blank	0.000	00:35
Control Test	0.078	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 05/17/2023 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LEVERIDGE,M
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 05/10/2023 04:06:18

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.029	04:01
Breath Volume	1.832	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.028	04:04
Breath Volume	2.207	04:04
Air Blank	0.000	04:05
Control Test	0.078	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 05/17/2023 20:45