

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PENA,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: RIVERA,CASEY  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 04/17/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 212641  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/07/2023 06:24:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:15 |
| Air Blank          | 0.000 | 06:15 |
| Control Test       | 0.080 | 06:16 |
| Air Blank          | 0.000 | 06:16 |
| Subject Sample # 1 | 0.228 | 06:19 |
| Breath Volume      | 1.839 | 06:19 |
| Air Blank          | 0.000 | 06:20 |
| Air Blank          | 0.000 | 06:22 |
| Subject Sample # 2 | 0.227 | 06:23 |
| Breath Volume      | 1.500 | 06:23 |
| Air Blank          | 0.000 | 06:23 |
| Control Test       | 0.079 | 06:24 |
| Air Blank          | 0.000 | 06:24 |
| Diagnostics Check  | OK    | 06:24 |

---

Data Download Date/Time: 05/16/2023 15:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORENCY,ALAIN,B  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: TILLERY,JORDAN  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 04/17/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 212641  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:57

SUBJECT'S LAST NAME: FLORENTINA

DATE/TIME OF TEST: 05/07/2023 07:31:40

---

|                   |           |       |                 |
|-------------------|-----------|-------|-----------------|
| Diagnostics Check | OK        | 07:25 |                 |
| Air Blank         | 0.000     | 07:25 |                 |
| Control Test      | 0.080     | 07:25 |                 |
| Air Blank         | 0.000     | 07:26 |                 |
| Subject Sample    | # 1 0.113 | 07:29 | Volume Not Met  |
| Breath Volume     | 0.910     | 07:29 |                 |
| Air Blank         |           | 07:29 | Improper Sample |
| Air Blank         | 0.000     | 07:30 |                 |
| Control Test      | 0.085     | 07:30 |                 |
| Air Blank         | 0.000     | 07:31 |                 |
| Diagnostics Check | OK        | 07:31 |                 |

---

Data Download Date/Time: 05/16/2023 15:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORENCY,ALAIN,B  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: TILLERY,JORDAN  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 04/17/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 212641  
DRY GAS EXPIRY DATE: 10/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:57

SUBJECT'S LAST NAME: FLORENTIA

DATE/TIME OF TEST: 05/07/2023 07:42:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:35 |
| Air Blank          | 0.000 | 07:35 |
| Control Test       | 0.078 | 07:35 |
| Air Blank          | 0.000 | 07:36 |
| Subject Sample # 1 | 0.171 | 07:36 |
| Breath Volume      | 1.914 | 07:36 |
| Air Blank          | 0.000 | 07:37 |
| Air Blank          | 0.000 | 07:39 |
| Subject Sample # 2 | 0.154 | 07:40 |
| Breath Volume      | 1.671 | 07:40 |
| Air Blank          | 0.000 | 07:41 |
| Control Test       | 0.083 | 07:41 |
| Air Blank          | 0.000 | 07:42 |
| Diagnostics Check  | OK    | 07:42 |

---

Data Download Date/Time: 05/16/2023 15:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASTOR,JOSE,F  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: RIOS,RICHARD  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1506449  
DRY GAS EXPIRY DATE: 06/10/2024  
VIOLATION CODE: DUI  
ARREST TIME: 17:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 05/27/2023 18:54:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:46 |
| Air Blank          | 0.000 | 18:46 |
| Control Test       | 0.079 | 18:46 |
| Air Blank          | 0.000 | 18:47 |
| Subject Sample # 1 | 0.274 | 18:49 |
| Breath Volume      | 1.292 | 18:49 |
| Air Blank          | 0.000 | 18:50 |
| Air Blank          | 0.000 | 18:52 |
| Subject Sample # 2 | 0.285 | 18:53 |
| Breath Volume      | 1.792 | 18:53 |
| Air Blank          | 0.000 | 18:53 |
| Control Test       | 0.077 | 18:54 |
| Air Blank          | 0.000 | 18:54 |
| Diagnostics Check  | OK    | 18:54 |

---

Data Download Date/Time: 05/31/2023 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: BATTAGLIA,P  
ARRESTING OFFICER AGENCY: North Bay Village PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SHEA

DATE/TIME OF TEST: 05/13/2023 00:20:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:12 |
| Air Blank          | 0.000 | 00:12 |
| Control Test       | 0.078 | 00:13 |
| Air Blank          | 0.000 | 00:13 |
| Subject Sample # 1 | 0.190 | 00:15 |
| Breath Volume      | 2.187 | 00:15 |
| Air Blank          | 0.000 | 00:16 |
| Air Blank          | 0.000 | 00:17 |
| Subject Sample # 2 | 0.175 | 00:18 |
| Breath Volume      | 1.589 | 00:18 |
| Air Blank          | 0.000 | 00:19 |
| Control Test       | 0.076 | 00:19 |
| Air Blank          | 0.000 | 00:19 |
| Diagnostics Check  | OK    | 00:20 |

---

Data Download Date/Time: 06/02/2023 08:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PALACIOS,L  
ARRESTING OFFICER AGENCY: North Bay Village PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 05/13/2023 05:44:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:38 |
| Air Blank          | 0.000 | 05:38 |
| Control Test       | 0.078 | 05:38 |
| Air Blank          | 0.000 | 05:39 |
| Subject Sample # 1 | 0.154 | 05:39 |
| Breath Volume      | 1.804 | 05:39 |
| Air Blank          | 0.000 | 05:40 |
| Air Blank          | 0.000 | 05:42 |
| Subject Sample # 2 | 0.150 | 05:43 |
| Breath Volume      | 2.617 | 05:43 |
| Air Blank          | 0.000 | 05:43 |
| Control Test       | 0.078 | 05:43 |
| Air Blank          | 0.000 | 05:44 |
| Diagnostics Check  | OK    | 05:44 |

---

Data Download Date/Time: 06/02/2023 08:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: BAKEWELL,Z  
ARRESTING OFFICER AGENCY: Miami Shores PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BENNETT JR

DATE/TIME OF TEST: 05/14/2023 00:27:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:20 |
| Air Blank          | 0.000 | 00:21 |
| Control Test       | 0.078 | 00:21 |
| Air Blank          | 0.000 | 00:22 |
| Subject Sample # 1 | 0.146 | 00:22 |
| Breath Volume      | 2.335 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Air Blank          | 0.000 | 00:25 |
| Subject Sample # 2 | 0.145 | 00:25 |
| Breath Volume      | 2.289 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Control Test       | 0.078 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Diagnostics Check  | OK    | 00:27 |

---

Data Download Date/Time: 06/02/2023 08:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PALACIOS,L  
ARRESTING OFFICER AGENCY: North Bay Village PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: JARAMILLO OROZCO

DATE/TIME OF TEST: 05/14/2023 02:24:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:17 |
| Air Blank          | 0.000 | 02:18 |
| Control Test       | 0.078 | 02:18 |
| Air Blank          | 0.000 | 02:19 |
| Subject Sample # 1 | 0.000 | 02:19 |
| Breath Volume      | 1.492 | 02:19 |
| Air Blank          | 0.000 | 02:20 |
| Air Blank          | 0.000 | 02:22 |
| Subject Sample # 2 | 0.000 | 02:22 |
| Breath Volume      | 2.617 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Control Test       | 0.078 | 02:23 |
| Air Blank          | 0.000 | 02:24 |
| Diagnostics Check  | OK    | 02:24 |

---

Data Download Date/Time: 06/02/2023 08:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: DELGADO,M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: CANDELA

DATE/TIME OF TEST: 05/20/2023 23:28:06

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:24 |                      |
| Air Blank         | 0.000 | 23:25 |                      |
| Control Test      | 0.077 | 23:25 |                      |
| Air Blank         | 0.000 | 23:26 |                      |
| Subject Sample    | # 1   | 23:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:26 |                      |
| Air Blank         | 0.000 | 23:27 |                      |
| Control Test      | 0.078 | 23:27 |                      |
| Air Blank         | 0.000 | 23:28 |                      |
| Diagnostics Check | OK    | 23:28 |                      |

---

Data Download Date/Time: 06/02/2023 08:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSADO,FREDDY,A  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: WILLIAMS,JASON  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: CINQUINO

DATE/TIME OF TEST: 05/01/2023 21:26:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:24 |                      |
| Air Blank         | 0.000 | 21:24 |                      |
| Control Test      | 0.082 | 21:24 |                      |
| Air Blank         | 0.000 | 21:25 |                      |
| Subject Sample    | # 1   | 21:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:25 |                      |
| Air Blank         | 0.000 | 21:25 |                      |
| Control Test      | 0.082 | 21:26 |                      |
| Air Blank         | 0.000 | 21:26 |                      |
| Diagnostics Check | OK    | 21:26 |                      |

---

Data Download Date/Time: 05/11/2023 07:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COOPER,HENRIQUE,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: COOPER,HENRIQUE  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: REYESURBINA

DATE/TIME OF TEST: 05/08/2023 02:41:45

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 02:36 |            |
| Air Blank         | 0.000     | 02:36 |            |
| Control Test      | 0.080     | 02:37 |            |
| Air Blank         | 0.000     | 02:37 |            |
| Subject Sample    | # 1 0.177 | 02:38 |            |
| Breath Volume     | 1.582     | 02:38 |            |
| Air Blank         | 0.000     | 02:38 |            |
| Air Blank         | 0.000     | 02:40 |            |
| Subject Sample    | # 2 0.195 | 02:40 |            |
| Breath Volume     | 1.281     | 02:40 |            |
| Air Blank         |           | 02:41 | RFI Detect |
| Air Blank         | 0.000     | 02:41 |            |

---

Data Download Date/Time: 05/11/2023 07:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKEWELL,ZACHARY,R  
BREATH TEST OPERATOR AGENCY: Miami Shores PD  
ARRESTING OFFICER: BAKEWELL,ZACHARY  
ARRESTING OFFICER AGENCY: Miami Shores PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/14/2023 16:55:31

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 16:52 |                      |
| Air Blank         | 0.000 | 16:53 |                      |
| Control Test      | 0.081 | 16:53 |                      |
| Air Blank         | 0.000 | 16:53 |                      |
| Subject Sample    | # 1   | 16:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 16:54 |                      |
| Air Blank         | 0.000 | 16:54 |                      |
| Control Test      | 0.081 | 16:54 |                      |
| Air Blank         | 0.000 | 16:55 |                      |
| Diagnostics Check | OK    | 16:55 |                      |

---

Data Download Date/Time: 06/02/2023 08:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANGEL LUIS, YESTER, O  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AOECH, GERMAN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 13:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: YESTER ORTIZ

DATE/TIME OF TEST: 05/14/2023 17:09:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:06 |                      |
| Air Blank         | 0.000 | 17:07 |                      |
| Control Test      | 0.081 | 17:07 |                      |
| Air Blank         | 0.000 | 17:08 |                      |
| Subject Sample    | # 1   | 17:08 | Subject Test Refused |
| Breath Volume     | 0.000 | 17:08 |                      |
| Air Blank         | 0.000 | 17:08 |                      |
| Control Test      | 0.081 | 17:09 |                      |
| Air Blank         | 0.000 | 17:09 |                      |
| Diagnostics Check | OK    | 17:09 |                      |

---

Data Download Date/Time: 06/02/2023 08:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTANA, IAN, O  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: WILLIAMS, JASON  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DIEURISTIL

DATE/TIME OF TEST: 05/15/2023 04:28:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Control Test       | 0.081 | 04:21 |
| Air Blank          | 0.000 | 04:22 |
| Subject Sample # 1 | 0.116 | 04:23 |
| Breath Volume      | 1.796 | 04:23 |
| Air Blank          | 0.000 | 04:24 |
| Air Blank          | 0.000 | 04:26 |
| Subject Sample # 2 | 0.113 | 04:26 |
| Breath Volume      | 1.675 | 04:26 |
| Air Blank          | 0.000 | 04:27 |
| Control Test       | 0.078 | 04:27 |
| Air Blank          | 0.000 | 04:28 |
| Diagnostics Check  | OK    | 04:28 |

---

Data Download Date/Time: 06/02/2023 08:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FANJUL,MAXIMO,D  
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD  
ARRESTING OFFICER: FANJUL,MAXIMO  
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ROUCO

DATE/TIME OF TEST: 05/17/2023 00:00:12

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 23:48 |                    |
| Air Blank         | 0.000 | 23:49 |                    |
| Control Test      | 0.081 | 23:49 |                    |
| Air Blank         | 0.000 | 23:50 |                    |
| Subject Sample    | # 1   | 23:53 | No Sample Provided |
| Breath Volume     | 0.000 | 23:53 |                    |
| Air Blank         | 0.000 | 23:53 |                    |
| Air Blank         | 0.000 | 23:55 |                    |
| Subject Sample    | # 2   | 23:58 | No Sample Provided |
| Breath Volume     | 0.000 | 23:58 |                    |
| Air Blank         | 0.000 | 23:59 |                    |
| Control Test      | 0.082 | 23:59 |                    |
| Air Blank         | 0.000 | 00:00 |                    |
| Diagnostics Check | OK    | 00:00 |                    |

Data Download Date/Time: 06/02/2023 08:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,CELINA,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CHERRELUS,QUINCY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 09:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:31

SUBJECT'S LAST NAME: STGILLES

DATE/TIME OF TEST: 05/18/2023 09:53:20

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 09:52 |            |
| Air Blank         | 0.000 | 09:52 |            |
| Control Test      | 0.000 | 09:52 | RFI Detect |
| Air Blank         | 0.000 | 09:53 |            |

---

Data Download Date/Time: 06/02/2023 08:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,CELINA,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CHERRELUS,QUINCEY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 09:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:31

SUBJECT'S LAST NAME: STGILLES

DATE/TIME OF TEST: 05/18/2023 10:01:13

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 09:54 |            |
| Air Blank         | 0.000     | 09:55 |            |
| Control Test      | 0.081     | 09:55 |            |
| Air Blank         | 0.000     | 09:56 |            |
| Subject Sample    | # 1 0.000 | 09:56 |            |
| Breath Volume     | 1.675     | 09:56 |            |
| Air Blank         | 0.000     | 09:57 |            |
| Air Blank         | 0.000     | 09:59 |            |
| Subject Sample    | # 2 0.000 | 09:59 |            |
| Breath Volume     | 1.601     | 09:59 |            |
| Air Blank         | 0.000     | 10:00 |            |
| Control Test      | 0.081     | 10:00 |            |
| Air Blank         |           | 10:00 | RFI Detect |
| Air Blank         | 0.000     | 10:01 |            |

---

Data Download Date/Time: 06/02/2023 08:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATTAGLIA,PAUL,M  
BREATH TEST OPERATOR AGENCY: North Bay Village PD  
ARRESTING OFFICER: RODRIGUEZ,JASON  
ARRESTING OFFICER AGENCY: North Bay Village PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 17:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: MANTILLA

DATE/TIME OF TEST: 05/26/2023 18:44:04

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:40 |                      |
| Air Blank         | 0.000 | 18:41 |                      |
| Control Test      | 0.079 | 18:41 |                      |
| Air Blank         | 0.000 | 18:42 |                      |
| Subject Sample    | # 1   | 18:42 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:42 |                      |
| Air Blank         | 0.000 | 18:43 |                      |
| Control Test      | 0.079 | 18:43 |                      |
| Air Blank         | 0.000 | 18:43 |                      |
| Diagnostics Check | OK    | 18:44 |                      |

---

Data Download Date/Time: 06/02/2023 08:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PERMUY-ESTRADA

DATE/TIME OF TEST: 05/13/2023 02:36:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Control Test       | 0.078 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Subject Sample # 1 | 0.213 | 02:31 |
| Breath Volume      | 2.566 | 02:31 |
| Air Blank          | 0.000 | 02:32 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 2 | 0.209 | 02:34 |
| Breath Volume      | 3.449 | 02:34 |
| Air Blank          | 0.000 | 02:35 |
| Control Test       | 0.079 | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Diagnostics Check  | OK    | 02:36 |

---

Data Download Date/Time: 06/02/2023 08:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTANA,JORGE,F  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: AFANADOR

DATE/TIME OF TEST: 05/26/2023 01:09:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:02 |
| Air Blank          | 0.000 | 01:02 |
| Control Test       | 0.078 | 01:03 |
| Air Blank          | 0.000 | 01:03 |
| Subject Sample # 1 | 0.189 | 01:04 |
| Breath Volume      | 1.625 | 01:04 |
| Air Blank          | 0.000 | 01:05 |
| Air Blank          | 0.000 | 01:06 |
| Subject Sample # 2 | 0.177 | 01:07 |
| Breath Volume      | 2.929 | 01:07 |
| Air Blank          | 0.000 | 01:08 |
| Control Test       | 0.078 | 01:08 |
| Air Blank          | 0.000 | 01:09 |
| Diagnostics Check  | OK    | 01:09 |

---

Data Download Date/Time: 06/02/2023 08:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: ORUE

DATE/TIME OF TEST: 05/27/2023 06:46:56

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 06:45 |                           |
| Air Blank         | 0.000 | 06:46 |                           |
| Control Test      | 0.000 | 06:46 | Control Outside Tolerance |
| Air Blank         | 0.000 | 06:46 |                           |

---

Data Download Date/Time: 06/02/2023 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: ORUE

DATE/TIME OF TEST: 05/27/2023 06:52:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 06:49 |                      |
| Air Blank         | 0.000 | 06:50 |                      |
| Control Test      | 0.078 | 06:50 |                      |
| Air Blank         | 0.000 | 06:51 |                      |
| Subject Sample    | # 1   | 06:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 06:51 |                      |
| Air Blank         | 0.000 | 06:51 |                      |
| Control Test      | 0.079 | 06:52 |                      |
| Air Blank         | 0.000 | 06:52 |                      |
| Diagnostics Check | OK    | 06:52 |                      |

---

Data Download Date/Time: 06/02/2023 08:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: GOMEZ-CORDOVA

DATE/TIME OF TEST: 05/28/2023 03:23:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.079 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Subject Sample # 1 | 0.089 | 03:18 |
| Breath Volume      | 1.289 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 2 | 0.089 | 03:21 |
| Breath Volume      | 1.734 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Control Test       | 0.078 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Diagnostics Check  | OK    | 03:23 |

---

Data Download Date/Time: 06/02/2023 08:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COOPER,HENRIQUE,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: COOPER,HENRIQUE  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A2  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CARDONA-PAVON

DATE/TIME OF TEST: 05/14/2023 01:30:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:24 |
| Air Blank          | 0.000 | 01:24 |
| Control Test       | 0.080 | 01:24 |
| Air Blank          | 0.000 | 01:25 |
| Subject Sample # 1 | 0.198 | 01:25 |
| Breath Volume      | 1.734 | 01:25 |
| Air Blank          | 0.000 | 01:26 |
| Air Blank          | 0.000 | 01:28 |
| Subject Sample # 2 | 0.199 | 01:28 |
| Breath Volume      | 1.175 | 01:28 |
| Air Blank          | 0.000 | 01:29 |
| Control Test       | 0.079 | 01:30 |
| Air Blank          | 0.000 | 01:30 |
| Diagnostics Check  | OK    | 01:30 |

---

Data Download Date/Time: 06/16/2023 09:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COOPER,HENRIQUE,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: COOPER,HENRIQUE  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A2  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: MENDOZA-CARRASCO

DATE/TIME OF TEST: 05/15/2023 03:43:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.080 | 03:38 |
| Air Blank          | 0.000 | 03:38 |
| Subject Sample # 1 | 0.099 | 03:39 |
| Breath Volume      | 2.074 | 03:39 |
| Air Blank          | 0.000 | 03:40 |
| Air Blank          | 0.000 | 03:41 |
| Subject Sample # 2 | 0.089 | 03:42 |
| Breath Volume      | 1.351 | 03:42 |
| Air Blank          | 0.000 | 03:42 |
| Control Test       | 0.079 | 03:43 |
| Air Blank          | 0.000 | 03:43 |
| Diagnostics Check  | OK    | 03:43 |

---

Data Download Date/Time: 06/16/2023 09:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PINEDA,MIGUEL,A  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: JACKSON,D  
ARRESTING OFFICER AGENCY: North Miami PD  
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A2  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SEAC

DATE/TIME OF TEST: 05/18/2023 04:00:26

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:48 |                    |
| Air Blank         | 0.000 | 03:48 |                    |
| Control Test      | 0.081 | 03:49 |                    |
| Air Blank         | 0.000 | 03:49 |                    |
| Subject Sample    | # 1   | 03:52 | No Sample Provided |
| Breath Volume     | 0.000 | 03:52 |                    |
| Air Blank         | 0.000 | 03:53 |                    |
| Air Blank         | 0.000 | 03:55 |                    |
| Subject Sample    | # 2   | 0.185 | 03:56              |
| Breath Volume     | 1.296 | 03:56 |                    |
| Air Blank         | 0.000 | 03:56 |                    |
| Air Blank         | 0.000 | 03:58 |                    |
| Subject Sample    | # 3   | 0.194 | 03:58              |
| Breath Volume     | 1.335 | 03:58 |                    |
| Air Blank         | 0.000 | 03:59 |                    |
| Control Test      | 0.081 | 03:59 |                    |
| Air Blank         | 0.000 | 04:00 |                    |
| Diagnostics Check | OK    | 04:00 |                    |

---

Data Download Date/Time: 06/16/2023 09:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAVES,JULIAN,  
BREATH TEST OPERATOR AGENCY: North Miami PD  
ARRESTING OFFICER: SMITH,PATRICK  
ARRESTING OFFICER AGENCY: North Miami PD  
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A2  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: MONESTIME

DATE/TIME OF TEST: 05/19/2023 05:23:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:17 |
| Air Blank          | 0.000 | 05:17 |
| Control Test       | 0.080 | 05:18 |
| Air Blank          | 0.000 | 05:18 |
| Subject Sample # 1 | 0.114 | 05:18 |
| Breath Volume      | 2.750 | 05:18 |
| Air Blank          | 0.000 | 05:19 |
| Air Blank          | 0.000 | 05:21 |
| Subject Sample # 2 | 0.101 | 05:21 |
| Breath Volume      | 2.125 | 05:21 |
| Air Blank          | 0.000 | 05:22 |
| Control Test       | 0.080 | 05:22 |
| Air Blank          | 0.000 | 05:23 |
| Diagnostics Check  | OK    | 05:23 |

---

Data Download Date/Time: 06/16/2023 09:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAVES,JULIAN,  
BREATH TEST OPERATOR AGENCY: North Miami PD  
ARRESTING OFFICER: CHAVES,JULIAN  
ARRESTING OFFICER AGENCY: North Miami PD  
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A2  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PONCE-APORTELA

DATE/TIME OF TEST: 05/21/2023 01:56:10

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:53 |                      |
| Air Blank         | 0.000 | 01:53 |                      |
| Control Test      | 0.080 | 01:53 |                      |
| Air Blank         | 0.000 | 01:54 |                      |
| Subject Sample    | # 1   | 01:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:54 |                      |
| Air Blank         | 0.000 | 01:55 |                      |
| Control Test      | 0.080 | 01:55 |                      |
| Air Blank         | 0.000 | 01:56 |                      |
| Diagnostics Check | OK    | 01:56 |                      |

---

Data Download Date/Time: 06/16/2023 09:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COOPER,HENRIQUE,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: COOPER,HENRIQUE  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A2  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: OROZCO-RAMIREZ

DATE/TIME OF TEST: 05/21/2023 04:37:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:31 |
| Air Blank          | 0.000 | 04:31 |
| Control Test       | 0.081 | 04:32 |
| Air Blank          | 0.000 | 04:32 |
| Subject Sample # 1 | 0.205 | 04:33 |
| Breath Volume      | 2.300 | 04:33 |
| Air Blank          | 0.000 | 04:33 |
| Air Blank          | 0.000 | 04:35 |
| Subject Sample # 2 | 0.185 | 04:35 |
| Breath Volume      | 2.410 | 04:35 |
| Air Blank          | 0.000 | 04:36 |
| Control Test       | 0.079 | 04:36 |
| Air Blank          | 0.000 | 04:37 |
| Diagnostics Check  | OK    | 04:37 |

---

Data Download Date/Time: 06/16/2023 09:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ODUM,ICE,D  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: ODUM,ICE  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A2  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CERMENO-ALVAREZ

DATE/TIME OF TEST: 05/22/2023 01:37:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.080 | 01:31 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 1 | 0.262 | 01:32 |
| Breath Volume      | 3.050 | 01:32 |
| Air Blank          | 0.000 | 01:33 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 2 | 0.255 | 01:35 |
| Breath Volume      | 1.839 | 01:35 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.078 | 01:36 |
| Air Blank          | 0.000 | 01:36 |
| Diagnostics Check  | OK    | 01:37 |

---

Data Download Date/Time: 06/16/2023 09:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COOPER,HENRIQUE,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: COOPER,HENRIQUE  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A2  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: AYALA-AYALA

DATE/TIME OF TEST: 05/29/2023 03:32:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:23 |
| Air Blank          | 0.000 | 03:23 |
| Control Test       | 0.081 | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Subject Sample # 1 | 0.170 | 03:25 |
| Breath Volume      | 1.585 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Air Blank          | 0.000 | 03:28 |
| Subject Sample # 2 | 0.173 | 03:30 |
| Breath Volume      | 2.640 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.077 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Diagnostics Check  | OK    | 03:32 |

---

Data Download Date/Time: 06/16/2023 09:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEWIS,TIMOTHY,B

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: LEWIS,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 2402328118

DRY GAS EXPIRY DATE: 01/20/2025

VIOLATION CODE: DUI

ARREST TIME: 12:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: CRIDISO

DATE/TIME OF TEST: 05/11/2023 13:11:14

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 12:59 |                |
| Air Blank         | 0.000     | 12:59 |                |
| Control Test      | 0.080     | 13:00 |                |
| Air Blank         | 0.000     | 13:00 |                |
| Subject Sample    | # 1 0.308 | 13:03 | Volume Not Met |
| Breath Volume     | 0.097     | 13:03 |                |
| Air Blank         | 0.000     | 13:04 |                |
| Air Blank         | 0.000     | 13:06 |                |
| Subject Sample    | # 2 0.322 | 13:09 | Volume Not Met |
| Breath Volume     | 0.035     | 13:09 |                |
| Air Blank         | 0.000     | 13:10 |                |
| Control Test      | 0.081     | 13:10 |                |
| Air Blank         | 0.000     | 13:11 |                |
| Diagnostics Check | OK        | 13:11 |                |

---

Data Download Date/Time: 06/20/2023 16:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LEWIS,TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 12:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: CRIDISO

DATE/TIME OF TEST: 05/11/2023 13:19:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 13:14 |                      |
| Air Blank         | 0.000 | 13:15 |                      |
| Control Test      | 0.079 | 13:15 |                      |
| Air Blank         | 0.000 | 13:15 |                      |
| Subject Sample    | # 1   | 13:17 | Subject Test Refused |
| Breath Volume     | 0.332 | 13:17 |                      |
| Air Blank         | 0.000 | 13:18 |                      |
| Control Test      | 0.079 | 13:18 |                      |
| Air Blank         | 0.000 | 13:19 |                      |
| Diagnostics Check | OK    | 13:19 |                      |

---

Data Download Date/Time: 06/20/2023 16:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUSBANDS, TYLER, J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HUSBANDS, TYLER  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 05/01/2023 21:02:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:56 |
| Air Blank          | 0.000 | 20:56 |
| Control Test       | 0.081 | 20:56 |
| Air Blank          | 0.000 | 20:57 |
| Subject Sample # 1 | 0.098 | 20:57 |
| Breath Volume      | 1.554 | 20:57 |
| Air Blank          | 0.000 | 20:58 |
| Air Blank          | 0.000 | 21:00 |
| Subject Sample # 2 | 0.105 | 21:01 |
| Breath Volume      | 2.074 | 21:01 |
| Air Blank          | 0.000 | 21:02 |
| Control Test       | 0.081 | 21:02 |
| Air Blank          | 0.000 | 21:02 |
| Diagnostics Check  | OK    | 21:02 |

---

Data Download Date/Time: 05/10/2023 12:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUSBANDS, TYLER, J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HUSBANDS, TYLER  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: KUK

DATE/TIME OF TEST: 05/02/2023 02:04:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:58 |
| Air Blank          | 0.000 | 01:58 |
| Control Test       | 0.081 | 01:58 |
| Air Blank          | 0.000 | 01:59 |
| Subject Sample # 1 | 0.235 | 01:59 |
| Breath Volume      | 1.398 | 01:59 |
| Air Blank          | 0.000 | 02:00 |
| Air Blank          | 0.000 | 02:02 |
| Subject Sample # 2 | 0.238 | 02:02 |
| Breath Volume      | 1.332 | 02:02 |
| Air Blank          | 0.000 | 02:03 |
| Control Test       | 0.081 | 02:03 |
| Air Blank          | 0.000 | 02:04 |
| Diagnostics Check  | OK    | 02:04 |

---

Data Download Date/Time: 05/10/2023 12:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,OFC B,  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: HOPPE,OFC A  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 12:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: COWLES

DATE/TIME OF TEST: 05/15/2023 01:00:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:52 |
| Air Blank          | 0.000 | 00:53 |
| Control Test       | 0.080 | 00:53 |
| Air Blank          | 0.000 | 00:54 |
| Subject Sample # 1 | 0.143 | 00:56 |
| Breath Volume      | 1.773 | 00:56 |
| Air Blank          | 0.000 | 00:56 |
| Air Blank          | 0.000 | 00:58 |
| Subject Sample # 2 | 0.142 | 00:59 |
| Breath Volume      | 1.410 | 00:59 |
| Air Blank          | 0.000 | 00:59 |
| Control Test       | 0.080 | 01:00 |
| Air Blank          | 0.000 | 01:00 |
| Diagnostics Check  | OK    | 01:00 |

---

Data Download Date/Time: 06/14/2023 10:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUSBANDS, TYLER, J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HUSBANDS, TYLER  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 05/16/2023 00:44:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:41 |                      |
| Air Blank         | 0.000 | 00:41 |                      |
| Control Test      | 0.080 | 00:42 |                      |
| Air Blank         | 0.000 | 00:42 |                      |
| Subject Sample    | # 1   | 00:42 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:42 |                      |
| Air Blank         | 0.000 | 00:43 |                      |
| Control Test      | 0.081 | 00:43 |                      |
| Air Blank         | 0.000 | 00:44 |                      |
| Diagnostics Check | OK    | 00:44 |                      |

---

Data Download Date/Time: 06/14/2023 10:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MITCHELL,ALEXANDRA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 17:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: AMERES

DATE/TIME OF TEST: 05/17/2023 18:14:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:07 |
| Air Blank          | 0.000 | 18:08 |
| Control Test       | 0.081 | 18:08 |
| Air Blank          | 0.000 | 18:09 |
| Subject Sample # 1 | 0.000 | 18:09 |
| Breath Volume      | 1.265 | 18:09 |
| Air Blank          | 0.000 | 18:10 |
| Air Blank          | 0.000 | 18:12 |
| Subject Sample # 2 | 0.000 | 18:12 |
| Breath Volume      | 1.421 | 18:12 |
| Air Blank          | 0.000 | 18:13 |
| Control Test       | 0.080 | 18:13 |
| Air Blank          | 0.000 | 18:14 |
| Diagnostics Check  | OK    | 18:14 |

---

Data Download Date/Time: 06/14/2023 10:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AHART,DEMITRI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 05/22/2023 04:24:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:16 |
| Air Blank          | 0.000 | 04:17 |
| Control Test       | 0.081 | 04:17 |
| Air Blank          | 0.000 | 04:18 |
| Subject Sample # 1 | 0.201 | 04:18 |
| Breath Volume      | 1.343 | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Air Blank          | 0.000 | 04:21 |
| Subject Sample # 2 | 0.194 | 04:22 |
| Breath Volume      | 1.453 | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Control Test       | 0.080 | 04:23 |
| Air Blank          | 0.000 | 04:23 |
| Diagnostics Check  | OK    | 04:24 |

---

Data Download Date/Time: 06/14/2023 10:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MITCHELL,ALEXANDRA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 05/31/2023 20:27:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:21 |
| Air Blank          | 0.000 | 20:21 |
| Control Test       | 0.081 | 20:21 |
| Air Blank          | 0.000 | 20:22 |
| Subject Sample # 1 | 0.165 | 20:22 |
| Breath Volume      | 2.457 | 20:22 |
| Air Blank          | 0.000 | 20:23 |
| Air Blank          | 0.000 | 20:25 |
| Subject Sample # 2 | 0.151 | 20:25 |
| Breath Volume      | 1.425 | 20:25 |
| Air Blank          | 0.000 | 20:26 |
| Control Test       | 0.081 | 20:26 |
| Air Blank          | 0.000 | 20:27 |
| Diagnostics Check  | OK    | 20:27 |

---

Data Download Date/Time: 06/14/2023 10:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ZSIDO,DANIELLE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/01/2023 00:48:30

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:45 |                      |
| Air Blank         | 0.000 | 00:45 |                      |
| Control Test      | 0.080 | 00:45 |                      |
| Air Blank         | 0.000 | 00:46 |                      |
| Subject Sample    | # 1   | 00:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:47 |                      |
| Air Blank         | 0.000 | 00:47 |                      |
| Control Test      | 0.081 | 00:47 |                      |
| Air Blank         | 0.000 | 00:48 |                      |
| Diagnostics Check | OK    | 00:48 |                      |

---

Data Download Date/Time: 05/10/2023 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUNDH,DEP,K  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SKALKO,DEP  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 05/01/2023 01:38:35

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:35 |                      |
| Air Blank         | 0.000 | 01:36 |                      |
| Control Test      | 0.081 | 01:36 |                      |
| Air Blank         | 0.000 | 01:36 |                      |
| Subject Sample    | # 1   | 01:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:37 |                      |
| Air Blank         | 0.000 | 01:37 |                      |
| Control Test      | 0.081 | 01:37 |                      |
| Air Blank         | 0.000 | 01:38 |                      |
| Diagnostics Check | OK    | 01:38 |                      |

---

Data Download Date/Time: 05/10/2023 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,BRYAN,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LEIB,MELANIE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: FLAHERTY

DATE/TIME OF TEST: 05/01/2023 06:51:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:45 |
| Air Blank          | 0.000 | 06:45 |
| Control Test       | 0.080 | 06:46 |
| Air Blank          | 0.000 | 06:46 |
| Subject Sample # 1 | 0.000 | 06:47 |
| Breath Volume      | 3.558 | 06:47 |
| Air Blank          | 0.000 | 06:47 |
| Air Blank          | 0.000 | 06:49 |
| Subject Sample # 2 | 0.000 | 06:50 |
| Breath Volume      | 3.378 | 06:50 |
| Air Blank          | 0.000 | 06:50 |
| Control Test       | 0.080 | 06:50 |
| Air Blank          | 0.000 | 06:51 |
| Diagnostics Check  | OK    | 06:51 |

---

Data Download Date/Time: 05/10/2023 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,CRISTIAN,D  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: DONOHUE,STEVEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/03/2023 00:27:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:18 |
| Air Blank          | 0.000 | 00:18 |
| Control Test       | 0.082 | 00:18 |
| Air Blank          | 0.000 | 00:19 |
| Subject Sample # 1 | 0.283 | 00:22 |
| Breath Volume      | 2.906 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Air Blank          | 0.000 | 00:24 |
| Subject Sample # 2 | 0.264 | 00:25 |
| Breath Volume      | 2.617 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Control Test       | 0.079 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Diagnostics Check  | OK    | 00:27 |

---

Data Download Date/Time: 05/10/2023 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: NEWTON,MICHAEL  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: UNGER

DATE/TIME OF TEST: 05/03/2023 02:19:12

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:16 |                      |
| Air Blank         | 0.000 | 02:16 |                      |
| Control Test      | 0.080 | 02:17 |                      |
| Air Blank         | 0.000 | 02:17 |                      |
| Subject Sample    | # 1   | 02:17 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:17 |                      |
| Air Blank         | 0.000 | 02:18 |                      |
| Control Test      | 0.081 | 02:18 |                      |
| Air Blank         | 0.000 | 02:19 |                      |
| Diagnostics Check | OK    | 02:19 |                      |

---

Data Download Date/Time: 05/10/2023 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CHURCHILL,CAMERON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 17:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: SPAETH

DATE/TIME OF TEST: 05/03/2023 18:29:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:23 |
| Air Blank          | 0.000 | 18:23 |
| Control Test       | 0.081 | 18:24 |
| Air Blank          | 0.000 | 18:24 |
| Subject Sample # 1 | 0.000 | 18:24 |
| Breath Volume      | 2.320 | 18:24 |
| Air Blank          | 0.000 | 18:25 |
| Air Blank          | 0.000 | 18:27 |
| Subject Sample # 2 | 0.000 | 18:27 |
| Breath Volume      | 2.164 | 18:27 |
| Air Blank          | 0.000 | 18:28 |
| Control Test       | 0.081 | 18:28 |
| Air Blank          | 0.000 | 18:29 |
| Diagnostics Check  | OK    | 18:29 |

---

Data Download Date/Time: 05/10/2023 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: YEARGIN, DEP GENESIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: GRIGGS

DATE/TIME OF TEST: 05/04/2023 02:00:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:54 |
| Air Blank          | 0.000 | 01:54 |
| Control Test       | 0.080 | 01:55 |
| Air Blank          | 0.000 | 01:55 |
| Subject Sample # 1 | 0.162 | 01:56 |
| Breath Volume      | 3.332 | 01:56 |
| Air Blank          | 0.000 | 01:56 |
| Air Blank          | 0.000 | 01:58 |
| Subject Sample # 2 | 0.158 | 01:59 |
| Breath Volume      | 2.132 | 01:59 |
| Air Blank          | 0.000 | 01:59 |
| Control Test       | 0.080 | 02:00 |
| Air Blank          | 0.000 | 02:00 |
| Diagnostics Check  | OK    | 02:00 |

---

Data Download Date/Time: 05/10/2023 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GAMBEL,Z,E  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GAMBEL,Z  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 05/04/2023 03:39:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:36 |                      |
| Air Blank         | 0.000 | 03:37 |                      |
| Control Test      | 0.081 | 03:37 |                      |
| Air Blank         | 0.000 | 03:38 |                      |
| Subject Sample    | # 1   | 03:38 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:38 |                      |
| Air Blank         | 0.000 | 03:38 |                      |
| Control Test      | 0.080 | 03:39 |                      |
| Air Blank         | 0.000 | 03:39 |                      |
| Diagnostics Check | OK    | 03:39 |                      |

---

Data Download Date/Time: 05/10/2023 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: YEARGIN,DEP GENESIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/04/2023 04:52:55

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:49 |                      |
| Air Blank         | 0.000 | 04:50 |                      |
| Control Test      | 0.080 | 04:50 |                      |
| Air Blank         | 0.000 | 04:51 |                      |
| Subject Sample    | # 1   | 04:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:51 |                      |
| Air Blank         | 0.000 | 04:52 |                      |
| Control Test      | 0.080 | 04:52 |                      |
| Air Blank         | 0.000 | 04:52 |                      |
| Diagnostics Check | OK    | 04:52 |                      |

---

Data Download Date/Time: 05/10/2023 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY W.,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILTSE,DEPUTY W.  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 05/04/2023 20:43:47

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:40 |                      |
| Air Blank         | 0.000 | 20:41 |                      |
| Control Test      | 0.080 | 20:41 |                      |
| Air Blank         | 0.000 | 20:42 |                      |
| Subject Sample    | # 1   | 20:42 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:42 |                      |
| Air Blank         | 0.000 | 20:42 |                      |
| Control Test      | 0.081 | 20:43 |                      |
| Air Blank         | 0.000 | 20:43 |                      |
| Diagnostics Check | OK    | 20:43 |                      |

---

Data Download Date/Time: 05/10/2023 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: DAVIS,B  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 05/05/2023 02:22:06

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:19 |                      |
| Air Blank         | 0.000 | 02:19 |                      |
| Control Test      | 0.080 | 02:19 |                      |
| Air Blank         | 0.000 | 02:20 |                      |
| Subject Sample    | # 1   | 02:20 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:20 |                      |
| Air Blank         | 0.000 | 02:21 |                      |
| Control Test      | 0.081 | 02:21 |                      |
| Air Blank         | 0.000 | 02:22 |                      |
| Diagnostics Check | OK    | 02:22 |                      |

---

Data Download Date/Time: 05/10/2023 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: YEARGIN,DEP GENESIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LOMBERTO

DATE/TIME OF TEST: 05/05/2023 03:14:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:08 |
| Air Blank          | 0.000 | 03:08 |
| Control Test       | 0.080 | 03:08 |
| Air Blank          | 0.000 | 03:09 |
| Subject Sample # 1 | 0.000 | 03:09 |
| Breath Volume      | 1.824 | 03:09 |
| Air Blank          | 0.000 | 03:10 |
| Air Blank          | 0.000 | 03:12 |
| Subject Sample # 2 | 0.000 | 03:13 |
| Breath Volume      | 1.851 | 03:13 |
| Air Blank          | 0.000 | 03:13 |
| Control Test       | 0.080 | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Diagnostics Check  | OK    | 03:14 |

---

Data Download Date/Time: 05/10/2023 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MORRIS,BRIDGETT  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 05/06/2023 01:40:58

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:36 |                      |
| Air Blank         | 0.000 | 01:37 |                      |
| Control Test      | 0.081 | 01:37 |                      |
| Air Blank         | 0.000 | 01:37 |                      |
| Subject Sample    | # 1   | 01:39 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:39 |                      |
| Air Blank         | 0.000 | 01:40 |                      |
| Control Test      | 0.081 | 01:40 |                      |
| Air Blank         | 0.000 | 01:40 |                      |
| Diagnostics Check | OK    | 01:40 |                      |

---

Data Download Date/Time: 05/10/2023 11:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANEY,DAMON,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LANEY,DAMON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: RAMIREZ CAMPOS

DATE/TIME OF TEST: 05/06/2023 03:19:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Control Test       | 0.081 | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Subject Sample # 1 | 0.190 | 03:15 |
| Breath Volume      | 2.492 | 03:15 |
| Air Blank          | 0.000 | 03:15 |
| Air Blank          | 0.000 | 03:17 |
| Subject Sample # 2 | 0.188 | 03:18 |
| Breath Volume      | 1.335 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Control Test       | 0.080 | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Diagnostics Check  | OK    | 03:19 |

---

Data Download Date/Time: 05/10/2023 11:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KATSAREAS,NIKOLAOS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CARLSON,S  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: KANIA

DATE/TIME OF TEST: 05/06/2023 08:00:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:53 |
| Air Blank          | 0.000 | 07:54 |
| Control Test       | 0.081 | 07:54 |
| Air Blank          | 0.000 | 07:55 |
| Subject Sample # 1 | 0.162 | 07:55 |
| Breath Volume      | 1.445 | 07:55 |
| Air Blank          | 0.000 | 07:56 |
| Air Blank          | 0.000 | 07:58 |
| Subject Sample # 2 | 0.181 | 07:58 |
| Breath Volume      | 1.320 | 07:58 |
| Air Blank          | 0.000 | 07:59 |
| Control Test       | 0.080 | 07:59 |
| Air Blank          | 0.000 | 08:00 |
| Diagnostics Check  | OK    | 08:00 |

---

Data Download Date/Time: 05/10/2023 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: DONOHUE,S  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: PRUITT

DATE/TIME OF TEST: 05/07/2023 01:30:50

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:28 |                      |
| Air Blank         | 0.000 | 01:28 |                      |
| Control Test      | 0.081 | 01:28 |                      |
| Air Blank         | 0.000 | 01:29 |                      |
| Subject Sample    | # 1   | 01:29 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:29 |                      |
| Air Blank         | 0.000 | 01:29 |                      |
| Control Test      | 0.081 | 01:30 |                      |
| Air Blank         | 0.000 | 01:30 |                      |
| Diagnostics Check | OK    | 01:30 |                      |

Data Download Date/Time: 05/10/2023 11:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MITCHELL,ALEXANDRA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: MARTINEZ CORONA

DATE/TIME OF TEST: 05/07/2023 04:20:35

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:17 |                      |
| Air Blank         | 0.000 | 04:18 |                      |
| Control Test      | 0.081 | 04:18 |                      |
| Air Blank         | 0.000 | 04:18 |                      |
| Subject Sample    | # 1   | 04:19 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:19 |                      |
| Air Blank         | 0.000 | 04:19 |                      |
| Control Test      | 0.081 | 04:19 |                      |
| Air Blank         | 0.000 | 04:20 |                      |
| Diagnostics Check | OK    | 04:20 |                      |

---

Data Download Date/Time: 05/10/2023 11:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: YEARGIN, DEP GENESIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: WEISENBERG

DATE/TIME OF TEST: 05/07/2023 06:12:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:05 |
| Air Blank          | 0.000 | 06:06 |
| Control Test       | 0.081 | 06:06 |
| Air Blank          | 0.000 | 06:07 |
| Subject Sample # 1 | 0.000 | 06:08 |
| Breath Volume      | 1.480 | 06:08 |
| Air Blank          | 0.000 | 06:08 |
| Air Blank          | 0.000 | 06:10 |
| Subject Sample # 2 | 0.000 | 06:10 |
| Breath Volume      | 1.789 | 06:10 |
| Air Blank          | 0.000 | 06:11 |
| Control Test       | 0.081 | 06:11 |
| Air Blank          | 0.000 | 06:12 |
| Diagnostics Check  | OK    | 06:12 |

---

Data Download Date/Time: 05/10/2023 11:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELSHANS,DEAN,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TORRES,ALVIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: B620108648080

DATE/TIME OF TEST: 05/07/2023 21:46:18

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 21:40 |            |
| Air Blank         | 0.000     | 21:40 |            |
| Control Test      | 0.080     | 21:40 |            |
| Air Blank         | 0.000     | 21:41 |            |
| Subject Sample    | # 1 0.229 | 21:41 |            |
| Breath Volume     | 2.214     | 21:41 |            |
| Air Blank         | 0.000     | 21:42 |            |
| Air Blank         | 0.000     | 21:44 |            |
| Subject Sample    | # 2 0.223 | 21:44 |            |
| Breath Volume     | 1.843     | 21:44 |            |
| Air Blank         | 0.000     | 21:45 |            |
| Control Test      | 0.081     | 21:45 |            |
| Air Blank         |           | 21:45 | RFI Detect |
| Air Blank         | 0.000     | 21:46 |            |

---

Data Download Date/Time: 05/10/2023 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELSHANS,DEAN,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: NEWTON,MICHAEL  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: KIMMEL

DATE/TIME OF TEST: 05/07/2023 22:35:27

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:26 |                      |
| Air Blank         | 0.000 | 22:27 |                      |
| Control Test      | 0.081 | 22:27 |                      |
| Air Blank         | 0.000 | 22:28 |                      |
| Subject Sample    | # 1   | 22:31 | No Sample Provided   |
| Breath Volume     | 0.000 | 22:31 |                      |
| Air Blank         | 0.000 | 22:31 |                      |
| Air Blank         | 0.000 | 22:33 |                      |
| Subject Sample    | # 2   | 22:34 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:34 |                      |
| Air Blank         | 0.000 | 22:34 |                      |
| Control Test      | 0.081 | 22:34 |                      |
| Air Blank         | 0.000 | 22:35 |                      |
| Diagnostics Check | OK    | 22:35 |                      |

Data Download Date/Time: 05/10/2023 11:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELSHANS,DEAN,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HICKS,JASON  
ARRESTING OFFICER AGENCY: Indian Shores PD  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: VOYTEK

DATE/TIME OF TEST: 05/07/2023 23:50:05

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:45 |                      |
| Air Blank         | 0.000 | 23:45 |                      |
| Control Test      | 0.081 | 23:45 |                      |
| Air Blank         | 0.000 | 23:46 |                      |
| Subject Sample    | # 1   | 23:48 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:48 |                      |
| Air Blank         | 0.000 | 23:49 |                      |
| Control Test      | 0.081 | 23:49 |                      |
| Air Blank         | 0.000 | 23:50 |                      |
| Diagnostics Check | OK    | 23:50 |                      |

---

Data Download Date/Time: 05/10/2023 11:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUNDH,DEP,K  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LUNDH,DEP  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: LECA RODRIGUEZ

DATE/TIME OF TEST: 05/08/2023 04:48:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:40 |
| Air Blank          | 0.000 | 04:41 |
| Control Test       | 0.081 | 04:41 |
| Air Blank          | 0.000 | 04:42 |
| Subject Sample # 1 | 0.216 | 04:43 |
| Breath Volume      | 2.175 | 04:43 |
| Air Blank          | 0.000 | 04:44 |
| Air Blank          | 0.000 | 04:45 |
| Subject Sample # 2 | 0.209 | 04:46 |
| Breath Volume      | 1.613 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Control Test       | 0.079 | 04:47 |
| Air Blank          | 0.000 | 04:48 |
| Diagnostics Check  | OK    | 04:48 |

---

Data Download Date/Time: 05/10/2023 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,BRYAN,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MILLER,HOWARD  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 15:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 05/08/2023 16:23:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 16:20 |                      |
| Air Blank         | 0.000 | 16:21 |                      |
| Control Test      | 0.081 | 16:21 |                      |
| Air Blank         | 0.000 | 16:21 |                      |
| Subject Sample    | # 1   | 16:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 16:22 |                      |
| Air Blank         | 0.000 | 16:22 |                      |
| Control Test      | 0.080 | 16:22 |                      |
| Air Blank         | 0.000 | 16:23 |                      |
| Diagnostics Check | OK    | 16:23 |                      |

---

Data Download Date/Time: 05/10/2023 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CAPPIELLO,S  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: STOWERS

DATE/TIME OF TEST: 05/12/2023 00:05:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:59 |
| Air Blank          | 0.000 | 23:59 |
| Control Test       | 0.081 | 23:59 |
| Air Blank          | 0.000 | 00:00 |
| Subject Sample # 1 | 0.115 | 00:01 |
| Breath Volume      | 1.574 | 00:01 |
| Air Blank          | 0.000 | 00:01 |
| Air Blank          | 0.000 | 00:03 |
| Subject Sample # 2 | 0.113 | 00:04 |
| Breath Volume      | 1.531 | 00:04 |
| Air Blank          | 0.000 | 00:04 |
| Control Test       | 0.080 | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Diagnostics Check  | OK    | 00:05 |

---

Data Download Date/Time: 06/14/2023 09:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: NEWTON,M  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: PALESKY

DATE/TIME OF TEST: 05/12/2023 01:12:46

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:09 |                      |
| Air Blank         | 0.000 | 01:10 |                      |
| Control Test      | 0.081 | 01:10 |                      |
| Air Blank         | 0.000 | 01:11 |                      |
| Subject Sample    | # 1   | 01:11 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:11 |                      |
| Air Blank         | 0.000 | 01:11 |                      |
| Control Test      | 0.081 | 01:12 |                      |
| Air Blank         | 0.000 | 01:12 |                      |
| Diagnostics Check | OK    | 01:12 |                      |

Data Download Date/Time: 06/14/2023 09:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GUZMAN,R  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: PARDO-CIFUENTES

DATE/TIME OF TEST: 05/12/2023 02:25:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:17 |
| Air Blank          | 0.000 | 02:18 |
| Control Test       | 0.081 | 02:18 |
| Air Blank          | 0.000 | 02:19 |
| Subject Sample # 1 | 0.125 | 02:20 |
| Breath Volume      | 1.523 | 02:20 |
| Air Blank          | 0.000 | 02:20 |
| Air Blank          | 0.000 | 02:22 |
| Subject Sample # 2 | 0.128 | 02:23 |
| Breath Volume      | 1.441 | 02:23 |
| Air Blank          | 0.000 | 02:24 |
| Control Test       | 0.081 | 02:24 |
| Air Blank          | 0.000 | 02:25 |
| Diagnostics Check  | OK    | 02:25 |

---

Data Download Date/Time: 06/14/2023 09:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GRIFFIN,K  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: SHAYHORN

DATE/TIME OF TEST: 05/12/2023 20:19:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:12 |
| Air Blank          | 0.000 | 20:13 |
| Control Test       | 0.081 | 20:13 |
| Air Blank          | 0.000 | 20:14 |
| Subject Sample # 1 | 0.000 | 20:14 |
| Breath Volume      | 1.816 | 20:14 |
| Air Blank          | 0.000 | 20:15 |
| Air Blank          | 0.000 | 20:17 |
| Subject Sample # 2 | 0.000 | 20:17 |
| Breath Volume      | 1.871 | 20:17 |
| Air Blank          | 0.000 | 20:18 |
| Control Test       | 0.081 | 20:18 |
| Air Blank          | 0.000 | 20:18 |
| Diagnostics Check  | OK    | 20:19 |

---

Data Download Date/Time: 06/14/2023 09:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CHURCHILL,CAMERON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: KEIPER

DATE/TIME OF TEST: 05/13/2023 21:13:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:07 |
| Air Blank          | 0.000 | 21:07 |
| Control Test       | 0.081 | 21:08 |
| Air Blank          | 0.000 | 21:08 |
| Subject Sample # 1 | 0.000 | 21:09 |
| Breath Volume      | 3.171 | 21:09 |
| Air Blank          | 0.000 | 21:09 |
| Air Blank          | 0.000 | 21:11 |
| Subject Sample # 2 | 0.000 | 21:12 |
| Breath Volume      | 2.507 | 21:12 |
| Air Blank          | 0.000 | 21:12 |
| Control Test       | 0.080 | 21:12 |
| Air Blank          | 0.000 | 21:13 |
| Diagnostics Check  | OK    | 21:13 |

---

Data Download Date/Time: 06/14/2023 09:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: JUNE,M  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 05/13/2023 22:05:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:02 |                      |
| Air Blank         | 0.000 | 22:02 |                      |
| Control Test      | 0.081 | 22:03 |                      |
| Air Blank         | 0.000 | 22:03 |                      |
| Subject Sample    | # 1   | 22:03 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:03 |                      |
| Air Blank         | 0.000 | 22:04 |                      |
| Control Test      | 0.080 | 22:04 |                      |
| Air Blank         | 0.000 | 22:05 |                      |
| Diagnostics Check | OK    | 22:05 |                      |

---

Data Download Date/Time: 06/14/2023 09:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MITCHELL,ALEXANDRA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: COLINDRES

DATE/TIME OF TEST: 05/13/2023 23:23:58

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:20 |                      |
| Air Blank         | 0.000 | 23:21 |                      |
| Control Test      | 0.081 | 23:21 |                      |
| Air Blank         | 0.000 | 23:22 |                      |
| Subject Sample    | # 1   | 23:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:22 |                      |
| Air Blank         | 0.000 | 23:23 |                      |
| Control Test      | 0.081 | 23:23 |                      |
| Air Blank         | 0.000 | 23:23 |                      |
| Diagnostics Check | OK    | 23:23 |                      |

---

Data Download Date/Time: 06/14/2023 09:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: NEWTON,T  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: PATRICK

DATE/TIME OF TEST: 05/14/2023 01:59:28

---

|                   |       |             |                    |
|-------------------|-------|-------------|--------------------|
| Diagnostics Check | OK    | 01:45       |                    |
| Air Blank         | 0.000 | 01:45       |                    |
| Control Test      | 0.080 | 01:45       |                    |
| Air Blank         | 0.000 | 01:46       |                    |
| Subject Sample    | # 1   | 01:49       | No Sample Provided |
| Breath Volume     | 0.000 | 01:49       |                    |
| Air Blank         | 0.000 | 01:50       |                    |
| Air Blank         | 0.000 | 01:52       |                    |
| Subject Sample    | # 2   | 0.129 01:54 |                    |
| Breath Volume     | 1.699 | 01:54       |                    |
| Air Blank         | 0.000 | 01:54       |                    |
| Air Blank         | 0.000 | 01:56       |                    |
| Subject Sample    | # 3   | 0.130 01:57 |                    |
| Breath Volume     | 2.437 | 01:57       |                    |
| Air Blank         | 0.000 | 01:58       |                    |
| Control Test      | 0.081 | 01:58       |                    |
| Air Blank         | 0.000 | 01:59       |                    |
| Diagnostics Check | OK    | 01:59       |                    |

---

Data Download Date/Time: 06/14/2023 09:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY W.,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GOODPASTOR,DEPUTY K  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/14/2023 23:38:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:31 |
| Air Blank          | 0.000 | 23:31 |
| Control Test       | 0.081 | 23:31 |
| Air Blank          | 0.000 | 23:32 |
| Subject Sample # 1 | 0.178 | 23:33 |
| Breath Volume      | 1.117 | 23:33 |
| Air Blank          | 0.000 | 23:34 |
| Air Blank          | 0.000 | 23:35 |
| Subject Sample # 2 | 0.195 | 23:36 |
| Breath Volume      | 2.414 | 23:36 |
| Air Blank          | 0.000 | 23:37 |
| Control Test       | 0.081 | 23:37 |
| Air Blank          | 0.000 | 23:38 |
| Diagnostics Check  | OK    | 23:38 |

---

Data Download Date/Time: 06/14/2023 09:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY W.,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SKEIRIK,DEPUTY D.  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/15/2023 00:47:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:40 |
| Air Blank          | 0.000 | 00:40 |
| Control Test       | 0.080 | 00:41 |
| Air Blank          | 0.000 | 00:41 |
| Subject Sample # 1 | 0.259 | 00:42 |
| Breath Volume      | 2.703 | 00:42 |
| Air Blank          | 0.000 | 00:42 |
| Air Blank          | 0.000 | 00:44 |
| Subject Sample # 2 | 0.257 | 00:45 |
| Breath Volume      | 2.234 | 00:45 |
| Air Blank          | 0.000 | 00:46 |
| Control Test       | 0.080 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Diagnostics Check  | OK    | 00:47 |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TAMAN,KELLEY  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: SANCHEZ CRUZ

DATE/TIME OF TEST: 05/15/2023 06:10:29

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 06:07 |                      |
| Air Blank         | 0.000 | 06:08 |                      |
| Control Test      | 0.081 | 06:08 |                      |
| Air Blank         | 0.000 | 06:08 |                      |
| Subject Sample    | # 1   | 06:09 | Subject Test Refused |
| Breath Volume     | 0.000 | 06:09 |                      |
| Air Blank         | 0.000 | 06:09 |                      |
| Control Test      | 0.081 | 06:09 |                      |
| Air Blank         | 0.000 | 06:10 |                      |
| Diagnostics Check | OK    | 06:10 |                      |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/16/2023 02:36:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:28 |
| Air Blank          | 0.000 | 02:28 |
| Control Test       | 0.080 | 02:29 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 1 | 0.102 | 02:31 |
| Breath Volume      | 2.371 | 02:31 |
| Air Blank          | 0.000 | 02:32 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 2 | 0.102 | 02:34 |
| Breath Volume      | 2.878 | 02:34 |
| Air Blank          | 0.000 | 02:35 |
| Control Test       | 0.080 | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Diagnostics Check  | OK    | 02:36 |

---

Data Download Date/Time: 06/14/2023 09:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DARLING,NATHAN,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: STILES,MEGAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 16:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 05/17/2023 17:45:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:39 |
| Air Blank          | 0.000 | 17:39 |
| Control Test       | 0.081 | 17:39 |
| Air Blank          | 0.000 | 17:40 |
| Subject Sample # 1 | 0.000 | 17:40 |
| Breath Volume      | 2.687 | 17:40 |
| Air Blank          | 0.000 | 17:41 |
| Air Blank          | 0.000 | 17:43 |
| Subject Sample # 2 | 0.000 | 17:43 |
| Breath Volume      | 2.671 | 17:43 |
| Air Blank          | 0.000 | 17:44 |
| Control Test       | 0.081 | 17:44 |
| Air Blank          | 0.000 | 17:45 |
| Diagnostics Check  | OK    | 17:45 |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: NEWTON,THOMAS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: STACHELSKY

DATE/TIME OF TEST: 05/18/2023 04:07:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:01 |
| Air Blank          | 0.000 | 04:02 |
| Control Test       | 0.081 | 04:02 |
| Air Blank          | 0.000 | 04:02 |
| Subject Sample # 1 | 0.127 | 04:03 |
| Breath Volume      | 1.242 | 04:03 |
| Air Blank          | 0.000 | 04:03 |
| Air Blank          | 0.000 | 04:05 |
| Subject Sample # 2 | 0.127 | 04:06 |
| Breath Volume      | 1.312 | 04:06 |
| Air Blank          | 0.000 | 04:06 |
| Control Test       | 0.081 | 04:07 |
| Air Blank          | 0.000 | 04:07 |
| Diagnostics Check  | OK    | 04:07 |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLE,DANIEL,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: COLE,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: JOYAL

DATE/TIME OF TEST: 05/18/2023 04:55:33

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:52 |                      |
| Air Blank         | 0.000 | 04:53 |                      |
| Control Test      | 0.081 | 04:53 |                      |
| Air Blank         | 0.000 | 04:53 |                      |
| Subject Sample    | # 1   | 04:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:54 |                      |
| Air Blank         | 0.000 | 04:54 |                      |
| Control Test      | 0.081 | 04:54 |                      |
| Air Blank         | 0.000 | 04:55 |                      |
| Diagnostics Check | OK    | 04:55 |                      |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: FOGLER TORTORA,JENNIFER  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 05/18/2023 05:35:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:28 |
| Air Blank          | 0.000 | 05:29 |
| Control Test       | 0.081 | 05:29 |
| Air Blank          | 0.000 | 05:29 |
| Subject Sample # 1 | 0.000 | 05:30 |
| Breath Volume      | 1.285 | 05:30 |
| Air Blank          | 0.000 | 05:31 |
| Air Blank          | 0.000 | 05:33 |
| Subject Sample # 2 | 0.000 | 05:34 |
| Breath Volume      | 2.160 | 05:34 |
| Air Blank          | 0.000 | 05:34 |
| Control Test       | 0.081 | 05:35 |
| Air Blank          | 0.000 | 05:35 |
| Diagnostics Check  | OK    | 05:35 |

---

Data Download Date/Time: 06/14/2023 09:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GOODPASTOR,K  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 05/18/2023 22:28:50

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:25 |                      |
| Air Blank         | 0.000 | 22:25 |                      |
| Control Test      | 0.080 | 22:26 |                      |
| Air Blank         | 0.000 | 22:26 |                      |
| Subject Sample    | # 1   | 22:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:27 |                      |
| Air Blank         | 0.000 | 22:27 |                      |
| Control Test      | 0.080 | 22:28 |                      |
| Air Blank         | 0.000 | 22:28 |                      |
| Diagnostics Check | OK    | 22:28 |                      |

---

Data Download Date/Time: 06/14/2023 09:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HUGHES,A  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: STEFANIK

DATE/TIME OF TEST: 05/19/2023 01:10:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:04 |
| Air Blank          | 0.000 | 01:04 |
| Control Test       | 0.079 | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Subject Sample # 1 | 0.187 | 01:06 |
| Breath Volume      | 1.144 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Air Blank          | 0.000 | 01:08 |
| Subject Sample # 2 | 0.201 | 01:09 |
| Breath Volume      | 1.339 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Control Test       | 0.080 | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Diagnostics Check  | OK    | 01:10 |

---

Data Download Date/Time: 06/14/2023 09:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DONOHUE,LAWRENCE  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/19/2023 08:41:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:35 |
| Air Blank          | 0.000 | 08:35 |
| Control Test       | 0.080 | 08:36 |
| Air Blank          | 0.000 | 08:36 |
| Subject Sample # 1 | 0.101 | 08:37 |
| Breath Volume      | 2.851 | 08:37 |
| Air Blank          | 0.000 | 08:37 |
| Air Blank          | 0.000 | 08:39 |
| Subject Sample # 2 | 0.096 | 08:40 |
| Breath Volume      | 1.781 | 08:40 |
| Air Blank          | 0.000 | 08:41 |
| Control Test       | 0.080 | 08:41 |
| Air Blank          | 0.000 | 08:41 |
| Diagnostics Check  | OK    | 08:41 |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELSHANS,DEAN,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BJURHOLM,JACOB  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: BALLOS

DATE/TIME OF TEST: 05/19/2023 23:48:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:46 |                      |
| Air Blank         | 0.000 | 23:46 |                      |
| Control Test      | 0.081 | 23:46 |                      |
| Air Blank         | 0.000 | 23:47 |                      |
| Subject Sample    | # 1   | 23:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:47 |                      |
| Air Blank         | 0.000 | 23:48 |                      |
| Control Test      | 0.080 | 23:48 |                      |
| Air Blank         | 0.000 | 23:48 |                      |
| Diagnostics Check | OK    | 23:48 |                      |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELSHANS,DEAN,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GRAHAM,DEANGELO  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: VAN HEUSEN

DATE/TIME OF TEST: 05/20/2023 02:54:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.081 | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Subject Sample # 1 | 0.060 | 02:50 |
| Breath Volume      | 4.453 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Air Blank          | 0.000 | 02:52 |
| Subject Sample # 2 | 0.060 | 02:53 |
| Breath Volume      | 4.476 | 02:53 |
| Air Blank          | 0.000 | 02:53 |
| Control Test       | 0.080 | 02:54 |
| Air Blank          | 0.000 | 02:54 |
| Diagnostics Check  | OK    | 02:54 |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CHURCHILL,CAMERON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BLANCHER

DATE/TIME OF TEST: 05/21/2023 01:34:06

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:30 |                      |
| Air Blank         | 0.000 | 01:31 |                      |
| Control Test      | 0.080 | 01:31 |                      |
| Air Blank         | 0.000 | 01:32 |                      |
| Subject Sample    | # 1   | 01:32 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:32 |                      |
| Air Blank         | 0.000 | 01:33 |                      |
| Control Test      | 0.080 | 01:33 |                      |
| Air Blank         | 0.000 | 01:34 |                      |
| Diagnostics Check | OK    | 01:34 |                      |

---

Data Download Date/Time: 06/14/2023 09:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: COYLE,TIMOTHY  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: PARKE

DATE/TIME OF TEST: 05/21/2023 19:27:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:18 |
| Air Blank          | 0.000 | 19:19 |
| Control Test       | 0.081 | 19:19 |
| Air Blank          | 0.000 | 19:19 |
| Subject Sample # 1 | 0.147 | 19:21 |
| Breath Volume      | 1.167 | 19:21 |
| Air Blank          | 0.000 | 19:22 |
| Air Blank          | 0.000 | 19:24 |
| Subject Sample # 2 | 0.147 | 19:25 |
| Breath Volume      | 1.558 | 19:25 |
| Air Blank          | 0.000 | 19:26 |
| Control Test       | 0.080 | 19:26 |
| Air Blank          | 0.000 | 19:27 |
| Diagnostics Check  | OK    | 19:27 |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,BRYAN,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SIMMS,NICHOLAS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: HAGER

DATE/TIME OF TEST: 05/22/2023 08:51:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:45 |
| Air Blank          | 0.000 | 08:45 |
| Control Test       | 0.080 | 08:46 |
| Air Blank          | 0.000 | 08:46 |
| Subject Sample # 1 | 0.000 | 08:47 |
| Breath Volume      | 1.261 | 08:47 |
| Air Blank          | 0.000 | 08:47 |
| Air Blank          | 0.000 | 08:49 |
| Subject Sample # 2 | 0.000 | 08:50 |
| Breath Volume      | 1.625 | 08:50 |
| Air Blank          | 0.000 | 08:50 |
| Control Test       | 0.080 | 08:50 |
| Air Blank          | 0.000 | 08:51 |
| Diagnostics Check  | OK    | 08:51 |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SKALKO,RYAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: LEBLANC

DATE/TIME OF TEST: 05/22/2023 21:02:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:55 |
| Air Blank          | 0.000 | 20:56 |
| Control Test       | 0.081 | 20:56 |
| Air Blank          | 0.000 | 20:57 |
| Subject Sample # 1 | 0.314 | 20:57 |
| Breath Volume      | 1.867 | 20:57 |
| Air Blank          | 0.000 | 20:58 |
| Air Blank          | 0.000 | 21:00 |
| Subject Sample # 2 | 0.310 | 21:00 |
| Breath Volume      | 1.734 | 21:00 |
| Air Blank          | 0.000 | 21:01 |
| Control Test       | 0.080 | 21:01 |
| Air Blank          | 0.000 | 21:02 |
| Diagnostics Check  | OK    | 21:02 |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: JUNE,MAX  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/23/2023 05:32:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:26 |
| Air Blank          | 0.000 | 05:26 |
| Control Test       | 0.081 | 05:26 |
| Air Blank          | 0.000 | 05:27 |
| Subject Sample # 1 | 0.000 | 05:27 |
| Breath Volume      | 1.714 | 05:27 |
| Air Blank          | 0.000 | 05:28 |
| Air Blank          | 0.000 | 05:30 |
| Subject Sample # 2 | 0.000 | 05:30 |
| Breath Volume      | 1.746 | 05:30 |
| Air Blank          | 0.000 | 05:31 |
| Control Test       | 0.080 | 05:31 |
| Air Blank          | 0.000 | 05:32 |
| Diagnostics Check  | OK    | 05:32 |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORGAN,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DORGAN,STEPHEN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MUCCI

DATE/TIME OF TEST: 05/24/2023 01:50:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:40 |
| Air Blank          | 0.000 | 01:41 |
| Control Test       | 0.080 | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Subject Sample # 1 | 0.171 | 01:44 |
| Breath Volume      | 2.507 | 01:44 |
| Air Blank          | 0.000 | 01:45 |
| Air Blank          | 0.000 | 01:47 |
| Subject Sample # 2 | 0.164 | 01:48 |
| Breath Volume      | 2.464 | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.077 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Diagnostics Check  | OK    | 01:50 |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ZSIDO,DANIELLE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 05/24/2023 02:18:04

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:15 |                      |
| Air Blank         | 0.000 | 02:15 |                      |
| Control Test      | 0.080 | 02:16 |                      |
| Air Blank         | 0.000 | 02:16 |                      |
| Subject Sample    | # 1   | 02:16 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:16 |                      |
| Air Blank         | 0.000 | 02:17 |                      |
| Control Test      | 0.079 | 02:17 |                      |
| Air Blank         | 0.000 | 02:17 |                      |
| Diagnostics Check | OK    | 02:18 |                      |

---

Data Download Date/Time: 06/14/2023 09:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GRIGGS,J  
ARRESTING OFFICER AGENCY: Indian Shores PD  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: KOVACS

DATE/TIME OF TEST: 05/24/2023 03:53:35

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:50 |                      |
| Air Blank         | 0.000 | 03:51 |                      |
| Control Test      | 0.081 | 03:51 |                      |
| Air Blank         | 0.000 | 03:52 |                      |
| Subject Sample    | # 1   | 03:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:52 |                      |
| Air Blank         | 0.000 | 03:52 |                      |
| Control Test      | 0.080 | 03:52 |                      |
| Air Blank         | 0.000 | 03:53 |                      |
| Diagnostics Check | OK    | 03:53 |                      |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KATSAREAS,NIKOLAOS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BENTLEY,T  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 08:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 05/25/2023 09:52:07

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 09:36 |                |
| Air Blank         | 0.000     | 09:37 |                |
| Control Test      | 0.081     | 09:37 |                |
| Air Blank         | 0.000     | 09:38 |                |
| Subject Sample    | # 1 0.000 | 09:41 | Volume Not Met |
| Breath Volume     | 0.179     | 09:41 |                |
| Air Blank         | 0.000     | 09:41 |                |
| Air Blank         | 0.000     | 09:43 |                |
| Subject Sample    | # 2 0.206 | 09:44 |                |
| Breath Volume     | 1.269     | 09:44 |                |
| Air Blank         | 0.000     | 09:45 |                |
| Air Blank         | 0.000     | 09:47 |                |
| Subject Sample    | # 3 0.206 | 09:50 | Volume Not Met |
| Breath Volume     | 0.484     | 09:50 |                |
| Air Blank         | 0.000     | 09:51 |                |
| Control Test      | 0.080     | 09:51 |                |
| Air Blank         | 0.000     | 09:52 |                |
| Diagnostics Check | OK        | 09:52 |                |

---

Data Download Date/Time: 06/14/2023 09:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KATSAREAS,NIKOLAOS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GUINN,R  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 09:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:49

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 05/25/2023 10:18:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:11 |
| Air Blank          | 0.000 | 10:12 |
| Control Test       | 0.080 | 10:12 |
| Air Blank          | 0.000 | 10:13 |
| Subject Sample # 1 | 0.000 | 10:13 |
| Breath Volume      | 3.386 | 10:13 |
| Air Blank          | 0.000 | 10:14 |
| Air Blank          | 0.000 | 10:16 |
| Subject Sample # 2 | 0.000 | 10:16 |
| Breath Volume      | 1.851 | 10:16 |
| Air Blank          | 0.000 | 10:17 |
| Control Test       | 0.080 | 10:17 |
| Air Blank          | 0.000 | 10:17 |
| Diagnostics Check  | OK    | 10:18 |

---

Data Download Date/Time: 06/14/2023 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: YEARGIN,DEP GENESIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: GONZALEZ MEDINA

DATE/TIME OF TEST: 05/26/2023 04:30:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:23 |
| Air Blank          | 0.000 | 04:23 |
| Control Test       | 0.081 | 04:23 |
| Air Blank          | 0.000 | 04:24 |
| Subject Sample # 1 | 0.163 | 04:25 |
| Breath Volume      | 1.402 | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Air Blank          | 0.000 | 04:27 |
| Subject Sample # 2 | 0.153 | 04:29 |
| Breath Volume      | 1.570 | 04:29 |
| Air Blank          | 0.000 | 04:29 |
| Control Test       | 0.080 | 04:29 |
| Air Blank          | 0.000 | 04:30 |
| Diagnostics Check  | OK    | 04:30 |

---

Data Download Date/Time: 06/14/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANEY,DAMON,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: RICOTTILLI,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 11:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:29

SUBJECT'S LAST NAME: KAUFERLEIN

DATE/TIME OF TEST: 05/26/2023 13:31:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:25 |
| Air Blank          | 0.000 | 13:26 |
| Control Test       | 0.081 | 13:26 |
| Air Blank          | 0.000 | 13:26 |
| Subject Sample # 1 | 0.000 | 13:27 |
| Breath Volume      | 1.617 | 13:27 |
| Air Blank          | 0.000 | 13:27 |
| Air Blank          | 0.000 | 13:29 |
| Subject Sample # 2 | 0.000 | 13:30 |
| Breath Volume      | 1.949 | 13:30 |
| Air Blank          | 0.000 | 13:30 |
| Control Test       | 0.081 | 13:31 |
| Air Blank          | 0.000 | 13:31 |
| Diagnostics Check  | OK    | 13:31 |

---

Data Download Date/Time: 06/14/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, OFC B,  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: RODRIGUEZ, OFC B  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: GASTON

DATE/TIME OF TEST: 05/26/2023 20:56:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:48 |
| Air Blank          | 0.000 | 20:48 |
| Control Test       | 0.081 | 20:48 |
| Air Blank          | 0.000 | 20:49 |
| Subject Sample # 1 | 0.000 | 20:51 |
| Breath Volume      | 2.679 | 20:51 |
| Air Blank          | 0.000 | 20:51 |
| Air Blank          | 0.000 | 20:53 |
| Subject Sample # 2 | 0.000 | 20:54 |
| Breath Volume      | 2.820 | 20:54 |
| Air Blank          | 0.000 | 20:55 |
| Control Test       | 0.082 | 20:55 |
| Air Blank          | 0.000 | 20:56 |
| Diagnostics Check  | OK    | 20:56 |

---

Data Download Date/Time: 06/14/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SEVOSTJANOV,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: ROTH

DATE/TIME OF TEST: 05/27/2023 03:00:43

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:48 |                    |
| Air Blank         | 0.000 | 02:49 |                    |
| Control Test      | 0.080 | 02:49 |                    |
| Air Blank         | 0.000 | 02:50 |                    |
| Subject Sample    | # 1   | 02:53 | No Sample Provided |
| Breath Volume     | 0.000 | 02:53 |                    |
| Air Blank         | 0.000 | 02:53 |                    |
| Air Blank         | 0.000 | 02:55 |                    |
| Subject Sample    | # 2   | 0.157 | 02:56              |
| Breath Volume     | 2.433 | 02:56 |                    |
| Air Blank         | 0.000 | 02:56 |                    |
| Air Blank         | 0.000 | 02:58 |                    |
| Subject Sample    | # 3   | 0.153 | 02:59              |
| Breath Volume     | 2.023 | 02:59 |                    |
| Air Blank         | 0.000 | 02:59 |                    |
| Control Test      | 0.080 | 03:00 |                    |
| Air Blank         | 0.000 | 03:00 |                    |
| Diagnostics Check | OK    | 03:00 |                    |

Data Download Date/Time: 06/14/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BRENNON,C  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MOUNCE

DATE/TIME OF TEST: 05/27/2023 03:31:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:24 |
| Air Blank          | 0.000 | 03:25 |
| Control Test       | 0.080 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Subject Sample # 1 | 0.000 | 03:27 |
| Breath Volume      | 2.324 | 03:27 |
| Air Blank          | 0.000 | 03:27 |
| Air Blank          | 0.000 | 03:29 |
| Subject Sample # 2 | 0.000 | 03:30 |
| Breath Volume      | 2.121 | 03:30 |
| Air Blank          | 0.000 | 03:30 |
| Control Test       | 0.080 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Diagnostics Check  | OK    | 03:31 |

---

Data Download Date/Time: 06/14/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: YEARGIN, DEP GENESIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 05/27/2023 03:59:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Control Test       | 0.080 | 03:53 |
| Air Blank          | 0.000 | 03:53 |
| Subject Sample # 1 | 0.166 | 03:54 |
| Breath Volume      | 1.882 | 03:54 |
| Air Blank          | 0.000 | 03:55 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 2 | 0.161 | 03:57 |
| Breath Volume      | 1.585 | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Control Test       | 0.080 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Diagnostics Check  | OK    | 03:59 |

---

Data Download Date/Time: 06/14/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: JENKINS,K  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SCAVELLI

DATE/TIME OF TEST: 05/27/2023 04:25:24

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:22 |                      |
| Air Blank         | 0.000 | 04:23 |                      |
| Control Test      | 0.080 | 04:23 |                      |
| Air Blank         | 0.000 | 04:23 |                      |
| Subject Sample    | # 1   | 04:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:24 |                      |
| Air Blank         | 0.000 | 04:24 |                      |
| Control Test      | 0.080 | 04:24 |                      |
| Air Blank         | 0.000 | 04:25 |                      |
| Diagnostics Check | OK    | 04:25 |                      |

---

Data Download Date/Time: 06/14/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GAMBEL,Z,E  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GAMBEL,Z  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: CULMER

DATE/TIME OF TEST: 05/27/2023 21:37:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:29 |
| Air Blank          | 0.000 | 21:30 |
| Control Test       | 0.080 | 21:30 |
| Air Blank          | 0.000 | 21:30 |
| Subject Sample # 1 | 0.223 | 21:32 |
| Breath Volume      | 2.718 | 21:32 |
| Air Blank          | 0.000 | 21:33 |
| Air Blank          | 0.000 | 21:35 |
| Subject Sample # 2 | 0.222 | 21:36 |
| Breath Volume      | 1.636 | 21:36 |
| Air Blank          | 0.000 | 21:36 |
| Control Test       | 0.080 | 21:36 |
| Air Blank          | 0.000 | 21:37 |
| Diagnostics Check  | OK    | 21:37 |

---

Data Download Date/Time: 06/14/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CHURCHILL,CAMERON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 05/28/2023 00:25:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:16 |
| Air Blank          | 0.000 | 00:17 |
| Control Test       | 0.077 | 00:17 |
| Air Blank          | 0.000 | 00:17 |
| Subject Sample # 1 | 0.258 | 00:18 |
| Breath Volume      | 1.644 | 00:18 |
| Air Blank          | 0.000 | 00:18 |
| Air Blank          | 0.000 | 00:20 |
| Subject Sample # 2 | 0.288 | 00:21 |
| Breath Volume      | 1.480 | 00:21 |
| Air Blank          | 0.000 | 00:21 |
| Air Blank          | 0.000 | 00:23 |
| Subject Sample # 3 | 0.282 | 00:24 |
| Breath Volume      | 1.539 | 00:24 |
| Air Blank          | 0.000 | 00:24 |
| Control Test       | 0.080 | 00:25 |
| Air Blank          | 0.000 | 00:25 |
| Diagnostics Check  | OK    | 00:25 |

---

Data Download Date/Time: 06/14/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: FERGUSON,B  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/28/2023 01:53:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:51 |                      |
| Air Blank         | 0.000 | 01:51 |                      |
| Control Test      | 0.080 | 01:51 |                      |
| Air Blank         | 0.000 | 01:52 |                      |
| Subject Sample    | # 1   | 01:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:52 |                      |
| Air Blank         | 0.000 | 01:53 |                      |
| Control Test      | 0.080 | 01:53 |                      |
| Air Blank         | 0.000 | 01:53 |                      |
| Diagnostics Check | OK    | 01:53 |                      |

---

Data Download Date/Time: 06/14/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORO,JARED,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MILLER,HOWARD  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 09:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:59

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/28/2023 10:27:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:21 |
| Air Blank          | 0.000 | 10:21 |
| Control Test       | 0.082 | 10:21 |
| Air Blank          | 0.000 | 10:22 |
| Subject Sample # 1 | 0.065 | 10:23 |
| Breath Volume      | 2.683 | 10:23 |
| Air Blank          | 0.000 | 10:24 |
| Air Blank          | 0.000 | 10:25 |
| Subject Sample # 2 | 0.061 | 10:26 |
| Breath Volume      | 2.605 | 10:26 |
| Air Blank          | 0.000 | 10:27 |
| Control Test       | 0.080 | 10:27 |
| Air Blank          | 0.000 | 10:27 |
| Diagnostics Check  | OK    | 10:27 |

---

Data Download Date/Time: 06/14/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: YEARGIN, DEP GENESIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 05/28/2023 21:21:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:14 |
| Air Blank          | 0.000 | 21:15 |
| Control Test       | 0.081 | 21:15 |
| Air Blank          | 0.000 | 21:16 |
| Subject Sample # 1 | 0.000 | 21:17 |
| Breath Volume      | 2.562 | 21:17 |
| Air Blank          | 0.000 | 21:17 |
| Air Blank          | 0.000 | 21:19 |
| Subject Sample # 2 | 0.000 | 21:20 |
| Breath Volume      | 2.058 | 21:20 |
| Air Blank          | 0.000 | 21:20 |
| Control Test       | 0.080 | 21:20 |
| Air Blank          | 0.000 | 21:21 |
| Diagnostics Check  | OK    | 21:21 |

---

Data Download Date/Time: 06/14/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: YEARGIN,DEP GENESIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SHRESTHA

DATE/TIME OF TEST: 05/29/2023 01:56:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:50 |
| Air Blank          | 0.000 | 01:50 |
| Control Test       | 0.081 | 01:51 |
| Air Blank          | 0.000 | 01:51 |
| Subject Sample # 1 | 0.000 | 01:51 |
| Breath Volume      | 2.968 | 01:51 |
| Air Blank          | 0.000 | 01:52 |
| Air Blank          | 0.000 | 01:54 |
| Subject Sample # 2 | 0.000 | 01:54 |
| Breath Volume      | 2.777 | 01:54 |
| Air Blank          | 0.000 | 01:55 |
| Control Test       | 0.080 | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Diagnostics Check  | OK    | 01:56 |

---

Data Download Date/Time: 06/14/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: DAVIS, BRITT  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 05/29/2023 04:17:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:10 |
| Air Blank          | 0.000 | 04:10 |
| Control Test       | 0.080 | 04:11 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 1 | 0.085 | 04:12 |
| Breath Volume      | 1.886 | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Air Blank          | 0.000 | 04:14 |
| Subject Sample # 2 | 0.081 | 04:15 |
| Breath Volume      | 2.292 | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Control Test       | 0.080 | 04:16 |
| Air Blank          | 0.000 | 04:17 |
| Diagnostics Check  | OK    | 04:17 |

---

Data Download Date/Time: 06/14/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SWIER, DEP HAYDEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: VALENTI

DATE/TIME OF TEST: 05/29/2023 05:17:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:10 |
| Air Blank          | 0.000 | 05:11 |
| Control Test       | 0.079 | 05:11 |
| Air Blank          | 0.000 | 05:11 |
| Subject Sample # 1 | 0.185 | 05:12 |
| Breath Volume      | 2.480 | 05:12 |
| Air Blank          | 0.000 | 05:13 |
| Air Blank          | 0.000 | 05:14 |
| Subject Sample # 2 | 0.183 | 05:15 |
| Breath Volume      | 2.437 | 05:15 |
| Air Blank          | 0.000 | 05:16 |
| Control Test       | 0.080 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Diagnostics Check  | OK    | 05:17 |

---

Data Download Date/Time: 06/14/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CARLSON, DEP SEAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 15:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: SUCH

DATE/TIME OF TEST: 05/29/2023 16:36:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:30 |
| Air Blank          | 0.000 | 16:31 |
| Control Test       | 0.080 | 16:31 |
| Air Blank          | 0.000 | 16:31 |
| Subject Sample # 1 | 0.000 | 16:32 |
| Breath Volume      | 2.250 | 16:32 |
| Air Blank          | 0.000 | 16:32 |
| Air Blank          | 0.000 | 16:34 |
| Subject Sample # 2 | 0.000 | 16:35 |
| Breath Volume      | 2.125 | 16:35 |
| Air Blank          | 0.000 | 16:35 |
| Control Test       | 0.081 | 16:36 |
| Air Blank          | 0.000 | 16:36 |
| Diagnostics Check  | OK    | 16:36 |

---

Data Download Date/Time: 06/14/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TACIA, STEVEN, M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SCOTT, KRISTOPHER  
ARRESTING OFFICER AGENCY: Treasure Island PD  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 17:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: CROUSE

DATE/TIME OF TEST: 05/29/2023 18:58:26

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 18:45 |                |
| Air Blank         | 0.000     | 18:46 |                |
| Control Test      | 0.081     | 18:46 |                |
| Air Blank         | 0.000     | 18:46 |                |
| Subject Sample    | # 1 0.171 | 18:48 |                |
| Breath Volume     | 1.335     | 18:48 |                |
| Air Blank         | 0.000     | 18:48 |                |
| Air Blank         | 0.000     | 18:50 |                |
| Subject Sample    | # 2 0.146 | 18:53 | Volume Not Met |
| Breath Volume     | 0.703     | 18:53 |                |
| Air Blank         | 0.000     | 18:54 |                |
| Air Blank         | 0.000     | 18:56 |                |
| Subject Sample    | # 3 0.171 | 18:56 |                |
| Breath Volume     | 1.308     | 18:56 |                |
| Air Blank         | 0.000     | 18:57 |                |
| Control Test      | 0.080     | 18:57 |                |
| Air Blank         | 0.000     | 18:58 |                |
| Diagnostics Check | OK        | 18:58 |                |

---

Data Download Date/Time: 06/14/2023 09:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GRAVER,RYAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG205902  
DRY GAS EXPIRY DATE: 02/28/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 05/30/2023 22:23:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:17 |
| Air Blank          | 0.000 | 22:17 |
| Control Test       | 0.080 | 22:17 |
| Air Blank          | 0.000 | 22:18 |
| Subject Sample # 1 | 0.294 | 22:19 |
| Breath Volume      | 2.523 | 22:19 |
| Air Blank          | 0.000 | 22:19 |
| Air Blank          | 0.000 | 22:21 |
| Subject Sample # 2 | 0.293 | 22:22 |
| Breath Volume      | 1.789 | 22:22 |
| Air Blank          | 0.000 | 22:22 |
| Control Test       | 0.080 | 22:23 |
| Air Blank          | 0.000 | 22:23 |
| Diagnostics Check  | OK    | 22:23 |

---

Data Download Date/Time: 06/14/2023 09:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

|                                                  |                                   |
|--------------------------------------------------|-----------------------------------|
| INSTRUMENT REGISTERED TO: FFWCC                  | Date or Time Changed: No          |
| SOFTWARE VERSION: 8100.27                        | DRY GAS STANDARD LOT #: 402469736 |
| BREATH TEST OPERATOR: BROOKS,JOSEPH,P            | DRY GAS EXPIRY DATE: 06/16/2025   |
| BREATH TEST OPERATOR AGENCY: FFWCC (Panama City) | VIOLATION CODE: BUI               |
| ARRESTING OFFICER: OLIVER,WILLIAM                | ARREST TIME: 17:08                |
| ARRESTING OFFICER AGENCY: FFWCC (Panama City)    | GENDER: Male                      |
| LAST AGENCY INSPECTION: 04/24/2023               | OBSERVATION PERIOD BEGAN: 17:10   |

SUBJECT'S LAST NAME: BROUSSARD      DATE/TIME OF TEST: 05/01/2023      17:50:09

---

|                         |       |       |
|-------------------------|-------|-------|
| Diagnostics Check       | OK    | 17:43 |
| Air Blank               | 0.000 | 17:43 |
| Control Test            | 0.082 | 17:44 |
| Air Blank               | 0.000 | 17:44 |
| Subject Sample      # 1 | 0.108 | 17:45 |
| Breath Volume           | 2.152 | 17:45 |
| Air Blank               | 0.000 | 17:45 |
| Air Blank               | 0.000 | 17:47 |
| Subject Sample      # 2 | 0.107 | 17:48 |
| Breath Volume           | 2.148 | 17:48 |
| Air Blank               | 0.000 | 17:49 |
| Control Test            | 0.082 | 17:49 |
| Air Blank               | 0.000 | 17:50 |
| Diagnostics Check       | OK    | 17:50 |

---

Data Download Date/Time: 05/15/2023 16:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MENDOZA,JORGE,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: PRITCHARD,THOMAS  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402469736  
DRY GAS EXPIRY DATE: 06/16/2025  
VIOLATION CODE: BUI  
ARREST TIME: 14:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 05/13/2023 15:33:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:26 |
| Air Blank          | 0.000 | 15:27 |
| Control Test       | 0.083 | 15:27 |
| Air Blank          | 0.000 | 15:27 |
| Subject Sample # 1 | 0.138 | 15:29 |
| Breath Volume      | 2.316 | 15:29 |
| Air Blank          | 0.000 | 15:29 |
| Air Blank          | 0.000 | 15:31 |
| Subject Sample # 2 | 0.135 | 15:32 |
| Breath Volume      | 1.656 | 15:32 |
| Air Blank          | 0.000 | 15:33 |
| Control Test       | 0.082 | 15:33 |
| Air Blank          | 0.000 | 15:33 |
| Diagnostics Check  | OK    | 15:33 |

---

Data Download Date/Time: 05/15/2023 16:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POWELL,MICHAEL,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BAZYLAK,DAVID  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 402469736  
DRY GAS EXPIRY DATE: 06/16/2025  
VIOLATION CODE: BUI  
ARREST TIME: 16:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: OSTERLAND

DATE/TIME OF TEST: 05/13/2023 17:26:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:20 |
| Air Blank          | 0.000 | 17:21 |
| Control Test       | 0.082 | 17:21 |
| Air Blank          | 0.000 | 17:22 |
| Subject Sample # 1 | 0.128 | 17:22 |
| Breath Volume      | 1.484 | 17:22 |
| Air Blank          | 0.000 | 17:23 |
| Air Blank          | 0.000 | 17:24 |
| Subject Sample # 2 | 0.130 | 17:25 |
| Breath Volume      | 1.894 | 17:25 |
| Air Blank          | 0.000 | 17:25 |
| Control Test       | 0.080 | 17:26 |
| Air Blank          | 0.000 | 17:26 |
| Diagnostics Check  | OK    | 17:26 |

---

Data Download Date/Time: 05/15/2023 16:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PRITCHARD,THOMAS,D

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: SENST,CASEY

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402469736

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 19:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: AMERSON

DATE/TIME OF TEST: 05/13/2023 19:52:05

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:49 |                      |
| Air Blank         | 0.000 | 19:49 |                      |
| Control Test      | 0.082 | 19:49 |                      |
| Air Blank         | 0.000 | 19:50 |                      |
| Subject Sample    | # 1   | 19:50 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:50 |                      |
| Air Blank         | 0.000 | 19:51 |                      |
| Control Test      | 0.083 | 19:51 |                      |
| Air Blank         | 0.000 | 19:51 |                      |
| Diagnostics Check | OK    | 19:52 |                      |

---

Data Download Date/Time: 05/15/2023 16:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARTLETT,SHELTON,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: PRITCHARD,ASHLEY  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402469736  
DRY GAS EXPIRY DATE: 06/16/2025  
VIOLATION CODE: BUI  
ARREST TIME: 19:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 05/13/2023 20:12:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:04 |
| Air Blank          | 0.000 | 20:04 |
| Control Test       | 0.081 | 20:04 |
| Air Blank          | 0.000 | 20:05 |
| Subject Sample # 1 | 0.136 | 20:07 |
| Breath Volume      | 2.488 | 20:07 |
| Air Blank          | 0.000 | 20:08 |
| Air Blank          | 0.000 | 20:10 |
| Subject Sample # 2 | 0.134 | 20:10 |
| Breath Volume      | 2.980 | 20:10 |
| Air Blank          | 0.000 | 20:11 |
| Control Test       | 0.082 | 20:11 |
| Air Blank          | 0.000 | 20:11 |
| Diagnostics Check  | OK    | 20:12 |

---

Data Download Date/Time: 05/15/2023 16:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

|                                                  |                                   |
|--------------------------------------------------|-----------------------------------|
| INSTRUMENT REGISTERED TO: FFWCC                  | Date or Time Changed: No          |
| SOFTWARE VERSION: 8100.27                        | DRY GAS STANDARD LOT #: 402469736 |
| BREATH TEST OPERATOR: PRITCHARD,THOMAS,D         | DRY GAS EXPIRY DATE: 06/16/2025   |
| BREATH TEST OPERATOR AGENCY: FFWCC (Panama City) | VIOLATION CODE: BUI               |
| ARRESTING OFFICER: MURSCH,THEODORE               | ARREST TIME: 15:20                |
| ARRESTING OFFICER AGENCY: FFWCC (Panama City)    | GENDER: Male                      |
| LAST AGENCY INSPECTION: 04/24/2023               | OBSERVATION PERIOD BEGAN: 15:45   |

|                             |                                        |
|-----------------------------|----------------------------------------|
| SUBJECT'S LAST NAME: BRYANT | DATE/TIME OF TEST: 05/14/2023 16:12:32 |
|-----------------------------|----------------------------------------|

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:06 |
| Air Blank          | 0.000 | 16:06 |
| Control Test       | 0.083 | 16:07 |
| Air Blank          | 0.000 | 16:07 |
| Subject Sample # 1 | 0.013 | 16:07 |
| Breath Volume      | 2.101 | 16:07 |
| Air Blank          | 0.000 | 16:08 |
| Air Blank          | 0.000 | 16:10 |
| Subject Sample # 2 | 0.014 | 16:11 |
| Breath Volume      | 1.820 | 16:11 |
| Air Blank          | 0.000 | 16:11 |
| Control Test       | 0.081 | 16:11 |
| Air Blank          | 0.000 | 16:12 |
| Diagnostics Check  | OK    | 16:12 |

---

Data Download Date/Time: 05/15/2023 16:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRITCHARD, THOMAS, D  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: PRITCHARD, THOMAS  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402469736  
DRY GAS EXPIRY DATE: 06/16/2025  
VIOLATION CODE: BUI  
ARREST TIME: 18:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/14/2023 19:18:51

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:16 |                      |
| Air Blank         | 0.000 | 19:16 |                      |
| Control Test      | 0.083 | 19:16 |                      |
| Air Blank         | 0.000 | 19:17 |                      |
| Subject Sample    | # 1   | 19:17 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:17 |                      |
| Air Blank         | 0.000 | 19:17 |                      |
| Control Test      | 0.080 | 19:18 |                      |
| Air Blank         | 0.000 | 19:18 |                      |
| Diagnostics Check | OK    | 19:18 |                      |

---

Data Download Date/Time: 05/15/2023 16:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MURSCH,THEODORE,R

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: MURSCH,THEODORE

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402469736

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 18:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/14/2023 19:35:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:24 |                      |
| Air Blank         | 0.000 | 19:25 |                      |
| Control Test      | 0.082 | 19:25 |                      |
| Air Blank         | 0.000 | 19:26 |                      |
| Subject Sample    | # 1   | 19:29 | No Sample Provided   |
| Breath Volume     | 0.000 | 19:29 |                      |
| Air Blank         | 0.000 | 19:29 |                      |
| Air Blank         | 0.000 | 19:31 |                      |
| Subject Sample    | # 2   | 19:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:33 |                      |
| Air Blank         | 0.000 | 19:34 |                      |
| Control Test      | 0.081 | 19:34 |                      |
| Air Blank         | 0.000 | 19:35 |                      |
| Diagnostics Check | OK    | 19:35 |                      |

---

Data Download Date/Time: 05/15/2023 16:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROOKS,JOSEPH,P  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: PRICHARD,ASHLEY  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402469736  
DRY GAS EXPIRY DATE: 06/16/2025  
VIOLATION CODE: BUI  
ARREST TIME: 16:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: BUCK

DATE/TIME OF TEST: 05/19/2023 17:14:50

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:06 |                      |
| Air Blank         | 0.000 | 17:06 |                      |
| Control Test      | 0.081 | 17:07 |                      |
| Air Blank         | 0.000 | 17:07 |                      |
| Subject Sample    | # 1   | 17:10 | No Sample Provided   |
| Breath Volume     | 0.000 | 17:10 |                      |
| Air Blank         | 0.000 | 17:11 |                      |
| Air Blank         | 0.000 | 17:13 |                      |
| Subject Sample    | # 2   | 17:13 | Subject Test Refused |
| Breath Volume     | 0.000 | 17:13 |                      |
| Air Blank         | 0.000 | 17:13 |                      |
| Control Test      | 0.082 | 17:14 |                      |
| Air Blank         | 0.000 | 17:14 |                      |
| Diagnostics Check | OK    | 17:14 |                      |

---

Data Download Date/Time: 06/16/2023 09:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROOKS,JOSEPH,P

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: BASFORD,NICOLE

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402469736

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 14:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:47

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/20/2023 15:16:31

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 15:13 |                      |
| Air Blank         | 0.000 | 15:13 |                      |
| Control Test      | 0.081 | 15:14 |                      |
| Air Blank         | 0.000 | 15:14 |                      |
| Subject Sample    | # 1   | 15:15 | Subject Test Refused |
| Breath Volume     | 0.000 | 15:15 |                      |
| Air Blank         | 0.000 | 15:15 |                      |
| Control Test      | 0.082 | 15:15 |                      |
| Air Blank         | 0.000 | 15:16 |                      |
| Diagnostics Check | OK    | 15:16 |                      |

---

Data Download Date/Time: 06/16/2023 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROOKS,JOSEPH,P  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BASFORD,TRAVIS  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402469736  
DRY GAS EXPIRY DATE: 06/16/2025  
VIOLATION CODE: BUI  
ARREST TIME: 16:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:21

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 05/20/2023 16:49:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:42 |
| Air Blank          | 0.000 | 16:43 |
| Control Test       | 0.082 | 16:43 |
| Air Blank          | 0.000 | 16:44 |
| Subject Sample # 1 | 0.163 | 16:44 |
| Breath Volume      | 3.480 | 16:44 |
| Air Blank          | 0.000 | 16:45 |
| Air Blank          | 0.000 | 16:47 |
| Subject Sample # 2 | 0.162 | 16:47 |
| Breath Volume      | 3.324 | 16:47 |
| Air Blank          | 0.000 | 16:48 |
| Control Test       | 0.082 | 16:48 |
| Air Blank          | 0.000 | 16:49 |
| Diagnostics Check  | OK    | 16:49 |

---

Data Download Date/Time: 06/16/2023 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROOKS,JOSEPH,P  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: STINSON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402469736  
DRY GAS EXPIRY DATE: 06/16/2025  
VIOLATION CODE: BUI  
ARREST TIME: 16:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: DURMISEVIC

DATE/TIME OF TEST: 05/20/2023 17:34:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:27 |
| Air Blank          | 0.000 | 17:28 |
| Control Test       | 0.079 | 17:28 |
| Air Blank          | 0.000 | 17:29 |
| Subject Sample # 1 | 0.195 | 17:29 |
| Breath Volume      | 2.382 | 17:29 |
| Air Blank          | 0.000 | 17:30 |
| Air Blank          | 0.000 | 17:32 |
| Subject Sample # 2 | 0.179 | 17:32 |
| Breath Volume      | 1.437 | 17:32 |
| Air Blank          | 0.000 | 17:33 |
| Control Test       | 0.079 | 17:33 |
| Air Blank          | 0.000 | 17:34 |
| Diagnostics Check  | OK    | 17:34 |

---

Data Download Date/Time: 06/16/2023 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

|                                                  |                                   |
|--------------------------------------------------|-----------------------------------|
| INSTRUMENT REGISTERED TO: FFWCC                  | Date or Time Changed: No          |
| SOFTWARE VERSION: 8100.27                        | DRY GAS STANDARD LOT #: 402469736 |
| BREATH TEST OPERATOR: SENST,CASEY,J              | DRY GAS EXPIRY DATE: 06/16/2025   |
| BREATH TEST OPERATOR AGENCY: FFWCC (Panama City) | VIOLATION CODE: BUI               |
| ARRESTING OFFICER: PRITCHARD,THOMAS              | ARREST TIME: 17:06                |
| ARRESTING OFFICER AGENCY: FFWCC (Panama City)    | GENDER: Male                      |
| LAST AGENCY INSPECTION: 05/15/2023               | OBSERVATION PERIOD BEGAN: 17:18   |

|                               |                               |          |
|-------------------------------|-------------------------------|----------|
| SUBJECT'S LAST NAME: DESANTIS | DATE/TIME OF TEST: 05/20/2023 | 17:47:13 |
|-------------------------------|-------------------------------|----------|

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 17:40 |            |
| Air Blank         | 0.000     | 17:40 |            |
| Control Test      | 0.079     | 17:40 |            |
| Air Blank         | 0.000     | 17:41 |            |
| Subject Sample    | # 1 0.130 | 17:42 |            |
| Breath Volume     | 2.351     | 17:42 |            |
| Air Blank         | 0.000     | 17:43 |            |
| Air Blank         | 0.000     | 17:44 |            |
| Subject Sample    | # 2 0.137 | 17:45 |            |
| Breath Volume     | 3.207     | 17:45 |            |
| Air Blank         | 0.000     | 17:46 |            |
| Control Test      | 0.081     | 17:46 |            |
| Air Blank         |           | 17:46 | RFI Detect |
| Air Blank         | 0.000     | 17:47 |            |

---

Data Download Date/Time: 06/16/2023 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROOKS,JOSEPH,P  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BRITTON,CORBIN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402469736  
DRY GAS EXPIRY DATE: 06/16/2025  
VIOLATION CODE: BUI  
ARREST TIME: 20:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: DAILY

DATE/TIME OF TEST: 05/20/2023 21:26:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:19 |
| Air Blank          | 0.000 | 21:20 |
| Control Test       | 0.082 | 21:20 |
| Air Blank          | 0.000 | 21:21 |
| Subject Sample # 1 | 0.150 | 21:21 |
| Breath Volume      | 1.199 | 21:21 |
| Air Blank          | 0.000 | 21:22 |
| Air Blank          | 0.000 | 21:24 |
| Subject Sample # 2 | 0.146 | 21:24 |
| Breath Volume      | 3.039 | 21:24 |
| Air Blank          | 0.000 | 21:25 |
| Control Test       | 0.081 | 21:25 |
| Air Blank          | 0.000 | 21:26 |
| Diagnostics Check  | OK    | 21:26 |

---

Data Download Date/Time: 06/16/2023 09:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRAGG,DEVIN,M

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: BRAGG,DEVIN

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402469736

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 18:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: RIGSBY

DATE/TIME OF TEST: 05/25/2023 19:04:24

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:01 |                      |
| Air Blank         | 0.000 | 19:01 |                      |
| Control Test      | 0.083 | 19:02 |                      |
| Air Blank         | 0.000 | 19:02 |                      |
| Subject Sample    | # 1   | 19:03 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:03 |                      |
| Air Blank         | 0.000 | 19:03 |                      |
| Control Test      | 0.084 | 19:03 |                      |
| Air Blank         | 0.000 | 19:04 |                      |
| Diagnostics Check | OK    | 19:04 |                      |

---

Data Download Date/Time: 06/16/2023 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

|                                                  |                                   |
|--------------------------------------------------|-----------------------------------|
| INSTRUMENT REGISTERED TO: FFWCC                  | Date or Time Changed: No          |
| SOFTWARE VERSION: 8100.27                        | DRY GAS STANDARD LOT #: 402469736 |
| BREATH TEST OPERATOR: BAZYLAK,DAVID,C            | DRY GAS EXPIRY DATE: 06/16/2025   |
| BREATH TEST OPERATOR AGENCY: FFWCC (Panama City) | VIOLATION CODE: BUI               |
| ARRESTING OFFICER: POWELL,MICHAEL                | ARREST TIME: 17:15                |
| ARRESTING OFFICER AGENCY: FFWCC (Panama City)    | GENDER: Male                      |
| LAST AGENCY INSPECTION: 05/15/2023               | OBSERVATION PERIOD BEGAN: 17:15   |

|                           |                               |          |
|---------------------------|-------------------------------|----------|
| SUBJECT'S LAST NAME: JUDD | DATE/TIME OF TEST: 05/26/2023 | 17:42:16 |
|---------------------------|-------------------------------|----------|

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:36 |
| Air Blank          | 0.000 | 17:36 |
| Control Test       | 0.082 | 17:36 |
| Air Blank          | 0.000 | 17:37 |
| Subject Sample # 1 | 0.202 | 17:37 |
| Breath Volume      | 1.410 | 17:37 |
| Air Blank          | 0.000 | 17:38 |
| Air Blank          | 0.000 | 17:40 |
| Subject Sample # 2 | 0.210 | 17:40 |
| Breath Volume      | 1.253 | 17:40 |
| Air Blank          | 0.000 | 17:41 |
| Control Test       | 0.082 | 17:41 |
| Air Blank          | 0.000 | 17:42 |
| Diagnostics Check  | OK    | 17:42 |

---

Data Download Date/Time: 06/16/2023 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARTLETT,SHELTON,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BARTLETT,SHELTON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402469736  
DRY GAS EXPIRY DATE: 06/16/2025  
VIOLATION CODE: BUI  
ARREST TIME: 15:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: SLUDER

DATE/TIME OF TEST: 05/27/2023 16:00:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:53 |
| Air Blank          | 0.000 | 15:53 |
| Control Test       | 0.081 | 15:53 |
| Air Blank          | 0.000 | 15:54 |
| Subject Sample # 1 | 0.158 | 15:55 |
| Breath Volume      | 4.472 | 15:55 |
| Air Blank          | 0.000 | 15:56 |
| Air Blank          | 0.000 | 15:58 |
| Subject Sample # 2 | 0.154 | 15:58 |
| Breath Volume      | 3.164 | 15:58 |
| Air Blank          | 0.000 | 15:59 |
| Control Test       | 0.082 | 15:59 |
| Air Blank          | 0.000 | 16:00 |
| Diagnostics Check  | OK    | 16:00 |

---

Data Download Date/Time: 06/16/2023 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BAZYLAK,DAVID,C

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: BAZYLAK,DAVID

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402469736

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 16:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/27/2023 16:52:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 16:43 |                      |
| Air Blank         | 0.000 | 16:43 |                      |
| Control Test      | 0.083 | 16:44 |                      |
| Air Blank         | 0.000 | 16:44 |                      |
| Subject Sample    | # 1   | 16:47 | No Sample Provided   |
| Breath Volume     | 0.000 | 16:47 |                      |
| Air Blank         | 0.000 | 16:48 |                      |
| Air Blank         | 0.000 | 16:50 |                      |
| Subject Sample    | # 2   | 16:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 16:51 |                      |
| Air Blank         | 0.000 | 16:51 |                      |
| Control Test      | 0.083 | 16:51 |                      |
| Air Blank         | 0.000 | 16:52 |                      |
| Diagnostics Check | OK    | 16:52 |                      |

---

Data Download Date/Time: 06/16/2023 09:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROOKS,JOSEPH,P

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: PRITCHARD,THOMAS

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402469736

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 16:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: MALLIN

DATE/TIME OF TEST: 05/27/2023 17:25:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:18 |
| Air Blank          | 0.000 | 17:19 |
| Control Test       | 0.084 | 17:19 |
| Air Blank          | 0.000 | 17:20 |
| Subject Sample # 1 | 0.121 | 17:20 |
| Breath Volume      | 3.457 | 17:20 |
| Air Blank          | 0.000 | 17:21 |
| Air Blank          | 0.000 | 17:23 |
| Subject Sample # 2 | 0.119 | 17:23 |
| Breath Volume      | 3.179 | 17:23 |
| Air Blank          | 0.000 | 17:24 |
| Control Test       | 0.083 | 17:24 |
| Air Blank          | 0.000 | 17:25 |
| Diagnostics Check  | OK    | 17:25 |

---

Data Download Date/Time: 06/16/2023 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOLNAR,JARROD,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MOLNAR,JARROD  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402469736  
DRY GAS EXPIRY DATE: 06/16/2025  
VIOLATION CODE: BUI  
ARREST TIME: 17:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/27/2023 17:42:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:39 |                      |
| Air Blank         | 0.000 | 17:40 |                      |
| Control Test      | 0.083 | 17:40 |                      |
| Air Blank         | 0.000 | 17:40 |                      |
| Subject Sample    | # 1   | 17:41 | Subject Test Refused |
| Breath Volume     | 0.000 | 17:41 |                      |
| Air Blank         | 0.000 | 17:41 |                      |
| Control Test      | 0.080 | 17:41 |                      |
| Air Blank         | 0.000 | 17:42 |                      |
| Diagnostics Check | OK    | 17:42 |                      |

---

Data Download Date/Time: 06/16/2023 09:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STINSON,ERIC,S

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: DENNIS,CASEY

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402469736

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 17:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/27/2023 18:03:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:56 |
| Air Blank          | 0.000 | 17:56 |
| Control Test       | 0.085 | 17:56 |
| Air Blank          | 0.000 | 17:57 |
| Subject Sample # 1 | 0.162 | 17:59 |
| Breath Volume      | 2.054 | 17:59 |
| Air Blank          | 0.000 | 17:59 |
| Air Blank          | 0.000 | 18:01 |
| Subject Sample # 2 | 0.168 | 18:02 |
| Breath Volume      | 2.839 | 18:02 |
| Air Blank          | 0.000 | 18:02 |
| Control Test       | 0.082 | 18:02 |
| Air Blank          | 0.000 | 18:03 |
| Diagnostics Check  | OK    | 18:03 |

---

Data Download Date/Time: 06/16/2023 09:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: OLIVER, WILLIAM, H

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: OLIVER, WILLIAM

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402469736

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 18:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 05/27/2023 18:35:20

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:32 |                      |
| Air Blank         | 0.000 | 18:32 |                      |
| Control Test      | 0.080 | 18:32 |                      |
| Air Blank         | 0.000 | 18:33 |                      |
| Subject Sample    | # 1   | 18:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:33 |                      |
| Air Blank         | 0.000 | 18:34 |                      |
| Control Test      | 0.082 | 18:34 |                      |
| Air Blank         | 0.000 | 18:35 |                      |
| Diagnostics Check | OK    | 18:35 |                      |

Data Download Date/Time: 06/16/2023 09:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

|                                                  |                                   |
|--------------------------------------------------|-----------------------------------|
| INSTRUMENT REGISTERED TO: FFWCC                  | Date or Time Changed: No          |
| SOFTWARE VERSION: 8100.27                        | DRY GAS STANDARD LOT #: 402469736 |
| BREATH TEST OPERATOR: PRITCHARD,THOMAS,D         | DRY GAS EXPIRY DATE: 06/16/2025   |
| BREATH TEST OPERATOR AGENCY: FFWCC (Panama City) | VIOLATION CODE: BUI               |
| ARRESTING OFFICER: PRITCHARD,THOMAS              | ARREST TIME: 15:39                |
| ARRESTING OFFICER AGENCY: FFWCC (Panama City)    | GENDER: Male                      |
| LAST AGENCY INSPECTION: 05/15/2023               | OBSERVATION PERIOD BEGAN: 15:50   |

|                              |                               |          |
|------------------------------|-------------------------------|----------|
| SUBJECT'S LAST NAME: BECKNER | DATE/TIME OF TEST: 05/28/2023 | 16:16:41 |
|------------------------------|-------------------------------|----------|

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:10 |
| Air Blank          | 0.000 | 16:10 |
| Control Test       | 0.083 | 16:11 |
| Air Blank          | 0.000 | 16:11 |
| Subject Sample # 1 | 0.105 | 16:12 |
| Breath Volume      | 1.203 | 16:12 |
| Air Blank          | 0.000 | 16:12 |
| Air Blank          | 0.000 | 16:14 |
| Subject Sample # 2 | 0.100 | 16:15 |
| Breath Volume      | 1.460 | 16:15 |
| Air Blank          | 0.000 | 16:15 |
| Control Test       | 0.080 | 16:16 |
| Air Blank          | 0.000 | 16:16 |
| Diagnostics Check  | OK    | 16:16 |

---

Data Download Date/Time: 06/16/2023 09:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

|                                                 |                                   |
|-------------------------------------------------|-----------------------------------|
| INSTRUMENT REGISTERED TO: FFWCC                 | Date or Time Changed: No          |
| SOFTWARE VERSION: 8100.27                       | DRY GAS STANDARD LOT #: 402469736 |
| BREATH TEST OPERATOR: STINSON,ERIC,S            | DRY GAS EXPIRY DATE: 06/16/2025   |
| BREATH TEST OPERATOR AGENCY: Okaloosa County SO | VIOLATION CODE: BUI               |
| ARRESTING OFFICER: POWELL,MIKE                  | ARREST TIME: 17:30                |
| ARRESTING OFFICER AGENCY: Okaloosa County SO    | GENDER: Male                      |
| LAST AGENCY INSPECTION: 05/15/2023              | OBSERVATION PERIOD BEGAN: 17:34   |

|                               |                               |          |
|-------------------------------|-------------------------------|----------|
| SUBJECT'S LAST NAME: GRIFFITH | DATE/TIME OF TEST: 05/28/2023 | 18:03:57 |
|-------------------------------|-------------------------------|----------|

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:55 |                      |
| Air Blank         | 0.000 | 17:55 |                      |
| Control Test      | 0.084 | 17:56 |                      |
| Air Blank         | 0.000 | 17:56 |                      |
| Subject Sample    | # 1   | 17:59 | No Sample Provided   |
| Breath Volume     | 0.000 | 17:59 |                      |
| Air Blank         | 0.000 | 18:00 |                      |
| Air Blank         | 0.000 | 18:02 |                      |
| Subject Sample    | # 2   | 18:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:02 |                      |
| Air Blank         | 0.000 | 18:03 |                      |
| Control Test      | 0.084 | 18:03 |                      |
| Air Blank         | 0.000 | 18:03 |                      |
| Diagnostics Check | OK    | 18:03 |                      |

---

Data Download Date/Time: 06/16/2023 09:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BAZYLAK,DAVID,C

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: BAZYLAK,DAVID

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402469736

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 17:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: GALLOWAY

DATE/TIME OF TEST: 05/28/2023 18:21:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:13 |
| Air Blank          | 0.000 | 18:14 |
| Control Test       | 0.081 | 18:14 |
| Air Blank          | 0.000 | 18:15 |
| Subject Sample # 1 | 0.163 | 18:16 |
| Breath Volume      | 1.878 | 18:16 |
| Air Blank          | 0.000 | 18:17 |
| Air Blank          | 0.000 | 18:19 |
| Subject Sample # 2 | 0.154 | 18:20 |
| Breath Volume      | 1.621 | 18:20 |
| Air Blank          | 0.000 | 18:20 |
| Control Test       | 0.081 | 18:21 |
| Air Blank          | 0.000 | 18:21 |
| Diagnostics Check  | OK    | 18:21 |

---

Data Download Date/Time: 06/16/2023 09:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARTLETT,SHELTON,T

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: BRAGG,DEVIN

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402469736

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 18:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 05/28/2023 18:53:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:49 |                      |
| Air Blank         | 0.000 | 18:50 |                      |
| Control Test      | 0.082 | 18:50 |                      |
| Air Blank         | 0.000 | 18:50 |                      |
| Subject Sample    | # 1   | 18:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:52 |                      |
| Air Blank         | 0.000 | 18:52 |                      |
| Control Test      | 0.082 | 18:52 |                      |
| Air Blank         | 0.000 | 18:53 |                      |
| Diagnostics Check | OK    | 18:53 |                      |

---

Data Download Date/Time: 06/16/2023 09:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MURSCH,THEODORE,R  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MURSCH,THEODORE  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402469736  
DRY GAS EXPIRY DATE: 06/16/2025  
VIOLATION CODE: BUI  
ARREST TIME: 20:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: HASSMANN

DATE/TIME OF TEST: 05/28/2023 21:35:42

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 21:24 |                    |
| Air Blank         | 0.000 | 21:24 |                    |
| Control Test      | 0.084 | 21:25 |                    |
| Air Blank         | 0.000 | 21:25 |                    |
| Subject Sample    | # 1   | 21:28 | No Sample Provided |
| Breath Volume     | 0.000 | 21:28 |                    |
| Air Blank         | 0.000 | 21:29 |                    |
| Air Blank         | 0.000 | 21:31 |                    |
| Subject Sample    | # 2   | 21:34 | No Sample Provided |
| Breath Volume     | 0.000 | 21:34 |                    |
| Air Blank         | 0.000 | 21:34 |                    |
| Control Test      | 0.082 | 21:35 |                    |
| Air Blank         | 0.000 | 21:35 |                    |
| Diagnostics Check | OK    | 21:35 |                    |

---

Data Download Date/Time: 06/16/2023 09:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BASFORD,NICOLE,M

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: BASFORD,NICOLE

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: YEAGER

DATE/TIME OF TEST: 05/06/2023 19:08:50

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:04 |                      |
| Air Blank         | 0.000 | 19:05 |                      |
| Control Test      | 0.080 | 19:05 |                      |
| Air Blank         | 0.000 | 19:06 |                      |
| Subject Sample    | # 1   | 19:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:07 |                      |
| Air Blank         | 0.000 | 19:07 |                      |
| Control Test      | 0.079 | 19:08 |                      |
| Air Blank         | 0.000 | 19:08 |                      |
| Diagnostics Check | OK    | 19:08 |                      |

---

Data Download Date/Time: 05/15/2023 16:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAYCHOFF, COLLIN, T  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: GOODFELLOW, JESSICA  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: BUI  
ARREST TIME: 16:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/28/2023 17:46:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:39 |
| Air Blank          | 0.000 | 17:40 |
| Control Test       | 0.082 | 17:40 |
| Air Blank          | 0.000 | 17:41 |
| Subject Sample # 1 | 0.043 | 17:41 |
| Breath Volume      | 2.558 | 17:41 |
| Air Blank          | 0.000 | 17:42 |
| Air Blank          | 0.000 | 17:44 |
| Subject Sample # 2 | 0.042 | 17:44 |
| Breath Volume      | 2.597 | 17:44 |
| Air Blank          | 0.000 | 17:45 |
| Control Test       | 0.080 | 17:45 |
| Air Blank          | 0.000 | 17:46 |
| Diagnostics Check  | OK    | 17:46 |

---

Data Download Date/Time: 06/23/2023 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAYCHOFF,COLLIN,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: STARLING,BRENNAN  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: BUI  
ARREST TIME: 18:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: HAMM

DATE/TIME OF TEST: 05/28/2023 19:07:04

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 19:05 |            |
| Air Blank         | 0.000 | 19:05 |            |
| Control Test      | 0.081 | 19:06 |            |
| Air Blank         |       | 19:06 | RFI Detect |
| Air Blank         | 0.000 | 19:07 |            |

---

Data Download Date/Time: 06/23/2023 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PULASKI,DAMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 16:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: KULL

DATE/TIME OF TEST: 05/27/2023 17:19:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:13 |
| Air Blank          | 0.000 | 17:13 |
| Control Test       | 0.079 | 17:13 |
| Air Blank          | 0.000 | 17:14 |
| Subject Sample # 1 | 0.138 | 17:14 |
| Breath Volume      | 1.402 | 17:14 |
| Air Blank          | 0.000 | 17:15 |
| Air Blank          | 0.000 | 17:17 |
| Subject Sample # 2 | 0.142 | 17:18 |
| Breath Volume      | 2.125 | 17:18 |
| Air Blank          | 0.000 | 17:18 |
| Control Test       | 0.079 | 17:19 |
| Air Blank          | 0.000 | 17:19 |
| Diagnostics Check  | OK    | 17:19 |

---

Data Download Date/Time: 06/20/2023 13:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: BROWN,JASON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 16:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: MUCHLER

DATE/TIME OF TEST: 05/27/2023 17:31:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:25 |
| Air Blank          | 0.000 | 17:25 |
| Control Test       | 0.078 | 17:26 |
| Air Blank          | 0.000 | 17:26 |
| Subject Sample # 1 | 0.144 | 17:27 |
| Breath Volume      | 1.515 | 17:27 |
| Air Blank          | 0.000 | 17:27 |
| Air Blank          | 0.000 | 17:29 |
| Subject Sample # 2 | 0.154 | 17:30 |
| Breath Volume      | 2.953 | 17:30 |
| Air Blank          | 0.000 | 17:30 |
| Control Test       | 0.079 | 17:31 |
| Air Blank          | 0.000 | 17:31 |
| Diagnostics Check  | OK    | 17:31 |

---

Data Download Date/Time: 06/20/2023 13:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: MOSCHIANO,DAVID  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 19:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: LEO

DATE/TIME OF TEST: 05/27/2023 19:34:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:28 |
| Air Blank          | 0.000 | 19:28 |
| Control Test       | 0.078 | 19:29 |
| Air Blank          | 0.000 | 19:29 |
| Subject Sample # 1 | 0.088 | 19:30 |
| Breath Volume      | 1.425 | 19:30 |
| Air Blank          | 0.000 | 19:30 |
| Air Blank          | 0.000 | 19:32 |
| Subject Sample # 2 | 0.086 | 19:32 |
| Breath Volume      | 2.261 | 19:32 |
| Air Blank          | 0.000 | 19:33 |
| Control Test       | 0.078 | 19:33 |
| Air Blank          | 0.000 | 19:34 |
| Diagnostics Check  | OK    | 19:34 |

---

Data Download Date/Time: 06/20/2023 13:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PETRU,MIROS  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 19:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: STACHOWIAK

DATE/TIME OF TEST: 05/27/2023 20:14:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:07 |
| Air Blank          | 0.000 | 20:07 |
| Control Test       | 0.078 | 20:07 |
| Air Blank          | 0.000 | 20:08 |
| Subject Sample # 1 | 0.169 | 20:09 |
| Breath Volume      | 1.398 | 20:09 |
| Air Blank          | 0.000 | 20:10 |
| Air Blank          | 0.000 | 20:11 |
| Subject Sample # 2 | 0.185 | 20:12 |
| Breath Volume      | 1.769 | 20:12 |
| Air Blank          | 0.000 | 20:13 |
| Control Test       | 0.078 | 20:13 |
| Air Blank          | 0.000 | 20:13 |
| Diagnostics Check  | OK    | 20:14 |

---

Data Download Date/Time: 06/20/2023 13:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PETRU,MIROS  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: LARSEN

DATE/TIME OF TEST: 05/27/2023 22:29:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:22 |
| Air Blank          | 0.000 | 22:23 |
| Control Test       | 0.078 | 22:23 |
| Air Blank          | 0.000 | 22:24 |
| Subject Sample # 1 | 0.121 | 22:24 |
| Breath Volume      | 1.238 | 22:24 |
| Air Blank          | 0.000 | 22:25 |
| Air Blank          | 0.000 | 22:27 |
| Subject Sample # 2 | 0.138 | 22:28 |
| Breath Volume      | 1.671 | 22:28 |
| Air Blank          | 0.000 | 22:28 |
| Control Test       | 0.078 | 22:28 |
| Air Blank          | 0.000 | 22:29 |
| Diagnostics Check  | OK    | 22:29 |

---

Data Download Date/Time: 06/20/2023 13:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: MOSCHIANO,DAVID  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: FREYRE

DATE/TIME OF TEST: 05/27/2023 23:16:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:10 |
| Air Blank          | 0.000 | 23:10 |
| Control Test       | 0.078 | 23:10 |
| Air Blank          | 0.000 | 23:11 |
| Subject Sample # 1 | 0.036 | 23:12 |
| Breath Volume      | 2.496 | 23:12 |
| Air Blank          | 0.000 | 23:12 |
| Air Blank          | 0.000 | 23:14 |
| Subject Sample # 2 | 0.032 | 23:15 |
| Breath Volume      | 1.214 | 23:15 |
| Air Blank          | 0.000 | 23:15 |
| Control Test       | 0.078 | 23:15 |
| Air Blank          | 0.000 | 23:16 |
| Diagnostics Check  | OK    | 23:16 |

---

Data Download Date/Time: 06/20/2023 13:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PULASKI,DAMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 20:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 05/28/2023 21:13:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:07 |
| Air Blank          | 0.000 | 21:07 |
| Control Test       | 0.078 | 21:08 |
| Air Blank          | 0.000 | 21:08 |
| Subject Sample # 1 | 0.117 | 21:09 |
| Breath Volume      | 2.898 | 21:09 |
| Air Blank          | 0.000 | 21:09 |
| Air Blank          | 0.000 | 21:11 |
| Subject Sample # 2 | 0.118 | 21:12 |
| Breath Volume      | 2.851 | 21:12 |
| Air Blank          | 0.000 | 21:12 |
| Control Test       | 0.078 | 21:13 |
| Air Blank          | 0.000 | 21:13 |
| Diagnostics Check  | OK    | 21:13 |

---

Data Download Date/Time: 06/20/2023 13:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: SCINTA,MICHAEL  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 23:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: CONARD-ROBERTS

DATE/TIME OF TEST: 05/29/2023 00:21:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:15 |
| Air Blank          | 0.000 | 00:15 |
| Control Test       | 0.078 | 00:15 |
| Air Blank          | 0.000 | 00:16 |
| Subject Sample # 1 | 0.187 | 00:17 |
| Breath Volume      | 2.027 | 00:17 |
| Air Blank          | 0.000 | 00:17 |
| Air Blank          | 0.000 | 00:19 |
| Subject Sample # 2 | 0.192 | 00:20 |
| Breath Volume      | 2.890 | 00:20 |
| Air Blank          | 0.000 | 00:20 |
| Control Test       | 0.078 | 00:21 |
| Air Blank          | 0.000 | 00:21 |
| Diagnostics Check  | OK    | 00:21 |

---

Data Download Date/Time: 06/20/2023 13:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVIS,REECE,H  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: ALVIS,REECE  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 18:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: KLIMES

DATE/TIME OF TEST: 05/29/2023 19:40:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:34 |
| Air Blank          | 0.000 | 19:35 |
| Control Test       | 0.078 | 19:35 |
| Air Blank          | 0.000 | 19:35 |
| Subject Sample # 1 | 0.102 | 19:36 |
| Breath Volume      | 1.996 | 19:36 |
| Air Blank          | 0.000 | 19:36 |
| Air Blank          | 0.000 | 19:38 |
| Subject Sample # 2 | 0.105 | 19:39 |
| Breath Volume      | 3.437 | 19:39 |
| Air Blank          | 0.000 | 19:39 |
| Control Test       | 0.078 | 19:40 |
| Air Blank          | 0.000 | 19:40 |
| Diagnostics Check  | OK    | 19:40 |

---

Data Download Date/Time: 06/20/2023 13:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: SWINDELL,SEAN  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 18:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: BONDAR

DATE/TIME OF TEST: 05/29/2023 19:53:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:46 |
| Air Blank          | 0.000 | 19:47 |
| Control Test       | 0.078 | 19:47 |
| Air Blank          | 0.000 | 19:47 |
| Subject Sample # 1 | 0.106 | 19:48 |
| Breath Volume      | 2.460 | 19:48 |
| Air Blank          | 0.000 | 19:49 |
| Air Blank          | 0.000 | 19:51 |
| Subject Sample # 2 | 0.102 | 19:51 |
| Breath Volume      | 2.734 | 19:51 |
| Air Blank          | 0.000 | 19:52 |
| Control Test       | 0.077 | 19:52 |
| Air Blank          | 0.000 | 19:53 |
| Diagnostics Check  | OK    | 19:53 |

---

Data Download Date/Time: 06/20/2023 13:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: MOSCHIANO,DAVID  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 19:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: DOWNEY

DATE/TIME OF TEST: 05/29/2023 20:48:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:41 |
| Air Blank          | 0.000 | 20:42 |
| Control Test       | 0.078 | 20:42 |
| Air Blank          | 0.000 | 20:43 |
| Subject Sample # 1 | 0.199 | 20:43 |
| Breath Volume      | 1.812 | 20:43 |
| Air Blank          | 0.000 | 20:44 |
| Air Blank          | 0.000 | 20:46 |
| Subject Sample # 2 | 0.198 | 20:46 |
| Breath Volume      | 4.082 | 20:46 |
| Air Blank          | 0.000 | 20:47 |
| Control Test       | 0.078 | 20:47 |
| Air Blank          | 0.000 | 20:48 |
| Diagnostics Check  | OK    | 20:48 |

---

Data Download Date/Time: 06/20/2023 13:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLEY,KYLE,C  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: POLEY,KYLE  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: BUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 05/27/2023 18:32:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:26 |
| Air Blank          | 0.000 | 18:26 |
| Control Test       | 0.081 | 18:26 |
| Air Blank          | 0.000 | 18:27 |
| Subject Sample # 1 | 0.126 | 18:28 |
| Breath Volume      | 2.472 | 18:28 |
| Air Blank          | 0.000 | 18:28 |
| Air Blank          | 0.000 | 18:30 |
| Subject Sample # 2 | 0.125 | 18:30 |
| Breath Volume      | 2.636 | 18:30 |
| Air Blank          | 0.000 | 18:31 |
| Control Test       | 0.081 | 18:31 |
| Air Blank          | 0.000 | 18:32 |
| Diagnostics Check  | OK    | 18:32 |

---

Data Download Date/Time: 06/28/2023 13:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLEY,KYLE,C  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: EDGE,DANIEL  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: BUI  
ARREST TIME: 18:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/27/2023 22:45:40

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 22:42 |            |
| Air Blank         | 0.000 | 22:42 |            |
| Control Test      | 0.083 | 22:43 |            |
| Air Blank         | 0.000 | 22:43 |            |
| Subject Sample    | # 1   | 22:44 | RFI Detect |
| Breath Volume     | 0.394 | 22:44 |            |
| Air Blank         | 0.000 | 22:44 |            |
| Control Test      | 0.081 | 22:44 |            |
| Air Blank         |       | 22:45 | RFI Detect |
| Air Blank         | 0.000 | 22:45 |            |

---

Data Download Date/Time: 06/28/2023 13:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLEY,KYLE,C  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: EDGE,DANIEL  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: BUI  
ARREST TIME: 18:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/27/2023 22:54:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:48 |
| Air Blank          | 0.000 | 22:49 |
| Control Test       | 0.082 | 22:49 |
| Air Blank          | 0.000 | 22:49 |
| Subject Sample # 1 | 0.049 | 22:50 |
| Breath Volume      | 4.292 | 22:50 |
| Air Blank          | 0.000 | 22:51 |
| Air Blank          | 0.000 | 22:52 |
| Subject Sample # 2 | 0.042 | 22:53 |
| Breath Volume      | 2.472 | 22:53 |
| Air Blank          | 0.000 | 22:54 |
| Control Test       | 0.081 | 22:54 |
| Air Blank          | 0.000 | 22:54 |
| Diagnostics Check  | OK    | 22:54 |

---

Data Download Date/Time: 06/28/2023 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OROL,MANUEL,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: OROL,MANUEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402310313  
DRY GAS EXPIRY DATE: 01/15/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: ORELLANA

DATE/TIME OF TEST: 05/28/2023 01:18:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:16 |                      |
| Air Blank         | 0.000 | 01:16 |                      |
| Control Test      | 0.080 | 01:16 |                      |
| Air Blank         | 0.000 | 01:17 |                      |
| Subject Sample    | # 1   | 01:17 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:17 |                      |
| Air Blank         | 0.000 | 01:17 |                      |
| Control Test      | 0.079 | 01:18 |                      |
| Air Blank         | 0.000 | 01:18 |                      |
| Diagnostics Check | OK    | 01:18 |                      |

---

Data Download Date/Time: 05/31/2023 21:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,DAVID,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: BLAKE,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: BOURGEOIS

DATE/TIME OF TEST: 05/15/2023 02:48:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:41 |
| Air Blank          | 0.000 | 02:42 |
| Control Test       | 0.081 | 02:42 |
| Air Blank          | 0.000 | 02:43 |
| Subject Sample # 1 | 0.102 | 02:43 |
| Breath Volume      | 1.382 | 02:43 |
| Air Blank          | 0.000 | 02:44 |
| Air Blank          | 0.000 | 02:46 |
| Subject Sample # 2 | 0.100 | 02:46 |
| Breath Volume      | 1.324 | 02:46 |
| Air Blank          | 0.000 | 02:47 |
| Control Test       | 0.081 | 02:47 |
| Air Blank          | 0.000 | 02:48 |
| Diagnostics Check  | OK    | 02:48 |

---

Data Download Date/Time: 05/16/2023 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: SMITH, TRAEVONNE  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/25/2023 01:34:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.081 | 01:28 |
| Air Blank          | 0.000 | 01:29 |
| Subject Sample # 1 | 0.160 | 01:29 |
| Breath Volume      | 2.601 | 01:29 |
| Air Blank          | 0.000 | 01:30 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 2 | 0.173 | 01:33 |
| Breath Volume      | 2.542 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Control Test       | 0.081 | 01:34 |
| Air Blank          | 0.000 | 01:34 |
| Diagnostics Check  | OK    | 01:34 |

---

Data Download Date/Time: 06/21/2023 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOLFER,BROOKS,A  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WOLFER,BROOKS  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 05/27/2023 23:26:11

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 23:14 |                |
| Air Blank         | 0.000     | 23:14 |                |
| Control Test      | 0.081     | 23:15 |                |
| Air Blank         | 0.000     | 23:15 |                |
| Subject Sample    | # 1 0.069 | 23:18 | Volume Not Met |
| Breath Volume     | 0.019     | 23:18 |                |
| Air Blank         | 0.000     | 23:19 |                |
| Air Blank         | 0.000     | 23:21 |                |
| Subject Sample    | # 2 0.071 | 23:24 | Volume Not Met |
| Breath Volume     | 0.187     | 23:24 |                |
| Air Blank         | 0.000     | 23:25 |                |
| Control Test      | 0.080     | 23:25 |                |
| Air Blank         | 0.000     | 23:26 |                |
| Diagnostics Check | OK        | 23:26 |                |

---

Data Download Date/Time: 06/21/2023 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOLFER,BROOKS,A  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WOLFER,BROOKS  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 05/27/2023 23:36:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:28 |
| Air Blank          | 0.000 | 23:29 |
| Control Test       | 0.080 | 23:29 |
| Air Blank          | 0.000 | 23:30 |
| Subject Sample # 1 | 0.087 | 23:30 |
| Breath Volume      | 1.468 | 23:30 |
| Air Blank          | 0.000 | 23:31 |
| Air Blank          | 0.000 | 23:32 |
| Subject Sample # 2 | 0.089 | 23:35 |
| Breath Volume      | 1.394 | 23:35 |
| Air Blank          | 0.000 | 23:35 |
| Control Test       | 0.080 | 23:36 |
| Air Blank          | 0.000 | 23:36 |
| Diagnostics Check  | OK    | 23:36 |

---

Data Download Date/Time: 06/21/2023 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREENUP, RONNY,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: GREETHAM, CAMRON  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A3  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: SIDOR

DATE/TIME OF TEST: 05/29/2023 20:31:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:24 |
| Air Blank          | 0.000 | 20:25 |
| Control Test       | 0.081 | 20:25 |
| Air Blank          | 0.000 | 20:26 |
| Subject Sample # 1 | 0.114 | 20:26 |
| Breath Volume      | 1.882 | 20:26 |
| Air Blank          | 0.000 | 20:27 |
| Air Blank          | 0.000 | 20:29 |
| Subject Sample # 2 | 0.104 | 20:29 |
| Breath Volume      | 1.480 | 20:29 |
| Air Blank          | 0.000 | 20:30 |
| Control Test       | 0.081 | 20:30 |
| Air Blank          | 0.000 | 20:31 |
| Diagnostics Check  | OK    | 20:31 |

---

Data Download Date/Time: 06/21/2023 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,T,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: GUTIERREZ,I  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/03/2023 22:24:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:17 |
| Air Blank          | 0.000 | 22:18 |
| Control Test       | 0.080 | 22:18 |
| Air Blank          | 0.000 | 22:19 |
| Subject Sample # 1 | 0.163 | 22:19 |
| Breath Volume      | 2.851 | 22:19 |
| Air Blank          | 0.000 | 22:20 |
| Air Blank          | 0.000 | 22:22 |
| Subject Sample # 2 | 0.163 | 22:22 |
| Breath Volume      | 3.058 | 22:22 |
| Air Blank          | 0.000 | 22:23 |
| Control Test       | 0.081 | 22:23 |
| Air Blank          | 0.000 | 22:24 |
| Diagnostics Check  | OK    | 22:24 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENRRIQUEZ,A,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SILVA,L  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: ESTEBANVALOR

DATE/TIME OF TEST: 05/06/2023 01:37:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.081 | 01:31 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 1 | 0.172 | 01:32 |
| Breath Volume      | 1.949 | 01:32 |
| Air Blank          | 0.000 | 01:33 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 2 | 0.171 | 01:35 |
| Breath Volume      | 1.949 | 01:35 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.080 | 01:36 |
| Air Blank          | 0.000 | 01:37 |
| Diagnostics Check  | OK    | 01:37 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARGUELLES, ABRAHAM, A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ARGUELLES, ABRAHAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TOLEDO GARCIA

DATE/TIME OF TEST: 05/06/2023 03:50:44

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 03:43 |            |
| Air Blank         | 0.000     | 03:44 |            |
| Control Test      | 0.080     | 03:44 |            |
| Air Blank         | 0.000     | 03:44 |            |
| Subject Sample    | # 1 0.152 | 03:45 |            |
| Breath Volume     | 1.296     | 03:45 |            |
| Air Blank         | 0.000     | 03:46 |            |
| Air Blank         | 0.000     | 03:47 |            |
| Subject Sample    | # 2       | 03:49 | RFI Detect |
| Breath Volume     | 0.593     | 03:49 |            |
| Air Blank         | 0.000     | 03:49 |            |
| Control Test      | 0.078     | 03:50 |            |
| Air Blank         | 0.000     | 03:50 |            |
| Diagnostics Check | OK        | 03:50 |            |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARGUELLES, ABRAHAM, A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ARGUELLES, ABRAHAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TOLEDO GARCIA

DATE/TIME OF TEST: 05/06/2023 04:00:11

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:54 |                |
| Air Blank         | 0.000     | 03:54 |                |
| Control Test      | 0.078     | 03:54 |                |
| Air Blank         | 0.000     | 03:55 |                |
| Subject Sample    | # 1 0.124 | 03:58 | Volume Not Met |
| Breath Volume     | 0.496     | 03:58 |                |
| Air Blank         |           | 03:58 | RFI Detect     |
| Air Blank         | 0.000     | 03:59 |                |
| Control Test      | 0.079     | 03:59 |                |
| Air Blank         | 0.000     | 04:00 |                |
| Diagnostics Check | OK        | 04:00 |                |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARGUELLES, ABRAHAM, A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ARGUELLES, ABRAHAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TOLEDO GARCIA

DATE/TIME OF TEST: 05/06/2023 04:04:52

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 04:04 |            |
| Air Blank         |    | 04:04 | RFI Detect |
| Air Blank         |    | 04:04 | RFI Detect |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARGUELLES, ABRAHAM, A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ARGUELLES, ABRAHAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TOLEDO GARCIA

DATE/TIME OF TEST: 05/06/2023 04:08:11

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 04:07 |            |
| Air Blank         |    | 04:08 | RFI Detect |
| Air Blank         |    | 04:08 | RFI Detect |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARGUELLES, ABRAHAM, A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ARGUELLES, ABRAHAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TOLEDO GARCIA

DATE/TIME OF TEST: 05/06/2023 04:13:21

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 04:12 |            |
| Air Blank         |       | 04:12 | RFI Detect |
| Air Blank         | 0.000 | 04:13 |            |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARGUELLES, ABRAHAM, A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ARGUELLES, ABRAHAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TOLEDO GARCIA

DATE/TIME OF TEST: 05/06/2023 04:19:47

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 04:19 |            |
| Air Blank         |    | 04:19 | RFI Detect |
| Air Blank         |    | 04:19 | RFI Detect |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENRRIQUEZ,A,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: ESPIN,A  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: MORALESTREGO

DATE/TIME OF TEST: 05/07/2023 01:24:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:17 |
| Air Blank          | 0.000 | 01:18 |
| Control Test       | 0.080 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Subject Sample # 1 | 0.157 | 01:19 |
| Breath Volume      | 2.312 | 01:19 |
| Air Blank          | 0.000 | 01:20 |
| Air Blank          | 0.000 | 01:22 |
| Subject Sample # 2 | 0.168 | 01:22 |
| Breath Volume      | 2.964 | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Control Test       | 0.079 | 01:23 |
| Air Blank          | 0.000 | 01:24 |
| Diagnostics Check  | OK    | 01:24 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,S,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: VATH,L  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 3392208A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: HOUSE

DATE/TIME OF TEST: 05/07/2023 14:35:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:28 |
| Air Blank          | 0.000 | 14:29 |
| Control Test       | 0.081 | 14:29 |
| Air Blank          | 0.000 | 14:30 |
| Subject Sample # 1 | 0.345 | 14:30 |
| Breath Volume      | 2.242 | 14:30 |
| Air Blank          | 0.000 | 14:31 |
| Air Blank          | 0.000 | 14:33 |
| Subject Sample # 2 | 0.328 | 14:33 |
| Breath Volume      | 1.570 | 14:33 |
| Air Blank          | 0.000 | 14:34 |
| Control Test       | 0.080 | 14:34 |
| Air Blank          | 0.000 | 14:35 |
| Diagnostics Check  | OK    | 14:35 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORIETY,G,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: VELEZRODRIGUEZ,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 05/12/2023 02:57:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.081 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 1 | 0.125 | 02:51 |
| Breath Volume      | 1.593 | 02:51 |
| Air Blank          | 0.000 | 02:52 |
| Air Blank          | 0.000 | 02:54 |
| Subject Sample # 2 | 0.121 | 02:55 |
| Breath Volume      | 1.710 | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Control Test       | 0.080 | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Diagnostics Check  | OK    | 02:57 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUDASH, MATTHEW, B  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: NA, NA  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: OTHER  
ARREST TIME: 18:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: MALAV-LEE

DATE/TIME OF TEST: 05/12/2023 18:57:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:49 |
| Air Blank          | 0.000 | 18:50 |
| Control Test       | 0.080 | 18:50 |
| Air Blank          | 0.000 | 18:51 |
| Subject Sample # 1 | 0.045 | 18:52 |
| Breath Volume      | 1.812 | 18:52 |
| Air Blank          | 0.000 | 18:53 |
| Air Blank          | 0.000 | 18:55 |
| Subject Sample # 2 | 0.044 | 18:56 |
| Breath Volume      | 2.078 | 18:56 |
| Air Blank          | 0.000 | 18:56 |
| Control Test       | 0.080 | 18:56 |
| Air Blank          | 0.000 | 18:57 |
| Diagnostics Check  | OK    | 18:57 |

---

Data Download Date/Time: 06/02/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MENNINI,LILLIANA  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: GINTHER

DATE/TIME OF TEST: 05/12/2023 21:01:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:55 |
| Air Blank          | 0.000 | 20:56 |
| Control Test       | 0.081 | 20:56 |
| Air Blank          | 0.000 | 20:57 |
| Subject Sample # 1 | 0.032 | 20:57 |
| Breath Volume      | 1.750 | 20:57 |
| Air Blank          | 0.000 | 20:58 |
| Air Blank          | 0.000 | 20:59 |
| Subject Sample # 2 | 0.031 | 21:00 |
| Breath Volume      | 1.250 | 21:00 |
| Air Blank          | 0.000 | 21:00 |
| Control Test       | 0.080 | 21:01 |
| Air Blank          | 0.000 | 21:01 |
| Diagnostics Check  | OK    | 21:01 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,E,W  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: HILL,G  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: MEZA

DATE/TIME OF TEST: 05/14/2023 05:17:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Control Test       | 0.080 | 05:12 |
| Air Blank          | 0.000 | 05:13 |
| Subject Sample # 1 | 0.150 | 05:13 |
| Breath Volume      | 2.574 | 05:13 |
| Air Blank          | 0.000 | 05:14 |
| Air Blank          | 0.000 | 05:15 |
| Subject Sample # 2 | 0.149 | 05:16 |
| Breath Volume      | 2.757 | 05:16 |
| Air Blank          | 0.000 | 05:16 |
| Control Test       | 0.080 | 05:17 |
| Air Blank          | 0.000 | 05:17 |
| Diagnostics Check  | OK    | 05:17 |

---

Data Download Date/Time: 06/02/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,T,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: VALENTINE,E  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: NORTH

DATE/TIME OF TEST: 05/14/2023 21:46:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:39 |
| Air Blank          | 0.000 | 21:39 |
| Control Test       | 0.081 | 21:40 |
| Air Blank          | 0.000 | 21:40 |
| Subject Sample # 1 | 0.228 | 21:41 |
| Breath Volume      | 3.644 | 21:41 |
| Air Blank          | 0.000 | 21:42 |
| Air Blank          | 0.000 | 21:44 |
| Subject Sample # 2 | 0.225 | 21:44 |
| Breath Volume      | 2.488 | 21:44 |
| Air Blank          | 0.000 | 21:45 |
| Control Test       | 0.080 | 21:45 |
| Air Blank          | 0.000 | 21:46 |
| Diagnostics Check  | OK    | 21:46 |

---

Data Download Date/Time: 06/02/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AGUILAR,JOHN,M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PALMER,R  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: PIERRE PAUL

DATE/TIME OF TEST: 05/15/2023 01:20:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:14 |
| Air Blank          | 0.000 | 01:14 |
| Control Test       | 0.080 | 01:15 |
| Air Blank          | 0.000 | 01:15 |
| Subject Sample # 1 | 0.253 | 01:16 |
| Breath Volume      | 2.855 | 01:16 |
| Air Blank          | 0.000 | 01:16 |
| Air Blank          | 0.000 | 01:18 |
| Subject Sample # 2 | 0.249 | 01:19 |
| Breath Volume      | 2.792 | 01:19 |
| Air Blank          | 0.000 | 01:19 |
| Control Test       | 0.080 | 01:20 |
| Air Blank          | 0.000 | 01:20 |
| Diagnostics Check  | OK    | 01:20 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JAKUBEK, THOMAS, T  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: ACUNTO, A  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: WHITTAKER

DATE/TIME OF TEST: 05/17/2023 05:55:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:49 |
| Air Blank          | 0.000 | 05:50 |
| Control Test       | 0.081 | 05:50 |
| Air Blank          | 0.000 | 05:51 |
| Subject Sample # 1 | 0.369 | 05:51 |
| Breath Volume      | 2.542 | 05:51 |
| Air Blank          | 0.000 | 05:52 |
| Air Blank          | 0.000 | 05:53 |
| Subject Sample # 2 | 0.384 | 05:54 |
| Breath Volume      | 1.921 | 05:54 |
| Air Blank          | 0.000 | 05:54 |
| Control Test       | 0.077 | 05:55 |
| Air Blank          | 0.000 | 05:55 |
| Diagnostics Check  | OK    | 05:55 |

---

Data Download Date/Time: 06/02/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOPS,C,M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BLOMQUIST,M  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: SUTTER

DATE/TIME OF TEST: 05/20/2023 21:45:14

---

|                   |       |       |               |
|-------------------|-------|-------|---------------|
| Diagnostics Check | OK    | 21:42 |               |
| Air Blank         | 0.000 | 21:42 |               |
| Control Test      | 0.081 | 21:42 |               |
| Air Blank         | 0.000 | 21:43 |               |
| Subject Sample    | # 1   | 21:43 | Slope Not Met |
| Breath Volume     | 1.640 | 21:43 |               |
| Air Blank         | 0.000 | 21:44 |               |
| Control Test      | 0.080 | 21:44 |               |
| Air Blank         | 0.000 | 21:45 |               |
| Diagnostics Check | OK    | 21:45 |               |

---

Data Download Date/Time: 06/02/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOPS,C,M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BLOMQUIST,M  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: SUTTER

DATE/TIME OF TEST: 05/20/2023 21:54:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:48 |
| Air Blank          | 0.000 | 21:48 |
| Control Test       | 0.080 | 21:49 |
| Air Blank          | 0.000 | 21:49 |
| Subject Sample # 1 | 0.190 | 21:50 |
| Breath Volume      | 2.355 | 21:50 |
| Air Blank          | 0.000 | 21:50 |
| Air Blank          | 0.000 | 21:52 |
| Subject Sample # 2 | 0.190 | 21:53 |
| Breath Volume      | 1.410 | 21:53 |
| Air Blank          | 0.000 | 21:53 |
| Control Test       | 0.081 | 21:53 |
| Air Blank          | 0.000 | 21:54 |
| Diagnostics Check  | OK    | 21:54 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENRRIQUEZ,A,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SILVA,L  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: ROSAS NIETO

DATE/TIME OF TEST: 05/20/2023 23:19:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:12 |
| Air Blank          | 0.000 | 23:13 |
| Control Test       | 0.080 | 23:13 |
| Air Blank          | 0.000 | 23:14 |
| Subject Sample # 1 | 0.196 | 23:14 |
| Breath Volume      | 2.000 | 23:14 |
| Air Blank          | 0.000 | 23:15 |
| Air Blank          | 0.000 | 23:17 |
| Subject Sample # 2 | 0.194 | 23:17 |
| Breath Volume      | 1.832 | 23:17 |
| Air Blank          | 0.000 | 23:18 |
| Control Test       | 0.080 | 23:18 |
| Air Blank          | 0.000 | 23:19 |
| Diagnostics Check  | OK    | 23:19 |

---

Data Download Date/Time: 06/02/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUDASH, MATTHEW, B  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PEREZ, M  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: MONDESIR

DATE/TIME OF TEST: 05/24/2023 01:10:58

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 00:57 |                |
| Air Blank         | 0.000     | 00:58 |                |
| Control Test      | 0.080     | 00:58 |                |
| Air Blank         | 0.000     | 00:59 |                |
| Subject Sample    | # 1 0.135 | 01:02 | Volume Not Met |
| Breath Volume     | 0.277     | 01:02 |                |
| Air Blank         | 0.000     | 01:02 |                |
| Air Blank         | 0.000     | 01:04 |                |
| Subject Sample    | # 2 0.143 | 01:06 |                |
| Breath Volume     | 1.437     | 01:06 |                |
| Air Blank         | 0.000     | 01:06 |                |
| Air Blank         | 0.000     | 01:08 |                |
| Subject Sample    | # 3 0.146 | 01:09 |                |
| Breath Volume     | 1.343     | 01:09 |                |
| Air Blank         | 0.000     | 01:10 |                |
| Control Test      | 0.079     | 01:10 |                |
| Air Blank         | 0.000     | 01:10 |                |
| Diagnostics Check | OK        | 01:10 |                |

---

Data Download Date/Time: 06/02/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,J,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: HERNANDEZ,J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 07:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:08

SUBJECT'S LAST NAME: ABDUL-HAQQ

DATE/TIME OF TEST: 05/26/2023 13:42:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:36 |
| Air Blank          | 0.000 | 13:36 |
| Control Test       | 0.080 | 13:37 |
| Air Blank          | 0.000 | 13:37 |
| Subject Sample # 1 | 0.000 | 13:38 |
| Breath Volume      | 1.828 | 13:38 |
| Air Blank          | 0.000 | 13:38 |
| Air Blank          | 0.000 | 13:40 |
| Subject Sample # 2 | 0.000 | 13:41 |
| Breath Volume      | 2.285 | 13:41 |
| Air Blank          | 0.000 | 13:41 |
| Control Test       | 0.080 | 13:41 |
| Air Blank          | 0.000 | 13:42 |
| Diagnostics Check  | OK    | 13:42 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PECORELLA,R,K  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BELARDO,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ESPANA,MANUEL

DATE/TIME OF TEST: 05/27/2023 01:22:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:15 |
| Air Blank          | 0.000 | 01:16 |
| Control Test       | 0.080 | 01:16 |
| Air Blank          | 0.000 | 01:17 |
| Subject Sample # 1 | 0.163 | 01:17 |
| Breath Volume      | 2.769 | 01:17 |
| Air Blank          | 0.000 | 01:18 |
| Air Blank          | 0.000 | 01:20 |
| Subject Sample # 2 | 0.159 | 01:20 |
| Breath Volume      | 2.925 | 01:20 |
| Air Blank          | 0.000 | 01:21 |
| Control Test       | 0.079 | 01:21 |
| Air Blank          | 0.000 | 01:22 |
| Diagnostics Check  | OK    | 01:22 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: WILLIAMS,ZACHARY  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 05/27/2023 21:26:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:17 |
| Air Blank          | 0.000 | 21:17 |
| Control Test       | 0.080 | 21:18 |
| Air Blank          | 0.000 | 21:18 |
| Subject Sample # 1 | 0.088 | 21:20 |
| Breath Volume      | 1.117 | 21:20 |
| Air Blank          | 0.000 | 21:20 |
| Air Blank          | 0.000 | 21:22 |
| Subject Sample # 2 | 0.077 | 21:25 |
| Breath Volume      | 1.105 | 21:25 |
| Air Blank          | 0.000 | 21:25 |
| Control Test       | 0.079 | 21:26 |
| Air Blank          | 0.000 | 21:26 |
| Diagnostics Check  | OK    | 21:26 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNSBERGER,LISA,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: ALLEN,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/28/2023 03:25:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:18 |
| Air Blank          | 0.000 | 03:18 |
| Control Test       | 0.081 | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Subject Sample # 1 | 0.087 | 03:20 |
| Breath Volume      | 1.703 | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Air Blank          | 0.000 | 03:23 |
| Subject Sample # 2 | 0.093 | 03:24 |
| Breath Volume      | 2.042 | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.080 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Diagnostics Check  | OK    | 03:25 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON,AUSTIN,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: VELEZ,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 05/28/2023 06:40:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:33 |
| Air Blank          | 0.000 | 06:33 |
| Control Test       | 0.080 | 06:34 |
| Air Blank          | 0.000 | 06:34 |
| Subject Sample # 1 | 0.144 | 06:35 |
| Breath Volume      | 1.964 | 06:35 |
| Air Blank          | 0.000 | 06:35 |
| Air Blank          | 0.000 | 06:37 |
| Subject Sample # 2 | 0.146 | 06:38 |
| Breath Volume      | 1.371 | 06:38 |
| Air Blank          | 0.000 | 06:39 |
| Control Test       | 0.080 | 06:39 |
| Air Blank          | 0.000 | 06:40 |
| Diagnostics Check  | OK    | 06:40 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,JAMES,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: ABROMAITIS,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: FERREY

DATE/TIME OF TEST: 05/28/2023 07:04:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:58 |
| Air Blank          | 0.000 | 06:59 |
| Control Test       | 0.080 | 06:59 |
| Air Blank          | 0.000 | 06:59 |
| Subject Sample # 1 | 0.087 | 07:00 |
| Breath Volume      | 1.804 | 07:00 |
| Air Blank          | 0.000 | 07:01 |
| Air Blank          | 0.000 | 07:02 |
| Subject Sample # 2 | 0.085 | 07:03 |
| Breath Volume      | 1.925 | 07:03 |
| Air Blank          | 0.000 | 07:03 |
| Control Test       | 0.080 | 07:04 |
| Air Blank          | 0.000 | 07:04 |
| Diagnostics Check  | OK    | 07:04 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOWNSEND,ROBERT,B  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CAUDILL,E  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: LASHLEY

DATE/TIME OF TEST: 05/28/2023 15:26:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:20 |
| Air Blank          | 0.000 | 15:20 |
| Control Test       | 0.080 | 15:20 |
| Air Blank          | 0.000 | 15:21 |
| Subject Sample # 1 | 0.188 | 15:22 |
| Breath Volume      | 2.593 | 15:22 |
| Air Blank          | 0.000 | 15:22 |
| Air Blank          | 0.000 | 15:24 |
| Subject Sample # 2 | 0.183 | 15:25 |
| Breath Volume      | 2.648 | 15:25 |
| Air Blank          | 0.000 | 15:25 |
| Control Test       | 0.080 | 15:26 |
| Air Blank          | 0.000 | 15:26 |
| Diagnostics Check  | OK    | 15:26 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUDASH, MATTHEW, B  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PATTEN, R  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MATIAS

DATE/TIME OF TEST: 05/29/2023 00:34:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Control Test       | 0.081 | 00:27 |
| Air Blank          | 0.000 | 00:28 |
| Subject Sample # 1 | 0.194 | 00:29 |
| Breath Volume      | 2.109 | 00:29 |
| Air Blank          | 0.000 | 00:30 |
| Air Blank          | 0.000 | 00:32 |
| Subject Sample # 2 | 0.186 | 00:33 |
| Breath Volume      | 2.078 | 00:33 |
| Air Blank          | 0.000 | 00:33 |
| Control Test       | 0.080 | 00:34 |
| Air Blank          | 0.000 | 00:34 |
| Diagnostics Check  | OK    | 00:34 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AGUILAR,JOHN,M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PURSLEY,B.  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: HERNANDEZ-MOSQUERA

DATE/TIME OF TEST: 05/29/2023 05:00:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:51 |
| Air Blank          | 0.000 | 04:51 |
| Control Test       | 0.080 | 04:51 |
| Air Blank          | 0.000 | 04:52 |
| Subject Sample # 1 | 0.030 | 04:54 |
| Breath Volume      | 1.308 | 04:54 |
| Air Blank          | 0.000 | 04:55 |
| Air Blank          | 0.000 | 04:56 |
| Subject Sample # 2 | 0.029 | 04:58 |
| Breath Volume      | 1.742 | 04:58 |
| Air Blank          | 0.000 | 04:59 |
| Control Test       | 0.080 | 04:59 |
| Air Blank          | 0.000 | 04:59 |
| Diagnostics Check  | OK    | 05:00 |

---

Data Download Date/Time: 06/02/2023 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ALEJANDRO,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: YOREWITCH,KYLE  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 15:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: GOMEZ PEREZ

DATE/TIME OF TEST: 05/29/2023 17:31:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:24 |
| Air Blank          | 0.000 | 17:25 |
| Control Test       | 0.080 | 17:25 |
| Air Blank          | 0.000 | 17:26 |
| Subject Sample # 1 | 0.162 | 17:26 |
| Breath Volume      | 2.128 | 17:26 |
| Air Blank          | 0.000 | 17:27 |
| Air Blank          | 0.000 | 17:29 |
| Subject Sample # 2 | 0.144 | 17:29 |
| Breath Volume      | 1.363 | 17:29 |
| Air Blank          | 0.000 | 17:30 |
| Control Test       | 0.079 | 17:30 |
| Air Blank          | 0.000 | 17:30 |
| Diagnostics Check  | OK    | 17:31 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENRRIQUEZ,A,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MACIQUES,R  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 05/29/2023 22:34:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:25 |
| Air Blank          | 0.000 | 22:25 |
| Control Test       | 0.080 | 22:25 |
| Air Blank          | 0.000 | 22:26 |
| Subject Sample # 1 | 0.214 | 22:29 |
| Breath Volume      | 1.351 | 22:29 |
| Air Blank          | 0.000 | 22:30 |
| Air Blank          | 0.000 | 22:31 |
| Subject Sample # 2 | 0.206 | 22:32 |
| Breath Volume      | 1.988 | 22:32 |
| Air Blank          | 0.000 | 22:33 |
| Control Test       | 0.079 | 22:33 |
| Air Blank          | 0.000 | 22:34 |
| Diagnostics Check  | OK    | 22:34 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAHAGIAN,TARA,K  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: LITKA,A  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ROCHE

DATE/TIME OF TEST: 05/30/2023 02:51:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:44 |
| Air Blank          | 0.000 | 02:44 |
| Control Test       | 0.080 | 02:45 |
| Air Blank          | 0.000 | 02:45 |
| Subject Sample # 1 | 0.143 | 02:46 |
| Breath Volume      | 2.406 | 02:46 |
| Air Blank          | 0.000 | 02:47 |
| Air Blank          | 0.000 | 02:49 |
| Subject Sample # 2 | 0.150 | 02:50 |
| Breath Volume      | 2.007 | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.080 | 02:51 |
| Air Blank          | 0.000 | 02:51 |
| Diagnostics Check  | OK    | 02:51 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWELL, ADAM, J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: VELEZ, A  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: GOSEY

DATE/TIME OF TEST: 05/30/2023 23:20:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:13 |
| Air Blank          | 0.000 | 23:13 |
| Control Test       | 0.080 | 23:14 |
| Air Blank          | 0.000 | 23:14 |
| Subject Sample # 1 | 0.141 | 23:15 |
| Breath Volume      | 2.371 | 23:15 |
| Air Blank          | 0.000 | 23:16 |
| Air Blank          | 0.000 | 23:18 |
| Subject Sample # 2 | 0.126 | 23:18 |
| Breath Volume      | 1.589 | 23:18 |
| Air Blank          | 0.000 | 23:19 |
| Control Test       | 0.079 | 23:19 |
| Air Blank          | 0.000 | 23:20 |
| Diagnostics Check  | OK    | 23:20 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: STEARNS,M  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: DEFFLEY

DATE/TIME OF TEST: 05/31/2023 22:27:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:21 |
| Air Blank          | 0.000 | 22:21 |
| Control Test       | 0.080 | 22:22 |
| Air Blank          | 0.000 | 22:22 |
| Subject Sample # 1 | 0.198 | 22:23 |
| Breath Volume      | 2.105 | 22:23 |
| Air Blank          | 0.000 | 22:23 |
| Air Blank          | 0.000 | 22:25 |
| Subject Sample # 2 | 0.196 | 22:26 |
| Breath Volume      | 2.351 | 22:26 |
| Air Blank          | 0.000 | 22:26 |
| Control Test       | 0.080 | 22:27 |
| Air Blank          | 0.000 | 22:27 |
| Diagnostics Check  | OK    | 22:27 |

---

Data Download Date/Time: 06/02/2023 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARGUELLES, ABRAHAM, A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ARGUELLES, ABRAHAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06022080A4  
DRY GAS EXPIRY DATE: 04/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 08:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: RAMIREZ HERNANDEZ

DATE/TIME OF TEST: 05/02/2023 10:08:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:02 |
| Air Blank          | 0.000 | 10:02 |
| Control Test       | 0.078 | 10:03 |
| Air Blank          | 0.000 | 10:03 |
| Subject Sample # 1 | 0.187 | 10:04 |
| Breath Volume      | 1.757 | 10:04 |
| Air Blank          | 0.000 | 10:04 |
| Air Blank          | 0.000 | 10:06 |
| Subject Sample # 2 | 0.188 | 10:07 |
| Breath Volume      | 2.218 | 10:07 |
| Air Blank          | 0.000 | 10:08 |
| Control Test       | 0.078 | 10:08 |
| Air Blank          | 0.000 | 10:08 |
| Diagnostics Check  | OK    | 10:08 |

---

Data Download Date/Time: 05/03/2023 16:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARNOLD,TRAVIS,J  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: MARTINEZ,JOHNATHON  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06022080A4  
DRY GAS EXPIRY DATE: 04/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 14:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:54

SUBJECT'S LAST NAME: PERDOMO GUZMAN

DATE/TIME OF TEST: 05/02/2023 15:30:58

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 15:22 |                      |
| Air Blank         | 0.000     | 15:22 |                      |
| Control Test      | 0.079     | 15:23 |                      |
| Air Blank         | 0.000     | 15:23 |                      |
| Subject Sample    | # 1 0.156 | 15:26 | Volume Not Met       |
| Breath Volume     | 0.203     | 15:26 |                      |
| Air Blank         | 0.000     | 15:27 |                      |
| Air Blank         | 0.000     | 15:29 |                      |
| Subject Sample    | # 2       | 15:29 | Subject Test Refused |
| Breath Volume     | 0.000     | 15:29 |                      |
| Air Blank         | 0.000     | 15:30 |                      |
| Control Test      | 0.077     | 15:30 |                      |
| Air Blank         | 0.000     | 15:30 |                      |
| Diagnostics Check | OK        | 15:30 |                      |

---

Data Download Date/Time: 05/03/2023 16:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARGUELLES, ABRAHAM, A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ARGUELLES, ABRAHAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06022080A4  
DRY GAS EXPIRY DATE: 04/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/20/2023 16:04:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:57 |
| Air Blank          | 0.000 | 15:58 |
| Control Test       | 0.077 | 15:58 |
| Air Blank          | 0.000 | 15:58 |
| Subject Sample # 1 | 0.286 | 15:59 |
| Breath Volume      | 2.601 | 15:59 |
| Air Blank          | 0.000 | 16:00 |
| Air Blank          | 0.000 | 16:01 |
| Subject Sample # 2 | 0.283 | 16:02 |
| Breath Volume      | 2.546 | 16:02 |
| Air Blank          | 0.000 | 16:03 |
| Control Test       | 0.077 | 16:03 |
| Air Blank          | 0.000 | 16:04 |
| Diagnostics Check  | OK    | 16:04 |

---

Data Download Date/Time: 06/17/2023 09:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCNEIL,DARRIN,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: WATSON,PATRICK  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ULRICK

DATE/TIME OF TEST: 05/05/2023 00:31:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:25 |
| Air Blank          | 0.000 | 00:25 |
| Control Test       | 0.081 | 00:26 |
| Air Blank          | 0.000 | 00:26 |
| Subject Sample # 1 | 0.000 | 00:27 |
| Breath Volume      | 2.363 | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Air Blank          | 0.000 | 00:29 |
| Subject Sample # 2 | 0.000 | 00:30 |
| Breath Volume      | 2.625 | 00:30 |
| Air Blank          | 0.000 | 00:30 |
| Control Test       | 0.082 | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Diagnostics Check  | OK    | 00:31 |

---

Data Download Date/Time: 05/19/2023 13:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOUIS-PIERRE,RUBEN,  
BREATH TEST OPERATOR AGENCY: Clewiston PD  
ARRESTING OFFICER: LOUIS-PIERRE,RUBEN  
ARRESTING OFFICER AGENCY: Clewiston PD  
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: NICOLAS

DATE/TIME OF TEST: 05/05/2023 21:36:02

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 21:31 |                    |
| Air Blank         | 0.000 | 21:32 |                    |
| Control Test      | 0.081 | 21:32 |                    |
| Air Blank         | 0.000 | 21:32 |                    |
| Subject Sample    | # 1   | 21:34 | Interferent Detect |
| Breath Volume     | 1.636 | 21:34 |                    |
| Air Blank         |       | 21:35 | Purge Fail         |
| Air Blank         |       | 21:36 | Purge Fail         |

---

Data Download Date/Time: 05/19/2023 13:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOUIS-PIERRE,RUBEN,  
BREATH TEST OPERATOR AGENCY: Clewiston PD  
ARRESTING OFFICER: LOUIS-PIERRE,RUBEN  
ARRESTING OFFICER AGENCY: Clewiston PD  
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: NICOLAS

DATE/TIME OF TEST: 05/05/2023 21:43:32

---

|                   |           |       |              |
|-------------------|-----------|-------|--------------|
| Diagnostics Check | OK        | 21:40 |              |
| Air Blank         | 0.000     | 21:41 |              |
| Control Test      | 0.081     | 21:41 |              |
| Air Blank         | 0.000     | 21:42 |              |
| Subject Sample    | # 1 0.000 | 21:42 |              |
| Breath Volume     | 1.789     | 21:42 |              |
| Air Blank         |           | 21:43 | Ambient Fail |
| Air Blank         | 0.000     | 21:43 |              |

---

Data Download Date/Time: 05/19/2023 13:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: THIMOTHEE,CHARLES  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 05/07/2023 04:23:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:19 |                      |
| Air Blank         | 0.000 | 04:19 |                      |
| Control Test      | 0.081 | 04:19 |                      |
| Air Blank         | 0.000 | 04:20 |                      |
| Subject Sample    | # 1   | 04:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:22 |                      |
| Air Blank         | 0.000 | 04:22 |                      |
| Control Test      | 0.082 | 04:23 |                      |
| Air Blank         | 0.000 | 04:23 |                      |
| Diagnostics Check | OK    | 04:23 |                      |

---

Data Download Date/Time: 05/19/2023 13:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCNEIL,DARRIN,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: DREW,LIAM  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: VILLAZON

DATE/TIME OF TEST: 05/14/2023 03:38:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.082 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Subject Sample # 1 | 0.151 | 03:34 |
| Breath Volume      | 3.371 | 03:34 |
| Air Blank          | 0.000 | 03:34 |
| Air Blank          | 0.000 | 03:36 |
| Subject Sample # 2 | 0.150 | 03:37 |
| Breath Volume      | 2.589 | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.082 | 03:38 |
| Air Blank          | 0.000 | 03:38 |
| Diagnostics Check  | OK    | 03:38 |

---

Data Download Date/Time: 05/19/2023 13:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AFONSO,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: HULL,GARY  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 05/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 05/21/2023 05:53:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:46 |
| Air Blank          | 0.000 | 05:47 |
| Control Test       | 0.082 | 05:47 |
| Air Blank          | 0.000 | 05:48 |
| Subject Sample # 1 | 0.031 | 05:49 |
| Breath Volume      | 2.476 | 05:49 |
| Air Blank          | 0.000 | 05:49 |
| Air Blank          | 0.000 | 05:51 |
| Subject Sample # 2 | 0.030 | 05:52 |
| Breath Volume      | 2.234 | 05:52 |
| Air Blank          | 0.000 | 05:53 |
| Control Test       | 0.083 | 05:53 |
| Air Blank          | 0.000 | 05:53 |
| Diagnostics Check  | OK    | 05:53 |

---

Data Download Date/Time: 06/30/2023 10:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AFONSO,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: CASTRO,JUAN  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 05/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HAAPANEN

DATE/TIME OF TEST: 05/26/2023 03:15:34

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:12 |                      |
| Air Blank         | 0.000 | 03:13 |                      |
| Control Test      | 0.082 | 03:13 |                      |
| Air Blank         | 0.000 | 03:13 |                      |
| Subject Sample    | # 1   | 03:14 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:14 |                      |
| Air Blank         | 0.000 | 03:14 |                      |
| Control Test      | 0.081 | 03:14 |                      |
| Air Blank         | 0.000 | 03:15 |                      |
| Diagnostics Check | OK    | 03:15 |                      |

---

Data Download Date/Time: 06/30/2023 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEWITT,JARED,S  
BREATH TEST OPERATOR AGENCY: Franklin County SO  
ARRESTING OFFICER: LINDERMAN,TYLER  
ARRESTING OFFICER AGENCY: Franklin County SO  
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: MILLENDER

DATE/TIME OF TEST: 05/05/2023 21:40:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:33 |
| Air Blank          | 0.000 | 21:34 |
| Control Test       | 0.080 | 21:34 |
| Air Blank          | 0.000 | 21:34 |
| Subject Sample # 1 | 0.000 | 21:36 |
| Breath Volume      | 2.292 | 21:36 |
| Air Blank          | 0.000 | 21:36 |
| Air Blank          | 0.000 | 21:38 |
| Subject Sample # 2 | 0.000 | 21:39 |
| Breath Volume      | 2.144 | 21:39 |
| Air Blank          | 0.000 | 21:39 |
| Control Test       | 0.081 | 21:39 |
| Air Blank          | 0.000 | 21:40 |
| Diagnostics Check  | OK    | 21:40 |

---

Data Download Date/Time: 05/10/2023 12:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEWITT,JARED,S  
BREATH TEST OPERATOR AGENCY: Franklin County SO  
ARRESTING OFFICER: HEWITT,JARED  
ARRESTING OFFICER AGENCY: Franklin County SO  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 05/26/2023 00:43:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:37 |
| Air Blank          | 0.000 | 00:37 |
| Control Test       | 0.080 | 00:38 |
| Air Blank          | 0.000 | 00:38 |
| Subject Sample # 1 | 0.204 | 00:39 |
| Breath Volume      | 3.093 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Subject Sample # 2 | 0.198 | 00:42 |
| Breath Volume      | 3.562 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Control Test       | 0.081 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Diagnostics Check  | OK    | 00:43 |

---

Data Download Date/Time: 06/12/2023 15:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,PAUL,G  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: YOUNG,PAUL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 18:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/27/2023 20:03:30

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:00 |                      |
| Air Blank         | 0.000 | 20:01 |                      |
| Control Test      | 0.081 | 20:01 |                      |
| Air Blank         | 0.000 | 20:01 |                      |
| Subject Sample    | # 1   | 20:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:02 |                      |
| Air Blank         | 0.000 | 20:02 |                      |
| Control Test      | 0.081 | 20:02 |                      |
| Air Blank         | 0.000 | 20:03 |                      |
| Diagnostics Check | OK    | 20:03 |                      |

---

Data Download Date/Time: 06/12/2023 15:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STALLINGS,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: STALLINGS,JOSHUA  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 19:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: BRISBIN

DATE/TIME OF TEST: 05/27/2023 20:56:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:50 |
| Air Blank          | 0.000 | 20:50 |
| Control Test       | 0.081 | 20:51 |
| Air Blank          | 0.000 | 20:51 |
| Subject Sample # 1 | 0.136 | 20:51 |
| Breath Volume      | 3.054 | 20:51 |
| Air Blank          | 0.000 | 20:52 |
| Air Blank          | 0.000 | 20:54 |
| Subject Sample # 2 | 0.137 | 20:54 |
| Breath Volume      | 3.433 | 20:54 |
| Air Blank          | 0.000 | 20:55 |
| Control Test       | 0.080 | 20:55 |
| Air Blank          | 0.000 | 20:56 |
| Diagnostics Check  | OK    | 20:56 |

---

Data Download Date/Time: 06/12/2023 15:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILL,JULIA,D  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: WILL,JULIA  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HUANG

DATE/TIME OF TEST: 05/28/2023 00:06:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:59 |
| Air Blank          | 0.000 | 23:59 |
| Control Test       | 0.081 | 23:59 |
| Air Blank          | 0.000 | 00:00 |
| Subject Sample # 1 | 0.029 | 00:01 |
| Breath Volume      | 1.140 | 00:01 |
| Air Blank          | 0.000 | 00:02 |
| Air Blank          | 0.000 | 00:04 |
| Subject Sample # 2 | 0.027 | 00:05 |
| Breath Volume      | 1.796 | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Control Test       | 0.082 | 00:06 |
| Air Blank          | 0.000 | 00:06 |
| Diagnostics Check  | OK    | 00:06 |

---

Data Download Date/Time: 06/12/2023 15:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAMOS,TOMAS,G  
BREATH TEST OPERATOR AGENCY: Apalachicola PD  
ARRESTING OFFICER: TURNER,JOHN  
ARRESTING OFFICER AGENCY: Apalachicola PD  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 05/28/2023 04:25:30

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 04:20 |            |
| Air Blank         | 0.000     | 04:21 |            |
| Control Test      | 0.081     | 04:21 |            |
| Air Blank         | 0.000     | 04:22 |            |
| Subject Sample    | # 1 0.256 | 04:23 |            |
| Breath Volume     | 2.753     | 04:23 |            |
| Air Blank         |           | 04:24 | Purge Fail |
| Air Blank         |           | 04:25 | Purge Fail |

---

Data Download Date/Time: 06/12/2023 15:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAMOS,TOMAS,G  
BREATH TEST OPERATOR AGENCY: Apalachicola PD  
ARRESTING OFFICER: TURNER,JOHN  
ARRESTING OFFICER AGENCY: Apalachicola PD  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 05/28/2023 04:34:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:27 |
| Air Blank          | 0.000 | 04:28 |
| Control Test       | 0.076 | 04:28 |
| Air Blank          | 0.000 | 04:29 |
| Subject Sample # 1 | 0.240 | 04:29 |
| Breath Volume      | 1.285 | 04:29 |
| Air Blank          | 0.000 | 04:30 |
| Air Blank          | 0.000 | 04:32 |
| Subject Sample # 2 | 0.226 | 04:32 |
| Breath Volume      | 1.343 | 04:32 |
| Air Blank          | 0.000 | 04:33 |
| Control Test       | 0.076 | 04:33 |
| Air Blank          | 0.000 | 04:34 |
| Diagnostics Check  | OK    | 04:34 |

---

Data Download Date/Time: 06/12/2023 15:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAXEY,JORDAN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MAGAHA,JARRED  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BENN

DATE/TIME OF TEST: 05/04/2023 04:00:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.080 | 03:51 |
| Air Blank          | 0.000 | 03:51 |
| Subject Sample # 1 | 0.206 | 03:52 |
| Breath Volume      | 1.695 | 03:52 |
| Air Blank          | 0.000 | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Subject Sample # 2 | 0.178 | 03:55 |
| Breath Volume      | 1.402 | 03:55 |
| Air Blank          | 0.000 | 03:55 |
| Air Blank          | 0.000 | 03:57 |
| Subject Sample # 3 | 0.194 | 03:58 |
| Breath Volume      | 3.250 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Control Test       | 0.076 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Diagnostics Check  | OK    | 04:00 |

---

Data Download Date/Time: 05/11/2023 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON, GERALD, N  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: HARRISON, GERALD  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BRINEGAR

DATE/TIME OF TEST: 05/06/2023 00:50:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:42 |
| Air Blank          | 0.000 | 00:42 |
| Control Test       | 0.080 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 1 | 0.088 | 00:46 |
| Breath Volume      | 1.347 | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Air Blank          | 0.000 | 00:48 |
| Subject Sample # 2 | 0.080 | 00:49 |
| Breath Volume      | 1.398 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Control Test       | 0.079 | 00:50 |
| Air Blank          | 0.000 | 00:50 |
| Diagnostics Check  | OK    | 00:50 |

---

Data Download Date/Time: 05/11/2023 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COPELAND,MICHAEL,C  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: COPELAND,MICHAEL  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: HASAN

DATE/TIME OF TEST: 05/06/2023 03:53:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:43 |
| Air Blank          | 0.000 | 03:44 |
| Control Test       | 0.080 | 03:44 |
| Air Blank          | 0.000 | 03:44 |
| Subject Sample # 1 | 0.206 | 03:45 |
| Breath Volume      | 1.714 | 03:45 |
| Air Blank          | 0.000 | 03:45 |
| Air Blank          | 0.000 | 03:47 |
| Subject Sample # 2 | 0.177 | 03:48 |
| Breath Volume      | 1.820 | 03:48 |
| Air Blank          | 0.000 | 03:49 |
| Air Blank          | 0.000 | 03:51 |
| Subject Sample # 3 | 0.199 | 03:51 |
| Breath Volume      | 2.011 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Control Test       | 0.079 | 03:52 |
| Air Blank          | 0.000 | 03:53 |
| Diagnostics Check  | OK    | 03:53 |

---

Data Download Date/Time: 05/11/2023 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUSH, TYLER, C  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: GREENE, NATHAN  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: BIFFI

DATE/TIME OF TEST: 05/08/2023 05:12:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:06 |
| Air Blank          | 0.000 | 05:07 |
| Control Test       | 0.079 | 05:07 |
| Air Blank          | 0.000 | 05:08 |
| Subject Sample # 1 | 0.200 | 05:08 |
| Breath Volume      | 1.613 | 05:08 |
| Air Blank          | 0.000 | 05:09 |
| Air Blank          | 0.000 | 05:10 |
| Subject Sample # 2 | 0.198 | 05:11 |
| Breath Volume      | 1.742 | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Control Test       | 0.079 | 05:12 |
| Air Blank          | 0.000 | 05:12 |
| Diagnostics Check  | OK    | 05:12 |

---

Data Download Date/Time: 05/11/2023 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REAGLE,ASHLEY,R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: REAGLE,ASHLEY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MCCLLOUD

DATE/TIME OF TEST: 05/13/2023 01:34:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:28 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.080 | 01:29 |
| Air Blank          | 0.000 | 01:29 |
| Subject Sample # 1 | 0.219 | 01:30 |
| Breath Volume      | 1.714 | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 2 | 0.224 | 01:33 |
| Breath Volume      | 1.433 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Control Test       | 0.080 | 01:34 |
| Air Blank          | 0.000 | 01:34 |
| Diagnostics Check  | OK    | 01:34 |

---

Data Download Date/Time: 06/06/2023 11:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COPELAND,MICHAEL,C  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: COPELAND,MICHAEL  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: LILLY

DATE/TIME OF TEST: 05/13/2023 02:56:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.079 | 02:51 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 1 | 0.188 | 02:52 |
| Breath Volume      | 1.273 | 02:52 |
| Air Blank          | 0.000 | 02:52 |
| Air Blank          | 0.000 | 02:54 |
| Subject Sample # 2 | 0.187 | 02:54 |
| Breath Volume      | 2.921 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Control Test       | 0.079 | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Diagnostics Check  | OK    | 02:56 |

---

Data Download Date/Time: 06/06/2023 11:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHELPS,WILLIAM,D  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MUETHE,M  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: PINEDA ARIAS

DATE/TIME OF TEST: 05/13/2023 05:50:21

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 05:33 |                |
| Air Blank         | 0.000     | 05:34 |                |
| Control Test      | 0.079     | 05:34 |                |
| Air Blank         | 0.000     | 05:35 |                |
| Subject Sample    | # 1 0.169 | 05:38 | Volume Not Met |
| Breath Volume     | 0.464     | 05:38 |                |
| Air Blank         | 0.000     | 05:38 |                |
| Air Blank         | 0.000     | 05:40 |                |
| Subject Sample    | # 2 0.150 | 05:43 |                |
| Breath Volume     | 1.972     | 05:43 |                |
| Air Blank         | 0.000     | 05:43 |                |
| Air Blank         | 0.000     | 05:45 |                |
| Subject Sample    | # 3 0.159 | 05:48 | Volume Not Met |
| Breath Volume     | 0.265     | 05:48 |                |
| Air Blank         | 0.000     | 05:49 |                |
| Control Test      | 0.076     | 05:49 |                |
| Air Blank         | 0.000     | 05:50 |                |
| Diagnostics Check | OK        | 05:50 |                |

Data Download Date/Time: 06/06/2023 11:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHELPS, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: BROWN, D  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 05/13/2023 06:05:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:59 |
| Air Blank          | 0.000 | 05:59 |
| Control Test       | 0.078 | 06:00 |
| Air Blank          | 0.000 | 06:00 |
| Subject Sample # 1 | 0.142 | 06:01 |
| Breath Volume      | 2.730 | 06:01 |
| Air Blank          | 0.000 | 06:02 |
| Air Blank          | 0.000 | 06:03 |
| Subject Sample # 2 | 0.128 | 06:04 |
| Breath Volume      | 2.648 | 06:04 |
| Air Blank          | 0.000 | 06:05 |
| Control Test       | 0.078 | 06:05 |
| Air Blank          | 0.000 | 06:05 |
| Diagnostics Check  | OK    | 06:05 |

---

Data Download Date/Time: 06/06/2023 11:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHELPS, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROTEM, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: BOUIE

DATE/TIME OF TEST: 05/13/2023 06:21:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:14 |
| Air Blank          | 0.000 | 06:14 |
| Control Test       | 0.079 | 06:14 |
| Air Blank          | 0.000 | 06:15 |
| Subject Sample # 1 | 0.000 | 06:16 |
| Breath Volume      | 1.265 | 06:16 |
| Air Blank          | 0.000 | 06:17 |
| Air Blank          | 0.000 | 06:19 |
| Subject Sample # 2 | 0.000 | 06:19 |
| Breath Volume      | 2.199 | 06:19 |
| Air Blank          | 0.000 | 06:20 |
| Control Test       | 0.079 | 06:20 |
| Air Blank          | 0.000 | 06:20 |
| Diagnostics Check  | OK    | 06:21 |

---

Data Download Date/Time: 06/06/2023 11:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KISTNER,CRAIG,K  
BREATH TEST OPERATOR AGENCY: FL State University  
ARRESTING OFFICER: DIAZ,DIANA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 05/17/2023 03:24:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.080 | 03:17 |
| Air Blank          | 0.000 | 03:17 |
| Subject Sample # 1 | 0.182 | 03:18 |
| Breath Volume      | 1.125 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 2 | 0.187 | 03:22 |
| Breath Volume      | 1.542 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Control Test       | 0.080 | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Diagnostics Check  | OK    | 03:24 |

---

Data Download Date/Time: 06/06/2023 11:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,ALEXANDER,M  
BREATH TEST OPERATOR AGENCY: FL State University  
ARRESTING OFFICER: BELL,JACOB  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 05/18/2023 05:12:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:05 |
| Air Blank          | 0.000 | 05:05 |
| Control Test       | 0.078 | 05:05 |
| Air Blank          | 0.000 | 05:06 |
| Subject Sample # 1 | 0.189 | 05:07 |
| Breath Volume      | 3.507 | 05:07 |
| Air Blank          | 0.000 | 05:08 |
| Air Blank          | 0.000 | 05:09 |
| Subject Sample # 2 | 0.189 | 05:10 |
| Breath Volume      | 2.992 | 05:10 |
| Air Blank          | 0.000 | 05:11 |
| Control Test       | 0.078 | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Diagnostics Check  | OK    | 05:12 |

---

Data Download Date/Time: 06/06/2023 11:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MASHBURN, TONY  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 10:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: BRADSHAW

DATE/TIME OF TEST: 05/19/2023 13:06:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:00 |
| Air Blank          | 0.000 | 13:00 |
| Control Test       | 0.078 | 13:00 |
| Air Blank          | 0.000 | 13:01 |
| Subject Sample # 1 | 0.210 | 13:02 |
| Breath Volume      | 2.417 | 13:02 |
| Air Blank          | 0.000 | 13:02 |
| Air Blank          | 0.000 | 13:04 |
| Subject Sample # 2 | 0.200 | 13:05 |
| Breath Volume      | 2.699 | 13:05 |
| Air Blank          | 0.000 | 13:05 |
| Control Test       | 0.079 | 13:05 |
| Air Blank          | 0.000 | 13:06 |
| Diagnostics Check  | OK    | 13:06 |

---

Data Download Date/Time: 06/06/2023 11:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHELPS, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: QUINN, D  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/20/2023 04:18:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Control Test       | 0.081 | 04:12 |
| Air Blank          | 0.000 | 04:13 |
| Subject Sample # 1 | 0.150 | 04:13 |
| Breath Volume      | 1.117 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Air Blank          | 0.000 | 04:16 |
| Subject Sample # 2 | 0.147 | 04:16 |
| Breath Volume      | 1.304 | 04:16 |
| Air Blank          | 0.000 | 04:17 |
| Control Test       | 0.081 | 04:17 |
| Air Blank          | 0.000 | 04:18 |
| Diagnostics Check  | OK    | 04:18 |

---

Data Download Date/Time: 06/06/2023 11:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MATLOCK,LOGAN,R  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: MAGEAU,NATHAN  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CRUMPTON

DATE/TIME OF TEST: 05/27/2023 03:03:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:57 |
| Air Blank          | 0.000 | 02:57 |
| Control Test       | 0.080 | 02:57 |
| Air Blank          | 0.000 | 02:58 |
| Subject Sample # 1 | 0.281 | 02:58 |
| Breath Volume      | 1.742 | 02:58 |
| Air Blank          | 0.000 | 02:59 |
| Air Blank          | 0.000 | 03:01 |
| Subject Sample # 2 | 0.300 | 03:01 |
| Breath Volume      | 2.015 | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Control Test       | 0.080 | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Diagnostics Check  | OK    | 03:03 |

---

Data Download Date/Time: 06/06/2023 11:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REAGLE,ASHLEY,R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: REAGLE,ASHLEY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A4  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/28/2023 02:17:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:11 |
| Air Blank          | 0.000 | 02:11 |
| Control Test       | 0.080 | 02:12 |
| Air Blank          | 0.000 | 02:12 |
| Subject Sample # 1 | 0.192 | 02:13 |
| Breath Volume      | 2.968 | 02:13 |
| Air Blank          | 0.000 | 02:13 |
| Air Blank          | 0.000 | 02:15 |
| Subject Sample # 2 | 0.186 | 02:15 |
| Breath Volume      | 2.605 | 02:15 |
| Air Blank          | 0.000 | 02:16 |
| Control Test       | 0.080 | 02:16 |
| Air Blank          | 0.000 | 02:17 |
| Diagnostics Check  | OK    | 02:17 |

---

Data Download Date/Time: 06/06/2023 11:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REVELL,DENNIS,E  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: RITTMAN,DESMOND  
ARRESTING OFFICER AGENCY: Quincy PD  
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 05/07/2023 23:02:43

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 22:54 |                      |
| Air Blank         | 0.000     | 22:54 |                      |
| Control Test      | 0.084     | 22:55 |                      |
| Air Blank         | 0.000     | 22:55 |                      |
| Subject Sample    | # 1 0.202 | 22:58 | Volume Not Met       |
| Breath Volume     | 0.335     | 22:58 |                      |
| Air Blank         | 0.000     | 22:59 |                      |
| Air Blank         | 0.000     | 23:01 |                      |
| Subject Sample    | # 2       | 23:01 | Subject Test Refused |
| Breath Volume     | 0.000     | 23:01 |                      |
| Air Blank         | 0.000     | 23:01 |                      |
| Control Test      | 0.082     | 23:02 |                      |
| Air Blank         | 0.000     | 23:02 |                      |
| Diagnostics Check | OK        | 23:02 |                      |

---

Data Download Date/Time: 05/16/2023 15:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ, RESEAN, R  
BREATH TEST OPERATOR AGENCY: Midway PD  
ARRESTING OFFICER: CRUZ, RESEAN  
ARRESTING OFFICER AGENCY: Midway PD  
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/28/2023 00:12:02

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:07 |                      |
| Air Blank         | 0.000 | 00:08 |                      |
| Control Test      | 0.084 | 00:08 |                      |
| Air Blank         | 0.000 | 00:09 |                      |
| Subject Sample    | # 1   | 00:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:10 |                      |
| Air Blank         | 0.000 | 00:11 |                      |
| Control Test      | 0.083 | 00:11 |                      |
| Air Blank         | 0.000 | 00:11 |                      |
| Diagnostics Check | OK    | 00:12 |                      |

---

Data Download Date/Time: 06/08/2023 12:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LECKINGER, MICHAEL, E  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: LECKINGER, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 12:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:54

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 05/13/2023 13:20:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:14 |
| Air Blank          | 0.000 | 13:14 |
| Control Test       | 0.081 | 13:15 |
| Air Blank          | 0.000 | 13:15 |
| Subject Sample # 1 | 0.106 | 13:16 |
| Breath Volume      | 1.300 | 13:16 |
| Air Blank          | 0.000 | 13:17 |
| Air Blank          | 0.000 | 13:18 |
| Subject Sample # 2 | 0.109 | 13:19 |
| Breath Volume      | 1.367 | 13:19 |
| Air Blank          | 0.000 | 13:19 |
| Control Test       | 0.081 | 13:20 |
| Air Blank          | 0.000 | 13:20 |
| Diagnostics Check  | OK    | 13:20 |

---

Data Download Date/Time: 06/12/2023 15:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REAGLE,ASHLEY,R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: REAGLE,ASHLEY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 05/10/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 05/14/2023 03:25:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:19 |
| Air Blank          | 0.000 | 03:20 |
| Control Test       | 0.082 | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 1 | 0.134 | 03:21 |
| Breath Volume      | 3.437 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Subject Sample # 2 | 0.137 | 03:24 |
| Breath Volume      | 2.871 | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.082 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Diagnostics Check  | OK    | 03:25 |

---

Data Download Date/Time: 06/12/2023 15:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: NUSCIS,NICHOLAS  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BREYER

DATE/TIME OF TEST: 05/01/2023 00:34:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:28 |
| Air Blank          | 0.000 | 00:28 |
| Control Test       | 0.082 | 00:28 |
| Air Blank          | 0.000 | 00:29 |
| Subject Sample # 1 | 0.169 | 00:29 |
| Breath Volume      | 1.234 | 00:29 |
| Air Blank          | 0.000 | 00:30 |
| Air Blank          | 0.000 | 00:32 |
| Subject Sample # 2 | 0.167 | 00:33 |
| Breath Volume      | 1.554 | 00:33 |
| Air Blank          | 0.000 | 00:33 |
| Control Test       | 0.083 | 00:34 |
| Air Blank          | 0.000 | 00:34 |
| Diagnostics Check  | OK    | 00:34 |

---

Data Download Date/Time: 05/19/2023 12:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: ARROYO,ALEJANDRO  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/01/2023 01:27:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:20 |
| Air Blank          | 0.000 | 01:21 |
| Control Test       | 0.083 | 01:21 |
| Air Blank          | 0.000 | 01:22 |
| Subject Sample # 1 | 0.246 | 01:22 |
| Breath Volume      | 2.406 | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Air Blank          | 0.000 | 01:25 |
| Subject Sample # 2 | 0.243 | 01:26 |
| Breath Volume      | 2.449 | 01:26 |
| Air Blank          | 0.000 | 01:26 |
| Control Test       | 0.082 | 01:27 |
| Air Blank          | 0.000 | 01:27 |
| Diagnostics Check  | OK    | 01:27 |

---

Data Download Date/Time: 05/19/2023 12:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HANSELL,ROBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: DANKER

DATE/TIME OF TEST: 05/01/2023 23:48:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:42 |
| Air Blank          | 0.000 | 23:42 |
| Control Test       | 0.083 | 23:43 |
| Air Blank          | 0.000 | 23:43 |
| Subject Sample # 1 | 0.000 | 23:44 |
| Breath Volume      | 1.179 | 23:44 |
| Air Blank          | 0.000 | 23:44 |
| Air Blank          | 0.000 | 23:46 |
| Subject Sample # 2 | 0.000 | 23:47 |
| Breath Volume      | 1.597 | 23:47 |
| Air Blank          | 0.000 | 23:47 |
| Control Test       | 0.083 | 23:47 |
| Air Blank          | 0.000 | 23:48 |
| Diagnostics Check  | OK    | 23:48 |

---

Data Download Date/Time: 05/19/2023 12:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HELLER,MICHAEL  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 05/02/2023 00:26:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:20 |
| Air Blank          | 0.000 | 00:20 |
| Control Test       | 0.083 | 00:20 |
| Air Blank          | 0.000 | 00:21 |
| Subject Sample # 1 | 0.020 | 00:22 |
| Breath Volume      | 2.976 | 00:22 |
| Air Blank          | 0.000 | 00:22 |
| Air Blank          | 0.000 | 00:24 |
| Subject Sample # 2 | 0.015 | 00:25 |
| Breath Volume      | 1.914 | 00:25 |
| Air Blank          | 0.000 | 00:25 |
| Control Test       | 0.082 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Diagnostics Check  | OK    | 00:26 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: GRIEVE,ALEX  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: RESENDIZ-FRANCO

DATE/TIME OF TEST: 05/02/2023 23:25:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:18 |
| Air Blank          | 0.000 | 23:18 |
| Control Test       | 0.083 | 23:18 |
| Air Blank          | 0.000 | 23:19 |
| Subject Sample # 1 | 0.218 | 23:20 |
| Breath Volume      | 2.222 | 23:20 |
| Air Blank          | 0.000 | 23:21 |
| Air Blank          | 0.000 | 23:22 |
| Subject Sample # 2 | 0.200 | 23:23 |
| Breath Volume      | 2.027 | 23:23 |
| Air Blank          | 0.000 | 23:24 |
| Control Test       | 0.080 | 23:24 |
| Air Blank          | 0.000 | 23:25 |
| Diagnostics Check  | OK    | 23:25 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,CARLOS,I  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: WHOBREY,SPENCER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: RODRIGUEZRODRIGUEZ

DATE/TIME OF TEST: 05/05/2023 01:01:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:55 |
| Air Blank          | 0.000 | 00:56 |
| Control Test       | 0.083 | 00:56 |
| Air Blank          | 0.000 | 00:57 |
| Subject Sample # 1 | 0.176 | 00:57 |
| Breath Volume      | 2.613 | 00:57 |
| Air Blank          | 0.000 | 00:58 |
| Air Blank          | 0.000 | 00:59 |
| Subject Sample # 2 | 0.177 | 01:00 |
| Breath Volume      | 1.914 | 01:00 |
| Air Blank          | 0.000 | 01:01 |
| Control Test       | 0.082 | 01:01 |
| Air Blank          | 0.000 | 01:01 |
| Diagnostics Check  | OK    | 01:01 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCANLAN,JAMES,M  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HANSELL,ROBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/05/2023 03:36:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Control Test       | 0.083 | 03:29 |
| Air Blank          | 0.000 | 03:29 |
| Subject Sample # 1 | 0.115 | 03:31 |
| Breath Volume      | 1.652 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Air Blank          | 0.000 | 03:34 |
| Subject Sample # 2 | 0.107 | 03:34 |
| Breath Volume      | 1.906 | 03:34 |
| Air Blank          | 0.000 | 03:35 |
| Control Test       | 0.080 | 03:35 |
| Air Blank          | 0.000 | 03:36 |
| Diagnostics Check  | OK    | 03:36 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HANSELL,ROBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: ATEHORTUA

DATE/TIME OF TEST: 05/05/2023 23:02:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:55 |
| Air Blank          | 0.000 | 22:55 |
| Control Test       | 0.083 | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Subject Sample # 1 | 0.000 | 22:57 |
| Breath Volume      | 1.261 | 22:57 |
| Air Blank          | 0.000 | 22:57 |
| Air Blank          | 0.000 | 22:59 |
| Subject Sample # 2 | 0.000 | 23:00 |
| Breath Volume      | 1.472 | 23:00 |
| Air Blank          | 0.000 | 23:01 |
| Control Test       | 0.082 | 23:01 |
| Air Blank          | 0.000 | 23:02 |
| Diagnostics Check  | OK    | 23:02 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: BEEKMAN,JOSH  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: FREYTES-GONZALEZ

DATE/TIME OF TEST: 05/06/2023 01:55:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.082 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Subject Sample # 1 | 0.118 | 01:50 |
| Breath Volume      | 2.308 | 01:50 |
| Air Blank          | 0.000 | 01:51 |
| Air Blank          | 0.000 | 01:53 |
| Subject Sample # 2 | 0.113 | 01:53 |
| Breath Volume      | 1.285 | 01:53 |
| Air Blank          | 0.000 | 01:54 |
| Control Test       | 0.080 | 01:54 |
| Air Blank          | 0.000 | 01:55 |
| Diagnostics Check  | OK    | 01:55 |

---

Data Download Date/Time: 05/19/2023 12:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: ROMERO-PEREZ,MICHAEL  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: COLLADO-ALMONTE

DATE/TIME OF TEST: 05/07/2023 00:32:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Control Test       | 0.083 | 00:27 |
| Air Blank          | 0.000 | 00:28 |
| Subject Sample # 1 | 0.017 | 00:28 |
| Breath Volume      | 3.925 | 00:28 |
| Air Blank          | 0.000 | 00:29 |
| Air Blank          | 0.000 | 00:30 |
| Subject Sample # 2 | 0.016 | 00:31 |
| Breath Volume      | 4.238 | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.082 | 00:32 |
| Air Blank          | 0.000 | 00:32 |
| Diagnostics Check  | OK    | 00:32 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HANSELL,ROBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: CARRILLO-PULIDO

DATE/TIME OF TEST: 05/07/2023 01:42:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:36 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.082 | 01:36 |
| Air Blank          | 0.000 | 01:37 |
| Subject Sample # 1 | 0.234 | 01:38 |
| Breath Volume      | 1.824 | 01:38 |
| Air Blank          | 0.000 | 01:38 |
| Air Blank          | 0.000 | 01:40 |
| Subject Sample # 2 | 0.247 | 01:41 |
| Breath Volume      | 3.464 | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.082 | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Diagnostics Check  | OK    | 01:42 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,CHRISTIAN,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: DEJESUS,JOSEPH  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/07/2023 07:44:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:37 |
| Air Blank          | 0.000 | 07:37 |
| Control Test       | 0.082 | 07:37 |
| Air Blank          | 0.000 | 07:38 |
| Subject Sample # 1 | 0.189 | 07:39 |
| Breath Volume      | 2.355 | 07:39 |
| Air Blank          | 0.000 | 07:40 |
| Air Blank          | 0.000 | 07:41 |
| Subject Sample # 2 | 0.180 | 07:43 |
| Breath Volume      | 2.671 | 07:43 |
| Air Blank          | 0.000 | 07:43 |
| Control Test       | 0.080 | 07:44 |
| Air Blank          | 0.000 | 07:44 |
| Diagnostics Check  | OK    | 07:44 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: MADRIGAL,KRISTYANN  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 05/08/2023 00:56:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:50 |
| Air Blank          | 0.000 | 00:51 |
| Control Test       | 0.083 | 00:51 |
| Air Blank          | 0.000 | 00:52 |
| Subject Sample # 1 | 0.063 | 00:52 |
| Breath Volume      | 4.183 | 00:52 |
| Air Blank          | 0.000 | 00:53 |
| Air Blank          | 0.000 | 00:54 |
| Subject Sample # 2 | 0.061 | 00:55 |
| Breath Volume      | 4.359 | 00:55 |
| Air Blank          | 0.000 | 00:55 |
| Control Test       | 0.081 | 00:56 |
| Air Blank          | 0.000 | 00:56 |
| Diagnostics Check  | OK    | 00:56 |

---

Data Download Date/Time: 05/19/2023 12:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HELLER,MICHAEL  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CARABALLO-OSTOLAZA

DATE/TIME OF TEST: 05/08/2023 03:46:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:40 |
| Air Blank          | 0.000 | 03:41 |
| Control Test       | 0.083 | 03:41 |
| Air Blank          | 0.000 | 03:42 |
| Subject Sample # 1 | 0.000 | 03:42 |
| Breath Volume      | 2.808 | 03:42 |
| Air Blank          | 0.000 | 03:42 |
| Air Blank          | 0.000 | 03:44 |
| Subject Sample # 2 | 0.000 | 03:45 |
| Breath Volume      | 2.960 | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.081 | 03:46 |
| Air Blank          | 0.000 | 03:46 |
| Diagnostics Check  | OK    | 03:46 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTO,BRANDON,J  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HELLER,MICHAEL  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 05/12/2023 05:18:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:10 |
| Air Blank          | 0.000 | 05:11 |
| Control Test       | 0.083 | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Subject Sample # 1 | 0.102 | 05:13 |
| Breath Volume      | 2.710 | 05:13 |
| Air Blank          | 0.000 | 05:14 |
| Air Blank          | 0.000 | 05:15 |
| Subject Sample # 2 | 0.086 | 05:16 |
| Breath Volume      | 1.308 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Control Test       | 0.080 | 05:17 |
| Air Blank          | 0.000 | 05:17 |
| Diagnostics Check  | OK    | 05:18 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: NUSCIS,NICHOLAS  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: MAYSONET-CABRERA

DATE/TIME OF TEST: 05/12/2023 23:39:56

---

|                   |       |       |                  |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK    | 23:39 |                  |
| Air Blank         |       | 23:39 | Sequence Aborted |
| Air Blank         | 0.000 | 23:39 |                  |

---

Data Download Date/Time: 05/19/2023 12:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: NUSCIS,NICHOLAS  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: MAYSONET-CABRERA

DATE/TIME OF TEST: 05/12/2023 23:55:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:48 |
| Air Blank          | 0.000 | 23:49 |
| Control Test       | 0.082 | 23:49 |
| Air Blank          | 0.000 | 23:49 |
| Subject Sample # 1 | 0.211 | 23:50 |
| Breath Volume      | 1.332 | 23:50 |
| Air Blank          | 0.000 | 23:51 |
| Air Blank          | 0.000 | 23:53 |
| Subject Sample # 2 | 0.212 | 23:53 |
| Breath Volume      | 1.292 | 23:53 |
| Air Blank          | 0.000 | 23:54 |
| Control Test       | 0.082 | 23:54 |
| Air Blank          | 0.000 | 23:55 |
| Diagnostics Check  | OK    | 23:55 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERRON,BENJAMIN,R  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: GARCIA,CORINA  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402330433  
DRY GAS EXPIRY DATE: 01/24/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/13/2023 04:19:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:11 |
| Air Blank          | 0.000 | 04:11 |
| Control Test       | 0.082 | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Subject Sample # 1 | 0.183 | 04:14 |
| Breath Volume      | 1.445 | 04:14 |
| Air Blank          | 0.000 | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Subject Sample # 2 | 0.172 | 04:18 |
| Breath Volume      | 1.437 | 04:18 |
| Air Blank          | 0.000 | 04:18 |
| Control Test       | 0.080 | 04:19 |
| Air Blank          | 0.000 | 04:19 |
| Diagnostics Check  | OK    | 04:19 |

---

Data Download Date/Time: 05/19/2023 12:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREENWALL,ROBERT,W  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: RAMIREZ,JOSE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MARTINEZ BERRIO

DATE/TIME OF TEST: 05/14/2023 02:11:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:03 |
| Air Blank          | 0.000 | 02:03 |
| Control Test       | 0.083 | 02:04 |
| Air Blank          | 0.000 | 02:04 |
| Subject Sample # 1 | 0.236 | 02:06 |
| Breath Volume      | 1.917 | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Air Blank          | 0.000 | 02:08 |
| Subject Sample # 2 | 0.223 | 02:09 |
| Breath Volume      | 1.976 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Control Test       | 0.082 | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Diagnostics Check  | OK    | 02:11 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: ARROYO,ALEJANDRO  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 05/14/2023 02:22:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:15 |
| Air Blank          | 0.000 | 02:15 |
| Control Test       | 0.082 | 02:15 |
| Air Blank          | 0.000 | 02:16 |
| Subject Sample # 1 | 0.062 | 02:17 |
| Breath Volume      | 2.519 | 02:17 |
| Air Blank          | 0.000 | 02:18 |
| Air Blank          | 0.000 | 02:20 |
| Subject Sample # 2 | 0.064 | 02:20 |
| Breath Volume      | 3.078 | 02:20 |
| Air Blank          | 0.000 | 02:21 |
| Control Test       | 0.082 | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Diagnostics Check  | OK    | 02:22 |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: WHOBREY,SPENCER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 14:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: PEREZ-COLON

DATE/TIME OF TEST: 05/14/2023 15:54:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:47 |
| Air Blank          | 0.000 | 15:48 |
| Control Test       | 0.082 | 15:48 |
| Air Blank          | 0.000 | 15:49 |
| Subject Sample # 1 | 0.172 | 15:49 |
| Breath Volume      | 2.609 | 15:49 |
| Air Blank          | 0.000 | 15:50 |
| Air Blank          | 0.000 | 15:52 |
| Subject Sample # 2 | 0.172 | 15:52 |
| Breath Volume      | 1.933 | 15:52 |
| Air Blank          | 0.000 | 15:53 |
| Control Test       | 0.082 | 15:53 |
| Air Blank          | 0.000 | 15:54 |
| Diagnostics Check  | OK    | 15:54 |

---

Data Download Date/Time: 05/19/2023 12:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTO,BRANDON,J  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: NIETO,MICHAEL  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: MAJOR

DATE/TIME OF TEST: 05/16/2023 04:46:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:42 |                      |
| Air Blank         | 0.000 | 04:42 |                      |
| Control Test      | 0.083 | 04:42 |                      |
| Air Blank         | 0.000 | 04:43 |                      |
| Subject Sample    | # 1   | 04:44 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:44 |                      |
| Air Blank         | 0.000 | 04:45 |                      |
| Control Test      | 0.084 | 04:45 |                      |
| Air Blank         | 0.000 | 04:46 |                      |
| Diagnostics Check | OK    | 04:46 |                      |

---

Data Download Date/Time: 05/19/2023 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTO,BRANDON,J  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HELLER,MICHAEL  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 05/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: MORENO DURAN

DATE/TIME OF TEST: 05/20/2023 05:33:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:27 |
| Air Blank          | 0.000 | 05:27 |
| Control Test       | 0.083 | 05:28 |
| Air Blank          | 0.000 | 05:28 |
| Subject Sample # 1 | 0.170 | 05:29 |
| Breath Volume      | 1.839 | 05:29 |
| Air Blank          | 0.000 | 05:29 |
| Air Blank          | 0.000 | 05:31 |
| Subject Sample # 2 | 0.183 | 05:32 |
| Breath Volume      | 1.164 | 05:32 |
| Air Blank          | 0.000 | 05:32 |
| Control Test       | 0.081 | 05:33 |
| Air Blank          | 0.000 | 05:33 |
| Diagnostics Check  | OK    | 05:33 |

---

Data Download Date/Time: 06/21/2023 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: WHOBREY,SPENCER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 05/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 12:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:14

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/20/2023 12:46:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:40 |
| Air Blank          | 0.000 | 12:41 |
| Control Test       | 0.083 | 12:41 |
| Air Blank          | 0.000 | 12:41 |
| Subject Sample # 1 | 0.000 | 12:42 |
| Breath Volume      | 1.972 | 12:42 |
| Air Blank          | 0.000 | 12:43 |
| Air Blank          | 0.000 | 12:45 |
| Subject Sample # 2 | 0.000 | 12:45 |
| Breath Volume      | 1.757 | 12:45 |
| Air Blank          | 0.000 | 12:45 |
| Control Test       | 0.083 | 12:46 |
| Air Blank          | 0.000 | 12:46 |
| Diagnostics Check  | OK    | 12:46 |

---

Data Download Date/Time: 06/21/2023 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRODTMAN,ERIC,J  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: GARCIA,CORINA  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 05/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: B  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: VILORIA-DELMAR

DATE/TIME OF TEST: 05/24/2023 00:27:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:21 |
| Air Blank          | 0.000 | 00:22 |
| Control Test       | 0.083 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Subject Sample # 1 | 0.155 | 00:23 |
| Breath Volume      | 1.855 | 00:23 |
| Air Blank          | 0.000 | 00:24 |
| Air Blank          | 0.000 | 00:25 |
| Subject Sample # 2 | 0.149 | 00:26 |
| Breath Volume      | 1.589 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Control Test       | 0.082 | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Diagnostics Check  | OK    | 00:27 |

---

Data Download Date/Time: 06/21/2023 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HANSELL,ROBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 05/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: URBINA

DATE/TIME OF TEST: 05/25/2023 02:46:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:40 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.083 | 02:41 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 1 | 0.000 | 02:42 |
| Breath Volume      | 3.582 | 02:42 |
| Air Blank          | 0.000 | 02:42 |
| Air Blank          | 0.000 | 02:44 |
| Subject Sample # 2 | 0.000 | 02:45 |
| Breath Volume      | 3.582 | 02:45 |
| Air Blank          | 0.000 | 02:45 |
| Control Test       | 0.081 | 02:46 |
| Air Blank          | 0.000 | 02:46 |
| Diagnostics Check  | OK    | 02:46 |

---

Data Download Date/Time: 06/21/2023 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERRON,BENJAMIN,R  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: DEOLEO,GEOVANNI  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 05/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402330433  
DRY GAS EXPIRY DATE: 01/24/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: MARTELLA

DATE/TIME OF TEST: 05/26/2023 23:08:32

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:56 |                |
| Air Blank         | 0.000     | 22:57 |                |
| Control Test      | 0.083     | 22:57 |                |
| Air Blank         | 0.000     | 22:57 |                |
| Subject Sample    | # 1 0.154 | 23:01 | Volume Not Met |
| Breath Volume     | 0.117     | 23:01 |                |
| Air Blank         | 0.000     | 23:01 |                |
| Air Blank         | 0.000     | 23:03 |                |
| Subject Sample    | # 2 0.150 | 23:06 | Volume Not Met |
| Breath Volume     | 0.140     | 23:06 |                |
| Air Blank         | 0.000     | 23:07 |                |
| Control Test      | 0.079     | 23:07 |                |
| Air Blank         | 0.000     | 23:08 |                |
| Diagnostics Check | OK        | 23:08 |                |

---

Data Download Date/Time: 06/21/2023 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: NUSCIS,NICHOLAS  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 05/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: SANCHEZ-MELENDZ

DATE/TIME OF TEST: 05/28/2023 02:33:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Control Test       | 0.083 | 02:27 |
| Air Blank          | 0.000 | 02:27 |
| Subject Sample # 1 | 0.164 | 02:28 |
| Breath Volume      | 2.460 | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Air Blank          | 0.000 | 02:31 |
| Subject Sample # 2 | 0.161 | 02:31 |
| Breath Volume      | 1.964 | 02:31 |
| Air Blank          | 0.000 | 02:32 |
| Control Test       | 0.083 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Diagnostics Check  | OK    | 02:33 |

---

Data Download Date/Time: 06/21/2023 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: MELENDEZ-MUNIZ,WILLIE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 05/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 05/28/2023 04:06:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Control Test       | 0.083 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Subject Sample # 1 | 0.181 | 04:01 |
| Breath Volume      | 2.062 | 04:01 |
| Air Blank          | 0.000 | 04:02 |
| Air Blank          | 0.000 | 04:04 |
| Subject Sample # 2 | 0.171 | 04:05 |
| Breath Volume      | 1.769 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Control Test       | 0.082 | 04:06 |
| Air Blank          | 0.000 | 04:06 |
| Diagnostics Check  | OK    | 04:06 |

---

Data Download Date/Time: 06/21/2023 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,CHRISTIAN,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: NAVARINO,DENNIS  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 05/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 13:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:24

SUBJECT'S LAST NAME: MONROUZEAU

DATE/TIME OF TEST: 05/29/2023 16:00:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:51 |
| Air Blank          | 0.000 | 15:51 |
| Control Test       | 0.084 | 15:51 |
| Air Blank          | 0.000 | 15:52 |
| Subject Sample # 1 | 0.000 | 15:55 |
| Breath Volume      | 3.636 | 15:55 |
| Air Blank          | 0.000 | 15:56 |
| Air Blank          | 0.000 | 15:58 |
| Subject Sample # 2 | 0.000 | 15:58 |
| Breath Volume      | 2.382 | 15:58 |
| Air Blank          | 0.000 | 15:59 |
| Control Test       | 0.083 | 15:59 |
| Air Blank          | 0.000 | 16:00 |
| Diagnostics Check  | OK    | 16:00 |

---

Data Download Date/Time: 06/21/2023 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: GUILBE,JOSE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 05/19/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 05/30/2023 21:02:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:55 |
| Air Blank          | 0.000 | 20:56 |
| Control Test       | 0.084 | 20:56 |
| Air Blank          | 0.000 | 20:57 |
| Subject Sample # 1 | 0.309 | 20:57 |
| Breath Volume      | 2.328 | 20:57 |
| Air Blank          | 0.000 | 20:58 |
| Air Blank          | 0.000 | 21:00 |
| Subject Sample # 2 | 0.304 | 21:00 |
| Breath Volume      | 2.277 | 21:00 |
| Air Blank          | 0.000 | 21:01 |
| Control Test       | 0.081 | 21:01 |
| Air Blank          | 0.000 | 21:02 |
| Diagnostics Check  | OK    | 21:02 |

---

Data Download Date/Time: 06/21/2023 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZIMMERMAN,J,  
BREATH TEST OPERATOR AGENCY: Port Richey PD  
ARRESTING OFFICER: ZIMMERMAN,J  
ARRESTING OFFICER AGENCY: Port Richey PD  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402370890  
DRY GAS EXPIRY DATE: 05/09/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: CATURIA

DATE/TIME OF TEST: 05/06/2023 00:10:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:04 |
| Air Blank          | 0.000 | 00:05 |
| Control Test       | 0.077 | 00:05 |
| Air Blank          | 0.000 | 00:06 |
| Subject Sample # 1 | 0.162 | 00:06 |
| Breath Volume      | 1.800 | 00:06 |
| Air Blank          | 0.000 | 00:07 |
| Air Blank          | 0.000 | 00:08 |
| Subject Sample # 2 | 0.156 | 00:09 |
| Breath Volume      | 1.281 | 00:09 |
| Air Blank          | 0.000 | 00:09 |
| Control Test       | 0.075 | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Diagnostics Check  | OK    | 00:10 |

---

Data Download Date/Time: 06/13/2023 20:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W  
BREATH TEST OPERATOR AGENCY: Port Richey PD  
ARRESTING OFFICER: JASPER,A  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402370890  
DRY GAS EXPIRY DATE: 05/09/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 05/07/2023 00:44:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:37 |
| Air Blank          | 0.000 | 00:38 |
| Control Test       | 0.077 | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Subject Sample # 1 | 0.210 | 00:39 |
| Breath Volume      | 3.019 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Air Blank          | 0.000 | 00:42 |
| Subject Sample # 2 | 0.203 | 00:42 |
| Breath Volume      | 2.957 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Control Test       | 0.078 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Diagnostics Check  | OK    | 00:44 |

---

Data Download Date/Time: 06/13/2023 20:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARAMIELLO,C,A  
BREATH TEST OPERATOR AGENCY: Port Richey PD  
ARRESTING OFFICER: CARAMIELLO,C  
ARRESTING OFFICER AGENCY: Port Richey PD  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402370890  
DRY GAS EXPIRY DATE: 05/09/2025  
VIOLATION CODE: DUI  
ARREST TIME: 12:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/09/2023 00:54:29

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 00:42 |                |
| Air Blank         | 0.000     | 00:43 |                |
| Control Test      | 0.078     | 00:43 |                |
| Air Blank         | 0.000     | 00:44 |                |
| Subject Sample    | # 1 0.050 | 00:47 | Volume Not Met |
| Breath Volume     | 0.074     | 00:47 |                |
| Air Blank         | 0.000     | 00:47 |                |
| Air Blank         | 0.000     | 00:49 |                |
| Subject Sample    | # 2 0.052 | 00:52 | Volume Not Met |
| Breath Volume     | 1.031     | 00:52 |                |
| Air Blank         | 0.000     | 00:53 |                |
| Control Test      | 0.080     | 00:53 |                |
| Air Blank         | 0.000     | 00:54 |                |
| Diagnostics Check | OK        | 00:54 |                |

---

Data Download Date/Time: 06/13/2023 20:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMELMAN, TYLER,  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: JASPER, ANDREW  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402370890  
DRY GAS EXPIRY DATE: 05/09/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GOHEL

DATE/TIME OF TEST: 05/10/2023 02:59:58

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 02:45 |                |
| Air Blank         | 0.000     | 02:46 |                |
| Control Test      | 0.077     | 02:46 |                |
| Air Blank         | 0.000     | 02:46 |                |
| Subject Sample    | # 1 0.142 | 02:50 | Volume Not Met |
| Breath Volume     | 0.808     | 02:50 |                |
| Air Blank         | 0.000     | 02:51 |                |
| Air Blank         | 0.000     | 02:52 |                |
| Subject Sample    | # 2 0.136 | 02:55 |                |
| Breath Volume     | 2.160     | 02:55 |                |
| Air Blank         | 0.000     | 02:55 |                |
| Air Blank         | 0.000     | 02:57 |                |
| Subject Sample    | # 3 0.146 | 02:58 |                |
| Breath Volume     | 1.214     | 02:58 |                |
| Air Blank         | 0.000     | 02:59 |                |
| Control Test      | 0.076     | 02:59 |                |
| Air Blank         | 0.000     | 02:59 |                |
| Diagnostics Check | OK        | 02:59 |                |

Data Download Date/Time: 06/13/2023 20:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMELMAN, TYLER,  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: HAMMELMAN, TYLER  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402370890  
DRY GAS EXPIRY DATE: 05/09/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ROSENCRANS

DATE/TIME OF TEST: 05/14/2023 03:23:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:17 |
| Air Blank          | 0.000 | 03:17 |
| Control Test       | 0.078 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Subject Sample # 1 | 0.116 | 03:18 |
| Breath Volume      | 2.347 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 2 | 0.113 | 03:22 |
| Breath Volume      | 1.199 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Control Test       | 0.078 | 03:23 |
| Air Blank          | 0.000 | 03:23 |
| Diagnostics Check  | OK    | 03:23 |

---

Data Download Date/Time: 06/13/2023 20:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W  
BREATH TEST OPERATOR AGENCY: Port Richey PD  
ARRESTING OFFICER: CURBELO,A  
ARRESTING OFFICER AGENCY: New Port Richey PD  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402370890  
DRY GAS EXPIRY DATE: 05/09/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RUBY

DATE/TIME OF TEST: 05/16/2023 04:39:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:33 |
| Air Blank          | 0.000 | 04:33 |
| Control Test       | 0.078 | 04:33 |
| Air Blank          | 0.000 | 04:34 |
| Subject Sample # 1 | 0.000 | 04:34 |
| Breath Volume      | 1.363 | 04:34 |
| Air Blank          | 0.000 | 04:35 |
| Air Blank          | 0.000 | 04:37 |
| Subject Sample # 2 | 0.000 | 04:38 |
| Breath Volume      | 1.546 | 04:38 |
| Air Blank          | 0.000 | 04:38 |
| Control Test       | 0.078 | 04:38 |
| Air Blank          | 0.000 | 04:39 |
| Diagnostics Check  | OK    | 04:39 |

---

Data Download Date/Time: 06/13/2023 20:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLLINS,THOMAS,J  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: MORRISON III,ARTHUR  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402370890  
DRY GAS EXPIRY DATE: 05/09/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: MARINI

DATE/TIME OF TEST: 05/19/2023 20:46:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:39 |
| Air Blank          | 0.000 | 20:40 |
| Control Test       | 0.078 | 20:40 |
| Air Blank          | 0.000 | 20:40 |
| Subject Sample # 1 | 0.145 | 20:42 |
| Breath Volume      | 1.761 | 20:42 |
| Air Blank          | 0.000 | 20:42 |
| Air Blank          | 0.000 | 20:44 |
| Subject Sample # 2 | 0.150 | 20:45 |
| Breath Volume      | 1.441 | 20:45 |
| Air Blank          | 0.000 | 20:45 |
| Control Test       | 0.077 | 20:46 |
| Air Blank          | 0.000 | 20:46 |
| Diagnostics Check  | OK    | 20:46 |

---

Data Download Date/Time: 06/13/2023 20:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCONNELL,MCCONNELL,W  
BREATH TEST OPERATOR AGENCY: Port Richey PD  
ARRESTING OFFICER: BERLYNE,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402370890  
DRY GAS EXPIRY DATE: 05/09/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: AREU

DATE/TIME OF TEST: 05/25/2023 19:52:11

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 19:39 |                |
| Air Blank         | 0.000     | 19:39 |                |
| Control Test      | 0.078     | 19:39 |                |
| Air Blank         | 0.000     | 19:40 |                |
| Subject Sample    | # 1 0.174 | 19:43 | Volume Not Met |
| Breath Volume     | 0.265     | 19:43 |                |
| Air Blank         | 0.000     | 19:44 |                |
| Air Blank         | 0.000     | 19:46 |                |
| Subject Sample    | # 2 0.183 | 19:47 |                |
| Breath Volume     | 2.468     | 19:47 |                |
| Air Blank         | 0.000     | 19:48 |                |
| Air Blank         | 0.000     | 19:49 |                |
| Subject Sample    | # 3 0.173 | 19:50 |                |
| Breath Volume     | 1.292     | 19:50 |                |
| Air Blank         | 0.000     | 19:51 |                |
| Control Test      | 0.076     | 19:51 |                |
| Air Blank         | 0.000     | 19:52 |                |
| Diagnostics Check | OK        | 19:52 |                |

---

Data Download Date/Time: 06/13/2023 20:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLLINS,THOMAS,J  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: COLLINS,THOMAS  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402370890  
DRY GAS EXPIRY DATE: 05/09/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 05/30/2023 00:24:23

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:20 |                      |
| Air Blank         | 0.000 | 00:20 |                      |
| Control Test      | 0.078 | 00:21 |                      |
| Air Blank         | 0.000 | 00:21 |                      |
| Subject Sample    | # 1   | 00:23 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:23 |                      |
| Air Blank         | 0.000 | 00:23 |                      |
| Control Test      | 0.078 | 00:23 |                      |
| Air Blank         | 0.000 | 00:24 |                      |
| Diagnostics Check | OK    | 00:24 |                      |

Data Download Date/Time: 06/13/2023 20:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAFLEUR,MATTHEW,W  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: LAFLEUR,MATTHEW  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 05/04/2023 02:23:42

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 02:12 |                |
| Air Blank         | 0.000     | 02:12 |                |
| Control Test      | 0.077     | 02:13 |                |
| Air Blank         | 0.000     | 02:13 |                |
| Subject Sample    | # 1 0.125 | 02:16 | Volume Not Met |
| Breath Volume     | 0.058     | 02:16 |                |
| Air Blank         | 0.000     | 02:17 |                |
| Air Blank         | 0.000     | 02:19 |                |
| Subject Sample    | # 2 0.147 | 02:22 | Volume Not Met |
| Breath Volume     | 0.015     | 02:22 |                |
| Air Blank         | 0.000     | 02:22 |                |
| Control Test      | 0.077     | 02:23 |                |
| Air Blank         | 0.000     | 02:23 |                |
| Diagnostics Check | OK        | 02:23 |                |

---

Data Download Date/Time: 05/29/2023 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BULLARD,SEAN,M  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: BULLARD,SEAN  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2982108A3  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WEBB IV

DATE/TIME OF TEST: 05/05/2023 01:54:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:48 |
| Air Blank          | 0.000 | 01:48 |
| Control Test       | 0.077 | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Subject Sample # 1 | 0.165 | 01:49 |
| Breath Volume      | 3.046 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Air Blank          | 0.000 | 01:52 |
| Subject Sample # 2 | 0.162 | 01:52 |
| Breath Volume      | 3.324 | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Control Test       | 0.076 | 01:53 |
| Air Blank          | 0.000 | 01:54 |
| Diagnostics Check  | OK    | 01:54 |

---

Data Download Date/Time: 05/29/2023 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAFLEUR, MATTHEW, W  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: LAFLEUR, MATTHEW  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CORRALES

DATE/TIME OF TEST: 05/05/2023 02:28:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Control Test       | 0.076 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Subject Sample # 1 | 0.138 | 02:24 |
| Breath Volume      | 3.773 | 02:24 |
| Air Blank          | 0.000 | 02:24 |
| Air Blank          | 0.000 | 02:26 |
| Subject Sample # 2 | 0.136 | 02:27 |
| Breath Volume      | 3.488 | 02:27 |
| Air Blank          | 0.000 | 02:27 |
| Control Test       | 0.076 | 02:28 |
| Air Blank          | 0.000 | 02:28 |
| Diagnostics Check  | OK    | 02:28 |

---

Data Download Date/Time: 05/29/2023 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEAT,MATTHEW,J  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: TEAT,MATTHEW  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: BLASKI

DATE/TIME OF TEST: 05/06/2023 03:03:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:57 |
| Air Blank          | 0.000 | 02:57 |
| Control Test       | 0.077 | 02:58 |
| Air Blank          | 0.000 | 02:58 |
| Subject Sample # 1 | 0.000 | 02:59 |
| Breath Volume      | 3.507 | 02:59 |
| Air Blank          | 0.000 | 02:59 |
| Air Blank          | 0.000 | 03:01 |
| Subject Sample # 2 | 0.000 | 03:02 |
| Breath Volume      | 3.437 | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Control Test       | 0.078 | 03:03 |
| Air Blank          | 0.000 | 03:03 |
| Diagnostics Check  | OK    | 03:03 |

---

Data Download Date/Time: 05/29/2023 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A  
BREATH TEST OPERATOR AGENCY: Cocoa PD  
ARRESTING OFFICER: ROSA #42,E  
ARRESTING OFFICER AGENCY: Cocoa PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 04:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SANCHEZ-ROSARIO

DATE/TIME OF TEST: 05/07/2023 04:53:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Control Test       | 0.077 | 04:47 |
| Air Blank          | 0.000 | 04:47 |
| Subject Sample # 1 | 0.094 | 04:49 |
| Breath Volume      | 3.417 | 04:49 |
| Air Blank          | 0.000 | 04:49 |
| Air Blank          | 0.000 | 04:51 |
| Subject Sample # 2 | 0.091 | 04:52 |
| Breath Volume      | 3.359 | 04:52 |
| Air Blank          | 0.000 | 04:52 |
| Control Test       | 0.075 | 04:53 |
| Air Blank          | 0.000 | 04:53 |
| Diagnostics Check  | OK    | 04:53 |

---

Data Download Date/Time: 05/29/2023 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: HILL,ROBERT  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: HOEPPNER

DATE/TIME OF TEST: 05/10/2023 18:59:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:53 |
| Air Blank          | 0.000 | 18:53 |
| Control Test       | 0.077 | 18:54 |
| Air Blank          | 0.000 | 18:54 |
| Subject Sample # 1 | 0.112 | 18:54 |
| Breath Volume      | 1.628 | 18:54 |
| Air Blank          | 0.000 | 18:55 |
| Air Blank          | 0.000 | 18:57 |
| Subject Sample # 2 | 0.118 | 18:58 |
| Breath Volume      | 2.289 | 18:58 |
| Air Blank          | 0.000 | 18:58 |
| Control Test       | 0.076 | 18:58 |
| Air Blank          | 0.000 | 18:59 |
| Diagnostics Check  | OK    | 18:59 |

---

Data Download Date/Time: 05/29/2023 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLSEN, IAN, N  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: ROSA, EMILIO  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: YOROR

DATE/TIME OF TEST: 05/11/2023 02:35:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Control Test       | 0.076 | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Subject Sample # 1 | 0.184 | 02:31 |
| Breath Volume      | 2.218 | 02:31 |
| Air Blank          | 0.000 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Subject Sample # 2 | 0.183 | 02:34 |
| Breath Volume      | 1.468 | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.075 | 02:35 |
| Air Blank          | 0.000 | 02:35 |
| Diagnostics Check  | OK    | 02:35 |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: HILL,ROBERT  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/11/2023 04:00:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Control Test       | 0.076 | 03:54 |
| Air Blank          | 0.000 | 03:55 |
| Subject Sample # 1 | 0.109 | 03:55 |
| Breath Volume      | 2.953 | 03:55 |
| Air Blank          | 0.000 | 03:56 |
| Air Blank          | 0.000 | 03:58 |
| Subject Sample # 2 | 0.106 | 03:58 |
| Breath Volume      | 3.574 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Control Test       | 0.076 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Diagnostics Check  | OK    | 04:00 |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAFLEUR, MATTHEW, W  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: LAFLEUR, MATTHEW  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: JEVITT

DATE/TIME OF TEST: 05/13/2023 04:12:46

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 04:09 |            |
| Air Blank         | 0.000 | 04:10 |            |
| Control Test      | 0.077 | 04:10 |            |
| Air Blank         | 0.000 | 04:11 |            |
| Subject Sample    | # 1   | 04:11 | RFI Detect |
| Breath Volume     | 0.000 | 04:11 |            |
| Air Blank         | 0.000 | 04:11 |            |
| Control Test      | 0.077 | 04:12 |            |
| Air Blank         | 0.000 | 04:12 |            |
| Diagnostics Check | OK    | 04:12 |            |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAFLEUR, MATTHEW, W  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: LAFLEUR, MATTHEW  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: JEVITT

DATE/TIME OF TEST: 05/13/2023 04:23:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:15 |
| Air Blank          | 0.000 | 04:15 |
| Control Test       | 0.077 | 04:16 |
| Air Blank          | 0.000 | 04:16 |
| Subject Sample # 1 | 0.154 | 04:18 |
| Breath Volume      | 1.597 | 04:18 |
| Air Blank          | 0.000 | 04:18 |
| Air Blank          | 0.000 | 04:20 |
| Subject Sample # 2 | 0.170 | 04:21 |
| Breath Volume      | 3.019 | 04:21 |
| Air Blank          | 0.000 | 04:22 |
| Control Test       | 0.076 | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Diagnostics Check  | OK    | 04:23 |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: HILL,ROBERT  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CLOUS

DATE/TIME OF TEST: 05/14/2023 04:42:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:36 |
| Air Blank          | 0.000 | 04:36 |
| Control Test       | 0.077 | 04:36 |
| Air Blank          | 0.000 | 04:37 |
| Subject Sample # 1 | 0.000 | 04:37 |
| Breath Volume      | 3.789 | 04:37 |
| Air Blank          | 0.000 | 04:38 |
| Air Blank          | 0.000 | 04:40 |
| Subject Sample # 2 | 0.000 | 04:41 |
| Breath Volume      | 4.566 | 04:41 |
| Air Blank          | 0.000 | 04:41 |
| Control Test       | 0.077 | 04:42 |
| Air Blank          | 0.000 | 04:42 |
| Diagnostics Check  | OK    | 04:42 |

---

Data Download Date/Time: 05/29/2023 10:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLSEN, IAN, N  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: ROSA, EMILIO  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: COLANGELO

DATE/TIME OF TEST: 05/20/2023 02:34:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:28 |
| Air Blank          | 0.000 | 02:28 |
| Control Test       | 0.076 | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 1 | 0.176 | 02:29 |
| Breath Volume      | 3.320 | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Air Blank          | 0.000 | 02:32 |
| Subject Sample # 2 | 0.171 | 02:32 |
| Breath Volume      | 2.851 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Control Test       | 0.075 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Diagnostics Check  | OK    | 02:34 |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLSEN, IAN, N  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: ROSA, EMILIO  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 04:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 05/20/2023 04:33:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:26 |
| Air Blank          | 0.000 | 04:27 |
| Control Test       | 0.077 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Subject Sample # 1 | 0.243 | 04:28 |
| Breath Volume      | 3.714 | 04:28 |
| Air Blank          | 0.000 | 04:29 |
| Air Blank          | 0.000 | 04:31 |
| Subject Sample # 2 | 0.240 | 04:31 |
| Breath Volume      | 2.527 | 04:31 |
| Air Blank          | 0.000 | 04:32 |
| Control Test       | 0.075 | 04:32 |
| Air Blank          | 0.000 | 04:33 |
| Diagnostics Check  | OK    | 04:33 |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMES,R,  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: ROSA,E  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 05/21/2023 02:49:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:42 |
| Air Blank          | 0.000 | 02:42 |
| Control Test       | 0.077 | 02:43 |
| Air Blank          | 0.000 | 02:43 |
| Subject Sample # 1 | 0.126 | 02:45 |
| Breath Volume      | 2.398 | 02:45 |
| Air Blank          | 0.000 | 02:45 |
| Air Blank          | 0.000 | 02:47 |
| Subject Sample # 2 | 0.128 | 02:48 |
| Breath Volume      | 1.464 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.075 | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Diagnostics Check  | OK    | 02:49 |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMES,R,  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: MIRANDA,J  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/21/2023 03:19:47

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 03:05 |                           |
| Air Blank         | 0.000     | 03:06 |                           |
| Control Test      | 0.075     | 03:06 |                           |
| Air Blank         | 0.000     | 03:06 |                           |
| Subject Sample    | # 1 0.248 | 03:10 | Volume Not Met            |
| Breath Volume     | 0.769     | 03:10 |                           |
| Air Blank         | 0.000     | 03:10 |                           |
| Air Blank         | 0.000     | 03:12 |                           |
| Subject Sample    | # 2 0.251 | 03:15 |                           |
| Breath Volume     | 4.105     | 03:15 |                           |
| Air Blank         | 0.000     | 03:15 |                           |
| Air Blank         | 0.000     | 03:17 |                           |
| Subject Sample    | # 3 0.245 | 03:18 |                           |
| Breath Volume     | 1.523     | 03:18 |                           |
| Air Blank         | 0.000     | 03:18 |                           |
| Control Test      | 0.074     | 03:19 | Control Outside Tolerance |
| Air Blank         |           | 03:19 | RFI Detect                |

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEACH,BRANDON,J  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: LEACH,BRANDON  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 08:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:32

SUBJECT'S LAST NAME: TROILA

DATE/TIME OF TEST: 05/21/2023 09:08:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:02 |
| Air Blank          | 0.000 | 09:02 |
| Control Test       | 0.077 | 09:03 |
| Air Blank          | 0.000 | 09:03 |
| Subject Sample # 1 | 0.000 | 09:04 |
| Breath Volume      | 3.242 | 09:04 |
| Air Blank          | 0.000 | 09:04 |
| Air Blank          | 0.000 | 09:06 |
| Subject Sample # 2 | 0.000 | 09:07 |
| Breath Volume      | 3.386 | 09:07 |
| Air Blank          | 0.000 | 09:07 |
| Control Test       | 0.077 | 09:07 |
| Air Blank          | 0.000 | 09:08 |
| Diagnostics Check  | OK    | 09:08 |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: HILL,ROBERT  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: WALL

DATE/TIME OF TEST: 05/22/2023 23:31:48

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 23:20 |                    |
| Air Blank         | 0.000 | 23:20 |                    |
| Control Test      | 0.076 | 23:21 |                    |
| Air Blank         | 0.000 | 23:21 |                    |
| Subject Sample    | # 1   | 23:24 | No Sample Provided |
| Breath Volume     | 0.000 | 23:24 |                    |
| Air Blank         | 0.000 | 23:25 |                    |
| Air Blank         | 0.000 | 23:27 |                    |
| Subject Sample    | # 2   | 23:30 | No Sample Provided |
| Breath Volume     | 0.000 | 23:30 |                    |
| Air Blank         | 0.000 | 23:30 |                    |
| Control Test      | 0.077 | 23:31 |                    |
| Air Blank         | 0.000 | 23:31 |                    |
| Diagnostics Check | OK    | 23:31 |                    |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLSEN, IAN, N  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: ROSA, EMILIO  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 05/26/2023 00:10:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:04 |
| Air Blank          | 0.000 | 00:04 |
| Control Test       | 0.077 | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Subject Sample # 1 | 0.145 | 00:05 |
| Breath Volume      | 2.421 | 00:05 |
| Air Blank          | 0.000 | 00:06 |
| Air Blank          | 0.000 | 00:08 |
| Subject Sample # 2 | 0.143 | 00:08 |
| Breath Volume      | 2.117 | 00:08 |
| Air Blank          | 0.000 | 00:09 |
| Control Test       | 0.077 | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Diagnostics Check  | OK    | 00:10 |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLSEN, IAN, N  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: ROSA, EMILIO  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/12/2033  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: FARLEY

DATE/TIME OF TEST: 05/26/2023 02:43:13

---

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 02:27 |                    |
| Air Blank         | 0.000     | 02:27 |                    |
| Control Test      | 0.076     | 02:28 |                    |
| Air Blank         | 0.000     | 02:28 |                    |
| Subject Sample    | # 1 0.185 | 02:30 |                    |
| Breath Volume     | 1.320     | 02:30 |                    |
| Air Blank         | 0.000     | 02:31 |                    |
| Air Blank         | 0.000     | 02:33 |                    |
| Subject Sample    | # 2       | 02:36 | No Sample Provided |
| Breath Volume     | 0.000     | 02:36 |                    |
| Air Blank         | 0.000     | 02:36 |                    |
| Air Blank         | 0.000     | 02:38 |                    |
| Subject Sample    | # 3       | 02:41 | No Sample Provided |
| Breath Volume     | 0.000     | 02:41 |                    |
| Air Blank         | 0.000     | 02:42 |                    |
| Control Test      | 0.076     | 02:42 |                    |
| Air Blank         | 0.000     | 02:43 |                    |
| Diagnostics Check | OK        | 02:43 |                    |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: HILL,ROBERT  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 05/27/2023 01:24:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Control Test       | 0.077 | 01:19 |
| Air Blank          | 0.000 | 01:19 |
| Subject Sample # 1 | 0.155 | 01:20 |
| Breath Volume      | 4.492 | 01:20 |
| Air Blank          | 0.000 | 01:20 |
| Air Blank          | 0.000 | 01:22 |
| Subject Sample # 2 | 0.146 | 01:23 |
| Breath Volume      | 2.949 | 01:23 |
| Air Blank          | 0.000 | 01:24 |
| Control Test       | 0.076 | 01:24 |
| Air Blank          | 0.000 | 01:24 |
| Diagnostics Check  | OK    | 01:24 |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: HILL,ROBERT  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: HANFI

DATE/TIME OF TEST: 05/27/2023 05:17:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:11 |
| Air Blank          | 0.000 | 05:11 |
| Control Test       | 0.077 | 05:12 |
| Air Blank          | 0.000 | 05:12 |
| Subject Sample # 1 | 0.103 | 05:13 |
| Breath Volume      | 4.550 | 05:13 |
| Air Blank          | 0.000 | 05:13 |
| Air Blank          | 0.000 | 05:15 |
| Subject Sample # 2 | 0.098 | 05:16 |
| Breath Volume      | 4.191 | 05:16 |
| Air Blank          | 0.000 | 05:16 |
| Control Test       | 0.076 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Diagnostics Check  | OK    | 05:17 |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAFLEUR, MATTHEW, W  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: LAFLEUR, MATTHEW  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/28/2023 02:47:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:41 |
| Air Blank          | 0.000 | 02:41 |
| Control Test       | 0.077 | 02:42 |
| Air Blank          | 0.000 | 02:42 |
| Subject Sample # 1 | 0.203 | 02:43 |
| Breath Volume      | 2.703 | 02:43 |
| Air Blank          | 0.000 | 02:43 |
| Air Blank          | 0.000 | 02:45 |
| Subject Sample # 2 | 0.198 | 02:46 |
| Breath Volume      | 1.457 | 02:46 |
| Air Blank          | 0.000 | 02:46 |
| Control Test       | 0.076 | 02:47 |
| Air Blank          | 0.000 | 02:47 |
| Diagnostics Check  | OK    | 02:47 |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: HILL,ROBERT  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1  
DRY GAS EXPIRY DATE: 12/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: ARMOUR

DATE/TIME OF TEST: 05/28/2023 21:12:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:06 |
| Air Blank          | 0.000 | 21:06 |
| Control Test       | 0.076 | 21:07 |
| Air Blank          | 0.000 | 21:07 |
| Subject Sample # 1 | 0.142 | 21:07 |
| Breath Volume      | 2.734 | 21:07 |
| Air Blank          | 0.000 | 21:08 |
| Air Blank          | 0.000 | 21:10 |
| Subject Sample # 2 | 0.130 | 21:10 |
| Breath Volume      | 1.996 | 21:10 |
| Air Blank          | 0.000 | 21:11 |
| Control Test       | 0.076 | 21:11 |
| Air Blank          | 0.000 | 21:12 |
| Diagnostics Check  | OK    | 21:12 |

---

Data Download Date/Time: 05/29/2023 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WIGGINS, TYLER, D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: PROVENCHER, JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: HARDMAN

DATE/TIME OF TEST: 05/06/2023 23:38:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:32 |
| Air Blank          | 0.000 | 23:32 |
| Control Test       | 0.079 | 23:33 |
| Air Blank          | 0.000 | 23:33 |
| Subject Sample # 1 | 0.121 | 23:34 |
| Breath Volume      | 1.960 | 23:34 |
| Air Blank          | 0.000 | 23:34 |
| Air Blank          | 0.000 | 23:36 |
| Subject Sample # 2 | 0.119 | 23:37 |
| Breath Volume      | 1.886 | 23:37 |
| Air Blank          | 0.000 | 23:37 |
| Control Test       | 0.079 | 23:38 |
| Air Blank          | 0.000 | 23:38 |
| Diagnostics Check  | OK    | 23:38 |

---

Data Download Date/Time: 05/25/2023 11:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEDERMAN,CASEY,L  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: VAN ZILE,JASON  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33321080A2  
DRY GAS EXPIRY DATE: 01/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: GORE

DATE/TIME OF TEST: 05/16/2023 22:53:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:47 |
| Air Blank          | 0.000 | 22:47 |
| Control Test       | 0.080 | 22:47 |
| Air Blank          | 0.000 | 22:48 |
| Subject Sample # 1 | 0.139 | 22:48 |
| Breath Volume      | 2.539 | 22:48 |
| Air Blank          | 0.000 | 22:49 |
| Air Blank          | 0.000 | 22:51 |
| Subject Sample # 2 | 0.143 | 22:52 |
| Breath Volume      | 3.281 | 22:52 |
| Air Blank          | 0.000 | 22:52 |
| Control Test       | 0.079 | 22:53 |
| Air Blank          | 0.000 | 22:53 |
| Diagnostics Check  | OK    | 22:53 |

---

Data Download Date/Time: 05/25/2023 11:41