

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULLES,ALBERTO,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: PULLES,ALBERTO
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: CADENAS

DATE/TIME OF TEST: 05/06/2023 04:36:42

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.079	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1 0.074	04:29	Volume Not Met
Breath Volume	0.074	04:29	
Air Blank	0.000	04:30	
Air Blank	0.000	04:31	
Subject Sample	# 2 0.085	04:35	Volume Not Met
Breath Volume	0.347	04:35	
Air Blank	0.000	04:35	
Control Test	0.079	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 05/22/2023 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULLES,ALBERTO,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: PULLES,ALBERTO
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: CADENAS

DATE/TIME OF TEST: 05/06/2023 04:46:47

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.078	04:40	
Air Blank	0.000	04:40	
Subject Sample # 1	0.085	04:42	
Breath Volume	1.683	04:42	
Air Blank	0.000	04:42	
Air Blank	0.000	04:44	
Subject Sample # 2		04:45	Subject Test Refused
Breath Volume	0.132	04:45	
Air Blank	0.000	04:45	
Control Test	0.076	04:46	RFI Detect
Air Blank	0.000	04:46	

Data Download Date/Time: 05/22/2023 20:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEANDRE,GERALD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LEANDRE,GERALD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: CHALES AGUILAR

DATE/TIME OF TEST: 05/14/2023 01:36:35

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:34	
Control Test	0.078	01:34	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:35	Subject Test Refused
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Control Test	0.078	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 06/09/2023 07:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: RIVERA-MORALES

DATE/TIME OF TEST: 05/14/2023 03:41:10

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.078	03:38	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Control Test	0.077	03:40	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 06/09/2023 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: WAITE,JASON
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: ELVIR-PEREZ

DATE/TIME OF TEST: 05/14/2023 05:29:27

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.078	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.165	05:24
Breath Volume	2.179	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.154	05:27
Breath Volume	1.234	05:27
Air Blank	0.000	05:28
Control Test	0.077	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 06/09/2023 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEANDRE,GERALD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LEANDRE,GERALD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BLAISE

DATE/TIME OF TEST: 05/21/2023 00:30:21

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.077	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.243	00:25
Breath Volume	1.917	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.247	00:28
Breath Volume	2.054	00:28
Air Blank	0.000	00:29
Control Test	0.076	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 06/09/2023 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: ORTEGA BACILA

DATE/TIME OF TEST: 05/21/2023 01:59:57

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.077	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.077	01:59	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 06/09/2023 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURBELO,CLAUDIA,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: CURBELO,CLAUDIA
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: FARINAS RAMIREZ

DATE/TIME OF TEST: 05/21/2023 03:40:03

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.076	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.146	03:35
Breath Volume	1.914	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.138	03:38
Breath Volume	1.437	03:38
Air Blank	0.000	03:39
Control Test	0.075	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:40

Data Download Date/Time: 06/09/2023 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,ALEXIS,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TANIS,MIKE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: CAMACHO SANCHEZ

DATE/TIME OF TEST: 05/21/2023 20:04:11

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:01	
Control Test	0.078	20:02	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:02	Subject Test Refused
Breath Volume	0.000	20:02	
Air Blank	0.000	20:03	
Control Test	0.077	20:03	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

Data Download Date/Time: 06/09/2023 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,ALEXIS,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GARCIA,ALEXIS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: TUCK

DATE/TIME OF TEST: 05/23/2023 17:35:14

Diagnostics Check	OK	17:32	
Air Blank	0.000	17:32	
Control Test	0.077	17:33	
Air Blank	0.000	17:33	
Subject Sample	# 1	17:33	Subject Test Refused
Breath Volume	0.000	17:33	
Air Blank	0.000	17:34	
Control Test	0.078	17:34	
Air Blank	0.000	17:35	
Diagnostics Check	OK	17:35	

Data Download Date/Time: 06/09/2023 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,WILFREDO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ORTEGA,WILFREDO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: TREJO GARCIA

DATE/TIME OF TEST: 05/28/2023 13:04:32

Diagnostics Check	OK	13:01	
Air Blank	0.000	13:02	
Control Test	0.077	13:02	
Air Blank	0.000	13:02	
Subject Sample	# 1	13:03	Subject Test Refused
Breath Volume	0.000	13:03	
Air Blank	0.000	13:03	
Control Test	0.078	13:03	
Air Blank	0.000	13:04	
Diagnostics Check	OK	13:04	

Data Download Date/Time: 06/09/2023 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,TERRELL,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: JONES,TERRELL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: REYNALDS

DATE/TIME OF TEST: 05/28/2023 22:26:27

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.077	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 1	22:18	No Sample Provided
Breath Volume	0.000	22:18	
Air Blank	0.000	22:19	
Air Blank	0.000	22:21	
Subject Sample	# 2	0.206	22:21
Breath Volume	2.015	22:21	
Air Blank	0.000	22:22	
Air Blank	0.000	22:24	
Subject Sample	# 3	0.200	22:24
Breath Volume	1.675	22:24	
Air Blank	0.000	22:25	
Control Test	0.078	22:25	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:26	

Data Download Date/Time: 06/09/2023 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: OROZCO

DATE/TIME OF TEST: 05/29/2023 17:50:15

Diagnostics Check	OK	17:43
Air Blank	0.000	17:44
Control Test	0.078	17:44
Air Blank	0.000	17:45
Subject Sample # 1	0.205	17:45
Breath Volume	3.531	17:45
Air Blank	0.000	17:46
Air Blank	0.000	17:47
Subject Sample # 2	0.201	17:48
Breath Volume	2.925	17:48
Air Blank	0.000	17:49
Control Test	0.076	17:49
Air Blank	0.000	17:50
Diagnostics Check	OK	17:50

Data Download Date/Time: 06/09/2023 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 05/01/2023 00:56:15

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.137	00:51
Breath Volume	1.484	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.143	00:54
Breath Volume	2.714	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 05/23/2023 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: REINA MOLINA

DATE/TIME OF TEST: 05/05/2023 23:37:12

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.175	23:32
Breath Volume	2.101	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.180	23:35
Breath Volume	1.445	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 05/23/2023 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 05/06/2023 02:45:49

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.171	02:41
Breath Volume	2.398	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.168	02:44
Breath Volume	2.320	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 05/23/2023 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERTD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: GURR

DATE/TIME OF TEST: 05/07/2023 02:13:08

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.080	02:11	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 05/23/2023 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: COMER

DATE/TIME OF TEST: 05/07/2023 03:37:20

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.080	03:35	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:36	
Control Test	0.079	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 05/23/2023 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 05/08/2023 02:57:29

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.117	02:53
Breath Volume	2.261	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.118	02:56
Breath Volume	2.277	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 05/23/2023 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/09/2023 02:07:02

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.137	02:02
Breath Volume	2.015	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.133	02:05
Breath Volume	1.648	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:07

Data Download Date/Time: 05/23/2023 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SWIFT

DATE/TIME OF TEST: 05/10/2023 01:16:17

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.161	01:12
Breath Volume	1.214	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.160	01:14
Breath Volume	1.265	01:14
Air Blank	0.000	01:15
Control Test	0.079	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 05/23/2023 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PONCE DIAZ

DATE/TIME OF TEST: 05/29/2023 00:31:56

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:30	Subject Test Refused
Breath Volume	0.000	00:30	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 06/14/2023 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,B,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MANDRICK,T
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 05/28/2023 22:19:51

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:18	
Control Test	0.078	22:19	Interferent Detect
Air Blank	0.000	22:19	

Data Download Date/Time: 06/27/2023 00:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: SYS CHECK
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/30/2023 13:56:26

Diagnostics Check	OK	13:50
Air Blank	0.000	13:51
Control Test	0.078	13:51
Air Blank	0.000	13:51
Subject Sample # 1	0.000	13:52
Breath Volume	2.164	13:52
Air Blank	0.000	13:52
Air Blank	0.000	13:54
Subject Sample # 2	0.000	13:55
Breath Volume	1.878	13:55
Air Blank	0.000	13:55
Control Test	0.077	13:55
Air Blank	0.000	13:56
Diagnostics Check	OK	13:56

Data Download Date/Time: 06/27/2023 00:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: SYS CHECK
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/30/2023 14:08:15

Diagnostics Check	OK	14:04	
Air Blank	0.000	14:05	
Control Test	0.078	14:05	
Air Blank	0.000	14:06	
Subject Sample	# 1	14:06	Slope Not Met
Breath Volume	1.101	14:06	
Air Blank	0.000	14:07	
Control Test	0.077	14:07	
Air Blank	0.000	14:08	
Diagnostics Check	OK	14:08	

Data Download Date/Time: 06/27/2023 00:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUMANS,S,E
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: DYAL,T
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 05/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0602108082
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: CRAFT

DATE/TIME OF TEST: 05/31/2023 19:38:24

Diagnostics Check	OK	19:31
Air Blank	0.000	19:31
Control Test	0.081	19:32
Air Blank	0.000	19:32
Subject Sample # 1	0.223	19:33
Breath Volume	1.640	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.235	19:36
Breath Volume	2.593	19:36
Air Blank	0.000	19:37
Control Test	0.080	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 06/01/2023 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,DALTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 07:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:43

SUBJECT'S LAST NAME: NEGRETE

DATE/TIME OF TEST: 05/20/2023 09:10:55

Diagnostics Check	OK	09:04
Air Blank	0.000	09:05
Control Test	0.081	09:05
Air Blank	0.000	09:06
Subject Sample # 1	0.111	09:06
Breath Volume	1.910	09:06
Air Blank	0.000	09:07
Air Blank	0.000	09:08
Subject Sample # 2	0.111	09:09
Breath Volume	2.699	09:09
Air Blank	0.000	09:10
Control Test	0.080	09:10
Air Blank	0.000	09:10
Diagnostics Check	OK	09:10

Data Download Date/Time: 05/26/2023 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/06/2023 23:09:17

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:05	No Sample Provided
Breath Volume	0.000	23:05	
Air Blank	0.000	23:05	
Air Blank	0.000	23:07	
Subject Sample	# 2	23:07	Subject Test Refused
Breath Volume	0.000	23:07	
Air Blank	0.000	23:08	
Control Test	0.079	23:08	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 06/30/2023 07:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,GLEN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 05/20/2023 00:42:18

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.076	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.000	00:38
Breath Volume	1.343	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.000	00:40
Breath Volume	1.425	00:40
Air Blank	0.000	00:41
Control Test	0.075	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 06/30/2023 07:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: LOVELL,TRENTON
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SAPP

DATE/TIME OF TEST: 05/26/2023 00:12:02

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.075	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.086	00:07
Breath Volume	1.414	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.085	00:10
Breath Volume	1.636	00:10
Air Blank	0.000	00:11
Control Test	0.075	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:12

Data Download Date/Time: 06/30/2023 07:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD, GLENN, B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD, GLENN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/26/2023 22:14:14

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.076	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.015	22:09
Breath Volume	2.746	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.010	22:12
Breath Volume	1.835	22:12
Air Blank	0.000	22:13
Control Test	0.075	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 06/30/2023 07:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/27/2023 06:25:28

Diagnostics Check	OK	06:19
Air Blank	0.000	06:19
Control Test	0.076	06:20
Air Blank	0.000	06:20
Subject Sample # 1	0.148	06:21
Breath Volume	1.550	06:21
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.154	06:23
Breath Volume	1.625	06:23
Air Blank	0.000	06:24
Control Test	0.076	06:24
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 06/30/2023 07:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUDSWORTH,SGT,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,CPL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 05/28/2023 01:22:55

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.077	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.152	01:18
Breath Volume	2.152	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:20
Subject Sample # 2	0.139	01:21
Breath Volume	1.312	01:21
Air Blank	0.000	01:22
Control Test	0.076	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 06/30/2023 07:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: MACIAS,NELSON
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: TORINO

DATE/TIME OF TEST: 05/10/2023 00:47:27

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.239	00:43
Breath Volume	1.664	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.258	00:45
Breath Volume	1.714	00:45
Air Blank	0.000	00:46
Control Test	0.075	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 06/27/2023 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/27/2023 16:27:40

Diagnostics Check	OK	16:21
Air Blank	0.000	16:21
Control Test	0.079	16:22
Air Blank	0.000	16:22
Subject Sample # 1	0.181	16:23
Breath Volume	1.316	16:23
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.175	16:26
Breath Volume	1.234	16:26
Air Blank	0.000	16:26
Control Test	0.079	16:27
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

Data Download Date/Time: 06/27/2023 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, TYLER, M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CARSON, KERRY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 05/06/2023 03:48:57

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.059	03:44
Breath Volume	3.339	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.058	03:47
Breath Volume	3.300	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 05/08/2023 16:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODDO,ISIAH,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ODDO,ISIAH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: HARTSFIELD

DATE/TIME OF TEST: 05/08/2023 00:44:52

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.035	00:40
Breath Volume	1.812	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:42
Subject Sample # 2	0.034	00:43
Breath Volume	2.761	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 05/08/2023 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,PAUL,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RODRIGUEZ,PAUL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: GOMES

DATE/TIME OF TEST: 05/08/2023 01:37:45

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.080	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1 0.194	01:33	
Breath Volume	1.343	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample	# 2	01:36	RFI Detect
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Control Test	0.078	01:37	
Air Blank	0.000	01:37	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 05/08/2023 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,PAUL,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RODRIGUEZ,PAUL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: GOMES

DATE/TIME OF TEST: 05/08/2023 01:49:17

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.207	01:44
Breath Volume	2.902	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.194	01:47
Breath Volume	1.171	01:47
Air Blank	0.000	01:48
Control Test	0.077	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 05/08/2023 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORTENBERRY,ZACHARY,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FORTENBERRY,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 05/10/2023 01:34:57

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.079	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.141	01:30
Breath Volume	3.273	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.132	01:33
Breath Volume	3.519	01:33
Air Blank	0.000	01:34
Control Test	0.078	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 05/12/2023 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,SHAWN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KELLY,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 13:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/10/2023 15:33:04

Diagnostics Check	OK	15:26
Air Blank	0.000	15:27
Control Test	0.082	15:27
Air Blank	0.000	15:28
Subject Sample # 1	0.061	15:28
Breath Volume	1.277	15:28
Air Blank	0.000	15:29
Air Blank	0.000	15:31
Subject Sample # 2	0.065	15:31
Breath Volume	1.746	15:31
Air Blank	0.000	15:32
Control Test	0.082	15:32
Air Blank	0.000	15:32
Diagnostics Check	OK	15:33

Data Download Date/Time: 06/08/2023 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,PATRICK,W
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BLACKWOOD,PATRICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: PARENTE

DATE/TIME OF TEST: 05/13/2023 00:25:16

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.196	00:20
Breath Volume	1.515	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.181	00:23
Breath Volume	1.105	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 06/08/2023 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: PARENTE

DATE/TIME OF TEST: 05/13/2023 01:09:40

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.081	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.223	01:05
Breath Volume	3.285	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.223	01:08
Breath Volume	2.359	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 06/08/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELENZ,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ELENZ,TIMOTHY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 05/14/2023 23:14:14

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.181	23:09
Breath Volume	1.863	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:11
Subject Sample # 2	0.176	23:12
Breath Volume	1.746	23:12
Air Blank	0.000	23:13
Control Test	0.079	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 06/08/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CARSON,KERRY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: TRIGO

DATE/TIME OF TEST: 05/15/2023 23:27:25

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.194	23:22
Breath Volume	4.089	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.176	23:25
Breath Volume	1.679	23:25
Air Blank	0.000	23:26
Control Test	0.078	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 06/08/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 05/16/2023 20:13:39

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:10	
Control Test	0.082	20:11	
Air Blank	0.000	20:11	
Subject Sample	# 1	20:12	Subject Test Refused
Breath Volume	0.000	20:12	
Air Blank	0.000	20:12	
Control Test	0.081	20:13	
Air Blank	0.000	20:13	
Diagnostics Check	OK	20:13	

Data Download Date/Time: 06/08/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: DAVIS,NATHANIEL
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/19/2023 21:47:58

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:42	
Control Test	0.082	21:42	
Air Blank	0.000	21:42	
Subject Sample	# 1 0.156	21:43	
Breath Volume	1.171	21:43	
Air Blank	0.000	21:44	
Air Blank	0.000	21:45	
Subject Sample	# 2 0.152	21:46	
Breath Volume	2.875	21:46	
Air Blank	0.000	21:47	
Control Test	0.000	21:47	RFI Detect
Air Blank	0.000	21:47	

Data Download Date/Time: 06/08/2023 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/20/2023 02:39:21

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.082	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:37	Interferent Detect
Breath Volume	1.546	02:37	
Air Blank	0.000	02:38	
Control Test	0.082	02:38	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 06/08/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/20/2023 02:47:31

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.115	02:43
Breath Volume	2.765	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.108	02:46
Breath Volume	1.441	02:46
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 06/08/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,PAUL,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RODRIGUEZ,PAUL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 05/20/2023 18:09:34

Diagnostics Check	OK	18:00	
Air Blank	0.000	18:01	
Control Test	0.082	18:01	
Air Blank	0.000	18:02	
Subject Sample	# 1	18:05	No Sample Provided
Breath Volume	0.000	18:05	
Air Blank	0.000	18:05	
Air Blank	0.000	18:07	
Subject Sample	# 2	18:08	Subject Test Refused
Breath Volume	0.000	18:08	
Air Blank	0.000	18:08	
Control Test	0.082	18:08	
Air Blank	0.000	18:09	
Diagnostics Check	OK	18:09	

Data Download Date/Time: 06/08/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BICKNELL,BRANDON,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CARSON,KERRY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 05/21/2023 00:11:41

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.082	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.000	00:07
Breath Volume	2.140	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.000	00:10
Breath Volume	1.652	00:10
Air Blank	0.000	00:10
Control Test	0.081	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 06/08/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BICKNELL,BRANDON,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WELCH,BLAKE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: ZIEL

DATE/TIME OF TEST: 05/21/2023 00:46:45

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.081	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.078	00:41
Breath Volume	1.527	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.077	00:45
Breath Volume	1.109	00:45
Air Blank	0.000	00:45
Control Test	0.081	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 06/08/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WHITE,TRAVIS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: KNEEN

DATE/TIME OF TEST: 05/21/2023 04:47:14

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.140	04:42
Breath Volume	3.003	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.136	04:45
Breath Volume	2.207	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 06/08/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBIN,BRITTON,L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: CORBIN,BRITTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: FRID

DATE/TIME OF TEST: 05/21/2023 19:49:56

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.081	19:44
Air Blank	0.000	19:44
Subject Sample # 1	0.255	19:45
Breath Volume	2.492	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:47
Subject Sample # 2	0.235	19:48
Breath Volume	1.148	19:48
Air Blank	0.000	19:49
Control Test	0.081	19:49
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 06/08/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYD,J,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOYD,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:28

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 05/22/2023 13:58:19

Diagnostics Check	OK	13:49
Air Blank	0.000	13:49
Control Test	0.081	13:50
Air Blank	0.000	13:50
Subject Sample # 1	0.344	13:51
Breath Volume	1.640	13:51
Air Blank	0.000	13:52
Air Blank	0.000	13:54
Subject Sample # 2	0.363	13:56
Breath Volume	1.761	13:56
Air Blank	0.000	13:57
Control Test	0.080	13:57
Air Blank	0.000	13:58
Diagnostics Check	OK	13:58

Data Download Date/Time: 06/08/2023 15:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYD,J,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOYD,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: ARSENAULT

DATE/TIME OF TEST: 05/28/2023 15:50:21

Diagnostics Check	OK	15:47	
Air Blank	0.000	15:47	
Control Test	0.082	15:48	
Air Blank	0.000	15:48	
Subject Sample	# 1	15:49	Subject Test Refused
Breath Volume	0.000	15:49	
Air Blank	0.000	15:49	
Control Test	0.082	15:49	
Air Blank	0.000	15:50	
Diagnostics Check	OK	15:50	

Data Download Date/Time: 06/08/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: PLYMALE

DATE/TIME OF TEST: 05/29/2023 01:43:24

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.082	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.188	01:39
Breath Volume	1.703	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.183	01:41
Breath Volume	1.785	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 06/08/2023 15:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODDO,ISAIAH,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BRUCE,HEATHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: WYNN

DATE/TIME OF TEST: 05/30/2023 21:28:38

Diagnostics Check	OK	21:23	
Air Blank	0.000	21:24	
Control Test	0.082	21:24	
Air Blank	0.000	21:25	
Subject Sample	# 1	21:27	Subject Test Refused
Breath Volume	0.000	21:27	
Air Blank	0.000	21:27	
Control Test	0.082	21:27	
Air Blank		21:28	RFI Detect
Air Blank	0.000	21:28	

Data Download Date/Time: 06/08/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEW,REECE,K

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: DEW,REECE

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1346146

DRY GAS EXPIRY DATE: 07/07/2023

VIOLATION CODE: DUI

ARREST TIME: 10:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: DOMINGO

DATE/TIME OF TEST: 05/13/2023 23:17:12

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:06	
Control Test	0.081	23:06	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:10	No Sample Provided
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Air Blank	0.000	23:12	
Subject Sample	# 2	23:15	No Sample Provided
Breath Volume	0.000	23:15	
Air Blank	0.000	23:16	
Control Test	0.080	23:16	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 06/30/2023 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HERNDON,DAVID,E

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: HERNDON,DAVID

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 1346146

DRY GAS EXPIRY DATE: 07/07/2023

VIOLATION CODE: DUI

ARREST TIME: 19:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: MOSELEY

DATE/TIME OF TEST: 05/27/2023 21:00:13

Diagnostics Check	OK	20:57	
Air Blank	0.000	20:57	
Control Test	0.081	20:58	
Air Blank	0.000	20:58	
Subject Sample	# 1	20:58	Subject Test Refused
Breath Volume	0.000	20:58	
Air Blank	0.000	20:59	
Control Test	0.081	20:59	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:00	

Data Download Date/Time: 06/30/2023 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,LUCAS,R
BREATH TEST OPERATOR AGENCY: Madison PD
ARRESTING OFFICER: SMITH,LUCAS
ARRESTING OFFICER AGENCY: Madison PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: EWDARDS

DATE/TIME OF TEST: 05/08/2023 20:25:16

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.077	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.200	20:20
Breath Volume	2.304	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.195	20:23
Breath Volume	2.339	20:23
Air Blank	0.000	20:24
Control Test	0.077	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 05/22/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,RYAN,O
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: DAVIS,RYAN
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/16/2023 17:19:29

Diagnostics Check	OK	17:14	
Air Blank	0.000	17:15	
Control Test	0.078	17:15	
Air Blank	0.000	17:16	
Subject Sample	# 1	17:18	Sequence Aborted
Breath Volume	0.609	17:18	
Air Blank	0.000	17:19	

Data Download Date/Time: 05/22/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,RYAN,O
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: DAVIS,RYAN
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/16/2023 17:35:48

Diagnostics Check	OK	17:23	
Air Blank	0.000	17:23	
Control Test	0.077	17:23	
Air Blank	0.000	17:24	
Subject Sample	# 1 0.227	17:27	Volume Not Met
Breath Volume	0.527	17:27	
Air Blank	0.000	17:28	
Air Blank	0.000	17:29	
Subject Sample	# 2 0.274	17:31	
Breath Volume	1.765	17:31	
Air Blank	0.000	17:31	
Air Blank	0.000	17:33	
Subject Sample	# 3 0.278	17:34	
Breath Volume	1.371	17:34	
Air Blank	0.000	17:34	
Control Test	0.073	17:35	Control Outside Tolerance
Air Blank	0.000	17:35	

Data Download Date/Time: 05/22/2023 15:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: NAVARRO,JAKE
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: BUCK

DATE/TIME OF TEST: 05/03/2023 00:08:19

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.185	00:03
Breath Volume	1.671	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.183	00:06
Breath Volume	1.281	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARGROVE, ERIK, A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MANSON, TIMOTHY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 05/06/2023 11:40:56

Diagnostics Check	OK	11:34
Air Blank	0.000	11:34
Control Test	0.081	11:34
Air Blank	0.000	11:35
Subject Sample # 1	0.000	11:36
Breath Volume	3.953	11:36
Air Blank	0.000	11:36
Air Blank	0.000	11:38
Subject Sample # 2	0.000	11:39
Breath Volume	4.621	11:39
Air Blank	0.000	11:40
Control Test	0.080	11:40
Air Blank	0.000	11:40
Diagnostics Check	OK	11:40

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FURMAN,FREDDIE,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 05/06/2023 18:16:48

Diagnostics Check	OK	18:10
Air Blank	0.000	18:11
Control Test	0.081	18:11
Air Blank	0.000	18:12
Subject Sample # 1	0.000	18:12
Breath Volume	3.816	18:12
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.000	18:15
Breath Volume	5.226	18:15
Air Blank	0.000	18:15
Control Test	0.080	18:16
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 05/10/2023 04:26:25

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.194	04:21
Breath Volume	2.015	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.196	04:24
Breath Volume	3.921	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: DENTON

DATE/TIME OF TEST: 05/12/2023 00:05:16

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.081	23:58
Air Blank	0.000	23:58
Subject Sample # 1	0.222	00:00
Breath Volume	1.605	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:02
Subject Sample # 2	0.235	00:03
Breath Volume	1.792	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 05/13/2023 00:10:38

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:08	
Control Test	0.080	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:09	
Control Test	0.080	00:10	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: WAHL

DATE/TIME OF TEST: 05/13/2023 16:39:00

Diagnostics Check	OK	16:32
Air Blank	0.000	16:33
Control Test	0.080	16:33
Air Blank	0.000	16:33
Subject Sample # 1	0.145	16:34
Breath Volume	2.238	16:34
Air Blank	0.000	16:34
Air Blank	0.000	16:36
Subject Sample # 2	0.148	16:37
Breath Volume	2.722	16:37
Air Blank	0.000	16:38
Control Test	0.080	16:38
Air Blank	0.000	16:38
Diagnostics Check	OK	16:39

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: HARTUNG, NICHILOS
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: LOPEZ HERNANDEZ

DATE/TIME OF TEST: 05/14/2023 06:22:20

Diagnostics Check	OK	06:16
Air Blank	0.000	06:16
Control Test	0.081	06:16
Air Blank	0.000	06:17
Subject Sample # 1	0.190	06:18
Breath Volume	1.671	06:18
Air Blank	0.000	06:18
Air Blank	0.000	06:20
Subject Sample # 2	0.185	06:20
Breath Volume	1.742	06:20
Air Blank	0.000	06:21
Control Test	0.080	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARTER,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PETTIJOHN,PHILIP
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 05/14/2023 20:52:34

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.080	20:45
Air Blank	0.000	20:45
Subject Sample # 1	0.208	20:47
Breath Volume	1.882	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.215	20:51
Breath Volume	2.000	20:51
Air Blank	0.000	20:51
Control Test	0.079	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 05/20/2023 02:38:15

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.304	02:32
Breath Volume	1.250	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.292	02:36
Breath Volume	1.761	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,DAVID,W
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: NELSON,DAVID
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: BOGGESS

DATE/TIME OF TEST: 05/20/2023 08:26:23

Diagnostics Check	OK	08:19
Air Blank	0.000	08:20
Control Test	0.080	08:20
Air Blank	0.000	08:21
Subject Sample # 1	0.049	08:21
Breath Volume	4.003	08:21
Air Blank	0.000	08:22
Air Blank	0.000	08:24
Subject Sample # 2	0.046	08:24
Breath Volume	3.449	08:24
Air Blank	0.000	08:25
Control Test	0.080	08:25
Air Blank	0.000	08:26
Diagnostics Check	OK	08:26

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY,ALISHA,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: CONLEY,ALISHA
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PARADISE

DATE/TIME OF TEST: 05/27/2023 03:47:27

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:36	
Subject Sample	# 1 0.171	03:40	Volume Not Met
Breath Volume	0.445	03:40	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 2 0.152	03:45	Volume Not Met
Breath Volume	0.082	03:45	
Air Blank	0.000	03:46	
Control Test	0.079	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,FORREST,H
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALLEN,FORREST
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 05/27/2023 04:11:58

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.161	04:07
Breath Volume	2.070	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.153	04:10
Breath Volume	1.304	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CASEY,TYLER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 05/27/2023 21:03:20

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.148	20:58
Breath Volume	1.343	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.159	21:01
Breath Volume	1.332	21:01
Air Blank	0.000	21:02
Control Test	0.081	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARTER,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: FOSTER,ALEX
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: FRIAS AVALOS

DATE/TIME OF TEST: 05/28/2023 20:54:55

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.080	20:45
Air Blank	0.000	20:46
Subject Sample # 1	0.284	20:47
Breath Volume	1.367	20:47
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.321	20:50
Breath Volume	3.726	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 3	0.320	20:53
Breath Volume	2.347	20:53
Air Blank	0.000	20:54
Control Test	0.079	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORE,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: GORE,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: BUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 05/29/2023 19:43:31

Diagnostics Check	OK	19:36
Air Blank	0.000	19:37
Control Test	0.081	19:37
Air Blank	0.000	19:38
Subject Sample # 1	0.145	19:38
Breath Volume	2.824	19:38
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.147	19:42
Breath Volume	4.097	19:42
Air Blank	0.000	19:42
Control Test	0.080	19:42
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ, JUAN, C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HOGAN, JEREMY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: HENRICKS DAVIS

DATE/TIME OF TEST: 05/31/2023 02:32:04

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.133	02:27
Breath Volume	3.660	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.136	02:30
Breath Volume	2.191	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:32

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARIN,ERIN,M
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BRAXTON,KYNLEY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: BOHON

DATE/TIME OF TEST: 05/31/2023 07:40:49

Diagnostics Check	OK	07:32
Air Blank	0.000	07:33
Control Test	0.080	07:33
Air Blank	0.000	07:33
Subject Sample # 1	0.275	07:35
Breath Volume	2.050	07:35
Air Blank	0.000	07:36
Air Blank	0.000	07:38
Subject Sample # 2	0.283	07:39
Breath Volume	2.300	07:39
Air Blank	0.000	07:39
Control Test	0.078	07:40
Air Blank	0.000	07:40
Diagnostics Check	OK	07:40

Data Download Date/Time: 06/26/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: COKONOUGH, ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 05/06/2023 05:49:35

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:38	
Control Test	0.081	05:38	
Air Blank	0.000	05:39	
Subject Sample	# 1	05:42	No Sample Provided
Breath Volume	0.000	05:42	
Air Blank	0.000	05:43	
Air Blank	0.000	05:45	
Subject Sample	# 2	05:48	No Sample Provided
Breath Volume	0.000	05:48	
Air Blank	0.000	05:48	
Control Test	0.081	05:48	
Air Blank	0.000	05:49	
Diagnostics Check	OK	05:49	

Data Download Date/Time: 05/31/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY, WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: FORT

DATE/TIME OF TEST: 05/07/2023 18:52:56

Diagnostics Check	OK	18:37	
Air Blank	0.000	18:38	
Control Test	0.080	18:38	
Air Blank	0.000	18:38	
Subject Sample	# 1 0.025	18:40	
Breath Volume	1.179	18:40	
Air Blank	0.000	18:40	
Air Blank	0.000	18:42	
Subject Sample	# 2 0.021	18:45	Volume Not Met
Breath Volume	0.988	18:45	
Air Blank	0.000	18:46	
Air Blank	0.000	18:48	
Subject Sample	# 3 0.019	18:51	Volume Not Met
Breath Volume	0.210	18:51	
Air Blank	0.000	18:52	
Control Test	0.081	18:52	
Air Blank	0.000	18:52	
Diagnostics Check	OK	18:52	

Data Download Date/Time: 05/31/2023 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY, WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: FORT

DATE/TIME OF TEST: 05/07/2023 19:14:31

Diagnostics Check	OK	18:59	
Air Blank	0.000	18:59	
Control Test	0.079	18:59	
Air Blank	0.000	19:00	
Subject Sample	# 1 0.016	19:03	Volume Not Met
Breath Volume	0.585	19:03	
Air Blank	0.000	19:04	
Air Blank	0.000	19:05	
Subject Sample	# 2 0.016	19:08	
Breath Volume	1.414	19:08	
Air Blank	0.000	19:08	
Air Blank	0.000	19:10	
Subject Sample	# 3 0.015	19:13	
Breath Volume	1.296	19:13	
Air Blank	0.000	19:13	
Control Test	0.079	19:13	
Air Blank	0.000	19:14	
Diagnostics Check	OK	19:14	

Data Download Date/Time: 05/31/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS, BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: WIMBERLEY

DATE/TIME OF TEST: 05/07/2023 21:08:26

Diagnostics Check	OK	20:57	
Air Blank	0.000	20:57	
Control Test	0.080	20:57	
Air Blank	0.000	20:58	
Subject Sample	# 1	21:01	No Sample Provided
Breath Volume	0.000	21:01	
Air Blank	0.000	21:01	
Air Blank	0.000	21:03	
Subject Sample	# 2	21:07	No Sample Provided
Breath Volume	0.000	21:07	
Air Blank	0.000	21:07	
Control Test	0.081	21:07	
Air Blank	0.000	21:08	
Diagnostics Check	OK	21:08	

Data Download Date/Time: 05/31/2023 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUCHNIR,EUGENE,V
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: KOUCHNIR,EUGENE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CLEMENT

DATE/TIME OF TEST: 05/09/2023 03:16:15

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.097	03:11
Breath Volume	3.574	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.102	03:14
Breath Volume	3.660	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 05/31/2023 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/10/2023 02:19:03

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.259	02:11
Breath Volume	1.203	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.232	02:14
Breath Volume	1.761	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 3	0.245	02:17
Breath Volume	1.554	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:19

Data Download Date/Time: 05/31/2023 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 05/14/2023 01:23:31

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.081	01:19	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.081	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 05/31/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: COKONOUGH,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: NOVAK

DATE/TIME OF TEST: 05/19/2023 20:48:44

Diagnostics Check	OK	20:42
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.277	20:44
Breath Volume	1.992	20:44
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.284	20:47
Breath Volume	2.335	20:47
Air Blank	0.000	20:47
Control Test	0.079	20:48
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 05/31/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: NADERA,ANTONIO
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 05/20/2023 02:08:18

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.200	02:03
Breath Volume	1.316	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:05
Subject Sample # 2	0.189	02:06
Breath Volume	1.128	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 05/31/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SENN, PHILIP
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: BAME

DATE/TIME OF TEST: 05/25/2023 22:18:20

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:07	
Control Test	0.079	22:07	
Air Blank	0.000	22:08	
Subject Sample	# 1	22:11	No Sample Provided
Breath Volume	0.000	22:11	
Air Blank	0.000	22:11	
Air Blank	0.000	22:13	
Subject Sample	# 2	22:16	No Sample Provided
Breath Volume	0.000	22:16	
Air Blank	0.000	22:17	
Control Test	0.080	22:17	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 05/31/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,SCOTT,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: KENNEDY,SCOTT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:17

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 05/28/2023 09:52:14

Diagnostics Check	OK	09:45
Air Blank	0.000	09:46
Control Test	0.080	09:46
Air Blank	0.000	09:47
Subject Sample # 1	0.000	09:47
Breath Volume	2.691	09:47
Air Blank	0.000	09:48
Air Blank	0.000	09:50
Subject Sample # 2	0.000	09:50
Breath Volume	1.652	09:50
Air Blank	0.000	09:51
Control Test	0.081	09:51
Air Blank	0.000	09:52
Diagnostics Check	OK	09:52

Data Download Date/Time: 05/31/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCALIA, ADAM, C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DAVIS, ALEX
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 05/07/2023 02:34:06

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.081	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:27	No Sample Provided
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample	# 2	02:32	No Sample Provided
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.082	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 05/31/2023 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAITHER,PHILLIP
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:27

SUBJECT'S LAST NAME: CRIGGER

DATE/TIME OF TEST: 05/11/2023 12:05:42

Diagnostics Check	OK	11:58	
Air Blank	0.000	11:59	
Control Test	0.081	11:59	
Air Blank	0.000	12:00	
Subject Sample	# 1 0.289	12:00	
Breath Volume	2.144	12:00	
Air Blank	0.000	12:01	
Air Blank	0.000	12:03	
Subject Sample	# 2	12:04	Subject Test Refused
Breath Volume	0.457	12:04	
Air Blank	0.000	12:04	
Control Test	0.080	12:05	
Air Blank	0.000	12:05	
Diagnostics Check	OK	12:05	

Data Download Date/Time: 05/31/2023 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BROADDUS,AARON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: WIACEK

DATE/TIME OF TEST: 05/11/2023 23:05:11

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.081	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.000	23:00
Breath Volume	2.828	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.000	23:03
Breath Volume	2.617	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 05/31/2023 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOST,CRAIG,R
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: STONE,MICHAEL
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 05/12/2023 18:43:43

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:41	
Control Test	0.081	18:41	
Air Blank	0.000	18:42	
Subject Sample	# 1	18:42	Subject Test Refused
Breath Volume	0.000	18:42	
Air Blank	0.000	18:42	
Control Test	0.082	18:43	
Air Blank	0.000	18:43	
Diagnostics Check	OK	18:43	

Data Download Date/Time: 05/31/2023 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: TREVINO

DATE/TIME OF TEST: 05/12/2023 18:50:36

Diagnostics Check	OK	18:47	
Air Blank	0.000	18:48	
Control Test	0.081	18:48	
Air Blank	0.000	18:48	
Subject Sample	# 1	18:49	Subject Test Refused
Breath Volume	0.000	18:49	
Air Blank	0.000	18:49	
Control Test	0.082	18:49	
Air Blank	0.000	18:50	
Diagnostics Check	OK	18:50	

Data Download Date/Time: 05/31/2023 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAFFA,AMELIA,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: HOWARD,AUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/14/2023 02:22:25

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.132	02:17
Breath Volume	2.496	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.126	02:20
Breath Volume	3.562	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 05/31/2023 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAFFA,AMELIA,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: HOWARD,AUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/14/2023 02:35:29

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.119	02:30
Breath Volume	3.351	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.119	02:33
Breath Volume	3.000	02:33
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 05/31/2023 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALAS,LUIS,E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: AMSDEN

DATE/TIME OF TEST: 05/20/2023 18:58:22

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:47	
Control Test	0.081	18:47	
Air Blank	0.000	18:48	
Subject Sample	# 1	18:51	No Sample Provided
Breath Volume	0.000	18:51	
Air Blank	0.000	18:51	
Air Blank	0.000	18:53	
Subject Sample	# 2	18:57	No Sample Provided
Breath Volume	0.000	18:57	
Air Blank	0.000	18:57	
Control Test	0.082	18:57	
Air Blank	0.000	18:58	
Diagnostics Check	OK	18:58	

Data Download Date/Time: 05/31/2023 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: BLAICH

DATE/TIME OF TEST: 05/23/2023 13:25:28

Diagnostics Check	OK	13:18
Air Blank	0.000	13:18
Control Test	0.081	13:18
Air Blank	0.000	13:19
Subject Sample # 1	0.213	13:20
Breath Volume	1.335	13:20
Air Blank	0.000	13:21
Air Blank	0.000	13:22
Subject Sample # 2	0.226	13:23
Breath Volume	2.105	13:23
Air Blank	0.000	13:24
Control Test	0.080	13:24
Air Blank	0.000	13:25
Diagnostics Check	OK	13:25

Data Download Date/Time: 05/31/2023 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PUGH,CHAISE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 05/25/2023 08:04:16

Diagnostics Check	OK	07:58
Air Blank	0.000	07:58
Control Test	0.081	07:58
Air Blank	0.000	07:59
Subject Sample # 1	0.149	07:59
Breath Volume	3.460	07:59
Air Blank	0.000	08:00
Air Blank	0.000	08:02
Subject Sample # 2	0.148	08:02
Breath Volume	3.523	08:02
Air Blank	0.000	08:03
Control Test	0.081	08:03
Air Blank	0.000	08:04
Diagnostics Check	OK	08:04

Data Download Date/Time: 05/31/2023 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MARIE

DATE/TIME OF TEST: 05/27/2023 02:02:55

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.103	01:58
Breath Volume	3.863	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:00
Subject Sample # 2	0.099	02:01
Breath Volume	3.531	02:01
Air Blank	0.000	02:02
Control Test	0.081	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 05/31/2023 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402335621
DRY GAS EXPIRY DATE: 02/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MACEDO

DATE/TIME OF TEST: 05/06/2023 00:33:27

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.079	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.185	00:28
Breath Volume	1.929	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.185	00:31
Breath Volume	2.097	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 05/16/2023 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: URBINA,JOSE,I
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: CASTANEDA,R
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402335621
DRY GAS EXPIRY DATE: 02/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: URBINA

DATE/TIME OF TEST: 05/06/2023 02:48:38

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank	0.000	02:47	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Control Test	0.080	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 05/16/2023 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402335621
DRY GAS EXPIRY DATE: 02/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/07/2023 00:38:59

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:36	
Control Test	0.080	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:37	Subject Test Refused
Breath Volume	0.000	00:37	
Air Blank	0.000	00:38	
Control Test	0.080	00:38	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 05/16/2023 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANAICK,KAYLA,N
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: FRNAKLIN,DANIEL
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402335621
DRY GAS EXPIRY DATE: 02/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/12/2023 00:48:21

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.115	00:41
Breath Volume	1.132	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:43
Subject Sample # 2	0.120	00:46
Breath Volume	1.476	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 05/16/2023 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTANEDA,RAFAEL,
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: CASTANEDA,RAFAEL
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402335621
DRY GAS EXPIRY DATE: 02/14/2025
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: ARELLANO

DATE/TIME OF TEST: 05/20/2023 04:57:23

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:49	
Control Test	0.080	04:49	
Air Blank	0.000	04:50	
Subject Sample	# 1	04:53	No Sample Provided
Breath Volume	0.000	04:53	
Air Blank	0.000	04:53	
Air Blank	0.000	04:55	
Subject Sample	# 2	04:56	Subject Test Refused
Breath Volume	0.000	04:56	
Air Blank	0.000	04:56	
Control Test	0.080	04:56	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 06/08/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402335621
DRY GAS EXPIRY DATE: 02/14/2025
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 05/20/2023 05:48:51

Diagnostics Check	OK	05:39
Air Blank	0.000	05:39
Control Test	0.080	05:40
Air Blank	0.000	05:40
Subject Sample # 1	0.165	05:41
Breath Volume	1.941	05:41
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.137	05:44
Breath Volume	2.015	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 3	0.159	05:47
Breath Volume	2.542	05:47
Air Blank	0.000	05:47
Control Test	0.079	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 06/08/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTANEDA,RAFAEL,
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: PITTS,DALTON
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402335621
DRY GAS EXPIRY DATE: 02/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MCKINSTER

DATE/TIME OF TEST: 05/21/2023 01:16:15

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.160	01:11
Breath Volume	3.398	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.159	01:14
Breath Volume	1.972	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 06/08/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402335621
DRY GAS EXPIRY DATE: 02/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JOINER

DATE/TIME OF TEST: 05/21/2023 03:24:51

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Control Test	0.080	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 06/08/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402335621
DRY GAS EXPIRY DATE: 02/14/2025
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 05/24/2023 19:24:53

Diagnostics Check	OK	19:18
Air Blank	0.000	19:19
Control Test	0.079	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.058	19:20
Breath Volume	1.628	19:20
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.056	19:23
Breath Volume	1.753	19:23
Air Blank	0.000	19:23
Control Test	0.080	19:24
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

Data Download Date/Time: 06/08/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402335621
DRY GAS EXPIRY DATE: 02/14/2025
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 05/25/2023 20:40:54

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.079	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.268	20:35
Breath Volume	2.351	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 2	0.272	20:39
Breath Volume	3.281	20:39
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 06/08/2023 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402335621
DRY GAS EXPIRY DATE: 02/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: LOWRY

DATE/TIME OF TEST: 05/27/2023 01:08:55

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.176	01:04
Breath Volume	2.121	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 2	0.175	01:07
Breath Volume	1.593	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 06/08/2023 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THERIAULT,TANNER,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MARTIN,NICHOLAS
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: CARLOS

DATE/TIME OF TEST: 05/03/2023 00:39:37

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.217	00:35
Breath Volume	1.449	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.215	00:38
Breath Volume	1.632	00:38
Air Blank	0.000	00:38
Control Test	0.080	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 05/31/2023 21:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHUNN,PETER,E
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: QUAGLIA,ERIC
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 05/05/2023 03:41:45

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.370	03:37
Breath Volume	2.398	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:39
Subject Sample # 2	0.370	03:40
Breath Volume	2.195	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 05/31/2023 21:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TAYLOR,ASHLEY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:13

SUBJECT'S LAST NAME: TAULBEE

DATE/TIME OF TEST: 05/11/2023 13:43:07

Diagnostics Check	OK	13:36
Air Blank	0.000	13:36
Control Test	0.080	13:37
Air Blank	0.000	13:37
Subject Sample # 1	0.192	13:38
Breath Volume	1.503	13:38
Air Blank	0.000	13:38
Air Blank	0.000	13:40
Subject Sample # 2	0.183	13:41
Breath Volume	2.179	13:41
Air Blank	0.000	13:42
Control Test	0.080	13:42
Air Blank	0.000	13:43
Diagnostics Check	OK	13:43

Data Download Date/Time: 05/31/2023 21:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 05/21/2023 20:54:45

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.079	20:49
Air Blank	0.000	20:49
Subject Sample # 1	0.014	20:50
Breath Volume	2.167	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.015	20:53
Breath Volume	3.046	20:53
Air Blank	0.000	20:53
Control Test	0.080	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 05/31/2023 21:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MASCETTA,JOHN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: MILONE

DATE/TIME OF TEST: 05/23/2023 19:46:13

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.080	19:40
Air Blank	0.000	19:41
Subject Sample # 1	0.251	19:41
Breath Volume	2.453	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.250	19:44
Breath Volume	2.007	19:44
Air Blank	0.000	19:45
Control Test	0.078	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 05/31/2023 21:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/24/2023 15:44:20

Diagnostics Check	OK	15:38
Air Blank	0.000	15:38
Control Test	0.080	15:38
Air Blank	0.000	15:39
Subject Sample # 1	0.046	15:39
Breath Volume	1.574	15:39
Air Blank	0.000	15:40
Air Blank	0.000	15:42
Subject Sample # 2	0.048	15:42
Breath Volume	2.089	15:42
Air Blank	0.000	15:43
Control Test	0.079	15:43
Air Blank	0.000	15:44
Diagnostics Check	OK	15:44

Data Download Date/Time: 05/31/2023 21:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,FABIAN,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GILBERT,MATT
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: MCLEAN

DATE/TIME OF TEST: 05/28/2023 01:09:33

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.192	01:04
Breath Volume	2.160	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.190	01:07
Breath Volume	2.136	01:07
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 05/31/2023 21:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRADY,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: FAUSTINO

DATE/TIME OF TEST: 05/03/2023 22:48:26

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.081	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.000	22:44
Breath Volume	2.414	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.000	22:47
Breath Volume	2.300	22:47
Air Blank	0.000	22:47
Control Test	0.082	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 05/10/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYALA,K,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: ROMERO,C
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HARLAN

DATE/TIME OF TEST: 05/03/2023 23:22:34

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.081	23:21	
Air Blank		23:22	RFI Detect
Air Blank	0.000	23:22	

Data Download Date/Time: 05/10/2023 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYALA,K,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: ROMERO,C
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HARLAN

DATE/TIME OF TEST: 05/03/2023 23:26:39

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:24	
Control Test	0.081	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 1	23:25	Subject Test Refused
Breath Volume	0.000	23:25	
Air Blank	0.000	23:25	
Control Test	0.081	23:26	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 05/10/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOLLEY,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: CORREIA

DATE/TIME OF TEST: 05/04/2023 03:53:36

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.082	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.164	03:49
Breath Volume	1.308	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.177	03:52
Breath Volume	1.921	03:52
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 05/10/2023 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARCARO,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: KOMARNICKI

DATE/TIME OF TEST: 05/04/2023 17:26:32

Diagnostics Check	OK	17:21	
Air Blank	0.000	17:22	
Control Test	0.081	17:22	
Air Blank	0.000	17:23	
Subject Sample	# 1	17:25	Subject Test Refused
Breath Volume	0.000	17:25	
Air Blank	0.000	17:25	
Control Test	0.081	17:25	
Air Blank	0.000	17:26	
Diagnostics Check	OK	17:26	

Data Download Date/Time: 05/10/2023 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARCARO,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: KOMARNICKI

DATE/TIME OF TEST: 05/04/2023 17:51:41

Diagnostics Check	OK	17:43	
Air Blank	0.000	17:43	
Control Test	0.081	17:44	
Air Blank	0.000	17:44	
Subject Sample	# 1	17:47	No Sample Provided
Breath Volume	0.000	17:47	
Air Blank	0.000	17:48	
Air Blank	0.000	17:50	
Subject Sample	# 2	17:50	Subject Test Refused
Breath Volume	0.000	17:50	
Air Blank	0.000	17:50	
Control Test	0.082	17:51	
Air Blank	0.000	17:51	
Diagnostics Check	OK	17:51	

Data Download Date/Time: 05/10/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MACIAS,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/04/2023 18:50:30

Diagnostics Check	OK	18:43
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:44
Subject Sample # 1	0.156	18:45
Breath Volume	2.984	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:47
Subject Sample # 2	0.153	18:48
Breath Volume	3.246	18:48
Air Blank	0.000	18:49
Control Test	0.081	18:49
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

Data Download Date/Time: 05/10/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BELL,B
ARRESTING OFFICER AGENCY: Indian River Shores
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: VANATTA

DATE/TIME OF TEST: 05/04/2023 22:08:18

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.081	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.000	22:04
Breath Volume	2.316	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.000	22:06
Breath Volume	1.757	22:06
Air Blank	0.000	22:07
Control Test	0.081	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 05/10/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,RICHARD,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GONZALEZ,ALI
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MEEK

DATE/TIME OF TEST: 05/06/2023 01:11:46

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.133	01:07
Breath Volume	1.394	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.128	01:10
Breath Volume	1.664	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,RICHARD,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,DAVID
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DAMES

DATE/TIME OF TEST: 05/06/2023 03:28:42

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.081	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:27	
Control Test	0.082	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,RICHARD,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ELLIOT,KEIRSHA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: GUADALUPE

DATE/TIME OF TEST: 05/06/2023 04:07:20

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.000	04:02
Breath Volume	1.125	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.000	04:05
Breath Volume	1.210	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: BENNEFIELD

DATE/TIME OF TEST: 05/06/2023 06:32:17

Diagnostics Check	OK	06:25
Air Blank	0.000	06:25
Control Test	0.081	06:26
Air Blank	0.000	06:26
Subject Sample # 1	0.123	06:27
Breath Volume	1.613	06:27
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample # 2	0.120	06:30
Breath Volume	1.156	06:30
Air Blank	0.000	06:31
Control Test	0.080	06:31
Air Blank	0.000	06:32
Diagnostics Check	OK	06:32

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,RICHARD,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOON,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 05/06/2023 22:56:28

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:53	
Control Test	0.081	22:54	
Air Blank	0.000	22:54	
Subject Sample	# 1	22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:55	
Control Test	0.081	22:55	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MCFARLAND

DATE/TIME OF TEST: 05/07/2023 01:15:33

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:13	
Control Test	0.081	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Control Test	0.081	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 05/07/2023 02:13:30

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.081	02:11	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.081	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CAOQUETTE,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HEEN

DATE/TIME OF TEST: 05/07/2023 03:10:13

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.082	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.082	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GONZALEZ,A
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BISCHOFF

DATE/TIME OF TEST: 05/07/2023 22:08:20

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.082	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.039	22:04
Breath Volume	1.214	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.040	22:06
Breath Volume	1.718	22:06
Air Blank	0.000	22:07
Control Test	0.081	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,R,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SELDES,Z
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GARCIA-RODAS

DATE/TIME OF TEST: 05/07/2023 23:58:37

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.082	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.159	23:53
Breath Volume	2.253	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.157	23:57
Breath Volume	1.480	23:57
Air Blank	0.000	23:57
Control Test	0.082	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ROMER

DATE/TIME OF TEST: 05/08/2023 02:25:48

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.082	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:24	
Control Test	0.082	02:25	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/09/2023 00:12:02

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.081	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.140	00:07
Breath Volume	4.335	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:09
Subject Sample # 2	0.141	00:10
Breath Volume	3.929	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:12

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: HERNANDEZ-BROWN

DATE/TIME OF TEST: 05/09/2023 23:53:19

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:48	
Control Test	0.081	23:48	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:52	
Control Test	0.082	23:52	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 05/10/2023 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ANTOSIA, E
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: GIRGENTI

DATE/TIME OF TEST: 05/10/2023 23:48:54

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:46	
Control Test	0.081	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:47	
Control Test	0.081	23:48	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 06/07/2023 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEACH,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: CONWAY

DATE/TIME OF TEST: 05/13/2023 01:01:21

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.081	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	01:00	
Control Test	0.082	01:00	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 06/07/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MARTIN,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HENDERSHOT

DATE/TIME OF TEST: 05/13/2023 02:52:21

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.081	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:47	No Sample Provided
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 2	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:51	
Control Test	0.082	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 06/07/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MARTIN,Q
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: ELLOWITZ

DATE/TIME OF TEST: 05/13/2023 19:15:37

Diagnostics Check	OK	19:09
Air Blank	0.000	19:09
Control Test	0.082	19:10
Air Blank	0.000	19:10
Subject Sample # 1	0.000	19:11
Breath Volume	2.929	19:11
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.000	19:14
Breath Volume	3.117	19:14
Air Blank	0.000	19:14
Control Test	0.081	19:15
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 06/07/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LOSTY,S
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: RICHARD

DATE/TIME OF TEST: 05/14/2023 01:18:10

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.082	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Control Test	0.081	01:17	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 06/07/2023 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CONSALO,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: SHAUGHNESSY

DATE/TIME OF TEST: 05/14/2023 23:27:32

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:25	
Control Test	0.081	23:25	
Air Blank	0.000	23:25	
Subject Sample	# 1	23:26	Subject Test Refused
Breath Volume	0.000	23:26	
Air Blank	0.000	23:26	
Control Test	0.082	23:26	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time: 06/07/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GONZALEZ, A
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: SHEETS

DATE/TIME OF TEST: 05/16/2023 21:32:10

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:29	
Control Test	0.081	21:30	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:30	Subject Test Refused
Breath Volume	0.000	21:30	
Air Blank	0.000	21:31	
Control Test	0.082	21:31	
Air Blank	0.000	21:32	
Diagnostics Check	OK	21:32	

Data Download Date/Time: 06/07/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCROGGIN,JOHN,C
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SCROGGIN,JOHN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 13:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 05/18/2023 14:43:03

Diagnostics Check	OK	14:36	
Air Blank	0.000	14:36	
Control Test	0.081	14:36	
Air Blank	0.000	14:37	
Subject Sample	# 1 0.191	14:38	
Breath Volume	4.437	14:38	
Air Blank	0.000	14:39	
Air Blank	0.000	14:40	
Subject Sample	# 2	14:41	RFI Detect
Breath Volume	4.250	14:41	
Air Blank	0.000	14:42	
Control Test	0.079	14:42	
Air Blank	0.000	14:42	
Diagnostics Check	OK	14:43	

Data Download Date/Time: 06/07/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCROGGIN,JOHN,C
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SCROGGIN,JOHN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 13:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 05/18/2023 14:52:34

Diagnostics Check	OK	14:46
Air Blank	0.000	14:46
Control Test	0.080	14:46
Air Blank	0.000	14:47
Subject Sample # 1	0.190	14:48
Breath Volume	4.289	14:48
Air Blank	0.000	14:48
Air Blank	0.000	14:50
Subject Sample # 2	0.177	14:50
Breath Volume	2.605	14:50
Air Blank	0.000	14:51
Control Test	0.079	14:51
Air Blank	0.000	14:52
Diagnostics Check	OK	14:52

Data Download Date/Time: 06/07/2023 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SANTANA,L
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/19/2023 05:21:07

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:18	
Control Test	0.081	05:19	
Air Blank	0.000	05:19	
Subject Sample	# 1	05:19	Subject Test Refused
Breath Volume	0.000	05:19	
Air Blank	0.000	05:20	
Control Test	0.081	05:20	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 06/07/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE, B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: POIRIER

DATE/TIME OF TEST: 05/20/2023 03:28:08

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.253	03:22
Breath Volume	1.671	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.248	03:26
Breath Volume	1.960	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 06/07/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SAUCIER,M
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 05/20/2023 03:50:47

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.109	03:46
Breath Volume	1.394	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.101	03:49
Breath Volume	2.269	03:49
Air Blank	0.000	03:49
Control Test	0.081	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 06/07/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON, M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SIGMON

DATE/TIME OF TEST: 05/21/2023 22:40:53

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.081	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.130	22:36
Breath Volume	1.167	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:38
Subject Sample # 2	0.111	22:39
Breath Volume	1.679	22:39
Air Blank	0.000	22:39
Control Test	0.081	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 06/07/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KACHELE, M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: REGINA

DATE/TIME OF TEST: 05/22/2023 03:48:22

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.081	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Control Test	0.081	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 06/07/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LOSTY,S
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/23/2023 20:58:02

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.081	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.153	20:53
Breath Volume	1.585	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.152	20:56
Breath Volume	1.656	20:56
Air Blank	0.000	20:57
Control Test	0.079	20:57
Air Blank	0.000	20:57
Diagnostics Check	OK	20:58

Data Download Date/Time: 06/07/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CHOVINARD,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: CIANCIO

DATE/TIME OF TEST: 05/26/2023 21:34:18

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.081	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.000	21:30
Breath Volume	1.503	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.000	21:32
Breath Volume	1.738	21:32
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 06/07/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MARTINEZ,O
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: MAZZELLA

DATE/TIME OF TEST: 05/27/2023 22:00:16

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.259	21:55
Breath Volume	3.382	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.250	21:58
Breath Volume	2.531	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 06/07/2023 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: THOMAS,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/28/2023 02:34:19

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.081	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.081	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 06/07/2023 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MARTIN,Q
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SHARKEY

DATE/TIME OF TEST: 05/28/2023 22:36:26

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:34	
Control Test	0.081	22:34	
Air Blank	0.000	22:34	
Subject Sample	# 1	22:35	Subject Test Refused
Breath Volume	0.000	22:35	
Air Blank	0.000	22:35	
Control Test	0.081	22:35	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

Data Download Date/Time: 06/07/2023 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LOSTY,S
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: CHAMPAGNE

DATE/TIME OF TEST: 05/29/2023 00:05:37

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.081	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Control Test	0.081	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 06/07/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI, OFC. M., L
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI, OFC. M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 05/12/2023 23:31:31

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:20	
Control Test	0.079	23:20	
Air Blank	0.000	23:21	
Subject Sample	# 1 0.156	23:24	Volume Not Met
Breath Volume	0.382	23:24	
Air Blank	0.000	23:25	
Air Blank	0.000	23:26	
Subject Sample	# 2 0.163	23:29	Volume Not Met
Breath Volume	0.121	23:29	
Air Blank	0.000	23:30	
Control Test	0.077	23:30	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 05/15/2023 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI, OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI, OFC. M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 05/12/2023 23:40:43

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.077	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.165	23:35
Breath Volume	1.394	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.151	23:39
Breath Volume	1.875	23:39
Air Blank	0.000	23:39
Control Test	0.078	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 05/15/2023 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 05/13/2023 00:04:32

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.078	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.000	00:00
Breath Volume	2.410	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.000	00:03
Breath Volume	2.058	00:03
Air Blank	0.000	00:03
Control Test	0.078	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 05/15/2023 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYTH,GARY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SMYTH,GARY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CHOJNOWSKI

DATE/TIME OF TEST: 05/15/2023 21:41:49

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.079	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.024	21:37
Breath Volume	1.937	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:39
Subject Sample # 2	0.025	21:40
Breath Volume	1.292	21:40
Air Blank	0.000	21:40
Control Test	0.079	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 06/23/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: TAMAYO

DATE/TIME OF TEST: 05/15/2023 22:39:46

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.077	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.253	22:35
Breath Volume	2.441	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.259	22:38
Breath Volume	2.550	22:38
Air Blank	0.000	22:38
Control Test	0.078	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 06/23/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MCCALLISTER

DATE/TIME OF TEST: 05/17/2023 20:23:32

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.077	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.230	20:19
Breath Volume	2.117	20:19
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 2	0.218	20:21
Breath Volume	1.562	20:21
Air Blank	0.000	20:22
Control Test	0.076	20:22
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

Data Download Date/Time: 06/23/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: MCALISTER

DATE/TIME OF TEST: 05/18/2023 22:20:29

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.000	22:16
Breath Volume	2.964	22:16
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.000	22:19
Breath Volume	2.875	22:19
Air Blank	0.000	22:19
Control Test	0.079	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 06/23/2023 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: DESAI

DATE/TIME OF TEST: 05/19/2023 03:54:36

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.214	03:50
Breath Volume	2.730	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.201	03:53
Breath Volume	2.062	03:53
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 06/23/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERNER,N
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 05/19/2023 04:52:10

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.079	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.165	04:45
Breath Volume	1.386	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.169	04:50
Breath Volume	1.628	04:50
Air Blank	0.000	04:51
Control Test	0.080	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 06/23/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHADL,CAMERON,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SCHADL,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: CERVANTES

DATE/TIME OF TEST: 05/20/2023 01:09:00

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.000	01:04
Breath Volume	2.457	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.000	01:07
Breath Volume	3.343	01:07
Air Blank	0.000	01:08
Control Test	0.078	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:09

Data Download Date/Time: 06/23/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PONCE-CRUZ

DATE/TIME OF TEST: 05/21/2023 01:35:24

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.102	01:30
Breath Volume	2.960	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.101	01:33
Breath Volume	4.171	01:33
Air Blank	0.000	01:34
Control Test	0.079	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 06/23/2023 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,S,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PACHECO,E
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SOUZA

DATE/TIME OF TEST: 05/23/2023 01:39:14

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.077	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.097	01:34
Breath Volume	2.398	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.092	01:37
Breath Volume	2.210	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 06/23/2023 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BALES,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:43

SUBJECT'S LAST NAME: WACHOB

DATE/TIME OF TEST: 05/24/2023 10:22:55

Diagnostics Check	OK	10:16
Air Blank	0.000	10:17
Control Test	0.079	10:17
Air Blank	0.000	10:18
Subject Sample # 1	0.234	10:18
Breath Volume	1.878	10:18
Air Blank	0.000	10:19
Air Blank	0.000	10:21
Subject Sample # 2	0.235	10:21
Breath Volume	1.570	10:21
Air Blank	0.000	10:22
Control Test	0.078	10:22
Air Blank	0.000	10:22
Diagnostics Check	OK	10:22

Data Download Date/Time: 06/23/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: ENGLAND

DATE/TIME OF TEST: 05/24/2023 21:36:26

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.078	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.132	21:31
Breath Volume	1.546	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.128	21:34
Breath Volume	1.433	21:34
Air Blank	0.000	21:35
Control Test	0.079	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 06/23/2023 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GOAD,C
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: IAGOVINO

DATE/TIME OF TEST: 05/26/2023 01:18:12

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.151	01:13
Breath Volume	2.371	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.148	01:16
Breath Volume	2.277	01:16
Air Blank	0.000	01:17
Control Test	0.078	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 06/23/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHAWB,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 05/26/2023 01:30:31

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.000	01:26
Breath Volume	3.894	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.000	01:29
Breath Volume	3.683	01:29
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 06/23/2023 12:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHADL,CAMERON,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SCHADL,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/26/2023 01:56:52

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.166	01:52
Breath Volume	1.812	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:54
Subject Sample # 2	0.184	01:55
Breath Volume	2.234	01:55
Air Blank	0.000	01:55
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 06/23/2023 12:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHADL,CAMERON,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SCHADL,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/26/2023 02:07:09

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.078	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.174	02:02
Breath Volume	1.574	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.175	02:05
Breath Volume	1.867	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 06/23/2023 12:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: SEUBERT

DATE/TIME OF TEST: 05/26/2023 04:50:47

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.077	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.180	04:46
Breath Volume	2.945	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:48
Subject Sample # 2	0.174	04:49
Breath Volume	2.417	04:49
Air Blank	0.000	04:49
Control Test	0.078	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 06/23/2023 12:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SURRAN,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIETRON,M
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: LEIGH

DATE/TIME OF TEST: 05/27/2023 15:38:54

Diagnostics Check	OK	15:26	
Air Blank	0.000	15:26	
Control Test	0.080	15:27	
Air Blank	0.000	15:27	
Subject Sample	# 1 0.000	15:30	Volume Not Met
Breath Volume	0.867	15:30	
Air Blank	0.000	15:31	
Air Blank	0.000	15:33	
Subject Sample	# 2 0.000	15:34	
Breath Volume	1.480	15:34	
Air Blank	0.000	15:35	
Air Blank	0.000	15:37	
Subject Sample	# 3 0.000	15:37	
Breath Volume	1.605	15:37	
Air Blank	0.000	15:37	
Control Test	0.078	15:38	
Air Blank	0.000	15:38	
Diagnostics Check	OK	15:38	

Data Download Date/Time: 06/23/2023 12:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GARRARD

DATE/TIME OF TEST: 05/28/2023 02:05:03

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.207	01:57
Breath Volume	3.476	01:57
Air Blank	0.000	01:58
Air Blank	0.000	01:59
Subject Sample # 2	0.228	02:00
Breath Volume	4.890	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 3	0.220	02:03
Breath Volume	3.195	02:03
Air Blank	0.000	02:04
Control Test	0.076	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:05

Data Download Date/Time: 06/23/2023 12:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: SCALA

DATE/TIME OF TEST: 05/28/2023 02:32:08

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.078	02:30	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:31	RFI Detect
Breath Volume	0.000	02:31	
Air Blank		02:31	RFI Detect
Air Blank	0.000	02:32	

Data Download Date/Time: 06/23/2023 12:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: SCALA

DATE/TIME OF TEST: 05/28/2023 02:41:51

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:34	
Control Test	0.078	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 1 0.338	02:37	
Breath Volume	1.453	02:37	
Air Blank	0.000	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 2 0.329	02:40	
Breath Volume	1.253	02:40	
Air Blank	0.000	02:40	
Control Test	0.074	02:41	Control Outside Tolerance
Air Blank	0.000	02:41	

Data Download Date/Time: 06/23/2023 12:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: SCALA

DATE/TIME OF TEST: 05/28/2023 03:00:54

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:46	
Control Test	0.077	02:47	
Air Blank	0.000	02:47	
Subject Sample	# 1 0.305	02:50	
Breath Volume	1.273	02:50	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 2 0.305	02:55	Volume Not Met
Breath Volume	0.988	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 3 0.313	02:59	
Breath Volume	1.183	02:59	
Air Blank	0.000	02:59	
Control Test	0.075	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 06/23/2023 12:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,S,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LEVERIDGE,M
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CRUZ-PENA

DATE/TIME OF TEST: 05/29/2023 03:36:36

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.080	03:25	
Air Blank	0.000	03:26	
Subject Sample	# 1 0.000	03:29	Volume Not Met
Breath Volume	0.132	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 2	03:35	No Sample Provided
Breath Volume	0.000	03:35	
Air Blank	0.000	03:35	
Control Test	0.079	03:35	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 06/23/2023 12:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,S,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: HARKAY,G
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 05/29/2023 04:15:04

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.079	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.196	04:10
Breath Volume	1.335	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.200	04:13
Breath Volume	2.531	04:13
Air Blank	0.000	04:14
Control Test	0.078	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:15

Data Download Date/Time: 06/23/2023 12:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYTH,GARY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRUNO,N
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: DAVISON

DATE/TIME OF TEST: 05/30/2023 23:38:41

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.078	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.159	23:34
Breath Volume	2.167	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.161	23:37
Breath Volume	1.367	23:37
Air Blank	0.000	23:37
Control Test	0.079	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 06/23/2023 12:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SMYTH,GARY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 05/01/2023 21:17:06

Diagnostics Check	Fail	21:16	Voltage/Current Test
Air Blank	0.000	21:17	

Data Download Date/Time: 05/16/2023 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SMYTH,GARY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 05/01/2023 21:27:14

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:19	
Control Test	0.079	21:19	
Air Blank	0.000	21:20	
Subject Sample	# 1	21:23	No Sample Provided
Breath Volume	0.000	21:23	
Air Blank	0.000	21:23	
Air Blank	0.000	21:25	
Subject Sample	# 2	21:25	Subject Test Refused
Breath Volume	0.000	21:25	
Air Blank	0.000	21:26	
Control Test	0.079	21:26	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:27	

Data Download Date/Time: 05/16/2023 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: CONNOLLY

DATE/TIME OF TEST: 05/07/2023 17:02:10

Diagnostics Check	OK	16:52	
Air Blank	0.000	16:53	
Control Test	0.077	16:53	
Air Blank	0.000	16:54	
Subject Sample	# 1 0.252	16:57	Volume Not Met
Breath Volume	0.062	16:57	
Air Blank	0.000	16:57	
Air Blank	0.000	16:59	
Subject Sample	# 2	17:00	Subject Test Refused
Breath Volume	0.000	17:00	
Air Blank	0.000	17:01	
Control Test	0.078	17:01	
Air Blank	0.000	17:02	
Diagnostics Check	OK	17:02	

Data Download Date/Time: 05/16/2023 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRECKLE,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: BAMACA-LOPEZ

DATE/TIME OF TEST: 05/08/2023 02:36:08

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.174	02:31
Breath Volume	2.789	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.179	02:34
Breath Volume	2.457	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 05/16/2023 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: LEONHARDT

DATE/TIME OF TEST: 05/24/2023 20:29:27

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.079	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.182	20:24
Breath Volume	1.171	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:26
Subject Sample # 2	0.173	20:27
Breath Volume	1.289	20:27
Air Blank	0.000	20:28
Control Test	0.078	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 06/23/2023 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, WILLIAM, S
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCKENZIE, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: BLOW

DATE/TIME OF TEST: 05/08/2023 10:55:51

Diagnostics Check	OK	10:50
Air Blank	0.000	10:50
Control Test	0.082	10:50
Air Blank	0.000	10:51
Subject Sample # 1	0.000	10:51
Breath Volume	3.011	10:51
Air Blank	0.000	10:52
Air Blank	0.000	10:54
Subject Sample # 2	0.000	10:54
Breath Volume	2.718	10:54
Air Blank	0.000	10:54
Control Test	0.082	10:55
Air Blank	0.000	10:55
Diagnostics Check	OK	10:55

Data Download Date/Time: 05/31/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, EDWIN,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HOON, ED
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: MICHEAL

DATE/TIME OF TEST: 05/11/2023 11:31:50

Diagnostics Check	OK	11:31	
Air Blank		11:31	RFI Detect
Air Blank	0.000	11:31	

Data Download Date/Time: 05/31/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,EDWIN,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RODRIGUEZ,EDWIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/11/2023 11:40:26

Diagnostics Check	OK	11:38	
Air Blank	0.000	11:39	
Control Test	0.081	11:39	
Air Blank	0.000	11:39	
		11:40	RFI Detect
Air Blank		11:40	RFI Detect

Data Download Date/Time: 05/31/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,EDWIN,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RODRIGUEZ,EDWIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/11/2023 11:48:40

Diagnostics Check	OK	11:42
Air Blank	0.000	11:43
Control Test	0.081	11:43
Air Blank	0.000	11:44
Subject Sample # 1	0.000	11:44
Breath Volume	2.929	11:44
Air Blank	0.000	11:44
Air Blank	0.000	11:46
Subject Sample # 2	0.000	11:47
Breath Volume	3.402	11:47
Air Blank	0.000	11:47
Control Test	0.082	11:48
Air Blank	0.000	11:48
Diagnostics Check	OK	11:48

Data Download Date/Time: 05/31/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: DOUGHNUT

DATE/TIME OF TEST: 05/12/2023 08:46:52

Diagnostics Check	OK	08:40
Air Blank	0.000	08:41
Control Test	0.081	08:41
Air Blank	0.000	08:42
Subject Sample # 1	0.000	08:42
Breath Volume	2.351	08:42
Air Blank	0.000	08:43
Air Blank	0.000	08:45
Subject Sample # 2	0.000	08:45
Breath Volume	1.628	08:45
Air Blank	0.000	08:45
Control Test	0.082	08:46
Air Blank	0.000	08:46
Diagnostics Check	OK	08:46

Data Download Date/Time: 05/31/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF,ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 05/12/2023 08:58:25

Diagnostics Check	OK	08:52
Air Blank	0.000	08:52
Control Test	0.082	08:53
Air Blank	0.000	08:53
Subject Sample # 1	0.000	08:54
Breath Volume	1.492	08:54
Air Blank	0.000	08:54
Air Blank	0.000	08:56
Subject Sample # 2	0.000	08:57
Breath Volume	2.472	08:57
Air Blank	0.000	08:57
Control Test	0.082	08:57
Air Blank	0.000	08:58
Diagnostics Check	OK	08:58

Data Download Date/Time: 05/31/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LEON,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DE LEON,LUIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: WOLFF

DATE/TIME OF TEST: 05/12/2023 09:10:41

Diagnostics Check	OK	09:04
Air Blank	0.000	09:05
Control Test	0.082	09:05
Air Blank	0.000	09:06
Subject Sample # 1	0.000	09:06
Breath Volume	1.394	09:06
Air Blank	0.000	09:06
Air Blank	0.000	09:08
Subject Sample # 2	0.000	09:09
Breath Volume	1.468	09:09
Air Blank	0.000	09:09
Control Test	0.082	09:10
Air Blank	0.000	09:10
Diagnostics Check	OK	09:10

Data Download Date/Time: 05/31/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF,ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 05/12/2023 09:19:23

Diagnostics Check	OK	09:13
Air Blank	0.000	09:14
Control Test	0.081	09:14
Air Blank	0.000	09:14
Subject Sample # 1	0.000	09:15
Breath Volume	1.644	09:15
Air Blank	0.000	09:15
Air Blank	0.000	09:17
Subject Sample # 2	0.000	09:18
Breath Volume	1.406	09:18
Air Blank	0.000	09:18
Control Test	0.082	09:18
Air Blank	0.000	09:19
Diagnostics Check	OK	09:19

Data Download Date/Time: 05/31/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LEON,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DE LEON,LUIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: WOLFF

DATE/TIME OF TEST: 05/12/2023 09:28:31

Diagnostics Check	OK	09:22
Air Blank	0.000	09:22
Control Test	0.081	09:23
Air Blank	0.000	09:23
Subject Sample # 1	0.000	09:24
Breath Volume	1.261	09:24
Air Blank	0.000	09:24
Air Blank	0.000	09:26
Subject Sample # 2	0.000	09:27
Breath Volume	1.546	09:27
Air Blank	0.000	09:27
Control Test	0.082	09:27
Air Blank	0.000	09:28
Diagnostics Check	OK	09:28

Data Download Date/Time: 05/31/2023 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF,ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 05/12/2023 09:37:03

Diagnostics Check	OK	09:31
Air Blank	0.000	09:31
Control Test	0.081	09:32
Air Blank	0.000	09:32
Subject Sample # 1	0.000	09:32
Breath Volume	1.691	09:32
Air Blank	0.000	09:33
Air Blank	0.000	09:35
Subject Sample # 2	0.000	09:35
Breath Volume	1.312	09:35
Air Blank	0.000	09:36
Control Test	0.081	09:36
Air Blank	0.000	09:36
Diagnostics Check	OK	09:37

Data Download Date/Time: 05/31/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LEON,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DE LEON,LUIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: WOLFF

DATE/TIME OF TEST: 05/12/2023 09:46:01

Diagnostics Check	OK	09:40
Air Blank	0.000	09:40
Control Test	0.082	09:40
Air Blank	0.000	09:41
Subject Sample # 1	0.000	09:41
Breath Volume	1.593	09:41
Air Blank	0.000	09:42
Air Blank	0.000	09:44
Subject Sample # 2	0.000	09:44
Breath Volume	1.445	09:44
Air Blank	0.000	09:45
Control Test	0.081	09:45
Air Blank	0.000	09:45
Diagnostics Check	OK	09:46

Data Download Date/Time: 05/31/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF,ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 05/12/2023 09:55:18

Diagnostics Check	OK	09:49
Air Blank	0.000	09:50
Control Test	0.082	09:50
Air Blank	0.000	09:50
Subject Sample # 1	0.000	09:51
Breath Volume	1.644	09:51
Air Blank	0.000	09:51
Air Blank	0.000	09:53
Subject Sample # 2	0.000	09:53
Breath Volume	1.648	09:53
Air Blank	0.000	09:54
Control Test	0.083	09:54
Air Blank	0.000	09:55
Diagnostics Check	OK	09:55

Data Download Date/Time: 05/31/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LEON,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DE LEON,LUIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 05/12/2023 10:05:11

Diagnostics Check	OK	09:59
Air Blank	0.000	09:59
Control Test	0.082	10:00
Air Blank	0.000	10:00
Subject Sample # 1	0.000	10:01
Breath Volume	1.347	10:01
Air Blank	0.000	10:01
Air Blank	0.000	10:03
Subject Sample # 2	0.000	10:03
Breath Volume	1.566	10:03
Air Blank	0.000	10:04
Control Test	0.082	10:04
Air Blank	0.000	10:05
Diagnostics Check	OK	10:05

Data Download Date/Time: 05/31/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF,ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 05/12/2023 10:13:08

Diagnostics Check	OK	10:07
Air Blank	0.000	10:07
Control Test	0.082	10:08
Air Blank	0.000	10:08
Subject Sample # 1	0.000	10:09
Breath Volume	1.558	10:09
Air Blank	0.000	10:09
Air Blank	0.000	10:11
Subject Sample # 2	0.000	10:11
Breath Volume	1.378	10:11
Air Blank	0.000	10:12
Control Test	0.082	10:12
Air Blank	0.000	10:13
Diagnostics Check	OK	10:13

Data Download Date/Time: 05/31/2023 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LEON,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DE LEON,LUIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: WOLFF

DATE/TIME OF TEST: 05/12/2023 10:23:00

Diagnostics Check	OK	10:17
Air Blank	0.000	10:17
Control Test	0.081	10:18
Air Blank	0.000	10:18
Subject Sample # 1	0.000	10:18
Breath Volume	1.585	10:18
Air Blank	0.000	10:19
Air Blank	0.000	10:21
Subject Sample # 2	0.000	10:21
Breath Volume	1.679	10:21
Air Blank	0.000	10:22
Control Test	0.082	10:22
Air Blank	0.000	10:22
Diagnostics Check	OK	10:23

Data Download Date/Time: 05/31/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 05/01/2023 00:24:40

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.024	00:19
Breath Volume	1.324	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.026	00:23
Breath Volume	2.503	00:23
Air Blank	0.000	00:23
Control Test	0.080	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 05/03/2023 14:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRIEND,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: TAMBURELLO

DATE/TIME OF TEST: 05/01/2023 02:38:33

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.139	02:33
Breath Volume	2.777	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.137	02:37
Breath Volume	2.394	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 05/03/2023 14:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SITKO,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: DECATO

DATE/TIME OF TEST: 05/01/2023 04:06:05

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.123	04:01
Breath Volume	2.210	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.115	04:04
Breath Volume	1.941	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 05/03/2023 14:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRIEND,S
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HERNANDEZ RODRIGUEZ

DATE/TIME OF TEST: 05/09/2023 01:30:23

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.169	01:25
Breath Volume	1.992	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.177	01:28
Breath Volume	3.062	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 05/22/2023 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARCHMAN,MARCHMAN,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DEPUTY,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 05/11/2023 05:10:24

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.079	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.110	05:05
Breath Volume	1.660	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.109	05:08
Breath Volume	2.195	05:08
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 05/22/2023 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRADLEY,W
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCOTT,BRADLEY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 05/11/2023 11:32:31

Diagnostics Check	OK	11:26
Air Blank	0.000	11:26
Control Test	0.080	11:27
Air Blank	0.000	11:27
Subject Sample # 1	0.000	11:28
Breath Volume	1.582	11:28
Air Blank	0.000	11:28
Air Blank	0.000	11:30
Subject Sample # 2	0.000	11:31
Breath Volume	2.109	11:31
Air Blank	0.000	11:31
Control Test	0.080	11:31
Air Blank	0.000	11:32
Diagnostics Check	OK	11:32

Data Download Date/Time: 05/22/2023 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRADLEY,W
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCOTT,BRADLEY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: GARRICK

DATE/TIME OF TEST: 05/11/2023 13:30:15

Diagnostics Check	OK	13:24
Air Blank	0.000	13:24
Control Test	0.080	13:25
Air Blank	0.000	13:25
Subject Sample # 1	0.000	13:25
Breath Volume	1.445	13:25
Air Blank	0.000	13:26
Air Blank	0.000	13:28
Subject Sample # 2	0.000	13:28
Breath Volume	2.453	13:28
Air Blank	0.000	13:29
Control Test	0.080	13:29
Air Blank	0.000	13:30
Diagnostics Check	OK	13:30

Data Download Date/Time: 05/22/2023 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,C,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DEPUTY,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 05/12/2023 02:23:54

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.106	02:19
Breath Volume	1.445	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.110	02:22
Breath Volume	1.937	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 05/22/2023 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DEPUTY,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 05/12/2023 04:13:23

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.018	04:12	RFI Detect
Air Blank	0.000	04:13	

Data Download Date/Time: 05/22/2023 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DEP,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 05/12/2023 04:20:51

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.000	04:16
Breath Volume	2.281	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.000	04:19
Breath Volume	2.113	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 05/22/2023 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 05:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 05/12/2023 06:19:52

Diagnostics Check	OK	06:13
Air Blank	0.000	06:14
Control Test	0.080	06:14
Air Blank	0.000	06:15
Subject Sample # 1	0.000	06:15
Breath Volume	2.527	06:15
Air Blank	0.000	06:16
Air Blank	0.000	06:18
Subject Sample # 2	0.000	06:18
Breath Volume	2.523	06:18
Air Blank	0.000	06:18
Control Test	0.080	06:19
Air Blank	0.000	06:19
Diagnostics Check	OK	06:19

Data Download Date/Time: 05/22/2023 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 05/12/2023 10:52:20

Diagnostics Check	OK	10:46
Air Blank	0.000	10:46
Control Test	0.080	10:47
Air Blank	0.000	10:47
Subject Sample # 1	0.000	10:48
Breath Volume	2.187	10:48
Air Blank	0.000	10:48
Air Blank	0.000	10:50
Subject Sample # 2	0.000	10:50
Breath Volume	2.121	10:50
Air Blank	0.000	10:51
Control Test	0.080	10:51
Air Blank	0.000	10:52
Diagnostics Check	OK	10:52

Data Download Date/Time: 05/22/2023 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BOWN

DATE/TIME OF TEST: 05/13/2023 02:55:05

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.163	02:50
Breath Volume	2.769	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample # 2	0.162	02:53
Breath Volume	1.343	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 05/22/2023 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SKOLNICK,LUIS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:44

SUBJECT'S LAST NAME: WAHL

DATE/TIME OF TEST: 05/14/2023 14:12:23

Diagnostics Check	OK	14:05
Air Blank	0.000	14:05
Control Test	0.079	14:06
Air Blank	0.000	14:06
Subject Sample # 1	0.272	14:07
Breath Volume	1.355	14:07
Air Blank	0.000	14:08
Air Blank	0.000	14:10
Subject Sample # 2	0.265	14:10
Breath Volume	1.265	14:10
Air Blank	0.000	14:11
Control Test	0.079	14:11
Air Blank	0.000	14:12
Diagnostics Check	OK	14:12

Data Download Date/Time: 05/22/2023 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRIEND,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: SALVADOR

DATE/TIME OF TEST: 05/14/2023 21:47:26

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.132	21:42
Breath Volume	1.578	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.132	21:45
Breath Volume	1.812	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 05/22/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LAKE,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: CUETO

DATE/TIME OF TEST: 05/15/2023 00:17:51

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.000	00:13
Breath Volume	1.796	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.000	00:16
Breath Volume	1.312	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 05/22/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 05/18/2023 12:23:17

Diagnostics Check	OK	12:17
Air Blank	0.000	12:17
Control Test	0.080	12:18
Air Blank	0.000	12:18
Subject Sample # 1	0.000	12:19
Breath Volume	2.980	12:19
Air Blank	0.000	12:19
Air Blank	0.000	12:21
Subject Sample # 2	0.000	12:21
Breath Volume	3.382	12:21
Air Blank	0.000	12:22
Control Test	0.080	12:22
Air Blank	0.000	12:23
Diagnostics Check	OK	12:23

Data Download Date/Time: 05/22/2023 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANO,MARDEE,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GRAY,A
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 05/20/2023 12:59:34

Diagnostics Check	OK	12:53
Air Blank	0.000	12:53
Control Test	0.080	12:53
Air Blank	0.000	12:54
Subject Sample # 1	0.177	12:54
Breath Volume	3.078	12:54
Air Blank	0.000	12:55
Air Blank	0.000	12:57
Subject Sample # 2	0.184	12:58
Breath Volume	2.355	12:58
Air Blank	0.000	12:58
Control Test	0.080	12:58
Air Blank	0.000	12:59
Diagnostics Check	OK	12:59

Data Download Date/Time: 05/22/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COLON, RAMONE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: KIMBALL

DATE/TIME OF TEST: 05/21/2023 01:12:15

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.221	01:07
Breath Volume	1.957	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.217	01:10
Breath Volume	2.062	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 05/22/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAZZARESE,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COLON,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: GROFF

DATE/TIME OF TEST: 05/22/2023 00:22:37

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.000	00:18
Breath Volume	1.210	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.000	00:21
Breath Volume	1.414	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 05/22/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO, MONICA, S
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PAILLE, SGT
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:13

SUBJECT'S LAST NAME: GUTIERREZ-GONZALEZ

DATE/TIME OF TEST: 05/23/2023 16:39:53

Diagnostics Check	OK	16:33
Air Blank	0.000	16:34
Control Test	0.079	16:34
Air Blank	0.000	16:34
Subject Sample # 1	0.257	16:35
Breath Volume	2.570	16:35
Air Blank	0.000	16:35
Air Blank	0.000	16:37
Subject Sample # 2	0.255	16:38
Breath Volume	1.519	16:38
Air Blank	0.000	16:38
Control Test	0.079	16:39
Air Blank	0.000	16:39
Diagnostics Check	OK	16:39

Data Download Date/Time: 06/29/2023 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANO,M,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MANO,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: GARROD

DATE/TIME OF TEST: 05/24/2023 13:20:59

Diagnostics Check	OK	13:14
Air Blank	0.000	13:15
Control Test	0.080	13:15
Air Blank	0.000	13:16
Subject Sample # 1	0.000	13:16
Breath Volume	1.816	13:16
Air Blank	0.000	13:17
Air Blank	0.000	13:19
Subject Sample # 2	0.000	13:19
Breath Volume	1.785	13:19
Air Blank	0.000	13:20
Control Test	0.080	13:20
Air Blank	0.000	13:20
Diagnostics Check	OK	13:20

Data Download Date/Time: 06/29/2023 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRIEND,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 05/27/2023 00:48:38

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.269	00:43
Breath Volume	1.437	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.262	00:47
Breath Volume	1.308	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 06/29/2023 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: KELLEHER

DATE/TIME OF TEST: 05/27/2023 01:47:00

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.185	01:42
Breath Volume	2.460	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:44
Subject Sample # 2	0.170	01:45
Breath Volume	1.988	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:47

Data Download Date/Time: 06/29/2023 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SALISBURY,OFFC
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 05/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DIEGUEZ

DATE/TIME OF TEST: 05/27/2023 03:28:23

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.129	03:22
Breath Volume	2.089	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.129	03:26
Breath Volume	2.164	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 06/29/2023 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAMPMANN,JOSHUA,Q
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: KAMPMANN,JOSHUA
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: BUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: SHOENECK

DATE/TIME OF TEST: 05/27/2023 18:17:26

Diagnostics Check	OK	18:11
Air Blank	0.000	18:11
Control Test	0.080	18:11
Air Blank	0.000	18:12
Subject Sample # 1	0.145	18:13
Breath Volume	1.699	18:13
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.146	18:15
Breath Volume	1.765	18:15
Air Blank	0.000	18:16
Control Test	0.079	18:16
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

Data Download Date/Time: 06/29/2023 07:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PARISH,G
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/20/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GROVES

DATE/TIME OF TEST: 05/28/2023 04:47:33

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.170	04:42
Breath Volume	2.484	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.165	04:46
Breath Volume	2.558	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 06/29/2023 07:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/29/2023 03:46:58

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.242	03:42
Breath Volume	3.917	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.233	03:45
Breath Volume	3.738	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 06/29/2023 07:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRADLEY,W
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCOTT,BRADLEY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 05/30/2023 07:09:53

Diagnostics Check	OK	07:04
Air Blank	0.000	07:04
Control Test	0.080	07:04
Air Blank	0.000	07:05
Subject Sample # 1	0.000	07:05
Breath Volume	2.019	07:05
Air Blank	0.000	07:06
Air Blank	0.000	07:08
Subject Sample # 2	0.000	07:08
Breath Volume	1.359	07:08
Air Blank	0.000	07:08
Control Test	0.081	07:09
Air Blank	0.000	07:09
Diagnostics Check	OK	07:09

Data Download Date/Time: 06/29/2023 07:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWOCHO,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: SCRIBNER

DATE/TIME OF TEST: 05/30/2023 13:42:57

Diagnostics Check	OK	13:35
Air Blank	0.000	13:35
Control Test	0.080	13:35
Air Blank	0.000	13:36
Subject Sample # 1	0.000	13:36
Breath Volume	2.320	13:36
Air Blank	0.000	13:37
Air Blank	0.000	13:39
Subject Sample # 2	0.000	13:41
Breath Volume	2.640	13:41
Air Blank	0.000	13:42
Control Test	0.080	13:42
Air Blank	0.000	13:42
Diagnostics Check	OK	13:42

Data Download Date/Time: 06/29/2023 07:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILD,JACOB
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: BUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: O'NEILL

DATE/TIME OF TEST: 05/28/2023 21:50:25

Diagnostics Check	OK	21:43
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:44
Subject Sample # 1	0.000	21:46
Breath Volume	2.976	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.000	21:49
Breath Volume	1.503	21:49
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 06/21/2023 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO, WILFRED
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: LABAT

DATE/TIME OF TEST: 05/05/2023 10:10:37

Diagnostics Check	OK	10:04
Air Blank	0.000	10:05
Control Test	0.085	10:05
Air Blank	0.000	10:05
Subject Sample # 1	0.198	10:06
Breath Volume	1.546	10:06
Air Blank	0.000	10:06
Air Blank	0.000	10:08
Subject Sample # 2	0.200	10:09
Breath Volume	2.464	10:09
Air Blank	0.000	10:09
Control Test	0.085	10:10
Air Blank	0.000	10:10
Diagnostics Check	OK	10:10

Data Download Date/Time: 06/21/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCGRAW, LOGAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: DETRESMERCADO

DATE/TIME OF TEST: 05/05/2023 16:31:46

Diagnostics Check	OK	16:25
Air Blank	0.000	16:25
Control Test	0.084	16:25
Air Blank	0.000	16:26
Subject Sample # 1	0.234	16:27
Breath Volume	1.503	16:27
Air Blank	0.000	16:27
Air Blank	0.000	16:29
Subject Sample # 2	0.232	16:30
Breath Volume	2.734	16:30
Air Blank	0.000	16:30
Control Test	0.083	16:31
Air Blank	0.000	16:31
Diagnostics Check	OK	16:31

Data Download Date/Time: 06/21/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: RAMIREZ CRUZ

DATE/TIME OF TEST: 05/05/2023 23:43:26

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.084	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.167	23:36
Breath Volume	1.476	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.165	23:41
Breath Volume	2.484	23:41
Air Blank	0.000	23:42
Control Test	0.083	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 06/21/2023 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JENNY, STEVEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: LOPEZ DE LEON

DATE/TIME OF TEST: 05/06/2023 02:11:44

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.084	02:00	
Air Blank	0.000	02:01	
Subject Sample	# 1 0.197	02:04	Volume Not Met
Breath Volume	0.511	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample	# 2 0.182	02:10	Volume Not Met
Breath Volume	0.386	02:10	
Air Blank	0.000	02:10	
Control Test	0.080	02:11	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 06/21/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HANSTINE,AUSTIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BERNADEL

DATE/TIME OF TEST: 05/06/2023 03:59:31

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.084	03:48	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:52	No Sample Provided
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample	# 2	03:58	No Sample Provided
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Control Test	0.084	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 06/21/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PREDELUS,SMITH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: MUSAEV

DATE/TIME OF TEST: 05/06/2023 06:37:17

Diagnostics Check	OK	06:31
Air Blank	0.000	06:31
Control Test	0.083	06:32
Air Blank	0.000	06:32
Subject Sample # 1	0.121	06:32
Breath Volume	1.386	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:35
Subject Sample # 2	0.127	06:35
Breath Volume	2.234	06:35
Air Blank	0.000	06:36
Control Test	0.083	06:36
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

Data Download Date/Time: 06/21/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JUDAH,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: DENT

DATE/TIME OF TEST: 05/06/2023 09:22:26

Diagnostics Check	OK	09:15
Air Blank	0.000	09:15
Control Test	0.084	09:15
Air Blank	0.000	09:16
Subject Sample # 1	0.172	09:18
Breath Volume	1.621	09:18
Air Blank	0.000	09:18
Air Blank	0.000	09:20
Subject Sample # 2	0.177	09:20
Breath Volume	1.488	09:20
Air Blank	0.000	09:21
Control Test	0.083	09:21
Air Blank	0.000	09:22
Diagnostics Check	OK	09:22

Data Download Date/Time: 06/21/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: VASON

DATE/TIME OF TEST: 05/07/2023 03:15:34

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.084	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.190	03:11
Breath Volume	2.433	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.190	03:14
Breath Volume	2.347	03:14
Air Blank	0.000	03:14
Control Test	0.084	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 06/21/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAURENT, WILFRID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: CARRASCOVILLARROEL

DATE/TIME OF TEST: 05/07/2023 06:57:46

Diagnostics Check	OK	06:51
Air Blank	0.000	06:51
Control Test	0.079	06:51
Air Blank	0.000	06:52
Subject Sample # 1	0.131	06:53
Breath Volume	1.531	06:53
Air Blank	0.000	06:54
Air Blank	0.000	06:55
Subject Sample # 2	0.137	06:56
Breath Volume	1.644	06:56
Air Blank	0.000	06:56
Control Test	0.083	06:57
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

Data Download Date/Time: 06/21/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WRINKLE,MIRANDA
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 05/07/2023 20:45:28

Diagnostics Check	OK	20:39
Air Blank	0.000	20:40
Control Test	0.084	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.209	20:41
Breath Volume	1.503	20:41
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.223	20:44
Breath Volume	2.437	20:44
Air Blank	0.000	20:44
Control Test	0.084	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 06/21/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI,ZACH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/31/2023 06:20:03

Diagnostics Check	OK	06:13
Air Blank	0.000	06:14
Control Test	0.084	06:14
Air Blank	0.000	06:14
Subject Sample # 1	0.136	06:15
Breath Volume	2.160	06:15
Air Blank	0.000	06:15
Air Blank	0.000	06:17
Subject Sample # 2	0.138	06:18
Breath Volume	2.792	06:18
Air Blank	0.000	06:19
Control Test	0.084	06:19
Air Blank	0.000	06:19
Diagnostics Check	OK	06:20

Data Download Date/Time: 06/21/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,MAXWELL,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: JAVOEN

DATE/TIME OF TEST: 05/27/2023 02:36:57

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.080	02:29	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:32	No Sample Provided
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Air Blank	0.000	02:35	
Subject Sample	# 2	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 06/18/2023 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,MAXWELL,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MAHER,W
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/26/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: BUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CANTWELL

DATE/TIME OF TEST: 05/27/2023 21:44:26

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.080	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:40	No Sample Provided
Breath Volume	0.000	21:40	
Air Blank	0.000	21:40	
Air Blank	0.000	21:42	
Subject Sample	# 2	21:43	Subject Test Refused
Breath Volume	0.000	21:43	
Air Blank	0.000	21:43	
Control Test	0.081	21:43	
Air Blank	0.000	21:44	
Diagnostics Check	OK	21:44	

Data Download Date/Time: 06/18/2023 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,MAXWELL,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WHITLEY

DATE/TIME OF TEST: 05/28/2023 05:50:29

Diagnostics Check	OK	05:43
Air Blank	0.000	05:44
Control Test	0.080	05:44
Air Blank	0.000	05:44
Subject Sample # 1	0.093	05:46
Breath Volume	1.753	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.094	05:49
Breath Volume	1.765	05:49
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 06/18/2023 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,MAXWELL,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HEATH,M
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: BUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 05/28/2023 21:34:31

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:26	
Control Test	0.081	21:26	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:30	No Sample Provided
Breath Volume	0.000	21:30	
Air Blank	0.000	21:30	
Air Blank	0.000	21:32	
Subject Sample	# 2	21:33	Subject Test Refused
Breath Volume	0.000	21:33	
Air Blank	0.000	21:33	
Control Test	0.081	21:33	
Air Blank	0.000	21:34	
Diagnostics Check	OK	21:34	

Data Download Date/Time: 06/18/2023 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FOREMAN,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:12

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 05/30/2023 12:46:32

Diagnostics Check	OK	12:38	
Air Blank	0.000	12:39	
Control Test	0.080	12:39	
Air Blank	0.000	12:39	
Subject Sample # 1	0.000	12:42	
Breath Volume	1.277	12:42	
Air Blank	0.000	12:43	
Air Blank	0.000	12:45	
Subject Sample # 2	0.000	12:45	
Breath Volume	1.492	12:45	
Air Blank		12:46	Improper Sample
Air Blank	0.000	12:46	

Data Download Date/Time: 06/18/2023 22:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: CHAPPELL,HANNAH
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 05/28/2023 06:45:44

Diagnostics Check	OK	06:38
Air Blank	0.000	06:39
Control Test	0.081	06:39
Air Blank	0.000	06:40
Subject Sample # 1	0.110	06:40
Breath Volume	2.304	06:40
Air Blank	0.000	06:41
Air Blank	0.000	06:43
Subject Sample # 2	0.112	06:43
Breath Volume	2.968	06:43
Air Blank	0.000	06:44
Control Test	0.079	06:45
Air Blank	0.000	06:45
Diagnostics Check	OK	06:45

Data Download Date/Time: 06/30/2023 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MAGLIANO,ANDRES,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: FRASER,JAMES

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 16:53

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 05/04/2023 17:23:07

Diagnostics Check	OK	17:20	
Air Blank	0.000	17:20	
Control Test	0.082	17:20	
Air Blank	0.000	17:21	
Subject Sample	# 1	17:21	Subject Test Refused
Breath Volume	0.000	17:21	
Air Blank	0.000	17:22	
Control Test	0.081	17:22	
Air Blank	0.000	17:23	
Diagnostics Check	OK	17:23	

Data Download Date/Time: 05/29/2023 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,ANGEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CLARK,CARL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/30/2023 03:23:21

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.155	03:18
Breath Volume	1.726	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.152	03:21
Breath Volume	2.644	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 06/19/2023 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: SIA

DATE/TIME OF TEST: 05/12/2023 02:02:37

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.080	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:01	Interferent Detect
Breath Volume	1.832	02:01	
Air Blank	0.000	02:01	
Control Test	0.080	02:02	
Air Blank		02:02	RFI Detect
Air Blank		02:02	RFI Detect

Data Download Date/Time: 06/01/2023 16:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: SIA

DATE/TIME OF TEST: 05/12/2023 02:07:13

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.080	02:04	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:05	Interferent Detect
Breath Volume	1.644	02:05	
Air Blank	0.000	02:06	
Control Test	0.080	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 06/01/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: SIA

DATE/TIME OF TEST: 05/12/2023 02:14:36

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.144	02:09
Breath Volume	1.402	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.150	02:13
Breath Volume	1.945	02:13
Air Blank	0.000	02:13
Control Test	0.082	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 06/01/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: COUCH

DATE/TIME OF TEST: 05/18/2023 20:56:39

Diagnostics Check	OK	20:55	
Air Blank		20:56	Ambient Fail
Air Blank	0.000	20:56	

Data Download Date/Time: 06/01/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: COUCH

DATE/TIME OF TEST: 05/18/2023 20:59:16

Diagnostics Check	OK	20:58	
Air Blank		20:58	Ambient Fail
Air Blank	0.000	20:59	

Data Download Date/Time: 06/01/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: COUCH

DATE/TIME OF TEST: 05/18/2023 21:01:56

Diagnostics Check	OK	21:00	
Air Blank		21:01	Ambient Fail
Air Blank	0.000	21:01	

Data Download Date/Time: 06/01/2023 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: LUPINACCI

DATE/TIME OF TEST: 05/20/2023 23:10:34

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:07	
Control Test	0.081	23:08	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:09	Interferent Detect
Breath Volume	2.929	23:09	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

Data Download Date/Time: 06/01/2023 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: LUPINACCI

DATE/TIME OF TEST: 05/20/2023 23:13:31

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:12	
Control Test	0.076	23:13	Interferent Detect
Air Blank	0.000	23:13	

Data Download Date/Time: 06/01/2023 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: LUPINACCI

DATE/TIME OF TEST: 05/20/2023 23:21:18

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.115	23:16
Breath Volume	1.718	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.117	23:19
Breath Volume	1.691	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 06/01/2023 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 05/21/2023 02:37:18

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.166	02:32
Breath Volume	3.406	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.171	02:35
Breath Volume	1.320	02:35
Air Blank	0.000	02:36
Control Test	0.083	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 06/01/2023 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: VILLANUEVA

DATE/TIME OF TEST: 05/21/2023 02:43:03

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:42	RFI Detect
Breath Volume	0.000	02:42	
Air Blank		02:42	RFI Detect
Air Blank	0.000	02:43	

Data Download Date/Time: 06/01/2023 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: VILLANUEVA

DATE/TIME OF TEST: 05/21/2023 02:51:06

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.113	02:46
Breath Volume	2.027	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample # 2	0.104	02:49
Breath Volume	1.578	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:51

Data Download Date/Time: 06/01/2023 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: SERAFIN

DATE/TIME OF TEST: 05/31/2023 22:51:53

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:47	
Control Test	0.082	22:47	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:49	Interferent Detect
Breath Volume	2.335	22:49	
Air Blank	0.000	22:50	
Control Test	0.079	22:50	
Air Blank	0.000	22:51	
Diagnostics Check	Fail	22:51	Analytical Stability Test
Air Blank	0.000	22:51	

Data Download Date/Time: 06/01/2023 16:42