

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CASTILLO CRUZ

DATE/TIME OF TEST: 05/20/2023 03:29:18

Diagnostics Check	Fail	03:28	Voltage/Current Test
Air Blank	0.000	03:29	

Data Download Date/Time: 06/12/2023 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CASTILLO CRUZ

DATE/TIME OF TEST: 05/20/2023 03:38:14

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.155	03:33
Breath Volume	1.945	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.135	03:36
Breath Volume	1.746	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 06/12/2023 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/19/2023 22:52:44

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.081	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.186	22:47
Breath Volume	1.949	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:49
Subject Sample # 2	0.172	22:51
Breath Volume	1.929	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 06/12/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: GOMEZ PEREZ

DATE/TIME OF TEST: 05/20/2023 02:21:34

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.081	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:19	Slope Not Met
Breath Volume	2.019	02:19	
Air Blank	0.000	02:20	
Control Test	0.081	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 06/12/2023 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: GOMEZ PEREZ

DATE/TIME OF TEST: 05/20/2023 02:31:24

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.173	02:26
Breath Volume	3.421	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.178	02:29
Breath Volume	3.554	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 06/12/2023 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: STEFANCIK

DATE/TIME OF TEST: 05/20/2023 20:37:34

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.080	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.226	20:28
Breath Volume	2.781	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:30
Subject Sample # 2	0.197	20:31
Breath Volume	1.425	20:31
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 3	0.215	20:35
Breath Volume	2.023	20:35
Air Blank	0.000	20:36
Control Test	0.080	20:36
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 06/12/2023 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: REINA RUIZ

DATE/TIME OF TEST: 05/21/2023 02:14:14

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.081	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.142	02:09
Breath Volume	2.812	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample # 2	0.141	02:12
Breath Volume	3.375	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 06/12/2023 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: MULRAIN

DATE/TIME OF TEST: 05/30/2023 20:07:18

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.082	20:00
Air Blank	0.000	20:01
Subject Sample # 1	0.209	20:02
Breath Volume	1.746	20:02
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.199	20:05
Breath Volume	2.542	20:05
Air Blank	0.000	20:06
Control Test	0.081	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 06/12/2023 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: DAHME

DATE/TIME OF TEST: 05/01/2023 02:40:04

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.170	02:35
Breath Volume	2.277	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.177	02:38
Breath Volume	2.660	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:40

Data Download Date/Time: 05/22/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: FONTANA

DATE/TIME OF TEST: 05/04/2023 22:55:14

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.081	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.200	22:50
Breath Volume	1.667	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.199	22:53
Breath Volume	2.089	22:53
Air Blank	0.000	22:54
Control Test	0.079	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 05/22/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MARSAN,NECOLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: ZUCKERMAN

DATE/TIME OF TEST: 05/08/2023 21:23:20

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.080	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.252	21:18
Breath Volume	2.757	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.233	21:21
Breath Volume	2.585	21:21
Air Blank	0.000	21:22
Control Test	0.078	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 05/22/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MUNOZ,RAFAEL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 05/09/2023 03:18:09

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.000	03:13
Breath Volume	1.546	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.000	03:16
Breath Volume	2.234	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 05/22/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,DARIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARPER,DARIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BERTHA 2

DATE/TIME OF TEST: 05/11/2023 23:39:13

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:35	
Control Test	0.080	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:37	Subject Test Refused
Breath Volume	0.000	23:37	
Air Blank	0.000	23:38	
Control Test	0.080	23:38	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 05/22/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTLEY, BRANDEN, C
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LABRADA, ALEXIS
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 05/12/2023 15:48:59

Diagnostics Check	OK	15:46	
Air Blank	0.000	15:46	
Control Test	0.080	15:46	
Air Blank	0.000	15:47	
Subject Sample	# 1	15:47	Subject Test Refused
Breath Volume	0.000	15:47	
Air Blank	0.000	15:48	
Control Test	0.080	15:48	
Air Blank	0.000	15:48	
Diagnostics Check	OK	15:48	

Data Download Date/Time: 05/22/2023 09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SCHRAGER,O
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ISMAILI

DATE/TIME OF TEST: 05/13/2023 17:13:28

Diagnostics Check	OK	17:10	
Air Blank	0.000	17:11	
Control Test	0.080	17:11	
Air Blank	0.000	17:11	
Subject Sample	# 1	17:12	Subject Test Refused
Breath Volume	0.000	17:12	
Air Blank	0.000	17:12	
Control Test	0.080	17:12	
Air Blank	0.000	17:13	
Diagnostics Check	OK	17:13	

Data Download Date/Time: 05/22/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,C
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: PAULSEN

DATE/TIME OF TEST: 05/14/2023 02:30:25

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.081	02:26
Breath Volume	1.601	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.084	02:28
Breath Volume	2.113	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 05/22/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MARSAN,NECOLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GOLDEN

DATE/TIME OF TEST: 05/14/2023 02:50:17

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.250	02:45
Breath Volume	2.078	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.265	02:48
Breath Volume	1.371	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 05/22/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SCHMDT,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: PRIEST

DATE/TIME OF TEST: 05/14/2023 19:15:17

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.080	19:09
Air Blank	0.000	19:09
Subject Sample # 1	0.154	19:10
Breath Volume	2.058	19:10
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.142	19:13
Breath Volume	1.277	19:13
Air Blank	0.000	19:14
Control Test	0.077	19:14
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 05/22/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,DAVID,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: NGUYEN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: POZEK

DATE/TIME OF TEST: 05/20/2023 00:52:36

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Subject Sample # 1	0.261	00:51	
Breath Volume	1.355	00:51	
Air Blank		00:51	RFI Detect
Air Blank		00:52	RFI Detect
Air Blank	0.000	00:52	

Data Download Date/Time: 05/22/2023 09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,DAVID,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: NGUYEN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: POZEK

DATE/TIME OF TEST: 05/20/2023 01:07:48

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.080	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1 0.251	00:56	
Breath Volume	1.640	00:56	
Air Blank	0.000	00:57	
Air Blank	0.000	00:59	
Subject Sample	# 2 0.231	01:02	Volume Not Met
Breath Volume	0.496	01:02	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample	# 3 0.250	01:06	
Breath Volume	2.039	01:06	
Air Blank	0.000	01:06	
Control Test	0.078	01:07	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 05/22/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,C
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 05/23/2023 00:56:19

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.081	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.173	00:51
Breath Volume	2.687	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.170	00:54
Breath Volume	2.281	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 06/23/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HALLORAN

DATE/TIME OF TEST: 05/27/2023 01:18:43

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.081	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.198	01:14
Breath Volume	3.632	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.206	01:17
Breath Volume	3.507	01:17
Air Blank	0.000	01:17
Control Test	0.080	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 06/23/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI, PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: LOCKHART

DATE/TIME OF TEST: 05/28/2023 01:31:59

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.209	01:27
Breath Volume	2.902	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.214	01:30
Breath Volume	2.761	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 06/23/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,DAVID,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BIFANO,ANTHONY
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 05/31/2023 03:50:06

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.081	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1 0.113	03:48	
Breath Volume	1.496	03:48	
Air Blank		03:48	RFI Detect
Air Blank	0.000	03:49	
Control Test	0.080	03:49	
Air Blank	0.000	03:50	
Diagnostics Check	OK	03:50	

Data Download Date/Time: 06/23/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,DAVID,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BIFANO,ANTHONY
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 05/31/2023 03:59:11

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.104	03:54
Breath Volume	2.214	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.099	03:57
Breath Volume	2.199	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 06/23/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KIANA,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: PEEK

DATE/TIME OF TEST: 05/01/2023 18:47:27

Diagnostics Check	OK	18:44	
Air Blank	0.000	18:45	
Control Test	0.080	18:45	
Air Blank	0.000	18:45	
Subject Sample	# 1	18:46	Subject Test Refused
Breath Volume	0.000	18:46	
Air Blank	0.000	18:46	
Control Test	0.080	18:46	
Air Blank	0.000	18:47	
Diagnostics Check	OK	18:47	

Data Download Date/Time: 05/22/2023 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNE,ASHTON,D
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: GEGEN

DATE/TIME OF TEST: 05/13/2023 20:50:15

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:47	
Control Test	0.079	20:48	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:48	Subject Test Refused
Breath Volume	0.000	20:48	
Air Blank	0.000	20:49	
Control Test	0.079	20:49	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

Data Download Date/Time: 05/22/2023 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: OROPESA

DATE/TIME OF TEST: 05/19/2023 03:32:18

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:29	
Control Test	0.080	03:29	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 05/22/2023 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA RUA,BRYAN,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CALHOUN,KEVIN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: HAJISALEM

DATE/TIME OF TEST: 05/20/2023 18:48:46

Diagnostics Check	OK	18:45	
Air Blank	0.000	18:46	
Control Test	0.080	18:46	
Air Blank	0.000	18:47	
Subject Sample	# 1	18:47	Subject Test Refused
Breath Volume	0.000	18:47	
Air Blank	0.000	18:47	
Control Test	0.080	18:48	
Air Blank	0.000	18:48	
Diagnostics Check	OK	18:48	

Data Download Date/Time: 05/22/2023 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,JAMES,C
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: RODRIGUEZ,JAMES
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 10:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:51

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 05/03/2023 11:30:35

Diagnostics Check	OK	11:27	
Air Blank	0.000	11:28	
Control Test	0.081	11:28	
Air Blank	0.000	11:29	
Subject Sample	# 1	11:29	Subject Test Refused
Breath Volume	0.000	11:29	
Air Blank	0.000	11:29	
Control Test	0.080	11:29	
Air Blank	0.000	11:30	
Diagnostics Check	OK	11:30	

Data Download Date/Time: 05/31/2023 23:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 05/12/2023 01:15:43

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.169	01:09
Breath Volume	2.304	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.184	01:14
Breath Volume	1.253	01:14
Air Blank	0.000	01:14
Control Test	0.081	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 05/31/2023 23:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: AYALA MARTIN

DATE/TIME OF TEST: 05/14/2023 04:28:23

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.082	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.067	04:24
Breath Volume	2.339	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.065	04:26
Breath Volume	2.871	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 05/31/2023 23:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALCON, IVAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BERNABE, DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 05/20/2023 14:40:22

Diagnostics Check	OK	14:34	
Air Blank	0.000	14:34	
Control Test	0.081	14:35	
Air Blank	0.000	14:35	
Subject Sample	# 1 0.175	14:35	
Breath Volume	2.804	14:35	
Air Blank	0.000	14:36	
Air Blank	0.000	14:38	
Subject Sample	# 2	14:38	Slope Not Met
Breath Volume	1.117	14:38	
Air Blank	0.000	14:39	
Control Test	0.081	14:39	
Air Blank	0.000	14:40	
Diagnostics Check	OK	14:40	

Data Download Date/Time: 05/31/2023 23:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALCON, IVAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BERNABE, DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 05/20/2023 14:52:14

Diagnostics Check	OK	14:45
Air Blank	0.000	14:45
Control Test	0.081	14:46
Air Blank	0.000	14:46
Subject Sample # 1	0.168	14:47
Breath Volume	2.570	14:47
Air Blank	0.000	14:48
Air Blank	0.000	14:50
Subject Sample # 2	0.169	14:50
Breath Volume	1.792	14:50
Air Blank	0.000	14:51
Control Test	0.081	14:51
Air Blank	0.000	14:52
Diagnostics Check	OK	14:52

Data Download Date/Time: 05/31/2023 23:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: VINCENT,CARL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: JULIEN

DATE/TIME OF TEST: 05/21/2023 00:55:35

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.270	00:50
Breath Volume	2.363	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:52
Subject Sample # 2	0.265	00:53
Breath Volume	2.355	00:53
Air Blank	0.000	00:54
Control Test	0.077	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 05/31/2023 23:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLORZANO,DYLAN,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SACKMANN,ANDREW
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: PEREZ ARTIAGA

DATE/TIME OF TEST: 05/21/2023 17:41:48

Diagnostics Check	OK	17:30
Air Blank	0.000	17:30
Control Test	0.081	17:30
Air Blank	0.000	17:31
Subject Sample # 1	0.222	17:34
Breath Volume	1.636	17:34
Air Blank	0.000	17:34
Air Blank	0.000	17:36
Subject Sample # 2	0.245	17:36
Breath Volume	1.914	17:36
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 3	0.238	17:40
Breath Volume	1.484	17:40
Air Blank	0.000	17:40
Control Test	0.080	17:41
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

Data Download Date/Time: 05/31/2023 23:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WROY,WALNER,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GARCIA,MANUEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 05/24/2023 20:58:52

Diagnostics Check	OK	20:51
Air Blank	0.000	20:52
Control Test	0.081	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.148	20:54
Breath Volume	1.906	20:54
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.149	20:57
Breath Volume	2.261	20:57
Air Blank	0.000	20:57
Control Test	0.078	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 05/31/2023 23:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: SALAZAR ESCORICA

DATE/TIME OF TEST: 05/27/2023 05:05:59

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:56	
Control Test	0.081	04:57	
Air Blank	0.000	04:57	
Subject Sample	# 1 0.190	04:58	
Breath Volume	2.269	04:58	
Air Blank	0.000	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 2 0.163	05:02	
Breath Volume	1.464	05:02	
Air Blank	0.000	05:03	
Air Blank	0.000	05:04	
Subject Sample	# 3 0.167	05:05	
Breath Volume	1.632	05:05	
Air Blank		05:05	RFI Detect
Air Blank	0.000	05:05	

Data Download Date/Time: 05/31/2023 23:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POST,JACOB,S
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: POST,JACOB
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: CHACON PORTILLO

DATE/TIME OF TEST: 05/27/2023 21:09:08

Diagnostics Check	OK	20:59
Air Blank	0.000	21:00
Control Test	0.081	21:00
Air Blank	0.000	21:00
Subject Sample # 1	0.195	21:01
Breath Volume	2.324	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.161	21:04
Breath Volume	3.550	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 3	0.184	21:07
Breath Volume	1.906	21:07
Air Blank	0.000	21:08
Control Test	0.081	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 05/31/2023 23:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIVERA,VICTOR,J

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: RIVERA,VICTOR

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 04/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402231696

DRY GAS EXPIRY DATE: 10/07/2024

VIOLATION CODE: DUI

ARREST TIME: 21:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/03/2023 21:39:06

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.081	21:35	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:37	Subject Test Refused
Breath Volume	0.000	21:37	
Air Blank	0.000	21:38	
Control Test	0.082	21:38	
Air Blank	0.000	21:39	
Diagnostics Check	OK	21:39	

Data Download Date/Time: 05/30/2023 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN BAPTISTE,ALBENS,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402231696
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: BORGE MEJIA

DATE/TIME OF TEST: 05/06/2023 00:21:38

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.082	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.165	00:17
Breath Volume	2.394	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.164	00:19
Breath Volume	2.445	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 05/30/2023 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN BAPTISTE,ALBENS,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402231696
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: LUSAMBO

DATE/TIME OF TEST: 05/13/2023 03:09:27

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.210	03:02
Breath Volume	1.386	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.213	03:07
Breath Volume	1.328	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 05/30/2023 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: Yes
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402231696
BREATH TEST OPERATOR: TAVERAS, TYLER, J	DRY GAS EXPIRY DATE: 10/07/2024
BREATH TEST OPERATOR AGENCY: FHP Troop D	VIOLATION CODE: DUI
ARRESTING OFFICER: TAVERAS, TYLER	ARREST TIME: 02:49
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Male
LAST AGENCY INSPECTION: 04/22/2023	OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ELYSEE	DATE/TIME OF TEST: 05/14/2023 03:41:26
-----------------------------	--

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.177	03:36
Breath Volume	2.675	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.185	03:39
Breath Volume	2.234	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 05/30/2023 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402231696
BREATH TEST OPERATOR: RIVERA,VICTOR,J	DRY GAS EXPIRY DATE: 10/07/2024
BREATH TEST OPERATOR AGENCY: FHP Troop D	VIOLATION CODE: DUI
ARRESTING OFFICER: SANTOS,JOSEPH	ARREST TIME: 01:46
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Male
LAST AGENCY INSPECTION: 04/22/2023	OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: GALARZA	DATE/TIME OF TEST: 05/20/2023 02:15:21
------------------------------	--

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.172	02:10
Breath Volume	3.460	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.170	02:13
Breath Volume	3.113	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 05/30/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN BAPTISTE,ALBENS,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402231696
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: PEREZ FLORES

DATE/TIME OF TEST: 05/27/2023 02:06:55

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.083	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.170	02:02
Breath Volume	3.343	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:04
Subject Sample # 2	0.173	02:05
Breath Volume	3.257	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 05/30/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SADOWSKI,ERICKA,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SADOWSKI,ERICKA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402231696
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/27/2023 02:54:29

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.080	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1 0.204	02:49	
Breath Volume	2.636	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 2 0.194	02:52	
Breath Volume	2.421	02:52	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank		02:54	RFI Detect
Air Blank	0.000	02:54	

Data Download Date/Time: 05/30/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SADOWSKI,ERICKA,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SADOWSKI,ERICKA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402231696
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/27/2023 03:04:22

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.198	02:59
Breath Volume	1.792	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.200	03:02
Breath Volume	3.050	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 05/30/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: COATES

DATE/TIME OF TEST: 05/05/2023 23:31:42

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.080	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1 0.179	23:27	
Breath Volume	3.753	23:27	
Air Blank	0.000	23:28	
Air Blank	0.000	23:30	
Subject Sample	# 2	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:30	
Control Test	0.079	23:31	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 05/15/2023 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: SILVA RIOS

DATE/TIME OF TEST: 05/07/2023 02:01:05

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.122	01:56
Breath Volume	1.875	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.114	01:59
Breath Volume	2.082	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 05/15/2023 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/07/2023 04:48:05

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.131	04:43
Breath Volume	1.925	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:45
Subject Sample # 2	0.125	04:46
Breath Volume	1.625	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 05/15/2023 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GONZALEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: KINLAW

DATE/TIME OF TEST: 05/06/2023 01:00:26

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.157	00:55
Breath Volume	2.554	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample # 2	0.164	00:58
Breath Volume	1.597	00:58
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 05/29/2023 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GONZALEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 05/21/2023 01:39:55

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.158	01:35
Breath Volume	1.679	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.152	01:38
Breath Volume	2.218	01:38
Air Blank	0.000	01:38
Control Test	0.078	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 05/29/2023 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GONZALEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WHITTINGTON

DATE/TIME OF TEST: 05/27/2023 03:57:44

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.122	03:53
Breath Volume	2.062	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.121	03:56
Breath Volume	2.179	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 05/29/2023 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GONZALEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: NOFSINGER JR

DATE/TIME OF TEST: 05/28/2023 03:41:03

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.116	03:36
Breath Volume	2.394	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.116	03:39
Breath Volume	2.941	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:41

Data Download Date/Time: 05/29/2023 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: LOGAN,ZACHARY,S	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: LOGAN,ZACHARY	ARREST TIME: 21:39
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Female
LAST AGENCY INSPECTION: 05/29/2023	OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BAILEY	DATE/TIME OF TEST: 05/30/2023	22:07:20
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.117	22:02
Breath Volume	2.164	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.113	22:05
Breath Volume	1.851	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 06/22/2023 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 05/20/2023 03:28:04

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.207	03:23
Breath Volume	2.093	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.203	03:26
Breath Volume	2.644	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:28

Data Download Date/Time: 05/31/2023 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOBOSIEWICZ,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/05/2023 01:51:55

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.080	01:45	
Air Blank	0.000	01:45	
Subject Sample	# 1 0.170	01:47	
Breath Volume	2.363	01:47	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 2	01:50	Interferent Detect
Breath Volume	1.847	01:50	
Air Blank	0.000	01:50	
Control Test	0.082	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 05/29/2023 03:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOBOSIEWICZ,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/05/2023 02:14:14

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.081	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1 0.177	02:12	
Breath Volume	1.640	02:12	
Air Blank		02:13	Purge Fail
Air Blank	0.000	02:14	

Data Download Date/Time: 05/29/2023 03:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 05/06/2023 01:56:09

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:53	
Control Test	0.081	01:53	
Air Blank	0.000	01:54	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:55	
Control Test	0.080	01:55	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 05/29/2023 03:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/09/2023 01:07:10

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.018	01:02
Breath Volume	1.589	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:04
Subject Sample # 2	0.019	01:05
Breath Volume	2.523	01:05
Air Blank	0.000	01:06
Control Test	0.081	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 05/29/2023 03:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: MONTGOMERY,KENNETH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24229080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/23/2023 02:03:17

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.113	01:58
Breath Volume	1.812	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.108	02:01
Breath Volume	1.367	02:01
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 06/01/2023 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: MONTGOMERY, KENNETH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24229080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 05/27/2023 00:39:25

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.082	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.095	00:34
Breath Volume	2.648	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.096	00:37
Breath Volume	2.984	00:37
Air Blank	0.000	00:38
Control Test	0.081	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 06/01/2023 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: ENGLAND

DATE/TIME OF TEST: 05/27/2023 22:37:27

Diagnostics Check	OK	22:36	
Air Blank		22:36	RFI Detect
Air Blank	0.000	22:37	

Data Download Date/Time: 06/01/2023 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: MONTGOMERY,KENNETH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24229080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: ENGLAND

DATE/TIME OF TEST: 05/27/2023 22:46:28

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.083	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.124	22:42
Breath Volume	3.203	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.118	22:44
Breath Volume	1.140	22:44
Air Blank	0.000	22:45
Control Test	0.082	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 06/01/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BERHAMAJ

DATE/TIME OF TEST: 05/28/2023 01:45:03

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.082	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.033	01:40
Breath Volume	1.617	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.032	01:43
Breath Volume	1.742	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:45

Data Download Date/Time: 06/01/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONIE, J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/28/2023 22:13:26

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:10	
Control Test	0.081	22:11	
Air Blank	0.000	22:11	
Subject Sample	# 1	22:12	Subject Test Refused
Breath Volume	0.000	22:12	
Air Blank	0.000	22:12	
Control Test	0.080	22:12	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

Data Download Date/Time: 06/01/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY, KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: ABJORNSON

DATE/TIME OF TEST: 05/28/2023 23:56:09

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.082	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.087	23:51
Breath Volume	3.906	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:53
Subject Sample # 2	0.091	23:54
Breath Volume	4.082	23:54
Air Blank	0.000	23:55
Control Test	0.082	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 06/01/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: RICHWINE,C
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 05/29/2023 01:12:03

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.082	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.142	01:07
Breath Volume	1.394	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.137	01:10
Breath Volume	1.597	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:12

Data Download Date/Time: 06/01/2023 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/27/2023 03:46:57

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.079	03:42	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:46	
Control Test	0.079	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 06/08/2023 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WARE ROGERS,BRANDON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: ABREU

DATE/TIME OF TEST: 05/05/2023 03:13:50

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:02	
Control Test	0.079	03:03	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:06	No Sample Provided
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
Subject Sample	# 2	03:12	No Sample Provided
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Control Test	0.079	03:13	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: CHUNXIA

DATE/TIME OF TEST: 05/06/2023 02:48:34

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.124	02:44
Breath Volume	1.984	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.123	02:47
Breath Volume	2.273	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: THORNBRUGH,E
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: LORENZO

DATE/TIME OF TEST: 05/06/2023 20:50:28

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:39	
Control Test	0.079	20:39	
Air Blank	0.000	20:40	
Subject Sample	# 1 0.065	20:43	Volume Not Met
Breath Volume	0.636	20:43	
Air Blank	0.000	20:43	
Air Blank	0.000	20:45	
Subject Sample	# 2 0.085	20:48	Volume Not Met
Breath Volume	0.062	20:48	
Air Blank	0.000	20:49	
Control Test	0.078	20:49	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: THORNBRUGH,E
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: LORENZO

DATE/TIME OF TEST: 05/06/2023 21:00:24

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.077	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.078	20:54
Breath Volume	2.253	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.075	20:58
Breath Volume	1.718	20:58
Air Blank	0.000	20:59
Control Test	0.078	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KETCHAM,JENNIFER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: FENNER

DATE/TIME OF TEST: 05/07/2023 01:54:46

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.304	01:50
Breath Volume	2.359	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.302	01:53
Breath Volume	2.437	01:53
Air Blank	0.000	01:53
Control Test	0.078	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODDY,AARON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RODDY,AARON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: BURR GIANGRANDI

DATE/TIME OF TEST: 05/11/2023 20:25:24

Diagnostics Check	OK	20:22	
Air Blank	0.000	20:23	
Control Test	0.079	20:23	
Air Blank	0.000	20:23	
Subject Sample	# 1	20:24	Subject Test Refused
Breath Volume	0.000	20:24	
Air Blank	0.000	20:24	
Control Test	0.079	20:24	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:25	

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRICKE,KENNETH,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CURRY,CONNOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: BRODHAGEN

DATE/TIME OF TEST: 05/13/2023 03:19:30

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.186	03:14
Breath Volume	2.945	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.181	03:17
Breath Volume	3.074	03:17
Air Blank	0.000	03:18
Control Test	0.078	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: IBARRA ROJAS

DATE/TIME OF TEST: 05/14/2023 04:24:25

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.193	04:19
Breath Volume	3.675	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.185	04:22
Breath Volume	2.773	04:22
Air Blank	0.000	04:23
Control Test	0.078	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA',C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MOORE,FRANCIS
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/14/2023 20:16:23

Diagnostics Check	OK	20:13	
Air Blank	0.000	20:13	
Control Test	0.079	20:14	
Air Blank	0.000	20:14	
Subject Sample	# 1	20:15	Subject Test Refused
Breath Volume	0.000	20:15	
Air Blank	0.000	20:15	
Control Test	0.079	20:15	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:16	

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: NDIAYE

DATE/TIME OF TEST: 05/15/2023 21:00:31

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:52	
Control Test	0.079	20:52	
Air Blank	0.000	20:53	
Subject Sample	# 1	20:56	No Sample Provided
Breath Volume	0.000	20:56	
Air Blank	0.000	20:56	
Air Blank	0.000	20:58	
Subject Sample	# 2	20:59	Subject Test Refused
Breath Volume	0.000	20:59	
Air Blank	0.000	20:59	
Control Test	0.079	20:59	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:00	

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODDY,AARON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RODDY,AARON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 13:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:06

SUBJECT'S LAST NAME: IDLOUGSSIYR

DATE/TIME OF TEST: 05/16/2023 14:33:04

Diagnostics Check	OK	14:27
Air Blank	0.000	14:27
Control Test	0.079	14:27
Air Blank	0.000	14:28
Subject Sample # 1	0.067	14:28
Breath Volume	1.582	14:28
Air Blank	0.000	14:29
Air Blank	0.000	14:31
Subject Sample # 2	0.063	14:31
Breath Volume	1.339	14:31
Air Blank	0.000	14:32
Control Test	0.079	14:32
Air Blank	0.000	14:32
Diagnostics Check	OK	14:33

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ROJAS BURGOS,JESUS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BONTO

DATE/TIME OF TEST: 05/17/2023 02:33:38

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:26	No Sample Provided
Breath Volume	0.000	02:26	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample	# 2	02:32	No Sample Provided
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Control Test	0.079	02:33	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BLANTON,TYLER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: SCHULTE

DATE/TIME OF TEST: 05/17/2023 02:52:21

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.197	02:47
Breath Volume	3.789	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.205	02:50
Breath Volume	3.871	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: JAMISON

DATE/TIME OF TEST: 05/18/2023 17:31:14

Diagnostics Check	OK	17:28	
Air Blank	0.000	17:29	
Control Test	0.080	17:29	
Air Blank	0.000	17:30	
Subject Sample	# 1	17:30	Subject Test Refused
Breath Volume	0.000	17:30	
Air Blank		17:30	RFI Detect
Air Blank	0.000	17:31	

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BARTH

DATE/TIME OF TEST: 05/20/2023 03:57:30

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:55	
Control Test	0.080	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.079	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANI,JAYESH,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: TELLIER,JEREMEY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:14

SUBJECT'S LAST NAME: MANI

DATE/TIME OF TEST: 05/20/2023 07:52:36

Diagnostics Check	OK	07:45
Air Blank	0.000	07:46
Control Test	0.078	07:46
Air Blank	0.000	07:47
Subject Sample # 1	0.142	07:48
Breath Volume	1.472	07:48
Air Blank	0.000	07:48
Air Blank	0.000	07:50
Subject Sample # 2	0.147	07:51
Breath Volume	1.488	07:51
Air Blank	0.000	07:51
Control Test	0.077	07:52
Air Blank	0.000	07:52
Diagnostics Check	OK	07:52

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/20/2023 19:17:45

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.080	19:09
Air Blank	0.000	19:10
Subject Sample # 1	0.158	19:12
Breath Volume	1.757	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:14
Subject Sample # 2	0.167	19:16
Breath Volume	1.457	19:16
Air Blank	0.000	19:16
Control Test	0.078	19:17
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ZITO,ANTHONY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: ROSARIO MONSANTO

DATE/TIME OF TEST: 05/22/2023 00:32:17

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.079	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1 0.126	00:25	Volume Not Met
Breath Volume	0.218	00:25	
Air Blank	0.000	00:26	
Air Blank	0.000	00:28	
Subject Sample	# 2	00:30	Subject Test Refused
Breath Volume	0.085	00:30	
Air Blank	0.000	00:31	
Control Test	0.078	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPIRITO,GEORGE,W
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SPIRITO,GEORGE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 05/23/2023 16:10:45

Diagnostics Check	OK	16:04
Air Blank	0.000	16:04
Control Test	0.078	16:05
Air Blank	0.000	16:05
Subject Sample # 1	0.050	16:06
Breath Volume	2.066	16:06
Air Blank	0.000	16:06
Air Blank	0.000	16:08
Subject Sample # 2	0.051	16:09
Breath Volume	2.246	16:09
Air Blank	0.000	16:09
Control Test	0.079	16:10
Air Blank	0.000	16:10
Diagnostics Check	OK	16:10

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: THORNBRUGH,E
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: JANOWITZ

DATE/TIME OF TEST: 05/25/2023 01:13:33

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.079	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,BRIAN,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: COBB,AUSTIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: BUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: O'BRYAN

DATE/TIME OF TEST: 05/26/2023 23:35:55

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:34	
Control Test	0.079	23:35	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,AUSTIN,C
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARPENTER,BENJAMIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: BUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: DEHANAS

DATE/TIME OF TEST: 05/27/2023 20:54:37

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:51	
Control Test	0.079	20:51	
Air Blank	0.000	20:52	
Subject Sample	# 1	20:53	Subject Test Refused
Breath Volume	0.000	20:53	
Air Blank	0.000	20:53	
Control Test	0.079	20:54	
Air Blank	0.000	20:54	
Diagnostics Check	OK	20:54	

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/27/2023 22:05:13

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.071	22:00
Breath Volume	1.207	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:02
Subject Sample # 2	0.074	22:03
Breath Volume	2.003	22:03
Air Blank	0.000	22:04
Control Test	0.078	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 06/08/2023 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,AUSTIN,C
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SAPP,BRIAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: BUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/28/2023 20:04:55

Diagnostics Check	OK	19:59	
Air Blank	0.000	20:00	
Control Test	0.080	20:00	
Air Blank	0.000	20:01	
Subject Sample	# 1	20:03	Subject Test Refused
Breath Volume	0.000	20:03	
Air Blank	0.000	20:04	
Control Test	0.080	20:04	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

Data Download Date/Time: 06/08/2023 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 05/29/2023 22:33:17

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:22	
Control Test	0.079	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1	22:26	No Sample Provided
Breath Volume	0.000	22:26	
Air Blank	0.000	22:26	
Air Blank	0.000	22:28	
Subject Sample	# 2	22:31	No Sample Provided
Breath Volume	0.000	22:31	
Air Blank	0.000	22:32	
Control Test	0.080	22:32	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 06/08/2023 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: RAYBURN

DATE/TIME OF TEST: 05/18/2023 22:44:40

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.079	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.000	22:40
Breath Volume	3.019	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.000	22:43
Breath Volume	3.347	22:43
Air Blank	0.000	22:43
Control Test	0.078	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 06/12/2023 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BIBEY

DATE/TIME OF TEST: 05/22/2023 02:48:24

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.319	02:42
Breath Volume	1.328	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.316	02:46
Breath Volume	1.890	02:46
Air Blank	0.000	02:47
Control Test	0.077	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 06/12/2023 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: APPEGATE

DATE/TIME OF TEST: 05/24/2023 20:32:42

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.080	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.199	20:27
Breath Volume	2.097	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:29
Subject Sample # 2	0.199	20:31
Breath Volume	1.679	20:31
Air Blank	0.000	20:31
Control Test	0.078	20:32
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 06/12/2023 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 05/26/2023 01:29:19

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.080	01:27	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Control Test	0.080	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 06/12/2023 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 05/27/2023 02:33:58

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.198	02:29
Breath Volume	2.816	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.198	02:32
Breath Volume	2.882	02:32
Air Blank	0.000	02:33
Control Test	0.078	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 06/12/2023 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: KEEL

DATE/TIME OF TEST: 05/29/2023 19:12:59

Diagnostics Check	OK	19:02
Air Blank	0.000	19:03
Control Test	0.080	19:03
Air Blank	0.000	19:04
Subject Sample # 1	0.209	19:04
Breath Volume	1.242	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.173	19:08
Breath Volume	1.218	19:08
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 3	0.188	19:11
Breath Volume	1.234	19:11
Air Blank	0.000	19:12
Control Test	0.079	19:12
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 06/12/2023 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: JONES,KAVIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: ROBERT

DATE/TIME OF TEST: 05/02/2023 01:19:31

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:09	
Control Test	0.079	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1 0.143	01:13	Volume Not Met
Breath Volume	0.183	01:13	
Air Blank	0.000	01:14	
Air Blank	0.000	01:16	
Subject Sample	# 2	01:17	Subject Test Refused
Breath Volume	0.015	01:17	
Air Blank	0.000	01:18	
Control Test	0.077	01:18	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 05/10/2023 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MELTON,EMILY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 05/05/2023 15:17:20

Diagnostics Check	OK	15:10
Air Blank	0.000	15:11
Control Test	0.080	15:11
Air Blank	0.000	15:12
Subject Sample # 1	0.223	15:12
Breath Volume	1.617	15:12
Air Blank	0.000	15:13
Air Blank	0.000	15:14
Subject Sample # 2	0.235	15:15
Breath Volume	1.570	15:15
Air Blank	0.000	15:16
Control Test	0.078	15:16
Air Blank	0.000	15:17
Diagnostics Check	OK	15:17

Data Download Date/Time: 05/10/2023 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 01:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: BUIE

DATE/TIME OF TEST: 05/20/2023 01:38:46

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.083	01:33
Breath Volume	1.871	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.085	01:37
Breath Volume	2.308	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 06/01/2023 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: FARLEY,JOSEPH,E	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: FARLEY,JOSEPH	ARREST TIME: 23:53
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 05/17/2023	OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: WATSON DATE/TIME OF TEST: 05/25/2023 00:08:54

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.081	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.151	00:04
Breath Volume	1.300	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:06
Subject Sample # 2	0.142	00:07
Breath Volume	1.273	00:07
Air Blank	0.000	00:07
Control Test	0.079	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 06/01/2023 13:12

Subject Test Electronic Data

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: WEIS

DATE/TIME OF TEST: 05/27/2023 02:29:47

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.082	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.049	02:25
Breath Volume	1.679	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.047	02:28
Breath Volume	1.699	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 06/01/2023 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: FARLEY,JOSEPH,E	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: FARLEY,JOSEPH	ARREST TIME: 00:38
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Female
LAST AGENCY INSPECTION: 05/17/2023	OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MENDEZ DATE/TIME OF TEST: 05/28/2023 00:51:19

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.063	00:46
Breath Volume	2.273	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.062	00:49
Breath Volume	1.976	00:49
Air Blank	0.000	00:50
Control Test	0.081	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 06/01/2023 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: HUEBNER

DATE/TIME OF TEST: 05/28/2023 02:35:00

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.145	02:30
Breath Volume	2.957	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:32
Subject Sample # 2	0.136	02:33
Breath Volume	2.304	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:35

Data Download Date/Time: 06/01/2023 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALLEN,JUSTIN,R

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: WALLEN,JUSTIN

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1355293

DRY GAS EXPIRY DATE: 07/29/2023

VIOLATION CODE: DUI

ARREST TIME: 21:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/20/2023 21:35:00

Diagnostics Check	OK	21:33	
Air Blank	0.000	21:34	
Control Test	0.074	21:34	Control Outside Tolerance
Air Blank	0.000	21:35	

Data Download Date/Time: 06/14/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PEREZ,D
ARRESTING OFFICER AGENCY: KEY BISCAYNE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: DELCID

DATE/TIME OF TEST: 05/13/2023 00:58:14

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.139	00:53
Breath Volume	1.835	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.151	00:56
Breath Volume	1.398	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 06/02/2023 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PEGUERO,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: TRINIDAD RAMIREZ

DATE/TIME OF TEST: 05/19/2023 03:59:07

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.000	03:54
Breath Volume	1.429	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.000	03:57
Breath Volume	1.921	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 06/02/2023 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CONCEPCION

DATE/TIME OF TEST: 05/20/2023 03:41:32

Diagnostics Check	Fail	03:41	Internal Printer Test
Air Blank	0.000	03:41	

Data Download Date/Time: 06/02/2023 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CONCEPCION

DATE/TIME OF TEST: 05/20/2023 03:47:41

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:45	
Control Test	0.079	03:45	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Control Test	0.078	03:47	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 06/02/2023 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: IRVINE,C
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: PAREJA

DATE/TIME OF TEST: 05/27/2023 02:11:05

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.080	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 06/02/2023 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: IRVINE,C
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: PAREJA

DATE/TIME OF TEST: 05/27/2023 02:23:22

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.162	02:18
Breath Volume	1.507	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.161	02:21
Breath Volume	1.691	02:21
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 06/02/2023 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: HERMES

DATE/TIME OF TEST: 05/03/2023 17:12:44

Diagnostics Check	OK	17:02
Air Blank	0.000	17:03
Control Test	0.080	17:03
Air Blank	0.000	17:03
Subject Sample # 1	0.366	17:05
Breath Volume	1.597	17:05
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample # 2	0.388	17:08
Breath Volume	1.804	17:08
Air Blank	0.000	17:08
Air Blank	0.000	17:10
Subject Sample # 3	0.351	17:11
Breath Volume	1.125	17:11
Air Blank	0.000	17:11
Control Test	0.078	17:12
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

Data Download Date/Time: 05/23/2023 23:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: WOODYD

DATE/TIME OF TEST: 05/03/2023 19:35:05

Diagnostics Check	OK	19:25
Air Blank	0.000	19:26
Control Test	0.080	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.237	19:29
Breath Volume	1.195	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.238	19:33
Breath Volume	1.484	19:33
Air Blank	0.000	19:34
Control Test	0.079	19:34
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 05/23/2023 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 05/05/2023 03:08:07

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.173	03:03
Breath Volume	2.718	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:05
Subject Sample # 2	0.159	03:06
Breath Volume	3.125	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 05/23/2023 23:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 05/10/2023 00:44:30

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.120	00:39
Breath Volume	1.449	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.129	00:42
Breath Volume	2.761	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 05/23/2023 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ENGEL,ANTHONY
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/05/2023 19:28:20

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.080	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.153	19:23
Breath Volume	2.699	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.156	19:26
Breath Volume	1.910	19:26
Air Blank	0.000	19:27
Control Test	0.080	19:27
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 06/07/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: LOPEZ FABIAN

DATE/TIME OF TEST: 05/06/2023 04:26:53

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.144	04:22
Breath Volume	1.796	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.156	04:25
Breath Volume	1.652	04:25
Air Blank	0.000	04:25
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 06/07/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: TOBAR

DATE/TIME OF TEST: 05/12/2023 18:23:17

Diagnostics Check	OK	18:17
Air Blank	0.000	18:17
Control Test	0.080	18:17
Air Blank	0.000	18:18
Subject Sample # 1	0.030	18:18
Breath Volume	2.406	18:18
Air Blank	0.000	18:19
Air Blank	0.000	18:21
Subject Sample # 2	0.029	18:21
Breath Volume	2.597	18:21
Air Blank	0.000	18:22
Control Test	0.081	18:22
Air Blank	0.000	18:23
Diagnostics Check	OK	18:23

Data Download Date/Time: 06/07/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RICHARDSON,ALEXANDER
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 05/13/2023 01:49:25

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.186	01:44
Breath Volume	2.085	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.185	01:47
Breath Volume	2.269	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 06/07/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BENINCASA,VINCENT
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: ARMBAND

DATE/TIME OF TEST: 05/13/2023 04:56:51

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.189	04:52
Breath Volume	2.074	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.184	04:55
Breath Volume	2.960	04:55
Air Blank	0.000	04:55
Control Test	0.078	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 06/07/2023 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MENDOZA,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 10:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: OLIVEIRA

DATE/TIME OF TEST: 05/13/2023 13:09:07

Diagnostics Check	OK	12:55	
Air Blank	0.000	12:56	
Control Test	0.080	12:56	
Air Blank	0.000	12:57	
Subject Sample	# 1 0.120	13:00	Volume Not Met
Breath Volume	0.097	13:00	
Air Blank	0.000	13:01	
Air Blank	0.000	13:02	
Subject Sample	# 2 0.168	13:04	No .020 agreement
Breath Volume	1.156	13:04	
Air Blank	0.000	13:04	
Air Blank	0.000	13:06	
Subject Sample	# 3 0.196	13:07	No .020 agreement
Breath Volume	1.554	13:07	
Air Blank	0.000	13:08	
Control Test	0.080	13:08	
Air Blank	0.000	13:09	
Diagnostics Check	OK	13:09	

Data Download Date/Time: 06/07/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MENDOZA,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 10:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: OLIVEIRA

DATE/TIME OF TEST: 05/13/2023 13:18:46

Diagnostics Check	OK	13:11
Air Blank	0.000	13:12
Control Test	0.079	13:12
Air Blank	0.000	13:13
Subject Sample # 1	0.167	13:14
Breath Volume	1.386	13:14
Air Blank	0.000	13:14
Air Blank	0.000	13:16
Subject Sample # 2	0.180	13:17
Breath Volume	1.800	13:17
Air Blank	0.000	13:17
Control Test	0.080	13:18
Air Blank	0.000	13:18
Diagnostics Check	OK	13:18

Data Download Date/Time: 06/07/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FANELLI,MICHELE
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 11:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 05/13/2023 14:33:01

Diagnostics Check	OK	14:26
Air Blank	0.000	14:27
Control Test	0.080	14:27
Air Blank	0.000	14:27
Subject Sample # 1	0.249	14:28
Breath Volume	3.179	14:28
Air Blank	0.000	14:29
Air Blank	0.000	14:30
Subject Sample # 2	0.237	14:31
Breath Volume	2.640	14:31
Air Blank	0.000	14:32
Control Test	0.080	14:32
Air Blank	0.000	14:32
Diagnostics Check	OK	14:33

Data Download Date/Time: 06/07/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NUNEZ,J
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: BARRON

DATE/TIME OF TEST: 05/14/2023 05:52:49

Diagnostics Check	OK	05:45
Air Blank	0.000	05:46
Control Test	0.080	05:46
Air Blank	0.000	05:47
Subject Sample # 1	0.145	05:48
Breath Volume	1.453	05:48
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.140	05:51
Breath Volume	1.472	05:51
Air Blank	0.000	05:51
Control Test	0.080	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 06/07/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOLDBERG,JOSHUA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: AVILES ACOSTA

DATE/TIME OF TEST: 05/15/2023 08:07:57

Diagnostics Check	OK	08:01
Air Blank	0.000	08:01
Control Test	0.080	08:02
Air Blank	0.000	08:02
Subject Sample # 1	0.000	08:03
Breath Volume	3.367	08:03
Air Blank	0.000	08:04
Air Blank	0.000	08:06
Subject Sample # 2	0.000	08:06
Breath Volume	1.593	08:06
Air Blank	0.000	08:07
Control Test	0.080	08:07
Air Blank	0.000	08:07
Diagnostics Check	OK	08:07

Data Download Date/Time: 06/07/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: STULL-HINKLE

DATE/TIME OF TEST: 05/15/2023 23:02:14

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.247	22:56
Breath Volume	1.289	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.247	23:00
Breath Volume	1.171	23:00
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 06/07/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 05/16/2023 05:25:08

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:14	
Control Test	0.080	05:14	
Air Blank	0.000	05:15	
Subject Sample	# 1 0.000	05:18	Volume Not Met
Breath Volume	0.183	05:18	
Air Blank	0.000	05:18	
Air Blank	0.000	05:20	
Subject Sample	# 2 0.000	05:23	Volume Not Met
Breath Volume	0.156	05:23	
Air Blank	0.000	05:24	
Control Test	0.081	05:24	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 05/16/2023 05:35:29

Diagnostics Check	OK	05:27	
Air Blank	0.000	05:27	
Control Test	0.079	05:27	
Air Blank	0.000	05:28	
Subject Sample	# 1 0.000	05:31	Volume Not Met
Breath Volume	0.132	05:31	
Air Blank	0.000	05:31	
Air Blank	0.000	05:33	
Subject Sample	# 2	05:34	Subject Test Refused
Breath Volume	0.000	05:34	
Air Blank	0.000	05:34	
Control Test	0.078	05:34	
Air Blank	0.000	05:35	
Diagnostics Check	OK	05:35	

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BRAGG,R
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/21/2023 04:28:04

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.210	04:22
Breath Volume	1.687	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.215	04:26
Breath Volume	2.000	04:26
Air Blank	0.000	04:27
Control Test	0.078	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:28

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALTUNAY,MOHAMMET
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: TORRES QUINTANA

DATE/TIME OF TEST: 05/21/2023 05:30:32

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:15	
Control Test	0.079	05:15	
Air Blank	0.000	05:16	
Subject Sample	# 1 0.198	05:19	Volume Not Met
Breath Volume	0.140	05:19	
Air Blank	0.000	05:20	
Air Blank	0.000	05:21	
Subject Sample	# 2 0.188	05:23	No .020 agreement
Breath Volume	1.343	05:23	
Air Blank	0.000	05:23	
Air Blank	0.000	05:25	
Subject Sample	# 3 0.216	05:28	No .020 agreement
Breath Volume	2.636	05:28	
Air Blank	0.000	05:29	
Control Test	0.078	05:29	
Air Blank	0.000	05:30	
Diagnostics Check	OK	05:30	

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALTUNAY,MOHAMMET
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: TORRES QUINTANA

DATE/TIME OF TEST: 05/21/2023 05:39:48

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.077	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.205	05:35
Breath Volume	1.515	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:37
Subject Sample # 2	0.189	05:38
Breath Volume	1.753	05:38
Air Blank	0.000	05:38
Control Test	0.077	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RADANOF,BARRY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:09

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 05/22/2023 09:39:20

Diagnostics Check	OK	09:32
Air Blank	0.000	09:32
Control Test	0.079	09:32
Air Blank	0.000	09:33
Subject Sample # 1	0.229	09:34
Breath Volume	1.605	09:34
Air Blank	0.000	09:35
Air Blank	0.000	09:37
Subject Sample # 2	0.229	09:37
Breath Volume	2.046	09:37
Air Blank	0.000	09:38
Control Test	0.079	09:38
Air Blank	0.000	09:39
Diagnostics Check	OK	09:39

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: EVARISTE

DATE/TIME OF TEST: 05/24/2023 02:50:09

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.020	02:45
Breath Volume	1.695	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.019	02:48
Breath Volume	1.359	02:48
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'DOR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: TALMADGE

DATE/TIME OF TEST: 05/24/2023 19:00:57

Diagnostics Check	OK	18:54
Air Blank	0.000	18:54
Control Test	0.080	18:55
Air Blank	0.000	18:55
Subject Sample # 1	0.235	18:56
Breath Volume	2.007	18:56
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.237	18:59
Breath Volume	2.718	18:59
Air Blank	0.000	19:00
Control Test	0.080	19:00
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: PALTOO

DATE/TIME OF TEST: 05/25/2023 00:32:42

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.139	00:27
Breath Volume	1.464	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.139	00:31
Breath Volume	1.300	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: SZUR

DATE/TIME OF TEST: 05/26/2023 19:50:28

Diagnostics Check	OK	19:44
Air Blank	0.000	19:44
Control Test	0.081	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.165	19:45
Breath Volume	3.835	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.174	19:48
Breath Volume	3.843	19:48
Air Blank	0.000	19:49
Control Test	0.080	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEE,JOHN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: SANCHEZ ALVAREZ

DATE/TIME OF TEST: 05/26/2023 20:40:41

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.081	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.096	20:36
Breath Volume	2.312	20:36
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 2	0.098	20:39
Breath Volume	2.292	20:39
Air Blank	0.000	20:39
Control Test	0.079	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RICHARDSON,ALEXANDER
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 05/27/2023 01:43:43

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.167	01:39
Breath Volume	1.199	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.164	01:42
Breath Volume	1.121	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SANTOS ABAD

DATE/TIME OF TEST: 05/27/2023 03:24:59

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.146	03:20
Breath Volume	2.007	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.157	03:23
Breath Volume	1.718	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FERNANDEZ,JOHN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 05/27/2023 03:52:53

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.194	03:47
Breath Volume	1.390	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.192	03:51
Breath Volume	1.335	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: VELASQUEZ ALVA

DATE/TIME OF TEST: 05/30/2023 01:35:57

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.230	01:31
Breath Volume	1.988	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:33
Subject Sample # 2	0.212	01:34
Breath Volume	1.367	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 06/07/2023 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 05/06/2023 02:25:52

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.079	02:13	
Air Blank	0.000	02:14	
Subject Sample	# 1 0.210	02:17	Volume Not Met
Breath Volume	0.578	02:17	
Air Blank	0.000	02:18	
Air Blank	0.000	02:20	
Subject Sample	# 2 0.226	02:21	
Breath Volume	1.550	02:21	
Air Blank	0.000	02:21	
Air Blank	0.000	02:23	
Subject Sample	# 3 0.228	02:24	
Breath Volume	1.417	02:24	
Air Blank	0.000	02:24	
Control Test	0.077	02:25	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 05/27/2023 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: SOZA

DATE/TIME OF TEST: 05/06/2023 20:37:35

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.078	20:30
Air Blank	0.000	20:30
Subject Sample # 1	0.147	20:32
Breath Volume	1.570	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:34
Subject Sample # 2	0.148	20:35
Breath Volume	1.242	20:35
Air Blank	0.000	20:36
Control Test	0.077	20:36
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 05/27/2023 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: CREAGAN

DATE/TIME OF TEST: 05/06/2023 23:10:44

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.078	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.117	23:05
Breath Volume	1.832	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.108	23:09
Breath Volume	1.367	23:09
Air Blank	0.000	23:09
Control Test	0.077	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 05/27/2023 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 05/07/2023 03:51:11

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.078	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.291	03:45
Breath Volume	1.644	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.290	03:49
Breath Volume	1.546	03:49
Air Blank	0.000	03:50
Control Test	0.076	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 05/27/2023 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: SPARIOSU

DATE/TIME OF TEST: 05/14/2023 03:44:08

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.000	03:39
Breath Volume	1.605	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.000	03:42
Breath Volume	1.496	03:42
Air Blank	0.000	03:43
Control Test	0.078	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 05/27/2023 19:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHM,JONATHAN,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: PINZON

DATE/TIME OF TEST: 05/22/2023 03:11:44

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.077	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1 0.166	03:03	Volume Not Met
Breath Volume	0.050	03:03	
Air Blank	0.000	03:04	
Air Blank	0.000	03:06	
Subject Sample	# 2 0.168	03:06	
Breath Volume	1.781	03:06	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
Subject Sample	# 3 0.171	03:10	
Breath Volume	1.816	03:10	
Air Blank	0.000	03:10	
Control Test	0.076	03:11	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 05/27/2023 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHM,JONATHAN,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: TABACH

DATE/TIME OF TEST: 05/28/2023 04:16:00

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.078	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.223	04:11
Breath Volume	2.085	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.228	04:14
Breath Volume	2.160	04:14
Air Blank	0.000	04:15
Control Test	0.077	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:16

Data Download Date/Time: 06/27/2023 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILAR,J,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: VELEZ,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ROMERO GARCIA

DATE/TIME OF TEST: 05/05/2023 01:43:58

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.077	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.177	01:39
Breath Volume	1.750	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.179	01:42
Breath Volume	1.445	01:42
Air Blank	0.000	01:43
Control Test	0.077	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 06/02/2023 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARGUELLES, ABRAHAM, A
BREATH TEST OPERATOR AGENCY: FHP TROOP F
ARRESTING OFFICER: ARGUELLES, ABRAHAM
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TOLEDO GARCIA

DATE/TIME OF TEST: 05/06/2023 04:14:44

Diagnostics Check	OK	04:14	
Air Blank		04:14	RFI Detect
Air Blank		04:14	RFI Detect

Data Download Date/Time: 06/02/2023 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: DAVILA MORALES

DATE/TIME OF TEST: 05/04/2023 16:31:38

Diagnostics Check	OK	16:28	
Air Blank	0.000	16:29	
Control Test	0.081	16:29	
Air Blank	0.000	16:30	
Subject Sample	# 1	16:30	Subject Test Refused
Breath Volume	0.000	16:30	
Air Blank	0.000	16:30	
Control Test	0.081	16:30	
Air Blank	0.000	16:31	
Diagnostics Check	OK	16:31	

Data Download Date/Time: 05/12/2023 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: NAVARRO,ROBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: CALLEJA

DATE/TIME OF TEST: 05/06/2023 00:19:08

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:16	
Control Test	0.081	00:16	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:17	Subject Test Refused
Breath Volume	0.000	00:17	
Air Blank	0.000	00:18	
Control Test	0.080	00:18	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:19	

Data Download Date/Time: 05/12/2023 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: UNLILASHVILI

DATE/TIME OF TEST: 05/01/2023 04:00:55

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.162	03:56
Breath Volume	2.144	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.160	03:59
Breath Volume	3.121	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 05/02/2023 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 05/03/2023 02:15:34

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.181	02:11
Breath Volume	1.941	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.186	02:13
Breath Volume	1.753	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 06/05/2023 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: LUKHI

DATE/TIME OF TEST: 05/04/2023 01:51:45

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.179	01:47
Breath Volume	3.371	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.184	01:50
Breath Volume	3.437	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 06/05/2023 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 05/04/2023 03:05:19

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.190	03:00
Breath Volume	3.707	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.183	03:03
Breath Volume	2.167	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: GIPSON

DATE/TIME OF TEST: 05/05/2023 00:33:28

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.166	00:28
Breath Volume	1.445	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.164	00:31
Breath Volume	2.226	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 05/05/2023 04:15:53

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:11	
Control Test	0.079	04:11	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:14	Subject Test Refused
Breath Volume	0.000	04:14	
Air Blank	0.000	04:14	
Control Test	0.079	04:15	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: PARKER JR

DATE/TIME OF TEST: 05/06/2023 00:25:56

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.013	00:21
Breath Volume	1.382	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.012	00:24
Breath Volume	1.511	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: RAYMOND

DATE/TIME OF TEST: 05/06/2023 04:15:28

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.195	04:09
Breath Volume	1.414	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.184	04:13
Breath Volume	1.460	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: GARCIA JIMENEZ

DATE/TIME OF TEST: 05/06/2023 05:03:51

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.123	04:59
Breath Volume	1.996	04:59
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.116	05:02
Breath Volume	1.421	05:02
Air Blank	0.000	05:02
Control Test	0.079	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: GOMEZ DIAZ

DATE/TIME OF TEST: 05/07/2023 04:14:48

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.188	04:10
Breath Volume	1.871	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.189	04:13
Breath Volume	1.664	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MOUSAW

DATE/TIME OF TEST: 05/07/2023 05:38:52

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.079	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.190	05:33
Breath Volume	2.300	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.177	05:37
Breath Volume	3.042	05:37
Air Blank	0.000	05:37
Control Test	0.078	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: OSORIO

DATE/TIME OF TEST: 05/09/2023 01:45:54

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.052	01:41
Breath Volume	1.167	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.056	01:44
Breath Volume	2.160	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: PETRY

DATE/TIME OF TEST: 05/11/2023 22:57:00

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.080	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.285	22:52
Breath Volume	2.683	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:54
Subject Sample # 2	0.292	22:55
Breath Volume	2.441	22:55
Air Blank	0.000	22:56
Control Test	0.078	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:57

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/12/2023 02:27:29

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.116	02:22
Breath Volume	1.707	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.122	02:25
Breath Volume	2.003	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 05/13/2023 04:20:57

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.079	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 1 0.126	04:13	Volume Not Met
Breath Volume	0.585	04:13	
Air Blank	0.000	04:14	
Air Blank	0.000	04:15	
Subject Sample	# 2 0.157	04:16	
Breath Volume	1.332	04:16	
Air Blank	0.000	04:17	
Air Blank	0.000	04:18	
Subject Sample	# 3 0.152	04:19	
Breath Volume	1.488	04:19	
Air Blank	0.000	04:20	
Control Test	0.078	04:20	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 05/13/2023 04:41:52

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.149	04:37
Breath Volume	1.941	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.141	04:40
Breath Volume	1.953	04:40
Air Blank	0.000	04:40
Control Test	0.078	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 06/05/2023 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DAVILA GAITAN

DATE/TIME OF TEST: 05/14/2023 02:17:09

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.166	02:12
Breath Volume	1.328	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.183	02:15
Breath Volume	2.449	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 06/05/2023 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: HERNANDEZ SANTIZ

DATE/TIME OF TEST: 05/14/2023 04:26:58

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.149	04:22
Breath Volume	1.699	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.154	04:25
Breath Volume	1.496	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 06/05/2023 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MCCLLOUD

DATE/TIME OF TEST: 05/15/2023 03:25:09

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.079	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:18	No Sample Provided
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample	# 2	03:23	No Sample Provided
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Control Test	0.079	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 06/05/2023 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LOPEZ,C
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:31

SUBJECT'S LAST NAME: CASTILLO SALAZAR

DATE/TIME OF TEST: 05/15/2023 08:04:08

Diagnostics Check	OK	07:51	
Air Blank	0.000	07:51	
Control Test	0.079	07:52	
Air Blank	0.000	07:52	
Subject Sample	# 1 0.184	07:55	Volume Not Met
Breath Volume	0.910	07:55	
Air Blank	0.000	07:56	
Air Blank	0.000	07:58	
Subject Sample	# 2 0.195	07:58	
Breath Volume	2.011	07:58	
Air Blank	0.000	07:59	
Air Blank	0.000	08:01	
Subject Sample	# 3 0.186	08:02	
Breath Volume	1.765	08:02	
Air Blank	0.000	08:03	
Control Test	0.078	08:03	
Air Blank	0.000	08:04	
Diagnostics Check	OK	08:04	

Data Download Date/Time: 06/05/2023 12:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CLAUDIO SEPULVEDA

DATE/TIME OF TEST: 05/17/2023 04:06:49

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.218	04:02
Breath Volume	1.250	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.213	04:05
Breath Volume	2.148	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 06/05/2023 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PENA,E
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: MARQUEZ-VEGA

DATE/TIME OF TEST: 05/18/2023 07:38:47

Diagnostics Check	OK	07:32
Air Blank	0.000	07:33
Control Test	0.079	07:33
Air Blank	0.000	07:33
Subject Sample # 1	0.000	07:34
Breath Volume	3.472	07:34
Air Blank	0.000	07:34
Air Blank	0.000	07:36
Subject Sample # 2	0.000	07:37
Breath Volume	3.464	07:37
Air Blank	0.000	07:37
Control Test	0.079	07:38
Air Blank	0.000	07:38
Diagnostics Check	OK	07:38

Data Download Date/Time: 06/05/2023 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: GILLESPIE

DATE/TIME OF TEST: 05/18/2023 23:11:31

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:08	
Control Test	0.079	23:08	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Control Test	0.079	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 06/05/2023 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBERTS, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: CARL

DATE/TIME OF TEST: 05/19/2023 02:58:38

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.000	02:54
Breath Volume	2.433	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.000	02:57
Breath Volume	1.457	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 06/05/2023 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLOOD,KRISTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: ASSELIN

DATE/TIME OF TEST: 05/19/2023 18:02:43

Diagnostics Check	OK	17:55
Air Blank	0.000	17:55
Control Test	0.079	17:56
Air Blank	0.000	17:56
Subject Sample # 1	0.129	17:57
Breath Volume	3.796	17:57
Air Blank	0.000	17:58
Air Blank	0.000	18:00
Subject Sample # 2	0.117	18:01
Breath Volume	1.515	18:01
Air Blank	0.000	18:01
Control Test	0.079	18:02
Air Blank	0.000	18:02
Diagnostics Check	OK	18:02

Data Download Date/Time: 06/05/2023 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: BERG

DATE/TIME OF TEST: 05/19/2023 20:33:14

Diagnostics Check	OK	20:26
Air Blank	0.000	20:27
Control Test	0.079	20:27
Air Blank	0.000	20:27
Subject Sample # 1	0.091	20:28
Breath Volume	2.777	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.089	20:31
Breath Volume	2.601	20:31
Air Blank	0.000	20:32
Control Test	0.079	20:32
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 06/05/2023 12:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: OCONNER

DATE/TIME OF TEST: 05/20/2023 02:45:53

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.189	02:41
Breath Volume	2.121	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.190	02:44
Breath Volume	1.722	02:44
Air Blank	0.000	02:44
Control Test	0.078	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 06/05/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BRAZAS

DATE/TIME OF TEST: 05/20/2023 04:46:50

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.078	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.035	04:42
Breath Volume	2.800	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.034	04:45
Breath Volume	2.632	04:45
Air Blank	0.000	04:45
Control Test	0.079	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 06/05/2023 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 05/21/2023 03:06:44

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.153	03:01
Breath Volume	1.765	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.149	03:05
Breath Volume	1.687	03:05
Air Blank	0.000	03:05
Control Test	0.078	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 06/05/2023 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: BOLESH III

DATE/TIME OF TEST: 05/21/2023 04:16:53

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.079	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1 0.194	04:05	No .020 agreement
Breath Volume	1.531	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 2 0.219	04:11	Volume Not Met
Breath Volume	0.406	04:11	
Air Blank	0.000	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 3 0.221	04:15	No .020 agreement
Breath Volume	1.781	04:15	
Air Blank	0.000	04:15	
Control Test	0.077	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 06/05/2023 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: ROACH

DATE/TIME OF TEST: 05/21/2023 04:37:13

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:21	
Control Test	0.077	04:21	
Air Blank	0.000	04:21	
Subject Sample	# 1 0.094	04:24	
Breath Volume	1.378	04:24	
Air Blank	0.000	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 2 0.100	04:29	Volume Not Met
Breath Volume	0.105	04:29	
Air Blank	0.000	04:30	
Air Blank	0.000	04:32	
Subject Sample	# 3 0.099	04:35	Volume Not Met
Breath Volume	0.464	04:35	
Air Blank	0.000	04:36	
Control Test	0.077	04:36	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 06/05/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: ROACH

DATE/TIME OF TEST: 05/21/2023 04:52:03

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:39	
Control Test	0.077	04:39	
Air Blank	0.000	04:39	
Subject Sample	# 1 0.098	04:41	
Breath Volume	1.917	04:41	
Air Blank	0.000	04:42	
Air Blank	0.000	04:43	
Subject Sample	# 2 0.101	04:47	Volume Not Met
Breath Volume	0.054	04:47	
Air Blank	0.000	04:47	
Air Blank	0.000	04:49	
Subject Sample	# 3 0.095	04:50	
Breath Volume	2.085	04:50	
Air Blank	0.000	04:51	
Control Test	0.076	04:51	
Air Blank	0.000	04:51	
Diagnostics Check	OK	04:52	

Data Download Date/Time: 06/05/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: MCCARTHY LOEHR

DATE/TIME OF TEST: 05/21/2023 23:01:03

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:53	
Control Test	0.079	22:53	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:57	No Sample Provided
Breath Volume	0.000	22:57	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 2	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.079	23:00	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 06/05/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: GUZMAN DIAZ

DATE/TIME OF TEST: 05/21/2023 23:40:04

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.078	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.139	23:35
Breath Volume	1.582	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.153	23:38
Breath Volume	2.796	23:38
Air Blank	0.000	23:39
Control Test	0.077	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:40

Data Download Date/Time: 06/05/2023 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 05/22/2023 04:39:38

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.173	04:35
Breath Volume	2.531	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.171	04:38
Breath Volume	2.519	04:38
Air Blank	0.000	04:38
Control Test	0.079	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 06/05/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRUZ,N
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: LUU

DATE/TIME OF TEST: 05/22/2023 16:42:44

Diagnostics Check	OK	16:35
Air Blank	0.000	16:35
Control Test	0.079	16:36
Air Blank	0.000	16:36
Subject Sample # 1	0.174	16:38
Breath Volume	2.085	16:38
Air Blank	0.000	16:38
Air Blank	0.000	16:40
Subject Sample # 2	0.178	16:41
Breath Volume	2.332	16:41
Air Blank	0.000	16:41
Control Test	0.085	16:42
Air Blank	0.000	16:42
Diagnostics Check	OK	16:42

Data Download Date/Time: 06/05/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: RYALS

DATE/TIME OF TEST: 05/25/2023 01:42:20

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.080	01:30	
Air Blank	0.000	01:30	
Subject Sample	# 1 0.171	01:33	Volume Not Met
Breath Volume	0.148	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample	# 2 0.172	01:37	
Breath Volume	1.136	01:37	
Air Blank	0.000	01:38	
Air Blank	0.000	01:40	
Subject Sample	# 3 0.175	01:40	
Breath Volume	1.257	01:40	
Air Blank	0.000	01:41	
Control Test	0.077	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 06/05/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 05/25/2023 22:51:33

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.080	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.163	22:46
Breath Volume	1.808	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:48
Subject Sample # 2	0.160	22:49
Breath Volume	1.613	22:49
Air Blank	0.000	22:50
Control Test	0.079	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 06/05/2023 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 05/26/2023 03:00:55

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.104	02:56
Breath Volume	1.699	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.104	02:59
Breath Volume	2.007	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 06/05/2023 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ADAIR,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: MALLARD

DATE/TIME OF TEST: 05/26/2023 19:16:22

Diagnostics Check	OK	19:09
Air Blank	0.000	19:09
Control Test	0.079	19:09
Air Blank	0.000	19:10
Subject Sample # 1	0.151	19:11
Breath Volume	1.207	19:11
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample # 2	0.149	19:14
Breath Volume	1.628	19:14
Air Blank	0.000	19:15
Control Test	0.078	19:15
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

Data Download Date/Time: 06/05/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: YAROSH

DATE/TIME OF TEST: 05/27/2023 02:42:26

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.141	02:37
Breath Volume	2.242	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.141	02:40
Breath Volume	2.308	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/05/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: CROSON

DATE/TIME OF TEST: 05/27/2023 02:52:36

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.078	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.036	02:48
Breath Volume	1.984	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.036	02:51
Breath Volume	1.714	02:51
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 06/05/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: OGBAJI

DATE/TIME OF TEST: 05/27/2023 04:07:05

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.094	04:02
Breath Volume	1.527	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.093	04:05
Breath Volume	1.800	04:05
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 06/05/2023 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: GUNTER JR

DATE/TIME OF TEST: 05/27/2023 05:18:19

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.079	05:12
Air Blank	0.000	05:12
Subject Sample # 1	0.111	05:13
Breath Volume	2.136	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.102	05:16
Breath Volume	1.875	05:16
Air Blank	0.000	05:17
Control Test	0.078	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 06/05/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: HERRERRA ALVAREZ

DATE/TIME OF TEST: 05/27/2023 18:31:06

Diagnostics Check	OK	18:20	
Air Blank	0.000	18:20	
Control Test	0.080	18:20	
Air Blank	0.000	18:21	
Subject Sample	# 1 0.227	18:24	Volume Not Met
Breath Volume	1.542	18:24	
Air Blank	0.000	18:25	
Air Blank	0.000	18:27	
Subject Sample	# 2	18:29	Subject Test Refused
Breath Volume	0.000	18:29	
Air Blank	0.000	18:30	
Control Test	0.080	18:30	
Air Blank	0.000	18:31	
Diagnostics Check	OK	18:31	

Data Download Date/Time: 06/05/2023 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: PEREZ RUIZ

DATE/TIME OF TEST: 05/28/2023 00:53:40

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.205	00:44
Breath Volume	1.125	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:46
Subject Sample # 2	0.230	00:48
Breath Volume	1.515	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:50
Subject Sample # 3	0.223	00:51
Breath Volume	1.937	00:51
Air Blank	0.000	00:52
Control Test	0.077	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 06/05/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 05/28/2023 02:25:59

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.180	02:21
Breath Volume	2.675	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:23
Subject Sample # 2	0.183	02:24
Breath Volume	2.503	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 06/05/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: YANTIL ROJAS

DATE/TIME OF TEST: 05/28/2023 04:17:50

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.174	04:13
Breath Volume	1.679	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.161	04:16
Breath Volume	1.449	04:16
Air Blank	0.000	04:16
Control Test	0.078	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 06/05/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 05/28/2023 23:46:18

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.078	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.282	23:41
Breath Volume	3.160	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.275	23:44
Breath Volume	3.449	23:44
Air Blank	0.000	23:45
Control Test	0.078	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 06/05/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: ALARCON BATISTA

DATE/TIME OF TEST: 05/29/2023 04:44:23

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.207	04:39
Breath Volume	2.386	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:41
Subject Sample # 2	0.193	04:42
Breath Volume	2.707	04:42
Air Blank	0.000	04:43
Control Test	0.078	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 06/05/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/29/2023 05:17:00

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.078	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.160	05:12
Breath Volume	1.726	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:14
Subject Sample # 2	0.162	05:15
Breath Volume	1.937	05:15
Air Blank	0.000	05:16
Control Test	0.077	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:17

Data Download Date/Time: 06/05/2023 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/30/2023 00:26:56

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.079	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.155	00:21
Breath Volume	1.144	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.160	00:25
Breath Volume	2.515	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 06/05/2023 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/30/2023 02:17:28

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.240	02:12
Breath Volume	2.187	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.248	02:15
Breath Volume	1.968	02:15
Air Blank	0.000	02:16
Control Test	0.078	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 06/05/2023 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 05/30/2023 19:45:50

Diagnostics Check	OK	19:39
Air Blank	0.000	19:39
Control Test	0.080	19:39
Air Blank	0.000	19:40
Subject Sample # 1	0.288	19:41
Breath Volume	3.066	19:41
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.289	19:44
Breath Volume	3.015	19:44
Air Blank	0.000	19:44
Control Test	0.078	19:45
Air Blank	0.000	19:45
Diagnostics Check	OK	19:45

Data Download Date/Time: 06/05/2023 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: MINNS JR

DATE/TIME OF TEST: 05/31/2023 05:06:14

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.080	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.098	05:01
Breath Volume	1.742	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.102	05:04
Breath Volume	1.140	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 06/05/2023 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLLINS,CHASE,M

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: COLLINS,CHASE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 21:07

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 05/30/2023 22:54:22

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.081	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.000	22:50
Breath Volume	2.167	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.000	22:53
Breath Volume	2.339	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 06/28/2023 08:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: DELANEY,BILL
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 05/04/2023 20:27:17

Diagnostics Check	OK	20:17	
Air Blank	0.000	20:17	
Control Test	0.081	20:17	
Air Blank	0.000	20:18	
Subject Sample	# 1 0.260	20:18	
Breath Volume	2.847	20:18	
Air Blank	0.000	20:19	
Air Blank	0.000	20:21	
Subject Sample	# 2 0.238	20:22	
Breath Volume	2.570	20:22	
Air Blank	0.000	20:23	
Air Blank	0.000	20:25	
Subject Sample	# 3 0.209	20:25	
Breath Volume	2.609	20:25	
Air Blank	0.000	20:26	
Control Test	0.072	20:26	Control Outside Tolerance
Air Blank	0.000	20:27	

Data Download Date/Time: 06/02/2023 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: DELANEY,BILL
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 05/04/2023 20:42:11

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.079	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.250	20:37
Breath Volume	2.777	20:37
Air Blank	0.000	20:38
Air Blank	0.000	20:39
Subject Sample # 2	0.261	20:40
Breath Volume	3.113	20:40
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 06/02/2023 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY,JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/26/2023 23:49:19

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.181	23:44
Breath Volume	1.468	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.196	23:47
Breath Volume	2.050	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 06/02/2023 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PETERSON,DEVON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: LYDEN

DATE/TIME OF TEST: 05/01/2023 02:28:09

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.126	02:23
Breath Volume	2.406	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.129	02:26
Breath Volume	1.953	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 05/04/2023 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEMIEC,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Key Colony Beach PD
ARRESTING OFFICER: DULLY,STEPHEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: TURNER-FERRUCCI

DATE/TIME OF TEST: 05/02/2023 22:13:15

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:06	
Control Test	0.080	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 1 0.177	22:08	
Breath Volume	1.800	22:08	
Air Blank	0.000	22:09	
Air Blank	0.000	22:10	
Subject Sample	# 2 0.175	22:11	
Breath Volume	1.425	22:11	
Air Blank	0.000	22:11	
Control Test	0.079	22:12	
Air Blank		22:12	RFI Detect
Air Blank	0.000	22:13	

Data Download Date/Time: 05/04/2023 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEMIEC,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Key Colony Beach PD
ARRESTING OFFICER: DULLY,STEPHEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: TURNER-FERRUCCI

DATE/TIME OF TEST: 05/02/2023 23:05:50

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.080	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.175	23:01
Breath Volume	1.562	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.168	23:04
Breath Volume	1.675	23:04
Air Blank	0.000	23:04
Control Test	0.080	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 05/04/2023 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DEL CUETO,ERYKA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: CHARTIER

DATE/TIME OF TEST: 05/04/2023 05:26:07

Diagnostics Check	OK	05:18
Air Blank	0.000	05:18
Control Test	0.080	05:19
Air Blank	0.000	05:19
Subject Sample # 1	0.176	05:20
Breath Volume	1.281	05:20
Air Blank	0.000	05:21
Air Blank	0.000	05:22
Subject Sample # 2	0.174	05:24
Breath Volume	1.296	05:24
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 05/04/2023 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DEL CUETO,ERYKA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: FESS

DATE/TIME OF TEST: 05/12/2023 20:40:14

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.080	20:33
Air Blank	0.000	20:34
Subject Sample # 1	0.363	20:35
Breath Volume	3.054	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:37
Subject Sample # 2	0.346	20:38
Breath Volume	2.109	20:38
Air Blank	0.000	20:39
Control Test	0.078	20:39
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 06/08/2023 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/13/2023 03:20:25

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.079	03:18	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 06/08/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DEL CUETO,ERYKA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BUTTON

DATE/TIME OF TEST: 05/13/2023 04:42:36

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.166	04:37
Breath Volume	1.796	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.162	04:40
Breath Volume	2.796	04:40
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 06/08/2023 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GONZALEZ RODAS

DATE/TIME OF TEST: 05/14/2023 02:47:09

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 06/08/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BUNKER

DATE/TIME OF TEST: 05/19/2023 01:36:52

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.195	01:32
Breath Volume	2.589	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.196	01:35
Breath Volume	2.996	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 06/08/2023 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CURRY,CONNOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: STONEBURNER

DATE/TIME OF TEST: 05/19/2023 04:36:41

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:33	
Control Test	0.080	04:34	
Air Blank	0.000	04:34	
Subject Sample	# 1	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:35	
Control Test	0.080	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 06/08/2023 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: MIRABAL BARROSO

DATE/TIME OF TEST: 05/23/2023 05:40:36

Diagnostics Check	OK	05:37	
Air Blank	0.000	05:37	
Control Test	0.080	05:38	
Air Blank	0.000	05:38	
Subject Sample	# 1	05:39	Subject Test Refused
Breath Volume	0.000	05:39	
Air Blank	0.000	05:39	
Control Test	0.080	05:39	
Air Blank	0.000	05:40	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 06/08/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/26/2023 22:27:51

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.247	22:23
Breath Volume	2.285	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample # 2	0.256	22:26
Breath Volume	1.773	22:26
Air Blank	0.000	22:26
Control Test	0.078	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 06/08/2023 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FLANEGAN,BROOKE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/27/2023 00:10:55

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.080	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:09	
Control Test	0.080	00:10	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 06/08/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRAGG,GARRETT
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 05/29/2023 02:40:52

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Control Test	0.080	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 06/08/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINTON, BLAKE, E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MINTON, BLAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: ANDRES

DATE/TIME OF TEST: 05/01/2023 03:19:43

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.132	03:14
Breath Volume	1.496	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.123	03:17
Breath Volume	1.765	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 05/06/2022 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HERNANDEZ,IVAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/14/2023 22:33:28

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:28	
Control Test	0.079	22:28	
Air Blank	0.000	22:28	
Subject Sample	# 1 0.283	22:32	Volume Not Met
Breath Volume	0.070	22:32	
Air Blank		22:32	Purge Fail
Air Blank	0.000	22:33	

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HERNANDEZ,IVAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/14/2023 22:38:08

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.074	22:37	Control Outside Tolerance
Air Blank	0.000	22:38	

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HERNANDEZ,IVAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/14/2023 22:45:05

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.078	22:40	
Air Blank	0.000	22:40	
Subject Sample	# 1 0.257	22:43	Volume Not Met
Breath Volume	0.285	22:43	
Air Blank		22:44	Purge Fail
Air Blank	0.000	22:45	

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HERNANDEZ,IVAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/14/2023 22:58:50

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:47	
Control Test	0.075	22:47	
Air Blank	0.000	22:48	
Subject Sample	# 1 0.282	22:51	Volume Not Met
Breath Volume	0.445	22:51	
Air Blank	0.000	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 2 0.260	22:56	Volume Not Met
Breath Volume	0.050	22:56	
Air Blank	0.000	22:57	
Control Test	0.076	22:58	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,VICTORIA,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: PHILLIPS,VICTORIA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 05/14/2023 23:16:20

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.076	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.105	23:11
Breath Volume	2.453	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:13
Subject Sample # 2	0.110	23:14
Breath Volume	3.722	23:14
Air Blank	0.000	23:15
Control Test	0.078	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRINGER,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: STRINGER,MICHAEL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MARCUM

DATE/TIME OF TEST: 05/16/2023 01:23:27

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.272	01:14
Breath Volume	1.117	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:16
Subject Sample # 2	0.247	01:18
Breath Volume	2.328	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:20
Subject Sample # 3	0.271	01:21
Breath Volume	1.234	01:21
Air Blank	0.000	01:22
Control Test	0.075	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: MAKI

DATE/TIME OF TEST: 05/16/2023 18:00:12

Diagnostics Check	OK	17:52
Air Blank	0.000	17:53
Control Test	0.079	17:53
Air Blank	0.000	17:54
Subject Sample # 1	0.176	17:55
Breath Volume	1.421	17:55
Air Blank	0.000	17:56
Air Blank	0.000	17:57
Subject Sample # 2	0.173	17:58
Breath Volume	1.726	17:58
Air Blank	0.000	17:59
Control Test	0.078	17:59
Air Blank	0.000	18:00
Diagnostics Check	OK	18:00

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAINO,JOSHUA,V
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: ZAINO,JOSHUA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: NEWPORT

DATE/TIME OF TEST: 05/21/2023 00:28:01

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.305	00:23
Breath Volume	2.640	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:25
Subject Sample # 2	0.308	00:26
Breath Volume	3.039	00:26
Air Blank	0.000	00:27
Control Test	0.077	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:28

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAULKNER,JAMES,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FAULKNER,JAMES
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MANSFIELD

DATE/TIME OF TEST: 05/21/2023 00:57:58

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.079	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.122	00:53
Breath Volume	3.316	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample # 2	0.121	00:56
Breath Volume	4.101	00:56
Air Blank	0.000	00:57
Control Test	0.077	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUTRAL,WILLIAM,Z
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ECKELI,CONNOR
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 05/21/2023 03:27:55

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.361	03:23
Breath Volume	1.589	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:25
Subject Sample # 2	0.353	03:26
Breath Volume	1.210	03:26
Air Blank	0.000	03:26
Control Test	0.076	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFENBERG,ASHTON,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WELFENBERG,ASHTON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: MORALES FRANQUI

DATE/TIME OF TEST: 05/21/2023 04:02:14

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.048	03:57
Breath Volume	3.093	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 2	0.044	04:00
Breath Volume	2.574	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JESSE,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BROWN,JESSE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: AURICH

DATE/TIME OF TEST: 05/22/2023 04:43:06

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.152	04:38
Breath Volume	3.113	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:40
Subject Sample # 2	0.147	04:41
Breath Volume	2.953	04:41
Air Blank	0.000	04:42
Control Test	0.078	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/23/2023 00:29:25

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.158	00:24
Breath Volume	1.593	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:26
Subject Sample # 2	0.140	00:27
Breath Volume	1.199	00:27
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 05/24/2023 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DUFOUR,NOAH,I

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: DUFOUR,NOAH

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 01:35

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 05/01/2023 02:39:39

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.081	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 1 0.093	02:32	Volume Not Met
Breath Volume	0.015	02:32	
Air Blank	0.000	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 2 0.077	02:38	Volume Not Met
Breath Volume	0.250	02:38	
Air Blank	0.000	02:38	
Control Test	0.079	02:39	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 05/30/2023 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DUFOUR,NOAH,I

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: DUFOUR,NOAH

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 01:35

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 05/01/2023 03:01:13

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.079	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1 0.087	02:48	
Breath Volume	1.800	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 2 0.086	02:54	Volume Not Met
Breath Volume	0.136	02:54	
Air Blank	0.000	02:54	
Air Blank	0.000	02:56	
Subject Sample	# 3 0.079	02:59	Volume Not Met
Breath Volume	0.269	02:59	
Air Blank	0.000	03:00	
Control Test	0.077	03:00	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 05/30/2023 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 03:18

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 05/06/2023 03:56:22

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.226	03:51
Breath Volume	1.976	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.223	03:54
Breath Volume	2.351	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 05/30/2023 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PINO,RICHARD,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: WALSTON,CHRISTOPHER

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 03:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 05/07/2023 04:15:21

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.253	04:10
Breath Volume	3.843	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.243	04:13
Breath Volume	3.179	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1408275
BREATH TEST OPERATOR: GAFKJEN,THOMAS,D	DRY GAS EXPIRY DATE: 11/19/2023
BREATH TEST OPERATOR AGENCY: Pensacola PD	VIOLATION CODE: DUI
ARRESTING OFFICER: GAFKJEN,THOMAS	ARREST TIME: 05:50
ARRESTING OFFICER AGENCY: Pensacola PD	GENDER: Male
LAST AGENCY INSPECTION: 04/25/2023	OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: NICHOLSON	DATE/TIME OF TEST: 05/11/2023	06:59:14
--------------------------------	-------------------------------	----------

Diagnostics Check	OK	06:52
Air Blank	0.000	06:53
Control Test	0.081	06:53
Air Blank	0.000	06:53
Subject Sample # 1	0.000	06:54
Breath Volume	2.804	06:54
Air Blank	0.000	06:55
Air Blank	0.000	06:57
Subject Sample # 2	0.000	06:57
Breath Volume	2.117	06:57
Air Blank	0.000	06:58
Control Test	0.080	06:58
Air Blank	0.000	06:59
Diagnostics Check	OK	06:59

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 00:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 05/12/2023 01:13:40

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.081	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.155	01:09
Breath Volume	2.218	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.153	01:12
Breath Volume	2.695	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MARTINEZ,HENDRICK
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 14:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:31

SUBJECT'S LAST NAME: FARMAN

DATE/TIME OF TEST: 05/13/2023 14:57:24

Diagnostics Check	OK	14:51
Air Blank	0.000	14:51
Control Test	0.081	14:52
Air Blank	0.000	14:52
Subject Sample # 1	0.000	14:53
Breath Volume	3.894	14:53
Air Blank	0.000	14:53
Air Blank	0.000	14:55
Subject Sample # 2	0.000	14:56
Breath Volume	3.953	14:56
Air Blank	0.000	14:56
Control Test	0.080	14:56
Air Blank	0.000	14:57
Diagnostics Check	OK	14:57

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1408275
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J	DRY GAS EXPIRY DATE: 11/19/2023
BREATH TEST OPERATOR AGENCY: Pensacola PD	VIOLATION CODE: DUI
ARRESTING OFFICER: MARTINEZ,HENDRICK	ARREST TIME: 12:12
ARRESTING OFFICER AGENCY: Pensacola PD	GENDER: Female
LAST AGENCY INSPECTION: 04/25/2023	OBSERVATION PERIOD BEGAN: 12:28

SUBJECT'S LAST NAME: WATKINS DATE/TIME OF TEST: 05/14/2023 12:56:34

Diagnostics Check	OK	12:48
Air Blank	0.000	12:49
Control Test	0.078	12:49
Air Blank	0.000	12:49
Subject Sample # 1	0.276	12:51
Breath Volume	1.613	12:51
Air Blank	0.000	12:52
Air Blank	0.000	12:54
Subject Sample # 2	0.258	12:54
Breath Volume	1.476	12:54
Air Blank	0.000	12:55
Control Test	0.075	12:55
Air Blank	0.000	12:56
Diagnostics Check	OK	12:56

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PINO,RICHARD,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: PINO,RICHARD

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 01:36

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/17/2023 02:20:47

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.081	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.188	02:16
Breath Volume	3.000	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.188	02:19
Breath Volume	2.105	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLLINS,CHASE,M

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: COLLINS,CHASE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 15:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 05/17/2023 16:15:10

Diagnostics Check	OK	16:08
Air Blank	0.000	16:09
Control Test	0.080	16:09
Air Blank	0.000	16:10
Subject Sample # 1	0.137	16:10
Breath Volume	1.378	16:10
Air Blank	0.000	16:11
Air Blank	0.000	16:13
Subject Sample # 2	0.126	16:13
Breath Volume	1.535	16:13
Air Blank	0.000	16:14
Control Test	0.080	16:14
Air Blank	0.000	16:15
Diagnostics Check	OK	16:15

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASWELL,AUSTIN,J

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: HASWELL,AUSTIN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 03:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: SEWELL-MCCLURG

DATE/TIME OF TEST: 05/18/2023 04:09:51

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.194	04:05
Breath Volume	1.953	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.187	04:08
Breath Volume	1.898	04:08
Air Blank	0.000	04:08
Control Test	0.079	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DUFOUR,NOAH,I

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: DODSON,EVAN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 19:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/22/2023 19:46:47

Diagnostics Check	OK	19:40
Air Blank	0.000	19:40
Control Test	0.080	19:40
Air Blank	0.000	19:41
Subject Sample # 1	0.305	19:41
Breath Volume	1.546	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.297	19:45
Breath Volume	1.281	19:45
Air Blank	0.000	19:45
Control Test	0.079	19:46
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 21:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 05/24/2023 22:07:17

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.135	22:02
Breath Volume	3.167	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.135	22:05
Breath Volume	2.906	22:05
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MARTINEZ,HENDRICK
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: DEMERS

DATE/TIME OF TEST: 05/25/2023 02:21:55

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.081	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.119	02:17
Breath Volume	1.933	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.113	02:20
Breath Volume	1.894	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1408275
BREATH TEST OPERATOR: COTTON,PIERCE,A	DRY GAS EXPIRY DATE: 11/19/2023
BREATH TEST OPERATOR AGENCY: Pensacola PD	VIOLATION CODE: DUI
ARRESTING OFFICER: COTTON,PIERCE	ARREST TIME: 21:47
ARRESTING OFFICER AGENCY: Pensacola PD	GENDER: Male
LAST AGENCY INSPECTION: 04/25/2023	OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: CARLINE	DATE/TIME OF TEST: 05/25/2023	22:33:36
------------------------------	-------------------------------	----------

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:16	
Control Test	0.081	22:17	
Air Blank	0.000	22:17	
Subject Sample	# 1 0.222	22:20	
Breath Volume	1.703	22:20	
Air Blank	0.000	22:21	
Air Blank	0.000	22:23	
Subject Sample	# 2 0.186	22:26	Volume Not Met
Breath Volume	0.082	22:26	
Air Blank	0.000	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 3 0.204	22:31	Volume Not Met
Breath Volume	0.039	22:31	
Air Blank	0.000	22:32	
Control Test	0.077	22:32	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COTTON,PIERCE,A

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: COTTON,PIERCE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 01:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: PERDOMO-GOMEZ

DATE/TIME OF TEST: 05/30/2023 02:16:37

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.207	02:11
Breath Volume	1.324	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.189	02:14
Breath Volume	1.636	02:14
Air Blank	0.000	02:15
Control Test	0.081	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 05/30/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: DEGROVE

DATE/TIME OF TEST: 05/12/2023 00:11:19

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.109	00:06
Breath Volume	2.371	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.110	00:09
Breath Volume	2.660	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 06/01/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: MCCANDLESS,MATTHEW

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 02:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BROWNE

DATE/TIME OF TEST: 05/12/2023 02:56:19

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.078	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.066	02:51
Breath Volume	2.363	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.061	02:54
Breath Volume	1.304	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 06/01/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 02:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/13/2023 02:22:04

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.079	02:20	
Air Blank	0.000	02:20	
Subject Sample	# 1 0.097	02:21	
Breath Volume	1.949	02:21	
Air Blank		02:21	RFI Detect
Air Blank		02:21	RFI Detect
Air Blank	0.000	02:22	

Data Download Date/Time: 06/01/2023 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 02:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/13/2023 02:30:12

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.079	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1 0.100	02:24	
Breath Volume	1.414	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 2 0.090	02:28	
Breath Volume	1.433	02:28	
Air Blank	0.000	02:28	
Control Test	0.080	02:29	
Air Blank	0.000	02:29	
Diagnostics Check	Fail	02:29	Internal Printer Test
Air Blank	0.000	02:30	

Data Download Date/Time: 06/01/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/13/2023 02:38:52

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.078	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.093	02:33
Breath Volume	1.996	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.089	02:37
Breath Volume	1.261	02:37
Air Blank	0.000	02:37
Control Test	0.078	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 06/01/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: GREEN,J,A	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GREEN,J	ARREST TIME: 01:21
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/27/2023	OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: WILLEY DATE/TIME OF TEST: 05/14/2023 01:33:53

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.078	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.120	01:29
Breath Volume	1.902	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.120	01:32
Breath Volume	2.796	01:32
Air Blank	0.000	01:32
Control Test	0.083	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 06/01/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: TARGOVE

DATE/TIME OF TEST: 05/17/2023 23:00:50

Diagnostics Check	Fail	23:00	Internal Printer Test
Air Blank	0.000	23:00	

Data Download Date/Time: 06/01/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: GREEN,J,A	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GREEN,J	ARREST TIME: 22:52
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/27/2023	OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: TARGOVE	DATE/TIME OF TEST: 05/17/2023	23:08:13
------------------------------	-------------------------------	----------

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.078	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.008	23:03
Breath Volume	2.769	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.000	23:06
Breath Volume	2.964	23:06
Air Blank	0.000	23:07
Control Test	0.081	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 06/01/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: GROVES

DATE/TIME OF TEST: 05/20/2023 02:29:52

Diagnostics Check	Fail	02:29	Internal Printer Test
Air Blank	0.000	02:29	

Data Download Date/Time: 06/01/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: ELLIS,ALEXANDER,	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ELLIS,ALEXANDER	ARREST TIME: 02:19
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/27/2023	OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GROVES DATE/TIME OF TEST: 05/20/2023 02:34:29

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:33	RFI Detect
Breath Volume	0.000	02:33	
Air Blank	0.000	02:33	
Control Test	0.080	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 06/01/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GROVES

DATE/TIME OF TEST: 05/20/2023 02:42:03

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.072	02:37
Breath Volume	2.554	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample # 2	0.074	02:40
Breath Volume	2.765	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/01/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 05/31/2023 23:21:55

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.078	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.116	23:17
Breath Volume	1.847	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:19
Subject Sample # 2	0.112	23:20
Breath Volume	1.460	23:20
Air Blank	0.000	23:20
Control Test	0.079	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 06/01/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RIGGINS,BRENDAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ENNIS

DATE/TIME OF TEST: 05/01/2023 00:46:51

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.078	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.153	00:42
Breath Volume	3.097	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:44
Subject Sample # 2	0.155	00:45
Breath Volume	2.402	00:45
Air Blank	0.000	00:45
Control Test	0.078	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 06/01/2023 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAHNSEN,A,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MORAN,S
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 05/04/2023 04:07:26

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.135	04:03
Breath Volume	1.417	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.141	04:05
Breath Volume	1.968	04:05
Air Blank	0.000	04:06
Control Test	0.076	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 06/01/2023 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALLOUN,SAMUEL,C
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RICHWINE,CHARLES
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MESH

DATE/TIME OF TEST: 05/05/2023 02:38:38

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.170	02:34
Breath Volume	1.335	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.159	02:37
Breath Volume	1.718	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CORREA,JONATHAN,E

BREATH TEST OPERATOR AGENCY: St. Augustine PD

ARRESTING OFFICER: CORREA,JONATHAN

ARRESTING OFFICER AGENCY: St. Augustine PD

LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211543

DRY GAS EXPIRY DATE: 01/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: KALOW

DATE/TIME OF TEST: 05/05/2023 22:57:27

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.079	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.189	22:53
Breath Volume	2.726	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.194	22:55
Breath Volume	2.960	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: LOVELACE

DATE/TIME OF TEST: 05/06/2023 00:46:31

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.079	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.144	00:42
Breath Volume	1.531	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.136	00:45
Breath Volume	1.101	00:45
Air Blank	0.000	00:45
Control Test	0.079	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: SCHRAUFNAGL

DATE/TIME OF TEST: 05/06/2023 04:50:52

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.079	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.153	04:46
Breath Volume	2.097	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.152	04:49
Breath Volume	1.160	04:49
Air Blank	0.000	04:49
Control Test	0.077	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: MCCANDLESS,MATTHEW

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211543

DRY GAS EXPIRY DATE: 01/05/2024

VIOLATION CODE: DUI

ARREST TIME: 19:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: HABERKORN

DATE/TIME OF TEST: 05/06/2023 21:00:51

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:47	
Control Test	0.079	20:47	
Air Blank	0.000	20:48	
Subject Sample	# 1 0.176	20:50	
Breath Volume	2.179	20:50	
Air Blank	0.000	20:51	
Air Blank	0.000	20:52	
Subject Sample	# 2 0.171	20:56	Volume Not Met
Breath Volume	0.738	20:56	
Air Blank	0.000	20:56	
Air Blank	0.000	20:58	
Subject Sample	# 3	20:59	Subject Test Refused
Breath Volume	0.000	20:59	
Air Blank	0.000	20:59	
Control Test	0.076	21:00	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:00	

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: MYERS,BENJAMIN,P	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: KELLY,GRIFFEN	ARREST TIME: 02:39
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/27/2023	OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BROWN DATE/TIME OF TEST: 05/07/2023 03:41:44

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.256	03:34
Breath Volume	1.324	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample # 2	0.249	03:39
Breath Volume	1.222	03:39
Air Blank	0.000	03:40
Control Test	0.076	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MCCANDLESS,MATTHEW	ARREST TIME: 03:23
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/27/2023	OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: MAGDALENO RODRIGUEZ	DATE/TIME OF TEST: 05/07/2023 04:12:56
--	--

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.077	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.108	04:07
Breath Volume	1.250	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.102	04:11
Breath Volume	1.531	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MCCANDLESS,MATTHEW	ARREST TIME: 22:06
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/27/2023	OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: RAMDASS	DATE/TIME OF TEST: 05/07/2023 23:35:17
------------------------------	--

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:24	
Control Test	0.080	23:24	
Air Blank	0.000	23:24	
Subject Sample	# 1 0.108	23:28	Volume Not Met
Breath Volume	0.765	23:28	
Air Blank	0.000	23:28	
Air Blank	0.000	23:30	
Subject Sample	# 2 0.106	23:33	Volume Not Met
Breath Volume	1.097	23:33	
Air Blank	0.000	23:34	
Control Test	0.079	23:34	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: KELLY,GRIFFIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 05/07/2023 23:58:04

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:45	
Control Test	0.078	23:45	
Air Blank	0.000	23:46	
Subject Sample	# 1 0.162	23:49	Volume Not Met
Breath Volume	0.378	23:49	
Air Blank	0.000	23:50	
Air Blank	0.000	23:51	
Subject Sample	# 2 0.188	23:52	
Breath Volume	1.496	23:52	
Air Blank	0.000	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 3 0.178	23:56	
Breath Volume	1.554	23:56	
Air Blank	0.000	23:57	
Control Test	0.076	23:57	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 05/08/2023 18:46:20

Diagnostics Check	OK	18:39
Air Blank	0.000	18:40
Control Test	0.078	18:40
Air Blank	0.000	18:41
Subject Sample # 1	0.000	18:42
Breath Volume	1.324	18:42
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.000	18:44
Breath Volume	1.726	18:44
Air Blank	0.000	18:45
Control Test	0.080	18:45
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MORAN,SEAN
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: ZILCH

DATE/TIME OF TEST: 05/12/2023 23:32:49

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:18	
Control Test	0.080	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 1 0.071	23:22	Volume Not Met
Breath Volume	0.171	23:22	
Air Blank	0.000	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 2 0.078	23:28	
Breath Volume	1.335	23:28	
Air Blank	0.000	23:28	
Air Blank	0.000	23:30	
Subject Sample	# 3 0.074	23:31	
Breath Volume	1.460	23:31	
Air Blank	0.000	23:31	
Control Test	0.078	23:32	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RIGGINS,BRENDAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: GANDHI

DATE/TIME OF TEST: 05/13/2023 00:54:56

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.174	00:50
Breath Volume	2.503	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:52
Subject Sample # 2	0.171	00:53
Breath Volume	2.929	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONTENOT,DOMINIC,M
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: RICHWINE,CHARLES
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: DINGEE

DATE/TIME OF TEST: 05/13/2023 17:14:04

Diagnostics Check	OK	17:07
Air Blank	0.000	17:07
Control Test	0.080	17:08
Air Blank	0.000	17:08
Subject Sample # 1	0.135	17:09
Breath Volume	1.312	17:09
Air Blank	0.000	17:09
Air Blank	0.000	17:11
Subject Sample # 2	0.132	17:12
Breath Volume	1.363	17:12
Air Blank	0.000	17:13
Control Test	0.079	17:13
Air Blank	0.000	17:13
Diagnostics Check	OK	17:14

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: CRESPO

DATE/TIME OF TEST: 05/14/2023 00:45:41

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.000	00:41
Breath Volume	2.136	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.000	00:44
Breath Volume	1.968	00:44
Air Blank	0.000	00:44
Control Test	0.079	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALLOUN,SAMUEL,C
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CLOUTIER,AUSTIN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: ELIAS

DATE/TIME OF TEST: 05/15/2023 02:27:17

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.033	02:22
Breath Volume	1.710	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.033	02:25
Breath Volume	2.054	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 06/01/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALLOUN,SAMUEL,C
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: TOWNSEND,T
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MCNAMARA

DATE/TIME OF TEST: 05/15/2023 03:00:58

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.080	02:49	
Air Blank	0.000	02:49	
Subject Sample	# 1 0.044	02:52	Volume Not Met
Breath Volume	0.843	02:52	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample	# 2 0.042	02:56	
Breath Volume	2.441	02:56	
Air Blank	0.000	02:57	
Air Blank	0.000	02:59	
Subject Sample	# 3 0.040	02:59	
Breath Volume	1.886	02:59	
Air Blank	0.000	03:00	
Control Test	0.079	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGOLD,K,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MANGOLD,K
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: LEGGETT

DATE/TIME OF TEST: 05/18/2023 03:40:56

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.119	03:36
Breath Volume	2.488	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:38
Subject Sample # 2	0.117	03:39
Breath Volume	2.816	03:39
Air Blank	0.000	03:40
Control Test	0.077	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 06/01/2023 16:09

Subject Test Electronic Data

DATE/TIME OF TEST: 05/18/2023 21:44:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: RIGGINS,BRENDAN	ARREST TIME: 22:44
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Female
LAST AGENCY INSPECTION: 04/27/2023	OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: CAMACHOFIGUEROA	DATE/TIME OF TEST: 05/21/2023	23:47:37
--------------------------------------	-------------------------------	----------

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.235	23:43
Breath Volume	2.457	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.231	23:46
Breath Volume	2.269	23:46
Air Blank	0.000	23:46
Control Test	0.078	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 05/22/2023 01:46:15

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:40	
Control Test	0.080	01:41	
Air Blank	0.000	01:41	
Subject Sample # 1	0.095	01:42	
Breath Volume	1.707	01:42	
Air Blank	0.000	01:43	
Air Blank	0.000	01:45	
Subject Sample # 2		01:45	Sequence Aborted
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORRE,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 05/22/2023 02:14:22

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.093	02:09
Breath Volume	2.097	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.085	02:12
Breath Volume	1.246	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLER,JOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: SCHEPPS

DATE/TIME OF TEST: 05/24/2023 00:41:07

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.247	00:36
Breath Volume	2.125	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.251	00:39
Breath Volume	2.613	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: JAROUSSE,DANIELLE	ARREST TIME: 17:57
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/27/2023	OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: CLARKSON	DATE/TIME OF TEST: 05/24/2023	18:35:40
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	18:26
Air Blank	0.000	18:27
Control Test	0.079	18:27
Air Blank	0.000	18:28
Subject Sample # 1	0.127	18:29
Breath Volume	1.207	18:29
Air Blank	0.000	18:30
Air Blank	0.000	18:32
Subject Sample # 2	0.127	18:34
Breath Volume	2.468	18:34
Air Blank	0.000	18:34
Control Test	0.079	18:35
Air Blank	0.000	18:35
Diagnostics Check	OK	18:35

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: KRAUSE

DATE/TIME OF TEST: 05/24/2023 23:33:38

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.052	23:29
Breath Volume	1.464	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.051	23:32
Breath Volume	1.128	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYMANS,MORNE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:59

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 05/25/2023 13:34:06

Diagnostics Check	OK	13:26
Air Blank	0.000	13:27
Control Test	0.080	13:27
Air Blank	0.000	13:27
Subject Sample # 1	0.255	13:29
Breath Volume	2.101	13:29
Air Blank	0.000	13:30
Air Blank	0.000	13:31
Subject Sample # 2	0.268	13:32
Breath Volume	2.019	13:32
Air Blank	0.000	13:33
Control Test	0.078	13:33
Air Blank	0.000	13:34
Diagnostics Check	OK	13:34

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: GATCHELL,TYLER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:09

SUBJECT'S LAST NAME: SPINALE

DATE/TIME OF TEST: 05/25/2023 15:37:20

Diagnostics Check	OK	15:31
Air Blank	0.000	15:31
Control Test	0.080	15:31
Air Blank	0.000	15:32
Subject Sample # 1	0.089	15:32
Breath Volume	1.531	15:32
Air Blank	0.000	15:33
Air Blank	0.000	15:35
Subject Sample # 2	0.098	15:35
Breath Volume	2.914	15:35
Air Blank	0.000	15:36
Control Test	0.080	15:36
Air Blank	0.000	15:37
Diagnostics Check	OK	15:37

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: HINES,MIKE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 05/25/2023 21:21:06

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.081	21:15
Air Blank	0.000	21:16
Subject Sample # 1	0.074	21:16
Breath Volume	2.378	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.074	21:19
Breath Volume	2.265	21:19
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: ALMAGUER,BRANDON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 05/25/2023 21:55:07

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.272	21:50
Breath Volume	2.437	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:52
Subject Sample # 2	0.284	21:53
Breath Volume	2.289	21:53
Air Blank	0.000	21:54
Control Test	0.079	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: HOWELL,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/25/2023 23:12:58

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.186	23:08
Breath Volume	3.632	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.195	23:11
Breath Volume	3.570	23:11
Air Blank	0.000	23:12
Control Test	0.079	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: KIDD

DATE/TIME OF TEST: 05/26/2023 00:25:18

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.240	00:19
Breath Volume	1.179	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:21
Subject Sample # 2	0.239	00:23
Breath Volume	1.136	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 06/01/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HUGHES,JARRON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: GROGAN

DATE/TIME OF TEST: 05/27/2023 22:02:49

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.266	21:57
Breath Volume	2.125	21:57
Air Blank	0.000	21:58
Air Blank	0.000	21:59
Subject Sample # 2	0.256	22:01
Breath Volume	1.582	22:01
Air Blank	0.000	22:01
Control Test	0.078	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 06/01/2023 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: NEMECEK

DATE/TIME OF TEST: 05/29/2023 19:15:16

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.079	19:09
Air Blank	0.000	19:10
Subject Sample # 1	0.134	19:10
Breath Volume	1.246	19:10
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample # 2	0.147	19:13
Breath Volume	3.062	19:13
Air Blank	0.000	19:14
Control Test	0.080	19:14
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 06/01/2023 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: MCCANDLESS,MATTHEW

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211543

DRY GAS EXPIRY DATE: 01/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:42

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: CORREA

DATE/TIME OF TEST: 05/29/2023 23:35:12

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.117	23:29
Breath Volume	2.050	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.124	23:33
Breath Volume	1.800	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 06/01/2023 16:10

Subject Test Electronic Data

DATE/TIME OF TEST: 05/30/2023 03:09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: SOLORZANO,DAISY
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: LUBERT

DATE/TIME OF TEST: 05/02/2023 23:04:03

Diagnostics Check	OK	22:59	
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:02	Subject Test Refused
Breath Volume	0.000	23:02	
Air Blank	0.000	23:03	
Control Test	0.079	23:03	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:04	

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PEREZ GODINEZ

DATE/TIME OF TEST: 05/03/2023 22:03:09

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.078	21:58
Breath Volume	2.289	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:00
Subject Sample # 2	0.073	22:01
Breath Volume	1.242	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MAYS,DARRYL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: GASTELBONDO

DATE/TIME OF TEST: 05/06/2023 02:05:16

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.146	02:00
Breath Volume	1.542	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:02
Subject Sample # 2	0.137	02:03
Breath Volume	1.878	02:03
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: GABRIEL VASQUEZ

DATE/TIME OF TEST: 05/10/2023 22:40:00

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.181	22:35
Breath Volume	2.039	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:37
Subject Sample # 2	0.182	22:38
Breath Volume	2.175	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:40

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: AGUILAR,WALTER
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: AVILES VELASQUEZ

DATE/TIME OF TEST: 05/11/2023 20:11:26

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:03	
Control Test	0.080	20:03	
Air Blank	0.000	20:04	
Subject Sample	# 1	20:07	No Sample Provided
Breath Volume	0.000	20:07	
Air Blank	0.000	20:07	
Air Blank	0.000	20:09	
Subject Sample	# 2	20:10	Subject Test Refused
Breath Volume	0.000	20:10	
Air Blank	0.000	20:10	
Control Test	0.079	20:10	
Air Blank	0.000	20:11	
Diagnostics Check	OK	20:11	

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,CARLOS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: JIMENEZ DOMINGO

DATE/TIME OF TEST: 05/18/2023 21:15:19

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:09
Subject Sample # 1	0.177	21:10
Breath Volume	1.851	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:12
Subject Sample # 2	0.170	21:13
Breath Volume	1.683	21:13
Air Blank	0.000	21:14
Control Test	0.078	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUNEZ GODINEZ,SANTOS,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: QUIROGA,DIEGO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: FUNEZ GODINEZ

DATE/TIME OF TEST: 05/20/2023 23:35:28

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.263	23:30
Breath Volume	2.007	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.259	23:33
Breath Volume	3.046	23:33
Air Blank	0.000	23:34
Control Test	0.078	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON, LOUIS, A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROBISON, LOUIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 09:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:41

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 05/22/2023 10:25:14

Diagnostics Check	OK	10:24	
Air Blank	0.000	10:24	
Control Test	0.079	10:24	
Air Blank		10:25	Improper Sample
Air Blank		10:25	Improper Sample

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON, LOUIS, A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROBISON, LOUIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 09:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:41

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 05/22/2023 10:28:47

Diagnostics Check	OK	10:27	
Air Blank		10:28	Ambient Fail
Air Blank	0.000	10:28	

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON, LOUIS, A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROBISON, LOUIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 09:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:41

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 05/22/2023 10:37:59

Diagnostics Check	OK	10:31
Air Blank	0.000	10:31
Control Test	0.078	10:32
Air Blank	0.000	10:32
Subject Sample # 1	0.228	10:33
Breath Volume	1.921	10:33
Air Blank	0.000	10:34
Air Blank	0.000	10:35
Subject Sample # 2	0.219	10:36
Breath Volume	3.625	10:36
Air Blank	0.000	10:37
Control Test	0.077	10:37
Air Blank	0.000	10:37
Diagnostics Check	OK	10:37

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ARGUETA PEREZ

DATE/TIME OF TEST: 05/25/2023 22:48:53

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.210	22:43
Breath Volume	1.597	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.193	22:46
Breath Volume	1.105	22:46
Air Blank	0.000	22:47
Control Test	0.075	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIROGA,DIEGO,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MALDONADO OXLAJ

DATE/TIME OF TEST: 05/27/2023 01:35:01

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.080	01:27	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:30	No Sample Provided
Breath Volume	0.000	01:30	
Air Blank	0.000	01:31	
Air Blank	0.000	01:33	
Subject Sample	# 2	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.081	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROCA,LUIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: VASQUEZ DOMINGO

DATE/TIME OF TEST: 05/28/2023 00:13:55

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.233	00:08
Breath Volume	2.152	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:10
Subject Sample # 2	0.224	00:12
Breath Volume	2.117	00:12
Air Blank	0.000	00:12
Control Test	0.076	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN,SCHELISA,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: SOLORZANO,D
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 05/29/2023 22:21:19

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.208	22:16
Breath Volume	1.855	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.201	22:19
Breath Volume	2.730	22:19
Air Blank	0.000	22:20
Control Test	0.079	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 05/30/2023 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARTINEZ,NICOLE
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/01/2023 04:09:42

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.340	04:04
Breath Volume	1.242	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.331	04:07
Breath Volume	1.898	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 05/03/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: LIGMAN

DATE/TIME OF TEST: 05/01/2023 23:57:55

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.110	23:53
Breath Volume	2.667	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.107	23:56
Breath Volume	2.906	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 05/03/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HERBEL

DATE/TIME OF TEST: 05/02/2023 00:32:02

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.236	00:27
Breath Volume	2.597	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.226	00:30
Breath Volume	1.328	00:30
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:32

Data Download Date/Time: 05/03/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 05/02/2023 04:37:22

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.190	04:32
Breath Volume	3.039	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.187	04:35
Breath Volume	3.085	04:35
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 05/03/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GIL,RAMON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: CHARLESTON

DATE/TIME OF TEST: 05/07/2023 03:13:18

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.221	03:08
Breath Volume	1.726	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.226	03:11
Breath Volume	1.851	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REVOREDO,JULIO
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MAQUEIRA

DATE/TIME OF TEST: 05/08/2023 01:04:54

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:53	
Control Test	0.081	00:53	
Air Blank	0.000	00:53	
Subject Sample	# 1 0.113	00:57	Volume Not Met
Breath Volume	0.703	00:57	
Air Blank	0.000	00:57	
Air Blank	0.000	00:59	
Subject Sample	# 2 0.132	00:59	
Breath Volume	1.351	00:59	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 3 0.146	01:03	
Breath Volume	2.230	01:03	
Air Blank	0.000	01:03	
Control Test	0.079	01:04	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: JOSEPH JR

DATE/TIME OF TEST: 05/08/2023 04:50:15

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.198	04:44
Breath Volume	1.484	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:46
Subject Sample # 2	0.183	04:48
Breath Volume	2.242	04:48
Air Blank	0.000	04:49
Control Test	0.078	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/12/2023 02:03:38

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.082	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.000	01:59
Breath Volume	2.710	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.000	02:02
Breath Volume	1.605	02:02
Air Blank	0.000	02:02
Control Test	0.081	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GUTIERREZ ARIAS

DATE/TIME OF TEST: 05/14/2023 01:45:12

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.193	01:40
Breath Volume	1.675	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.201	01:43
Breath Volume	1.746	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: TENJO CASTRO

DATE/TIME OF TEST: 05/14/2023 03:18:33

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.087	03:13
Breath Volume	3.308	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.088	03:16
Breath Volume	3.109	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RINCON

DATE/TIME OF TEST: 05/14/2023 04:41:18

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.167	04:36
Breath Volume	2.687	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.169	04:39
Breath Volume	2.847	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ENGEL,ANTHONY
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: CYPRESS

DATE/TIME OF TEST: 05/17/2023 20:15:16

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.081	20:09
Air Blank	0.000	20:09
Subject Sample # 1	0.236	20:10
Breath Volume	3.097	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:12
Subject Sample # 2	0.232	20:13
Breath Volume	2.250	20:13
Air Blank	0.000	20:14
Control Test	0.080	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: NECKOWITZ

DATE/TIME OF TEST: 05/17/2023 23:13:46

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.081	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.051	23:09
Breath Volume	1.871	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.051	23:12
Breath Volume	2.085	23:12
Air Blank	0.000	23:12
Control Test	0.080	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HALES,ANDREW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: TOLEDO FILHO

DATE/TIME OF TEST: 05/18/2023 22:19:19

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.081	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.146	22:14
Breath Volume	2.230	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.145	22:17
Breath Volume	2.808	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ANDRADE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: HONORE

DATE/TIME OF TEST: 05/19/2023 01:00:08

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.114	00:55
Breath Volume	1.449	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample # 2	0.115	00:58
Breath Volume	1.484	00:58
Air Blank	0.000	00:59
Control Test	0.078	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALTUNAY,MUHAMMED
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 05/19/2023 02:19:14

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.261	02:10
Breath Volume	1.562	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.239	02:13
Breath Volume	2.238	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 3	0.251	02:17
Breath Volume	1.527	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BRICENO,OSVALDO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:03

SUBJECT'S LAST NAME: NICOLAS

DATE/TIME OF TEST: 05/19/2023 09:40:29

Diagnostics Check	OK	09:33
Air Blank	0.000	09:33
Control Test	0.081	09:33
Air Blank	0.000	09:34
Subject Sample # 1	0.176	09:35
Breath Volume	2.265	09:35
Air Blank	0.000	09:36
Air Blank	0.000	09:38
Subject Sample # 2	0.168	09:38
Breath Volume	2.242	09:38
Air Blank	0.000	09:39
Control Test	0.079	09:39
Air Blank	0.000	09:40
Diagnostics Check	OK	09:40

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BENINCASA,VINCENT
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 05/20/2023 04:47:04

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.086	04:42
Breath Volume	1.398	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 2	0.075	04:45
Breath Volume	1.339	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: SALES DOMINGO

DATE/TIME OF TEST: 05/21/2023 19:10:57

Diagnostics Check	OK	19:03
Air Blank	0.000	19:03
Control Test	0.082	19:04
Air Blank	0.000	19:04
Subject Sample # 1	0.222	19:04
Breath Volume	1.488	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.220	19:09
Breath Volume	2.300	19:09
Air Blank	0.000	19:09
Control Test	0.081	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SIERRA,NICHOLAS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: DELVA

DATE/TIME OF TEST: 05/25/2023 18:29:52

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.081	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.260	18:25
Breath Volume	3.375	18:25
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.257	18:28
Breath Volume	3.175	18:28
Air Blank	0.000	18:28
Control Test	0.079	18:29
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: GALLIGAN

DATE/TIME OF TEST: 05/25/2023 20:03:45

Diagnostics Check	OK	19:56
Air Blank	0.000	19:57
Control Test	0.081	19:57
Air Blank	0.000	19:58
Subject Sample # 1	0.115	19:58
Breath Volume	1.785	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.107	20:02
Breath Volume	1.898	20:02
Air Blank	0.000	20:02
Control Test	0.080	20:03
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RUSSELL,AMON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 05/26/2023 00:23:37

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.081	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.093	00:18
Breath Volume	2.582	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.092	00:21
Breath Volume	2.015	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 06/07/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: EMERENCIANO,MATHEUS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ZULETA RAMIREZ

DATE/TIME OF TEST: 05/26/2023 01:48:00

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.000	01:43
Breath Volume	3.546	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.000	01:46
Breath Volume	3.687	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:48

Data Download Date/Time: 06/07/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/26/2023 02:27:21

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.140	02:22
Breath Volume	1.757	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.137	02:25
Breath Volume	1.574	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 06/07/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ECHARRY GONZALEZ

DATE/TIME OF TEST: 05/28/2023 04:44:32

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.081	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.239	04:39
Breath Volume	2.585	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.240	04:42
Breath Volume	1.671	04:42
Air Blank	0.000	04:43
Control Test	0.077	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 06/07/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: ARGUETA MARTINEZ

DATE/TIME OF TEST: 05/29/2023 00:20:49

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.081	00:15
Air Blank	0.000	00:15
Subject Sample # 1	0.125	00:16
Breath Volume	1.339	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.127	00:19
Breath Volume	1.699	00:19
Air Blank	0.000	00:19
Control Test	0.080	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 06/07/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CALEB,A
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: ELLISON,TIMOTHY
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: BURGETT

DATE/TIME OF TEST: 05/14/2023 16:25:57

Diagnostics Check	OK	16:19
Air Blank	0.000	16:20
Control Test	0.078	16:20
Air Blank	0.000	16:20
Subject Sample # 1	0.210	16:21
Breath Volume	2.511	16:21
Air Blank	0.000	16:21
Air Blank	0.000	16:23
Subject Sample # 2	0.218	16:24
Breath Volume	2.394	16:24
Air Blank	0.000	16:25
Control Test	0.076	16:25
Air Blank	0.000	16:25
Diagnostics Check	OK	16:25

Data Download Date/Time: 06/01/2023 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: VO,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 05/16/2023 19:12:34

Diagnostics Check	OK	19:11	
Air Blank	0.000	19:11	
Control Test	0.000	19:12	RFI Detect
Air Blank	0.000	19:12	

Data Download Date/Time: 06/01/2023 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: VO,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 05/16/2023 19:20:30

Diagnostics Check	OK	19:14
Air Blank	0.000	19:14
Control Test	0.079	19:15
Air Blank	0.000	19:15
Subject Sample # 1	0.000	19:16
Breath Volume	2.550	19:16
Air Blank	0.000	19:16
Air Blank	0.000	19:18
Subject Sample # 2	0.000	19:19
Breath Volume	1.406	19:19
Air Blank	0.000	19:19
Control Test	0.078	19:19
Air Blank	0.000	19:20
Diagnostics Check	OK	19:20

Data Download Date/Time: 06/01/2023 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNSED, DAVID, J
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: GILL, BENJAMIN
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: BUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 05/20/2023 22:10:14

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:58	
Control Test	0.078	21:58	
Air Blank	0.000	21:59	
Subject Sample	# 1 0.206	22:02	Volume Not Met
Breath Volume	0.000	22:02	
Air Blank	0.000	22:03	
Air Blank	0.000	22:05	
Subject Sample	# 2 0.207	22:05	
Breath Volume	2.203	22:05	
Air Blank	0.000	22:06	
Air Blank	0.000	22:08	
Subject Sample	# 3 0.207	22:08	
Breath Volume	2.378	22:08	
Air Blank	0.000	22:09	
Control Test	0.077	22:09	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

Data Download Date/Time: 06/01/2023 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WEBSTER JR

DATE/TIME OF TEST: 05/27/2023 02:52:37

Diagnostics Check	OK	02:52	
Air Blank		02:52	RFI Detect
Air Blank	0.000	02:52	

Data Download Date/Time: 07/01/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 05/27/2023 02:55:35

Diagnostics Check	OK	02:54	
Air Blank		02:55	RFI Detect
Air Blank	0.000	02:55	

Data Download Date/Time: 07/01/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: WEBSTER JR

DATE/TIME OF TEST: 05/27/2023 02:57:45

Diagnostics Check	OK	02:57	
Air Blank		02:57	RFI Detect
Air Blank	0.000	02:57	

Data Download Date/Time: 07/01/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WEBSTER JR

DATE/TIME OF TEST: 05/27/2023 02:59:47

Diagnostics Check	OK	02:59	
Air Blank		02:59	RFI Detect
Air Blank		02:59	RFI Detect

Data Download Date/Time: 07/01/2023 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WEBSTER JR

DATE/TIME OF TEST: 05/27/2023 03:02:05

Diagnostics Check	OK	03:01	
Air Blank		03:01	RFI Detect
Air Blank		03:02	RFI Detect

Data Download Date/Time: 07/01/2023 17:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,DAVIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WEBSTER JR

DATE/TIME OF TEST: 05/27/2023 03:22:53

Diagnostics Check	OK	03:22	
Air Blank		03:22	RFI Detect
Air Blank		03:22	RFI Detect

Data Download Date/Time: 07/01/2023 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WEBSTER JR

DATE/TIME OF TEST: 05/27/2023 03:24:21

Diagnostics Check	OK	03:24	
Air Blank		03:24	RFI Detect
Air Blank		03:24	RFI Detect

Data Download Date/Time: 07/01/2023 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WEBSTER JR

DATE/TIME OF TEST: 05/27/2023 03:32:51

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.200	03:28
Breath Volume	2.746	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.196	03:31
Breath Volume	2.058	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 07/01/2023 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, JOSHUA, H
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HARRISON, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: DEBORD

DATE/TIME OF TEST: 05/28/2023 03:37:55

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.081	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1 0.132	03:33	
Breath Volume	3.410	03:33	
Air Blank	0.000	03:34	
Air Blank	0.000	03:36	
Subject Sample	# 2	03:36	RFI Detect
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Control Test	0.081	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 07/01/2023 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, JOSHUA, H
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HARRISON, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: DEBORD

DATE/TIME OF TEST: 05/28/2023 03:46:17

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.125	03:41
Breath Volume	2.695	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.128	03:44
Breath Volume	3.265	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 07/01/2023 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAN,DONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MORALES MORALES

DATE/TIME OF TEST: 05/20/2023 00:03:25

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.188	23:58
Breath Volume	2.984	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.184	00:01
Breath Volume	2.699	00:01
Air Blank	0.000	00:02
Control Test	0.078	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 06/08/2023 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: MCMILLAN

DATE/TIME OF TEST: 05/26/2023 18:11:07

Diagnostics Check	OK	18:04
Air Blank	0.000	18:04
Control Test	0.080	18:05
Air Blank	0.000	18:05
Subject Sample # 1	0.165	18:06
Breath Volume	4.136	18:06
Air Blank	0.000	18:07
Air Blank	0.000	18:08
Subject Sample # 2	0.163	18:09
Breath Volume	3.980	18:09
Air Blank	0.000	18:10
Control Test	0.079	18:10
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 06/08/2023 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 05/28/2023 02:45:58

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.181	02:41
Breath Volume	2.667	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:43
Subject Sample # 2	0.173	02:44
Breath Volume	2.613	02:44
Air Blank	0.000	02:44
Control Test	0.078	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 06/08/2023 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1482048
BREATH TEST OPERATOR: LYNCH,CHAD,J	DRY GAS EXPIRY DATE: 04/21/2024
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: LYNCH,CHAD	ARREST TIME: 05:11
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Male
LAST AGENCY INSPECTION: 05/16/2023	OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: YOUNG	DATE/TIME OF TEST: 05/29/2023	06:10:40
----------------------------	-------------------------------	----------

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:01	
Control Test	0.079	06:01	
Air Blank	0.000	06:02	
Subject Sample	# 1 0.124	06:05	Volume Not Met
Breath Volume	0.160	06:05	
Air Blank	0.000	06:06	
Air Blank	0.000	06:07	
Subject Sample	# 2	06:08	Subject Test Refused
Breath Volume	0.367	06:08	
Air Blank	0.000	06:09	
Control Test	0.078	06:10	
Air Blank	0.000	06:10	
Diagnostics Check	OK	06:10	

Data Download Date/Time: 06/14/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: RIVERA CRUZ

DATE/TIME OF TEST: 05/13/2023 23:35:58

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.079	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.078	23:31
Breath Volume	2.031	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.086	23:34
Breath Volume	2.234	23:34
Air Blank	0.000	23:35
Control Test	0.079	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 07/05/2023 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENDON,D,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: DORTA MARISCAL

DATE/TIME OF TEST: 05/15/2023 04:14:33

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.118	04:09
Breath Volume	1.519	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.116	04:12
Breath Volume	1.792	04:12
Air Blank	0.000	04:13
Control Test	0.077	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 07/05/2023 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: PHIPPS III

DATE/TIME OF TEST: 05/19/2023 19:17:49

Diagnostics Check	OK	19:07
Air Blank	0.000	19:07
Control Test	0.079	19:08
Air Blank	0.000	19:08
Subject Sample # 1	0.334	19:09
Breath Volume	1.785	19:09
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample # 2	0.365	19:13
Breath Volume	2.535	19:13
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 3	0.385	19:16
Breath Volume	2.492	19:16
Air Blank	0.000	19:16
Control Test	0.078	19:17
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 07/05/2023 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,IVAN,K
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,IVAN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: LOPEZ MARTINEZ

DATE/TIME OF TEST: 05/23/2023 03:21:11

Diagnostics Check	Fail	03:20	Internal Printer Test
Air Blank	0.000	03:21	

Data Download Date/Time: 07/05/2023 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,IVAN,K
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,IVAN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: LOPEZ MARTINEZ

DATE/TIME OF TEST: 05/23/2023 03:30:53

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.183	03:26
Breath Volume	1.750	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.177	03:29
Breath Volume	2.273	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 07/05/2023 09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS, ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: DAVIS JR

DATE/TIME OF TEST: 05/27/2023 00:43:44

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.080	00:38
Breath Volume	2.179	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.081	00:42
Breath Volume	2.597	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 07/05/2023 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENDON,D,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: BOGGS

DATE/TIME OF TEST: 05/27/2023 19:50:29

Diagnostics Check	OK	19:44
Air Blank	0.000	19:44
Control Test	0.079	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.114	19:45
Breath Volume	4.554	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.116	19:48
Breath Volume	4.937	19:48
Air Blank	0.000	19:49
Control Test	0.080	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 07/05/2023 09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKE,B,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: SANCHEZ INFANTE

DATE/TIME OF TEST: 05/28/2023 04:50:23

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.080	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.122	04:45
Breath Volume	2.488	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.118	04:48
Breath Volume	1.519	04:48
Air Blank	0.000	04:49
Control Test	0.081	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 07/05/2023 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENDON,D,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: AREVALOS RIVERO

DATE/TIME OF TEST: 05/28/2023 23:20:28

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.079	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.200	23:16
Breath Volume	1.242	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.217	23:18
Breath Volume	1.714	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 07/05/2023 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: KNIEPS

DATE/TIME OF TEST: 05/29/2023 00:58:28

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.131	00:53
Breath Volume	1.832	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample # 2	0.118	00:56
Breath Volume	1.886	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 07/05/2023 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: CARACCIOLO

DATE/TIME OF TEST: 05/02/2023 02:29:16

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.066	02:24
Breath Volume	2.757	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.063	02:27
Breath Volume	1.800	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 05/02/2023 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BERG,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: BUNN

DATE/TIME OF TEST: 05/28/2023 14:34:25

Diagnostics Check	OK	14:28
Air Blank	0.000	14:28
Control Test	0.082	14:29
Air Blank	0.000	14:29
Subject Sample # 1	0.036	14:30
Breath Volume	1.125	14:30
Air Blank	0.000	14:30
Air Blank	0.000	14:32
Subject Sample # 2	0.037	14:32
Breath Volume	1.328	14:32
Air Blank	0.000	14:33
Control Test	0.081	14:33
Air Blank	0.000	14:34
Diagnostics Check	OK	14:34

Data Download Date/Time: 06/05/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 05/02/2023 23:19:56

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.078	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.114	23:13
Breath Volume	1.417	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:15
Subject Sample # 2	0.110	23:18
Breath Volume	1.250	23:18
Air Blank	0.000	23:18
Control Test	0.080	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 06/05/2023 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: DAURO

DATE/TIME OF TEST: 05/03/2023 04:30:36

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:29	
Control Test	0.074	04:29	Control Outside Tolerance
Air Blank	0.000	04:30	

Data Download Date/Time: 06/05/2023 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: DAURO

DATE/TIME OF TEST: 05/03/2023 04:39:17

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.077	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.122	04:34
Breath Volume	2.027	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:36
Subject Sample # 2	0.121	04:37
Breath Volume	1.597	04:37
Air Blank	0.000	04:38
Control Test	0.078	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 06/05/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ESTRELLA MATOS

DATE/TIME OF TEST: 05/06/2023 03:55:05

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.077	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.143	03:50
Breath Volume	2.308	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:52
Subject Sample # 2	0.134	03:53
Breath Volume	2.015	03:53
Air Blank	0.000	03:54
Control Test	0.077	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 06/05/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SURIEL

DATE/TIME OF TEST: 05/07/2023 02:00:45

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.076	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.204	01:55
Breath Volume	3.062	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.190	01:58
Breath Volume	3.175	01:58
Air Blank	0.000	01:59
Control Test	0.076	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 06/05/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 05/07/2023 22:26:09

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.078	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.315	22:21
Breath Volume	2.718	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:23
Subject Sample # 2	0.307	22:24
Breath Volume	1.949	22:24
Air Blank	0.000	22:25
Control Test	0.075	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 06/05/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: STRUNK

DATE/TIME OF TEST: 05/10/2023 00:31:40

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.078	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.130	00:26
Breath Volume	2.718	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.119	00:29
Breath Volume	1.550	00:29
Air Blank	0.000	00:30
Control Test	0.077	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 06/05/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 05/12/2023 02:20:24

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.077	02:14	
Air Blank	0.000	02:15	
Subject Sample	# 1 0.123	02:15	
Breath Volume	1.691	02:15	
Air Blank	0.000	02:16	
Air Blank	0.000	02:17	
Subject Sample	# 2 0.118	02:18	
Breath Volume	2.312	02:18	
Air Blank	0.000	02:19	
Control Test	0.077	02:19	
Air Blank		02:19	RFI Detect
Air Blank	0.000	02:20	

Data Download Date/Time: 06/05/2023 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MEJIA LOPEZ

DATE/TIME OF TEST: 05/17/2023 01:30:59

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.077	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.061	01:26
Breath Volume	3.214	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.055	01:29
Breath Volume	1.703	01:29
Air Blank	0.000	01:29
Control Test	0.078	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 06/05/2023 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 05/20/2023 03:12:38

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.078	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.107	03:07
Breath Volume	2.535	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.108	03:10
Breath Volume	1.777	03:10
Air Blank	0.000	03:11
Control Test	0.076	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 06/05/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HIDALGO HERNANDEZ

DATE/TIME OF TEST: 05/22/2023 00:02:01

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:48	
Control Test	0.078	23:49	
Air Blank	0.000	23:49	
Subject Sample	# 2 0.087	23:52	Volume Not Met
Breath Volume	0.765	23:52	
Air Blank	0.000	23:53	
Air Blank	0.000	23:55	
Subject Sample	# 3 0.094	23:56	
Breath Volume	1.378	23:56	
Air Blank	0.000	23:57	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.097	00:00	
Breath Volume	1.253	00:00	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 06/05/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA, MARCANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/26/2023 03:32:09

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.076	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.223	03:27
Breath Volume	2.414	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.235	03:30
Breath Volume	1.878	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 06/05/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT, JUSTIN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PERDOMO ARRIECHE

DATE/TIME OF TEST: 05/27/2023 22:00:58

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.077	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.168	21:56
Breath Volume	3.113	21:56
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.168	21:59
Breath Volume	2.906	21:59
Air Blank	0.000	21:59
Control Test	0.079	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 06/05/2023 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: MONTOYO CHAVEZ

DATE/TIME OF TEST: 05/01/2023 00:57:15

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.081	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.131	00:52
Breath Volume	2.398	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.127	00:55
Breath Volume	2.164	00:55
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 05/02/2023 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: DILLARD

DATE/TIME OF TEST: 05/01/2023 04:20:03

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.081	04:08	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:11	No Sample Provided
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Air Blank	0.000	04:14	
Subject Sample	# 2	0.164	04:15
Breath Volume	1.636	04:15	
Air Blank	0.000	04:15	
Air Blank	0.000	04:17	
Subject Sample	# 3	0.170	04:18
Breath Volume	1.593	04:18	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 05/02/2023 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 05/03/2023 03:45:32

Diagnostics Check	Fail	03:45	Internal Printer Test
Air Blank	0.000	03:45	

Data Download Date/Time: 06/13/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 05/03/2023 03:54:27

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.082	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.247	03:49
Breath Volume	2.050	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.254	03:52
Breath Volume	2.445	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 06/13/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: BOND

DATE/TIME OF TEST: 05/17/2023 23:27:43

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.165	23:22
Breath Volume	1.234	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:24
Subject Sample # 2	0.152	23:26
Breath Volume	1.171	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 06/13/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BEAVER

DATE/TIME OF TEST: 05/19/2023 03:43:33

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.056	03:38
Breath Volume	2.781	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.051	03:42
Breath Volume	1.515	03:42
Air Blank	0.000	03:42
Control Test	0.082	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 06/13/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 05/01/2023 05:19:59

Diagnostics Check	OK	05:12
Air Blank	0.000	05:13
Control Test	0.080	05:13
Air Blank	0.000	05:14
Subject Sample # 1	0.161	05:15
Breath Volume	1.605	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:17
Subject Sample # 2	0.162	05:18
Breath Volume	1.734	05:18
Air Blank	0.000	05:19
Control Test	0.080	05:19
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 05/02/2023 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/17/2023 04:41:35

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.000	04:37
Breath Volume	2.226	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.000	04:40
Breath Volume	1.902	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 06/05/2023 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 05/19/2023 02:53:18

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.000	02:48
Breath Volume	1.757	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.000	02:51
Breath Volume	1.812	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 06/05/2023 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: HAWKS

DATE/TIME OF TEST: 05/25/2023 02:14:28

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.071	02:09
Breath Volume	2.640	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.069	02:12
Breath Volume	1.605	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 06/05/2023 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/28/2023 03:36:04

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.121	03:31
Breath Volume	1.929	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.126	03:34
Breath Volume	1.945	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:36

Data Download Date/Time: 06/05/2023 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: TOCCO

DATE/TIME OF TEST: 05/28/2023 05:38:52

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.079	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.113	05:34
Breath Volume	2.281	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:36
Subject Sample # 2	0.113	05:37
Breath Volume	2.152	05:37
Air Blank	0.000	05:37
Control Test	0.080	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 06/05/2023 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 05/03/2023 21:38:28

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:21	
Control Test	0.080	21:22	
Air Blank	0.000	21:22	
Subject Sample	# 1 0.172	21:25	
Breath Volume	1.410	21:25	
Air Blank	0.000	21:26	
Air Blank	0.000	21:27	
Subject Sample	# 2 0.125	21:30	Volume Not Met
Breath Volume	0.019	21:30	
Air Blank	0.000	21:31	
Air Blank	0.000	21:33	
Subject Sample	# 3 0.153	21:36	Volume Not Met
Breath Volume	0.253	21:36	
Air Blank	0.000	21:37	
Control Test	0.074	21:37	Control Outside Tolerance
Air Blank	0.000	21:38	

Data Download Date/Time: 05/25/2023 01:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JACOB,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: LOWERY,JOHN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ENGELS

DATE/TIME OF TEST: 05/12/2023 23:09:00

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.081	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.132	23:03
Breath Volume	1.167	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.128	23:07
Breath Volume	1.187	23:07
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:09

Data Download Date/Time: 05/25/2023 01:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEDESCH,LAURA,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: DIEDESCH,LAURA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 05/20/2023 03:44:58

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.079	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1 0.140	03:37	Volume Not Met
Breath Volume	0.140	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 2 0.133	03:43	Volume Not Met
Breath Volume	0.273	03:43	
Air Blank	0.000	03:44	
Control Test	0.077	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 05/25/2023 01:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,ROBERT,D
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: AUTREY,TREVOR
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: STEMPLE

DATE/TIME OF TEST: 05/27/2023 22:12:48

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:10	
Control Test	0.080	22:10	
Air Blank	0.000	22:11	
Subject Sample	# 1	22:11	Subject Test Refused
Breath Volume	0.000	22:11	
Air Blank	0.000	22:11	
Control Test	0.080	22:12	
Air Blank	0.000	22:12	
Diagnostics Check	OK	22:12	

Data Download Date/Time: 06/14/2023 04:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBE,BRIAN,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: SCOTT,ROBERT
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MAKI

DATE/TIME OF TEST: 05/28/2023 04:17:29

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.248	04:12
Breath Volume	3.679	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.244	04:15
Breath Volume	1.882	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 06/14/2023 04:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PETTIFER,JAMES,P

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILLIAMS,CLINT

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 15:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: DUNN JR'

DATE/TIME OF TEST: 05/29/2023 16:22:40

Diagnostics Check	OK	16:15
Air Blank	0.000	16:16
Control Test	0.078	16:16
Air Blank	0.000	16:17
Subject Sample # 1	0.120	16:17
Breath Volume	2.703	16:17
Air Blank	0.000	16:18
Air Blank	0.000	16:20
Subject Sample # 2	0.116	16:20
Breath Volume	3.921	16:20
Air Blank	0.000	16:21
Control Test	0.078	16:22
Air Blank	0.000	16:22
Diagnostics Check	OK	16:22

Data Download Date/Time: 06/20/2023 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PETTIFER,JAMES,P

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILKINS,MALAKI

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 15:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: VANVAERENBERGH

DATE/TIME OF TEST: 05/29/2023 16:33:55

Diagnostics Check	OK	16:27
Air Blank	0.000	16:27
Control Test	0.078	16:27
Air Blank	0.000	16:28
Subject Sample # 1	0.101	16:29
Breath Volume	1.628	16:29
Air Blank	0.000	16:29
Air Blank	0.000	16:31
Subject Sample # 2	0.105	16:32
Breath Volume	1.792	16:32
Air Blank	0.000	16:32
Control Test	0.077	16:33
Air Blank	0.000	16:33
Diagnostics Check	OK	16:33

Data Download Date/Time: 06/20/2023 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 15922080A4
BREATH TEST OPERATOR: PETTIFER,JAMES,P	DRY GAS EXPIRY DATE: 08/05/2024
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)	VIOLATION CODE: DUI
ARRESTING OFFICER: WILSON,NICK	ARREST TIME: 16:48
ARRESTING OFFICER AGENCY: Tampa PD	GENDER: Male
LAST AGENCY INSPECTION: 05/23/2023	OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: CRISTESCU	DATE/TIME OF TEST: 05/29/2023	17:46:29
--------------------------------	-------------------------------	----------

Diagnostics Check	OK	17:40
Air Blank	0.000	17:40
Control Test	0.078	17:41
Air Blank	0.000	17:41
Subject Sample # 1	0.000	17:42
Breath Volume	3.183	17:42
Air Blank	0.000	17:42
Air Blank	0.000	17:44
Subject Sample # 2	0.000	17:45
Breath Volume	3.589	17:45
Air Blank	0.000	17:45
Control Test	0.078	17:45
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

Data Download Date/Time: 06/20/2023 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PROUTY,EDWARD
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: BUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: CHADBOURNE

DATE/TIME OF TEST: 05/29/2023 19:56:29

Diagnostics Check	OK	19:49	
Air Blank	0.000	19:49	
Control Test	0.078	19:50	
Air Blank	0.000	19:50	
Subject Sample	# 1 0.226	19:51	
Breath Volume	1.632	19:51	
Air Blank	0.000	19:51	
Air Blank	0.000	19:53	
Subject Sample	# 2 0.229	19:54	
Breath Volume	1.562	19:54	
Air Blank	0.000	19:54	
Control Test	0.078	19:55	
Air Blank	0.000	19:55	
Diagnostics Check	Fail	19:56	Internal Printer Test
Air Blank	0.000	19:56	

Data Download Date/Time: 06/20/2023 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ, OFC B,

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: 1515, HOPPE

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 15:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: LECUYER

DATE/TIME OF TEST: 05/28/2023 15:50:51

Diagnostics Check	OK	15:43
Air Blank	0.000	15:44
Control Test	0.081	15:44
Air Blank	0.000	15:45
Subject Sample # 1	0.051	15:45
Breath Volume	1.410	15:45
Air Blank	0.000	15:46
Air Blank	0.000	15:48
Subject Sample # 2	0.048	15:49
Breath Volume	1.558	15:49
Air Blank	0.000	15:49
Control Test	0.080	15:50
Air Blank	0.000	15:50
Diagnostics Check	OK	15:50

Data Download Date/Time: 06/20/2023 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALVIS,REECE,H

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: ALVIS,REECE

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: BUI

ARREST TIME: 18:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 05/28/2023 18:35:07

Diagnostics Check	OK	18:32	
Air Blank	0.000	18:32	
Control Test	0.081	18:32	
Air Blank	0.000	18:33	
Subject Sample	# 1	18:33	RFI Detect
Breath Volume	0.000	18:33	
Air Blank	0.000	18:34	
Control Test	0.080	18:34	
Air Blank		18:34	RFI Detect
Air Blank	0.000	18:35	

Data Download Date/Time: 06/20/2023 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALVIS,REECE,H

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: ALVIS,REECE

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 05/28/2023 18:47:13

Diagnostics Check	OK	18:45	
Air Blank	0.000	18:46	
Control Test	0.080	18:46	
Air Blank		18:46	RFI Detect
Air Blank	0.000	18:47	

Data Download Date/Time: 06/20/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 15421080A1
BREATH TEST OPERATOR: ALVIS,REECE,H	DRY GAS EXPIRY DATE: 08/05/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)	VIOLATION CODE: BUI
ARRESTING OFFICER: ALVIS,REECE	ARREST TIME: 18:00
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)	GENDER: Male
LAST AGENCY INSPECTION: 05/05/2023	OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: GORDON	DATE/TIME OF TEST: 05/28/2023 18:50:24
-----------------------------	--

Diagnostics Check	OK	18:49	
Air Blank		18:49	RFI Detect
Air Blank	0.000	18:50	

Data Download Date/Time: 06/20/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PETTIFER,JAMES,P

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: ALVIS,REECE

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: DUI

ARREST TIME: 18:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 05/28/2023 18:56:54

Diagnostics Check	OK	18:52	
Air Blank	0.000	18:53	
Control Test	0.080	18:53	
Air Blank	0.000	18:54	
Subject Sample	# 1 0.125	18:55	
Breath Volume	3.601	18:55	
Air Blank		18:55	RFI Detect
Air Blank	0.000	18:55	
Control Test	0.080	18:56	
Air Blank	0.000	18:56	
Diagnostics Check	OK	18:56	

Data Download Date/Time: 06/20/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 15421080A1
BREATH TEST OPERATOR: PETTIFER',JAMES,P	DRY GAS EXPIRY DATE: 08/05/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)	VIOLATION CODE: BUI
ARRESTING OFFICER: ALVIS,REECE	ARREST TIME: 18:00
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)	GENDER: Male
LAST AGENCY INSPECTION: 05/05/2023	OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: GORDON	DATE/TIME OF TEST: 05/28/2023 19:05:07
-----------------------------	--

Diagnostics Check	OK	18:58
Air Blank	0.000	18:59
Control Test	0.080	18:59
Air Blank	0.000	19:00
Subject Sample # 1	0.112	19:00
Breath Volume	1.582	19:00
Air Blank	0.000	19:01
Air Blank	0.000	19:02
Subject Sample # 2	0.126	19:03
Breath Volume	2.839	19:03
Air Blank	0.000	19:04
Control Test	0.080	19:04
Air Blank	0.000	19:04
Diagnostics Check	OK	19:05

Data Download Date/Time: 06/20/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ,BENNETT,

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: RODRIGUEZ,OFC B

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: HILDABRAND

DATE/TIME OF TEST: 05/28/2023 19:25:30

Diagnostics Check	OK	19:14	
Air Blank	0.000	19:15	
Control Test	0.080	19:15	
Air Blank	0.000	19:15	
Subject Sample	# 1 0.143	19:19	Volume Not Met
Breath Volume	0.203	19:19	
Air Blank	0.000	19:19	
Air Blank	0.000	19:21	
Subject Sample	# 2 0.137	19:24	Volume Not Met
Breath Volume	0.707	19:24	
Air Blank		19:24	RFI Detect
Air Blank	0.000	19:25	

Data Download Date/Time: 06/20/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,OFC B,
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: RODRIGUEZ,OFC B
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: HILDABRAND

DATE/TIME OF TEST: 05/28/2023 19:30:08

Diagnostics Check	Fail	19:29	Internal Printer Test
Air Blank	0.000	19:30	

Data Download Date/Time: 06/20/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ, OFC B,

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: 1845, RODRIGUEZ

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: HILDABRAND

DATE/TIME OF TEST: 05/28/2023 19:32:45

Diagnostics Check

Fail

19:32

Internal Printer Test

Air Blank

0.000

19:32

Data Download Date/Time: 06/20/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ, OFC B,

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: RODRIGUEZ, OFC B

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: HILDABRAND

DATE/TIME OF TEST: 05/28/2023 19:40:28

Diagnostics Check	OK	19:37	
Air Blank	0.000	19:37	
Control Test	0.080	19:38	
Air Blank	0.000	19:38	
Subject Sample	# 1	19:39	RFI Detect
Breath Volume	0.000	19:39	
Air Blank	0.000	19:39	
Control Test	0.080	19:39	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 06/20/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ,OFC B,

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILLIAMS,OFC C

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 05/05/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: EITEL

DATE/TIME OF TEST: 05/28/2023 19:42:52

Diagnostics Check

Fail

19:42

Internal Printer Test

Air Blank

0.000

19:42

Data Download Date/Time: 06/20/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAWRENCE,JOSHUA,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: LAWRENCE,JOSHUA

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG129201

DRY GAS EXPIRY DATE: 10/19/2023

VIOLATION CODE: BUI

ARREST TIME: 18:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/13/2023 19:56:25

Diagnostics Check

Fail

19:55

Internal Printer Test

Air Blank

0.000

19:56

Data Download Date/Time: 05/22/2023 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAWRENCE,JOSHUA,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: LAWRENCE,JOSHUA

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 04/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG129201

DRY GAS EXPIRY DATE: 10/19/2023

VIOLATION CODE: BUI

ARREST TIME: 18:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/13/2023 20:04:55

Diagnostics Check	OK	19:58
Air Blank	0.000	19:58
Control Test	0.079	19:59
Air Blank	0.000	19:59
Subject Sample # 1	0.079	20:00
Breath Volume	1.988	20:00
Air Blank	0.000	20:00
Air Blank	0.000	20:02
Subject Sample # 2	0.070	20:03
Breath Volume	1.867	20:03
Air Blank	0.000	20:03
Control Test	0.079	20:04
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 05/22/2023 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAWRENCE,JOSHUA,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: THOMAS,MEGAN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG310302

DRY GAS EXPIRY DATE: 04/13/2025

VIOLATION CODE: BUI

ARREST TIME: 19:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: SHNEIDER

DATE/TIME OF TEST: 05/27/2023 20:26:07

Diagnostics Check	OK	20:23	
Air Blank	0.000	20:23	
Control Test	0.080	20:23	
Air Blank	0.000	20:24	
Subject Sample	# 1	20:24	Subject Test Refused
Breath Volume	0.000	20:24	
Air Blank	0.000	20:25	
Control Test	0.082	20:25	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:26	

Data Download Date/Time: 06/20/2023 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NORTH,DUANE,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: THOMAS,MEGAN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 15:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:13

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 05/06/2023 16:37:32

Diagnostics Check

Fail

16:37

Internal Printer Test

Air Blank

0.000

16:37

Data Download Date/Time: 05/16/2023 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NORTH,DUANE,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: THOMAS,MEGAN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 15:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:13

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 05/06/2023 16:46:51

Diagnostics Check	OK	16:40
Air Blank	0.000	16:40
Control Test	0.082	16:40
Air Blank	0.000	16:41
Subject Sample # 1	0.126	16:42
Breath Volume	2.179	16:42
Air Blank	0.000	16:43
Air Blank	0.000	16:44
Subject Sample # 2	0.117	16:45
Breath Volume	2.488	16:45
Air Blank	0.000	16:45
Control Test	0.081	16:46
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

Data Download Date/Time: 05/16/2023 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NORTH,DUANE,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: MARROQUIN,RICHARD

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 15:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:56

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 05/06/2023 16:58:07

Diagnostics Check	OK	16:51
Air Blank	0.000	16:51
Control Test	0.082	16:52
Air Blank	0.000	16:52
Subject Sample # 1	0.000	16:53
Breath Volume	2.468	16:53
Air Blank	0.000	16:54
Air Blank	0.000	16:56
Subject Sample # 2	0.000	16:56
Breath Volume	1.371	16:56
Air Blank	0.000	16:57
Control Test	0.081	16:57
Air Blank	0.000	16:58
Diagnostics Check	OK	16:58

Data Download Date/Time: 05/16/2023 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NORTH,DUANE,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: EDSON,MAXWELL

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 17:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: HEAD

DATE/TIME OF TEST: 05/06/2023 17:55:59

Diagnostics Check	OK	17:49	
Air Blank	0.000	17:50	
Control Test	0.081	17:50	
Air Blank	0.000	17:50	
Subject Sample	# 1 0.000	17:51	
Breath Volume	3.796	17:51	
Air Blank	0.000	17:51	
Air Blank	0.000	17:53	
Subject Sample	# 2 0.000	17:54	
Breath Volume	3.078	17:54	
Air Blank	0.000	17:54	
Control Test	0.081	17:55	
Air Blank		17:55	RFI Detect
Air Blank	0.000	17:55	

Data Download Date/Time: 05/16/2023 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NORTH,DUANE,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: CHRISTMAS,FOLIN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 05/06/2023 18:51:12

Diagnostics Check	OK	18:50	
Air Blank		18:51	RFI Detect
Air Blank		18:51	RFI Detect

Data Download Date/Time: 05/16/2023 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NORTH,DUANE,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: CHRISTMAS,FOLIN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 05/06/2023 18:56:19

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:53	
Control Test	0.082	18:54	
Air Blank	0.000	18:54	
Subject Sample	# 1	18:54	Subject Test Refused
Breath Volume	0.000	18:54	
Air Blank	0.000	18:55	
Control Test	0.082	18:55	
Air Blank	0.000	18:56	
Diagnostics Check	OK	18:56	

Data Download Date/Time: 05/16/2023 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: GORE,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/07/2023 18:08:46

Diagnostics Check	OK	18:02	
Air Blank	0.000	18:02	
Control Test	0.077	18:03	
Air Blank	0.000	18:03	
Subject Sample	# 1 0.122	18:04	
Breath Volume	2.785	18:04	
Air Blank	0.000	18:05	
Air Blank	0.000	18:07	
Subject Sample	# 2 0.119	18:07	
Breath Volume	2.324	18:07	
Air Blank		18:08	Ambient Fail
Air Blank	0.000	18:08	

Data Download Date/Time: 05/15/2023 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVER, WILLIAM, H
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: OLIVER, WILLIAM
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/20/2023 17:55:08

Diagnostics Check	OK	17:48
Air Blank	0.000	17:49
Control Test	0.076	17:49
Air Blank	0.000	17:49
Subject Sample # 1	0.050	17:50
Breath Volume	2.539	17:50
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 2	0.052	17:53
Breath Volume	1.164	17:53
Air Blank	0.000	17:54
Control Test	0.076	17:54
Air Blank	0.000	17:55
Diagnostics Check	OK	17:55

Data Download Date/Time: 06/16/2023 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWER,RACHEL,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BOWER,RACHEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/20/2023 19:10:30

Diagnostics Check	OK	19:09	
Air Blank		19:10	RFI Detect
Air Blank	0.000	19:10	

Data Download Date/Time: 06/16/2023 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWER,RACHEL,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BOWER,RACHEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/20/2023 19:20:00

Diagnostics Check	OK	19:14
Air Blank	0.000	19:14
Control Test	0.077	19:14
Air Blank	0.000	19:15
Subject Sample # 1	0.144	19:15
Breath Volume	3.316	19:15
Air Blank	0.000	19:16
Air Blank	0.000	19:18
Subject Sample # 2	0.136	19:18
Breath Volume	3.683	19:18
Air Blank	0.000	19:19
Control Test	0.076	19:19
Air Blank	0.000	19:19
Diagnostics Check	OK	19:20

Data Download Date/Time: 06/16/2023 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURSCH,THEODORE,R
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MURSCH,THEODORE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 05/20/2023 19:30:24

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:27	
Control Test	0.076	19:28	
Air Blank	0.000	19:28	
Subject Sample	# 1	19:29	Subject Test Refused
Breath Volume	0.000	19:29	
Air Blank	0.000	19:29	
Control Test	0.076	19:29	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:30	

Data Download Date/Time: 06/16/2023 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORE,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: GORE,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 05/27/2023 16:25:27

Diagnostics Check	Fail	16:24	Internal Printer Test
Air Blank	0.000	16:25	

Data Download Date/Time: 06/16/2023 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORE,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: GORE,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 05/27/2023 16:35:11

Diagnostics Check	OK	16:27
Air Blank	0.000	16:28
Control Test	0.078	16:28
Air Blank	0.000	16:29
Subject Sample # 1	0.229	16:29
Breath Volume	2.203	16:29
Air Blank	0.000	16:30
Air Blank	0.000	16:32
Subject Sample # 2	0.226	16:33
Breath Volume	2.585	16:33
Air Blank	0.000	16:34
Control Test	0.078	16:34
Air Blank	0.000	16:35
Diagnostics Check	OK	16:35

Data Download Date/Time: 06/16/2023 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD, TRAVIS, L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: JARRELL, CHRISTIAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: BYERS

DATE/TIME OF TEST: 05/27/2023 17:54:36

Diagnostics Check	OK	17:51	
Air Blank	0.000	17:52	
Control Test	0.077	17:52	
Air Blank	0.000	17:52	
Subject Sample	# 1	17:53	RFI Detect
Breath Volume	0.000	17:53	
Air Blank	0.000	17:53	
Control Test	0.077	17:53	
Air Blank	0.000	17:54	
Diagnostics Check	OK	17:54	

Data Download Date/Time: 06/16/2023 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 15421080A1
BREATH TEST OPERATOR: BASFORD,TRAVIS,L	DRY GAS EXPIRY DATE: 08/05/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: JARRELL,CHRISTIAN	ARREST TIME: 17:20
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/15/2023	OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: BYERS	DATE/TIME OF TEST: 05/27/2023	18:00:11
----------------------------	-------------------------------	----------

Diagnostics Check	OK	17:56	
Air Blank	0.000	17:57	
Control Test	0.077	17:57	
Air Blank	0.000	17:58	
Subject Sample	# 1	17:58	Subject Test Refused
Breath Volume	0.000	17:58	
Air Blank	0.000	17:58	
Control Test	0.077	17:59	
Air Blank		17:59	Ambient Fail
Air Blank	0.000	18:00	

Data Download Date/Time: 06/16/2023 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD, TRAVIS, L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RICE, ZACH
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 05/28/2023 18:46:41

Diagnostics Check	OK	18:39
Air Blank	0.000	18:39
Control Test	0.078	18:40
Air Blank	0.000	18:40
Subject Sample # 1	0.155	18:42
Breath Volume	3.949	18:42
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.162	18:45
Breath Volume	3.421	18:45
Air Blank	0.000	18:45
Control Test	0.078	18:46
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 06/16/2023 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUSTIN,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: HAUGE,CAMERON
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 15:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 05/28/2023 16:20:55

Diagnostics Check	OK	16:15	
Air Blank	0.000	16:15	
Control Test	0.078	16:16	
Air Blank	0.000	16:16	
Subject Sample	# 1	16:19	Subject Test Refused
Breath Volume	0.000	16:19	
Air Blank	0.000	16:19	
Control Test	0.080	16:20	
Air Blank	0.000	16:20	
Diagnostics Check	OK	16:20	

Data Download Date/Time: 06/23/2023 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILMORE,RICHARD,L

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: AUSTIN,NICHOLAS

ARRESTING OFFICER AGENCY: FFWCC (Lake City)

LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: LONGSTRETH

DATE/TIME OF TEST: 05/28/2023 19:48:47

Diagnostics Check	OK	19:45	
Air Blank	0.000	19:46	
Control Test	0.078	19:46	
Air Blank	0.000	19:47	
Subject Sample	# 1	19:47	Subject Test Refused
Breath Volume	0.000	19:47	
Air Blank	0.000	19:47	
Control Test	0.078	19:48	
Air Blank	0.000	19:48	
Diagnostics Check	OK	19:48	

Data Download Date/Time: 06/23/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JOVANOVA, MICHAEL, N

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: JOVANOVA, MICHAEL

ARRESTING OFFICER AGENCY: FFWCC (Lake City)

LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 16:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: COCHRANE

DATE/TIME OF TEST: 05/27/2023 17:11:00

Diagnostics Check	OK	17:04
Air Blank	0.000	17:04
Control Test	0.079	17:05
Air Blank	0.000	17:05
Subject Sample # 1	0.050	17:06
Breath Volume	2.972	17:06
Air Blank	0.000	17:07
Air Blank	0.000	17:08
Subject Sample # 2	0.050	17:09
Breath Volume	1.480	17:09
Air Blank	0.000	17:10
Control Test	0.081	17:10
Air Blank	0.000	17:10
Diagnostics Check	OK	17:11

Data Download Date/Time: 06/23/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,THOMAS,B
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: REID,THOMAS
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: BLOCH

DATE/TIME OF TEST: 05/27/2023 18:17:44

Diagnostics Check	Fail	18:17	Internal Printer Test
Air Blank	0.000	18:17	

Data Download Date/Time: 06/23/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,MICHAEL,
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: PIETROBURGO,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: GRAU

DATE/TIME OF TEST: 05/28/2023 18:16:22

Diagnostics Check	OK	18:09
Air Blank	0.000	18:10
Control Test	0.081	18:10
Air Blank	0.000	18:11
Subject Sample # 1	0.136	18:11
Breath Volume	2.011	18:11
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.132	18:14
Breath Volume	1.976	18:14
Air Blank	0.000	18:15
Control Test	0.081	18:15
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 06/23/2023 12:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,CALEB,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALMERTON,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/06/2023 04:11:06

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.082	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Control Test	0.080	04:10	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNELL,C,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SNELL,C
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: FOHLBROOK

DATE/TIME OF TEST: 05/06/2023 09:59:37

Diagnostics Check	OK	09:53
Air Blank	0.000	09:53
Control Test	0.081	09:54
Air Blank	0.000	09:54
Subject Sample # 1	0.000	09:55
Breath Volume	3.847	09:55
Air Blank	0.000	09:55
Air Blank	0.000	09:57
Subject Sample # 2	0.000	09:58
Breath Volume	3.371	09:58
Air Blank	0.000	09:58
Control Test	0.080	09:58
Air Blank	0.000	09:59
Diagnostics Check	OK	09:59

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECERRA,J
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HERNANDEZ CRUZ

DATE/TIME OF TEST: 05/06/2023 22:49:47

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.081	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.232	22:44
Breath Volume	2.140	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.236	22:47
Breath Volume	2.082	22:47
Air Blank	0.000	22:48
Control Test	0.078	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRETT,R
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:52

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 05/08/2023 13:16:28

Diagnostics Check	OK	13:13	
Air Blank	0.000	13:13	
Control Test	0.081	13:14	
Air Blank	0.000	13:14	
Subject Sample	# 1	13:15	Subject Test Refused
Breath Volume	0.000	13:15	
Air Blank	0.000	13:15	
Control Test	0.081	13:15	
Air Blank	0.000	13:16	
Diagnostics Check	OK	13:16	

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GILLIARD,S
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SOWARDS

DATE/TIME OF TEST: 05/09/2023 02:25:11

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:22	
Control Test	0.082	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:23	Subject Test Refused
Breath Volume	0.000	02:23	
Air Blank	0.000	02:24	
Control Test	0.081	02:24	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ALLRED,C
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MASTEN

DATE/TIME OF TEST: 05/10/2023 02:12:28

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.124	02:07
Breath Volume	1.320	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:09
Subject Sample # 2	0.120	02:10
Breath Volume	1.246	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: FORBES,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 05/11/2023 19:51:49

Diagnostics Check	OK	19:44
Air Blank	0.000	19:45
Control Test	0.081	19:45
Air Blank	0.000	19:45
Subject Sample # 1	0.000	19:46
Breath Volume	1.671	19:46
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.000	19:50
Breath Volume	1.964	19:50
Air Blank	0.000	19:50
Control Test	0.081	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,C,H
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROBERTS,P
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: FRANKELE

DATE/TIME OF TEST: 05/11/2023 22:17:47

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:10	
Control Test	0.081	22:11	
Air Blank	0.000	22:11	
Subject Sample # 1	0.140	22:12	
Breath Volume	3.132	22:12	
Air Blank	0.000	22:13	
Air Blank	0.000	22:15	
Subject Sample # 2		22:16	Subject Test Refused
Breath Volume	0.074	22:16	
Air Blank	0.000	22:16	
Control Test	0.080	22:17	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:17	

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Sebring PD
ARRESTING OFFICER: PANTOJA,YOSMANY
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MABINS

DATE/TIME OF TEST: 05/13/2023 03:21:50

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.252	03:17
Breath Volume	2.703	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:19
Subject Sample # 2	0.235	03:20
Breath Volume	2.621	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,J,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ALLRED,C
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:32

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/13/2023 07:00:11

Diagnostics Check	OK	06:52
Air Blank	0.000	06:52
Control Test	0.081	06:53
Air Blank	0.000	06:53
Subject Sample # 1	0.238	06:54
Breath Volume	1.332	06:54
Air Blank	0.000	06:55
Air Blank	0.000	06:57
Subject Sample # 2	0.221	06:58
Breath Volume	1.167	06:58
Air Blank	0.000	06:59
Control Test	0.081	06:59
Air Blank	0.000	07:00
Diagnostics Check	OK	07:00

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DETTMAN,A
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: PEREZ VERDUGO

DATE/TIME OF TEST: 05/14/2023 03:00:53

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:58	
Control Test	0.081	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.081	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PANTOJA,YOSMANY
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: TORRES JIMENEZ

DATE/TIME OF TEST: 05/14/2023 03:28:32

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: KAMAU,TARIG
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 05/14/2023 04:02:41

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.222	03:57
Breath Volume	1.359	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.234	04:00
Breath Volume	1.464	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNELL,C,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SNELL,C
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 09:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: HERIBERTO

DATE/TIME OF TEST: 05/16/2023 09:51:28

Diagnostics Check	OK	09:45
Air Blank	0.000	09:45
Control Test	0.081	09:46
Air Blank	0.000	09:46
Subject Sample # 1	0.000	09:47
Breath Volume	2.656	09:47
Air Blank	0.000	09:47
Air Blank	0.000	09:49
Subject Sample # 2	0.000	09:50
Breath Volume	2.816	09:50
Air Blank	0.000	09:50
Control Test	0.080	09:50
Air Blank	0.000	09:51
Diagnostics Check	OK	09:51

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARTIN,BRIAN
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: KIGER

DATE/TIME OF TEST: 05/17/2023 09:54:18

Diagnostics Check	OK	09:50	
Air Blank	0.000	09:51	
Control Test	0.081	09:51	
Air Blank	0.000	09:52	
Subject Sample	# 1	09:52	Subject Test Refused
Breath Volume	0.000	09:52	
Air Blank	0.000	09:53	
Control Test	0.080	09:53	
Air Blank	0.000	09:54	
Diagnostics Check	OK	09:54	

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: TAYLOR,AUSTIN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 05/20/2023 20:13:13

Diagnostics Check	OK	20:06
Air Blank	0.000	20:06
Control Test	0.081	20:06
Air Blank	0.000	20:07
Subject Sample # 1	0.187	20:07
Breath Volume	1.863	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.187	20:11
Breath Volume	2.046	20:11
Air Blank	0.000	20:12
Control Test	0.080	20:12
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,A,K
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BUTLER,A
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: ZERO TOL
ARREST TIME: 12:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 05/25/2023 13:15:29

Diagnostics Check	OK	13:09
Air Blank	0.000	13:10
Control Test	0.081	13:10
Air Blank	0.000	13:10
Subject Sample # 1	0.000	13:11
Breath Volume	2.027	13:11
Air Blank	0.000	13:11
Air Blank	0.000	13:13
Subject Sample # 2	0.000	13:14
Breath Volume	2.078	13:14
Air Blank	0.000	13:14
Control Test	0.080	13:14
Air Blank	0.000	13:15
Diagnostics Check	OK	13:15

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,REBECCA,C
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,R
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: KIRCHMAN

DATE/TIME OF TEST: 05/25/2023 17:26:33

Diagnostics Check	OK	17:23	
Air Blank	0.000	17:23	
Control Test	0.080	17:24	
Air Blank	0.000	17:24	
Subject Sample	# 1	17:25	Subject Test Refused
Breath Volume	0.000	17:25	
Air Blank	0.000	17:25	
Control Test	0.081	17:25	
Air Blank	0.000	17:26	
Diagnostics Check	OK	17:26	

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,CALEB,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: RYBACKI-SUER

DATE/TIME OF TEST: 05/29/2023 23:58:30

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.081	23:56	
Air Blank	0.000	23:56	
Subject Sample	# 1	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:57	
Control Test	0.080	23:57	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALMERTON,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: GOMEZ BADILLO

DATE/TIME OF TEST: 05/30/2023 19:08:28

Diagnostics Check	OK	19:01
Air Blank	0.000	19:02
Control Test	0.080	19:02
Air Blank	0.000	19:03
Subject Sample # 1	0.164	19:03
Breath Volume	2.785	19:03
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.160	19:06
Breath Volume	2.867	19:06
Air Blank	0.000	19:07
Control Test	0.079	19:07
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

Data Download Date/Time: 06/13/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,ERICK,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: LAGUNAS,JEWEL
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/05/2023 14:29:21

Diagnostics Check	OK	14:17
Air Blank	0.000	14:17
Control Test	0.080	14:17
Air Blank	0.000	14:18
Subject Sample # 1	0.315	14:19
Breath Volume	1.960	14:19
Air Blank	0.000	14:20
Air Blank	0.000	14:22
Subject Sample # 2	0.279	14:24
Breath Volume	1.511	14:24
Air Blank	0.000	14:25
Air Blank	0.000	14:27
Subject Sample # 3	0.294	14:27
Breath Volume	1.289	14:27
Air Blank	0.000	14:28
Control Test	0.079	14:28
Air Blank	0.000	14:29
Diagnostics Check	OK	14:29

Data Download Date/Time: 06/28/2023 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,ERICK,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: KOBS,ANDREW
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: GRINE

DATE/TIME OF TEST: 05/29/2023 20:53:14

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.037	20:48
Breath Volume	2.191	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.031	20:51
Breath Volume	1.750	20:51
Air Blank	0.000	20:52
Control Test	0.079	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 06/28/2023 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: ABRAHAM

DATE/TIME OF TEST: 05/01/2023 20:21:49

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.078	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.000	20:17
Breath Volume	2.449	20:17
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.000	20:20
Breath Volume	2.726	20:20
Air Blank	0.000	20:20
Control Test	0.077	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 06/01/2023 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 05/21/2023 17:46:30

Diagnostics Check	Fail	17:46	Internal Printer Test
Air Blank	0.000	17:46	

Data Download Date/Time: 06/01/2023 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/21/2023 22:33:26

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.077	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.000	22:28
Breath Volume	1.425	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.000	22:32
Breath Volume	1.785	22:32
Air Blank	0.000	22:32
Control Test	0.077	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 06/01/2023 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,D
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: GANDHI

DATE/TIME OF TEST: 05/31/2023 22:12:33

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.078	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.000	22:08
Breath Volume	2.210	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.000	22:11
Breath Volume	2.460	22:11
Air Blank	0.000	22:11
Control Test	0.077	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 06/01/2023 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: SERAFIN

DATE/TIME OF TEST: 05/31/2023 22:55:09

Diagnostics Check	Fail	22:54	Internal Printer Test
Air Blank	0.000	22:55	

Data Download Date/Time: 06/01/2023 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: SERAFIN

DATE/TIME OF TEST: 05/31/2023 23:04:36

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.078	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.040	22:58
Breath Volume	1.773	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.041	23:03
Breath Volume	1.289	23:03
Air Blank	0.000	23:03
Control Test	0.078	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 06/01/2023 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO	ARREST TIME: 02:28
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Male
LAST AGENCY INSPECTION: 04/13/2023	OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MOJENA	DATE/TIME OF TEST: 05/14/2023 03:21:25
-----------------------------	--

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.078	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.078	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 05/31/2023 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ACOSTA,CAROLINA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: UTRERA CEDENO

DATE/TIME OF TEST: 05/20/2023 03:04:47

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.146	02:59
Breath Volume	2.425	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample # 2	0.138	03:03
Breath Volume	2.382	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 05/31/2023 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PENA,CHRISTOPHER,L

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: FERNANDEZ,ANTHONY

ARRESTING OFFICER AGENCY: Coral Gables PD

LAST AGENCY INSPECTION: 04/13/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 01:34

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GADDY

DATE/TIME OF TEST: 05/28/2023 03:15:49

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:13	
Control Test	0.078	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:14	
Control Test	0.080	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 05/31/2023 11:13