

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAPELA,JORDAN,A

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: CAPELA,JORDAN

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 05/17/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 03:22

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HOLLOMAN

DATE/TIME OF TEST: 06/01/2023 03:37:13

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.055	03:32
Breath Volume	1.933	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.059	03:35
Breath Volume	1.742	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 06/01/2023 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,MATTHEW,M

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: SMITH,MATTHEW

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 06/04/2023 02:46:02

Diagnostics Check	OK	02:44	
Air Blank		02:45	Ambient Fail
Air Blank	0.000	02:46	

Data Download Date/Time: 07/18/2023 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,MATTHEW,M

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: SMITH,MATTHEW

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 06/04/2023 02:49:35

Diagnostics Check	OK	02:47	
Air Blank		02:48	Ambient Fail
Air Blank	0.000	02:49	

Data Download Date/Time: 07/18/2023 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,MATTHEW,M

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: SMITH,MATTHEW

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 06/04/2023 02:53:01

Diagnostics Check	OK	02:51	
Air Blank		02:52	Ambient Fail
Air Blank		02:53	Purge Fail

Data Download Date/Time: 07/18/2023 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,MATTHEW,M

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: SMITH,MATTHEW

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 06/04/2023 02:57:01

Diagnostics Check	OK	02:55	
Air Blank		02:56	Ambient Fail
Air Blank	0.000	02:57	

Data Download Date/Time: 07/18/2023 14:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORZO,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: JOHNSTON,S
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: WELLER

DATE/TIME OF TEST: 06/04/2023 02:11:50

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.252	02:06
Breath Volume	1.566	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.250	02:10
Breath Volume	2.410	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 06/26/2023 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBIN,BRITTON,L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BOWER,RACHEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: DUNFORD

DATE/TIME OF TEST: 06/04/2023 19:05:03

Diagnostics Check	OK	18:56	
Air Blank	0.000	18:57	
Control Test	0.080	18:57	
Air Blank	0.000	18:57	
Subject Sample	# 1	19:01	No Sample Provided
Breath Volume	0.000	19:01	
Air Blank	0.000	19:01	
Air Blank	0.000	19:03	
Subject Sample	# 2	19:03	Subject Test Refused
Breath Volume	0.000	19:03	
Air Blank	0.000	19:04	
Control Test	0.080	19:04	
Air Blank	0.000	19:04	
Diagnostics Check	OK	19:05	

Data Download Date/Time: 06/07/2023 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORMLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GORMLEY, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: GROSHONG

DATE/TIME OF TEST: 06/10/2023 15:31:54

Diagnostics Check	OK	15:25
Air Blank	0.000	15:25
Control Test	0.081	15:26
Air Blank	0.000	15:26
Subject Sample # 1	0.285	15:27
Breath Volume	3.460	15:27
Air Blank	0.000	15:27
Air Blank	0.000	15:29
Subject Sample # 2	0.285	15:30
Breath Volume	3.546	15:30
Air Blank	0.000	15:30
Control Test	0.082	15:31
Air Blank	0.000	15:31
Diagnostics Check	OK	15:31

Data Download Date/Time: 07/05/2023 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLTON,JUSTIN,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOLTON,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/11/2023 19:30:19

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.081	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.323	19:22
Breath Volume	1.414	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.302	19:25
Breath Volume	1.878	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 3	0.312	19:28
Breath Volume	1.460	19:28
Air Blank	0.000	19:29
Control Test	0.079	19:29
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 07/05/2023 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BURT

DATE/TIME OF TEST: 06/18/2023 05:18:33

Diagnostics Check	OK	05:10
Air Blank	0.000	05:11
Control Test	0.081	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.129	05:13
Breath Volume	1.941	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.123	05:16
Breath Volume	1.738	05:16
Air Blank	0.000	05:17
Control Test	0.078	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 07/05/2023 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/18/2023 21:07:40

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:05	
Control Test	0.082	21:05	
Air Blank	0.000	21:06	
Subject Sample	# 1	21:06	Subject Test Refused
Breath Volume	0.000	21:06	
Air Blank	0.000	21:06	
Control Test	0.082	21:07	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time: 07/05/2023 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 06/07/2023
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: BAUMAN

DATE/TIME OF TEST: 06/20/2023 00:08:31

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.081	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.153	00:03
Breath Volume	2.820	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.153	00:06
Breath Volume	2.843	00:06
Air Blank	0.000	00:07
Control Test	0.081	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 07/05/2023 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,JAVON,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAEZ,JORDI
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 06/08/2023 05:04:44

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.279	04:59
Breath Volume	1.593	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.281	05:03
Breath Volume	1.230	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 07/11/2023 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BURKS,LEXI
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 06/10/2023 19:38:53

Diagnostics Check	OK	19:32
Air Blank	0.000	19:32
Control Test	0.080	19:33
Air Blank	0.000	19:33
Subject Sample # 1	0.216	19:34
Breath Volume	1.339	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:36
Subject Sample # 2	0.223	19:37
Breath Volume	1.636	19:37
Air Blank	0.000	19:37
Control Test	0.080	19:38
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 07/11/2023 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABDALE,JAMES,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: CARROLL,FRANK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: DENEGRI

DATE/TIME OF TEST: 06/15/2023 13:48:27

Diagnostics Check	OK	13:42
Air Blank	0.000	13:43
Control Test	0.080	13:43
Air Blank	0.000	13:43
Subject Sample # 1	0.000	13:44
Breath Volume	1.785	13:44
Air Blank	0.000	13:44
Air Blank	0.000	13:46
Subject Sample # 2	0.000	13:47
Breath Volume	2.437	13:47
Air Blank	0.000	13:47
Control Test	0.080	13:47
Air Blank	0.000	13:48
Diagnostics Check	OK	13:48

Data Download Date/Time: 07/11/2023 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MILLER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/16/2023 04:14:38

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:09	
Control Test	0.080	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 1 0.187	04:13	Volume Not Met
Breath Volume	0.195	04:13	
Air Blank		04:14	Ambient Fail
Air Blank	0.000	04:14	

Data Download Date/Time: 07/11/2023 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MILLER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/16/2023 04:23:33

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.206	04:18
Breath Volume	1.230	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.210	04:21
Breath Volume	1.394	04:21
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 07/11/2023 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRANDALL,DOUGLAS,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CRANDALL,DOUGLAS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MIRANDA CUEVAS

DATE/TIME OF TEST: 06/18/2023 03:45:32

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.214	03:39
Breath Volume	1.347	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.211	03:43
Breath Volume	1.726	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 07/11/2023 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: FLAHERTY,ZACHARY
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 06/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 06/19/2023 02:18:19

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.079	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:11	No Sample Provided
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2	02:16	No Sample Provided
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Control Test	0.081	02:17	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 07/11/2023 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRABHAM,CHARLES,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BRABHAM,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 06/24/2023 00:57:14

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.080	00:52	
Air Blank	0.000	00:52	
Subject Sample	# 1 0.161	00:53	
Breath Volume	1.824	00:53	
Air Blank	0.000	00:53	
Air Blank	0.000	00:55	
Subject Sample	# 2 0.149	00:56	
Breath Volume	2.175	00:56	
Air Blank		00:56	RFI Detect
Air Blank	0.000	00:57	

Data Download Date/Time: 07/11/2023 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINS,JAMES,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: MCCARY

DATE/TIME OF TEST: 06/04/2023 17:41:42

Diagnostics Check	OK	17:35
Air Blank	0.000	17:35
Control Test	0.079	17:36
Air Blank	0.000	17:36
Subject Sample # 1	0.000	17:37
Breath Volume	2.937	17:37
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 2	0.000	17:40
Breath Volume	3.492	17:40
Air Blank	0.000	17:40
Control Test	0.080	17:41
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

Data Download Date/Time: 06/29/2023 19:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: SELVIDGE

DATE/TIME OF TEST: 06/06/2023 17:35:47

Diagnostics Check	OK	17:35	
Air Blank		17:35	RFI Detect
Air Blank	0.000	17:35	

Data Download Date/Time: 06/29/2023 19:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: SELVIDGE

DATE/TIME OF TEST: 06/06/2023 17:43:40

Diagnostics Check	OK	17:37
Air Blank	0.000	17:38
Control Test	0.080	17:38
Air Blank	0.000	17:39
Subject Sample # 1	0.000	17:39
Breath Volume	2.554	17:39
Air Blank	0.000	17:39
Air Blank	0.000	17:41
Subject Sample # 2	0.000	17:42
Breath Volume	3.218	17:42
Air Blank	0.000	17:42
Control Test	0.080	17:43
Air Blank	0.000	17:43
Diagnostics Check	OK	17:43

Data Download Date/Time: 06/29/2023 19:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITCHARD, THOMAS, D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: PRITCHARD, THOMAS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 06/19/2023 01:38:05

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.178	01:33
Breath Volume	4.531	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:35
Subject Sample # 2	0.179	01:36
Breath Volume	3.609	01:36
Air Blank	0.000	01:37
Control Test	0.077	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 06/29/2023 19:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO, WESTON, C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HOWARD, ELLIOTT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ELHEMAMY

DATE/TIME OF TEST: 06/24/2023 02:45:59

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.080	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:45	
Control Test	0.079	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 06/29/2023 19:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEYES,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GEORGE,AMANDA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 13:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:11

SUBJECT'S LAST NAME: LANGLEY

DATE/TIME OF TEST: 06/26/2023 13:43:25

Diagnostics Check	OK	13:40	
Air Blank	0.000	13:41	
Control Test	0.079	13:41	
Air Blank	0.000	13:41	
Subject Sample	# 1	13:42	Subject Test Refused
Breath Volume	0.000	13:42	
Air Blank	0.000	13:42	
Control Test	0.079	13:42	
Air Blank	0.000	13:43	
Diagnostics Check	OK	13:43	

Data Download Date/Time: 06/29/2023 19:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELAN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PHELAN,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: ALLBRITTON

DATE/TIME OF TEST: 06/10/2023 03:50:30

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.161	03:45
Breath Volume	4.132	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.162	03:48
Breath Volume	4.093	03:48
Air Blank	0.000	03:49
Control Test	0.077	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 06/25/2023 00:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 06/11/2023 03:37:21

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.080	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:36	RFI Detect
Breath Volume	0.000	03:36	
Air Blank		03:36	RFI Detect
Air Blank	0.000	03:37	

Data Download Date/Time: 06/25/2023 00:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 06/11/2023 03:45:20

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.078	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.195	03:40
Breath Volume	2.601	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:42
Subject Sample # 2	0.194	03:43
Breath Volume	3.335	03:43
Air Blank	0.000	03:44
Control Test	0.078	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 06/25/2023 00:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: VOUGHT

DATE/TIME OF TEST: 06/24/2023 03:54:49

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.101	03:50
Breath Volume	4.761	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.098	03:53
Breath Volume	4.433	03:53
Air Blank	0.000	03:53
Control Test	0.078	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 06/25/2023 00:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: GAMBLE

DATE/TIME OF TEST: 06/25/2023 01:56:24

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.078	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.178	01:51
Breath Volume	1.738	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample # 2	0.177	01:54
Breath Volume	1.132	01:54
Air Blank	0.000	01:55
Control Test	0.078	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARKE,GREGORY,A
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: HODGES,NICHOLAS
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 07:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: MCCLUSKEY

DATE/TIME OF TEST: 06/26/2023 08:34:10

Diagnostics Check	OK	08:27
Air Blank	0.000	08:27
Control Test	0.080	08:28
Air Blank	0.000	08:28
Subject Sample # 1	0.273	08:29
Breath Volume	3.011	08:29
Air Blank	0.000	08:30
Air Blank	0.000	08:31
Subject Sample # 2	0.272	08:32
Breath Volume	2.074	08:32
Air Blank	0.000	08:33
Control Test	0.078	08:33
Air Blank	0.000	08:34
Diagnostics Check	OK	08:34

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON, TYLER, L
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: HARMON, TYLER
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:21

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 06/29/2023 13:12:44

Diagnostics Check	OK	13:12	
Air Blank		13:12	RFI Detect
Air Blank	0.000	13:12	

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON, TYLER, L
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: HARMON, TYLER
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:21

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 06/29/2023 13:20:41

Diagnostics Check	OK	13:14
Air Blank	0.000	13:14
Control Test	0.079	13:15
Air Blank	0.000	13:15
Subject Sample # 1	0.017	13:16
Breath Volume	1.769	13:16
Air Blank	0.000	13:16
Air Blank	0.000	13:18
Subject Sample # 2	0.015	13:19
Breath Volume	2.773	13:19
Air Blank	0.000	13:19
Control Test	0.078	13:20
Air Blank	0.000	13:20
Diagnostics Check	OK	13:20

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLHARD,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VILLHARD,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:18

SUBJECT'S LAST NAME: MARQUEZ BURGOS

DATE/TIME OF TEST: 06/07/2023 14:39:04

Diagnostics Check	OK	14:38	
Air Blank		14:38	RFI Detect
Air Blank	0.000	14:39	

Data Download Date/Time: 07/05/2023 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLHARD,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VILLHARD,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:18

SUBJECT'S LAST NAME: MARQUEZ BURGOS

DATE/TIME OF TEST: 06/07/2023 14:42:04

Diagnostics Check	OK	14:41	
Air Blank		14:41	RFI Detect
Air Blank	0.000	14:42	

Data Download Date/Time: 07/05/2023 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLHARD,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VILLHARD,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:18

SUBJECT'S LAST NAME: MARQUEZ BURGOS

DATE/TIME OF TEST: 06/07/2023 14:59:58

Diagnostics Check	OK	14:53	
Air Blank	0.000	14:54	
Control Test	0.081	14:54	
Air Blank	0.000	14:55	
Subject Sample	# 1 0.099	14:58	Volume Not Met
Breath Volume	0.015	14:58	
Air Blank		14:59	Purge Fail
Air Blank		14:59	Purge Fail

Data Download Date/Time: 07/05/2023 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLHARD,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VILLHARD,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:18

SUBJECT'S LAST NAME: MARQUEZ BURGOS

DATE/TIME OF TEST: 06/07/2023 15:03:23

Diagnostics Check	OK	15:02	
Air Blank		15:02	Ambient Fail
Air Blank		15:03	Purge Fail

Data Download Date/Time: 07/05/2023 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLHARD,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VILLHARD,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:18

SUBJECT'S LAST NAME: MARQUEZ BURGOS

DATE/TIME OF TEST: 06/07/2023 15:25:52

Diagnostics Check	OK	15:24	
Air Blank		15:25	Ambient Fail
Air Blank	0.000	15:25	

Data Download Date/Time: 07/05/2023 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLHARD,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VILLHARD,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:18

SUBJECT'S LAST NAME: MARQUEZ BURGOS

DATE/TIME OF TEST: 06/07/2023 17:38:15

Diagnostics Check	OK	17:36	
Air Blank	0.000	17:37	
Control Test	0.062	17:37	Interferent Detect
Air Blank	0.000	17:38	

Data Download Date/Time: 07/05/2023 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, TYLER, M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WYCHUNAS, REIDE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 06/08/2023 21:28:10

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:27	
Control Test	0.073	21:27	Control Outside Tolerance
Air Blank	0.000	21:28	

Data Download Date/Time: 07/05/2023 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, TYLER, M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WYCHUNAS, REIDE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 06/08/2023 21:32:15

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:31	
Control Test	0.074	21:31	Control Outside Tolerance
Air Blank	0.000	21:32	

Data Download Date/Time: 07/05/2023 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LORD,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: ONG

DATE/TIME OF TEST: 06/04/2023 14:47:50

Diagnostics Check	OK	14:41
Air Blank	0.000	14:42
Control Test	0.081	14:42
Air Blank	0.000	14:42
Subject Sample # 1	0.000	14:43
Breath Volume	1.769	14:43
Air Blank	0.000	14:43
Air Blank	0.000	14:45
Subject Sample # 2	0.000	14:46
Breath Volume	2.441	14:46
Air Blank	0.000	14:46
Control Test	0.078	14:47
Air Blank	0.000	14:47
Diagnostics Check	OK	14:47

Data Download Date/Time: 07/11/2023 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEMMONS,LAQUINTA,M
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PLEMMONS,LAQUINTA
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: GASKINS

DATE/TIME OF TEST: 06/05/2023 00:02:58

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.079	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 2 0.148	23:58	
Breath Volume	2.175	23:58	
Air Blank	0.000	23:58	
Air Blank	0.000	00:00	
Subject Sample	# 1 0.161	00:00	
Breath Volume	3.238	00:00	
Air Blank	0.000	00:01	
Control Test	0.078	00:02	
Air Blank		00:02	RFI Detect
Air Blank	0.000	00:02	

Data Download Date/Time: 07/11/2023 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 06/09/2023 22:16:40

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.287	22:11
Breath Volume	1.585	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.290	22:14
Breath Volume	1.500	22:14
Air Blank	0.000	22:15
Control Test	0.077	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 07/11/2023 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MARKMILLER,JOHN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 06/10/2023 01:49:45

Diagnostics Check	OK	01:49	
Air Blank		01:49	RFI Detect
Air Blank	0.000	01:49	

Data Download Date/Time: 07/11/2023 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MARKMILLER,JOHN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 06/10/2023 02:00:59

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.080	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1 0.123	01:56	Volume Not Met
Breath Volume	0.281	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 2	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	02:00	
Control Test	0.079	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 07/11/2023 07:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: EDOUARD,DONALD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/10/2023 07:54:16

Diagnostics Check	OK	07:47
Air Blank	0.000	07:48
Control Test	0.080	07:48
Air Blank	0.000	07:49
Subject Sample # 1	0.136	07:49
Breath Volume	3.675	07:49
Air Blank	0.000	07:50
Air Blank	0.000	07:52
Subject Sample # 2	0.134	07:52
Breath Volume	3.226	07:52
Air Blank	0.000	07:53
Control Test	0.078	07:53
Air Blank	0.000	07:54
Diagnostics Check	OK	07:54

Data Download Date/Time: 07/11/2023 07:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRO,MARCEL,
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: ELLIOT,NIGEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 06/13/2023 23:02:49

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.172	22:58
Breath Volume	3.128	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.172	23:01
Breath Volume	3.218	23:01
Air Blank	0.000	23:01
Control Test	0.078	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 07/11/2023 07:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,JAMES,R
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CASON,JAMES
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: EMERY

DATE/TIME OF TEST: 06/23/2023 23:48:03

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.234	23:43
Breath Volume	1.175	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:45
Subject Sample # 2	0.236	23:46
Breath Volume	1.507	23:46
Air Blank	0.000	23:47
Control Test	0.078	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:48

Data Download Date/Time: 07/11/2023 07:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 06/24/2023 12:39:49

Diagnostics Check	OK	12:33
Air Blank	0.000	12:33
Control Test	0.079	12:34
Air Blank	0.000	12:34
Subject Sample # 1	0.276	12:35
Breath Volume	2.914	12:35
Air Blank	0.000	12:35
Air Blank	0.000	12:37
Subject Sample # 2	0.274	12:38
Breath Volume	2.539	12:38
Air Blank	0.000	12:38
Control Test	0.078	12:39
Air Blank	0.000	12:39
Diagnostics Check	OK	12:39

Data Download Date/Time: 07/11/2023 07:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: PASLEY,JACOB
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024023281
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: POSEY-HUCKABEE

DATE/TIME OF TEST: 06/06/2023 00:38:33

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.082	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.146	00:34
Breath Volume	1.687	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.146	00:36
Breath Volume	1.320	00:36
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 06/21/2023 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/04/2023 23:54:01

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.083	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.155	23:49
Breath Volume	1.652	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:51
Subject Sample # 2	0.162	23:52
Breath Volume	1.800	23:52
Air Blank	0.000	23:53
Control Test	0.081	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:54

Data Download Date/Time: 06/20/2023 15:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,BRETT,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WAGNER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CLIFTON

DATE/TIME OF TEST: 06/10/2023 02:53:44

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.082	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.082	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 06/20/2023 15:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 06/23/2023 03:17:14

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.082	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.096	03:12
Breath Volume	1.871	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.092	03:15
Breath Volume	1.574	03:15
Air Blank	0.000	03:16
Control Test	0.082	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 07/03/2023 19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHERS, WILLIAM, A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MATHERS, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: WINTER

DATE/TIME OF TEST: 06/24/2023 05:28:46

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:26	
Control Test	0.082	05:26	
Air Blank	0.000	05:27	
Subject Sample	# 1 0.178	05:27	
Breath Volume	1.191	05:27	
Air Blank		05:28	RFI Detect
Air Blank		05:28	RFI Detect
Air Blank		05:28	RFI Detect

Data Download Date/Time: 07/03/2023 19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHERS, WILLIAM, A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MATHERS, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: WINTER

DATE/TIME OF TEST: 06/24/2023 05:38:38

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.080	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.170	05:33
Breath Volume	1.710	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.170	05:36
Breath Volume	1.140	05:36
Air Blank	0.000	05:37
Control Test	0.078	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 07/03/2023 19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CAROS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 06/27/2023 04:50:46

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.082	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1 0.264	04:49	
Breath Volume	1.703	04:49	
Air Blank		04:50	Purge Fail
Air Blank	0.000	04:50	

Data Download Date/Time: 07/03/2023 19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 06/27/2023 05:00:07

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.076	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.240	04:55
Breath Volume	1.355	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.220	04:58
Breath Volume	1.503	04:58
Air Blank	0.000	04:59
Control Test	0.076	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 07/03/2023 19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RADCLIFF, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: COPEs

DATE/TIME OF TEST: 06/28/2023 22:32:26

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:19	
Control Test	0.082	22:19	
Air Blank	0.000	22:20	
Subject Sample	# 1 0.218	22:22	
Breath Volume	2.011	22:22	
Air Blank	0.000	22:23	
Air Blank	0.000	22:24	
Subject Sample	# 2 0.207	22:27	Volume Not Met
Breath Volume	0.863	22:27	
Air Blank	0.000	22:28	
Air Blank	0.000	22:30	
Subject Sample	# 3 0.206	22:30	
Breath Volume	1.886	22:30	
Air Blank	0.000	22:31	
Control Test	0.074	22:31	Control Outside Tolerance
Air Blank	0.000	22:32	

Data Download Date/Time: 07/03/2023 19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RADCLIFF, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: COPEs

DATE/TIME OF TEST: 06/28/2023 23:03:07

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:55	
Control Test	0.077	22:55	
Air Blank	0.000	22:56	
Subject Sample	# 1 0.210	22:56	
Breath Volume	1.207	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 2 0.197	23:01	
Breath Volume	1.621	23:01	
Air Blank		23:02	Purge Fail
Air Blank	0.000	23:03	

Data Download Date/Time: 07/03/2023 19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Holmes County SO
ARRESTING OFFICER: WELCH,WAYLON
ARRESTING OFFICER AGENCY: Holmes County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BROWNELL

DATE/TIME OF TEST: 06/22/2023 20:29:24

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:21	
Control Test	0.080	20:21	
Air Blank	0.000	20:22	
Subject Sample	# 1	20:25	No Sample Provided
Breath Volume	0.000	20:25	
Air Blank	0.000	20:25	
Air Blank	0.000	20:27	
Subject Sample	# 2	20:28	Subject Test Refused
Breath Volume	0.000	20:28	
Air Blank	0.000	20:28	
Control Test	0.080	20:28	
Air Blank	0.000	20:29	
Diagnostics Check	OK	20:29	

Data Download Date/Time: 07/11/2023 18:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,DALLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: FAUCETT

DATE/TIME OF TEST: 06/26/2023 08:34:09

Diagnostics Check	OK	08:30	
Air Blank	0.000	08:30	
Control Test	0.082	08:30	
Air Blank	0.000	08:31	
Subject Sample	# 1	08:32	Subject Test Refused
Breath Volume	0.000	08:32	
Air Blank	0.000	08:33	
Control Test	0.081	08:33	
Air Blank	0.000	08:34	
Diagnostics Check	OK	08:34	

Data Download Date/Time: 07/11/2023 18:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,DONALD,R
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SELLERS,DONALD
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BETTS

DATE/TIME OF TEST: 06/01/2023 02:14:08

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.081	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.086	02:07
Breath Volume	1.238	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.084	02:12
Breath Volume	1.253	02:12
Air Blank	0.000	02:13
Control Test	0.082	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 06/05/2023 03:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,CAMERON,D
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: HART,DYLAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 06/22/2023 02:28:30

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.025	02:24
Breath Volume	2.453	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.023	02:27
Breath Volume	1.949	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 07/07/2023 00:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,DONALD,R
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SELLERS,DONALD
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/04/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GREGG

DATE/TIME OF TEST: 06/23/2023 21:46:37

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.080	21:41
Air Blank	0.000	21:41
Subject Sample # 1	0.156	21:42
Breath Volume	1.761	21:42
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.154	21:45
Breath Volume	1.554	21:45
Air Blank	0.000	21:45
Control Test	0.080	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 07/07/2023 00:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,DONALD,R
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SELLERS,DONALD
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 06/25/2023 23:25:11

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.079	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1 0.131	23:23	
Breath Volume	1.312	23:23	
Air Blank		23:24	Purge Fail
Air Blank	0.000	23:25	

Data Download Date/Time: 07/07/2023 00:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,DONALD,R
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SELLERS,DONALD
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 06/25/2023 23:36:55

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.079	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.131	23:30
Breath Volume	1.398	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:32
Subject Sample # 2	0.133	23:35
Breath Volume	1.578	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 07/07/2023 00:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,DYLAN,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: HART,DYLAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: ILLICH

DATE/TIME OF TEST: 06/26/2023 19:04:34

Diagnostics Check	OK	18:57
Air Blank	0.000	18:58
Control Test	0.080	18:58
Air Blank	0.000	18:59
Subject Sample # 1	0.000	18:59
Breath Volume	1.300	18:59
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample # 2	0.000	19:03
Breath Volume	1.519	19:03
Air Blank	0.000	19:03
Control Test	0.080	19:03
Air Blank	0.000	19:04
Diagnostics Check	OK	19:04

Data Download Date/Time: 07/07/2023 00:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRENNON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SABATINE,MATTHEW
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 06/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 06/30/2023 14:36:39

Diagnostics Check	OK	14:29
Air Blank	0.000	14:30
Control Test	0.079	14:30
Air Blank	0.000	14:30
Subject Sample # 1	0.037	14:32
Breath Volume	1.945	14:32
Air Blank	0.000	14:32
Air Blank	0.000	14:34
Subject Sample # 2	0.035	14:35
Breath Volume	2.445	14:35
Air Blank	0.000	14:35
Control Test	0.080	14:36
Air Blank	0.000	14:36
Diagnostics Check	OK	14:36

Data Download Date/Time: 07/07/2023 00:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: FRANK

DATE/TIME OF TEST: 06/03/2023 04:01:42

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.085	03:57
Breath Volume	2.191	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.081	04:00
Breath Volume	3.113	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 06/28/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILKEY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GILKEY,JOSHUA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 06/04/2023 17:32:22

Diagnostics Check	OK	17:25	
Air Blank	0.000	17:26	
Control Test	0.082	17:26	
Air Blank	0.000	17:26	
Subject Sample	# 1 0.292	17:27	
Breath Volume	3.019	17:27	
Air Blank	0.000	17:28	
Air Blank	0.000	17:29	
Subject Sample	# 2 0.296	17:30	
Breath Volume	2.800	17:30	
Air Blank	0.000	17:31	
Control Test	0.080	17:31	
Air Blank		17:31	RFI Detect
Air Blank	0.000	17:32	

Data Download Date/Time: 06/28/2023 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAKE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: MCKELVEY

DATE/TIME OF TEST: 06/08/2023 20:41:49

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.082	20:36
Air Blank	0.000	20:36
Subject Sample # 1	0.169	20:37
Breath Volume	3.277	20:37
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.156	20:40
Breath Volume	3.058	20:40
Air Blank	0.000	20:40
Control Test	0.080	20:41
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 06/28/2023 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTLE,ALICIA,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CASTLE,ALICIA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: BUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 06/08/2023 21:22:19

Diagnostics Check	OK	21:15
Air Blank	0.000	21:16
Control Test	0.081	21:16
Air Blank	0.000	21:17
Subject Sample # 1	0.161	21:17
Breath Volume	3.808	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.158	21:20
Breath Volume	3.101	21:20
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 06/28/2023 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAKE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: BLOCKER

DATE/TIME OF TEST: 06/08/2023 23:55:41

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.081	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.159	23:47
Breath Volume	2.347	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample # 2	0.137	23:50
Breath Volume	1.355	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 3	0.147	23:54
Breath Volume	1.925	23:54
Air Blank	0.000	23:54
Control Test	0.079	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 06/28/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MOOTY

DATE/TIME OF TEST: 06/10/2023 23:27:30

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.082	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.184	23:22
Breath Volume	1.902	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:24
Subject Sample # 2	0.174	23:25
Breath Volume	1.296	23:25
Air Blank	0.000	23:26
Control Test	0.078	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 06/28/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPUTO,TAYLOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: CAPUTO,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/11/2023 00:20:05

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.082	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.226	00:15
Breath Volume	1.898	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:17
Subject Sample # 2	0.221	00:18
Breath Volume	1.781	00:18
Air Blank	0.000	00:19
Control Test	0.077	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 06/28/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 06/12/2023 02:42:47

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.081	02:38	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:40	RFI Detect
Breath Volume	0.488	02:40	
Air Blank	0.000	02:41	
Control Test	0.079	02:42	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 06/28/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 06/12/2023 02:51:16

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.078	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.201	02:46
Breath Volume	2.105	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample # 2	0.194	02:49
Breath Volume	2.808	02:49
Air Blank	0.000	02:50
Control Test	0.075	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 06/28/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAKE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: VASQUEZ RAMIREZ

DATE/TIME OF TEST: 06/16/2023 23:49:59

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.081	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.237	23:45
Breath Volume	2.292	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:47
Subject Sample # 2	0.233	23:48
Breath Volume	1.917	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 06/28/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAKE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: LOPEZ GARCIA

DATE/TIME OF TEST: 06/17/2023 21:52:00

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.082	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.217	21:47
Breath Volume	2.375	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.221	21:50
Breath Volume	2.339	21:50
Air Blank	0.000	21:51
Control Test	0.081	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:52

Data Download Date/Time: 06/28/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,CASANDRA,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MAGURA,AUSTIN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MCLEAN

DATE/TIME OF TEST: 06/18/2023 00:39:44

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.050	00:35
Breath Volume	2.308	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.048	00:38
Breath Volume	2.078	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 06/28/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BAPTISTE,MANOUCHEKA
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: DIAZ LOPEZ

DATE/TIME OF TEST: 06/19/2023 01:02:49

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:58	
Control Test	0.082	00:58	
Air Blank	0.000	00:59	
Subject Sample	# 1	01:00	RFI Detect
Breath Volume	1.242	01:00	
Air Blank	0.000	01:01	
Control Test	0.082	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 06/28/2023 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BAPTISTE,MANOUCHEKA
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DIAZ LOPEZ

DATE/TIME OF TEST: 06/19/2023 01:28:41

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.081	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.169	01:22
Breath Volume	1.144	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.169	01:26
Breath Volume	1.847	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 06/28/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPUTO,TAYLOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: CAPUTO,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 06/25/2023 23:07:45

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:54	
Control Test	0.080	22:54	
Air Blank	0.000	22:54	
Subject Sample	# 1 0.148	22:58	Volume Not Met
Breath Volume	0.066	22:58	
Air Blank	0.000	22:58	
Air Blank	0.000	23:00	
Subject Sample	# 2 0.124	23:01	
Breath Volume	1.296	23:01	
Air Blank	0.000	23:02	
Air Blank	0.000	23:04	
Subject Sample	# 3 0.130	23:05	
Breath Volume	1.683	23:05	
Air Blank	0.000	23:06	
Control Test	0.077	23:07	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 06/28/2023 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTHOLOMEW, MARK, A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BARTHOLOMEW, MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 06/26/2023 19:36:01

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:33	
Control Test	0.082	19:33	
Air Blank	0.000	19:34	
Subject Sample	# 1	19:34	RFI Detect
Breath Volume	0.152	19:34	
Air Blank	0.000	19:35	
Control Test	0.081	19:35	
Air Blank	0.000	19:35	
Diagnostics Check	OK	19:36	

Data Download Date/Time: 06/28/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CASSOLA

DATE/TIME OF TEST: 06/30/2023 23:57:03

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.129	23:52
Breath Volume	1.574	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.126	23:55
Breath Volume	1.375	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:57

Data Download Date/Time: 07/26/2023 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GREEN,OMARI
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 06/01/2023 05:04:38

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.080	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.169	05:00
Breath Volume	2.050	05:00
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.163	05:03
Breath Volume	1.960	05:03
Air Blank	0.000	05:03
Control Test	0.079	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 06/16/2023 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GREEN,OMARI
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/02/2023 00:08:12

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:55	
Control Test	0.080	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 3 0.172	23:59	Volume Not Met
Breath Volume	0.167	23:59	
Air Blank	0.000	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.188	00:03	
Breath Volume	1.257	00:03	
Air Blank	0.000	00:04	
Air Blank	0.000	00:05	
Subject Sample	# 2 0.193	00:06	
Breath Volume	1.203	00:06	
Air Blank	0.000	00:07	
Control Test	0.078	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 06/16/2023 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TAYLOR,OCTAVIOUS
ARRESTING OFFICER AGENCY: Davenport PD
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: AVILES FAJARDO

DATE/TIME OF TEST: 06/05/2023 01:35:01

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.141	01:30
Breath Volume	1.558	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.155	01:33
Breath Volume	1.742	01:33
Air Blank	0.000	01:34
Control Test	0.079	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:35

Data Download Date/Time: 06/16/2023 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: WONDERS

DATE/TIME OF TEST: 06/06/2023 21:41:56

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.196	21:36
Breath Volume	1.156	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.192	21:40
Breath Volume	1.394	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 06/16/2023 18:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: FLOREZ

DATE/TIME OF TEST: 06/10/2023 00:43:24

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.159	00:38
Breath Volume	1.386	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.163	00:41
Breath Volume	1.390	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 06/16/2023 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/13/2023 02:52:22

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.133	02:47
Breath Volume	1.250	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.131	02:50
Breath Volume	1.347	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 06/16/2023 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: SKYBERG

DATE/TIME OF TEST: 06/15/2023 11:49:52

Diagnostics Check	OK	11:43
Air Blank	0.000	11:44
Control Test	0.080	11:44
Air Blank	0.000	11:45
Subject Sample # 1	0.000	11:45
Breath Volume	1.960	11:45
Air Blank	0.000	11:46
Air Blank	0.000	11:48
Subject Sample # 2	0.000	11:48
Breath Volume	1.988	11:48
Air Blank	0.000	11:48
Control Test	0.080	11:49
Air Blank	0.000	11:49
Diagnostics Check	OK	11:49

Data Download Date/Time: 06/16/2023 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 06/16/2023 08:52:51

Diagnostics Check	OK	08:46
Air Blank	0.000	08:47
Control Test	0.080	08:47
Air Blank	0.000	08:47
Subject Sample # 1	0.000	08:48
Breath Volume	2.535	08:48
Air Blank	0.000	08:49
Air Blank	0.000	08:51
Subject Sample # 2	0.000	08:51
Breath Volume	2.609	08:51
Air Blank	0.000	08:51
Control Test	0.081	08:52
Air Blank	0.000	08:52
Diagnostics Check	OK	08:52

Data Download Date/Time: 06/16/2023 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: CASTRO,JOHN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: ALEJANDRE

DATE/TIME OF TEST: 06/17/2023 03:49:53

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.131	03:45
Breath Volume	1.417	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:47
Subject Sample # 2	0.130	03:48
Breath Volume	1.347	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 06/22/2023 15:17:24

Diagnostics Check	OK	15:11
Air Blank	0.000	15:11
Control Test	0.080	15:12
Air Blank	0.000	15:12
Subject Sample # 1	0.000	15:13
Breath Volume	2.265	15:13
Air Blank	0.000	15:13
Air Blank	0.000	15:15
Subject Sample # 2	0.000	15:16
Breath Volume	3.660	15:16
Air Blank	0.000	15:16
Control Test	0.080	15:16
Air Blank	0.000	15:17
Diagnostics Check	OK	15:17

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: GILLESPIE

DATE/TIME OF TEST: 06/27/2023 11:15:46

Diagnostics Check	OK	11:09
Air Blank	0.000	11:10
Control Test	0.080	11:10
Air Blank	0.000	11:11
Subject Sample # 1	0.000	11:11
Breath Volume	2.453	11:11
Air Blank	0.000	11:11
Air Blank	0.000	11:13
Subject Sample # 2	0.000	11:14
Breath Volume	2.449	11:14
Air Blank	0.000	11:14
Control Test	0.079	11:15
Air Blank	0.000	11:15
Diagnostics Check	OK	11:15

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 06/29/2023 10:15:07

Diagnostics Check	OK	10:09
Air Blank	0.000	10:09
Control Test	0.080	10:09
Air Blank	0.000	10:10
Subject Sample # 1	0.000	10:10
Breath Volume	2.164	10:10
Air Blank	0.000	10:11
Air Blank	0.000	10:13
Subject Sample # 2	0.000	10:13
Breath Volume	2.539	10:13
Air Blank	0.000	10:14
Control Test	0.080	10:14
Air Blank	0.000	10:15
Diagnostics Check	OK	10:15

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: VILLA

DATE/TIME OF TEST: 06/29/2023 11:38:56

Diagnostics Check	OK	11:32
Air Blank	0.000	11:33
Control Test	0.080	11:33
Air Blank	0.000	11:34
Subject Sample # 1	0.000	11:34
Breath Volume	1.714	11:34
Air Blank	0.000	11:35
Air Blank	0.000	11:37
Subject Sample # 2	0.000	11:37
Breath Volume	2.199	11:37
Air Blank	0.000	11:38
Control Test	0.080	11:38
Air Blank	0.000	11:38
Diagnostics Check	OK	11:38

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 06/29/2023 12:32:15

Diagnostics Check	OK	12:26
Air Blank	0.000	12:26
Control Test	0.080	12:26
Air Blank	0.000	12:27
Subject Sample # 1	0.000	12:27
Breath Volume	2.007	12:27
Air Blank	0.000	12:28
Air Blank	0.000	12:30
Subject Sample # 2	0.000	12:30
Breath Volume	2.296	12:30
Air Blank	0.000	12:31
Control Test	0.080	12:31
Air Blank	0.000	12:32
Diagnostics Check	OK	12:32

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: TOOMEY

DATE/TIME OF TEST: 06/29/2023 14:29:49

Diagnostics Check	OK	14:23
Air Blank	0.000	14:24
Control Test	0.081	14:24
Air Blank	0.000	14:24
Subject Sample # 1	0.000	14:25
Breath Volume	1.445	14:25
Air Blank	0.000	14:26
Air Blank	0.000	14:28
Subject Sample # 2	0.000	14:28
Breath Volume	1.617	14:28
Air Blank	0.000	14:28
Control Test	0.078	14:29
Air Blank	0.000	14:29
Diagnostics Check	OK	14:29

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: ANGLIN

DATE/TIME OF TEST: 06/30/2023 08:40:38

Diagnostics Check	OK	08:34
Air Blank	0.000	08:35
Control Test	0.081	08:35
Air Blank	0.000	08:36
Subject Sample # 1	0.000	08:36
Breath Volume	1.332	08:36
Air Blank	0.000	08:36
Air Blank	0.000	08:38
Subject Sample # 2	0.000	08:39
Breath Volume	1.507	08:39
Air Blank	0.000	08:39
Control Test	0.081	08:40
Air Blank	0.000	08:40
Diagnostics Check	OK	08:40

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: ASSINI

DATE/TIME OF TEST: 06/30/2023 11:03:46

Diagnostics Check	OK	10:57
Air Blank	0.000	10:58
Control Test	0.080	10:58
Air Blank	0.000	10:59
Subject Sample # 1	0.000	10:59
Breath Volume	4.355	10:59
Air Blank	0.000	10:59
Air Blank	0.000	11:01
Subject Sample # 2	0.000	11:02
Breath Volume	4.429	11:02
Air Blank	0.000	11:02
Control Test	0.080	11:03
Air Blank	0.000	11:03
Diagnostics Check	OK	11:03

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: BLACKMON

DATE/TIME OF TEST: 06/30/2023 12:01:38

Diagnostics Check	OK	11:55
Air Blank	0.000	11:56
Control Test	0.080	11:56
Air Blank	0.000	11:56
Subject Sample # 1	0.000	11:57
Breath Volume	2.601	11:57
Air Blank	0.000	11:57
Air Blank	0.000	11:59
Subject Sample # 2	0.000	12:00
Breath Volume	2.195	12:00
Air Blank	0.000	12:00
Control Test	0.079	12:01
Air Blank	0.000	12:01
Diagnostics Check	OK	12:01

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: FONSECA

DATE/TIME OF TEST: 06/30/2023 12:14:21

Diagnostics Check	OK	12:08
Air Blank	0.000	12:08
Control Test	0.080	12:08
Air Blank	0.000	12:09
Subject Sample # 1	0.000	12:10
Breath Volume	1.472	12:10
Air Blank	0.000	12:10
Air Blank	0.000	12:12
Subject Sample # 2	0.000	12:12
Breath Volume	1.953	12:12
Air Blank	0.000	12:13
Control Test	0.080	12:13
Air Blank	0.000	12:14
Diagnostics Check	OK	12:14

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: GRANT'PERKINS

DATE/TIME OF TEST: 06/30/2023 13:10:02

Diagnostics Check	OK	13:04
Air Blank	0.000	13:04
Control Test	0.080	13:04
Air Blank	0.000	13:05
Subject Sample # 1	0.000	13:05
Breath Volume	1.761	13:05
Air Blank	0.000	13:06
Air Blank	0.000	13:08
Subject Sample # 2	0.000	13:08
Breath Volume	1.187	13:08
Air Blank	0.000	13:09
Control Test	0.079	13:09
Air Blank	0.000	13:09
Diagnostics Check	OK	13:10

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RITTER,ALEC,D

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: RITTER,ALEC

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30822080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 13:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:39

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 06/02/2023 15:06:09

Diagnostics Check	OK	14:59
Air Blank	0.000	14:59
Control Test	0.083	14:59
Air Blank	0.000	15:00
Subject Sample # 1	0.202	15:01
Breath Volume	3.101	15:01
Air Blank	0.000	15:01
Air Blank	0.000	15:03
Subject Sample # 2	0.194	15:04
Breath Volume	2.539	15:04
Air Blank	0.000	15:05
Control Test	0.081	15:05
Air Blank	0.000	15:06
Diagnostics Check	OK	15:06

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: BLACKMON

DATE/TIME OF TEST: 06/04/2023 03:12:36

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.082	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.217	03:07
Breath Volume	1.417	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.221	03:10
Breath Volume	2.210	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PETERS,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: KINNE

DATE/TIME OF TEST: 06/04/2023 03:52:56

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.129	03:47
Breath Volume	1.402	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.130	03:51
Breath Volume	1.347	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,DAVID,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAMOS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: KABACEVIC

DATE/TIME OF TEST: 06/04/2023 15:43:08

Diagnostics Check	OK	15:35
Air Blank	0.000	15:36
Control Test	0.082	15:36
Air Blank	0.000	15:36
Subject Sample # 1	0.136	15:38
Breath Volume	2.414	15:38
Air Blank	0.000	15:38
Air Blank	0.000	15:40
Subject Sample # 2	0.140	15:41
Breath Volume	2.914	15:41
Air Blank	0.000	15:42
Control Test	0.081	15:42
Air Blank	0.000	15:43
Diagnostics Check	OK	15:43

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: JUSTINO

DATE/TIME OF TEST: 06/05/2023 02:06:23

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.082	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:05	
Control Test	0.081	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: URLACHER

DATE/TIME OF TEST: 06/06/2023 23:01:16

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.083	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.245	22:56
Breath Volume	1.906	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.248	22:59
Breath Volume	1.730	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PETERS,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: GRIEVES

DATE/TIME OF TEST: 06/07/2023 21:57:40

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.083	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.000	21:53
Breath Volume	2.187	21:53
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.000	21:56
Breath Volume	1.832	21:56
Air Blank	0.000	21:56
Control Test	0.082	21:57
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MERCIER

DATE/TIME OF TEST: 06/08/2023 03:17:41

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.082	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.122	03:13
Breath Volume	1.628	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.116	03:16
Breath Volume	2.429	03:16
Air Blank	0.000	03:16
Control Test	0.081	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: MONTILLA

DATE/TIME OF TEST: 06/08/2023 07:42:47

Diagnostics Check	OK	07:36
Air Blank	0.000	07:36
Control Test	0.081	07:36
Air Blank	0.000	07:37
Subject Sample # 1	0.158	07:38
Breath Volume	2.320	07:38
Air Blank	0.000	07:39
Air Blank	0.000	07:40
Subject Sample # 2	0.159	07:41
Breath Volume	1.718	07:41
Air Blank	0.000	07:41
Control Test	0.082	07:42
Air Blank	0.000	07:42
Diagnostics Check	OK	07:42

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: CRAWLEY

DATE/TIME OF TEST: 06/10/2023 16:32:24

Diagnostics Check	OK	16:25
Air Blank	0.000	16:26
Control Test	0.083	16:26
Air Blank	0.000	16:27
Subject Sample # 1	0.000	16:27
Breath Volume	3.222	16:27
Air Blank	0.000	16:28
Air Blank	0.000	16:30
Subject Sample # 2	0.000	16:31
Breath Volume	3.386	16:31
Air Blank	0.000	16:31
Control Test	0.082	16:31
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POULIN,D,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROZELL,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: PREACHER

DATE/TIME OF TEST: 06/11/2023 16:07:35

Diagnostics Check	OK	16:00	
Air Blank	0.000	16:01	
Control Test	0.084	16:01	
Air Blank	0.000	16:02	
Subject Sample	# 1 0.198	16:02	
Breath Volume	1.960	16:02	
Air Blank	0.000	16:03	
Air Blank	0.000	16:05	
Subject Sample	# 2	16:06	Subject Test Refused
Breath Volume	0.000	16:06	
Air Blank	0.000	16:06	
Control Test	0.082	16:06	
Air Blank	0.000	16:07	
Diagnostics Check	OK	16:07	

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, COLLEEN, E
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE, R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: HASAN

DATE/TIME OF TEST: 06/12/2023 14:42:05

Diagnostics Check	OK	14:35
Air Blank	0.000	14:36
Control Test	0.084	14:36
Air Blank	0.000	14:36
Subject Sample # 1	0.000	14:37
Breath Volume	1.863	14:37
Air Blank	0.000	14:38
Air Blank	0.000	14:40
Subject Sample # 2	0.000	14:40
Breath Volume	1.582	14:40
Air Blank	0.000	14:41
Control Test	0.082	14:41
Air Blank	0.000	14:42
Diagnostics Check	OK	14:42

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: REID,HUNTER
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: BREAKER

DATE/TIME OF TEST: 06/17/2023 03:29:37

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.083	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.156	03:24
Breath Volume	1.726	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.160	03:27
Breath Volume	1.863	03:27
Air Blank	0.000	03:28
Control Test	0.083	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 06/21/2023 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SCHMIDT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 06/17/2023 04:16:37

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.083	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.133	04:10
Breath Volume	1.960	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.133	04:14
Breath Volume	2.117	04:14
Air Blank	0.000	04:15
Control Test	0.082	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 06/21/2023 19:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SIUZDAK,COLTON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: SUCHOMSKI

DATE/TIME OF TEST: 06/17/2023 21:56:17

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.084	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.016	21:50
Breath Volume	2.636	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.014	21:54
Breath Volume	1.394	21:54
Air Blank	0.000	21:55
Control Test	0.083	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 06/21/2023 19:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 06/20/2023 20:58:56

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:55	
Control Test	0.083	20:56	
Air Blank	0.000	20:56	
Subject Sample	# 1	20:57	Subject Test Refused
Breath Volume	0.000	20:57	
Air Blank	0.000	20:58	
Control Test	0.082	20:58	
Air Blank	0.000	20:58	
Diagnostics Check	OK	20:58	

Data Download Date/Time: 06/21/2023 19:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: GIMONDO

DATE/TIME OF TEST: 06/29/2023 20:07:21

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.084	20:00
Air Blank	0.000	20:01
Subject Sample # 1	0.167	20:02
Breath Volume	1.128	20:02
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.165	20:05
Breath Volume	1.242	20:05
Air Blank	0.000	20:06
Control Test	0.082	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: REID,HUNTER
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: PLUSKOTA

DATE/TIME OF TEST: 06/30/2023 21:58:21

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.084	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.052	21:53
Breath Volume	2.093	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.048	21:56
Breath Volume	1.179	21:56
Air Blank	0.000	21:57
Control Test	0.083	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BOTELHO,ISAC
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HURLEY

DATE/TIME OF TEST: 06/01/2023 01:43:07

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.208	01:38
Breath Volume	1.414	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.226	01:41
Breath Volume	2.101	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,VICTORIA,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SIMONSEN,ERIK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CRANSTON

DATE/TIME OF TEST: 06/02/2023 03:33:02

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.078	03:30	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.078	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERT,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PEREZ,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: THEMUDA

DATE/TIME OF TEST: 06/03/2023 01:31:53

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.079	01:29	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Control Test	0.078	01:31	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 06/03/2023 03:26:21

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.203	03:21
Breath Volume	1.140	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample # 2	0.196	03:24
Breath Volume	1.417	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLANSON,JACE,W
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: BRAUN,DANIELLE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ZAY

DATE/TIME OF TEST: 06/03/2023 03:40:41

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.225	03:36
Breath Volume	2.300	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.222	03:38
Breath Volume	2.023	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO,KENNETH,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PRIETO,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 06/03/2023 16:27:36

Diagnostics Check	OK	16:21
Air Blank	0.000	16:22
Control Test	0.082	16:22
Air Blank	0.000	16:22
Subject Sample # 1	0.000	16:23
Breath Volume	3.015	16:23
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.000	16:26
Breath Volume	3.148	16:26
Air Blank	0.000	16:26
Control Test	0.080	16:27
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: URDANETA ALVARADO

DATE/TIME OF TEST: 06/04/2023 03:27:07

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.078	03:22
Breath Volume	1.699	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.079	03:25
Breath Volume	1.507	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CETIN,TRAVIS,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: CETIN,TRAVIS
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 13:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: BLANCHETTE

DATE/TIME OF TEST: 06/04/2023 14:57:34

Diagnostics Check	OK	14:51
Air Blank	0.000	14:51
Control Test	0.081	14:51
Air Blank	0.000	14:52
Subject Sample # 1	0.238	14:52
Breath Volume	2.519	14:52
Air Blank	0.000	14:53
Air Blank	0.000	14:55
Subject Sample # 2	0.235	14:55
Breath Volume	2.542	14:55
Air Blank	0.000	14:56
Control Test	0.079	14:56
Air Blank	0.000	14:57
Diagnostics Check	OK	14:57

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETTS,J,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: PREYER

DATE/TIME OF TEST: 06/05/2023 22:35:58

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.081	22:30
Air Blank	0.000	22:30
Subject Sample # 1	0.012	22:31
Breath Volume	1.937	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.010	22:34
Breath Volume	1.968	22:34
Air Blank	0.000	22:35
Control Test	0.081	22:35
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 06/27/2023 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CATHERR

DATE/TIME OF TEST: 06/09/2023 23:00:12

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.142	22:55
Breath Volume	1.113	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.140	22:58
Breath Volume	1.226	22:58
Air Blank	0.000	22:59
Control Test	0.081	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 06/10/2023 01:48:34

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.176	01:43
Breath Volume	1.984	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:45
Subject Sample # 2	0.168	01:47
Breath Volume	2.773	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: PASCO

DATE/TIME OF TEST: 06/11/2023 03:45:59

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.127	03:41
Breath Volume	1.835	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.129	03:44
Breath Volume	3.488	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVILA,VICTOR,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: DAVILA,VICTOR
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 06/11/2023 20:25:20

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:11	
Control Test	0.081	20:11	
Air Blank	0.000	20:12	
Subject Sample	# 1 0.077	20:15	Volume Not Met
Breath Volume	0.273	20:15	
Air Blank	0.000	20:16	
Air Blank	0.000	20:17	
Subject Sample	# 2 0.082	20:20	
Breath Volume	1.421	20:20	
Air Blank	0.000	20:21	
Air Blank	0.000	20:22	
Subject Sample	# 3 0.089	20:23	
Breath Volume	1.515	20:23	
Air Blank	0.000	20:24	
Control Test	0.080	20:24	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:25	

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HERNANDEZ,IVAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 06/12/2023 01:30:52

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.220	01:25
Breath Volume	1.277	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.233	01:29
Breath Volume	2.812	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFENBERG,ASHTON,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WELFENBERG,ASHTON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: APPLEHANS

DATE/TIME OF TEST: 06/14/2023 03:04:44

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.082	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.000	02:59
Breath Volume	1.964	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.000	03:03
Breath Volume	1.644	03:03
Air Blank	0.000	03:03
Control Test	0.081	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUTRAL,WILLIAM,Z
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FUTRAL,WILLIAM
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402378118
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 06/17/2023 04:25:44

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.080	04:22	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Control Test	0.080	04:25	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GOMEZ PEREZ

DATE/TIME OF TEST: 06/18/2023 01:40:00

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.112	01:35
Breath Volume	1.226	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.110	01:38
Breath Volume	1.593	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:40

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: PACHECO-SERRANO

DATE/TIME OF TEST: 06/18/2023 03:50:14

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.188	03:45
Breath Volume	2.531	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.192	03:48
Breath Volume	2.191	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECCUE,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BECCUE,JOSHUA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 06/20/2023 01:37:01

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.082	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:35	Subject Test Refused
Breath Volume	0.000	01:35	
Air Blank	0.000	01:36	
Control Test	0.081	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,SPENSER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KENNEDY,SPENSER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: TRENTER

DATE/TIME OF TEST: 06/21/2023 18:41:51

Diagnostics Check	OK	18:36	
Air Blank	0.000	18:36	
Control Test	0.080	18:37	
Air Blank	0.000	18:37	
Subject Sample	# 1	18:40	Subject Test Refused
Breath Volume	0.000	18:40	
Air Blank	0.000	18:40	
Control Test	0.080	18:41	
Air Blank	0.000	18:41	
Diagnostics Check	OK	18:41	

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SAINES

DATE/TIME OF TEST: 06/21/2023 23:52:48

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.081	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1 0.120	23:45	Volume Not Met
Breath Volume	0.824	23:45	
Air Blank	0.000	23:46	
Air Blank	0.000	23:48	
Subject Sample	# 2 0.108	23:51	Volume Not Met
Breath Volume	0.667	23:51	
Air Blank	0.000	23:51	
Control Test	0.080	23:52	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SAINES

DATE/TIME OF TEST: 06/22/2023 00:06:40

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.123	00:02
Breath Volume	1.335	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.124	00:05
Breath Volume	1.593	00:05
Air Blank	0.000	00:05
Control Test	0.079	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,SAMANTHA,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FRENCH,SAMANTHA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 06/22/2023 22:33:15

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.080	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 1 0.191	22:25	Volume Not Met
Breath Volume	0.128	22:25	
Air Blank	0.000	22:26	
Air Blank	0.000	22:28	
Subject Sample	# 2 0.198	22:28	
Breath Volume	2.230	22:28	
Air Blank	0.000	22:29	
Air Blank	0.000	22:31	
Subject Sample	# 3 0.188	22:31	
Breath Volume	1.414	22:31	
Air Blank	0.000	22:32	
Control Test	0.080	22:32	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASLEY,M,
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: EASLEY,M
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: LAMBCKE

DATE/TIME OF TEST: 06/23/2023 02:53:26

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.288	02:48
Breath Volume	2.242	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.297	02:51
Breath Volume	3.070	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SHUMAN,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 06/24/2023 01:49:18

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.081	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.127	01:44
Breath Volume	4.609	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.124	01:47
Breath Volume	3.941	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 06/24/2023 03:37:32

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.143	03:32
Breath Volume	1.250	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.147	03:35
Breath Volume	3.292	03:35
Air Blank	0.000	03:36
Control Test	0.082	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZEUSKI,JIMMY,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ZEUSKI,JIMMY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/25/2023 00:57:24

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.082	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.173	00:52
Breath Volume	2.324	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.174	00:55
Breath Volume	1.890	00:55
Air Blank	0.000	00:56
Control Test	0.082	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SHUMAN,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/25/2023 04:47:28

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.082	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.120	04:42
Breath Volume	2.039	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.127	04:45
Breath Volume	2.511	04:45
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUIS,ALYSSA,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ROUIS,ALYSSA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: CALIGUIRE

DATE/TIME OF TEST: 06/25/2023 05:47:30

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:42	
Control Test	0.082	05:42	
Air Blank	0.000	05:43	
Subject Sample	# 1 0.000	05:43	
Breath Volume	3.027	05:43	
Air Blank	0.000	05:43	
Air Blank	0.000	05:45	
Subject Sample	# 2	05:46	RFI Detect
Breath Volume	1.000	05:46	
Air Blank	0.000	05:46	
Control Test	0.081	05:46	
Air Blank	0.000	05:47	
Diagnostics Check	OK	05:47	

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUIS,ALYSSA,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ROUIS,ALYSSA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: CALIGUIRE

DATE/TIME OF TEST: 06/25/2023 05:56:03

Diagnostics Check	OK	05:50
Air Blank	0.000	05:50
Control Test	0.082	05:50
Air Blank	0.000	05:51
Subject Sample # 1	0.000	05:51
Breath Volume	3.519	05:51
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample # 2	0.000	05:54
Breath Volume	3.312	05:54
Air Blank	0.000	05:55
Control Test	0.082	05:55
Air Blank	0.000	05:55
Diagnostics Check	OK	05:56

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SHUMAN,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 06/26/2023 03:15:41

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.082	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.133	03:10
Breath Volume	1.832	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.129	03:14
Breath Volume	1.832	03:14
Air Blank	0.000	03:14
Control Test	0.083	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 06/27/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TAPANES,CODY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: KADEL

DATE/TIME OF TEST: 06/29/2023 22:13:58

Diagnostics Check	OK	22:07
Air Blank	0.000	22:07
Control Test	0.080	22:08
Air Blank	0.000	22:08
Subject Sample # 1	0.167	22:09
Breath Volume	1.730	22:09
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.169	22:12
Breath Volume	1.261	22:12
Air Blank	0.000	22:13
Control Test	0.078	22:13
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 07/24/2023 19:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRINKLE,BRITTANY,N
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SANTAMARIA,J
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 06/30/2023 23:33:58

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:33	
Control Test	0.078	23:33	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:33	

Data Download Date/Time: 07/24/2023 19:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURO,JAMES,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HEWW,HEWITT
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: R120-793-78-043-0

DATE/TIME OF TEST: 06/03/2023 14:58:05

Diagnostics Check	OK	14:49	
Air Blank	0.000	14:50	
Control Test	0.079	14:50	
Air Blank	0.000	14:50	
Subject Sample	# 1	14:54	No Sample Provided
Breath Volume	0.000	14:54	
Air Blank	0.000	14:54	
Air Blank	0.000	14:56	
Subject Sample	# 2	14:56	Subject Test Refused
Breath Volume	0.000	14:56	
Air Blank	0.000	14:57	
Control Test	0.079	14:57	
Air Blank	0.000	14:58	
Diagnostics Check	OK	14:58	

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SAAYMAN,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 06/04/2023 02:03:04

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1 0.172	01:55	Volume Not Met
Breath Volume	0.136	01:55	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample	# 2 0.168	01:58	
Breath Volume	2.410	01:58	
Air Blank	0.000	01:58	
Air Blank	0.000	02:00	
Subject Sample	# 3 0.166	02:01	
Breath Volume	2.753	02:01	
Air Blank	0.000	02:02	
Control Test	0.079	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURO,JAMES,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,KYLE
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: SUGGS

DATE/TIME OF TEST: 06/04/2023 11:25:17

Diagnostics Check	OK	11:16
Air Blank	0.000	11:16
Control Test	0.080	11:17
Air Blank	0.000	11:17
Subject Sample # 1	0.255	11:18
Breath Volume	2.105	11:18
Air Blank	0.000	11:18
Air Blank	0.000	11:20
Subject Sample # 2	0.234	11:20
Breath Volume	2.300	11:20
Air Blank	0.000	11:21
Air Blank	0.000	11:23
Subject Sample # 3	0.238	11:23
Breath Volume	2.281	11:23
Air Blank	0.000	11:24
Control Test	0.078	11:24
Air Blank	0.000	11:25
Diagnostics Check	OK	11:25

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURO,JAMES,M
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: JOVANOVA,M
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 06/04/2023 17:20:19

Diagnostics Check	OK	17:14
Air Blank	0.000	17:14
Control Test	0.080	17:14
Air Blank	0.000	17:15
Subject Sample # 1	0.000	17:16
Breath Volume	2.214	17:16
Air Blank	0.000	17:16
Air Blank	0.000	17:18
Subject Sample # 2	0.000	17:18
Breath Volume	1.777	17:18
Air Blank	0.000	17:19
Control Test	0.079	17:19
Air Blank	0.000	17:20
Diagnostics Check	OK	17:20

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SMITH,JOSH
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: HINTERSEHR

DATE/TIME OF TEST: 06/10/2023 01:17:54

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:09	
Control Test	0.080	01:10	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:13	No Sample Provided
Breath Volume	0.000	01:13	
Air Blank	0.000	01:14	
Air Blank	0.000	01:16	
Subject Sample	# 2	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.079	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 06/10/2023 19:47:00

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.080	19:40
Air Blank	0.000	19:40
Subject Sample # 1	0.168	19:42
Breath Volume	1.660	19:42
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.163	19:45
Breath Volume	1.195	19:45
Air Blank	0.000	19:46
Control Test	0.079	19:46
Air Blank	0.000	19:46
Diagnostics Check	OK	19:47

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 06/12/2023 22:47:45

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.140	22:42
Breath Volume	1.718	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.132	22:46
Breath Volume	1.253	22:46
Air Blank	0.000	22:46
Control Test	0.079	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURO,JAMES,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KNIGHT,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: SCHMITT

DATE/TIME OF TEST: 06/18/2023 07:46:34

Diagnostics Check	OK	07:40
Air Blank	0.000	07:40
Control Test	0.080	07:41
Air Blank	0.000	07:41
Subject Sample # 1	0.155	07:42
Breath Volume	2.253	07:42
Air Blank	0.000	07:42
Air Blank	0.000	07:44
Subject Sample # 2	0.150	07:45
Breath Volume	2.921	07:45
Air Blank	0.000	07:45
Control Test	0.079	07:45
Air Blank	0.000	07:46
Diagnostics Check	OK	07:46

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: RAMOS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 06/18/2023 19:28:55

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:15	
Control Test	0.080	19:16	
Air Blank	0.000	19:16	
Subject Sample	# 1 0.087	19:19	Volume Not Met
Breath Volume	0.753	19:19	
Air Blank	0.000	19:20	
Air Blank	0.000	19:22	
Subject Sample	# 2 0.095	19:24	
Breath Volume	1.554	19:24	
Air Blank	0.000	19:24	
Air Blank	0.000	19:26	
Subject Sample	# 3 0.098	19:27	
Breath Volume	1.589	19:27	
Air Blank	0.000	19:27	
Control Test	0.077	19:28	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:28	

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 06/18/2023 20:04:33

Diagnostics Check	OK	19:52	
Air Blank	0.000	19:53	
Control Test	0.079	19:53	
Air Blank	0.000	19:54	
Subject Sample	# 1 0.117	19:57	Volume Not Met
Breath Volume	0.222	19:57	
Air Blank	0.000	19:57	
Air Blank	0.000	19:59	
Subject Sample	# 2 0.116	20:02	Volume Not Met
Breath Volume	0.234	20:02	
Air Blank	0.000	20:03	
Control Test	0.078	20:03	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, ANTHONY, T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: MCDONOLD, D
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 06/24/2023 03:39:51

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.172	03:35
Breath Volume	2.167	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.154	03:38
Breath Volume	1.492	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ,G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: ENNEN

DATE/TIME OF TEST: 06/26/2023 21:16:39

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.079	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.112	21:12
Breath Volume	2.300	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.110	21:15
Breath Volume	1.917	21:15
Air Blank	0.000	21:15
Control Test	0.079	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, ANTHONY, T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ADOLF, C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: ALLER

DATE/TIME OF TEST: 06/28/2023 22:20:52

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:07	
Control Test	0.078	22:07	
Air Blank	0.000	22:08	
Subject Sample	# 1 0.197	22:11	Volume Not Met
Breath Volume	0.355	22:11	
Air Blank	0.000	22:11	
Air Blank	0.000	22:13	
Subject Sample	# 2 0.272	22:14	
Breath Volume	1.339	22:14	
Air Blank	0.000	22:15	
Air Blank	0.000	22:16	
Subject Sample	# 3 0.255	22:19	
Breath Volume	2.636	22:19	
Air Blank	0.000	22:19	
Control Test	0.078	22:20	
Air Blank	0.000	22:20	
Diagnostics Check	OK	22:20	

Data Download Date/Time: 06/29/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,TIM
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: TESSMER

DATE/TIME OF TEST: 06/29/2023 16:55:33

Diagnostics Check	OK	16:48
Air Blank	0.000	16:49
Control Test	0.078	16:49
Air Blank	0.000	16:49
Subject Sample # 1	0.172	16:50
Breath Volume	1.582	16:50
Air Blank	0.000	16:51
Air Blank	0.000	16:53
Subject Sample # 2	0.160	16:54
Breath Volume	1.851	16:54
Air Blank	0.000	16:54
Control Test	0.078	16:54
Air Blank	0.000	16:55
Diagnostics Check	OK	16:55

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, ANTHONY, T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DEP, DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 06/30/2023 03:25:25

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.197	03:21
Breath Volume	1.316	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.191	03:23
Breath Volume	1.117	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOVER,DAVID,W
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: HOOVER,DAVID
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 06/01/2023 00:38:34

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.079	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.169	00:30
Breath Volume	2.851	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.191	00:33
Breath Volume	1.976	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 3	0.173	00:36
Breath Volume	1.750	00:36
Air Blank	0.000	00:37
Control Test	0.078	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREANO,JACK,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: ANDREANO,JACK
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MALLETT

DATE/TIME OF TEST: 06/01/2023 03:24:10

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Control Test	0.079	03:23	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: SILVERIO,A
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 06/03/2023 00:13:16

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.000	00:08
Breath Volume	1.265	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.000	00:11
Breath Volume	1.359	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILDEMAN-STUART,N
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: LATIMORE

DATE/TIME OF TEST: 06/03/2023 02:34:02

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.080	02:33	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LATIMIER,OFFICER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 06/04/2023 02:52:18

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:35	
Control Test	0.080	02:35	
Air Blank	0.000	02:36	
Subject Sample	# 1 0.157	02:39	Volume Not Met
Breath Volume	0.878	02:39	
Air Blank	0.000	02:40	
Air Blank	0.000	02:41	
Subject Sample	# 2 0.161	02:44	
Breath Volume	1.195	02:44	
Air Blank	0.000	02:45	
Air Blank	0.000	02:47	
Subject Sample	# 3 0.133	02:50	Volume Not Met
Breath Volume	0.289	02:50	
Air Blank	0.000	02:51	
Control Test	0.076	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LATIMIER,J
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 06/04/2023 03:01:27

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.159	02:56
Breath Volume	1.292	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.172	02:59
Breath Volume	2.402	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: PITTMAN,T
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:16

SUBJECT'S LAST NAME: WINTERS

DATE/TIME OF TEST: 06/04/2023 14:53:50

Diagnostics Check	OK	14:50	
Air Blank	0.000	14:51	
Control Test	0.080	14:51	
Air Blank	0.000	14:52	
Subject Sample	# 1	14:52	Subject Test Refused
Breath Volume	0.000	14:52	
Air Blank	0.000	14:52	
Control Test	0.080	14:53	
Air Blank	0.000	14:53	
Diagnostics Check	OK	14:53	

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: POTH,LAUREN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 06/09/2023 22:54:13

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:51	
Control Test	0.080	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 1	22:52	RFI Detect
Breath Volume	0.000	22:52	
Air Blank	0.000	22:53	
Control Test	0.080	22:53	
Air Blank	0.000	22:54	
Diagnostics Check	OK	22:54	

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: POTH,LAUREN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 06/09/2023 23:04:31

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.132	22:59
Breath Volume	3.847	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.137	23:02
Breath Volume	3.921	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: PEREZ,C
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 06/12/2023 00:55:13

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.180	00:50
Breath Volume	1.156	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:52
Subject Sample # 2	0.179	00:53
Breath Volume	1.664	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BANNICK,COURTNEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: BRUZA

DATE/TIME OF TEST: 06/12/2023 14:59:23

Diagnostics Check	OK	14:52
Air Blank	0.000	14:53
Control Test	0.079	14:53
Air Blank	0.000	14:54
Subject Sample # 1	0.048	14:54
Breath Volume	1.132	14:54
Air Blank	0.000	14:55
Air Blank	0.000	14:57
Subject Sample # 2	0.050	14:57
Breath Volume	1.164	14:57
Air Blank	0.000	14:58
Control Test	0.080	14:58
Air Blank	0.000	14:59
Diagnostics Check	OK	14:59

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONDAHL,CARL,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PERALTA,BRYAN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 06/17/2023 01:27:33

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.080	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.239	01:22
Breath Volume	1.394	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.251	01:25
Breath Volume	1.371	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: TUFTS,KEVIN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SORRELL

DATE/TIME OF TEST: 06/17/2023 04:48:26

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:46	
Control Test	0.080	04:46	
Air Blank	0.000	04:46	
Subject Sample	# 1	04:47	Subject Test Refused
Breath Volume	0.000	04:47	
Air Blank	0.000	04:47	
Control Test	0.079	04:47	
Air Blank	0.000	04:48	
Diagnostics Check	OK	04:48	

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,RYAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MOONEY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 06/17/2023 22:53:46

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.080	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.185	22:49
Breath Volume	2.390	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.172	22:52
Breath Volume	2.675	22:52
Air Blank	0.000	22:52
Control Test	0.080	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONDAHL,CARL,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ROBERT,BRANDON
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: RIVAS REYES

DATE/TIME OF TEST: 06/17/2023 23:03:32

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:02	Subject Test Refused
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Control Test	0.080	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILDEMAN-STUART,N
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MCLEA

DATE/TIME OF TEST: 06/18/2023 03:40:14

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.080	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:39	
Control Test	0.080	03:39	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONDAHL,CARL,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ROBERT,BRANDON
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: GINGERELLA

DATE/TIME OF TEST: 06/18/2023 20:28:01

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.298	20:23
Breath Volume	1.378	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.316	20:26
Breath Volume	1.726	20:26
Air Blank	0.000	20:27
Control Test	0.077	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:28

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MONTALVO,LUIS
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 06/19/2023 00:47:52

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.201	00:43
Breath Volume	2.261	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.198	00:46
Breath Volume	2.113	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONDAHL,CARL,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: SCHNEIDER,CODY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: HIBBERT

DATE/TIME OF TEST: 06/19/2023 04:03:31

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.197	03:58
Breath Volume	2.039	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.196	04:01
Breath Volume	2.101	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: SYLDOR,JEAN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: ROBILLARD

DATE/TIME OF TEST: 06/19/2023 13:25:22

Diagnostics Check	OK	13:20	
Air Blank	0.000	13:21	
Control Test	0.080	13:21	
Air Blank	0.000	13:22	
Subject Sample	# 1	13:23	Subject Test Refused
Breath Volume	0.000	13:23	
Air Blank	0.000	13:24	
Control Test	0.080	13:24	
Air Blank	0.000	13:25	
Diagnostics Check	OK	13:25	

Data Download Date/Time: 06/21/2023 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,ROBERT,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MONTALVO,LUIS
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/23/2023 00:06:00

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.079	00:02	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:05	
Control Test	0.080	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,ROBERT,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MONTALVO,LOUIS
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/23/2023 00:34:31

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.079	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.000	00:29
Breath Volume	1.839	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.000	00:33
Breath Volume	2.589	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: NGALA

DATE/TIME OF TEST: 06/23/2023 02:18:07

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.191	02:13
Breath Volume	1.347	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:15
Subject Sample # 2	0.186	02:16
Breath Volume	1.171	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: BULTRON,CARLOS
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/25/2023 02:19:43

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:18	
Control Test	0.079	02:18	
Air Blank	0.000	02:19	
		02:19	RFI Detect
Air Blank	0.000	02:19	

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: BULTRON,CARLOS
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/25/2023 02:58:06

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.315	02:52
Breath Volume	2.183	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.320	02:56
Breath Volume	2.015	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WILDEMAN-STUART,JORDAN

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SHIFFLETT

DATE/TIME OF TEST: 06/04/2023 03:58:12

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.079	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 06/28/2023 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WILDEMAN-STUART,JORDAN

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:49

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/08/2023 00:49:53

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.079	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.078	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 06/28/2023 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SOMMERSDORF,JOHN,G

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: SOMMERSDORF,JOHN

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 13:28

GENDER: Female

OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 06/08/2023 14:26:41

Diagnostics Check	OK	14:20
Air Blank	0.000	14:21
Control Test	0.079	14:21
Air Blank	0.000	14:21
Subject Sample # 1	0.118	14:22
Breath Volume	1.441	14:22
Air Blank	0.000	14:22
Air Blank	0.000	14:24
Subject Sample # 2	0.118	14:25
Breath Volume	1.824	14:25
Air Blank	0.000	14:25
Control Test	0.079	14:26
Air Blank	0.000	14:26
Diagnostics Check	OK	14:26

Data Download Date/Time: 06/28/2023 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLCHARRAN,DILON,C
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: DOLCHARRAN,DILON
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: LEDASMA

DATE/TIME OF TEST: 06/10/2023 22:14:30

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:12	
Control Test	0.080	22:12	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:13	
Control Test	0.080	22:13	
Air Blank	0.000	22:14	
Diagnostics Check	OK	22:14	

Data Download Date/Time: 06/28/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: LENFANT, BEN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 05/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 06/22/2023 19:47:14

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:44	
Control Test	0.079	19:45	
Air Blank	0.000	19:45	
Subject Sample	# 1	19:45	Subject Test Refused
Breath Volume	0.000	19:45	
Air Blank	0.000	19:46	
Control Test	0.080	19:46	
Air Blank	0.000	19:47	
Diagnostics Check	OK	19:47	

Data Download Date/Time: 06/28/2023 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOLCHARRAN,DILON,C

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: CROWE,SCOTT

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BOULIEU

DATE/TIME OF TEST: 06/29/2023 22:57:22

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:55	
Control Test	0.078	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:56	
Control Test	0.078	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 07/27/2023 14:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELHO,ISAC,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BOTELHO,ISAC
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MCKNIGHT

DATE/TIME OF TEST: 06/30/2023 00:09:17

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.082	00:07	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:08	
Control Test	0.082	00:08	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 07/27/2023 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: HINCAPIE

DATE/TIME OF TEST: 06/05/2023 22:48:06

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.083	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.170	22:40
Breath Volume	1.191	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.173	22:46
Breath Volume	1.746	22:46
Air Blank	0.000	22:47
Control Test	0.082	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 07/05/2023 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBINSON,G
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 06/07/2023 17:25:35

Diagnostics Check	OK	17:18
Air Blank	0.000	17:18
Control Test	0.083	17:19
Air Blank	0.000	17:19
Subject Sample # 1	0.000	17:20
Breath Volume	2.460	17:20
Air Blank	0.000	17:20
Air Blank	0.000	17:22
Subject Sample # 2	0.000	17:24
Breath Volume	2.214	17:24
Air Blank	0.000	17:24
Control Test	0.082	17:24
Air Blank	0.000	17:25
Diagnostics Check	OK	17:25

Data Download Date/Time: 07/05/2023 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/07/2023 23:16:26

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.083	23:10	
Air Blank	0.000	23:10	
Subject Sample	# 1 0.124	23:11	
Breath Volume	2.617	23:11	
Air Blank	0.000	23:12	
Air Blank	0.000	23:14	
Subject Sample	# 2 0.128	23:14	
Breath Volume	2.433	23:14	
Air Blank	0.000	23:15	
Control Test	0.104	23:15	Interferent Detect
Air Blank		23:16	Purge Fail

Data Download Date/Time: 07/05/2023 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: YILMAZ

DATE/TIME OF TEST: 06/08/2023 04:03:25

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.084	04:00	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:01	Interferent Detect
Breath Volume	3.058	04:01	
Air Blank	0.000	04:02	
Control Test	0.053	04:02	Interferent Detect
Air Blank	0.000	04:03	

Data Download Date/Time: 07/05/2023 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARVER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: BAUSCHER

DATE/TIME OF TEST: 06/09/2023 23:53:39

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.083	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.097	23:49
Breath Volume	2.265	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.089	23:52
Breath Volume	2.035	23:52
Air Blank	0.000	23:52
Control Test	0.082	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 07/05/2023 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 06/11/2023 05:51:50

Diagnostics Check	OK	05:45
Air Blank	0.000	05:45
Control Test	0.083	05:45
Air Blank	0.000	05:46
Subject Sample # 1	0.140	05:47
Breath Volume	1.308	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:49
Subject Sample # 2	0.143	05:50
Breath Volume	3.097	05:50
Air Blank	0.000	05:50
Control Test	0.082	05:51
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 07/05/2023 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: DUCKER

DATE/TIME OF TEST: 06/12/2023 02:29:38

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.083	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.187	02:24
Breath Volume	1.660	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.184	02:27
Breath Volume	1.421	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 07/05/2023 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 06/16/2023 03:10:36

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.083	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.163	03:05
Breath Volume	1.531	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:07
Subject Sample # 2	0.180	03:08
Breath Volume	1.472	03:08
Air Blank	0.000	03:09
Control Test	0.082	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 07/05/2023 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: BONO MORALES

DATE/TIME OF TEST: 06/16/2023 04:31:58

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.082	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.155	04:27
Breath Volume	3.062	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample # 2	0.155	04:30
Breath Volume	3.015	04:30
Air Blank	0.000	04:31
Control Test	0.082	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 07/05/2023 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 06/17/2023 01:52:36

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.083	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.149	01:48
Breath Volume	1.792	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.158	01:50
Breath Volume	1.804	01:50
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 07/05/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 06/17/2023 04:08:08

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.082	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.038	04:03
Breath Volume	1.171	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.036	04:06
Breath Volume	1.289	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 07/05/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 06/17/2023 04:54:17

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.083	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.133	04:49
Breath Volume	2.027	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.135	04:52
Breath Volume	1.617	04:52
Air Blank	0.000	04:53
Control Test	0.081	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 07/05/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/22/2023 02:41:03

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.083	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.106	02:36
Breath Volume	2.031	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:38
Subject Sample # 2	0.106	02:39
Breath Volume	1.699	02:39
Air Blank	0.000	02:40
Control Test	0.082	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:41

Data Download Date/Time: 07/05/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 06/24/2023 03:59:07

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.023	03:54
Breath Volume	1.980	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.021	03:57
Breath Volume	1.726	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 07/05/2023 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: BUNTON

DATE/TIME OF TEST: 06/25/2023 04:35:21

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.084	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.047	04:30
Breath Volume	3.738	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.047	04:33
Breath Volume	3.015	04:33
Air Blank	0.000	04:34
Control Test	0.082	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 07/05/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: RAMIREZ NOA

DATE/TIME OF TEST: 06/25/2023 05:08:57

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.081	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.207	05:04
Breath Volume	2.339	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:06
Subject Sample # 2	0.198	05:07
Breath Volume	2.132	05:07
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 07/05/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: ALEJANDRO

DATE/TIME OF TEST: 06/26/2023 05:44:49

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.083	05:36
Air Blank	0.000	05:36
Subject Sample # 1	0.114	05:39
Breath Volume	1.144	05:39
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.121	05:43
Breath Volume	1.441	05:43
Air Blank	0.000	05:43
Control Test	0.081	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 07/05/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HOLLORAN,JAMES
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: COSTELLO

DATE/TIME OF TEST: 06/10/2023 17:39:13

Diagnostics Check	OK	17:32
Air Blank	0.000	17:33
Control Test	0.079	17:33
Air Blank	0.000	17:34
Subject Sample # 1	0.179	17:34
Breath Volume	3.097	17:34
Air Blank	0.000	17:35
Air Blank	0.000	17:37
Subject Sample # 2	0.173	17:37
Breath Volume	2.710	17:37
Air Blank	0.000	17:38
Control Test	0.078	17:38
Air Blank	0.000	17:39
Diagnostics Check	OK	17:39

Data Download Date/Time: 07/26/2023 22:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BRACKNEY

DATE/TIME OF TEST: 06/06/2023 03:16:38

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.213	03:12
Breath Volume	3.476	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.220	03:15
Breath Volume	3.714	03:15
Air Blank	0.000	03:15
Control Test	0.077	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 06/08/2023 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 06/11/2023 02:16:17

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:05	
Control Test	0.077	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1 0.110	02:09	Volume Not Met
Breath Volume	0.015	02:09	
Air Blank	0.000	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 2	02:14	No Sample Provided
Breath Volume	0.000	02:14	
Air Blank	0.000	02:15	
Control Test	0.076	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 06/19/2023 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 06/11/2023 02:21:33

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:19	
Control Test	0.078	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Control Test	0.078	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 06/19/2023 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402609610
DRY GAS EXPIRY DATE: 12/09/2025
VIOLATION CODE: DUI
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: VILLEDAS VASQUEZ

DATE/TIME OF TEST: 06/11/2023 09:47:23

Diagnostics Check	OK	09:41
Air Blank	0.000	09:41
Control Test	0.078	09:41
Air Blank	0.000	09:42
Subject Sample # 1	0.199	09:42
Breath Volume	3.585	09:42
Air Blank	0.000	09:43
Air Blank	0.000	09:45
Subject Sample # 2	0.207	09:45
Breath Volume	2.976	09:45
Air Blank	0.000	09:46
Control Test	0.078	09:46
Air Blank	0.000	09:47
Diagnostics Check	OK	09:47

Data Download Date/Time: 06/19/2023 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: JACK

DATE/TIME OF TEST: 06/14/2023 22:32:27

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.078	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.266	22:27
Breath Volume	1.230	22:27
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.252	22:30
Breath Volume	1.636	22:30
Air Blank	0.000	22:31
Control Test	0.078	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 06/28/2023 07:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: ACKER

DATE/TIME OF TEST: 06/25/2023 20:34:12

Diagnostics Check	OK	20:27
Air Blank	0.000	20:28
Control Test	0.078	20:28
Air Blank	0.000	20:29
Subject Sample # 1	0.000	20:29
Breath Volume	3.093	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.000	20:32
Breath Volume	3.523	20:32
Air Blank	0.000	20:33
Control Test	0.078	20:33
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 06/28/2023 07:16