

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MCLEMORE,MICHAEL	ARREST TIME: 00:15
ARRESTING OFFICER AGENCY: Winter Springs PD	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: STACKS	DATE/TIME OF TEST: 07/21/2023 01:18:53
-----------------------------	--

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.248	01:14
Breath Volume	2.167	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.248	01:17
Breath Volume	2.750	01:17
Air Blank	0.000	01:17
Control Test	0.077	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 07/24/2023 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: KRUPINSKI,LUKE	ARREST TIME: 01:07
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: SANTIAGO	DATE/TIME OF TEST: 07/22/2023 02:21:49
-------------------------------	--

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.123	02:17
Breath Volume	2.710	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:19
Subject Sample # 2	0.117	02:20
Breath Volume	3.535	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 07/24/2023 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: DO,LENNI

ARRESTING OFFICER AGENCY: Lake Mary PD

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 23:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 07/22/2023 02:36:18

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:28	
Control Test	0.079	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 1 0.121	02:32	Volume Not Met
Breath Volume	0.277	02:32	
Air Blank	0.000	02:32	
Air Blank	0.000	02:34	
Subject Sample	# 2	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:35	
Control Test	0.079	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 07/24/2023 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GRUNAT,JERRY	ARREST TIME: 02:34
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: HERRERA VAZQUZ	DATE/TIME OF TEST: 07/22/2023	04:30:17
-------------------------------------	-------------------------------	----------

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.153	04:25
Breath Volume	2.843	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.151	04:28
Breath Volume	1.382	04:28
Air Blank	0.000	04:29
Control Test	0.078	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 07/24/2023 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASTRO,JONATHAN,A

BREATH TEST OPERATOR AGENCY: Winter Springs PD

ARRESTING OFFICER: PEREZ,CHRISTIAN

ARRESTING OFFICER AGENCY: Winter Springs PD

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 17:32

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: LAWRANCE

DATE/TIME OF TEST: 07/22/2023 19:02:27

Diagnostics Check	OK	19:01	
Air Blank	0.000	19:02	
Control Test	0.000	19:02	RFI Detect
Air Blank		19:02	RFI Detect

Data Download Date/Time: 07/24/2023 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASTRO,JONATHAN,A

BREATH TEST OPERATOR AGENCY: Winter Springs PD

ARRESTING OFFICER: PEREZ,CHRISTIAN

ARRESTING OFFICER AGENCY: Winter Springs PD

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 17:32

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: LAWRANCE

DATE/TIME OF TEST: 07/22/2023 19:08:34

Diagnostics Check	OK	19:05	
Air Blank	0.000	19:06	
Control Test	0.078	19:06	
Air Blank	0.000	19:06	
Subject Sample	# 1	19:07	RFI Detect
Breath Volume	0.000	19:07	
Air Blank	0.000	19:07	
Control Test	0.078	19:07	
Air Blank	0.000	19:08	
Diagnostics Check	OK	19:08	

Data Download Date/Time: 07/24/2023 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASTRO,DAWN,E

BREATH TEST OPERATOR AGENCY: Winter Springs PD

ARRESTING OFFICER: PEREZ,CHISTIAN

ARRESTING OFFICER AGENCY: Winter Springs PD

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 17:32

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: LAWRANCE

DATE/TIME OF TEST: 07/22/2023 19:13:51

Diagnostics Check	OK	19:11	
Air Blank	0.000	19:12	
Control Test	0.079	19:12	
Air Blank	0.000	19:12	
Subject Sample	# 1	19:13	RFI Detect
Breath Volume	0.000	19:13	
Air Blank		19:13	RFI Detect
Air Blank	0.000	19:13	

Data Download Date/Time: 07/24/2023 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASTRO,JONATHAN,A

BREATH TEST OPERATOR AGENCY: Winter Springs PD

ARRESTING OFFICER: PEREZ,CHRISTIAN

ARRESTING OFFICER AGENCY: Winter Springs PD

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 17:32

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: LAWRANCE

DATE/TIME OF TEST: 07/22/2023 19:18:47

Diagnostics Check	OK	19:16	
Air Blank	0.000	19:16	
Control Test	0.079	19:17	
Air Blank	0.000	19:17	
Subject Sample	# 1	19:18	RFI Detect
Breath Volume	0.000	19:18	
Air Blank		19:18	RFI Detect
Air Blank	0.000	19:18	

Data Download Date/Time: 07/24/2023 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CASTRO,JONATHAN,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Winter Springs PD	VIOLATION CODE: DUI
ARRESTING OFFICER: PEREZ,CHRISTIAN	ARREST TIME: 17:32
ARRESTING OFFICER AGENCY: Winter Springs PD	GENDER: Female
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: LAWRANCE DATE/TIME OF TEST: 07/22/2023 19:29:28

Diagnostics Check	OK	19:21
Air Blank	0.000	19:22
Control Test	0.079	19:22
Air Blank	0.000	19:22
Subject Sample # 1	0.196	19:24
Breath Volume	1.679	19:24
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.213	19:28
Breath Volume	2.214	19:28
Air Blank	0.000	19:28
Control Test	0.078	19:28
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 07/24/2023 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GUILLEN-LAVA,F	ARREST TIME: 18:56
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: JENKINS	DATE/TIME OF TEST: 07/23/2023	03:12:11
------------------------------	-------------------------------	----------

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.080	02:57	
Air Blank	0.000	02:57	
Subject Sample	# 1 0.058	02:59	
Breath Volume	1.394	02:59	
Air Blank	0.000	03:00	
Air Blank	0.000	03:01	
Subject Sample	# 2 0.050	03:05	Volume Not Met
Breath Volume	1.027	03:05	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 3 0.053	03:10	Volume Not Met
Breath Volume	1.054	03:10	
Air Blank	0.000	03:11	
Control Test	0.078	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 07/24/2023 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GUILLEN-LAVA,F	ARREST TIME: 18:56
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: JENKINS	DATE/TIME OF TEST: 07/23/2023 03:33:51
------------------------------	--

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.079	03:19	
Air Blank	0.000	03:20	
Subject Sample # 1	0.055	03:21	
Breath Volume	1.378	03:21	
Air Blank	0.000	03:21	
Air Blank	0.000	03:23	
Subject Sample # 2	0.046	03:26	Volume Not Met
Breath Volume	0.328	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample # 3	0.044	03:32	Volume Not Met
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.079	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 07/24/2023 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CHINEA,TAMY,

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: GRUNAT,J

ARRESTING OFFICER AGENCY: Seminole County SO

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 02:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: HIGHSMITH

DATE/TIME OF TEST: 07/23/2023 04:22:40

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.080	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.103	04:17
Breath Volume	1.835	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.097	04:21
Breath Volume	1.386	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 07/24/2023 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANROMPA,V,
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: SANDOVAL,R
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: APODACA

DATE/TIME OF TEST: 07/23/2023 05:53:47

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.080	05:47
Air Blank	0.000	05:48
Subject Sample # 1	0.155	05:49
Breath Volume	1.488	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.157	05:52
Breath Volume	1.179	05:52
Air Blank	0.000	05:52
Control Test	0.079	05:53
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 07/24/2023 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: SIMMENS,ADAM,J	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: FRANZE,FRANK	ARREST TIME: 10:30
ARRESTING OFFICER AGENCY: Oviedo PD	GENDER: Female
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: VRLJANOVIC	DATE/TIME OF TEST: 07/25/2023 11:33:19
---------------------------------	--

Diagnostics Check	OK	11:26
Air Blank	0.000	11:26
Control Test	0.080	11:26
Air Blank	0.000	11:27
Subject Sample # 1	0.129	11:28
Breath Volume	2.031	11:28
Air Blank	0.000	11:28
Air Blank	0.000	11:30
Subject Sample # 2	0.127	11:31
Breath Volume	1.730	11:31
Air Blank	0.000	11:32
Control Test	0.079	11:32
Air Blank	0.000	11:33
Diagnostics Check	OK	11:33

Data Download Date/Time: 08/03/2023 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SMITH,LOGAN	ARREST TIME: 01:47
ARRESTING OFFICER AGENCY: Oviedo PD	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BUERK	DATE/TIME OF TEST: 07/26/2023	02:46:13
----------------------------	-------------------------------	----------

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.159	02:41
Breath Volume	1.714	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.149	02:44
Breath Volume	1.988	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 08/03/2023 04:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NOLAND,THOMAS,E

BREATH TEST OPERATOR AGENCY: Lake Mary PD

ARRESTING OFFICER: ROCHE,MARK

ARRESTING OFFICER AGENCY: Seminole County SO

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 14:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: GHAZZAWIEH

DATE/TIME OF TEST: 07/26/2023 16:21:42

Diagnostics Check	OK	16:15
Air Blank	0.000	16:15
Control Test	0.080	16:16
Air Blank	0.000	16:16
Subject Sample # 1	0.000	16:17
Breath Volume	1.328	16:17
Air Blank	0.000	16:17
Air Blank	0.000	16:19
Subject Sample # 2	0.000	16:20
Breath Volume	1.136	16:20
Air Blank	0.000	16:20
Control Test	0.080	16:21
Air Blank	0.000	16:21
Diagnostics Check	OK	16:21

Data Download Date/Time: 08/03/2023 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: RODRIGUEZ,E	ARREST TIME: 00:39
ARRESTING OFFICER AGENCY: Altamonte Springs PD	GENDER: Female
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: KLINGER	DATE/TIME OF TEST: 07/27/2023	01:50:44
------------------------------	-------------------------------	----------

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.098	01:45
Breath Volume	1.609	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.100	01:49
Breath Volume	2.011	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 08/03/2023 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BETHAM,KEITH,D	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SERRANO,JIMMY	ARREST TIME: 22:52
ARRESTING OFFICER AGENCY: Oviedo PD	GENDER: Female
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: FATIMA	DATE/TIME OF TEST: 07/28/2023 00:16:10
-----------------------------	--

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:05	
Control Test	0.080	00:05	
Air Blank	0.000	00:06	
Subject Sample	# 1 0.231	00:09	Volume Not Met
Breath Volume	0.093	00:09	
Air Blank	0.000	00:09	
Air Blank	0.000	00:11	
Subject Sample	# 2	00:14	No Sample Provided
Breath Volume	0.000	00:14	
Air Blank	0.000	00:15	
Control Test	0.078	00:15	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 08/03/2023 04:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: DIAL,VICTORIA

ARRESTING OFFICER AGENCY: Altamonte Springs PD

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 00:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: FREDDA

DATE/TIME OF TEST: 07/28/2023 01:48:59

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.068	01:44
Breath Volume	1.761	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.063	01:47
Breath Volume	1.125	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 08/03/2023 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PRICE,CHASE	ARREST TIME: 01:06
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: CARLEY	DATE/TIME OF TEST: 07/29/2023	02:53:56
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.141	02:49
Breath Volume	1.921	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.134	02:52
Breath Volume	2.628	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 08/03/2023 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: KOUROUMOUSIS,GEORGE	ARREST TIME: 04:59
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: SATONES	DATE/TIME OF TEST: 07/29/2023 06:34:10
------------------------------	--

Diagnostics Check	OK	06:27
Air Blank	0.000	06:28
Control Test	0.080	06:28
Air Blank	0.000	06:29
Subject Sample # 1	0.156	06:29
Breath Volume	2.886	06:29
Air Blank	0.000	06:30
Air Blank	0.000	06:32
Subject Sample # 2	0.142	06:32
Breath Volume	2.605	06:32
Air Blank	0.000	06:33
Control Test	0.080	06:33
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

Data Download Date/Time: 08/03/2023 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SMITH,LOGAN	ARREST TIME: 00:59
ARRESTING OFFICER AGENCY: Oviedo PD	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: YOUNG	DATE/TIME OF TEST: 07/30/2023 02:03:41
----------------------------	--

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.257	01:59
Breath Volume	2.128	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.253	02:02
Breath Volume	1.625	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 08/03/2023 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PITTA,C	ARREST TIME: 03:46
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: SALISBURY	DATE/TIME OF TEST: 07/30/2023 05:09:48
--------------------------------	--

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.079	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.245	05:05
Breath Volume	1.703	05:05
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.235	05:08
Breath Volume	1.277	05:08
Air Blank	0.000	05:08
Control Test	0.079	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 08/03/2023 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: DRACH,S	ARREST TIME: 00:46
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MOREHEAD	DATE/TIME OF TEST: 07/31/2023	02:03:41
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.229	01:59
Breath Volume	1.195	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.212	02:02
Breath Volume	1.667	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 08/03/2023 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,GEORGE,
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: MICHAELS,GLEN
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 06/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: TERRELL

DATE/TIME OF TEST: 07/14/2023 02:15:17

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.264	02:10
Breath Volume	4.007	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.264	02:13
Breath Volume	3.492	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 08/01/2023 00:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HADDOCKS,KEVIN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 07/25/2023 16:49:01

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.079	16:41
Air Blank	0.000	16:41
Subject Sample # 1	0.290	16:44
Breath Volume	1.230	16:44
Air Blank	0.000	16:44
Air Blank	0.000	16:46
Subject Sample # 2	0.290	16:47
Breath Volume	1.707	16:47
Air Blank	0.000	16:48
Control Test	0.078	16:48
Air Blank	0.000	16:48
Diagnostics Check	OK	16:49

Data Download Date/Time: 08/15/2023 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,BOBBY,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARRISON,BOBBY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: FLOREA

DATE/TIME OF TEST: 07/03/2023 01:59:14

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:45	
Control Test	0.080	01:45	
Air Blank	0.000	01:46	
Subject Sample	# 1 0.211	01:49	Volume Not Met
Breath Volume	0.992	01:49	
Air Blank	0.000	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 2 0.200	01:53	
Breath Volume	1.773	01:53	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample	# 3 0.211	01:57	
Breath Volume	2.437	01:57	
Air Blank	0.000	01:58	
Control Test	0.076	01:58	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 07/05/2023 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLTON,JUSTIN,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOLTON,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: DUKAS

DATE/TIME OF TEST: 07/03/2023 21:06:55

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.082	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.000	21:02
Breath Volume	1.117	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.000	21:05
Breath Volume	1.164	21:05
Air Blank	0.000	21:05
Control Test	0.081	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 07/05/2023 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOMAN,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HOMAN,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: BUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 07/01/2023 19:37:59

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:35	
Control Test	0.080	19:35	
Air Blank	0.000	19:36	
Subject Sample	# 1	19:36	Subject Test Refused
Breath Volume	0.000	19:36	
Air Blank	0.000	19:37	
Control Test	0.079	19:37	
Air Blank	0.000	19:37	
Diagnostics Check	OK	19:37	

Data Download Date/Time: 07/11/2023 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASLEY,JACOB,W
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: PASLEY,JACOB
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: FRENCH-POU

DATE/TIME OF TEST: 07/08/2023 05:58:05

Diagnostics Check	OK	05:51
Air Blank	0.000	05:52
Control Test	0.080	05:52
Air Blank	0.000	05:52
Subject Sample # 1	0.165	05:53
Breath Volume	2.593	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.172	05:56
Breath Volume	1.824	05:56
Air Blank	0.000	05:57
Control Test	0.080	05:57
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 07/11/2023 18:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LEIGHA,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: JONES,LEIGHA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:03

SUBJECT'S LAST NAME: MORENO MARADIAGA

DATE/TIME OF TEST: 07/04/2023 11:44:52

Diagnostics Check	OK	11:36
Air Blank	0.000	11:37
Control Test	0.081	11:37
Air Blank	0.000	11:37
Subject Sample # 1	0.206	11:39
Breath Volume	1.695	11:39
Air Blank	0.000	11:39
Air Blank	0.000	11:41
Subject Sample # 2	0.213	11:43
Breath Volume	1.222	11:43
Air Blank	0.000	11:43
Control Test	0.080	11:44
Air Blank	0.000	11:44
Diagnostics Check	OK	11:44

Data Download Date/Time: 07/11/2023 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TOBY,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FLAHERTY,ZACHARY
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 06/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: GADDY

DATE/TIME OF TEST: 07/06/2023 19:55:37

Diagnostics Check	OK	19:48
Air Blank	0.000	19:49
Control Test	0.080	19:49
Air Blank	0.000	19:49
Subject Sample # 1	0.154	19:50
Breath Volume	2.382	19:50
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.149	19:54
Breath Volume	2.453	19:54
Air Blank	0.000	19:54
Control Test	0.079	19:55
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 07/11/2023 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABDALE,JAMES,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: ROBINSON,RICKY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: HATCH

DATE/TIME OF TEST: 07/11/2023 19:09:09

Diagnostics Check	OK	19:02
Air Blank	0.000	19:02
Control Test	0.081	19:03
Air Blank	0.000	19:03
Subject Sample # 1	0.232	19:04
Breath Volume	1.851	19:04
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.226	19:07
Breath Volume	1.726	19:07
Air Blank	0.000	19:08
Control Test	0.079	19:08
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 08/22/2023 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TOBY,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CLEMENTS,TOBY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: WIDEMAN

DATE/TIME OF TEST: 07/11/2023 20:18:26

Diagnostics Check	OK	20:12
Air Blank	0.000	20:12
Control Test	0.080	20:12
Air Blank	0.000	20:13
Subject Sample # 1	0.227	20:14
Breath Volume	2.242	20:14
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.223	20:16
Breath Volume	2.265	20:16
Air Blank	0.000	20:17
Control Test	0.079	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 08/22/2023 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: GONZALEZ,JOSE
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: DILLARD

DATE/TIME OF TEST: 07/13/2023 04:21:18

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.183	04:16
Breath Volume	2.027	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.167	04:19
Breath Volume	1.925	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 08/22/2023 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: SAPUNDJIEFF,THOMAS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: PEDROSA

DATE/TIME OF TEST: 07/14/2023 00:34:39

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.097	00:30
Breath Volume	1.375	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.092	00:33
Breath Volume	1.414	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 08/22/2023 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LEIGHA,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: JONES,LEIGHA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: BETHEL

DATE/TIME OF TEST: 07/17/2023 17:29:40

Diagnostics Check	OK	17:23
Air Blank	0.000	17:23
Control Test	0.079	17:24
Air Blank	0.000	17:24
Subject Sample # 1	0.243	17:25
Breath Volume	1.671	17:25
Air Blank	0.000	17:25
Air Blank	0.000	17:27
Subject Sample # 2	0.248	17:28
Breath Volume	1.367	17:28
Air Blank	0.000	17:28
Control Test	0.079	17:29
Air Blank	0.000	17:29
Diagnostics Check	OK	17:29

Data Download Date/Time: 08/22/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/18/2023 19:47:57

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.080	19:42
Air Blank	0.000	19:42
Subject Sample # 1	0.180	19:43
Breath Volume	1.785	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.183	19:46
Breath Volume	2.476	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 08/22/2023 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: GONZALEZ,JOSE
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: EBANKS

DATE/TIME OF TEST: 07/24/2023 03:17:29

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.178	03:12
Breath Volume	1.847	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.170	03:15
Breath Volume	1.894	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 08/22/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TOBY,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CLEMENTS,TOBY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/30/2023 00:17:28

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:15	
Control Test	0.081	00:15	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:16	
Control Test	0.081	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 08/22/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUTTERLY,BROOKE,L
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MINCK,ISAAC
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: GADDY

DATE/TIME OF TEST: 07/30/2023 02:18:14

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.222	02:13
Breath Volume	2.058	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.223	02:16
Breath Volume	3.371	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 08/22/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: TELLE,MADDISON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 07/30/2023 04:31:55

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.132	04:27
Breath Volume	2.710	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample # 2	0.138	04:30
Breath Volume	1.984	04:30
Air Blank	0.000	04:30
Control Test	0.081	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 08/22/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AULTMAN,BRENDON,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SKELLY,RACHEL
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: MARCHAN

DATE/TIME OF TEST: 07/01/2023 07:05:31

Diagnostics Check	OK	06:57	
Air Blank	0.000	06:58	
Control Test	0.080	06:58	
Air Blank	0.000	06:58	
Subject Sample	# 1 0.154	07:00	
Breath Volume	1.671	07:00	
Air Blank	0.000	07:01	
Air Blank	0.000	07:02	
Subject Sample	# 2	07:03	Subject Test Refused
Breath Volume	0.421	07:03	
Air Blank	0.000	07:04	
Control Test	0.080	07:04	
Air Blank		07:05	RFI Detect
Air Blank	0.000	07:05	

Data Download Date/Time: 07/13/2023 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRAGUE,JONATHON,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SPRAGUE,JONATHON
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: SECKLER

DATE/TIME OF TEST: 07/01/2023 23:15:22

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:12	
Control Test	0.079	23:13	
Air Blank	0.000	23:13	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:14	
Control Test	0.079	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 07/13/2023 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DEBORD

DATE/TIME OF TEST: 07/09/2023 01:17:20

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.125	01:12
Breath Volume	3.015	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.131	01:15
Breath Volume	2.386	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 07/13/2023 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DEWIT,HAROLD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 07/14/2023 23:43:16

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.000	23:38
Breath Volume	1.687	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.000	23:41
Breath Volume	1.964	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 08/09/2023 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: LUJAN

DATE/TIME OF TEST: 07/16/2023 05:14:59

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:09
Subject Sample # 1	0.146	05:10
Breath Volume	1.976	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:12
Subject Sample # 2	0.142	05:13
Breath Volume	1.671	05:13
Air Blank	0.000	05:14
Control Test	0.079	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 08/09/2023 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: EDGE

DATE/TIME OF TEST: 07/18/2023 00:41:53

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:40	Subject Test Refused
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Control Test	0.079	00:41	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 08/09/2023 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 07/19/2023 16:28:59

Diagnostics Check	OK	16:26	
Air Blank	0.000	16:26	
Control Test	0.079	16:26	
Air Blank	0.000	16:27	
Subject Sample	# 1	16:27	Subject Test Refused
Breath Volume	0.000	16:27	
Air Blank	0.000	16:28	
Control Test	0.079	16:28	
Air Blank	0.000	16:28	
Diagnostics Check	OK	16:28	

Data Download Date/Time: 08/09/2023 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 07/29/2023 02:35:08

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.080	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1 0.062	02:33	
Breath Volume	1.359	02:33	
Air Blank		02:33	RFI Detect
Air Blank	0.000	02:34	
Control Test	0.079	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 08/09/2023 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 07/29/2023 02:39:43

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.078	02:37	
Air Blank	0.000	02:37	
Subject Sample	# 1 0.062	02:38	
Breath Volume	1.511	02:38	
Air Blank		02:38	RFI Detect
Air Blank	0.000	02:38	
Control Test	0.080	02:39	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 08/09/2023 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALENCIA, GUILLERMO, S
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: WILLIAMS, K
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 07/01/2023 00:45:32

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.282	00:39
Breath Volume	2.714	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.264	00:43
Breath Volume	1.324	00:43
Air Blank	0.000	00:44
Control Test	0.076	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARX,CHRISTOHER,P
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WILLIAMS,K
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 07/01/2023 23:39:47

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.079	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.169	23:35
Breath Volume	1.460	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.160	23:38
Breath Volume	1.589	23:38
Air Blank	0.000	23:38
Control Test	0.079	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LORD,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BILLIOT

DATE/TIME OF TEST: 07/03/2023 02:20:02

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.087	02:15
Breath Volume	1.320	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.084	02:18
Breath Volume	1.214	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:20

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARX,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: O'CONNERY,N
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 07/05/2023 20:55:41

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:49	
Control Test	0.080	20:49	
Air Blank	0.000	20:49	
Subject Sample # 1	0.199	20:51	
Breath Volume	1.906	20:51	
Air Blank	0.000	20:52	
Air Blank		20:53	Improper Sample
Air Blank	0.000	20:54	
Control Test	0.081	20:55	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARX,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: O'CONNERY,N
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 07/05/2023 21:06:20

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.079	21:00
Air Blank	0.000	21:01
Subject Sample # 1	0.207	21:01
Breath Volume	1.691	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.204	21:04
Breath Volume	1.988	21:04
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,ANGELA,F
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: STOFEL,WESLEY
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 07/07/2023 18:39:32

Diagnostics Check	OK	18:32
Air Blank	0.000	18:33
Control Test	0.079	18:33
Air Blank	0.000	18:34
Subject Sample # 1	0.211	18:34
Breath Volume	3.000	18:34
Air Blank	0.000	18:35
Air Blank	0.000	18:37
Subject Sample # 2	0.210	18:37
Breath Volume	2.640	18:37
Air Blank	0.000	18:38
Control Test	0.079	18:38
Air Blank	0.000	18:39
Diagnostics Check	OK	18:39

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: SMITH JR,LORNE
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BENITES

DATE/TIME OF TEST: 07/13/2023 01:08:55

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.221	01:00
Breath Volume	2.390	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.199	01:04
Breath Volume	2.390	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 3	0.208	01:07
Breath Volume	3.617	01:07
Air Blank	0.000	01:07
Control Test	0.077	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,EDWARD,E
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WILLIAMS,K
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 07/16/2023 00:20:32

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:08	
Control Test	0.080	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1 0.130	00:12	Volume Not Met
Breath Volume	0.625	00:12	
Air Blank	0.000	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 2 0.126	00:15	
Breath Volume	2.058	00:15	
Air Blank	0.000	00:16	
Air Blank	0.000	00:17	
Subject Sample	# 3 0.131	00:18	
Breath Volume	1.812	00:18	
Air Blank	0.000	00:19	
Control Test	0.078	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,JUSTIN,B
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: BONK,JONATHAN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/16/2023 07:00:07

Diagnostics Check	OK	06:52
Air Blank	0.000	06:52
Control Test	0.079	06:53
Air Blank	0.000	06:53
Subject Sample # 1	0.075	06:55
Breath Volume	2.238	06:55
Air Blank	0.000	06:56
Air Blank	0.000	06:57
Subject Sample # 2	0.067	06:58
Breath Volume	1.566	06:58
Air Blank	0.000	06:59
Control Test	0.077	06:59
Air Blank	0.000	07:00
Diagnostics Check	OK	07:00

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSKO, KRISTOPHER, R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO, KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BECERRA

DATE/TIME OF TEST: 07/22/2023 03:03:45

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.044	02:59
Breath Volume	2.410	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.041	03:02
Breath Volume	2.371	03:02
Air Blank	0.000	03:02
Control Test	0.079	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: STAHL

DATE/TIME OF TEST: 07/22/2023 08:59:52

Diagnostics Check	OK	08:59	
Air Blank		08:59	RFI Detect
Air Blank	0.000	08:59	

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: STAHL

DATE/TIME OF TEST: 07/22/2023 09:07:27

Diagnostics Check	OK	09:01
Air Blank	0.000	09:01
Control Test	0.080	09:02
Air Blank	0.000	09:02
Subject Sample # 1	0.169	09:03
Breath Volume	1.312	09:03
Air Blank	0.000	09:03
Air Blank	0.000	09:05
Subject Sample # 2	0.170	09:05
Breath Volume	1.226	09:05
Air Blank	0.000	09:06
Control Test	0.078	09:06
Air Blank	0.000	09:07
Diagnostics Check	OK	09:07

Data Download Date/Time: 07/28/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: ANDRADE SOTO

DATE/TIME OF TEST: 07/14/2023 04:39:52

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.234	04:34
Breath Volume	3.117	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.226	04:38
Breath Volume	3.085	04:38
Air Blank	0.000	04:38
Control Test	0.078	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 07/29/2023 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRO,MARCEL,
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: ELLIOTT,NIGEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 07/15/2023 16:30:33

Diagnostics Check	OK	16:23
Air Blank	0.000	16:24
Control Test	0.079	16:24
Air Blank	0.000	16:25
Subject Sample # 1	0.192	16:25
Breath Volume	2.074	16:25
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.196	16:28
Breath Volume	2.167	16:28
Air Blank	0.000	16:29
Control Test	0.075	16:29
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 07/29/2023 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GUERRY, C
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:02

SUBJECT'S LAST NAME: BALTZ

DATE/TIME OF TEST: 07/19/2023 11:30:20

Diagnostics Check	OK	11:24
Air Blank	0.000	11:24
Control Test	0.080	11:24
Air Blank	0.000	11:25
Subject Sample # 1	0.013	11:25
Breath Volume	2.179	11:25
Air Blank	0.000	11:26
Air Blank	0.000	11:28
Subject Sample # 2	0.014	11:28
Breath Volume	3.093	11:28
Air Blank	0.000	11:29
Control Test	0.080	11:29
Air Blank	0.000	11:30
Diagnostics Check	OK	11:30

Data Download Date/Time: 07/29/2023 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,NATHAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SUMNER,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: SPRADLEY

DATE/TIME OF TEST: 07/26/2023 23:26:46

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.080	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.158	23:22
Breath Volume	4.355	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.154	23:25
Breath Volume	4.230	23:25
Air Blank	0.000	23:25
Control Test	0.078	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 07/29/2023 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSKO,KRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/27/2023 02:39:46

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.080	02:27	
Air Blank	0.000	02:28	
Subject Sample	# 1 0.041	02:31	Volume Not Met
Breath Volume	0.312	02:31	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 2 0.039	02:34	
Breath Volume	1.328	02:34	
Air Blank	0.000	02:35	
Air Blank	0.000	02:36	
Subject Sample	# 3 0.038	02:38	
Breath Volume	1.613	02:38	
Air Blank	0.000	02:38	
Control Test	0.078	02:39	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 07/29/2023 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CASON, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DECOURCY

DATE/TIME OF TEST: 07/29/2023 00:07:38

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.160	00:03
Breath Volume	2.468	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.152	00:05
Breath Volume	1.488	00:05
Air Blank	0.000	00:06
Control Test	0.082	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 08/15/2023 20:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402328118
BREATH TEST OPERATOR: PASLEY,JACOB,W	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: Washington County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: KELLY,JOHN	ARREST TIME: 23:09
ARRESTING OFFICER AGENCY: Washington County SO	GENDER: Male
LAST AGENCY INSPECTION: 07/07/2023	OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: BEST	DATE/TIME OF TEST: 07/15/2023 00:29:12
---------------------------	--

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.081	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.000	00:25
Breath Volume	3.125	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.000	00:27
Breath Volume	3.390	00:27
Air Blank	0.000	00:28
Control Test	0.081	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 08/17/2023 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: ELDER

DATE/TIME OF TEST: 07/01/2023 03:10:33

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.082	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.082	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 07/03/2023 19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: WHITAKER

DATE/TIME OF TEST: 07/04/2023 05:58:39

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.083	05:51
Air Blank	0.000	05:52
Subject Sample # 1	0.163	05:53
Breath Volume	3.785	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:55
Subject Sample # 2	0.163	05:56
Breath Volume	4.246	05:56
Air Blank	0.000	05:57
Control Test	0.080	05:58
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 08/13/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RADCLIFF, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: TALLENT

DATE/TIME OF TEST: 07/09/2023 01:41:38

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.082	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.150	01:36
Breath Volume	3.609	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.142	01:39
Breath Volume	3.949	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 08/13/2023 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: HOOKER

DATE/TIME OF TEST: 07/09/2023 17:31:04

Diagnostics Check	OK	17:22
Air Blank	0.000	17:23
Control Test	0.083	17:23
Air Blank	0.000	17:23
Subject Sample # 1	0.294	17:25
Breath Volume	1.605	17:25
Air Blank	0.000	17:26
Air Blank	0.000	17:27
Subject Sample # 2	0.292	17:28
Breath Volume	1.593	17:28
Air Blank	0.000	17:30
Control Test	0.078	17:30
Air Blank	0.000	17:30
Diagnostics Check	OK	17:31

Data Download Date/Time: 08/13/2023 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JONES,DALLAS,L

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: JONES,DALLAS

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 2402328118

DRY GAS EXPIRY DATE: 01/20/2025

VIOLATION CODE: DUI

ARREST TIME: 02:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 07/02/2023 03:09:38

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.081	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.286	03:01
Breath Volume	1.511	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.323	03:04
Breath Volume	1.359	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 3	0.328	03:07
Breath Volume	2.214	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 07/11/2023 18:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/06/2023 00:44:51

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.081	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:43	Slope Not Met
Breath Volume	2.507	00:43	
Air Blank	0.000	00:43	
Control Test	0.082	00:44	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 07/11/2023 18:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 2402328118
BREATH TEST OPERATOR: HENSON,JAY,C	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: HENSON,JAY	ARREST TIME: 23:57
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2023	OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MITCHELL	DATE/TIME OF TEST: 07/06/2023	00:54:10
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.081	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.298	00:49
Breath Volume	3.812	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.307	00:52
Breath Volume	3.734	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 07/11/2023 18:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 2402328118
BREATH TEST OPERATOR: CHASON,HUNTER,B	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: CHASON,HUNTER	ARREST TIME: 01:59
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Female
LAST AGENCY INSPECTION: 07/11/2023	OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: HENDRIX	DATE/TIME OF TEST: 07/16/2023	03:46:46
------------------------------	-------------------------------	----------

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.080	03:42
Breath Volume	2.660	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.082	03:45
Breath Volume	3.101	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 08/17/2023 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,DALLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: JUNGBLOM

DATE/TIME OF TEST: 07/16/2023 19:15:10

Diagnostics Check	OK	19:11	
Air Blank	0.000	19:11	
Control Test	0.080	19:12	
Air Blank	0.000	19:12	
Subject Sample	# 1	19:13	Subject Test Refused
Breath Volume	0.000	19:13	
Air Blank	0.000	19:14	
Control Test	0.081	19:14	
Air Blank	0.000	19:15	
Diagnostics Check	OK	19:15	

Data Download Date/Time: 08/17/2023 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRENNON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SABATINE,MATTHEW
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 06/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 12:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: WINFREE

DATE/TIME OF TEST: 07/01/2023 13:37:29

Diagnostics Check	OK	13:26	
Air Blank	0.000	13:26	
Control Test	0.080	13:26	
Air Blank	0.000	13:27	
Subject Sample	# 1 0.007	13:30	Volume Not Met
Breath Volume	0.769	13:30	
Air Blank	0.000	13:31	
Air Blank	0.000	13:32	
Subject Sample	# 2 0.000	13:36	Volume Not Met
Breath Volume	0.464	13:36	
Air Blank	0.000	13:36	
Control Test	0.081	13:36	
Air Blank	0.000	13:37	
Diagnostics Check	OK	13:37	

Data Download Date/Time: 07/07/2023 00:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRENNON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: GOMILLION,LANCE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 07:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: WALDRON

DATE/TIME OF TEST: 07/02/2023 07:49:51

Diagnostics Check	OK	07:43
Air Blank	0.000	07:43
Control Test	0.080	07:43
Air Blank	0.000	07:44
Subject Sample # 1	0.000	07:45
Breath Volume	2.617	07:45
Air Blank	0.000	07:45
Air Blank	0.000	07:47
Subject Sample # 2	0.000	07:48
Breath Volume	2.128	07:48
Air Blank	0.000	07:48
Control Test	0.080	07:49
Air Blank	0.000	07:49
Diagnostics Check	OK	07:49

Data Download Date/Time: 07/07/2023 00:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: THOMPSON,WILLIAM
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: LAWERENCE

DATE/TIME OF TEST: 07/06/2023 21:59:58

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.079	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.000	21:55
Breath Volume	1.281	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.000	21:58
Breath Volume	2.074	21:58
Air Blank	0.000	21:59
Control Test	0.081	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 07/07/2023 00:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS,BLAINE,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: AKIN,STERLING
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 07:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: CASON

DATE/TIME OF TEST: 07/22/2023 08:10:22

Diagnostics Check	OK	08:04
Air Blank	0.000	08:04
Control Test	0.079	08:05
Air Blank	0.000	08:05
Subject Sample # 1	0.000	08:06
Breath Volume	1.222	08:06
Air Blank	0.000	08:06
Air Blank	0.000	08:08
Subject Sample # 2	0.000	08:09
Breath Volume	1.367	08:09
Air Blank	0.000	08:09
Control Test	0.079	08:09
Air Blank	0.000	08:10
Diagnostics Check	OK	08:10

Data Download Date/Time: 08/06/2023 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRENNON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: PIAGET,PHILIP
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/24/2023 08:16:27

Diagnostics Check	OK	08:14	
Air Blank	0.000	08:14	
Control Test	0.080	08:14	
Air Blank	0.000	08:15	
Subject Sample	# 1	08:15	RFI Detect
Breath Volume	0.000	08:15	
Air Blank		08:15	RFI Detect
Air Blank	0.000	08:16	

Data Download Date/Time: 08/06/2023 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRENNON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: PIAGET,PHILIP
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/24/2023 08:20:07

Diagnostics Check	OK	08:19	
Air Blank		08:19	RFI Detect
Air Blank	0.000	08:20	

Data Download Date/Time: 08/06/2023 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRENNON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: PIAGET,PHILIP
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/24/2023 08:27:56

Diagnostics Check	OK	08:22	
Air Blank	0.000	08:23	
Control Test	0.080	08:23	
Air Blank	0.000	08:24	
Subject Sample # 1	0.000	08:24	
Breath Volume	2.300	08:24	
Air Blank	0.000	08:24	
Air Blank		08:26	RFI Detect
Air Blank	0.000	08:27	
Control Test	0.080	08:27	
Air Blank		08:27	RFI Detect
Air Blank	0.000	08:27	

Data Download Date/Time: 08/06/2023 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,WESLEY,A
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: SABATINE,MATTHEW
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 07/29/2023 20:58:32

Diagnostics Check	OK	20:51
Air Blank	0.000	20:52
Control Test	0.080	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.144	20:53
Breath Volume	2.046	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.138	20:56
Breath Volume	2.167	20:56
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 08/06/2023 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: WATSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: GARCIA MARTINEZ

DATE/TIME OF TEST: 07/30/2023 16:23:33

Diagnostics Check	OK	16:20	
Air Blank	0.000	16:21	
Control Test	0.080	16:21	
Air Blank	0.000	16:22	
Subject Sample	# 1	16:22	Subject Test Refused
Breath Volume	0.000	16:22	
Air Blank	0.000	16:22	
Control Test	0.079	16:22	
Air Blank	0.000	16:23	
Diagnostics Check	OK	16:23	

Data Download Date/Time: 08/06/2023 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: MEMON,MATTHEW
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 07/01/2023 01:18:09

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.250	01:13
Breath Volume	1.503	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample # 2	0.264	01:16
Breath Volume	1.777	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 07/26/2023 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAKE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/02/2023 23:58:10

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.081	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.199	23:53
Breath Volume	1.371	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.208	23:56
Breath Volume	1.902	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 07/26/2023 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA, ADAEL,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ACOSTA, ADAEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: TROCHEZ LOPEZ

DATE/TIME OF TEST: 07/09/2023 00:27:31

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.081	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.203	00:21
Breath Volume	1.914	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.198	00:25
Breath Volume	1.824	00:25
Air Blank	0.000	00:26
Control Test	0.077	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 07/26/2023 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPUTO,TAYLOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: CAPUTO,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MOISE

DATE/TIME OF TEST: 07/13/2023 03:11:26

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.191	03:07
Breath Volume	1.511	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.192	03:09
Breath Volume	1.871	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 07/26/2023 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTO,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: RESTO,LUIS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MUNOZ MARTINEZ

DATE/TIME OF TEST: 07/14/2023 01:36:04

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.081	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.192	01:31
Breath Volume	2.722	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:33
Subject Sample # 2	0.186	01:34
Breath Volume	1.535	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:36

Data Download Date/Time: 07/26/2023 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MCWILLIAMS,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:23

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/14/2023 08:50:07

Diagnostics Check	OK	08:43
Air Blank	0.000	08:44
Control Test	0.081	08:44
Air Blank	0.000	08:45
Subject Sample # 1	0.103	08:45
Breath Volume	2.410	08:45
Air Blank	0.000	08:46
Air Blank	0.000	08:48
Subject Sample # 2	0.108	08:48
Breath Volume	2.667	08:48
Air Blank	0.000	08:49
Control Test	0.081	08:49
Air Blank	0.000	08:50
Diagnostics Check	OK	08:50

Data Download Date/Time: 07/26/2023 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTO,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: RESTO,LUS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: MARINHO DE ARAUJO

DATE/TIME OF TEST: 07/16/2023 06:51:01

Diagnostics Check	OK	06:44
Air Blank	0.000	06:45
Control Test	0.081	06:45
Air Blank	0.000	06:45
Subject Sample # 1	0.211	06:46
Breath Volume	1.824	06:46
Air Blank	0.000	06:47
Air Blank	0.000	06:49
Subject Sample # 2	0.198	06:49
Breath Volume	1.320	06:49
Air Blank	0.000	06:50
Control Test	0.080	06:50
Air Blank	0.000	06:50
Diagnostics Check	OK	06:51

Data Download Date/Time: 07/26/2023 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: RAMIREZ GARCIA

DATE/TIME OF TEST: 07/17/2023 03:18:19

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1 0.139	03:14	Slope Not Level
Breath Volume	1.203	03:14	
Air Blank	0.000	03:14	
Air Blank	0.000	03:16	
Subject Sample	# 2	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Control Test	0.078	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 07/26/2023 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: BAEZA LANDIN

DATE/TIME OF TEST: 07/23/2023 01:56:12

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.082	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.187	01:51
Breath Volume	3.445	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.190	01:54
Breath Volume	3.511	01:54
Air Blank	0.000	01:55
Control Test	0.081	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 07/26/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MARTINEZ PEREZ

DATE/TIME OF TEST: 07/29/2023 02:23:10

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.194	02:18
Breath Volume	1.671	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.196	02:21
Breath Volume	2.066	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPUTO,TAYLOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: CAPUTO,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 07/31/2023 01:56:41

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.081	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.186	01:52
Breath Volume	2.113	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.195	01:55
Breath Volume	2.761	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: SAMS

DATE/TIME OF TEST: 07/01/2023 03:01:36

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.081	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Control Test	0.080	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 07/06/2023 03:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: DOMINGUEZ-SUAREZ

DATE/TIME OF TEST: 07/07/2023 10:48:46

Diagnostics Check	OK	10:42
Air Blank	0.000	10:43
Control Test	0.080	10:43
Air Blank	0.000	10:43
Subject Sample # 1	0.000	10:44
Breath Volume	1.968	10:44
Air Blank	0.000	10:44
Air Blank	0.000	10:46
Subject Sample # 2	0.000	10:47
Breath Volume	1.515	10:47
Air Blank	0.000	10:47
Control Test	0.080	10:48
Air Blank	0.000	10:48
Diagnostics Check	OK	10:48

Data Download Date/Time: 08/11/2023 17:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/09/2023 22:20:02

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.000	22:15
Breath Volume	2.167	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.000	22:18
Breath Volume	2.128	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:20

Data Download Date/Time: 08/11/2023 17:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: GILLESPIE

DATE/TIME OF TEST: 07/10/2023 11:13:25

Diagnostics Check	OK	11:07
Air Blank	0.000	11:07
Control Test	0.081	11:08
Air Blank	0.000	11:08
Subject Sample # 1	0.000	11:09
Breath Volume	2.421	11:09
Air Blank	0.000	11:09
Air Blank	0.000	11:11
Subject Sample # 2	0.000	11:12
Breath Volume	2.308	11:12
Air Blank	0.000	11:12
Control Test	0.080	11:12
Air Blank	0.000	11:13
Diagnostics Check	OK	11:13

Data Download Date/Time: 08/11/2023 17:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: SKRIVER

DATE/TIME OF TEST: 07/13/2023 10:56:58

Diagnostics Check	OK	10:50
Air Blank	0.000	10:51
Control Test	0.079	10:51
Air Blank	0.000	10:52
Subject Sample # 1	0.000	10:52
Breath Volume	4.531	10:52
Air Blank	0.000	10:53
Air Blank	0.000	10:55
Subject Sample # 2	0.000	10:55
Breath Volume	4.890	10:55
Air Blank	0.000	10:56
Control Test	0.079	10:56
Air Blank	0.000	10:56
Diagnostics Check	OK	10:56

Data Download Date/Time: 08/11/2023 17:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: SKYBERG

DATE/TIME OF TEST: 07/13/2023 11:31:07

Diagnostics Check	OK	11:25
Air Blank	0.000	11:25
Control Test	0.079	11:26
Air Blank	0.000	11:26
Subject Sample # 1	0.000	11:26
Breath Volume	1.511	11:26
Air Blank	0.000	11:27
Air Blank	0.000	11:29
Subject Sample # 2	0.000	11:29
Breath Volume	1.781	11:29
Air Blank	0.000	11:30
Control Test	0.078	11:30
Air Blank	0.000	11:31
Diagnostics Check	OK	11:31

Data Download Date/Time: 08/11/2023 17:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: GALINDO-RIVERA

DATE/TIME OF TEST: 07/13/2023 12:43:27

Diagnostics Check	OK	12:37
Air Blank	0.000	12:37
Control Test	0.080	12:38
Air Blank	0.000	12:38
Subject Sample # 1	0.000	12:39
Breath Volume	2.390	12:39
Air Blank	0.000	12:39
Air Blank	0.000	12:41
Subject Sample # 2	0.000	12:42
Breath Volume	2.574	12:42
Air Blank	0.000	12:42
Control Test	0.080	12:42
Air Blank	0.000	12:43
Diagnostics Check	OK	12:43

Data Download Date/Time: 08/11/2023 17:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 07/13/2023 16:40:18

Diagnostics Check	OK	16:34
Air Blank	0.000	16:34
Control Test	0.080	16:34
Air Blank	0.000	16:35
Subject Sample # 1	0.000	16:36
Breath Volume	1.851	16:36
Air Blank	0.000	16:36
Air Blank	0.000	16:38
Subject Sample # 2	0.000	16:38
Breath Volume	2.875	16:38
Air Blank	0.000	16:39
Control Test	0.079	16:39
Air Blank	0.000	16:40
Diagnostics Check	OK	16:40

Data Download Date/Time: 08/11/2023 17:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: HERNANDEZ-VILLALON

DATE/TIME OF TEST: 07/14/2023 09:29:53

Diagnostics Check	OK	09:23
Air Blank	0.000	09:24
Control Test	0.080	09:24
Air Blank	0.000	09:25
Subject Sample # 1	0.000	09:25
Breath Volume	1.796	09:25
Air Blank	0.000	09:26
Air Blank	0.000	09:28
Subject Sample # 2	0.000	09:28
Breath Volume	1.414	09:28
Air Blank	0.000	09:28
Control Test	0.080	09:29
Air Blank	0.000	09:29
Diagnostics Check	OK	09:29

Data Download Date/Time: 08/11/2023 17:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: JABLONSKI

DATE/TIME OF TEST: 07/14/2023 10:07:21

Diagnostics Check	OK	10:01
Air Blank	0.000	10:01
Control Test	0.081	10:01
Air Blank	0.000	10:02
Subject Sample # 1	0.000	10:03
Breath Volume	4.179	10:03
Air Blank	0.000	10:03
Air Blank	0.000	10:05
Subject Sample # 2	0.000	10:05
Breath Volume	4.089	10:05
Air Blank	0.000	10:06
Control Test	0.080	10:06
Air Blank	0.000	10:07
Diagnostics Check	OK	10:07

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/14/2023 10:33:32

Diagnostics Check	OK	10:27
Air Blank	0.000	10:28
Control Test	0.080	10:28
Air Blank	0.000	10:28
Subject Sample # 1	0.000	10:29
Breath Volume	3.312	10:29
Air Blank	0.000	10:29
Air Blank	0.000	10:31
Subject Sample # 2	0.000	10:32
Breath Volume	2.593	10:32
Air Blank	0.000	10:32
Control Test	0.080	10:32
Air Blank	0.000	10:33
Diagnostics Check	OK	10:33

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 07/14/2023 12:20:19

Diagnostics Check	OK	12:14
Air Blank	0.000	12:14
Control Test	0.081	12:15
Air Blank	0.000	12:15
Subject Sample # 1	0.000	12:16
Breath Volume	2.589	12:16
Air Blank	0.000	12:16
Air Blank	0.000	12:18
Subject Sample # 2	0.000	12:18
Breath Volume	2.445	12:18
Air Blank	0.000	12:19
Control Test	0.080	12:19
Air Blank	0.000	12:20
Diagnostics Check	OK	12:20

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 07/20/2023 08:45:24

Diagnostics Check	OK	08:39
Air Blank	0.000	08:39
Control Test	0.080	08:40
Air Blank	0.000	08:40
Subject Sample # 1	0.000	08:41
Breath Volume	2.644	08:41
Air Blank	0.000	08:41
Air Blank	0.000	08:43
Subject Sample # 2	0.000	08:44
Breath Volume	2.531	08:44
Air Blank	0.000	08:44
Control Test	0.080	08:44
Air Blank	0.000	08:45
Diagnostics Check	OK	08:45

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: TOOMEY

DATE/TIME OF TEST: 07/20/2023 10:18:47

Diagnostics Check	OK	10:12
Air Blank	0.000	10:12
Control Test	0.080	10:13
Air Blank	0.000	10:13
Subject Sample # 1	0.000	10:14
Breath Volume	2.953	10:14
Air Blank	0.000	10:14
Air Blank	0.000	10:16
Subject Sample # 2	0.000	10:17
Breath Volume	2.828	10:17
Air Blank	0.000	10:17
Control Test	0.080	10:18
Air Blank	0.000	10:18
Diagnostics Check	OK	10:18

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 07/20/2023 15:18:12

Diagnostics Check	OK	15:11
Air Blank	0.000	15:12
Control Test	0.079	15:12
Air Blank	0.000	15:13
Subject Sample # 1	0.000	15:13
Breath Volume	3.890	15:13
Air Blank	0.000	15:14
Air Blank	0.000	15:16
Subject Sample # 2	0.000	15:16
Breath Volume	3.398	15:16
Air Blank	0.000	15:17
Control Test	0.079	15:17
Air Blank	0.000	15:18
Diagnostics Check	OK	15:18

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 07/21/2023 08:30:46

Diagnostics Check	OK	08:24
Air Blank	0.000	08:25
Control Test	0.080	08:25
Air Blank	0.000	08:26
Subject Sample # 1	0.000	08:26
Breath Volume	3.031	08:26
Air Blank	0.000	08:26
Air Blank	0.000	08:28
Subject Sample # 2	0.000	08:29
Breath Volume	3.074	08:29
Air Blank	0.000	08:29
Control Test	0.080	08:30
Air Blank	0.000	08:30
Diagnostics Check	OK	08:30

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/21/2023 09:11:55

Diagnostics Check	OK	09:05
Air Blank	0.000	09:06
Control Test	0.080	09:06
Air Blank	0.000	09:07
Subject Sample # 1	0.000	09:07
Breath Volume	2.960	09:07
Air Blank	0.000	09:08
Air Blank	0.000	09:10
Subject Sample # 2	0.000	09:10
Breath Volume	2.968	09:10
Air Blank	0.000	09:11
Control Test	0.079	09:11
Air Blank	0.000	09:11
Diagnostics Check	OK	09:11

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: ANGLIN

DATE/TIME OF TEST: 07/21/2023 09:28:44

Diagnostics Check	OK	09:22
Air Blank	0.000	09:23
Control Test	0.080	09:23
Air Blank	0.000	09:24
Subject Sample # 1	0.000	09:24
Breath Volume	1.828	09:24
Air Blank	0.000	09:25
Air Blank	0.000	09:27
Subject Sample # 2	0.000	09:27
Breath Volume	1.328	09:27
Air Blank	0.000	09:27
Control Test	0.079	09:28
Air Blank	0.000	09:28
Diagnostics Check	OK	09:28

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: ASSINI

DATE/TIME OF TEST: 07/21/2023 12:05:43

Diagnostics Check	OK	11:59
Air Blank	0.000	12:00
Control Test	0.079	12:00
Air Blank	0.000	12:01
Subject Sample # 1	0.000	12:01
Breath Volume	4.285	12:01
Air Blank	0.000	12:01
Air Blank	0.000	12:03
Subject Sample # 2	0.000	12:04
Breath Volume	3.890	12:04
Air Blank	0.000	12:04
Control Test	0.079	12:05
Air Blank	0.000	12:05
Diagnostics Check	OK	12:05

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 06/01/2014
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: GRANT-PERKINS

DATE/TIME OF TEST: 07/21/2023 12:19:51

Diagnostics Check	OK	12:13
Air Blank	0.000	12:14
Control Test	0.080	12:14
Air Blank	0.000	12:14
Subject Sample # 1	0.000	12:15
Breath Volume	1.980	12:15
Air Blank	0.000	12:16
Air Blank	0.000	12:18
Subject Sample # 2	0.000	12:18
Breath Volume	2.656	12:18
Air Blank	0.000	12:18
Control Test	0.080	12:19
Air Blank	0.000	12:19
Diagnostics Check	OK	12:19

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MCINTYRE

DATE/TIME OF TEST: 07/23/2023 02:57:23

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.081	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.081	02:52
Breath Volume	1.601	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.074	02:55
Breath Volume	1.562	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MARRERO MALDONADO

DATE/TIME OF TEST: 07/23/2023 04:17:49

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.080	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.280	04:12
Breath Volume	1.250	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample # 2	0.290	04:16
Breath Volume	1.437	04:16
Air Blank	0.000	04:16
Control Test	0.078	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: HERNANDEZ SANLUIS

DATE/TIME OF TEST: 07/25/2023 18:48:42

Diagnostics Check	OK	18:42
Air Blank	0.000	18:42
Control Test	0.080	18:43
Air Blank	0.000	18:43
Subject Sample # 1	0.194	18:44
Breath Volume	1.414	18:44
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 2	0.190	18:47
Breath Volume	1.824	18:47
Air Blank	0.000	18:47
Control Test	0.080	18:48
Air Blank	0.000	18:48
Diagnostics Check	OK	18:48

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10912080A1
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MERISTIN

DATE/TIME OF TEST: 07/30/2023 05:37:15

Diagnostics Check	OK	05:34	
Air Blank	0.000	05:34	
Control Test	0.081	05:35	
Air Blank	0.000	05:35	
Subject Sample	# 1	05:35	Subject Test Refused
Breath Volume	0.000	05:35	
Air Blank	0.000	05:36	
Control Test	0.080	05:36	
Air Blank	0.000	05:37	
Diagnostics Check	OK	05:37	

Data Download Date/Time: 08/11/2023 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERKLIN, LEYON, G
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: KERKLIN, LEYON
ARRESTING OFFICER AGENCY: Hamilton County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/17/2023 00:17:18

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.082	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.142	00:11
Breath Volume	1.746	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:13
Subject Sample # 2	0.153	00:15
Breath Volume	2.144	00:15
Air Blank	0.000	00:16
Control Test	0.084	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 07/30/2023 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: TORRES RENDON

DATE/TIME OF TEST: 07/31/2023 00:51:27

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:48	
Control Test	0.077	00:48	
Air Blank	0.000	00:49	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:50	
Control Test	0.077	00:50	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 08/08/2023 11:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: REID,HUNTER
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 07/01/2023 22:42:30

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.083	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.000	22:38
Breath Volume	2.882	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.000	22:41
Breath Volume	2.675	22:41
Air Blank	0.000	22:41
Control Test	0.082	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROZZELL,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: MCCANDLESS

DATE/TIME OF TEST: 07/03/2023 20:50:11

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.081	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.129	20:45
Breath Volume	1.316	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.134	20:48
Breath Volume	1.671	20:48
Air Blank	0.000	20:49
Control Test	0.081	20:49
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 07/04/2023 19:11:58

Diagnostics Check	OK	19:08	
Air Blank	0.000	19:09	
Control Test	0.082	19:09	
Air Blank	0.000	19:10	
Subject Sample	# 1	19:10	Subject Test Refused
Breath Volume	0.000	19:10	
Air Blank	0.000	19:11	
Control Test	0.081	19:11	
Air Blank	0.000	19:11	
Diagnostics Check	OK	19:11	

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREZKOP,ZACH,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 07/04/2023 23:03:36

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.083	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.075	22:58
Breath Volume	1.214	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.078	23:01
Breath Volume	1.312	23:01
Air Blank	0.000	23:02
Control Test	0.083	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/05/2023 00:47:34

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.082	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.182	00:42
Breath Volume	2.738	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.173	00:45
Breath Volume	2.191	00:45
Air Blank	0.000	00:46
Control Test	0.083	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREZKOP,ZACHARY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: TINLEY

DATE/TIME OF TEST: 07/05/2023 01:53:51

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.082	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.083	01:53	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PETERS,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/06/2023 02:08:17

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.083	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.132	02:03
Breath Volume	3.347	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.131	02:06
Breath Volume	2.671	02:06
Air Blank	0.000	02:07
Control Test	0.082	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 07/07/2023 16:42:38

Diagnostics Check	OK	16:32	
Air Blank	0.000	16:33	
Control Test	0.082	16:33	
Air Blank	0.000	16:34	
Subject Sample	# 1 0.274	16:35	
Breath Volume	1.292	16:35	
Air Blank	0.000	16:36	
Air Blank	0.000	16:38	
Subject Sample	# 2	16:40	Slope Not Met
Breath Volume	0.515	16:40	
Air Blank	0.000	16:41	
Control Test	0.081	16:41	
Air Blank	0.000	16:42	
Diagnostics Check	OK	16:42	

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: LENTINI

DATE/TIME OF TEST: 07/08/2023 23:08:50

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.083	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.095	23:04
Breath Volume	1.152	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:06
Subject Sample # 2	0.097	23:07
Breath Volume	1.632	23:07
Air Blank	0.000	23:07
Control Test	0.081	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:24

SUBJECT'S LAST NAME: CREECH

DATE/TIME OF TEST: 07/11/2023 13:50:26

Diagnostics Check	OK	13:44
Air Blank	0.000	13:44
Control Test	0.083	13:44
Air Blank	0.000	13:45
Subject Sample # 1	0.000	13:46
Breath Volume	2.218	13:46
Air Blank	0.000	13:46
Air Blank	0.000	13:48
Subject Sample # 2	0.000	13:49
Breath Volume	1.890	13:49
Air Blank	0.000	13:49
Control Test	0.082	13:49
Air Blank	0.000	13:50
Diagnostics Check	OK	13:50

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BERLYNE,ROGER,D

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: BERLYNE,ROGER

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30822080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ROMERO CASTILLO

DATE/TIME OF TEST: 07/15/2023 01:41:56

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:40	
Control Test	0.083	01:41	
Air Blank		01:41	RFI Detect
Air Blank	0.000	01:41	

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,ROGER,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ROMERO CASTILLO

DATE/TIME OF TEST: 07/15/2023 01:51:09

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.082	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.063	01:46
Breath Volume	1.785	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.064	01:49
Breath Volume	2.269	01:49
Air Blank	0.000	01:50
Control Test	0.082	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: PRATZ

DATE/TIME OF TEST: 07/15/2023 22:06:39

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.083	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.000	22:02
Breath Volume	1.125	22:02
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.000	22:05
Breath Volume	1.996	22:05
Air Blank	0.000	22:05
Control Test	0.082	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: CASTONGUAY

DATE/TIME OF TEST: 07/15/2023 23:42:59

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:30	
Control Test	0.083	23:31	
Air Blank	0.000	23:31	
Subject Sample	# 1 0.161	23:34	Volume Not Met
Breath Volume	0.121	23:34	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample	# 2 0.184	23:37	
Breath Volume	3.386	23:37	
Air Blank	0.000	23:38	
Air Blank	0.000	23:40	
Subject Sample	# 3 0.174	23:41	
Breath Volume	2.343	23:41	
Air Blank	0.000	23:42	
Control Test	0.082	23:42	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, DOMINIQUE, L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HOWARD, DOMINIQUE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: VARGAS PACHECO

DATE/TIME OF TEST: 07/16/2023 02:39:46

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.082	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.168	02:34
Breath Volume	1.796	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.166	02:38
Breath Volume	1.707	02:38
Air Blank	0.000	02:38
Control Test	0.081	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,ROGER,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: PRIDGEN

DATE/TIME OF TEST: 07/16/2023 02:55:12

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.082	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.110	02:50
Breath Volume	1.812	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample # 2	0.103	02:53
Breath Volume	1.183	02:53
Air Blank	0.000	02:54
Control Test	0.083	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 07/19/2023 19:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SIUZDAK,C
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 07/16/2023 23:20:33

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:18	
Control Test	0.083	23:18	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.082	23:19	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 07/19/2023 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POULIN,DANIELLE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: POULIN,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: FABER

DATE/TIME OF TEST: 07/17/2023 19:47:37

Diagnostics Check	OK	19:38
Air Blank	0.000	19:39
Control Test	0.083	19:39
Air Blank	0.000	19:40
Subject Sample # 1	0.227	19:41
Breath Volume	1.773	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.222	19:45
Breath Volume	1.941	19:45
Air Blank	0.000	19:46
Control Test	0.081	19:46
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 07/19/2023 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DILLON,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,DILLON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WERKMEISTER

DATE/TIME OF TEST: 07/23/2023 03:25:35

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.124	03:20
Breath Volume	1.339	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.119	03:24
Breath Volume	2.160	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 07/27/2023 02:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RIDDLE,BRIANNA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: BULLARD

DATE/TIME OF TEST: 07/23/2023 20:38:06

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:35	
Control Test	0.081	20:36	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:36	Subject Test Refused
Breath Volume	0.000	20:36	
Air Blank	0.000	20:37	
Control Test	0.080	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 07/27/2023 02:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,MALLORY,O
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 07/24/2023 20:18:52

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.080	20:12
Air Blank	0.000	20:13
Subject Sample # 1	0.000	20:14
Breath Volume	1.792	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.000	20:17
Breath Volume	1.429	20:17
Air Blank	0.000	20:17
Control Test	0.080	20:18
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 07/27/2023 02:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BADIO,GOLVENS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BADIO,GOLVENS
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: SABB

DATE/TIME OF TEST: 07/25/2023 19:26:32

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.080	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.229	19:22
Breath Volume	3.011	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.234	19:24
Breath Volume	2.898	19:24
Air Blank	0.000	19:25
Control Test	0.080	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 07/27/2023 02:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: URONU,PATRICK,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: URONU,PATRICK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CHI

DATE/TIME OF TEST: 07/27/2023 02:47:00

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.081	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1 0.125	02:36	Volume Not Met
Breath Volume	0.988	02:36	
Air Blank	0.000	02:37	
Air Blank	0.000	02:38	
Subject Sample	# 2 0.134	02:40	
Breath Volume	1.800	02:40	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 3 0.129	02:45	
Breath Volume	1.656	02:45	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 08/23/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JESSICA,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HAMMAD,JIHAD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CAVAGNARO

DATE/TIME OF TEST: 07/27/2023 19:44:59

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:43	
Control Test	0.081	19:44	
Air Blank		19:44	RFI Detect
Air Blank	0.000	19:44	

Data Download Date/Time: 08/23/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JESSICA,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HAMMAD,JIHAD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CAVAGNARO

DATE/TIME OF TEST: 07/27/2023 19:50:40

Diagnostics Check	OK	19:48	
Air Blank	0.000	19:49	
Control Test	0.081	19:49	
Air Blank	0.000	19:50	
		19:50	RFI Detect
Air Blank	0.000	19:50	

Data Download Date/Time: 08/23/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JESSICA,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HAMMAD,JIHAD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CARAGNARO

DATE/TIME OF TEST: 07/27/2023 19:56:12

Diagnostics Check	OK	19:52	
Air Blank	0.000	19:52	
Control Test	0.081	19:52	
Air Blank	0.000	19:53	
Subject Sample # 1	0.000	19:54	
Breath Volume	2.324	19:54	
Air Blank		19:54	RFI Detect
Air Blank	0.000	19:55	
Control Test	0.082	19:55	
Air Blank	0.000	19:56	
Diagnostics Check	OK	19:56	

Data Download Date/Time: 08/23/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JESSICA,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HAMMAD,JIHAD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CAVAGNARO

DATE/TIME OF TEST: 07/27/2023 20:16:27

Diagnostics Check	OK	20:05	
Air Blank	0.000	20:05	
Control Test	0.082	20:05	
Air Blank	0.000	20:06	
Subject Sample # 1	0.000	20:07	
Breath Volume	1.851	20:07	
Air Blank	0.000	20:07	
Air Blank	0.000	20:09	
Subject Sample # 2	0.000	20:12	Volume Not Met
Breath Volume	0.226	20:12	
Air Blank	0.000	20:13	
Air Blank	0.000	20:15	
Subject Sample # 3		20:15	Sequence Aborted
Breath Volume	0.000	20:15	
Air Blank	0.000	20:16	

Data Download Date/Time: 08/23/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 07/30/2023 03:42:27

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.080	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1 0.154	03:34	Volume Not Met
Breath Volume	0.273	03:34	
Air Blank	0.000	03:35	
Air Blank	0.000	03:37	
Subject Sample	# 2 0.174	03:37	
Breath Volume	1.453	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 3 0.168	03:40	
Breath Volume	1.347	03:40	
Air Blank	0.000	03:41	
Control Test	0.080	03:41	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 08/23/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLANSON,JACE,W
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: ALLANSON,JACE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: BRANDENBURG

DATE/TIME OF TEST: 07/30/2023 20:27:16

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:24	
Control Test	0.080	20:25	
Air Blank	0.000	20:25	
Subject Sample	# 1	20:25	Subject Test Refused
Breath Volume	0.000	20:25	
Air Blank	0.000	20:26	
Control Test	0.081	20:26	
Air Blank	0.000	20:27	
Diagnostics Check	OK	20:27	

Data Download Date/Time: 08/23/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: REID,H
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/19/2023 20:41:02

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.078	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.195	20:36
Breath Volume	3.191	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:38
Subject Sample # 2	0.191	20:39
Breath Volume	2.550	20:39
Air Blank	0.000	20:40
Control Test	0.078	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:41

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATCH,VIRGINIA,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 07/22/2023 18:09:33

Diagnostics Check	OK	18:01
Air Blank	0.000	18:01
Control Test	0.078	18:02
Air Blank	0.000	18:02
Subject Sample # 1	0.259	18:03
Breath Volume	1.339	18:03
Air Blank	0.000	18:04
Air Blank	0.000	18:06
Subject Sample # 2	0.248	18:07
Breath Volume	1.292	18:07
Air Blank	0.000	18:08
Control Test	0.077	18:08
Air Blank	0.000	18:09
Diagnostics Check	OK	18:09

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MARCUS,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 07/22/2023 23:25:08

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:13	
Control Test	0.079	23:14	
Air Blank	0.000	23:14	
Subject Sample	# 1 0.020	23:17	Volume Not Met
Breath Volume	0.015	23:17	
Air Blank	0.000	23:18	
Air Blank	0.000	23:20	
Subject Sample	# 2 0.022	23:23	Volume Not Met
Breath Volume	0.535	23:23	
Air Blank	0.000	23:24	
Control Test	0.080	23:24	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MARCUS,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 07/23/2023 01:58:06

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.156	01:53
Breath Volume	2.101	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:55
Subject Sample # 2	0.150	01:56
Breath Volume	2.046	01:56
Air Blank	0.000	01:57
Control Test	0.078	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VESPER,MITCHELL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 07/25/2023 00:31:41

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.079	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.152	00:26
Breath Volume	1.332	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.163	00:30
Breath Volume	1.406	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,JOHN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: GREER

DATE/TIME OF TEST: 07/25/2023 01:26:18

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.094	01:21
Breath Volume	1.617	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.088	01:24
Breath Volume	1.343	01:24
Air Blank	0.000	01:25
Control Test	0.078	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BERLYNE,ROGER,D

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: BERLYNE,R

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30822080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 20:53

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MURDZA

DATE/TIME OF TEST: 07/28/2023 21:48:16

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.079	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.233	21:43
Breath Volume	2.480	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.218	21:46
Breath Volume	1.980	21:46
Air Blank	0.000	21:47
Control Test	0.077	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VESPER,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/29/2023 01:28:05

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.202	01:23
Breath Volume	2.531	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.190	01:26
Breath Volume	2.593	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SIUZDAK,C
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 07/29/2023 18:48:52

Diagnostics Check	OK	18:45	
Air Blank	0.000	18:46	
Control Test	0.079	18:46	
Air Blank	0.000	18:47	
Subject Sample	# 1	18:47	Subject Test Refused
Breath Volume	0.000	18:47	
Air Blank	0.000	18:47	
Control Test	0.079	18:48	
Air Blank	0.000	18:48	
Diagnostics Check	OK	18:48	

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WARE,G
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:41

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 07/01/2023 13:05:17

Diagnostics Check	OK	13:02	
Air Blank	0.000	13:02	
Control Test	0.079	13:03	
Air Blank	0.000	13:03	
Subject Sample	# 1	13:03	Subject Test Refused
Breath Volume	0.000	13:03	
Air Blank	0.000	13:04	
Control Test	0.079	13:04	
Air Blank	0.000	13:05	
Diagnostics Check	OK	13:05	

Data Download Date/Time: 07/24/2023 19:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ADKINS,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:51

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 07/02/2023 08:18:28

Diagnostics Check	OK	08:11
Air Blank	0.000	08:12
Control Test	0.079	08:12
Air Blank	0.000	08:12
Subject Sample # 1	0.114	08:13
Breath Volume	2.054	08:13
Air Blank	0.000	08:14
Air Blank	0.000	08:16
Subject Sample # 2	0.111	08:16
Breath Volume	1.585	08:16
Air Blank	0.000	08:17
Control Test	0.079	08:17
Air Blank	0.000	08:18
Diagnostics Check	OK	08:18

Data Download Date/Time: 07/24/2023 19:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRINKLE,BRITTANY,N
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: KRUGER

DATE/TIME OF TEST: 07/03/2023 01:27:53

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.079	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:26	
Control Test	0.078	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 07/24/2023 19:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN, DEMERIS, P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RITTER, ALEC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/03/2023 21:59:40

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:56	
Control Test	0.079	21:56	
Air Blank	0.000	21:57	
Subject Sample	# 1	21:57	Subject Test Refused
Breath Volume	0.000	21:57	
Air Blank	0.000	21:58	
Control Test	0.078	21:59	
Air Blank	0.000	21:59	
Diagnostics Check	OK	21:59	

Data Download Date/Time: 07/24/2023 19:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HOYT,D
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 07/03/2023 23:34:57

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:32	
Control Test	0.080	23:32	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:34	
Control Test	0.079	23:34	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 07/24/2023 19:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN,DEMERIS,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SARGENT,BRENT
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: HUGHART

DATE/TIME OF TEST: 07/04/2023 05:05:41

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.078	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.176	05:00
Breath Volume	2.308	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.170	05:04
Breath Volume	2.570	05:04
Air Blank	0.000	05:04
Control Test	0.078	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 07/24/2023 19:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HATCH,V
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: DEON

DATE/TIME OF TEST: 07/04/2023 23:44:33

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.079	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.155	23:39
Breath Volume	2.531	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.153	23:42
Breath Volume	2.714	23:42
Air Blank	0.000	23:43
Control Test	0.078	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 07/24/2023 19:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: CRAMER,K
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: GARCIA MARTINEZ

DATE/TIME OF TEST: 07/05/2023 00:20:33

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.079	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.179	00:15
Breath Volume	1.828	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.180	00:18
Breath Volume	2.726	00:18
Air Blank	0.000	00:19
Control Test	0.078	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 07/24/2023 19:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: SNUGGS

DATE/TIME OF TEST: 07/06/2023 23:35:29

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:32	
Control Test	0.080	23:33	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:34	
Control Test	0.078	23:34	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 07/24/2023 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL, TYLER, C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING, J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: REICH

DATE/TIME OF TEST: 07/14/2023 23:52:14

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.079	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.175	23:47
Breath Volume	1.718	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample # 2	0.158	23:50
Breath Volume	1.421	23:50
Air Blank	0.000	23:51
Control Test	0.077	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 07/24/2023 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL, TYLER, C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING, J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: NOTARO

DATE/TIME OF TEST: 07/16/2023 02:02:08

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.079	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1 0.190	01:50	
Breath Volume	2.179	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 2 0.148	01:56	Volume Not Met
Breath Volume	0.375	01:56	
Air Blank	0.000	01:56	
Air Blank	0.000	01:58	
Subject Sample	# 3	02:00	Subject Test Refused
Breath Volume	0.406	02:00	
Air Blank	0.000	02:01	
Control Test	0.077	02:01	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 07/24/2023 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDZINSKI,BRANDON,E
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BRAVE,STEVIE
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LOPEZ FELIPE

DATE/TIME OF TEST: 07/17/2023 01:12:25

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.113	01:07
Breath Volume	2.917	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.112	01:10
Breath Volume	2.691	01:10
Air Blank	0.000	01:11
Control Test	0.078	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 07/24/2023 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,THOMAS,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MURCHINSON,RONNIE
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: MCCLOY

DATE/TIME OF TEST: 07/20/2023 20:17:04

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:12	
Control Test	0.078	20:12	
Air Blank	0.000	20:13	
Subject Sample	# 1	20:15	Subject Test Refused
Breath Volume	0.000	20:15	
Air Blank	0.000	20:16	
Control Test	0.077	20:16	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:17	

Data Download Date/Time: 07/24/2023 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TAPANES,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/21/2023 22:12:37

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.214	22:07
Breath Volume	2.441	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.212	22:10
Breath Volume	2.722	22:10
Air Blank	0.000	22:11
Control Test	0.079	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 07/24/2023 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TAPANES,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: ARZU

DATE/TIME OF TEST: 07/22/2023 22:44:21

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:41	
Control Test	0.079	22:42	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:43	
Control Test	0.078	22:43	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:44	

Data Download Date/Time: 07/24/2023 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 07/28/2023 23:58:06

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.000	23:53
Breath Volume	2.703	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.000	23:56
Breath Volume	2.136	23:56
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 08/25/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WILSON,D
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:43

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/30/2023 12:11:27

Diagnostics Check	OK	12:04
Air Blank	0.000	12:04
Control Test	0.079	12:04
Air Blank	0.000	12:05
Subject Sample # 1	0.294	12:06
Breath Volume	1.601	12:06
Air Blank	0.000	12:07
Air Blank	0.000	12:09
Subject Sample # 2	0.294	12:09
Breath Volume	1.308	12:09
Air Blank	0.000	12:10
Control Test	0.079	12:10
Air Blank	0.000	12:11
Diagnostics Check	OK	12:11

Data Download Date/Time: 08/25/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: PETERSON,W
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: SOONDAR

DATE/TIME OF TEST: 07/31/2023 19:36:33

Diagnostics Check	OK	19:30
Air Blank	0.000	19:30
Control Test	0.079	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.124	19:31
Breath Volume	2.882	19:31
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample # 2	0.124	19:35
Breath Volume	2.742	19:35
Air Blank	0.000	19:35
Control Test	0.079	19:35
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 08/25/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,THOMAS,B
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: REID,THOMAS
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 07/01/2023 21:30:12

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.079	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.145	21:25
Breath Volume	1.808	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.162	21:28
Breath Volume	2.027	21:28
Air Blank	0.000	21:29
Control Test	0.078	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ELIZABETH,A
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: REID,THOMAS
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HUGAR

DATE/TIME OF TEST: 07/03/2023 00:25:16

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.171	00:20
Breath Volume	1.492	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.182	00:23
Breath Volume	1.937	00:23
Air Blank	0.000	00:24
Control Test	0.078	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANDRE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HATCH,VIRGINIA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/04/2023 04:56:20

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.166	04:51
Breath Volume	1.902	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.163	04:54
Breath Volume	2.277	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ADOLF,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: CLIBURN

DATE/TIME OF TEST: 07/04/2023 19:14:30

Diagnostics Check	OK	19:02	
Air Blank	0.000	19:03	
Control Test	0.079	19:03	
Air Blank	0.000	19:04	
Subject Sample	# 1 0.182	19:07	Volume Not Met
Breath Volume	0.316	19:07	
Air Blank	0.000	19:08	
Air Blank	0.000	19:09	
Subject Sample	# 2 0.174	19:12	Volume Not Met
Breath Volume	0.949	19:12	
Air Blank	0.000	19:13	
Control Test	0.076	19:13	
Air Blank	0.000	19:14	
Diagnostics Check	OK	19:14	

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,RON
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: DIMICHELE

DATE/TIME OF TEST: 07/09/2023 12:28:47

Diagnostics Check	OK	12:25	
Air Blank	0.000	12:26	
Control Test	0.080	12:26	
Air Blank	0.000	12:27	
Subject Sample	# 1	12:27	Subject Test Refused
Breath Volume	0.000	12:27	
Air Blank	0.000	12:27	
Control Test	0.080	12:28	
Air Blank	0.000	12:28	
Diagnostics Check	OK	12:28	

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TWITTY,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 07/14/2023 02:25:43

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.155	02:21
Breath Volume	1.695	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.144	02:24
Breath Volume	1.304	02:24
Air Blank	0.000	02:24
Control Test	0.079	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINSTRY,MICHAEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MCKINSTRY,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: WADSWORTH

DATE/TIME OF TEST: 07/15/2023 20:12:06

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:10	
Control Test	0.080	20:11	
Air Blank		20:11	RFI Detect
Air Blank	0.000	20:12	

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINSTRY,MICHAEL,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MCKINSTRY,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: W326432836310

DATE/TIME OF TEST: 07/15/2023 20:18:08

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:15	
Control Test	0.079	20:16	
Air Blank	0.000	20:16	
Subject Sample	# 1	20:16	Subject Test Refused
Breath Volume	0.000	20:16	
Air Blank	0.000	20:17	
Control Test	0.080	20:17	
Air Blank	0.000	20:18	
Diagnostics Check	OK	20:18	

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINSTRY, MICHAEL, B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BROWNING, J
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DIDION

DATE/TIME OF TEST: 07/15/2023 20:30:07

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:24	
Control Test	0.080	20:25	
Air Blank	0.000	20:25	
Subject Sample	# 1	20:28	Subject Test Refused
Breath Volume	0.000	20:28	
Air Blank	0.000	20:29	
Control Test	0.079	20:29	
Air Blank	0.000	20:30	
Diagnostics Check	OK	20:30	

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MALIK

DATE/TIME OF TEST: 07/20/2023 00:46:41

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.080	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:45	
Control Test	0.080	00:46	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GITTO,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: RICHIE

DATE/TIME OF TEST: 07/22/2023 00:05:05

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.073	23:59
Breath Volume	1.164	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.072	00:03
Breath Volume	1.664	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:05

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:32

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 07/22/2023 12:00:07

Diagnostics Check	OK	11:53
Air Blank	0.000	11:53
Control Test	0.079	11:53
Air Blank	0.000	11:54
Subject Sample # 1	0.000	11:55
Breath Volume	1.433	11:55
Air Blank	0.000	11:55
Air Blank	0.000	11:57
Subject Sample # 2	0.000	11:58
Breath Volume	1.582	11:58
Air Blank	0.000	11:59
Control Test	0.079	11:59
Air Blank	0.000	12:00
Diagnostics Check	OK	12:00

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: ROUSE

DATE/TIME OF TEST: 07/22/2023 17:37:40

Diagnostics Check	OK	17:31
Air Blank	0.000	17:31
Control Test	0.079	17:32
Air Blank	0.000	17:32
Subject Sample # 1	0.115	17:33
Breath Volume	1.390	17:33
Air Blank	0.000	17:33
Air Blank	0.000	17:35
Subject Sample # 2	0.115	17:36
Breath Volume	1.382	17:36
Air Blank	0.000	17:36
Control Test	0.079	17:37
Air Blank	0.000	17:37
Diagnostics Check	OK	17:37

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:02

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 07/23/2023 09:29:03

Diagnostics Check	OK	09:22
Air Blank	0.000	09:22
Control Test	0.079	09:23
Air Blank	0.000	09:23
Subject Sample # 1	0.000	09:24
Breath Volume	1.691	09:24
Air Blank	0.000	09:24
Air Blank	0.000	09:26
Subject Sample # 2	0.000	09:27
Breath Volume	2.027	09:27
Air Blank	0.000	09:28
Control Test	0.080	09:28
Air Blank	0.000	09:28
Diagnostics Check	OK	09:29

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, ANTHONY, T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CARTER, S
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: NERCESSIAN

DATE/TIME OF TEST: 07/23/2023 20:12:33

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:59	
Control Test	0.080	19:59	
Air Blank	0.000	20:00	
Subject Sample	# 1 0.143	20:03	Volume Not Met
Breath Volume	0.164	20:03	
Air Blank	0.000	20:04	
Air Blank	0.000	20:05	
Subject Sample	# 2 0.195	20:07	
Breath Volume	1.253	20:07	
Air Blank	0.000	20:08	
Air Blank	0.000	20:10	
Subject Sample	# 3 0.192	20:11	
Breath Volume	1.218	20:11	
Air Blank	0.000	20:11	
Control Test	0.077	20:11	
Air Blank	0.000	20:12	
Diagnostics Check	OK	20:12	

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,KYLE
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: GAUTHEIR

DATE/TIME OF TEST: 07/26/2023 22:00:15

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.079	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.009	21:55
Breath Volume	1.703	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.008	21:58
Breath Volume	2.046	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,KYLE
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/27/2023 02:30:29

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.270	02:22
Breath Volume	1.339	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 2	0.296	02:25
Breath Volume	1.265	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 3	0.291	02:28
Breath Volume	1.207	02:28
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 07/27/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: JACKSON,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: POSEY

DATE/TIME OF TEST: 07/29/2023 20:58:20

Diagnostics Check	OK	20:51
Air Blank	0.000	20:52
Control Test	0.080	20:52
Air Blank	0.000	20:53
Subject Sample # 1	0.000	20:53
Breath Volume	3.085	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.000	20:56
Breath Volume	2.671	20:56
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 09/05/2023 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BURWELL,M
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CLATER

DATE/TIME OF TEST: 07/29/2023 23:36:29

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.080	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:35	
Control Test	0.081	23:35	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 09/05/2023 09:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HEWITT,B
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 07/30/2023 04:37:35

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:35	
Control Test	0.080	04:35	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:36	Subject Test Refused
Breath Volume	0.000	04:36	
Air Blank	0.000	04:36	
Control Test	0.082	04:36	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 09/05/2023 09:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 07/31/2023 18:00:04

Diagnostics Check	OK	17:49	
Air Blank	0.000	17:49	
Control Test	0.080	17:49	
Air Blank	0.000	17:50	
Subject Sample	# 1 0.096	17:53	Volume Not Met
Breath Volume	0.437	17:53	
Air Blank	0.000	17:54	
Air Blank	0.000	17:56	
Subject Sample	# 2	17:58	Subject Test Refused
Breath Volume	0.171	17:58	
Air Blank	0.000	17:59	
Control Test	0.082	17:59	
Air Blank	0.000	17:59	
Diagnostics Check	OK	18:00	

Data Download Date/Time: 09/05/2023 09:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 07/31/2023 22:43:02

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.197	22:37
Breath Volume	1.589	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.208	22:41
Breath Volume	1.308	22:41
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:43

Data Download Date/Time: 09/05/2023 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,ROBERT,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MONTALVO,LUIS
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BUESO CALDERON

DATE/TIME OF TEST: 07/01/2023 02:12:37

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.078	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:05	No Sample Provided
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 2	02:11	No Sample Provided
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Control Test	0.079	02:12	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,ROBERT,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MONTALVO,LUIS
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BUESO CALDERON

DATE/TIME OF TEST: 07/01/2023 02:23:03

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:19	
Control Test	0.080	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:22	
Control Test	0.080	02:22	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,JOSEPH,
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MORGAN,JOSEPH
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: VUKEL

DATE/TIME OF TEST: 07/01/2023 13:38:06

Diagnostics Check	OK	13:27	
Air Blank	0.000	13:28	
Control Test	0.080	13:28	
Air Blank	0.000	13:29	
Subject Sample # 1	0.301	13:30	
Breath Volume	2.062	13:30	
Air Blank	0.000	13:31	
Air Blank	0.000	13:32	
Subject Sample # 2	0.278	13:33	
Breath Volume	2.363	13:33	
Air Blank	0.000	13:34	
Air Blank	0.000	13:36	
Subject Sample # 3	0.295	13:37	
Breath Volume	1.640	13:37	
Air Blank		13:37	RFI Detect
Air Blank	0.000	13:38	

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRONK,A,
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: CRONK,A
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: TERRELL

DATE/TIME OF TEST: 07/04/2023 20:57:20

Diagnostics Check	OK	20:54	
Air Blank	0.000	20:54	
Control Test	0.079	20:55	
Air Blank	0.000	20:55	
Subject Sample	# 1	20:55	Subject Test Refused
Breath Volume	0.000	20:55	
Air Blank	0.000	20:56	
Control Test	0.080	20:56	
Air Blank	0.000	20:57	
Diagnostics Check	OK	20:57	

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,STEPHEN,V
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: MARINACCI,JOSEPH
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: TOWMSEND

DATE/TIME OF TEST: 07/05/2023 00:36:16

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.080	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:29	No Sample Provided
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 2	00:34	No Sample Provided
Breath Volume	0.000	00:34	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASIMIRO,B,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: NIEVES,B
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HOWE

DATE/TIME OF TEST: 07/05/2023 02:17:37

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.103	02:13
Breath Volume	1.687	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.100	02:16
Breath Volume	1.828	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWNSBERGER, BLAINE, T
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: ORINGDERFF, DARLA
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 07/06/2023 17:18:48

Diagnostics Check	OK	17:11
Air Blank	0.000	17:11
Control Test	0.080	17:11
Air Blank	0.000	17:12
Subject Sample # 1	0.187	17:14
Breath Volume	1.234	17:14
Air Blank	0.000	17:15
Air Blank	0.000	17:16
Subject Sample # 2	0.182	17:17
Breath Volume	1.578	17:17
Air Blank	0.000	17:17
Control Test	0.078	17:18
Air Blank	0.000	17:18
Diagnostics Check	OK	17:18

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS, DOUG, J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BARKER, D
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: VILA

DATE/TIME OF TEST: 07/08/2023 02:24:51

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.157	02:18
Breath Volume	1.292	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.152	02:23
Breath Volume	1.296	02:23
Air Blank	0.000	02:23
Control Test	0.079	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERDT, ANDRU, J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: RICHETTI, KALEIGH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: THIGPEN

DATE/TIME OF TEST: 07/11/2023 02:08:15

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.160	02:03
Breath Volume	1.519	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.161	02:06
Breath Volume	2.527	02:06
Air Blank	0.000	02:07
Control Test	0.078	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BANNICK,COURTNEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: CODDING

DATE/TIME OF TEST: 07/11/2023 16:30:39

Diagnostics Check	OK	16:23
Air Blank	0.000	16:24
Control Test	0.079	16:24
Air Blank	0.000	16:24
Subject Sample # 1	0.000	16:25
Breath Volume	1.703	16:25
Air Blank	0.000	16:25
Air Blank	0.000	16:27
Subject Sample # 2	0.000	16:29
Breath Volume	1.320	16:29
Air Blank	0.000	16:29
Control Test	0.080	16:30
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: PARISH,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: FISH

DATE/TIME OF TEST: 07/13/2023 01:14:08

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank		01:13	RFI Detect
Air Blank		01:14	RFI Detect

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,BROOKE,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: GOTT,CHRISTOPHER
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: MEDLEY

DATE/TIME OF TEST: 07/15/2023 01:10:01

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.185	01:05
Breath Volume	2.867	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:07
Subject Sample # 2	0.176	01:08
Breath Volume	1.851	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:10

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BANNICK,COURTNEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: BEIGHLE HINE

DATE/TIME OF TEST: 07/15/2023 18:38:01

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.080	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.139	18:33
Breath Volume	1.105	18:33
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.135	18:36
Breath Volume	1.152	18:36
Air Blank	0.000	18:37
Control Test	0.080	18:37
Air Blank	0.000	18:37
Diagnostics Check	OK	18:38

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILDEMAN-STUART,NATASH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 07/15/2023 18:46:06

Diagnostics Check	OK	18:43	
Air Blank	0.000	18:43	
Control Test	0.079	18:44	
Air Blank	0.000	18:44	
Subject Sample	# 1	18:44	Subject Test Refused
Breath Volume	0.000	18:44	
Air Blank	0.000	18:45	
Control Test	0.079	18:45	
Air Blank	0.000	18:46	
Diagnostics Check	OK	18:46	

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LUTFLF
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: BRITT

DATE/TIME OF TEST: 07/16/2023 05:03:42

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.192	04:59
Breath Volume	3.746	04:59
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.187	05:02
Breath Volume	3.332	05:02
Air Blank	0.000	05:02
Control Test	0.079	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIMER,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LATIMER,JUSTIN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/17/2023 00:05:08

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.080	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.222	00:00
Breath Volume	3.250	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:02
Subject Sample # 2	0.216	00:03
Breath Volume	3.394	00:03
Air Blank	0.000	00:04
Control Test	0.076	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH, LAUREN, A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: MORALES, K
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: WOLFORD

DATE/TIME OF TEST: 07/18/2023 04:13:33

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.079	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1 0.126	04:05	Volume Not Met
Breath Volume	0.570	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 2 0.161	04:08	
Breath Volume	1.343	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 3 0.156	04:11	
Breath Volume	1.886	04:11	
Air Blank	0.000	04:12	
Control Test	0.077	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: JONES,D
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: BOUCHARD

DATE/TIME OF TEST: 07/19/2023 18:39:05

Diagnostics Check	OK	18:36	
Air Blank	0.000	18:36	
Control Test	0.080	18:37	
Air Blank	0.000	18:37	
Subject Sample	# 1	18:37	Subject Test Refused
Breath Volume	0.000	18:37	
Air Blank	0.000	18:38	
Control Test	0.080	18:38	
Air Blank	0.000	18:38	
Diagnostics Check	OK	18:39	

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,BROOKE,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: MASON,BROOKE
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MARQUEZ SANCHEZ

DATE/TIME OF TEST: 07/20/2023 20:22:36

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.079	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.233	20:18
Breath Volume	2.140	20:18
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.235	20:21
Breath Volume	2.140	20:21
Air Blank	0.000	20:21
Control Test	0.078	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,BROOKE,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: CHANDLER,HASON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: PORTILLO

DATE/TIME OF TEST: 07/21/2023 04:21:40

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.118	04:17
Breath Volume	2.820	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.119	04:20
Breath Volume	3.890	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH, LAUREN, A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: NIEVES, B
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: CHECHO

DATE/TIME OF TEST: 07/22/2023 03:31:02

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	
Control Test	0.080	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 07/25/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSENKARSKI,MAX,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: OSENKARSKI,MAX
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: STEWART MOORE

DATE/TIME OF TEST: 07/26/2023 21:32:43

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.268	21:28
Breath Volume	3.527	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.283	21:31
Breath Volume	2.843	21:31
Air Blank	0.000	21:31
Control Test	0.079	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIMER,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LATIMER,JUSTIN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DE LA ROSA

DATE/TIME OF TEST: 07/29/2023 02:43:38

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.079	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONDAHL,CARL,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CUTLIFF

DATE/TIME OF TEST: 07/29/2023 03:35:14

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.200	03:28
Breath Volume	1.324	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.201	03:33
Breath Volume	2.144	03:33
Air Blank	0.000	03:34
Control Test	0.077	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,L,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: FARNELL

DATE/TIME OF TEST: 07/31/2023 02:32:23

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.079	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 07/02/2023 21:16:51

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.077	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.240	21:12
Breath Volume	2.753	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:14
Subject Sample # 2	0.233	21:15
Breath Volume	1.523	21:15
Air Blank	0.000	21:15
Control Test	0.077	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 07/27/2023 14:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BLUM,S,K

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: COFFEY,W

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: LUDOLPH

DATE/TIME OF TEST: 07/03/2023 23:00:58

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:58	
Control Test	0.077	22:58	
Air Blank	0.000	22:59	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.077	23:00	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUM,S,K
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: COFFEY,W
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: LUDOLPH

DATE/TIME OF TEST: 07/03/2023 23:17:47

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.078	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.139	23:13
Breath Volume	1.253	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:15
Subject Sample # 2	0.147	23:16
Breath Volume	1.542	23:16
Air Blank	0.000	23:16
Control Test	0.077	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: DUBOSE

DATE/TIME OF TEST: 07/05/2023 22:21:32

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.077	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.288	22:17
Breath Volume	1.937	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.287	22:20
Breath Volume	1.726	22:20
Air Blank	0.000	22:20
Control Test	0.075	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOLCHARRAN,DILON,C

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: DOLCHARRAN,DILON

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 07/09/2023 04:02:00

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.079	03:59	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Control Test	0.078	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: HELLER

DATE/TIME OF TEST: 07/11/2023 09:06:59

Diagnostics Check	OK	08:55	
Air Blank	0.000	08:55	
Control Test	0.079	08:56	
Air Blank	0.000	08:56	
Subject Sample	# 1 0.000	08:59	Volume Not Met
Breath Volume	0.550	08:59	
Air Blank	0.000	09:00	
Air Blank	0.000	09:02	
Subject Sample	# 2 0.000	09:02	
Breath Volume	1.968	09:02	
Air Blank	0.000	09:03	
Air Blank	0.000	09:05	
Subject Sample	# 3 0.000	09:05	
Breath Volume	1.390	09:05	
Air Blank	0.000	09:06	
Control Test	0.078	09:06	
Air Blank	0.000	09:06	
Diagnostics Check	OK	09:06	

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PRESOIR

DATE/TIME OF TEST: 07/15/2023 03:12:25

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.105	03:07
Breath Volume	1.152	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.104	03:10
Breath Volume	1.179	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: NOSTRAMO

DATE/TIME OF TEST: 07/16/2023 01:30:31

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.078	01:19	
Air Blank	0.000	01:20	
Subject Sample	# 1 0.058	01:23	Volume Not Met
Breath Volume	0.000	01:23	
Air Blank	0.000	01:24	
Air Blank	0.000	01:26	
Subject Sample	# 2	01:29	No Sample Provided
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Control Test	0.079	01:29	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCLAIN, DAVID, R

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: MCCLAIN, DAVID

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: NOSTRAMO

DATE/TIME OF TEST: 07/16/2023 01:42:51

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:33	
Control Test	0.079	01:34	
Air Blank	0.000	01:34	
Subject Sample	# 1 0.000	01:37	Volume Not Met
Breath Volume	0.000	01:37	
Air Blank	0.000	01:38	
Air Blank	0.000	01:40	
Subject Sample	# 2	01:41	Subject Test Refused
Breath Volume	0.000	01:41	
Air Blank	0.000	01:41	
Control Test	0.079	01:42	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:11

SUBJECT'S LAST NAME: REGAN

DATE/TIME OF TEST: 07/16/2023 13:37:55

Diagnostics Check	OK	13:31
Air Blank	0.000	13:32
Control Test	0.078	13:32
Air Blank	0.000	13:33
Subject Sample # 1	0.309	13:33
Breath Volume	3.343	13:33
Air Blank	0.000	13:34
Air Blank	0.000	13:36
Subject Sample # 2	0.317	13:36
Breath Volume	2.023	13:36
Air Blank	0.000	13:37
Control Test	0.079	13:37
Air Blank	0.000	13:37
Diagnostics Check	OK	13:37

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCLAIN,DAVID,R

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: MCCLAIN,DAVID

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: WERTHMULLER

DATE/TIME OF TEST: 07/16/2023 21:42:46

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:39	
Control Test	0.080	21:39	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:41	Subject Test Refused
Breath Volume	0.000	21:41	
Air Blank	0.000	21:41	
Control Test	0.080	21:42	
Air Blank	0.000	21:42	
Diagnostics Check	OK	21:42	

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 07/17/2023 02:02:29

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.078	01:55	
Air Blank	0.000	01:55	
Subject Sample	# 1 0.203	01:58	Volume Not Met
Breath Volume	1.035	01:58	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2	02:01	Sequence Aborted
Breath Volume	0.015	02:01	
Air Blank	0.000	02:02	

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: D'AGOSTINO,STEFANO,V
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: D'AGOSTINO,STEFANO
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: ESCALONA MORALES

DATE/TIME OF TEST: 07/23/2023 00:19:55

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.080	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.189	00:15
Breath Volume	1.710	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:17
Subject Sample # 2	0.182	00:18
Breath Volume	2.113	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COFFEY-SHADE,W,T

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: COFFEY-SHADE,W

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MERCED

DATE/TIME OF TEST: 07/27/2023 01:35:16

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.080	01:33	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.079	01:34	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 07/27/2023 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,DAVID,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN,DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: ZUNIGA MORALES

DATE/TIME OF TEST: 07/29/2023 03:03:28

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.077	02:58
Breath Volume	1.988	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.077	03:02
Breath Volume	1.332	03:02
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 08/23/2023 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: ARRIAGA, IKE
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: UNION

DATE/TIME OF TEST: 07/29/2023 04:56:41

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.250	04:51
Breath Volume	1.214	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.242	04:55
Breath Volume	1.550	04:55
Air Blank	0.000	04:55
Control Test	0.077	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 08/23/2023 14:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: BLANKENBAKER

DATE/TIME OF TEST: 07/01/2023 05:49:44

Diagnostics Check	OK	05:48	
Air Blank	0.000	05:48	
Control Test	0.083	05:49	
Air Blank		05:49	RFI Detect
Air Blank	0.000	05:49	

Data Download Date/Time: 07/27/2023 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: BLANKENBAKER

DATE/TIME OF TEST: 07/01/2023 05:57:46

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.082	05:52
Air Blank	0.000	05:52
Subject Sample # 1	0.116	05:53
Breath Volume	1.363	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:55
Subject Sample # 2	0.122	05:56
Breath Volume	1.574	05:56
Air Blank	0.000	05:56
Control Test	0.081	05:57
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 07/27/2023 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SOTO CASTELLANO

DATE/TIME OF TEST: 07/03/2023 03:15:28

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.083	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.129	03:10
Breath Volume	2.203	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.126	03:13
Breath Volume	1.539	03:13
Air Blank	0.000	03:14
Control Test	0.082	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 07/27/2023 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/04/2023 02:56:03

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.082	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1 0.132	02:48	Volume Not Met
Breath Volume	0.953	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 2 0.139	02:51	
Breath Volume	1.359	02:51	
Air Blank	0.000	02:51	
Air Blank	0.000	02:53	
Subject Sample	# 3 0.145	02:54	
Breath Volume	1.281	02:54	
Air Blank	0.000	02:55	
Control Test	0.081	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 07/27/2023 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABBE,KYLER,S
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SHUMAN,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ONEILL

DATE/TIME OF TEST: 07/04/2023 03:48:56

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.215	03:44
Breath Volume	2.101	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.214	03:47
Breath Volume	2.019	03:47
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 07/27/2023 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAINO,J,V
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: PEREZ,A
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 07/06/2023 02:01:14

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.083	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.192	01:56
Breath Volume	1.488	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.196	01:59
Breath Volume	1.867	01:59
Air Blank	0.000	02:00
Control Test	0.081	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 07/27/2023 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,VICTORIA,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SIMONSEN,ERIK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: DANDREANO

DATE/TIME OF TEST: 07/08/2023 04:40:26

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.177	04:35
Breath Volume	1.988	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.172	04:38
Breath Volume	2.257	04:38
Air Blank	0.000	04:39
Control Test	0.082	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 07/27/2023 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: KOTEWA

DATE/TIME OF TEST: 07/09/2023 00:09:40

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.082	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.107	00:04
Breath Volume	2.179	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.113	00:08
Breath Volume	1.898	00:08
Air Blank	0.000	00:08
Control Test	0.082	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 07/27/2023 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABBE,KYLER,S
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LABBE,KYLER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/09/2023 02:27:33

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.151	02:23
Breath Volume	1.210	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.156	02:25
Breath Volume	1.812	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 07/27/2023 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEMLER,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SEMLER,JOSEPH
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 07/09/2023 04:53:18

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.163	04:48
Breath Volume	1.101	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:50
Subject Sample # 2	0.152	04:51
Breath Volume	1.324	04:51
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 07/27/2023 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIPP,JORDAN,E
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ECK,ENIAH
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MAYNARD

DATE/TIME OF TEST: 07/09/2023 05:16:53

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.238	05:12
Breath Volume	1.445	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:14
Subject Sample # 2	0.236	05:15
Breath Volume	1.746	05:15
Air Blank	0.000	05:15
Control Test	0.078	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 07/27/2023 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JESSICA,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SHUMAN,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 08:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:43

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 07/09/2023 09:35:56

Diagnostics Check	OK	09:30
Air Blank	0.000	09:30
Control Test	0.081	09:30
Air Blank	0.000	09:31
Subject Sample # 1	0.188	09:31
Breath Volume	1.609	09:31
Air Blank	0.000	09:32
Air Blank	0.000	09:34
Subject Sample # 2	0.173	09:34
Breath Volume	1.320	09:34
Air Blank	0.000	09:35
Control Test	0.081	09:35
Air Blank	0.000	09:35
Diagnostics Check	OK	09:35

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BRIGHTMON

DATE/TIME OF TEST: 07/10/2023 04:37:58

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.000	04:33
Breath Volume	2.144	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.000	04:36
Breath Volume	1.441	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HERNANDEZ,IVAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 07/12/2023 03:57:36

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.083	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.184	03:53
Breath Volume	3.484	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.179	03:56
Breath Volume	2.640	03:56
Air Blank	0.000	03:56
Control Test	0.082	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKINNER,KAILEIGH,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: SKINNER,KAILEIGH
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 07/13/2023 02:45:53

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.082	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.000	02:41
Breath Volume	1.332	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.000	02:44
Breath Volume	2.207	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:41

SUBJECT'S LAST NAME: MCCLANE

DATE/TIME OF TEST: 07/13/2023 10:14:21

Diagnostics Check	OK	10:02	
Air Blank	0.000	10:03	
Control Test	0.082	10:03	
Air Blank	0.000	10:04	
Subject Sample	# 1 0.000	10:07	Volume Not Met
Breath Volume	0.718	10:07	
Air Blank	0.000	10:07	
Air Blank	0.000	10:09	
Subject Sample	# 2 0.000	10:13	Volume Not Met
Breath Volume	0.359	10:13	
Air Blank	0.000	10:13	
Control Test	0.081	10:13	
Air Blank	0.000	10:14	
Diagnostics Check	OK	10:14	

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUTRAL,WILLIAM,Z
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FUTRAL,WILLIAM
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 07/15/2023 03:24:09

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.082	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Control Test	0.082	03:23	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,SPENSER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: EKELI,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: MALAGON

DATE/TIME OF TEST: 07/15/2023 20:10:25

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.082	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.184	20:05
Breath Volume	3.054	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.196	20:08
Breath Volume	3.238	20:08
Air Blank	0.000	20:09
Control Test	0.081	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: RICO HERNANDEZ

DATE/TIME OF TEST: 07/16/2023 04:04:28

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.082	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.236	03:59
Breath Volume	1.421	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.243	04:02
Breath Volume	1.402	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAGLAL,BRANDON,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: JAGLAL,BRANDON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: WEICHMAN

DATE/TIME OF TEST: 07/19/2023 23:38:36

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.082	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:31	No Sample Provided
Breath Volume	0.000	23:31	
Air Blank	0.000	23:32	
Air Blank	0.000	23:33	
Subject Sample	# 2	23:37	No Sample Provided
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Control Test	0.082	23:38	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAGLAL,BRANDON,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: JAGLAL,BRANDON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: WEICHMAN

DATE/TIME OF TEST: 07/19/2023 23:46:58

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.102	23:42
Breath Volume	3.222	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.103	23:45
Breath Volume	3.398	23:45
Air Blank	0.000	23:46
Control Test	0.081	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADDOCK,CHARLES,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PADDOCK,CHARLES
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: PALUSHI

DATE/TIME OF TEST: 07/20/2023 01:54:09

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.153	01:49
Breath Volume	1.859	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.164	01:52
Breath Volume	3.281	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: LOMBARDI

DATE/TIME OF TEST: 07/22/2023 01:25:08

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.084	01:19
Breath Volume	1.453	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:21
Subject Sample # 2	0.086	01:23
Breath Volume	1.378	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWER,DALTON,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: OWER,DALTON
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MCDANIELS

DATE/TIME OF TEST: 07/23/2023 00:45:20

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:33	
Control Test	0.081	00:33	
Air Blank	0.000	00:34	
Subject Sample	# 1	00:37	No Sample Provided
Breath Volume	0.000	00:37	
Air Blank	0.000	00:38	
Air Blank	0.000	00:39	
Subject Sample	# 2	0.191	00:40
Breath Volume	1.921	00:40	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 3	0.219	00:43
Breath Volume	1.851	00:43	
Air Blank		00:44	Purge Fail
Air Blank		00:45	Purge Fail

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWER,D,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: OWER,D
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: MCDANIELS

DATE/TIME OF TEST: 07/23/2023 00:48:51

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:48	
Control Test	0.000	00:48	Control Outside Tolerance
Air Blank		00:48	Ambient Fail

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWER,D,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: OWER,D
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: MCDANIELS

DATE/TIME OF TEST: 07/23/2023 00:51:47

Diagnostics Check	Fail	00:51	DSP Test
Air Blank		00:51	Ambient Fail

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTIAN,O
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GONZALEZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/23/2023 01:23:54

Diagnostics Check	Fail	01:23	DSP Test
Air Blank		01:23	Ambient Fail

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTIAN,O
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GONZALEZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/23/2023 01:28:07

Diagnostics Check	Fail	01:27	DSP Test
Air Blank		01:28	Ambient Fail

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTIAN,O
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GONZALEZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/23/2023 01:35:35

Diagnostics Check	Fail	01:35	DSP Test
Air Blank		01:35	Ambient Fail

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTIAN,O
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GONZALEZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/23/2023 01:56:14

Diagnostics Check	Fail	01:55	DSP Test
Air Blank		01:56	Ambient Fail

Data Download Date/Time: 07/27/2023 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: BRANCATO

DATE/TIME OF TEST: 07/01/2023 05:03:25

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.083	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.134	04:58
Breath Volume	2.808	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.119	05:01
Breath Volume	1.175	05:01
Air Blank	0.000	05:02
Control Test	0.082	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 07/05/2023 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/05/2023 01:05:08

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.082	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:59	No Sample Provided
Breath Volume	0.000	00:59	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 2	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:04	
Control Test	0.081	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 07/05/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: EUGENE

DATE/TIME OF TEST: 07/05/2023 04:20:14

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.082	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.017	04:15
Breath Volume	3.160	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.017	04:18
Breath Volume	2.523	04:18
Air Blank	0.000	04:19
Control Test	0.081	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 07/05/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MBAYE

DATE/TIME OF TEST: 07/06/2023 02:22:37

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.082	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.137	02:17
Breath Volume	2.144	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.132	02:20
Breath Volume	2.097	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/06/2023 04:55:18

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.081	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.115	04:50
Breath Volume	1.574	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.118	04:53
Breath Volume	1.640	04:53
Air Blank	0.000	04:54
Control Test	0.083	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: FARLIN

DATE/TIME OF TEST: 07/07/2023 23:33:50

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.084	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.148	23:29
Breath Volume	1.898	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.149	23:32
Breath Volume	1.953	23:32
Air Blank	0.000	23:32
Control Test	0.082	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MUNOZ CASTIBLANCO

DATE/TIME OF TEST: 07/08/2023 04:04:25

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.083	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.168	03:59
Breath Volume	2.613	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.157	04:02
Breath Volume	2.140	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: COMM VEHIC
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: HERNANDEZ JR

DATE/TIME OF TEST: 07/08/2023 05:01:09

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.082	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.153	04:56
Breath Volume	3.535	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 2	0.156	04:59
Breath Volume	3.574	04:59
Air Blank	0.000	05:00
Control Test	0.081	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIRALDO, J
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SHEPELEV

DATE/TIME OF TEST: 07/09/2023 03:10:41

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.083	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.000	03:06
Breath Volume	2.648	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.000	03:09
Breath Volume	1.378	03:09
Air Blank	0.000	03:09
Control Test	0.082	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: PRAJAPATI

DATE/TIME OF TEST: 07/09/2023 04:25:45

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.083	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.101	04:20
Breath Volume	2.011	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.095	04:24
Breath Volume	1.910	04:24
Air Blank	0.000	04:24
Control Test	0.081	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: FAKOLADE

DATE/TIME OF TEST: 07/09/2023 05:41:52

Diagnostics Check	OK	05:34
Air Blank	0.000	05:35
Control Test	0.082	05:35
Air Blank	0.000	05:35
Subject Sample # 1	0.162	05:36
Breath Volume	2.632	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.160	05:40
Breath Volume	1.707	05:40
Air Blank	0.000	05:40
Control Test	0.082	05:41
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/10/2023 01:29:30

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.083	01:27	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:28	Subject Test Refused
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Control Test	0.080	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEYERS,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 07/13/2023 03:41:00

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.083	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.000	03:36
Breath Volume	1.816	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.000	03:39
Breath Volume	3.058	03:39
Air Blank	0.000	03:40
Control Test	0.082	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:41

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 07/14/2023 02:23:07

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.082	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.146	02:18
Breath Volume	1.648	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.140	02:21
Breath Volume	1.394	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BARBIE

DATE/TIME OF TEST: 07/14/2023 04:28:31

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.082	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.054	04:23
Breath Volume	2.949	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.057	04:26
Breath Volume	3.625	04:26
Air Blank	0.000	04:27
Control Test	0.082	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 07/14/2023 06:38:31

Diagnostics Check	OK	06:30
Air Blank	0.000	06:30
Control Test	0.082	06:30
Air Blank	0.000	06:31
Subject Sample # 1	0.182	06:33
Breath Volume	1.519	06:33
Air Blank	0.000	06:34
Air Blank	0.000	06:35
Subject Sample # 2	0.183	06:36
Breath Volume	1.433	06:36
Air Blank	0.000	06:37
Control Test	0.081	06:37
Air Blank	0.000	06:38
Diagnostics Check	OK	06:38

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/15/2023 00:39:56

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.083	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.133	00:35
Breath Volume	1.722	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:37
Subject Sample # 2	0.142	00:38
Breath Volume	1.765	00:38
Air Blank	0.000	00:39
Control Test	0.081	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HOLCOMB

DATE/TIME OF TEST: 07/15/2023 00:57:38

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.082	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.112	00:52
Breath Volume	3.253	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.108	00:56
Breath Volume	2.812	00:56
Air Blank	0.000	00:56
Control Test	0.082	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/15/2023 03:22:30

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.083	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.175	03:17
Breath Volume	2.125	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.181	03:20
Breath Volume	2.238	03:20
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 07/15/2023 22:33:26

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.083	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.246	22:28
Breath Volume	1.500	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.250	22:31
Breath Volume	1.847	22:31
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: FLORES JIMENEZ

DATE/TIME OF TEST: 07/16/2023 00:32:17

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.082	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.078	00:27
Breath Volume	2.097	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.074	00:30
Breath Volume	1.875	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SAMPER GUITIERREZ

DATE/TIME OF TEST: 07/16/2023 02:44:35

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.265	02:39
Breath Volume	2.820	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.250	02:42
Breath Volume	2.261	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DOE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 07/16/2023 17:13:07

Diagnostics Check	OK	17:11	
Air Blank	0.000	17:12	
Control Test	0.087	17:12	Control Outside Tolerance
Air Blank	0.000	17:13	

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DOE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 07/16/2023 17:28:30

Diagnostics Check	OK	17:14	
Air Blank	0.000	17:14	
Control Test	0.082	17:14	
Air Blank	0.000	17:15	
Subject Sample	# 1 0.000	17:15	
Breath Volume	2.257	17:15	
Air Blank	0.000	17:16	
Air Blank	0.000	17:18	
Subject Sample	# 2	17:21	No Sample Provided
Breath Volume	0.000	17:21	
Air Blank	0.000	17:21	
Air Blank	0.000	17:23	
Subject Sample	# 3	17:27	No Sample Provided
Breath Volume	0.000	17:27	
Air Blank	0.000	17:27	
Control Test	0.082	17:27	
Air Blank	0.000	17:28	
Diagnostics Check	OK	17:28	

Data Download Date/Time: 08/01/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: ALICEA

DATE/TIME OF TEST: 07/16/2023 23:13:07

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.082	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.115	23:06
Breath Volume	1.246	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:08
Subject Sample # 2	0.112	23:11
Breath Volume	1.656	23:11
Air Blank	0.000	23:12
Control Test	0.081	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASI,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: AJISHEGIRI

DATE/TIME OF TEST: 07/17/2023 01:40:45

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.082	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.091	01:36
Breath Volume	2.140	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.090	01:39
Breath Volume	1.750	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: HENLEY

DATE/TIME OF TEST: 07/17/2023 03:14:48

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.083	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.201	03:10
Breath Volume	2.773	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.201	03:13
Breath Volume	2.453	03:13
Air Blank	0.000	03:13
Control Test	0.081	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PFEUFFER

DATE/TIME OF TEST: 07/18/2023 01:57:32

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.193	01:52
Breath Volume	1.882	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.197	01:55
Breath Volume	1.480	01:55
Air Blank	0.000	01:56
Control Test	0.082	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TANGEMAN,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: DERBY

DATE/TIME OF TEST: 07/18/2023 16:08:59

Diagnostics Check	OK	15:53	
Air Blank	0.000	15:53	
Control Test	0.083	15:53	
Air Blank	0.000	15:54	
Subject Sample	# 1 0.209	15:57	Slope Not Level
Breath Volume	2.414	15:57	
Air Blank	0.000	15:58	
Air Blank	0.000	16:00	
Subject Sample	# 2 0.236	16:01	
Breath Volume	1.265	16:01	
Air Blank	0.000	16:02	
Air Blank	0.000	16:04	
Subject Sample	# 3 0.229	16:07	Volume Not Met
Breath Volume	0.039	16:07	
Air Blank	0.000	16:08	
Control Test	0.081	16:08	
Air Blank	0.000	16:08	
Diagnostics Check	OK	16:08	

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TANGEMAN,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: DERBY

DATE/TIME OF TEST: 07/18/2023 16:28:20

Diagnostics Check	OK	16:13	
Air Blank	0.000	16:13	
Control Test	0.079	16:13	
Air Blank	0.000	16:14	
Subject Sample # 1	0.223	16:15	
Breath Volume	1.468	16:15	
Air Blank	0.000	16:16	
Air Blank	0.000	16:17	
Subject Sample # 2	0.196	16:21	Volume Not Met
Breath Volume	0.863	16:21	
Air Blank	0.000	16:22	
Air Blank	0.000	16:23	
Subject Sample # 3	0.224	16:26	
Breath Volume	1.429	16:26	
Air Blank	0.000	16:27	
Control Test	0.079	16:27	
Air Blank	0.000	16:28	
Diagnostics Check	OK	16:28	

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: BORENSTEIN

DATE/TIME OF TEST: 07/20/2023 00:07:52

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.082	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.154	00:02
Breath Volume	1.640	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.165	00:06
Breath Volume	2.023	00:06
Air Blank	0.000	00:06
Control Test	0.082	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: NOONAN

DATE/TIME OF TEST: 07/20/2023 03:27:01

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.082	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.150	03:21
Breath Volume	1.828	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.149	03:25
Breath Volume	1.140	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:27

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: PANDA

DATE/TIME OF TEST: 07/20/2023 04:27:58

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.082	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:23	No Sample Provided
Breath Volume	0.000	04:23	
Air Blank	0.000	04:24	
Air Blank	0.000	04:26	
Subject Sample	# 2	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:27	
Control Test	0.082	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: CLAVELO TARARONA

DATE/TIME OF TEST: 07/21/2023 04:40:07

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.083	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.056	04:35
Breath Volume	2.839	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.056	04:38
Breath Volume	2.203	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MARJI

DATE/TIME OF TEST: 07/21/2023 22:35:16

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.082	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.000	22:30
Breath Volume	3.261	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.000	22:33
Breath Volume	3.480	22:33
Air Blank	0.000	22:34
Control Test	0.082	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 07/22/2023 02:05:16

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.083	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.000	02:00
Breath Volume	2.328	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.000	02:03
Breath Volume	2.371	02:03
Air Blank	0.000	02:04
Control Test	0.082	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/22/2023 03:22:03

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.083	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.056	03:17
Breath Volume	1.390	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:19
Subject Sample # 2	0.060	03:20
Breath Volume	1.460	03:20
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: JARAMILLO JARAMILLO

DATE/TIME OF TEST: 07/22/2023 03:58:32

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.082	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.094	03:53
Breath Volume	2.265	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.095	03:56
Breath Volume	2.195	03:56
Air Blank	0.000	03:57
Control Test	0.082	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HARIG

DATE/TIME OF TEST: 07/23/2023 03:05:21

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.084	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.222	03:00
Breath Volume	3.500	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.225	03:03
Breath Volume	2.859	03:03
Air Blank	0.000	03:04
Control Test	0.082	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SANTANA, M
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 07/23/2023 04:52:45

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.083	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.079	04:48
Breath Volume	2.585	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.077	04:51
Breath Volume	2.472	04:51
Air Blank	0.000	04:51
Control Test	0.082	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: ISAAC

DATE/TIME OF TEST: 07/24/2023 04:13:48

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.083	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.000	04:09
Breath Volume	2.359	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.000	04:12
Breath Volume	1.703	04:12
Air Blank	0.000	04:12
Control Test	0.081	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/24/2023 21:08:28

Diagnostics Check	OK	20:59
Air Blank	0.000	20:59
Control Test	0.084	21:00
Air Blank	0.000	21:00
Subject Sample # 1	0.022	21:02
Breath Volume	1.914	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.020	21:06
Breath Volume	1.500	21:06
Air Blank	0.000	21:07
Control Test	0.082	21:07
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 07/24/2023 22:21:49

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.083	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.000	22:17
Breath Volume	3.710	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.000	22:20
Breath Volume	1.375	22:20
Air Blank	0.000	22:20
Control Test	0.082	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: DOVER

DATE/TIME OF TEST: 07/26/2023 00:42:19

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.083	00:31	
Air Blank	0.000	00:31	
Subject Sample	# 1 0.168	00:34	Volume Not Met
Breath Volume	0.136	00:34	
Air Blank	0.000	00:35	
Air Blank	0.000	00:37	
Subject Sample	# 2 0.179	00:40	Volume Not Met
Breath Volume	0.050	00:40	
Air Blank	0.000	00:41	
Control Test	0.080	00:41	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: DOVER

DATE/TIME OF TEST: 07/26/2023 01:00:08

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:45	
Control Test	0.080	00:46	
Air Blank	0.000	00:46	
Subject Sample	# 1 0.160	00:47	
Breath Volume	1.246	00:47	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 2 0.173	00:52	Volume Not Met
Breath Volume	0.238	00:52	
Air Blank	0.000	00:53	
Air Blank	0.000	00:55	
Subject Sample	# 3 0.168	00:58	Volume Not Met
Breath Volume	0.292	00:58	
Air Blank	0.000	00:59	
Control Test	0.080	00:59	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: CASTRO SIERRA

DATE/TIME OF TEST: 07/27/2023 04:16:04

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.082	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.000	04:11
Breath Volume	2.230	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.007	04:14
Breath Volume	3.097	04:14
Air Blank	0.000	04:15
Control Test	0.082	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:16

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO, E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: NAVA MARCOS

DATE/TIME OF TEST: 07/28/2023 01:02:06

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.083	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.131	00:57
Breath Volume	2.171	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.130	01:00
Breath Volume	1.941	01:00
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: OSORIO

DATE/TIME OF TEST: 07/28/2023 02:58:50

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.082	02:55	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.080	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: RAMSINGH

DATE/TIME OF TEST: 07/28/2023 04:36:08

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.082	04:25	
Air Blank	0.000	04:25	
Subject Sample	# 1 0.167	04:28	Volume Not Met
Breath Volume	0.128	04:28	
Air Blank	0.000	04:29	
Air Blank	0.000	04:31	
Subject Sample	# 2 0.168	04:34	Volume Not Met
Breath Volume	0.015	04:34	
Air Blank	0.000	04:35	
Control Test	0.081	04:35	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 08/01/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: RAMSINGH

DATE/TIME OF TEST: 07/28/2023 04:53:20

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:45	
Control Test	0.081	04:45	
Air Blank	0.000	04:46	
Subject Sample	# 1 0.156	04:49	Volume Not Met
Breath Volume	0.230	04:49	
Air Blank	0.000	04:50	
Air Blank	0.000	04:51	
Subject Sample	# 2	04:51	Subject Test Refused
Breath Volume	0.000	04:51	
Air Blank	0.000	04:52	
Control Test	0.079	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 08/01/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO, E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: SENGUN

DATE/TIME OF TEST: 07/28/2023 05:37:04

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.081	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.187	05:32
Breath Volume	3.062	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:34
Subject Sample # 2	0.179	05:35
Breath Volume	1.800	05:35
Air Blank	0.000	05:36
Control Test	0.081	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:37

Data Download Date/Time: 08/01/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/29/2023 03:19:38

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.083	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.129	03:14
Breath Volume	1.570	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.129	03:17
Breath Volume	1.367	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 08/01/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: CEXIL

DATE/TIME OF TEST: 07/29/2023 06:08:59

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.083	06:02
Air Blank	0.000	06:02
Subject Sample # 1	0.018	06:03
Breath Volume	1.523	06:03
Air Blank	0.000	06:04
Air Blank	0.000	06:06
Subject Sample # 2	0.019	06:07
Breath Volume	1.453	06:07
Air Blank	0.000	06:08
Control Test	0.082	06:08
Air Blank	0.000	06:08
Diagnostics Check	OK	06:08

Data Download Date/Time: 08/01/2023 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ROZELL

DATE/TIME OF TEST: 07/29/2023 22:53:23

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:40	
Control Test	0.084	22:41	
Air Blank	0.000	22:41	
Subject Sample	# 1 0.140	22:44	Volume Not Met
Breath Volume	0.472	22:44	
Air Blank	0.000	22:45	
Air Blank	0.000	22:47	
Subject Sample	# 2 0.177	22:48	
Breath Volume	1.542	22:48	
Air Blank	0.000	22:48	
Air Blank	0.000	22:50	
Subject Sample	# 3 0.177	22:51	
Breath Volume	1.285	22:51	
Air Blank	0.000	22:52	
Control Test	0.081	22:52	
Air Blank	0.000	22:53	
Diagnostics Check	OK	22:53	

Data Download Date/Time: 08/01/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: POLISE

DATE/TIME OF TEST: 07/30/2023 03:27:56

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.082	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.096	03:23
Breath Volume	1.828	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:25
Subject Sample # 2	0.101	03:26
Breath Volume	2.554	03:26
Air Blank	0.000	03:26
Control Test	0.083	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 08/01/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: AL HAJALI EL CHARANI

DATE/TIME OF TEST: 07/30/2023 05:54:11

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.083	05:48
Air Blank	0.000	05:48
Subject Sample # 1	0.177	05:49
Breath Volume	1.492	05:49
Air Blank	0.000	05:50
Air Blank	0.000	05:52
Subject Sample # 2	0.176	05:52
Breath Volume	1.144	05:52
Air Blank	0.000	05:53
Control Test	0.083	05:53
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

Data Download Date/Time: 08/01/2023 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BRECHBILL

DATE/TIME OF TEST: 07/28/2023 04:12:29

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.229	04:07
Breath Volume	2.058	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.223	04:10
Breath Volume	2.082	04:10
Air Blank	0.000	04:11
Control Test	0.078	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 08/08/2023 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,ELISHA,D
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: WENCES,SEFE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: BUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 07/02/2023 17:00:13

Diagnostics Check	OK	16:51
Air Blank	0.000	16:52
Control Test	0.080	16:52
Air Blank	0.000	16:53
Subject Sample # 1	0.165	16:53
Breath Volume	1.566	16:53
Air Blank	0.000	16:54
Air Blank	0.000	16:55
Subject Sample # 2	0.158	16:58
Breath Volume	1.699	16:58
Air Blank	0.000	16:59
Control Test	0.079	16:59
Air Blank	0.000	17:00
Diagnostics Check	OK	17:00

Data Download Date/Time: 07/26/2023 22:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/09/2023 23:36:45

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.156	23:32
Breath Volume	2.726	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.158	23:35
Breath Volume	3.085	23:35
Air Blank	0.000	23:35
Control Test	0.081	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 08/07/2023 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/12/2023 17:46:29

Diagnostics Check	OK	17:39
Air Blank	0.000	17:40
Control Test	0.080	17:40
Air Blank	0.000	17:41
Subject Sample # 1	0.000	17:42
Breath Volume	1.292	17:42
Air Blank	0.000	17:42
Air Blank	0.000	17:44
Subject Sample # 2	0.000	17:45
Breath Volume	1.281	17:45
Air Blank	0.000	17:45
Control Test	0.081	17:45
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

Data Download Date/Time: 08/07/2023 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/23/2023 20:46:40

Diagnostics Check	OK	20:36	
Air Blank	0.000	20:37	
Control Test	0.080	20:37	
Air Blank	0.000	20:38	
Subject Sample	# 1	20:41	No Sample Provided
Breath Volume	0.000	20:41	
Air Blank	0.000	20:41	
Air Blank	0.000	20:43	
Subject Sample	# 2	20:45	Subject Test Refused
Breath Volume	0.000	20:45	
Air Blank	0.000	20:45	
Control Test	0.080	20:46	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:46	

Data Download Date/Time: 08/07/2023 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COZAD,AUNDREA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: FARAH

DATE/TIME OF TEST: 07/23/2023 21:42:15

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:39	
Control Test	0.080	21:39	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:41	
Control Test	0.079	21:41	
Air Blank	0.000	21:42	
Diagnostics Check	OK	21:42	

Data Download Date/Time: 08/07/2023 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COZAD,AUNDREA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: FARAH

DATE/TIME OF TEST: 07/23/2023 21:52:12

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.228	21:47
Breath Volume	2.554	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.223	21:50
Breath Volume	1.472	21:50
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 08/07/2023 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/23/2023 22:03:06

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.078	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.189	21:58
Breath Volume	1.953	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.197	22:01
Breath Volume	2.269	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 08/07/2023 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:56

SUBJECT'S LAST NAME: USSERY

DATE/TIME OF TEST: 07/30/2023 16:23:04

Diagnostics Check	OK	16:16
Air Blank	0.000	16:16
Control Test	0.080	16:16
Air Blank	0.000	16:17
Subject Sample # 1	0.270	16:18
Breath Volume	1.402	16:18
Air Blank	0.000	16:18
Air Blank	0.000	16:20
Subject Sample # 2	0.277	16:21
Breath Volume	1.281	16:21
Air Blank	0.000	16:22
Control Test	0.079	16:22
Air Blank	0.000	16:22
Diagnostics Check	OK	16:23

Data Download Date/Time: 08/07/2023 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: SANCHEZ RODRIGUEZ

DATE/TIME OF TEST: 07/30/2023 17:35:28

Diagnostics Check	OK	17:28
Air Blank	0.000	17:29
Control Test	0.080	17:29
Air Blank	0.000	17:29
Subject Sample # 1	0.158	17:30
Breath Volume	2.203	17:30
Air Blank	0.000	17:30
Air Blank	0.000	17:32
Subject Sample # 2	0.152	17:33
Breath Volume	3.703	17:33
Air Blank	0.000	17:34
Control Test	0.080	17:34
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 08/07/2023 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATLEY,KENDALL,L
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: HATLEY,KENDALL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/31/2023 04:06:47

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.171	04:02
Breath Volume	3.023	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.168	04:05
Breath Volume	2.449	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 08/07/2023 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA, WILLIAM
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: LUNA OLVERA

DATE/TIME OF TEST: 07/05/2023 22:22:35

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.080	22:20	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:21	
Control Test	0.080	22:22	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: DEL CUETO,ERYKA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: SPECHT SOTO

DATE/TIME OF TEST: 07/10/2023 01:09:47

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:08	RFI Detect
Breath Volume	0.000	01:08	
Air Blank	0.000	01:08	
Control Test	0.080	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: DEL CUETO,ERYKA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: SPECHT SOTO

DATE/TIME OF TEST: 07/10/2023 01:16:47

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.080	01:12	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Control Test	0.080	01:16	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,MEGHAN,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: POPE,NICHOLAS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: REINFANDT

DATE/TIME OF TEST: 07/11/2023 23:23:57

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.000	23:19
Breath Volume	4.007	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.000	23:22
Breath Volume	3.535	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433249
BREATH TEST OPERATOR: NIEMIEC,NICHOLAS,J	DRY GAS EXPIRY DATE: 01/08/2024
BREATH TEST OPERATOR AGENCY: KEY COLONY BEACH PD	VIOLATION CODE: DUI
ARRESTING OFFICER: FOSTER,MEGHAN	ARREST TIME: 22:44
ARRESTING OFFICER AGENCY: MONROE COUNTY SO	GENDER: Female
LAST AGENCY INSPECTION: 07/05/2023	OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: ROBERTS	DATE/TIME OF TEST: 07/15/2023	23:53:27
------------------------------	-------------------------------	----------

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1 0.200	23:40	
Breath Volume	1.230	23:40	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample	# 2 0.147	23:46	Volume Not Met
Breath Volume	0.187	23:46	
Air Blank	0.000	23:46	
Air Blank	0.000	23:48	
Subject Sample	# 3 0.162	23:51	Volume Not Met
Breath Volume	0.054	23:51	
Air Blank	0.000	23:52	
Control Test	0.078	23:52	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NIEMIEC,NICHOLAS,J

BREATH TEST OPERATOR AGENCY: KEY COLONY BEACH PD

ARRESTING OFFICER: GUERRERO MARTINEZ,KEVEN

ARRESTING OFFICER AGENCY: MONROE COUNTY SO

LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433249

DRY GAS EXPIRY DATE: 01/08/2024

VIOLATION CODE: DUI

ARREST TIME: 22:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CASTELLANO HERNANDEZ

DATE/TIME OF TEST: 07/16/2023 00:51:37

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:45
Subject Sample	# 1 0.158	00:47
Breath Volume	2.800	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample	# 2 0.160	00:50
Breath Volume	2.652	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BLEICH

DATE/TIME OF TEST: 07/18/2023 21:39:48

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.080	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.159	21:35
Breath Volume	1.789	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.159	21:38
Breath Volume	2.171	21:38
Air Blank	0.000	21:38
Control Test	0.080	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: DEL CUETO,ERYKA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: O'DONNELL

DATE/TIME OF TEST: 07/18/2023 23:23:46

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.080	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:22	
Control Test	0.080	23:23	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILROY,ZACHARY,I
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GILROY,ZACHARY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/21/2023 19:15:34

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.081	19:06
Air Blank	0.000	19:07
Subject Sample # 1	0.212	19:07
Breath Volume	2.464	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.253	19:11
Breath Volume	2.671	19:11
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 3	0.229	19:13
Breath Volume	3.558	19:13
Air Blank	0.000	19:14
Control Test	0.080	19:14
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: NALE

DATE/TIME OF TEST: 07/24/2023 04:14:40

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.137	04:10
Breath Volume	3.019	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.126	04:13
Breath Volume	2.160	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/27/2023 00:45:19

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.081	00:43	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Control Test	0.081	00:44	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRERO-MARTINEZ,KEVEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 07/29/2023 05:46:45

Diagnostics Check	OK	05:43	
Air Blank	0.000	05:44	
Control Test	0.080	05:44	
Air Blank	0.000	05:44	
Subject Sample	# 1	05:45	Subject Test Refused
Breath Volume	0.000	05:45	
Air Blank	0.000	05:45	
Control Test	0.080	05:46	
Air Blank	0.000	05:46	
Diagnostics Check	OK	05:46	

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: JERNIGAN,DAVID
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: CHAPLIN

DATE/TIME OF TEST: 07/29/2023 23:37:13

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.233	23:32
Breath Volume	1.667	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.236	23:35
Breath Volume	1.949	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: POPE,NICHOLAS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: POLLOCK

DATE/TIME OF TEST: 07/30/2023 19:53:40

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:51	
Control Test	0.081	19:51	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:52	Subject Test Refused
Breath Volume	0.000	19:52	
Air Blank	0.000	19:52	
Control Test	0.080	19:53	
Air Blank	0.000	19:53	
Diagnostics Check	OK	19:53	

Data Download Date/Time: 08/03/2023 08:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MCARTHUR

DATE/TIME OF TEST: 07/31/2023 22:57:20

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:44	
Control Test	0.080	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 1 0.217	22:46	
Breath Volume	1.277	22:46	
Air Blank	0.000	22:47	
Air Blank	0.000	22:49	
Subject Sample	# 2 0.171	22:52	Volume Not Met
Breath Volume	0.105	22:52	
Air Blank	0.000	22:53	
Air Blank	0.000	22:54	
Subject Sample	# 3 0.197	22:55	
Breath Volume	1.152	22:55	
Air Blank	0.000	22:56	
Control Test	0.078	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 08/03/2023 08:00