

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS,E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: GRUNAT,JERRY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: DECEMBRE

DATE/TIME OF TEST: 08/01/2023 18:05:56

Diagnostics Check	OK	17:59
Air Blank	0.000	17:59
Control Test	0.079	17:59
Air Blank	0.000	18:00
Subject Sample # 1	0.000	18:01
Breath Volume	1.246	18:01
Air Blank	0.000	18:01
Air Blank	0.000	18:03
Subject Sample # 2	0.000	18:04
Breath Volume	1.468	18:04
Air Blank	0.000	18:05
Control Test	0.079	18:05
Air Blank	0.000	18:05
Diagnostics Check	OK	18:05

Data Download Date/Time: 08/03/2023 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BETHAM,KEITH,D	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: FONSECA,INES	ARREST TIME: 00:03
ARRESTING OFFICER AGENCY: Oviedo PD	GENDER: Female
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GAMBUZA	DATE/TIME OF TEST: 08/04/2023 01:06:13
------------------------------	--

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.040	01:01
Breath Volume	1.914	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.042	01:04
Breath Volume	2.664	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 08/24/2023 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HAMLIN,WILLIAM	ARREST TIME: 01:09
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SELLNEYER	DATE/TIME OF TEST: 08/05/2023	04:51:11
--------------------------------	-------------------------------	----------

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.079	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.151	04:46
Breath Volume	2.507	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.153	04:49
Breath Volume	3.222	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 08/24/2023 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MOXLEY,MICHAEL	ARREST TIME: 03:01
ARRESTING OFFICER AGENCY: Casselberry PD	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: TANIS	DATE/TIME OF TEST: 08/06/2023 04:07:40
----------------------------	--

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.221	04:02
Breath Volume	2.851	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.223	04:06
Breath Volume	2.511	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 08/24/2023 05:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CHINEA,TAMY,

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: BRYANT,STEVEN

ARRESTING OFFICER AGENCY: Longwood PD

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 01:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: GUILLEN MARMOL

DATE/TIME OF TEST: 08/07/2023 02:56:05

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:50
Air Blank	0.000	02:50
Subject Sample	# 1 0.264	02:51
Breath Volume	1.800	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample	# 2 0.262	02:54
Breath Volume	2.218	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 08/24/2023 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PRICE,CHASE	ARREST TIME: 18:38
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Female
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: MCFARLAND	DATE/TIME OF TEST: 08/07/2023	20:35:50
--------------------------------	-------------------------------	----------

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:32	
Control Test	0.079	20:32	
Air Blank	0.000	20:33	
Subject Sample	# 1	20:34	Interferent Detect
Breath Volume	2.109	20:34	
Air Blank	0.000	20:34	
Control Test	0.079	20:35	
Air Blank	0.000	20:35	
Diagnostics Check	OK	20:35	

Data Download Date/Time: 08/24/2023 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PRICE,CHASE	ARREST TIME: 18:38
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Female
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: MCFARLAND	DATE/TIME OF TEST: 08/07/2023	20:52:51
--------------------------------	-------------------------------	----------

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:49	
Control Test	0.080	20:49	
Air Blank	0.000	20:50	
Subject Sample	# 1	20:51	Interferent Detect
Breath Volume	2.281	20:51	
Air Blank	0.000	20:51	
Control Test	0.080	20:52	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 08/24/2023 05:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CHINEA,TAMY,

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: PRICE,CHASE

ARRESTING OFFICER AGENCY: Seminole County SO

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 18:38

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MCFARLAND

DATE/TIME OF TEST: 08/08/2023 00:51:19

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:48	
Control Test	0.080	00:48	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:49	Interferent Detect
Breath Volume	1.742	00:49	
Air Blank	0.000	00:50	
Control Test	0.080	00:50	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 08/24/2023 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: HIDEROA,KEVIN,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD	VIOLATION CODE: DUI
ARRESTING OFFICER: MORRISSEY,ZACHARY	ARREST TIME: 06:46
ARRESTING OFFICER AGENCY: Altamonte Springs PD	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: RIVAS	DATE/TIME OF TEST: 08/08/2023	11:37:08
----------------------------	-------------------------------	----------

Diagnostics Check	OK	11:30
Air Blank	0.000	11:31
Control Test	0.079	11:31
Air Blank	0.000	11:32
Subject Sample # 1	0.077	11:32
Breath Volume	1.207	11:32
Air Blank	0.000	11:33
Air Blank	0.000	11:35
Subject Sample # 2	0.075	11:35
Breath Volume	1.683	11:35
Air Blank	0.000	11:36
Control Test	0.080	11:36
Air Blank	0.000	11:37
Diagnostics Check	OK	11:37

Data Download Date/Time: 08/24/2023 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: JOHNSON,MELISSA	ARREST TIME: 03:44
ARRESTING OFFICER AGENCY: Casselberry PD	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: PHILLIP	DATE/TIME OF TEST: 08/09/2023 05:08:45
------------------------------	--

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.079	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.153	05:03
Breath Volume	1.199	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.162	05:07
Breath Volume	1.730	05:07
Air Blank	0.000	05:07
Control Test	0.079	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 08/24/2023 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: BANOS,FRANK	ARREST TIME: 02:00
ARRESTING OFFICER AGENCY: Maitland PD	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: OSORIO	DATE/TIME OF TEST: 08/13/2023 04:03:31
-----------------------------	--

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.076	03:59
Breath Volume	2.332	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.079	04:02
Breath Volume	3.257	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 08/24/2023 05:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CHINEA,TAMY,

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: GIOIELLI,ANTHONY

ARRESTING OFFICER AGENCY: Longwood PD

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 02:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: KARRI

DATE/TIME OF TEST: 08/13/2023 04:36:43

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.106	04:31
Breath Volume	1.660	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.100	04:35
Breath Volume	1.425	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 08/24/2023 05:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TEEHAN,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Altamonte Springs PD

ARRESTING OFFICER: TEEHAN,MATTHEW

ARRESTING OFFICER AGENCY: Altamonte Springs PD

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 07:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:26

SUBJECT'S LAST NAME: HUSSEY

DATE/TIME OF TEST: 08/13/2023 07:56:13

Diagnostics Check	OK	07:49
Air Blank	0.000	07:50
Control Test	0.079	07:50
Air Blank	0.000	07:51
Subject Sample # 1	0.000	07:51
Breath Volume	3.605	07:51
Air Blank	0.000	07:52
Air Blank	0.000	07:54
Subject Sample # 2	0.000	07:54
Breath Volume	3.050	07:54
Air Blank	0.000	07:55
Control Test	0.081	07:55
Air Blank	0.000	07:56
Diagnostics Check	OK	07:56

Data Download Date/Time: 08/24/2023 05:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOUCK,ZACHARY
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 08/13/2023 18:03:58

Diagnostics Check	OK	17:56
Air Blank	0.000	17:56
Control Test	0.079	17:57
Air Blank	0.000	17:57
Subject Sample # 1	0.135	17:59
Breath Volume	1.664	17:59
Air Blank	0.000	18:00
Air Blank	0.000	18:02
Subject Sample # 2	0.137	18:02
Breath Volume	1.136	18:02
Air Blank	0.000	18:03
Control Test	0.079	18:03
Air Blank	0.000	18:03
Diagnostics Check	OK	18:03

Data Download Date/Time: 08/24/2023 05:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PENA, ULYSSES, J

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: PENA, ULYSSES

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 17:18

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: MERILLO

DATE/TIME OF TEST: 08/13/2023 21:27:44

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.080	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.208	21:23
Breath Volume	2.460	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.216	21:26
Breath Volume	2.359	21:26
Air Blank	0.000	21:26
Control Test	0.077	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 08/24/2023 05:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PRICE,CHASE	ARREST TIME: 01:02
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ECHEVARRIA MORALES	DATE/TIME OF TEST: 08/14/2023 02:14:06
---	--

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.000	02:09
Breath Volume	2.515	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.000	02:12
Breath Volume	1.710	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 08/24/2023 05:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ANDREWS,GARY	ARREST TIME: 02:43
ARRESTING OFFICER AGENCY: Maitland PD	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: KHAIR EDDIN	DATE/TIME OF TEST: 08/14/2023 03:55:22
----------------------------------	--

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.156	03:50
Breath Volume	1.484	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.136	03:53
Breath Volume	1.578	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 08/24/2023 05:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIDDLE,JAMES,A

BREATH TEST OPERATOR AGENCY: Lake Mary PD

ARRESTING OFFICER: BITTIKOFER,PEGGY

ARRESTING OFFICER AGENCY: Winter Springs PD

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 21:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 08/14/2023 22:19:44

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.080	22:15	
Air Blank	0.000	22:15	
Subject Sample	# 1	22:18	RFI Detect
Breath Volume	0.742	22:18	
Air Blank	0.000	22:18	
Control Test	0.081	22:19	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

Data Download Date/Time: 08/24/2023 05:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIDDLE,JAMES,A

BREATH TEST OPERATOR AGENCY: Lake Mary PD

ARRESTING OFFICER: BITTIKOFER,PEGGY

ARRESTING OFFICER AGENCY: Winter Springs PD

LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 21:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 08/14/2023 22:30:14

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.237	22:25
Breath Volume	1.890	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.249	22:28
Breath Volume	1.898	22:28
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 08/24/2023 05:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SMITH,J	ARREST TIME: 23:47
ARRESTING OFFICER AGENCY: Maitland PD	GENDER: Female
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: KOSKY	DATE/TIME OF TEST: 08/15/2023	01:05:02
----------------------------	-------------------------------	----------

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.198	00:59
Breath Volume	1.828	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.197	01:03
Breath Volume	1.648	01:03
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:05

Data Download Date/Time: 08/24/2023 05:04

Subject Test Electronic Data

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GARCIA-CHOW,LUIS	ARREST TIME: 04:40
ARRESTING OFFICER AGENCY: Oviedo PD	GENDER: Male
LAST AGENCY INSPECTION: 07/20/2023	OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: BOWIE	DATE/TIME OF TEST: 08/19/2023 05:35:28
----------------------------	--

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.081	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.159	05:31
Breath Volume	3.707	05:31
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.160	05:34
Breath Volume	3.875	05:34
Air Blank	0.000	05:34
Control Test	0.080	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 08/24/2023 05:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: KRUPA,KYLE
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: PRYOR

DATE/TIME OF TEST: 08/24/2023 21:59:24

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:48	
Control Test	0.080	21:48	
Air Blank	0.000	21:49	
Subject Sample	# 1	21:52	No Sample Provided
Breath Volume	0.000	21:52	
Air Blank	0.000	21:52	
Air Blank	0.000	21:54	
Subject Sample	# 2	21:58	No Sample Provided
Breath Volume	0.000	21:58	
Air Blank	0.000	21:58	
Control Test	0.081	21:58	
Air Blank	0.000	21:59	
Diagnostics Check	OK	21:59	

Data Download Date/Time: 09/06/2023 01:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BETHAM,KEITH,D

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: CILENTI,ANTHONY

ARRESTING OFFICER AGENCY: Maitland PD

LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 02:59

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 08/26/2023 04:48:39

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:47	
Control Test	0.080	04:48	RFI Detect
Air Blank	0.000	04:48	

Data Download Date/Time: 09/06/2023 01:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BETHAM,KEITH,D

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: CILENTI,ANTHONY

ARRESTING OFFICER AGENCY: Maitland PD

LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 02:59

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 08/26/2023 04:51:41

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:50	
Control Test	0.080	04:51	
Air Blank		04:51	RFI Detect
Air Blank		04:51	RFI Detect

Data Download Date/Time: 09/06/2023 01:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BETHAM,KEITH,D	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: CILENTI,ANTHONY	ARREST TIME: 02:59
ARRESTING OFFICER AGENCY: Maitland PD	GENDER: Female
LAST AGENCY INSPECTION: 08/23/2023	OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: DOYLE	DATE/TIME OF TEST: 08/26/2023	04:59:58
----------------------------	-------------------------------	----------

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.081	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.199	04:55
Breath Volume	1.660	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.185	04:58
Breath Volume	1.460	04:58
Air Blank	0.000	04:59
Control Test	0.081	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 09/06/2023 02:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BLAZINA,MEGAN,N	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ADAMS,J	ARREST TIME: 17:08
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 08/23/2023	OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: JUAREZ	DATE/TIME OF TEST: 08/27/2023	18:38:19
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.080	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.200	18:33
Breath Volume	4.144	18:33
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.202	18:36
Breath Volume	2.308	18:36
Air Blank	0.000	18:37
Control Test	0.080	18:37
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 09/06/2023 02:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: CHINEA,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: VARGAS,JULISSA	ARREST TIME: 00:09
ARRESTING OFFICER AGENCY: Oviedo PD	GENDER: Male
LAST AGENCY INSPECTION: 08/23/2023	OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: ARICO	DATE/TIME OF TEST: 08/28/2023 01:24:23
----------------------------	--

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.175	01:19
Breath Volume	1.488	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.186	01:22
Breath Volume	1.812	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 09/06/2023 02:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NOLAND,THOMAS,E

BREATH TEST OPERATOR AGENCY: Lake Mary PD

ARRESTING OFFICER: CAVALLERO,TRAVIS

ARRESTING OFFICER AGENCY: Casselberry PD

LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 05:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 08/30/2023 06:45:37

Diagnostics Check	OK	06:38
Air Blank	0.000	06:38
Control Test	0.080	06:39
Air Blank	0.000	06:39
Subject Sample # 1	0.263	06:40
Breath Volume	1.710	06:40
Air Blank	0.000	06:41
Air Blank	0.000	06:43
Subject Sample # 2	0.255	06:44
Breath Volume	1.371	06:44
Air Blank	0.000	06:44
Control Test	0.079	06:45
Air Blank	0.000	06:45
Diagnostics Check	OK	06:45

Data Download Date/Time: 09/06/2023 02:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: WATSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: DEHOFF

DATE/TIME OF TEST: 08/13/2023 16:19:19

Diagnostics Check	OK	16:13
Air Blank	0.000	16:13
Control Test	0.079	16:13
Air Blank	0.000	16:14
Subject Sample # 1	0.220	16:14
Breath Volume	2.390	16:14
Air Blank	0.000	16:15
Air Blank	0.000	16:17
Subject Sample # 2	0.227	16:17
Breath Volume	2.847	16:17
Air Blank	0.000	16:18
Control Test	0.078	16:18
Air Blank	0.000	16:19
Diagnostics Check	OK	16:19

Data Download Date/Time: 08/15/2023 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLANDA, RICARDO, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HOLANDA, RICARDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 08/20/2023 02:03:30

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.190	01:57
Breath Volume	1.226	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.178	02:01
Breath Volume	1.585	02:01
Air Blank	0.000	02:02
Control Test	0.081	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 09/13/2023 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLANDA, RICARDO, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HOLANDA, RICARDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MCWILLIAMS

DATE/TIME OF TEST: 08/24/2023 03:45:57

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.091	03:41
Breath Volume	1.261	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.089	03:44
Breath Volume	2.332	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 09/13/2023 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,LIBBEE,S
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: JABLANSKI,JOSHUA
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 08/05/2023 02:24:18

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.131	02:19
Breath Volume	2.554	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.129	02:22
Breath Volume	1.828	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 08/22/2023 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SHELAINE,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: SAPUNDJIEFF,THOMAS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/06/2023 04:19:44

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.181	04:14
Breath Volume	1.496	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.180	04:18
Breath Volume	2.042	04:18
Air Blank	0.000	04:18
Control Test	0.080	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 08/22/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CARLTON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DUGGER,MARK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WILKERSON

DATE/TIME OF TEST: 08/07/2023 02:08:10

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.080	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:55	No Sample Provided
Breath Volume	0.000	01:55	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample	# 2	0.155	02:00
Breath Volume	1.234	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:03	
Subject Sample	# 3	0.118	02:06
Breath Volume	0.441	02:06	Volume Not Met
Air Blank	0.000	02:07	
Control Test	0.079	02:07	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 08/22/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MINCK,ISAAC
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: SKUNTZ

DATE/TIME OF TEST: 08/13/2023 03:20:14

Diagnostics Check	OK	03:19	
Air Blank		03:19	RFI Detect
Air Blank	0.000	03:20	

Data Download Date/Time: 08/22/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MINCK,ISAAC
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: SKUNTZ

DATE/TIME OF TEST: 08/13/2023 03:33:29

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:26	No Sample Provided
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 2	03:32	No Sample Provided
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.081	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 08/22/2023 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,JORDI,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CRANDALL,DOUGLAS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MUSCARELLA

DATE/TIME OF TEST: 08/17/2023 22:31:10

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:26	
Control Test	0.079	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1	22:29	Subject Test Refused
Breath Volume	0.000	22:29	
Air Blank	0.000	22:30	
Control Test	0.081	22:30	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 08/22/2023 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,LIBBEE,S
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: WILLIAMS,STEPHANIE
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220644
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BYARS

DATE/TIME OF TEST: 08/19/2023 03:48:46

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.068	03:48	Control Outside Tolerance
Air Blank	0.000	03:48	

Data Download Date/Time: 08/22/2023 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,LIBBEE,S
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: WILLIAMS,STEPHANIE
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220644
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: BYARS

DATE/TIME OF TEST: 08/19/2023 04:37:47

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.081	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:30	No Sample Provided
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2	04:36	No Sample Provided
Breath Volume	0.000	04:36	
Air Blank	0.000	04:36	
Control Test	0.080	04:37	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 08/22/2023 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LEIGHA,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: JONES,LEIGHA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: TORRES DE JESUS

DATE/TIME OF TEST: 08/24/2023 19:56:32

Diagnostics Check	OK	19:53	
Air Blank	0.000	19:54	
Control Test	0.080	19:54	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:55	Subject Test Refused
Breath Volume	0.000	19:55	
Air Blank	0.000	19:55	
Control Test	0.080	19:55	
Air Blank	0.000	19:56	
Diagnostics Check	OK	19:56	

Data Download Date/Time: 09/19/2023 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLANES,MILAGROS,O
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: LLANES,MILAGROS
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: URTNOWSKI

DATE/TIME OF TEST: 08/25/2023 03:57:02

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.215	03:52
Breath Volume	4.382	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:54
Subject Sample # 2	0.218	03:55
Breath Volume	5.320	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:57

Data Download Date/Time: 09/19/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CARLTON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LAMAR,JOSHUA
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MAZZOTTA

DATE/TIME OF TEST: 08/30/2023 00:53:53

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.219	00:46
Breath Volume	2.171	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.196	00:49
Breath Volume	1.558	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 3	0.232	00:52
Breath Volume	3.558	00:52
Air Blank	0.000	00:52
Control Test	0.081	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 09/19/2023 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: HAMILTON,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: DARROUGH

DATE/TIME OF TEST: 08/05/2023 01:44:36

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1 0.117	01:35	Volume Not Met
Breath Volume	0.242	01:35	
Air Blank	0.000	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 2 0.127	01:39	
Breath Volume	2.660	01:39	
Air Blank	0.000	01:39	
Air Blank	0.000	01:41	
Subject Sample	# 3 0.137	01:42	
Breath Volume	1.906	01:42	
Air Blank	0.000	01:43	
Control Test	0.078	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 08/09/2023 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINS,JAMES,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: NINA-LIRANZO

DATE/TIME OF TEST: 08/13/2023 05:53:36

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.079	05:48
Air Blank	0.000	05:48
Subject Sample # 1	0.143	05:49
Breath Volume	1.941	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.157	05:52
Breath Volume	2.511	05:52
Air Blank	0.000	05:52
Control Test	0.079	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 08/30/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,ALEXANDR,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FISCHER,ALEXANDR
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: GORDLEY

DATE/TIME OF TEST: 08/16/2023 16:03:37

Diagnostics Check	OK	16:02	
Air Blank		16:03	RFI Detect
Air Blank	0.000	16:03	

Data Download Date/Time: 08/30/2023 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,ALEXANDR,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FISCHER,ALEXANDR
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: GORDLEY

DATE/TIME OF TEST: 08/16/2023 16:08:14

Diagnostics Check	OK	16:06	
Air Blank	0.000	16:07	
Control Test	0.007	16:07	Control Outside Tolerance
Air Blank	0.000	16:08	

Data Download Date/Time: 08/30/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,ALEXANDR,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FISCHER,ALEXANDR
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: GORDLEY

DATE/TIME OF TEST: 08/16/2023 16:10:48

Diagnostics Check	Fail	16:10	Internal Printer Test
Air Blank	0.000	16:10	

Data Download Date/Time: 08/30/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,ALEXANDR,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FISCHER,ALEXANDR
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: GORDLEY

DATE/TIME OF TEST: 08/16/2023 16:29:04

Diagnostics Check	OK	16:23
Air Blank	0.000	16:23
Control Test	0.079	16:23
Air Blank	0.000	16:24
Subject Sample # 1	0.000	16:24
Breath Volume	3.265	16:24
Air Blank	0.000	16:25
Air Blank	0.000	16:27
Subject Sample # 2	0.000	16:27
Breath Volume	4.023	16:27
Air Blank	0.000	16:28
Control Test	0.080	16:28
Air Blank	0.000	16:28
Diagnostics Check	OK	16:29

Data Download Date/Time: 08/30/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,ALEXANDR,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CULBERTSON,TYLER
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: WILLMAN

DATE/TIME OF TEST: 08/17/2023 17:12:34

Diagnostics Check	OK	17:06
Air Blank	0.000	17:06
Control Test	0.079	17:07
Air Blank	0.000	17:07
Subject Sample # 1	0.201	17:08
Breath Volume	3.660	17:08
Air Blank	0.000	17:08
Air Blank	0.000	17:10
Subject Sample # 2	0.210	17:10
Breath Volume	2.316	17:10
Air Blank	0.000	17:11
Control Test	0.078	17:11
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

Data Download Date/Time: 08/30/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: ESCOTO

DATE/TIME OF TEST: 08/19/2023 07:32:11

Diagnostics Check	OK	07:26
Air Blank	0.000	07:26
Control Test	0.080	07:26
Air Blank	0.000	07:27
Subject Sample # 1	0.180	07:27
Breath Volume	1.722	07:27
Air Blank	0.000	07:28
Air Blank	0.000	07:30
Subject Sample # 2	0.172	07:30
Breath Volume	1.550	07:30
Air Blank	0.000	07:31
Control Test	0.079	07:31
Air Blank	0.000	07:32
Diagnostics Check	OK	07:32

Data Download Date/Time: 08/30/2023 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: LESLIE

DATE/TIME OF TEST: 08/27/2023 22:38:58

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.078	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.152	22:34
Breath Volume	2.714	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:36
Subject Sample # 2	0.148	22:37
Breath Volume	2.781	22:37
Air Blank	0.000	22:38
Control Test	0.077	22:38
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 08/30/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,AMANDA,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MILLER,JERRY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO325341-8
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 08/28/2023 03:03:17

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:59	No Sample Provided
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample	# 2	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Control Test	0.079	03:02	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 08/30/2023 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 08/06/2023 01:02:02

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.000	00:57
Breath Volume	1.882	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.000	01:00
Breath Volume	1.609	01:00
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:02

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: LADNER

DATE/TIME OF TEST: 08/08/2023 18:36:07

Diagnostics Check	OK	18:26
Air Blank	0.000	18:27
Control Test	0.079	18:27
Air Blank	0.000	18:28
Subject Sample # 1	0.269	18:31
Breath Volume	1.671	18:31
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.261	18:34
Breath Volume	1.339	18:34
Air Blank	0.000	18:35
Control Test	0.077	18:35
Air Blank	0.000	18:36
Diagnostics Check	OK	18:36

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: YECKLEY

DATE/TIME OF TEST: 08/09/2023 03:40:22

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.082	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.000	03:36
Breath Volume	2.628	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.000	03:39
Breath Volume	3.007	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: OVERMAN

DATE/TIME OF TEST: 08/10/2023 23:32:36

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.081	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.000	23:28
Breath Volume	3.574	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.000	23:31
Breath Volume	1.875	23:31
Air Blank	0.000	23:31
Control Test	0.081	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WHITE,TRAVIS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/11/2023 23:55:38

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.182	23:48
Breath Volume	1.707	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.184	23:53
Breath Volume	2.429	23:53
Air Blank	0.000	23:54
Control Test	0.078	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 08/13/2023 03:53:06

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.082	03:52	
Air Blank		03:52	RFI Detect
Air Blank	0.000	03:53	

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 08/13/2023 03:57:18

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.081	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:56	
Control Test	0.082	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BRUCE,HEATHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/14/2023 00:52:15

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:49	
Control Test	0.080	00:50	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.081	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIM,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 14:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:42

SUBJECT'S LAST NAME: HARBIN JOHNSON

DATE/TIME OF TEST: 08/14/2023 15:15:50

Diagnostics Check	OK	15:09	
Air Blank	0.000	15:10	
Control Test	0.080	15:10	
Air Blank	0.000	15:11	
Subject Sample # 1	0.000	15:12	
Breath Volume	1.500	15:12	
Air Blank	0.000	15:13	
Air Blank	0.000	15:15	
Subject Sample # 2		15:15	Sequence Aborted
Breath Volume	0.000	15:15	
Air Blank	0.000	15:15	

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: PLYMALE

DATE/TIME OF TEST: 08/16/2023 00:44:23

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:36	
Control Test	0.080	00:36	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:40	No Sample Provided
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Air Blank	0.000	00:42	
Subject Sample	# 2	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:43	
Control Test	0.080	00:43	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MAYNARD,AARON
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SMITHA

DATE/TIME OF TEST: 08/16/2023 01:20:16

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.077	01:15	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:18	Subject Test Refused
Breath Volume	0.000	01:18	
Air Blank	0.000	01:19	
Control Test	0.077	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: PROB/PAROL
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 08/17/2023 23:11:17

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.139	23:05
Breath Volume	2.105	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.138	23:09
Breath Volume	1.824	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 08/18/2023 21:23:05

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.080	21:16
Air Blank	0.000	21:17
Subject Sample # 1	0.114	21:18
Breath Volume	1.714	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:20
Subject Sample # 2	0.115	21:21
Breath Volume	1.925	21:21
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEIB,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BAINES

DATE/TIME OF TEST: 08/19/2023 01:08:40

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: HOUSE

DATE/TIME OF TEST: 08/19/2023 21:52:09

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.081	21:45
Air Blank	0.000	21:46
Subject Sample # 1	0.000	21:46
Breath Volume	1.371	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.000	21:50
Breath Volume	2.058	21:50
Air Blank	0.000	21:51
Control Test	0.081	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 08/19/2023 23:27:02

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.088	23:22
Breath Volume	2.074	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:24
Subject Sample # 2	0.091	23:25
Breath Volume	2.023	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:27

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: GOYNES

DATE/TIME OF TEST: 08/20/2023 01:16:16

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:13	
Control Test	0.081	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:15	
Control Test	0.080	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORTENBERRY,ZACHARY,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FORTENBERRY,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 08/20/2023 01:41:36

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:37
Breath Volume	1.398	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.000	01:40
Breath Volume	1.441	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 08/20/2023 06:10:29

Diagnostics Check	OK	06:03
Air Blank	0.000	06:04
Control Test	0.081	06:04
Air Blank	0.000	06:05
Subject Sample # 1	0.155	06:06
Breath Volume	1.464	06:06
Air Blank	0.000	06:06
Air Blank	0.000	06:08
Subject Sample # 2	0.153	06:08
Breath Volume	1.992	06:08
Air Blank	0.000	06:09
Control Test	0.080	06:09
Air Blank	0.000	06:10
Diagnostics Check	OK	06:10

Data Download Date/Time: 09/05/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, TYLER, M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BLACKWOOD, PATRICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: MOHAMED BOUNE

DATE/TIME OF TEST: 08/22/2023 23:01:15

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:53	
Control Test	0.081	22:53	
Air Blank	0.000	22:54	
Subject Sample	# 1 0.184	22:54	
Breath Volume	1.734	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:57	
Subject Sample	# 2	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, TYLER, M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WYCHUNAS, REIDE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 08/23/2023 01:14:22

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.080	01:12	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:13	
Control Test	0.081	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 08/23/2023 02:26:31

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.019	02:21
Breath Volume	1.355	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.017	02:25
Breath Volume	1.332	02:25
Air Blank	0.000	02:25
Control Test	0.081	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER, MONICA, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MILLENDER, MONICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: FABEL

DATE/TIME OF TEST: 08/24/2023 03:13:06

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.049	03:08
Breath Volume	1.808	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.047	03:11
Breath Volume	3.050	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NASH,JUSTIN,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NASH,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 10:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:26

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 08/25/2023 12:52:47

Diagnostics Check	OK	12:46
Air Blank	0.000	12:46
Control Test	0.082	12:47
Air Blank	0.000	12:47
Subject Sample # 1	0.000	12:48
Breath Volume	2.382	12:48
Air Blank	0.000	12:48
Air Blank	0.000	12:50
Subject Sample # 2	0.000	12:51
Breath Volume	1.265	12:51
Air Blank	0.000	12:51
Control Test	0.082	12:52
Air Blank	0.000	12:52
Diagnostics Check	OK	12:52

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODDO,ISIAH,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ODDO,ISIAH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GUIDO

DATE/TIME OF TEST: 08/26/2023 00:47:21

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.081	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.045	00:42
Breath Volume	3.187	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.045	00:45
Breath Volume	3.195	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: CRAVEY

DATE/TIME OF TEST: 08/26/2023 17:43:22

Diagnostics Check	OK	17:36
Air Blank	0.000	17:36
Control Test	0.081	17:37
Air Blank	0.000	17:37
Subject Sample # 1	0.000	17:38
Breath Volume	1.625	17:38
Air Blank	0.000	17:38
Air Blank	0.000	17:40
Subject Sample # 2	0.000	17:42
Breath Volume	1.800	17:42
Air Blank	0.000	17:42
Control Test	0.081	17:42
Air Blank	0.000	17:43
Diagnostics Check	OK	17:43

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: ONEIL

DATE/TIME OF TEST: 08/26/2023 22:17:59

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.081	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.230	22:13
Breath Volume	2.480	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:15
Subject Sample # 2	0.217	22:16
Breath Volume	2.722	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODDO,ISAIAH,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ODDO,ISAIAH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MAYNARD

DATE/TIME OF TEST: 08/26/2023 22:54:11

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:51	
Control Test	0.080	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 1	22:52	Subject Test Refused
Breath Volume	0.000	22:52	
Air Blank	0.000	22:53	
Control Test	0.079	22:53	
Air Blank	0.000	22:54	
Diagnostics Check	OK	22:54	

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BICKNELL,BRANDON,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BRUCE,HEATHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: MORIE

DATE/TIME OF TEST: 08/27/2023 00:36:06

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:24	
Control Test	0.079	00:25	
Air Blank	0.000	00:25	
Subject Sample	# 1 0.115	00:28	Volume Not Met
Breath Volume	0.308	00:28	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 2 0.108	00:31	
Breath Volume	1.679	00:31	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 3 0.111	00:34	
Breath Volume	1.832	00:34	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, TYLER, M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JONES, TYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ROLLO

DATE/TIME OF TEST: 08/27/2023 01:44:10

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRILL,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MERRILL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/28/2023 15:28:23

Diagnostics Check	OK	15:21
Air Blank	0.000	15:22
Control Test	0.080	15:22
Air Blank	0.000	15:22
Subject Sample # 1	0.233	15:23
Breath Volume	1.472	15:23
Air Blank	0.000	15:24
Air Blank	0.000	15:25
Subject Sample # 2	0.228	15:26
Breath Volume	1.722	15:26
Air Blank	0.000	15:27
Control Test	0.079	15:27
Air Blank	0.000	15:28
Diagnostics Check	OK	15:28

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: HUNTINGTON

DATE/TIME OF TEST: 08/30/2023 01:58:31

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.198	01:53
Breath Volume	2.199	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.190	01:56
Breath Volume	2.003	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 09/05/2023 17:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSKO,KRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 08/05/2023 05:49:25

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.080	05:43
Air Blank	0.000	05:44
Subject Sample # 1	0.116	05:44
Breath Volume	1.640	05:44
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.118	05:47
Breath Volume	1.304	05:47
Air Blank	0.000	05:48
Control Test	0.078	05:48
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 08/29/2023 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/05/2023 23:03:05

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.079	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.184	22:58
Breath Volume	2.449	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:00
Subject Sample # 2	0.188	23:01
Breath Volume	2.359	23:01
Air Blank	0.000	23:02
Control Test	0.077	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 08/29/2023 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 08/06/2023 22:12:05

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:09	
Control Test	0.080	22:09	
Air Blank	0.000	22:09	
Subject Sample	# 1	22:10	Subject Test Refused
Breath Volume	0.000	22:10	
Air Blank	0.000	22:11	
Control Test	0.078	22:11	
Air Blank	0.000	22:12	
Diagnostics Check	OK	22:12	

Data Download Date/Time: 08/29/2023 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: NICKELSON

DATE/TIME OF TEST: 08/08/2023 23:22:35

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.000	23:18
Breath Volume	3.347	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.000	23:21
Breath Volume	2.960	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 08/29/2023 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEDJESKE,JAMES,R
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: TEDJESKE,JAMES
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 08/17/2023 02:47:20

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.079	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.080	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 08/29/2023 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HANSEN,ERIC
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: TUNSIL

DATE/TIME OF TEST: 08/19/2023 03:21:29

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.154	03:16
Breath Volume	1.207	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:18
Subject Sample # 2	0.154	03:19
Breath Volume	2.195	03:19
Air Blank	0.000	03:20
Control Test	0.078	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 08/29/2023 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ROBINS

DATE/TIME OF TEST: 08/20/2023 01:25:07

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.112	01:20
Breath Volume	1.699	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.109	01:23
Breath Volume	1.820	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 08/29/2023 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MANSFIELD

DATE/TIME OF TEST: 08/20/2023 03:58:24

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.078	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.164	03:53
Breath Volume	2.582	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.159	03:56
Breath Volume	2.593	03:56
Air Blank	0.000	03:57
Control Test	0.077	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 08/29/2023 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: BUNTON

DATE/TIME OF TEST: 08/20/2023 04:16:20

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.079	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:06	No Sample Provided
Breath Volume	0.000	04:06	
Air Blank	0.000	04:07	
Air Blank	0.000	04:09	
Subject Sample	# 2	0.203	No .020 agreement
Breath Volume	2.371	04:10	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample	# 3	0.179	No .020 agreement
Breath Volume	1.261	04:14	
Air Blank	0.000	04:15	
Control Test	0.078	04:15	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 08/29/2023 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSKO,KRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: ALVAREZ SALCEDO

DATE/TIME OF TEST: 08/20/2023 14:45:25

Diagnostics Check	OK	14:39
Air Blank	0.000	14:39
Control Test	0.079	14:39
Air Blank	0.000	14:40
Subject Sample # 1	0.000	14:41
Breath Volume	2.375	14:41
Air Blank	0.000	14:41
Air Blank	0.000	14:43
Subject Sample # 2	0.000	14:44
Breath Volume	2.406	14:44
Air Blank	0.000	14:44
Control Test	0.079	14:44
Air Blank	0.000	14:45
Diagnostics Check	OK	14:45

Data Download Date/Time: 08/29/2023 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: WHITEHEAD

DATE/TIME OF TEST: 08/20/2023 23:06:28

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.264	22:58
Breath Volume	2.449	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.231	23:01
Breath Volume	1.359	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 3	0.236	23:04
Breath Volume	1.300	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 08/29/2023 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: GUERRY,CHAD
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:36

SUBJECT'S LAST NAME: ARROYO-VAZQUEZ

DATE/TIME OF TEST: 08/23/2023 12:05:41

Diagnostics Check	OK	11:56
Air Blank	0.000	11:57
Control Test	0.079	11:57
Air Blank	0.000	11:58
Subject Sample # 1	0.269	11:59
Breath Volume	1.718	11:59
Air Blank	0.000	11:59
Air Blank	0.000	12:01
Subject Sample # 2	0.279	12:03
Breath Volume	1.488	12:03
Air Blank	0.000	12:04
Control Test	0.077	12:04
Air Blank	0.000	12:05
Diagnostics Check	OK	12:05

Data Download Date/Time: 08/29/2023 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEMMONS, LAQUINTA, M
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: COOK, SAMUEL
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:38

SUBJECT'S LAST NAME: STARLING

DATE/TIME OF TEST: 08/26/2023 10:40:55

Diagnostics Check	OK	10:31
Air Blank	0.000	10:31
Control Test	0.080	10:31
Air Blank	0.000	10:32
Subject Sample # 1	0.269	10:34
Breath Volume	1.687	10:34
Air Blank	0.000	10:35
Air Blank	0.000	10:37
Subject Sample # 2	0.273	10:38
Breath Volume	1.968	10:38
Air Blank	0.000	10:39
Control Test	0.076	10:40
Air Blank	0.000	10:40
Diagnostics Check	OK	10:40

Data Download Date/Time: 08/29/2023 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRBY,CHAD,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KIRBY,CHAD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 08/26/2023 22:32:06

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.000	22:27
Breath Volume	1.164	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.000	22:30
Breath Volume	2.582	22:30
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 08/29/2023 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: SMITH JR,LORNE
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 07/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: RAULERSON

DATE/TIME OF TEST: 08/01/2023 17:38:58

Diagnostics Check	OK	17:32	
Air Blank	0.000	17:32	
Control Test	0.082	17:32	
Air Blank	0.000	17:33	
Subject Sample	# 1 0.113	17:34	
Breath Volume	2.828	17:34	
Air Blank	0.000	17:35	
Air Blank	0.000	17:36	
Subject Sample	# 2 0.106	17:38	
Breath Volume	2.101	17:38	
Air Blank		17:38	Improper Sample
Air Blank	0.000	17:38	

Data Download Date/Time: 08/15/2023 20:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: SMITH JR,LORNE
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 07/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: RAULERSON

DATE/TIME OF TEST: 08/01/2023 17:47:57

Diagnostics Check	OK	17:41
Air Blank	0.000	17:41
Control Test	0.081	17:42
Air Blank	0.000	17:42
Subject Sample # 1	0.113	17:43
Breath Volume	3.355	17:43
Air Blank	0.000	17:44
Air Blank	0.000	17:45
Subject Sample # 2	0.108	17:46
Breath Volume	3.085	17:46
Air Blank	0.000	17:47
Control Test	0.077	17:47
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 08/15/2023 20:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,NATHAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SUMNER,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: HUMPHRIES

DATE/TIME OF TEST: 08/02/2023 05:47:45

Diagnostics Check	OK	05:41
Air Blank	0.000	05:41
Control Test	0.083	05:42
Air Blank	0.000	05:42
Subject Sample # 1	0.126	05:43
Breath Volume	3.183	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:45
Subject Sample # 2	0.120	05:46
Breath Volume	3.027	05:46
Air Blank	0.000	05:46
Control Test	0.081	05:47
Air Blank	0.000	05:47
Diagnostics Check	OK	05:47

Data Download Date/Time: 08/15/2023 20:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: SMITH JR,LORNE
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 07/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 08/06/2023 15:27:29

Diagnostics Check	OK	15:24	
Air Blank	0.000	15:24	
Control Test	0.082	15:24	
Air Blank	0.000	15:25	
Subject Sample	# 1	15:26	Subject Test Refused
Breath Volume	0.000	15:26	
Air Blank	0.000	15:26	
Control Test	0.082	15:26	
Air Blank	0.000	15:27	
Diagnostics Check	OK	15:27	

Data Download Date/Time: 08/15/2023 20:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LORD,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623825
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 08/13/2023 00:47:09

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.081	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.169	00:42
Breath Volume	1.417	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.173	00:45
Breath Volume	1.906	00:45
Air Blank	0.000	00:46
Control Test	0.082	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 08/15/2023 20:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARX,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WILLIAMS,KYLE
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 07/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DIAZ-RODRIGUEZ

DATE/TIME OF TEST: 08/13/2023 02:13:33

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.082	02:09	
Air Blank	0.000	02:10	
Subject Sample	# 1 0.203	02:11	
Breath Volume	2.953	02:11	
Air Blank		02:12	Purge Fail
Air Blank		02:13	Purge Fail

Data Download Date/Time: 08/15/2023 20:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARX,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WILLIAMS,KYLE
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 07/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DIAZ-RODRIGUEZ

DATE/TIME OF TEST: 08/13/2023 02:26:35

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.210	02:21
Breath Volume	2.636	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.202	02:24
Breath Volume	2.371	02:24
Air Blank	0.000	02:25
Control Test	0.081	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 08/15/2023 20:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 08/13/2023 04:10:49

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.063	04:06
Breath Volume	2.230	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:08
Subject Sample # 2	0.064	04:09
Breath Volume	1.742	04:09
Air Blank	0.000	04:09
Control Test	0.083	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 08/15/2023 20:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: PIERCE,MARK
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 07/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:32

SUBJECT'S LAST NAME: STAHL

DATE/TIME OF TEST: 08/13/2023 07:00:38

Diagnostics Check	OK	06:52	
Air Blank	0.000	06:52	
Control Test	0.082	06:52	
Air Blank	0.000	06:53	
Subject Sample	# 1	06:56	No Sample Provided
Breath Volume	0.000	06:56	
Air Blank	0.000	06:57	
Air Blank	0.000	06:59	
Subject Sample	# 2	06:59	Subject Test Refused
Breath Volume	0.000	06:59	
Air Blank	0.000	06:59	
Control Test	0.082	07:00	
Air Blank	0.000	07:00	
Diagnostics Check	OK	07:00	

Data Download Date/Time: 08/15/2023 20:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: SMITH JR,LORNE
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 08/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623825
DRY GAS EXPIRY DATE: 12/25/2025
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 08/20/2023 01:14:34

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.082	01:12	
Air Blank	0.000	01:12	
Subject Sample # 1	0.223	01:13	
Breath Volume	2.171	01:13	
Air Blank		01:14	Purge Fail
Air Blank	0.000	01:14	

Data Download Date/Time: 09/25/2023 23:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: SMITH JR,LORNE
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 08/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623825
DRY GAS EXPIRY DATE: 12/25/2025
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: HILARIO

DATE/TIME OF TEST: 08/20/2023 01:25:18

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.211	01:19
Breath Volume	1.187	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.208	01:23
Breath Volume	1.207	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 09/25/2023 23:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: PIERCE,MARK
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 08/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623825
DRY GAS EXPIRY DATE: 12/25/2025
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 08/28/2023 02:28:59

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.083	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:24	No Sample Provided
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 2	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Control Test	0.083	02:28	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 09/25/2023 23:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,ANGELA,F
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: HOWARD,ROGER
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 08/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623825
DRY GAS EXPIRY DATE: 12/25/2025
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 08/29/2023 21:43:10

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:40	
Control Test	0.083	21:41	
Air Blank	0.000	21:41	
Subject Sample	# 1	21:41	Subject Test Refused
Breath Volume	0.000	21:41	
Air Blank	0.000	21:42	
Control Test	0.083	21:42	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 09/25/2023 23:23

Subject Test Electronic Data

DATE/TIME OF TEST: 08/06/2023 21:06:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: THOMAS,JON
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: KYNARD

DATE/TIME OF TEST: 08/17/2023 01:08:10

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.170	01:03
Breath Volume	2.613	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:05
Subject Sample # 2	0.180	01:06
Breath Volume	4.917	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 08/17/2023 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CRUZ,ADAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: MCELVAIN

DATE/TIME OF TEST: 08/04/2023 23:05:45

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.083	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.000	23:01
Breath Volume	2.710	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.000	23:04
Breath Volume	2.664	23:04
Air Blank	0.000	23:04
Control Test	0.082	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 08/13/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,Y
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: JASSO

DATE/TIME OF TEST: 08/16/2023 01:48:47

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.083	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.185	01:40
Breath Volume	1.269	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample # 2	0.159	01:43
Breath Volume	1.617	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 3	0.181	01:46
Breath Volume	1.531	01:46
Air Blank	0.000	01:47
Control Test	0.077	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 09/21/2023 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SADLER

DATE/TIME OF TEST: 08/17/2023 02:36:12

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.082	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.206	02:30
Breath Volume	2.101	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.211	02:34
Breath Volume	1.675	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 09/21/2023 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,JR,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SOLOMON,JR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 08/11/2023 22:54:01

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.116	22:49
Breath Volume	4.441	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.112	22:52
Breath Volume	4.054	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:54

Data Download Date/Time: 08/17/2023 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 08/17/2023 06:37:07

Diagnostics Check	OK	06:29
Air Blank	0.000	06:29
Control Test	0.081	06:30
Air Blank	0.000	06:30
Subject Sample # 1	0.173	06:31
Breath Volume	2.800	06:31
Air Blank	0.000	06:31
Air Blank	0.000	06:33
Subject Sample # 2	0.156	06:35
Breath Volume	1.792	06:35
Air Blank	0.000	06:36
Control Test	0.081	06:36
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

Data Download Date/Time: 08/17/2023 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: TATE,AYDEN
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: HEFFEL

DATE/TIME OF TEST: 08/18/2023 22:08:43

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.116	22:03
Breath Volume	1.222	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.120	22:07
Breath Volume	1.140	22:07
Air Blank	0.000	22:07
Control Test	0.081	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 09/27/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 2402328118
BREATH TEST OPERATOR: THOMPSON,JOSEPH,M	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: Washington County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PASLEY,JACOB	ARREST TIME: 00:12
ARRESTING OFFICER AGENCY: Washington County SO	GENDER: Male
LAST AGENCY INSPECTION: 08/17/2023	OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BOSTON	DATE/TIME OF TEST: 08/20/2023 01:01:38
-----------------------------	--

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:53	
Control Test	0.081	00:54	
Air Blank	0.000	00:54	
Subject Sample # 1	0.135	00:57	Volume Not Met
Breath Volume	0.031	00:57	
Air Blank	0.000	00:58	
Air Blank		01:00	RFI Detect
Air Blank	0.000	01:00	
Control Test	0.083	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 09/27/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 2402328118
BREATH TEST OPERATOR: THOMPSON,JOSEPH,M	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: Washington County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PASLEY,JACOB	ARREST TIME: 00:12
ARRESTING OFFICER AGENCY: Washington County SO	GENDER: Male
LAST AGENCY INSPECTION: 08/17/2023	OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BOSTON	DATE/TIME OF TEST: 08/20/2023 01:14:08
-----------------------------	--

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.080	01:05	
Air Blank	0.000	01:05	
Subject Sample	# 1 0.130	01:08	Volume Not Met
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample	# 2	01:12	Subject Test Refused
Breath Volume	0.761	01:12	
Air Blank	0.000	01:13	
Control Test	0.080	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 09/27/2023 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 2402328118
BREATH TEST OPERATOR: THOMPSON,JOSEPH,M	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: Washington County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: THOMPSON,JOSEPH	ARREST TIME: 01:18
ARRESTING OFFICER AGENCY: Washington County SO	GENDER: Female
LAST AGENCY INSPECTION: 08/17/2023	OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: CUSHMAN	DATE/TIME OF TEST: 08/21/2023 01:46:03
------------------------------	--

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Control Test	0.080	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 09/27/2023 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402328118
BREATH TEST OPERATOR: HOMAN,MICHAEL,J	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: HICKS,ROBERT	ARREST TIME: 19:06
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 08/17/2023	OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: BUTLER	DATE/TIME OF TEST: 08/26/2023 19:47:00
-----------------------------	--

Diagnostics Check	OK	19:46	
Air Blank		19:46	RFI Detect
Air Blank		19:47	RFI Detect

Data Download Date/Time: 09/27/2023 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 2402328118
BREATH TEST OPERATOR: HOMAN,MICHAEL,J	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: HICKS,ROBERT	ARREST TIME: 19:06
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 08/17/2023	OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: BUTLER	DATE/TIME OF TEST: 08/26/2023 19:55:02
-----------------------------	--

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:50	
Control Test	0.081	19:51	
Air Blank	0.000	19:51	
Subject Sample # 1	0.044	19:51	
Breath Volume	1.308	19:51	
Air Blank	0.000	19:52	
Air Blank		19:54	Ambient Fail
Air Blank	0.000	19:55	

Data Download Date/Time: 09/27/2023 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 2402328118
BREATH TEST OPERATOR: HOMAN,MICHAEL,J	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: HICKS,ROBERT	ARREST TIME: 19:06
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 08/17/2023	OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: BUTLER	DATE/TIME OF TEST: 08/26/2023	20:05:04
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.079	19:58
Air Blank	0.000	19:58
Subject Sample # 1	0.045	19:59
Breath Volume	1.636	19:59
Air Blank	0.000	20:00
Air Blank	0.000	20:02
Subject Sample # 2	0.041	20:03
Breath Volume	3.894	20:03
Air Blank	0.000	20:04
Control Test	0.079	20:04
Air Blank	0.000	20:04
Diagnostics Check	OK	20:05

Data Download Date/Time: 09/27/2023 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 07:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 08/27/2023 09:44:52

Diagnostics Check	OK	09:38
Air Blank	0.000	09:38
Control Test	0.080	09:39
Air Blank	0.000	09:39
Subject Sample # 1	0.064	09:40
Breath Volume	4.492	09:40
Air Blank	0.000	09:40
Air Blank	0.000	09:42
Subject Sample # 2	0.065	09:43
Breath Volume	4.761	09:43
Air Blank	0.000	09:43
Control Test	0.082	09:44
Air Blank	0.000	09:44
Diagnostics Check	OK	09:44

Data Download Date/Time: 09/27/2023 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 2402328118
BREATH TEST OPERATOR: PASLEY,JACOB,W	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: Washington County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PASLEY,JACOB	ARREST TIME: 06:30
ARRESTING OFFICER AGENCY: Washington County SO	GENDER: Male
LAST AGENCY INSPECTION: 08/17/2023	OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: LU	DATE/TIME OF TEST: 08/29/2023 19:49:42
-------------------------	--

Diagnostics Check	OK	19:43
Air Blank	0.000	19:44
Control Test	0.081	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.000	19:45
Breath Volume	1.621	19:45
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.000	19:48
Breath Volume	1.937	19:48
Air Blank	0.000	19:48
Control Test	0.081	19:49
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 09/27/2023 19:13

Subject Test Electronic Data

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS,BLAINE,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: AKIN,STERLING
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: MIESTCHOVICH

DATE/TIME OF TEST: 08/05/2023 16:18:48

Diagnostics Check	OK	16:11
Air Blank	0.000	16:11
Control Test	0.080	16:11
Air Blank	0.000	16:12
Subject Sample # 1	0.000	16:13
Breath Volume	1.535	16:13
Air Blank	0.000	16:13
Air Blank	0.000	16:15
Subject Sample # 2	0.000	16:17
Breath Volume	2.019	16:17
Air Blank	0.000	16:17
Control Test	0.081	16:18
Air Blank	0.000	16:18
Diagnostics Check	OK	16:18

Data Download Date/Time: 08/06/2023 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUBITSCHKE,LEONARD,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MCENEANEY,SEAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ANTOINE

DATE/TIME OF TEST: 08/01/2023 01:48:33

Diagnostics Check	OK	01:48	
Air Blank		01:48	RFI Detect
Air Blank		01:48	RFI Detect

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUBITSCHEK,LEONARD,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MCENEANEY,SEAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ANTOINE

DATE/TIME OF TEST: 08/01/2023 01:51:52

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
		01:51	RFI Detect
Air Blank		01:51	RFI Detect

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUBITSCHKE,LEONARD,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MCENEANEY,SEAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ANTOINE

DATE/TIME OF TEST: 08/01/2023 01:54:42

Diagnostics Check	OK	01:54	
Air Blank		01:54	RFI Detect
Air Blank	0.000	01:54	

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUBITSCHK,LEONARD,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MCENEANEY,SEAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ANTOINE

DATE/TIME OF TEST: 08/01/2023 02:03:50

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.332	01:58
Breath Volume	3.640	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.320	02:02
Breath Volume	4.648	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 08/05/2023 01:41:46

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.044	01:37
Breath Volume	2.558	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.041	01:40
Breath Volume	2.707	01:40
Air Blank	0.000	01:40
Control Test	0.081	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPUTO,TAYLOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: CAPUTO,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ALBERT

DATE/TIME OF TEST: 08/05/2023 01:49:25

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.081	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:48	Sequence Aborted
Breath Volume	0.000	01:48	
Air Blank	0.000	01:49	

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPUTO,TAYLOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: CAPUTO,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ALBERT

DATE/TIME OF TEST: 08/05/2023 02:16:24

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:12	
Control Test	0.082	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.081	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUBITSCHEK,LEONARD,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: KUBITSCHEK,LEONARD
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: VOSATKA

DATE/TIME OF TEST: 08/07/2023 18:06:39

Diagnostics Check	OK	17:57
Air Blank	0.000	17:57
Control Test	0.081	17:58
Air Blank	0.000	17:58
Subject Sample # 1	0.327	17:59
Breath Volume	1.851	17:59
Air Blank	0.000	17:59
Air Blank	0.000	18:01
Subject Sample # 2	0.279	18:01
Breath Volume	1.191	18:01
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 3	0.318	18:05
Breath Volume	2.710	18:05
Air Blank	0.000	18:05
Control Test	0.078	18:06
Air Blank	0.000	18:06
Diagnostics Check	OK	18:06

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/11/2023 22:17:48

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.081	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.194	22:12
Breath Volume	1.632	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.192	22:16
Breath Volume	1.269	22:16
Air Blank	0.000	22:16
Control Test	0.079	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: RUTHERFORD,ROBERT
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: SHINER

DATE/TIME OF TEST: 08/15/2023 04:31:22

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.176	04:25
Breath Volume	1.812	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.184	04:29
Breath Volume	1.816	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEGO,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: WADE,TYLER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 08/17/2023 21:43:25

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.208	21:38
Breath Volume	2.214	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.201	21:41
Breath Volume	2.175	21:41
Air Blank	0.000	21:42
Control Test	0.079	21:42
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAZIANO,STEVEN,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GRAZIANO,STEVEN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: RECINOS-CASTILLO

DATE/TIME OF TEST: 08/18/2023 17:12:51

Diagnostics Check	OK	17:05
Air Blank	0.000	17:06
Control Test	0.082	17:06
Air Blank	0.000	17:07
Subject Sample # 1	0.155	17:08
Breath Volume	1.199	17:08
Air Blank	0.000	17:09
Air Blank	0.000	17:10
Subject Sample # 2	0.171	17:11
Breath Volume	2.816	17:11
Air Blank	0.000	17:11
Control Test	0.080	17:12
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: HEYEN

DATE/TIME OF TEST: 08/19/2023 01:10:01

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.081	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.000	01:05
Breath Volume	1.679	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.000	01:08
Breath Volume	1.796	01:08
Air Blank	0.000	01:09
Control Test	0.082	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:10

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,KYLE,S
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: RODRIGUES,RUI
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:46

SUBJECT'S LAST NAME: VAN WIEREN

DATE/TIME OF TEST: 08/19/2023 16:14:12

Diagnostics Check	OK	16:07
Air Blank	0.000	16:07
Control Test	0.082	16:08
Air Blank	0.000	16:08
Subject Sample # 1	0.270	16:09
Breath Volume	2.902	16:09
Air Blank	0.000	16:10
Air Blank	0.000	16:11
Subject Sample # 2	0.278	16:12
Breath Volume	2.214	16:12
Air Blank	0.000	16:13
Control Test	0.081	16:13
Air Blank	0.000	16:14
Diagnostics Check	OK	16:14

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: POPOLIZIO

DATE/TIME OF TEST: 08/20/2023 03:41:59

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.150	03:36
Breath Volume	1.496	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.142	03:40
Breath Volume	1.257	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUBITSCHEK,LEONARD,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: KUBITSCHEK,LEONARD
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/20/2023 18:10:27

Diagnostics Check	OK	18:04
Air Blank	0.000	18:04
Control Test	0.081	18:05
Air Blank	0.000	18:05
Subject Sample # 1	0.172	18:06
Breath Volume	2.093	18:06
Air Blank	0.000	18:06
Air Blank	0.000	18:08
Subject Sample # 2	0.163	18:08
Breath Volume	3.996	18:08
Air Blank	0.000	18:09
Control Test	0.079	18:09
Air Blank	0.000	18:10
Diagnostics Check	OK	18:10

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA,ARMANDO,R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA,ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: COTHRAN

DATE/TIME OF TEST: 08/21/2023 22:33:34

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:22	
Control Test	0.081	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1 0.169	22:26	Volume Not Met
Breath Volume	0.015	22:26	
Air Blank	0.000	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 2 0.132	22:31	Volume Not Met
Breath Volume	0.578	22:31	
Air Blank	0.000	22:32	
Control Test	0.077	22:32	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA,ARMANDO,R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA,ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: COTHRAN

DATE/TIME OF TEST: 08/21/2023 22:44:14

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.078	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.151	22:39
Breath Volume	1.437	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:41
Subject Sample # 2	0.155	22:42
Breath Volume	1.484	22:42
Air Blank	0.000	22:43
Control Test	0.076	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILKEY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GILKEY,JOSHUA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 08/23/2023 19:44:24

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.082	19:38
Air Blank	0.000	19:38
Subject Sample # 1	0.147	19:39
Breath Volume	1.718	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.142	19:42
Breath Volume	1.222	19:42
Air Blank	0.000	19:43
Control Test	0.081	19:43
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 08/25/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA,ARMANDO,R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA,ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ANTONIO VENTURA

DATE/TIME OF TEST: 08/27/2023 00:36:58

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.082	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.090	00:32
Breath Volume	1.886	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.085	00:35
Breath Volume	1.792	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 09/27/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAKE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HESTER,JAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/27/2023 04:07:32

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.081	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1 0.141	04:02	
Breath Volume	2.761	04:02	
Air Blank	0.000	04:03	
Air Blank	0.000	04:05	
Subject Sample	# 2 0.141	04:05	
Breath Volume	2.617	04:05	
Air Blank	0.000	04:06	
Control Test	0.081	04:06	
Air Blank		04:07	RFI Detect
Air Blank	0.000	04:07	

Data Download Date/Time: 09/27/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,NATHAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SUMNER,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: LAGOS MONCADA

DATE/TIME OF TEST: 08/20/2023 01:07:14

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.081	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.141	01:02
Breath Volume	2.464	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:04
Subject Sample # 2	0.144	01:05
Breath Volume	1.757	01:05
Air Blank	0.000	01:06
Control Test	0.081	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 08/29/2023 01:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,NATHAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SUMNER,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: TYRE

DATE/TIME OF TEST: 08/24/2023 02:38:02

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.082	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.164	02:33
Breath Volume	1.480	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.166	02:36
Breath Volume	2.222	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:38

Data Download Date/Time: 08/29/2023 01:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YAKUBSIN,JAMES,R
BREATH TEST OPERATOR AGENCY: Dixie County SO
ARRESTING OFFICER: YAKUBSIN,JAMES
ARRESTING OFFICER AGENCY: Dixie County SO
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: APPLEBY

DATE/TIME OF TEST: 08/02/2023 00:08:49

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.078	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.120	00:04
Breath Volume	2.410	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.109	00:07
Breath Volume	1.183	00:07
Air Blank	0.000	00:07
Control Test	0.077	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 08/08/2023 11:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YAKUBSIN,JAMES,R
BREATH TEST OPERATOR AGENCY: Dixie County SO
ARRESTING OFFICER: YAKUBSIN,JAMES
ARRESTING OFFICER AGENCY: Dixie County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MISSILDINE

DATE/TIME OF TEST: 08/11/2023 03:12:55

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.079	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.078	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 09/29/2023 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAULKNER,JAMES,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FAULKNER,JAMES
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: JEAN MARY

DATE/TIME OF TEST: 08/02/2023 23:01:57

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.080	22:55
Air Blank	0.000	22:55
Subject Sample # 1	0.000	22:57
Breath Volume	1.546	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.000	23:00
Breath Volume	1.300	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 08/23/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CARRION

DATE/TIME OF TEST: 08/04/2023 23:33:56

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:29	
Control Test	0.081	23:29	
Air Blank	0.000	23:30	
Subject Sample	# 1	23:32	RFI Detect
Breath Volume	0.644	23:32	
Air Blank		23:33	RFI Detect
Air Blank	0.000	23:33	

Data Download Date/Time: 08/23/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CARRION

DATE/TIME OF TEST: 08/04/2023 23:39:20

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.078	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:37	RFI Detect
Breath Volume	0.160	23:37	
Air Blank	0.000	23:38	
Control Test	0.079	23:38	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 08/23/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZEZUSKI,JIMMY,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ZEZUSKI,JIMMY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: PACIOREK

DATE/TIME OF TEST: 08/04/2023 23:56:33

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.105	23:51
Breath Volume	2.714	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.102	23:54
Breath Volume	1.769	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 08/23/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROZZELL,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 08/02/2023 00:18:03

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:15	
Control Test	0.080	00:15	
Air Blank	0.000	00:16	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:17	
Control Test	0.080	00:17	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 08/02/2023 19:07:48

Diagnostics Check	OK	18:55	
Air Blank	0.000	18:56	
Control Test	0.079	18:56	
Air Blank	0.000	18:57	
Subject Sample	# 1 0.283	19:00	Volume Not Met
Breath Volume	0.105	19:00	
Air Blank	0.000	19:01	
Air Blank	0.000	19:02	
Subject Sample	# 2 0.298	19:06	Volume Not Met
Breath Volume	0.554	19:06	
Air Blank	0.000	19:06	
Control Test	0.077	19:07	
Air Blank	0.000	19:07	
Diagnostics Check	OK	19:07	

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 08/02/2023 19:17:41

Diagnostics Check	OK	19:11	
Air Blank	0.000	19:12	
Control Test	0.076	19:12	
Air Blank	0.000	19:12	
Subject Sample	# 1	19:15	Subject Test Refused
Breath Volume	0.085	19:15	
Air Blank	0.000	19:16	
Control Test	0.077	19:17	
Air Blank	0.000	19:17	
Diagnostics Check	OK	19:17	

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,ROGER,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: COOPMAN

DATE/TIME OF TEST: 08/04/2023 00:26:03

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.179	00:21
Breath Volume	1.308	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.176	00:24
Breath Volume	1.691	00:24
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:26

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 08/05/2023 01:43:09

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.212	01:38
Breath Volume	1.453	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.208	01:41
Breath Volume	1.402	01:41
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POULIN,D,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: POULIN,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 08/06/2023 17:32:26

Diagnostics Check	OK	17:21	
Air Blank	0.000	17:21	
Control Test	0.079	17:21	
Air Blank	0.000	17:22	
Subject Sample	# 1	17:25	No Sample Provided
Breath Volume	0.000	17:25	
Air Blank	0.000	17:25	
Air Blank	0.000	17:27	
Subject Sample	# 2	17:31	No Sample Provided
Breath Volume	0.000	17:31	
Air Blank	0.000	17:31	
Control Test	0.079	17:31	
Air Blank	0.000	17:32	
Diagnostics Check	OK	17:32	

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: DOWLING

DATE/TIME OF TEST: 08/06/2023 20:38:30

Diagnostics Check	OK	20:32
Air Blank	0.000	20:32
Control Test	0.079	20:32
Air Blank	0.000	20:33
Subject Sample # 1	0.174	20:34
Breath Volume	2.074	20:34
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.177	20:36
Breath Volume	2.503	20:36
Air Blank	0.000	20:37
Control Test	0.078	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 08/13/2023 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MARAGH

DATE/TIME OF TEST: 08/07/2023 03:43:31

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.079	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.079	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 08/13/2023 19:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 08/07/2023 21:36:40

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:27	
Control Test	0.080	21:28	
Air Blank	0.000	21:28	
Subject Sample	# 1 0.277	21:31	Volume Not Met
Breath Volume	0.019	21:31	
Air Blank	0.000	21:32	
Air Blank	0.000	21:34	
Subject Sample	# 2	21:35	Subject Test Refused
Breath Volume	0.062	21:35	
Air Blank	0.000	21:35	
Control Test	0.076	21:36	
Air Blank	0.000	21:36	
Diagnostics Check	OK	21:36	

Data Download Date/Time: 08/13/2023 19:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: HETHERINGTON

DATE/TIME OF TEST: 08/07/2023 23:48:59

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.173	23:44
Breath Volume	2.492	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.168	23:47
Breath Volume	1.320	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 08/13/2023 19:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,ROGER,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: HERNANDEZ ACEVEDO

DATE/TIME OF TEST: 08/12/2023 22:24:37

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.079	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.000	22:20
Breath Volume	1.863	22:20
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.000	22:23
Breath Volume	2.785	22:23
Air Blank	0.000	22:23
Control Test	0.079	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 08/13/2023 19:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VESPER,MITCHELL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: CHUMLEY

DATE/TIME OF TEST: 08/12/2023 23:26:27

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:15	
Control Test	0.079	23:15	
Air Blank	0.000	23:16	
Subject Sample	# 1 0.175	23:19	Volume Not Met
Breath Volume	0.382	23:19	
Air Blank	0.000	23:19	
Air Blank	0.000	23:21	
Subject Sample	# 2 0.182	23:24	Volume Not Met
Breath Volume	0.437	23:24	
Air Blank	0.000	23:25	
Control Test	0.077	23:25	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 08/13/2023 19:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BERLYNE,R,D

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: BERLYNE,R

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 07/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30822080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 08/13/2023 01:42:00

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.078	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:41	
Control Test	0.078	01:41	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 08/13/2023 19:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VESPER,MITCHELL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 08/13/2023 23:07:06

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.080	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.234	23:02
Breath Volume	1.460	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:04
Subject Sample # 2	0.234	23:05
Breath Volume	1.960	23:05
Air Blank	0.000	23:06
Control Test	0.078	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: GIDDENS

DATE/TIME OF TEST: 08/14/2023 20:57:20

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.080	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.000	20:53
Breath Volume	4.363	20:53
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.000	20:55
Breath Volume	3.363	20:55
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 08/14/2023 22:39:30

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.079	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Control Test	0.079	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 08/15/2023 01:28:31

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.219	01:23
Breath Volume	2.621	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:25
Subject Sample # 2	0.209	01:26
Breath Volume	2.273	01:26
Air Blank	0.000	01:27
Control Test	0.077	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,ROGER,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ELDER

DATE/TIME OF TEST: 08/18/2023 01:52:52

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.000	01:48
Breath Volume	1.722	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.000	01:51
Breath Volume	1.878	01:51
Air Blank	0.000	01:51
Control Test	0.080	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MARCUS,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: DRAWE

DATE/TIME OF TEST: 08/20/2023 02:04:45

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.134	02:00
Breath Volume	1.835	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.127	02:03
Breath Volume	2.230	02:03
Air Blank	0.000	02:03
Control Test	0.079	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MARCUS,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: RIVERA TORRES

DATE/TIME OF TEST: 08/20/2023 02:38:23

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.107	02:33
Breath Volume	2.417	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.110	02:36
Breath Volume	3.277	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MARCUS,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEVER,I
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HERNANDEZ RUBI

DATE/TIME OF TEST: 08/20/2023 03:08:43

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.079	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.079	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MARCUS,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MCNEELEY,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CAMINO

DATE/TIME OF TEST: 08/20/2023 03:41:57

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.152	03:37
Breath Volume	1.761	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.164	03:40
Breath Volume	2.707	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: REID,HUNTER
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: BRUZDA

DATE/TIME OF TEST: 08/21/2023 19:00:12

Diagnostics Check	OK	18:50
Air Blank	0.000	18:50
Control Test	0.079	18:50
Air Blank	0.000	18:51
Subject Sample # 1	0.309	18:54
Breath Volume	1.156	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.324	18:58
Breath Volume	1.628	18:58
Air Blank	0.000	18:59
Control Test	0.078	18:59
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: REID,HUNTER
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: PAWLEY

DATE/TIME OF TEST: 08/22/2023 19:59:57

Diagnostics Check	OK	19:48	
Air Blank	0.000	19:49	
Control Test	0.080	19:49	
Air Blank	0.000	19:49	
Subject Sample	# 1	19:52	No Sample Provided
Breath Volume	0.000	19:52	
Air Blank	0.000	19:53	
Air Blank	0.000	19:55	
Subject Sample	# 2	19:58	No Sample Provided
Breath Volume	0.000	19:58	
Air Blank	0.000	19:59	
Control Test	0.080	19:59	
Air Blank	0.000	19:59	
Diagnostics Check	OK	19:59	

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VESPER,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/22/2023 23:19:17

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.080	23:16	
Air Blank	0.000	23:17	
Subject Sample	# 1	23:17	Subject Test Refused
Breath Volume	0.000	23:17	
Air Blank	0.000	23:18	
Control Test	0.079	23:18	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PERISEE

DATE/TIME OF TEST: 08/24/2023 03:43:51

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.079	03:36	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:39	No Sample Provided
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 2	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.079	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,ROGER,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: MURWIN

DATE/TIME OF TEST: 08/25/2023 23:50:45

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.000	23:46
Breath Volume	2.441	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.000	23:49
Breath Volume	1.570	23:49
Air Blank	0.000	23:49
Control Test	0.080	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,ROGER,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: KOSTER

DATE/TIME OF TEST: 08/26/2023 23:50:18

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.040	23:45
Breath Volume	1.328	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.036	23:48
Breath Volume	1.234	23:48
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SIUZDAK,C
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: GERNHART

DATE/TIME OF TEST: 08/28/2023 00:35:47

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:33	
Control Test	0.080	00:33	
Air Blank	0.000	00:34	
Subject Sample	# 1	00:34	Subject Test Refused
Breath Volume	0.000	00:34	
Air Blank	0.000	00:34	
Control Test	0.079	00:35	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 09/29/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL, TYLER, C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TAPONES, C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 08/26/2023 00:37:49

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.080	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1 0.159	00:30	Volume Not Met
Breath Volume	0.191	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 2 0.150	00:36	Volume Not Met
Breath Volume	0.679	00:36	
Air Blank	0.000	00:36	
Control Test	0.078	00:37	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 09/29/2023 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL, TYLER, C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:48

SUBJECT'S LAST NAME: SNEPERGER

DATE/TIME OF TEST: 08/26/2023 00:55:18

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.078	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.197	00:50
Breath Volume	1.835	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.188	00:53
Breath Volume	1.160	00:53
Air Blank	0.000	00:54
Control Test	0.078	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 09/29/2023 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRINKLE,BRITTANY,N
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LANE,K
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: MERRITT

DATE/TIME OF TEST: 08/26/2023 07:05:18

Diagnostics Check	OK	06:59	
Air Blank	0.000	06:59	
Control Test	0.080	06:59	
Air Blank	0.000	07:00	
Subject Sample	# 1 0.000	07:01	
Breath Volume	1.789	07:01	
Air Blank	0.000	07:01	
Air Blank	0.000	07:03	
Subject Sample	# 2 0.000	07:04	
Breath Volume	1.863	07:04	
Air Blank	0.000	07:05	
		07:05	RFI Detect
Air Blank		07:05	RFI Detect

Data Download Date/Time: 09/29/2023 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL, TYLER, C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING, J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 08/27/2023 02:16:32

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.129	02:11
Breath Volume	2.558	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.126	02:14
Breath Volume	2.324	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 09/29/2023 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: VARVARO

DATE/TIME OF TEST: 08/27/2023 21:48:21

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.078	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.000	21:43
Breath Volume	2.437	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.000	21:46
Breath Volume	2.332	21:46
Air Blank	0.000	21:47
Control Test	0.078	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 09/29/2023 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 08/28/2023 00:50:43

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:48	
Control Test	0.080	00:48	
Air Blank	0.000	00:49	
Subject Sample	# 1	00:49	Subject Test Refused
Breath Volume	0.000	00:49	
Air Blank	0.000	00:49	
Control Test	0.078	00:50	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 09/29/2023 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN, DEMERIS, P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SUKHDEO, A
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ROMERO MENDEZ

DATE/TIME OF TEST: 08/29/2023 03:37:48

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.125	03:32
Breath Volume	2.468	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.120	03:36
Breath Volume	2.101	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 09/29/2023 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NORDLE,L
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:18

SUBJECT'S LAST NAME: WHITEHEAD

DATE/TIME OF TEST: 08/29/2023 09:42:16

Diagnostics Check	OK	09:39	
Air Blank	0.000	09:39	
Control Test	0.080	09:40	
Air Blank	0.000	09:40	
Subject Sample	# 1	09:40	Subject Test Refused
Breath Volume	0.000	09:40	
Air Blank	0.000	09:41	
Control Test	0.078	09:41	
Air Blank	0.000	09:42	
Diagnostics Check	OK	09:42	

Data Download Date/Time: 09/29/2023 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:17

SUBJECT'S LAST NAME: CRISPIN BLAS

DATE/TIME OF TEST: 08/29/2023 13:43:26

Diagnostics Check	OK	13:37
Air Blank	0.000	13:37
Control Test	0.079	13:38
Air Blank	0.000	13:38
Subject Sample # 1	0.269	13:38
Breath Volume	1.835	13:38
Air Blank	0.000	13:39
Air Blank	0.000	13:41
Subject Sample # 2	0.284	13:41
Breath Volume	1.929	13:41
Air Blank	0.000	13:42
Control Test	0.077	13:42
Air Blank	0.000	13:43
Diagnostics Check	OK	13:43

Data Download Date/Time: 09/29/2023 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ROSARISO,F
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 08/01/2023 22:33:40

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:30	
Control Test	0.079	22:31	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:32	
Control Test	0.080	22:33	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 08/25/2023 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRINKLE,BRITTANY,N
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HANSEN,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: DIGIACOMO

DATE/TIME OF TEST: 08/02/2023 16:37:42

Diagnostics Check	OK	16:31
Air Blank	0.000	16:31
Control Test	0.080	16:31
Air Blank	0.000	16:32
Subject Sample # 1	0.200	16:33
Breath Volume	1.644	16:33
Air Blank	0.000	16:33
Air Blank	0.000	16:35
Subject Sample # 2	0.199	16:36
Breath Volume	1.656	16:36
Air Blank	0.000	16:36
Control Test	0.080	16:37
Air Blank	0.000	16:37
Diagnostics Check	OK	16:37

Data Download Date/Time: 08/25/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDZINSKI,BRANDON,E
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARRIS,P
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: LAMOREAUX

DATE/TIME OF TEST: 08/03/2023 03:12:02

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.287	03:07
Breath Volume	1.988	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample # 2	0.282	03:10
Breath Volume	2.390	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:12

Data Download Date/Time: 08/25/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,ERNEST,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LONEY,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 08/07/2023 21:56:41

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:54	
Control Test	0.079	21:54	
Air Blank	0.000	21:55	
Subject Sample	# 1	21:55	Subject Test Refused
Breath Volume	0.000	21:55	
Air Blank	0.000	21:55	
Control Test	0.079	21:56	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time: 08/25/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: TEMAJGABRIEL

DATE/TIME OF TEST: 08/08/2023 06:23:15

Diagnostics Check	OK	06:16
Air Blank	0.000	06:16
Control Test	0.079	06:16
Air Blank	0.000	06:17
Subject Sample # 1	0.069	06:18
Breath Volume	2.378	06:18
Air Blank	0.000	06:18
Air Blank	0.000	06:20
Subject Sample # 2	0.071	06:21
Breath Volume	2.863	06:21
Air Blank	0.000	06:22
Control Test	0.079	06:22
Air Blank	0.000	06:23
Diagnostics Check	OK	06:23

Data Download Date/Time: 08/25/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HANNON,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:44

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 08/09/2023 13:11:49

Diagnostics Check	OK	13:04
Air Blank	0.000	13:05
Control Test	0.079	13:05
Air Blank	0.000	13:06
Subject Sample # 1	0.079	13:07
Breath Volume	1.906	13:07
Air Blank	0.000	13:08
Air Blank	0.000	13:09
Subject Sample # 2	0.085	13:10
Breath Volume	2.281	13:10
Air Blank	0.000	13:10
Control Test	0.078	13:11
Air Blank	0.000	13:11
Diagnostics Check	OK	13:11

Data Download Date/Time: 08/25/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,ERNEST,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING,JONATHAN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 08/13/2023 00:52:05

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.078	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.000	00:47
Breath Volume	1.378	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:50
Breath Volume	1.324	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 08/25/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TAPANES,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: MENO

DATE/TIME OF TEST: 08/13/2023 02:23:43

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.140	02:18
Breath Volume	2.035	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.139	02:22
Breath Volume	1.621	02:22
Air Blank	0.000	02:22
Control Test	0.077	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 08/25/2023 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REED,ERNEST,L

BREATH TEST OPERATOR AGENCY: Sumter County SO

ARRESTING OFFICER: GOEHRING,JONATHAN

ARRESTING OFFICER AGENCY: Sumter County SO

LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 221303

DRY GAS EXPIRY DATE: 01/05/2025

VIOLATION CODE: DUI

ARREST TIME: 23:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: CORBIN

DATE/TIME OF TEST: 08/18/2023 00:36:32

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.079	00:34	
Air Blank	0.000	00:34	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 08/25/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN, DEMERIS, P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: KJEKSTAD, ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: HARRELSON

DATE/TIME OF TEST: 08/19/2023 05:37:10

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.080	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.064	05:32
Breath Volume	1.332	05:32
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.064	05:35
Breath Volume	1.234	05:35
Air Blank	0.000	05:36
Control Test	0.079	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 08/25/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN,DEMERIS,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ROSARIO,FOUSTO
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 08/19/2023 20:47:21

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.079	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.275	20:42
Breath Volume	1.921	20:42
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.280	20:45
Breath Volume	1.886	20:45
Air Blank	0.000	20:46
Control Test	0.079	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 08/25/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GUINANE,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/20/2023 19:02:51

Diagnostics Check	OK	18:51	
Air Blank	0.000	18:51	
Control Test	0.079	18:52	
Air Blank	0.000	18:52	
Subject Sample	# 1 0.000	18:55	Volume Not Met
Breath Volume	0.792	18:55	
Air Blank	0.000	18:56	
Air Blank	0.000	18:58	
Subject Sample	# 2 0.000	19:01	Volume Not Met
Breath Volume	0.015	19:01	
Air Blank	0.000	19:01	
Control Test	0.078	19:02	
Air Blank	0.000	19:02	
Diagnostics Check	OK	19:02	

Data Download Date/Time: 08/25/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN, DEMERIS, P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SURIEL, ENMANUEL
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/20/2023 23:20:45

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.222	23:16
Breath Volume	3.347	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.225	23:19
Breath Volume	3.359	23:19
Air Blank	0.000	23:19
Control Test	0.078	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 08/25/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:18

SUBJECT'S LAST NAME: PARROTT

DATE/TIME OF TEST: 08/22/2023 06:45:29

Diagnostics Check	OK	06:38
Air Blank	0.000	06:38
Control Test	0.079	06:39
Air Blank	0.000	06:39
Subject Sample # 1	0.000	06:40
Breath Volume	4.371	06:40
Air Blank	0.000	06:41
Air Blank	0.000	06:43
Subject Sample # 2	0.000	06:44
Breath Volume	4.000	06:44
Air Blank	0.000	06:44
Control Test	0.079	06:44
Air Blank	0.000	06:45
Diagnostics Check	OK	06:45

Data Download Date/Time: 08/25/2023 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: KROSS

DATE/TIME OF TEST: 08/23/2023 00:32:57

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.080	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Control Test	0.080	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 08/25/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN, DEMERIS, P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: PETERSON, WALTER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 08/23/2023 23:38:22

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.161	23:32
Breath Volume	1.625	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.158	23:36
Breath Volume	1.878	23:36
Air Blank	0.000	23:37
Control Test	0.078	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 08/25/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POULIN,DANIELLE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: AMATO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MCWHORTER

DATE/TIME OF TEST: 08/02/2023 04:16:25

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.142	04:11
Breath Volume	2.035	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.138	04:14
Breath Volume	2.550	04:14
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 09/05/2023 09:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVIS,JIM
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: RAMDOOLAR

DATE/TIME OF TEST: 08/06/2023 03:23:36

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.081	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:19	No Sample Provided
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample	# 2	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.082	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 09/05/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ,G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: SONSTEGARD

DATE/TIME OF TEST: 08/08/2023 19:06:34

Diagnostics Check	OK	18:58
Air Blank	0.000	18:58
Control Test	0.081	18:58
Air Blank	0.000	18:59
Subject Sample # 1	0.196	19:00
Breath Volume	1.746	19:00
Air Blank	0.000	19:01
Air Blank	0.000	19:03
Subject Sample # 2	0.193	19:04
Breath Volume	2.031	19:04
Air Blank	0.000	19:05
Control Test	0.079	19:05
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 09/05/2023 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: KNAPP

DATE/TIME OF TEST: 08/09/2023 18:20:36

Diagnostics Check	OK	18:13
Air Blank	0.000	18:13
Control Test	0.081	18:13
Air Blank	0.000	18:14
Subject Sample # 1	0.120	18:15
Breath Volume	2.046	18:15
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.115	18:18
Breath Volume	1.976	18:18
Air Blank	0.000	18:19
Control Test	0.080	18:19
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

Data Download Date/Time: 09/05/2023 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ADOLF,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: DUTTON

DATE/TIME OF TEST: 08/10/2023 07:36:16

Diagnostics Check	OK	07:21	
Air Blank	0.000	07:22	
Control Test	0.081	07:22	
Air Blank	0.000	07:23	
Subject Sample	# 1 0.287	07:26	Volume Not Met
Breath Volume	0.382	07:26	
Air Blank	0.000	07:27	
Air Blank	0.000	07:28	
Subject Sample	# 2 0.308	07:31	
Breath Volume	1.691	07:31	
Air Blank	0.000	07:32	
Air Blank	0.000	07:34	
Subject Sample	# 3 0.327	07:34	
Breath Volume	1.726	07:34	
Air Blank	0.000	07:35	
Control Test	0.077	07:35	
Air Blank	0.000	07:36	
Diagnostics Check	OK	07:36	

Data Download Date/Time: 09/05/2023 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: HOFF

DATE/TIME OF TEST: 08/10/2023 15:27:32

Diagnostics Check	OK	15:15
Air Blank	0.000	15:15
Control Test	0.080	15:16
Air Blank	0.000	15:16
Subject Sample # 1	0.287	15:18
Breath Volume	1.519	15:18
Air Blank	0.000	15:19
Air Blank	0.000	15:20
Subject Sample # 2	0.246	15:22
Breath Volume	3.031	15:22
Air Blank	0.000	15:23
Air Blank	0.000	15:25
Subject Sample # 3	0.254	15:25
Breath Volume	1.394	15:25
Air Blank	0.000	15:26
Control Test	0.078	15:26
Air Blank	0.000	15:27
Diagnostics Check	OK	15:27

Data Download Date/Time: 09/05/2023 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 08/12/2023 19:04:37

Diagnostics Check	OK	19:01	
Air Blank	0.000	19:02	
Control Test	0.081	19:02	
Air Blank	0.000	19:03	
Subject Sample	# 1	19:03	Subject Test Refused
Breath Volume	0.000	19:03	
Air Blank	0.000	19:03	
Control Test	0.081	19:04	
Air Blank	0.000	19:04	
Diagnostics Check	OK	19:04	

Data Download Date/Time: 09/05/2023 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SAAYMAN,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 08/13/2023 00:33:00

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.079	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.175	00:28
Breath Volume	1.257	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.182	00:31
Breath Volume	1.210	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:33

Data Download Date/Time: 09/05/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 08/18/2023 15:32:45

Diagnostics Check	OK	15:26
Air Blank	0.000	15:26
Control Test	0.082	15:27
Air Blank	0.000	15:27
Subject Sample # 1	0.179	15:28
Breath Volume	1.628	15:28
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.176	15:31
Breath Volume	1.980	15:31
Air Blank	0.000	15:31
Control Test	0.081	15:32
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

Data Download Date/Time: 09/05/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ,GEOFFREY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: HECKMAN

DATE/TIME OF TEST: 08/18/2023 22:11:06

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.162	22:06
Breath Volume	1.675	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:08
Subject Sample # 2	0.170	22:09
Breath Volume	1.226	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 09/05/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BEEMER

DATE/TIME OF TEST: 08/20/2023 01:42:02

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.226	01:37
Breath Volume	2.062	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.213	01:40
Breath Volume	1.316	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:42

Data Download Date/Time: 09/05/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOVANOVA, MICHAEL, N
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: JOVANOVA, MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CARRERAS

DATE/TIME OF TEST: 08/20/2023 03:16:16

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.013	03:11
Breath Volume	1.539	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.012	03:14
Breath Volume	3.070	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 09/05/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: TYZBIR

DATE/TIME OF TEST: 08/31/2023 21:26:41

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.081	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.213	21:22
Breath Volume	1.640	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.214	21:25
Breath Volume	1.199	21:25
Air Blank	0.000	21:25
Control Test	0.080	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 09/05/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILLIAMS,AUSTIN
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 08/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: CRISP

DATE/TIME OF TEST: 08/31/2023 12:17:06

Diagnostics Check	OK	12:10
Air Blank	0.000	12:11
Control Test	0.079	12:11
Air Blank	0.000	12:12
Subject Sample # 1	0.335	12:12
Breath Volume	1.210	12:12
Air Blank	0.000	12:13
Air Blank	0.000	12:15
Subject Sample # 2	0.319	12:15
Breath Volume	1.796	12:15
Air Blank	0.000	12:16
Control Test	0.080	12:16
Air Blank	0.000	12:16
Diagnostics Check	OK	12:17

Data Download Date/Time: 09/22/2023 06:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BANNICK,COURTNEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 08/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/31/2023 14:49:50

Diagnostics Check	OK	14:43
Air Blank	0.000	14:43
Control Test	0.079	14:44
Air Blank	0.000	14:44
Subject Sample # 1	0.000	14:45
Breath Volume	1.703	14:45
Air Blank	0.000	14:45
Air Blank	0.000	14:47
Subject Sample # 2	0.000	14:48
Breath Volume	1.316	14:48
Air Blank	0.000	14:48
Control Test	0.080	14:49
Air Blank	0.000	14:49
Diagnostics Check	OK	14:49

Data Download Date/Time: 09/22/2023 06:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,STEPHEN,V
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: TORRES,DANIEL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: MANDRY

DATE/TIME OF TEST: 08/02/2023 00:49:53

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.079	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORR,ERICA,F
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: RICHETTI,KALEIGH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 08/02/2023 03:02:07

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.153	02:57
Breath Volume	3.203	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.147	03:00
Breath Volume	2.214	03:00
Air Blank	0.000	03:01
Control Test	0.077	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: SYBOODEEN

DATE/TIME OF TEST: 08/04/2023 06:23:24

Diagnostics Check	OK	06:22	
Air Blank	0.000	06:22	
Control Test	0.000	06:22	RFI Detect
Air Blank	0.000	06:23	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: SYBOODEEN

DATE/TIME OF TEST: 08/04/2023 06:33:09

Diagnostics Check	OK	06:30	
Air Blank	0.000	06:30	
Control Test	0.080	06:31	
Air Blank	0.000	06:31	
Subject Sample	# 1	06:31	Subject Test Refused
Breath Volume	0.000	06:31	
Air Blank	0.000	06:32	
Control Test	0.079	06:32	
Air Blank	0.000	06:33	
Diagnostics Check	OK	06:33	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,C,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ,C
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: HEARD

DATE/TIME OF TEST: 08/06/2023 04:39:24

Diagnostics Check	OK	04:38	
Air Blank		04:39	RFI Detect
Air Blank		04:39	RFI Detect

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,C,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ,C
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: HEARD

DATE/TIME OF TEST: 08/06/2023 04:53:32

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.250	04:48
Breath Volume	1.347	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:50
Subject Sample # 2	0.245	04:51
Breath Volume	2.199	04:51
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREANO, JACK,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: DIAZ, WILFREDO
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:01

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 08/06/2023 09:25:06

Diagnostics Check	OK	09:22	
Air Blank	0.000	09:22	
Control Test	0.079	09:22	
Air Blank	0.000	09:23	
Subject Sample	# 1	09:23	Subject Test Refused
Breath Volume	0.000	09:23	
Air Blank	0.000	09:24	
Control Test	0.079	09:24	
Air Blank	0.000	09:25	
Diagnostics Check	OK	09:25	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WORFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: LAFAVOR

DATE/TIME OF TEST: 08/08/2023 12:15:17

Diagnostics Check	OK	12:05
Air Blank	0.000	12:06
Control Test	0.079	12:06
Air Blank	0.000	12:06
Subject Sample # 1	0.282	12:07
Breath Volume	1.679	12:07
Air Blank	0.000	12:08
Air Blank	0.000	12:09
Subject Sample # 2	0.259	12:10
Breath Volume	1.148	12:10
Air Blank	0.000	12:11
Air Blank	0.000	12:13
Subject Sample # 3	0.258	12:13
Breath Volume	1.433	12:13
Air Blank	0.000	12:14
Control Test	0.079	12:14
Air Blank	0.000	12:15
Diagnostics Check	OK	12:15

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY, WAYNE, G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PERRY, WAYNE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 08/11/2023 18:34:57

Diagnostics Check	OK	18:21	
Air Blank	0.000	18:21	
Control Test	0.080	18:22	
Air Blank	0.000	18:22	
Subject Sample	# 1 0.225	18:25	Volume Not Met
Breath Volume	0.902	18:25	
Air Blank	0.000	18:26	
Air Blank	0.000	18:28	
Subject Sample	# 2 0.257	18:29	No .020 agreement
Breath Volume	1.472	18:29	
Air Blank	0.000	18:29	
Air Blank	0.000	18:31	
Subject Sample	# 3 0.286	18:33	No .020 agreement
Breath Volume	2.265	18:33	
Air Blank	0.000	18:34	
Control Test	0.078	18:34	
Air Blank	0.000	18:34	
Diagnostics Check	OK	18:34	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONDAHL,CARL,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/12/2023 01:45:39

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:40	
Control Test	0.078	01:40	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Control Test	0.079	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILDEMAN-STUART,N
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: BRACCO

DATE/TIME OF TEST: 08/13/2023 02:31:44

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.106	02:27
Breath Volume	1.929	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.114	02:30
Breath Volume	2.886	02:30
Air Blank	0.000	02:30
Control Test	0.078	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UMANA,RAUL,J
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: UMANA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:34

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/13/2023 10:01:33

Diagnostics Check	OK	09:55
Air Blank	0.000	09:55
Control Test	0.080	09:55
Air Blank	0.000	09:56
Subject Sample # 1	0.178	09:57
Breath Volume	1.121	09:57
Air Blank	0.000	09:57
Air Blank	0.000	09:59
Subject Sample # 2	0.169	09:59
Breath Volume	1.847	09:59
Air Blank	0.000	10:00
Control Test	0.080	10:00
Air Blank	0.000	10:01
Diagnostics Check	OK	10:01

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERDT, ANDRU, J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: CINTRON-CARABALLO, JEAN
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: BERG

DATE/TIME OF TEST: 08/14/2023 01:09:46

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.080	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.164	01:05
Breath Volume	1.539	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.163	01:08
Breath Volume	1.398	01:08
Air Blank	0.000	01:08
Control Test	0.078	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,L,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: RICHETTI,K
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 08/14/2023 22:13:34

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.080	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.126	22:08
Breath Volume	1.648	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.131	22:11
Breath Volume	1.546	22:11
Air Blank	0.000	22:12
Control Test	0.078	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COODY,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: COODY,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: ESPARZA ARMENTA

DATE/TIME OF TEST: 08/19/2023 05:54:56

Diagnostics Check	OK	05:53	
Air Blank	0.000	05:53	
Control Test	0.080	05:53	
Air Blank	0.000	05:54	
		05:54	RFI Detect
Air Blank	0.000	05:54	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COODY,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: COODY,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: ESPARZA ARMENTA

DATE/TIME OF TEST: 08/19/2023 06:03:15

Diagnostics Check	OK	05:56
Air Blank	0.000	05:57
Control Test	0.080	05:57
Air Blank	0.000	05:58
Subject Sample # 1	0.140	05:58
Breath Volume	1.308	05:58
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.136	06:01
Breath Volume	1.378	06:01
Air Blank	0.000	06:02
Control Test	0.079	06:02
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COODY,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: COODY,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: LASTRES

DATE/TIME OF TEST: 08/21/2023 02:05:34

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.104	01:59
Breath Volume	1.625	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.112	02:03
Breath Volume	1.355	02:03
Air Blank	0.000	02:04
Control Test	0.077	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUBICEK,JORDAN,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ROUBICEK,JORDAN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: SCHOLTEN

DATE/TIME OF TEST: 08/21/2023 22:32:35

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:30	
Control Test	0.079	22:30	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Control Test	0.079	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,ROBERT,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: HART,ROBERT
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 08/22/2023 23:20:44

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:16	
Control Test	0.080	23:16	
Air Blank	0.000	23:17	
Subject Sample	# 1	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.080	23:20	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: SCHNEIDER,CODY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 08/26/2023 03:23:48

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.167	03:18
Breath Volume	3.593	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.160	03:22
Breath Volume	3.945	03:22
Air Blank	0.000	03:22
Control Test	0.079	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CLARK,JACOB
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 08/26/2023 05:11:18

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.129	05:06
Breath Volume	3.332	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.127	05:09
Breath Volume	2.988	05:09
Air Blank	0.000	05:10
Control Test	0.079	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,BROOKE,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: GOTT,CHRISTOPHER
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: WYTRWAL

DATE/TIME OF TEST: 08/27/2023 00:32:57

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.080	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Control Test	0.080	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,BROOKE,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: MASON,BROOKE
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CORTEZ DE LA ROSA

DATE/TIME OF TEST: 08/27/2023 02:38:54

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:37	
Control Test	0.079	02:38	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 08/30/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILDEMAN-STUART,NATASH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: MONZALVO HERNANDEZ

DATE/TIME OF TEST: 08/27/2023 22:21:01

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.204	22:16
Breath Volume	3.371	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:18
Subject Sample # 2	0.191	22:19
Breath Volume	2.261	22:19
Air Blank	0.000	22:20
Control Test	0.075	22:20
Air Blank	0.000	22:20
Diagnostics Check	OK	22:21

Data Download Date/Time: 08/30/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLO,DEREK,J
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: POLO,DEREK
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 08/28/2023 22:27:55

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.079	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.266	22:23
Breath Volume	3.808	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample # 2	0.261	22:26
Breath Volume	3.757	22:26
Air Blank	0.000	22:26
Control Test	0.077	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 08/30/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,L,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: TORRES,V
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 08/29/2023 21:19:57

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.080	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.052	21:15
Breath Volume	2.023	21:15
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.052	21:18
Breath Volume	1.589	21:18
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 08/30/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: RENGIFO REBOLLEDO

DATE/TIME OF TEST: 08/04/2023 01:19:06

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.079	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.220	01:10
Breath Volume		1.207	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:12
Subject Sample	# 2	0.242	01:14
Breath Volume		1.296	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 3	0.216	01:17
Breath Volume		1.160	01:17
Air Blank		0.000	01:18
Control Test		0.078	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 08/23/2023 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COFFEY-SHADE, W, T

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: HILLIKER, C

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 08/05/2023 01:20:46

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.079	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.123	01:16
Breath Volume	3.917	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:18
Subject Sample # 2	0.117	01:19
Breath Volume	2.242	01:19
Air Blank	0.000	01:19
Control Test	0.079	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 08/23/2023 14:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCLAIN,DAVID,R

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: MCCLAIN,DAVID

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:17

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MELENDEZ ZAYAS

DATE/TIME OF TEST: 08/12/2023 01:41:27

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.080	01:39	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 08/23/2023 14:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COFFEY-SHADE, W, T

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: POST, H

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: HERNANDEZ-MARTINEZ

DATE/TIME OF TEST: 08/19/2023 04:32:45

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:30	
Control Test	0.080	04:30	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:31	
Control Test	0.080	04:32	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 08/23/2023 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COFFEY-SHADE, W, T

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: POST, H

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29821080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 08/20/2023 02:26:12

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.157	02:21
Breath Volume	3.292	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.158	02:24
Breath Volume	3.246	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 08/23/2023 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COFFEY-SHADE, W, T

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: POST, H

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A5

DRY GAS EXPIRY DATE: 06/05/2025

VIOLATION CODE: DUI

ARREST TIME: 02:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: ROBLERO-HERNANDEZ

DATE/TIME OF TEST: 08/24/2023 03:17:18

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.214	03:10
Breath Volume	1.421	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.243	03:13
Breath Volume	2.792	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 3	0.204	03:15
Breath Volume	1.632	03:15
Air Blank	0.000	03:16
Control Test	0.076	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 09/27/2023 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/04/2023 04:57:25

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:54	
Control Test	0.084	04:54	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:55	Subject Test Refused
Breath Volume	0.000	04:55	
Air Blank	0.000	04:56	
Control Test	0.082	04:56	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 08/04/2023 23:20:44

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.084	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.000	23:16
Breath Volume	2.023	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.000	23:19
Breath Volume	1.238	23:19
Air Blank	0.000	23:19
Control Test	0.083	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: RODRIGUEZ LOPEZ

DATE/TIME OF TEST: 08/05/2023 04:18:10

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.084	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.162	04:12
Breath Volume	1.832	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.159	04:16
Breath Volume	1.722	04:16
Air Blank	0.000	04:17
Control Test	0.082	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: GUERRERO MARULANDA

DATE/TIME OF TEST: 08/06/2023 00:26:08

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.083	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.199	00:21
Breath Volume	2.636	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:23
Subject Sample # 2	0.194	00:24
Breath Volume	2.535	00:24
Air Blank	0.000	00:25
Control Test	0.082	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 08/07/2023 02:26:15

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.083	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.158	02:21
Breath Volume	1.527	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.165	02:24
Breath Volume	1.761	02:24
Air Blank	0.000	02:25
Control Test	0.081	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: VARGAS CARAPIA

DATE/TIME OF TEST: 08/07/2023 04:30:20

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.083	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.148	04:25
Breath Volume	1.851	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.135	04:28
Breath Volume	1.488	04:28
Air Blank	0.000	04:29
Control Test	0.082	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: FERRER

DATE/TIME OF TEST: 08/09/2023 00:45:21

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.084	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.032	00:40
Breath Volume	1.234	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.033	00:43
Breath Volume	1.292	00:43
Air Blank	0.000	00:44
Control Test	0.083	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 08/10/2023 03:34:05

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.084	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.135	03:29
Breath Volume	2.027	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.131	03:32
Breath Volume	1.878	03:32
Air Blank	0.000	03:33
Control Test	0.083	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 08/12/2023 01:17:19

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.084	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.125	01:12
Breath Volume	2.078	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.128	01:15
Breath Volume	1.808	01:15
Air Blank	0.000	01:16
Control Test	0.082	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/12/2023 03:47:02

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.084	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1 0.102	03:35	Volume Not Met
Breath Volume	0.441	03:35	
Air Blank	0.000	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 2 0.125	03:39	
Breath Volume	1.437	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 3 0.107	03:45	Volume Not Met
Breath Volume	0.281	03:45	
Air Blank	0.000	03:46	
Control Test	0.082	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/12/2023 03:58:53

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.124	03:54
Breath Volume	1.554	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.120	03:57
Breath Volume	1.726	03:57
Air Blank	0.000	03:57
Control Test	0.083	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 08/12/2023 04:48:14

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.084	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.229	04:43
Breath Volume	1.273	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.231	04:46
Breath Volume	2.789	04:46
Air Blank	0.000	04:47
Control Test	0.083	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KELLEY,L
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: MACKALL

DATE/TIME OF TEST: 08/13/2023 01:21:01

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.083	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.141	01:15
Breath Volume	1.531	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.137	01:19
Breath Volume	1.269	01:19
Air Blank	0.000	01:20
Control Test	0.083	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:21

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: COMM VEHIC
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: QUINTERO

DATE/TIME OF TEST: 08/14/2023 04:09:02

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.083	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.147	04:04
Breath Volume	3.988	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:06
Subject Sample # 2	0.145	04:07
Breath Volume	3.910	04:07
Air Blank	0.000	04:08
Control Test	0.082	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:09

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: CARTAGENA-CAMPOS

DATE/TIME OF TEST: 08/20/2023 03:33:24

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.084	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.108	03:28
Breath Volume	1.207	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.109	03:31
Breath Volume	1.246	03:31
Air Blank	0.000	03:32
Control Test	0.083	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MAQUEIRA MORFA

DATE/TIME OF TEST: 08/20/2023 04:09:20

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.083	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.078	04:04
Breath Volume	1.960	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.078	04:07
Breath Volume	2.156	04:07
Air Blank	0.000	04:08
Control Test	0.083	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 08/21/2023 03:38:08

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.083	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.244	03:33
Breath Volume	1.910	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.231	03:36
Breath Volume	1.605	03:36
Air Blank	0.000	03:37
Control Test	0.082	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/21/2023 04:09:51

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.220	04:03
Breath Volume	1.253	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:05
Subject Sample # 2	0.215	04:08
Breath Volume	2.570	04:08
Air Blank	0.000	04:08
Control Test	0.082	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: GANNON

DATE/TIME OF TEST: 08/25/2023 03:56:09

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.083	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.167	03:51
Breath Volume	2.316	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.162	03:54
Breath Volume	1.992	03:54
Air Blank	0.000	03:55
Control Test	0.082	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GATEWOOD

DATE/TIME OF TEST: 08/26/2023 03:43:17

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.084	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.082	03:38
Breath Volume	2.472	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.082	03:41
Breath Volume	2.324	03:41
Air Blank	0.000	03:42
Control Test	0.082	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: LANDRY

DATE/TIME OF TEST: 08/30/2023 19:41:15

Diagnostics Check	OK	19:34
Air Blank	0.000	19:34
Control Test	0.082	19:35
Air Blank	0.000	19:35
Subject Sample # 1	0.159	19:36
Breath Volume	1.660	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:38
Subject Sample # 2	0.166	19:39
Breath Volume	1.675	19:39
Air Blank	0.000	19:40
Control Test	0.082	19:40
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 09/05/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BISHOCK

DATE/TIME OF TEST: 08/07/2023 21:41:55

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.078	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.008	21:37
Breath Volume	3.234	21:37
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.009	21:40
Breath Volume	3.285	21:40
Air Blank	0.000	21:40
Control Test	0.078	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 08/08/2023 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: WELSH

DATE/TIME OF TEST: 08/18/2023 20:40:31

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.078	20:34
Air Blank	0.000	20:35
Subject Sample # 1	0.239	20:35
Breath Volume	1.140	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 2	0.241	20:38
Breath Volume	1.996	20:38
Air Blank	0.000	20:39
Control Test	0.078	20:39
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 09/08/2023 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HUDAK

DATE/TIME OF TEST: 08/20/2023 00:26:36

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.078	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.000	00:22
Breath Volume	1.683	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.000	00:25
Breath Volume	1.808	00:25
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 09/08/2023 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: NORMANDY

DATE/TIME OF TEST: 08/02/2023 18:51:41

Diagnostics Check	OK	18:42
Air Blank	0.000	18:43
Control Test	0.080	18:43
Air Blank	0.000	18:44
Subject Sample # 1	0.308	18:46
Breath Volume	1.203	18:46
Air Blank	0.000	18:47
Air Blank	0.000	18:48
Subject Sample # 2	0.290	18:50
Breath Volume	1.226	18:50
Air Blank	0.000	18:50
Control Test	0.077	18:51
Air Blank	0.000	18:51
Diagnostics Check	OK	18:51

Data Download Date/Time: 08/07/2023 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ROBERT,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: GOODSON,LEANDRA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 12:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:57

SUBJECT'S LAST NAME: GARCIA MELGAR

DATE/TIME OF TEST: 08/03/2023 13:26:36

Diagnostics Check	OK	13:19
Air Blank	0.000	13:20
Control Test	0.079	13:20
Air Blank	0.000	13:21
Subject Sample # 1	0.218	13:22
Breath Volume	1.957	13:22
Air Blank	0.000	13:22
Air Blank	0.000	13:24
Subject Sample # 2	0.233	13:25
Breath Volume	1.527	13:25
Air Blank	0.000	13:25
Control Test	0.079	13:26
Air Blank	0.000	13:26
Diagnostics Check	OK	13:26

Data Download Date/Time: 08/07/2023 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COZAD,AUNDREA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: IRVIN

DATE/TIME OF TEST: 08/09/2023 05:55:02

Diagnostics Check	OK	05:51	
Air Blank	0.000	05:52	
Control Test	0.080	05:52	
Air Blank	0.000	05:52	
Subject Sample	# 1	05:53	Subject Test Refused
Breath Volume	0.000	05:53	
Air Blank	0.000	05:54	
Control Test	0.080	05:54	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:55	

Data Download Date/Time: 09/07/2023 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COZAD,AUNDREA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: IRVIN

DATE/TIME OF TEST: 08/09/2023 06:17:53

Diagnostics Check	OK	06:11
Air Blank	0.000	06:12
Control Test	0.079	06:12
Air Blank	0.000	06:13
Subject Sample # 1	0.126	06:13
Breath Volume	3.226	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:16
Subject Sample # 2	0.125	06:16
Breath Volume	3.855	06:16
Air Blank	0.000	06:16
Control Test	0.081	06:17
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 09/07/2023 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LERNER,MATTHEW,H
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ANDREWS,SHAAMAR
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: ICKES

DATE/TIME OF TEST: 08/12/2023 19:00:37

Diagnostics Check	OK	18:54
Air Blank	0.000	18:55
Control Test	0.079	18:55
Air Blank	0.000	18:55
Subject Sample # 1	0.169	18:56
Breath Volume	2.628	18:56
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.171	18:59
Breath Volume	2.511	18:59
Air Blank	0.000	18:59
Control Test	0.079	19:00
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 09/07/2023 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELFREY,PAUL,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: LAING

DATE/TIME OF TEST: 08/13/2023 05:47:27

Diagnostics Check	OK	05:44	
Air Blank	0.000	05:44	
Control Test	0.080	05:45	
Air Blank	0.000	05:45	
Subject Sample	# 1	05:46	Subject Test Refused
Breath Volume	0.000	05:46	
Air Blank	0.000	05:46	
Control Test	0.080	05:46	
Air Blank	0.000	05:47	
Diagnostics Check	OK	05:47	

Data Download Date/Time: 09/07/2023 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COZAD,AUNDREA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 08/20/2023 05:30:26

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.080	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.135	05:26
Breath Volume	2.042	05:26
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.139	05:28
Breath Volume	1.699	05:28
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 09/07/2023 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: SANSONE

DATE/TIME OF TEST: 08/06/2023 22:52:53

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.078	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.170	22:48
Breath Volume	3.234	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:50
Subject Sample # 2	0.154	22:51
Breath Volume	1.335	22:51
Air Blank	0.000	22:51
Control Test	0.077	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 08/29/2023 09:43