

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: SIMMENS,ADAM,J	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: NOLAND,THOMAS	ARREST TIME: 23:09
ARRESTING OFFICER AGENCY: Lake Mary PD	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: WOODS	DATE/TIME OF TEST: 10/01/2023 00:09:31
----------------------------	--

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.225	00:04
Breath Volume	4.132	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.217	00:07
Breath Volume	4.714	00:07
Air Blank	0.000	00:08
Control Test	0.079	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 10/12/2023 05:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: BEHR,CHRISTINA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 13:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:53

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/01/2023 14:32:07

Diagnostics Check	OK	14:20	
Air Blank	0.000	14:21	
Control Test	0.079	14:21	
Air Blank	0.000	14:22	
Subject Sample	# 1	14:25	No Sample Provided
Breath Volume	0.000	14:25	
Air Blank	0.000	14:25	
Air Blank	0.000	14:27	
Subject Sample	# 2	14:30	No Sample Provided
Breath Volume	0.000	14:30	
Air Blank	0.000	14:31	
Control Test	0.080	14:31	
Air Blank	0.000	14:32	
Diagnostics Check	OK	14:32	

Data Download Date/Time: 10/12/2023 05:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: GRUNAT,JERRY

ARRESTING OFFICER AGENCY: Seminole County SO

LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 01:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/02/2023 02:38:21

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.013	02:34
Breath Volume	2.519	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.013	02:36
Breath Volume	2.402	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 10/12/2023 05:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ADAMS,JUSTIN	ARREST TIME: 00:43
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Female
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: NEWCOME HERMES	DATE/TIME OF TEST: 10/07/2023	02:06:10
-------------------------------------	-------------------------------	----------

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.079	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.000	02:01
Breath Volume	1.343	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.000	02:04
Breath Volume	1.460	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 10/12/2023 05:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: REDICK,D	ARREST TIME: 03:05
ARRESTING OFFICER AGENCY: Altamonte Springs PD	GENDER: Female
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SAURI	DATE/TIME OF TEST: 10/07/2023	05:09:08
----------------------------	-------------------------------	----------

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.079	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.132	05:04
Breath Volume	1.523	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.123	05:07
Breath Volume	2.964	05:07
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 10/12/2023 05:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: BLACKWELL,R	ARREST TIME: 00:52
ARRESTING OFFICER AGENCY: Maitland PD	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: JONES	DATE/TIME OF TEST: 10/08/2023 02:01:39
----------------------------	--

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.155	01:57
Breath Volume	2.304	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.146	02:00
Breath Volume	2.230	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 10/12/2023 05:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MOXLEY,M	ARREST TIME: 03:21
ARRESTING OFFICER AGENCY: Casselberry PD	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: EHMEN	DATE/TIME OF TEST: 10/10/2023 05:30:12
----------------------------	--

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.157	05:25
Breath Volume	2.027	05:25
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.146	05:28
Breath Volume	1.949	05:28
Air Blank	0.000	05:29
Control Test	0.079	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 10/12/2023 05:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: JUSINO,KARLO	ARREST TIME: 20:35
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Female
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: DIAZ	DATE/TIME OF TEST: 10/10/2023	23:12:06
---------------------------	-------------------------------	----------

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.080	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.341	23:07
Breath Volume	2.250	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.354	23:10
Breath Volume	2.031	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 10/12/2023 05:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: LOUISSAINT,R	ARREST TIME: 23:11
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: MORENO LARA	DATE/TIME OF TEST: 10/11/2023 00:22:38
----------------------------------	--

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.194	00:17
Breath Volume	1.886	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.190	00:21
Breath Volume	1.402	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 10/12/2023 05:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PRICE,CHASE	ARREST TIME: 01:20
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: MARTINEZ VARGAS	DATE/TIME OF TEST: 10/12/2023 02:55:46
--------------------------------------	--

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.163	02:51
Breath Volume	1.503	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 2	0.154	02:54
Breath Volume	1.464	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 10/12/2023 05:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BETHAM,KEITH,D	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: BRYANT,STEVEN	ARREST TIME: 23:46
ARRESTING OFFICER AGENCY: Longwood PD	GENDER: Female
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: MCDONALD	DATE/TIME OF TEST: 10/14/2023	00:44:57
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.171	00:40
Breath Volume	2.667	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:42
Subject Sample # 2	0.174	00:43
Breath Volume	2.753	00:43
Air Blank	0.000	00:44
Control Test	0.079	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 10/19/2023 05:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: NOLAND,THOMAS	ARREST TIME: 02:13
ARRESTING OFFICER AGENCY: Lake Mary PD	GENDER: Female
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ALVAREZ	DATE/TIME OF TEST: 10/14/2023 03:34:49
------------------------------	--

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.158	03:30
Breath Volume	2.574	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.153	03:33
Breath Volume	1.804	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 10/19/2023 05:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: RILEY,BRANDON	ARREST TIME: 03:00
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MCGHEE	DATE/TIME OF TEST: 10/14/2023 04:32:30
-----------------------------	--

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.175	04:27
Breath Volume	4.792	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.174	04:31
Breath Volume	4.015	04:31
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 10/19/2023 05:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GRUNAT,JERRY	ARREST TIME: 02:21
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MCCONNELL	DATE/TIME OF TEST: 10/15/2023 03:58:13
--------------------------------	--

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.160	03:53
Breath Volume	2.031	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.157	03:56
Breath Volume	1.585	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 10/19/2023 05:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: BLACKWELL,ROBERT	ARREST TIME: 02:34
ARRESTING OFFICER AGENCY: Maitland PD	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GONZALEZ BENITEZ	DATE/TIME OF TEST: 10/15/2023	04:18:06
---------------------------------------	-------------------------------	----------

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.095	04:13
Breath Volume	2.449	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.091	04:16
Breath Volume	1.683	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 10/19/2023 05:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SINGH,TERRANCE	ARREST TIME: 03:53
ARRESTING OFFICER AGENCY: Altamonte Springs PD	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: BONILLA	DATE/TIME OF TEST: 10/15/2023 04:58:31
------------------------------	--

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.079	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.163	04:54
Breath Volume	3.156	04:54
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.166	04:57
Breath Volume	1.750	04:57
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 10/19/2023 05:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DASQUE,DARREN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: MUCKERSON

DATE/TIME OF TEST: 10/15/2023 05:10:02

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.188	05:05
Breath Volume	2.550	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.186	05:08
Breath Volume	3.593	05:08
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:10

Data Download Date/Time: 10/19/2023 05:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GANT,JARRET	ARREST TIME: 04:23
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: RAMOS-RODRIGUEZ	DATE/TIME OF TEST: 10/15/2023 05:29:34
--------------------------------------	--

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.079	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.069	05:24
Breath Volume	1.652	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.070	05:28
Breath Volume	2.121	05:28
Air Blank	0.000	05:28
Control Test	0.079	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 10/19/2023 05:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Winter Springs PD
ARRESTING OFFICER: ESPINOSA,A
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 10/15/2023 06:50:35

Diagnostics Check	OK	06:47	
Air Blank	0.000	06:48	
Control Test	0.080	06:48	
Air Blank	0.000	06:49	
Subject Sample	# 1	06:49	Subject Test Refused
Breath Volume	0.000	06:49	
Air Blank	0.000	06:49	
Control Test	0.080	06:50	
Air Blank	0.000	06:50	
Diagnostics Check	OK	06:50	

Data Download Date/Time: 10/19/2023 05:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: MACTYE,DOUGLAS
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: DERUSSO

DATE/TIME OF TEST: 10/19/2023 15:35:30

Diagnostics Check	OK	15:24	
Air Blank	0.000	15:24	
Control Test	0.080	15:24	
Air Blank	0.000	15:25	
Subject Sample	# 1	15:28	No Sample Provided
Breath Volume	0.000	15:28	
Air Blank	0.000	15:29	
Air Blank	0.000	15:31	
Subject Sample	# 2	15:34	No Sample Provided
Breath Volume	0.000	15:34	
Air Blank	0.000	15:34	
Control Test	0.081	15:34	
Air Blank	0.000	15:35	
Diagnostics Check	OK	15:35	

Data Download Date/Time: 10/26/2023 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HIPPLER,ALYCIA	ARREST TIME: 04:36
ARRESTING OFFICER AGENCY: Altamonte Springs PD	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: RODRIGUEZ	DATE/TIME OF TEST: 10/20/2023 05:28:02
--------------------------------	--

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.222	05:23
Breath Volume	1.832	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.221	05:26
Breath Volume	2.343	05:26
Air Blank	0.000	05:27
Control Test	0.079	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:28

Data Download Date/Time: 10/26/2023 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PRICE,CHASE	ARREST TIME: 03:32
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Female
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: BROOME	DATE/TIME OF TEST: 10/21/2023 05:01:50
-----------------------------	--

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.160	04:57
Breath Volume	2.773	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.157	05:00
Breath Volume	2.960	05:00
Air Blank	0.000	05:00
Control Test	0.080	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 10/26/2023 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: WATT,ANTONIO	ARREST TIME: 21:03
ARRESTING OFFICER AGENCY: Altamonte Springs PD	GENDER: Female
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BROWN	DATE/TIME OF TEST: 10/21/2023 22:27:11
----------------------------	--

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.269	22:22
Breath Volume	2.910	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.270	22:25
Breath Volume	3.136	22:25
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 10/26/2023 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HORNE,CLINTEN	ARREST TIME: 00:22
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Female
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: RUIZ	DATE/TIME OF TEST: 10/22/2023 01:26:10
---------------------------	--

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.179	01:21
Breath Volume	1.222	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.186	01:24
Breath Volume	2.710	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 10/26/2023 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: IRIZARRY,CHRISTOPHER	ARREST TIME: 00:49
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: BELL JR	DATE/TIME OF TEST: 10/22/2023 02:04:54
------------------------------	--

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.079	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1 0.222	01:54	No .020 agreement
Breath Volume	1.679	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample	# 2 0.245	01:57	No .020 agreement
Breath Volume	1.199	01:57	
Air Blank	0.000	01:58	
Air Blank	0.000	02:00	
Subject Sample	# 3	02:03	No Sample Provided
Breath Volume	0.000	02:03	
Air Blank	0.000	02:04	
Control Test	0.079	02:04	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 10/26/2023 03:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SMITH,LOGAN	ARREST TIME: 04:25
ARRESTING OFFICER AGENCY: Oviedo PD	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: FISH	DATE/TIME OF TEST: 10/22/2023 05:07:32
---------------------------	--

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.079	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.039	05:03
Breath Volume	4.296	05:03
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.038	05:06
Breath Volume	3.554	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 10/26/2023 03:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: CAVALLERO,TRAVIS	ARREST TIME: 00:45
ARRESTING OFFICER AGENCY: Casselberry PD	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: ANDINO	DATE/TIME OF TEST: 10/23/2023 01:49:30
-----------------------------	--

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.117	01:45
Breath Volume	2.023	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.117	01:48
Breath Volume	1.394	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 10/26/2023 03:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HIDALGO,A	ARREST TIME: 02:53
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: HERNANDEZ OLGUIN	DATE/TIME OF TEST: 10/23/2023	03:58:57
---------------------------------------	-------------------------------	----------

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.136	03:54
Breath Volume	1.675	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.128	03:57
Breath Volume	1.492	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 10/26/2023 03:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KOURI,TAMY,

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: BARBOSA,L

ARRESTING OFFICER AGENCY: Altamonte Springs PD

LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 20:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: PEREZ DELGADO

DATE/TIME OF TEST: 10/23/2023 22:37:43

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.080	22:30
Air Blank	0.000	22:30
Subject Sample	# 1 0.241	22:32
Breath Volume	3.207	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample	# 2 0.240	22:36
Breath Volume	2.019	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 10/26/2023 03:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MARSHALL,R	ARREST TIME: 23:27
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ESPINOZA	DATE/TIME OF TEST: 10/24/2023 01:05:34
-------------------------------	--

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.080	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.234	01:01
Breath Volume	1.972	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.231	01:04
Breath Volume	2.839	01:04
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 10/26/2023 03:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MOORE,A	ARREST TIME: 23:56
ARRESTING OFFICER AGENCY: Maitland PD	GENDER: Male
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: NICOLAS	DATE/TIME OF TEST: 10/24/2023	01:48:09
------------------------------	-------------------------------	----------

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.000	01:43
Breath Volume	3.148	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.000	01:46
Breath Volume	2.425	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 10/26/2023 03:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: UCELO,B	ARREST TIME: 22:51
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Female
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: KENNON	DATE/TIME OF TEST: 10/25/2023 00:11:34
-----------------------------	--

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.129	00:06
Breath Volume	1.890	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.127	00:10
Breath Volume	1.746	00:10
Air Blank	0.000	00:10
Control Test	0.081	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 10/26/2023 03:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MERCADO,G	ARREST TIME: 22:46
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Female
LAST AGENCY INSPECTION: 10/25/2023	OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: RUFFIN	DATE/TIME OF TEST: 10/26/2023	00:41:09
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.197	00:36
Breath Volume	1.640	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.193	00:39
Breath Volume	1.539	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 10/26/2023 03:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ADAMS,JUSTIN	ARREST TIME: 02:25
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 10/25/2023	OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SMITH	DATE/TIME OF TEST: 10/27/2023 03:44:26
----------------------------	--

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.080	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.140	03:40
Breath Volume	2.968	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.141	03:42
Breath Volume	3.363	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 11/22/2023 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: NOLAND,T	ARREST TIME: 00:10
ARRESTING OFFICER AGENCY: Lake Mary PD	GENDER: Male
LAST AGENCY INSPECTION: 10/25/2023	OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: POWELL JR.	DATE/TIME OF TEST: 10/29/2023 01:35:57
---------------------------------	--

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.075	01:31
Breath Volume	2.351	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:33
Subject Sample # 2	0.079	01:34
Breath Volume	1.203	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 11/22/2023 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: RODRIGUEZ,J	ARREST TIME: 01:15
ARRESTING OFFICER AGENCY: Casselberry PD	GENDER: Female
LAST AGENCY INSPECTION: 10/25/2023	OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: WILLIS	DATE/TIME OF TEST: 10/29/2023 02:41:50
-----------------------------	--

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.159	02:37
Breath Volume	1.558	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.152	02:40
Breath Volume	1.691	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 11/22/2023 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: VAN ROMPA,V	ARREST TIME: 21:39
ARRESTING OFFICER AGENCY: Lake Mary PD	GENDER: Male
LAST AGENCY INSPECTION: 10/25/2023	OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: PEREZ LOPEZ	DATE/TIME OF TEST: 10/29/2023	22:53:57
----------------------------------	-------------------------------	----------

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.077	22:49
Breath Volume	2.140	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.074	22:52
Breath Volume	2.078	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 11/22/2023 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SMALLEY,R	ARREST TIME: 22:28
ARRESTING OFFICER AGENCY: Winter Springs PD	GENDER: Male
LAST AGENCY INSPECTION: 10/25/2023	OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ROSILLO ARRIOLA	DATE/TIME OF TEST: 10/29/2023 23:52:54
--------------------------------------	--

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.185	23:48
Breath Volume	2.503	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.194	23:51
Breath Volume	1.808	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 11/22/2023 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MATEO,Y	ARREST TIME: 23:18
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Male
LAST AGENCY INSPECTION: 10/25/2023	OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: SHANKS	DATE/TIME OF TEST: 10/31/2023 00:35:17
-----------------------------	--

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.150	00:30
Breath Volume	2.316	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.142	00:33
Breath Volume	1.898	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 11/22/2023 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEWART, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEWART, WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: TRUEHILL

DATE/TIME OF TEST: 10/01/2023 03:35:39

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.081	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:34	
Control Test	0.080	03:35	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 10/31/2023 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,AMANDA,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MINNIFIELD,SAM
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 10/07/2023 02:50:29

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.081	02:37	
Air Blank	0.000	02:38	
Subject Sample	# 1 0.090	02:41	Volume Not Met
Breath Volume	0.652	02:41	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample	# 2 0.095	02:45	
Breath Volume	1.800	02:45	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample	# 3 0.094	02:48	
Breath Volume	2.679	02:48	
Air Blank	0.000	02:49	
Control Test	0.077	02:49	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 10/31/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 10/14/2023 02:26:53

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.195	02:22
Breath Volume	2.308	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.194	02:25
Breath Volume	2.429	02:25
Air Blank	0.000	02:25
Control Test	0.081	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SECCHIARI,CLARK
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 10/15/2023 16:49:16

Diagnostics Check	OK	16:42
Air Blank	0.000	16:43
Control Test	0.083	16:43
Air Blank	0.000	16:44
Subject Sample # 1	0.251	16:44
Breath Volume	1.925	16:44
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.246	16:47
Breath Volume	2.257	16:47
Air Blank	0.000	16:48
Control Test	0.079	16:48
Air Blank	0.000	16:49
Diagnostics Check	OK	16:49

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,TIMYRON,N
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GLASS,JAMES
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 10:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: LEFLER

DATE/TIME OF TEST: 10/16/2023 11:14:31

Diagnostics Check	OK	11:07
Air Blank	0.000	11:08
Control Test	0.082	11:08
Air Blank	0.000	11:09
Subject Sample # 1	0.226	11:09
Breath Volume	1.355	11:09
Air Blank	0.000	11:10
Air Blank	0.000	11:12
Subject Sample # 2	0.231	11:12
Breath Volume	1.851	11:12
Air Blank	0.000	11:13
Control Test	0.081	11:13
Air Blank	0.000	11:14
Diagnostics Check	OK	11:14

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,BRITTANY,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/18/2023 18:02:39

Diagnostics Check	OK	17:54
Air Blank	0.000	17:55
Control Test	0.082	17:55
Air Blank	0.000	17:55
Subject Sample # 1	0.224	17:57
Breath Volume	1.625	17:57
Air Blank	0.000	17:58
Air Blank	0.000	18:00
Subject Sample # 2	0.206	18:00
Breath Volume	1.429	18:00
Air Blank	0.000	18:01
Control Test	0.077	18:02
Air Blank	0.000	18:02
Diagnostics Check	OK	18:02

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLTON,JUSTIN,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOLTON,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: RICHEY

DATE/TIME OF TEST: 10/23/2023 21:52:43

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.079	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.197	21:48
Breath Volume	1.972	21:48
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.193	21:51
Breath Volume	1.207	21:51
Air Blank	0.000	21:51
Control Test	0.081	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDD,RACHEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUDD,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 10/27/2023 20:26:36

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.082	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.000	20:22
Breath Volume	1.906	20:22
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.000	20:25
Breath Volume	1.871	20:25
Air Blank	0.000	20:25
Control Test	0.083	20:25
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORCE,TAYLOR,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FORCE,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: GOULD

DATE/TIME OF TEST: 10/28/2023 03:17:25

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.082	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.173	03:12
Breath Volume	1.636	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.167	03:15
Breath Volume	1.621	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDD,RACHEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUDD,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 10/28/2023 21:25:59

Diagnostics Check	OK	21:14	
Air Blank	0.000	21:14	
Control Test	0.081	21:15	
Air Blank	0.000	21:15	
Subject Sample	# 1 0.234	21:18	Volume Not Met
Breath Volume	0.175	21:18	
Air Blank	0.000	21:19	
Air Blank	0.000	21:21	
Subject Sample	# 2 0.226	21:24	Volume Not Met
Breath Volume	0.015	21:24	
Air Blank	0.000	21:24	
Control Test	0.079	21:25	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORCE,TAYLOR,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FORCE,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/29/2023 03:46:30

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.071	03:42
Breath Volume	2.113	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.073	03:44
Breath Volume	2.273	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 10/29/2023 18:51:07

Diagnostics Check	OK	18:43
Air Blank	0.000	18:43
Control Test	0.081	18:44
Air Blank	0.000	18:44
Subject Sample # 1	0.167	18:46
Breath Volume	1.125	18:46
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample # 2	0.167	18:49
Breath Volume	1.335	18:49
Air Blank	0.000	18:50
Control Test	0.080	18:50
Air Blank	0.000	18:51
Diagnostics Check	OK	18:51

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SECHIARRI,CLARK
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/29/2023 20:46:36

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.081	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.247	20:42
Breath Volume	1.757	20:42
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.238	20:44
Breath Volume	1.941	20:44
Air Blank	0.000	20:45
Control Test	0.079	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LRRODRIGUEZ,LUIS,R
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: FULTON,DAKOTA
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: KELIGOND

DATE/TIME OF TEST: 10/31/2023 04:25:57

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:22	
Control Test	0.082	04:22	
Air Blank	0.000	04:23	
Subject Sample	# 1	04:25	Sequence Aborted
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAVES,TIMOTHY,F
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: PARENTE,JOSEPH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 08:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 10/31/2023 09:25:42

Diagnostics Check	OK	09:17
Air Blank	0.000	09:17
Control Test	0.081	09:18
Air Blank	0.000	09:18
Subject Sample # 1	0.321	09:21
Breath Volume	3.312	09:21
Air Blank	0.000	09:21
Air Blank	0.000	09:23
Subject Sample # 2	0.311	09:24
Breath Volume	1.828	09:24
Air Blank	0.000	09:24
Control Test	0.078	09:25
Air Blank	0.000	09:25
Diagnostics Check	OK	09:25

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSTICE,JOSEPH,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JUSTICE,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 13:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:42

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/31/2023 14:08:29

Diagnostics Check	OK	14:02
Air Blank	0.000	14:02
Control Test	0.082	14:03
Air Blank	0.000	14:03
Subject Sample # 1	0.278	14:04
Breath Volume	1.425	14:04
Air Blank	0.000	14:04
Air Blank	0.000	14:06
Subject Sample # 2	0.274	14:06
Breath Volume	1.589	14:06
Air Blank	0.000	14:07
Control Test	0.079	14:07
Air Blank	0.000	14:08
Diagnostics Check	OK	14:08

Data Download Date/Time: 11/21/2023 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WILSON,MALCOLM
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DINKLA

DATE/TIME OF TEST: 10/03/2023 03:45:18

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.000	03:41
Breath Volume	3.988	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.000	03:43
Breath Volume	2.886	03:43
Air Blank	0.000	03:44
Control Test	0.081	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 10/10/2023 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLANES,MILAGROS,O
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: LLANES,MILAGROS
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MCCAFFREY

DATE/TIME OF TEST: 10/06/2023 03:22:01

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 10/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,LIBBEE,S
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: WILLIAMS,STEPHANIE
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/14/2023 06:11:30

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:00	
Control Test	0.081	06:00	
Air Blank	0.000	06:01	
Subject Sample	# 1	06:04	No Sample Provided
Breath Volume	0.000	06:04	
Air Blank	0.000	06:05	
Air Blank	0.000	06:07	
Subject Sample	# 2	06:10	No Sample Provided
Breath Volume	0.000	06:10	
Air Blank	0.000	06:10	
Control Test	0.082	06:10	
Air Blank	0.000	06:11	
Diagnostics Check	OK	06:11	

Data Download Date/Time: 11/21/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SALTER,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PICKENS

DATE/TIME OF TEST: 10/15/2023 04:21:18

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.082	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.196	04:16
Breath Volume	2.855	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.198	04:19
Breath Volume	3.433	04:19
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 11/21/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LEIGHA,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: JONES,LEIGHA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ALEGRIA DAVILA

DATE/TIME OF TEST: 10/15/2023 20:52:53

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.082	20:47
Air Blank	0.000	20:47
Subject Sample # 1	0.276	20:48
Breath Volume	2.023	20:48
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.282	20:51
Breath Volume	1.468	20:51
Air Blank	0.000	20:51
Control Test	0.081	20:52
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 11/21/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MILLER,ETHAN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/18/2023 00:47:50

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:36	
Control Test	0.082	00:37	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:40	No Sample Provided
Breath Volume	0.000	00:40	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 2	00:46	No Sample Provided
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Control Test	0.081	00:47	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 11/21/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MINCK,ISAAC
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 10/21/2023 02:02:59

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.081	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 1 0.188	01:50	
Breath Volume	1.984	01:50	
Air Blank	0.000	01:50	
Air Blank	0.000	01:52	
Subject Sample	# 2 0.128	01:55	Volume Not Met
Breath Volume	1.015	01:55	
Air Blank	0.000	01:56	
Air Blank	0.000	01:58	
Subject Sample	# 3 0.193	02:01	Volume Not Met
Breath Volume	0.203	02:01	
Air Blank	0.000	02:02	
Control Test	0.079	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 11/21/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MINCK,ISAAC
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 10/21/2023 02:20:26

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:09	
Control Test	0.079	02:09	
Air Blank	0.000	02:09	
Subject Sample	# 1 0.198	02:13	Volume Not Met
Breath Volume	0.363	02:13	
Air Blank	0.000	02:13	
Air Blank	0.000	02:15	
Subject Sample	# 2 0.177	02:18	Volume Not Met
Breath Volume	0.210	02:18	
Air Blank	0.000	02:19	
Control Test	0.078	02:19	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 11/21/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MINCK,ISAAC
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 10/21/2023 02:56:29

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.078	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.173	02:50
Breath Volume	1.894	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.169	02:54
Breath Volume	2.792	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 11/21/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MINCK,ISAAC
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: GADDY

DATE/TIME OF TEST: 10/22/2023 20:09:33

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:58	
Control Test	0.081	19:58	
Air Blank	0.000	19:59	
Subject Sample	# 1	20:02	No Sample Provided
Breath Volume	0.000	20:02	
Air Blank	0.000	20:03	
Air Blank	0.000	20:05	
Subject Sample	# 2	20:08	No Sample Provided
Breath Volume	0.000	20:08	
Air Blank	0.000	20:08	
Control Test	0.082	20:08	
Air Blank	0.000	20:09	
Diagnostics Check	OK	20:09	

Data Download Date/Time: 11/21/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WELCH,CHAD
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: MCRAE

DATE/TIME OF TEST: 10/27/2023 17:02:59

Diagnostics Check	OK	16:56
Air Blank	0.000	16:57
Control Test	0.080	16:57
Air Blank	0.000	16:58
Subject Sample # 1	0.000	16:58
Breath Volume	2.378	16:58
Air Blank	0.000	16:59
Air Blank	0.000	17:01
Subject Sample # 2	0.000	17:01
Breath Volume	2.203	17:01
Air Blank	0.000	17:02
Control Test	0.081	17:02
Air Blank	0.000	17:02
Diagnostics Check	OK	17:02

Data Download Date/Time: 11/21/2023 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: UPSHAW,JASON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/29/2023 01:52:53

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.081	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 11/21/2023 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: ALTMAN,MICHAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: MCDAVID

DATE/TIME OF TEST: 10/29/2023 04:40:01

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.137	04:35
Breath Volume	1.390	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:37
Subject Sample # 2	0.140	04:38
Breath Volume	2.882	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:40

Data Download Date/Time: 11/21/2023 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HOWARD,N
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 10/14/2023 18:22:52

Diagnostics Check	OK	18:16
Air Blank	0.000	18:17
Control Test	0.080	18:17
Air Blank	0.000	18:18
Subject Sample # 1	0.000	18:18
Breath Volume	2.210	18:18
Air Blank	0.000	18:19
Air Blank	0.000	18:21
Subject Sample # 2	0.000	18:21
Breath Volume	2.804	18:21
Air Blank	0.000	18:21
Control Test	0.079	18:22
Air Blank	0.000	18:22
Diagnostics Check	OK	18:22

Data Download Date/Time: 10/25/2023 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOONE,RODNEY,E
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: BOONE,RODNEY
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 10/27/2023 17:23:55

Diagnostics Check	OK	17:12	
Air Blank	0.000	17:12	
Control Test	0.079	17:12	
Air Blank	0.000	17:13	
Subject Sample	# 1 0.189	17:16	Volume Not Met
Breath Volume	0.050	17:16	
Air Blank	0.000	17:17	
Air Blank	0.000	17:18	
Subject Sample	# 2 0.159	17:22	Volume Not Met
Breath Volume	0.035	17:22	
Air Blank	0.000	17:22	
Control Test	0.078	17:23	
Air Blank	0.000	17:23	
Diagnostics Check	OK	17:23	

Data Download Date/Time: 11/25/2023 23:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, TYLER, M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WYCHUNAS, REIDE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: DEATON

DATE/TIME OF TEST: 10/06/2023 21:25:24

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:22	
Control Test	0.081	21:23	
Air Blank	0.000	21:23	
Subject Sample	# 1	21:24	Subject Test Refused
Breath Volume	0.000	21:24	
Air Blank	0.000	21:24	
Control Test	0.081	21:24	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 11/05/2023 20:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODDO,ISAIAH,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CARSON,KERRY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/13/2023 04:49:54

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.000	04:45
Breath Volume	1.148	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.000	04:48
Breath Volume	1.214	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 11/05/2023 20:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: SCHMOLKE

DATE/TIME OF TEST: 10/14/2023 01:28:18

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.080	01:25	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:27	
Control Test	0.081	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: SHERRION

DATE/TIME OF TEST: 10/14/2023 18:56:22

Diagnostics Check	OK	18:45	
Air Blank	0.000	18:46	
Control Test	0.078	18:46	
Air Blank	0.000	18:47	
Subject Sample	# 1 0.276	18:47	No .020 agreement
Breath Volume	1.886	18:47	
Air Blank	0.000	18:48	
Air Blank	0.000	18:50	
Subject Sample	# 2 0.244	18:50	No .020 agreement
Breath Volume	2.234	18:50	
Air Blank	0.000	18:51	
Air Blank	0.000	18:53	
Subject Sample	# 3 0.220	18:54	No .020 agreement
Breath Volume	1.746	18:54	
Air Blank	0.000	18:55	
Control Test	0.080	18:55	
Air Blank	0.000	18:56	
Diagnostics Check	OK	18:56	

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: SPELL

DATE/TIME OF TEST: 10/14/2023 21:40:30

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:29	
Control Test	0.081	21:29	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:33	No Sample Provided
Breath Volume	0.000	21:33	
Air Blank	0.000	21:33	
Air Blank	0.000	21:35	
Subject Sample	# 2	21:39	No Sample Provided
Breath Volume	0.000	21:39	
Air Blank	0.000	21:39	
Control Test	0.080	21:39	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ELMORE

DATE/TIME OF TEST: 10/15/2023 01:45:08

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.122	01:40
Breath Volume	2.257	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.120	01:43
Breath Volume	2.746	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BLACKWOOD,PATRICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BUNKER

DATE/TIME OF TEST: 10/15/2023 02:05:13

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.172	02:00
Breath Volume	2.257	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.175	02:03
Breath Volume	2.828	02:03
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAHL,RANDY,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAHL,RANDY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: PYLAND

DATE/TIME OF TEST: 10/15/2023 16:57:27

Diagnostics Check	OK	16:51
Air Blank	0.000	16:51
Control Test	0.081	16:52
Air Blank	0.000	16:52
Subject Sample # 1	0.000	16:53
Breath Volume	1.175	16:53
Air Blank	0.000	16:53
Air Blank	0.000	16:55
Subject Sample # 2	0.000	16:56
Breath Volume	2.605	16:56
Air Blank	0.000	16:56
Control Test	0.081	16:56
Air Blank	0.000	16:57
Diagnostics Check	OK	16:57

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISANSAL,HARMON,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEISANSAL,HARMON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RABEN

DATE/TIME OF TEST: 10/20/2023 00:27:23

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.129	00:22
Breath Volume	3.058	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.130	00:25
Breath Volume	3.367	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JESSICA,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NELSON,JESSICA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/20/2023 03:54:30

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:53	
Control Test	0.080	03:53	
Air Blank		03:54	RFI Detect
Air Blank	0.000	03:54	

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JESSICA,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NELSON,JESSICA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/20/2023 03:57:48

Diagnostics Check	OK	03:57	
Air Blank		03:57	RFI Detect
Air Blank		03:57	RFI Detect

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JESSICA,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NELSON,JESSICA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/20/2023 04:03:23

Diagnostics Check	OK	03:59	
Air Blank	0.000	04:00	
Control Test	0.077	04:00	
Air Blank	0.000	04:01	
Subject Sample	# 1 0.196	04:01	
Breath Volume	3.394	04:01	
Air Blank		04:02	RFI Detect
Air Blank	0.000	04:02	
Control Test	0.081	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: CATO

DATE/TIME OF TEST: 10/20/2023 23:14:30

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.199	23:09
Breath Volume	2.101	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.197	23:12
Breath Volume	1.492	23:12
Air Blank	0.000	23:13
Control Test	0.078	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/21/2023 01:32:00

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.166	01:27
Breath Volume	2.562	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.160	01:30
Breath Volume	2.339	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:32

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,DYLAN,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: HART,DYLAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 10/21/2023 05:01:52

Diagnostics Check	OK	05:01	
Air Blank		05:01	RFI Detect
Air Blank	0.000	05:01	

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,DYLAN,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: HART,DYLAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 10/21/2023 05:10:35

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.081	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.168	05:06
Breath Volume	1.589	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.165	05:09
Breath Volume	1.562	05:09
Air Blank	0.000	05:09
Control Test	0.078	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 11/05/2023 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOSLEY,KYLE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 12:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 10/21/2023 13:07:15

Diagnostics Check	OK	13:06	
Air Blank	0.000	13:07	
		13:07	RFI Detect
Air Blank		13:07	RFI Detect

Data Download Date/Time: 11/05/2023 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOSLEY,KYLE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 12:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 10/21/2023 13:19:17

Diagnostics Check	OK	13:12
Air Blank	0.000	13:13
Control Test	0.081	13:13
Air Blank	0.000	13:13
Subject Sample # 1	0.299	13:14
Breath Volume	1.445	13:14
Air Blank	0.000	13:15
Air Blank	0.000	13:17
Subject Sample # 2	0.288	13:17
Breath Volume	1.710	13:17
Air Blank	0.000	13:18
Control Test	0.076	13:18
Air Blank	0.000	13:19
Diagnostics Check	OK	13:19

Data Download Date/Time: 11/05/2023 20:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODDO, ISIAH, J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ODDO, ISIAH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 10/22/2023 00:11:08

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.080	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1 0.127	00:03	Volume Not Met
Breath Volume	0.253	00:03	
Air Blank	0.000	00:04	
Air Blank	0.000	00:06	
Subject Sample	# 2 0.119	00:09	Volume Not Met
Breath Volume	0.386	00:09	
Air Blank	0.000	00:10	
Control Test	0.079	00:10	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 11/05/2023 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODDO,ISIAH,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ODDO,ISIAH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 10/22/2023 00:28:31

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:14	
Control Test	0.078	00:14	
Air Blank	0.000	00:15	
Subject Sample	# 1 0.136	00:18	Volume Not Met
Breath Volume	0.234	00:18	
Air Blank	0.000	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 2 0.131	00:22	
Breath Volume	1.371	00:22	
Air Blank	0.000	00:23	
Air Blank	0.000	00:24	
Subject Sample	# 3 0.143	00:26	
Breath Volume	1.503	00:26	
Air Blank	0.000	00:27	
Control Test	0.078	00:27	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 11/05/2023 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPORCO, DAVID, A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES, ZACARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: COZART

DATE/TIME OF TEST: 10/22/2023 02:27:00

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.108	02:22
Breath Volume	3.777	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 2	0.104	02:25
Breath Volume	3.609	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:27

Data Download Date/Time: 11/05/2023 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: ASPER

DATE/TIME OF TEST: 10/24/2023 23:06:06

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.079	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.169	23:01
Breath Volume	2.554	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.175	23:04
Breath Volume	1.875	23:04
Air Blank	0.000	23:05
Control Test	0.078	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 11/05/2023 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODDO,ISIAH,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ODDO,ISIAH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WALDEN

DATE/TIME OF TEST: 10/25/2023 20:47:22

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.080	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.073	20:42
Breath Volume	2.210	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.068	20:45
Breath Volume	1.417	20:45
Air Blank	0.000	20:46
Control Test	0.079	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 11/05/2023 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,DYLAN,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: HART,DYLAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 10/26/2023 01:57:26

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1 0.202	01:52	
Breath Volume	4.007	01:52	
Air Blank	0.000	01:53	
Air Blank	0.000	01:55	
Subject Sample	# 2	01:55	Slope Not Met
Breath Volume	1.207	01:55	
Air Blank	0.000	01:56	
Control Test	0.080	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 11/05/2023 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,DYLAN,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: HART,DYLAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 10/26/2023 02:07:13

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.172	02:01
Breath Volume	1.593	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.190	02:05
Breath Volume	3.375	02:05
Air Blank	0.000	02:06
Control Test	0.078	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 11/05/2023 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 11:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:53

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 10/29/2023 13:20:21

Diagnostics Check	OK	13:13
Air Blank	0.000	13:14
Control Test	0.081	13:14
Air Blank	0.000	13:15
Subject Sample # 1	0.171	13:15
Breath Volume	1.589	13:15
Air Blank	0.000	13:16
Air Blank	0.000	13:18
Subject Sample # 2	0.176	13:18
Breath Volume	1.394	13:18
Air Blank	0.000	13:19
Control Test	0.078	13:19
Air Blank	0.000	13:20
Diagnostics Check	OK	13:20

Data Download Date/Time: 11/05/2023 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLAUGHTER,JASON,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SLAUGHTER,JASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: ELLSWORTH

DATE/TIME OF TEST: 10/29/2023 19:22:36

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.082	19:16
Air Blank	0.000	19:16
Subject Sample # 1	0.158	19:18
Breath Volume	1.757	19:18
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.145	19:20
Breath Volume	1.507	19:20
Air Blank	0.000	19:21
Control Test	0.078	19:22
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

Data Download Date/Time: 11/05/2023 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSKO,KRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: KENNERLEY

DATE/TIME OF TEST: 10/01/2023 14:39:35

Diagnostics Check	OK	14:32
Air Blank	0.000	14:33
Control Test	0.076	14:33
Air Blank	0.000	14:34
Subject Sample # 1	0.194	14:35
Breath Volume	1.605	14:35
Air Blank	0.000	14:35
Air Blank	0.000	14:37
Subject Sample # 2	0.191	14:38
Breath Volume	1.429	14:38
Air Blank	0.000	14:38
Control Test	0.075	14:38
Air Blank	0.000	14:39
Diagnostics Check	OK	14:39

Data Download Date/Time: 10/24/2023 17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRBY, CHAD, W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACLAREN, DEREK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: SILAS

DATE/TIME OF TEST: 10/03/2023 19:06:44

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.077	19:01
Air Blank	0.000	19:01
Subject Sample # 1	0.000	19:02
Breath Volume	2.968	19:02
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.000	19:05
Breath Volume	3.050	19:05
Air Blank	0.000	19:05
Control Test	0.077	19:06
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 10/24/2023 17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: JACKSON 3

DATE/TIME OF TEST: 10/06/2023 02:35:55

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.077	02:31	
Air Blank	0.000	02:31	
Subject Sample	# 1 0.158	02:32	
Breath Volume	1.855	02:32	
Air Blank	0.000	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 2 0.150	02:35	
Breath Volume	2.847	02:35	
Air Blank		02:35	RFI Detect
Air Blank	0.000	02:35	

Data Download Date/Time: 10/24/2023 17:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FARIS

DATE/TIME OF TEST: 10/14/2023 02:34:13

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.201	02:26
Breath Volume	2.328	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.228	02:29
Breath Volume	3.859	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 3	0.220	02:32
Breath Volume	2.046	02:32
Air Blank	0.000	02:33
Control Test	0.075	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 10/24/2023 17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HANSEN,ERIC
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: ARONSON

DATE/TIME OF TEST: 10/15/2023 03:56:42

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.078	03:45	
Air Blank	0.000	03:46	
Subject Sample	# 1 0.154	03:49	Volume Not Met
Breath Volume	0.035	03:49	
Air Blank	0.000	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 2 0.165	03:55	Volume Not Met
Breath Volume	0.437	03:55	
Air Blank	0.000	03:55	
Control Test	0.073	03:56	Control Outside Tolerance
Air Blank	0.000	03:56	

Data Download Date/Time: 10/24/2023 17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: POLVERE

DATE/TIME OF TEST: 10/17/2023 15:23:59

Diagnostics Check	OK	15:23	
Air Blank		15:23	RFI Detect
Air Blank	0.000	15:23	

Data Download Date/Time: 10/24/2023 17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: POLVERE

DATE/TIME OF TEST: 10/17/2023 15:27:03

Diagnostics Check	OK	15:25	
Air Blank	0.000	15:26	
Control Test	0.078	15:26	
Air Blank		15:26	RFI Detect
Air Blank	0.000	15:27	

Data Download Date/Time: 10/24/2023 17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: POLVERE

DATE/TIME OF TEST: 10/17/2023 15:34:58

Diagnostics Check	OK	15:29	
Air Blank	0.000	15:29	
Control Test	0.079	15:29	
Air Blank	0.000	15:30	
Subject Sample	# 1 0.355	15:30	
Breath Volume	1.843	15:30	
Air Blank	0.000	15:31	
Air Blank	0.000	15:33	
Subject Sample	# 2 0.340	15:33	
Breath Volume	1.746	15:33	
Air Blank	0.000	15:34	
		15:34	RFI Detect
Air Blank	0.000	15:34	

Data Download Date/Time: 10/24/2023 17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRO,MARCEL,
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PLEMMONS,LAQUINTA
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: AVERY

DATE/TIME OF TEST: 10/17/2023 18:16:17

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.078	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.000	18:11
Breath Volume	1.839	18:11
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.000	18:14
Breath Volume	2.246	18:14
Air Blank	0.000	18:15
Control Test	0.078	18:15
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 10/24/2023 17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: RIVERA CASTILLO

DATE/TIME OF TEST: 10/19/2023 23:17:21

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.077	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.151	23:12
Breath Volume	1.410	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:14
Subject Sample # 2	0.141	23:15
Breath Volume	1.105	23:15
Air Blank	0.000	23:16
Control Test	0.077	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 10/24/2023 17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: CROOKER

DATE/TIME OF TEST: 10/28/2023 10:29:14

Diagnostics Check	OK	10:23	
Air Blank	0.000	10:23	
Control Test	0.077	10:23	
Air Blank	0.000	10:24	
Subject Sample # 1	0.132	10:24	
Breath Volume	1.792	10:24	
Air Blank	0.000	10:25	
Air Blank	0.000	10:27	
Subject Sample # 2	0.135	10:27	
Breath Volume	1.820	10:27	
Air Blank	0.000	10:28	
Control Test	0.074	10:28	Control Outside Tolerance
Air Blank	0.000	10:29	

Data Download Date/Time: 11/29/2023 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSKO,KRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DICKS

DATE/TIME OF TEST: 10/29/2023 03:28:58

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.077	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.110	03:24
Breath Volume	1.476	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.107	03:27
Breath Volume	1.488	03:27
Air Blank	0.000	03:28
Control Test	0.077	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 11/29/2023 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIXON, WESLEY, B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DIXON, WESLEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/14/2023 00:48:16

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.078	00:42	
Air Blank	0.000	00:43	
Subject Sample # 1	0.162	00:43	
Breath Volume	3.460	00:43	
Air Blank	0.000	00:44	
Air Blank	0.000	00:46	
Subject Sample # 2	0.156	00:46	
Breath Volume	3.257	00:46	
Air Blank	0.000	00:47	
Control Test	0.074	00:47	Control Outside Tolerance
Air Blank	0.000	00:48	

Data Download Date/Time: 11/14/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 06:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: MASSINGILL

DATE/TIME OF TEST: 10/01/2023 07:33:15

Diagnostics Check	OK	07:26
Air Blank	0.000	07:27
Control Test	0.082	07:27
Air Blank	0.000	07:27
Subject Sample # 1	0.099	07:28
Breath Volume	2.175	07:28
Air Blank	0.000	07:29
Air Blank	0.000	07:30
Subject Sample # 2	0.094	07:31
Breath Volume	2.593	07:31
Air Blank	0.000	07:32
Control Test	0.081	07:32
Air Blank	0.000	07:33
Diagnostics Check	OK	07:33

Data Download Date/Time: 10/18/2023 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 10/04/2023 05:57:18

Diagnostics Check	OK	05:48
Air Blank	0.000	05:49
Control Test	0.082	05:49
Air Blank	0.000	05:50
Subject Sample # 1	0.086	05:52
Breath Volume	2.027	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:54
Subject Sample # 2	0.077	05:55
Breath Volume	2.324	05:55
Air Blank	0.000	05:56
Control Test	0.080	05:56
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 10/18/2023 01:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: GREENFIELD

DATE/TIME OF TEST: 10/07/2023 21:22:08

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:19	
Control Test	0.083	21:19	
Air Blank	0.000	21:19	
Subject Sample	# 1	21:20	Subject Test Refused
Breath Volume	0.000	21:20	
Air Blank	0.000	21:21	
Control Test	0.082	21:21	
Air Blank	0.000	21:22	
Diagnostics Check	OK	21:22	

Data Download Date/Time: 10/18/2023 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 10/14/2023 02:12:39

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.083	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.193	02:07
Breath Volume	2.351	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.195	02:10
Breath Volume	2.730	02:10
Air Blank	0.000	02:11
Control Test	0.081	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 10/18/2023 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/17/2023 21:30:21

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.083	21:24
Air Blank	0.000	21:24
Subject Sample # 1	0.181	21:25
Breath Volume	2.367	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:27
Subject Sample # 2	0.174	21:28
Breath Volume	1.871	21:28
Air Blank	0.000	21:29
Control Test	0.080	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 10/18/2023 01:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MELTON,EMILY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 10/22/2023 04:22:38

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.083	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.119	04:14
Breath Volume	1.167	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample # 2	0.143	04:17
Breath Volume	1.937	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:19
Subject Sample # 3	0.139	04:20
Breath Volume	2.738	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 11/19/2023 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: PACKARD

DATE/TIME OF TEST: 10/28/2023 02:38:46

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.082	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.196	02:33
Breath Volume	2.214	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.204	02:36
Breath Volume	3.574	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 11/19/2023 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: QUINTON

DATE/TIME OF TEST: 10/29/2023 00:15:26

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.082	00:13	
Air Blank	0.000	00:13	
Subject Sample	# 1	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:14	
Control Test	0.083	00:14	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 11/19/2023 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RADCLIFF, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 10/30/2023 18:47:12

Diagnostics Check	OK	18:40
Air Blank	0.000	18:41
Control Test	0.081	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.168	18:42
Breath Volume	1.386	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:44
Subject Sample # 2	0.176	18:45
Breath Volume	2.183	18:45
Air Blank	0.000	18:46
Control Test	0.081	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 11/19/2023 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDY,JOE,E
BREATH TEST OPERATOR AGENCY: Holmes County SO
ARRESTING OFFICER: SHIRAH,MIKAELA
ARRESTING OFFICER AGENCY: Holmes County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: POLAND

DATE/TIME OF TEST: 10/01/2023 02:47:11

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.192	02:42
Breath Volume	1.320	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.193	02:45
Breath Volume	1.875	02:45
Air Blank	0.000	02:46
Control Test	0.082	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 10/23/2023 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 2402328118
BREATH TEST OPERATOR: HARDY,JOE,E	DRY GAS EXPIRY DATE: 01/20/2025
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: RECORD,GEORGE	ARREST TIME: 02:26
ARRESTING OFFICER AGENCY: Bonifay PD	GENDER: Female
LAST AGENCY INSPECTION: 09/27/2023	OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: REYNOLDS	DATE/TIME OF TEST: 10/14/2023 02:56:12
-------------------------------	--

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.274	02:51
Breath Volume	1.320	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 2	0.254	02:54
Breath Volume	2.648	02:54
Air Blank	0.000	02:55
Control Test	0.082	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 10/23/2023 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WELCH,WAYLON
ARRESTING OFFICER AGENCY: Holmes County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:46

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 10/16/2023 14:13:49

Diagnostics Check	OK	14:06
Air Blank	0.000	14:06
Control Test	0.080	14:07
Air Blank	0.000	14:07
Subject Sample # 1	0.248	14:08
Breath Volume	2.523	14:08
Air Blank	0.000	14:09
Air Blank	0.000	14:10
Subject Sample # 2	0.241	14:12
Breath Volume	3.777	14:12
Air Blank	0.000	14:12
Control Test	0.078	14:13
Air Blank	0.000	14:13
Diagnostics Check	OK	14:13

Data Download Date/Time: 10/23/2023 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Holmes County SO
ARRESTING OFFICER: SHIRAH,MIKAYLA
ARRESTING OFFICER AGENCY: Holmes County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: RICHMOND

DATE/TIME OF TEST: 10/28/2023 00:56:13

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:51	No Sample Provided
Breath Volume	0.000	00:51	
Air Blank	0.000	00:51	
Air Blank	0.000	00:53	
Subject Sample	# 2	00:55	Subject Test Refused
Breath Volume	0.000	00:55	
Air Blank		00:55	RFI Detect
Air Blank	0.000	00:56	

Data Download Date/Time: 11/29/2023 00:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Holmes County SO
ARRESTING OFFICER: SHIRAH,MIKAELA
ARRESTING OFFICER AGENCY: Holmes County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: RICHMOND

DATE/TIME OF TEST: 10/28/2023 00:50:47

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.082	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:45	No Sample Provided
Breath Volume	0.000	00:45	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 2	00:49	Subject Test Refused
Breath Volume	0.000	00:49	
Air Blank	0.000	00:49	
Control Test	0.080	00:50	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 11/29/2023 00:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: STUCKEY

DATE/TIME OF TEST: 10/15/2023 00:16:49

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.110	00:12
Breath Volume	3.761	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.108	00:15
Breath Volume	3.800	00:15
Air Blank	0.000	00:15
Control Test	0.081	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 12/10/2023 23:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,ERIC,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: FLORES,PEARL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: KREJCI

DATE/TIME OF TEST: 10/01/2023 18:32:23

Diagnostics Check	OK	18:24
Air Blank	0.000	18:24
Control Test	0.082	18:24
Air Blank	0.000	18:25
Subject Sample # 1	0.187	18:27
Breath Volume	1.152	18:27
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.189	18:30
Breath Volume	1.550	18:30
Air Blank	0.000	18:31
Control Test	0.080	18:31
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBA,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WALKER,RONALD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: NESTOR

DATE/TIME OF TEST: 10/02/2023 04:25:25

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.082	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.184	04:21
Breath Volume	1.738	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.182	04:23
Breath Volume	2.425	04:23
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ANZALONE,JEFFREY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 10/05/2023 22:57:18

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.082	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.171	22:52
Breath Volume	1.507	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.174	22:55
Breath Volume	2.007	22:55
Air Blank	0.000	22:56
Control Test	0.081	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBA,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ALBA,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: FRAZER

DATE/TIME OF TEST: 10/06/2023 02:09:57

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.082	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.170	02:05
Breath Volume	1.183	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.190	02:08
Breath Volume	2.726	02:08
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: PARRETT,AUSTIN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: VELASQUEZ MENDEZ

DATE/TIME OF TEST: 10/09/2023 22:46:28

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.082	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.250	22:40
Breath Volume	1.140	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.237	22:44
Breath Volume	2.578	22:44
Air Blank	0.000	22:45
Control Test	0.081	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: SANCHEZ RIVAS

DATE/TIME OF TEST: 10/10/2023 04:55:15

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.082	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.193	04:50
Breath Volume	2.851	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.186	04:53
Breath Volume	2.988	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: DIODATO

DATE/TIME OF TEST: 10/10/2023 21:58:12

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.082	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.000	21:53
Breath Volume	1.781	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.000	21:56
Breath Volume	3.109	21:56
Air Blank	0.000	21:57
Control Test	0.081	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAZIANO,STEVEN,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GRAZIANO,STEVEN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: PASCUAL

DATE/TIME OF TEST: 10/11/2023 19:10:47

Diagnostics Check	OK	19:09	
Air Blank	0.000	19:09	
Control Test	0.082	19:10	Interferent Detect
Air Blank	0.000	19:10	

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAZIANO,STEVEN,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GRAZIANO,STEVEN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: PASCUAL

DATE/TIME OF TEST: 10/11/2023 19:17:14

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:16	
Control Test	0.081	19:16	Interferent Detect
Air Blank	0.000	19:17	

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAZIANO,STEVEN,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GRAZIANO,STEVEN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: P240000834560

DATE/TIME OF TEST: 10/11/2023 19:25:58

Diagnostics Check	OK	19:24	
Air Blank	0.000	19:25	
Control Test	0.081	19:25	Interferent Detect
Air Blank	0.000	19:25	

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAZIANO,STEVEN,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GRAZIANO,STEVEN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: PASCUAL

DATE/TIME OF TEST: 10/11/2023 19:49:35

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:43	
Control Test	0.082	19:43	
Air Blank	0.000	19:44	
Subject Sample # 1	0.242	19:45	
Breath Volume	2.246	19:45	
Air Blank	0.000	19:46	
Air Blank	0.000	19:47	
Subject Sample # 2	0.235	19:48	
Breath Volume	1.929	19:48	
Air Blank	0.000	19:48	
Control Test	0.078	19:49	Interferent Detect
Air Blank	0.000	19:49	

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: VALENTIN,CARLOS
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: DAMICO

DATE/TIME OF TEST: 10/14/2023 04:47:30

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.131	04:42
Breath Volume	2.679	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.130	04:45
Breath Volume	2.593	04:45
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBA,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ALBA,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: FOLK

DATE/TIME OF TEST: 10/14/2023 21:10:16

Diagnostics Check	OK	21:03
Air Blank	0.000	21:04
Control Test	0.082	21:04
Air Blank	0.000	21:05
Subject Sample # 1	0.185	21:05
Breath Volume	2.933	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.198	21:08
Breath Volume	3.394	21:08
Air Blank	0.000	21:09
Control Test	0.081	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MARTIN,NICHOLAS
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/15/2023 05:44:55

Diagnostics Check	OK	05:38
Air Blank	0.000	05:39
Control Test	0.082	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.190	05:40
Breath Volume	2.640	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:42
Subject Sample # 2	0.195	05:43
Breath Volume	2.097	05:43
Air Blank	0.000	05:44
Control Test	0.082	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITES,DEJAS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: COLGAN,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 10/17/2023 03:00:51

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.083	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.037	02:56
Breath Volume	1.898	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:58
Subject Sample # 2	0.036	02:59
Breath Volume	2.031	02:59
Air Blank	0.000	02:59
Control Test	0.082	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAKE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GIBBONS,ROBERT
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 10/17/2023 22:28:02

Diagnostics Check	OK	22:13	
Air Blank	0.000	22:14	
Control Test	0.082	22:14	
Air Blank	0.000	22:14	
Subject Sample	# 1 0.243	22:17	
Breath Volume	2.031	22:17	
Air Blank	0.000	22:18	
Air Blank	0.000	22:19	
Subject Sample	# 2 0.255	22:23	Volume Not Met
Breath Volume	0.390	22:23	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample	# 3 0.257	22:26	
Breath Volume	1.492	22:26	
Air Blank	0.000	22:27	
Control Test	0.079	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABARBERA,SUMMER,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TOMCZYK,STENLY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: HRONIS

DATE/TIME OF TEST: 10/21/2023 19:17:03

Diagnostics Check	OK	19:09	
Air Blank	0.000	19:10	
Control Test	0.082	19:10	
Air Blank	0.000	19:11	
Subject Sample # 1	0.240	19:13	
Breath Volume	2.742	19:13	
Air Blank	0.000	19:14	
Air Blank		19:15	Improper Sample
Air Blank	0.000	19:16	
Control Test	0.083	19:16	
Air Blank	0.000	19:16	
Diagnostics Check	OK	19:17	

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABARBERA,SUMMER,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TOMCZYK,STENLY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: HRONIS

DATE/TIME OF TEST: 10/21/2023 19:28:30

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:19	
Control Test	0.081	19:20	
Air Blank	0.000	19:20	
Subject Sample # 1	0.221	19:23	
Breath Volume	2.050	19:23	
Air Blank	0.000	19:23	
Air Blank	0.000	19:25	
Subject Sample # 2		19:28	Sequence Aborted
Breath Volume	0.000	19:28	
Air Blank	0.000	19:28	

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,CASANDRA,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: SOUTHER,AUSTIN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 10/22/2023 03:02:40

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.082	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.157	02:58
Breath Volume	1.167	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.163	03:01
Breath Volume	2.597	03:01
Air Blank	0.000	03:01
Control Test	0.082	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 10/26/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINK,ZACHARY,I
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FINK,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MILONE

DATE/TIME OF TEST: 10/28/2023 22:07:51

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.082	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.206	22:03
Breath Volume	1.765	22:03
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.205	22:06
Breath Volume	1.539	22:06
Air Blank	0.000	22:06
Control Test	0.082	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 11/29/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LOUIS,RODNESHIA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 10/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SCHUMANN

DATE/TIME OF TEST: 10/29/2023 23:55:46

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.082	23:45	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:48	No Sample Provided
Breath Volume	0.000	23:48	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample	# 2	23:54	No Sample Provided
Breath Volume	0.000	23:54	
Air Blank	0.000	23:54	
Control Test	0.081	23:55	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 11/29/2023 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LOUIS,RODNESHIA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 10/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SCHUMANN

DATE/TIME OF TEST: 10/30/2023 00:09:56

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:58	
Control Test	0.082	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.123	00:02	Volume Not Met
Breath Volume	0.125	00:02	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample	# 2 0.130	00:08	Volume Not Met
Breath Volume	0.437	00:08	
Air Blank	0.000	00:09	
Control Test	0.080	00:09	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 11/29/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEW,REECE,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DEW,REECE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 14:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 10/12/2023 17:05:36

Diagnostics Check	OK	16:59
Air Blank	0.000	16:59
Control Test	0.081	17:00
Air Blank	0.000	17:00
Subject Sample # 1	0.187	17:01
Breath Volume	1.871	17:01
Air Blank	0.000	17:01
Air Blank	0.000	17:03
Subject Sample # 2	0.192	17:04
Breath Volume	2.292	17:04
Air Blank	0.000	17:04
Control Test	0.081	17:05
Air Blank	0.000	17:05
Diagnostics Check	OK	17:05

Data Download Date/Time: 11/28/2023 19:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/24/2023 00:16:47

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.082	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.000	00:12
Breath Volume	2.058	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.000	00:15
Breath Volume	2.105	00:15
Air Blank	0.000	00:15
Control Test	0.082	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 10/30/2023 02:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,LANDON,R
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: BURNETTE,TYLER
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: ROSENDALE

DATE/TIME OF TEST: 10/27/2023 20:47:36

Diagnostics Check	OK	20:41	
Air Blank	0.000	20:41	
Control Test	0.082	20:42	
Air Blank	0.000	20:42	
Subject Sample	# 1 0.290	20:43	
Breath Volume	1.210	20:43	
Air Blank	0.000	20:43	
Air Blank	0.000	20:45	
Subject Sample	# 2	20:46	Slope Not Met
Breath Volume	1.195	20:46	
Air Blank	0.000	20:46	
Control Test	0.081	20:46	
Air Blank	0.000	20:47	
Diagnostics Check	OK	20:47	

Data Download Date/Time: 10/30/2023 02:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,LANDON,R
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: BURNETTE,TYLER
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: ROSENDALE

DATE/TIME OF TEST: 10/27/2023 20:56:59

Diagnostics Check	OK	20:50
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.273	20:52
Breath Volume	1.121	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.281	20:55
Breath Volume	1.304	20:55
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 10/30/2023 02:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEASEL,KELSEY,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: BARKER,ROBERT
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CHAVARRIA

DATE/TIME OF TEST: 10/26/2023 00:26:58

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.080	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1 0.133	00:24	Volume Not Met
Breath Volume	0.097	00:24	
Air Blank		00:25	Improper Sample
Air Blank	0.000	00:26	
Control Test	0.078	00:26	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

Data Download Date/Time: 11/27/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,BRITTANY,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ECK,ENIAH
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: DEBLIECK

DATE/TIME OF TEST: 10/27/2023 20:22:43

Diagnostics Check	OK	20:13
Air Blank	0.000	20:14
Control Test	0.081	20:14
Air Blank	0.000	20:15
Subject Sample # 1	0.304	20:18
Breath Volume	1.109	20:18
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.292	20:20
Breath Volume	1.503	20:20
Air Blank	0.000	20:21
Control Test	0.078	20:22
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 11/27/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,V,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SIMONSEN,E
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 10/28/2023 03:30:31

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.148	03:26
Breath Volume	1.484	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.149	03:28
Breath Volume	1.484	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 11/27/2023 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 10/29/2023 03:40:42

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.223	03:36
Breath Volume	2.187	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.220	03:39
Breath Volume	1.503	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 11/27/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SIMONSEN,ERIK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 10/29/2023 03:52:33

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.190	03:47
Breath Volume	3.957	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.193	03:50
Breath Volume	3.437	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 11/27/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PATTERSON,HANNA
ARRESTING OFFICER AGENCY: Dunnellon PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 10/29/2023 05:07:48

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.080	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.168	05:03
Breath Volume	2.746	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:05
Subject Sample # 2	0.156	05:06
Breath Volume	1.308	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 11/27/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COUNTS,MELISSA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COUNTS,MELISSA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: RUSHING

DATE/TIME OF TEST: 10/30/2023 16:45:21

Diagnostics Check	OK	16:34
Air Blank	0.000	16:35
Control Test	0.080	16:35
Air Blank	0.000	16:35
Subject Sample # 1	0.177	16:38
Breath Volume	1.546	16:38
Air Blank	0.000	16:39
Air Blank	0.000	16:41
Subject Sample # 2	0.169	16:43
Breath Volume	1.109	16:43
Air Blank	0.000	16:44
Control Test	0.080	16:44
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 11/27/2023 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SILVERSTONE

DATE/TIME OF TEST: 10/01/2023 01:21:46

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.079	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.019	01:17
Breath Volume	1.535	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.016	01:20
Breath Volume	1.320	01:20
Air Blank	0.000	01:20
Control Test	0.078	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: KOEHLER

DATE/TIME OF TEST: 10/01/2023 03:06:56

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.123	03:02
Breath Volume	1.832	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.118	03:05
Breath Volume	1.527	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: CARLISLE

DATE/TIME OF TEST: 10/01/2023 20:42:18

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:39	
Control Test	0.080	20:40	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:40	Subject Test Refused
Breath Volume	0.000	20:40	
Air Blank	0.000	20:41	
Control Test	0.080	20:41	
Air Blank	0.000	20:42	
Diagnostics Check	OK	20:42	

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,R,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/03/2023 02:40:35

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1 0.174	02:33	
Breath Volume	1.546	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 2	02:38	Subject Test Refused
Breath Volume	0.187	02:38	
Air Blank	0.000	02:39	
Control Test	0.079	02:39	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 10/03/2023 15:55:40

Diagnostics Check	OK	15:47
Air Blank	0.000	15:47
Control Test	0.080	15:48
Air Blank	0.000	15:48
Subject Sample # 1	0.000	15:51
Breath Volume	2.515	15:51
Air Blank	0.000	15:51
Air Blank	0.000	15:53
Subject Sample # 2	0.000	15:54
Breath Volume	2.046	15:54
Air Blank	0.000	15:54
Control Test	0.080	15:55
Air Blank	0.000	15:55
Diagnostics Check	OK	15:55

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SIUZDDAK,C
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: GILLER

DATE/TIME OF TEST: 10/06/2023 17:36:44

Diagnostics Check	OK	17:33	
Air Blank	0.000	17:34	
Control Test	0.080	17:34	
Air Blank	0.000	17:34	
Subject Sample	# 1	17:35	Subject Test Refused
Breath Volume	0.000	17:35	
Air Blank	0.000	17:35	
Control Test	0.080	17:36	
Air Blank	0.000	17:36	
Diagnostics Check	OK	17:36	

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,JOHNATHAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BALCZUKIEWICZ

DATE/TIME OF TEST: 10/07/2023 03:07:57

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.080	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 1 0.138	03:00	Volume Not Met
Breath Volume	1.054	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 2 0.147	03:03	
Breath Volume	1.203	03:03	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 3 0.151	03:06	
Breath Volume	1.109	03:06	
Air Blank	0.000	03:07	
Control Test	0.079	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BROWNLOW

DATE/TIME OF TEST: 10/07/2023 19:36:37

Diagnostics Check	OK	19:30
Air Blank	0.000	19:30
Control Test	0.080	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.056	19:32
Breath Volume	1.984	19:32
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample # 2	0.056	19:35
Breath Volume	1.714	19:35
Air Blank	0.000	19:35
Control Test	0.080	19:36
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ORTIZ,B
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/08/2023 04:07:41

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.214	04:02
Breath Volume	2.531	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.214	04:06
Breath Volume	1.152	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: COCCHI

DATE/TIME OF TEST: 10/09/2023 19:18:05

Diagnostics Check	OK	19:10
Air Blank	0.000	19:11
Control Test	0.079	19:11
Air Blank	0.000	19:11
Subject Sample # 1	0.181	19:13
Breath Volume	1.265	19:13
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample # 2	0.174	19:16
Breath Volume	1.132	19:16
Air Blank	0.000	19:17
Control Test	0.078	19:17
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VESPER,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: SIEVERS

DATE/TIME OF TEST: 10/11/2023 20:17:33

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:14	
Control Test	0.080	20:15	
Air Blank	0.000	20:15	
Subject Sample	# 1	20:16	Subject Test Refused
Breath Volume	0.000	20:16	
Air Blank	0.000	20:16	
Control Test	0.079	20:16	
Air Blank	0.000	20:17	
Diagnostics Check	OK	20:17	

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SLUDIAK,C
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: G125013825090

DATE/TIME OF TEST: 10/11/2023 23:12:05

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.205	23:07
Breath Volume	1.628	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.195	23:10
Breath Volume	1.437	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VESPER,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/13/2023 03:50:50

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.079	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:45	No Sample Provided
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Air Blank	0.000	03:47	
Subject Sample	# 2	03:49	Subject Test Refused
Breath Volume	0.515	03:49	
Air Blank	0.000	03:49	
Control Test	0.079	03:50	
Air Blank	0.000	03:50	
Diagnostics Check	OK	03:50	

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MARCUS,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 10/15/2023 02:49:10

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.080	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:48	
Control Test	0.078	02:48	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VESPER,MITCHELL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 10/16/2023 22:16:03

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.077	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.108	22:11
Breath Volume	1.343	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:13
Subject Sample # 2	0.111	22:14
Breath Volume	1.117	22:14
Air Blank	0.000	22:15
Control Test	0.077	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:16

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VESPER,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 10/21/2023 21:27:03

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.080	21:21
Air Blank	0.000	21:21
Subject Sample # 1	0.209	21:22
Breath Volume	1.988	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:24
Subject Sample # 2	0.204	21:25
Breath Volume	2.031	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:27

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: WELSH

DATE/TIME OF TEST: 10/22/2023 02:40:21

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:37	
Control Test	0.078	02:37	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:38	Subject Test Refused
Breath Volume	0.000	02:38	
Air Blank	0.000	02:39	
Control Test	0.079	02:39	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VESPER,MITCHELL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: OVERBEY

DATE/TIME OF TEST: 10/22/2023 22:41:43

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:39	
Control Test	0.080	22:39	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:40	
Control Test	0.079	22:41	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 10/30/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 10/23/2023 20:47:50

Diagnostics Check	OK	20:36	
Air Blank	0.000	20:36	
Control Test	0.080	20:36	
Air Blank	0.000	20:37	
Subject Sample	# 1 0.166	20:40	Volume Not Met
Breath Volume	0.796	20:40	
Air Blank	0.000	20:41	
Air Blank	0.000	20:43	
Subject Sample	# 2 0.158	20:46	Volume Not Met
Breath Volume	0.777	20:46	
Air Blank	0.000	20:46	
Control Test	0.079	20:47	
Air Blank	0.000	20:47	
Diagnostics Check	OK	20:47	

Data Download Date/Time: 10/30/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: REID,H
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: ELDREDGE

DATE/TIME OF TEST: 10/25/2023 20:02:16

Diagnostics Check	OK	19:59	
Air Blank	0.000	19:59	
Control Test	0.080	20:00	
Air Blank	0.000	20:00	
Subject Sample	# 1	20:00	Subject Test Refused
Breath Volume	0.000	20:00	
Air Blank	0.000	20:01	
Control Test	0.080	20:01	
Air Blank	0.000	20:02	
Diagnostics Check	OK	20:02	

Data Download Date/Time: 10/30/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: ZAWADA

DATE/TIME OF TEST: 10/26/2023 20:00:32

Diagnostics Check	OK	19:54
Air Blank	0.000	19:54
Control Test	0.080	19:54
Air Blank	0.000	19:55
Subject Sample # 1	0.160	19:56
Breath Volume	1.738	19:56
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.161	19:58
Breath Volume	1.679	19:58
Air Blank	0.000	19:59
Control Test	0.079	19:59
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 10/30/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MARCUS,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: LECHLAK

DATE/TIME OF TEST: 10/28/2023 00:40:16

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.080	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 10/30/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MARCUS,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/29/2023 00:37:53

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.079	00:35	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:36	
Control Test	0.079	00:37	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 10/30/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREZKOP,ZACHARY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FRAYER,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: COCHRANE

DATE/TIME OF TEST: 10/29/2023 07:11:19

Diagnostics Check	OK	07:04
Air Blank	0.000	07:04
Control Test	0.079	07:04
Air Blank	0.000	07:05
Subject Sample # 1	0.082	07:06
Breath Volume	1.800	07:06
Air Blank	0.000	07:07
Air Blank	0.000	07:09
Subject Sample # 2	0.082	07:09
Breath Volume	2.523	07:09
Air Blank	0.000	07:10
Control Test	0.078	07:10
Air Blank	0.000	07:11
Diagnostics Check	OK	07:11

Data Download Date/Time: 10/30/2023 20:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREZKOP,ZACHARY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COOPER,W
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:16

SUBJECT'S LAST NAME: ALFORD

DATE/TIME OF TEST: 10/29/2023 09:39:43

Diagnostics Check	OK	09:36	
Air Blank	0.000	09:37	
Control Test	0.079	09:37	
Air Blank	0.000	09:38	
Subject Sample	# 1	09:38	Subject Test Refused
Breath Volume	0.000	09:38	
Air Blank	0.000	09:38	
Control Test	0.079	09:39	
Air Blank	0.000	09:39	
Diagnostics Check	OK	09:39	

Data Download Date/Time: 10/30/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: DARBY

DATE/TIME OF TEST: 10/29/2023 19:55:32

Diagnostics Check	OK	19:48
Air Blank	0.000	19:49
Control Test	0.079	19:49
Air Blank	0.000	19:49
Subject Sample # 1	0.000	19:50
Breath Volume	3.152	19:50
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.000	19:54
Breath Volume	3.097	19:54
Air Blank	0.000	19:54
Control Test	0.080	19:54
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 10/30/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,ROGER,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: MULLINGS

DATE/TIME OF TEST: 10/31/2023 00:08:40

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.080	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1 0.106	00:01	Volume Not Met
Breath Volume	0.339	00:01	
Air Blank	0.000	00:02	
Air Blank	0.000	00:03	
Subject Sample	# 2 0.118	00:07	Volume Not Met
Breath Volume	0.316	00:07	
Air Blank	0.000	00:07	
Control Test	0.078	00:08	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 11/28/2023 22:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN, DEMERIS, P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SAPP, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: KEEFE

DATE/TIME OF TEST: 10/02/2023 00:58:52

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.079	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.252	00:54
Breath Volume	1.605	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.237	00:57
Breath Volume	1.562	00:57
Air Blank	0.000	00:57
Control Test	0.078	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 10/26/2023 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GEOHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 10/02/2023 23:40:53

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.079	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.279	23:35
Breath Volume	1.121	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.291	23:39
Breath Volume	2.535	23:39
Air Blank	0.000	23:39
Control Test	0.079	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 10/26/2023 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ROSARIO,F
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:43

SUBJECT'S LAST NAME: BRYON

DATE/TIME OF TEST: 10/05/2023 13:11:24

Diagnostics Check	OK	13:03
Air Blank	0.000	13:04
Control Test	0.079	13:04
Air Blank	0.000	13:05
Subject Sample # 1	0.306	13:06
Breath Volume	1.675	13:06
Air Blank	0.000	13:06
Air Blank	0.000	13:08
Subject Sample # 2	0.303	13:09
Breath Volume	1.386	13:09
Air Blank	0.000	13:10
Control Test	0.079	13:10
Air Blank	0.000	13:11
Diagnostics Check	OK	13:11

Data Download Date/Time: 10/26/2023 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GUINANE,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: GUYETT

DATE/TIME OF TEST: 10/05/2023 17:32:38

Diagnostics Check	OK	17:25
Air Blank	0.000	17:26
Control Test	0.079	17:26
Air Blank	0.000	17:26
Subject Sample # 1	0.000	17:28
Breath Volume	1.734	17:28
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.000	17:31
Breath Volume	2.363	17:31
Air Blank	0.000	17:31
Control Test	0.079	17:32
Air Blank	0.000	17:32
Diagnostics Check	OK	17:32

Data Download Date/Time: 10/26/2023 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GEOHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 10/06/2023 19:40:45

Diagnostics Check	OK	19:33	
Air Blank	0.000	19:34	
Control Test	0.079	19:34	
Air Blank	0.000	19:35	
Subject Sample	# 1 0.327	19:36	
Breath Volume	1.308	19:36	
Air Blank	0.000	19:36	
Air Blank	0.000	19:38	
Subject Sample	# 2	19:39	Slope Not Met
Breath Volume	0.839	19:39	
Air Blank	0.000	19:39	
Control Test	0.078	19:40	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 10/26/2023 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GEOHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 10/06/2023 19:49:27

Diagnostics Check	OK	19:42
Air Blank	0.000	19:43
Control Test	0.079	19:43
Air Blank	0.000	19:43
Subject Sample # 1	0.319	19:44
Breath Volume	1.589	19:44
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.319	19:47
Breath Volume	1.941	19:47
Air Blank	0.000	19:48
Control Test	0.078	19:48
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 10/26/2023 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GEOHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: BEAMER

DATE/TIME OF TEST: 10/07/2023 20:05:12

Diagnostics Check	OK	19:58
Air Blank	0.000	19:58
Control Test	0.079	19:59
Air Blank	0.000	19:59
Subject Sample # 1	0.225	20:00
Breath Volume	2.136	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.232	20:03
Breath Volume	1.453	20:03
Air Blank	0.000	20:04
Control Test	0.078	20:04
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 10/26/2023 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SANTAMARIA,J
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: PHAN

DATE/TIME OF TEST: 10/08/2023 23:47:18

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.081	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.071	23:42
Breath Volume	1.621	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.066	23:45
Breath Volume	1.277	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 10/26/2023 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BRAVE,S
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: BRUNNER

DATE/TIME OF TEST: 10/12/2023 19:06:33

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:54	
Control Test	0.079	18:54	
Air Blank	0.000	18:54	
Subject Sample	# 1 0.000	18:55	
Breath Volume	1.250	18:55	
Air Blank	0.000	18:56	
Air Blank	0.000	18:58	
Subject Sample	# 2 0.000	19:01	Volume Not Met
Breath Volume	0.812	19:01	
Air Blank	0.000	19:01	
Air Blank	0.000	19:03	
Subject Sample	# 3 0.000	19:05	
Breath Volume	1.378	19:05	
Air Blank	0.000	19:05	
Control Test	0.079	19:05	
Air Blank	0.000	19:06	
Diagnostics Check	OK	19:06	

Data Download Date/Time: 10/26/2023 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GEORING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: BRANTLEY

DATE/TIME OF TEST: 10/13/2023 22:00:57

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.079	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.224	21:56
Breath Volume	3.699	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:58
Subject Sample # 2	0.216	21:59
Breath Volume	1.808	21:59
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 10/26/2023 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MALESKE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: SCHAUFLE

DATE/TIME OF TEST: 10/14/2023 01:52:25

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.313	01:47
Breath Volume	2.707	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.308	01:50
Breath Volume	2.527	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 10/26/2023 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: REILLY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 10/14/2023 22:15:42

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.190	22:07
Breath Volume	1.894	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.217	22:10
Breath Volume	1.695	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 3	0.206	22:14
Breath Volume	2.769	22:14
Air Blank	0.000	22:14
Control Test	0.079	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 10/26/2023 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MALESKE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: MAGUIRE

DATE/TIME OF TEST: 10/14/2023 22:27:47

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:25	
Control Test	0.079	22:25	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:26	
Control Test	0.080	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 10/26/2023 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: FROELICH,T
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 10/15/2023 02:58:58

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.000	02:54
Breath Volume	1.351	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.000	02:57
Breath Volume	1.628	02:57
Air Blank	0.000	02:58
Control Test	0.085	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 10/26/2023 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Wildwood PD
ARRESTING OFFICER: LEMANWICZ,K
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: CHILEL

DATE/TIME OF TEST: 10/15/2023 21:11:46

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:05	
Control Test	0.078	21:06	
Air Blank	0.000	21:06	
Subject Sample	# 1 0.249	21:07	
Breath Volume	1.328	21:07	
Air Blank	0.000	21:08	
Air Blank	0.000	21:10	
Subject Sample	# 2 0.249	21:10	
Breath Volume	1.375	21:10	
Air Blank		21:11	RFI Detect
Air Blank	0.000	21:11	

Data Download Date/Time: 10/26/2023 08:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REED,ERNEST,L

BREATH TEST OPERATOR AGENCY: Sumter County SO

ARRESTING OFFICER: SCHMIDT,ANDREW

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 221303

DRY GAS EXPIRY DATE: 01/05/2025

VIOLATION CODE: DUI

ARREST TIME: 14:01

GENDER: Female

OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: LEFLER

DATE/TIME OF TEST: 10/17/2023 15:03:19

Diagnostics Check	OK	15:00	
Air Blank	0.000	15:00	
Control Test	0.080	15:01	
Air Blank	0.000	15:01	
Subject Sample	# 1	15:01	Subject Test Refused
Breath Volume	0.000	15:01	
Air Blank	0.000	15:02	
Control Test	0.080	15:02	
Air Blank	0.000	15:03	
Diagnostics Check	OK	15:03	

Data Download Date/Time: 10/26/2023 08:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,MARCUS,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 10/20/2023 23:27:10

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:24	
Control Test	0.079	23:25	
Air Blank	0.000	23:25	
Subject Sample	# 1	23:25	Subject Test Refused
Breath Volume	0.000	23:25	
Air Blank	0.000	23:26	
Control Test	0.079	23:26	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time: 10/26/2023 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDZINSKI,BRANDON,E
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: COLQUITT

DATE/TIME OF TEST: 10/21/2023 16:02:26

Diagnostics Check	OK	15:56
Air Blank	0.000	15:56
Control Test	0.080	15:57
Air Blank	0.000	15:57
Subject Sample # 1	0.187	15:58
Breath Volume	2.476	15:58
Air Blank	0.000	15:58
Air Blank	0.000	16:00
Subject Sample # 2	0.184	16:00
Breath Volume	3.222	16:00
Air Blank	0.000	16:01
Control Test	0.080	16:01
Air Blank	0.000	16:02
Diagnostics Check	OK	16:02

Data Download Date/Time: 10/26/2023 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,MARCUS,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SANTAMARIA,J
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/22/2023 02:26:57

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.218	02:21
Breath Volume	1.585	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.207	02:25
Breath Volume	2.277	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 10/26/2023 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,MARCUS,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ALLEN,G
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/22/2023 03:27:25

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.184	03:22
Breath Volume	1.824	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.176	03:25
Breath Volume	1.738	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 10/26/2023 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,MARCUS,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HIGHTOWER,TUCKER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: CHRISTIANSEN

DATE/TIME OF TEST: 10/26/2023 22:52:05

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.080	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.194	22:46
Breath Volume	2.171	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.196	22:50
Breath Volume	2.371	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 11/21/2023 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LEMANOWICZ,K
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 10/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: LORENZ

DATE/TIME OF TEST: 10/28/2023 02:47:08

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.079	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 11/21/2023 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NORDLE,L
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: APPEGATE

DATE/TIME OF TEST: 10/29/2023 02:13:38

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.175	02:08
Breath Volume	1.371	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.186	02:12
Breath Volume	1.105	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 11/21/2023 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MALESKE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HERNANDEZ MORALES

DATE/TIME OF TEST: 10/29/2023 04:10:37

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.041	04:05
Breath Volume	1.125	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.039	04:09
Breath Volume	1.246	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 11/21/2023 16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SAPP,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 10/01/2023 04:58:05

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:55	
Control Test	0.081	04:56	
Air Blank	0.000	04:56	
Subject Sample	# 1	04:56	Subject Test Refused
Breath Volume	0.000	04:56	
Air Blank	0.000	04:57	
Control Test	0.081	04:57	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 10/23/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 10/05/2023 15:57:02

Diagnostics Check	OK	15:54	
Air Blank	0.000	15:54	
Control Test	0.080	15:54	
Air Blank	0.000	15:55	
Subject Sample	# 1	15:55	Subject Test Refused
Breath Volume	0.000	15:55	
Air Blank	0.000	15:56	
Control Test	0.080	15:56	
Air Blank	0.000	15:56	
Diagnostics Check	OK	15:57	

Data Download Date/Time: 10/23/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 10/08/2023 22:14:23

Diagnostics Check	OK	22:07
Air Blank	0.000	22:08
Control Test	0.080	22:08
Air Blank	0.000	22:08
Subject Sample # 1	0.266	22:09
Breath Volume	1.402	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:11
Subject Sample # 2	0.248	22:12
Breath Volume	1.253	22:12
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 10/23/2023 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TWITTY,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MORRISSEY

DATE/TIME OF TEST: 10/11/2023 03:16:10

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.094	03:11
Breath Volume	1.210	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.087	03:14
Breath Volume	1.292	03:14
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 10/23/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOVANOVA, MICHAEL, N
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: JOVANOVA, MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: MACHADO

DATE/TIME OF TEST: 10/12/2023 19:17:40

Diagnostics Check	OK	19:14	
Air Blank	0.000	19:15	
Control Test	0.081	19:15	
Air Blank	0.000	19:16	
Subject Sample	# 1	19:16	Subject Test Refused
Breath Volume	0.000	19:16	
Air Blank	0.000	19:16	
Control Test	0.080	19:17	
Air Blank	0.000	19:17	
Diagnostics Check	OK	19:17	

Data Download Date/Time: 10/23/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: MARTINEZ ROCHEZ

DATE/TIME OF TEST: 10/13/2023 19:50:18

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.080	19:43
Air Blank	0.000	19:44
Subject Sample # 1	0.265	19:45
Breath Volume	1.742	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.251	19:48
Breath Volume	1.863	19:48
Air Blank	0.000	19:49
Control Test	0.079	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 10/23/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOVANOVA, MICHAEL, N
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: JOVANOVA, MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: KEARNS

DATE/TIME OF TEST: 10/14/2023 00:56:19

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.000	00:51
Breath Volume	1.664	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.000	00:54
Breath Volume	1.210	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 10/23/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 10/14/2023 19:28:05

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:25	
Control Test	0.081	19:26	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:26	Subject Test Refused
Breath Volume	0.000	19:26	
Air Blank	0.000	19:27	
Control Test	0.080	19:27	
Air Blank	0.000	19:27	
Diagnostics Check	OK	19:28	

Data Download Date/Time: 10/23/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: SAPP,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 10/15/2023 03:49:36

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:47	
Control Test	0.081	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:48	
Control Test	0.080	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 10/23/2023 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: CADEGIANI

DATE/TIME OF TEST: 10/21/2023 23:21:41

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:19	
Control Test	0.081	23:19	
Air Blank	0.000	23:20	
Subject Sample	# 1	23:20	Subject Test Refused
Breath Volume	0.000	23:20	
Air Blank	0.000	23:20	
Control Test	0.080	23:21	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 10/23/2023 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: TOWNSLEY

DATE/TIME OF TEST: 10/26/2023 02:29:58

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.080	02:27	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:29	
Control Test	0.081	02:29	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 11/30/2023 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HANEY,M
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: LOGGINS

DATE/TIME OF TEST: 10/31/2023 20:12:57

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:10	
Control Test	0.081	20:10	
Air Blank	0.000	20:11	
Subject Sample	# 1	20:11	Subject Test Refused
Breath Volume	0.000	20:11	
Air Blank	0.000	20:12	
Control Test	0.080	20:12	
Air Blank	0.000	20:12	
Diagnostics Check	OK	20:12	

Data Download Date/Time: 11/30/2023 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BANNICK,COURTNEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CHILDS

DATE/TIME OF TEST: 10/03/2023 02:59:04

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.081	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:58	
Control Test	0.081	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DOUG,J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: NUNEZ RUIZ

DATE/TIME OF TEST: 10/04/2023 07:00:26

Diagnostics Check	OK	06:46	
Air Blank	0.000	06:47	
Control Test	0.080	06:47	
Air Blank	0.000	06:48	
Subject Sample	# 1 0.090	06:51	Volume Not Met
Breath Volume	0.035	06:51	
Air Blank	0.000	06:51	
Air Blank	0.000	06:53	
Subject Sample	# 2 0.084	06:54	
Breath Volume	1.648	06:54	
Air Blank	0.000	06:54	
Air Blank	0.000	06:56	
Subject Sample	# 3 0.080	06:58	
Breath Volume	1.421	06:58	
Air Blank	0.000	06:59	
Control Test	0.082	06:59	
Air Blank	0.000	07:00	
Diagnostics Check	OK	07:00	

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DOUG,J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: LAWLES,BILLY
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 10/04/2023 07:43:21

Diagnostics Check	OK	07:36
Air Blank	0.000	07:37
Control Test	0.080	07:37
Air Blank	0.000	07:38
Subject Sample # 1	0.103	07:38
Breath Volume	2.406	07:38
Air Blank	0.000	07:39
Air Blank	0.000	07:41
Subject Sample # 2	0.101	07:41
Breath Volume	2.476	07:41
Air Blank	0.000	07:42
Control Test	0.081	07:42
Air Blank	0.000	07:43
Diagnostics Check	OK	07:43

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LAWLES,BILLY
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: KOPELAKIS

DATE/TIME OF TEST: 10/05/2023 11:20:39

Diagnostics Check	OK	11:14
Air Blank	0.000	11:14
Control Test	0.080	11:15
Air Blank	0.000	11:15
Subject Sample # 1	0.243	11:16
Breath Volume	2.308	11:16
Air Blank	0.000	11:16
Air Blank	0.000	11:18
Subject Sample # 2	0.240	11:19
Breath Volume	2.929	11:19
Air Blank	0.000	11:19
Control Test	0.080	11:20
Air Blank	0.000	11:20
Diagnostics Check	OK	11:20

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: RODRIGUEZ,E
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: ROSARIO JR

DATE/TIME OF TEST: 10/06/2023 21:17:43

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.081	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.141	21:11
Breath Volume	1.468	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.135	21:16
Breath Volume	1.140	21:16
Air Blank	0.000	21:16
Control Test	0.083	21:17
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: ROOMSBURG,R
ARRESTING OFFICER AGENCY: Umatilla PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: BARNARD

DATE/TIME OF TEST: 10/06/2023 23:58:32

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.081	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:57	
Control Test	0.080	23:57	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: BULTRON,CARLOS
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: LIMON

DATE/TIME OF TEST: 10/07/2023 05:07:16

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:52	
Control Test	0.081	04:52	
Air Blank	0.000	04:52	
Subject Sample	# 1 0.239	04:54	
Breath Volume	1.632	04:54	
Air Blank	0.000	04:55	
Air Blank	0.000	04:57	
Subject Sample	# 2	05:00	No Sample Provided
Breath Volume	0.000	05:00	
Air Blank	0.000	05:00	
Air Blank	0.000	05:02	
Subject Sample	# 3	05:05	No Sample Provided
Breath Volume	0.000	05:05	
Air Blank	0.000	05:06	
Control Test	0.080	05:06	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,BROOKE,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: MASON,BROOKE
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:11

SUBJECT'S LAST NAME: ROBLERO ROBLERO

DATE/TIME OF TEST: 10/07/2023 09:38:11

Diagnostics Check	OK	09:31
Air Blank	0.000	09:32
Control Test	0.080	09:32
Air Blank	0.000	09:33
Subject Sample # 1	0.251	09:33
Breath Volume	3.062	09:33
Air Blank	0.000	09:34
Air Blank	0.000	09:36
Subject Sample # 2	0.240	09:36
Breath Volume	3.441	09:36
Air Blank	0.000	09:37
Control Test	0.081	09:37
Air Blank	0.000	09:38
Diagnostics Check	OK	09:38

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: ALLEN,BREANNA
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: VERNON

DATE/TIME OF TEST: 10/07/2023 23:40:04

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:29	
Control Test	0.081	23:29	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:33	No Sample Provided
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Air Blank	0.000	23:35	
Subject Sample	# 2	23:38	No Sample Provided
Breath Volume	0.000	23:38	
Air Blank	0.000	23:39	
Control Test	0.080	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:40	

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BANNICK,COURTNEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: FINOL SOSA

DATE/TIME OF TEST: 10/08/2023 01:31:21

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:19	
Control Test	0.081	01:19	
Air Blank	0.000	01:20	
Subject Sample	# 1 0.183	01:23	Volume Not Met
Breath Volume	0.277	01:23	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample	# 2 0.177	01:26	
Breath Volume	1.343	01:26	
Air Blank	0.000	01:26	
Air Blank	0.000	01:28	
Subject Sample	# 3 0.185	01:29	
Breath Volume	1.808	01:29	
Air Blank	0.000	01:30	
Control Test	0.081	01:30	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,S
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:07

SUBJECT'S LAST NAME: RUBIO

DATE/TIME OF TEST: 10/08/2023 07:38:55

Diagnostics Check	OK	07:31
Air Blank	0.000	07:31
Control Test	0.081	07:32
Air Blank	0.000	07:32
Subject Sample # 1	0.151	07:33
Breath Volume	1.742	07:33
Air Blank	0.000	07:34
Air Blank	0.000	07:36
Subject Sample # 2	0.160	07:37
Breath Volume	1.433	07:37
Air Blank	0.000	07:38
Control Test	0.082	07:38
Air Blank	0.000	07:38
Diagnostics Check	OK	07:38

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,J,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MINOSO,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: RABUCK

DATE/TIME OF TEST: 10/12/2023 00:24:57

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:22	
Control Test	0.080	00:22	
Air Blank	0.000	00:23	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:24	
Control Test	0.080	00:24	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINS,JOSH,T
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: PETTIS,JESSE
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: BRANEM

DATE/TIME OF TEST: 10/15/2023 01:39:38

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.236	01:35
Breath Volume	1.546	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.236	01:38
Breath Volume	2.761	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COODY,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: COODY,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: GOYENS

DATE/TIME OF TEST: 10/21/2023 08:11:31

Diagnostics Check	OK	08:08	
Air Blank	0.000	08:09	
Control Test	0.081	08:09	
Air Blank	0.000	08:09	
Subject Sample	# 1	08:10	Subject Test Refused
Breath Volume	0.000	08:10	
Air Blank	0.000	08:10	
Control Test	0.081	08:10	
Air Blank	0.000	08:11	
Diagnostics Check	OK	08:11	

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 10/21/2023 19:59:28

Diagnostics Check	OK	19:52
Air Blank	0.000	19:53
Control Test	0.081	19:53
Air Blank	0.000	19:54
Subject Sample # 1	0.202	19:55
Breath Volume	2.109	19:55
Air Blank	0.000	19:55
Air Blank	0.000	19:57
Subject Sample # 2	0.200	19:57
Breath Volume	2.000	19:57
Air Blank	0.000	19:58
Control Test	0.082	19:58
Air Blank	0.000	19:59
Diagnostics Check	OK	19:59

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: PITTMAN,T
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 10/22/2023 02:37:08

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.081	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:30	No Sample Provided
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Air Blank	0.000	02:32	
Subject Sample	# 2	02:35	No Sample Provided
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.081	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: LERMONT ALVAREZ

DATE/TIME OF TEST: 10/22/2023 03:07:43

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.154	03:02
Breath Volume	1.289	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.158	03:06
Breath Volume	1.578	03:06
Air Blank	0.000	03:06
Control Test	0.082	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,JOSEPH,
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: TODARO,V
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 10/22/2023 04:08:43

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.081	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.072	04:04
Breath Volume	2.617	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.071	04:07
Breath Volume	2.839	04:07
Air Blank	0.000	04:07
Control Test	0.082	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERDT,ANDRU,J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: CINTRON,JEAN
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: ROTH

DATE/TIME OF TEST: 10/22/2023 18:21:08

Diagnostics Check	OK	18:14
Air Blank	0.000	18:15
Control Test	0.081	18:15
Air Blank	0.000	18:16
Subject Sample # 1	0.087	18:16
Breath Volume	1.609	18:16
Air Blank	0.000	18:17
Air Blank	0.000	18:19
Subject Sample # 2	0.087	18:19
Breath Volume	1.582	18:19
Air Blank	0.000	18:20
Control Test	0.081	18:20
Air Blank	0.000	18:21
Diagnostics Check	OK	18:21

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: ALLEN,B
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: W320432636250

DATE/TIME OF TEST: 10/22/2023 22:24:50

Diagnostics Check	OK	22:13	
Air Blank	0.000	22:13	
Control Test	0.080	22:14	
Air Blank	0.000	22:14	
Subject Sample	# 1	22:17	No Sample Provided
Breath Volume	0.000	22:17	
Air Blank	0.000	22:18	
Air Blank	0.000	22:20	
Subject Sample	# 2	22:23	No Sample Provided
Breath Volume	0.000	22:23	
Air Blank	0.000	22:23	
Control Test	0.081	22:24	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 10/25/2023 07:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUBICEK,JORDAN,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: FEAGANS,DUSTIN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MIKELL

DATE/TIME OF TEST: 10/29/2023 03:39:11

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:27	
Control Test	0.081	03:28	
Air Blank	0.000	03:28	
Subject Sample	# 1 0.169	03:31	Volume Not Met
Breath Volume	0.148	03:31	
Air Blank	0.000	03:32	
Air Blank	0.000	03:34	
Subject Sample	# 2 0.165	03:37	Volume Not Met
Breath Volume	0.117	03:37	
Air Blank	0.000	03:38	
Control Test	0.079	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 11/28/2023 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUBICEK,JORDAN,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ROUBICEK,JORDAN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: M240530654100

DATE/TIME OF TEST: 10/29/2023 04:02:30

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.000	03:58
Breath Volume	3.320	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.000	04:01
Breath Volume	2.609	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 11/28/2023 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: TORRES,V
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: BALDEO

DATE/TIME OF TEST: 10/31/2023 01:51:14

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.195	01:46
Breath Volume	1.480	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.191	01:49
Breath Volume	1.183	01:49
Air Blank	0.000	01:50
Control Test	0.081	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 11/28/2023 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: GOMES-DEBRITO

DATE/TIME OF TEST: 10/04/2023 01:29:08

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.081	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.253	01:24
Breath Volume	2.835	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.235	01:27
Breath Volume	1.664	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 10/06/2023 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLCHARRAN,DILON,C
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: BRAVERMAN,JACOB
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: DURRE

DATE/TIME OF TEST: 10/05/2023 02:57:38

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.081	02:57	RFI Detect
Air Blank	0.000	02:57	

Data Download Date/Time: 10/06/2023 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLCHARRAN,DILON,C
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: BRAVERMAN,JACOB
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: DURRE

DATE/TIME OF TEST: 10/05/2023 03:03:37

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Control Test	0.082	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 10/06/2023 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 10/12/2023 02:43:45

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.078	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.120	02:38
Breath Volume	1.152	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.133	02:41
Breath Volume	1.660	02:41
Air Blank	0.000	02:42
Control Test	0.081	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 10/24/2023 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: TYSON-JOHNSON

DATE/TIME OF TEST: 10/17/2023 02:48:51

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.077	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.046	02:43
Breath Volume	1.109	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.045	02:47
Breath Volume	1.429	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 10/24/2023 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: QUEEN

DATE/TIME OF TEST: 10/19/2023 18:13:03

Diagnostics Check	OK	18:06
Air Blank	0.000	18:07
Control Test	0.079	18:07
Air Blank	0.000	18:08
Subject Sample # 1	0.000	18:08
Breath Volume	2.511	18:08
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.000	18:11
Breath Volume	3.226	18:11
Air Blank	0.000	18:12
Control Test	0.079	18:12
Air Blank	0.000	18:12
Diagnostics Check	OK	18:13

Data Download Date/Time: 10/19/2023 22:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: ROWAN

DATE/TIME OF TEST: 10/28/2023 23:13:13

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:01	
Control Test	0.080	23:02	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:05	No Sample Provided
Breath Volume	0.000	23:05	
Air Blank	0.000	23:06	
Air Blank	0.000	23:08	
Subject Sample	# 2	23:11	No Sample Provided
Breath Volume	0.000	23:11	
Air Blank	0.000	23:12	
Control Test	0.080	23:12	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

Data Download Date/Time: 11/06/2023 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: CARRERO VELAZQUEZ

DATE/TIME OF TEST: 10/05/2023 03:59:50

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.183	03:54
Breath Volume	1.527	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.186	03:58
Breath Volume	1.687	03:58
Air Blank	0.000	03:58
Control Test	0.079	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 10/18/2023 11:27