

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,ANDREW,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: CORTES,RAUL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: GALLIANO

DATE/TIME OF TEST: 10/25/2023 21:46:36

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:44	
Control Test	0.081	21:44	
Air Blank	0.000	21:45	
Subject Sample	# 1	21:45	Subject Test Refused
Breath Volume	0.000	21:45	
Air Blank	0.000	21:45	
Control Test	0.080	21:46	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

Data Download Date/Time: 11/21/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISNEY,WHITNEY,L
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: GONZALEZ,JUSTIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/27/2023 14:40:52

Diagnostics Check	OK	14:33
Air Blank	0.000	14:34
Control Test	0.080	14:34
Air Blank	0.000	14:35
Subject Sample # 1	0.226	14:36
Breath Volume	1.632	14:36
Air Blank	0.000	14:36
Air Blank	0.000	14:38
Subject Sample # 2	0.219	14:39
Breath Volume	3.460	14:39
Air Blank	0.000	14:39
Control Test	0.078	14:40
Air Blank	0.000	14:40
Diagnostics Check	OK	14:40

Data Download Date/Time: 11/21/2023 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISNEY,WHITNEY,L
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: BAEZ,JOSEPH
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: ARMENTEROS

DATE/TIME OF TEST: 10/27/2023 15:46:39

Diagnostics Check	OK	15:43	
Air Blank	0.000	15:44	
Control Test	0.080	15:44	
Air Blank	0.000	15:45	
Subject Sample	# 1	15:45	Subject Test Refused
Breath Volume	0.000	15:45	
Air Blank	0.000	15:45	
Control Test	0.080	15:46	
Air Blank	0.000	15:46	
Diagnostics Check	OK	15:46	

Data Download Date/Time: 11/21/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,ALEXIA,S
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: BRAVO,GEORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BARON ORTIZ

DATE/TIME OF TEST: 10/28/2023 03:08:15

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.143	03:03
Breath Volume	1.656	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:05
Subject Sample # 2	0.125	03:06
Breath Volume	1.195	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 11/21/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,ALEXIA,S
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: REYES,JOEL
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: HERMIDA

DATE/TIME OF TEST: 10/28/2023 05:07:19

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.081	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.000	05:02
Breath Volume	4.199	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.000	05:05
Breath Volume	2.539	05:05
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 11/21/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,ALEXIA,S
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: REYES,JOEL
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: HERMIDA

DATE/TIME OF TEST: 10/28/2023 05:39:01

Diagnostics Check	OK	05:35	
Air Blank	0.000	05:36	
Control Test	0.081	05:36	
Air Blank	0.000	05:37	
Subject Sample	# 1	05:37	RFI Detect
Breath Volume	3.515	05:37	
Air Blank	0.000	05:38	
Control Test	0.081	05:38	
Air Blank	0.000	05:38	
Diagnostics Check	OK	05:39	

Data Download Date/Time: 11/21/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,ALEXIA,S
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: REYES,JOEL
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: HERMIDA

DATE/TIME OF TEST: 10/28/2023 06:07:01

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.081	06:01
Air Blank	0.000	06:02
Subject Sample # 1	0.000	06:02
Breath Volume	1.832	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.000	06:05
Breath Volume	2.152	06:05
Air Blank	0.000	06:06
Control Test	0.080	06:06
Air Blank	0.000	06:06
Diagnostics Check	OK	06:07

Data Download Date/Time: 11/21/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/29/2023 04:28:12

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.212	04:23
Breath Volume	3.253	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 2	0.202	04:26
Breath Volume	2.152	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 11/21/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUNEZ,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: NUNEZ,BENJAMIN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: SOLANO

DATE/TIME OF TEST: 10/20/2023 20:57:49

Diagnostics Check	OK	20:54	
Air Blank	0.000	20:55	
Control Test	0.080	20:55	
Air Blank	0.000	20:56	
Subject Sample	# 1	20:56	Subject Test Refused
Breath Volume	0.000	20:56	
Air Blank	0.000	20:56	
Control Test	0.079	20:57	
Air Blank	0.000	20:57	
Diagnostics Check	OK	20:57	

Data Download Date/Time: 11/16/2023 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOSA,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: HOPKINS,DALE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/22/2023 19:33:33

Diagnostics Check	OK	19:20	
Air Blank	0.000	19:21	
Control Test	0.079	19:21	
Air Blank	0.000	19:21	
Subject Sample	# 1 0.123	19:25	Volume Not Met
Breath Volume	0.781	19:25	
Air Blank	0.000	19:25	
Air Blank	0.000	19:27	
Subject Sample	# 2 0.133	19:28	
Breath Volume	1.105	19:28	
Air Blank	0.000	19:28	
Air Blank	0.000	19:30	
Subject Sample	# 3 0.124	19:32	
Breath Volume	2.289	19:32	
Air Blank	0.000	19:32	
Control Test	0.079	19:32	
Air Blank	0.000	19:33	
Diagnostics Check	OK	19:33	

Data Download Date/Time: 11/16/2023 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: LINS,GIOVANE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: NYEI

DATE/TIME OF TEST: 10/28/2023 02:30:26

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.181	02:26
Breath Volume	2.742	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.175	02:28
Breath Volume	1.484	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 11/16/2023 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,MATTHEW,T
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: RUDOLPH,MATTHEW
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: MEKHTIYEVA

DATE/TIME OF TEST: 10/04/2023 17:45:26

Diagnostics Check	OK	17:42	
Air Blank	0.000	17:43	
Control Test	0.079	17:43	
Air Blank	0.000	17:43	
Subject Sample	# 1	17:44	Subject Test Refused
Breath Volume	0.000	17:44	
Air Blank	0.000	17:44	
Control Test	0.079	17:44	
Air Blank	0.000	17:45	
Diagnostics Check	OK	17:45	

Data Download Date/Time: 10/05/2023 04:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BREMER,DENISE
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: CERON

DATE/TIME OF TEST: 10/29/2023 01:18:14

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.079	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Control Test	0.079	01:17	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 11/03/2023 05:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGO,TOMAS,F
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD
ARRESTING OFFICER: REGO,TOMAS
ARRESTING OFFICER AGENCY: Bay Harbor Island PD
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: ALMARAL

DATE/TIME OF TEST: 10/29/2023 22:37:15

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.178	22:32
Breath Volume	2.023	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.178	22:35
Breath Volume	1.953	22:35
Air Blank	0.000	22:36
Control Test	0.078	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 11/03/2023 05:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,LILIAN,
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD
ARRESTING OFFICER: REGO,TOMAS
ARRESTING OFFICER AGENCY: Bay Harbor Island PD
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/30/2023 04:24:18

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.080	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:22	Subject Test Refused
Breath Volume	0.000	04:22	
Air Blank	0.000	04:23	
Control Test	0.080	04:23	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 11/03/2023 05:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSTROV,VANESSA,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: RUIZ,DANIEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 10/12/2023 03:46:40

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.079	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1 0.110	03:40	Volume Not Met
Breath Volume	0.625	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 2	03:45	Subject Test Refused
Breath Volume	0.246	03:45	
Air Blank	0.000	03:45	
Control Test	0.077	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 10/22/2023 04:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ-LOPEZ,RICHARD,A
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: GOMEZ-LOPEZ,RICHARD
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: RUBIO-MADARIAGA

DATE/TIME OF TEST: 10/14/2023 04:34:29

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.119	04:30
Breath Volume	1.691	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.117	04:32
Breath Volume	2.003	04:32
Air Blank	0.000	04:33
Control Test	0.077	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 10/22/2023 04:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODUM,ICE,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: ODUM,ICE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 10/15/2023 00:28:29

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.080	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1 0.164	00:25	Volume Not Met
Breath Volume	0.726	00:25	
Air Blank	0.000	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 2	00:28	Sequence Aborted
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	

Data Download Date/Time: 10/22/2023 04:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODUM,ICE,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: ODUM,ICE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 10/15/2023 00:36:27

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.079	00:34	
Air Blank	0.000	00:34	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 10/22/2023 04:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,HENRIQUE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: COOPER,HENRIQUE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 10/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CARDONA-PAVON

DATE/TIME OF TEST: 10/23/2023 02:47:40

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:45	
Control Test	0.080	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.079	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 11/17/2023 05:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROCHA,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: ALI BAEZ

DATE/TIME OF TEST: 10/04/2023 05:32:20

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:30	
Control Test	0.080	05:30	
Air Blank	0.000	05:30	
Subject Sample	# 1	05:30	Subject Test Refused
Breath Volume	0.000	05:30	
Air Blank	0.000	05:31	
Control Test	0.080	05:31	
Air Blank	0.000	05:32	
Diagnostics Check	OK	05:32	

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:23

SUBJECT'S LAST NAME: PILKENTON

DATE/TIME OF TEST: 10/04/2023 13:49:28

Diagnostics Check	OK	13:46	
Air Blank	0.000	13:46	
Control Test	0.080	13:46	
Air Blank	0.000	13:47	
Subject Sample	# 1	13:48	Subject Test Refused
Breath Volume	0.000	13:48	
Air Blank	0.000	13:48	
Control Test	0.080	13:48	
Air Blank	0.000	13:49	
Diagnostics Check	OK	13:49	

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FIGUEROA,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: GONSKA

DATE/TIME OF TEST: 10/05/2023 02:20:45

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.194	02:16
Breath Volume	1.527	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.197	02:19
Breath Volume	1.375	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RYAN,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 10/05/2023 03:39:56

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.214	03:34
Breath Volume	2.343	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.198	03:38
Breath Volume	1.773	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,K,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 12:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: VEGAFRUSHOIN

DATE/TIME OF TEST: 10/06/2023 01:05:41

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.079	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.365	01:01
Breath Volume	2.164	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.350	01:04
Breath Volume	2.105	01:04
Air Blank	0.000	01:04
Control Test	0.079	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: DELROSARIO-CAMACH

DATE/TIME OF TEST: 10/07/2023 00:24:44

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:22	
Control Test	0.080	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.080	00:24	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,DEP REINA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: TAYLOR-WELLS

DATE/TIME OF TEST: 10/07/2023 02:01:35

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.080	01:59	
Air Blank	0.000	01:59	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.080	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 10/07/2023 21:44:55

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.236	21:40
Breath Volume	2.109	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:42
Subject Sample # 2	0.232	21:43
Breath Volume	1.855	21:43
Air Blank	0.000	21:44
Control Test	0.079	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SLOANE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MARQUES

DATE/TIME OF TEST: 10/08/2023 01:52:49

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.080	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.079	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MATRANGA

DATE/TIME OF TEST: 10/08/2023 02:42:04

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.226	02:32
Breath Volume	1.980	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.201	02:36
Breath Volume	1.593	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 3	0.222	02:40
Breath Volume	1.394	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:42

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GONZALEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: FORRESTER

DATE/TIME OF TEST: 10/08/2023 22:28:08

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.080	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.000	22:23
Breath Volume	2.808	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.000	22:26
Breath Volume	1.582	22:26
Air Blank	0.000	22:27
Control Test	0.080	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAEDKE,KELLY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SUAREZ PEREZ

DATE/TIME OF TEST: 10/09/2023 03:19:52

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.080	03:17	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.081	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIDY,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKEIRIK,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: DUNDORE

DATE/TIME OF TEST: 10/10/2023 02:18:54

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.000	02:14
Breath Volume	1.343	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.000	02:17
Breath Volume	1.175	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAVER,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: QUIROZ

DATE/TIME OF TEST: 10/11/2023 01:53:45

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.080	01:53	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: GROVES

DATE/TIME OF TEST: 10/11/2023 23:16:05

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.093	23:11
Breath Volume	2.738	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.088	23:14
Breath Volume	1.937	23:14
Air Blank	0.000	23:15
Control Test	0.079	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,STEPHAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,STEPHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: DEKOVIC

DATE/TIME OF TEST: 10/13/2023 06:19:05

Diagnostics Check	OK	06:13
Air Blank	0.000	06:13
Control Test	0.080	06:13
Air Blank	0.000	06:14
Subject Sample # 1	0.000	06:14
Breath Volume	2.886	06:14
Air Blank	0.000	06:15
Air Blank	0.000	06:17
Subject Sample # 2	0.000	06:17
Breath Volume	3.234	06:17
Air Blank	0.000	06:18
Control Test	0.079	06:18
Air Blank	0.000	06:19
Diagnostics Check	OK	06:19

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: TUREK

DATE/TIME OF TEST: 10/13/2023 19:49:25

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.079	19:43
Air Blank	0.000	19:44
Subject Sample # 1	0.000	19:44
Breath Volume	1.730	19:44
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.000	19:48
Breath Volume	1.789	19:48
Air Blank	0.000	19:48
Control Test	0.080	19:48
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SEVOSTJANOV,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: LANIER

DATE/TIME OF TEST: 10/13/2023 23:56:32

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.140	23:51
Breath Volume	1.347	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:53
Subject Sample # 2	0.142	23:54
Breath Volume	1.906	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BECERRO-HORMIGA

DATE/TIME OF TEST: 10/14/2023 00:22:43

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.079	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.244	00:18
Breath Volume	3.265	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.249	00:21
Breath Volume	3.093	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WHEELER,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MAZZONE

DATE/TIME OF TEST: 10/14/2023 01:29:23

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.080	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:28	Subject Test Refused
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Control Test	0.080	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: CAMILO

DATE/TIME OF TEST: 10/14/2023 02:30:00

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.148	02:25
Breath Volume	2.636	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:27
Subject Sample # 2	0.144	02:28
Breath Volume	2.296	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:30

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SHOULTA,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 10/14/2023 22:17:21

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:11
Subject Sample # 1	0.165	22:12
Breath Volume	1.347	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.164	22:15
Breath Volume	1.363	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 10/18/2023 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: ALVELO KUILAN

DATE/TIME OF TEST: 10/15/2023 00:21:43

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.059	00:17
Breath Volume	1.429	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.061	00:20
Breath Volume	1.191	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 10/18/2023 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JUNE,DEPUTY M.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BAGGERLY

DATE/TIME OF TEST: 10/15/2023 21:35:59

Diagnostics Check	OK	21:31	
Air Blank	0.000	21:32	
Control Test	0.080	21:32	
Air Blank	0.000	21:33	
Subject Sample	# 1	21:34	Subject Test Refused
Breath Volume	0.000	21:34	
Air Blank	0.000	21:35	
Control Test	0.080	21:35	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

Data Download Date/Time: 10/18/2023 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RYAN,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: DECUIR

DATE/TIME OF TEST: 10/16/2023 03:37:58

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.113	03:33
Breath Volume	2.417	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.110	03:36
Breath Volume	2.285	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 10/18/2023 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 10/17/2023 03:47:09

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.000	03:42
Breath Volume	2.156	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.000	03:45
Breath Volume	2.632	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 10/18/2023 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 10/17/2023 22:43:51

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.000	22:39
Breath Volume	2.238	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.000	22:42
Breath Volume	2.398	22:42
Air Blank	0.000	22:42
Control Test	0.080	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 10/18/2023 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPICER,NICHOLAS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DUNLAP

DATE/TIME OF TEST: 10/18/2023 01:34:07

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:33	
Control Test	0.080	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 10/18/2023 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATSAREAS,NIKOLAOS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORTORA,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:08

SUBJECT'S LAST NAME: KAPRIELIAN

DATE/TIME OF TEST: 10/20/2023 13:34:35

Diagnostics Check	OK	13:28
Air Blank	0.000	13:29
Control Test	0.079	13:29
Air Blank	0.000	13:29
Subject Sample # 1	0.000	13:30
Breath Volume	2.582	13:30
Air Blank	0.000	13:30
Air Blank	0.000	13:32
Subject Sample # 2	0.000	13:33
Breath Volume	2.800	13:33
Air Blank	0.000	13:33
Control Test	0.079	13:34
Air Blank	0.000	13:34
Diagnostics Check	OK	13:34

Data Download Date/Time: 11/08/2023 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: VLAHAKIS

DATE/TIME OF TEST: 10/21/2023 11:09:40

Diagnostics Check	OK	11:03
Air Blank	0.000	11:03
Control Test	0.079	11:04
Air Blank	0.000	11:04
Subject Sample # 1	0.000	11:05
Breath Volume	3.441	11:05
Air Blank	0.000	11:05
Air Blank	0.000	11:07
Subject Sample # 2	0.000	11:08
Breath Volume	3.632	11:08
Air Blank	0.000	11:08
Control Test	0.081	11:09
Air Blank	0.000	11:09
Diagnostics Check	OK	11:09

Data Download Date/Time: 11/08/2023 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: MADDOX

DATE/TIME OF TEST: 10/22/2023 04:18:29

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.080	04:16	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:17	Subject Test Refused
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Control Test	0.080	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 11/08/2023 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DOUGHERTY,ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 10/22/2023 05:11:47

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.214	05:07
Breath Volume	2.332	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.203	05:10
Breath Volume	1.664	05:10
Air Blank	0.000	05:10
Control Test	0.079	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 11/08/2023 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTIN,JEFFERY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/23/2023 03:01:32

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	02:59	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Control Test	0.080	03:00	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 11/08/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVANOVICLM,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IVANOVICLM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 14:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 10/24/2023 16:34:19

Diagnostics Check	OK	16:27
Air Blank	0.000	16:28
Control Test	0.079	16:28
Air Blank	0.000	16:29
Subject Sample # 1	0.362	16:29
Breath Volume	1.214	16:29
Air Blank	0.000	16:30
Air Blank	0.000	16:32
Subject Sample # 2	0.354	16:32
Breath Volume	1.738	16:32
Air Blank	0.000	16:33
Control Test	0.079	16:33
Air Blank	0.000	16:34
Diagnostics Check	OK	16:34

Data Download Date/Time: 11/08/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: UVALDO REYES

DATE/TIME OF TEST: 10/28/2023 01:29:19

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.000	01:25
Breath Volume	1.234	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.000	01:27
Breath Volume	1.246	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 11/08/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: WALDHALM

DATE/TIME OF TEST: 10/28/2023 04:08:13

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.000	04:04
Breath Volume	2.082	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.000	04:06
Breath Volume	2.343	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 11/08/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GOODPASTER,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 10/28/2023 21:47:56

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.079	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.165	21:43
Breath Volume	1.238	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:45
Subject Sample # 2	0.175	21:46
Breath Volume	1.449	21:46
Air Blank	0.000	21:47
Control Test	0.079	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 11/08/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 10/29/2023 02:04:22

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.078	01:59
Breath Volume	1.746	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.076	02:02
Breath Volume	1.628	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 11/08/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: GROOMS

DATE/TIME OF TEST: 10/29/2023 04:45:49

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.080	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.175	04:41
Breath Volume	2.160	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.175	04:44
Breath Volume	1.714	04:44
Air Blank	0.000	04:44
Control Test	0.079	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 11/08/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: MABEE

DATE/TIME OF TEST: 10/29/2023 23:53:29

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.000	23:49
Breath Volume	2.222	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.000	23:52
Breath Volume	2.296	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 11/08/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: DANNY

DATE/TIME OF TEST: 10/30/2023 00:26:40

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.249	00:22
Breath Volume	1.562	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.241	00:25
Breath Volume	1.796	00:25
Air Blank	0.000	00:25
Control Test	0.079	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 11/08/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,STEPHAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: TAZZIOLI

DATE/TIME OF TEST: 10/30/2023 21:22:02

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:11	
Control Test	0.080	21:11	
Air Blank	0.000	21:11	
Subject Sample	# 1	21:15	No Sample Provided
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Air Blank	0.000	21:17	
Subject Sample	# 2	21:20	No Sample Provided
Breath Volume	0.000	21:20	
Air Blank	0.000	21:21	
Control Test	0.080	21:21	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:22	

Data Download Date/Time: 11/08/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SEVOSTJANOV,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BRANGARD

DATE/TIME OF TEST: 10/01/2023 21:27:06

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.081	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.046	21:22
Breath Volume	1.894	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.043	21:25
Breath Volume	1.941	21:25
Air Blank	0.000	21:26
Control Test	0.080	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 10/18/2023 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BANNISTER,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: HARDWICK

DATE/TIME OF TEST: 10/02/2023 21:17:47

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.081	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.181	21:13
Breath Volume	2.335	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:15
Subject Sample # 2	0.184	21:16
Breath Volume	1.621	21:16
Air Blank	0.000	21:16
Control Test	0.078	21:17
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 10/18/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RENNIE,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: PEREZ JUTIZ

DATE/TIME OF TEST: 10/04/2023 03:45:13

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.082	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Control Test	0.081	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 10/18/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: SHERIFF

DATE/TIME OF TEST: 10/04/2023 21:49:15

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.081	21:43
Air Blank	0.000	21:44
Subject Sample # 1	0.000	21:44
Breath Volume	1.937	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.000	21:47
Breath Volume	1.980	21:47
Air Blank	0.000	21:48
Control Test	0.081	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 10/18/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAKA,ALTIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DORGAN,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: GALINDO

DATE/TIME OF TEST: 10/07/2023 06:20:23

Diagnostics Check	OK	06:13
Air Blank	0.000	06:14
Control Test	0.081	06:14
Air Blank	0.000	06:15
Subject Sample # 1	0.100	06:15
Breath Volume	1.515	06:15
Air Blank	0.000	06:16
Air Blank	0.000	06:18
Subject Sample # 2	0.100	06:18
Breath Volume	1.546	06:18
Air Blank	0.000	06:19
Control Test	0.081	06:19
Air Blank	0.000	06:20
Diagnostics Check	OK	06:20

Data Download Date/Time: 10/18/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EVANGELOUS,NICHOLAS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: ALLISON

DATE/TIME OF TEST: 10/13/2023 20:00:34

Diagnostics Check	OK	19:54
Air Blank	0.000	19:54
Control Test	0.080	19:54
Air Blank	0.000	19:55
Subject Sample # 1	0.202	19:55
Breath Volume	2.394	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.199	19:58
Breath Volume	1.937	19:58
Air Blank	0.000	19:59
Control Test	0.081	19:59
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 10/18/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRADY,J
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: DAVILA HORNA

DATE/TIME OF TEST: 10/14/2023 00:07:06

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Subject Sample # 3	0.222	23:58
Breath Volume	1.605	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.201	00:01
Breath Volume	2.367	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.189	00:05
Breath Volume	1.507	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 10/18/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 10/14/2023 00:34:32

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.080	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.212	00:29
Breath Volume	2.875	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.208	00:32
Breath Volume	1.898	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 10/18/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: PANZO CALIHUA

DATE/TIME OF TEST: 10/14/2023 02:51:22

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.129	02:46
Breath Volume	1.367	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample # 2	0.124	02:49
Breath Volume	2.148	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 10/18/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUSCH,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 10/15/2023 17:09:56

Diagnostics Check	OK	17:07	
Air Blank	0.000	17:07	
Control Test	0.081	17:07	
Air Blank	0.000	17:08	
Subject Sample	# 1	17:08	Subject Test Refused
Breath Volume	0.000	17:08	
Air Blank	0.000	17:09	
Control Test	0.081	17:09	
Air Blank	0.000	17:09	
Diagnostics Check	OK	17:09	

Data Download Date/Time: 10/18/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TRADOR,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: ABRAMS

DATE/TIME OF TEST: 10/17/2023 21:21:08

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:15
Subject Sample # 1	0.204	21:16
Breath Volume	1.734	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:18
Subject Sample # 2	0.195	21:19
Breath Volume	1.269	21:19
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 10/18/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PLUMB,JASON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: WHITAKER

DATE/TIME OF TEST: 10/18/2023 16:59:00

Diagnostics Check	OK	16:56	
Air Blank	0.000	16:56	
Control Test	0.079	16:56	
Air Blank	0.000	16:57	
Subject Sample	# 1	16:57	Subject Test Refused
Breath Volume	0.000	16:57	
Air Blank	0.000	16:58	
Control Test	0.080	16:58	
Air Blank	0.000	16:58	
Diagnostics Check	OK	16:59	

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/20/2023 02:07:51

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.000	02:03
Breath Volume	2.328	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.000	02:06
Breath Volume	1.921	02:06
Air Blank	0.000	02:06
Control Test	0.080	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: DEETZ

DATE/TIME OF TEST: 10/21/2023 02:03:36

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.080	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.080	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUREUX,MATT,L
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: LAMOUREUX,MATT
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: BUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: SAMS

DATE/TIME OF TEST: 10/22/2023 20:00:58

Diagnostics Check	OK	19:54
Air Blank	0.000	19:55
Control Test	0.080	19:55
Air Blank	0.000	19:56
Subject Sample # 1	0.000	19:56
Breath Volume	2.476	19:56
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.000	19:59
Breath Volume	1.578	19:59
Air Blank	0.000	20:00
Control Test	0.079	20:00
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAVER,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SANTERAMO

DATE/TIME OF TEST: 10/23/2023 22:12:06

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.078	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.058	22:07
Breath Volume	2.292	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.053	22:10
Breath Volume	1.945	22:10
Air Blank	0.000	22:11
Control Test	0.076	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PENNAMON,D'ANDRE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MICHAUD

DATE/TIME OF TEST: 10/23/2023 22:39:35

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.079	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Control Test	0.079	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STACHNIK,DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: HACKETT

DATE/TIME OF TEST: 10/24/2023 20:22:03

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.078	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.000	20:17
Breath Volume	2.171	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.000	20:20
Breath Volume	1.925	20:20
Air Blank	0.000	20:21
Control Test	0.077	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:22

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUTH,TIMOTHY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: ANANTH

DATE/TIME OF TEST: 10/24/2023 22:24:08

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.079	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.161	22:19
Breath Volume	1.804	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:21
Subject Sample # 2	0.167	22:22
Breath Volume	2.203	22:22
Air Blank	0.000	22:23
Control Test	0.075	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 10/26/2023 01:49:42

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.136	01:43
Breath Volume	2.250	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.142	01:48
Breath Volume	1.550	01:48
Air Blank	0.000	01:48
Control Test	0.078	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAVER,DEP R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:41

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 10/28/2023 07:12:05

Diagnostics Check	OK	07:03
Air Blank	0.000	07:03
Control Test	0.079	07:04
Air Blank	0.000	07:04
Subject Sample # 1	0.134	07:06
Breath Volume	1.296	07:06
Air Blank	0.000	07:07
Air Blank	0.000	07:09
Subject Sample # 2	0.133	07:10
Breath Volume	1.113	07:10
Air Blank	0.000	07:11
Control Test	0.079	07:11
Air Blank	0.000	07:12
Diagnostics Check	OK	07:12

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STOFFAN,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: HOOPER

DATE/TIME OF TEST: 10/29/2023 21:47:50

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:45	
Control Test	0.080	21:45	
Air Blank	0.000	21:45	
Subject Sample	# 1	21:46	Subject Test Refused
Breath Volume	0.000	21:46	
Air Blank	0.000	21:46	
Control Test	0.080	21:47	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COYLE,TIMOTHY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: TYSON

DATE/TIME OF TEST: 10/31/2023 17:41:29

Diagnostics Check	OK	17:35
Air Blank	0.000	17:35
Control Test	0.080	17:35
Air Blank	0.000	17:36
Subject Sample # 1	0.000	17:37
Breath Volume	1.261	17:37
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 2	0.000	17:40
Breath Volume	2.023	17:40
Air Blank	0.000	17:40
Control Test	0.078	17:40
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VILLA,ALAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: VANDERMOLEN

DATE/TIME OF TEST: 10/31/2023 19:50:54

Diagnostics Check	OK	19:43
Air Blank	0.000	19:44
Control Test	0.079	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.132	19:45
Breath Volume	1.859	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.134	19:49
Breath Volume	1.675	19:49
Air Blank	0.000	19:49
Control Test	0.078	19:50
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 11/08/2023 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 10/22/2023 00:29:33

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.080	00:26	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.081	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 11/08/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTLETT,SHELTON,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BARTLETT,SHELTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: OTHER
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 10/14/2023 20:30:55

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.082	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.000	20:26
Breath Volume	4.226	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.000	20:29
Breath Volume	4.273	20:29
Air Blank	0.000	20:30
Control Test	0.082	20:30
Air Blank	0.000	20:30
Diagnostics Check	OK	20:30

Data Download Date/Time: 11/26/2023 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,EMILY,A
BREATH TEST OPERATOR AGENCY: FHP TROOP A EAST
ARRESTING OFFICER: MELTON,EMILY
ARRESTING OFFICER AGENCY: FHP TROOP A EAST
LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212952
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 10/02/2023 04:25:14

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.089	04:15
Breath Volume	1.937	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.019	04:20
Breath Volume	1.347	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 3	0.088	04:23
Breath Volume	3.660	04:23
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 10/09/2023 01:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAMMER,DALTON,J
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: SWIFT,MICHAEL
ARRESTING OFFICER AGENCY: GULF COUNTY SO
LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212952
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: RAINES

DATE/TIME OF TEST: 10/06/2023 17:31:33

Diagnostics Check	OK	17:22
Air Blank	0.000	17:22
Control Test	0.079	17:23
Air Blank	0.000	17:23
Subject Sample # 1	0.181	17:24
Breath Volume	1.824	17:24
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 2	0.211	17:27
Breath Volume	2.468	17:27
Air Blank	0.000	17:27
Air Blank	0.000	17:29
Subject Sample # 3	0.204	17:30
Breath Volume	2.300	17:30
Air Blank	0.000	17:30
Control Test	0.079	17:30
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 10/09/2023 01:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP TROOP A EAST
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP TROOP A EAST
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212952
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: LAKE

DATE/TIME OF TEST: 10/11/2023 01:20:23

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:08	
Air Blank	0.000	01:08	
Subject Sample	# 1 0.101	01:11	Volume Not Met
Breath Volume	0.449	01:11	
Air Blank	0.000	01:12	
Air Blank	0.000	01:14	
Subject Sample	# 2 0.111	01:14	
Breath Volume	1.707	01:14	
Air Blank	0.000	01:15	
Air Blank	0.000	01:17	
Subject Sample	# 3 0.114	01:18	
Breath Volume	1.507	01:18	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 11/28/2023 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAMMER,DALTON,J
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: SWIFT,MICHAEL
ARRESTING OFFICER AGENCY: GULF COUNTY SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212952
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: DALTON

DATE/TIME OF TEST: 10/14/2023 01:08:50

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.121	01:03
Breath Volume	2.781	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.123	01:07
Breath Volume	3.457	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 11/28/2023 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODINEZ,JAMIE,Z
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: SWIFT,MICHAEL
ARRESTING OFFICER AGENCY: GULF COUNTY SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212952
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: GODINEZ

DATE/TIME OF TEST: 10/15/2023 00:57:35

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.256	00:53
Breath Volume	2.207	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.253	00:56
Breath Volume	1.933	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 11/28/2023 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,JERRY,W
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: LEE,JERRY
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212952
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: AARON

DATE/TIME OF TEST: 10/16/2023 22:56:50

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.000	22:52
Breath Volume	2.414	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.000	22:55
Breath Volume	2.675	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 11/28/2023 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNODLE,KADE,I
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: CORREATORRES

DATE/TIME OF TEST: 10/02/2023 03:10:38

Diagnostics Check	OK	03:10	
Air Blank		03:10	RFI Detect
Air Blank	0.000	03:10	

Data Download Date/Time: 10/10/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNODLE,KADE,I
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: CORREATORRES

DATE/TIME OF TEST: 10/02/2023 03:20:21

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.196	03:15
Breath Volume	1.265	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.186	03:18
Breath Volume	1.265	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 10/10/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESTES,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: MARES SOLIS

DATE/TIME OF TEST: 10/02/2023 05:32:22

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.079	05:27
Air Blank	0.000	05:27
Subject Sample # 1	0.062	05:28
Breath Volume	3.437	05:28
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.062	05:30
Breath Volume	2.875	05:30
Air Blank	0.000	05:31
Control Test	0.080	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 10/10/2023 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECORELLA,R,K
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VELEZ,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 10/05/2023 00:26:32

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.079	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.242	00:18
Breath Volume	2.757	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.265	00:21
Breath Volume	2.625	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:23
Subject Sample # 3	0.265	00:24
Breath Volume	2.679	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 10/10/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VAZQUEZ,D
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: VERAS

DATE/TIME OF TEST: 10/07/2023 00:04:41

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.191	00:00
Breath Volume	2.785	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.197	00:03
Breath Volume	3.105	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 10/10/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SCOTT,E
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 10/07/2023 10:32:35

Diagnostics Check	OK	10:26
Air Blank	0.000	10:27
Control Test	0.080	10:27
Air Blank	0.000	10:27
Subject Sample # 1	0.000	10:28
Breath Volume	2.640	10:28
Air Blank	0.000	10:28
Air Blank	0.000	10:30
Subject Sample # 2	0.000	10:31
Breath Volume	1.542	10:31
Air Blank	0.000	10:31
Control Test	0.080	10:31
Air Blank	0.000	10:32
Diagnostics Check	OK	10:32

Data Download Date/Time: 10/10/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MYKHASKIV,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: REUMBERTO

DATE/TIME OF TEST: 10/08/2023 05:51:54

Diagnostics Check	OK	05:45
Air Blank	0.000	05:46
Control Test	0.080	05:46
Air Blank	0.000	05:47
Subject Sample # 1	0.068	05:47
Breath Volume	1.929	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.066	05:50
Breath Volume	2.035	05:50
Air Blank	0.000	05:50
Control Test	0.079	05:51
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 10/10/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FAREZ ESPIN,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: KEARNS

DATE/TIME OF TEST: 10/10/2023 21:18:05

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.079	21:12
Air Blank	0.000	21:12
Subject Sample # 1	0.180	21:13
Breath Volume	2.027	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:15
Subject Sample # 2	0.177	21:16
Breath Volume	2.660	21:16
Air Blank	0.000	21:17
Control Test	0.080	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLOAT,LUCAS,G
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HILL,G
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/12/2023 04:06:05

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.247	04:01
Breath Volume	2.875	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.239	04:04
Breath Volume	1.183	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOMANDO,TINA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BENDER,L
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FURLONG

DATE/TIME OF TEST: 10/12/2023 23:31:36

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.155	23:26
Breath Volume	1.554	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.151	23:30
Breath Volume	1.699	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,AUSTIN,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GOODMAN,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ROSSANO

DATE/TIME OF TEST: 10/15/2023 02:51:31

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.176	02:47
Breath Volume	3.535	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.180	02:49
Breath Volume	3.136	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAMPBELL,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 10/16/2023 20:50:32

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:48	
Control Test	0.079	20:48	
Air Blank	0.000	20:49	
Subject Sample	# 1	20:49	RFI Detect
Breath Volume	0.824	20:49	
Air Blank		20:49	RFI Detect
Air Blank	0.000	20:50	

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAMPBELL,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 10/16/2023 20:59:32

Diagnostics Check	OK	20:52
Air Blank	0.000	20:52
Control Test	0.080	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.247	20:54
Breath Volume	1.316	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.240	20:57
Breath Volume	1.257	20:57
Air Blank	0.000	20:58
Control Test	0.077	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA,R,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELISSAINT,M
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/19/2023 02:31:17

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.249	02:26
Breath Volume	2.742	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.240	02:29
Breath Volume	2.601	02:29
Air Blank	0.000	02:30
Control Test	0.078	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAHAGIAN,TARA,K
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JOHNSON,T
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/20/2023 23:47:29

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.000	23:43
Breath Volume	1.601	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.000	23:46
Breath Volume	2.734	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL, ADAM, J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HERNANDEZ, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CAMALLERIS

DATE/TIME OF TEST: 10/23/2023 01:27:55

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.130	01:23
Breath Volume	1.652	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:25
Subject Sample # 2	0.134	01:26
Breath Volume	1.394	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECORELLA,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GRAY,J
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: PARR

DATE/TIME OF TEST: 10/25/2023 03:40:24

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.079	03:36	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.289	03:38	
Air Blank	0.000	03:39	
Control Test	0.081	03:39	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNODLE,KADE,I
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ELIASSAINT,MADOCHEE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 10/25/2023 05:08:43

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:05	
Control Test	0.079	05:05	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:08	Sequence Aborted
Breath Volume	0.000	05:08	
Air Blank	0.000	05:08	

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,AUSTIN,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GUTIERREZ,I
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: GENEVRO

DATE/TIME OF TEST: 10/27/2023 22:07:30

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.214	22:02
Breath Volume	1.511	22:02
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.210	22:05
Breath Volume	2.910	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDASH,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ALLEN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/28/2023 04:29:56

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.079	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.108	04:25
Breath Volume	2.332	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.100	04:28
Breath Volume	1.910	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA,R,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELIASSAINT,M
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: PINTA

DATE/TIME OF TEST: 10/29/2023 05:25:50

Diagnostics Check	OK	05:19
Air Blank	0.000	05:20
Control Test	0.080	05:20
Air Blank	0.000	05:21
Subject Sample # 1	0.000	05:21
Breath Volume	3.792	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.000	05:24
Breath Volume	3.640	05:24
Air Blank	0.000	05:24
Control Test	0.080	05:25
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,S,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SPEICH,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: AYALAGARCIA

DATE/TIME OF TEST: 10/31/2023 00:20:20

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.254	00:15
Breath Volume	1.808	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.261	00:18
Breath Volume	1.800	00:18
Air Blank	0.000	00:19
Control Test	0.078	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STANLEY,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MCMULLAN

DATE/TIME OF TEST: 10/31/2023 00:49:11

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.078	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1 0.177	00:41	Volume Not Met
Breath Volume	0.960	00:41	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 2 0.174	00:44	
Breath Volume	1.703	00:44	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample	# 3 0.172	00:47	
Breath Volume	1.117	00:47	
Air Blank	0.000	00:48	
Control Test	0.076	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 11/10/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDASH,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GUTIERREZ,I
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/09/2023 20:53:56

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.220	20:49
Breath Volume	1.667	20:49
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.229	20:52
Breath Volume	2.449	20:52
Air Blank	0.000	20:53
Control Test	0.079	20:53
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 10/10/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,TRAVIS,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: WATSON,PATRICK
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0602280A4
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: RAMIREZ CASIANO

DATE/TIME OF TEST: 10/06/2023 06:58:23

Diagnostics Check	OK	06:55	
Air Blank	0.000	06:56	
Control Test	0.076	06:56	
Air Blank	0.000	06:56	
Subject Sample	# 1	06:57	Subject Test Refused
Breath Volume	0.000	06:57	
Air Blank	0.000	06:57	
Control Test	0.077	06:57	
Air Blank	0.000	06:58	
Diagnostics Check	OK	06:58	

Data Download Date/Time: 10/15/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS-PIERRE,RUBEN,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: BARRIENTOS,KENNY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 10/08/2023 05:34:28

Diagnostics Check	OK	05:27	
Air Blank	0.000	05:27	
Control Test	0.076	05:28	
Air Blank	0.000	05:28	
Subject Sample	# 1 0.248	05:29	
Breath Volume	1.351	05:29	
Air Blank	0.000	05:30	
Air Blank	0.000	05:31	
Subject Sample	# 2 0.269	05:33	
Breath Volume	1.476	05:33	
Air Blank		05:34	Purge Fail
Air Blank	0.000	05:34	

Data Download Date/Time: 10/13/2023 14:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLVERA,JESUS,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: RIVERA,SAMUEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 10/08/2023 19:01:40

Diagnostics Check	OK	18:58	
Air Blank	0.000	18:59	
Control Test	0.082	18:59	
Air Blank	0.000	19:00	
Subject Sample	# 1	19:00	Subject Test Refused
Breath Volume	0.000	19:00	
Air Blank	0.000	19:00	
Control Test	0.081	19:01	
Air Blank	0.000	19:01	
Diagnostics Check	OK	19:01	

Data Download Date/Time: 10/13/2023 14:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMCZYK, ANDREW, A
BREATH TEST OPERATOR AGENCY: Jefferson County SO
ARRESTING OFFICER: ADAMCZYK, ANDREW
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 09/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: SHIVER

DATE/TIME OF TEST: 10/02/2023 16:25:17

Diagnostics Check	OK	16:18
Air Blank	0.000	16:19
Control Test	0.080	16:19
Air Blank	0.000	16:20
Subject Sample # 1	0.000	16:20
Breath Volume	1.757	16:20
Air Blank	0.000	16:21
Air Blank	0.000	16:23
Subject Sample # 2	0.000	16:23
Breath Volume	1.664	16:23
Air Blank	0.000	16:24
Control Test	0.079	16:24
Air Blank	0.000	16:25
Diagnostics Check	OK	16:25

Data Download Date/Time: 10/16/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHEA,COLBY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROSMINI,JOE
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 09/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MAGSAM

DATE/TIME OF TEST: 10/10/2023 22:27:31

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.080	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.000	22:23
Breath Volume	2.015	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.000	22:26
Breath Volume	1.890	22:26
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 10/16/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORRIVEAU, CHAD, R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SMITH, STEWART
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 10/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: HUME FRYE

DATE/TIME OF TEST: 10/25/2023 18:16:01

Diagnostics Check	OK	18:09
Air Blank	0.000	18:10
Control Test	0.080	18:10
Air Blank	0.000	18:10
Subject Sample # 1	0.313	18:11
Breath Volume	1.578	18:11
Air Blank	0.000	18:12
Air Blank	0.000	18:13
Subject Sample # 2	0.312	18:14
Breath Volume	2.269	18:14
Air Blank	0.000	18:15
Control Test	0.079	18:15
Air Blank	0.000	18:15
Diagnostics Check	OK	18:16

Data Download Date/Time: 11/27/2023 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: VAUGHN,A
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 09/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/06/2023 20:11:47

Diagnostics Check	OK	20:05
Air Blank	0.000	20:05
Control Test	0.079	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.187	20:07
Breath Volume	2.027	20:07
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.204	20:10
Breath Volume	1.960	20:10
Air Blank	0.000	20:10
Control Test	0.079	20:11
Air Blank	0.000	20:11
Diagnostics Check	OK	20:11

Data Download Date/Time: 10/09/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,R,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BROWN,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/15/2023 01:49:27

Diagnostics Check	OK	01:49	
Air Blank		01:49	RFI Detect
Air Blank		01:49	RFI Detect

Data Download Date/Time: 11/14/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,R,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BROWN,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/15/2023 01:58:25

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.080	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.125	01:53
Breath Volume	1.730	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.125	01:56
Breath Volume	2.035	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 11/14/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSH, TYLER, C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: COURCHESNE, CAMRON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: PAREDES

DATE/TIME OF TEST: 10/18/2023 01:19:30

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:17	
Control Test	0.080	01:17	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:18	Subject Test Refused
Breath Volume	0.000	01:18	
Air Blank	0.000	01:18	
Control Test	0.080	01:18	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 11/14/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CLARK,AARON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: PETTY

DATE/TIME OF TEST: 10/18/2023 04:11:43

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.119	04:07
Breath Volume	2.152	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.124	04:10
Breath Volume	2.757	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 11/14/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS, WILLIAM, D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: YORK, C
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/20/2023 02:55:01

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.156	02:50
Breath Volume	1.726	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.160	02:53
Breath Volume	2.160	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:55

Data Download Date/Time: 11/14/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSH, TYLER, C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: VAZQUEZ, SAMUEL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 10/24/2023 23:12:45

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:08	
Control Test	0.079	23:08	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:11	Subject Test Refused
Breath Volume	0.000	23:11	
Air Blank	0.000	23:11	
Control Test	0.080	23:12	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 11/14/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG, JOSHUA, G
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MCCABE, PATRICK
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 10/25/2023 06:41:46

Diagnostics Check	OK	06:31	
Air Blank	0.000	06:32	
Control Test	0.081	06:32	
Air Blank	0.000	06:33	
Subject Sample	# 1	06:36	No Sample Provided
Breath Volume	0.000	06:36	
Air Blank	0.000	06:36	
Air Blank	0.000	06:38	
Subject Sample	# 2	06:40	Subject Test Refused
Breath Volume	0.000	06:40	
Air Blank	0.000	06:40	
Control Test	0.080	06:41	
Air Blank	0.000	06:41	
Diagnostics Check	OK	06:41	

Data Download Date/Time: 11/14/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAINES,ROBERT,L
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: GAINES,ROBERT
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 10/29/2023 03:08:29

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:07	RFI Detect
Breath Volume	0.000	03:07	
Air Blank		03:08	RFI Detect
Air Blank	0.000	03:08	

Data Download Date/Time: 11/14/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAINES,ROBERT,L
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: GAINES,ROBERT
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 10/29/2023 03:19:36

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.218	03:14
Breath Volume	1.632	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.229	03:18
Breath Volume	1.738	03:18
Air Blank	0.000	03:18
Control Test	0.077	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 11/14/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNOR,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: O'CONNOR,STEPHEN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 10/29/2023 04:22:23

Diagnostics Check	OK	04:21	
Air Blank		04:22	RFI Detect
Air Blank		04:22	RFI Detect

Data Download Date/Time: 11/14/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNOR,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: O'CONNOR,STEPHEN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 10/29/2023 04:24:44

Diagnostics Check	OK	04:24	
Air Blank		04:24	RFI Detect
Air Blank		04:24	RFI Detect

Data Download Date/Time: 11/14/2023 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAINES,CRAIG,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GAINES,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 10/04/2023 20:20:43

Diagnostics Check	OK	20:12	
Air Blank	0.000	20:12	
Control Test	0.081	20:13	
Air Blank	0.000	20:13	
Subject Sample	# 1 0.222	20:16	Volume Not Met
Breath Volume	0.503	20:16	
Air Blank	0.000	20:17	
Air Blank	0.000	20:19	
Subject Sample	# 2	20:19	Subject Test Refused
Breath Volume	0.000	20:19	
Air Blank	0.000	20:19	
Control Test	0.076	20:20	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 10/11/2023 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFORD,STERLING,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SAFFORD,STERLING
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/06/2023 23:31:31

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.137	23:27
Breath Volume	1.710	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.132	23:29
Breath Volume	2.105	23:29
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 10/11/2023 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ, RESEAN, R
BREATH TEST OPERATOR AGENCY: Midway PD
ARRESTING OFFICER: PIERRE, EDGARD
ARRESTING OFFICER AGENCY: Midway PD
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HERRING

DATE/TIME OF TEST: 10/21/2023 01:45:47

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Subject Sample # 1	0.173	01:42	
Breath Volume	2.144	01:42	
Air Blank	0.000	01:43	
Air Blank		01:45	Ambient Fail
Air Blank	0.000	01:45	

Data Download Date/Time: 11/21/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORAN,ETHAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FORAN,ETHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: COUSNARD

DATE/TIME OF TEST: 10/21/2023 16:37:23

Diagnostics Check	OK	16:30
Air Blank	0.000	16:30
Control Test	0.082	16:31
Air Blank	0.000	16:31
Subject Sample # 1	0.146	16:32
Breath Volume	2.093	16:32
Air Blank	0.000	16:33
Air Blank	0.000	16:35
Subject Sample # 2	0.145	16:35
Breath Volume	1.105	16:35
Air Blank	0.000	16:36
Control Test	0.081	16:36
Air Blank	0.000	16:37
Diagnostics Check	OK	16:37

Data Download Date/Time: 11/21/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFORD,STERLING,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SAFFORD,STERLING
ARRESTING OFFICER AGENCY: Gadsden County SO
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: CHASON

DATE/TIME OF TEST: 10/22/2023 06:13:07

Diagnostics Check	OK	06:06
Air Blank	0.000	06:06
Control Test	0.081	06:06
Air Blank	0.000	06:07
Subject Sample # 1	0.179	06:08
Breath Volume	3.144	06:08
Air Blank	0.000	06:09
Air Blank	0.000	06:10
Subject Sample # 2	0.178	06:11
Breath Volume	2.203	06:11
Air Blank	0.000	06:12
Control Test	0.079	06:12
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

Data Download Date/Time: 11/21/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALLOW, CHARLES, S
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MALLOW, CHARLES
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 09/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: BUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: ELLINOR

DATE/TIME OF TEST: 10/04/2023 21:23:08

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:12	
Control Test	0.081	21:12	
Air Blank	0.000	21:13	
Subject Sample	# 1	21:16	No Sample Provided
Breath Volume	0.000	21:16	
Air Blank	0.000	21:16	
Air Blank	0.000	21:18	
Subject Sample	# 2	21:21	No Sample Provided
Breath Volume	0.000	21:21	
Air Blank	0.000	21:22	
Control Test	0.081	21:22	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 10/19/2023 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASS,JEFFREY,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: RIVERS,AUSTIN
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 09/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 10/13/2023 05:43:23

Diagnostics Check	OK	05:37	
Air Blank	0.000	05:37	
Control Test	0.081	05:37	
Air Blank	0.000	05:38	
Subject Sample	# 1	05:41	Interferent Detect
Breath Volume	1.429	05:41	
Air Blank	0.000	05:42	
Control Test	0.084	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 10/19/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASS,JEFFREY,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: RIVERS,AUSTIN
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 09/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 10/13/2023 05:54:30

Diagnostics Check	OK	05:46	
Air Blank	0.000	05:47	
Control Test	0.081	05:47	
Air Blank	0.000	05:47	
Subject Sample	# 1 0.082	05:48	
Breath Volume	1.164	05:48	
Air Blank	0.000	05:49	
Air Blank	0.000	05:50	
Subject Sample	# 2	05:53	Interferent Detect
Breath Volume	1.703	05:53	
Air Blank		05:53	Purge Fail
Air Blank	0.000	05:54	

Data Download Date/Time: 10/19/2023 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASS,JEFFREY,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: RIVERS,AUSTIN
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 09/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 10/13/2023 06:07:01

Diagnostics Check	OK	06:02	
Air Blank	0.000	06:02	
Control Test	0.080	06:03	
Air Blank	0.000	06:03	
Subject Sample	# 1	06:05	Interferent Detect
Breath Volume	1.480	06:05	
Air Blank		06:06	Purge Fail
Air Blank	0.000	06:07	

Data Download Date/Time: 10/19/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: RIVERS,AUSTIN
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 09/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 10/13/2023 06:17:20

Diagnostics Check	OK	06:10
Air Blank	0.000	06:10
Control Test	0.081	06:10
Air Blank	0.000	06:11
Subject Sample # 1	0.075	06:12
Breath Volume	1.378	06:12
Air Blank	0.000	06:13
Air Blank	0.000	06:14
Subject Sample # 2	0.075	06:15
Breath Volume	1.652	06:15
Air Blank	0.000	06:16
Control Test	0.081	06:16
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 10/19/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,DUSTIN,A
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: TRUJILLO,ERNESTO
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 09/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: SKEENS

DATE/TIME OF TEST: 10/14/2023 22:17:24

Diagnostics Check	OK	22:13	
Air Blank	0.000	22:14	
Control Test	0.082	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 1	22:16	Subject Test Refused
Breath Volume	0.000	22:16	
Air Blank	0.000	22:16	
Control Test	0.082	22:16	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:17	

Data Download Date/Time: 10/19/2023 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: TOLES,CHARLES
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GODFREY

DATE/TIME OF TEST: 10/30/2023 02:35:39

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.082	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.081	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 11/14/2023 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIDEAUX, ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: KRISTON

DATE/TIME OF TEST: 10/14/2023 00:35:36

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:24	
Control Test	0.078	00:24	
Air Blank	0.000	00:25	
Subject Sample	# 1 0.081	00:28	Volume Not Met
Breath Volume	0.656	00:28	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 2 0.095	00:34	Volume Not Met
Breath Volume	0.460	00:34	
Air Blank	0.000	00:34	
Control Test	0.076	00:35	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 11/08/2023 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIDEAUX, ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: KRISTON

DATE/TIME OF TEST: 10/14/2023 00:42:58

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.078	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.097	00:38
Breath Volume	1.800	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.093	00:41
Breath Volume	1.675	00:41
Air Blank	0.000	00:42
Control Test	0.078	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 11/08/2023 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANIEL JR, CONNELL
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 10/15/2023 23:04:52

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.078	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.147	23:00
Breath Volume	1.308	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample # 2	0.152	23:03
Breath Volume	1.507	23:03
Air Blank	0.000	23:03
Control Test	0.078	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 11/08/2023 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CUEVAS, JOSE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: FORTUNE

DATE/TIME OF TEST: 10/28/2023 02:49:28

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.078	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.130	02:45
Breath Volume	2.058	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.130	02:48
Breath Volume	1.812	02:48
Air Blank	0.000	02:48
Control Test	0.078	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 11/08/2023 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHIPLEY, DANIEL
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 10/28/2023 23:13:24

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.078	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.245	23:09
Breath Volume	2.980	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.245	23:11
Breath Volume	2.488	23:11
Air Blank	0.000	23:12
Control Test	0.078	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 11/08/2023 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VAZQUEZ,RAUL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: CARDONA HERNANDEZ

DATE/TIME OF TEST: 10/01/2023 00:17:55

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.083	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.138	00:11
Breath Volume	1.507	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.138	00:16
Breath Volume	2.496	00:16
Air Blank	0.000	00:17
Control Test	0.083	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILAN,VICTOR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MILAN,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 11:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: ORGANISTA SANCHEZ

DATE/TIME OF TEST: 10/01/2023 01:31:08

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.083	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.171	01:26
Breath Volume	3.320	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:28
Subject Sample # 2	0.173	01:29
Breath Volume	3.378	01:29
Air Blank	0.000	01:30
Control Test	0.083	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FLORES-ALDRICH

DATE/TIME OF TEST: 10/01/2023 03:07:50

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.083	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.000	03:03
Breath Volume	2.796	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.000	03:06
Breath Volume	1.773	03:06
Air Blank	0.000	03:06
Control Test	0.083	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BLASSINGAME,ALISSA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/01/2023 03:27:01

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.083	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.178	03:22
Breath Volume	2.328	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.179	03:25
Breath Volume	1.910	03:25
Air Blank	0.000	03:26
Control Test	0.083	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:27

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: LOPEZ-RODRIGUEZ

DATE/TIME OF TEST: 10/02/2023 03:59:25

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.083	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.171	03:54
Breath Volume	2.488	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.158	03:57
Breath Volume	1.804	03:57
Air Blank	0.000	03:58
Control Test	0.082	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: SOTO ORTIZ

DATE/TIME OF TEST: 10/07/2023 00:37:06

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.083	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.125	00:32
Breath Volume	1.972	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.134	00:35
Breath Volume	2.265	00:35
Air Blank	0.000	00:36
Control Test	0.083	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: ELDER,MICHEAL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: GUZMAN AMARO

DATE/TIME OF TEST: 10/07/2023 01:32:19

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.083	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.173	01:27
Breath Volume	1.777	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.167	01:30
Breath Volume	1.542	01:30
Air Blank	0.000	01:31
Control Test	0.083	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ROSARIO-MELO

DATE/TIME OF TEST: 10/08/2023 03:28:51

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.084	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.149	03:24
Breath Volume	2.164	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.149	03:27
Breath Volume	2.414	03:27
Air Blank	0.000	03:27
Control Test	0.083	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/09/2023 00:45:36

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.084	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.220	00:40
Breath Volume	1.449	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.219	00:44
Breath Volume	1.367	00:44
Air Blank	0.000	00:44
Control Test	0.083	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SCANLAN,JAMES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: COSTA

DATE/TIME OF TEST: 10/09/2023 19:14:44

Diagnostics Check	OK	19:11	
Air Blank	0.000	19:11	
Control Test	0.083	19:11	
Air Blank	0.000	19:12	
Subject Sample	# 1 0.117	19:13	
Breath Volume	2.339	19:13	
Air Blank		19:13	RFI Detect
Air Blank	0.000	19:13	
Control Test	0.082	19:14	
Air Blank	0.000	19:14	
Diagnostics Check	OK	19:14	

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SCANLAN,JAMES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: COSTA

DATE/TIME OF TEST: 10/09/2023 19:22:24

Diagnostics Check	OK	19:16
Air Blank	0.000	19:16
Control Test	0.084	19:17
Air Blank	0.000	19:17
Subject Sample # 1	0.111	19:17
Breath Volume	2.277	19:17
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.107	19:20
Breath Volume	2.347	19:20
Air Blank	0.000	19:21
Control Test	0.084	19:21
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: JACKSON,ALLEN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/12/2023 05:30:32

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.084	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.124	05:26
Breath Volume	1.593	05:26
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.123	05:29
Breath Volume	1.378	05:29
Air Blank	0.000	05:29
Control Test	0.083	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GARCIA,CORINA
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FARIA-PEREZ

DATE/TIME OF TEST: 10/14/2023 03:47:51

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.084	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.062	03:43
Breath Volume	2.437	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.068	03:46
Breath Volume	3.500	03:46
Air Blank	0.000	03:46
Control Test	0.083	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,SHAWN,N
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WOODS,SHAWN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 07:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:02

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 10/14/2023 10:45:46

Diagnostics Check	OK	10:39
Air Blank	0.000	10:39
Control Test	0.083	10:40
Air Blank	0.000	10:40
Subject Sample # 1	0.139	10:41
Breath Volume	1.261	10:41
Air Blank	0.000	10:42
Air Blank	0.000	10:43
Subject Sample # 2	0.138	10:44
Breath Volume	1.878	10:44
Air Blank	0.000	10:44
Control Test	0.083	10:45
Air Blank	0.000	10:45
Diagnostics Check	OK	10:45

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: GUADALUPE, GABRIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DAVILA VELAZQUEZ

DATE/TIME OF TEST: 10/15/2023 01:40:13

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.084	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.199	01:34
Breath Volume	2.519	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.185	01:38
Breath Volume	1.554	01:38
Air Blank	0.000	01:39
Control Test	0.078	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 10/23/2023 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: NIEVES, CHARLIE
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: QUILES FUENTES

DATE/TIME OF TEST: 10/15/2023 04:24:30

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:20	
Control Test	0.083	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:22	RFI Detect
Breath Volume	1.062	04:22	
Air Blank	0.000	04:23	
Control Test	0.083	04:23	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 10/23/2023 15:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: NIEVES, CHARLIE
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: QUILES FUNETES

DATE/TIME OF TEST: 10/15/2023 04:35:43

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.081	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.103	04:30
Breath Volume	1.890	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.096	04:34
Breath Volume	2.117	04:34
Air Blank	0.000	04:34
Control Test	0.082	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 10/23/2023 15:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VAZQUEZ,ELIEZER
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: PACHECHO-RODRIGUEZ

DATE/TIME OF TEST: 10/16/2023 01:47:38

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.083	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.194	01:43
Breath Volume	1.664	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.195	01:46
Breath Volume	1.804	01:46
Air Blank	0.000	01:46
Control Test	0.081	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 10/23/2023 15:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/19/2023 23:59:33

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.084	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.038	23:55
Breath Volume	1.433	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.037	23:58
Breath Volume	1.691	23:58
Air Blank	0.000	23:58
Control Test	0.084	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 10/23/2023 15:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: NIEVES, CHARLIE
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: RODRIGUEZ SOTOMAYOR

DATE/TIME OF TEST: 10/20/2023 03:42:32

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.084	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.190	03:37
Breath Volume	3.242	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.200	03:40
Breath Volume	4.246	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 10/23/2023 15:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILAN,VICTOR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MILAN,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: VICHES

DATE/TIME OF TEST: 10/22/2023 00:43:00

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.083	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.182	00:38
Breath Volume	1.968	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample # 2	0.181	00:41
Breath Volume	2.183	00:41
Air Blank	0.000	00:42
Control Test	0.083	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:43

Data Download Date/Time: 10/23/2023 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAMINSKI,KELSEY,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: CLERMONT

DATE/TIME OF TEST: 10/22/2023 06:14:51

Diagnostics Check	OK	06:11	
Air Blank	0.000	06:12	
Control Test	0.083	06:12	
Air Blank	0.000	06:13	
Subject Sample	# 1	06:13	Subject Test Refused
Breath Volume	0.000	06:13	
Air Blank	0.000	06:13	
Control Test	0.083	06:14	
Air Blank	0.000	06:14	
Diagnostics Check	OK	06:14	

Data Download Date/Time: 10/23/2023 15:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: HELLER,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: WILLITS

DATE/TIME OF TEST: 10/23/2023 02:45:44

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.084	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.164	02:40
Breath Volume	2.789	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.160	02:44
Breath Volume	2.722	02:44
Air Blank	0.000	02:44
Control Test	0.083	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 10/23/2023 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GLADDEN

DATE/TIME OF TEST: 10/24/2023 01:45:44

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.083	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1 0.035	01:33	Volume Not Met
Breath Volume	0.929	01:33	
Air Blank	0.000	01:33	
Air Blank	0.000	01:35	
Subject Sample	# 2 0.036	01:38	
Breath Volume	1.300	01:38	
Air Blank	0.000	01:39	
Air Blank	0.000	01:41	
Subject Sample	# 3 0.033	01:44	Volume Not Met
Breath Volume	0.882	01:44	
Air Blank	0.000	01:44	
Control Test	0.082	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 11/22/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: GANT

DATE/TIME OF TEST: 10/24/2023 21:19:30

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.084	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.000	21:15
Breath Volume	2.296	21:15
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.000	21:18
Breath Volume	2.277	21:18
Air Blank	0.000	21:18
Control Test	0.084	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 11/22/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ-FELICIANO,DERICK
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 10/25/2023 02:48:38

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.084	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.000	02:44
Breath Volume	3.234	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.000	02:47
Breath Volume	2.875	02:47
Air Blank	0.000	02:47
Control Test	0.083	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 11/22/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ARROYO,ALEJANDRO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: RIVERA OQUENDO

DATE/TIME OF TEST: 10/26/2023 03:50:54

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.083	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.139	03:46
Breath Volume	1.566	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.124	03:49
Breath Volume	1.527	03:49
Air Blank	0.000	03:50
Control Test	0.083	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 11/22/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,CARLOS,I
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TURNER,RODNEY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: JAMESON

DATE/TIME OF TEST: 10/26/2023 20:32:15

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.083	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.162	20:27
Breath Volume	2.210	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.171	20:30
Breath Volume	1.925	20:30
Air Blank	0.000	20:31
Control Test	0.083	20:31
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 11/22/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: PEREZ-LOPEZ

DATE/TIME OF TEST: 10/28/2023 04:47:46

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.083	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.209	04:40
Breath Volume	1.144	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.209	04:46
Breath Volume	1.226	04:46
Air Blank	0.000	04:46
Control Test	0.082	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 11/22/2023 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ARCHIBALD

DATE/TIME OF TEST: 10/29/2023 01:18:22

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.082	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.186	01:13
Breath Volume	1.472	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.190	01:16
Breath Volume	1.597	01:16
Air Blank	0.000	01:17
Control Test	0.083	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 11/22/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: ROSADO-BAEZ

DATE/TIME OF TEST: 10/29/2023 23:30:03

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.083	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.070	23:25
Breath Volume	2.156	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:27
Subject Sample # 2	0.069	23:28
Breath Volume	1.882	23:28
Air Blank	0.000	23:29
Control Test	0.083	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:30

Data Download Date/Time: 11/22/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GUADALUPE,GABRIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: VASQUEZ-MORALES

DATE/TIME OF TEST: 10/29/2023 23:48:14

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.083	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.123	23:43
Breath Volume	1.160	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:45
Subject Sample # 2	0.122	23:46
Breath Volume	1.160	23:46
Air Blank	0.000	23:47
Control Test	0.083	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 11/22/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 10/30/2023 02:11:59

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.084	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.000	02:07
Breath Volume	2.101	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.000	02:10
Breath Volume	2.699	02:10
Air Blank	0.000	02:11
Control Test	0.084	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 11/22/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ-FELCIANO,DERICK
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/30/2023 03:27:23

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.084	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.148	03:22
Breath Volume	2.542	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.151	03:25
Breath Volume	1.617	03:25
Air Blank	0.000	03:26
Control Test	0.083	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 11/22/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,BRIAN,Z
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,BRIAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 10/04/2023 03:06:38

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.000	03:01
Breath Volume	2.449	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.000	03:05
Breath Volume	1.683	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 10/30/2023 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASPER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: JASPER,ANDREW
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 10:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:41

SUBJECT'S LAST NAME: VERSAGGI

DATE/TIME OF TEST: 10/07/2023 11:12:45

Diagnostics Check	OK	11:06
Air Blank	0.000	11:06
Control Test	0.079	11:07
Air Blank	0.000	11:07
Subject Sample # 1	0.000	11:08
Breath Volume	1.304	11:08
Air Blank	0.000	11:08
Air Blank	0.000	11:10
Subject Sample # 2	0.000	11:11
Breath Volume	1.414	11:11
Air Blank	0.000	11:11
Control Test	0.079	11:12
Air Blank	0.000	11:12
Diagnostics Check	OK	11:12

Data Download Date/Time: 10/30/2023 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,BRIAN,Z
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WIGGINS,AUSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/21/2023 00:52:06

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.079	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.163	00:47
Breath Volume	1.949	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:49
Subject Sample # 2	0.160	00:50
Breath Volume	1.578	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 10/30/2023 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMELMAN, TYLER,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAMMELMAN, TYLER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: FRIEND

DATE/TIME OF TEST: 10/23/2023 23:02:53

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.065	22:58
Breath Volume	1.285	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.061	23:01
Breath Volume	2.656	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 10/30/2023 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONNEMANN, MARK, W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HORNSTEIN, RANDY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: TOWASNICKI

DATE/TIME OF TEST: 10/29/2023 04:21:35

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:16	
Control Test	0.079	04:17	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:19	RFI Detect
Breath Volume	0.015	04:19	
Air Blank	0.000	04:20	
Control Test	0.079	04:20	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 10/30/2023 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONNEMANN, MARK, W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HORNSTEIN, RANDY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: TOWASNICKI

DATE/TIME OF TEST: 10/29/2023 04:35:38

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:24	
Control Test	0.077	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 1 0.178	04:28	Volume Not Met
Breath Volume	0.109	04:28	
Air Blank	0.000	04:29	
Air Blank	0.000	04:30	
Subject Sample	# 2 0.169	04:34	Volume Not Met
Breath Volume	0.191	04:34	
Air Blank	0.000	04:34	
Control Test	0.078	04:35	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 10/30/2023 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONNEMANN, MARK, W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HORNSTEIN, RANDY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: TOWASNICKI

DATE/TIME OF TEST: 10/29/2023 04:54:37

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:43	
Control Test	0.078	04:43	
Air Blank	0.000	04:44	
Subject Sample	# 1	04:47	No Sample Provided
Breath Volume	0.000	04:47	
Air Blank	0.000	04:47	
Air Blank	0.000	04:49	
Subject Sample	# 2	04:53	Volume Not Met
Breath Volume	0.328	04:53	
Air Blank	0.000	04:53	
Control Test	0.078	04:53	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 10/30/2023 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/03/2023 02:40:01

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.000	02:35
Breath Volume	1.472	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.000	02:38
Breath Volume	2.664	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:40

Data Download Date/Time: 10/23/2023 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: JAROSZEWSKI
PEREIRA,THIAGO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: KAYLOR

DATE/TIME OF TEST: 10/03/2023 18:47:09

Diagnostics Check	OK	18:40
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.139	18:42
Breath Volume	4.769	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.134	18:45
Breath Volume	5.015	18:45
Air Blank	0.000	18:46
Control Test	0.079	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 10/23/2023 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BURGOS VERGARA

DATE/TIME OF TEST: 10/20/2023 23:49:19

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.166	23:44
Breath Volume	2.988	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.159	23:47
Breath Volume	1.289	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 10/23/2023 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLYTH,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: HENRIQUEZ,ENRIQUE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 10/22/2023 22:43:37

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.080	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.146	22:38
Breath Volume	1.187	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:40
Subject Sample # 2	0.151	22:42
Breath Volume	2.027	22:42
Air Blank	0.000	22:42
Control Test	0.077	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 10/23/2023 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 042
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: FRITZ

DATE/TIME OF TEST: 10/01/2023 03:43:43

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.008	03:39
Breath Volume	3.519	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.000	03:42
Breath Volume	3.441	03:42
Air Blank	0.000	03:42
Control Test	0.077	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 11/02/2023 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLARD,SEAN,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BULLARD,SEAN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 12/12/2033
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: VORONOVICH

DATE/TIME OF TEST: 10/10/2023 03:02:58

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.077	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.152	02:58
Breath Volume	1.632	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.158	03:01
Breath Volume	1.851	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 11/02/2023 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 042
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: RECTOR

DATE/TIME OF TEST: 10/15/2023 00:57:08

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.078	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:50	No Sample Provided
Breath Volume	0.000	00:50	
Air Blank	0.000	00:50	
Air Blank	0.000	00:52	
Subject Sample	# 2	00:55	No Sample Provided
Breath Volume	0.000	00:55	
Air Blank	0.000	00:56	
Control Test	0.079	00:56	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

Data Download Date/Time: 11/02/2023 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAT, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: MORRIS, KIENEN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2
DRY GAS EXPIRY DATE: 12/12/2033
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MONTROSE

DATE/TIME OF TEST: 10/27/2023 00:11:52

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.122	00:07
Breath Volume	1.589	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.126	00:10
Breath Volume	1.753	00:10
Air Blank	0.000	00:10
Control Test	0.078	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 11/02/2023 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 042
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 10/28/2023 21:24:50

Diagnostics Check	OK	21:18
Air Blank	0.000	21:19
Control Test	0.077	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.233	21:20
Breath Volume	4.179	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:22
Subject Sample # 2	0.223	21:23
Breath Volume	2.199	21:23
Air Blank	0.000	21:23
Control Test	0.076	21:24
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 11/02/2023 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 042
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: ROSA ORTIZ

DATE/TIME OF TEST: 10/29/2023 02:05:54

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.078	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.269	02:01
Breath Volume	1.652	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:03
Subject Sample # 2	0.269	02:04
Breath Volume	2.335	02:04
Air Blank	0.000	02:05
Control Test	0.075	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 11/02/2023 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BRANTLEY

DATE/TIME OF TEST: 10/05/2023 04:10:40

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.084	04:01	
Air Blank	0.000	04:01	
Subject Sample # 1	0.182	04:04	
Breath Volume	1.542	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:07	
Subject Sample # 2	0.180	04:08	
Breath Volume	2.687	04:08	
Air Blank	0.000	04:09	
Control Test	0.082	04:09	
Air Blank	0.000	04:10	
Diagnostics Check	Fail	04:10	Temperature Regulation Test
Air Blank	0.000	04:10	

Data Download Date/Time: 11/07/2023 18:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BRANTLEY

DATE/TIME OF TEST: 10/05/2023 04:18:32

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.083	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:17	Subject Test Refused
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Control Test	0.083	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 11/07/2023 18:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BRANTLEY

DATE/TIME OF TEST: 10/05/2023 04:46:24

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.083	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.174	04:41
Breath Volume	2.296	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.162	04:44
Breath Volume	1.500	04:44
Air Blank	0.000	04:45
Control Test	0.083	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 11/07/2023 18:51