

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN, VALERIE, C
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MARQUEZ, GABRIEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: KWON

DATE/TIME OF TEST: 10/07/2023 05:23:10

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.080	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.161	05:18
Breath Volume	1.300	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:20
Subject Sample # 2	0.168	05:21
Breath Volume	2.332	05:21
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 12/19/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MEZINORD,VILINOR
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 10/11/2023 02:50:56

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.067	02:46
Breath Volume	2.777	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.074	02:49
Breath Volume	2.921	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 12/19/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TORRES,LYDIA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: SHEPARD

DATE/TIME OF TEST: 10/15/2023 16:57:32

Diagnostics Check	OK	16:47	
Air Blank	0.000	16:47	
Control Test	0.081	16:48	
Air Blank	0.000	16:48	
Subject Sample	# 1 0.246	16:51	
Breath Volume	1.640	16:51	
Air Blank	0.000	16:52	
Air Blank	0.000	16:53	
Subject Sample	# 2	16:55	Subject Test Refused
Breath Volume	0.109	16:55	
Air Blank	0.000	16:56	
Control Test	0.078	16:56	
Air Blank	0.000	16:57	
Diagnostics Check	OK	16:57	

Data Download Date/Time: 12/19/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: PREISS,ISABELLA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: KUSHNER

DATE/TIME OF TEST: 10/18/2023 19:22:42

Diagnostics Check	OK	19:16
Air Blank	0.000	19:16
Control Test	0.080	19:16
Air Blank	0.000	19:17
Subject Sample # 1	0.226	19:18
Breath Volume	2.781	19:18
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.225	19:20
Breath Volume	2.593	19:20
Air Blank	0.000	19:21
Control Test	0.080	19:22
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

Data Download Date/Time: 12/19/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: PREISS,ISABELLA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 10/20/2023 01:21:10

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.000	01:16
Breath Volume	3.835	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.000	01:19
Breath Volume	3.976	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 12/19/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN, VALERIE, C
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: COLON, JOSEPH
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:32

SUBJECT'S LAST NAME: GROTTON

DATE/TIME OF TEST: 10/21/2023 07:00:57

Diagnostics Check	OK	06:54
Air Blank	0.000	06:54
Control Test	0.080	06:55
Air Blank	0.000	06:55
Subject Sample # 1	0.167	06:56
Breath Volume	1.726	06:56
Air Blank	0.000	06:56
Air Blank	0.000	06:58
Subject Sample # 2	0.173	06:59
Breath Volume	1.597	06:59
Air Blank	0.000	07:00
Control Test	0.079	07:00
Air Blank	0.000	07:00
Diagnostics Check	OK	07:00

Data Download Date/Time: 12/19/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLE,DARRELL,E
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ROLLE,DARRELL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: BUI
ARREST TIME: 09:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:43

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 10/22/2023 10:25:09

Diagnostics Check	OK	10:13	
Air Blank	0.000	10:13	
Control Test	0.080	10:13	
Air Blank	0.000	10:14	
Subject Sample	# 1 0.163	10:17	Volume Not Met
Breath Volume	0.167	10:17	
Air Blank	0.000	10:18	
Air Blank	0.000	10:20	
Subject Sample	# 2 0.191	10:23	Volume Not Met
Breath Volume	0.550	10:23	
Air Blank	0.000	10:24	
Control Test	0.077	10:24	
Air Blank	0.000	10:25	
Diagnostics Check	OK	10:25	

Data Download Date/Time: 12/19/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GRABILL,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/24/2023 03:22:05

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.097	03:15
Breath Volume	1.679	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.091	03:20
Breath Volume	1.781	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 12/19/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WOODALL,HANK
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: BEAL

DATE/TIME OF TEST: 10/29/2023 21:43:46

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.186	21:39
Breath Volume	1.953	21:39
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.186	21:42
Breath Volume	2.277	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 12/19/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLISON,PATRICK,W
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: ALLISON,PATRICK
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: HIBBS

DATE/TIME OF TEST: 10/10/2023 16:58:59

Diagnostics Check	OK	16:50	
Air Blank	0.000	16:50	
Control Test	0.079	16:50	
Air Blank	0.000	16:51	
Subject Sample	# 1 0.261	16:54	Volume Not Met
Breath Volume	0.351	16:54	
Air Blank	0.000	16:55	
Air Blank	0.000	16:57	
Subject Sample	# 2	16:57	Subject Test Refused
Breath Volume	0.000	16:57	
Air Blank	0.000	16:58	
Control Test	0.078	16:58	
Air Blank	0.000	16:58	
Diagnostics Check	OK	16:58	

Data Download Date/Time: 10/31/2023 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WALTER,CRAIG
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: LEDY

DATE/TIME OF TEST: 10/13/2023 21:56:04

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.303	21:47
Breath Volume	2.308	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.276	21:51
Breath Volume	1.175	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:53
Subject Sample # 3	0.289	21:54
Breath Volume	2.289	21:54
Air Blank	0.000	21:55
Control Test	0.077	21:55
Air Blank	0.000	21:55
Diagnostics Check	OK	21:56

Data Download Date/Time: 10/31/2023 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,LAMAR,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: RIDEOUT,TIMOTHY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: WINFREY

DATE/TIME OF TEST: 10/15/2023 21:36:28

Diagnostics Check	OK	21:33	
Air Blank	0.000	21:34	
Control Test	0.080	21:34	
Air Blank	0.000	21:34	
Subject Sample	# 1	21:35	Subject Test Refused
Breath Volume	0.000	21:35	
Air Blank	0.000	21:35	
Control Test	0.080	21:35	
Air Blank	0.000	21:36	
Diagnostics Check	OK	21:36	

Data Download Date/Time: 10/31/2023 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: LOVELESS,PAYTON
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 10/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: GOMEZ CASTANEDA

DATE/TIME OF TEST: 10/29/2023 02:53:11

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.079	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 10/31/2023 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE,KYLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: TROXELL

DATE/TIME OF TEST: 10/01/2023 23:33:32

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.077	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.113	23:28
Breath Volume	1.812	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.110	23:32
Breath Volume	1.652	23:32
Air Blank	0.000	23:32
Control Test	0.078	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 10/30/2023 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LULU,ABD,U
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LIPTAK,LAURINE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: GOMEZ PUCHADES

DATE/TIME OF TEST: 10/04/2023 19:02:12

Diagnostics Check	OK	18:55
Air Blank	0.000	18:56
Control Test	0.078	18:56
Air Blank	0.000	18:57
Subject Sample # 1	0.272	18:57
Breath Volume	3.187	18:57
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.284	19:00
Breath Volume	1.890	19:00
Air Blank	0.000	19:01
Control Test	0.077	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: LEONE

DATE/TIME OF TEST: 10/04/2023 22:02:51

Diagnostics Check	OK	21:59	
Air Blank	0.000	22:00	
Control Test	0.079	22:00	
Air Blank	0.000	22:00	
Subject Sample	# 1	22:02	Subject Test Refused
Breath Volume	0.000	22:02	
Air Blank		22:02	RFI Detect
Air Blank	0.000	22:02	

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: LEONE

DATE/TIME OF TEST: 10/04/2023 22:08:02

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:05	
Control Test	0.079	22:05	
Air Blank	0.000	22:06	
Subject Sample	# 1	22:06	Subject Test Refused
Breath Volume	0.000	22:06	
Air Blank	0.000	22:07	
Control Test	0.079	22:07	
Air Blank	0.000	22:07	
Diagnostics Check	OK	22:08	

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ARTEAGA

DATE/TIME OF TEST: 10/06/2023 03:38:09

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.078	03:35	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Control Test	0.078	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICHMAN,MICHELLE,K
BREATH TEST OPERATOR AGENCY: Bunnell PD
ARRESTING OFFICER: WICHMAN,MICHELLE
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: F260293648490

DATE/TIME OF TEST: 10/08/2023 21:13:43

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.078	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.151	21:09
Breath Volume	1.398	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.150	21:12
Breath Volume	1.851	21:12
Air Blank	0.000	21:12
Control Test	0.079	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,D,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: QUINTIERI,D
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: SAMUEL

DATE/TIME OF TEST: 10/09/2023 21:58:01

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.078	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.132	21:53
Breath Volume	1.406	21:53
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.125	21:56
Breath Volume	1.351	21:56
Air Blank	0.000	21:57
Control Test	0.078	21:57
Air Blank	0.000	21:57
Diagnostics Check	OK	21:58

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: TOWNE

DATE/TIME OF TEST: 10/10/2023 20:29:02

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.079	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.161	20:24
Breath Volume	1.132	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:26
Subject Sample # 2	0.165	20:27
Breath Volume	1.683	20:27
Air Blank	0.000	20:28
Control Test	0.078	20:28
Air Blank	0.000	20:28
Diagnostics Check	OK	20:29

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: RUFFALO

DATE/TIME OF TEST: 10/12/2023 19:48:41

Diagnostics Check	OK	19:42
Air Blank	0.000	19:42
Control Test	0.078	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.108	19:44
Breath Volume	2.566	19:44
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.106	19:47
Breath Volume	3.152	19:47
Air Blank	0.000	19:47
Control Test	0.078	19:48
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LULU,ABD,U
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: FANSLER,MICHEAL
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 10/14/2023 23:30:45

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.078	23:20	
Air Blank	0.000	23:21	
Subject Sample	# 1 0.157	23:24	Volume Not Met
Breath Volume	0.539	23:24	
Air Blank	0.000	23:25	
Air Blank	0.000	23:27	
Subject Sample	# 2	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Control Test	0.075	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MARSAN,NECOLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 10/15/2023 01:33:14

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.077	01:31	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.077	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MARVEL

DATE/TIME OF TEST: 10/15/2023 02:46:34

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.078	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.079	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MARSAN,NECOLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: MORETTI

DATE/TIME OF TEST: 10/15/2023 21:03:40

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
Control Test	0.078	21:01	
Air Blank	0.000	21:02	
Subject Sample	# 1	21:02	Subject Test Refused
Breath Volume	0.000	21:02	
Air Blank	0.000	21:02	
Control Test	0.078	21:03	
Air Blank	0.000	21:03	
Diagnostics Check	OK	21:03	

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI,PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: EVANS-GREEN

DATE/TIME OF TEST: 10/21/2023 01:11:47

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.078	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.117	01:07
Breath Volume	2.085	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.116	01:10
Breath Volume	1.734	01:10
Air Blank	0.000	01:10
Control Test	0.078	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 10/30/2023 15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIFANO,ANTHONY,S
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KEATON,RAY
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 10/22/2023 01:27:52

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.078	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.165	01:20
Breath Volume	1.835	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample # 2	0.187	01:23
Breath Volume	2.070	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:25
Subject Sample # 3	0.180	01:26
Breath Volume	1.816	01:26
Air Blank	0.000	01:26
Control Test	0.076	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 10/30/2023 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISS,ADAM,B
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BISS,ADAM
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: DILLARD

DATE/TIME OF TEST: 10/22/2023 02:23:08

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.078	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:22	
Control Test	0.079	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 10/30/2023 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LULU,ABD,U
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SCHRAGER,OLIVIA
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 10/27/2023 19:56:10

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.078	19:50
Air Blank	0.000	19:50
Subject Sample # 1	0.171	19:51
Breath Volume	1.976	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.171	19:54
Breath Volume	1.273	19:54
Air Blank	0.000	19:55
Control Test	0.076	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 10/30/2023 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LULU,ABD,U
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PRENTISS,HAYLEIGH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/27/2023 20:31:58

Diagnostics Check	OK	20:29	
Air Blank	0.000	20:29	
Control Test	0.078	20:29	
Air Blank	0.000	20:30	
Subject Sample	# 1	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:31	
Control Test	0.078	20:31	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 10/30/2023 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: HERRELL

DATE/TIME OF TEST: 10/28/2023 01:07:34

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.079	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:06	Subject Test Refused
Breath Volume	0.000	01:06	
Air Blank	0.000	01:06	
Control Test	0.079	01:06	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 10/30/2023 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/29/2023 04:52:10

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.079	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.141	04:47
Breath Volume	4.347	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.139	04:50
Breath Volume	4.214	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 10/30/2023 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/02/2023 01:39:30

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.143	01:34
Breath Volume	1.886	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.139	01:37
Breath Volume	1.531	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 10/09/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MCKUZES

DATE/TIME OF TEST: 10/02/2023 04:19:21

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.138	04:13
Breath Volume	1.394	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.136	04:17
Breath Volume	1.570	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 10/09/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: DELISI

DATE/TIME OF TEST: 10/04/2023 01:45:38

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Subject Sample # 1	0.000	01:43	
Breath Volume	2.113	01:43	
Air Blank		01:44	RFI Detect
Air Blank	0.000	01:44	
Control Test	0.080	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 10/09/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: DELISI

DATE/TIME OF TEST: 10/04/2023 01:53:27

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.000	01:49
Breath Volume	1.289	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.000	01:52
Breath Volume	1.160	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 10/09/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: CHATELAIN

DATE/TIME OF TEST: 10/06/2023 23:47:52

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:33	
Control Test	0.080	23:33	
Air Blank	0.000	23:33	
Subject Sample	# 1 0.046	23:34	
Breath Volume	1.273	23:34	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample	# 2 0.040	23:40	Volume Not Met
Breath Volume	1.027	23:40	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample	# 3 0.040	23:46	Volume Not Met
Breath Volume	0.500	23:46	
Air Blank	0.000	23:46	
Control Test	0.076	23:47	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 10/09/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: CHATELAIN

DATE/TIME OF TEST: 10/06/2023 23:58:45

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.078	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.040	23:53
Breath Volume	1.246	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.038	23:57
Breath Volume	1.175	23:57
Air Blank	0.000	23:57
Control Test	0.078	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 10/09/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANDRE,
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: MARCHESE,JOSEPH
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: WANECK

DATE/TIME OF TEST: 10/29/2023 18:27:02

Diagnostics Check	OK	18:20
Air Blank	0.000	18:21
Control Test	0.079	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.192	18:22
Breath Volume	1.566	18:22
Air Blank	0.000	18:23
Air Blank	0.000	18:24
Subject Sample # 2	0.185	18:25
Breath Volume	1.265	18:25
Air Blank	0.000	18:26
Control Test	0.079	18:26
Air Blank	0.000	18:26
Diagnostics Check	OK	18:27

Data Download Date/Time: 12/01/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WONSCH,JESSE,W
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BEYER,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 10/01/2023 21:14:19

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.082	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.175	21:09
Breath Volume	1.117	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.174	21:12
Breath Volume	1.195	21:12
Air Blank	0.000	21:13
Control Test	0.081	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 10/09/2023 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD, N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DIXON, JAMES
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: JUSTICE

DATE/TIME OF TEST: 10/08/2023 23:32:38

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.082	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.246	23:28
Breath Volume	1.441	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.250	23:31
Breath Volume	1.773	23:31
Air Blank	0.000	23:31
Control Test	0.080	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 10/09/2023 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSH, TYLER, C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: VAZQUEZ, SAMUEL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 10/24/2023 23:06:50

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.082	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.131	23:02
Breath Volume	1.570	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.138	23:05
Breath Volume	2.187	23:05
Air Blank	0.000	23:05
Control Test	0.082	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 11/14/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNOR,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: O'CONNOR,STEPHEN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 10/29/2023 04:34:03

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.082	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.183	04:29
Breath Volume	2.050	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.179	04:32
Breath Volume	2.179	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:34

Data Download Date/Time: 11/14/2023 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: OGLE,SHAWN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: PALANZA

DATE/TIME OF TEST: 10/20/2023 21:49:41

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:47	
Control Test	0.081	21:47	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:48	
Control Test	0.080	21:49	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 10/23/2023 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GARNER,MICHAEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: LOHMAN

DATE/TIME OF TEST: 10/21/2023 16:50:01

Diagnostics Check	OK	16:45	
Air Blank	0.000	16:46	
Control Test	0.080	16:46	
Air Blank	0.000	16:47	
Subject Sample	# 1	16:48	Subject Test Refused
Breath Volume	0.000	16:48	
Air Blank	0.000	16:49	
Control Test	0.081	16:49	
Air Blank	0.000	16:49	
Diagnostics Check	OK	16:50	

Data Download Date/Time: 10/23/2023 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PABON,G,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: MORALES,K

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 04:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: CRUMBO

DATE/TIME OF TEST: 10/01/2023 04:59:53

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.080	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.000	04:55
Breath Volume	2.558	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.000	04:58
Breath Volume	2.703	04:58
Air Blank	0.000	04:58
Control Test	0.078	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 10/27/2023 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: APONTE,A

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 17:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: MEUTH

DATE/TIME OF TEST: 10/02/2023 18:24:43

Diagnostics Check	OK	18:11	
Air Blank	0.000	18:11	
Control Test	0.080	18:12	
Air Blank	0.000	18:12	
Subject Sample	# 1 0.258	18:15	Volume Not Met
Breath Volume	0.015	18:15	
Air Blank	0.000	18:16	
Air Blank	0.000	18:18	
Subject Sample	# 2 0.334	18:20	
Breath Volume	1.519	18:20	
Air Blank	0.000	18:21	
Air Blank	0.000	18:22	
		18:23	Improper Sample
Air Blank	0.000	18:23	
Control Test	0.076	18:24	
Air Blank	0.000	18:24	
Diagnostics Check	OK	18:24	

Data Download Date/Time: 10/27/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: APONTE,A

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 17:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: MEUTH

DATE/TIME OF TEST: 10/02/2023 18:41:32

Diagnostics Check	OK	18:27	
Air Blank	0.000	18:28	
Control Test	0.079	18:28	
Air Blank	0.000	18:29	
Subject Sample	# 1 0.331	18:31	
Breath Volume	1.296	18:31	
Air Blank	0.000	18:31	
Air Blank	0.000	18:33	
Subject Sample	# 2 0.344	18:36	Volume Not Met
Breath Volume	0.519	18:36	
Air Blank	0.000	18:37	
Air Blank	0.000	18:39	
Subject Sample	# 3 0.355	18:39	
Breath Volume	2.480	18:39	
Air Blank	0.000	18:40	
Control Test	0.074	18:40	Control Outside Tolerance
Air Blank	0.000	18:41	

Data Download Date/Time: 10/27/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PABON,G,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BARKER,D

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 23:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 10/04/2023 23:58:11

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.166	23:53
Breath Volume	1.792	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.168	23:56
Breath Volume	1.910	23:56
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 10/27/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: DION,B

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:28

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 10/07/2023 01:19:05

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.089	01:14
Breath Volume	1.984	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:16
Subject Sample # 2	0.093	01:17
Breath Volume	2.539	01:17
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 10/27/2023 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LATIMER,J,A

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: LATIMER,J

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:21

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MCNEAL

DATE/TIME OF TEST: 10/08/2023 01:04:52

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:02	
Control Test	0.080	01:02	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:03	
Control Test	0.081	01:04	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 10/27/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: MACAVOY,JASON

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 10/21/2023 01:54:16

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.079	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.080	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 10/27/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BUSHONG,NICOLE

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MENDEZ LOPEZ

DATE/TIME OF TEST: 10/26/2023 02:21:59

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.101	02:17
Breath Volume	2.000	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.100	02:20
Breath Volume	1.355	02:20
Air Blank	0.000	02:21
Control Test	0.078	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 10/27/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TUFTS,KEVIN,S

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: APONTE,ALEX

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 15:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/26/2023 16:07:40

Diagnostics Check	OK	16:01
Air Blank	0.000	16:01
Control Test	0.080	16:01
Air Blank	0.000	16:02
Subject Sample # 1	0.289	16:03
Breath Volume	2.136	16:03
Air Blank	0.000	16:03
Air Blank	0.000	16:05
Subject Sample # 2	0.298	16:06
Breath Volume	2.386	16:06
Air Blank	0.000	16:06
Control Test	0.079	16:07
Air Blank	0.000	16:07
Diagnostics Check	OK	16:07

Data Download Date/Time: 10/27/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LATIMER,J,A

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: LATIMER,J

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/26/2023 23:22:18

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.079	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.161	23:17
Breath Volume	1.117	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.166	23:20
Breath Volume	1.531	23:20
Air Blank	0.000	23:21
Control Test	0.077	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 10/27/2023 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REASONER,DEBBIE,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: JACKSON,J
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 10/29/2023 08:01:40

Diagnostics Check	OK	07:54
Air Blank	0.000	07:54
Control Test	0.081	07:55
Air Blank	0.000	07:55
Subject Sample # 1	0.067	07:56
Breath Volume	2.585	07:56
Air Blank	0.000	07:57
Air Blank	0.000	07:59
Subject Sample # 2	0.063	08:00
Breath Volume	3.582	08:00
Air Blank	0.000	08:00
Control Test	0.081	08:01
Air Blank	0.000	08:01
Diagnostics Check	OK	08:01

Data Download Date/Time: 11/28/2023 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,R,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: PENA,R
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: RODRIGUEZ PIZANA

DATE/TIME OF TEST: 10/01/2023 01:59:31

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.191	01:54
Breath Volume	1.933	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.184	01:57
Breath Volume	2.558	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 10/18/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: HYLTON

DATE/TIME OF TEST: 10/03/2023 03:15:31

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.089	03:10
Breath Volume	3.117	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.085	03:14
Breath Volume	3.785	03:14
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 10/18/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: VASQUEZ IRAHETA

DATE/TIME OF TEST: 10/06/2023 23:58:50

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:55	
Control Test	0.080	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 1	23:57	Interferent Detect
Breath Volume	2.328	23:57	
Air Blank	0.000	23:57	
Control Test	0.079	23:58	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 10/18/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: VASQUEZ IRAHETA

DATE/TIME OF TEST: 10/07/2023 00:08:51

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.123	00:04
Breath Volume	2.472	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.118	00:07
Breath Volume	1.984	00:07
Air Blank	0.000	00:07
Control Test	0.081	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 10/18/2023 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: BAIJNATH

DATE/TIME OF TEST: 10/07/2023 21:01:13

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.080	20:55
Air Blank	0.000	20:55
Subject Sample # 1	0.248	20:56
Breath Volume	2.265	20:56
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.235	20:59
Breath Volume	2.550	20:59
Air Blank	0.000	21:00
Control Test	0.079	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 10/18/2023 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: GUZMAN ROCHA

DATE/TIME OF TEST: 10/07/2023 21:37:19

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.095	21:32
Breath Volume	1.339	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.097	21:35
Breath Volume	1.613	21:35
Air Blank	0.000	21:36
Control Test	0.080	21:36
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 10/18/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: MARINEZ

DATE/TIME OF TEST: 10/08/2023 02:42:54

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.100	02:38
Breath Volume	1.781	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample # 2	0.096	02:41
Breath Volume	1.238	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 10/18/2023 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,J,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: TORRES,V
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 10/08/2023 18:37:27

Diagnostics Check	OK	18:32	
Air Blank	0.000	18:32	
Control Test	0.080	18:32	
Air Blank	0.000	18:33	
Subject Sample	# 1	18:36	Subject Test Refused
Breath Volume	0.000	18:36	
Air Blank	0.000	18:36	
Control Test	0.080	18:36	
Air Blank	0.000	18:37	
Diagnostics Check	OK	18:37	

Data Download Date/Time: 10/18/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MAVES,JEFFREY
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MUNOZ CORTES

DATE/TIME OF TEST: 10/11/2023 01:21:28

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.166	01:17
Breath Volume	1.355	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.146	01:19
Breath Volume	1.257	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 10/18/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: EASTRIDGE,JAMES
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: JEAN LOUIS

DATE/TIME OF TEST: 10/11/2023 01:41:32

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.235	01:36
Breath Volume	2.296	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.231	01:40
Breath Volume	2.628	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 10/18/2023 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MAVES,JEFFREY
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: GURRERO MUNOZ

DATE/TIME OF TEST: 10/15/2023 05:15:29

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:03	
Control Test	0.080	05:04	
Air Blank	0.000	05:04	
Subject Sample	# 1	05:07	No Sample Provided
Breath Volume	0.000	05:07	
Air Blank	0.000	05:08	
Air Blank	0.000	05:10	
Subject Sample	# 2	0.201	05:11
Breath Volume	1.515	05:11	
Air Blank	0.000	05:11	
Air Blank	0.000	05:13	
Subject Sample	# 3	0.184	05:13
Breath Volume	2.117	05:13	
Air Blank	0.000	05:14	
Control Test	0.080	05:14	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 10/18/2023 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAFA,YUSUF,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ARAFA,YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 10/01/2023 22:02:20

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.221	21:57
Breath Volume	4.066	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.220	22:00
Breath Volume	3.792	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 10/06/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: DAVIS,LEANDRA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BOLES

DATE/TIME OF TEST: 10/12/2023 02:27:59

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.000	02:23
Breath Volume	2.628	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.000	02:26
Breath Volume	3.382	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATLEY,KENDALL,L
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: HATLEY,KENDALL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BOWKER

DATE/TIME OF TEST: 10/13/2023 02:26:50

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.081	02:16	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:19	No Sample Provided
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample	# 2	02:25	No Sample Provided
Breath Volume	0.000	02:25	
Air Blank	0.000	02:25	
Control Test	0.080	02:26	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: SHAPIRO

DATE/TIME OF TEST: 10/13/2023 04:26:23

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.180	04:18
Breath Volume	1.867	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.152	04:21
Breath Volume	1.527	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 3	0.156	04:24
Breath Volume	1.144	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: DAVIS,LANDRE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: FAIN

DATE/TIME OF TEST: 10/17/2023 22:41:03

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.081	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.332	22:33
Breath Volume	1.128	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.370	22:36
Breath Volume	2.660	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 3	0.359	22:39
Breath Volume	2.812	22:39
Air Blank	0.000	22:40
Control Test	0.081	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:41

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/22/2023 02:56:39

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.102	02:52
Breath Volume	3.148	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.098	02:55
Breath Volume	3.164	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELCHER,ANDREW,N
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MELCHER,ANDREW
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 12:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: WORRELL

DATE/TIME OF TEST: 10/24/2023 14:06:57

Diagnostics Check	OK	14:00
Air Blank	0.000	14:00
Control Test	0.080	14:01
Air Blank	0.000	14:01
Subject Sample # 1	0.221	14:02
Breath Volume	1.445	14:02
Air Blank	0.000	14:02
Air Blank	0.000	14:04
Subject Sample # 2	0.212	14:05
Breath Volume	1.234	14:05
Air Blank	0.000	14:06
Control Test	0.081	14:06
Air Blank	0.000	14:06
Diagnostics Check	OK	14:06

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/24/2023 18:50:33

Diagnostics Check	OK	18:43
Air Blank	0.000	18:43
Control Test	0.080	18:43
Air Blank	0.000	18:44
Subject Sample # 1	0.191	18:45
Breath Volume	1.242	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:47
Subject Sample # 2	0.199	18:48
Breath Volume	2.710	18:48
Air Blank	0.000	18:49
Control Test	0.081	18:49
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 10/25/2023 02:41:20

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.182	02:36
Breath Volume	1.437	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.191	02:39
Breath Volume	1.613	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MARSHALL,LEWIS
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: HADLEY

DATE/TIME OF TEST: 10/26/2023 23:02:06

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.081	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.109	22:57
Breath Volume	1.363	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.111	23:00
Breath Volume	1.488	23:00
Air Blank	0.000	23:01
Control Test	0.081	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATLEY,KENDALL,L
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: HATLEY,KENDALL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: HUTCHISON

DATE/TIME OF TEST: 10/26/2023 23:32:45

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.162	23:28
Breath Volume	2.402	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.161	23:31
Breath Volume	2.093	23:31
Air Blank	0.000	23:31
Control Test	0.080	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 10/27/2023 20:42:39

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:31	
Control Test	0.080	20:31	
Air Blank	0.000	20:32	
Subject Sample	# 1 0.119	20:35	Volume Not Met
Breath Volume	0.097	20:35	
Air Blank	0.000	20:36	
Air Blank	0.000	20:37	
Subject Sample	# 2 0.117	20:41	Volume Not Met
Breath Volume	0.101	20:41	
Air Blank	0.000	20:41	
Control Test	0.081	20:42	
Air Blank	0.000	20:42	
Diagnostics Check	OK	20:42	

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 10/27/2023 20:55:28

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:44	
Control Test	0.079	20:44	
Air Blank	0.000	20:45	
Subject Sample	# 1 0.126	20:48	Volume Not Met
Breath Volume	0.050	20:48	
Air Blank	0.000	20:48	
Air Blank	0.000	20:50	
Subject Sample	# 2 0.133	20:53	Volume Not Met
Breath Volume	0.152	20:53	
Air Blank	0.000	20:54	
Control Test	0.080	20:54	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 10/27/2023 21:05:05

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.132	21:00
Breath Volume	1.464	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:02
Subject Sample # 2	0.130	21:03
Breath Volume	1.167	21:03
Air Blank	0.000	21:04
Control Test	0.079	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:05

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ANDREWS,SHAMAAR
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 10/28/2023 01:40:25

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.078	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.099	01:35
Breath Volume	1.320	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.100	01:38
Breath Volume	1.527	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 11/06/2023 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: NOWALK

DATE/TIME OF TEST: 10/28/2023 18:36:48

Diagnostics Check	OK	18:29
Air Blank	0.000	18:30
Control Test	0.080	18:30
Air Blank	0.000	18:30
Subject Sample # 1	0.337	18:32
Breath Volume	2.730	18:32
Air Blank	0.000	18:32
Air Blank	0.000	18:34
Subject Sample # 2	0.343	18:35
Breath Volume	2.683	18:35
Air Blank	0.000	18:35
Control Test	0.080	18:36
Air Blank	0.000	18:36
Diagnostics Check	OK	18:36

Data Download Date/Time: 11/06/2023 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/02/2023 05:58:39

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.079	05:52
Air Blank	0.000	05:52
Subject Sample # 1	0.129	05:53
Breath Volume	1.429	05:53
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.140	05:57
Breath Volume	1.773	05:57
Air Blank	0.000	05:57
Control Test	0.078	05:58
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 10/15/2023 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CALASSO,ANTHONY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 10/03/2023 06:09:57

Diagnostics Check	OK	05:55	
Air Blank	0.000	05:55	
Control Test	0.078	05:56	
Air Blank	0.000	05:56	
Subject Sample	# 1 0.146	05:57	
Breath Volume	1.492	05:57	
Air Blank	0.000	05:57	
Air Blank	0.000	05:59	
Subject Sample	# 2 0.108	06:02	Volume Not Met
Breath Volume	0.546	06:02	
Air Blank	0.000	06:03	
Air Blank	0.000	06:05	
Subject Sample	# 3 0.117	06:08	Volume Not Met
Breath Volume	0.433	06:08	
Air Blank	0.000	06:09	
Control Test	0.078	06:09	
Air Blank	0.000	06:09	
Diagnostics Check	OK	06:09	

Data Download Date/Time: 10/15/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PAUGA,JOHN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: MARLER

DATE/TIME OF TEST: 10/05/2023 13:32:15

Diagnostics Check	OK	13:25
Air Blank	0.000	13:26
Control Test	0.080	13:26
Air Blank	0.000	13:27
Subject Sample # 1	0.169	13:27
Breath Volume	2.449	13:27
Air Blank	0.000	13:28
Air Blank	0.000	13:30
Subject Sample # 2	0.156	13:30
Breath Volume	1.582	13:30
Air Blank	0.000	13:31
Control Test	0.079	13:31
Air Blank	0.000	13:32
Diagnostics Check	OK	13:32

Data Download Date/Time: 10/15/2023 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINGRY,HOBIE,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MANSON,TIMOTHY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 12:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: SPILLERS

DATE/TIME OF TEST: 10/11/2023 13:27:19

Diagnostics Check	OK	13:18
Air Blank	0.000	13:18
Control Test	0.079	13:19
Air Blank	0.000	13:19
Subject Sample # 1	0.163	13:20
Breath Volume	1.390	13:20
Air Blank	0.000	13:20
Air Blank	0.000	13:22
Subject Sample # 2	0.175	13:25
Breath Volume	1.910	13:25
Air Blank	0.000	13:26
Control Test	0.079	13:26
Air Blank	0.000	13:27
Diagnostics Check	OK	13:27

Data Download Date/Time: 10/15/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MCQUAY,ZACHARY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/11/2023 23:57:29

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.232	23:52
Breath Volume	1.734	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.222	23:56
Breath Volume	1.796	23:56
Air Blank	0.000	23:56
Control Test	0.077	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 10/15/2023 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL, NEIL, F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCQUAY, ZACHARY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: LAFONTAINE

DATE/TIME OF TEST: 10/13/2023 01:41:48

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.153	01:36
Breath Volume	2.429	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.140	01:40
Breath Volume	2.699	01:40
Air Blank	0.000	01:40
Control Test	0.078	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 10/15/2023 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINGRY,HOBIE,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HOLMES,KYLE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/19/2023 01:39:10

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.113	01:34
Breath Volume	1.699	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.104	01:37
Breath Volume	1.464	01:37
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 11/14/2023 01:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: DE LA CRUZ,JUAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: ROJAS FALCON

DATE/TIME OF TEST: 10/20/2023 23:25:07

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.169	23:20
Breath Volume	2.281	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:22
Subject Sample # 2	0.166	23:23
Breath Volume	1.511	23:23
Air Blank	0.000	23:24
Control Test	0.077	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 11/14/2023 01:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/21/2023 02:02:17

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.079	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:01	Sequence Aborted
Breath Volume	0.000	02:01	
Air Blank	0.000	02:02	

Data Download Date/Time: 11/14/2023 01:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/15/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/21/2023 02:15:25

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.079	02:11	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:13	Subject Test Refused
Breath Volume	0.882	02:13	
Air Blank	0.000	02:14	
Control Test	0.079	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 11/14/2023 01:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/21/2023 02:30:58

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.078	02:23	
Air Blank	0.000	02:24	
Subject Sample # 1	0.103	02:25	
Breath Volume	1.292	02:25	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample # 2	0.106	02:29	
Breath Volume	1.816	02:29	
Air Blank	0.000	02:30	
		02:30	RFI Detect
Air Blank	0.000	02:30	

Data Download Date/Time: 11/14/2023 01:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/21/2023 02:44:38

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.097	02:39
Breath Volume	1.628	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.095	02:43
Breath Volume	1.101	02:43
Air Blank	0.000	02:43
Control Test	0.078	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 11/14/2023 01:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,CALE,V
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: LUCUS,KOBI
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 07:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:42

SUBJECT'S LAST NAME: MENDEZ PEREZ

DATE/TIME OF TEST: 10/22/2023 08:15:43

Diagnostics Check	OK	08:02	
Air Blank	0.000	08:02	
Control Test	0.078	08:02	
Air Blank	0.000	08:03	
Subject Sample	# 1 0.156	08:06	Volume Not Met
Breath Volume	1.011	08:06	
Air Blank	0.000	08:07	
Air Blank	0.000	08:08	
Subject Sample	# 2 0.185	08:10	
Breath Volume	2.109	08:10	
Air Blank	0.000	08:11	
Air Blank	0.000	08:13	
Subject Sample	# 3 0.192	08:14	
Breath Volume	1.835	08:14	
Air Blank	0.000	08:14	
Control Test	0.075	08:15	
Air Blank	0.000	08:15	
Diagnostics Check	OK	08:15	

Data Download Date/Time: 11/14/2023 01:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HARRISON,MARAH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:08

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/26/2023 15:35:07

Diagnostics Check	OK	15:28
Air Blank	0.000	15:29
Control Test	0.079	15:29
Air Blank	0.000	15:30
Subject Sample # 1	0.270	15:30
Breath Volume	1.468	15:30
Air Blank	0.000	15:31
Air Blank	0.000	15:33
Subject Sample # 2	0.251	15:33
Breath Volume	1.109	15:33
Air Blank	0.000	15:34
Control Test	0.078	15:34
Air Blank	0.000	15:35
Diagnostics Check	OK	15:35

Data Download Date/Time: 11/14/2023 01:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: DE LA CRUZ,JUAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: HERNANDEZ BELTRAN

DATE/TIME OF TEST: 10/27/2023 23:46:14

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.246	23:41
Breath Volume	2.765	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.239	23:44
Breath Volume	1.921	23:44
Air Blank	0.000	23:45
Control Test	0.079	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 11/14/2023 01:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SASSER,PATRICK
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/29/2023 05:53:13

Diagnostics Check	OK	05:46
Air Blank	0.000	05:47
Control Test	0.079	05:47
Air Blank	0.000	05:47
Subject Sample # 1	0.098	05:48
Breath Volume	1.613	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.102	05:51
Breath Volume	2.261	05:51
Air Blank	0.000	05:52
Control Test	0.079	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 11/14/2023 01:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURLEY,JORDAN,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GREENUP,R
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: AGUILAR MARTINEZ

DATE/TIME OF TEST: 10/18/2023 04:06:19

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.079	03:58	
Air Blank	0.000	03:59	
Subject Sample # 1	0.194	04:01	
Breath Volume	2.214	04:01	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample # 2		04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.079	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 10/18/2023 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KAYLEE,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: OCAMPO MENENDEZ

DATE/TIME OF TEST: 10/19/2023 01:09:00

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.080	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1 0.103	01:01	Volume Not Met
Breath Volume	0.363	01:01	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample	# 2 0.114	01:07	Volume Not Met
Breath Volume	0.484	01:07	
Air Blank	0.000	01:08	
Control Test	0.079	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 11/21/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KAYLEE,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: OCAMPO MENENDEZ

DATE/TIME OF TEST: 10/19/2023 01:13:49

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:10	
Control Test	0.080	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:12	RFI Detect
Breath Volume	0.246	01:12	
Air Blank	0.000	01:12	
Control Test	0.079	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 11/21/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KAYLEE,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: OCAMPO MENENDEZ

DATE/TIME OF TEST: 10/19/2023 01:28:15

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:17	
Control Test	0.079	01:17	
Air Blank	0.000	01:17	
Subject Sample	# 1 0.107	01:21	Volume Not Met
Breath Volume	0.523	01:21	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 2 0.106	01:26	Volume Not Met
Breath Volume	0.597	01:26	
Air Blank	0.000	01:27	
Control Test	0.079	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 11/21/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JACKSON,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: JUAN PASCUAL

DATE/TIME OF TEST: 10/20/2023 01:21:37

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.079	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.332	01:17
Breath Volume	2.355	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.313	01:20
Breath Volume	1.777	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 11/21/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 10/20/2023 01:58:57

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.122	01:54
Breath Volume	2.464	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.129	01:57
Breath Volume	2.457	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 11/21/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: GAYA

DATE/TIME OF TEST: 10/21/2023 02:24:06

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:13	
Control Test	0.079	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1 0.110	02:16	Volume Not Met
Breath Volume	0.082	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample	# 2 0.119	02:22	Volume Not Met
Breath Volume	0.136	02:22	
Air Blank	0.000	02:23	
Control Test	0.078	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 11/21/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN,TIM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: BONAGURA

DATE/TIME OF TEST: 10/21/2023 20:28:11

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.079	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.000	20:23
Breath Volume	2.921	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.000	20:26
Breath Volume	2.332	20:26
Air Blank	0.000	20:27
Control Test	0.080	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 11/21/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: SALVADOR MORALES

DATE/TIME OF TEST: 10/22/2023 06:06:11

Diagnostics Check	OK	06:03	
Air Blank	0.000	06:03	
Control Test	0.080	06:04	
Air Blank	0.000	06:04	
Subject Sample	# 1	06:04	Subject Test Refused
Breath Volume	0.000	06:04	
Air Blank	0.000	06:05	
Control Test	0.080	06:05	
Air Blank	0.000	06:06	
Diagnostics Check	OK	06:06	

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KNAPP,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: HAM

DATE/TIME OF TEST: 10/23/2023 03:23:29

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.139	03:16
Breath Volume	2.050	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.144	03:21
Breath Volume	1.492	03:21
Air Blank	0.000	03:22
Control Test	0.078	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 10/24/2023 03:30:08

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.170	03:25
Breath Volume	1.593	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.174	03:28
Breath Volume	1.132	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ORCUTT,A
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: TERRELL

DATE/TIME OF TEST: 10/25/2023 03:43:44

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.177	03:39
Breath Volume	2.199	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.178	03:42
Breath Volume	2.335	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RUSS,ERIC
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: BLASI

DATE/TIME OF TEST: 10/25/2023 19:42:33

Diagnostics Check	OK	19:34
Air Blank	0.000	19:35
Control Test	0.080	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.016	19:37
Breath Volume	2.058	19:37
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.015	19:41
Breath Volume	2.332	19:41
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CIPOLLONE

DATE/TIME OF TEST: 10/26/2023 03:09:00

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.152	03:04
Breath Volume	1.273	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.157	03:07
Breath Volume	1.574	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:09

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRUM,JACOB,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DENNIS,JONATHAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: HILDRETH

DATE/TIME OF TEST: 10/26/2023 05:28:03

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.116	05:23
Breath Volume	1.121	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.109	05:26
Breath Volume	1.468	05:26
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:28

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RUSS,ERIC
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: HILTON

DATE/TIME OF TEST: 10/27/2023 00:15:30

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.151	00:10
Breath Volume	3.308	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.153	00:14
Breath Volume	2.312	00:14
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: LUNAN

DATE/TIME OF TEST: 10/27/2023 00:41:26

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.138	00:36
Breath Volume	1.554	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.136	00:39
Breath Volume	1.699	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CLUTZ

DATE/TIME OF TEST: 10/27/2023 03:08:05

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.295	03:00
Breath Volume	1.675	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.321	03:03
Breath Volume	2.300	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 3	0.335	03:06
Breath Volume	2.027	03:06
Air Blank	0.000	03:07
Control Test	0.076	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ARRIETA FERNANDEZ

DATE/TIME OF TEST: 10/27/2023 03:32:27

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.079	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1 0.155	03:25	Volume Not Met
Breath Volume	0.558	03:25	
Air Blank	0.000	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 2 0.161	03:30	Volume Not Met
Breath Volume	0.078	03:30	
Air Blank	0.000	03:31	
Control Test	0.072	03:31	Control Outside Tolerance
Air Blank	0.000	03:32	

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JACKSON,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: RAMBARRAN

DATE/TIME OF TEST: 10/27/2023 23:56:35

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.079	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 1 0.328	23:52	
Breath Volume	2.441	23:52	
Air Blank	0.000	23:52	
Air Blank	0.000	23:54	
Subject Sample	# 2	23:55	Slope Not Met
Breath Volume	1.292	23:55	
Air Blank	0.000	23:55	
Control Test	0.078	23:56	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JACKSON,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: RAMBARRAN

DATE/TIME OF TEST: 10/28/2023 00:04:14

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:59
Subject Sample # 2	0.318	23:59
Breath Volume	2.011	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.307	00:02
Breath Volume	2.117	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING, JONATHAN, L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: DEMPSEY, ALEXIS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: DUMENIGO HERNANDEZ

DATE/TIME OF TEST: 10/28/2023 03:16:40

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.178	03:12
Breath Volume	2.460	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.173	03:15
Breath Volume	2.207	03:15
Air Blank	0.000	03:15
Control Test	0.078	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ORCUTT,A
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: VALDEZ MORENO

DATE/TIME OF TEST: 10/28/2023 21:29:36

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.080	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.082	21:25
Breath Volume	2.812	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.083	21:28
Breath Volume	3.195	21:28
Air Blank	0.000	21:28
Control Test	0.080	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,CHEYANNE,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WOOD,REBECCA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: RUGGIERO

DATE/TIME OF TEST: 10/29/2023 00:04:01

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.176	23:59
Breath Volume	1.246	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.171	00:02
Breath Volume	2.082	00:02
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:04

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,CHEYANNE,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,BRITTANY
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BOSCH

DATE/TIME OF TEST: 10/29/2023 00:35:22

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:24	
Control Test	0.080	00:25	
Air Blank	0.000	00:25	
Subject Sample	# 1 0.217	00:28	Volume Not Met
Breath Volume	0.328	00:28	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 2	00:33	Subject Test Refused
Breath Volume	0.105	00:33	
Air Blank	0.000	00:34	
Control Test	0.078	00:34	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: EWALD

DATE/TIME OF TEST: 10/29/2023 01:07:13

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.078	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.164	01:02
Breath Volume	1.335	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.174	01:05
Breath Volume	1.468	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,FRANCISCO,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MARTINEZ,FRANCISCO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: DESOUZA

DATE/TIME OF TEST: 10/29/2023 04:52:54

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.080	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.147	04:48
Breath Volume	3.414	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.152	04:51
Breath Volume	3.429	04:51
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASH,CONNOR,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HASH,CONNOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: BURMAA

DATE/TIME OF TEST: 10/29/2023 06:12:56

Diagnostics Check	OK	06:05
Air Blank	0.000	06:06
Control Test	0.079	06:06
Air Blank	0.000	06:06
Subject Sample # 1	0.089	06:08
Breath Volume	2.269	06:08
Air Blank	0.000	06:09
Air Blank	0.000	06:11
Subject Sample # 2	0.102	06:11
Breath Volume	1.984	06:11
Air Blank	0.000	06:12
Control Test	0.078	06:12
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KUPKA,RIKKI
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: LOPEZ TELLEZ

DATE/TIME OF TEST: 10/29/2023 14:42:02

Diagnostics Check	OK	14:36
Air Blank	0.000	14:36
Control Test	0.080	14:36
Air Blank	0.000	14:37
Subject Sample # 1	0.112	14:37
Breath Volume	3.164	14:37
Air Blank	0.000	14:38
Air Blank	0.000	14:40
Subject Sample # 2	0.108	14:40
Breath Volume	2.617	14:40
Air Blank	0.000	14:41
Control Test	0.079	14:41
Air Blank	0.000	14:41
Diagnostics Check	OK	14:42

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: BRIONESARAUZ

DATE/TIME OF TEST: 10/29/2023 18:05:24

Diagnostics Check	OK	17:59
Air Blank	0.000	17:59
Control Test	0.079	18:00
Air Blank	0.000	18:00
Subject Sample # 1	0.179	18:01
Breath Volume	2.062	18:01
Air Blank	0.000	18:01
Air Blank	0.000	18:03
Subject Sample # 2	0.167	18:03
Breath Volume	1.332	18:03
Air Blank	0.000	18:04
Control Test	0.079	18:04
Air Blank	0.000	18:05
Diagnostics Check	OK	18:05

Data Download Date/Time: 11/21/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING, JONATHAN, L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: HERNANDEZ, YOLHEN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 10/29/2023 23:50:12

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.146	23:45
Breath Volume	2.066	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.142	23:48
Breath Volume	1.828	23:48
Air Blank	0.000	23:49
Control Test	0.078	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 11/21/2023 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MATERA,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ROYKO

DATE/TIME OF TEST: 10/30/2023 01:58:52

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.256	01:51
Breath Volume	3.210	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.232	01:54
Breath Volume	1.933	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 3	0.232	01:57
Breath Volume	2.027	01:57
Air Blank	0.000	01:57
Control Test	0.078	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 11/21/2023 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: ARANDA

DATE/TIME OF TEST: 10/30/2023 20:35:42

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.080	20:30
Air Blank	0.000	20:30
Subject Sample # 1	0.267	20:31
Breath Volume	1.992	20:31
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.266	20:34
Breath Volume	1.941	20:34
Air Blank	0.000	20:34
Control Test	0.078	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 11/21/2023 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: LOREDO

DATE/TIME OF TEST: 10/30/2023 22:28:41

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.079	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.142	22:23
Breath Volume	1.101	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.140	22:27
Breath Volume	1.210	22:27
Air Blank	0.000	22:27
Control Test	0.078	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 11/21/2023 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FERRARO,PAUL
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/31/2023 02:37:27

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.197	02:32
Breath Volume	1.207	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.212	02:35
Breath Volume	1.417	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 11/21/2023 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,EDWARD,E
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: FISHER,C
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: SHIRK

DATE/TIME OF TEST: 10/15/2023 21:36:05

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:30
Subject Sample # 1	0.084	21:31
Breath Volume	1.394	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:33
Subject Sample # 2	0.084	21:34
Breath Volume	1.863	21:34
Air Blank	0.000	21:35
Control Test	0.079	21:35
Air Blank	0.000	21:35
Diagnostics Check	OK	21:36

Data Download Date/Time: 10/24/2023 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,EDWARD,E
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: FISHER,C
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: LINGARD

DATE/TIME OF TEST: 10/16/2023 20:05:27

Diagnostics Check	OK	19:58
Air Blank	0.000	19:59
Control Test	0.079	19:59
Air Blank	0.000	20:00
Subject Sample # 1	0.281	20:00
Breath Volume	2.253	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.290	20:03
Breath Volume	1.621	20:03
Air Blank	0.000	20:04
Control Test	0.079	20:04
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 10/24/2023 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/27/2023 21:33:42

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:29	
Control Test	0.080	21:29	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:32	Subject Test Refused
Breath Volume	0.000	21:32	
Air Blank	0.000	21:32	
Control Test	0.080	21:33	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 11/30/2023 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STREIGOLD,DAVID,M
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: STREIGOLD,DAVID
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: MPAFE

DATE/TIME OF TEST: 10/29/2023 12:38:43

Diagnostics Check	OK	12:32
Air Blank	0.000	12:33
Control Test	0.079	12:33
Air Blank	0.000	12:33
Subject Sample # 1	0.000	12:34
Breath Volume	2.597	12:34
Air Blank	0.000	12:34
Air Blank	0.000	12:36
Subject Sample # 2	0.000	12:37
Breath Volume	2.531	12:37
Air Blank	0.000	12:37
Control Test	0.079	12:38
Air Blank	0.000	12:38
Diagnostics Check	OK	12:38

Data Download Date/Time: 11/30/2023 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: COE,M
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: MILSAP

DATE/TIME OF TEST: 10/28/2023 01:05:57

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.079	01:03	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Control Test	0.078	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 11/28/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUBICEK,JORDAN,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: FEAGANS,DUSTIN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MIKELL

DATE/TIME OF TEST: 10/29/2023 03:58:35

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.079	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1 0.170	03:52	Volume Not Met
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample	# 2	03:57	Subject Test Refused
Breath Volume	0.101	03:57	
Air Blank	0.000	03:57	
Control Test	0.075	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 11/28/2023 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: VANIS,HEATHER
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: MERCHANT

DATE/TIME OF TEST: 10/10/2023 22:43:44

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.081	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.153	22:39
Breath Volume	2.246	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:41
Subject Sample # 2	0.150	22:42
Breath Volume	2.011	22:42
Air Blank	0.000	22:42
Control Test	0.076	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 11/09/2023 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWEN,RYAN,D
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: VANIS,HEATHER
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 10/27/2023 23:09:07

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.117	23:03
Breath Volume	1.324	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.111	23:07
Breath Volume	1.382	23:07
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 11/09/2023 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,B
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: BROWN,ZECHARIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: RAWLINS

DATE/TIME OF TEST: 10/01/2023 19:27:25

Diagnostics Check	OK	19:21	
Air Blank	0.000	19:21	
Control Test	0.075	19:21	
Air Blank	0.000	19:22	
Subject Sample # 1	0.244	19:22	
Breath Volume	2.082	19:22	
Air Blank	0.000	19:23	
Air Blank	0.000	19:25	
Subject Sample # 2	0.234	19:25	
Breath Volume	2.011	19:25	
Air Blank	0.000	19:26	
Control Test	0.074	19:26	Control Outside Tolerance
Air Blank	0.000	19:27	

Data Download Date/Time: 10/23/2023 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGRAM,TRISTAN,N
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SWEENEY,P
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 10/08/2023 02:37:03

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.076	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.229	02:32
Breath Volume	3.660	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.234	02:35
Breath Volume	3.820	02:35
Air Blank	0.000	02:36
Control Test	0.076	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:37

Data Download Date/Time: 10/23/2023 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,B
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SWEENEY,PARKER
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/15/2023 01:39:57

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.076	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.261	01:32
Breath Volume	1.703	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.234	01:35
Breath Volume	1.261	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:37
Subject Sample # 3	0.245	01:38
Breath Volume	1.546	01:38
Air Blank	0.000	01:39
Control Test	0.075	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 10/23/2023 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGRAM,T,N
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: GUTHRIE,A
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: LUEDKE

DATE/TIME OF TEST: 10/19/2023 22:42:26

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.076	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.165	22:37
Breath Volume	2.621	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.170	22:40
Breath Volume	3.964	22:40
Air Blank	0.000	22:41
Control Test	0.076	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 10/23/2023 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIPPEON,DANIEL,P
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: RIPPEON,DANIEL
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 07:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:38

SUBJECT'S LAST NAME: DIEHL

DATE/TIME OF TEST: 10/21/2023 08:05:31

Diagnostics Check	OK	07:59	
Air Blank	0.000	07:59	
Control Test	0.076	07:59	
Air Blank	0.000	08:00	
Subject Sample # 1	0.145	08:00	
Breath Volume	3.269	08:00	
Air Blank	0.000	08:01	
Air Blank	0.000	08:03	
Subject Sample # 2	0.144	08:03	
Breath Volume	3.425	08:03	
Air Blank	0.000	08:04	
Control Test	0.074	08:04	Control Outside Tolerance
Air Blank	0.000	08:05	

Data Download Date/Time: 10/23/2023 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,B
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: GUTHRIE,ADAM
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: LUBERTOWICZ

DATE/TIME OF TEST: 10/29/2023 02:34:09

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.207	02:29
Breath Volume	3.484	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.199	02:32
Breath Volume	4.132	02:32
Air Blank	0.000	02:33
Control Test	0.076	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 11/21/2023 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,RICHARD,S
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: GAFFNEY,TRYSTYN
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/01/2023 05:21:58

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:19	
Control Test	0.080	05:19	
Air Blank	0.000	05:19	
Subject Sample	# 1	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:21	
Control Test	0.080	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 10/30/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,RICHARD,S
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: GAFFNEY,TRYSTYN
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/01/2023 05:32:09

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.080	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.132	05:27
Breath Volume	1.703	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.146	05:30
Breath Volume	2.277	05:30
Air Blank	0.000	05:31
Control Test	0.080	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 10/30/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFFNEY,CASEY,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: PHILIPS,JAMES
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: ADAMSON

DATE/TIME OF TEST: 10/08/2023 23:57:05

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.081	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.124	23:52
Breath Volume	3.027	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.118	23:55
Breath Volume	1.882	23:55
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 10/25/2023 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELODY, FRANCIS, D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: WELLS, KYLE
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/13/2023 20:56:26

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.081	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.253	20:51
Breath Volume	2.109	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.263	20:54
Breath Volume	1.410	20:54
Air Blank	0.000	20:55
Control Test	0.079	20:55
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 10/25/2023 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELODY,FRANCIS,D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: WELLS,KYLE
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 10/16/2023 01:08:46

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.082	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.181	01:03
Breath Volume	1.542	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.189	01:07
Breath Volume	1.937	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 10/25/2023 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFFNEY,CASEY,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: WELLS,KYLE
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 10/22/2023 23:23:07

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.212	23:18
Breath Volume	1.863	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.215	23:21
Breath Volume	1.871	23:21
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 10/25/2023 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERLAIN,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CHAMBERLAIN,JOSHUA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/14/2023 02:46:47

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.082	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.203	02:42
Breath Volume	2.156	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.200	02:45
Breath Volume	1.976	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 10/17/2023 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERLAIN,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DOMINICIS,L
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 10/29/2023 02:35:20

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.082	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.315	02:30
Breath Volume	2.847	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:32
Subject Sample # 2	0.315	02:33
Breath Volume	3.367	02:33
Air Blank	0.000	02:34
Control Test	0.082	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 11/30/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCGRAW,LOGAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: ELSTON

DATE/TIME OF TEST: 10/10/2023 19:02:14

Diagnostics Check	OK	18:52
Air Blank	0.000	18:53
Control Test	0.077	18:53
Air Blank	0.000	18:53
Subject Sample # 1	0.259	18:54
Breath Volume	2.003	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.293	18:57
Breath Volume	2.046	18:57
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 3	0.281	19:00
Breath Volume	1.921	19:00
Air Blank	0.000	19:01
Control Test	0.077	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 11/08/2023 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHOY,ANNELEYSIS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: CHOICE MUNOZ

DATE/TIME OF TEST: 10/11/2023 02:03:27

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.078	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.000	01:59
Breath Volume	3.937	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.000	02:02
Breath Volume	4.199	02:02
Air Blank	0.000	02:02
Control Test	0.077	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: YIM,RANDY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/11/2023 03:40:17

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.077	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.227	03:35
Breath Volume	2.105	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.227	03:38
Breath Volume	4.449	03:38
Air Blank	0.000	03:39
Control Test	0.076	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MIRANDA MAURICIO

DATE/TIME OF TEST: 10/12/2023 00:30:03

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.077	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.206	00:25
Breath Volume	2.113	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.215	00:28
Breath Volume	2.453	00:28
Air Blank	0.000	00:29
Control Test	0.077	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:30

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIESENFELT, JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: MELENDEZ MARTINEZ

DATE/TIME OF TEST: 10/12/2023 03:55:20

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.077	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.119	03:51
Breath Volume	1.507	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.119	03:53
Breath Volume	2.636	03:53
Air Blank	0.000	03:54
Control Test	0.077	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHEWS,ZACHARY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: CHERELUS

DATE/TIME OF TEST: 10/12/2023 22:38:47

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.077	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.000	22:34
Breath Volume	2.628	22:34
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.000	22:37
Breath Volume	2.609	22:37
Air Blank	0.000	22:37
Control Test	0.077	22:38
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: TELLEZ JIMENEZ

DATE/TIME OF TEST: 10/14/2023 00:25:30

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.077	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.157	00:21
Breath Volume	2.203	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.160	00:24
Breath Volume	2.257	00:24
Air Blank	0.000	00:24
Control Test	0.076	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUTFL,RAED
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: RUTLEDGE

DATE/TIME OF TEST: 10/14/2023 04:58:46

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.076	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.206	04:47
Breath Volume	1.160	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:49
Subject Sample # 2	0.182	04:52
Breath Volume	1.312	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 3	0.193	04:57
Breath Volume	1.820	04:57
Air Blank	0.000	04:57
Control Test	0.075	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CORDOBA TORRES

DATE/TIME OF TEST: 10/15/2023 04:20:40

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.077	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.108	04:16
Breath Volume	3.191	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.106	04:19
Breath Volume	3.175	04:19
Air Blank	0.000	04:19
Control Test	0.076	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REYES,RICARDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: RODRIGUEZSOLARES

DATE/TIME OF TEST: 10/16/2023 06:32:08

Diagnostics Check	OK	06:25
Air Blank	0.000	06:26
Control Test	0.077	06:26
Air Blank	0.000	06:26
Subject Sample # 1	0.160	06:27
Breath Volume	1.789	06:27
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample # 2	0.161	06:30
Breath Volume	1.781	06:30
Air Blank	0.000	06:31
Control Test	0.076	06:31
Air Blank	0.000	06:32
Diagnostics Check	OK	06:32

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DEVOSS,EZEKIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:44

SUBJECT'S LAST NAME: GOMEZPEREZ

DATE/TIME OF TEST: 10/16/2023 10:14:49

Diagnostics Check	OK	10:08
Air Blank	0.000	10:08
Control Test	0.077	10:09
Air Blank	0.000	10:09
Subject Sample # 1	0.147	10:10
Breath Volume	1.570	10:10
Air Blank	0.000	10:11
Air Blank	0.000	10:12
Subject Sample # 2	0.150	10:13
Breath Volume	1.390	10:13
Air Blank	0.000	10:13
Control Test	0.076	10:14
Air Blank	0.000	10:14
Diagnostics Check	OK	10:14

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: LEVINE

DATE/TIME OF TEST: 10/17/2023 22:46:54

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.077	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.118	22:42
Breath Volume	1.214	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:44
Subject Sample # 2	0.121	22:45
Breath Volume	1.730	22:45
Air Blank	0.000	22:46
Control Test	0.077	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KILLON,JONATHAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/18/2023 02:36:32

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.076	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Control Test	0.077	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAYDEN, BENJAMIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MAYBIN HILL

DATE/TIME OF TEST: 10/18/2023 03:01:39

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.076	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.213	02:57
Breath Volume	3.007	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.224	03:00
Breath Volume	3.425	03:00
Air Blank	0.000	03:00
Control Test	0.076	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHEWS,ZACHARY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: BENABE CARRASQUILLO

DATE/TIME OF TEST: 10/18/2023 06:22:20

Diagnostics Check	OK	06:16
Air Blank	0.000	06:16
Control Test	0.077	06:16
Air Blank	0.000	06:17
Subject Sample # 1	0.000	06:17
Breath Volume	2.367	06:17
Air Blank	0.000	06:18
Air Blank	0.000	06:20
Subject Sample # 2	0.000	06:20
Breath Volume	3.261	06:20
Air Blank	0.000	06:21
Control Test	0.077	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHOY,ANNELEYSIS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: WYSOCKI

DATE/TIME OF TEST: 10/19/2023 02:45:59

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:35	
Control Test	0.077	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:39	No Sample Provided
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Air Blank	0.000	02:41	
Subject Sample	# 2	02:44	No Sample Provided
Breath Volume	0.000	02:44	
Air Blank	0.000	02:45	
Control Test	0.077	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCCARTHY,MICHAEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: NURAKHMETOVA

DATE/TIME OF TEST: 10/19/2023 03:54:18

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.077	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:49	No Sample Provided
Breath Volume	0.000	03:49	
Air Blank	0.000	03:50	
Air Blank	0.000	03:52	
Subject Sample	# 2	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Control Test	0.077	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE, BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 10/19/2023 04:21:44

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.076	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.183	04:17
Breath Volume	3.718	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.185	04:20
Breath Volume	3.335	04:20
Air Blank	0.000	04:20
Control Test	0.077	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 10/19/2023 23:49:28

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.078	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.033	23:45
Breath Volume	1.820	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.033	23:48
Breath Volume	1.773	23:48
Air Blank	0.000	23:48
Control Test	0.077	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: QUEZADA VARGAS, JEAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/20/2023 01:06:52

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.077	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.133	01:02
Breath Volume	1.222	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.139	01:05
Breath Volume	1.257	01:05
Air Blank	0.000	01:05
Control Test	0.077	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GRISALES,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MILLA REYES

DATE/TIME OF TEST: 10/21/2023 01:15:50

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.076	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.119	01:11
Breath Volume	1.312	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.126	01:14
Breath Volume	1.992	01:14
Air Blank	0.000	01:14
Control Test	0.077	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PREDELUS,SMITH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MOEGER

DATE/TIME OF TEST: 10/21/2023 01:38:59

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.077	01:24	
Air Blank	0.000	01:25	
Subject Sample	# 1 0.140	01:28	
Breath Volume	2.437	01:28	
Air Blank	0.000	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 2 0.145	01:34	Volume Not Met
Breath Volume	0.335	01:34	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample	# 3 0.137	01:37	
Breath Volume	1.417	01:37	
Air Blank	0.000	01:38	
Control Test	0.074	01:38	Control Outside Tolerance
Air Blank	0.000	01:38	

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PREDELUS,SMITH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MOEGER

DATE/TIME OF TEST: 10/21/2023 01:47:51

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.076	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.141	01:42
Breath Volume	2.023	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.140	01:46
Breath Volume	1.464	01:46
Air Blank	0.000	01:46
Control Test	0.076	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIAZ, BRITZZY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: LICONA FERRERA

DATE/TIME OF TEST: 10/21/2023 03:59:25

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.076	03:48	
Air Blank	0.000	03:48	
Subject Sample	# 1 0.209	03:49	
Breath Volume	2.312	03:49	
Air Blank	0.000	03:49	
Air Blank	0.000	03:51	
Subject Sample	# 2 0.206	03:55	Volume Not Met
Breath Volume	0.054	03:55	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample	# 3 0.201	03:57	
Breath Volume	2.148	03:57	
Air Blank	0.000	03:58	
Control Test	0.075	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA, SAMUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: BLACKWELL

DATE/TIME OF TEST: 10/21/2023 10:01:45

Diagnostics Check	OK	09:55
Air Blank	0.000	09:56
Control Test	0.077	09:56
Air Blank	0.000	09:56
Subject Sample # 1	0.086	09:57
Breath Volume	3.226	09:57
Air Blank	0.000	09:57
Air Blank	0.000	09:59
Subject Sample # 2	0.090	10:00
Breath Volume	3.453	10:00
Air Blank	0.000	10:00
Control Test	0.075	10:01
Air Blank	0.000	10:01
Diagnostics Check	OK	10:01

Data Download Date/Time: 11/08/2023 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ORTIZ,PAUL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: HERNANDEZYHERNANDEZ

DATE/TIME OF TEST: 10/21/2023 16:16:47

Diagnostics Check	OK	16:10
Air Blank	0.000	16:11
Control Test	0.077	16:11
Air Blank	0.000	16:11
Subject Sample # 1	0.264	16:12
Breath Volume	2.960	16:12
Air Blank	0.000	16:12
Air Blank	0.000	16:14
Subject Sample # 2	0.276	16:15
Breath Volume	3.343	16:15
Air Blank	0.000	16:15
Control Test	0.076	16:16
Air Blank	0.000	16:16
Diagnostics Check	OK	16:16

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ESPINOSA, ANGEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 10/22/2023 02:42:50

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.076	02:29	
Air Blank	0.000	02:29	
Subject Sample	# 1 0.207	02:30	
Breath Volume	1.683	02:30	
Air Blank	0.000	02:30	
Air Blank	0.000	02:32	
Subject Sample	# 2	02:35	No Sample Provided
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Air Blank	0.000	02:38	
Subject Sample	# 3	02:41	No Sample Provided
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.076	02:42	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRELL,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 10/22/2023 03:02:28

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.077	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.212	02:57
Breath Volume	2.574	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.202	03:01
Breath Volume	2.957	03:01
Air Blank	0.000	03:01
Control Test	0.076	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATIAS, ROLANDO
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ACTE XOL

DATE/TIME OF TEST: 10/22/2023 04:02:49

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.077	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.135	03:58
Breath Volume	1.722	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.137	04:01
Breath Volume	1.566	04:01
Air Blank	0.000	04:01
Control Test	0.077	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RILEY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: SODEN

DATE/TIME OF TEST: 10/24/2023 03:06:03

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.077	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.175	03:01
Breath Volume	1.683	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.191	03:04
Breath Volume	1.800	03:04
Air Blank	0.000	03:05
Control Test	0.075	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:06

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIEBKNECHT,KEVIN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: CHARLAN

DATE/TIME OF TEST: 10/24/2023 18:13:15

Diagnostics Check	OK	18:01	
Air Blank	0.000	18:02	
Control Test	0.077	18:02	
Air Blank	0.000	18:03	
Subject Sample	# 1	18:06	No Sample Provided
Breath Volume	0.000	18:06	
Air Blank	0.000	18:06	
Air Blank	0.000	18:08	
Subject Sample	# 2	18:11	No Sample Provided
Breath Volume	0.000	18:11	
Air Blank	0.000	18:12	
Control Test	0.078	18:12	
Air Blank	0.000	18:13	
Diagnostics Check	OK	18:13	

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: FERNANDES

DATE/TIME OF TEST: 10/25/2023 02:18:17

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.077	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.093	02:13
Breath Volume	1.132	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.100	02:16
Breath Volume	2.355	02:16
Air Blank	0.000	02:17
Control Test	0.077	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ORTIZ,PAUL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: CINTO PEREZ

DATE/TIME OF TEST: 10/25/2023 20:54:12

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.077	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.192	20:49
Breath Volume	2.195	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.176	20:52
Breath Volume	1.464	20:52
Air Blank	0.000	20:53
Control Test	0.076	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARUNA, HEATHER
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 10/25/2023 21:38:14

Diagnostics Check	OK	21:23	
Air Blank	0.000	21:24	
Control Test	0.077	21:24	
Air Blank	0.000	21:25	
Subject Sample	# 1 0.317	21:25	
Breath Volume	1.210	21:25	
Air Blank	0.000	21:26	
Air Blank	0.000	21:27	
Subject Sample	# 2 0.280	21:31	Volume Not Met
Breath Volume	0.742	21:31	
Air Blank	0.000	21:31	
Air Blank	0.000	21:33	
Subject Sample	# 3 0.244	21:36	Volume Not Met
Breath Volume	0.656	21:36	
Air Blank	0.000	21:37	
Control Test	0.075	21:37	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 10/26/2023 03:49:19

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.076	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:45	No Sample Provided
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Air Blank	0.000	03:47	
Subject Sample	# 2	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.077	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCPHEE,WILLIS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:46

SUBJECT'S LAST NAME: GOTAY

DATE/TIME OF TEST: 10/26/2023 08:18:22

Diagnostics Check	OK	08:11
Air Blank	0.000	08:11
Control Test	0.077	08:12
Air Blank	0.000	08:12
Subject Sample # 1	0.133	08:13
Breath Volume	1.824	08:13
Air Blank	0.000	08:14
Air Blank	0.000	08:16
Subject Sample # 2	0.130	08:16
Breath Volume	1.355	08:16
Air Blank	0.000	08:17
Control Test	0.077	08:17
Air Blank	0.000	08:18
Diagnostics Check	OK	08:18

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:32

SUBJECT'S LAST NAME: RUIZGARCIA

DATE/TIME OF TEST: 10/26/2023 12:01:45

Diagnostics Check	OK	11:55
Air Blank	0.000	11:56
Control Test	0.077	11:56
Air Blank	0.000	11:57
Subject Sample # 1	0.000	11:57
Breath Volume	2.117	11:57
Air Blank	0.000	11:57
Air Blank	0.000	11:59
Subject Sample # 2	0.000	12:00
Breath Volume	1.734	12:00
Air Blank	0.000	12:00
Control Test	0.077	12:01
Air Blank	0.000	12:01
Diagnostics Check	OK	12:01

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ORTIZ,PAUL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: MEJIA CASTANON

DATE/TIME OF TEST: 10/27/2023 00:22:39

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.077	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.166	00:18
Breath Volume	1.996	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.164	00:21
Breath Volume	1.523	00:21
Air Blank	0.000	00:21
Control Test	0.076	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DEVOSS,EZEKIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/28/2023 00:57:17

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.077	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1 0.227	00:50	Volume Not Met
Breath Volume	0.089	00:50	
Air Blank	0.000	00:50	
Air Blank	0.000	00:52	
Subject Sample	# 2	00:55	No Sample Provided
Breath Volume	0.000	00:55	
Air Blank	0.000	00:56	
Control Test	0.074	00:56	Control Outside Tolerance
Air Blank	0.000	00:57	

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DEVOSS,EZEKIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/28/2023 01:06:17

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.076	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.223	01:01
Breath Volume	1.308	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.223	01:04
Breath Volume	1.632	01:04
Air Blank	0.000	01:05
Control Test	0.075	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/28/2023 02:36:52

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.077	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.058	02:32
Breath Volume	2.300	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.053	02:35
Breath Volume	1.746	02:35
Air Blank	0.000	02:35
Control Test	0.076	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WEIBEZAHL

DATE/TIME OF TEST: 10/29/2023 04:04:00

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:53	
Control Test	0.076	03:53	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:57	No Sample Provided
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 2	04:02	No Sample Provided
Breath Volume	0.000	04:02	
Air Blank	0.000	04:03	
Control Test	0.078	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GRAYALES,ASIA
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: IGNOT ROSARIO

DATE/TIME OF TEST: 10/29/2023 05:18:30

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.078	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.097	05:14
Breath Volume	1.691	05:14
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.102	05:17
Breath Volume	1.718	05:17
Air Blank	0.000	05:17
Control Test	0.077	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOSEPH,FRANKY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/29/2023 05:48:42

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.077	05:43
Air Blank	0.000	05:44
Subject Sample # 1	0.026	05:44
Breath Volume	2.183	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.025	05:47
Breath Volume	2.226	05:47
Air Blank	0.000	05:47
Control Test	0.077	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/29/2023 07:37:19

Diagnostics Check	OK	07:29
Air Blank	0.000	07:29
Control Test	0.077	07:30
Air Blank	0.000	07:30
Subject Sample # 1	0.143	07:31
Breath Volume	1.222	07:31
Air Blank	0.000	07:31
Air Blank	0.000	07:33
Subject Sample # 2	0.141	07:35
Breath Volume	2.886	07:35
Air Blank	0.000	07:36
Control Test	0.077	07:36
Air Blank	0.000	07:37
Diagnostics Check	OK	07:37

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PARRA,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: RIVERORONDON

DATE/TIME OF TEST: 10/29/2023 17:35:16

Diagnostics Check	OK	17:27
Air Blank	0.000	17:27
Control Test	0.077	17:27
Air Blank	0.000	17:28
Subject Sample # 1	0.109	17:30
Breath Volume	1.121	17:30
Air Blank	0.000	17:30
Air Blank	0.000	17:32
Subject Sample # 2	0.117	17:33
Breath Volume	1.667	17:33
Air Blank	0.000	17:34
Control Test	0.077	17:34
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ALVARADO PERAZA

DATE/TIME OF TEST: 10/30/2023 04:35:52

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.077	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.218	04:31
Breath Volume	1.984	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.221	04:34
Breath Volume	3.304	04:34
Air Blank	0.000	04:34
Control Test	0.076	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARUNA,HEATHER
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: LIGHTNER

DATE/TIME OF TEST: 10/31/2023 03:55:32

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.077	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.218	03:51
Breath Volume	2.746	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.225	03:54
Breath Volume	1.945	03:54
Air Blank	0.000	03:54
Control Test	0.076	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 11/08/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CLABAUGH,ARIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: TORRES OTERO

DATE/TIME OF TEST: 10/01/2023 03:19:06

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.172	03:14
Breath Volume	2.386	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.175	03:17
Breath Volume	2.531	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEWIS,BEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: FLORESRIVERA

DATE/TIME OF TEST: 10/01/2023 08:40:09

Diagnostics Check	OK	08:34
Air Blank	0.000	08:34
Control Test	0.083	08:34
Air Blank	0.000	08:35
Subject Sample # 1	0.157	08:35
Breath Volume	2.039	08:35
Air Blank	0.000	08:36
Air Blank	0.000	08:38
Subject Sample # 2	0.154	08:38
Breath Volume	1.511	08:38
Air Blank	0.000	08:39
Control Test	0.081	08:39
Air Blank	0.000	08:40
Diagnostics Check	OK	08:40

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTRO,FABIAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/01/2023 10:13:42

Diagnostics Check	OK	10:03
Air Blank	0.000	10:04
Control Test	0.082	10:04
Air Blank	0.000	10:05
Subject Sample # 1	0.149	10:07
Breath Volume	1.218	10:07
Air Blank	0.000	10:08
Air Blank	0.000	10:09
Subject Sample # 2	0.154	10:12
Breath Volume	1.425	10:12
Air Blank	0.000	10:12
Control Test	0.080	10:13
Air Blank	0.000	10:13
Diagnostics Check	OK	10:13

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CLABAUGH,ARIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 10/02/2023 01:46:40

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.082	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.000	01:42
Breath Volume	1.738	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.000	01:45
Breath Volume	1.968	01:45
Air Blank	0.000	01:45
Control Test	0.081	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TURNER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HOLLINSHEAD

DATE/TIME OF TEST: 10/04/2023 02:04:26

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.082	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.063	02:00
Breath Volume	1.472	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.066	02:02
Breath Volume	1.332	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BARBOUR

DATE/TIME OF TEST: 10/04/2023 22:54:33

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.083	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.000	22:50
Breath Volume	1.812	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.000	22:53
Breath Volume	1.515	22:53
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORDERO,PATRICIA
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: SANTIZO PEREZ

DATE/TIME OF TEST: 10/05/2023 00:45:02

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.082	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.170	00:40
Breath Volume	1.605	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.188	00:43
Breath Volume	2.574	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:45

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEBSTER, BRAD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/05/2023 04:48:00

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.082	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.192	04:43
Breath Volume	1.148	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.180	04:46
Breath Volume	1.238	04:46
Air Blank	0.000	04:47
Control Test	0.082	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:48

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BLACKWELL,ROBERT
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: FRANCHI-RAMOS

DATE/TIME OF TEST: 10/06/2023 01:20:31

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.082	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.117	01:15
Breath Volume	2.488	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.116	01:18
Breath Volume	1.742	01:18
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: YIM,RANDY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: CANDELARIA

DATE/TIME OF TEST: 10/06/2023 03:32:47

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.082	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.177	03:28
Breath Volume	4.117	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.177	03:31
Breath Volume	4.585	03:31
Air Blank	0.000	03:31
Control Test	0.081	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOSEPH,FRANKY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: CINTO BAMACA

DATE/TIME OF TEST: 10/06/2023 04:06:00

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.082	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.166	04:01
Breath Volume	2.988	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.164	04:04
Breath Volume	2.085	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:06

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BASUALDO,DENEILLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: JACINTO CRUZ

DATE/TIME OF TEST: 10/06/2023 04:31:04

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.081	04:25	
Air Blank	0.000	04:26	
Subject Sample # 1	0.177	04:26	
Breath Volume	1.601	04:26	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample # 2	0.160	04:29	
Breath Volume	2.300	04:29	
Air Blank	0.000	04:30	
Control Test	0.074	04:30	Control Outside Tolerance
Air Blank	0.000	04:31	

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BASUALDO,DENEILLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: JACINTO CRUZ

DATE/TIME OF TEST: 10/06/2023 04:39:58

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.175	04:35
Breath Volume	3.292	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:37
Subject Sample # 2	0.178	04:38
Breath Volume	3.109	04:38
Air Blank	0.000	04:39
Control Test	0.078	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DONALD,DOMINICK
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: HEFFELFINGER

DATE/TIME OF TEST: 10/06/2023 21:56:01

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.082	21:50
Air Blank	0.000	21:50
Subject Sample # 1	0.151	21:51
Breath Volume	2.429	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:53
Subject Sample # 2	0.148	21:54
Breath Volume	2.703	21:54
Air Blank	0.000	21:55
Control Test	0.081	21:55
Air Blank	0.000	21:55
Diagnostics Check	OK	21:56

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUGO,OSCAR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 10/07/2023 03:05:21

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.082	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.230	02:56
Breath Volume	1.355	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.206	03:00
Breath Volume	1.191	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 3	0.217	03:03
Breath Volume	1.570	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SELEY

DATE/TIME OF TEST: 10/07/2023 03:35:42

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.194	03:31
Breath Volume	2.015	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.192	03:34
Breath Volume	1.921	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORDERO,PATRICIA
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: USSERY

DATE/TIME OF TEST: 10/07/2023 04:12:40

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.082	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.000	04:08
Breath Volume	2.593	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.000	04:11
Breath Volume	2.750	04:11
Air Blank	0.000	04:11
Control Test	0.081	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: ESQUIVEL

DATE/TIME OF TEST: 10/07/2023 06:32:15

Diagnostics Check	OK	06:26
Air Blank	0.000	06:26
Control Test	0.082	06:26
Air Blank	0.000	06:27
Subject Sample # 1	0.112	06:27
Breath Volume	2.488	06:27
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample # 2	0.114	06:30
Breath Volume	3.558	06:30
Air Blank	0.000	06:31
Control Test	0.081	06:31
Air Blank	0.000	06:32
Diagnostics Check	OK	06:32

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIDDIQUI,RAFEEQ
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 10/07/2023 08:24:02

Diagnostics Check	OK	08:12	
Air Blank	0.000	08:13	
Control Test	0.081	08:13	
Air Blank	0.000	08:13	
Subject Sample	# 1	08:17	No Sample Provided
Breath Volume	0.000	08:17	
Air Blank	0.000	08:17	
Air Blank	0.000	08:19	
Subject Sample	# 2	08:22	No Sample Provided
Breath Volume	0.000	08:22	
Air Blank	0.000	08:23	
Control Test	0.081	08:23	
Air Blank	0.000	08:23	
Diagnostics Check	OK	08:24	

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CRANE,ALEX
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: SAFOS

DATE/TIME OF TEST: 10/08/2023 02:17:53

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.081	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.182	02:13
Breath Volume	3.136	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:15
Subject Sample # 2	0.169	02:16
Breath Volume	2.285	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ESPINOSA, ANGEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: TIMEAU

DATE/TIME OF TEST: 10/09/2023 03:03:38

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1 0.215	02:55	Volume Not Met
Breath Volume	0.085	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 2 0.209	02:58	
Breath Volume	1.578	02:58	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample	# 3 0.219	03:01	
Breath Volume	1.148	03:01	
Air Blank	0.000	03:02	
Control Test	0.078	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WHITE,RYAN
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 09/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BICKERSTAFF

DATE/TIME OF TEST: 10/10/2023 00:17:29

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.082	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:10	No Sample Provided
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Air Blank	0.000	00:12	
Subject Sample	# 2	00:16	No Sample Provided
Breath Volume	0.000	00:16	
Air Blank	0.000	00:16	
Control Test	0.081	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 10/10/2023 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BAYLIS,BRYAN,T

BREATH TEST OPERATOR AGENCY: Cocoa PD

ARRESTING OFFICER: HARTZELL,EDWARD

ARRESTING OFFICER AGENCY: Cocoa PD

LAST AGENCY INSPECTION: 10/03/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 25721080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SNEED

DATE/TIME OF TEST: 10/05/2023 01:57:25

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:55	
Control Test	0.080	01:55	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:56	
Control Test	0.080	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 11/02/2023 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAYLIS,BRYAN,T
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: BELLOSO

DATE/TIME OF TEST: 10/06/2023 05:11:42

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.079	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.000	05:06
Breath Volume	1.140	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.000	05:10
Breath Volume	1.480	05:10
Air Blank	0.000	05:10
Control Test	0.077	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 11/02/2023 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROVENCHER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: PROVENCHER,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: STRAUB

DATE/TIME OF TEST: 10/15/2023 19:00:05

Diagnostics Check	OK	18:48	
Air Blank	0.000	18:49	
Control Test	0.078	18:49	
Air Blank	0.000	18:49	
Subject Sample	# 1 0.139	18:52	Volume Not Met
Breath Volume	0.128	18:52	
Air Blank	0.000	18:53	
Air Blank	0.000	18:55	
Subject Sample	# 2 0.143	18:58	Volume Not Met
Breath Volume	0.160	18:58	
Air Blank	0.000	18:59	
Control Test	0.079	18:59	
Air Blank	0.000	19:00	
Diagnostics Check	OK	19:00	

Data Download Date/Time: 11/02/2023 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROVENCHER, JOSHUA, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: 3S02, PROVENCHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: STRAUB

DATE/TIME OF TEST: 10/15/2023 19:14:54

Diagnostics Check	OK	19:02	
Air Blank	0.000	19:02	
Control Test	0.078	19:02	
Air Blank	0.000	19:03	
Subject Sample	# 1 0.146	19:06	Volume Not Met
Breath Volume	0.632	19:06	
Air Blank	0.000	19:07	
Air Blank	0.000	19:09	
Subject Sample	# 2 0.148	19:10	
Breath Volume	1.816	19:10	
Air Blank	0.000	19:10	
Air Blank	0.000	19:12	
Subject Sample	# 3 0.144	19:13	
Breath Volume	2.644	19:13	
Air Blank	0.000	19:14	
Control Test	0.078	19:14	
Air Blank	0.000	19:14	
Diagnostics Check	OK	19:14	

Data Download Date/Time: 11/02/2023 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, TYRELL, J
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: BUCKLEY, DREW
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: RAYNER

DATE/TIME OF TEST: 10/04/2023 07:35:50

Diagnostics Check	OK	07:29
Air Blank	0.000	07:30
Control Test	0.079	07:30
Air Blank	0.000	07:31
Subject Sample # 1	0.136	07:31
Breath Volume	2.160	07:31
Air Blank	0.000	07:32
Air Blank	0.000	07:33
Subject Sample # 2	0.135	07:34
Breath Volume	3.074	07:34
Air Blank	0.000	07:34
Control Test	0.078	07:35
Air Blank	0.000	07:35
Diagnostics Check	OK	07:35

Data Download Date/Time: 10/24/2023 05:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MALAGON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MICKETT

DATE/TIME OF TEST: 10/24/2023 02:24:07

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.000	02:19
Breath Volume	1.960	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.000	02:22
Breath Volume	1.382	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 11/01/2023 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERINO,WESLEY,G

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAYCOCK,CODY

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 10/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29122080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:32

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: ZIMARDO

DATE/TIME OF TEST: 10/30/2023 01:55:11

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.176	01:50
Breath Volume	1.636	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.173	01:53
Breath Volume	1.316	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 11/28/2023 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: TRZASKA

DATE/TIME OF TEST: 10/20/2023 04:06:18

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.078	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.078	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 11/08/2023 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: HEARN

DATE/TIME OF TEST: 10/20/2023 23:50:19

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.078	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.139	23:45
Breath Volume	1.476	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:47
Subject Sample # 2	0.144	23:48
Breath Volume	1.656	23:48
Air Blank	0.000	23:49
Control Test	0.078	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 11/08/2023 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/21/2023 03:52:56

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.079	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.078	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 11/08/2023 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TACIA,STEVEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN,DUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: BUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 10/21/2023 23:30:40

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.078	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Control Test	0.078	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 11/08/2023 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 10/22/2023 00:26:29

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.078	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.061	00:21
Breath Volume	2.839	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.061	00:24
Breath Volume	2.570	00:24
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 11/08/2023 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 10/22/2023 01:54:14

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.078	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.079	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 11/08/2023 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TACIA,STEVEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TACIA,STEVEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: DALE

DATE/TIME OF TEST: 10/29/2023 01:53:19

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.077	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:52	
Control Test	0.078	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 11/08/2023 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/06/2023 23:36:26

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.080	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.000	23:32
Breath Volume	3.210	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.000	23:35
Breath Volume	2.371	23:35
Air Blank	0.000	23:35
Control Test	0.079	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 10/18/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: GIUMETTI

DATE/TIME OF TEST: 10/08/2023 01:23:33

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:20	
Control Test	0.080	01:20	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.080	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 10/18/2023 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: ARGENTO

DATE/TIME OF TEST: 10/27/2023 22:19:10

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.000	22:14
Breath Volume	2.574	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.000	22:17
Breath Volume	1.675	22:17
Air Blank	0.000	22:18
Control Test	0.079	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 11/08/2023 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: DAILY

DATE/TIME OF TEST: 10/29/2023 01:25:34

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.080	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:24	
Control Test	0.081	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 11/08/2023 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON',A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 10/29/2023 06:12:11

Diagnostics Check	OK	06:09	
Air Blank	0.000	06:09	
Control Test	0.080	06:10	
Air Blank	0.000	06:10	
Subject Sample	# 1	06:10	Subject Test Refused
Breath Volume	0.000	06:10	
Air Blank	0.000	06:11	
Control Test	0.080	06:11	
Air Blank	0.000	06:12	
Diagnostics Check	OK	06:12	

Data Download Date/Time: 11/08/2023 12:45